

Participant Summary

MONTH 4 OF YEAR 2023

SPEC: 40 IMMUNOLOGY			SUB: 120 GENERAL			GRP: 220 ASOT, RA, MONO, CRP				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-351	47	VITROS 560		15	3.865	0.224	5.80	3.190	4.540	
145	CRP - QUANTITATIVE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE	3 SD OR 10.00	{LOWER}			
2023-352	34	DADE DIMEN		16	1.488	0.028	1.88	1.400	1.570	
2023-352	37	BECKMAN		12	1.243	0.104	8.40	0.930	1.560	
2023-352	39	ROCHE COBA		12	1.242	0.048	3.90	1.100	1.390	
2023-352	42	ASI ARLING		21	1.260	0.420	33.33	0.000	2.520	
2023-352	47	VITROS 560		15	1.985	0.163	8.19	1.500	2.470	
2023-353	34	DADE DIMEN		16	0.113	0.060	53.20	0	0.290	
2023-353	37	BECKMAN		12	0.389	0.268	68.96	0	1.190	
2023-353	39	ROCHE COBA		12	0.254	0.103	40.33	0	0.560	
2023-353	42	ASI ARLING		21	0.300	0.300	100.00	0	1.200	
2023-353	47	VITROS 560		15	0.514	0.036	6.95	0.410	0.620	
2023-354	34	DADE DIMEN		16	1.318	0.022	1.70	1.250	1.390	
2023-354	37	BECKMAN		12	1.156	0.110	9.52	0.830	1.490	
2023-354	39	ROCHE COBA		12	1.164	0.036	3.13	1.060	1.270	
2023-354	42	ASI ARLING		21	1.170	0.354	30.23	0.110	2.230	
2023-354	47	VITROS 560		15	1.837	0.189	10.30	1.270	2.410	
2023-355	34	DADE DIMEN		16	0.049	0.010	19.66	0.020	0.080	
2023-355	37	BECKMAN		12	0.378	0.267	70.47	0	1.180	
2023-355	39	ROCHE COBA		12	0.253	0.104	41.22	0	0.570	
2023-355	42	ASI ARLING		21	0.300	0.300	100.00	0	1.200	
2023-355	47	VITROS 560		15	0.525	0.053	10.16	0.370	0.690	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
091 ANTISTREP.O-QUALITA.													
2023-351	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	31	31	31	100		
2023-352	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	31	29	29	100		
2023-352	384	ASI A.SCT.		001 001	001 001		002 - POSITIVE	31	2	29	0		
2023-353	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	31	31	31	100		
2023-354	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	31	31	31	100		
2023-355	384	ASI A.SCT.		002 002	002 002		002 - POSITIVE	31	31	31	100		
092 ANTISTREP.O-SEMI QUA													
2023-351	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	11	11	11	100		
2023-352	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	11	10	10	100		
2023-352	391	ASI SLIDE		014 014	014 014		016 - 200 IU/ml	11	1	10	0		
2023-353	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	11	11	11	100		
2023-354	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	11	11	11	100		
2023-355	391	ASI SLIDE		015 019	015 019		018 - 800 IU/ml	11	4	11	100		
2023-355	391	ASI SLIDE		015 019	015 019		017 - 400 IU/ml	11	7	11	100		
103 INF. MONO- QUALITAT.													
2023-351	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	162	162	162	100		
2023-351	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	23	23	23	100		

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY													SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score									
2023-351	400	AIM MONO		001	001	001	001	001 - NEGATIVE	24	24	24	100								
	103	INF. MONO-	QUALITAT.																	
2023-352	393	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	162	162	162	100								
2023-352	394	TECO DIAG		001	001	001	001	001 - NEGATIVE	23	23	23	100								
2023-352	400	AIM MONO		001	001	001	001	001 - NEGATIVE	24	23	23	100								
2023-352	400	AIM MONO		001	001	001	001	002 - POSITIVE	24	1	23	0								
2023-353	393	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	162	162	162	100								
2023-353	394	TECO DIAG		001	001	001	001	001 - NEGATIVE	23	23	23	100								
2023-353	400	AIM MONO		001	001	001	001	001 - NEGATIVE	24	24	24	100								
2023-354	393	ASI A.SCT.		002	002	002	002	002 - POSITIVE	162	162	162	100								
2023-354	394	TECO DIAG		002	002	002	002	002 - POSITIVE	23	23	23	100								
2023-354	400	AIM MONO		002	002	002	002	002 - POSITIVE	24	24	24	100								
2023-355	393	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	162	161	161	100								
2023-355	393	ASI A.SCT.		001	001	001	001	002 - POSITIVE	162	1	161	0								
2023-355	394	TECO DIAG		001	001	001	001	001 - NEGATIVE	23	23	23	100								
2023-355	400	AIM MONO		001	001	001	001	001 - NEGATIVE	24	24	24	100								
	105	RF FACTOR(RA)-	QUALI.																	
2023-351	400	TECO DIAG.		001	001	001	001	001 - NEGATIVE	40	40	40	100								
2023-351	402	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	143	138	138	100								
2023-351	402	ASI A.SCT.		001	001	001	001	002 - POSITIVE	143	5	138	0								
2023-351	406	AIM CRP RF		001	001	001	001	001 - NEGATIVE	69	69	69	100								
2023-351	408	STANBIO		001	001	001	001	001 - NEGATIVE	23	23	23	100								
2023-352	400	TECO DIAG.		002	002	002	002	002 - POSITIVE	40	40	40	100								
2023-352	402	ASI A.SCT.		002	002	002	002	002 - POSITIVE	143	143	143	100								

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-352	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	69	69	69	100	
2023-352	408	STANBIO		002 002	002 002		002 - POSITIVE	23	23	23	100	
105 RF FACTOR(RA)-QUALI.												
2023-353	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	40	40	40	100	
2023-353	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	143	143	143	100	
2023-353	406	AIM CRP RF		001 001	001 001		002 - POSITIVE	69	1	68	0	
2023-353	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	69	68	68	100	
2023-353	408	STANBIO		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2023-354	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	40	40	40	100	
2023-354	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	143	143	143	100	
2023-354	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	69	69	69	100	
2023-354	408	STANBIO		002 002	002 002		002 - POSITIVE	23	23	23	100	
2023-355	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	40	40	40	100	
2023-355	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	143	143	143	100	
2023-355	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	69	69	69	100	
2023-355	408	STANBIO		001 001	001 001		001 - NEGATIVE	23	23	23	100	
106 RF FAC(RA)-SEMI QUAN												
2023-351	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	13	13	13	100	
2023-352	407	ASI SLIDE		005 009	004 008		004 - 40-80 IU/m	13	2	13	100	
2023-352	407	ASI SLIDE		005 009	004 008		005 - 72-160 IU	13	3	13	100	
2023-352	407	ASI SLIDE		005 009	004 008		007 - 264-640 IU	13	4	13	100	
2023-352	407	ASI SLIDE		005 009	004 008		006 - 136-320 IU	13	3	13	100	
2023-352	407	ASI SLIDE		005 009	004 008		008 - 520-1280	13	1	13	100	
2023-353	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	13	13	13	100	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
106 RF FAC(RA)-SEMI QUAN												
2023-354	407	ASI SLIDE		003 007	003 007		008 - 520-1280	13	1	12	0	
2023-354	407	ASI SLIDE		003 007	003 007		006 - 136-320 IU	13	4	12	100	
2023-354	407	ASI SLIDE		003 007	003 007		003 - 24-40 IU/m	13	1	12	100	
2023-354	407	ASI SLIDE		003 007	003 007		005 - 72-160 IU	13	4	12	100	
2023-354	407	ASI SLIDE		003 007	003 007		004 - 40-80 IU/m	13	2	12	100	
2023-354	407	ASI SLIDE		003 007	003 007		007 - 264-640 IU	13	1	12	100	
2023-355	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	13	13	13	100	
144 CRP- QUALITATIVE												
2023-351	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	125	119	119	100	
2023-351	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	125	6	119	0	
2023-351	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	31	31	31	100	
2023-351	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	60	59	59	100	
2023-351	036	AIM CRP RF		002 002	002 002		001 - NEGATIVE	60	1	59	0	
2023-351	041	STANBIO		002 002	002 002		002 - POSITIVE	14	14	14	100	
2023-352	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	125	125	125	100	
2023-352	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	31	3	28	0	
2023-352	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	31	28	28	100	
2023-352	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	60	60	60	100	
2023-352	041	STANBIO		002 002	002 002		002 - POSITIVE	14	14	14	100	
2023-353	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	125	125	125	100	
2023-353	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	31	31	31	100	
2023-353	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	60	60	60	100	
2023-353	041	STANBIO		001 001	001 001		001 - NEGATIVE	14	14	14	100	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
144 CRP- QUALITATIVE											
2023-354	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	125	124	124	100
2023-354	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	125	1	124	0
2023-354	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	31	4	27	0
2023-354	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	31	27	27	100
2023-354	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	60	60	60	100
2023-354	041	STANBIO		002 002	002 002		002 - POSITIVE	14	14	14	100
2023-355	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	125	125	125	100
2023-355	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	31	31	31	100
2023-355	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	60	60	60	100
2023-355	041	STANBIO		001 001	001 001		001 - NEGATIVE	14	14	14	100

Participant Summary
MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
093 ANTI-HUMAN IMM. (HIV)												
2023-211	023 ORTHO			002 002	002 002		002 - POS/R	10	10	10	100	
2023-211	055 ALERE DETE			002 002	002 002		002 - POS/R	26	26	26	100	
2023-212	023 ORTHO			001 001	001 001		001 - NEG/NR	10	10	10	100	
2023-212	055 ALERE DETE			001 001	001 001		001 - NEG/NR	26	26	26	100	
2023-213	023 ORTHO			002 002	002 002		002 - POS/R	10	10	10	100	
2023-213	055 ALERE DETE			002 002	002 002		002 - POS/R	26	26	26	100	
2023-214	023 ORTHO			001 001	001 001		001 - NEG/NR	10	10	10	100	
2023-214	055 ALERE DETE			001 001	001 001		001 - NEG/NR	26	26	26	100	
2023-215	023 ORTHO			001 001	001 001		001 - NEG/NR	10	10	10	100	
2023-215	055 ALERE DETE			001 001	001 001		001 - NEG/NR	26	26	26	100	

Participant Summary
MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 247 MYCOPLASMA						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
548 MYCOPLASMA													
2023-121	001	IMMU. CARD		001 001	001 001		001 - Non Reacti	638	638	638	100		
2023-122	001	IMMU. CARD		002 002	002 002		002 - Reactive	638	637	637	100		
2023-122	001	IMMU. CARD		002 002	002 002		001 - Non Reacti	638	1	637	0		
2023-123	001	IMMU. CARD		001 001	001 001		001 - Non Reacti	638	638	638	100		
2023-124	001	IMMU. CARD		002 002	002 002		002 - Reactive	638	638	638	100		
2023-125	001	IMMU. CARD		001 001	001 001		001 - Non Reacti	638	638	638	100		

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
554 COVID-19 ANTIBODY												
2023-271	001 IgM	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-271	002 IgG	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-272	001 IgM	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-272	002 IgG	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-273	001 IgM	030 FASTEP		001 001	001 001		001 - Negative	33	33	33	100	
2023-273	002 IgG	030 FASTEP		001 001	001 001		001 - Negative	33	33	33	100	
2023-274	001 IgM	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-274	002 IgG	030 FASTEP		002 002	002 002		002 - Positive	33	33	33	100	
2023-275	001 IgM	030 FASTEP		001 001	001 001		001 - Negative	33	33	33	100	
2023-275	002 IgG	030 FASTEP		001 001	001 001		001 - Negative	33	33	33	100	

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
109 SYPHILLIS SERO.-QUAL												
2023-321	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR Non Rea	28	28	28	100	
2023-321	002 RPR	033 ASI		001 001	001 001		001 - NR Non Rea	538	538	538	100	
2023-321	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR Non Rea	66	66	66	100	
2023-321	002 RPR	053 AIM RPR		001 001	001 001		001 - NR Non Rea	94	94	94	100	
2023-321	002 RPR	057 Germain La		001 001	001 001		001 - NR Non Rea	27	27	27	100	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
109 SYPHILLIS SERO.-QUAL											
2023-322	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	28	28	28	100
2023-322	002 RPR	033 ASI		004 004	004 004		004 - React RPR	538	537	537	100
2023-322	002 RPR	033 ASI		004 004	004 004		001 - NR Non Rea	538	1	537	0
2023-322	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	66	66	100
2023-322	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	94	94	94	100
2023-322	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	27	27	100
2023-323	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR Non Rea	28	28	28	100
2023-323	002 RPR	033 ASI		001 001	001 001		001 - NR Non Rea	538	538	538	100
2023-323	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR Non Rea	66	66	66	100
2023-323	002 RPR	053 AIM RPR		001 001	001 001		001 - NR Non Rea	94	94	94	100
2023-323	002 RPR	057 Germain La		001 001	001 001		001 - NR Non Rea	27	27	27	100
2023-324	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	28	28	28	100
2023-324	002 RPR	033 ASI		004 004	004 004		004 - React RPR	538	538	538	100
2023-324	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	66	66	100
2023-324	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	94	93	93	100
2023-324	002 RPR	053 AIM RPR		004 004	004 004		001 - NR Non Rea	94	1	93	0
2023-324	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	27	27	100
2023-325	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	28	28	28	100
2023-325	002 RPR	033 ASI		004 004	004 004		004 - React RPR	538	538	538	100
2023-325	002 RPR	040 TECO DIAG.		004 004	004 004		001 - NR Non Rea	66	1	65	0
2023-325	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	65	65	100
2023-325	002 RPR	053 AIM RPR		004 004	004 004		002 - WR Weakly	94	1	92	0
2023-325	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	94	92	92	100
2023-325	002 RPR	053 AIM RPR		004 004	004 004		001 - NR Non Rea	94	1	92	0
2023-325	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	27	27	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
110 SYPHILLIS SERO-QUANT											
2023-321	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR	11	11	11	100
2023-321	002 RPR	033 ASI		001 001	001 001		001 - NR	247	247	247	100
2023-321	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR	15	15	15	100
2023-321	002 RPR	053 AIM RPR		001 001	001 001		001 - NR	19	19	19	100
2023-322	002 RPR	030 IMMUNOSTIC		003 005	003 005		004 - R- 2 DILS.	11	10	11	100
2023-322	002 RPR	030 IMMUNOSTIC		003 005	003 005		005 - R- 4 DILS.	11	1	11	100
2023-322	002 RPR	033 ASI		003 005	003 005		005 - R- 4 DILS.	247	79	244	100
2023-322	002 RPR	033 ASI		003 005	003 005		006 - R- 8 DILS.	247	3	244	0
2023-322	002 RPR	033 ASI		003 005	003 005		003 - R- 1 DIL.	247	5	244	100
2023-322	002 RPR	033 ASI		003 005	003 005		004 - R- 2 DILS.	247	160	244	100
2023-322	002 RPR	040 TECO DIAG.		003 005	003 005		005 - R- 4 DILS.	15	2	15	100
2023-322	002 RPR	040 TECO DIAG.		003 005	003 005		003 - R- 1 DIL.	15	1	15	100
2023-322	002 RPR	040 TECO DIAG.		003 005	003 005		004 - R- 2 DILS.	15	12	15	100
2023-322	002 RPR	053 AIM RPR		003 005	003 005		003 - R- 1 DIL.	19	6	19	100
2023-322	002 RPR	053 AIM RPR		003 005	003 005		004 - R- 2 DILS.	19	10	19	100
2023-322	002 RPR	053 AIM RPR		003 005	003 005		005 - R- 4 DILS.	19	3	19	100
2023-323	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR	11	11	11	100
2023-323	002 RPR	033 ASI		001 001	001 001		001 - NR	247	247	247	100
2023-323	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR	15	15	15	100
2023-323	002 RPR	053 AIM RPR		001 001	001 001		001 - NR	19	19	19	100
2023-324	002 RPR	030 IMMUNOSTIC		004 006	004 006		005 - R- 4 DILS.	11	10	11	100
2023-324	002 RPR	030 IMMUNOSTIC		004 006	004 006		006 - R- 8 DILS.	11	1	11	100
2023-324	002 RPR	033 ASI		004 006	004 006		004 - R- 2 DILS.	247	17	247	100
2023-324	002 RPR	033 ASI		004 006	004 006		005 - R- 4 DILS.	247	186	247	100
2023-324	002 RPR	033 ASI		004 006	004 006		006 - R- 8 DILS.	247	44	247	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-324	002 RPR	040 TECO DIAG.		004 006	004 006		004 - R- 2 DILS.	15	3	14	100
2023-324	002 RPR	040 TECO DIAG.		004 006	004 006		003 - R- 1 DIL.	15	1	14	0
2023-324	002 RPR	040 TECO DIAG.		004 006	004 006		005 - R- 4 DILS.	15	11	14	100
2023-324	002 RPR	053 AIM RPR		004 006	004 006		005 - R- 4 DILS.	19	10	19	100
2023-324	002 RPR	053 AIM RPR		004 006	004 006		004 - R- 2 DILS.	19	9	19	100
110 SYPHILLIS SERO-QUANT											
2023-325	002 RPR	030 IMMUNOSTIC		003 005	003 005		006 - R- 8 DILS.	11	1	10	0
2023-325	002 RPR	030 IMMUNOSTIC		003 005	003 005		005 - R- 4 DILS.	11	2	10	100
2023-325	002 RPR	030 IMMUNOSTIC		003 005	003 005		004 - R- 2 DILS.	11	8	10	100
2023-325	002 RPR	033 ASI		003 005	003 005		003 - R- 1 DIL.	247	1	247	100
2023-325	002 RPR	033 ASI		003 005	003 005		004 - R- 2 DILS.	247	164	247	100
2023-325	002 RPR	033 ASI		003 005	003 005		005 - R- 4 DILS.	247	82	247	100
2023-325	002 RPR	040 TECO DIAG.		003 005	003 005		005 - R- 4 DILS.	15	2	15	100
2023-325	002 RPR	040 TECO DIAG.		003 005	003 005		003 - R- 1 DIL.	15	2	15	100
2023-325	002 RPR	040 TECO DIAG.		003 005	003 005		004 - R- 2 DILS.	15	11	15	100
2023-325	002 RPR	053 AIM RPR		003 005	003 005		003 - R- 1 DIL.	19	7	19	100
2023-325	002 RPR	053 AIM RPR		003 005	003 005		004 - R- 2 DILS.	19	12	19	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 260 BACTER. -GRAM STAIN			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
111 BACTER. GRAM STAIN											
2023-221				004 004	004 004		004 - GRAM POS C	126	122	122	100
2023-221				004 004	004 004		002 - GRAM POS B	126	2	122	0
2023-221				004 004	004 004		003 - GRAM NEG C	126	2	122	0
2023-222				001 001	001 001		007 - GRAM NEG D	126	1	121	0
2023-222				001 001	001 001		002 - GRAM POS B	126	1	121	0
2023-222				001 001	001 001		005 - GRAM NEG C	126	3	121	0
2023-222				001 001	001 001		001 - GRAM NEG B	126	121	121	100
2023-223				001 001	001 001		005 - GRAM NEG C	126	2	123	0
2023-223				001 001	001 001		001 - GRAM NEG B	126	123	123	100
2023-223				001 001	001 001		004 - GRAM POS C	126	1	123	0
2023-224				001 001	001 001		001 - GRAM NEG B	126	123	123	100
2023-224				001 001	001 001		007 - GRAM NEG D	126	1	123	0
2023-224				001 001	001 001		002 - GRAM POS B	126	2	123	0
2023-225				004 004	004 004		002 - GRAM POS B	126	2	124	0
2023-225				004 004	004 004		004 - GRAM POS C	126	124	124	100

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
112 BACTER.-IDENTIFICAT.											
2023-241	001 AEROBIC	032 MICROSCAN			198 198		198 - Strep. aga	27	23	23	100
2023-241	001 AEROBIC	032 MICROSCAN			198 198		207 - Strep. mi.	27	3	23	0

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-241	001 AEROBIC	032 MICROSCAN			198 198		256 - No Org. IS	27	1	23	0
2023-241	001 AEROBIC	034 MICROSCAN			198 198		198 - Strep. aga	25	24	24	100
2023-241	001 AEROBIC	034 MICROSCAN			198 198		207 - Strep. mi.	25	1	24	0
2023-241	001 AEROBIC	046 MICROSCAN			198 198		198 - Strep. aga	31	29	29	100
2023-241	001 AEROBIC	046 MICROSCAN			198 198		207 - Strep. mi.	31	2	29	0
2023-241	001 AEROBIC	049 VITEK 2			198 198		259 - g+ cocci	40	1	38	0
2023-241	001 AEROBIC	049 VITEK 2			198 198		207 - Strep. mi.	40	1	38	0
2023-241	001 AEROBIC	049 VITEK 2			198 198		198 - Strep. aga	40	38	38	100
112 BACTER.-IDENTIFICAT.											
2023-242	001 AEROBIC	032 MICROSCAN			188 188		188 - Staphy. ep	48	24	48	100
2023-242	001 AEROBIC	032 MICROSCAN			187 187		187 - Staphy. au	48	24	48	100
2023-242	001 AEROBIC	034 MICROSCAN			188 188		068 - Esche coli	52	1	51	0
2023-242	001 AEROBIC	034 MICROSCAN			187 187		187 - Staphy. au	52	26	51	100
2023-242	001 AEROBIC	034 MICROSCAN			188 188		188 - Staphy. ep	52	25	51	100
2023-242	001 AEROBIC	042 VITEK(AMS)			187 187		187 - Staphy. au	12	6	12	100
2023-242	001 AEROBIC	042 VITEK(AMS)			188 188		188 - Staphy. ep	12	6	12	100
2023-242	001 AEROBIC	046 MICROSCAN			187 187		187 - Staphy. au	58	29	58	100
2023-242	001 AEROBIC	046 MICROSCAN			188 188		188 - Staphy. ep	58	29	58	100
2023-242	001 AEROBIC	049 VITEK 2			188 188		188 - Staphy. ep	76	38	75	100
2023-242	001 AEROBIC	049 VITEK 2			188 188		259 - g+ cocci	76	1	75	0
2023-242	001 AEROBIC	049 VITEK 2			187 187		187 - Staphy. au	76	37	75	100
2023-243	001 AEROBIC	001 CONVENCIO.			040 040		038 - Campy. sp.	16	3	7	100
2023-243	001 AEROBIC	001 CONVENCIO.			040 040		040 - Campy. jej	16	4	7	100
2023-243	001 AEROBIC	001 CONVENCIO.			040 040		251 - NO S.Sh.Y.	16	9	7	0
2023-243	001 AEROBIC	032 MICROSCAN			040 040		046 - Citrobacte	14	2	0	0
2023-243	001 AEROBIC	032 MICROSCAN			040 040		251 - NO S.Sh.Y.	14	12	0	0
2023-243	001 AEROBIC	034 MICROSCAN			040 040		251 - NO S.Sh.Y.	14	10	3	0
2023-243	001 AEROBIC	034 MICROSCAN			040 040		043 - Citrobacte	14	1	3	0

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-243	001 AEROBIC	034 MICROSCAN			040 040		040 - Campy. jej	14	3	3	100	
2023-243	001 AEROBIC	046 MICROSCAN			040 040		046 - Citrobacte	17	2	0	0	
2023-243	001 AEROBIC	046 MICROSCAN			040 040		188 - Staphy. ep	17	1	0	0	
2023-243	001 AEROBIC	046 MICROSCAN			040 040		043 - Citrobacte	17	1	0	0	
2023-243	001 AEROBIC	046 MICROSCAN			040 040		251 - NO S.Sh.Y.	17	13	0	0	
2023-243	001 AEROBIC	049 VITEK 2			040 040		040 - Campy. jej	24	4	4	100	
2023-243	001 AEROBIC	049 VITEK 2			040 040		251 - NO S.Sh.Y.	24	20	4	0	
112 BACTER.-IDENTIFICAT.												
2023-244	001 AEROBIC	032 MICROSCAN			076 076		076 - Ente/faeci	24	24	24	100	
2023-244	001 AEROBIC	034 MICROSCAN			076 076		076 - Ente/faeci	26	25	25	100	
2023-244	001 AEROBIC	034 MICROSCAN			076 076		075 - Enter faec	26	1	25	0	
2023-244	001 AEROBIC	046 MICROSCAN			076 076		076 - Ente/faeci	29	29	29	100	
2023-244	001 AEROBIC	049 VITEK 2			076 076		076 - Ente/faeci	39	39	39	100	
2023-245	001 AEROBIC	032 MICROSCAN			068 068		068 - Esche coli	24	24	24	100	
2023-245	001 AEROBIC	034 MICROSCAN			068 068		068 - Esche coli	26	26	26	100	
2023-245	001 AEROBIC	046 MICROSCAN			068 068		068 - Esche coli	29	29	29	100	
2023-245	001 AEROBIC	049 VITEK 2			068 068		068 - Esche coli	39	39	39	100	

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
113 BACT.ANTIM. SUSCEPT.												
2023-244	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		004 004		001 - SUSCEPTIB.	34	2	31	0	
2023-244	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		004 004		004 - RESISTANT	34	32	31	100	
2023-244	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		004 004		004 - RESISTANT	34	32	31	0	
2023-244	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	35	34	33	100	
2023-244	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	35	34	33	0	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-244	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		002 - INTERMEDI.	35	1	33	0
2023-244	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	34	31	30	100
2023-244	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	34	31	30	0
2023-244	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		004 - RESISTANT	34	1	30	0
2023-244	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		002 - INTERMEDI.	34	2	30	0
2023-244	004 MICRO SCAN	054 PENICILLIN	000 Not apply		004 004		004 - RESISTANT	21	19	18	100
2023-244	004 MICRO SCAN	054 PENICILLIN	000 Not apply		004 004		004 - RESISTANT	21	19	18	0
2023-244	004 MICRO SCAN	054 PENICILLIN	000 Not apply		004 004		001 - SUSCEPTIB.	21	2	18	0
2023-244	004 MICRO SCAN	062 TETRACYCLI	000 Not apply		004 004		001 - SUSCEPTIB.	22	1	20	0
2023-244	004 MICRO SCAN	062 TETRACYCLI	000 Not apply		004 004		004 - RESISTANT	22	21	20	100
2023-244	004 MICRO SCAN	062 TETRACYCLI	000 Not apply		004 004		004 - RESISTANT	22	21	20	0
2023-244	004 MICRO SCAN	068 VANCOMYCIN	000 Not apply		004 004		001 - SUSCEPTIB.	48	1	46	0
2023-244	004 MICRO SCAN	068 VANCOMYCIN	000 Not apply		004 004		004 - RESISTANT	48	47	46	100
2023-244	004 MICRO SCAN	068 VANCOMYCIN	000 Not apply		004 004		004 - RESISTANT	48	47	46	0
2023-244	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	62	62	61	100
2023-244	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	62	62	61	0
2023-244	004 MICRO SCAN	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	67	67	66	100
2023-244	004 MICRO SCAN	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	67	67	66	0
2023-244	004 MICRO SCAN	097 DAPTOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	100
2023-244	004 MICRO SCAN	097 DAPTOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	0
2023-244	006 VITEK SYS.	003 AMPICILLIN	000 Not apply		004 004		004 - RESISTANT	23	23	22	100
2023-244	006 VITEK SYS.	003 AMPICILLIN	000 Not apply		004 004		004 - RESISTANT	23	23	22	0
2023-244	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	100
2023-244	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	0
2023-244	006 VITEK SYS.	062 TETRACYCLI	000 Not apply		004 004		004 - RESISTANT	14	14	14	100
2023-244	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		004 004		004 - RESISTANT	33	32	31	100
2023-244	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		004 004		004 - RESISTANT	33	32	31	0
2023-244	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		004 004		001 - SUSCEPTIB.	33	1	31	0
2023-244	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	38	38	38	100
2023-244	006 VITEK SYS.	084 GENTA.-500	000 Not apply		001 001		001 - SUSCEPTIB.	15	15	15	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-244	006	VITEK SYS.	086	LINEZOLID	000	Not apply	001 001	001 - SUSCEPTIB.	41	41	41	100
2023-244	006	VITEK SYS.	090	STREPTOMY.	000	Not apply	001 001	001 - SUSCEPTIB.	12	12	12	100
113 BACT.ANTIM. SUSCEPT.												
2023-245	004	MICRO SCAN	001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	38	100
2023-245	004	MICRO SCAN	003	AMPICILLIN	000	Not apply	001 001	001 - SUSCEPTIB.	56	56	56	100
2023-245	004	MICRO SCAN	004	AMPIC/SULB	000	Not apply	001 001	001 - SUSCEPTIB.	19	19	19	100
2023-245	004	MICRO SCAN	007	AZTREONAM	000	Not apply	001 001	001 - SUSCEPTIB.	15	15	15	100
2023-245	004	MICRO SCAN	010	CEFAZOLIN	000	Not apply	001 001	001 - SUSCEPTIB.	39	39	39	100
2023-245	004	MICRO SCAN	020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	14	14	14	100
2023-245	004	MICRO SCAN	022	CEFTRIAXON	000	Not apply	001 001	001 - SUSCEPTIB.	10	10	10	100
2023-245	004	MICRO SCAN	029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	59	59	59	100
2023-245	004	MICRO SCAN	037	GENTAMICIN	000	Not apply	001 001	001 - SUSCEPTIB.	39	39	39	100
2023-245	004	MICRO SCAN	038	IMIPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	21	21	21	100
2023-245	004	MICRO SCAN	050	NITROFURAN	000	Not apply	001 001	001 - SUSCEPTIB.	55	55	55	100
2023-245	004	MICRO SCAN	065	TOBRAMYCIN	000	Not apply	001 001	001 - SUSCEPTIB.	47	47	47	100
2023-245	004	MICRO SCAN	067	TRIME/SULF	000	Not apply	001 001	001 - SUSCEPTIB.	59	59	59	100
2023-245	004	MICRO SCAN	072	CEFEPIME	000	Not apply	001 001	001 - SUSCEPTIB.	22	22	22	100
2023-245	004	MICRO SCAN	073	PIPER/TAZO	000	Not apply	001 001	001 - SUSCEPTIB.	15	15	15	100
2023-245	004	MICRO SCAN	080	LEVOFLOXA.	000	Not apply	001 001	001 - SUSCEPTIB.	39	39	39	100
2023-245	004	MICRO SCAN	096	ERTAPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	13	13	13	100
2023-245	006	VITEK SYS.	001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	12	12	12	100
2023-245	006	VITEK SYS.	002	AMOX/CLAVU	000	Not apply	001 001	001 - SUSCEPTIB.	10	10	10	100
2023-245	006	VITEK SYS.	003	AMPICILLIN	000	Not apply	001 001	001 - SUSCEPTIB.	33	33	32	100
2023-245	006	VITEK SYS.	003	AMPICILLIN	000	Not apply	001 001	001 - SUSCEPTIB.	33	33	32	0
2023-245	006	VITEK SYS.	004	AMPIC/SULB	000	Not apply	001 001	001 - SUSCEPTIB.	13	13	12	100
2023-245	006	VITEK SYS.	004	AMPIC/SULB	000	Not apply	001 001	001 - SUSCEPTIB.	13	13	12	0
2023-245	006	VITEK SYS.	010	CEFAZOLIN	000	Not apply	001 001	001 - SUSCEPTIB.	25	25	24	100
2023-245	006	VITEK SYS.	010	CEFAZOLIN	000	Not apply	001 001	001 - SUSCEPTIB.	25	25	24	0
2023-245	006	VITEK SYS.	020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	15	15	15	100
2023-245	006	VITEK SYS.	029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	37	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-245	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	38	38	37	0
2023-245	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	32	32	31	100
2023-245	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	32	32	31	0
2023-245	006 VITEK SYS.	038 IMIPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	20	20	20	100
2023-245	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	41	41	40	100
2023-245	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	41	41	40	0
2023-245	006 VITEK SYS.	065 TOBRAMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	23	23	23	100
2023-245	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	35	35	34	100
2023-245	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	35	35	34	0
2023-245	006 VITEK SYS.	072 CEFEPIME	000 Not apply		001 001		001 - SUSCEPTIB.	21	21	20	100
2023-245	006 VITEK SYS.	072 CEFEPIME	000 Not apply		001 001		001 - SUSCEPTIB.	21	21	20	0
2023-245	006 VITEK SYS.	073 PIPER/TAZO	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	18	100
2023-245	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	20	20	20	100
2023-245	006 VITEK SYS.	087 MEROPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	10	10	10	100
2023-245	006 VITEK SYS.	096 ERTAPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	10	10	10	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY		SUB: 140 BACTERIOLOGY				GRP: 263 BACTER.- DIRECT ANT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
147 BACTER.DIRECT ANTIG.											
2023-231				001 001	001 001		001 - NEGATIVE	12	12	12	100
2023-232				002 002	002 002		002 - POSITIVE	12	12	12	100
2023-233				001 001	001 001		001 - NEGATIVE	12	12	12	100
2023-234				002 002	002 002		002 - POSITIVE	12	12	12	100
2023-235				002 002	002 002		001 - NEGATIVE	12	1	11	0
2023-235				002 002	002 002		002 - POSITIVE	12	11	11	100
120 IDENT OF PARASITES											
SPE: 50 MICROBIOLOGY		SUB: 170 PARASITOLOGY				GRP: 350 ID OF PARASITES					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-311	001	E. histoly	002	CYSTS		001 001	003 - MODERATE	35	14	35	100
2023-311	001	E. histoly	002	CYSTS		001 001	002 - FEW	35	18	35	100
2023-311	001	E. histoly	002	CYSTS		001 001	004 - MANY	35	3	35	100
2023-311	003	E. coli	002	CYSTS		002 004	003 - MODERATE	54	26	54	100
2023-311	003	E. coli	002	CYSTS		002 004	002 - FEW	54	9	54	100
2023-311	003	E. coli	002	CYSTS		002 004	004 - MANY	54	19	54	100
2023-311	021	Taenia s/s	003	OVA OR EGG		002 004	003 - MODERATE	54	20	53	100
2023-311	021	Taenia s/s	003	OVA OR EGG		002 004	002 - FEW	54	28	53	100
2023-311	021	Taenia s/s	003	OVA OR EGG		002 004	004 - MANY	54	5	53	100
2023-311	021	Taenia s/s	003	OVA OR EGG		002 004	006 - PAR NOT RE	54	1	53	0

Participant Summary
MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY		SUB: 170 PARASITOLOGY			GRP: 350 ID OF PARASITES						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
120 IDENT OF PARASITES											
2023-312	050 NPF	005 NOT APPLY			005 005		005 - NOT APPLY	54	54	54	100
2023-313	003 E. coli	002 CYSTS			001 001		002 - FEW	11	11	11	100
2023-313	009 Giardia l.	002 CYSTS			002 004		003 - MODERATE	54	27	53	100
2023-313	009 Giardia l.	002 CYSTS			002 004		002 - FEW	54	11	53	100
2023-313	009 Giardia l.	002 CYSTS			002 004		004 - MANY	54	15	53	100
2023-313	009 Giardia l.	002 CYSTS			002 004		006 - PAR NOT RE	54	1	53	0
2023-313	024 Hymenolep.	003 OVA OR EGG			002 004		003 - MODERATE	54	16	54	100
2023-313	024 Hymenolep.	003 OVA OR EGG			002 004		002 - FEW	54	35	54	100
2023-313	024 Hymenolep.	003 OVA OR EGG			002 004		004 - MANY	54	3	54	100
2023-313	036 Blastocys.	002 CYSTS					003 - MODERATE	11	3	0	0
2023-313	036 Blastocys.	002 CYSTS					002 - FEW	11	8	0	0
2023-314	050 NPF	005 NOT APPLY			005 005		005 - NOT APPLY	54	46	46	100
2023-314	050 NPF	005 NOT APPLY			005 005		006 - PAR NOT RE	54	8	46	0
2023-315	031 Clonorchis	003 OVA OR EGG			005 005		002 - FEW	54	1	52	100
2023-315	031 Clonorchis	003 OVA OR EGG			005 005		005 - NOT APPLY	54	51	52	100
2023-315	031 Clonorchis	003 OVA OR EGG			005 005		006 - PAR NOT RE	54	2	52	0

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 360 PRES/ABS OF PARASITE			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
121 PRESENCE OR ABSENCE											
2023-311				002 002	002 002		002 - Presence	35	35	35	100
2023-312				001 001	001 001		001 - Absence	35	35	35	100
2023-313				002 002	002 002		002 - Presence	35	35	35	100
2023-314				001 001	001 001		001 - Absence	35	29	29	100
2023-314				001 001	001 001		002 - Presence	35	6	29	0
2023-315				002 002	002 002		002 - Presence	35	35	35	100
SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
176 INFLUENZA A&B QUALI											
2023-251	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	100	100	100	100
2023-251	001 INFLU. A	004 Sekisui Os		002 002	002 002		001 - NEGATIVE	122	2	120	0
2023-251	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	122	120	120	100
2023-251	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	47	46	46	100
2023-251	001 INFLU. A	006 Quidel Flu		002 002	002 002		001 - NEGATIVE	47	1	46	0
2023-251	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	23	23	23	100
2023-251	001 INFLU. A	008 Remel x/p		002 002	002 002		001 - NEGATIVE	13	1	12	0
2023-251	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	13	12	12	100
2023-251	001 INFLU. A	012 BIOSIG A+B			002 002		001 - NEGATIVE	314	67	247	0

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-251	001 INFLU. A	012 BIOSIG A+B			002 002		002 - POSITIVE	314	247	247	100	
2023-251	001 INFLU. A	016 STATUS FLU			002 002		002 - POSITIVE	133	106	106	100	
2023-251	001 INFLU. A	016 STATUS FLU			002 002		001 - NEGATIVE	133	27	106	0	
2023-251	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100	
2023-251	002 INFLU. B	004 Sekisui Os		001 001	001 001		002 - POSITIVE	122	1	121	0	
2023-251	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	121	121	100	
2023-251	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100	
2023-251	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2023-251	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100	
2023-251	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	313	313	313	100	
2023-251	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	133	133	133	100	
176	INFLUENZA A&B QUALI											
2023-252	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100	
2023-252	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	122	122	100	
2023-252	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100	
2023-252	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2023-252	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100	
2023-252	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	314	314	314	100	
2023-252	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	133	133	133	100	
2023-252	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100	
2023-252	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	122	122	100	
2023-252	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100	
2023-252	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2023-252	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100	
2023-252	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	312	312	312	100	
2023-252	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	132	132	132	100	
2023-253	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	100	100	100	100	
2023-253	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	122	121	121	100	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-253	001 INFLU. A	004 Sekisui Os		002 002	002 002		001 - NEGATIVE	122	1	121	0		
2023-253	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	47	47	47	100		
2023-253	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	23	23	23	100		
2023-253	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	13	13	13	100		
2023-253	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	314	313	313	100		
2023-253	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		001 - NEGATIVE	314	1	313	0		
2023-253	001 INFLU. A	016 STATUS FLU		002 002	002 002		002 - POSITIVE	133	133	133	100		
2023-253	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100		
2023-253	002 INFLU. B	004 Sekisui Os		001 001	001 001		002 - POSITIVE	122	1	121	0		
2023-253	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	121	121	100		
2023-253	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100		
2023-253	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	23	23	23	100		
2023-253	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100		
2023-253	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	313	313	313	100		
2023-253	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	133	133	133	100		
176	INFLUENZA A&B QUALI												
2023-254	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100		
2023-254	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	122	122	100		
2023-254	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100		
2023-254	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100		
2023-254	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100		
2023-254	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	315	315	315	100		
2023-254	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	132	132	132	100		
2023-254	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	99	99	99	100		
2023-254	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	122	122	122	100		
2023-254	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	47	47	47	100		
2023-254	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	22	21	21	100		
2023-254	002 INFLU. B	007 BD Veritor		001 001	001 001		002 - POSITIVE	22	1	21	0		
2023-254	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100		
2023-254	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	311	311	311	100		

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-254	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	133	133	133	100
	176 INFLUENZA A&B QUALI										
2023-255	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	98	98	98	100
2023-255	001 INFLU. A	004 Sekisui Os		001 001	001 001		002 - POSITIVE	121	1	120	0
2023-255	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	121	120	120	100
2023-255	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	48	48	48	100
2023-255	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	23	23	23	100
2023-255	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	13	13	13	100
2023-255	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	313	313	313	100
2023-255	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	132	132	132	100
2023-255	002 INFLU. B	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	101	100	100	100
2023-255	002 INFLU. B	003 Sofia CLIA		002 002	002 002		001 - NEGATIVE	101	1	100	0
2023-255	002 INFLU. B	004 Sekisui Os		002 002	002 002		002 - POSITIVE	121	120	120	100
2023-255	002 INFLU. B	004 Sekisui Os		002 002	002 002		001 - NEGATIVE	121	1	120	0
2023-255	002 INFLU. B	006 Quidel Flu		002 002	002 002		002 - POSITIVE	46	46	46	100
2023-255	002 INFLU. B	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100
2023-255	002 INFLU. B	008 Remel x/p		002 002	002 002		002 - POSITIVE	13	13	13	100
2023-255	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	312	312	312	100
2023-255	002 INFLU. B	016 STATUS FLU		002 002	002 002		002 - POSITIVE	132	132	132	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY			SUB: 180 Virology				GRP: 450 RSV Syncytial Virus					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
555 Resp Syncytial Virus												
2023-251	001	ALERE ABBO		001	001	001	001	001 - Neg	31	31	31	100
2023-251	002	SOFIA RSV		001	001	001	001	002 - Pos	29	1	28	0
2023-251	002	SOFIA RSV		001	001	001	001	001 - Neg	29	28	28	100
2023-252	001	ALERE ABBO		001	001	001	001	001 - Neg	31	31	31	100
2023-252	002	SOFIA RSV		001	001	001	001	001 - Neg	29	28	28	100
2023-252	002	SOFIA RSV		001	001	001	001	002 - Pos	29	1	28	0
2023-253	001	ALERE ABBO		001	001	001	001	001 - Neg	31	31	31	100
2023-253	002	SOFIA RSV		001	001	001	001	001 - Neg	29	29	29	100
2023-254	001	ALERE ABBO		002	002	002	002	001 - Neg	31	1	30	0
2023-254	001	ALERE ABBO		002	002	002	002	002 - Pos	31	30	30	100
2023-254	002	SOFIA RSV		002	002	002	002	002 - Pos	29	29	29	100
2023-255	001	ALERE ABBO		001	001	001	001	001 - Neg	31	30	30	100
2023-255	001	ALERE ABBO		001	001	001	001	002 - Pos	31	1	30	0
2023-255	002	SOFIA RSV		001	001	001	001	001 - Neg	29	28	28	100
2023-255	002	SOFIA RSV		001	001	001	001	002 - Pos	29	1	28	0

SPE: 50 MICROBIOLOGY			SUB: 180 Virology				GRP: 460 COVID 19 ANTIGEN					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
561 COVID 19 ANTIGEN												
2023-251				002	002		001 - Neg	22	1	21	0	
2023-251				002	002		002 - Pos	22	21	21	100	

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 460 COVID 19 ANTIGEN

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-251	001	Bin		002	002		002 - Pos	169	169	169	100
2023-251	002	CareS		002	002		002 - Pos	199	199	199	100
2023-251	003	BD Veritor		002	002		002 - Pos	26	26	26	100
2023-251	004	QUICKVUE		002	002		002 - Pos	47	47	47	100
2023-251	005	INDICAID		002	002		002 - Pos	66	66	66	100
2023-251	006	Clarity		002	002		001 - Neg	73	1	72	0
2023-251	006	Clarity		002	002		002 - Pos	73	72	72	100
2023-251	007	Sofia		002	002		002 - Pos	49	49	49	100
2023-251	008	GenBody		002	002		002 - Pos	20	20	20	100
2023-251	009	Celltrion		002	002		002 - Pos	50	50	50	100
2023-251	011	Status		002	002		002 - Pos	22	22	22	100
561 COVID 19 ANTIGEN											
2023-252				002	002		001 - Neg	22	1	21	0
2023-252				002	002		002 - Pos	22	21	21	100
2023-252	001	Bin		002	002		002 - Pos	169	169	169	100
2023-252	002	CareS		002	002		002 - Pos	199	199	199	100
2023-252	003	BD Veritor		002	002		002 - Pos	26	26	26	100
2023-252	004	QUICKVUE		002	002		002 - Pos	47	47	47	100
2023-252	005	INDICAID		002	002		002 - Pos	66	66	66	100
2023-252	006	Clarity		002	002		002 - Pos	73	72	72	100
2023-252	006	Clarity		002	002		001 - Neg	73	1	72	0
2023-252	007	Sofia		002	002		002 - Pos	49	49	49	100
2023-252	008	GenBody		002	002		002 - Pos	20	20	20	100
2023-252	009	Celltrion		002	002		002 - Pos	50	50	50	100
2023-252	011	Status		002	002		002 - Pos	22	22	22	100
2023-253				001	001		001 - Neg	22	22	22	100
2023-253	001	Bin		001	001		001 - Neg	169	169	169	100
2023-253	002	CareS		001	001		001 - Neg	199	199	199	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY		SUB: 180 Virology			GRP: 460 COVID 19 ANTIGEN						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-253	003	BD Veritor		001	001		001 - Neg	26	26	26	100
2023-253	004	QUICKVUE		001	001		001 - Neg	47	47	47	100
2023-253	005	INDICAID		001	001		001 - Neg	66	66	66	100
2023-253	006	Clarity		001	001		001 - Neg	73	73	73	100
2023-253	007	Sofia		001	001		001 - Neg	49	49	49	100
2023-253	008	GenBody		001	001		001 - Neg	20	20	20	100
2023-253	009	Celltrion		001	001		001 - Neg	50	50	50	100
2023-253	011	Status		001	001		001 - Neg	22	22	22	100
561 COVID 19 ANTIGEN											
2023-254				001	001		002 - Pos	22	1	21	0
2023-254				001	001		001 - Neg	22	21	21	100
2023-254	001	Bin		001	001		001 - Neg	169	169	169	100
2023-254	002	CareS		001	001		001 - Neg	199	199	199	100
2023-254	003	BD Veritor		001	001		001 - Neg	26	26	26	100
2023-254	004	QUICKVUE		001	001		001 - Neg	47	47	47	100
2023-254	005	INDICAID		001	001		001 - Neg	66	66	66	100
2023-254	006	Clarity		001	001		001 - Neg	73	72	72	100
2023-254	006	Clarity		001	001		002 - Pos	73	1	72	0
2023-254	007	Sofia		001	001		001 - Neg	49	49	49	100
2023-254	008	GenBody		001	001		001 - Neg	20	20	20	100
2023-254	009	Celltrion		001	001		001 - Neg	50	50	50	100
2023-254	011	Status		001	001		001 - Neg	22	22	22	100
2023-255				001	001		001 - Neg	22	22	22	100
2023-255	001	Bin		001	001		001 - Neg	169	169	169	100
2023-255	002	CareS		001	001		001 - Neg	199	199	199	100
2023-255	003	BD Veritor		001	001		001 - Neg	26	26	26	100
2023-255	004	QUICKVUE		001	001		001 - Neg	47	47	47	100
2023-255	005	INDICAID		001	001		001 - Neg	66	66	66	100

Participant Summary

MONTH 4 OF YEAR 2023

SPE: 50 MICROBIOLOGY			SUB: 180 Virology			GRP: 460 COVID 19 ANTIGEN						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-255	006	Clarity		001	001		001 - Neg	73	72	72	100	
2023-255	006	Clarity		001	001		002 - Pos	73	1	72	0	
2023-255	007	Sofia		001	001		001 - Neg	49	49	49	100	
2023-255	008	GenBody		001	001		001 - Neg	20	20	20	100	
2023-255	009	Celltrion		001	001		001 - Neg	50	50	50	100	
2023-255	011	Status		001	001		001 - Neg	22	22	22	100	