

Participant Summary
MONTH 12 OF YEAR 2022

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
159 BF RED CELL COUNT		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
2022-951				47	0.739	0.965	130.61	0	4.000		
2022-952				47	1038.391	136.732	13.17	628.000	1449.000		
2022-953				47	0.674	0.861	127.78	0	3.000		
2022-954				47	0.636	0.828	130.15	0	3.000		
2022-955				47	1053.065	127.551	12.11	670.000	1436.000		
160 BF WHITE CELL COUNT											
2022-951				46	6.178	2.014	32.61	0.000	12.000		
2022-952				46	266.250	36.479	13.70	157.000	376.000		
2022-953				46	6.682	1.578	23.61	2.000	11.000		
2022-954				46	6.721	1.604	23.87	2.000	12.000		
2022-955				46	280.093	35.715	12.75	173.000	387.000		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 225 HELICOBACTER PYLORI						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
155 HELICOBACTER PYLORI												
2022A131	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	24	24	24	100	
2022A132	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022A133	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	24	24	24	100	
2022A134	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022A135	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	24	24	24	100	

SPEC: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. LABS	PART	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
145 CRP - QUANTITATIVE											
			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}							
2022A101	34	DADE DIMEN		15		1.631	0.040	2.47	1.510	1.750	
2022A101	37	BECKMAN		12		1.422	0.094	6.60	1.140	1.700	
2022A101	39	ROCHE COBA		11		1.402	0.051	3.65	1.250	1.560	
2022A101	42	ASI ARLING		22		1.943	0.583	29.99	0.190	3.690	
2022A101	47	VITROS 560		15		2.359	0.151	6.39	1.910	2.810	
2022A102	34	DADE DIMEN		15		0.073	0.038	51.99	0	0.190	
2022A102	37	BECKMAN		12		0.390	0.267	68.52	0	1.190	

Participant Summary

MONTH 12 OF YEAR 2022

SPEC: 40 IMMUNOLOGY			SUB: 120 GENERAL			GRP: 220 ASOT, RA, MONO, CRP					
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022A102	39	ROCHE COBA		11	0.248	0.110	44.37	0	0.580		
2022A102	42	ASI ARLING		22	0.273	0.299	109.54	0	1.170		
2022A102	47	VITROS 560		15	0.500	0.000	0.00	0.500	0.500		
145 CRP - QUANTITATIVE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}							
2022A103	34	DADE DIMEN		15	3.199	0.106	3.31	2.880	3.520		
2022A103	37	BECKMAN		12	2.930	0.123	4.19	2.560	3.300		
2022A103	39	ROCHE COBA		11	2.966	0.113	3.80	2.630	3.300		
2022A103	42	ASI ARLING		22	2.536	1.084	42.75	0	5.790		
2022A103	47	VITROS 560		15	4.228	0.190	4.49	3.660	4.800		
2022A104	34	DADE DIMEN		15	1.547	0.063	4.07	1.360	1.740		
2022A104	37	BECKMAN		12	1.354	0.102	7.52	1.050	1.660		
2022A104	39	ROCHE COBA		11	1.375	0.037	2.71	1.260	1.490		
2022A104	42	ASI ARLING		22	1.691	0.590	34.89	0	3.460		
2022A104	47	VITROS 560		15	2.234	0.190	8.50	1.660	2.800		
2022A105	34	DADE DIMEN		15	0.061	0.039	63.63	0	0.180		
2022A105	37	BECKMAN		12	0.412	0.288	70.02	0	1.280		
2022A105	39	ROCHE COBA		11	0.246	0.114	46.19	0	0.590		
2022A105	42	ASI ARLING		22	0.273	0.299	109.54	0	1.170		
2022A105	47	VITROS 560		15	0.500	0.000	0.00	0.500	0.500		

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL			GRP: 220 ASOT, RA, MONO, CRP						
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART	# LABS	# "SF"	Score	
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
091 ANTISTREP.O-QUALITA.												
2022A101	384	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	32	32	32	100
2022A102	384	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	32	32	32	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
091 ANTISTREP.O-QUALITA.												
2022A103	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	32	32	32	100	
2022A104	384	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	32	32	32	100	
2022A105	384	ASI A.SCT.		002 002	002 002		002 - POSITIVE	32	32	32	100	
092 ANTISTREP.O-SEMI QUA												
2022A101	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	13	13	13	100	
2022A102	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	13	13	13	100	
2022A103	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	13	13	13	100	
2022A104	391	ASI SLIDE		014 014	014 014		014 - <100 IU/ml	13	13	13	100	
2022A105	391	ASI SLIDE		016 020	015 019		016 - 200 IU/ml	13	2	13	100	
2022A105	391	ASI SLIDE		016 020	015 019		017 - 400 IU/ml	13	4	13	100	
2022A105	391	ASI SLIDE		016 020	015 019		018 - 800 IU/ml	13	7	13	100	
103 INF. MONO- QUALITAT.												
2022A101	393	ASI A.SCT.		001 001	001 001		002 - POSITIVE	165	1	164	0	
2022A101	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	165	164	164	100	
2022A101	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2022A101	400	AIM MONO		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022A101	405	Acceava		001 001	001 001		001 - NEGATIVE	10	10	10	100	
2022A102	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	165	165	165	100	
2022A102	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	25	25	25	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022A102	400	AIM MONO		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022A102	405	Acceava		001 001	001 001		001 - NEGATIVE	10	10	10	100	
103 INF. MONO- QUALITAT.												
2022A103	393	ASI A.SCT.		001 001	001 001		002 - POSITIVE	165	1	164	0	
2022A103	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	165	164	164	100	
2022A103	394	TECO DIAG		001 001	001 001		002 - POSITIVE	25	1	24	0	
2022A103	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	25	24	24	100	
2022A103	400	AIM MONO		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022A103	405	Acceava		001 001	001 001		001 - NEGATIVE	10	10	10	100	
2022A104	393	ASI A.SCT.		002 002	002 002		002 - POSITIVE	165	163	163	100	
2022A104	393	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	165	2	163	0	
2022A104	394	TECO DIAG		002 002	002 002		002 - POSITIVE	25	25	25	100	
2022A104	400	AIM MONO		002 002	002 002		002 - POSITIVE	22	22	22	100	
2022A104	405	Acceava		002 002	002 002		002 - POSITIVE	10	10	10	100	
2022A105	393	ASI A.SCT.		001 001	001 001		002 - POSITIVE	165	2	163	0	
2022A105	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	165	163	163	100	
2022A105	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2022A105	400	AIM MONO		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022A105	405	Acceava		001 001	001 001		001 - NEGATIVE	10	10	10	100	
105 RF FACTOR(RA)-QUALI.												
2022A101	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	59	59	59	100	
2022A101	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	148	148	148	100	
2022A101	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	70	70	70	100	
2022A101	408	STANBIO		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022A102	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	59	59	59	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A102	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	148	147	147	100
2022A102	402	ASI A.SCT.		001 001	001 001		002 - POSITIVE	148	1	147	0
2022A102	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	70	65	65	100
2022A102	406	AIM CRP RF		001 001	001 001		002 - POSITIVE	70	5	65	0
2022A102	408	STANBIO		001 001	001 001		001 - NEGATIVE	24	24	24	100
105 RF FACTOR(RA)-QUALI.											
2022A103	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	59	58	58	100
2022A103	400	TECO DIAG.		001 001	001 001		002 - POSITIVE	59	1	58	0
2022A103	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	148	143	143	100
2022A103	402	ASI A.SCT.		001 001	001 001		002 - POSITIVE	148	5	143	0
2022A103	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	70	61	61	100
2022A103	406	AIM CRP RF		001 001	001 001		002 - POSITIVE	70	9	61	0
2022A103	408	STANBIO		001 001	001 001		001 - NEGATIVE	24	24	24	100
2022A104	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	59	59	59	100
2022A104	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	148	148	148	100
2022A104	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	70	70	70	100
2022A104	408	STANBIO		002 002	002 002		002 - POSITIVE	24	24	24	100
2022A105	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	59	59	59	100
2022A105	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	148	147	147	100
2022A105	402	ASI A.SCT.		001 001	001 001		002 - POSITIVE	148	1	147	0
2022A105	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	70	70	70	100
2022A105	408	STANBIO		001 001	001 001		001 - NEGATIVE	24	24	24	100
106 RF FAC(RA)-SEMI QUAN											
2022A101	402	TECO DIAG.		002 006	004 008		004 - 40-80 IU/m	14	14	14	100
2022A101	404	AIM CRP RF		003 007	004 008		004 - 40-80 IU/m	10	1	10	100
2022A101	404	AIM CRP RF		003 007	004 008		005 - 72-160 IU	10	4	10	100
2022A101	404	AIM CRP RF		003 007	004 008		006 - 136-320 IU	10	2	10	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY													SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP			
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score			
2022A101	404	AIM CRP RF					003	007	004	008			007 - 264-640 IU	10	3	10	100			
2022A101	407	ASI SLIDE					004	008	004	008			004 - 40-80 IU/m	22	2	22	100			
2022A101	407	ASI SLIDE					004	008	004	008			005 - 72-160 IU	22	5	22	100			
2022A101	407	ASI SLIDE					004	008	004	008			006 - 136-320 IU	22	8	22	100			
2022A101	407	ASI SLIDE					004	008	004	008			007 - 264-640 IU	22	6	22	100			
2022A101	407	ASI SLIDE					004	008	004	008			008 - 520-1280	22	1	22	100			
106 RF FAC(RA)-SEMI QUAN																				
2022A102	402	TECO DIAG.					001	001	001	001			001 - <8-20 IU/m	14	14	14	100			
2022A102	404	AIM CRP RF							001	001			001 - <8-20 IU/m	10	7	7	100			
2022A102	404	AIM CRP RF							001	001			002 - 8-20 IU/ml	10	3	7	0			
2022A102	407	ASI SLIDE					001	001	001	001			001 - <8-20 IU/m	22	22	22	100			
2022A103	402	TECO DIAG.					001	001	001	001			001 - <8-20 IU/m	14	14	14	100			
2022A103	404	AIM CRP RF							001	001			001 - <8-20 IU/m	10	7	7	100			
2022A103	404	AIM CRP RF							001	001			002 - 8-20 IU/ml	10	2	7	0			
2022A103	404	AIM CRP RF							001	001			003 - 24-40 IU/m	10	1	7	0			
2022A103	407	ASI SLIDE					001	001	001	001			001 - <8-20 IU/m	22	22	22	100			
2022A104	402	TECO DIAG.					002	005	004	008			002 - 8-20 IU/ml	14	1	14	100			
2022A104	402	TECO DIAG.					002	005	004	008			003 - 24-40 IU/m	14	13	14	100			
2022A104	404	AIM CRP RF					003	007	004	008			003 - 24-40 IU/m	10	1	10	100			
2022A104	404	AIM CRP RF					003	007	004	008			005 - 72-160 IU	10	5	10	100			
2022A104	404	AIM CRP RF					003	007	004	008			006 - 136-320 IU	10	3	10	100			
2022A104	404	AIM CRP RF					003	007	004	008			007 - 264-640 IU	10	1	10	100			
2022A104	407	ASI SLIDE					004	008	004	008			003 - 24-40 IU/m	22	1	21	0			
2022A104	407	ASI SLIDE					004	008	004	008			004 - 40-80 IU/m	22	3	21	100			
2022A104	407	ASI SLIDE					004	008	004	008			005 - 72-160 IU	22	4	21	100			
2022A104	407	ASI SLIDE					004	008	004	008			006 - 136-320 IU	22	12	21	100			
2022A104	407	ASI SLIDE					004	008	004	008			007 - 264-640 IU	22	2	21	100			

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
106 RF FAC(RA)-SEMI QUAN												
2022A105	402	TECO DIAG.		001 001	001 001		001 - <8-20 IU/m	14	14	14	100	
2022A105	404	AIM CRP RF		001 001	001 001		001 - <8-20 IU/m	10	10	10	100	
2022A105	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	22	22	22	100	
144 CRP- QUALITATIVE												
2022A101	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	131	131	131	100	
2022A101	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	39	39	39	100	
2022A101	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	54	54	54	100	
2022A101	041	STANBIO		002 002	002 002		002 - POSITIVE	15	15	15	100	
2022A102	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	131	131	131	100	
2022A102	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	39	39	39	100	
2022A102	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	54	54	54	100	
2022A102	041	STANBIO		001 001	001 001		001 - NEGATIVE	15	15	15	100	
2022A103	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	131	8	123	0	
2022A103	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	131	123	123	100	
2022A103	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	39	1	38	0	
2022A103	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	39	38	38	100	
2022A103	036	AIM CRP RF		002 002	002 002		001 - NEGATIVE	54	2	52	0	
2022A103	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	54	52	52	100	
2022A103	041	STANBIO		002 002	002 002		002 - POSITIVE	15	15	15	100	
2022A104	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	131	1	130	0	
2022A104	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	131	130	130	100	
2022A104	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	39	4	35	0	
2022A104	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	39	35	35	100	
2022A104	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	54	54	54	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY													SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score									
2022A104	041	STANBIO		002	002	002	002	002	-	POSITIVE	15	15	15	100						
144 CRP- QUALITATIVE																				
2022A105	026	ASI A.SCT.		001	001	001	001	001	-	NEGATIVE	131	130	130	100						
2022A105	026	ASI A.SCT.		001	001	001	001	002	-	POSITIVE	131	1	130	0						
2022A105	030	TECO DIAG.		001	001	001	001	001	-	NEGATIVE	39	39	39	100						
2022A105	036	AIM CRP RF		001	001	001	001	001	-	NEGATIVE	54	54	54	100						
2022A105	041	STANBIO		001	001	001	001	001	-	NEGATIVE	15	15	15	100						

SPE: 40 IMMUNOLOGY													SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC, HIV			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score									
093 ANTI-HUMAN IMM. (HIV)																				
2022-991	055	ALERE DETE		002	002	002	002	002	-	POS/R	29	29	29	100						
2022-992	055	ALERE DETE		002	002	002	002	002	-	POS/R	29	29	29	100						
2022-993	055	ALERE DETE		001	001	001	001	001	-	NEG/NR	29	29	29	100						
2022-994	055	ALERE DETE		002	002	002	002	002	-	POS/R	29	29	29	100						
2022-995	055	ALERE DETE		002	002	002	002	002	-	POS/R	29	29	29	100						

Participant Summary
MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 240 RUBELLA - QUALITAT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
107 RUBELLA-QUALITATIVE												
2022A081	363 ASI Rubell			002 002	002 002		002 - POS.OR IM.	10	10	10	100	
2022A082	363 ASI Rubell			001 001	001 001		001 - NEG-NOT IM	10	10	10	100	
2022A083	363 ASI Rubell			002 002	002 002		002 - POS.OR IM.	10	10	10	100	
2022A084	363 ASI Rubell			001 001	001 001		001 - NEG-NOT IM	10	10	10	100	
2022A085	363 ASI Rubell			002 002	002 002		002 - POS.OR IM.	10	10	10	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 247 MYCOPLASMA				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
548 MYCOPLASMA												
2022-901	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	562	2	560	0
2022-901	001	IMMU. CARD		002	002	002	002	002 - Reactive	562	560	560	100
2022-902	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	562	560	560	100
2022-902	001	IMMU. CARD		001	001	001	001	002 - Reactive	562	2	560	0
2022-903	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	562	562	562	100
2022-904	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	562	1	561	0
2022-904	001	IMMU. CARD		002	002	002	002	002 - Reactive	562	561	561	100
2022-905	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	562	1	561	0
2022-905	001	IMMU. CARD		002	002	002	002	002 - Reactive	562	561	561	100

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
554 COVID-19 ANTBODY													
2022A051	001	IgM	002	HEALGEN	002	002	002	002	001 - Negative	32	1	31	0
2022A051	001	IgM	002	HEALGEN	002	002	002	002	002 - Positive	32	31	31	100
2022A051	001	IgM	016	CLARITY	002	002	002	002	002 - Positive	11	11	11	100
2022A051	001	IgM	030	FASTEP	002	002	002	002	002 - Positive	70	70	70	100
2022A051	001	IgM	031	ASSURE COV	002	002	002	002	002 - Positive	14	14	14	100
2022A051	002	IgG	002	HEALGEN	002	002	002	002	002 - Positive	32	32	32	100
2022A051	002	IgG	016	CLARITY	002	002	002	002	002 - Positive	10	10	10	100
2022A051	002	IgG	030	FASTEP	002	002	002	002	002 - Positive	70	70	70	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A051	002 IgG	031 ASSURE COV		002 002	002 002		002 - Positive	14	14	14	100		
554 COVID-19 ANTBODY													
2022A052	001 IgM	002 HEALGEN		001 001	001 001		001 - Negative	32	32	32	100		
2022A052	001 IgM	016 CLARITY		001 001	001 001		001 - Negative	10	10	10	100		
2022A052	001 IgM	030 FASTEP		001 001	001 001		001 - Negative	70	70	70	100		
2022A052	001 IgM	031 ASSURE COV		001 001	001 001		001 - Negative	14	14	14	100		
2022A052	002 IgG	002 HEALGEN		001 001	001 001		001 - Negative	32	32	32	100		
2022A052	002 IgG	016 CLARITY		001 001	001 001		001 - Negative	10	10	10	100		
2022A052	002 IgG	030 FASTEP		001 001	001 001		001 - Negative	70	70	70	100		
2022A052	002 IgG	031 ASSURE COV		001 001	001 001		001 - Negative	14	14	14	100		
2022A053	001 IgM	002 HEALGEN		002 002	002 002		001 - Negative	32	1	31	0		
2022A053	001 IgM	002 HEALGEN		002 002	002 002		002 - Positive	32	31	31	100		
2022A053	001 IgM	016 CLARITY		002 002	002 002		002 - Positive	10	10	10	100		
2022A053	001 IgM	030 FASTEP		002 002	002 002		001 - Negative	70	2	68	0		
2022A053	001 IgM	030 FASTEP		002 002	002 002		002 - Positive	70	68	68	100		
2022A053	001 IgM	031 ASSURE COV		002 002	002 002		002 - Positive	14	14	14	100		
2022A053	002 IgG	002 HEALGEN		002 002	002 002		002 - Positive	32	32	32	100		
2022A053	002 IgG	016 CLARITY		002 002	002 002		002 - Positive	10	10	10	100		
2022A053	002 IgG	030 FASTEP		002 002	002 002		002 - Positive	70	70	70	100		
2022A053	002 IgG	031 ASSURE COV		002 002	002 002		002 - Positive	14	14	14	100		
2022A054	001 IgM	002 HEALGEN		002 002	002 002		001 - Negative	32	1	31	0		
2022A054	001 IgM	002 HEALGEN		002 002	002 002		002 - Positive	32	31	31	100		
2022A054	001 IgM	016 CLARITY		002 002	002 002		002 - Positive	10	10	10	100		
2022A054	001 IgM	030 FASTEP		002 002	002 002		001 - Negative	69	2	67	0		
2022A054	001 IgM	030 FASTEP		002 002	002 002		002 - Positive	69	67	67	100		
2022A054	001 IgM	031 ASSURE COV		002 002	002 002		002 - Positive	14	14	14	100		
2022A054	002 IgG	002 HEALGEN		002 002	002 002		002 - Positive	32	32	32	100		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A054	002 IgG	016 CLARITY		002 002	002 002		002 - Positive	10	10	10	100		
2022A054	002 IgG	030 FASTEP		002 002	002 002		002 - Positive	71	71	71	100		
2022A054	002 IgG	031 ASSURE COV		002 002	002 002		002 - Positive	14	14	14	100		
554 COVID-19 ANTBODY													
2022A055	001 IgM	002 HEALGEN		001 001	001 001		001 - Negative	32	32	32	100		
2022A055	001 IgM	030 FASTEP		001 001	001 001		001 - Negative	70	70	70	100		
2022A055	001 IgM	031 ASSURE COV		001 001	001 001		001 - Negative	14	14	14	100		
2022A055	002 IgG	002 HEALGEN		001 001	001 001		001 - Negative	32	32	32	100		
2022A055	002 IgG	016 CLARITY		001 001	001 001		001 - Negative	10	10	10	100		
2022A055	002 IgG	030 FASTEP		001 001	001 001		001 - Negative	70	70	70	100		
2022A055	002 IgG	031 ASSURE COV		001 001	001 001		001 - Negative	14	14	14	100		

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
109 SYPHILLIS SERO.-QUAL													
2022A071	002 RPR	028 BLOKIT-RPR		004 004	004 004		004 - React RPR	10	10	10	100		
2022A071	002 RPR	030 IMMUNOSTIC		004 004	004 004		001 - NR Non Rea	33	1	32	0		
2022A071	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	33	32	32	100		
2022A071	002 RPR	033 ASI		004 004	004 004		001 - NR Non Rea	536	1	535	0		
2022A071	002 RPR	033 ASI		004 004	004 004		004 - React RPR	536	535	535	100		
2022A071	002 RPR	040 TECO DIAG.		004 004	004 004		001 - NR Non Rea	66	5	61	0		
2022A071	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	61	61	100		
2022A071	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	92	92	92	100		
2022A071	002 RPR	057 Germain La		004 004	004 004		001 - NR Non Rea	27	2	25	0		
2022A071	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	25	25	100		
2022A072	002 RPR	028 BLOKIT-RPR		001 001	001 001		001 - NR Non Rea	10	10	10	100		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A072	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR Non Rea	33	31	31	100		
2022A072	002 RPR	030 IMMUNOSTIC		001 001	001 001		004 - React RPR	33	2	31	0		
2022A072	002 RPR	033 ASI		001 001	001 001		001 - NR Non Rea	536	534	534	100		
2022A072	002 RPR	033 ASI		001 001	001 001		004 - React RPR	536	2	534	0		
2022A072	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR Non Rea	66	66	66	100		
2022A072	002 RPR	053 AIM RPR		001 001	001 001		001 - NR Non Rea	92	92	92	100		
2022A072	002 RPR	057 Germain La		001 001	001 001		001 - NR Non Rea	27	26	26	100		
2022A072	002 RPR	057 Germain La		001 001	001 001		004 - React RPR	27	1	26	0		
109 SYPHILLIS SERO.-QUAL													
2022A073	002 RPR	028 BLOKIT-RPR		004 004	004 004		004 - React RPR	10	10	10	100		
2022A073	002 RPR	030 IMMUNOSTIC		004 004	004 004		001 - NR Non Rea	33	2	31	0		
2022A073	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	33	31	31	100		
2022A073	002 RPR	033 ASI		004 004	004 004		004 - React RPR	536	536	536	100		
2022A073	002 RPR	040 TECO DIAG.		004 004	004 004		001 - NR Non Rea	66	1	65	0		
2022A073	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	65	65	100		
2022A073	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	92	92	92	100		
2022A073	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	27	27	100		
2022A074	002 RPR	028 BLOKIT-RPR		004 004	004 004		004 - React RPR	10	10	10	100		
2022A074	002 RPR	030 IMMUNOSTIC		004 004	004 004		001 - NR Non Rea	33	2	31	0		
2022A074	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	33	31	31	100		
2022A074	002 RPR	033 ASI		004 004	004 004		002 - WR Weakly	536	1	533	0		
2022A074	002 RPR	033 ASI		004 004	004 004		001 - NR Non Rea	536	2	533	0		
2022A074	002 RPR	033 ASI		004 004	004 004		004 - React RPR	536	533	533	100		
2022A074	002 RPR	040 TECO DIAG.		004 004	004 004		001 - NR Non Rea	66	2	64	0		
2022A074	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	66	64	64	100		
2022A074	002 RPR	053 AIM RPR		004 004	004 004		001 - NR Non Rea	92	1	91	0		
2022A074	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	92	91	91	100		
2022A074	002 RPR	057 Germain La		004 004	004 004		001 - NR Non Rea	27	1	26	0		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A074	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	27	26	26	100
109 SYPHILLIS SERO.-QUAL											
2022A075	002 RPR	028 BLOKIT-RPR		001 001	001 001		001 - NR Non Rea	10	10	10	100
2022A075	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR Non Rea	33	31	31	100
2022A075	002 RPR	030 IMMUNOSTIC		001 001	001 001		004 - React RPR	33	2	31	0
2022A075	002 RPR	033 ASI		001 001	001 001		001 - NR Non Rea	536	534	534	100
2022A075	002 RPR	033 ASI		001 001	001 001		004 - React RPR	536	2	534	0
2022A075	002 RPR	040 TECO DIAG.		001 001	001 001		002 - WR Weakly	66	1	65	0
2022A075	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR Non Rea	66	65	65	100
2022A075	002 RPR	053 AIM RPR		001 001	001 001		001 - NR Non Rea	92	91	91	100
2022A075	002 RPR	053 AIM RPR		001 001	001 001		004 - React RPR	92	1	91	0
2022A075	002 RPR	057 Germain La		001 001	001 001		001 - NR Non Rea	27	26	26	100
2022A075	002 RPR	057 Germain La		001 001	001 001		004 - React RPR	27	1	26	0
110 SYPHILLIS SERO-QUANT											
2022A071	002 RPR	030 IMMUNOSTIC		003 005	003 005		001 - NR	13	1	12	0
2022A071	002 RPR	030 IMMUNOSTIC		003 005	003 005		004 - R- 2 DILS.	13	11	12	100
2022A071	002 RPR	030 IMMUNOSTIC		003 005	003 005		005 - R- 4 DILS.	13	1	12	100
2022A071	002 RPR	033 ASI		003 005	003 005	003 005	003 - R- 1 DIL.	245	4	244	100
2022A071	002 RPR	033 ASI		003 005	003 005	003 005	004 - R- 2 DILS.	245	198	244	100
2022A071	002 RPR	033 ASI		003 005	003 005	003 005	005 - R- 4 DILS.	245	42	244	100
2022A071	002 RPR	033 ASI		003 005	003 005	003 005	008 - R- 32 DILS	245	1	244	0
2022A071	002 RPR	040 TECO DIAG.		003 005	003 005		001 - NR	16	1	15	0
2022A071	002 RPR	040 TECO DIAG.		003 005	003 005		003 - R- 1 DIL.	16	1	15	100
2022A071	002 RPR	040 TECO DIAG.		003 005	003 005		004 - R- 2 DILS.	16	13	15	100
2022A071	002 RPR	040 TECO DIAG.		003 005	003 005		005 - R- 4 DILS.	16	1	15	100
2022A071	002 RPR	053 AIM RPR		003 005	003 005		003 - R- 1 DIL.	19	8	19	100
2022A071	002 RPR	053 AIM RPR		003 005	003 005		004 - R- 2 DILS.	19	11	19	100
2022A072	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR	13	13	13	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY									
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A072	002	RPR	033	ASI			001	001	001	001	001	001	001 - NR	245	244	244	100
2022A072	002	RPR	033	ASI			001	001	001	001	001	001	003 - R- 1 DIL.	245	1	244	0
2022A072	002	RPR	040	TECO DIAG.			001	001	001	001			001 - NR	16	16	16	100
2022A072	002	RPR	053	AIM RPR			001	001	001	001			001 - NR	19	19	19	100
110 SYPHILLIS SERO-QUANT																	
2022A073	002	RPR	030	IMMUNOSTIC			004	006	004	006			004 - R- 2 DILS.	13	3	13	100
2022A073	002	RPR	030	IMMUNOSTIC			004	006	004	006			005 - R- 4 DILS.	13	9	13	100
2022A073	002	RPR	030	IMMUNOSTIC			004	006	004	006			006 - R- 8 DILS.	13	1	13	100
2022A073	002	RPR	033	ASI			004	006	004	006	004	006	004 - R- 2 DILS.	245	24	245	100
2022A073	002	RPR	033	ASI			004	006	004	006	004	006	005 - R- 4 DILS.	245	204	245	100
2022A073	002	RPR	033	ASI			004	006	004	006	004	006	006 - R- 8 DILS.	245	17	245	100
2022A073	002	RPR	040	TECO DIAG.			004	006	004	006			004 - R- 2 DILS.	16	6	16	100
2022A073	002	RPR	040	TECO DIAG.			004	006	004	006			005 - R- 4 DILS.	16	10	16	100
2022A073	002	RPR	053	AIM RPR			004	006	004	006			004 - R- 2 DILS.	19	9	19	100
2022A073	002	RPR	053	AIM RPR			004	006	004	006			005 - R- 4 DILS.	19	10	19	100
2022A074	002	RPR	030	IMMUNOSTIC			003	005	003	005			003 - R- 1 DIL.	13	2	13	100
2022A074	002	RPR	030	IMMUNOSTIC			003	005	003	005			004 - R- 2 DILS.	13	8	13	100
2022A074	002	RPR	030	IMMUNOSTIC			003	005	003	005			005 - R- 4 DILS.	13	3	13	100
2022A074	002	RPR	033	ASI			003	005	003	005	003	005	001 - NR	245	2	243	0
2022A074	002	RPR	033	ASI			003	005	003	005	003	005	003 - R- 1 DIL.	245	8	243	100
2022A074	002	RPR	033	ASI			003	005	003	005	003	005	004 - R- 2 DILS.	245	205	243	100
2022A074	002	RPR	033	ASI			003	005	003	005	003	005	005 - R- 4 DILS.	245	30	243	100
2022A074	002	RPR	040	TECO DIAG.			003	005	003	005			003 - R- 1 DIL.	16	1	16	100
2022A074	002	RPR	040	TECO DIAG.			003	005	003	005			004 - R- 2 DILS.	16	13	16	100
2022A074	002	RPR	040	TECO DIAG.			003	005	003	005			005 - R- 4 DILS.	16	2	16	100
2022A074	002	RPR	053	AIM RPR			003	005	003	005			003 - R- 1 DIL.	19	5	19	100
2022A074	002	RPR	053	AIM RPR			003	005	003	005			004 - R- 2 DILS.	19	12	19	100
2022A074	002	RPR	053	AIM RPR			003	005	003	005			005 - R- 4 DILS.	19	2	19	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY									
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
110 SYPHILLIS SERO-QUANT																	
2022A075	002	RPR	030	IMMUNOSTIC			001	001	001	001			001 - NR	13	13	13	100
2022A075	002	RPR	033	ASI			001	001	001	001	001	001	001 - NR	245	243	243	100
2022A075	002	RPR	033	ASI			001	001	001	001	001	001	004 - R- 2 DILS.	245	2	243	0
2022A075	002	RPR	040	TECO DIAG.			001	001	001	001			001 - NR	16	16	16	100
2022A075	002	RPR	053	AIM RPR			001	001	001	001			001 - NR	19	19	19	100
111 BACTER. GRAM STAIN																	
SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 260 BACTER. -GRAM STAIN									
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A001							002	002	002	002			004 - GRAM POS C	98	8	79	0
2022A001							002	002	002	002			006 - GRAM POS C	98	3	79	0
2022A001							002	002	002	002			008 - GRAM POS C	98	1	79	0
2022A001							002	002	002	002			001 - GRAM NEG B	98	7	79	0
2022A001							002	002	002	002			002 - GRAM POS B	98	79	79	100
2022A002							003	003	003	003			003 - GRAM NEG C	98	95	95	100
2022A002							003	003	003	003			004 - GRAM POS C	98	2	95	0
2022A002							003	003	003	003			001 - GRAM NEG B	98	1	95	0
2022A003							003	003	003	003			003 - GRAM NEG C	98	94	94	100
2022A003							003	003	003	003			004 - GRAM POS C	98	2	94	0
2022A003							003	003	003	003			001 - GRAM NEG B	98	2	94	0
2022A004							002	002	002	002			004 - GRAM POS C	98	4	80	0
2022A004							002	002	002	002			005 - GRAM NEG C	98	1	80	0

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 260 BACTER. -GRAM STAIN				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022A004				002 002	002 002		006 - GRAM POS C	98	5	80	0	
2022A004				002 002	002 002		008 - GRAM POS C	98	1	80	0	
2022A004				002 002	002 002		001 - GRAM NEG B	98	7	80	0	
2022A004				002 002	002 002		002 - GRAM POS B	98	80	80	100	
111 BACTER. GRAM STAIN												
2022A005				002 002	002 002		004 - GRAM POS C	98	1	82	0	
2022A005				002 002	002 002		006 - GRAM POS C	98	8	82	0	
2022A005				002 002	002 002		008 - GRAM POS C	98	1	82	0	
2022A005				002 002	002 002		009 - GRAM POS B	98	1	82	0	
2022A005				002 002	002 002		001 - GRAM NEG B	98	5	82	0	
2022A005				002 002	002 002		002 - GRAM POS B	98	82	82	100	

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
112 BACTER. -IDENTIFICAT.												
2022A021	001 AEROBIC	001 CONVENCIO.			075 075		075 - Enter faec	20	4	14	100	
2022A021	001 AEROBIC	001 CONVENCIO.			075 075		120 - Neisse. sp	20	5	14	0	
2022A021	001 AEROBIC	001 CONVENCIO.			075 075		256 - No Org. IS	20	1	14	0	
2022A021	001 AEROBIC	001 CONVENCIO.			075 075		259 - g+ cocci	20	1	14	0	
2022A021	001 AEROBIC	001 CONVENCIO.			075 075		260 - g - cocci	20	3	14	0	
2022A021	001 AEROBIC	001 CONVENCIO.			121 121		120 - Neisse. sp	20	5	14	100	
2022A021	001 AEROBIC	001 CONVENCIO.			121 121		121 - Neisse. go	20	6	14	100	
2022A021	001 AEROBIC	032 MICROSCAN			075 075		075 - Enter faec	23	22	23	100	
2022A021	001 AEROBIC	032 MICROSCAN			121 121		121 - Neisse. go	23	1	23	100	
2022A021	001 AEROBIC	034 MICROSCAN			075 075		075 - Enter faec	26	20	26	100	
2022A021	001 AEROBIC	034 MICROSCAN			121 121		120 - Neisse. sp	26	1	26	100	
2022A021	001 AEROBIC	034 MICROSCAN			121 121		121 - Neisse. go	26	5	26	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A021	001 AEROBIC	046 MICROSCAN			075 075		075 - Enter faec	30	29	30	100
2022A021	001 AEROBIC	046 MICROSCAN			121 121		121 - Naisse. go	30	1	30	100
2022A021	001 AEROBIC	049 VITEK 2			075 075		075 - Enter faec	45	37	45	100
2022A021	001 AEROBIC	049 VITEK 2			121 121		120 - Naisse. sp	45	1	45	100
2022A021	001 AEROBIC	049 VITEK 2			121 121		121 - Naisse. go	45	7	45	100
112 BACTER.-IDENTIFICAT.											
2022A022	001 AEROBIC	001 CONVENCIO.			096 096		092 - Haemo. sp.	25	9	18	100
2022A022	001 AEROBIC	001 CONVENCIO.			096 096		095 - Haemo. he.	25	1	18	0
2022A022	001 AEROBIC	001 CONVENCIO.			096 096		096 - Haemo. in.	25	2	18	100
2022A022	001 AEROBIC	001 CONVENCIO.			187 187		256 - No Org. IS	25	1	18	0
2022A022	001 AEROBIC	001 CONVENCIO.			187 187		258 - g -bacilli	25	1	18	0
2022A022	001 AEROBIC	001 CONVENCIO.			187 187		259 - g+ cocci	25	2	18	0
2022A022	001 AEROBIC	001 CONVENCIO.			187 187		260 - g - cocci	25	2	18	0
2022A022	001 AEROBIC	001 CONVENCIO.			187 187		187 - Staphy. au	25	5	18	100
2022A022	001 AEROBIC	001 CONVENCIO.			188 188		188 - Staphy. ep	25	2	18	100
2022A022	001 AEROBIC	032 MICROSCAN			096 096		075 - Enter faec	48	1	47	0
2022A022	001 AEROBIC	032 MICROSCAN			187 187		187 - Staphy. au	48	23	47	100
2022A022	001 AEROBIC	032 MICROSCAN			188 188		188 - Staphy. ep	48	24	47	100
2022A022	001 AEROBIC	034 MICROSCAN			096 096		096 - Haemo. in.	53	5	53	100
2022A022	001 AEROBIC	034 MICROSCAN			187 187		187 - Staphy. au	53	25	53	100
2022A022	001 AEROBIC	034 MICROSCAN			188 188		188 - Staphy. ep	53	23	53	100
2022A022	001 AEROBIC	042 VITEK(AMS)			187 187		187 - Staphy. au	10	5	10	100
2022A022	001 AEROBIC	042 VITEK(AMS)			188 188		188 - Staphy. ep	10	5	10	100
2022A022	001 AEROBIC	046 MICROSCAN			096 096		092 - Haemo. sp.	63	2	62	100
2022A022	001 AEROBIC	046 MICROSCAN			096 096		096 - Haemo. in.	63	2	62	100
2022A022	001 AEROBIC	046 MICROSCAN			096 096		189 - Staphy. he	63	1	62	0
2022A022	001 AEROBIC	046 MICROSCAN			187 187		187 - Staphy. au	63	30	62	100
2022A022	001 AEROBIC	046 MICROSCAN			188 188		188 - Staphy. ep	63	28	62	100
2022A022	001 AEROBIC	049 VITEK 2			096 096		092 - Haemo. sp.	89	2	89	100
2022A022	001 AEROBIC	049 VITEK 2			096 096		096 - Haemo. in.	89	8	89	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A022	001 AEROBIC	049 VITEK 2			187 187		187 - Staphy. au	89	41	89	100
2022A022	001 AEROBIC	049 VITEK 2			188 188		188 - Staphy. ep	89	38	89	100
112 BACTER.-IDENTIFICAT.											
2022A023	001 AEROBIC	032 MICROSCAN			068 068		251 - NO S.Sh.Y.	21	4	17	0
2022A023	001 AEROBIC	032 MICROSCAN			068 068		067 - Escher, sp	21	1	17	100
2022A023	001 AEROBIC	032 MICROSCAN			068 068		068 - Esche coli	21	16	17	100
2022A023	001 AEROBIC	034 MICROSCAN			068 068		251 - NO S.Sh.Y.	24	5	19	0
2022A023	001 AEROBIC	034 MICROSCAN			068 068		068 - Esche coli	24	19	19	100
2022A023	001 AEROBIC	046 MICROSCAN			068 068		251 - NO S.Sh.Y.	29	7	22	0
2022A023	001 AEROBIC	046 MICROSCAN			068 068		067 - Escher, sp	29	1	22	100
2022A023	001 AEROBIC	046 MICROSCAN			068 068		068 - Esche coli	29	21	22	100
2022A023	001 AEROBIC	049 VITEK 2			068 068		251 - NO S.Sh.Y.	40	3	37	0
2022A023	001 AEROBIC	049 VITEK 2			068 068		067 - Escher, sp	40	2	37	100
2022A023	001 AEROBIC	049 VITEK 2			068 068		068 - Esche coli	40	35	37	100
2022A024	001 AEROBIC	032 MICROSCAN			046 046		043 - Citrobacte	24	4	24	100
2022A024	001 AEROBIC	032 MICROSCAN			046 046		046 - Citrobacte	24	20	24	100
2022A024	001 AEROBIC	034 MICROSCAN			046 046		043 - Citrobacte	25	6	25	100
2022A024	001 AEROBIC	034 MICROSCAN			046 046		046 - Citrobacte	25	19	25	100
2022A024	001 AEROBIC	046 MICROSCAN			046 046		043 - Citrobacte	31	1	31	100
2022A024	001 AEROBIC	046 MICROSCAN			046 046		046 - Citrobacte	31	30	31	100
2022A024	001 AEROBIC	049 VITEK 2			046 046		046 - Citrobacte	42	42	42	100
2022A025	001 AEROBIC	032 MICROSCAN			075 075		075 - Enter faec	24	24	24	100
2022A025	001 AEROBIC	034 MICROSCAN			075 075		075 - Enter faec	25	25	25	100
2022A025	001 AEROBIC	046 MICROSCAN			075 075		075 - Enter faec	30	30	30	100
2022A025	001 AEROBIC	049 VITEK 2			075 075		075 - Enter faec	42	42	42	100

Participant Summary
MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
113 BACT.ANTIM. SUSCEPT.												
2022A024	004 MICRO SCAN	001 AMIKACIN	000 Not apply		001 001		001 - SUSCEPTIB.	51	51	51	100	
2022A024	004 MICRO SCAN	004 AMPIC/SULB	000 Not apply		001 001		001 - SUSCEPTIB.	20	20	20	100	
2022A024	004 MICRO SCAN	007 AZTREONAM	000 Not apply		001 001		001 - SUSCEPTIB.	11	11	11	100	
2022A024	004 MICRO SCAN	017 CEFOTAXIME	000 Not apply		001 001		001 - SUSCEPTIB.	13	13	13	100	
2022A024	004 MICRO SCAN	020 CEFTAZIDIM	000 Not apply		001 001		001 - SUSCEPTIB.	12	12	12	100	
2022A024	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	65	65	65	100	
2022A024	004 MICRO SCAN	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	55	55	55	100	
2022A024	004 MICRO SCAN	038 IMIPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	24	24	24	100	
2022A024	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	57	57	57	100	
2022A024	004 MICRO SCAN	062 TETRACYCLI	000 Not apply		001 001		001 - SUSCEPTIB.	13	13	13	100	
2022A024	004 MICRO SCAN	065 TOBRAMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	55	55	55	100	
2022A024	004 MICRO SCAN	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	64	64	64	100	
2022A024	004 MICRO SCAN	072 CEFEPIME	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	34	100	
2022A024	004 MICRO SCAN	073 PIPER/TAZO	000 Not apply		001 001		001 - SUSCEPTIB.	19	19	19	100	
2022A024	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	43	43	43	100	
2022A024	004 MICRO SCAN	087 MEROPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	10	10	10	100	
2022A024	004 MICRO SCAN	096 ERTAPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	15	15	15	100	
2022A024	006 VITEK SYS.	001 AMIKACIN	000 Not apply		001 001		001 - SUSCEPTIB.	15	15	15	100	
2022A024	006 VITEK SYS.	020 CEFTAZIDIM	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	18	100	
2022A024	006 VITEK SYS.	022 CEFTRIAXON	000 Not apply		001 001		001 - SUSCEPTIB.	12	12	12	100	
2022A024	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	43	43	43	100	
2022A024	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	32	32	32	100	
2022A024	006 VITEK SYS.	038 IMIPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	27	27	27	100	
2022A024	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	42	100	
2022A024	006 VITEK SYS.	065 TOBRAMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	29	29	29	100	
2022A024	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	38	38	38	100	
2022A024	006 VITEK SYS.	072 CEFEPIME	000 Not apply		001 001		001 - SUSCEPTIB.	29	29	29	100	
2022A024	006 VITEK SYS.	073 PIPER/TAZO	000 Not apply		001 001		001 - SUSCEPTIB.	10	10	10	100	
2022A024	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	22	22	22	100	

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A024	006 VITEK SYS.	096 ERTAPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	19	19	19	100		
113 BACT.ANTIM. SUSCEPT.													
2022A025	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	66	66	66	100		
2022A025	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	64	64	64	100		
2022A025	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		004 - RESISTANT	56	1	55	0		
2022A025	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	56	55	55	100		
2022A025	004 MICRO SCAN	054 PENICILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	34	100		
2022A025	004 MICRO SCAN	068 VANCOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	54	54	54	100		
2022A025	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	53	53	53	100		
2022A025	004 MICRO SCAN	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	53	53	53	100		
2022A025	004 MICRO SCAN	097 DAPTOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	31	31	31	100		
2022A025	006 VITEK SYS.	003 AMPICILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	39	39	39	100		
2022A025	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	41	41	40	100		
2022A025	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	41	41	40	0		
2022A025	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	100		
2022A025	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	0		
2022A025	006 VITEK SYS.	054 PENICILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	12	12	12	100		
2022A025	006 VITEK SYS.	062 TETRACYCLI	000 Not apply		004 004		004 - RESISTANT	11	11	11	100		
2022A025	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	41	41	41	100		
2022A025	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	34	100		
2022A025	006 VITEK SYS.	084 GENTA.-500	000 Not apply		001 001		001 - SUSCEPTIB.	10	10	10	100		
2022A025	006 VITEK SYS.	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	36	36	36	100		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY		SUB: 140 BACTERIOLOGY				GRP: 263 BACTER.- DIRECT ANT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
147 BACTER.DIRECT ANTIG.											
2022A011				002 002	002 002		002 - POSITIVE	10	10	10	100
2022A012				001 001	001 001		001 - NEGATIVE	10	10	10	100
2022A013				002 002	002 002		002 - POSITIVE	10	10	10	100
2022A014				001 001	001 001		001 - NEGATIVE	10	9	9	100
2022A014				001 001	001 001		002 - POSITIVE	10	1	9	0
2022A015				002 002	002 002		002 - POSITIVE	10	10	10	100

SPE: 50 MICROBIOLOGY		SUB: 170 PARASITOLOGY				GRP: 350 ID OF PARASITES					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
120 IDENT OF PARASITES											
2022A061	013	Balantidi.	001	TROPHOZOIT		002 004	006 - PAR NOT RE	67	9	58	0
2022A061	013	Balantidi.	001	TROPHOZOIT		002 004	002 - FEW	67	2	58	100
2022A061	013	Balantidi.	001	TROPHOZOIT		002 004	003 - MODERATE	67	17	58	100
2022A061	013	Balantidi.	001	TROPHOZOIT		002 004	004 - MANY	67	39	58	100
2022A061	013	Balantidi.	002	CYSTS		001 001	002 - FEW	15	5	15	100
2022A061	013	Balantidi.	002	CYSTS		001 001	003 - MODERATE	15	4	15	100
2022A061	013	Balantidi.	002	CYSTS		001 001	004 - MANY	15	6	15	100
2022A061	014	Enterobius	003	OVA OR EGG		002 004	006 - PAR NOT RE	67	7	60	0
2022A061	014	Enterobius	003	OVA OR EGG		002 004	002 - FEW	67	57	60	100
2022A061	014	Enterobius	003	OVA OR EGG		002 004	003 - MODERATE	67	3	60	100
2022A061	018	Strongylo.	004	LARVAE		002 004	006 - PAR NOT RE	67	6	61	0

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 350 ID OF PARASITES			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A061	018	Strongylo.	004	LARVAE	002	004	002 - FEW	67	58	61	100
2022A061	018	Strongylo.	004	LARVAE	002	004	003 - MODERATE	67	3	61	100
120 IDENT OF PARASITES											
2022A062	033	Paragonim.	003	OVA OR EGG	002	004	002 - FEW	67	38	67	100
2022A062	033	Paragonim.	003	OVA OR EGG	002	004	003 - MODERATE	67	28	67	100
2022A062	033	Paragonim.	003	OVA OR EGG	002	004	004 - MANY	67	1	67	100
2022A063	001	E. histoly	002	CYSTS	001	001	002 - FEW	27	14	27	100
2022A063	001	E. histoly	002	CYSTS	001	001	003 - MODERATE	27	12	27	100
2022A063	001	E. histoly	002	CYSTS	001	001	004 - MANY	27	1	27	100
2022A063	003	E. coli	002	CYSTS	002	004	006 - PAR NOT RE	67	2	65	0
2022A063	003	E. coli	002	CYSTS	002	004	002 - FEW	67	2	65	100
2022A063	003	E. coli	002	CYSTS	002	004	003 - MODERATE	67	20	65	100
2022A063	003	E. coli	002	CYSTS	002	004	004 - MANY	67	43	65	100
2022A063	005	Ioda. buts	002	CYSTS	001	001	002 - FEW	17	13	17	100
2022A063	005	Ioda. buts	002	CYSTS	001	001	003 - MODERATE	17	4	17	100
2022A063	028	Schistoso.	003	OVA OR EGG	001	001	002 - FEW	14	14	14	100
2022A064	091	No trichro	005	NOT APPLY	001	001	005 - NOT APPLY	64	64	64	100
2022A065	015	Trichuris	003	OVA OR EGG	005	005	002 - FEW	67	2	67	100
2022A065	015	Trichuris	003	OVA OR EGG	005	005	005 - NOT APPLY	67	65	67	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 360 PRES/ABS OF PARASITE			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
121 PRESENCE OR ABSENCE											
2022A061				002 002	002 002		002 - Presence	35	35	35	100
2022A062				002 002	002 002		001 - Absence	34	1	33	0
2022A062				002 002	002 002		002 - Presence	34	33	33	100
2022A063				002 002	002 002		002 - Presence	35	35	35	100
2022A064				003 003	003 003		003 - NO TRICHRO	34	34	34	100
2022A065				002 002	002 002		002 - Presence	34	34	34	100

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
176 INFLUENZA A&B QUALI											
2022A031	001 INFLU. A	001 BINAX NOW		002 002	002 002		002 - POSITIVE	14	14	14	100
2022A031	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	95	95	95	100
2022A031	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	108	108	108	100
2022A031	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	53	53	53	100
2022A031	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	25	25	25	100
2022A031	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	17	17	17	100
2022A031	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	350	350	350	100
2022A031	001 INFLU. A	016 STATUS FLU		002 002	002 002		002 - POSITIVE	62	62	62	100
2022A031	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100
2022A031	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	95	95	95	100
2022A031	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	108	108	108	100

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A031	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100		
2022A031	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A031	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A031	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	350	350	350	100		
2022A031	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	62	62	62	100		
176	INFLUENZA A&B QUALI												
2022A032	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A032	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	94	94	94	100		
2022A032	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	107	107	107	100		
2022A032	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100		
2022A032	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A032	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A032	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	350	350	350	100		
2022A032	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	63	63	63	100		
2022A032	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A032	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	95	95	95	100		
2022A032	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	108	108	108	100		
2022A032	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100		
2022A032	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A032	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A032	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	350	350	350	100		
2022A032	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	61	61	61	100		
2022A033	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A033	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	96	96	96	100		
2022A033	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	108	108	108	100		
2022A033	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100		
2022A033	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A033	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022A033	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	350	350	350	100		
2022A033	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	63	63	63	100		
2022A033	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A033	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	93	93	93	100		
2022A033	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	108	108	108	100		
2022A033	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	54	54	54	100		
2022A033	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A033	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A033	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	350	350	350	100		
2022A033	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	61	61	61	100		
176	INFLUENZA A&B QUALI												
2022A034	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A034	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	95	95	95	100		
2022A034	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	110	110	110	100		
2022A034	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	52	52	52	100		
2022A034	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A034	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A034	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	352	352	352	100		
2022A034	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	62	62	62	100		
2022A034	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		
2022A034	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	95	95	95	100		
2022A034	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	107	107	107	100		
2022A034	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100		
2022A034	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100		
2022A034	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100		
2022A034	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	348	348	348	100		
2022A034	002 INFLU. B	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	62	62	62	100		
2022A035	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	14	14	14	100		

Participant Summary

MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A & B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022A035	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	95	95	95	100
2022A035	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	108	108	108	100
2022A035	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	53	53	53	100
2022A035	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	25	25	25	100
2022A035	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	17	17	17	100
2022A035	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	349	349	349	100
2022A035	001 INFLU. A	016 STATUS FLU		001 001	001 001		001 - NEGATIVE	63	63	63	100
2022A035	002 INFLU. B	001 BINAX NOW		002 002	002 002		002 - POSITIVE	14	14	14	100
2022A035	002 INFLU. B	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	95	95	95	100
2022A035	002 INFLU. B	004 Sekisui Os		002 002	002 002		002 - POSITIVE	108	108	108	100
2022A035	002 INFLU. B	006 Quidel Flu		002 002	002 002		002 - POSITIVE	53	53	53	100
2022A035	002 INFLU. B	007 BD Veritor		002 002	002 002		002 - POSITIVE	25	25	25	100
2022A035	002 INFLU. B	008 Remel x/p		002 002	002 002		002 - POSITIVE	17	17	17	100
2022A035	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	350	350	350	100
2022A035	002 INFLU. B	016 STATUS FLU		002 002	002 002		002 - POSITIVE	62	62	62	100

Participant Summary
MONTH 12 OF YEAR 2022

SPE: 50 MICROBIOLOGY		SUB: 180 Virology				GRP: 450 RSV Syncytial Virus						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
555 Resp Syncytial Virus												
2022A031	001	ALERE ABBO		001	001	001	001	001 - Neg	28	28	28	100
2022A031	002	SOFIA RSV		001	001	001	001	001 - Neg	26	26	26	100
2022A032	001	ALERE ABBO		001	001	001	001	001 - Neg	28	28	28	100
2022A032	002	SOFIA RSV		001	001	001	001	001 - Neg	26	26	26	100
2022A033	001	ALERE ABBO		001	001	001	001	001 - Neg	28	28	28	100
2022A033	002	SOFIA RSV		001	001	001	001	001 - Neg	26	26	26	100
2022A034	001	ALERE ABBO		002	002	002	002	001 - Neg	28	1	27	0
2022A034	001	ALERE ABBO		002	002	002	002	002 - Pos	28	27	27	100
2022A034	002	SOFIA RSV		002	002	002	002	002 - Pos	26	26	26	100
2022A035	001	ALERE ABBO		001	001	001	001	001 - Neg	28	28	28	100
2022A035	002	SOFIA RSV		001	001	001	001	001 - Neg	26	26	26	100