

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
001 ALAN AMINO	ALT/SGPT	CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %										
2020-001	5 ENZ. 37 C	38 SIEMENS DI	143 SIEMENS	14	181.786	4.459	2.45	145.000	218.000				
2020-001	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	15	167.600	9.803	5.85	134.000	201.000				
2020-001	5 ENZ. 37 C	212 S. D. EXL	143 SIEMENS	23	179.591	3.939	2.19	144.000	216.000				
2020-001	8 OX.OF NADH	165 VITROS	63 KODAK	11	160.636	7.678	4.78	129.000	193.000				
2020-001	8 OX.OF NADH	165 VITROS	155 ALTV	19	161.000	4.091	2.54	129.000	193.000				
2020-001	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	168.500	2.327	1.38	135.000	202.000				
2020-001	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	20	179.800	4.411	2.45	144.000	216.000				
2020-001	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	161.900	5.758	3.56	130.000	194.000				
2020-001	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	24	148.217	3.526	2.38	119.000	178.000				
2020-001	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	184.161	11.914	6.47	147.000	221.000				
2020-001	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	11	196.636	19.004	9.66	157.000	236.000				
2020-001	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	163.385	4.412	2.70	131.000	196.000				
2020-001	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	38	182.500	5.684	3.11	146.000	219.000				
2020-001	13 ALA &aKETO	165 VITROS	63 KODAK	10	157.500	3.471	2.20	126.000	189.000				
2020-001	13 ALA &aKETO	165 VITROS	155 ALTV	104	160.010	4.462	2.79	128.000	192.000				
2020-002	5 ENZ. 37 C	38 SIEMENS DI	143 SIEMENS	14	260.500	5.865	2.25	208.000	313.000				
2020-002	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	15	243.600	12.208	5.01	195.000	292.000				
2020-002	5 ENZ. 37 C	212 S. D. EXL	143 SIEMENS	23	256.773	6.755	2.63	205.000	308.000				
2020-002	8 OX.OF NADH	165 VITROS	63 KODAK	11	231.364	8.183	3.54	185.000	278.000				
2020-002	8 OX.OF NADH	165 VITROS	155 ALTV	19	231.211	5.672	2.45	185.000	277.000				
2020-002	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	240.667	3.815	1.59	193.000	289.000				
2020-002	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	20	259.750	7.327	2.82	208.000	312.000				
2020-002	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	235.567	8.139	3.46	188.000	283.000				
2020-002	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	24	215.609	4.489	2.08	172.000	259.000				
2020-002	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	263.129	18.397	6.99	211.000	316.000				
2020-002	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	11	290.182	30.051	10.36	232.000	348.000				
2020-002	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	236.320	5.034	2.13	189.000	284.000				
2020-002	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	38	261.342	8.421	3.22	209.000	314.000				

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-002	13 ALA &aKETO	165 VITROS	63 KODAK	10	227.600	5.970	2.62	182.000	273.000		
2020-002	13 ALA &aKETO	165 VITROS	155 ALTV	104	229.745	6.140	2.67	184.000	276.000		
001 ALAN AMINO ALT/SGPT		CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-003	5 ENZ. 37 C	38 SIEMENS DI	143 SIEMENS	14	79.714	3.304	4.15	64.000	96.000		
2020-003	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	15	71.800	4.750	6.62	57.000	86.000		
2020-003	5 ENZ. 37 C	212 S. D. EXL	143 SIEMENS	23	80.000	2.923	3.65	64.000	96.000		
2020-003	8 OX.OF NADH	165 VITROS	63 KODAK	11	70.364	5.381	7.65	56.000	84.000		
2020-003	8 OX.OF NADH	165 VITROS	155 ALTV	19	70.000	1.806	2.58	56.000	84.000		
2020-003	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	78.583	1.754	2.23	63.000	94.000		
2020-003	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	20	80.350	4.328	5.39	64.000	96.000		
2020-003	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	69.467	2.513	3.62	56.000	83.000		
2020-003	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	24	63.042	2.169	3.44	50.000	76.000		
2020-003	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	79.484	5.786	7.28	64.000	95.000		
2020-003	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	11	81.727	6.180	7.56	65.000	98.000		
2020-003	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	69.120	1.532	2.22	55.000	83.000		
2020-003	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	38	80.395	3.321	4.13	64.000	96.000		
2020-003	13 ALA &aKETO	165 VITROS	63 KODAK	10	68.800	2.441	3.55	55.000	83.000		
2020-003	13 ALA &aKETO	165 VITROS	155 ALTV	104	69.204	2.283	3.30	55.000	83.000		
2020-004	5 ENZ. 37 C	38 SIEMENS DI	143 SIEMENS	14	298.286	7.025	2.36	239.000	358.000		
2020-004	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	15	279.400	13.827	4.95	224.000	335.000		
2020-004	5 ENZ. 37 C	212 S. D. EXL	143 SIEMENS	23	295.227	6.193	2.10	236.000	354.000		
2020-004	8 OX.OF NADH	165 VITROS	63 KODAK	11	265.273	9.659	3.64	212.000	318.000		
2020-004	8 OX.OF NADH	165 VITROS	155 ALTV	19	262.895	7.239	2.75	210.000	315.000		
2020-004	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	274.333	4.110	1.50	219.000	329.000		
2020-004	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	20	297.400	7.749	2.61	238.000	357.000		
2020-004	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	268.069	7.216	2.69	214.000	322.000		
2020-004	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	24	246.375	8.817	3.58	197.000	296.000		
2020-004	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	302.194	19.630	6.50	242.000	363.000		
2020-004	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	11	334.455	38.855	11.62	268.000	401.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-004	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	271.400	5.973	2.20	217.000	326.000		
2020-004	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	38	298.342	10.022	3.36	239.000	358.000		
2020-004	13 ALA &aKETO	165 VITROS	63 KODAK	10	261.100	5.665	2.17	209.000	313.000		
2020-004	13 ALA &aKETO	165 VITROS	155 ALTV	104	263.235	6.833	2.60	211.000	316.000		
001 ALAN AMINO ALT/SGPT		CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-005	5 ENZ. 37 C	38 SIEMENS DI	143 SIEMENS	14	16.643	1.394	8.38	13.000	20.000		
2020-005	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	15	11.600	2.417	20.83	9.000	14.000		
2020-005	5 ENZ. 37 C	212 S. D. EXL	143 SIEMENS	23	16.227	2.194	13.52	13.000	19.000		
2020-005	8 OX.OF NADH	165 VITROS	63 KODAK	11	11.000	6.267	56.97	9.000	13.000		
2020-005	8 OX.OF NADH	165 VITROS	155 ALTV	19	9.056	0.229	2.53	7.000	11.000		
2020-005	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	23.083	1.256	5.44	18.000	28.000		
2020-005	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	20	16.300	2.532	15.53	13.000	20.000		
2020-005	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	10.733	0.964	8.98	9.000	13.000		
2020-005	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	24	10.417	0.954	9.16	8.000	13.000		
2020-005	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	13.935	2.327	16.70	11.000	17.000		
2020-005	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	11	11.455	1.827	15.95	9.000	14.000		
2020-005	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	10.538	0.796	7.55	8.000	13.000		
2020-005	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	38	16.079	2.044	12.71	13.000	19.000		
2020-005	13 ALA &aKETO	165 VITROS	63 KODAK	10	7.900	1.136	14.38	6.000	9.000		
2020-005	13 ALA &aKETO	165 VITROS	155 ALTV	104	8.202	1.121	13.67	7.000	10.000		
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-001	3 DYE BI-B.G	165 VITROS	63 KODAK	143	3.254	0.096	2.96	2.900	3.600		
2020-001	3 DYE BI-B.G	165 VITROS	63 KODAK	143	3.300	0.082	2.47	3.000	3.600		
2020-001	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	36	3.308	0.092	2.79	3.000	3.600		
2020-001	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	24	3.170	0.080	2.54	2.900	3.500		
2020-001	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	3.300	0.093	2.81	3.000	3.600		
2020-001	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	3.330	0.106	3.18	3.000	3.700		
2020-001	3 DYE BI-B.G	194 BIOLIS-24S	149 JAS	10	3.310	0.104	3.15	3.000	3.600		
2020-001	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	18	3.206	0.078	2.43	2.900	3.500		
2020-001	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	15	3.321	0.056	1.68	3.000	3.700		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
	2020-001	3 DYE BI-B.G	203 ROCHE	93 ROCHE	32	3.406	0.086	2.54	3.100	3.700	
	2020-001	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	2.997	0.077	2.57	2.700	3.300	
	2020-001	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	2.950	0.050	1.69	2.700	3.200	
	2020-001	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	60	2.990	0.122	4.07	2.700	3.300	
	2020-001	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	3.155	0.050	1.58	2.800	3.500	
002 ALBUMIN			CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %							
	2020-002	3 DYE BI-B.G	165 VITROS	63 KODAK	143	3.867	0.047	1.22	3.500	4.300	
	2020-002	3 DYE BI-B.G	165 VITROS	63 KODAK	143	4.049	0.114	2.81	3.600	4.500	
	2020-002	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	36	4.020	0.089	2.21	3.600	4.400	
	2020-002	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	24	3.900	0.078	2.00	3.500	4.300	
	2020-002	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	3.958	0.110	2.78	3.600	4.400	
	2020-002	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	3.986	0.099	2.48	3.600	4.400	
	2020-002	3 DYE BI-B.G	194 BIOLIS-24S	149 JAS	10	4.030	0.100	2.49	3.600	4.400	
	2020-002	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	18	3.906	0.064	1.64	3.500	4.300	
	2020-002	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	15	4.007	0.085	2.13	3.600	4.400	
	2020-002	3 DYE BI-B.G	203 ROCHE	93 ROCHE	32	4.132	0.109	2.64	3.700	4.500	
	2020-002	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	3.675	0.043	1.18	3.300	4.000	
	2020-002	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	3.700	0.133	3.58	3.300	4.100	
	2020-002	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	61	3.690	0.131	3.56	3.300	4.100	
	2020-002	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	3.855	0.078	2.03	3.500	4.200	
	2020-003	3 DYE BI-B.G	165 VITROS	63 KODAK	143	2.354	0.064	2.70	2.100	2.600	
	2020-003	3 DYE BI-B.G	165 VITROS	63 KODAK	143	2.433	0.047	1.94	2.200	2.700	
	2020-003	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	36	2.403	0.069	2.86	2.200	2.600	
	2020-003	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	24	2.322	0.051	2.18	2.100	2.600	
	2020-003	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	2.357	0.090	3.83	2.100	2.600	
	2020-003	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	2.476	0.085	3.45	2.200	2.700	
	2020-003	3 DYE BI-B.G	194 BIOLIS-24S	149 JAS	10	2.450	0.067	2.74	2.200	2.700	
	2020-003	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	18	2.339	0.068	2.90	2.100	2.600	
	2020-003	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	15	2.473	0.068	2.75	2.200	2.700	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-003	3 DYE BI-B.G	203 ROCHE	93 ROCHE	32	2.497	0.098	3.94	2.200	2.700		
2020-003	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	2.106	0.056	2.64	1.900	2.300		
2020-003	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	2.125	0.043	2.04	1.900	2.300		
2020-003	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	60	2.122	0.080	3.79	1.900	2.300		
2020-003	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	2.236	0.048	2.15	2.000	2.500		
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-004	3 DYE BI-B.G	165 VITROS	63 KODAK	143	4.233	0.094	2.23	3.800	4.700		
2020-004	3 DYE BI-B.G	165 VITROS	63 KODAK	143	4.445	0.145	3.25	4.000	4.900		
2020-004	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	36	4.340	0.113	2.60	3.900	4.800		
2020-004	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	24	4.221	0.141	3.35	3.800	4.600		
2020-004	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	4.264	0.125	2.93	3.800	4.700		
2020-004	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	4.271	0.148	3.48	3.800	4.700		
2020-004	3 DYE BI-B.G	194 BIOLIS-24S	149 JAS	10	4.360	0.080	1.83	3.900	4.800		
2020-004	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	18	4.241	0.069	1.63	3.800	4.700		
2020-004	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	15	4.280	0.117	2.72	3.900	4.700		
2020-004	3 DYE BI-B.G	203 ROCHE	93 ROCHE	32	4.478	0.119	2.66	4.000	4.900		
2020-004	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	3.950	0.050	1.27	3.600	4.300		
2020-004	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	33	4.038	0.114	2.82	3.600	4.400		
2020-004	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	60	4.028	0.116	2.87	3.600	4.400		
2020-004	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	4.200	0.074	1.76	3.800	4.600		
2020-005	3 DYE BI-B.G	165 VITROS	63 KODAK	143	1.714	0.050	2.92	1.500	1.900		
2020-005	3 DYE BI-B.G	165 VITROS	63 KODAK	143	1.833	0.047	2.57	1.700	2.000		
2020-005	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	36	1.753	0.060	3.42	1.600	1.900		
2020-005	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	24	1.663	0.063	3.81	1.500	1.800		
2020-005	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	1.771	0.070	3.95	1.600	1.900		
2020-005	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	34	1.827	0.057	3.09	1.600	2.000		
2020-005	3 DYE BI-B.G	194 BIOLIS-24S	149 JAS	10	1.830	0.046	2.50	1.600	2.000		
2020-005	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	18	1.711	0.066	3.84	1.500	1.900		
2020-005	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	15	1.820	0.065	3.59	1.600	2.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-005	3 DYE BI-B.G	203	ROCHE	93	ROCHE	32	1.816	0.079	4.38	1.600	2.000	
2020-005	4 DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	33	1.503	0.047	3.10	1.400	1.700	
2020-005	4 DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	33	1.525	0.043	2.84	1.400	1.700	
2020-005	4 DYE BI-B.P	212	S. D. EXL	143	SIEMENS	60	1.505	0.059	3.95	1.400	1.700	
2020-005	4 DYE BI-B.P	215	B.C DXC	134	BECKM.COUL	11	1.609	0.067	4.15	1.400	1.800	
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275		EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-001	3 COLORIMET.	184	INTEGRA	93	ROCHE	30	126.414	3.926	3.11	88.000	164.000	
2020-001	3 COLORIMET.	203	ROCHE	93	ROCHE	22	123.955	4.685	3.78	87.000	161.000	
2020-001	6 KINETIC 37	38	SIEMENS DI	143	SIEMENS	15	137.267	7.009	5.11	96.000	178.000	
2020-001	6 KINETIC 37	165	VITROS	63	KODAK	143	135.106	7.907	5.85	95.000	176.000	
2020-001	6 KINETIC 37	212	S. D. EXL	143	SIEMENS	25	133.667	6.829	5.11	94.000	174.000	
2020-001	6 KINETIC 37	212	S. D. EXL	143	SIEMENS	25	136.500	4.975	3.64	96.000	177.000	
2020-001	6 KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	109.083	8.450	7.75	76.000	142.000	
2020-001	8 BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	25	113.240	7.951	7.02	79.000	147.000	
2020-001	9 ENZYMAT.37	191	RX-DAYTONA	126	RANDOX	27	119.815	12.323	10.28	84.000	156.000	
2020-001	11 Modified I	38	SIEMENS DI	143	SIEMENS	17	133.471	5.019	3.76	93.000	174.000	
2020-001	11 Modified I	212	S. D. EXL	143	SIEMENS	31	135.433	4.500	3.32	95.000	176.000	
2020-002	3 COLORIMET.	184	INTEGRA	93	ROCHE	30	180.724	5.212	2.88	127.000	235.000	
2020-002	3 COLORIMET.	203	ROCHE	93	ROCHE	22	177.364	5.749	3.24	124.000	231.000	
2020-002	6 KINETIC 37	38	SIEMENS DI	143	SIEMENS	15	198.067	8.590	4.34	139.000	257.000	
2020-002	6 KINETIC 37	165	VITROS	63	KODAK	143	171.282	9.987	5.83	120.000	223.000	
2020-002	6 KINETIC 37	212	S. D. EXL	143	SIEMENS	25	193.500	8.372	4.33	135.000	252.000	
2020-002	6 KINETIC 37	212	S. D. EXL	143	SIEMENS	25	198.000	6.782	3.43	139.000	257.000	
2020-002	6 KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	155.083	12.366	7.97	109.000	202.000	
2020-002	8 BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	25	162.120	11.378	7.02	113.000	211.000	
2020-002	9 ENZYMAT.37	191	RX-DAYTONA	126	RANDOX	27	170.444	15.569	9.13	119.000	222.000	
2020-002	11 Modified I	38	SIEMENS DI	143	SIEMENS	17	193.000	8.182	4.24	135.000	251.000	
2020-002	11 Modified I	212	S. D. EXL	143	SIEMENS	31	194.100	5.211	2.68	136.000	252.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
003	ALKALINE PHOSPHATE	CLIA TEST ID #:	275	EVALUATION CRITERIA: TARGET VALUE 30 %									
2020-003	3	COLORIMET.	184 INTEGRA	93	ROCHE	30	59.621	1.690	2.83	42.000	78.000		
2020-003	3	COLORIMET.	203 ROCHE	93	ROCHE	22	58.318	2.991	5.13	41.000	76.000		
2020-003	6	KINETIC 37	38 SIEMENS DI	143	SIEMENS	15	65.267	4.328	6.63	46.000	85.000		
2020-003	6	KINETIC 37	165 VITROS	63	KODAK	143	73.655	4.592	6.23	52.000	96.000		
2020-003	6	KINETIC 37	212 S. D. EXL	143	SIEMENS	25	62.708	4.315	6.88	44.000	82.000		
2020-003	6	KINETIC 37	212 S. D. EXL	143	SIEMENS	25	61.250	4.815	7.86	43.000	80.000		
2020-003	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	12	50.833	4.259	8.38	36.000	66.000		
2020-003	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	25	52.920	3.621	6.84	37.000	69.000		
2020-003	9	ENZYMAT.37	191 RX-DAYTONA	126	RANDOX	27	57.481	6.262	10.89	40.000	75.000		
2020-003	11	Modified I	38 SIEMENS DI	143	SIEMENS	17	61.412	4.679	7.62	43.000	80.000		
2020-003	11	Modified I	212 S. D. EXL	143	SIEMENS	31	62.900	3.700	5.88	44.000	82.000		
2020-004	3	COLORIMET.	184 INTEGRA	93	ROCHE	30	207.483	6.595	3.18	145.000	270.000		
2020-004	3	COLORIMET.	203 ROCHE	93	ROCHE	22	204.227	6.640	3.25	143.000	265.000		
2020-004	6	KINETIC 37	38 SIEMENS DI	143	SIEMENS	15	227.267	10.963	4.82	159.000	295.000		
2020-004	6	KINETIC 37	165 VITROS	63	KODAK	143	185.401	10.256	5.53	130.000	241.000		
2020-004	6	KINETIC 37	212 S. D. EXL	143	SIEMENS	25	223.120	11.283	5.06	156.000	290.000		
2020-004	6	KINETIC 37	212 S. D. EXL	143	SIEMENS	25	228.000	5.477	2.40	160.000	296.000		
2020-004	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	12	180.000	12.213	6.79	126.000	234.000		
2020-004	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	25	187.720	14.858	7.92	131.000	244.000		
2020-004	9	ENZYMAT.37	191 RX-DAYTONA	126	RANDOX	27	193.222	17.933	9.28	135.000	251.000		
2020-004	11	Modified I	38 SIEMENS DI	143	SIEMENS	17	223.059	6.169	2.77	156.000	290.000		
2020-004	11	Modified I	212 S. D. EXL	143	SIEMENS	31	222.967	5.683	2.55	156.000	290.000		
2020-005	3	COLORIMET.	184 INTEGRA	93	ROCHE	30	17.867	0.921	5.16	13.000	23.000		
2020-005	3	COLORIMET.	203 ROCHE	93	ROCHE	22	17.136	1.179	6.88	12.000	22.000		
2020-005	6	KINETIC 37	38 SIEMENS DI	143	SIEMENS	15	19.600	3.116	15.90	14.000	25.000		
2020-005	6	KINETIC 37	165 VITROS	63	KODAK	143	29.106	2.349	8.07	20.000	38.000		
2020-005	6	KINETIC 37	212 S. D. EXL	143	SIEMENS	25	18.417	2.929	15.90	13.000	24.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-005	6 KINETIC 37	212 S. D. EXL	143 SIEMENS	25	20.500	3.354	16.36	14.000	27.000		
2020-005	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	14.917	1.977	13.26	10.000	19.000		
2020-005	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	25	15.125	1.333	8.81	11.000	20.000		
2020-005	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	27	16.778	2.455	14.63	12.000	22.000		
2020-005	11 Modified I	38 SIEMENS DI	143 SIEMENS	17	16.882	2.423	14.35	12.000	22.000		
2020-005	11 Modified I	212 S. D. EXL	143 SIEMENS	31	18.032	3.032	16.81	13.000	23.000		
004 AMYLASE		CLIA TEST ID #: 285	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-001	3 ENZYM. 37C	165 VITROS	63 KODAK	81	94.580	6.071	6.42	66.000	123.000		
2020-001	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	130.250	2.586	1.99	91.000	169.000		
2020-001	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	128.500	3.202	2.49	90.000	167.000		
2020-001	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	127.750	3.080	2.41	89.000	166.000		
2020-001	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	11	124.636	3.255	2.61	87.000	162.000		
2020-001	9 ENZ/COL.37	203 ROCHE	93 ROCHE	22	120.409	1.923	1.60	84.000	157.000		
2020-001	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	99.765	6.804	6.82	70.000	130.000		
2020-002	3 ENZYM. 37C	165 VITROS	63 KODAK	81	137.963	8.035	5.82	97.000	179.000		
2020-002	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	184.833	3.578	1.94	129.000	240.000		
2020-002	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	180.150	3.811	2.12	126.000	234.000		
2020-002	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	181.125	4.675	2.58	127.000	235.000		
2020-002	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	11	175.727	4.308	2.45	123.000	228.000		
2020-002	9 ENZ/COL.37	203 ROCHE	93 ROCHE	22	171.455	2.388	1.39	120.000	223.000		
2020-002	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	141.706	9.529	6.72	99.000	184.000		
2020-003	3 ENZYM. 37C	165 VITROS	63 KODAK	81	40.802	5.516	13.52	29.000	53.000		
2020-003	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	65.500	1.500	2.29	46.000	85.000		
2020-003	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	62.375	1.932	3.10	44.000	81.000		
2020-003	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	62.200	1.720	2.77	44.000	81.000		
2020-003	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	11	62.727	1.710	2.73	44.000	82.000		
2020-003	9 ENZ/COL.37	203 ROCHE	93 ROCHE	22	60.591	1.114	1.84	42.000	79.000		
2020-003	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	48.353	3.693	7.64	34.000	63.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
004	AMYLASE	CLIA TEST ID #: 285		EVALUATION CRITERIA: TARGET VALUE 30 %							
2020-004	3 ENZYM. 37C	165 VITROS	63 KODAK	81	156.815	8.177	5.21	110.000	204.000		
2020-004	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	210.333	3.009	1.43	147.000	273.000		
2020-004	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	205.550	4.914	2.39	144.000	267.000		
2020-004	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	208.625	7.105	3.41	146.000	271.000		
2020-004	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	11	201.000	7.839	3.90	141.000	261.000		
2020-004	9 ENZ/COL.37	203 ROCHE	93 ROCHE	22	194.682	2.960	1.52	136.000	253.000		
2020-004	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	161.588	10.944	6.77	113.000	210.000		
2020-005	3 ENZYM. 37C	165 VITROS	63 KODAK	81	30.000	0.000	0.00	21.000	39.000		
2020-005	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	23.667	0.850	3.59	17.000	31.000		
2020-005	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	21.750	1.090	5.01	15.000	28.000		
2020-005	4 HYDRO ACT.	212 S. D. EXL	143 SIEMENS	21	21.300	0.954	4.48	15.000	28.000		
2020-005	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	11	23.727	0.617	2.60	17.000	31.000		
2020-005	9 ENZ/COL.37	203 ROCHE	93 ROCHE	22	22.476	0.499	2.22	16.000	29.000		
2020-005	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	16.471	1.334	8.10	12.000	21.000		
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-001	1 ASPAR&KETO	165 VITROS	63 KODAK	136	227.074	7.408	3.26	182.000	272.000		
2020-001	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	12	202.667	9.186	4.53	162.000	243.000		
2020-001	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	33	193.848	6.416	3.31	155.000	233.000		
2020-001	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	193.879	6.052	3.12	155.000	233.000		
2020-001	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	210.667	10.339	4.91	169.000	253.000		
2020-001	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	188.545	2.903	1.54	151.000	226.000		
2020-001	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	196.194	5.733	2.92	157.000	235.000		
2020-001	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	170.080	7.293	4.29	136.000	204.000		
2020-001	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	210.645	13.540	6.43	169.000	253.000		
2020-001	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	200.800	14.274	7.11	161.000	241.000		
2020-001	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	221.900	18.539	8.35	178.000	266.000		
2020-001	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	217.000	1.633	0.75	174.000	260.000		
2020-001	11 MOD. IFCC	203 ROCHE	93 ROCHE	24	196.792	5.874	2.98	157.000	236.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
005 ASP. AMINO	AST/SGOT	CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE	20	%						
2020-002	1 ASPAR&KETO	165 VITROS	63 KODAK	136	341.200	9.730	2.85	273.000	409.000		
2020-002	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	12	284.833	15.916	5.59	228.000	342.000		
2020-002	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	33	276.576	8.224	2.97	221.000	332.000		
2020-002	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	293.000	24.913	8.50	234.000	352.000		
2020-002	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	274.224	8.052	2.94	219.000	329.000		
2020-002	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	265.636	5.104	1.92	213.000	319.000		
2020-002	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	277.258	7.849	2.83	222.000	333.000		
2020-002	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	241.280	10.208	4.23	193.000	290.000		
2020-002	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	297.226	20.019	6.74	238.000	357.000		
2020-002	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	290.600	13.735	4.73	232.000	349.000		
2020-002	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	305.333	4.028	1.32	244.000	366.000		
2020-002	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	318.100	27.223	8.56	254.000	382.000		
2020-002	11 MOD. IFCC	203 ROCHE	93 ROCHE	24	280.250	8.922	3.18	224.000	336.000		
2020-003	1 ASPAR&KETO	165 VITROS	63 KODAK	136	102.163	3.158	3.09	82.000	123.000		
2020-003	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	12	95.583	4.232	4.43	76.000	115.000		
2020-003	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	33	93.152	4.016	4.31	75.000	112.000		
2020-003	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	92.931	2.852	3.07	74.000	112.000		
2020-003	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	100.333	6.848	6.82	80.000	120.000		
2020-003	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	93.636	1.611	1.72	75.000	112.000		
2020-003	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	94.548	2.722	2.88	76.000	113.000		
2020-003	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	82.280	4.075	4.95	66.000	99.000		
2020-003	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	101.613	7.486	7.37	81.000	122.000		
2020-003	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	103.000	7.563	7.34	82.000	124.000		
2020-003	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	104.900	8.549	8.15	84.000	126.000		
2020-003	11 MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	106.000	1.633	1.54	85.000	127.000		
2020-003	11 MOD. IFCC	203 ROCHE	93 ROCHE	24	94.500	3.149	3.33	76.000	113.000		
2020-004	1 ASPAR&KETO	165 VITROS	63 KODAK	136	400.793	12.872	3.21	321.000	481.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-004	1	ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	12	324.083	19.172	5.92	259.000	389.000		
2020-004	5	ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	33	316.364	11.140	3.52	253.000	380.000		
2020-004	5	ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	339.333	12.284	3.62	271.000	407.000		
2020-004	5	ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	315.276	9.314	2.95	252.000	378.000		
2020-004	10	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	302.091	4.316	1.43	242.000	363.000		
2020-004	11	MOD. IFCC	184 INTEGRA	93 ROCHE	31	314.903	11.139	3.54	252.000	378.000		
2020-004	11	MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	274.680	11.811	4.30	220.000	330.000		
2020-004	11	MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	337.548	24.311	7.20	270.000	405.000		
2020-004	11	MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	275.000	121.970	44.35	220.000	330.000		
2020-004	11	MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	347.000	3.266	0.94	278.000	416.000		
2020-004	11	MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	363.200	35.735	9.84	291.000	436.000		
2020-004	11	MOD. IFCC	203 ROCHE	93 ROCHE	24	317.250	11.914	3.76	254.000	381.000		
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE 20 %									
2020-005	1	ASPAR&KETO	165 VITROS	63 KODAK	136	33.843	0.953	2.82	27.000	41.000		
2020-005	1	ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	12	28.083	2.139	7.62	22.000	34.000		
2020-005	5	ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	33	31.394	1.705	5.43	25.000	38.000		
2020-005	5	ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	31.052	1.766	5.69	25.000	37.000		
2020-005	5	ENZYMA.37C	212 S. D. EXL	143 SIEMENS	59	32.333	2.055	6.36	26.000	39.000		
2020-005	10	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	34.455	1.233	3.58	28.000	41.000		
2020-005	11	MOD. IFCC	184 INTEGRA	93 ROCHE	31	30.129	1.338	4.44	24.000	36.000		
2020-005	11	MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	26.560	1.359	5.12	21.000	32.000		
2020-005	11	MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	33.452	2.563	7.66	27.000	40.000		
2020-005	11	MOD. IFCC	191 RX-DAYTONA	126 RANDOX	31	90.200	110.424	122.42	72.000	108.000		
2020-005	11	MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	32.500	2.012	6.19	26.000	39.000		
2020-005	11	MOD. IFCC	199 Rx-IMOLA	126 RANDOX	10	37.667	3.771	10.01	30.000	45.000		
2020-005	11	MOD. IFCC	203 ROCHE	93 ROCHE	24	29.625	1.111	3.75	24.000	36.000		
006	BILIRUBIN, TOTAL	CLIA TEST ID #: 305										
2020-001	1	BICHR.NO B	194 BIOLIS-24S	149 JAS	10	1.913	0.206	10.77	1.530	2.300		
2020-001	3	E-M NO BK.	184 INTEGRA	93 ROCHE	11	1.978	0.039	1.95	1.580	2.370		
2020-001	4	J-G WITH B	38 SIEMENS DI	143 SIEMENS	32	2.227	0.083	3.73	1.780	2.670		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-001	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	22	2.369	0.094	3.95	1.900	2.840
2020-001	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	2.437	0.161	6.61	1.950	2.920
2020-001	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	52	2.264	0.096	4.26	1.810	2.720
2020-001	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	2.547	0.166	6.51	2.040	3.060
2020-001	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	2.308	0.171	7.39	1.850	2.770
2020-001	7	DIAZO-DYPH	165	VITROS	63	KODAK	140	2.276	0.142	6.24	1.820	2.730
2020-001	13	AZOBIL.	184	INTEGRA	93	ROCHE	17	2.061	0.084	4.06	1.650	2.470
2020-001	13	AZOBIL.	203	ROCHE	93	ROCHE	22	1.979	0.079	3.97	1.580	2.370
2020-001	18	COLORIMETR	203	ROCHE	93	ROCHE	10	1.972	0.048	2.44	1.580	2.370
006 BILIRUBIN, TOTAL			CLIA TEST ID #: 305			EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-002	1	BICHR.NO B	194	BIOLIS-24S	149	JAS	10	2.815	0.301	10.68	2.250	3.380
2020-002	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	3.004	0.086	2.86	2.400	3.600
2020-002	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	32	3.355	0.139	4.14	2.680	4.030
2020-002	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	22	3.478	0.148	4.24	2.780	4.170
2020-002	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	3.636	0.274	7.54	2.910	4.360
2020-002	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	52	3.395	0.142	4.19	2.720	4.070
2020-002	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	3.412	0.293	8.58	2.730	4.090
2020-002	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	3.809	0.171	4.49	3.050	4.570
2020-002	7	DIAZO-DYPH	165	VITROS	63	KODAK	140	3.374	0.177	5.26	2.700	4.050
2020-002	13	AZOBIL.	184	INTEGRA	93	ROCHE	17	3.085	0.135	4.37	2.470	3.700
2020-002	13	AZOBIL.	203	ROCHE	93	ROCHE	22	2.988	0.070	2.35	2.390	3.590
2020-002	18	COLORIMETR	203	ROCHE	93	ROCHE	10	2.987	0.063	2.12	2.390	3.580
2020-003	1	BICHR.NO B	194	BIOLIS-24S	149	JAS	10	0.802	0.081	10.07	0.640	0.960
2020-003	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	0.758	0.029	3.77	0.610	0.910
2020-003	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	32	0.851	0.040	4.66	0.680	1.020
2020-003	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	22	0.949	0.062	6.50	0.760	1.140
2020-003	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	0.944	0.098	10.41	0.760	1.130
2020-003	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	52	0.863	0.059	6.85	0.690	1.040
2020-003	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	1.124	0.169	15.07	0.900	1.350

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	0.914	0.077	8.41	0.730	1.100
2020-003	7	DIAZO-DYPH	165	VITROS	63	KODAK	140	0.870	0.088	10.14	0.700	1.040
2020-003	13	AZOBIL.	184	INTEGRA	93	ROCHE	17	0.776	0.035	4.57	0.620	0.930
2020-003	13	AZOBIL.	203	ROCHE	93	ROCHE	22	0.755	0.035	4.58	0.600	0.910
2020-003	18	COLORIMETR	203	ROCHE	93	ROCHE	10	0.750	0.025	3.32	0.600	0.900
006 BILIRUBIN, TOTAL			CLIA TEST ID #: 305			EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-004	1	BICHR.NO B	194	BIOLIS-24S	149	JAS	10	3.317	0.358	10.78	2.650	3.980
2020-004	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	3.498	0.078	2.24	2.800	4.200
2020-004	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	32	3.918	0.128	3.26	3.130	4.700
2020-004	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	22	4.054	0.172	4.25	3.240	4.860
2020-004	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	4.273	0.293	6.86	3.420	5.130
2020-004	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	52	4.005	0.165	4.11	3.200	4.810
2020-004	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	4.046	0.340	8.39	3.240	4.860
2020-004	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	4.429	0.195	4.40	3.540	5.320
2020-004	7	DIAZO-DYPH	165	VITROS	63	KODAK	140	3.944	0.218	5.52	3.160	4.730
2020-004	13	AZOBIL.	184	INTEGRA	93	ROCHE	17	3.610	0.149	4.12	2.890	4.330
2020-004	13	AZOBIL.	203	ROCHE	93	ROCHE	22	3.514	0.106	3.02	2.810	4.220
2020-004	18	COLORIMETR	203	ROCHE	93	ROCHE	10	3.491	0.058	1.66	2.790	4.190
2020-005	1	BICHR.NO B	194	BIOLIS-24S	149	JAS	10	0.112	0.031	27.61	0.090	0.130
2020-005	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	0.041	0.049	119.96	0.030	0.050
2020-005	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	32	0.092	0.021	22.49	0.070	0.110
2020-005	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	22	0.098	0.010	9.72	0.080	0.120
2020-005	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	0.121	0.042	35.02	0.100	0.150
2020-005	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	52	0.084	0.042	50.06	0.070	0.100
2020-005	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	0.251	0.088	35.19	0.200	0.300
2020-005	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	0.090	0.031	34.43	0.070	0.110
2020-005	7	DIAZO-DYPH	165	VITROS	63	KODAK	140	0.108	0.023	21.40	0.090	0.130
2020-005	13	AZOBIL.	184	INTEGRA	93	ROCHE	17	0.054	0.049	89.85	0.040	0.070
2020-005	13	AZOBIL.	203	ROCHE	93	ROCHE	22	0.129	0.045	34.48	0.100	0.150

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-005	18	COLORIMETR	203 ROCHE	93	ROCHE	10	0.145	0.015	10.34	0.120	0.170
010	CALCIUM, TOTAL		CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE	1.00					
2020-001	3	ARSEN.-III	165 VITROS	63	KODAK	158	8.815	0.156	1.77	7.800	9.800
2020-001	3	ARSEN.-III	186 BECK. COUL	76	AU-BECKMAN	27	8.707	0.249	2.86	7.700	9.700
2020-001	3	ARSEN.-III	191 RX-DAYTONA	126	RANDOX	36	9.158	0.295	3.22	8.200	10.200
2020-001	3	ARSEN.-III	194 BIOLIS-24S	149	JAS	11	8.891	0.315	3.54	7.900	9.900
2020-001	3	ARSEN.-III	197 ARCHITECT	1	ABBOTT	14	8.893	0.243	2.74	7.900	9.900
2020-001	3	ARSEN.-III	199 Rx-IMOLA	126	RANDOX	15	9.107	0.232	2.55	8.100	10.100
2020-001	8	CRESOLPHT.	38 SIEMENS DI	143	SIEMENS	33	8.634	0.174	2.02	7.600	9.600
2020-001	8	CRESOLPHT.	184 INTEGRA	93	ROCHE	18	8.878	0.272	3.06	7.900	9.900
2020-001	8	CRESOLPHT.	203 ROCHE	93	ROCHE	10	8.950	0.169	1.89	8.000	10.000
2020-001	8	CRESOLPHT.	212 S. D. EXL	143	SIEMENS	61	8.605	0.297	3.45	7.600	9.600
2020-001	14	ELECTRODE	215 B.C DXC	134	BECKM.COUL	12	8.600	0.178	2.07	7.600	9.600
2020-001	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	8.841	0.240	2.72	7.800	9.800
2020-001	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	9.380	0.426	4.54	8.400	10.400
2020-001	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	8.925	0.151	1.69	7.900	9.900
2020-001	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	8.722	0.266	3.05	7.700	9.700
2020-002	3	ARSEN.-III	165 VITROS	63	KODAK	158	10.244	0.176	1.72	9.200	11.200
2020-002	3	ARSEN.-III	186 BECK. COUL	76	AU-BECKMAN	27	10.319	0.217	2.10	9.300	11.300
2020-002	3	ARSEN.-III	191 RX-DAYTONA	126	RANDOX	36	10.572	0.355	3.36	9.600	11.600
2020-002	3	ARSEN.-III	194 BIOLIS-24S	149	JAS	11	10.455	0.306	2.92	9.500	11.500
2020-002	3	ARSEN.-III	197 ARCHITECT	1	ABBOTT	14	10.407	0.345	3.32	9.400	11.400
2020-002	3	ARSEN.-III	199 Rx-IMOLA	126	RANDOX	15	10.433	0.267	2.56	9.400	11.400
2020-002	8	CRESOLPHT.	38 SIEMENS DI	143	SIEMENS	33	9.997	0.205	2.05	9.000	11.000
2020-002	8	CRESOLPHT.	184 INTEGRA	93	ROCHE	18	10.372	0.318	3.06	9.400	11.400
2020-002	8	CRESOLPHT.	203 ROCHE	93	ROCHE	10	10.450	0.254	2.43	9.500	11.500
2020-002	8	CRESOLPHT.	212 S. D. EXL	143	SIEMENS	61	9.975	0.274	2.74	9.000	11.000
2020-002	14	ELECTRODE	215 B.C DXC	134	BECKM.COUL	12	9.933	0.189	1.90	8.900	10.900
2020-002	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	10.341	0.297	2.87	9.300	11.300

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-002	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	10.660	0.512	4.81	9.700	11.700			
2020-002	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	21	10.178	0.297	2.92	9.200	11.200			
2020-002	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	21	10.355	0.153	1.48	9.400	11.400			
010	CALCIUM, TOTAL	CLIA TEST ID #: 345	EVALUATION CRITERIA: TARGET VALUE	1.00								
2020-003	3 ARSEN.-III	165 VITROS	63 KODAK	158	7.110	0.140	1.96	6.100	8.100			
2020-003	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	27	7.044	0.191	2.71	6.000	8.000			
2020-003	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX	36	7.514	0.266	3.54	6.500	8.500			
2020-003	3 ARSEN.-III	194 BIOLIS-24S	149 JAS	11	7.173	0.181	2.53	6.200	8.200			
2020-003	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	14	7.157	0.195	2.73	6.200	8.200			
2020-003	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX	15	7.440	0.487	6.54	6.400	8.400			
2020-003	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	33	7.045	0.176	2.50	6.000	8.000			
2020-003	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	18	7.212	0.187	2.60	6.200	8.200			
2020-003	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	7.290	0.239	3.27	6.300	8.300			
2020-003	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	61	7.007	0.259	3.69	6.000	8.000			
2020-003	14 ELECTRODE	215 B.C DXC	134 BECKM.COUL	12	7.075	0.142	2.01	6.100	8.100			
2020-003	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	7.147	0.161	2.26	6.100	8.100			
2020-003	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	7.660	0.356	4.64	6.700	8.700			
2020-003	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	21	7.255	0.124	1.71	6.300	8.300			
2020-003	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	21	6.944	0.302	4.35	5.900	7.900			
2020-004	3 ARSEN.-III	165 VITROS	63 KODAK	158	10.879	0.202	1.86	9.900	11.900			
2020-004	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	27	11.048	0.335	3.03	10.000	12.000			
2020-004	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX	36	11.250	0.372	3.30	10.300	12.300			
2020-004	3 ARSEN.-III	194 BIOLIS-24S	149 JAS	11	11.155	0.271	2.43	10.200	12.200			
2020-004	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	14	11.157	0.394	3.53	10.200	12.200			
2020-004	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX	15	11.100	0.350	3.16	10.100	12.100			
2020-004	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	33	10.715	0.273	2.55	9.700	11.700			
2020-004	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	18	10.944	0.267	2.44	9.900	11.900			
2020-004	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	11.120	0.223	2.00	10.100	12.100			
2020-004	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	61	10.702	0.334	3.12	9.700	11.700			

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-004	14	ELECTRODE	215 B.C DXC	134	BECKM.COUL	12	10.575	0.196	1.86	9.600	11.600	
2020-004	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	11.006	0.267	2.42	10.000	12.000	
2020-004	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	10.300	1.788	17.36	9.300	11.300	
2020-004	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	10.767	0.302	2.80	9.800	11.800	
2020-004	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	11.025	0.187	1.69	10.000	12.000	
010	CALCIUM, TOTAL		CLIA TEST ID #: 345	EVALUATION CRITERIA: TARGET VALUE 1.00								
2020-005	3	ARSEN.-III	165 VITROS	63	KODAK	158	5.749	0.126	2.19	4.700	6.700	
2020-005	3	ARSEN.-III	186 BECK. COUL	76	AU-BECKMAN	27	5.707	0.149	2.61	4.700	6.700	
2020-005	3	ARSEN.-III	191 RX-DAYTONA	126	RANDOX	36	6.133	0.231	3.77	5.100	7.100	
2020-005	3	ARSEN.-III	194 BIOLIS-24S	149	JAS	11	5.782	0.164	2.84	4.800	6.800	
2020-005	3	ARSEN.-III	197 ARCHITECT	1	ABBOTT	14	5.829	0.179	3.07	4.800	6.800	
2020-005	3	ARSEN.-III	199 Rx-IMOLA	126	RANDOX	15	6.053	0.216	3.56	5.100	7.100	
2020-005	8	CRESOLPHT.	38 SIEMENS DI	143	SIEMENS	33	5.824	0.216	3.71	4.800	6.800	
2020-005	8	CRESOLPHT.	184 INTEGRA	93	ROCHE	18	5.872	0.199	3.40	4.900	6.900	
2020-005	8	CRESOLPHT.	203 ROCHE	93	ROCHE	10	5.930	0.200	3.38	4.900	6.900	
2020-005	8	CRESOLPHT.	212 S. D. EXL	143	SIEMENS	61	5.820	0.314	5.40	4.800	6.800	
2020-005	14	ELECTRODE	215 B.C DXC	134	BECKM.COUL	12	5.875	0.153	2.61	4.900	6.900	
2020-005	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	5.835	0.157	2.69	4.800	6.800	
2020-005	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	7.320	2.304	31.48	6.300	8.300	
2020-005	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	5.957	0.159	2.67	5.000	7.000	
2020-005	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	21	5.456	0.236	4.33	4.500	6.500	
011	CHLORIDE		CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2020-001	3	ISE DILUTE	184 INTEGRA	93	ROCHE	38	99.763	1.677	1.68	95.000	105.000	
2020-001	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	27	97.630	2.696	2.76	93.000	103.000	
2020-001	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	16	100.133	1.204	1.20	95.000	105.000	
2020-001	3	ISE DILUTE	203 ROCHE	93	ROCHE	34	95.471	2.003	2.10	91.000	100.000	
2020-001	3	ISE DILUTE	215 B.C DXC	134	BECKM.COUL	12	98.833	2.303	2.33	94.000	104.000	
2020-001	4	ISE UND-DI	165 VITROS	63	KODAK	156	99.155	1.512	1.52	94.000	104.000	
2020-001	4	ISE UND-DI	191 RX-DAYTONA	126	RANDOX	37	99.270	1.982	2.00	94.000	104.000	
2020-001	4	ISE UND-DI	194 BIOLIS-24S	149	JAS	14	99.857	1.552	1.55	95.000	105.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-001	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	16	97.875	1.932	1.97	93.000	103.000		
2020-001	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	99.677	1.489	1.49	95.000	105.000		
2020-001	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	49	99.625	1.423	1.43	95.000	105.000		
011 CHLORIDE		CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2020-002	3 ISE DILUTE	184 INTEGRA	93 ROCHE	38	106.368	1.979	1.86	101.000	112.000		
2020-002	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	27	103.154	1.915	1.86	98.000	108.000		
2020-002	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	16	104.933	1.062	1.01	100.000	110.000		
2020-002	3 ISE DILUTE	203 ROCHE	93 ROCHE	34	101.882	1.843	1.81	97.000	107.000		
2020-002	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	104.500	2.141	2.05	99.000	110.000		
2020-002	4 ISE UND-DI	165 VITROS	63 KODAK	156	104.948	1.703	1.62	100.000	110.000		
2020-002	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	37	105.514	2.332	2.21	100.000	111.000		
2020-002	4 ISE UND-DI	194 BIOLIS-24S	149 JAS	14	107.000	1.363	1.27	102.000	112.000		
2020-002	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	16	105.250	2.704	2.57	100.000	111.000		
2020-002	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	107.032	1.513	1.41	102.000	112.000		
2020-002	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	49	106.500	1.646	1.55	101.000	112.000		
2020-003	3 ISE DILUTE	184 INTEGRA	93 ROCHE	38	93.632	1.708	1.82	89.000	98.000		
2020-003	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	27	93.074	2.053	2.21	88.000	98.000		
2020-003	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	16	95.133	1.147	1.21	90.000	100.000		
2020-003	3 ISE DILUTE	203 ROCHE	93 ROCHE	34	90.353	2.027	2.24	86.000	95.000		
2020-003	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	94.000	1.633	1.74	89.000	99.000		
2020-003	4 ISE UND-DI	165 VITROS	63 KODAK	156	94.162	1.522	1.62	89.000	99.000		
2020-003	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	37	92.861	2.110	2.27	88.000	98.000		
2020-003	4 ISE UND-DI	194 BIOLIS-24S	149 JAS	14	93.643	1.674	1.79	89.000	98.000		
2020-003	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	16	91.875	2.204	2.40	87.000	96.000		
2020-003	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	92.806	1.731	1.86	88.000	97.000		
2020-003	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	49	92.771	1.649	1.78	88.000	97.000		
2020-004	3 ISE DILUTE	184 INTEGRA	93 ROCHE	38	109.184	2.338	2.14	104.000	115.000		
2020-004	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	27	105.654	2.449	2.32	100.000	111.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-004	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	16	107.000	1.506	1.41	102.000	112.000
2020-004	3	ISE DILUTE	203	ROCHE	93	ROCHE	34	104.727	1.958	1.87	99.000	110.000
2020-004	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	106.917	2.361	2.21	102.000	112.000
2020-004	4	ISE UND-DI	165	VITROS	63	KODAK	156	107.477	2.141	1.99	102.000	113.000
2020-004	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	37	108.351	2.327	2.15	103.000	114.000
2020-004	4	ISE UND-DI	194	BIOLIS-24S	149	JAS	14	110.357	2.091	1.89	105.000	116.000
2020-004	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	16	107.750	3.455	3.21	102.000	113.000
2020-004	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	109.774	1.717	1.56	104.000	115.000
2020-004	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	49	109.592	2.019	1.84	104.000	115.000
011	CHLORIDE	CLIA TEST ID #: 355				EVALUATION CRITERIA: TARGET VALUE 5 %						
2020-005	3	ISE DILUTE	184	INTEGRA	93	ROCHE	38	85.711	1.554	1.81	81.000	90.000
2020-005	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	27	85.778	1.988	2.32	81.000	90.000
2020-005	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	16	87.933	1.181	1.34	84.000	92.000
2020-005	3	ISE DILUTE	203	ROCHE	93	ROCHE	34	82.500	1.867	2.26	78.000	87.000
2020-005	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	87.083	1.935	2.22	83.000	91.000
2020-005	4	ISE UND-DI	165	VITROS	63	KODAK	156	86.871	1.573	1.81	83.000	91.000
2020-005	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	37	85.027	1.852	2.18	81.000	89.000
2020-005	4	ISE UND-DI	194	BIOLIS-24S	149	JAS	14	85.286	2.185	2.56	81.000	90.000
2020-005	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	16	84.188	1.944	2.31	80.000	88.000
2020-005	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	84.516	1.388	1.64	80.000	89.000
2020-005	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	49	84.313	1.277	1.51	80.000	89.000
012	CHOLESTEROL, TOTAL	CLIA TEST ID #: 365				EVALUATION CRITERIA: TARGET VALUE 10 %						
2020-001	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	34	155.294	5.834	3.76	140.000	171.000
2020-001	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	34	153.857	4.580	2.98	138.000	169.000
2020-001	1	ENZYMATIC	165	VITROS	63	KODAK	147	159.462	4.205	2.64	144.000	175.000
2020-001	1	ENZYMATIC	184	INTEGRA	93	ROCHE	21	167.000	5.640	3.38	150.000	184.000
2020-001	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	41	171.825	5.783	3.37	155.000	189.000
2020-001	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	41	169.222	8.456	5.00	152.000	186.000
2020-001	1	ENZYMATIC	194	BIOLIS-24S	149	JAS	10	168.400	6.666	3.96	152.000	185.000
2020-001	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	17	165.412	5.542	3.35	149.000	182.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-001	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	14	172.000	3.964	2.30	155.000	189.000			
2020-001	1 ENZYMATI	203 ROCHE	93 ROCHE	18	163.778	4.315	2.63	147.000	180.000			
2020-001	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	58	155.500	5.281	3.40	140.000	171.000			
2020-001	1 ENZYMATI	214 Envoy 500	150 ELITECH	10	163.300	13.379	8.19	147.000	180.000			
2020-001	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	164.083	2.812	1.71	148.000	180.000			
2020-001	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	24	161.217	4.191	2.60	145.000	177.000			
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %									
2020-002	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	176.143	5.488	3.12	159.000	194.000			
2020-002	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	176.353	7.376	4.18	159.000	194.000			
2020-002	1 ENZYMATI	165 VITROS	63 KODAK	147	189.745	5.170	2.72	171.000	209.000			
2020-002	1 ENZYMATI	184 INTEGRA	93 ROCHE	21	191.857	6.475	3.38	173.000	211.000			
2020-002	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	193.333	9.718	5.03	174.000	213.000			
2020-002	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	196.350	7.330	3.73	177.000	216.000			
2020-002	1 ENZYMATI	194 BIOLIS-24S	149 JAS	10	193.300	5.515	2.85	174.000	213.000			
2020-002	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	189.588	5.961	3.14	171.000	209.000			
2020-002	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	14	197.929	5.161	2.61	178.000	218.000			
2020-002	1 ENZYMATI	203 ROCHE	93 ROCHE	18	188.222	4.577	2.43	169.000	207.000			
2020-002	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	58	176.509	5.670	3.21	159.000	194.000			
2020-002	1 ENZYMATI	214 Envoy 500	150 ELITECH	10	182.400	25.827	14.16	164.000	201.000			
2020-002	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	187.417	3.774	2.01	169.000	206.000			
2020-002	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	24	184.750	6.207	3.36	166.000	203.000			
2020-003	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	130.529	4.571	3.50	117.000	144.000			
2020-003	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	128.000	3.703	2.89	115.000	141.000			
2020-003	1 ENZYMATI	165 VITROS	63 KODAK	147	127.490	3.136	2.46	115.000	140.000			
2020-003	1 ENZYMATI	184 INTEGRA	93 ROCHE	21	138.750	8.005	5.77	125.000	153.000			
2020-003	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	142.902	5.207	3.64	129.000	157.000			
2020-003	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	142.778	5.493	3.85	129.000	157.000			
2020-003	1 ENZYMATI	194 BIOLIS-24S	149 JAS	10	142.000	3.688	2.60	128.000	156.000			
2020-003	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	138.235	4.821	3.49	124.000	152.000			

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-003	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	14	141.643	4.081	2.88	127.000	156.000		
2020-003	1 ENZYMATI	203 ROCHE	93 ROCHE	18	137.167	3.547	2.59	123.000	151.000		
2020-003	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	58	130.589	4.459	3.41	118.000	144.000		
2020-003	1 ENZYMATI	214 Envoy 500	150 ELITECH	10	137.500	13.433	9.77	124.000	151.000		
2020-003	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	135.333	2.357	1.74	122.000	149.000		
2020-003	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	24	133.083	4.071	3.06	120.000	146.000		
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-004	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	183.000	5.952	3.25	165.000	201.000		
2020-004	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	186.909	5.900	3.16	168.000	206.000		
2020-004	1 ENZYMATI	165 VITROS	63 KODAK	147	202.680	5.848	2.89	182.000	223.000		
2020-004	1 ENZYMATI	184 INTEGRA	93 ROCHE	21	201.762	6.354	3.15	182.000	222.000		
2020-004	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	204.444	11.334	5.54	184.000	225.000		
2020-004	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	209.750	6.826	3.25	189.000	231.000		
2020-004	1 ENZYMATI	194 BIOLIS-24S	149 JAS	10	206.600	7.378	3.57	186.000	227.000		
2020-004	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	200.412	5.730	2.86	180.000	220.000		
2020-004	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	14	208.286	6.363	3.06	187.000	229.000		
2020-004	1 ENZYMATI	203 ROCHE	93 ROCHE	18	199.833	5.718	2.86	180.000	220.000		
2020-004	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	58	185.724	6.945	3.74	167.000	204.000		
2020-004	1 ENZYMATI	214 Envoy 500	150 ELITECH	10	199.400	13.894	6.97	179.000	219.000		
2020-004	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	199.667	4.210	2.11	180.000	220.000		
2020-004	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	24	196.250	7.001	3.57	177.000	216.000		
2020-005	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	111.265	4.053	3.64	100.000	122.000		
2020-005	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	34	110.714	3.614	3.26	100.000	122.000		
2020-005	1 ENZYMATI	165 VITROS	63 KODAK	147	103.945	3.054	2.94	94.000	114.000		
2020-005	1 ENZYMATI	184 INTEGRA	93 ROCHE	21	118.095	3.853	3.26	106.000	130.000		
2020-005	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	120.732	4.644	3.85	109.000	133.000		
2020-005	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	41	121.222	5.116	4.22	109.000	133.000		
2020-005	1 ENZYMATI	194 BIOLIS-24S	149 JAS	10	122.200	4.377	3.58	110.000	134.000		
2020-005	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	117.176	3.929	3.35	105.000	129.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-005	1 ENZYMATIC	199 Rx-IMOLA	126 RANDOX	14	118.786	3.426	2.88	107.000	131.000			
2020-005	1 ENZYMATIC	203 ROCHE	93 ROCHE	18	114.778	3.083	2.69	103.000	126.000			
2020-005	1 ENZYMATIC	212 S. D. EXL	143 SIEMENS	58	111.552	3.966	3.56	100.000	123.000			
2020-005	1 ENZYMATIC	214 Envoy 500	150 ELITECH	10	121.300	25.892	21.35	109.000	133.000			
2020-005	1 ENZYMATIC	215 B.C DXC	134 BECKM.COUL	12	113.417	1.605	1.42	102.000	125.000			
2020-005	8 ENZYMATIC	186 BECK. COUL	76 AU-BECKMAN	24	111.957	2.493	2.23	101.000	123.000			
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %									
2020-001	11 ENC.-COLO.	165 VITROS	63 KODAK	142	54.306	3.382	6.23	38.000	70.600			
2020-001	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	32	62.747	3.540	5.64	43.900	81.600			
2020-001	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	24	63.079	4.795	7.60	44.200	82.000			
2020-001	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	37	56.403	4.647	8.24	39.500	73.300			
2020-001	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	12	43.583	2.564	5.88	30.500	56.700			
2020-001	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	55.167	5.615	10.18	38.600	71.700			
2020-001	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	59.000	1.414	2.40	41.300	76.700			
2020-001	11 ENC.-COLO.	203 ROCHE	93 ROCHE	29	61.800	2.622	4.24	43.300	80.300			
2020-001	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	17	62.959	4.025	6.39	44.100	81.800			
2020-001	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	64.400	4.152	6.45	45.100	83.700			
2020-001	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	54.750	1.479	2.70	38.300	71.200			
2020-001	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	12	63.575	2.483	3.91	44.500	82.600			
2020-001	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	31	54.570	1.907	3.49	38.200	70.900			
2020-001	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	55	55.493	1.942	3.50	38.800	72.100			
2020-002	11 ENC.-COLO.	165 VITROS	63 KODAK	142	59.834	3.580	5.98	41.900	77.800			
2020-002	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	32	66.416	3.528	5.31	46.500	86.300			
2020-002	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	24	71.554	4.804	6.71	50.100	93.000			
2020-002	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	37	62.105	5.466	8.80	43.500	80.700			
2020-002	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	12	49.267	2.370	4.81	34.500	64.000			
2020-002	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	61.333	1.247	2.03	42.900	79.700			
2020-002	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	61.187	7.351	12.01	42.800	79.500			
2020-002	11 ENC.-COLO.	203 ROCHE	93 ROCHE	29	65.597	2.806	4.28	45.900	85.300			

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2020-002	12	HOM-LQ.S.D	197 ARCHITECT	1	ABBOTT	17	70.459	4.376	6.21	49.300	91.600	
2020-002	12	HOM-LQ.S.D	214 Envoy 500	150	ELITECH	10	60.750	1.639	2.70	42.500	79.000	
2020-002	12	HOM-LQ.S.D	214 Envoy 500	150	ELITECH	10	69.600	9.276	13.33	48.700	90.500	
2020-002	12	HOM-LQ.S.D	215 B.C DXC	134	BECKM.COUL	12	70.950	2.664	3.75	49.700	92.200	
2020-002	13	AMIN.W.E.Q	38 SIEMENS DI	143	SIEMENS	31	56.043	1.953	3.48	39.200	72.900	
2020-002	13	AMIN.W.E.Q	212 S. D. EXL	143	SIEMENS	55	56.950	2.149	3.77	39.900	74.000	
013	CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-003	11	ENC.-COLO.	165 VITROS	63	KODAK	142	49.505	2.919	5.90	34.700	64.400	
2020-003	11	ENC.-COLO.	184 INTEGRA	93	ROCHE	32	59.594	3.351	5.62	41.700	77.500	
2020-003	11	ENC.-COLO.	186 BECK. COUL	76	AU-BECKMAN	24	54.935	3.096	5.64	38.500	71.400	
2020-003	11	ENC.-COLO.	191 RX-DAYTONA	126	RANDOX	37	50.265	5.254	10.45	35.200	65.300	
2020-003	11	ENC.-COLO.	194 BIOLIS-24S	154	(IDG)	12	37.333	1.434	3.84	26.100	48.500	
2020-003	11	ENC.-COLO.	199 Rx-IMOLA	126	RANDOX	16	46.860	5.947	12.69	32.800	60.900	
2020-003	11	ENC.-COLO.	199 Rx-IMOLA	126	RANDOX	16	49.333	3.399	6.89	34.500	64.100	
2020-003	11	ENC.-COLO.	203 ROCHE	93	ROCHE	29	58.279	2.293	3.93	40.800	75.800	
2020-003	12	HOM-LQ.S.D	197 ARCHITECT	1	ABBOTT	17	54.888	3.512	6.40	38.400	71.400	
2020-003	12	HOM-LQ.S.D	214 Envoy 500	150	ELITECH	10	56.800	3.124	5.50	39.800	73.800	
2020-003	12	HOM-LQ.S.D	214 Envoy 500	150	ELITECH	10	50.000	0.707	1.41	35.000	65.000	
2020-003	12	HOM-LQ.S.D	215 B.C DXC	134	BECKM.COUL	12	55.708	1.504	2.70	39.000	72.400	
2020-003	13	AMIN.W.E.Q	38 SIEMENS DI	143	SIEMENS	31	53.550	2.430	4.54	37.500	69.600	
2020-003	13	AMIN.W.E.Q	212 S. D. EXL	143	SIEMENS	55	54.502	1.953	3.58	38.200	70.900	
2020-004	11	ENC.-COLO.	165 VITROS	63	KODAK	142	62.452	3.635	5.82	43.700	81.200	
2020-004	11	ENC.-COLO.	184 INTEGRA	93	ROCHE	32	67.888	3.653	5.38	47.500	88.300	
2020-004	11	ENC.-COLO.	186 BECK. COUL	76	AU-BECKMAN	24	75.278	4.207	5.59	52.700	97.900	
2020-004	11	ENC.-COLO.	191 RX-DAYTONA	126	RANDOX	37	65.905	5.131	7.78	46.100	85.700	
2020-004	11	ENC.-COLO.	194 BIOLIS-24S	154	(IDG)	12	51.417	3.252	6.33	36.000	66.800	
2020-004	11	ENC.-COLO.	199 Rx-IMOLA	126	RANDOX	16	65.667	3.682	5.61	46.000	85.400	
2020-004	11	ENC.-COLO.	199 Rx-IMOLA	126	RANDOX	16	64.840	6.793	10.48	45.400	84.300	
2020-004	11	ENC.-COLO.	203 ROCHE	93	ROCHE	29	67.334	2.565	3.81	47.100	87.500	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-004	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	17	74.318	4.438	5.97	52.000	96.600			
2020-004	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	63.750	1.479	2.32	44.600	82.900			
2020-004	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	76.200	5.212	6.84	53.300	99.100			
2020-004	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	12	75.950	2.267	2.99	53.200	98.700			
2020-004	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	31	56.694	2.224	3.92	39.700	73.700			
2020-004	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	55	57.417	2.281	3.97	40.200	74.600			
013	CHOLESTEROL, (HDL)	CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %									
2020-005	11 ENC.-COLO.	165 VITROS	63 KODAK	142	45.697	2.504	5.48	32.000	59.400			
2020-005	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	32	54.556	3.257	5.97	38.200	70.900			
2020-005	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	24	47.909	3.082	6.43	33.500	62.300			
2020-005	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	37	44.403	4.716	10.62	31.100	57.700			
2020-005	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	12	30.950	1.887	6.10	21.700	40.200			
2020-005	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	41.620	5.574	13.39	29.100	54.100			
2020-005	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	16	46.667	2.625	5.62	32.700	60.700			
2020-005	11 ENC.-COLO.	203 ROCHE	93 ROCHE	29	53.390	2.083	3.90	37.400	69.400			
2020-005	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	17	47.835	2.426	5.07	33.500	62.200			
2020-005	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	52.000	7.403	14.24	36.400	67.600			
2020-005	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	46.500	1.118	2.40	32.600	60.500			
2020-005	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	12	48.650	1.562	3.21	34.100	63.200			
2020-005	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	31	50.673	2.191	4.32	35.500	65.900			
2020-005	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	55	51.561	2.001	3.88	36.100	67.000			
014	CREATINE KINASE (CK)	CLIA TEST ID #: 385										
2020-001	3 KINETIC 37	165 VITROS	63 KODAK	74	237.027	16.897	7.13	166.000	308.000			
2020-001	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	264.833	14.033	5.30	185.000	344.000			
2020-001	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	12	244.417	12.196	4.99	171.000	318.000			
2020-001	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	304.909	6.934	2.27	213.000	396.000			
2020-001	8 Hexokinase	184 INTEGRA	93 ROCHE	13	289.000	8.566	2.96	202.000	376.000			
2020-001	8 Hexokinase	203 ROCHE	93 ROCHE	12	291.417	6.601	2.27	204.000	379.000			
2020-001	9 SIEMENS	212 S. D. EXL	143 SIEMENS	20	291.900	9.924	3.40	204.000	379.000			

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
014	CREATINE KINASE (CK)	CLIA TEST ID #:	385	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-002	3 KINETIC 37	165 VITROS	63 KODAK	74	320.583	19.183	5.98	224.000	417.000			
2020-002	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	382.611	19.822	5.18	268.000	497.000			
2020-002	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	12	354.583	20.254	5.71	248.000	461.000			
2020-002	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	434.818	11.052	2.54	304.000	565.000			
2020-002	8 Hexokinase	184 INTEGRA	93 ROCHE	13	416.462	14.042	3.37	292.000	541.000			
2020-002	8 Hexokinase	203 ROCHE	93 ROCHE	12	418.500	9.751	2.33	293.000	544.000			
2020-002	9 SIEMENS	212 S. D. EXL	143 SIEMENS	20	420.650	12.547	2.98	294.000	547.000			
2020-003	3 KINETIC 37	165 VITROS	63 KODAK	74	107.644	6.943	6.45	75.000	140.000			
2020-003	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	114.389	6.945	6.07	80.000	149.000			
2020-003	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	12	107.250	7.120	6.64	75.000	139.000			
2020-003	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	136.000	2.296	1.69	95.000	177.000			
2020-003	8 Hexokinase	184 INTEGRA	93 ROCHE	13	127.846	3.254	2.55	89.000	166.000			
2020-003	8 Hexokinase	203 ROCHE	93 ROCHE	12	127.583	3.427	2.69	89.000	166.000			
2020-003	9 SIEMENS	212 S. D. EXL	143 SIEMENS	20	126.500	6.095	4.82	89.000	164.000			
2020-004	3 KINETIC 37	165 VITROS	63 KODAK	74	354.153	19.883	5.61	248.000	460.000			
2020-004	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	439.722	28.087	6.39	308.000	572.000			
2020-004	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	12	409.667	22.171	5.41	287.000	533.000			
2020-004	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	496.000	10.189	2.05	347.000	645.000			
2020-004	8 Hexokinase	184 INTEGRA	93 ROCHE	13	473.615	12.456	2.63	332.000	616.000			
2020-004	8 Hexokinase	203 ROCHE	93 ROCHE	12	482.250	8.604	1.78	338.000	627.000			
2020-004	9 SIEMENS	212 S. D. EXL	143 SIEMENS	20	481.500	15.148	3.15	337.000	626.000			
2020-005	3 KINETIC 37	165 VITROS	63 KODAK	74	20.181	0.509	2.52	14.000	26.000			
2020-005	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	18.611	1.799	9.67	13.000	24.000			
2020-005	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	12	20.417	3.013	14.76	14.000	27.000			
2020-005	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	25.455	1.076	4.23	18.000	33.000			
2020-005	8 Hexokinase	184 INTEGRA	93 ROCHE	13	24.077	1.071	4.45	17.000	31.000			

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-005	8	Hexokinase	203 ROCHE	93	ROCHE	12	23.250	0.829	3.57	16.000	30.000
2020-005	9	SIEMENS	212 S. D. EXL	143	SIEMENS	20	21.579	2.324	10.77	15.000	28.000
016 CREATININE			CLIA TEST ID #: 405		EVALUATION CRITERIA: TARGET VALUE	15 % OR 0.30	{GREATER}				
2020-001	2	AK. PIC. N	191 RX-DAYTONA	126	RANDOX	10	3.203	0.199	6.21	2.720	3.680
2020-001	2	AK. PIC. N	194 BIOLIS-24S	154	(IDG)	11	3.567	0.067	1.87	3.030	4.100
2020-001	2	AK. PIC. N	194 BIOLIS-24S	154	(IDG)	11	3.583	0.062	1.74	3.050	4.120
2020-001	3	AK. PIC. L	191 RX-DAYTONA	126	RANDOX	26	2.990	0.187	6.26	2.540	3.440
2020-001	4	ENZYMATIC	165 VITROS	63	KODAK	13	3.371	0.144	4.27	2.870	3.880
2020-001	6	KIN. ALK P	38 SIEMENS DI	143	SIEMENS	20	3.559	0.110	3.09	3.020	4.090
2020-001	6	KIN. ALK P	184 INTEGRA	93	ROCHE	23	3.606	0.182	5.04	3.070	4.150
2020-001	6	KIN. ALK P	186 BECK. COUL	76	AU-BECKMAN	27	3.499	0.111	3.19	2.970	4.020
2020-001	6	KIN. ALK P	186 BECK. COUL	76	AU-BECKMAN	27	3.507	0.199	5.67	2.980	4.030
2020-001	6	KIN. ALK P	197 ARCHITECT	1	ABBOTT	17	3.679	0.121	3.29	3.130	4.230
2020-001	6	KIN. ALK P	203 ROCHE	93	ROCHE	22	3.544	0.092	2.58	3.010	4.080
2020-001	6	KIN. ALK P	212 S. D. EXL	143	SIEMENS	29	3.624	0.106	2.91	3.080	4.170
2020-001	6	KIN. ALK P	215 B.C DXC	134	BECKM.COUL	12	3.576	0.103	2.88	3.040	4.110
2020-001	8	IDMS-CREA	165 VITROS	63	KODAK	148	3.387	0.086	2.53	2.880	3.890
2020-001	11	IDMS Jaffe	212 S. D. EXL	143	SIEMENS	24	3.625	0.110	3.03	3.080	4.170
2020-002	2	AK. PIC. N	191 RX-DAYTONA	126	RANDOX	10	4.200	0.290	6.90	3.570	4.830
2020-002	2	AK. PIC. N	194 BIOLIS-24S	154	(IDG)	11	4.767	0.189	3.96	4.050	5.480
2020-002	2	AK. PIC. N	194 BIOLIS-24S	154	(IDG)	11	4.697	0.277	5.90	3.990	5.400
2020-002	3	AK. PIC. L	191 RX-DAYTONA	126	RANDOX	26	3.970	0.306	7.72	3.370	4.570
2020-002	4	ENZYMATIC	165 VITROS	63	KODAK	13	4.389	0.221	5.04	3.730	5.050
2020-002	6	KIN. ALK P	38 SIEMENS DI	143	SIEMENS	20	4.833	0.148	3.06	4.110	5.560
2020-002	6	KIN. ALK P	184 INTEGRA	93	ROCHE	23	4.742	0.231	4.86	4.030	5.450
2020-002	6	KIN. ALK P	186 BECK. COUL	76	AU-BECKMAN	27	4.643	0.197	4.25	3.950	5.340
2020-002	6	KIN. ALK P	186 BECK. COUL	76	AU-BECKMAN	27	4.741	0.156	3.29	4.030	5.450
2020-002	6	KIN. ALK P	197 ARCHITECT	1	ABBOTT	17	5.039	0.170	3.38	4.280	5.790
2020-002	6	KIN. ALK P	203 ROCHE	93	ROCHE	22	4.683	0.135	2.88	3.980	5.390

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-002	6 KIN. ALK P	212 S. D. EXL	143 SIEMENS	29	4.886	0.120	2.46	4.150	5.620		
2020-002	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	4.853	0.129	2.66	4.120	5.580		
2020-002	8 IDMS-CREA	165 VITROS	63 KODAK	148	4.409	0.127	2.88	3.750	5.070		
2020-002	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	24	4.892	0.123	2.50	4.160	5.630		
016 CREATININE		CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}								
2020-003	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	1.877	0.161	8.60	1.580	2.180		
2020-003	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	11	2.044	0.097	4.77	1.740	2.350		
2020-003	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	11	2.073	0.052	2.53	1.760	2.380		
2020-003	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	26	1.793	0.135	7.56	1.490	2.090		
2020-003	4 ENZYMATIC	165 VITROS	63 KODAK	13	1.950	0.048	2.47	1.650	2.250		
2020-003	6 KIN. ALK P	38 SIEMENS DI	143 SIEMENS	20	2.039	0.063	3.09	1.730	2.340		
2020-003	6 KIN. ALK P	184 INTEGRA	93 ROCHE	23	2.054	0.103	5.03	1.750	2.360		
2020-003	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	27	1.983	0.072	3.65	1.680	2.280		
2020-003	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	27	2.040	0.174	8.52	1.730	2.350		
2020-003	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	17	1.993	0.073	3.66	1.690	2.290		
2020-003	6 KIN. ALK P	203 ROCHE	93 ROCHE	22	2.022	0.060	2.96	1.720	2.330		
2020-003	6 KIN. ALK P	212 S. D. EXL	143 SIEMENS	29	2.067	0.080	3.88	1.760	2.380		
2020-003	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	2.043	0.063	3.10	1.740	2.350		
2020-003	8 IDMS-CREA	165 VITROS	63 KODAK	148	1.962	0.056	2.83	1.660	2.260		
2020-003	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	24	2.079	0.053	2.54	1.770	2.390		
2020-004	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	4.640	0.336	7.25	3.940	5.340		
2020-004	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	11	5.427	0.116	2.13	4.610	6.240		
2020-004	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	11	5.270	0.317	6.02	4.480	6.060		
2020-004	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	26	4.456	0.350	7.85	3.790	5.120		
2020-004	4 ENZYMATIC	165 VITROS	63 KODAK	13	4.828	0.209	4.33	4.100	5.550		
2020-004	6 KIN. ALK P	38 SIEMENS DI	143 SIEMENS	20	5.464	0.160	2.93	4.640	6.280		
2020-004	6 KIN. ALK P	184 INTEGRA	93 ROCHE	23	5.235	0.278	5.32	4.450	6.020		
2020-004	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	27	5.230	0.170	3.26	4.450	6.010		
2020-004	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	27	5.285	0.191	3.62	4.490	6.080		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-004	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	17	5.689	0.195	3.43	4.840	6.540
2020-004	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	5.290	0.166	3.15	4.500	6.080
2020-004	6	KIN. ALK P	212	S. D. EXL	143	SIEMENS	29	5.516	0.146	2.66	4.690	6.340
2020-004	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	12	5.468	0.144	2.64	4.650	6.290
2020-004	8	IDMS-CREA	165	VITROS	63	KODAK	148	4.835	0.142	2.94	4.110	5.560
2020-004	11	IDMS Jaffe	212	S. D. EXL	143	SIEMENS	24	5.530	0.159	2.87	4.700	6.360
016	CREATININE	CLIA TEST ID #: 405		EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}								
2020-005	2	AK. PIC. N	191	RX-DAYTONA	126	RANDOX	10	1.082	0.181	16.74	0.780	1.380
2020-005	2	AK. PIC. N	194	BIOLIS-24S	154	(IDG)	11	1.114	0.068	6.11	0.810	1.410
2020-005	2	AK. PIC. N	194	BIOLIS-24S	154	(IDG)	11	1.200	0.000	0.00	0.900	1.500
2020-005	3	AK. PIC. L	191	RX-DAYTONA	126	RANDOX	26	0.971	0.105	10.80	0.670	1.270
2020-005	4	ENZYMATIC	165	VITROS	63	KODAK	13	0.959	0.035	3.65	0.660	1.260
2020-005	6	KIN. ALK P	38	SIEMENS DI	143	SIEMENS	20	0.931	0.064	6.88	0.630	1.230
2020-005	6	KIN. ALK P	184	INTEGRA	93	ROCHE	23	1.040	0.064	6.15	0.740	1.340
2020-005	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	27	0.975	0.058	5.92	0.680	1.280
2020-005	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	27	1.097	0.144	13.16	0.800	1.400
2020-005	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	17	0.862	0.059	6.80	0.560	1.160
2020-005	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	0.973	0.030	3.06	0.670	1.270
2020-005	6	KIN. ALK P	212	S. D. EXL	143	SIEMENS	29	0.956	0.058	6.10	0.660	1.260
2020-005	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	12	1.033	0.054	5.26	0.730	1.330
2020-005	8	IDMS-CREA	165	VITROS	63	KODAK	148	0.962	0.044	4.53	0.660	1.260
2020-005	11	IDMS Jaffe	212	S. D. EXL	143	SIEMENS	24	0.938	0.048	5.11	0.640	1.240
017	GLUCOSE	CLIA TEST ID #: 415		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}								
2020-001	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	12	91.500	1.848	2.02	82.000	101.000
2020-001	3	GL. OXI. C	165	VITROS	63	KODAK	166	93.233	2.032	2.18	84.000	103.000
2020-001	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	12	99.667	2.593	2.60	90.000	110.000
2020-001	3	GL. OXI. C	214	Envoy 500	150	ELITECH	11	98.909	1.564	1.58	89.000	109.000
2020-001	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	32	98.375	3.080	3.13	89.000	108.000
2020-001	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	97.500	2.335	2.39	88.000	107.000
2020-001	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	27	94.444	2.973	3.15	85.000	104.000

Participant Summary
MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-001	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	99.897	4.373	4.38	90.000	110.000
2020-001	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	102.000	2.291	2.25	92.000	112.000
2020-001	5	HEXOKINASE	199	Rx-IMOLA	126	RANDOX	12	99.667	2.687	2.70	90.000	110.000
2020-001	5	HEXOKINASE	203	ROCHE	93	ROCHE	18	96.444	2.852	2.96	87.000	106.000
2020-001	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	59	99.517	3.024	3.04	90.000	109.000
2020-001	6	HEXOKI. CO	191	RX-DAYTONA	126	RANDOX	10	95.000	3.376	3.55	86.000	105.000
2020-001	12	HEXOKINASE	184	INTEGRA	93	ROCHE	27	97.259	2.716	2.79	88.000	107.000
2020-001	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	92.444	2.967	3.21	83.000	102.000
2020-001	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	95.500	2.031	2.13	86.000	105.000
017 GLUCOSE			CLIA TEST ID #: 415			EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2020-002	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	12	118.750	1.831	1.54	107.000	131.000
2020-002	3	GL. OXI. C	165	VITROS	63	KODAK	166	121.185	2.344	1.93	109.000	133.000
2020-002	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	12	128.333	3.793	2.96	116.000	141.000
2020-002	3	GL. OXI. C	214	Envoy 500	150	ELITECH	11	126.900	2.914	2.30	114.000	140.000
2020-002	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	32	126.438	4.054	3.21	114.000	139.000
2020-002	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	127.300	3.494	2.74	115.000	140.000
2020-002	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	27	121.148	3.274	2.70	109.000	133.000
2020-002	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	129.375	3.199	2.47	116.000	142.000
2020-002	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	124.893	5.073	4.06	112.000	137.000
2020-002	5	HEXOKINASE	199	Rx-IMOLA	126	RANDOX	12	125.917	3.174	2.52	113.000	139.000
2020-002	5	HEXOKINASE	203	ROCHE	93	ROCHE	18	123.882	2.398	1.94	111.000	136.000
2020-002	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	59	127.172	3.217	2.53	114.000	140.000
2020-002	6	HEXOKI. CO	191	RX-DAYTONA	126	RANDOX	10	121.000	6.309	5.21	109.000	133.000
2020-002	12	HEXOKINASE	184	INTEGRA	93	ROCHE	27	124.926	3.219	2.58	112.000	137.000
2020-002	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	118.882	2.541	2.14	107.000	131.000
2020-002	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	123.500	2.318	1.88	111.000	136.000
2020-003	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	12	59.417	1.115	1.88	53.000	65.000
2020-003	3	GL. OXI. C	165	VITROS	63	KODAK	166	57.976	1.822	3.14	52.000	64.000
2020-003	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	12	66.667	1.841	2.76	60.000	73.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	3	GL. OXI. C	214	Envoy 500	150	ELITECH	11	66.091	1.781	2.70	59.000	73.000
2020-003	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	32	64.677	1.532	2.37	58.000	71.000
2020-003	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	64.400	2.200	3.42	58.000	71.000
2020-003	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	27	62.074	1.884	3.04	56.000	68.000
2020-003	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	66.321	1.605	2.42	60.000	73.000
2020-003	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	66.375	1.728	2.60	60.000	73.000
2020-003	5	HEXOKINASE	199	Rx-IMOLA	126	RANDOX	12	66.417	1.656	2.49	60.000	73.000
2020-003	5	HEXOKINASE	203	ROCHE	93	ROCHE	18	62.722	2.103	3.35	56.000	69.000
2020-003	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	59	65.655	2.064	3.14	59.000	72.000
2020-003	6	HEXOKI. CO	191	RX-DAYTONA	126	RANDOX	10	64.200	1.887	2.94	58.000	71.000
2020-003	12	HEXOKINASE	184	INTEGRA	93	ROCHE	27	63.667	1.700	2.67	57.000	70.000
2020-003	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	59.500	2.062	3.46	54.000	66.000
2020-003	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	62.313	1.570	2.52	56.000	69.000
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2020-004	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	12	131.000	1.472	1.12	118.000	144.000
2020-004	3	GL. OXI. C	165	VITROS	63	KODAK	166	134.540	2.891	2.15	121.000	148.000
2020-004	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	12	144.083	3.378	2.34	130.000	158.000
2020-004	3	GL. OXI. C	214	Envoy 500	150	ELITECH	11	138.182	4.745	3.43	124.000	152.000
2020-004	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	32	139.594	4.703	3.37	126.000	154.000
2020-004	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	138.100	2.809	2.03	124.000	152.000
2020-004	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	27	134.704	4.479	3.32	121.000	148.000
2020-004	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	140.750	2.332	1.66	127.000	155.000
2020-004	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	29	139.107	5.492	3.95	125.000	153.000
2020-004	5	HEXOKINASE	199	Rx-IMOLA	126	RANDOX	12	138.250	3.538	2.56	124.000	152.000
2020-004	5	HEXOKINASE	203	ROCHE	93	ROCHE	18	138.722	4.356	3.14	125.000	153.000
2020-004	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	59	140.276	3.938	2.81	126.000	154.000
2020-004	6	HEXOKI. CO	191	RX-DAYTONA	126	RANDOX	10	134.400	4.005	2.98	121.000	148.000
2020-004	12	HEXOKINASE	184	INTEGRA	93	ROCHE	27	137.577	3.607	2.62	124.000	151.000
2020-004	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	130.222	3.938	3.02	117.000	143.000
2020-004	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	136.500	3.674	2.69	123.000	150.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
017	GLUCOSE	CLIA TEST ID #: 415		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}								
2020-005	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	37.750	1.299	3.44	32.000	44.000	
2020-005	3	GL. OXI. C	165 VITROS	63	KODAK	166	35.860	1.136	3.17	30.000	42.000	
2020-005	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	43.917	1.656	3.77	38.000	50.000	
2020-005	3	GL. OXI. C	214 Envoy 500	150	ELITECH	11	42.200	0.872	2.07	36.000	48.000	
2020-005	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	32	41.839	1.416	3.39	36.000	48.000	
2020-005	5	HEXOKINASE	184 INTEGRA	93	ROCHE	10	40.500	1.857	4.59	35.000	47.000	
2020-005	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	27	39.815	1.540	3.87	34.000	46.000	
2020-005	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	29	44.172	1.858	4.21	38.000	50.000	
2020-005	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	29	43.875	0.781	1.78	38.000	50.000	
2020-005	5	HEXOKINASE	199 Rx-IMOLA	126	RANDOX	12	43.750	1.422	3.25	38.000	50.000	
2020-005	5	HEXOKINASE	203 ROCHE	93	ROCHE	18	40.111	1.242	3.10	34.000	46.000	
2020-005	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	59	42.483	1.783	4.20	36.000	48.000	
2020-005	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	42.600	1.497	3.51	37.000	49.000	
2020-005	12	HEXOKINASE	184 INTEGRA	93	ROCHE	27	40.519	1.316	3.25	35.000	47.000	
2020-005	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	18	37.000	1.202	3.25	31.000	43.000	
2020-005	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	40.000	1.118	2.80	34.000	46.000	
018	IRON, TOTAL	CLIA TEST ID #: 425		EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-001	4	COULOMETR.	165 VITROS	63	KODAK	14	171.286	5.787	3.38	137.000	206.000	
2020-002	4	COULOMETR.	165 VITROS	63	KODAK	14	248.143	8.766	3.53	199.000	298.000	
2020-003	4	COULOMETR.	165 VITROS	63	KODAK	14	88.071	4.183	4.75	70.000	106.000	
2020-004	4	COULOMETR.	165 VITROS	63	KODAK	14	288.500	7.872	2.73	231.000	346.000	
2020-005	4	COULOMETR.	165 VITROS	63	KODAK	14	39.929	3.954	9.90	32.000	48.000	
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435										
2020-001	3	KINETIC-37	186 BECK. COUL	76	AU-BECKMAN	18	260.444	16.717	6.42	208.000	313.000	
2020-001	3	KINETIC-37	194 BIOLIS-24S	154	(IDG)	10	275.800	12.057	4.37	221.000	331.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-001	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	252.273	6.440	2.55	202.000 303.000	
2020-001	6	NADH	184	INTEGRA	93	ROCHE	15	327.467	10.923	3.34	262.000 393.000	
2020-001	6	NADH	203	ROCHE	93	ROCHE	22	318.500	9.576	3.01	255.000 382.000	
2020-001	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	13	299.308	6.787	2.27	239.000 359.000	
2020-001	7	SIEMENS 37	212	S. D. EXL	143	SIEMENS	20	304.350	12.322	4.05	243.000 365.000	
2020-001	9	LDH1	165	VITROS	63	KODAK	65	365.266	8.298	2.27	292.000 438.000	
019	LACTASE DEHY. (LDH)		CLIA	TEST ID #: 435		EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-002	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	367.889	24.481	6.65	294.000 441.000	
2020-002	3	KINETIC-37	194	BIOLIS-24S	154	(IDG)	10	391.600	16.936	4.32	313.000 470.000	
2020-002	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	351.364	7.240	2.06	281.000 422.000	
2020-002	6	NADH	184	INTEGRA	93	ROCHE	15	456.067	11.997	2.63	365.000 547.000	
2020-002	6	NADH	203	ROCHE	93	ROCHE	22	449.318	14.137	3.15	359.000 539.000	
2020-002	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	13	426.385	8.949	2.10	341.000 512.000	
2020-002	7	SIEMENS 37	212	S. D. EXL	143	SIEMENS	20	428.000	14.061	3.29	342.000 514.000	
2020-002	9	LDH1	165	VITROS	63	KODAK	65	511.250	15.372	3.01	409.000 614.000	
2020-003	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	126.444	8.591	6.79	101.000 152.000	
2020-003	3	KINETIC-37	194	BIOLIS-24S	154	(IDG)	10	134.900	5.856	4.34	108.000 162.000	
2020-003	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	123.000	2.697	2.19	98.000 148.000	
2020-003	6	NADH	184	INTEGRA	93	ROCHE	15	161.400	5.535	3.43	129.000 194.000	
2020-003	6	NADH	203	ROCHE	93	ROCHE	22	155.636	5.322	3.42	125.000 187.000	
2020-003	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	13	145.231	5.591	3.85	116.000 174.000	
2020-003	7	SIEMENS 37	212	S. D. EXL	143	SIEMENS	20	146.000	7.543	5.17	117.000 175.000	
2020-003	9	LDH1	165	VITROS	63	KODAK	65	176.921	5.174	2.92	142.000 212.000	
2020-004	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	420.333	28.152	6.70	336.000 504.000	
2020-004	3	KINETIC-37	194	BIOLIS-24S	154	(IDG)	10	449.200	23.224	5.17	359.000 539.000	
2020-004	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	397.545	9.811	2.47	318.000 477.000	
2020-004	6	NADH	184	INTEGRA	93	ROCHE	15	516.000	15.904	3.08	413.000 619.000	
2020-004	6	NADH	203	ROCHE	93	ROCHE	22	509.500	12.302	2.41	408.000 611.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-004	7	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	13	487.462	11.453	2.35	390.000	585.000
2020-004	7	SIEMENS	37	212	S. D. EXL	143	SIEMENS	20	486.950	16.262	3.34	390.000	584.000
2020-004	9	LDH1		165	VITROS	63	KODAK	65	578.815	20.444	3.53	463.000	695.000
019 LACTASE DEHY. (LDH)		CLIA TEST ID #: 435				EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-005	3	KINETIC-37		186	BECK. COUL	76	AU-BECKMAN	18	40.167	3.236	8.06	32.000	48.000
2020-005	3	KINETIC-37		194	BIOLIS-24S	154	(IDG)	10	43.100	3.833	8.89	34.000	52.000
2020-005	3	KINETIC-37		215	B.C DXC	134	BECKM.COUL	11	40.091	2.021	5.04	32.000	48.000
2020-005	6	NADH		184	INTEGRA	93	ROCHE	15	53.533	2.963	5.54	43.000	64.000
2020-005	6	NADH		203	ROCHE	93	ROCHE	22	50.455	1.751	3.47	40.000	61.000
2020-005	7	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	13	46.692	4.250	9.10	37.000	56.000
2020-005	7	SIEMENS	37	212	S. D. EXL	143	SIEMENS	20	47.000	3.449	7.34	38.000	56.000
2020-005	9	LDH1		165	VITROS	63	KODAK	65	51.109	4.780	9.35	41.000	61.000
021 MAGNESIUM		CLIA TEST ID #: 455				EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-001	2	CALMAGITE		215	B.C DXC	134	BECKM.COUL	11	3.236	0.260	8.04	2.400	4.000
2020-001	5	COLOR./K E		165	VITROS	63	KODAK	21	3.381	0.066	1.96	2.500	4.200
2020-002	2	CALMAGITE		215	B.C DXC	134	BECKM.COUL	11	4.155	0.239	5.75	3.100	5.200
2020-002	5	COLOR./K E		165	VITROS	63	KODAK	21	4.352	0.096	2.20	3.300	5.400
2020-003	2	CALMAGITE		215	B.C DXC	134	BECKM.COUL	11	2.020	0.075	3.70	1.500	2.500
2020-003	5	COLOR./K E		165	VITROS	63	KODAK	21	2.181	0.073	3.35	1.600	2.700
2020-004	2	CALMAGITE		215	B.C DXC	134	BECKM.COUL	11	4.609	0.219	4.76	3.500	5.800
2020-004	5	COLOR./K E		165	VITROS	63	KODAK	21	4.790	0.111	2.31	3.600	6.000
2020-005	2	CALMAGITE		215	B.C DXC	134	BECKM.COUL	11	1.290	0.054	4.17	1.000	1.600
2020-005	5	COLOR./K E		165	VITROS	63	KODAK	21	1.357	0.058	4.30	1.000	1.700
022 POTASSIUM		CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00							
2020-001	3	ISE DILUT.		184	INTEGRA	93	ROCHE	36	3.844	0.055	1.43	2.800	4.800
2020-001	3	ISE DILUT.		186	BECK. COUL	76	AU-BECKMAN	27	3.785	0.085	2.24	2.800	4.800

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-001	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	3.835	0.084	2.18	2.800	4.800	
2020-001	3 ISE DILUT.	203 ROCHE	93 ROCHE		35	3.866	0.086	2.22	2.900	4.900	
2020-001	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	3.725	0.083	2.23	2.700	4.700	
2020-001	4 ISE UNLIL.	165 VITROS	63 KODAK		159	3.809	0.072	1.88	2.800	4.800	
2020-001	4 ISE UNLIL.	191 RX-DAYTONA	126 RANDOX		35	3.834	0.083	2.15	2.800	4.800	
2020-001	4 ISE UNLIL.	194 BIOLIS-24S	149 JAS		14	3.736	0.117	3.14	2.700	4.700	
2020-001	4 ISE UNLIL.	199 Rx-IMOLA	126 RANDOX		16	3.800	0.087	2.28	2.800	4.800	
2020-001	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		31	3.781	0.064	1.70	2.800	4.800	
2020-001	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		50	3.780	0.064	1.69	2.800	4.800	
022 POTASSIUM		CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE 1.00								
2020-002	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	4.692	0.064	1.36	3.700	5.700	
2020-002	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		27	4.633	0.090	1.95	3.600	5.600	
2020-002	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	4.682	0.098	2.10	3.700	5.700	
2020-002	3 ISE DILUT.	203 ROCHE	93 ROCHE		35	4.743	0.090	1.90	3.700	5.700	
2020-002	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	4.600	0.091	1.98	3.600	5.600	
2020-002	4 ISE UNLIL.	165 VITROS	63 KODAK		159	4.686	0.090	1.93	3.700	5.700	
2020-002	4 ISE UNLIL.	191 RX-DAYTONA	126 RANDOX		35	4.671	0.086	1.84	3.700	5.700	
2020-002	4 ISE UNLIL.	194 BIOLIS-24S	149 JAS		14	4.400	0.141	3.21	3.400	5.400	
2020-002	4 ISE UNLIL.	199 Rx-IMOLA	126 RANDOX		16	4.619	0.113	2.45	3.600	5.600	
2020-002	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		31	4.665	0.078	1.68	3.700	5.700	
2020-002	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		50	4.635	0.082	1.77	3.600	5.600	
2020-003	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	2.809	0.028	1.00	1.800	3.800	
2020-003	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		27	2.833	0.082	2.88	1.800	3.800	
2020-003	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	2.800	0.077	2.74	1.800	3.800	
2020-003	3 ISE DILUT.	203 ROCHE	93 ROCHE		35	2.854	0.073	2.56	1.900	3.900	
2020-003	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	2.683	0.055	2.06	1.700	3.700	
2020-003	4 ISE UNLIL.	165 VITROS	63 KODAK		159	2.784	0.055	1.96	1.800	3.800	
2020-003	4 ISE UNLIL.	191 RX-DAYTONA	126 RANDOX		35	2.837	0.102	3.58	1.800	3.800	
2020-003	4 ISE UNLIL.	194 BIOLIS-24S	149 JAS		14	2.957	0.098	3.31	2.000	4.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLIX	16	2.838	0.070	2.45	1.800	3.800
2020-003	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	2.745	0.071	2.59	1.700	3.700
2020-003	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	2.735	0.052	1.89	1.700	3.700
022	POTASSIUM	CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00						
2020-004	3	ISE DILUT.	184	INTEGRA	93	ROCHE	36	5.085	0.081	1.59	4.100	6.100
2020-004	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	27	5.023	0.093	1.86	4.000	6.000
2020-004	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	17	5.076	0.094	1.85	4.100	6.100
2020-004	3	ISE DILUT.	203	ROCHE	93	ROCHE	35	5.160	0.107	2.08	4.200	6.200
2020-004	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	5.033	0.075	1.48	4.000	6.000
2020-004	4	ISE UNLIL.	165	VITROS	63	KODAK	159	5.106	0.108	2.11	4.100	6.100
2020-004	4	ISE UNLIL.	191	Rx-DAYTONA	126	RANLIX	35	5.043	0.125	2.48	4.000	6.000
2020-004	4	ISE UNLIL.	194	BIOLIS-24S	149	JAS	14	4.700	0.136	2.90	3.700	5.700
2020-004	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLIX	16	5.006	0.134	2.69	4.000	6.000
2020-004	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	5.077	0.079	1.56	4.100	6.100
2020-004	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	5.053	0.093	1.84	4.100	6.100
2020-005	3	ISE DILUT.	184	INTEGRA	93	ROCHE	36	2.097	0.044	2.10	1.100	3.100
2020-005	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	27	2.170	0.066	3.02	1.200	3.200
2020-005	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	17	2.082	0.038	1.83	1.100	3.100
2020-005	3	ISE DILUT.	203	ROCHE	93	ROCHE	35	2.134	0.079	3.71	1.100	3.100
2020-005	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	1.983	0.069	3.46	1.000	3.000
2020-005	4	ISE UNLIL.	165	VITROS	63	KODAK	159	2.066	0.053	2.59	1.100	3.100
2020-005	4	ISE UNLIL.	191	Rx-DAYTONA	126	RANLIX	35	2.160	0.107	4.97	1.200	3.200
2020-005	4	ISE UNLIL.	194	BIOLIS-24S	149	JAS	14	2.379	0.094	3.95	1.400	3.400
2020-005	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLIX	16	2.147	0.050	2.32	1.100	3.100
2020-005	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	2.010	0.053	2.64	1.000	3.000
2020-005	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	1.996	0.035	1.74	1.000	3.000
023	SODIUM	CLIA TEST ID #: 475				EVALUATION CRITERIA: TARGET VALUE 4.00						
2020-001	3	ISE DILUTE	165	VITROS	63	KODAK	10	148.000	2.757	1.86	144.000	152.000
2020-001	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	148.281	2.004	1.35	144.000	152.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-001	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	27	147.704	2.839	1.92	144.000	152.000
2020-001	3	ISE DILUTE	194	BIOLIS-24S	149	JAS	13	141.308	1.897	1.34	137.000	145.000
2020-001	3	ISE DILUTE	194	BIOLIS-24S	149	JAS	13	146.000	2.449	1.68	142.000	150.000
2020-001	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	15	150.467	1.408	0.94	146.000	154.000
2020-001	3	ISE DILUTE	203	ROCHE	93	ROCHE	26	149.320	2.034	1.36	145.000	153.000
2020-001	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	146.250	2.087	1.43	142.000	150.000
2020-001	4	ISE UNLIL.	165	VITROS	63	KODAK	148	149.528	1.958	1.31	146.000	154.000
2020-001	4	ISE UNLIL.	165	VITROS	63	KODAK	148	147.556	1.257	0.85	144.000	152.000
2020-001	4	ISE UNLIL.	191	RX-DAYTONA	126	RANLIL.	30	148.367	1.798	1.21	144.000	152.000
2020-001	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLIL.	15	147.200	2.007	1.36	143.000	151.000
2020-001	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	150.032	2.584	1.72	146.000	154.000
2020-001	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	149.250	2.046	1.37	145.000	153.000
2020-001	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	149.939	2.104	1.40	146.000	154.000
2020-001	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	151.333	0.943	0.62	147.000	155.000
023	SODIUM		CLIA TEST ID #: 475			EVALUATION CRITERIA: TARGET VALUE 4.00						
2020-002	3	ISE DILUTE	165	VITROS	63	KODAK	10	159.300	2.759	1.73	155.000	163.000
2020-002	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	157.000	2.151	1.37	153.000	161.000
2020-002	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	27	156.704	2.942	1.88	153.000	161.000
2020-002	3	ISE DILUTE	194	BIOLIS-24S	149	JAS	13	152.750	2.165	1.42	149.000	157.000
2020-002	3	ISE DILUTE	194	BIOLIS-24S	149	JAS	13	147.231	1.576	1.07	143.000	151.000
2020-002	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	15	158.400	1.583	1.00	154.000	162.000
2020-002	3	ISE DILUTE	203	ROCHE	93	ROCHE	26	158.577	2.405	1.52	155.000	163.000
2020-002	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	155.167	2.339	1.51	151.000	159.000
2020-002	4	ISE UNLIL.	165	VITROS	63	KODAK	148	158.333	1.247	0.79	154.000	162.000
2020-002	4	ISE UNLIL.	165	VITROS	63	KODAK	148	159.510	2.194	1.38	156.000	164.000
2020-002	4	ISE UNLIL.	191	RX-DAYTONA	126	RANLIL.	30	157.517	2.514	1.60	154.000	162.000
2020-002	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLIL.	15	157.067	2.863	1.82	153.000	161.000
2020-002	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	158.500	2.000	1.26	155.000	163.000
2020-002	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	31	158.133	2.045	1.29	154.000	162.000
2020-002	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	160.000	1.414	0.88	156.000	164.000
2020-002	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	50	157.878	2.273	1.44	154.000	162.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
023	SODIUM	CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE	4.00							
2020-003	3 ISE DILUTE	165 VITROS	63 KODAK		10	139.900	2.427	1.73	136.000	144.000	
2020-003	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	140.097	1.593	1.14	136.000	144.000	
2020-003	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		27	141.000	2.211	1.57	137.000	145.000	
2020-003	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		13	136.231	1.887	1.39	132.000	140.000	
2020-003	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		13	139.000	1.871	1.35	135.000	143.000	
2020-003	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	142.133	1.500	1.06	138.000	146.000	
2020-003	3 ISE DILUTE	203 ROCHE	93 ROCHE		26	142.077	1.880	1.32	138.000	146.000	
2020-003	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	138.167	1.404	1.02	134.000	142.000	
2020-003	4 ISE UNDIL.	165 VITROS	63 KODAK		148	140.000	2.000	1.43	136.000	144.000	
2020-003	4 ISE UNDIL.	165 VITROS	63 KODAK		148	140.000	1.700	1.21	136.000	144.000	
2020-003	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		30	139.655	1.987	1.42	136.000	144.000	
2020-003	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		15	138.733	1.692	1.22	135.000	143.000	
2020-003	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		31	142.710	2.203	1.54	139.000	147.000	
2020-003	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		31	142.000	1.732	1.22	138.000	146.000	
2020-003	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		50	142.857	2.070	1.45	139.000	147.000	
2020-003	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		50	142.667	1.886	1.32	139.000	147.000	
2020-004	3 ISE DILUTE	165 VITROS	63 KODAK		10	163.300	3.796	2.32	159.000	167.000	
2020-004	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	159.967	1.853	1.16	156.000	164.000	
2020-004	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		27	160.926	3.102	1.93	157.000	165.000	
2020-004	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		13	155.250	1.920	1.24	151.000	159.000	
2020-004	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		13	150.154	2.627	1.75	146.000	154.000	
2020-004	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	161.733	2.380	1.47	158.000	166.000	
2020-004	3 ISE DILUTE	203 ROCHE	93 ROCHE		26	162.462	2.406	1.48	158.000	166.000	
2020-004	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	158.750	1.738	1.09	155.000	163.000	
2020-004	4 ISE UNDIL.	165 VITROS	63 KODAK		148	158.556	2.266	1.43	155.000	163.000	
2020-004	4 ISE UNDIL.	165 VITROS	63 KODAK		148	163.731	2.576	1.57	160.000	168.000	
2020-004	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		30	161.033	2.429	1.51	157.000	165.000	
2020-004	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		15	160.467	3.462	2.16	156.000	164.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-004	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	159.250	3.231	2.03	155.000	163.000		
2020-004	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	161.742	3.111	1.92	158.000	166.000		
2020-004	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	50	160.333	9.463	5.90	156.000	164.000		
2020-004	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	50	161.460	2.662	1.65	157.000	165.000		
023	SODIUM	CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00							
2020-005	3 ISE DILUTE	165 VITROS	63 KODAK	10	128.700	2.002	1.56	125.000	133.000		
2020-005	3 ISE DILUTE	184 INTEGRA	93 ROCHE	32	129.531	1.695	1.31	126.000	134.000		
2020-005	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	27	130.731	1.788	1.37	127.000	135.000		
2020-005	3 ISE DILUTE	194 BIOLIS-24S	149 JAS	13	128.333	1.027	0.80	124.000	132.000		
2020-005	3 ISE DILUTE	194 BIOLIS-24S	149 JAS	13	130.250	1.479	1.14	126.000	134.000		
2020-005	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	130.000	2.033	1.56	126.000	134.000		
2020-005	3 ISE DILUTE	203 ROCHE	93 ROCHE	26	130.462	2.291	1.76	126.000	134.000		
2020-005	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	128.083	1.656	1.29	124.000	132.000		
2020-005	4 ISE UNLIL.	165 VITROS	63 KODAK	148	128.590	1.609	1.25	125.000	133.000		
2020-005	4 ISE UNLIL.	165 VITROS	63 KODAK	148	128.667	1.700	1.32	125.000	133.000		
2020-005	4 ISE UNLIL.	191 RX-DAYTONA	126 RANDOX	30	128.267	1.896	1.48	124.000	132.000		
2020-005	4 ISE UNLIL.	199 RX-IMOLA	126 RANDOX	15	127.933	1.611	1.26	124.000	132.000		
2020-005	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	132.258	2.257	1.71	128.000	136.000		
2020-005	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	31	131.625	2.288	1.74	128.000	136.000		
2020-005	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	50	132.061	1.754	1.33	128.000	136.000		
2020-005	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	50	140.667	13.695	9.74	137.000	145.000		
024	TOTAL PROTEIN	CLIA TEST ID #: 485		EVALUATION CRITERIA: TARGET VALUE 10 %							
2020-001	1 BIURET	38 SIEMENS DI	143 SIEMENS	33	5.248	0.102	1.94	4.700	5.800		
2020-001	1 BIURET	165 VITROS	63 KODAK	141	5.029	0.124	2.46	4.500	5.500		
2020-001	1 BIURET	184 INTEGRA	93 ROCHE	30	4.917	0.134	2.73	4.400	5.400		
2020-001	1 BIURET	186 BECK. COUL	76 AU-BECKMAN	23	4.965	0.134	2.70	4.500	5.500		
2020-001	1 BIURET	191 RX-DAYTONA	126 RANDOX	36	5.100	0.167	3.28	4.600	5.600		
2020-001	1 BIURET	191 RX-DAYTONA	126 RANDOX	36	5.133	0.094	1.84	4.600	5.600		
2020-001	1 BIURET	194 BIOLIS-24S	149 JAS	10	5.450	0.169	3.10	4.900	6.000		
2020-001	1 BIURET	197 ARCHITECT	1 ABBOTT	19	5.037	0.172	3.42	4.500	5.500		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-001	1	BIURET	199	Rx-IMOLA	126	RANDOX	14	5.064	0.123	2.43	4.600	5.600
2020-001	1	BIURET	203	ROCHE	93	ROCHE	27	5.096	0.109	2.14	4.600	5.600
2020-001	1	BIURET	212	S. D. EXL	143	SIEMENS	60	5.278	0.156	2.96	4.800	5.800
2020-001	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	5.092	0.132	2.59	4.600	5.600
024 TOTAL PROTEIN			CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-002	1	BIURET	38	SIEMENS DI	143	SIEMENS	33	6.439	0.158	2.45	5.800	7.100
2020-002	1	BIURET	165	VITROS	63	KODAK	141	6.115	0.140	2.29	5.500	6.700
2020-002	1	BIURET	184	INTEGRA	93	ROCHE	30	6.047	0.131	2.17	5.400	6.700
2020-002	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	23	6.074	0.165	2.71	5.500	6.700
2020-002	1	BIURET	191	RX-DAYTONA	126	RANDOX	36	6.300	0.000	0.00	5.700	6.900
2020-002	1	BIURET	191	RX-DAYTONA	126	RANDOX	36	6.275	0.210	3.35	5.600	6.900
2020-002	1	BIURET	194	BIOLIS-24S	149	JAS	10	6.810	0.221	3.25	6.100	7.500
2020-002	1	BIURET	197	ARCHITECT	1	ABBOTT	19	6.216	0.142	2.29	5.600	6.800
2020-002	1	BIURET	199	Rx-IMOLA	126	RANDOX	14	6.229	0.148	2.38	5.600	6.900
2020-002	1	BIURET	203	ROCHE	93	ROCHE	27	6.231	0.103	1.65	5.600	6.900
2020-002	1	BIURET	212	S. D. EXL	143	SIEMENS	60	6.464	0.183	2.83	5.800	7.100
2020-002	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	6.258	0.206	3.29	5.600	6.900
2020-003	1	BIURET	38	SIEMENS DI	143	SIEMENS	33	3.830	0.072	1.87	3.400	4.200
2020-003	1	BIURET	165	VITROS	63	KODAK	141	3.729	0.115	3.10	3.400	4.100
2020-003	1	BIURET	184	INTEGRA	93	ROCHE	30	3.620	0.101	2.80	3.300	4.000
2020-003	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	23	3.622	0.114	3.15	3.300	4.000
2020-003	1	BIURET	191	RX-DAYTONA	126	RANDOX	36	3.743	0.154	4.10	3.400	4.100
2020-003	1	BIURET	191	RX-DAYTONA	126	RANDOX	36	3.633	0.047	1.30	3.300	4.000
2020-003	1	BIURET	194	BIOLIS-24S	149	JAS	10	4.050	0.092	2.28	3.600	4.500
2020-003	1	BIURET	197	ARCHITECT	1	ABBOTT	19	3.658	0.127	3.47	3.300	4.000
2020-003	1	BIURET	199	Rx-IMOLA	126	RANDOX	14	3.679	0.086	2.34	3.300	4.000
2020-003	1	BIURET	203	ROCHE	93	ROCHE	27	3.763	0.119	3.16	3.400	4.100
2020-003	1	BIURET	212	S. D. EXL	143	SIEMENS	60	3.869	0.109	2.83	3.500	4.300
2020-003	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	3.700	0.115	3.12	3.300	4.100

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
024 TOTAL PROTEIN		CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-004	1	BIURET	38 SIEMENS DI	143 SIEMENS	33	7.033	0.168	2.39	6.300	7.700	
2020-004	1	BIURET	165 VITROS	63 KODAK	141	6.641	0.160	2.42	6.000	7.300	
2020-004	1	BIURET	184 INTEGRA	93 ROCHE	30	6.533	0.181	2.78	5.900	7.200	
2020-004	1	BIURET	186 BECK. COUL	76 AU-BECKMAN	23	6.630	0.183	2.76	6.000	7.300	
2020-004	1	BIURET	191 RX-DAYTONA	126 RANDOX	36	6.667	0.094	1.41	6.000	7.300	
2020-004	1	BIURET	191 RX-DAYTONA	126 RANDOX	36	6.800	0.225	3.31	6.100	7.500	
2020-004	1	BIURET	194 BIOLIS-24S	149 JAS	10	7.390	0.212	2.87	6.700	8.100	
2020-004	1	BIURET	197 ARCHITECT	1 ABBOTT	19	6.758	0.166	2.46	6.100	7.400	
2020-004	1	BIURET	199 Rx-IMOLA	126 RANDOX	14	6.757	0.180	2.67	6.100	7.400	
2020-004	1	BIURET	203 ROCHE	93 ROCHE	27	6.744	0.102	1.52	6.100	7.400	
2020-004	1	BIURET	212 S. D. EXL	143 SIEMENS	60	7.013	0.207	2.95	6.300	7.700	
2020-004	1	BIURET	215 B.C DXC	134 BECKM.COUL	12	6.792	0.206	3.03	6.100	7.500	
2020-005	1	BIURET	38 SIEMENS DI	143 SIEMENS	33	2.821	0.077	2.73	2.500	3.100	
2020-005	1	BIURET	165 VITROS	63 KODAK	141	2.766	0.109	3.93	2.500	3.000	
2020-005	1	BIURET	184 INTEGRA	93 ROCHE	30	2.673	0.100	3.73	2.400	2.900	
2020-005	1	BIURET	186 BECK. COUL	76 AU-BECKMAN	23	2.683	0.092	3.42	2.400	3.000	
2020-005	1	BIURET	191 RX-DAYTONA	126 RANDOX	36	2.743	0.096	3.52	2.500	3.000	
2020-005	1	BIURET	191 RX-DAYTONA	126 RANDOX	36	2.900	0.245	8.45	2.600	3.200	
2020-005	1	BIURET	194 BIOLIS-24S	149 JAS	10	3.000	0.118	3.94	2.700	3.300	
2020-005	1	BIURET	197 ARCHITECT	1 ABBOTT	19	2.644	0.060	2.26	2.400	2.900	
2020-005	1	BIURET	199 Rx-IMOLA	126 RANDOX	14	2.721	0.067	2.48	2.400	3.000	
2020-005	1	BIURET	203 ROCHE	93 ROCHE	27	2.789	0.083	2.98	2.500	3.100	
2020-005	1	BIURET	212 S. D. EXL	143 SIEMENS	60	2.842	0.099	3.50	2.600	3.100	
2020-005	1	BIURET	215 B.C DXC	134 BECKM.COUL	12	2.725	0.142	5.22	2.500	3.000	
025 TRIGLYCERIDES		CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-001	2	ENZ. COLO.	165 VITROS	63 KODAK	144	162.741	4.636	2.85	122.000	203.000	
2020-001	2	ENZ. COLO.	203 ROCHE	93 ROCHE	17	144.118	4.993	3.46	108.000	180.000	
2020-001	4	ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	30	135.897	3.487	2.57	102.000	170.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-001	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	57	136.250	3.572	2.62	102.000	170.000
2020-001	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	25	144.760	7.490	5.17	109.000	181.000
2020-001	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	22	141.636	6.183	4.37	106.000	177.000
2020-001	9	GPO TRIND.	199	Rx-IMOLA	126	RANDOX	10	140.000	4.025	2.87	105.000	175.000
2020-001	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	151.182	5.060	3.35	113.000	189.000
2020-001	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	27	146.333	5.235	3.58	110.000	183.000
2020-001	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	14	138.286	3.514	2.54	104.000	173.000
2020-001	10	ENZYM. GPO	203	ROCHE	93	ROCHE	17	142.941	2.313	1.62	107.000	179.000
2020-001	10	ENZYM. GPO	214	Envoy 500	150	ELITECH	10	141.600	6.726	4.75	106.000	177.000
025 TRIGLYCERIDES			CLIA TEST ID #: 495			EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-002	2	ENZ. COLO.	165	VITROS	63	KODAK	144	199.091	5.274	2.65	149.000	249.000
2020-002	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	171.813	4.275	2.49	129.000	215.000
2020-002	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	30	166.793	4.664	2.80	125.000	208.000
2020-002	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	57	166.482	3.836	2.30	125.000	208.000
2020-002	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	25	175.000	7.725	4.41	131.000	219.000
2020-002	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	22	168.864	7.307	4.33	127.000	211.000
2020-002	9	GPO TRIND.	199	Rx-IMOLA	126	RANDOX	10	167.400	4.758	2.84	126.000	209.000
2020-002	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	185.818	3.242	1.74	139.000	232.000
2020-002	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	27	174.654	5.771	3.30	131.000	218.000
2020-002	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	14	164.857	5.926	3.59	124.000	206.000
2020-002	10	ENZYM. GPO	203	ROCHE	93	ROCHE	17	171.882	3.628	2.11	129.000	215.000
2020-002	10	ENZYM. GPO	214	Envoy 500	150	ELITECH	10	160.700	26.822	16.69	121.000	201.000
2020-003	2	ENZ. COLO.	165	VITROS	63	KODAK	144	119.720	3.802	3.18	90.000	150.000
2020-003	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	109.188	2.627	2.41	82.000	136.000
2020-003	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	30	98.828	3.052	3.09	74.000	124.000
2020-003	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	57	99.875	3.257	3.26	75.000	125.000
2020-003	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	25	108.040	4.677	4.33	81.000	135.000
2020-003	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	22	106.000	5.641	5.32	80.000	133.000
2020-003	9	GPO TRIND.	199	Rx-IMOLA	126	RANDOX	10	104.500	2.579	2.47	78.000	131.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	115.273	3.595	3.12	86.000	144.000
2020-003	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	27	111.500	3.377	3.03	84.000	139.000
2020-003	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	14	104.500	2.922	2.80	78.000	131.000
2020-003	10	ENZYM. GPO	203	ROCHE	93	ROCHE	17	109.412	2.475	2.26	82.000	137.000
2020-003	10	ENZYM. GPO	214	Envoy 500	150	ELITECH	10	108.000	5.865	5.43	81.000	135.000
025 TRIGLYCERIDES		CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-004	2	ENZ. COLO.	165	VITROS	63	KODAK	144	216.653	6.647	3.07	162.000	271.000
2020-004	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	184.438	3.807	2.06	138.000	231.000
2020-004	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	30	181.103	4.389	2.42	136.000	226.000
2020-004	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	57	180.965	4.573	2.53	136.000	226.000
2020-004	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	25	190.360	8.827	4.64	143.000	238.000
2020-004	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	22	182.591	7.499	4.11	137.000	228.000
2020-004	9	GPO TRIND.	199	Rx-IMOLA	126	RANDOX	10	181.400	5.625	3.10	136.000	227.000
2020-004	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	199.545	3.056	1.53	150.000	249.000
2020-004	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	27	187.333	6.583	3.51	141.000	234.000
2020-004	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	14	176.714	6.029	3.41	133.000	221.000
2020-004	10	ENZYM. GPO	203	ROCHE	93	ROCHE	17	185.063	2.487	1.34	139.000	231.000
2020-004	10	ENZYM. GPO	214	Envoy 500	150	ELITECH	10	182.400	11.333	6.21	137.000	228.000
2020-005	2	ENZ. COLO.	165	VITROS	63	KODAK	144	88.113	2.680	3.04	66.000	110.000
2020-005	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	83.647	3.289	3.93	63.000	105.000
2020-005	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	30	71.367	2.983	4.18	54.000	89.000
2020-005	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	57	72.464	2.976	4.11	54.000	91.000
2020-005	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	25	81.920	4.127	5.04	61.000	102.000
2020-005	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	22	79.905	2.910	3.64	60.000	100.000
2020-005	9	GPO TRIND.	199	Rx-IMOLA	126	RANDOX	10	79.600	2.059	2.59	60.000	100.000
2020-005	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	87.818	2.167	2.47	66.000	110.000
2020-005	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	27	85.731	2.550	2.97	64.000	107.000
2020-005	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	14	79.286	2.657	3.35	59.000	99.000
2020-005	10	ENZYM. GPO	203	ROCHE	93	ROCHE	17	83.235	2.016	2.42	62.000	104.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-005	10 ENZYM. GPO	214	Envoy	500	150 ELITECH	10	91.500	31.043	33.93	69.000	114.000	
026	UREA NITROGEN (BUN)	CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}								
2020-001	2 UREAS.GLDH	38	SIEMENS	DI	143 SIEMENS	32	24.344	1.189	4.88	22.000	27.000	
2020-001	2 UREAS.GLDH	184	INTEGRA		93 ROCHE	39	23.949	0.876	3.66	22.000	26.000	
2020-001	2 UREAS.GLDH	186	BECK.	COUL	76 AU-BECKMAN	27	24.385	0.684	2.80	22.000	27.000	
2020-001	2 UREAS.GLDH	191	RX-DAYTONA		126 RANDOX	36	24.314	1.008	4.14	22.000	27.000	
2020-001	2 UREAS.GLDH	194	BIOLIS-24S		149 JAS	10	24.800	1.778	7.17	23.000	27.000	
2020-001	2 UREAS.GLDH	199	Rx-IMOLA		126 RANDOX	13	24.308	1.136	4.67	22.000	26.000	
2020-001	2 UREAS.GLDH	203	ROCHE		93 ROCHE	35	23.941	0.802	3.35	22.000	26.000	
2020-001	2 UREAS.GLDH	212	S. D. EXL		143 SIEMENS	61	24.533	1.072	4.37	22.000	27.000	
2020-001	8 UREASE QUI	165	VITROS		63 KODAK	161	21.814	0.652	2.99	20.000	24.000	
2020-002	2 UREAS.GLDH	38	SIEMENS	DI	143 SIEMENS	32	30.813	1.530	4.96	28.000	34.000	
2020-002	2 UREAS.GLDH	184	INTEGRA		93 ROCHE	39	30.410	1.055	3.47	28.000	33.000	
2020-002	2 UREAS.GLDH	186	BECK.	COUL	76 AU-BECKMAN	27	30.885	0.974	3.15	28.000	34.000	
2020-002	2 UREAS.GLDH	191	RX-DAYTONA		126 RANDOX	36	30.528	1.067	3.49	28.000	33.000	
2020-002	2 UREAS.GLDH	194	BIOLIS-24S		149 JAS	10	29.900	2.071	6.93	27.000	33.000	
2020-002	2 UREAS.GLDH	199	Rx-IMOLA		126 RANDOX	13	30.538	1.646	5.39	28.000	33.000	
2020-002	2 UREAS.GLDH	203	ROCHE		93 ROCHE	35	30.147	1.004	3.33	27.000	33.000	
2020-002	2 UREAS.GLDH	212	S. D. EXL		143 SIEMENS	61	31.033	1.278	4.12	28.000	34.000	
2020-002	8 UREASE QUI	165	VITROS		63 KODAK	161	27.925	0.771	2.76	25.000	30.000	
2020-003	2 UREAS.GLDH	38	SIEMENS	DI	143 SIEMENS	32	16.406	0.896	5.46	14.000	18.000	
2020-003	2 UREAS.GLDH	184	INTEGRA		93 ROCHE	39	16.282	0.638	3.92	14.000	18.000	
2020-003	2 UREAS.GLDH	186	BECK.	COUL	76 AU-BECKMAN	27	16.654	0.676	4.06	15.000	19.000	
2020-003	2 UREAS.GLDH	191	RX-DAYTONA		126 RANDOX	36	16.771	0.636	3.79	15.000	19.000	
2020-003	2 UREAS.GLDH	194	BIOLIS-24S		149 JAS	10	16.900	0.700	4.14	15.000	19.000	
2020-003	2 UREAS.GLDH	199	Rx-IMOLA		126 RANDOX	13	16.692	0.722	4.32	15.000	19.000	
2020-003	2 UREAS.GLDH	203	ROCHE		93 ROCHE	35	16.257	0.769	4.73	14.000	18.000	
2020-003	2 UREAS.GLDH	212	S. D. EXL		143 SIEMENS	61	16.650	0.891	5.35	15.000	19.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	8 UREASE QUI	165	VITROS	63	KODAK	161	14.880	0.520	3.49	13.000	17.000	
026	UREA NITROGEN (BUN)	CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}								
2020-004	2 UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	32	33.781	1.494	4.42	31.000	37.000	
2020-004	2 UREAS.GLDH	184	INTEGRA	93	ROCHE	39	33.513	1.217	3.63	30.000	37.000	
2020-004	2 UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	27	33.962	1.091	3.21	31.000	37.000	
2020-004	2 UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	33.056	1.268	3.84	30.000	36.000	
2020-004	2 UREAS.GLDH	194	BIOLIS-24S	149	JAS	10	33.800	1.778	5.26	31.000	37.000	
2020-004	2 UREAS.GLDH	199	Rx-IMOLA	126	RANDOX	13	33.231	1.625	4.89	30.000	36.000	
2020-004	2 UREAS.GLDH	203	ROCHE	93	ROCHE	35	33.600	1.224	3.64	31.000	37.000	
2020-004	2 UREAS.GLDH	212	S. D. EXL	143	SIEMENS	61	34.217	1.279	3.74	31.000	37.000	
2020-004	8 UREASE QUI	165	VITROS	63	KODAK	161	30.873	0.832	2.70	28.000	34.000	
2020-005	2 UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	32	10.938	0.747	6.83	9.000	13.000	
2020-005	2 UREAS.GLDH	184	INTEGRA	93	ROCHE	39	11.026	0.480	4.35	9.000	13.000	
2020-005	2 UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	27	11.192	0.556	4.97	9.000	13.000	
2020-005	2 UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	11.528	0.552	4.79	10.000	14.000	
2020-005	2 UREAS.GLDH	194	BIOLIS-24S	149	JAS	10	11.700	0.781	6.68	10.000	14.000	
2020-005	2 UREAS.GLDH	199	Rx-IMOLA	126	RANDOX	13	11.308	0.462	4.08	9.000	13.000	
2020-005	2 UREAS.GLDH	203	ROCHE	93	ROCHE	35	10.971	0.453	4.13	9.000	13.000	
2020-005	2 UREAS.GLDH	212	S. D. EXL	143	SIEMENS	61	11.424	0.906	7.93	9.000	13.000	
2020-005	8 UREASE QUI	165	VITROS	63	KODAK	161	10.013	0.418	4.18	8.000	12.000	
027	URIC ACID	CLIA TEST ID #: 515		EVALUATION CRITERIA: TARGET VALUE 17 %								
2020-001	2 URICASE	38	SIEMENS DI	143	SIEMENS	18	5.672	0.141	2.48	4.700	6.600	
2020-001	2 URICASE	165	VITROS	63	KODAK	96	5.732	0.139	2.42	4.800	6.700	
2020-001	2 URICASE	184	INTEGRA	93	ROCHE	19	5.895	0.170	2.88	4.900	6.900	
2020-001	2 URICASE	197	ARCHITECT	1	ABBOTT	14	5.821	0.126	2.17	4.800	6.800	
2020-001	2 URICASE	203	ROCHE	93	ROCHE	24	5.563	0.135	2.42	4.600	6.500	
2020-001	2 URICASE	212	S. D. EXL	143	SIEMENS	32	5.635	0.166	2.94	4.700	6.600	
2020-001	5 URICASE 37	186	BECK. COUL	76	AU-BECKMAN	17	5.769	0.110	1.91	4.800	6.700	
2020-001	5 URICASE 37	215	B.C DXC	134	BECKM.COUL	11	5.527	0.086	1.56	4.600	6.500	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
027 URIC ACID		CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %								
2020-002	2 URICASE	38 SIEMENS DI	143 SIEMENS		18	7.583	0.189	2.50	6.300	8.900	
2020-002	2 URICASE	165 VITROS	63 KODAK		96	7.593	0.192	2.52	6.300	8.900	
2020-002	2 URICASE	184 INTEGRA	93 ROCHE		19	7.932	0.239	3.01	6.600	9.300	
2020-002	2 URICASE	197 ARCHITECT	1 ABBOTT		14	7.721	0.094	1.22	6.400	9.000	
2020-002	2 URICASE	203 ROCHE	93 ROCHE		24	7.517	0.143	1.91	6.200	8.800	
2020-002	2 URICASE	212 S. D. EXL	143 SIEMENS		32	7.535	0.234	3.10	6.300	8.800	
2020-002	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		17	7.638	0.176	2.31	6.300	8.900	
2020-002	5 URICASE 37	215 B.C DXC	134 BECKM.COUL		11	7.291	0.124	1.70	6.100	8.500	
2020-003	2 URICASE	38 SIEMENS DI	143 SIEMENS		18	3.288	0.113	3.44	2.700	3.800	
2020-003	2 URICASE	165 VITROS	63 KODAK		96	3.326	0.087	2.62	2.800	3.900	
2020-003	2 URICASE	184 INTEGRA	93 ROCHE		19	3.463	0.109	3.14	2.900	4.100	
2020-003	2 URICASE	197 ARCHITECT	1 ABBOTT		14	3.457	0.090	2.61	2.900	4.000	
2020-003	2 URICASE	203 ROCHE	93 ROCHE		24	3.279	0.087	2.64	2.700	3.800	
2020-003	2 URICASE	212 S. D. EXL	143 SIEMENS		32	3.303	0.131	3.96	2.700	3.900	
2020-003	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		17	3.463	0.086	2.47	2.900	4.100	
2020-003	5 URICASE 37	215 B.C DXC	134 BECKM.COUL		11	3.309	0.067	2.02	2.700	3.900	
2020-004	2 URICASE	38 SIEMENS DI	143 SIEMENS		18	8.456	0.231	2.74	7.000	9.900	
2020-004	2 URICASE	165 VITROS	63 KODAK		96	8.482	0.227	2.67	7.000	9.900	
2020-004	2 URICASE	184 INTEGRA	93 ROCHE		19	8.805	0.307	3.49	7.300	10.300	
2020-004	2 URICASE	197 ARCHITECT	1 ABBOTT		14	8.614	0.168	1.96	7.100	10.100	
2020-004	2 URICASE	203 ROCHE	93 ROCHE		24	8.404	0.184	2.19	7.000	9.800	
2020-004	2 URICASE	212 S. D. EXL	143 SIEMENS		32	8.422	0.255	3.02	7.000	9.900	
2020-004	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		17	8.531	0.211	2.48	7.100	10.000	
2020-004	5 URICASE 37	215 B.C DXC	134 BECKM.COUL		11	8.082	0.147	1.81	6.700	9.500	
2020-005	2 URICASE	38 SIEMENS DI	143 SIEMENS		18	1.829	0.089	4.88	1.500	2.100	
2020-005	2 URICASE	165 VITROS	63 KODAK		96	1.681	0.071	4.23	1.400	2.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-005	2	URICASE	184	INTEGRA	93	ROCHE	19	1.867	0.067	3.57	1.500	2.200
2020-005	2	URICASE	197	ARCHITECT	1	ABBOTT	14	1.907	0.070	3.69	1.600	2.200
2020-005	2	URICASE	203	ROCHE	93	ROCHE	24	1.750	0.050	2.86	1.500	2.000
2020-005	2	URICASE	212	S. D. EXL	143	SIEMENS	32	1.829	0.105	5.76	1.500	2.100
2020-005	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	17	1.912	0.096	5.04	1.600	2.200
2020-005	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	11	1.800	0.060	3.35	1.500	2.100
135	GGT (GAMMA GLU. TRA)		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-001	5	KINETIC 37	165	VITROS	63	KODAK	23	291.000	8.753	3.01	233.000	349.000
2020-002	5	KINETIC 37	165	VITROS	63	KODAK	23	400.304	14.035	3.51	320.000	480.000
2020-003	5	KINETIC 37	165	VITROS	63	KODAK	23	130.696	4.154	3.18	105.000	157.000
2020-004	5	KINETIC 37	165	VITROS	63	KODAK	23	448.391	17.267	3.85	359.000	538.000
2020-005	5	KINETIC 37	165	VITROS	63	KODAK	23	17.318	1.103	6.37	14.000	21.000
136	LACTIC ACID											
2020-001	1	LACTATE-P.	165	VITROS	63	KODAK	17	2.771	0.057	2.06	2.200	3.300
2020-002	1	LACTATE-P.	165	VITROS	63	KODAK	17	3.835	0.076	1.99	3.100	4.600
2020-003	1	LACTATE-P.	165	VITROS	63	KODAK	17	1.359	0.049	3.62	1.100	1.600
2020-004	1	LACTATE-P.	165	VITROS	63	KODAK	17	4.312	0.083	1.93	3.400	5.200
2020-005	1	LACTATE-P.	165	VITROS	63	KODAK	17	0.529	0.046	8.61	0.400	0.600
137	LIPASE											
2020-001	1	COLORIMET.	203	ROCHE	93	ROCHE	12	44.500	3.571	8.02	36.000	53.000
2020-001	2	ENZIMATIC	165	VITROS	63	KODAK	80	420.863	13.415	3.19	337.000	505.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-001	2 ENZIMATIC	215 B.C DXC	134 BECKM.COUL		12	49.667	2.357	4.75	40.000	60.000	
2020-001	4 ENZ/COLOR	186 BECK. COUL	76 AU-BECKMAN		14	44.357	4.835	10.90	35.000	53.000	
2020-001	4 ENZ/COLOR	203 ROCHE	93 ROCHE		14	42.857	3.440	8.03	34.000	51.000	
2020-001	5 SIEMENS 37	212 S. D. EXL	143 SIEMENS		16	182.938	16.589	9.07	146.000	220.000	
137 LIPASE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-002	1 COLORIMET.	203 ROCHE	93 ROCHE		12	58.083	4.699	8.09	46.000	70.000	
2020-002	2 ENZIMATIC	165 VITROS	63 KODAK		80	561.100	18.146	3.23	449.000	673.000	
2020-002	2 ENZIMATIC	215 B.C DXC	134 BECKM.COUL		12	60.750	2.046	3.37	49.000	73.000	
2020-002	4 ENZ/COLOR	186 BECK. COUL	76 AU-BECKMAN		14	59.643	4.352	7.30	48.000	72.000	
2020-002	4 ENZ/COLOR	203 ROCHE	93 ROCHE		14	56.571	4.732	8.36	45.000	68.000	
2020-002	5 SIEMENS 37	212 S. D. EXL	143 SIEMENS		16	233.733	15.321	6.55	187.000	280.000	
2020-003	1 COLORIMET.	203 ROCHE	93 ROCHE		12	27.000	1.780	6.59	22.000	32.000	
2020-003	2 ENZIMATIC	165 VITROS	63 KODAK		80	236.113	9.893	4.19	189.000	283.000	
2020-003	2 ENZIMATIC	215 B.C DXC	134 BECKM.COUL		12	36.333	2.357	6.49	29.000	44.000	
2020-003	4 ENZ/COLOR	186 BECK. COUL	76 AU-BECKMAN		14	25.214	3.051	12.10	20.000	30.000	
2020-003	4 ENZ/COLOR	203 ROCHE	93 ROCHE		14	26.357	1.586	6.02	21.000	32.000	
2020-003	5 SIEMENS 37	212 S. D. EXL	143 SIEMENS		16	115.133	8.326	7.23	92.000	138.000	
2020-004	1 COLORIMET.	203 ROCHE	93 ROCHE		12	64.833	6.203	9.57	52.000	78.000	
2020-004	2 ENZIMATIC	165 VITROS	63 KODAK		80	622.288	20.155	3.24	498.000	747.000	
2020-004	2 ENZIMATIC	215 B.C DXC	134 BECKM.COUL		12	66.583	2.564	3.85	53.000	80.000	
2020-004	4 ENZ/COLOR	186 BECK. COUL	76 AU-BECKMAN		14	67.231	3.765	5.60	54.000	81.000	
2020-004	4 ENZ/COLOR	203 ROCHE	93 ROCHE		14	63.000	6.245	9.91	50.000	76.000	
2020-004	5 SIEMENS 37	212 S. D. EXL	143 SIEMENS		16	256.375	21.109	8.23	205.000	308.000	
2020-005	1 COLORIMET.	203 ROCHE	93 ROCHE		12	15.750	1.010	6.42	13.000	19.000	
2020-005	2 ENZIMATIC	165 VITROS	63 KODAK		80	106.388	6.078	5.71	85.000	128.000	
2020-005	2 ENZIMATIC	215 B.C DXC	134 BECKM.COUL		12	27.000	2.380	8.82	22.000	32.000	
2020-005	4 ENZ/COLOR	186 BECK. COUL	76 AU-BECKMAN		14	13.154	2.931	22.28	11.000	16.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-005	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	15.500	0.732	4.72	12.000	19.000
2020-005	5	SIEMENS 37	212	S. D. EXL	143	SIEMENS	16	73.813	9.983	13.52	59.000	89.000
138	PHOSPHORUS	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-001	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	17	3.265	0.145	4.45	2.600	3.900
2020-001	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	3.226	0.115	3.57	2.600	3.900
2020-001	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	18	3.061	0.101	3.29	2.400	3.700
2020-001	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	2.987	0.171	5.72	2.400	3.600
2020-001	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	3.086	0.131	4.23	2.500	3.700
2020-001	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	3.172	0.166	5.24	2.500	3.800
2020-001	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	3.800	0.169	4.45	3.000	4.600
2020-001	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	3.192	0.161	5.03	2.600	3.800
2020-001	2	END POINT	184	INTEGRA	93	ROCHE	10	3.080	0.098	3.18	2.500	3.700
2020-002	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	17	4.112	0.076	1.84	3.300	4.900
2020-002	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	4.002	0.118	2.94	3.200	4.800
2020-002	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	18	3.917	0.130	3.32	3.100	4.700
2020-002	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	3.913	0.268	6.85	3.100	4.700
2020-002	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	3.969	0.158	3.98	3.200	4.800
2020-002	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	4.786	0.136	2.83	3.800	5.700
2020-002	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	4.053	0.154	3.80	3.200	4.900
2020-002	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	4.167	0.201	4.83	3.300	5.000
2020-002	2	END POINT	184	INTEGRA	93	ROCHE	10	3.950	0.102	2.59	3.200	4.700
2020-003	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	17	2.112	0.096	4.56	1.700	2.500
2020-003	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	2.227	0.111	4.99	1.800	2.700
2020-003	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	18	2.028	0.099	4.88	1.600	2.400
2020-003	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	1.960	0.120	6.12	1.600	2.400
2020-003	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	2.052	0.113	5.52	1.600	2.500
2020-003	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	2.081	0.133	6.41	1.700	2.500
2020-003	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	2.614	0.136	5.18	2.100	3.100

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-003	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	2.050	0.050	2.44	1.600	2.500
2020-003	2	END POINT	184	INTEGRA	93	ROCHE	10	2.050	0.092	4.50	1.600	2.500
138	PHOSPHORUS	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-004	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	17	4.535	0.137	3.02	3.600	5.400
2020-004	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	4.378	0.144	3.29	3.500	5.300
2020-004	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	18	4.247	0.119	2.81	3.400	5.100
2020-004	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	4.320	0.290	6.72	3.500	5.200
2020-004	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	4.379	0.130	2.96	3.500	5.300
2020-004	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	5.271	0.128	2.42	4.200	6.300
2020-004	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	4.478	0.139	3.09	3.600	5.400
2020-004	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	4.608	0.214	4.64	3.700	5.500
2020-004	2	END POINT	184	INTEGRA	93	ROCHE	10	4.330	0.100	2.32	3.500	5.200
2020-005	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	17	1.344	0.050	3.69	1.100	1.600
2020-005	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	1.429	0.129	9.04	1.100	1.700
2020-005	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	18	1.283	0.083	6.49	1.000	1.500
2020-005	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	1.233	0.114	9.21	1.000	1.500
2020-005	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	1.271	0.092	7.24	1.000	1.500
2020-005	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	1.316	0.068	5.14	1.100	1.600
2020-005	1	PHOSPHOMO.	212	S. D. EXL	143	SIEMENS	33	1.786	0.112	6.30	1.400	2.100
2020-005	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	1.233	0.075	6.04	1.000	1.500
2020-005	2	END POINT	184	INTEGRA	93	ROCHE	10	1.290	0.083	6.44	1.000	1.500
169	CO2					EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-001	1	Generic Me	38	SIEMENS DI	143	SIEMENS	15	20.667	4.222	20.43	8.000	33.000
2020-001	1	Generic Me	165	VITROS	63	KODAK	59	17.362	1.459	8.40	13.000	22.000
2020-001	1	Generic Me	184	INTEGRA	93	ROCHE	13	17.923	1.269	7.08	14.000	22.000
2020-001	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	20	17.100	1.997	11.68	11.000	23.000
2020-001	1	Generic Me	203	ROCHE	93	ROCHE	15	16.333	2.211	13.54	10.000	23.000
2020-001	1	Generic Me	212	S. D. EXL	143	SIEMENS	19	19.211	2.440	12.70	12.000	27.000
2020-001	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	11	18.273	2.453	13.42	11.000	26.000

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-001	2 ENZYMATI	165 VITROS	63 KODAK	90	17.393	1.443	8.29	13.000	22.000		
2020-001	2 ENZYMATI	203 ROCHE	93 ROCHE	12	16.583	1.498	9.03	12.000	21.000		
2020-001	2 ENZYMATI	212 S. D. EXL	143 SIEMENS	33	19.094	2.403	12.58	12.000	26.000		
169 CO2		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-002	1 Generic Me	38 SIEMENS DI	143 SIEMENS	15	23.667	3.534	14.93	13.000	34.000		
2020-002	1 Generic Me	165 VITROS	63 KODAK	59	19.729	1.645	8.34	15.000	25.000		
2020-002	1 Generic Me	184 INTEGRA	93 ROCHE	13	20.000	1.519	7.60	15.000	25.000		
2020-002	1 Generic Me	186 BECK. COUL	76 AU-BECKMAN	20	20.000	2.449	12.25	13.000	27.000		
2020-002	1 Generic Me	203 ROCHE	93 ROCHE	15	18.667	2.650	14.20	11.000	27.000		
2020-002	1 Generic Me	212 S. D. EXL	143 SIEMENS	19	21.158	2.323	10.98	14.000	28.000		
2020-002	2 ENZYMATI	38 SIEMENS DI	143 SIEMENS	11	21.364	1.920	8.99	16.000	27.000		
2020-002	2 ENZYMATI	165 VITROS	63 KODAK	90	19.798	1.478	7.46	15.000	24.000		
2020-002	2 ENZYMATI	203 ROCHE	93 ROCHE	12	18.750	1.963	10.47	13.000	25.000		
2020-002	2 ENZYMATI	212 S. D. EXL	143 SIEMENS	33	21.563	2.318	10.75	15.000	29.000		
2020-003	1 Generic Me	38 SIEMENS DI	143 SIEMENS	15	17.600	3.861	21.94	6.000	29.000		
2020-003	1 Generic Me	165 VITROS	63 KODAK	59	13.864	1.455	10.49	9.000	18.000		
2020-003	1 Generic Me	184 INTEGRA	93 ROCHE	13	13.692	1.588	11.60	9.000	18.000		
2020-003	1 Generic Me	186 BECK. COUL	76 AU-BECKMAN	20	13.750	1.841	13.39	8.000	19.000		
2020-003	1 Generic Me	203 ROCHE	93 ROCHE	15	13.133	1.928	14.68	7.000	19.000		
2020-003	1 Generic Me	212 S. D. EXL	143 SIEMENS	19	15.316	2.153	14.06	9.000	22.000		
2020-003	2 ENZYMATI	38 SIEMENS DI	143 SIEMENS	11	15.273	2.178	14.26	9.000	22.000		
2020-003	2 ENZYMATI	165 VITROS	63 KODAK	90	13.922	1.392	10.00	10.000	18.000		
2020-003	2 ENZYMATI	203 ROCHE	93 ROCHE	12	13.333	1.650	12.37	8.000	18.000		
2020-003	2 ENZYMATI	212 S. D. EXL	143 SIEMENS	33	15.656	2.159	13.79	9.000	22.000		
2020-004	1 Generic Me	38 SIEMENS DI	143 SIEMENS	15	24.000	3.706	15.44	13.000	35.000		
2020-004	1 Generic Me	165 VITROS	63 KODAK	59	20.237	1.844	9.11	15.000	26.000		
2020-004	1 Generic Me	184 INTEGRA	93 ROCHE	13	19.769	2.081	10.53	14.000	26.000		
2020-004	1 Generic Me	186 BECK. COUL	76 AU-BECKMAN	20	20.750	2.681	12.92	13.000	29.000		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
	2020-004	1	Generic Me	203 ROCHE	93	ROCHE	15	19.733	2.932	14.86	11.000	29.000
	2020-004	1	Generic Me	212 S. D. EXL	143	SIEMENS	19	22.842	2.621	11.47	15.000	31.000
	2020-004	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	11	21.727	1.656	7.62	17.000	27.000
	2020-004	2	ENZYMATIC	165 VITROS	63	KODAK	90	20.398	1.662	8.15	15.000	25.000
	2020-004	2	ENZYMATIC	203 ROCHE	93	ROCHE	12	20.000	1.472	7.36	16.000	24.000
	2020-004	2	ENZYMATIC	212 S. D. EXL	143	SIEMENS	33	22.688	1.810	7.98	17.000	28.000
169 CO2			CLIA TEST ID #: 0			EVALUATION CRITERIA: TARGET VALUE 3 SD						
	2020-005	1	Generic Me	38 SIEMENS DI	143	SIEMENS	15	15.267	3.924	25.70	3.000	27.000
	2020-005	1	Generic Me	165 VITROS	63	KODAK	59	12.102	1.189	9.82	9.000	16.000
	2020-005	1	Generic Me	184 INTEGRA	93	ROCHE	13	11.846	1.875	15.83	6.000	17.000
	2020-005	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	20	11.500	1.565	13.61	7.000	16.000
	2020-005	1	Generic Me	203 ROCHE	93	ROCHE	15	11.133	1.543	13.86	7.000	16.000
	2020-005	1	Generic Me	212 S. D. EXL	143	SIEMENS	19	13.316	2.363	17.75	6.000	20.000
	2020-005	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	11	12.909	0.900	6.97	10.000	16.000
	2020-005	2	ENZYMATIC	165 VITROS	63	KODAK	90	11.911	1.371	11.51	8.000	16.000
	2020-005	2	ENZYMATIC	203 ROCHE	93	ROCHE	12	11.667	1.434	12.29	7.000	16.000
	2020-005	2	ENZYMATIC	212 S. D. EXL	143	SIEMENS	33	13.625	1.932	14.18	8.000	19.000
170 DIRECT BILIRUBIN						EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}						
	2020-001	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	16	0.675	0.075	11.11	0.500	0.900
	2020-001	4	JEND.-GROG	212 S. D. EXL	143	SIEMENS	23	0.687	0.045	6.52	0.500	0.800
	2020-001	11	SPECT. WIT	165 VITROS	63	KODAK	56	0.473	0.119	25.10	0.100	0.800
	2020-001	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.700	0.038	5.40	0.600	0.800
	2020-001	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	0.683	0.037	5.45	0.500	0.800
	2020-001	17	SIEMENS JE	212 S. D. EXL	143	SIEMENS	12	0.725	0.159	21.90	0.200	1.200
	2020-002	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	16	0.973	0.044	4.54	0.800	1.200
	2020-002	4	JEND.-GROG	212 S. D. EXL	143	SIEMENS	23	0.991	0.058	5.88	0.800	1.200
	2020-002	11	SPECT. WIT	165 VITROS	63	KODAK	56	0.904	0.116	12.89	0.600	1.300
	2020-002	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.993	0.046	4.61	0.800	1.200
	2020-002	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	0.983	0.055	5.62	0.800	1.200

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-002	17	SIEMENS JE	212	S. D. EXL	143	SIEMENS	12	0.991	0.090	9.08	0.700	1.300	
170	DIRECT BILIRUBIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}								
2020-003	4	JEND.-GROG	38	SIEMENS DI	143	SIEMENS	16	0.231	0.046	20.04	0.100	0.400	
2020-003	4	JEND.-GROG	212	S. D. EXL	143	SIEMENS	23	0.252	0.050	19.81	0.100	0.400	
2020-003	11	SPECT. WIT	165	VITROS	63	KODAK	56	0.030	0.081	272.72	0	0.300	
2020-003	13	AZOBIL.W/D	203	ROCHE	93	ROCHE	15	0.280	0.040	14.29	0.200	0.400	
2020-003	16	DIAZO ENDP	215	B.C DXC	134	BECKM.COUL	12	0.275	0.043	15.75	0.100	0.400	
2020-003	17	SIEMENS JE	212	S. D. EXL	143	SIEMENS	12	0.255	0.050	19.56	0.100	0.400	
2020-004	4	JEND.-GROG	38	SIEMENS DI	143	SIEMENS	16	1.107	0.044	4.00	0.900	1.300	
2020-004	4	JEND.-GROG	212	S. D. EXL	143	SIEMENS	23	1.135	0.070	6.15	0.900	1.400	
2020-004	11	SPECT. WIT	165	VITROS	63	KODAK	56	1.064	0.118	11.10	0.700	1.400	
2020-004	13	AZOBIL.W/D	203	ROCHE	93	ROCHE	15	1.064	0.139	13.10	0.600	1.500	
2020-004	16	DIAZO ENDP	215	B.C DXC	134	BECKM.COUL	12	1.117	0.055	4.95	0.900	1.300	
2020-004	17	SIEMENS JE	212	S. D. EXL	143	SIEMENS	12	1.125	0.116	10.34	0.800	1.500	
2020-005	4	JEND.-GROG	38	SIEMENS DI	143	SIEMENS	16	0.075	0.043	57.74	0	0.200	
2020-005	4	JEND.-GROG	212	S. D. EXL	143	SIEMENS	23	0.057	0.050	87.71	0	0.200	
2020-005	11	SPECT. WIT	165	VITROS	63	KODAK	56	0.000	0.000	0.00	0.000	0.000	
2020-005	13	AZOBIL.W/D	203	ROCHE	93	ROCHE	15	0.187	0.034	18.21	0.100	0.300	
2020-005	16	DIAZO ENDP	215	B.C DXC	134	BECKM.COUL	12	0.067	0.047	70.71	0	0.200	
2020-005	17	SIEMENS JE	212	S. D. EXL	143	SIEMENS	12	0.075	0.083	110.55	0	0.300	

SPEC: 10 CHEMISTRY		SUB: 20 BLOOD GASES				GRP: 30 BLOOD GASES							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335		EVALUATION CRITERIA: TARGET VALUE 8 % OR 5.00 {GREATER}								
2020-051	260	SIEMENS					11	77.245	2.638	3.42	71.100	83.400	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-051	265	ROCHE COBA		29	80.028	1.976	2.47	73.600	86.400	
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE 8 % OR 5.00 {GREATER}						
2020-052	260	SIEMENS		11	22.155	1.054	4.76	17.200	27.200	
2020-052	265	ROCHE COBA		29	24.021	0.591	2.46	19.000	29.000	
2020-053	260	SIEMENS		11	44.864	1.592	3.55	39.900	49.900	
2020-053	265	ROCHE COBA		29	48.486	0.785	1.62	43.500	53.500	
2020-054	260	SIEMENS		11	22.036	0.908	4.12	17.000	27.000	
2020-054	265	ROCHE COBA		29	23.862	0.432	1.81	18.900	28.900	
2020-055	260	SIEMENS		11	44.727	2.065	4.62	39.700	49.700	
2020-055	265	ROCHE COBA		29	48.304	0.983	2.03	43.300	53.300	
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE 0.04						
2020-051	260	SIEMENS		11	7.127	0.004	0.05	7.087	7.167	
2020-051	265	ROCHE COBA		29	7.115	0.007	0.10	7.075	7.155	
2020-052	260	SIEMENS		11	7.661	0.008	0.10	7.621	7.701	
2020-052	265	ROCHE COBA		29	7.612	0.005	0.06	7.572	7.652	
2020-053	260	SIEMENS		11	7.379	0.005	0.07	7.340	7.420	
2020-053	265	ROCHE COBA		29	7.350	0.006	0.08	7.310	7.390	
2020-054	260	SIEMENS		11	7.660	0.007	0.09	7.620	7.700	
2020-054	265	ROCHE COBA		29	7.613	0.005	0.06	7.573	7.653	
2020-055	260	SIEMENS		11	7.380	0.006	0.08	7.340	7.420	
2020-055	265	ROCHE COBA		29	7.350	0.006	0.08	7.310	7.390	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 20 BLOOD GASES		GRP: 30 BLOOD GASES						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
009	BLOOD GASES-pO2	CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-051	260 SIEMENS			11	95.536	4.213	4.41	82.900	108.200	
2020-051	265 ROCHE COBA			29	92.234	11.896	12.90	56.500	127.900	
2020-052	260 SIEMENS			11	148.055	10.383	7.01	116.900	179.200	
2020-052	265 ROCHE COBA			29	153.938	4.012	2.61	141.900	166.000	
2020-053	260 SIEMENS			11	125.773	4.621	3.67	111.900	139.600	
2020-053	265 ROCHE COBA			29	132.239	3.042	2.30	123.100	141.400	
2020-054	260 SIEMENS			11	148.555	4.390	2.96	135.400	161.700	
2020-054	265 ROCHE COBA			29	153.069	4.214	2.75	140.400	165.700	
2020-055	260 SIEMENS			11	127.355	5.176	4.06	111.800	142.900	
2020-055	265 ROCHE COBA			29	129.600	4.611	3.56	115.800	143.400	
<hr/>										
SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0								
2020-031	115 VITROS ECI			10	17.110	0.699	4.09	15.000	19.200	
2020-031	116 ROCHE COBA			17	22.053	1.461	6.62	17.700	26.400	
2020-031	123 ADVIA CEN.			10	23.930	1.327	5.54	19.900	27.900	
2020-031	125 B.C.ACCESS			11	23.545	2.080	8.83	17.300	29.800	
2020-032	115 VITROS ECI			10	6.350	0.233	3.68	5.600	7.100	
2020-032	116 ROCHE COBA			17	8.788	1.222	13.90	5.100	12.500	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-032	123	ADVIA CEN.		10	9.300	0.657	7.07	7.300	11.300
2020-032	125	B.C.ACCESS		11	9.082	0.785	8.64	6.700	11.400
029	FOLL. STIM. HOR.-FSH		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD				
2020-033	115	VITROS ECI		10	10.530	0.326	3.09	9.600	11.500
2020-033	116	ROCHE COBA		17	14.194	1.278	9.00	10.400	18.000
2020-033	123	ADVIA CEN.		10	15.540	1.036	6.66	12.400	18.600
2020-033	125	B.C.ACCESS		11	15.236	1.063	6.98	12.000	18.400
2020-034	115	VITROS ECI		10	22.080	0.792	3.59	19.700	24.500
2020-034	116	ROCHE COBA		17	28.259	1.666	5.90	23.300	33.300
2020-034	123	ADVIA CEN.		10	30.890	2.550	8.26	23.200	38.500
2020-034	125	B.C.ACCESS		11	31.064	2.258	7.27	24.300	37.800
2020-035	115	VITROS ECI		10	24.430	0.870	3.56	21.800	27.000
2020-035	116	ROCHE COBA		17	31.224	1.893	6.06	25.500	36.900
2020-035	123	ADVIA CEN.		10	33.400	3.632	10.88	22.500	44.300
2020-035	125	B.C.ACCESS		11	33.727	2.200	6.52	27.100	40.300
030	FREE THYROXINE		CLIA TEST ID #: 545						
2020-031	115	VITROS ECI		33	4.631	0.161	3.47	4.100	5.100
2020-031	116	ROCHE COBA		44	2.133	0.134	6.29	1.700	2.500
2020-031	125	B.C.ACCESS		41	1.810	0.090	4.97	1.500	2.100
2020-031	149	ARC i2000s		20	1.755	0.086	4.93	1.500	2.000
2020-031	160	SIEMENS EX		40	2.068	0.113	5.45	1.700	2.400
2020-032	115	VITROS ECI		33	1.458	0.095	6.55	1.200	1.700
2020-032	116	ROCHE COBA		44	0.860	0.069	7.98	0.700	1.100
2020-032	125	B.C.ACCESS		41	0.833	0.069	8.23	0.600	1.000
2020-032	149	ARC i2000s		20	0.700	0.000	0.00	0.700	0.700
2020-032	160	SIEMENS EX		40	0.903	0.057	6.31	0.700	1.100

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
030	FREE THYROXINE	CLIA TEST ID #: 545	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-033	115 VITROS ECI		33	2.809	0.142	5.06	2.400	3.200	
2020-033	116 ROCHE COBA		44	1.400	0.103	7.39	1.100	1.700	
2020-033	125 B.C.ACCESS		41	1.241	0.081	6.51	1.000	1.500	
2020-033	149 ARC i2000s		20	1.100	0.046	4.17	1.000	1.200	
2020-033	160 SIEMENS EX		40	1.355	0.080	5.94	1.100	1.600	
2020-034	115 VITROS ECI		33	5.730	0.204	3.56	5.100	6.300	
2020-034	116 ROCHE COBA		44	2.686	0.176	6.55	2.200	3.200	
2020-034	125 B.C.ACCESS		41	2.160	0.148	6.85	1.700	2.600	
2020-034	149 ARC i2000s		20	2.245	0.150	6.68	1.800	2.700	
2020-034	160 SIEMENS EX		40	2.590	0.132	5.09	2.200	3.000	
2020-035	115 VITROS ECI		33	6.127	0.158	2.58	5.700	6.600	
2020-035	116 ROCHE COBA		44	2.943	0.161	5.46	2.500	3.400	
2020-035	125 B.C.ACCESS		41	2.350	0.145	6.17	1.900	2.800	
2020-035	149 ARC i2000s		20	2.395	0.150	6.26	1.900	2.800	
2020-035	160 SIEMENS EX		40	2.825	0.122	4.32	2.500	3.200	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555							
2020-031	115 VITROS ECI		17	341.912	22.209	6.50	275.300	408.500	
2020-031	116 ROCHE COBA		20	315.160	15.677	4.97	268.100	362.200	
2020-031	125 B.C.ACCESS		11	327.227	12.733	3.89	289.000	365.400	
2020-032	115 VITROS ECI		17	2.400	0.000	0.00	2.400	2.400	
2020-032	116 ROCHE COBA		20	1.740	0.111	6.40	1.400	2.100	
2020-032	125 B.C.ACCESS		11	2.636	0.219	8.29	2.000	3.300	
2020-033	115 VITROS ECI		17	128.924	9.148	7.10	101.500	156.400	
2020-033	116 ROCHE COBA		20	118.260	6.676	5.65	98.200	138.300	
2020-033	125 B.C.ACCESS		11	125.327	5.309	4.24	109.400	141.300	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-034	115 VITROS ECI			17	509.006	33.001	6.48	410.000	608.000	
2020-034	116 ROCHE COBA			20	472.185	22.171	4.70	405.700	538.700	
2020-034	125 B.C.ACCESS			11	479.791	18.603	3.88	424.000	535.600	
2020-035	115 VITROS ECI			17	587.041	39.134	6.67	469.600	704.400	
2020-035	116 ROCHE COBA			20	532.395	54.409	10.22	369.200	695.600	
2020-035	125 B.C.ACCESS			11	550.473	23.002	4.18	481.500	619.500	
034	PROGESTERONE	CLIA TEST ID #: 0								
2020-031	116 ROCHE COBA			11	15.591	0.754	4.83	13.300	17.900	
2020-032	116 ROCHE COBA			11	0.500	0.095	19.07	0.200	0.800	
2020-033	116 ROCHE COBA			11	6.955	0.403	5.80	5.700	8.200	
2020-034	116 ROCHE COBA			11	21.718	0.695	3.20	19.600	23.800	
2020-035	116 ROCHE COBA			11	24.009	0.854	3.56	21.400	26.600	
035	PROLACTIN									
2020-031	116 ROCHE COBA			18	15.617	1.036	6.63	12.500	18.700	
2020-031	125 B.C.ACCESS			11	15.191	0.846	5.57	12.700	17.700	
2020-032	116 ROCHE COBA			18	3.383	0.177	5.24	2.900	3.900	
2020-032	125 B.C.ACCESS			11	2.773	0.200	7.23	2.200	3.400	
2020-033	116 ROCHE COBA			18	8.022	0.438	5.46	6.700	9.300	
2020-033	125 B.C.ACCESS			11	7.536	0.437	5.81	6.200	8.800	
2020-034	116 ROCHE COBA			18	21.967	1.075	4.90	18.700	25.200	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-034	125	B.C.ACCESS		11	20.555	1.164	5.66	17.100	24.000	
035	PROLACTIN		CLIA TEST ID #: 0							EVALUATION CRITERIA: TARGET VALUE 3 SD
2020-035	116	ROCHE COBA		18	24.967	1.412	5.65	20.700	29.200	
2020-035	125	B.C.ACCESS		11	22.955	1.160	5.05	19.500	26.400	
037	TESTOSTERONE									
2020-031	116	ROCHE COBA		11	144.900	290.915	200.77	0	1017.600	
2020-032	116	ROCHE COBA		11	19.000	38.275	201.45	0	133.800	
2020-033	116	ROCHE COBA		11	72.100	144.738	200.75	0	506.300	
2020-034	116	ROCHE COBA		11	192.100	385.596	200.73	0	1348.900	
2020-035	116	ROCHE COBA		11	213.891	429.976	201.03	0	1503.800	
039	THYRO. STI. HOR.-TSH		CLIA TEST ID #: 585							
2020-031	106	TOSOH AIA		13	5.962	0.560	9.39	4.300	7.600	
2020-031	115	VITROS ECI		38	4.947	0.209	4.22	4.300	5.600	
2020-031	116	ROCHE COBA		54	4.091	0.169	4.12	3.600	4.600	
2020-031	125	B.C.ACCESS		59	3.588	0.215	5.99	2.900	4.200	
2020-031	149	ARC i2000s		21	3.638	0.210	5.78	3.000	4.300	
2020-031	160	SIEMENS EX		67	3.467	0.219	6.33	2.800	4.100	
2020-032	106	TOSOH AIA		13	0.715	0.095	13.26	0.400	1.000	
2020-032	115	VITROS ECI		38	0.674	0.044	6.54	0.500	0.800	
2020-032	116	ROCHE COBA		54	0.843	0.053	6.31	0.700	1.000	
2020-032	125	B.C.ACCESS		59	0.617	0.042	6.77	0.500	0.700	
2020-032	149	ARC i2000s		21	0.586	0.047	7.97	0.400	0.700	
2020-032	160	SIEMENS EX		67	0.657	0.058	8.81	0.500	0.800	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-033	106 TOSOH AIA			13	2.746	0.206	7.51	2.100	3.400
2020-033	115 VITROS ECI			38	2.437	0.093	3.82	2.200	2.700
2020-033	116 ROCHE COBA			54	2.168	0.097	4.46	1.900	2.500
2020-033	125 B.C.ACCESS			59	1.753	0.106	6.07	1.400	2.100
2020-033	149 ARC i2000s			21	1.800	0.148	8.24	1.400	2.200
2020-033	160 SIEMENS EX			67	1.788	0.125	7.00	1.400	2.200
2020-034	106 TOSOH AIA			13	8.146	0.911	11.18	5.400	10.900
2020-034	115 VITROS ECI			38	6.968	0.304	4.36	6.100	7.900
2020-034	116 ROCHE COBA			54	5.485	0.280	5.10	4.600	6.300
2020-034	125 B.C.ACCESS			59	5.024	0.322	6.41	4.100	6.000
2020-034	149 ARC i2000s			21	5.076	0.326	6.43	4.100	6.100
2020-034	160 SIEMENS EX			67	4.778	0.255	5.34	4.000	5.500
2020-035	106 TOSOH AIA			13	9.338	0.646	6.92	7.400	11.300
2020-035	115 VITROS ECI			38	7.866	0.359	4.57	6.800	8.900
2020-035	116 ROCHE COBA			54	6.119	0.295	4.82	5.200	7.000
2020-035	125 B.C.ACCESS			59	5.742	0.353	6.15	4.700	6.800
2020-035	149 ARC i2000s			21	5.724	0.346	6.05	4.700	6.800
2020-035	160 SIEMENS EX			67	5.355	0.324	6.05	4.400	6.300
040	THYROXINE (T4)	CLIA TEST ID #: 595	EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}						
2020-031	32 DADE DIME.			26	9.858	0.577	5.86	7.900	11.800
2020-031	106 TOSOH AIA			10	8.360	0.721	8.63	6.700	10.000
2020-031	115 VITROS ECI			33	8.518	0.390	4.58	6.800	10.200
2020-031	116 ROCHE COBA			51	8.535	0.483	5.65	6.800	10.200
2020-031	125 B.C.ACCESS			50	8.424	0.652	7.74	6.700	10.100
2020-031	149 ARC i2000s			18	8.506	0.361	4.25	6.800	10.200
2020-031	160 SIEMENS EX			30	9.837	0.768	7.81	7.900	11.800
2020-031	162 Siemens Di			10	9.780	0.729	7.46	7.800	11.700

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
040 THYROXINE (T4)		CLIA TEST ID #: 595		EVALUATION CRITERIA: TARGET VALUE	20 % OR 1.00	{GREATER}		
2020-032	32	DADE DIME.	26	3.423	0.383	11.18	2.400	4.400
2020-032	106	TOSOH AIA	10	3.250	0.301	9.26	2.300	4.300
2020-032	115	VITROS ECI	33	3.300	0.135	4.09	2.300	4.300
2020-032	116	ROCHE COBA	51	3.392	0.190	5.59	2.400	4.400
2020-032	125	B.C.ACCESS	50	3.090	0.331	10.72	2.100	4.100
2020-032	149	ARC i2000s	18	2.767	0.219	7.90	1.800	3.800
2020-032	160	SIEMENS EX	30	3.537	0.492	13.92	2.500	4.500
2020-032	162	Siemens Di	10	3.430	0.496	14.46	2.400	4.400
2020-033	32	DADE DIME.	26	5.931	0.415	7.00	4.700	7.100
2020-033	106	TOSOH AIA	10	5.500	0.415	7.54	4.400	6.600
2020-033	115	VITROS ECI	33	5.494	0.271	4.93	4.400	6.600
2020-033	116	ROCHE COBA	51	5.457	0.306	5.61	4.400	6.500
2020-033	125	B.C.ACCESS	50	5.464	0.461	8.43	4.400	6.600
2020-033	149	ARC i2000s	18	5.024	0.290	5.77	4.000	6.000
2020-033	160	SIEMENS EX	30	5.997	0.586	9.78	4.800	7.200
2020-033	162	Siemens Di	10	5.930	0.398	6.71	4.700	7.100
2020-034	32	DADE DIME.	26	12.962	0.488	3.77	10.400	15.600
2020-034	106	TOSOH AIA	10	10.460	0.822	7.86	8.400	12.600
2020-034	115	VITROS ECI	33	10.909	0.592	5.43	8.700	13.100
2020-034	116	ROCHE COBA	51	11.040	0.458	4.15	8.800	13.200
2020-034	125	B.C.ACCESS	50	10.468	0.680	6.49	8.400	12.600
2020-034	149	ARC i2000s	18	11.417	0.669	5.86	9.100	13.700
2020-034	160	SIEMENS EX	30	12.940	0.904	6.99	10.400	15.500
2020-034	162	Siemens Di	10	12.800	0.854	6.68	10.200	15.400
2020-035	32	DADE DIME.	26	14.477	0.542	3.74	11.600	17.400
2020-035	106	TOSOH AIA	10	11.400	0.823	7.22	9.100	13.700

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-035	115	VITROS ECI		33	11.945	0.702	5.88	9.600	14.300	
2020-035	116	ROCHE COBA		51	12.155	0.572	4.71	9.700	14.600	
2020-035	125	B.C.ACCESS		50	11.255	0.712	6.33	9.000	13.500	
2020-035	149	ARC i2000s		18	12.750	0.876	6.87	10.200	15.300	
2020-035	160	SIEMENS EX		30	14.360	0.845	5.88	11.500	17.200	
2020-035	162	Siemens Di		10	14.210	0.955	6.72	11.400	17.100	
139	ESTRADIOL		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-031	116	ROCHE COBA		10	195.200	7.662	3.93	146.400	244.000	
2020-032	116	ROCHE COBA		10	31.990	4.724	14.77	24.000	40.000	
2020-033	116	ROCHE COBA		10	113.720	4.960	4.36	85.300	142.200	
2020-034	116	ROCHE COBA		10	245.360	11.634	4.74	184.000	306.700	
2020-035	116	ROCHE COBA		10	267.680	6.368	2.38	200.800	334.600	
141	FOLATE									
2020-031	115	VITROS ECI		12	5.100	0.460	9.02	3.800	6.400	
2020-031	116	ROCHE COBA		18	4.483	0.532	11.86	3.400	5.600	
2020-031	125	B.C.ACCESS		13	5.869	0.300	5.11	4.400	7.300	
2020-031	149	ARC i2000s		10	6.520	0.477	7.32	4.900	8.200	
2020-031	160	SIEMENS EX		10	3.680	0.379	10.30	2.800	4.600	
2020-031	160	SIEMENS EX		10	3.000	0.000	0.00	2.200	3.800	
2020-032	115	VITROS ECI		12	1.250	0.166	13.27	0.900	1.600	
2020-032	116	ROCHE COBA		18	2.000	0.000	0.00	1.500	2.500	
2020-032	125	B.C.ACCESS		13	0.685	0.257	37.50	0.500	0.900	
2020-032	149	ARC i2000s		10	1.030	0.644	62.48	0.800	1.300	
2020-033	115	VITROS ECI		12	2.383	0.182	7.63	1.800	3.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-033	116	ROCHE COBA		18	2.024	0.081	3.99	1.500	2.500
2020-033	125	B.C.ACCESS		13	2.654	0.224	8.44	2.000	3.300
2020-033	149	ARC i2000s		10	2.780	0.346	12.44	2.100	3.500
141	FOLATE		CLIA TEST ID #: 0						
									EVALUATION CRITERIA: TARGET VALUE 25 %
2020-034	115	VITROS ECI		12	8.008	0.817	10.20	6.000	10.000
2020-034	116	ROCHE COBA		18	7.094	0.731	10.30	5.300	8.900
2020-034	125	B.C.ACCESS		13	8.554	0.440	5.14	6.400	10.700
2020-034	149	ARC i2000s		10	10.150	0.984	9.70	7.600	12.700
2020-035	115	VITROS ECI		12	9.750	1.127	11.56	7.300	12.200
2020-035	116	ROCHE COBA		18	8.294	0.987	11.90	6.200	10.400
2020-035	125	B.C.ACCESS		13	9.731	0.680	6.99	7.300	12.200
2020-035	149	ARC i2000s		10	11.420	1.053	9.22	8.600	14.300
142	hLH (HUMAN LUT. HOR)								
2020-031	116	ROCHE COBA		17	27.682	1.349	4.87	20.800	34.600
2020-031	123	ADVIA CEN.		10	24.810	1.488	6.00	18.600	31.000
2020-031	125	B.C.ACCESS		11	19.736	1.151	5.83	14.800	24.700
2020-032	116	ROCHE COBA		17	6.512	0.377	5.79	4.900	8.100
2020-032	123	ADVIA CEN.		10	4.690	0.164	3.50	3.500	5.900
2020-032	125	B.C.ACCESS		11	4.727	0.191	4.04	3.500	5.900
2020-033	116	ROCHE COBA		17	15.276	0.757	4.96	11.500	19.100
2020-033	123	ADVIA CEN.		10	11.630	0.885	7.61	8.700	14.500
2020-033	125	B.C.ACCESS		11	10.873	0.564	5.19	8.200	13.600
2020-034	116	ROCHE COBA		17	37.288	2.051	5.50	28.000	46.600
2020-034	123	ADVIA CEN.		10	32.700	2.362	7.22	24.500	40.900
2020-034	125	B.C.ACCESS		11	25.945	1.036	3.99	19.500	32.400

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
142 hLH (HUMAN LUT. HOR)										
			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %						
	2020-035	116	ROCHE COBA		17	41.176	2.000	4.86	30.900	51.500
	2020-035	123	ADVIA CEN.		10	36.040	2.405	6.67	27.000	45.100
	2020-035	125	B.C.ACCESS		11	28.136	1.499	5.33	21.100	35.200

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)								
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFeree LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
031 hCG-QUALITATIVE												
2020-031	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100
2020-031	142	INSTANT VI		002	002	002	002	002 - POSITIVE	48	48	48	100
2020-031	150	AIM STEP		002	002	002	002	002 - POSITIVE	84	84	84	100
2020-031	164	HCG Alere		002	002	002	002	002 - POSITIVE	64	64	64	100
2020-031	165	Osom HCG C		002	002	002	002	002 - POSITIVE	18	18	18	100
2020-031	167	Tanner Sci		002	002	002	002	002 - POSITIVE	17	17	17	100
2020-032	113	BIO-SIGN		001	001	001	001	001 - NEGATIVE	21	21	21	100
2020-032	142	INSTANT VI		001	001	001	001	001 - NEGATIVE	48	47	47	100
2020-032	142	INSTANT VI		001	001	001	001	002 - POSITIVE	48	1	47	0
2020-032	150	AIM STEP		001	001	001	001	001 - NEGATIVE	84	81	81	100
2020-032	150	AIM STEP		001	001	001	001	002 - POSITIVE	84	3	81	0
2020-032	164	HCG Alere		001	001	001	001	001 - NEGATIVE	64	64	64	100
2020-032	165	Osom HCG C		001	001	001	001	001 - NEGATIVE	18	18	18	100
2020-032	167	Tanner Sci		001	001	001	001	001 - NEGATIVE	17	17	17	100
2020-033	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100
2020-033	142	INSTANT VI		002	002	002	002	002 - POSITIVE	48	48	48	100
2020-033	150	AIM STEP		002	002	002	002	002 - POSITIVE	84	84	84	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-033	164	HCG Alere		002 002	002 002		002 - POSITIVE	64	64	64	100	
2020-033	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-033	167	Tanner Sci		002 002	002 002		002 - POSITIVE	17	17	17	100	
031 hCG-QUALITATIVE												
2020-034	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-034	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-034	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	84	84	100	
2020-034	164	HCG Alere		002 002	002 002		002 - POSITIVE	64	64	64	100	
2020-034	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-034	167	Tanner Sci		002 002	002 002		002 - POSITIVE	17	17	17	100	
2020-035	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-035	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-035	150	AIM STEP		002 002	002 002		001 - NEGATIVE	84	1	83	0	
2020-035	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	83	83	100	
2020-035	164	HCG Alere		002 002	002 002		002 - POSITIVE	64	64	64	100	
2020-035	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-035	167	Tanner Sci		002 002	002 002		002 - POSITIVE	17	17	17	100	
SPE: 10 CHEMISTRY												
SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY								
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2020-021	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-021	083	STAMBIO		002 002	002 002		002 - POSITIVE	12	12	12	100	
2020-021	112	ONE STEP D		002 002	002 002		002 - POSITIVE	14	14	14	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 45 ENDOCRINOL.-hCG ONLY

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-021	114	INSTANT VI		002 002	002 002		002 - POSITIVE	122	122	122	100
2020-021	117	TECO DIAG.		002 002	002 002		002 - POSITIVE	10	10	10	100
2020-021	118	AIM STEP		002 002	002 002		002 - POSITIVE	146	146	146	100
2020-021	130	ALERE HCG		002 002	002 002		002 - POSITIVE	88	88	88	100
2020-021	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100
2020-021	133	FaStep		002 002	002 002		002 - POSITIVE	13	13	13	100
2020-021	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	63	63	63	100
151 hCG ONLY QUALITATIVE											
2020-022	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100
2020-022	083	STAMBIO		002 002	002 002		002 - POSITIVE	12	12	12	100
2020-022	112	ONE STEP D		002 002	002 002		002 - POSITIVE	14	14	14	100
2020-022	114	INSTANT VI		002 002	002 002		001 - NEGATIVE	122	1	121	0
2020-022	114	INSTANT VI		002 002	002 002		002 - POSITIVE	122	121	121	100
2020-022	117	TECO DIAG.		002 002	002 002		002 - POSITIVE	10	10	10	100
2020-022	118	AIM STEP		002 002	002 002		002 - POSITIVE	146	146	146	100
2020-022	130	ALERE HCG		002 002	002 002		002 - POSITIVE	88	88	88	100
2020-022	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100
2020-022	133	FaStep		002 002	002 002		002 - POSITIVE	13	13	13	100
2020-022	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	63	63	63	100
2020-023	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100
2020-023	083	STAMBIO		002 002	002 002		002 - POSITIVE	12	12	12	100
2020-023	112	ONE STEP D		002 002	002 002		002 - POSITIVE	14	14	14	100
2020-023	114	INSTANT VI		002 002	002 002		002 - POSITIVE	122	122	122	100
2020-023	117	TECO DIAG.		002 002	002 002		002 - POSITIVE	10	10	10	100
2020-023	118	AIM STEP		002 002	002 002		002 - POSITIVE	146	146	146	100
2020-023	130	ALERE HCG		002 002	002 002		002 - POSITIVE	88	88	88	100
2020-023	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100
2020-023	133	FaStep		002 002	002 002		002 - POSITIVE	13	13	13	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY													SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score									
2020-023	134	Tanner Sc.		002	002	002	002	002	-	POSITIVE	63	63	63	100						
151 hCG ONLY QUALITATIVE																				
2020-024	018	BIO SIGN		001	001	001	001	001	-	NEGATIVE	21	21	21	100						
2020-024	083	STAMBIO		001	001	001	001	001	-	NEGATIVE	12	12	12	100						
2020-024	112	ONE STEP D		001	001	001	001	001	-	NEGATIVE	14	14	14	100						
2020-024	114	INSTANT VI		001	001	001	001	001	-	NEGATIVE	122	121	121	100						
2020-024	114	INSTANT VI		001	001	001	001	001	002	POSITIVE	122	1	121	0						
2020-024	117	TECO DIAG.		001	001	001	001	001	-	NEGATIVE	10	10	10	100						
2020-024	118	AIM STEP		001	001	001	001	001	-	NEGATIVE	146	146	146	100						
2020-024	130	ALERE HCG		001	001	001	001	001	-	NEGATIVE	88	88	88	100						
2020-024	131	OSOM HCG		001	001	001	001	001	-	NEGATIVE	25	25	25	100						
2020-024	133	FaStep		001	001	001	001	001	-	NEGATIVE	13	13	13	100						
2020-024	134	Tanner Sc.		001	001	001	001	001	-	NEGATIVE	63	63	63	100						
2020-025	018	BIO SIGN		001	001	001	001	001	-	NEGATIVE	21	21	21	100						
2020-025	083	STAMBIO		001	001	001	001	001	-	NEGATIVE	12	12	12	100						
2020-025	112	ONE STEP D		001	001	001	001	001	-	NEGATIVE	14	14	14	100						
2020-025	114	INSTANT VI		001	001	001	001	001	-	NEGATIVE	122	121	121	100						
2020-025	114	INSTANT VI		001	001	001	001	001	002	POSITIVE	122	1	121	0						
2020-025	117	TECO DIAG.		001	001	001	001	001	-	NEGATIVE	10	10	10	100						
2020-025	118	AIM STEP		001	001	001	001	001	-	NEGATIVE	146	146	146	100						
2020-025	130	ALERE HCG		001	001	001	001	001	-	NEGATIVE	88	88	88	100						
2020-025	131	OSOM HCG		001	001	001	001	001	-	NEGATIVE	25	25	25	100						
2020-025	133	FaStep		001	001	001	001	001	-	NEGATIVE	13	13	13	100						
2020-025	134	Tanner Sc.		001	001	001	001	001	-	NEGATIVE	63	63	63	100						

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 55 ENDO.(T3 UPTAKE)

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
173 T3 UPTAKE		CLIA TEST ID #: 565	EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-031	1 %U	84 VITROS ECI	1 25-35	26	51.272	1.878	3.66	45.640	56.900		
2020-031	1 %U	91 B.C.ACCESS	2 35-45	27	48.196	2.461	5.11	40.810	55.580		
2020-031	2 TUV	95 Arc i2000s	3 0.7 - 1.2	13	0.917	0.051	5.55	0.760	1.070		
2020-031	4 %	32 DIMENSION	1 25-35	10	40.282	1.516	3.76	35.730	44.830		
2020-031	4 %	32 DIMENSION	2 35-45	40	40.973	1.611	3.93	36.140	45.810		
2020-031	4 %	91 B.C.ACCESS	2 35-45	10	47.610	4.670	9.81	33.600	61.620		
2020-031	5 TBI	32 DIMENSION	2 35-45	12	40.425	1.490	3.69	35.950	44.900		
2020-031	5 TBI	98 ROCHE COBA	4 0.8-1.3	47	0.737	0.061	8.31	0.550	0.920		
2020-032	1 %U	84 VITROS ECI	1 25-35	26	57.212	1.859	3.25	51.630	62.790		
2020-032	1 %U	91 B.C.ACCESS	2 35-45	27	52.933	1.824	3.44	47.460	58.400		
2020-032	2 TUV	95 Arc i2000s	3 0.7 - 1.2	13	0.627	0.032	5.17	0.530	0.720		
2020-032	4 %	32 DIMENSION	1 25-35	10	40.224	1.247	3.10	36.480	43.970		
2020-032	4 %	32 DIMENSION	2 35-45	40	41.230	1.857	4.50	35.660	46.800		
2020-032	4 %	91 B.C.ACCESS	2 35-45	10	52.060	2.141	4.11	45.640	58.480		
2020-032	5 TBI	32 DIMENSION	2 35-45	12	40.833	1.863	4.56	35.240	46.420		
2020-032	5 TBI	98 ROCHE COBA	4 0.8-1.3	47	0.872	0.049	5.67	0.720	1.020		
2020-033	1 %U	84 VITROS ECI	1 25-35	26	54.196	1.905	3.51	48.480	59.910		
2020-033	1 %U	91 B.C.ACCESS	2 35-45	27	50.963	2.044	4.01	44.830	57.100		
2020-033	2 TUV	95 Arc i2000s	3 0.7 - 1.2	13	0.758	0.046	6.12	0.620	0.900		
2020-033	4 %	32 DIMENSION	1 25-35	10	40.047	1.104	2.76	36.730	43.360		
2020-033	4 %	32 DIMENSION	2 35-45	40	40.743	1.875	4.60	35.120	46.370		
2020-033	4 %	91 B.C.ACCESS	2 35-45	10	50.750	2.049	4.04	44.600	56.900		
2020-033	5 TBI	32 DIMENSION	2 35-45	12	40.508	2.361	5.83	33.420	47.590		
2020-033	5 TBI	98 ROCHE COBA	4 0.8-1.3	47	0.821	0.053	6.51	0.660	0.980		
2020-034	1 %U	84 VITROS ECI	1 25-35	26	49.046	1.733	3.53	43.850	54.250		
2020-034	1 %U	91 B.C.ACCESS	2 35-45	27	47.915	1.569	3.28	43.210	52.620		

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-034	2 TUV	95	Arc i2000s	3	0.7 - 1.2	13	1.032	0.055	5.28	0.870	1.200	
2020-034	4 %	32	DIMENSION	1	25-35	10	40.366	1.100	2.72	37.070	43.670	
2020-034	4 %	32	DIMENSION	2	35-45	40	40.768	1.785	4.38	35.410	46.120	
2020-034	4 %	91	B.C.ACCESS	2	35-45	10	47.520	1.792	3.77	42.140	52.900	
2020-034	5 TBI	32	DIMENSION	2	35-45	12	40.683	2.199	5.41	34.090	47.280	
2020-034	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.683	0.059	8.69	0.500	0.860	
173 T3 UPTAKE		CLIA TEST ID #: 565				EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-035	1 %U	84	VITROS ECI	1	25-35	26	48.108	1.876	3.90	42.480	53.740	
2020-035	1 %U	91	B.C.ACCESS	2	35-45	27	47.148	2.219	4.71	40.490	53.810	
2020-035	2 TUV	95	Arc i2000s	3	0.7 - 1.2	13	1.072	0.056	5.20	0.900	1.240	
2020-035	4 %	32	DIMENSION	1	25-35	10	40.847	1.645	4.03	35.910	45.780	
2020-035	4 %	32	DIMENSION	2	35-45	40	40.868	1.901	4.65	35.160	46.570	
2020-035	4 %	91	B.C.ACCESS	2	35-45	10	46.430	2.366	5.10	39.330	53.530	
2020-035	5 TBI	32	DIMENSION	2	35-45	12	39.950	1.721	4.31	34.790	45.110	
2020-035	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.648	0.061	9.35	0.470	0.830	

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY				GRP: 60 TOXICOLOGY						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
043 DIGOXIN		CLIA TEST ID #: 645				EVALUATION CRITERIA: TARGET VALUE 20 % OR 0.20 {GREATER}						
2020-031	50 VITROS					14	0.921	0.086	9.33	0.700	1.100	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
043 DIGOXIN		CLIA TEST ID #: 645	EVALUATION CRITERIA: TARGET VALUE	20	% OR 0.20	{GREATER}				
2020-032	50 VITROS			14	0.500	0.065	13.09	0.300	0.700	
2020-033	50 VITROS			14	0.607	0.116	19.15	0.400	0.800	
2020-034	50 VITROS			14	1.179	0.115	9.72	0.900	1.400	
2020-035	50 VITROS			14	1.343	0.105	7.82	1.100	1.600	
045 FERRITIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%					
2020-031	50 VITROS			13	67.815	3.443	5.08	50.900	84.800	
2020-031	90 ROCHE COBA			15	113.047	6.607	5.84	84.800	141.300	
2020-032	50 VITROS			13	14.915	0.898	6.02	11.200	18.600	
2020-032	90 ROCHE COBA			15	23.307	1.413	6.06	17.500	29.100	
2020-033	50 VITROS			13	34.900	2.546	7.30	26.200	43.600	
2020-033	90 ROCHE COBA			15	58.473	3.112	5.32	43.900	73.100	
2020-034	50 VITROS			13	93.877	3.918	4.17	70.400	117.300	
2020-034	90 ROCHE COBA			15	161.080	7.634	4.74	120.800	201.400	
2020-035	50 VITROS			13	107.069	5.061	4.73	80.300	133.800	
2020-035	90 ROCHE COBA			15	180.100	10.674	5.93	135.100	225.100	
046 GENTAMICIN		CLIA TEST ID #: 665								
2020-031	50 VITROS			13	5.108	0.213	4.17	3.800	6.400	
2020-032	50 VITROS			13	0.600	0.000	0.00	0.500	0.800	
2020-033	50 VITROS			13	1.992	0.073	3.66	1.500	2.500	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
046	GENTAMICIN		CLIA TEST ID #: 665		EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-034	50 VITROS			13	7.715	0.293	3.80	5.800	9.600	
2020-035	50 VITROS			13	8.923	0.461	5.17	6.700	11.200	
050	PHENYTOIN		CLIA TEST ID #: 695							
2020-031	50 VITROS			15	11.580	0.524	4.53	8.700	14.500	
2020-031	92 ROCHE COBA			10	11.770	0.473	4.02	8.800	14.700	
2020-032	50 VITROS			15	3.620	2.702	74.64	2.700	4.500	
2020-032	92 ROCHE COBA			10	0.800	0.000	0.00	0.600	1.000	
2020-033	50 VITROS			15	3.873	0.211	5.45	2.900	4.800	
2020-033	92 ROCHE COBA			10	4.380	0.279	6.36	3.300	5.500	
2020-034	50 VITROS			15	17.607	0.857	4.87	13.200	22.000	
2020-034	92 ROCHE COBA			10	17.380	0.881	5.07	13.000	21.700	
2020-035	50 VITROS			15	20.473	0.827	4.04	15.400	25.600	
2020-035	92 ROCHE COBA			10	19.590	1.447	7.38	14.700	24.500	
054	THEOPHYLLINE		CLIA TEST ID #: 735							
2020-031	50 VITROS			13	23.523	0.585	2.49	17.600	29.400	
2020-032	50 VITROS			13	2.254	0.300	13.32	1.700	2.800	
2020-033	50 VITROS			13	10.800	0.413	3.83	8.100	13.500	
2020-034	50 VITROS			13	33.154	0.952	2.87	24.900	41.400	
2020-035	50 VITROS			13	37.000	1.030	2.78	27.800	46.300	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
127	VANCOMYCIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-031	50 VITROS			16	20.556	0.975	4.74	15.400	25.700	
2020-032	50 VITROS			16	5.000	0.000	0.00	3.800	6.300	
2020-033	50 VITROS			16	7.994	0.594	7.43	6.000	10.000	
2020-034	50 VITROS			16	31.294	1.389	4.44	23.500	39.100	
2020-035	50 VITROS			16	36.694	1.448	3.95	27.500	45.900	
SPEC: 10 CHEMISTRY				SUB: 50 URINALYSIS		GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
065	URINA-PROTEIN QUANT.				EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-061	58 VITROS			38	126.000	8.147	6.47	94.500	157.500	
2020-061	62 DADE DIM.			27	110.767	3.223	2.91	83.100	138.500	
2020-061	79 BECK. COUL			12	98.258	4.303	4.38	73.700	122.800	
2020-061	93 ARCHITECT			15	90.940	2.658	2.92	68.200	113.700	
2020-061	95 INTEGRA			13	86.554	10.788	12.46	64.900	108.200	
2020-061	113 ROCHE COBA			26	86.208	2.264	2.63	64.700	107.800	
2020-061	139 S. EXL.200			35	109.570	1.432	1.31	82.200	137.000	
2020-062	58 VITROS			38	9.216	0.576	6.25	4.200	14.200	
2020-062	62 DADE DIM.			27	24.244	1.362	5.62	18.200	30.300	
2020-062	79 BECK. COUL			12	5.567	2.262	40.64	0.600	10.600	
2020-062	93 ARCHITECT			15	7.213	1.229	17.04	2.200	12.200	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-062	95	INTEGRA		13	2.469	1.015	41.10	0	7.500	
2020-062	113	ROCHE COBA		26	3.808	0.581	15.26	0	8.800	
2020-062	139	S. EXL.200		35	24.154	2.034	8.42	18.100	30.200	
065 URINA-PROTEIN QUANT.			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-063	58	VITROS		38	99.079	5.783	5.84	74.300	123.800	
2020-063	62	DADE DIM.		27	100.969	1.376	1.36	75.700	126.200	
2020-063	79	BECK. COUL		12	93.008	4.159	4.47	69.800	116.300	
2020-063	93	ARCHITECT		15	89.067	2.910	3.27	66.800	111.300	
2020-063	95	INTEGRA		13	84.662	10.965	12.95	63.500	105.800	
2020-063	113	ROCHE COBA		26	84.723	2.558	3.02	63.500	105.900	
2020-063	139	S. EXL.200		35	101.349	2.372	2.34	76.000	126.700	
2020-064	58	VITROS		38	128.027	6.582	5.14	96.000	160.000	
2020-064	62	DADE DIM.		27	108.319	1.885	1.74	81.200	135.400	
2020-064	79	BECK. COUL		12	94.050	6.583	7.00	70.500	117.600	
2020-064	93	ARCHITECT		15	88.287	2.757	3.12	66.200	110.400	
2020-064	95	INTEGRA		13	85.485	10.065	11.77	64.100	106.900	
2020-064	113	ROCHE COBA		26	85.588	2.557	2.99	64.200	107.000	
2020-064	139	S. EXL.200		35	108.860	2.189	2.01	81.600	136.100	
2020-065	58	VITROS		38	7.703	0.513	6.66	2.700	12.700	
2020-065	62	DADE DIM.		27	15.978	1.218	7.62	11.000	21.000	
2020-065	79	BECK. COUL		12	5.350	0.439	8.20	0.400	10.400	
2020-065	93	ARCHITECT		15	7.620	1.818	23.86	2.600	12.600	
2020-065	95	INTEGRA		13	3.154	1.139	36.12	0	8.200	
2020-065	113	ROCHE COBA		26	4.017	0.493	12.27	0	9.000	
2020-065	139	S. EXL.200		35	15.820	1.646	10.41	10.800	20.800	
067 URI-SPECIFIC GRAVITY				EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-061	11	AMES NM10S		50	1.020	0.003	0.33	1.010	1.030	
2020-061	13	BAYER CLI		36	1.020	0.002	0.21	1.014	1.027	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-061	67	URISCAN	50	1.017	0.002	0.25	1.010	1.025
2020-061	71	ATLAS	21	1.026	0.002	0.17	1.021	1.032
2020-061	72	URS-9-TECO	106	1.020	0.004	0.35	1.009	1.030
2020-061	80	ROCHE (CHE	35	1.012	0.002	0.23	1.005	1.019
2020-061	89	AIMSTRIP	114	1.018	0.003	0.31	1.008	1.027
2020-061	94	CLINITEK	107	1.020	0.003	0.25	1.012	1.028
2020-061	102	ACON LAB	179	1.019	0.003	0.26	1.011	1.027
2020-061	117	URS 10 A	10	1.021	0.003	0.26	1.012	1.029
2020-061	131	iCHEM VELO	28	1.026	0.001	0.10	1.023	1.029
2020-061	135	Hybrido AU	12	1.025	0.001	0.09	1.022	1.028
2020-061	138	URS-10T	10	1.020	0.002	0.15	1.015	1.024
2020-061	140	MISSION A	63	1.019	0.003	0.28	1.010	1.027
2020-061	147	Mission120	11	1.020	0.004	0.36	1.009	1.031
067	URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD			
2020-062	11	AMES NM10S	50	1.011	0.004	0.39	0.999	1.023
2020-062	13	BAYER CLI	36	1.013	0.003	0.27	1.004	1.021
2020-062	67	URISCAN	50	1.011	0.004	0.38	1.000	1.023
2020-062	71	ATLAS	21	1.030	0.001	0.07	1.028	1.032
2020-062	72	URS-9-TECO	106	1.013	0.004	0.35	1.002	1.023
2020-062	80	ROCHE (CHE	35	1.013	0.003	0.29	1.004	1.022
2020-062	89	AIMSTRIP	114	1.011	0.004	0.36	1.000	1.022
2020-062	94	CLINITEK	107	1.012	0.003	0.29	1.003	1.021
2020-062	102	ACON LAB	179	1.015	0.004	0.44	1.001	1.028
2020-062	117	URS 10 A	10	1.016	0.003	0.27	1.007	1.024
2020-062	131	iCHEM VELO	28	1.025	0.001	0.09	1.022	1.028
2020-062	135	Hybrido AU	12	1.025	0.001	0.07	1.022	1.027
2020-062	138	URS-10T	10	1.012	0.002	0.23	1.005	1.018
2020-062	140	MISSION A	63	1.014	0.004	0.40	1.002	1.027
2020-062	147	Mission120	11	1.016	0.005	0.47	1.002	1.031

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
067	URI-SPECIFIC GRAVITY	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-063	11 AMES NM10S			50	1.024	0.002	0.22	1.017	1.030
2020-063	13 BAYER CLI			36	1.023	0.003	0.27	1.014	1.031
2020-063	67 URISCAN			50	1.025	0.004	0.38	1.013	1.037
2020-063	71 ATLAS			21	1.030	0.001	0.10	1.027	1.033
2020-063	72 URS-9-TECO			106	1.022	0.003	0.29	1.013	1.031
2020-063	80 ROCHE (CHE			35	1.016	0.002	0.22	1.010	1.023
2020-063	89 AIMSTRIP			114	1.022	0.002	0.24	1.014	1.029
2020-063	94 CLINITEK			107	1.024	0.002	0.23	1.017	1.031
2020-063	102 ACON LAB			179	1.023	0.003	0.26	1.015	1.032
2020-063	117 URS 10 A			10	1.022	0.003	0.32	1.012	1.032
2020-063	131 iCHEM VELO			28	1.024	0.001	0.10	1.022	1.027
2020-063	135 Hybrido AU			12	1.024	0.001	0.07	1.021	1.026
2020-063	138 URS-10T			10	1.023	0.003	0.33	1.012	1.033
2020-063	140 MISSION A			63	1.023	0.003	0.26	1.015	1.031
2020-063	147 Mission120			11	1.023	0.002	0.24	1.016	1.030
2020-064	11 AMES NM10S			50	1.015	0.004	0.35	1.005	1.026
2020-064	13 BAYER CLI			36	1.016	0.002	0.25	1.008	1.023
2020-064	67 URISCAN			50	1.014	0.002	0.21	1.007	1.020
2020-064	71 ATLAS			21	1.030	0.001	0.09	1.027	1.033
2020-064	72 URS-9-TECO			106	1.017	0.004	0.36	1.006	1.028
2020-064	80 ROCHE (CHE			35	1.014	0.002	0.22	1.007	1.020
2020-064	89 AIMSTRIP			114	1.014	0.003	0.34	1.003	1.024
2020-064	94 CLINITEK			107	1.016	0.003	0.28	1.007	1.024
2020-064	102 ACON LAB			179	1.016	0.004	0.42	1.004	1.029
2020-064	117 URS 10 A			10	1.018	0.004	0.39	1.006	1.030
2020-064	131 iCHEM VELO			28	1.028	0.001	0.10	1.025	1.031
2020-064	135 Hybrido AU			12	1.028	0.001	0.09	1.025	1.031
2020-064	138 URS-10T			10	1.017	0.003	0.33	1.007	1.027

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-064	140	MISSION A			63	1.016	0.004	0.42	1.003	1.029
2020-064	147	Mission120			11	1.020	0.000	0.00	1.020	1.020
067	URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-065	11	AMES NM10S			50	1.021	0.003	0.26	1.013	1.029
2020-065	13	BAYER CLI			36	1.021	0.004	0.39	1.009	1.033
2020-065	67	URISCAN			50	1.022	0.002	0.24	1.015	1.029
2020-065	71	ATLAS			21	1.031	0.001	0.14	1.027	1.035
2020-065	72	URS-9-TECO			106	1.021	0.004	0.43	1.008	1.035
2020-065	80	ROCHE (CHE			35	1.017	0.002	0.24	1.010	1.025
2020-065	89	AIMSTRIP			114	1.020	0.003	0.34	1.010	1.031
2020-065	94	CLINITEK			107	1.021	0.003	0.31	1.012	1.031
2020-065	102	ACON LAB			179	1.023	0.003	0.29	1.014	1.032
2020-065	117	URS 10 A			10	1.021	0.004	0.41	1.008	1.033
2020-065	131	iCHEM VELO			28	1.026	0.001	0.09	1.024	1.029
2020-065	135	Hybrido AU			12	1.026	0.001	0.08	1.023	1.028
2020-065	138	URS-10T			10	1.021	0.004	0.41	1.008	1.033
2020-065	140	MISSION A			63	1.022	0.003	0.26	1.014	1.030
2020-065	147	Mission120			11	1.020	0.000	0.00	1.020	1.020
166	MICROALBUMIN			EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-061	58	VITROS			33	833.058	29.645	3.56	624.800	1041.300
2020-061	58	VITROS			33	739.660	45.831	6.20	554.700	924.600
2020-061	62	DADE DIM.			36	817.244	40.754	4.99	612.900	1021.600
2020-061	79	BECK. COUL			16	791.947	53.112	6.71	594.000	989.900
2020-061	79	BECK. COUL			16	739.660	45.831	6.20	554.700	924.600
2020-061	85	Rx DAYTONA			22	663.767	39.387	5.93	497.800	829.700
2020-061	85	Rx DAYTONA			22	713.705	79.199	11.10	535.300	892.100
2020-061	93	ARCHITECT			18	801.906	33.276	4.15	601.400	1002.400
2020-061	95	INTEGRA			25	739.660	45.831	6.20	554.700	924.600
2020-061	95	INTEGRA			25	802.948	47.248	5.88	602.200	1003.700
2020-061	113	ROCHE COBA			23	769.291	20.017	2.60	577.000	961.600

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-061	139 S. EXL.200			47	805.146	44.496	5.53	603.900	1006.400	
2020-061	139 S. EXL.200			47	300.000	0.000	0.00	225.000	375.000	
166 MICROALBUMIN			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-062	58 VITROS			33	1.020	2.040	200.00	0	6.000	
2020-062	58 VITROS			33	6.032	0.177	2.93	1.000	11.000	
2020-062	62 DADE DIM.			36	1.906	0.866	45.43	0	6.900	
2020-062	79 BECK. COUL			16	1.020	2.040	200.00	0	6.000	
2020-062	79 BECK. COUL			16	2.231	2.775	124.36	0	7.200	
2020-062	85 Rx DAYTONA			22	1.627	2.149	132.04	0	6.600	
2020-062	85 Rx DAYTONA			22	3.000	0.000	0.00	0	8.000	
2020-062	93 ARCHITECT			18	5.467	1.622	29.67	0.500	10.500	
2020-062	95 INTEGRA			25	5.738	8.594	149.78	0.700	10.700	
2020-062	95 INTEGRA			25	1.020	2.040	200.00	0	6.000	
2020-062	113 ROCHE COBA			23	10.565	3.740	35.40	5.600	15.600	
2020-062	139 S. EXL.200			47	5.000	0.000	0.00	0.000	10.000	
2020-062	139 S. EXL.200			47	2.439	1.277	52.34	0	7.400	
2020-063	58 VITROS			33	817.994	30.307	3.71	613.500	1022.500	
2020-063	58 VITROS			33	696.000	49.702	7.14	522.000	870.000	
2020-063	62 DADE DIM.			36	791.074	33.739	4.26	593.300	988.800	
2020-063	79 BECK. COUL			16	776.273	53.728	6.92	582.200	970.300	
2020-063	79 BECK. COUL			16	696.000	49.702	7.14	522.000	870.000	
2020-063	85 Rx DAYTONA			22	662.533	47.852	7.22	496.900	828.200	
2020-063	85 Rx DAYTONA			22	697.238	77.920	11.18	522.900	871.500	
2020-063	93 ARCHITECT			18	794.856	34.759	4.37	596.100	993.600	
2020-063	95 INTEGRA			25	696.000	49.702	7.14	522.000	870.000	
2020-063	95 INTEGRA			25	780.328	37.411	4.79	585.200	975.400	
2020-063	113 ROCHE COBA			23	753.904	31.674	4.20	565.400	942.400	
2020-063	139 S. EXL.200			47	783.070	47.077	6.01	587.300	978.800	
2020-063	139 S. EXL.200			47	300.000	0.000	0.00	225.000	375.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
166	MICROALBUMIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	% OR 5.00	{GREATER}				
2020-064	58	VITROS		33	677.940	18.834	2.78	508.500	847.400	
2020-064	58	VITROS		33	780.277	38.327	4.91	585.200	975.300	
2020-064	62	DADE DIM.		36	782.434	35.654	4.56	586.800	978.000	
2020-064	79	BECK. COUL		16	677.940	18.834	2.78	508.500	847.400	
2020-064	79	BECK. COUL		16	751.460	44.291	5.89	563.600	939.300	
2020-064	85	Rx DAYTONA		22	699.732	76.613	10.95	524.800	874.700	
2020-064	85	Rx DAYTONA		22	631.900	40.388	6.39	473.900	789.900	
2020-064	93	ARCHITECT		18	754.744	33.802	4.48	566.100	943.400	
2020-064	95	INTEGRA		25	750.408	37.571	5.01	562.800	938.000	
2020-064	95	INTEGRA		25	677.940	18.834	2.78	508.500	847.400	
2020-064	113	ROCHE COBA		23	728.461	25.204	3.46	546.300	910.600	
2020-064	139	S. EXL.200		47	300.000	0.000	0.00	225.000	375.000	
2020-064	139	S. EXL.200		47	769.680	39.806	5.17	577.300	962.100	
2020-065	58	VITROS		33	6.032	0.177	2.93	1.000	11.000	
2020-065	58	VITROS		33	1.020	2.040	200.00	0	6.000	
2020-065	62	DADE DIM.		36	5.426	1.293	23.83	0.400	10.400	
2020-065	79	BECK. COUL		16	2.156	2.812	130.40	0	7.200	
2020-065	79	BECK. COUL		16	1.020	2.040	200.00	0	6.000	
2020-065	85	Rx DAYTONA		22	3.000	0.000	0.00	0	8.000	
2020-065	85	Rx DAYTONA		22	1.623	2.152	132.61	0	6.600	
2020-065	93	ARCHITECT		18	4.871	0.319	6.56	0	9.900	
2020-065	95	INTEGRA		25	1.020	2.040	200.00	0	6.000	
2020-065	95	INTEGRA		25	5.242	8.282	157.99	0.200	10.200	
2020-065	113	ROCHE COBA		23	10.565	3.740	35.40	5.600	15.600	
2020-065	139	S. EXL.200		47	5.372	1.630	30.34	0.400	10.400	
2020-065	139	S. EXL.200		47	5.000	0.000	0.00	0.000	10.000	
167	CREATININE									
2020-061	58	VITROS		46	220.264	7.969	3.62	165.200	275.300	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS			
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2020-061	62	DADE DIM.		31	239.817	10.049	4.19	179.900	299.800
2020-061	62	DADE DIM.		31	219.271	11.621	5.30	164.500	274.100
2020-061	79	BECK. COUL		14	210.393	14.335	6.81	157.800	263.000
2020-061	85	Rx DAYTONA		15	219.220	20.863	9.52	164.400	274.000
2020-061	93	ARCHITECT		17	223.724	7.130	3.19	167.800	279.700
2020-061	95	INTEGRA		23	242.605	8.907	3.67	182.000	303.300
2020-061	95	INTEGRA		23	238.667	3.399	1.42	179.000	298.300
2020-061	113	ROCHE COBA		26	237.400	7.266	3.06	178.100	296.800
2020-061	139	S. EXL.200		48	228.721	5.628	2.46	171.500	285.900
167	CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-062	58	VITROS		46	226.646	11.497	5.07	170.000	283.300
2020-062	62	DADE DIM.		31	217.745	11.539	5.30	163.300	272.200
2020-062	62	DADE DIM.		31	242.017	8.240	3.40	181.500	302.500
2020-062	79	BECK. COUL		14	211.707	14.292	6.75	158.800	264.600
2020-062	85	Rx DAYTONA		15	221.433	24.465	11.05	166.100	276.800
2020-062	93	ARCHITECT		17	222.812	8.455	3.79	167.100	278.500
2020-062	95	INTEGRA		23	240.667	4.497	1.87	180.500	300.800
2020-062	95	INTEGRA		23	243.709	11.185	4.59	182.800	304.600
2020-062	113	ROCHE COBA		26	236.604	8.119	3.43	177.500	295.800
2020-062	139	S. EXL.200		48	228.096	6.083	2.67	171.100	285.100
2020-063	58	VITROS		46	3.073	0.703	22.86	0	8.100
2020-063	62	DADE DIM.		31	8.400	3.578	42.59	3.400	13.400
2020-063	62	DADE DIM.		31	7.510	5.609	74.69	2.500	12.500
2020-063	79	BECK. COUL		14	0.885	1.524	172.23	0	5.900
2020-063	85	Rx DAYTONA		15	2.787	2.945	105.68	0	7.800
2020-063	93	ARCHITECT		17	4.863	0.328	6.74	0	9.900
2020-063	95	INTEGRA		23	0.945	1.277	135.06	0	5.900
2020-063	95	INTEGRA		23	13.000	0.000	0.00	8.000	18.000
2020-063	113	ROCHE COBA		26	3.558	1.244	34.97	0	8.600

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS					
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-063	139 S. EXL.200			48	8.510	5.870	68.98	3.500	13.500		
167 CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}								
2020-064	58 VITROS			46	221.924	8.524	3.84	166.400	277.400		
2020-064	62 DADE DIM.			31	213.177	10.321	4.84	159.900	266.500		
2020-064	62 DADE DIM.			31	234.000	8.204	3.51	175.500	292.500		
2020-064	79 BECK. COUL			14	196.714	15.079	7.67	147.500	245.900		
2020-064	85 Rx DAYTONA			15	218.460	24.319	11.13	163.800	273.100		
2020-064	93 ARCHITECT			17	218.782	6.367	2.91	164.100	273.500		
2020-064	95 INTEGRA			23	230.333	4.922	2.14	172.800	287.900		
2020-064	95 INTEGRA			23	234.891	8.132	3.46	176.200	293.600		
2020-064	113 ROCHE COBA			26	233.904	8.534	3.65	175.400	292.400		
2020-064	139 S. EXL.200			48	223.685	5.722	2.56	167.800	279.600		
2020-065	58 VITROS			46	3.218	0.229	7.11	0	8.200		
2020-065	62 DADE DIM.			31	8.400	3.578	42.59	3.400	13.400		
2020-065	62 DADE DIM.			31	7.513	5.606	74.61	2.500	12.500		
2020-065	79 BECK. COUL			14	0.785	1.569	200.01	0	5.800		
2020-065	85 Rx DAYTONA			15	1.100	1.240	112.72	0	6.100		
2020-065	93 ARCHITECT			17	4.863	0.328	6.74	0	9.900		
2020-065	95 INTEGRA			23	0.786	1.342	170.63	0	5.800		
2020-065	95 INTEGRA			23	13.000	0.000	0.00	8.000	18.000		
2020-065	113 ROCHE COBA			26	3.554	1.232	34.66	0	8.600		
2020-065	139 S. EXL.200			48	8.515	5.865	68.88	3.500	13.500		

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS						
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART #	LABS #	"SF"	Score	
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
057 URINALYSIS-BILIRUBIN												
2020-061	011 AMES NM10S			057	057	057	057	057 - NEGATIVE	50	50	50	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-061	013	BAYER CLI		057	057	057	057	057	- NEGATIVE	37	37	37	100
2020-061	067	URISCAN		057	057	057	057	057	- NEGATIVE	49	49	49	100
2020-061	071	ATLAS		057	057	057	057	057	- NEGATIVE	21	21	21	100
2020-061	072	URS-9-TECO		057	057	057	057	057	- NEGATIVE	106	106	106	100
2020-061	080	ROCHE (CHE		057	057	057	057	057	- NEGATIVE	34	34	34	100
2020-061	089	AIMSTRIP		057	057	057	057	057	- NEGATIVE	113	113	113	100
2020-061	094	CLINITEK		057	057	057	057	057	- NEGATIVE	106	106	106	100
2020-061	102	ACON LAB		057	057	057	057	057	- NEGATIVE	174	174	174	100
2020-061	117	URS 10 A		057	057	057	057	057	- NEGATIVE	10	10	10	100
2020-061	131	ICHEM VELO		057	057	057	057	057	- NEGATIVE	27	27	27	100
2020-061	135	Hybrido AU		057	057	057	057	057	- NEGATIVE	12	12	12	100
2020-061	138	URS-10T		057	057	057	057	057	- NEGATIVE	10	10	10	100
2020-061	140	MISSION A		057	057	057	057	057	- NEGATIVE	65	65	65	100
2020-061	147	Mission120		057	057	057	057	057	- NEGATIVE	11	11	11	100
2020-061	148	Mission500		057	057	057	057	057	- NEGATIVE	11	11	11	100
057 URINALYSIS-BILIRUBIN													
2020-062	011	AMES NM10S		057	057	057	057	057	- NEGATIVE	50	50	50	100
2020-062	013	BAYER CLI		057	057	057	057	057	- NEGATIVE	37	37	37	100
2020-062	067	URISCAN		057	057	057	057	057	- NEGATIVE	49	49	49	100
2020-062	071	ATLAS		057	057	057	057	057	- NEGATIVE	21	21	21	100
2020-062	072	URS-9-TECO		057	057	057	057	057	- NEGATIVE	106	106	106	100
2020-062	080	ROCHE (CHE		057	057	057	057	057	- NEGATIVE	34	34	34	100
2020-062	089	AIMSTRIP		057	057	057	057	057	- NEGATIVE	113	113	113	100
2020-062	094	CLINITEK		057	057	057	057	057	- NEGATIVE	106	106	106	100
2020-062	102	ACON LAB		057	057	057	057	057	- NEGATIVE	174	174	174	100
2020-062	117	URS 10 A		057	057	057	057	057	- NEGATIVE	10	10	10	100
2020-062	131	ICHEM VELO		057	057	057	057	057	- NEGATIVE	27	27	27	100
2020-062	135	Hybrido AU		057	057	057	057	057	- NEGATIVE	12	12	12	100
2020-062	138	URS-10T		057	057	057	057	057	- NEGATIVE	10	10	10	100
2020-062	140	MISSION A		057	057	057	057	057	- NEGATIVE	65	65	65	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	147	Mission120		057	057	057	057 - NEGATIVE	11	11	11	100	
2020-062	148	Mission500		057	057	057	057 - NEGATIVE	11	11	11	100	
057 URINALYSIS-BILIRUBIN												
2020-063	011	AMES NM10S		057	057	057	057 - NEGATIVE	50	50	50	100	
2020-063	013	BAYER CLI		057	057	057	057 - NEGATIVE	37	37	37	100	
2020-063	067	URISCAN		057	057	057	057 - NEGATIVE	49	49	49	100	
2020-063	071	ATLAS		057	057	057	057 - NEGATIVE	21	21	21	100	
2020-063	072	URS-9-TECO		057	057	057	057 - NEGATIVE	106	106	106	100	
2020-063	080	ROCHE (CHE		057	057	057	057 - NEGATIVE	34	34	34	100	
2020-063	089	AIMSTRIP		057	057	057	057 - NEGATIVE	113	113	113	100	
2020-063	094	CLINITEK		057	057	057	057 - NEGATIVE	106	106	106	100	
2020-063	102	ACON LAB		057	057	057	057 - NEGATIVE	174	173	173	100	
2020-063	102	ACON LAB		057	057	057	058 - TRACES, 1+	174	1	173	0	
2020-063	117	URS 10 A		057	057	057	057 - NEGATIVE	10	10	10	100	
2020-063	131	iCHEM VELO		057	057	057	057 - NEGATIVE	27	27	27	100	
2020-063	135	Hybrido AU		057	057	057	057 - NEGATIVE	12	12	12	100	
2020-063	138	URS-10T		057	057	057	057 - NEGATIVE	10	10	10	100	
2020-063	140	MISSION A		057	057	057	057 - NEGATIVE	65	65	65	100	
2020-063	147	Mission120		057	057	057	057 - NEGATIVE	11	11	11	100	
2020-063	148	Mission500		057	057	057	057 - NEGATIVE	11	11	11	100	
2020-064	011	AMES NM10S		058	060	058	058 - TRACES, 1+	50	3	50	100	
2020-064	011	AMES NM10S		058	060	058	059 - 2+	50	40	50	100	
2020-064	011	AMES NM10S		058	060	058	060 - 3+	50	7	50	100	
2020-064	013	BAYER CLI		058	060	058	057 - NEGATIVE	37	1	36	0	
2020-064	013	BAYER CLI		058	060	058	058 - TRACES, 1+	37	6	36	100	
2020-064	013	BAYER CLI		058	060	058	059 - 2+	37	28	36	100	
2020-064	013	BAYER CLI		058	060	058	060 - 3+	37	2	36	100	
2020-064	067	URISCAN		057	059	058	057 - NEGATIVE	49	2	49	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-064	067	URISCAN		057 059	058 060		058 - TRACES,1+	49	28	49	100
2020-064	067	URISCAN		057 059	058 060		059 - 2+	49	16	49	100
2020-064	067	URISCAN		057 059	058 060		060 - 3+	49	3	49	100
2020-064	071	ATLAS		057 059	058 060		058 - TRACES,1+	21	19	21	100
2020-064	071	ATLAS		057 059	058 060		059 - 2+	21	2	21	100
2020-064	072	URS-9-TECO		058 060	058 060		057 - NEGATIVE	106	3	103	0
2020-064	072	URS-9-TECO		058 060	058 060		058 - TRACES,1+	106	27	103	100
2020-064	072	URS-9-TECO		058 060	058 060		059 - 2+	106	51	103	100
2020-064	072	URS-9-TECO		058 060	058 060		060 - 3+	106	25	103	100
2020-064	080	ROCHE (CHE		057 059	058 060		058 - TRACES,1+	34	29	34	100
2020-064	080	ROCHE (CHE		057 059	058 060		059 - 2+	34	4	34	100
2020-064	080	ROCHE (CHE		057 059	058 060		060 - 3+	34	1	34	100
2020-064	089	AIMSTRIP		058 060	058 060		057 - NEGATIVE	113	2	111	0
2020-064	089	AIMSTRIP		058 060	058 060		058 - TRACES,1+	113	38	111	100
2020-064	089	AIMSTRIP		058 060	058 060		059 - 2+	113	59	111	100
2020-064	089	AIMSTRIP		058 060	058 060		060 - 3+	113	14	111	100
2020-064	094	CLINITEK		058 060	058 060		058 - TRACES,1+	106	13	106	100
2020-064	094	CLINITEK		058 060	058 060		059 - 2+	106	90	106	100
2020-064	094	CLINITEK		058 060	058 060		060 - 3+	106	3	106	100
2020-064	102	ACON LAB		057 059	058 060		057 - NEGATIVE	174	9	174	100
2020-064	102	ACON LAB		057 059	058 060		058 - TRACES,1+	174	111	174	100
2020-064	102	ACON LAB		057 059	058 060		059 - 2+	174	37	174	100
2020-064	102	ACON LAB		057 059	058 060		060 - 3+	174	17	174	100
2020-064	117	URS 10 A		058 060	058 060		059 - 2+	10	8	10	100
2020-064	117	URS 10 A		058 060	058 060		060 - 3+	10	2	10	100
2020-064	131	ICHEM VELO		057 057	058 060		057 - NEGATIVE	27	27	27	100
2020-064	135	Hybrido AU		058 060	058 060		059 - 2+	12	11	12	100
2020-064	135	Hybrido AU		058 060	058 060		060 - 3+	12	1	12	100
2020-064	138	URS-10T		058 060	058 060		058 - TRACES,1+	10	3	10	100
2020-064	138	URS-10T		058 060	058 060		059 - 2+	10	7	10	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-064	140	MISSION A		057 059	058 060		057 - NEGATIVE	65	4	65	100	
2020-064	140	MISSION A		057 059	058 060		058 - TRACES, 1+	65	47	65	100	
2020-064	140	MISSION A		057 059	058 060		059 - 2+	65	12	65	100	
2020-064	140	MISSION A		057 059	058 060		060 - 3+	65	2	65	100	
2020-064	147	Mission120		058 060	058 060		058 - TRACES, 1+	11	4	11	100	
2020-064	147	Mission120		058 060	058 060		059 - 2+	11	6	11	100	
2020-064	147	Mission120		058 060	058 060		060 - 3+	11	1	11	100	
2020-064	148	Mission500		057 059	058 060		058 - TRACES, 1+	11	9	11	100	
2020-064	148	Mission500		057 059	058 060		059 - 2+	11	2	11	100	
057 URINALYSIS-BILIRUBIN												
2020-065	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	50	50	50	100	
2020-065	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	37	37	37	100	
2020-065	067	URISCAN		057 057	057 057		057 - NEGATIVE	49	49	49	100	
2020-065	071	ATLAS		057 057	057 057		057 - NEGATIVE	21	21	21	100	
2020-065	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	106	106	106	100	
2020-065	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	34	34	34	100	
2020-065	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	113	112	112	100	
2020-065	089	AIMSTRIP		057 057	057 057		058 - TRACES, 1+	113	1	112	0	
2020-065	094	CLINITEK		057 057	057 057		057 - NEGATIVE	106	106	106	100	
2020-065	102	ACON LAB		057 057	057 057		057 - NEGATIVE	174	173	173	100	
2020-065	102	ACON LAB		057 057	057 057		058 - TRACES, 1+	174	1	173	0	
2020-065	117	URS 10 A		057 057	057 057		057 - NEGATIVE	10	10	10	100	
2020-065	131	ICHEM VELO		057 057	057 057		057 - NEGATIVE	27	27	27	100	
2020-065	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	12	12	12	100	
2020-065	138	URS-10T		057 057	057 057		057 - NEGATIVE	10	10	10	100	
2020-065	140	MISSION A		057 057	057 057		057 - NEGATIVE	65	65	65	100	
2020-065	147	Mission120		057 057	057 057		057 - NEGATIVE	11	11	11	100	
2020-065	148	Mission500		057 057	057 057		057 - NEGATIVE	11	11	11	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
058 URINALYSIS-GLUCOSE													
2020-061	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	50	50	50	100		
2020-061	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	36	36	36	100		
2020-061	067	URISCAN		073 073	073 073		073 - NEGATIVE	49	49	49	100		
2020-061	071	ATLAS		073 073	073 073		073 - NEGATIVE	21	21	21	100		
2020-061	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	108	108	108	100		
2020-061	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	35	35	35	100		
2020-061	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	113	113	113	100		
2020-061	094	CLINITEK		073 073	073 073		073 - NEGATIVE	106	105	105	100		
2020-061	094	CLINITEK		073 073	073 073		074 - TRACES	106	1	105	0		
2020-061	102	ACON LAB		073 073	073 073		073 - NEGATIVE	177	177	177	100		
2020-061	117	URS 10 A		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2020-061	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	28	27	27	100		
2020-061	131	iCHEM VELO		073 073	073 073		074 - TRACES	28	1	27	0		
2020-061	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	12	12	100		
2020-061	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2020-061	140	MISSION A		073 073	073 073		073 - NEGATIVE	63	63	63	100		
2020-061	147	Mission120		073 073	073 073		073 - NEGATIVE	11	11	11	100		
2020-061	148	Mission500		073 073	073 073		073 - NEGATIVE	11	11	11	100		
2020-062	011	AMES NM10S		076 078	075 077		073 - NEGATIVE	50	1	49	0		
2020-062	011	AMES NM10S		076 078	075 077		075 - 250mg/dl-+	50	2	49	100		
2020-062	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	50	16	49	100		
2020-062	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	50	30	49	100		
2020-062	011	AMES NM10S		076 078	075 077		078 - 2000mg/dl	50	1	49	100		
2020-062	013	BAYER CLI		076 078	075 077		075 - 250mg/dl-+	36	1	36	100		
2020-062	013	BAYER CLI		076 078	075 077		076 - 500mg/dl2+	36	16	36	100		
2020-062	013	BAYER CLI		076 078	075 077		077 - 1000mg/dl	36	19	36	100		
2020-062	067	URISCAN		077 078	075 077		075 - 250mg/dl-+	49	1	49	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-062	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	49	4	49	100
2020-062	067	URISCAN		077 078	075 077		077 - 1000mg/dl	49	15	49	100
2020-062	067	URISCAN		077 078	075 077		078 - 2000mg/dl	49	29	49	100
2020-062	071	ATLAS		075 077	075 077		076 - 500mg/dl2+	21	20	21	100
2020-062	071	ATLAS		075 077	075 077		077 - 1000mg/dl	21	1	21	100
2020-062	072	URS-9-TECO		075 077	075 077		075 - 250mg/dl-+	108	3	98	100
2020-062	072	URS-9-TECO		075 077	075 077		076 - 500mg/dl2+	108	48	98	100
2020-062	072	URS-9-TECO		075 077	075 077		077 - 1000mg/dl	108	47	98	100
2020-062	072	URS-9-TECO		075 077	075 077		078 - 2000mg/dl	108	10	98	0
2020-062	080	ROCHE (CHE		076 078	075 077		075 - 250mg/dl-+	35	1	35	100
2020-062	080	ROCHE (CHE		076 078	075 077		076 - 500mg/dl2+	35	1	35	100
2020-062	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	35	32	35	100
2020-062	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	35	1	35	100
2020-062	089	AIMSTRIP		075 077	075 077		074 - TRACES	113	1	110	0
2020-062	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	113	5	110	100
2020-062	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	113	67	110	100
2020-062	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	113	38	110	100
2020-062	089	AIMSTRIP		075 077	075 077		078 - 2000mg/dl	113	2	110	0
2020-062	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	106	39	106	100
2020-062	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	106	65	106	100
2020-062	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	106	2	106	100
2020-062	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	177	91	171	100
2020-062	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	177	78	171	100
2020-062	102	ACON LAB		075 077	075 077		078 - 2000mg/dl	177	5	171	0
2020-062	102	ACON LAB		075 077	075 077		073 - NEGATIVE	177	1	171	0
2020-062	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	177	2	171	100
2020-062	117	URS 10 A		075 077	075 077		075 - 250mg/dl-+	10	1	10	100
2020-062	117	URS 10 A		075 077	075 077		076 - 500mg/dl2+	10	6	10	100
2020-062	117	URS 10 A		075 077	075 077		077 - 1000mg/dl	10	3	10	100
2020-062	131	ICHEM VELO		075 077	075 077		076 - 500mg/dl2+	28	25	28	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	28	3	28	100	
2020-062	135	Hybrido AU		076 078	075 077		077 - 1000mg/dl	12	8	12	100	
2020-062	135	Hybrido AU		076 078	075 077		078 - 2000mg/dl	12	4	12	100	
2020-062	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	10	3	10	100	
2020-062	138	URS-10T		076 078	075 077		077 - 1000mg/dl	10	6	10	100	
2020-062	138	URS-10T		076 078	075 077		078 - 2000mg/dl	10	1	10	100	
2020-062	140	MISSION A		076 078	075 077		075 - 250mg/dl-+	63	1	63	100	
2020-062	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	63	22	63	100	
2020-062	140	MISSION A		076 078	075 077		077 - 1000mg/dl	63	35	63	100	
2020-062	140	MISSION A		076 078	075 077		078 - 2000mg/dl	63	5	63	100	
2020-062	147	Mission120		076 078	075 077		076 - 500mg/dl2+	11	2	11	100	
2020-062	147	Mission120		076 078	075 077		077 - 1000mg/dl	11	8	11	100	
2020-062	147	Mission120		076 078	075 077		078 - 2000mg/dl	11	1	11	100	
2020-062	148	Mission500		076 078	075 077		076 - 500mg/dl2+	11	4	11	100	
2020-062	148	Mission500		076 078	075 077		077 - 1000mg/dl	11	6	11	100	
2020-062	148	Mission500		076 078	075 077		078 - 2000mg/dl	11	1	11	100	
058 URINALYSIS-GLUCOSE												
2020-063	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	50	18	50	100	
2020-063	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	50	32	50	100	
2020-063	013	BAYER CLI		076 078	075 077		076 - 500mg/dl2+	36	15	36	100	
2020-063	013	BAYER CLI		076 078	075 077		077 - 1000mg/dl	36	21	36	100	
2020-063	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	49	3	49	100	
2020-063	067	URISCAN		077 078	075 077		077 - 1000mg/dl	49	12	49	100	
2020-063	067	URISCAN		077 078	075 077		078 - 2000mg/dl	49	34	49	100	
2020-063	071	ATLAS		076 078	075 077		077 - 1000mg/dl	21	20	21	100	
2020-063	071	ATLAS		076 078	075 077		078 - 2000mg/dl	21	1	21	100	
2020-063	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	108	44	108	100	
2020-063	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	108	55	108	100	
2020-063	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	108	9	108	100	
2020-063	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	35	34	35	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-063	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	35	1	35	100
2020-063	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	113	3	112	100
2020-063	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	113	78	112	100
2020-063	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	113	31	112	100
2020-063	089	AIMSTRIP		075 077	075 077		078 - 2000mg/dl	113	1	112	0
2020-063	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	106	1	106	100
2020-063	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	106	37	106	100
2020-063	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	106	67	106	100
2020-063	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	106	1	106	100
2020-063	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	177	1	173	100
2020-063	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	177	108	173	100
2020-063	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	177	64	173	100
2020-063	102	ACON LAB		075 077	075 077		078 - 2000mg/dl	177	4	173	0
2020-063	117	URS 10 A		076 078	075 077		076 - 500mg/dl2+	10	5	10	100
2020-063	117	URS 10 A		076 078	075 077		077 - 1000mg/dl	10	5	10	100
2020-063	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	28	25	28	100
2020-063	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	28	3	28	100
2020-063	135	Hybrido AU		077 078	075 077		077 - 1000mg/dl	12	5	12	100
2020-063	135	Hybrido AU		077 078	075 077		078 - 2000mg/dl	12	7	12	100
2020-063	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	10	2	10	100
2020-063	138	URS-10T		076 078	075 077		077 - 1000mg/dl	10	7	10	100
2020-063	138	URS-10T		076 078	075 077		078 - 2000mg/dl	10	1	10	100
2020-063	140	MISSION A		076 078	075 077		075 - 250mg/dl-+	63	1	63	100
2020-063	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	63	27	63	100
2020-063	140	MISSION A		076 078	075 077		077 - 1000mg/dl	63	34	63	100
2020-063	140	MISSION A		076 078	075 077		078 - 2000mg/dl	63	1	63	100
2020-063	147	Mission120		076 078	075 077		076 - 500mg/dl2+	11	1	11	100
2020-063	147	Mission120		076 078	075 077		077 - 1000mg/dl	11	9	11	100
2020-063	147	Mission120		076 078	075 077		078 - 2000mg/dl	11	1	11	100
2020-063	148	Mission500		076 078	075 077		076 - 500mg/dl2+	11	4	11	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-063	148	Mission500		076 078	075 077		077 - 1000mg/dl	11	5	11	100		
2020-063	148	Mission500		076 078	075 077		078 - 2000mg/dl	11	2	11	100		
058 URINALYSIS-GLUCOSE													
2020-064	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	50	50	50	100		
2020-064	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	36	36	36	100		
2020-064	067	URISCAN		073 073	073 073		073 - NEGATIVE	49	49	49	100		
2020-064	071	ATLAS		073 073	073 073		073 - NEGATIVE	21	21	21	100		
2020-064	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	108	105	105	100		
2020-064	072	URS-9-TECO		073 073	073 073		074 - TRACES	108	3	105	0		
2020-064	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	35	35	35	100		
2020-064	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	113	113	113	100		
2020-064	094	CLINITEK		073 073	073 073		073 - NEGATIVE	106	106	106	100		
2020-064	102	ACON LAB		073 073	073 073		073 - NEGATIVE	177	176	176	100		
2020-064	102	ACON LAB		073 073	073 073		077 - 1000mg/dl	177	1	176	0		
2020-064	117	URS 10 A		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2020-064	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	28	28	28	100		
2020-064	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	12	12	100		
2020-064	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2020-064	140	MISSION A		073 073	073 073		073 - NEGATIVE	63	63	63	100		
2020-064	147	Mission120		073 073	073 073		073 - NEGATIVE	11	11	11	100		
2020-064	148	Mission500		073 073	073 073		073 - NEGATIVE	11	11	11	100		
2020-065	011	AMES NM10S		074 076	074 076		073 - NEGATIVE	50	2	48	0		
2020-065	011	AMES NM10S		074 076	074 076		074 - TRACES	50	20	48	100		
2020-065	011	AMES NM10S		074 076	074 076		075 - 250mg/dl-+	50	27	48	100		
2020-065	011	AMES NM10S		074 076	074 076		076 - 500mg/dl2+	50	1	48	100		
2020-065	013	BAYER CLI		074 076	074 076		074 - TRACES	36	13	36	100		
2020-065	013	BAYER CLI		074 076	074 076		075 - 250mg/dl-+	36	23	36	100		
2020-065	067	URISCAN		075 077	074 076		074 - TRACES	49	1	48	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-065	067	URISCAN		075 077	074 076		075 - 250mg/dl-+	49	17	48	100
2020-065	067	URISCAN		075 077	074 076		076 - 500mg/dl2+	49	23	48	100
2020-065	067	URISCAN		075 077	074 076		077 - 1000mg/dl	49	7	48	100
2020-065	067	URISCAN		075 077	074 076		078 - 2000mg/dl	49	1	48	0
2020-065	071	ATLAS		076 078	074 076		076 - 500mg/dl2+	21	3	21	100
2020-065	071	ATLAS		076 078	074 076		077 - 1000mg/dl	21	17	21	100
2020-065	071	ATLAS		076 078	074 076		078 - 2000mg/dl	21	1	21	100
2020-065	072	URS-9-TECO		074 076	074 076		073 - NEGATIVE	108	6	100	0
2020-065	072	URS-9-TECO		074 076	074 076		074 - TRACES	108	33	100	100
2020-065	072	URS-9-TECO		074 076	074 076		075 - 250mg/dl-+	108	59	100	100
2020-065	072	URS-9-TECO		074 076	074 076		076 - 500mg/dl2+	108	8	100	100
2020-065	072	URS-9-TECO		074 076	074 076		077 - 1000mg/dl	108	2	100	0
2020-065	080	ROCHE (CHE		076 078	074 076		075 - 250mg/dl-+	35	1	35	100
2020-065	080	ROCHE (CHE		076 078	074 076		076 - 500mg/dl2+	35	1	35	100
2020-065	080	ROCHE (CHE		076 078	074 076		077 - 1000mg/dl	35	32	35	100
2020-065	080	ROCHE (CHE		076 078	074 076		078 - 2000mg/dl	35	1	35	100
2020-065	089	AIMSTRIP		074 076	074 076		074 - TRACES	113	26	113	100
2020-065	089	AIMSTRIP		074 076	074 076		075 - 250mg/dl-+	113	87	113	100
2020-065	094	CLINITEK		074 076	074 076		073 - NEGATIVE	106	1	105	0
2020-065	094	CLINITEK		074 076	074 076		074 - TRACES	106	38	105	100
2020-065	094	CLINITEK		074 076	074 076		075 - 250mg/dl-+	106	67	105	100
2020-065	102	ACON LAB		074 076	074 076		073 - NEGATIVE	177	1	176	0
2020-065	102	ACON LAB		074 076	074 076		074 - TRACES	177	24	176	100
2020-065	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	177	151	176	100
2020-065	102	ACON LAB		074 076	074 076		076 - 500mg/dl2+	177	1	176	100
2020-065	117	URS 10 A		074 076	074 076		073 - NEGATIVE	10	1	9	0
2020-065	117	URS 10 A		074 076	074 076		074 - TRACES	10	2	9	100
2020-065	117	URS 10 A		074 076	074 076		075 - 250mg/dl-+	10	7	9	100
2020-065	131	iCHEM VELO		075 077	074 076		076 - 500mg/dl2+	28	25	28	100
2020-065	131	iCHEM VELO		075 077	074 076		077 - 1000mg/dl	28	3	28	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	135	Hybrido AU		077 078	074 076		077 - 1000mg/dl	12	5	12	100	
2020-065	135	Hybrido AU		077 078	074 076		078 - 2000mg/dl	12	7	12	100	
2020-065	138	URS-10T		073 075	074 076		074 - TRACES	10	5	10	100	
2020-065	138	URS-10T		073 075	074 076		075 - 250mg/dl-+	10	4	10	100	
2020-065	138	URS-10T		073 075	074 076		076 - 500mg/dl2+	10	1	10	100	
2020-065	140	MISSION A		074 076	074 076		073 - NEGATIVE	63	1	62	0	
2020-065	140	MISSION A		074 076	074 076		074 - TRACES	63	7	62	100	
2020-065	140	MISSION A		074 076	074 076		075 - 250mg/dl-+	63	55	62	100	
2020-065	147	Mission120		074 076	074 076		074 - TRACES	11	2	11	100	
2020-065	147	Mission120		074 076	074 076		075 - 250mg/dl-+	11	9	11	100	
2020-065	148	Mission500		074 076	074 076		075 - 250mg/dl-+	11	11	11	100	
059 URINALYS.HG.OR BLOOD												
2020-061	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	3	50	100	
2020-061	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	50	47	50	100	
2020-061	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	37	37	37	100	
2020-061	067	URISCAN		068 069	068 069		066 - NEGATIVE	49	1	48	0	
2020-061	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	49	2	48	100	
2020-061	067	URISCAN		068 069	068 069		069 - POS.(3+))	49	46	48	100	
2020-061	071	ATLAS		068 069	068 069		069 - POS.(3+))	21	21	21	100	
2020-061	072	URS-9-TECO		068 069	068 069		066 - NEGATIVE	108	1	107	0	
2020-061	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	108	11	107	100	
2020-061	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	108	96	107	100	
2020-061	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	35	1	35	100	
2020-061	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	35	34	35	100	
2020-061	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	113	2	113	100	
2020-061	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	113	111	113	100	
2020-061	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	106	3	106	100	
2020-061	094	CLINITEK		068 069	068 069		069 - POS.(3+))	106	103	106	100	
2020-061	102	ACON LAB		068 069	068 069		066 - NEGATIVE	178	1	177	0	
2020-061	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	178	2	177	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-061	102	ACON LAB		068 069	068 069		069 - POS.(3+))	178	175	177	100		
2020-061	117	URS 10 A		068 069	068 069		069 - POS.(3+))	10	10	10	100		
2020-061	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	28	3	28	100		
2020-061	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	28	25	28	100		
2020-061	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	7	12	100		
2020-061	135	Hybrido AU		067 069	068 069		069 - POS.(3+))	12	5	12	100		
2020-061	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	10	10	100		
2020-061	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	63	3	63	100		
2020-061	140	MISSION A		068 069	068 069		069 - POS.(3+))	63	60	63	100		
2020-061	147	Mission120		068 069	068 069		069 - POS.(3+))	10	10	10	100		
2020-061	148	Mission500		068 069	068 069		069 - POS.(3+))	10	10	10	100		
059 URINALYS.HG.OR BLOOD													
2020-062	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	50	50	50	100		
2020-062	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	37	36	36	100		
2020-062	013	BAYER CLI		066 066	066 066		067 - TRACES	37	1	36	0		
2020-062	067	URISCAN		066 066	066 066		066 - NEGATIVE	49	49	49	100		
2020-062	071	ATLAS		066 066	066 066		066 - NEGATIVE	21	21	21	100		
2020-062	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	108	108	108	100		
2020-062	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	35	35	35	100		
2020-062	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	113	113	113	100		
2020-062	094	CLINITEK		066 066	066 066		066 - NEGATIVE	106	106	106	100		
2020-062	102	ACON LAB		066 066	066 066		066 - NEGATIVE	178	177	177	100		
2020-062	102	ACON LAB		066 066	066 066		069 - POS.(3+))	178	1	177	0		
2020-062	117	URS 10 A		066 066	066 066		066 - NEGATIVE	10	10	10	100		
2020-062	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	28	26	26	100		
2020-062	131	iCHEM VELO		066 066	066 066		068 - POS.(1-2+)	28	2	26	0		
2020-062	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	12	12	12	100		
2020-062	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100		
2020-062	140	MISSION A		066 066	066 066		066 - NEGATIVE	63	63	63	100		
2020-062	147	Mission120		066 066	066 066		066 - NEGATIVE	10	10	10	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	148	Mission500		066 066	066 066		066 - NEGATIVE	10	10	10	100	
059 URINALYS.HG.OR BLOOD												
2020-063	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	4	50	100	
2020-063	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	50	46	50	100	
2020-063	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	37	4	37	100	
2020-063	013	BAYER CLI		068 069	068 069		069 - POS.(3+)	37	33	37	100	
2020-063	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	49	2	49	100	
2020-063	067	URISCAN		068 069	068 069		069 - POS.(3+)	49	47	49	100	
2020-063	071	ATLAS		068 069	068 069		069 - POS.(3+)	21	21	21	100	
2020-063	072	URS-9-TECO		068 069	068 069		067 - TRACES	108	1	107	0	
2020-063	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	108	17	107	100	
2020-063	072	URS-9-TECO		068 069	068 069		069 - POS.(3+)	108	90	107	100	
2020-063	080	ROCHE (CHE		068 069	068 069		066 - NEGATIVE	35	1	33	0	
2020-063	080	ROCHE (CHE		068 069	068 069		067 - TRACES	35	1	33	0	
2020-063	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	35	1	33	100	
2020-063	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+)	35	32	33	100	
2020-063	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	113	1	112	0	
2020-063	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	113	10	112	100	
2020-063	089	AIMSTRIP		068 069	068 069		069 - POS.(3+)	113	102	112	100	
2020-063	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	106	3	106	100	
2020-063	094	CLINITEK		068 069	068 069		069 - POS.(3+)	106	103	106	100	
2020-063	102	ACON LAB		068 069	068 069		066 - NEGATIVE	178	1	177	0	
2020-063	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	178	6	177	100	
2020-063	102	ACON LAB		068 069	068 069		069 - POS.(3+)	178	171	177	100	
2020-063	117	URS 10 A		068 069	068 069		068 - POS.(1-2+)	10	2	10	100	
2020-063	117	URS 10 A		068 069	068 069		069 - POS.(3+)	10	8	10	100	
2020-063	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	28	3	28	100	
2020-063	131	iCHEM VELO		068 069	068 069		069 - POS.(3+)	28	25	28	100	
2020-063	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	7	12	100	
2020-063	135	Hybrido AU		067 069	068 069		069 - POS.(3+)	12	5	12	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-063	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	10	10	100
2020-063	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	63	4	63	100
2020-063	140	MISSION A		068 069	068 069		069 - POS.(3+))	63	59	63	100
2020-063	147	Mission120		068 069	068 069		069 - POS.(3+))	10	10	10	100
2020-063	148	Mission500		068 069	068 069		069 - POS.(3+))	10	10	10	100
059 URINALYS.HG.OR BLOOD											
2020-064	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	4	50	100
2020-064	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	50	46	50	100
2020-064	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	37	37	37	100
2020-064	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	49	4	49	100
2020-064	067	URISCAN		068 069	068 069		069 - POS.(3+))	49	45	49	100
2020-064	071	ATLAS		068 069	068 069		069 - POS.(3+))	21	21	21	100
2020-064	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	108	21	108	100
2020-064	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	108	87	108	100
2020-064	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	35	4	35	100
2020-064	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	35	31	35	100
2020-064	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	113	1	112	0
2020-064	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	113	5	112	100
2020-064	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	113	107	112	100
2020-064	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	106	6	106	100
2020-064	094	CLINITEK		068 069	068 069		069 - POS.(3+))	106	100	106	100
2020-064	102	ACON LAB		068 069	068 069		066 - NEGATIVE	178	1	177	0
2020-064	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	178	2	177	100
2020-064	102	ACON LAB		068 069	068 069		069 - POS.(3+))	178	175	177	100
2020-064	117	URS 10 A		068 069	068 069		068 - POS.(1-2+)	10	1	10	100
2020-064	117	URS 10 A		068 069	068 069		069 - POS.(3+))	10	9	10	100
2020-064	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	28	11	28	100
2020-064	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	28	17	28	100
2020-064	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	12	12	100
2020-064	138	URS-10T		068 069	068 069		066 - NEGATIVE	10	1	9	0

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-064	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	9	9	100	
2020-064	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	63	3	63	100	
2020-064	140	MISSION A		068 069	068 069		069 - POS.(3+))	63	60	63	100	
2020-064	147	Mission120		068 069	068 069		069 - POS.(3+))	10	10	10	100	
2020-064	148	Mission500		068 069	068 069		069 - POS.(3+))	10	10	10	100	
059 URINALYS.HG.OR BLOOD												
2020-065	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	50	49	49	100	
2020-065	011	AMES NM10S		066 066	066 066		069 - POS.(3+))	50	1	49	0	
2020-065	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	37	36	36	100	
2020-065	013	BAYER CLI		066 066	066 066		067 - TRACES	37	1	36	0	
2020-065	067	URISCAN		066 066	066 066		066 - NEGATIVE	49	49	49	100	
2020-065	071	ATLAS		066 066	066 066		066 - NEGATIVE	21	21	21	100	
2020-065	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	108	108	108	100	
2020-065	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	35	34	34	100	
2020-065	080	ROCHE (CHE		066 066	066 066		069 - POS.(3+))	35	1	34	0	
2020-065	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	113	112	112	100	
2020-065	089	AIMSTRIP		066 066	066 066		069 - POS.(3+))	113	1	112	0	
2020-065	094	CLINITEK		066 066	066 066		066 - NEGATIVE	106	105	105	100	
2020-065	094	CLINITEK		066 066	066 066		067 - TRACES	106	1	105	0	
2020-065	102	ACON LAB		066 066	066 066		066 - NEGATIVE	178	177	177	100	
2020-065	102	ACON LAB		066 066	066 066		069 - POS.(3+))	178	1	177	0	
2020-065	117	URS 10 A		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-065	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	28	27	27	100	
2020-065	131	iCHEM VELO		066 066	066 066		068 - POS.(1-2+)	28	1	27	0	
2020-065	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	12	11	11	100	
2020-065	135	Hybrido AU		066 066	066 066		068 - POS.(1-2+)	12	1	11	0	
2020-065	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-065	140	MISSION A		066 066	066 066		066 - NEGATIVE	63	63	63	100	
2020-065	147	Mission120		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-065	148	Mission500		066 066	066 066		066 - NEGATIVE	10	10	10	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
060 URINALYSIS-KETONES													
2020-061	011	AMES NM10S		090	090	090	090	090 - NEGATIVE	50	50	50	100	
2020-061	013	BAYER CLI		090	090	090	090	090 - NEGATIVE	36	36	36	100	
2020-061	067	URISCAN		090	090	090	090	090 - NEGATIVE	49	49	49	100	
2020-061	071	ATLAS		090	090	090	090	090 - NEGATIVE	21	21	21	100	
2020-061	072	URS-9-TECO		090	090	090	090	090 - NEGATIVE	109	109	109	100	
2020-061	080	ROCHE (CHE		090	090	090	090	090 - NEGATIVE	35	35	35	100	
2020-061	089	AIMSTRIP		090	090	090	090	090 - NEGATIVE	113	113	113	100	
2020-061	094	CLINITEK		090	090	090	090	090 - NEGATIVE	106	106	106	100	
2020-061	102	ACON LAB		090	090	090	090	090 - NEGATIVE	176	176	176	100	
2020-061	117	URS 10 A		090	090	090	090	090 - NEGATIVE	10	10	10	100	
2020-061	131	iCHEM VELO		090	090	090	090	090 - NEGATIVE	28	28	28	100	
2020-061	135	Hybrido AU		090	090	090	090	090 - NEGATIVE	12	12	12	100	
2020-061	138	URS-10T		090	090	090	090	090 - NEGATIVE	10	10	10	100	
2020-061	140	MISSION A		090	090	090	090	090 - NEGATIVE	63	63	63	100	
2020-061	147	Mission120		090	090	090	090	090 - NEGATIVE	11	11	11	100	
2020-061	148	Mission500		090	090	090	090	090 - NEGATIVE	11	11	11	100	
2020-062	011	AMES NM10S		090	090	090	090	090 - NEGATIVE	50	50	50	100	
2020-062	013	BAYER CLI		090	090	090	090	090 - NEGATIVE	36	36	36	100	
2020-062	067	URISCAN		090	090	090	090	090 - NEGATIVE	49	49	49	100	
2020-062	071	ATLAS		090	090	090	090	090 - NEGATIVE	21	21	21	100	
2020-062	072	URS-9-TECO		090	090	090	090	090 - NEGATIVE	109	108	108	100	
2020-062	072	URS-9-TECO		090	090	090	090	091 - TRACE	109	1	108	0	
2020-062	080	ROCHE (CHE		090	090	090	090	090 - NEGATIVE	35	35	35	100	
2020-062	089	AIMSTRIP		090	090	090	090	090 - NEGATIVE	113	113	113	100	
2020-062	094	CLINITEK		090	090	090	090	090 - NEGATIVE	106	105	105	100	
2020-062	094	CLINITEK		090	090	090	090	094 - LARGE-3-4+	106	1	105	0	
2020-062	102	ACON LAB		090	090	090	090	090 - NEGATIVE	176	176	176	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	117	URS 10 A		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-062	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	28	28	28	100	
2020-062	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100	
2020-062	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-062	140	MISSION A		090 090	090 090		090 - NEGATIVE	63	63	63	100	
2020-062	147	Mission120		090 090	090 090		090 - NEGATIVE	11	11	11	100	
2020-062	148	Mission500		090 090	090 090		090 - NEGATIVE	11	11	11	100	
060 URINALYSIS-KETONES												
2020-063	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	50	50	50	100	
2020-063	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	36	36	36	100	
2020-063	067	URISCAN		090 090	090 090		090 - NEGATIVE	49	49	49	100	
2020-063	071	ATLAS		090 090	090 090		090 - NEGATIVE	21	21	21	100	
2020-063	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	109	109	109	100	
2020-063	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	35	35	35	100	
2020-063	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	113	112	112	100	
2020-063	089	AIMSTRIP		090 090	090 090		094 - LARGE-3-4+	113	1	112	0	
2020-063	094	CLINITEK		090 090	090 090		090 - NEGATIVE	106	105	105	100	
2020-063	094	CLINITEK		090 090	090 090		094 - LARGE-3-4+	106	1	105	0	
2020-063	102	ACON LAB		090 090	090 090		090 - NEGATIVE	176	176	176	100	
2020-063	117	URS 10 A		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-063	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	28	28	28	100	
2020-063	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100	
2020-063	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-063	140	MISSION A		090 090	090 090		090 - NEGATIVE	63	63	63	100	
2020-063	147	Mission120		090 090	090 090		090 - NEGATIVE	11	11	11	100	
2020-063	148	Mission500		090 090	090 090		090 - NEGATIVE	11	11	11	100	
2020-064	011	AMES NM10S		093 094	093 094		093 - MOD. (2+)	50	1	50	100	
2020-064	011	AMES NM10S		093 094	093 094		094 - LARGE-3-4+	50	49	50	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-064	013	BAYER CLI		093 094	093 094		094 - LARGE-3-4+	36	36	36	100
2020-064	067	URISCAN		093 094	093 094		093 - MOD. (2+)	49	16	49	100
2020-064	067	URISCAN		093 094	093 094		094 - LARGE-3-4+	49	33	49	100
2020-064	071	ATLAS		093 094	093 094		094 - LARGE-3-4+	21	21	21	100
2020-064	072	URS-9-TECO		093 094	093 094		092 - SMALL (1+)	109	2	107	0
2020-064	072	URS-9-TECO		093 094	093 094		093 - MOD. (2+)	109	12	107	100
2020-064	072	URS-9-TECO		093 094	093 094		094 - LARGE-3-4+	109	95	107	100
2020-064	080	ROCHE (CHE		093 094	093 094		093 - MOD. (2+)	35	5	35	100
2020-064	080	ROCHE (CHE		093 094	093 094		094 - LARGE-3-4+	35	30	35	100
2020-064	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	113	1	113	100
2020-064	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	113	112	113	100
2020-064	094	CLINITEK		093 094	093 094		090 - NEGATIVE	106	1	105	0
2020-064	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	106	105	105	100
2020-064	102	ACON LAB		093 094	093 094		090 - NEGATIVE	176	2	174	0
2020-064	102	ACON LAB		093 094	093 094		093 - MOD. (2+)	176	4	174	100
2020-064	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	176	170	174	100
2020-064	117	URS 10 A		093 094	093 094		093 - MOD. (2+)	10	2	10	100
2020-064	117	URS 10 A		093 094	093 094		094 - LARGE-3-4+	10	8	10	100
2020-064	131	iCHEM VELO		092 094	093 094		093 - MOD. (2+)	28	18	28	100
2020-064	131	iCHEM VELO		092 094	093 094		094 - LARGE-3-4+	28	10	28	100
2020-064	135	Hybrido AU		093 094	093 094		094 - LARGE-3-4+	12	12	12	100
2020-064	138	URS-10T		093 094	093 094		092 - SMALL (1+)	10	1	9	0
2020-064	138	URS-10T		093 094	093 094		093 - MOD. (2+)	10	1	9	100
2020-064	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	10	8	9	100
2020-064	140	MISSION A		093 094	093 094		092 - SMALL (1+)	63	1	62	0
2020-064	140	MISSION A		093 094	093 094		093 - MOD. (2+)	63	2	62	100
2020-064	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	63	60	62	100
2020-064	147	Mission120		093 094	093 094		094 - LARGE-3-4+	11	11	11	100
2020-064	148	Mission500		093 094	093 094		094 - LARGE-3-4+	11	11	11	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
060 URINALYSIS-KETONES											
2020-065	011 AMES NM10S			092 094	092 094		090 - NEGATIVE	50	1	49	0
2020-065	011 AMES NM10S			092 094	092 094		092 - SMALL (1+)	50	2	49	100
2020-065	011 AMES NM10S			092 094	092 094		093 - MOD. (2+)	50	28	49	100
2020-065	011 AMES NM10S			092 094	092 094		094 - LARGE-3-4+	50	19	49	100
2020-065	013 BAYER CLI			092 094	092 094		092 - SMALL (1+)	36	3	36	100
2020-065	013 BAYER CLI			092 094	092 094		093 - MOD. (2+)	36	26	36	100
2020-065	013 BAYER CLI			092 094	092 094		094 - LARGE-3-4+	36	7	36	100
2020-065	067 URISCAN			092 094	092 094		092 - SMALL (1+)	49	2	49	100
2020-065	067 URISCAN			092 094	092 094		093 - MOD. (2+)	49	31	49	100
2020-065	067 URISCAN			092 094	092 094		094 - LARGE-3-4+	49	16	49	100
2020-065	071 ATLAS			092 094	092 094		093 - MOD. (2+)	21	21	21	100
2020-065	072 URS-9-TECO			093 094	092 094		091 - TRACE	109	2	107	0
2020-065	072 URS-9-TECO			093 094	092 094		092 - SMALL (1+)	109	3	107	100
2020-065	072 URS-9-TECO			093 094	092 094		093 - MOD. (2+)	109	41	107	100
2020-065	072 URS-9-TECO			093 094	092 094		094 - LARGE-3-4+	109	63	107	100
2020-065	080 ROCHE (CHE			093 094	092 094		092 - SMALL (1+)	35	6	35	100
2020-065	080 ROCHE (CHE			093 094	092 094		093 - MOD. (2+)	35	10	35	100
2020-065	080 ROCHE (CHE			093 094	092 094		094 - LARGE-3-4+	35	19	35	100
2020-065	089 AIMSTRIP			092 094	092 094		092 - SMALL (1+)	113	4	113	100
2020-065	089 AIMSTRIP			092 094	092 094		093 - MOD. (2+)	113	61	113	100
2020-065	089 AIMSTRIP			092 094	092 094		094 - LARGE-3-4+	113	48	113	100
2020-065	094 CLINITEK			092 094	092 094		092 - SMALL (1+)	106	8	106	100
2020-065	094 CLINITEK			092 094	092 094		093 - MOD. (2+)	106	86	106	100
2020-065	094 CLINITEK			092 094	092 094		094 - LARGE-3-4+	106	12	106	100
2020-065	102 ACON LAB			093 094	092 094		092 - SMALL (1+)	176	11	176	100
2020-065	102 ACON LAB			093 094	092 094		093 - MOD. (2+)	176	53	176	100
2020-065	102 ACON LAB			093 094	092 094		094 - LARGE-3-4+	176	112	176	100
2020-065	117 URS 10 A			092 094	092 094		093 - MOD. (2+)	10	6	10	100
2020-065	117 URS 10 A			092 094	092 094		094 - LARGE-3-4+	10	4	10	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	131	iCHEM VELO		092 094	092 094		092 - SMALL (1+)	28	11	28	100	
2020-065	131	iCHEM VELO		092 094	092 094		093 - MOD. (2+)	28	14	28	100	
2020-065	131	iCHEM VELO		092 094	092 094		094 - LARGE-3-4+	28	3	28	100	
2020-065	135	Hybrido AU		092 094	092 094		093 - MOD. (2+)	12	9	12	100	
2020-065	135	Hybrido AU		092 094	092 094		094 - LARGE-3-4+	12	3	12	100	
2020-065	138	URS-10T		093 094	092 094		092 - SMALL (1+)	10	2	10	100	
2020-065	138	URS-10T		093 094	092 094		093 - MOD. (2+)	10	3	10	100	
2020-065	138	URS-10T		093 094	092 094		094 - LARGE-3-4+	10	5	10	100	
2020-065	140	MISSION A		093 094	092 094		090 - NEGATIVE	63	1	62	0	
2020-065	140	MISSION A		093 094	092 094		092 - SMALL (1+)	63	1	62	100	
2020-065	140	MISSION A		093 094	092 094		093 - MOD. (2+)	63	11	62	100	
2020-065	140	MISSION A		093 094	092 094		094 - LARGE-3-4+	63	50	62	100	
2020-065	147	Mission120		093 094	092 094		093 - MOD. (2+)	11	1	11	100	
2020-065	147	Mission120		093 094	092 094		094 - LARGE-3-4+	11	10	11	100	
2020-065	148	Mission500		093 094	092 094		093 - MOD. (2+)	11	3	11	100	
2020-065	148	Mission500		093 094	092 094		094 - LARGE-3-4+	11	8	11	100	
061 URINALYSIS-NITRATE												
2020-061	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2020-061	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100	
2020-061	067	URISCAN		070 070	070 070		070 - NEGATIVE	49	49	49	100	
2020-061	071	ATLAS		070 070	070 070		070 - NEGATIVE	21	21	21	100	
2020-061	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	108	108	108	100	
2020-061	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	35	35	35	100	
2020-061	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	113	113	113	100	
2020-061	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100	
2020-061	102	ACON LAB		070 070	070 070		070 - NEGATIVE	176	176	176	100	
2020-061	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2020-061	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	28	28	28	100	
2020-061	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2020-061	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-061	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100		
2020-061	147	Mission120		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-061	148	Mission500		070 070	070 070		070 - NEGATIVE	11	11	11	100		
061 URINALYSIS-NITRATE													
2020-062	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100		
2020-062	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100		
2020-062	067	URISCAN		070 070	070 070		070 - NEGATIVE	49	49	49	100		
2020-062	071	ATLAS		070 070	070 070		070 - NEGATIVE	21	21	21	100		
2020-062	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	108	108	108	100		
2020-062	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	35	35	35	100		
2020-062	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	113	113	113	100		
2020-062	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100		
2020-062	102	ACON LAB		070 070	070 070		070 - NEGATIVE	176	176	176	100		
2020-062	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-062	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	28	28	28	100		
2020-062	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100		
2020-062	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-062	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100		
2020-062	147	Mission120		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-062	148	Mission500		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-063	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100		
2020-063	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100		
2020-063	067	URISCAN		070 070	070 070		070 - NEGATIVE	49	49	49	100		
2020-063	071	ATLAS		070 070	070 070		070 - NEGATIVE	21	21	21	100		
2020-063	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	108	108	108	100		
2020-063	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	35	35	35	100		
2020-063	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	113	113	113	100		
2020-063	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-063	102	ACON LAB		070 070	070 070		070 - NEGATIVE	176	176	176	100		
2020-063	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-063	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	28	28	28	100		
2020-063	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100		
2020-063	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-063	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100		
2020-063	147	Mission120		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-063	148	Mission500		070 070	070 070		070 - NEGATIVE	11	11	11	100		
061 URINALYSIS-NITRATE													
2020-064	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100		
2020-064	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100		
2020-064	067	URISCAN		070 070	070 070		070 - NEGATIVE	49	49	49	100		
2020-064	071	ATLAS		070 070	070 070		070 - NEGATIVE	21	21	21	100		
2020-064	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	108	107	107	100		
2020-064	072	URS-9-TECO		070 070	070 070		072 - POSITIVE	108	1	107	0		
2020-064	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	35	35	35	100		
2020-064	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	113	113	113	100		
2020-064	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100		
2020-064	102	ACON LAB		070 070	070 070		070 - NEGATIVE	176	173	173	100		
2020-064	102	ACON LAB		070 070	070 070		072 - POSITIVE	176	3	173	0		
2020-064	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-064	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	28	28	28	100		
2020-064	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100		
2020-064	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2020-064	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100		
2020-064	147	Mission120		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-064	148	Mission500		070 070	070 070		070 - NEGATIVE	11	11	11	100		
2020-065	011	AMES NM10S		071 072	071 072		070 - NEGATIVE	50	1	49	0		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	011	AMES NM10S		071 072	071 072		072 - POSITIVE	50	49	49	100	
2020-065	013	BAYER CLI		071 072	071 072		072 - POSITIVE	37	37	37	100	
2020-065	067	URISCAN		071 072	071 072		072 - POSITIVE	49	49	49	100	
2020-065	071	ATLAS		071 072	071 072		072 - POSITIVE	21	21	21	100	
2020-065	072	URS-9-TECO		071 072	071 072		071 - WEAK POS.	108	3	108	100	
2020-065	072	URS-9-TECO		071 072	071 072		072 - POSITIVE	108	105	108	100	
2020-065	080	ROCHE (CHE		071 072	071 072		072 - POSITIVE	35	35	35	100	
2020-065	089	AIMSTRIP		071 072	071 072		071 - WEAK POS.	113	1	113	100	
2020-065	089	AIMSTRIP		071 072	071 072		072 - POSITIVE	113	112	113	100	
2020-065	094	CLINITEK		071 072	071 072		072 - POSITIVE	106	106	106	100	
2020-065	102	ACON LAB		071 072	071 072		070 - NEGATIVE	176	2	174	0	
2020-065	102	ACON LAB		071 072	071 072		071 - WEAK POS.	176	1	174	100	
2020-065	102	ACON LAB		071 072	071 072		072 - POSITIVE	176	173	174	100	
2020-065	117	URS 10 A		071 072	071 072		072 - POSITIVE	10	10	10	100	
2020-065	131	iCHEM VELO		071 072	071 072		072 - POSITIVE	28	28	28	100	
2020-065	135	Hybrido AU		071 072	071 072		072 - POSITIVE	12	12	12	100	
2020-065	138	URS-10T		071 072	071 072		072 - POSITIVE	10	10	10	100	
2020-065	140	MISSION A		071 072	071 072		070 - NEGATIVE	63	3	60	0	
2020-065	140	MISSION A		071 072	071 072		072 - POSITIVE	63	60	60	100	
2020-065	147	Mission120		071 072	071 072		072 - POSITIVE	11	11	11	100	
2020-065	148	Mission500		071 072	071 072		072 - POSITIVE	11	11	11	100	
062 URINALYSIS-pH												
2020-061	011	AMES NM10S		085 088	084 088		086 - 7.0	50	22	50	100	
2020-061	011	AMES NM10S		085 088	084 088		087 - 7.5	50	28	50	100	
2020-061	013	BAYER CLI		085 088	084 088		086 - 7.0	37	16	37	100	
2020-061	013	BAYER CLI		085 088	084 088		087 - 7.5	37	21	37	100	
2020-061	067	URISCAN		084 088	084 088		085 - 6.5	49	9	49	100	
2020-061	067	URISCAN		084 088	084 088		086 - 7.0	49	39	49	100	
2020-061	067	URISCAN		084 088	084 088		087 - 7.5	49	1	49	100	
2020-061	071	ATLAS		084 088	084 088		086 - 7.0	21	20	21	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-061	071	ATLAS		084 088	084 088		087 - 7.5	21	1	21	100	
2020-061	072	URS-9-TECO		084 088	084 088		084 - 6.0	107	1	107	100	
2020-061	072	URS-9-TECO		084 088	084 088		085 - 6.5	107	4	107	100	
2020-061	072	URS-9-TECO		084 088	084 088		086 - 7.0	107	97	107	100	
2020-061	072	URS-9-TECO		084 088	084 088		087 - 7.5	107	5	107	100	
2020-061	080	ROCHE (CHE		084 088	084 088		086 - 7.0	35	35	35	100	
2020-061	089	AIMSTRIP		084 088	084 088		086 - 7.0	113	86	113	100	
2020-061	089	AIMSTRIP		084 088	084 088		087 - 7.5	113	27	113	100	
2020-061	094	CLINITEK		085 088	084 088		086 - 7.0	106	41	106	100	
2020-061	094	CLINITEK		085 088	084 088		087 - 7.5	106	65	106	100	
2020-061	102	ACON LAB		084 088	084 088		084 - 6.0	177	1	177	100	
2020-061	102	ACON LAB		084 088	084 088		085 - 6.5	177	4	177	100	
2020-061	102	ACON LAB		084 088	084 088		086 - 7.0	177	145	177	100	
2020-061	102	ACON LAB		084 088	084 088		087 - 7.5	177	27	177	100	
2020-061	117	URS 10 A		084 088	084 088		086 - 7.0	10	10	10	100	
2020-061	131	iCHEM VELO		084 088	084 088		086 - 7.0	28	28	28	100	
2020-061	135	Hybrido AU		085 088	084 088		086 - 7.0	12	5	12	100	
2020-061	135	Hybrido AU		085 088	084 088		087 - 7.5	12	7	12	100	
2020-061	138	URS-10T		084 088	084 088		085 - 6.5	10	1	10	100	
2020-061	138	URS-10T		084 088	084 088		086 - 7.0	10	9	10	100	
2020-061	140	MISSION A		084 088	084 088		086 - 7.0	63	61	63	100	
2020-061	140	MISSION A		084 088	084 088		087 - 7.5	63	2	63	100	
2020-061	147	Mission120		084 088	084 088		086 - 7.0	11	10	11	100	
2020-061	147	Mission120		084 088	084 088		087 - 7.5	11	1	11	100	
2020-061	148	Mission500		084 088	084 088		086 - 7.0	11	11	11	100	
062 URINALYSIS-pH												
2020-062	011	AMES NM10S		086 088	086 088		086 - 7.0	50	2	50	100	
2020-062	011	AMES NM10S		086 088	086 088		087 - 7.5	50	3	50	100	
2020-062	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	50	45	50	100	
2020-062	013	BAYER CLI		086 088	086 088		087 - 7.5	37	3	37	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-062	013	BAYER CLI		086 088	086 088		088 - 8.00R MORE	37	34	37	100		
2020-062	067	URISCAN		086 088	086 088		087 - 7.5	49	17	49	100		
2020-062	067	URISCAN		086 088	086 088		088 - 8.00R MORE	49	32	49	100		
2020-062	071	ATLAS		086 088	086 088		088 - 8.00R MORE	21	21	21	100		
2020-062	072	URS-9-TECO		085 088	086 088		084 - 6.0	107	1	106	0		
2020-062	072	URS-9-TECO		085 088	086 088		085 - 6.5	107	2	106	100		
2020-062	072	URS-9-TECO		085 088	086 088		086 - 7.0	107	4	106	100		
2020-062	072	URS-9-TECO		085 088	086 088		087 - 7.5	107	53	106	100		
2020-062	072	URS-9-TECO		085 088	086 088		088 - 8.00R MORE	107	47	106	100		
2020-062	080	ROCHE (CHE		086 088	086 088		088 - 8.00R MORE	35	35	35	100		
2020-062	089	AIMSTRIP		086 088	086 088		087 - 7.5	113	10	113	100		
2020-062	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	113	103	113	100		
2020-062	094	CLINITEK		086 088	086 088		087 - 7.5	106	10	106	100		
2020-062	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	106	96	106	100		
2020-062	102	ACON LAB		086 088	086 088		086 - 7.0	177	1	177	100		
2020-062	102	ACON LAB		086 088	086 088		087 - 7.5	177	8	177	100		
2020-062	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	177	168	177	100		
2020-062	117	URS 10 A		085 088	086 088		087 - 7.5	10	6	10	100		
2020-062	117	URS 10 A		085 088	086 088		088 - 8.00R MORE	10	4	10	100		
2020-062	131	iCHEM VELO		086 088	086 088		088 - 8.00R MORE	28	28	28	100		
2020-062	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	12	12	12	100		
2020-062	138	URS-10T		086 088	086 088		087 - 7.5	10	5	10	100		
2020-062	138	URS-10T		086 088	086 088		088 - 8.00R MORE	10	5	10	100		
2020-062	140	MISSION A		086 088	086 088		087 - 7.5	63	2	63	100		
2020-062	140	MISSION A		086 088	086 088		088 - 8.00R MORE	63	61	63	100		
2020-062	147	Mission120		086 088	086 088		087 - 7.5	11	1	11	100		
2020-062	147	Mission120		086 088	086 088		088 - 8.00R MORE	11	10	11	100		
2020-062	148	Mission500		086 088	086 088		087 - 7.5	11	1	11	100		
2020-062	148	Mission500		086 088	086 088		088 - 8.00R MORE	11	10	11	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
062 URINALYSIS-pH													
2020-063	011 AMES NM10S			082 086	082 086		083 - 5.5	50	7	50	100		
2020-063	011 AMES NM10S			082 086	082 086		084 - 6.0	50	43	50	100		
2020-063	013 BAYER CLI			082 086	082 086		083 - 5.5	37	3	37	100		
2020-063	013 BAYER CLI			082 086	082 086		084 - 6.0	37	34	37	100		
2020-063	067 URISCAN			081 085	082 086		082 - 5.0	49	14	48	100		
2020-063	067 URISCAN			081 085	082 086		083 - 5.5	49	27	48	100		
2020-063	067 URISCAN			081 085	082 086		084 - 6.0	49	7	48	100		
2020-063	067 URISCAN			081 085	082 086		087 - 7.5	49	1	48	0		
2020-063	071 ATLAS			082 086	082 086		084 - 6.0	21	21	21	100		
2020-063	072 URS-9-TECO			082 086	082 086		082 - 5.0	107	2	107	100		
2020-063	072 URS-9-TECO			082 086	082 086		083 - 5.5	107	12	107	100		
2020-063	072 URS-9-TECO			082 086	082 086		084 - 6.0	107	68	107	100		
2020-063	072 URS-9-TECO			082 086	082 086		085 - 6.5	107	25	107	100		
2020-063	080 ROCHE (CHE			082 086	082 086		082 - 5.0	35	4	35	100		
2020-063	080 ROCHE (CHE			082 086	082 086		084 - 6.0	35	31	35	100		
2020-063	089 AIMSTRIP			082 086	082 086		082 - 5.0	113	5	113	100		
2020-063	089 AIMSTRIP			082 086	082 086		083 - 5.5	113	6	113	100		
2020-063	089 AIMSTRIP			082 086	082 086		084 - 6.0	113	101	113	100		
2020-063	089 AIMSTRIP			082 086	082 086		085 - 6.5	113	1	113	100		
2020-063	094 CLINITEK			082 086	082 086		083 - 5.5	106	3	106	100		
2020-063	094 CLINITEK			082 086	082 086		084 - 6.0	106	103	106	100		
2020-063	102 ACON LAB			082 086	082 086		082 - 5.0	177	6	176	100		
2020-063	102 ACON LAB			082 086	082 086		083 - 5.5	177	2	176	100		
2020-063	102 ACON LAB			082 086	082 086		084 - 6.0	177	166	176	100		
2020-063	102 ACON LAB			082 086	082 086		085 - 6.5	177	2	176	100		
2020-063	102 ACON LAB			082 086	082 086		088 - 8.0OR MORE	177	1	176	0		
2020-063	117 URS 10 A			082 086	082 086		083 - 5.5	10	1	10	100		
2020-063	117 URS 10 A			082 086	082 086		084 - 6.0	10	6	10	100		
2020-063	117 URS 10 A			082 086	082 086		085 - 6.5	10	3	10	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-063	131	iCHEM VELO		082 086	082 086		082 - 5.0	28	4	28	100
2020-063	131	iCHEM VELO		082 086	082 086		084 - 6.0	28	24	28	100
2020-063	135	Hybrido AU		082 086	082 086		084 - 6.0	12	12	12	100
2020-063	138	URS-10T		082 086	082 086		083 - 5.5	10	3	10	100
2020-063	138	URS-10T		082 086	082 086		084 - 6.0	10	6	10	100
2020-063	138	URS-10T		082 086	082 086		085 - 6.5	10	1	10	100
2020-063	140	MISSION A		082 086	082 086		084 - 6.0	63	63	63	100
2020-063	147	Mission120		082 086	082 086		084 - 6.0	11	11	11	100
2020-063	148	Mission500		082 086	082 086		083 - 5.5	11	1	11	100
2020-063	148	Mission500		082 086	082 086		084 - 6.0	11	10	11	100
062 URINALYSIS-pH											
2020-064	011	AMES NM10S		086 088	086 088		085 - 6.5	50	1	49	0
2020-064	011	AMES NM10S		086 088	086 088		087 - 7.5	50	7	49	100
2020-064	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	50	42	49	100
2020-064	013	BAYER CLI		086 088	086 088		087 - 7.5	37	7	37	100
2020-064	013	BAYER CLI		086 088	086 088		088 - 8.0OR MORE	37	30	37	100
2020-064	067	URISCAN		085 088	086 088		087 - 7.5	49	38	49	100
2020-064	067	URISCAN		085 088	086 088		088 - 8.0OR MORE	49	11	49	100
2020-064	071	ATLAS		085 088	086 088		087 - 7.5	21	14	21	100
2020-064	071	ATLAS		085 088	086 088		088 - 8.0OR MORE	21	7	21	100
2020-064	072	URS-9-TECO		085 088	086 088		086 - 7.0	107	12	107	100
2020-064	072	URS-9-TECO		085 088	086 088		087 - 7.5	107	65	107	100
2020-064	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	107	30	107	100
2020-064	080	ROCHE (CHE		086 088	086 088		086 - 7.0	35	2	35	100
2020-064	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	35	33	35	100
2020-064	089	AIMSTRIP		086 088	086 088		087 - 7.5	113	22	113	100
2020-064	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	113	91	113	100
2020-064	094	CLINITEK		086 088	086 088		087 - 7.5	106	15	106	100
2020-064	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	106	91	106	100
2020-064	102	ACON LAB		086 088	086 088		084 - 6.0	177	1	176	0

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-064	102	ACON LAB		086 088	086 088		087 - 7.5	177	23	176	100	
2020-064	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	177	153	176	100	
2020-064	117	URS 10 A		085 088	086 088		086 - 7.0	10	2	10	100	
2020-064	117	URS 10 A		085 088	086 088		087 - 7.5	10	6	10	100	
2020-064	117	URS 10 A		085 088	086 088		088 - 8.00R MORE	10	2	10	100	
2020-064	131	iCHEM VELO		084 088	086 088		086 - 7.0	28	15	28	100	
2020-064	131	iCHEM VELO		084 088	086 088		088 - 8.00R MORE	28	13	28	100	
2020-064	135	Hybrido AU		086 088	086 088		087 - 7.5	12	1	12	100	
2020-064	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	12	11	12	100	
2020-064	138	URS-10T		085 088	086 088		087 - 7.5	10	7	10	100	
2020-064	138	URS-10T		085 088	086 088		088 - 8.00R MORE	10	3	10	100	
2020-064	140	MISSION A		086 088	086 088		087 - 7.5	63	14	63	100	
2020-064	140	MISSION A		086 088	086 088		088 - 8.00R MORE	63	49	63	100	
2020-064	147	Mission120		086 088	086 088		087 - 7.5	11	2	11	100	
2020-064	147	Mission120		086 088	086 088		088 - 8.00R MORE	11	9	11	100	
2020-064	148	Mission500		086 088	086 088		087 - 7.5	11	1	11	100	
2020-064	148	Mission500		086 088	086 088		088 - 8.00R MORE	11	10	11	100	
062 URINALYSIS-pH												
2020-065	011	AMES NM10S		080 084	080 084		082 - 5.0	50	25	49	100	
2020-065	011	AMES NM10S		080 084	080 084		083 - 5.5	50	15	49	100	
2020-065	011	AMES NM10S		080 084	080 084		084 - 6.0	50	9	49	100	
2020-065	011	AMES NM10S		080 084	080 084		087 - 7.5	50	1	49	0	
2020-065	013	BAYER CLI		080 084	080 084		082 - 5.0	37	17	37	100	
2020-065	013	BAYER CLI		080 084	080 084		083 - 5.5	37	16	37	100	
2020-065	013	BAYER CLI		080 084	080 084		084 - 6.0	37	4	37	100	
2020-065	067	URISCAN		080 084	080 084		082 - 5.0	49	42	49	100	
2020-065	067	URISCAN		080 084	080 084		083 - 5.5	49	7	49	100	
2020-065	071	ATLAS		080 084	080 084		082 - 5.0	21	13	21	100	
2020-065	071	ATLAS		080 084	080 084		083 - 5.5	21	8	21	100	
2020-065	072	URS-9-TECO		082 086	080 084		082 - 5.0	107	22	107	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-065	072	URS-9-TECO		082 086	080 084		083 - 5.5	107	20	107	100		
2020-065	072	URS-9-TECO		082 086	080 084		084 - 6.0	107	62	107	100		
2020-065	072	URS-9-TECO		082 086	080 084		085 - 6.5	107	3	107	100		
2020-065	080	ROCHE (CHE		080 084	080 084		082 - 5.0	35	35	35	100		
2020-065	089	AIMSTRIP		080 084	080 084		082 - 5.0	113	68	113	100		
2020-065	089	AIMSTRIP		080 084	080 084		083 - 5.5	113	29	113	100		
2020-065	089	AIMSTRIP		080 084	080 084		084 - 6.0	113	16	113	100		
2020-065	094	CLINITEK		081 085	080 084		082 - 5.0	106	31	106	100		
2020-065	094	CLINITEK		081 085	080 084		083 - 5.5	106	68	106	100		
2020-065	094	CLINITEK		081 085	080 084		084 - 6.0	106	7	106	100		
2020-065	102	ACON LAB		082 086	080 084		082 - 5.0	177	53	176	100		
2020-065	102	ACON LAB		082 086	080 084		083 - 5.5	177	40	176	100		
2020-065	102	ACON LAB		082 086	080 084		084 - 6.0	177	82	176	100		
2020-065	102	ACON LAB		082 086	080 084		085 - 6.5	177	1	176	100		
2020-065	102	ACON LAB		082 086	080 084		088 - 8.0OR MORE	177	1	176	0		
2020-065	117	URS 10 A		081 085	080 084		082 - 5.0	10	2	10	100		
2020-065	117	URS 10 A		081 085	080 084		083 - 5.5	10	6	10	100		
2020-065	117	URS 10 A		081 085	080 084		084 - 6.0	10	1	10	100		
2020-065	117	URS 10 A		081 085	080 084		085 - 6.5	10	1	10	100		
2020-065	131	iCHEM VELO		080 084	080 084		082 - 5.0	28	27	28	100		
2020-065	131	iCHEM VELO		080 084	080 084		083 - 5.5	28	1	28	100		
2020-065	135	Hybrido AU		080 084	080 084		082 - 5.0	12	12	12	100		
2020-065	138	URS-10T		082 086	080 084		082 - 5.0	10	3	10	100		
2020-065	138	URS-10T		082 086	080 084		083 - 5.5	10	3	10	100		
2020-065	138	URS-10T		082 086	080 084		084 - 6.0	10	4	10	100		
2020-065	140	MISSION A		081 085	080 084		083 - 5.5	63	33	62	100		
2020-065	140	MISSION A		081 085	080 084		084 - 6.0	63	22	62	100		
2020-065	140	MISSION A		081 085	080 084		088 - 8.0OR MORE	63	1	62	0		
2020-065	140	MISSION A		081 085	080 084		082 - 5.0	63	7	62	100		
2020-065	147	Mission120		081 085	080 084		082 - 5.0	11	4	11	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	147	Mission120		081 085	080 084		083 - 5.5	11	7	11	100	
2020-065	148	Mission500		082 086	080 084		082 - 5.0	11	1	11	100	
2020-065	148	Mission500		082 086	080 084		083 - 5.5	11	4	11	100	
2020-065	148	Mission500		082 086	080 084		084 - 6.0	11	6	11	100	
064 URIN-PROTEIN QUALI.												
2020-061	011	AMES NM10S		053 055	053 055		054 - 2+	51	37	51	100	
2020-061	011	AMES NM10S		053 055	053 055		053 - 1+	51	3	51	100	
2020-061	011	AMES NM10S		053 055	053 055		055 - 3+	51	11	51	100	
2020-061	013	BAYER CLI		053 055	053 055		054 - 2+	36	27	36	100	
2020-061	013	BAYER CLI		053 055	053 055		055 - 3+	36	9	36	100	
2020-061	067	URISCAN		053 055	053 055		054 - 2+	49	38	49	100	
2020-061	067	URISCAN		053 055	053 055		053 - 1+	49	9	49	100	
2020-061	067	URISCAN		053 055	053 055		055 - 3+	49	2	49	100	
2020-061	071	ATLAS		053 055	053 055		054 - 2+	21	18	21	100	
2020-061	071	ATLAS		053 055	053 055		053 - 1+	21	2	21	100	
2020-061	071	ATLAS		053 055	053 055		055 - 3+	21	1	21	100	
2020-061	072	URS-9-TECO		053 055	053 055		054 - 2+	108	71	106	100	
2020-061	072	URS-9-TECO		053 055	053 055		053 - 1+	108	2	106	100	
2020-061	072	URS-9-TECO		053 055	053 055		055 - 3+	108	33	106	100	
2020-061	072	URS-9-TECO		053 055	053 055		056 - 4+	108	2	106	0	
2020-061	080	ROCHE (CHE		053 055	053 055		054 - 2+	35	22	34	100	
2020-061	080	ROCHE (CHE		053 055	053 055		051 - NEGATIVE	35	1	34	0	
2020-061	080	ROCHE (CHE		053 055	053 055		055 - 3+	35	12	34	100	
2020-061	089	AIMSTRIP		053 055	053 055		054 - 2+	112	88	110	100	
2020-061	089	AIMSTRIP		053 055	053 055		053 - 1+	112	5	110	100	
2020-061	089	AIMSTRIP		053 055	053 055		055 - 3+	112	17	110	100	
2020-061	089	AIMSTRIP		053 055	053 055		056 - 4+	112	2	110	0	
2020-061	094	CLINITEK		053 055	053 055		054 - 2+	106	90	104	100	
2020-061	094	CLINITEK		053 055	053 055		053 - 1+	106	4	104	100	
2020-061	094	CLINITEK		053 055	053 055		055 - 3+	106	10	104	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-061	094	CLINITEK		053	055	053	055	056 - 4+	106	2	104	0
2020-061	102	ACON LAB		053	055	053	055	054 - 2+	177	128	175	100
2020-061	102	ACON LAB		053	055	053	055	052 - TRACES	177	2	175	0
2020-061	102	ACON LAB		053	055	053	055	053 - 1+	177	7	175	100
2020-061	102	ACON LAB		053	055	053	055	055 - 3+	177	40	175	100
2020-061	117	URS 10 A		053	055	053	055	054 - 2+	10	6	10	100
2020-061	117	URS 10 A		053	055	053	055	055 - 3+	10	4	10	100
2020-061	131	iCHEM VELO		053	055	053	055	054 - 2+	28	27	28	100
2020-061	131	iCHEM VELO		053	055	053	055	055 - 3+	28	1	28	100
2020-061	135	Hybrido AU		053	055	053	055	054 - 2+	12	11	12	100
2020-061	135	Hybrido AU		053	055	053	055	055 - 3+	12	1	12	100
2020-061	138	URS-10T		054	056	053	055	054 - 2+	10	3	10	100
2020-061	138	URS-10T		054	056	053	055	053 - 1+	10	2	10	100
2020-061	138	URS-10T		054	056	053	055	055 - 3+	10	4	10	100
2020-061	138	URS-10T		054	056	053	055	056 - 4+	10	1	10	100
2020-061	140	MISSION A		053	055	053	055	054 - 2+	63	55	63	100
2020-061	140	MISSION A		053	055	053	055	053 - 1+	63	3	63	100
2020-061	140	MISSION A		053	055	053	055	055 - 3+	63	5	63	100
2020-061	147	Mission120		053	055	053	055	054 - 2+	11	10	11	100
2020-061	147	Mission120		053	055	053	055	055 - 3+	11	1	11	100
2020-061	148	Mission500		053	055	053	055	054 - 2+	11	7	11	100
2020-061	148	Mission500		053	055	053	055	053 - 1+	11	1	11	100
2020-061	148	Mission500		053	055	053	055	055 - 3+	11	3	11	100
064 URIN-PROTEIN QUALI.												
2020-062	011	AMES NM10S		051	051	051	051	051 - NEGATIVE	51	51	51	100
2020-062	013	BAYER CLI		051	051	051	051	051 - NEGATIVE	36	36	36	100
2020-062	067	URISCAN		051	051	051	051	051 - NEGATIVE	49	48	48	100
2020-062	067	URISCAN		051	051	051	051	053 - 1+	49	1	48	0
2020-062	071	ATLAS		051	051	051	051	051 - NEGATIVE	21	21	21	100
2020-062	072	URS-9-TECO		051	051	051	051	051 - NEGATIVE	108	108	108	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	35	35	35	100	
2020-062	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	112	110	110	100	
2020-062	089	AIMSTRIP		051 051	051 051		052 - TRACES	112	2	110	0	
2020-062	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	106	106	100	
2020-062	102	ACON LAB		051 051	051 051		051 - NEGATIVE	177	172	172	100	
2020-062	102	ACON LAB		051 051	051 051		052 - TRACES	177	4	172	0	
2020-062	102	ACON LAB		051 051	051 051		055 - 3+	177	1	172	0	
2020-062	117	URS 10 A		051 051	051 051		051 - NEGATIVE	10	10	10	100	
2020-062	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	28	28	28	100	
2020-062	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100	
2020-062	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100	
2020-062	140	MISSION A		051 051	051 051		051 - NEGATIVE	63	62	62	100	
2020-062	140	MISSION A		051 051	051 051		052 - TRACES	63	1	62	0	
2020-062	147	Mission120		051 051	051 051		051 - NEGATIVE	11	11	11	100	
2020-062	148	Mission500		051 051	051 051		051 - NEGATIVE	11	11	11	100	
064 URIN-PROTEIN QUALI.												
2020-063	011	AMES NM10S		053 055	053 055		054 - 2+	51	43	51	100	
2020-063	011	AMES NM10S		053 055	053 055		053 - 1+	51	2	51	100	
2020-063	011	AMES NM10S		053 055	053 055		055 - 3+	51	6	51	100	
2020-063	013	BAYER CLI		053 055	053 055		054 - 2+	36	30	36	100	
2020-063	013	BAYER CLI		053 055	053 055		053 - 1+	36	2	36	100	
2020-063	013	BAYER CLI		053 055	053 055		055 - 3+	36	4	36	100	
2020-063	067	URISCAN		053 055	053 055		054 - 2+	49	31	49	100	
2020-063	067	URISCAN		053 055	053 055		053 - 1+	49	18	49	100	
2020-063	071	ATLAS		053 055	053 055		054 - 2+	21	19	21	100	
2020-063	071	ATLAS		053 055	053 055		053 - 1+	21	2	21	100	
2020-063	072	URS-9-TECO		053 055	053 055		054 - 2+	108	83	108	100	
2020-063	072	URS-9-TECO		053 055	053 055		053 - 1+	108	9	108	100	
2020-063	072	URS-9-TECO		053 055	053 055		055 - 3+	108	16	108	100	
2020-063	080	ROCHE (CHE		053 055	053 055		054 - 2+	35	26	35	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-063	080	ROCHE (CHE		053 055	053 055		055 - 3+	35	9	35	100	
2020-063	089	AIMSTRIP		053 055	053 055		054 - 2+	112	94	112	100	
2020-063	089	AIMSTRIP		053 055	053 055		053 - 1+	112	13	112	100	
2020-063	089	AIMSTRIP		053 055	053 055		055 - 3+	112	5	112	100	
2020-063	094	CLINITEK		053 055	053 055		054 - 2+	106	94	105	100	
2020-063	094	CLINITEK		053 055	053 055		053 - 1+	106	6	105	100	
2020-063	094	CLINITEK		053 055	053 055		055 - 3+	106	5	105	100	
2020-063	094	CLINITEK		053 055	053 055		056 - 4+	106	1	105	0	
2020-063	102	ACON LAB		053 055	053 055		054 - 2+	177	152	174	100	
2020-063	102	ACON LAB		053 055	053 055		051 - NEGATIVE	177	1	174	0	
2020-063	102	ACON LAB		053 055	053 055		052 - TRACES	177	2	174	0	
2020-063	102	ACON LAB		053 055	053 055		053 - 1+	177	10	174	100	
2020-063	102	ACON LAB		053 055	053 055		055 - 3+	177	12	174	100	
2020-063	117	URS 10 A		053 055	053 055		054 - 2+	10	9	10	100	
2020-063	117	URS 10 A		053 055	053 055		055 - 3+	10	1	10	100	
2020-063	131	iCHEM VELO		053 055	053 055		054 - 2+	28	27	28	100	
2020-063	131	iCHEM VELO		053 055	053 055		055 - 3+	28	1	28	100	
2020-063	135	Hybrido AU		053 055	053 055		054 - 2+	12	11	12	100	
2020-063	135	Hybrido AU		053 055	053 055		053 - 1+	12	1	12	100	
2020-063	138	URS-10T		053 055	053 055		054 - 2+	10	5	10	100	
2020-063	138	URS-10T		053 055	053 055		053 - 1+	10	3	10	100	
2020-063	138	URS-10T		053 055	053 055		055 - 3+	10	2	10	100	
2020-063	140	MISSION A		053 055	053 055		054 - 2+	63	51	63	100	
2020-063	140	MISSION A		053 055	053 055		053 - 1+	63	10	63	100	
2020-063	140	MISSION A		053 055	053 055		055 - 3+	63	2	63	100	
2020-063	147	Mission120		053 055	053 055		054 - 2+	11	6	11	100	
2020-063	147	Mission120		053 055	053 055		053 - 1+	11	4	11	100	
2020-063	147	Mission120		053 055	053 055		055 - 3+	11	1	11	100	
2020-063	148	Mission500		053 055	053 055		054 - 2+	11	8	11	100	
2020-063	148	Mission500		053 055	053 055		053 - 1+	11	2	11	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-063	148	Mission500		053 055	053 055		055 - 3+	11	1	11	100		
064 URIN-PROTEIN QUALI.													
2020-064	011	AMES NM10S		053 055	053 055		054 - 2+	51	37	51	100		
2020-064	011	AMES NM10S		053 055	053 055		053 - 1+	51	2	51	100		
2020-064	011	AMES NM10S		053 055	053 055		055 - 3+	51	12	51	100		
2020-064	013	BAYER CLI		053 055	053 055		054 - 2+	36	24	36	100		
2020-064	013	BAYER CLI		053 055	053 055		055 - 3+	36	12	36	100		
2020-064	067	URISCAN		053 055	053 055		054 - 2+	49	39	49	100		
2020-064	067	URISCAN		053 055	053 055		053 - 1+	49	3	49	100		
2020-064	067	URISCAN		053 055	053 055		055 - 3+	49	7	49	100		
2020-064	071	ATLAS		053 055	053 055		054 - 2+	21	19	21	100		
2020-064	071	ATLAS		053 055	053 055		053 - 1+	21	2	21	100		
2020-064	072	URS-9-TECO		053 055	053 055		054 - 2+	108	58	105	100		
2020-064	072	URS-9-TECO		053 055	053 055		053 - 1+	108	1	105	100		
2020-064	072	URS-9-TECO		053 055	053 055		055 - 3+	108	46	105	100		
2020-064	072	URS-9-TECO		053 055	053 055		056 - 4+	108	3	105	0		
2020-064	080	ROCHE (CHE		053 055	053 055		054 - 2+	35	24	35	100		
2020-064	080	ROCHE (CHE		053 055	053 055		053 - 1+	35	2	35	100		
2020-064	080	ROCHE (CHE		053 055	053 055		055 - 3+	35	9	35	100		
2020-064	089	AIMSTRIP		053 055	053 055		054 - 2+	112	79	109	100		
2020-064	089	AIMSTRIP		053 055	053 055		053 - 1+	112	4	109	100		
2020-064	089	AIMSTRIP		053 055	053 055		055 - 3+	112	26	109	100		
2020-064	089	AIMSTRIP		053 055	053 055		056 - 4+	112	3	109	0		
2020-064	094	CLINITEK		053 055	053 055		054 - 2+	106	87	104	100		
2020-064	094	CLINITEK		053 055	053 055		053 - 1+	106	4	104	100		
2020-064	094	CLINITEK		053 055	053 055		055 - 3+	106	13	104	100		
2020-064	094	CLINITEK		053 055	053 055		056 - 4+	106	2	104	0		
2020-064	102	ACON LAB		053 055	053 055		054 - 2+	177	134	173	100		
2020-064	102	ACON LAB		053 055	053 055		051 - NEGATIVE	177	2	173	0		
2020-064	102	ACON LAB		053 055	053 055		053 - 1+	177	3	173	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-064	102	ACON LAB		053 055	053 055		055 - 3+	177	36	173	100	
2020-064	102	ACON LAB		053 055	053 055		056 - 4+	177	2	173	0	
2020-064	117	URS 10 A		053 055	053 055		054 - 2+	10	6	10	100	
2020-064	117	URS 10 A		053 055	053 055		055 - 3+	10	4	10	100	
2020-064	131	iCHEM VELO		053 055	053 055		054 - 2+	28	25	28	100	
2020-064	131	iCHEM VELO		053 055	053 055		055 - 3+	28	3	28	100	
2020-064	135	Hybrido AU		053 055	053 055		054 - 2+	12	11	12	100	
2020-064	135	Hybrido AU		053 055	053 055		055 - 3+	12	1	12	100	
2020-064	138	URS-10T		053 055	053 055		054 - 2+	10	5	9	100	
2020-064	138	URS-10T		053 055	053 055		053 - 1+	10	1	9	100	
2020-064	138	URS-10T		053 055	053 055		055 - 3+	10	3	9	100	
2020-064	138	URS-10T		053 055	053 055		056 - 4+	10	1	9	0	
2020-064	140	MISSION A		053 055	053 055		054 - 2+	63	53	63	100	
2020-064	140	MISSION A		053 055	053 055		053 - 1+	63	1	63	100	
2020-064	140	MISSION A		053 055	053 055		055 - 3+	63	9	63	100	
2020-064	147	Mission120		053 055	053 055		054 - 2+	11	10	11	100	
2020-064	147	Mission120		053 055	053 055		055 - 3+	11	1	11	100	
2020-064	148	Mission500		053 055	053 055		054 - 2+	11	6	11	100	
2020-064	148	Mission500		053 055	053 055		055 - 3+	11	5	11	100	
064 URIN-PROTEIN QUALI.												
2020-065	011	AMES NM10S		051 051	051 051		054 - 2+	51	1	50	0	
2020-065	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	51	50	50	100	
2020-065	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	36	36	36	100	
2020-065	067	URISCAN		051 051	051 051		051 - NEGATIVE	49	49	49	100	
2020-065	071	ATLAS		051 051	051 051		051 - NEGATIVE	21	21	21	100	
2020-065	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	108	108	108	100	
2020-065	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	35	34	34	100	
2020-065	080	ROCHE (CHE		051 051	051 051		052 - TRACES	35	1	34	0	
2020-065	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	112	112	112	100	
2020-065	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	106	106	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-065	102	ACON LAB		051 051	051 051		051 - NEGATIVE	177	175	175	100		
2020-065	102	ACON LAB		051 051	051 051		052 - TRACES	177	1	175	0		
2020-065	102	ACON LAB		051 051	051 051		055 - 3+	177	1	175	0		
2020-065	117	URS 10 A		051 051	051 051		051 - NEGATIVE	10	10	10	100		
2020-065	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	28	28	28	100		
2020-065	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100		
2020-065	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100		
2020-065	140	MISSION A		051 051	051 051		051 - NEGATIVE	63	63	63	100		
2020-065	147	Mission120		051 051	051 051		051 - NEGATIVE	11	11	11	100		
2020-065	148	Mission500		051 051	051 051		051 - NEGATIVE	11	11	11	100		
066 URINAL.-UROBILINOGEN													
2020-061	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100		
2020-061	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100		
2020-061	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	49	49	49	100		
2020-061	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	21	21	21	100		
2020-061	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	108	108	108	100		
2020-061	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	35	35	35	100		
2020-061	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	112	112	100		
2020-061	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	107	107	100		
2020-061	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	175	174	174	100		
2020-061	102	ACON LAB		061 061	061 061		062 - 2	175	1	174	0		
2020-061	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	28	28	28	100		
2020-061	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100		
2020-061	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100		
2020-061	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100		
2020-061	147	Mission120		061 061	061 061		061 - NORMAL,NEG	11	11	11	100		
2020-061	148	Mission500		061 061	061 061		061 - NORMAL,NEG	11	11	11	100		
2020-062	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100		

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score			
2020-062	013	BAYER CLI		061	061	061	061	061	-	NORMAL,NEG	36	36	36	100
2020-062	067	URISCAN		061	061	061	061	061	-	NORMAL,NEG	49	49	49	100
2020-062	071	ATLAS		061	061	061	061	061	-	NORMAL,NEG	21	21	21	100
2020-062	072	URS-9-TECO		061	061	061	061	061	-	NORMAL,NEG	108	108	108	100
2020-062	080	ROCHE (CHE		061	061	061	061	061	-	NORMAL,NEG	35	35	35	100
2020-062	089	AIMSTRIP		061	061	061	061	061	-	NORMAL,NEG	112	112	112	100
2020-062	094	CLINITEK		061	061	061	061	061	-	NORMAL,NEG	107	107	107	100
2020-062	102	ACON LAB		061	061	061	061	061	-	NORMAL,NEG	175	174	174	100
2020-062	102	ACON LAB		061	061	061	061	062	-	2	175	1	174	0
2020-062	131	ICHEM VELO		061	061	061	061	061	-	NORMAL,NEG	28	28	28	100
2020-062	135	Hybrido AU		061	061	061	061	061	-	NORMAL,NEG	12	12	12	100
2020-062	138	URS-10T		061	061	061	061	061	-	NORMAL,NEG	10	10	10	100
2020-062	140	MISSION A		061	061	061	061	061	-	NORMAL,NEG	63	63	63	100
2020-062	147	Mission120		061	061	061	061	061	-	NORMAL,NEG	11	11	11	100
2020-062	148	Mission500		061	061	061	061	061	-	NORMAL,NEG	11	11	11	100
066 URINAL.-UROBILINOGEN														
2020-063	011	AMES NM10S		061	061	061	061	061	-	NORMAL,NEG	51	51	51	100
2020-063	013	BAYER CLI		061	061	061	061	061	-	NORMAL,NEG	36	36	36	100
2020-063	067	URISCAN		061	061	061	061	061	-	NORMAL,NEG	49	49	49	100
2020-063	071	ATLAS		061	061	061	061	061	-	NORMAL,NEG	21	21	21	100
2020-063	072	URS-9-TECO		061	061	061	061	061	-	NORMAL,NEG	108	108	108	100
2020-063	080	ROCHE (CHE		061	061	061	061	061	-	NORMAL,NEG	35	35	35	100
2020-063	089	AIMSTRIP		061	061	061	061	061	-	NORMAL,NEG	112	112	112	100
2020-063	094	CLINITEK		061	061	061	061	061	-	NORMAL,NEG	107	107	107	100
2020-063	102	ACON LAB		061	061	061	061	061	-	NORMAL,NEG	175	174	174	100
2020-063	102	ACON LAB		061	061	061	061	062	-	2	175	1	174	0
2020-063	131	ICHEM VELO		061	061	061	061	061	-	NORMAL,NEG	28	28	28	100
2020-063	135	Hybrido AU		061	061	061	061	061	-	NORMAL,NEG	12	12	12	100
2020-063	138	URS-10T		061	061	061	061	061	-	NORMAL,NEG	10	10	10	100
2020-063	140	MISSION A		061	061	061	061	061	-	NORMAL,NEG	63	63	63	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-063	147	Mission120		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-063	148	Mission500		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
066 URINAL.-UROBILINOGEN												
2020-064	011	AMES NM10S		062 064	061 063		061 - NORMAL,NEG	51	8	51	100	
2020-064	011	AMES NM10S		062 064	061 063		062 - 2	51	21	51	100	
2020-064	011	AMES NM10S		062 064	061 063		063 - 4	51	22	51	100	
2020-064	013	BAYER CLI		061 063	061 063		063 - 4	36	8	36	100	
2020-064	013	BAYER CLI		061 063	061 063		061 - NORMAL,NEG	36	12	36	100	
2020-064	013	BAYER CLI		061 063	061 063		062 - 2	36	16	36	100	
2020-064	067	URISCAN			061 063		061 - NORMAL,NEG	49	11	48	100	
2020-064	067	URISCAN			061 063		062 - 2	49	13	48	100	
2020-064	067	URISCAN			061 063		063 - 4	49	24	48	100	
2020-064	067	URISCAN			061 063		064 - 8	49	1	48	0	
2020-064	071	ATLAS		061 063	061 063		062 - 2	21	15	21	100	
2020-064	071	ATLAS		061 063	061 063		063 - 4	21	6	21	100	
2020-064	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	108	17	102	100	
2020-064	072	URS-9-TECO		061 063	061 063		062 - 2	108	62	102	100	
2020-064	072	URS-9-TECO		061 063	061 063		063 - 4	108	23	102	100	
2020-064	072	URS-9-TECO		061 063	061 063		064 - 8	108	6	102	0	
2020-064	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	35	29	35	100	
2020-064	080	ROCHE (CHE		061 061	061 063		062 - 2	35	6	35	100	
2020-064	089	AIMSTRIP			061 063		061 - NORMAL,NEG	112	62	112	100	
2020-064	089	AIMSTRIP			061 063		062 - 2	112	38	112	100	
2020-064	089	AIMSTRIP			061 063		063 - 4	112	12	112	100	
2020-064	094	CLINITEK		061 063	061 063		061 - NORMAL,NEG	107	16	107	100	
2020-064	094	CLINITEK		061 063	061 063		062 - 2	107	52	107	100	
2020-064	094	CLINITEK		061 063	061 063		063 - 4	107	39	107	100	
2020-064	102	ACON LAB			061 063		062 - 2	175	34	175	100	
2020-064	102	ACON LAB			061 063		063 - 4	175	2	175	100	
2020-064	102	ACON LAB			061 063		061 - NORMAL,NEG	175	139	175	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-064	131	iCHEM VELO		062 064	061 063		063 - 4	28	28	28	100	
2020-064	135	Hybrido AU		061 063	061 063		061 - NORMAL,NEG	12	3	12	100	
2020-064	135	Hybrido AU		061 063	061 063		062 - 2	12	9	12	100	
2020-064	138	URS-10T		061 063	061 063		061 - NORMAL,NEG	10	1	10	100	
2020-064	138	URS-10T		061 063	061 063		062 - 2	10	6	10	100	
2020-064	138	URS-10T		061 063	061 063		063 - 4	10	3	10	100	
2020-064	140	MISSION A		061 061	061 063		061 - NORMAL,NEG	63	51	63	100	
2020-064	140	MISSION A		061 061	061 063		062 - 2	63	11	63	100	
2020-064	140	MISSION A		061 061	061 063		063 - 4	63	1	63	100	
2020-064	147	Mission120			061 063		061 - NORMAL,NEG	11	8	11	100	
2020-064	147	Mission120			061 063		062 - 2	11	3	11	100	
2020-064	148	Mission500		061 061	061 063		061 - NORMAL,NEG	11	9	11	100	
2020-064	148	Mission500		061 061	061 063		062 - 2	11	2	11	100	
066 URINAL.-UROBILINOGEN												
2020-065	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2020-065	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2020-065	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	49	49	49	100	
2020-065	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	21	21	21	100	
2020-065	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	108	108	108	100	
2020-065	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	35	35	35	100	
2020-065	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	111	111	100	
2020-065	089	AIMSTRIP		061 061	061 061		062 - 2	112	1	111	0	
2020-065	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	107	107	100	
2020-065	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	175	175	175	100	
2020-065	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	28	28	28	100	
2020-065	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2020-065	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	9	9	100	
2020-065	138	URS-10T		061 061	061 061		062 - 2	10	1	9	0	
2020-065	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100	
2020-065	147	Mission120		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	148	Mission500		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
184 URINALYSIS LEUKOCYTE												
2020-061				001 001	001 001		001 - Negative	47	47	47	100	
2020-061	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100	
2020-061	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100	
2020-061	067	URISCAN		001 001	001 001		001 - Negative	47	47	47	100	
2020-061	071	ATLAS		001 001	001 001		001 - Negative	20	20	20	100	
2020-061	072	URS-9-TECO		001 001	001 001		001 - Negative	94	94	94	100	
2020-061	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100	
2020-061	089	AIMSTRIP		001 001	001 001		001 - Negative	104	103	103	100	
2020-061	089	AIMSTRIP		001 001	001 001		005 - Large	104	1	103	0	
2020-061	094	CLINITEK		001 001	001 001		001 - Negative	104	104	104	100	
2020-061	102	ACON LAB		001 001	001 001		001 - Negative	145	144	144	100	
2020-061	102	ACON LAB		001 001	001 001		002 - Traces	145	1	144	0	
2020-061	117	URS 10 A		001 001	001 001		001 - Negative	10	10	10	100	
2020-061	131	iCHEM VELO		001 001	001 001		001 - Negative	27	27	27	100	
2020-061	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100	
2020-061	138	URS-10T		001 001	001 001		001 - Negative	10	10	10	100	
2020-061	140	MISSION A		001 001	001 001		001 - Negative	74	74	74	100	
2020-061	147	Mission120		001 001	001 001		001 - Negative	12	12	12	100	
2020-062				001 001	001 001		001 - Negative	47	47	47	100	
2020-062	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100	
2020-062	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100	
2020-062	067	URISCAN		001 001	001 001		001 - Negative	47	47	47	100	
2020-062	071	ATLAS		001 001	001 001		001 - Negative	20	20	20	100	
2020-062	072	URS-9-TECO		001 001	001 001		001 - Negative	94	94	94	100	
2020-062	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100	
2020-062	089	AIMSTRIP		001 001	001 001		001 - Negative	104	104	104	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-062	094 CLINITEK			001 001	001 001		001 - Negative	104	104	104	100	
2020-062	102 ACON LAB			001 001	001 001		001 - Negative	145	145	145	100	
2020-062	117 URS 10 A			001 001	001 001		001 - Negative	10	10	10	100	
2020-062	131 iCHEM VELO			001 001	001 001		001 - Negative	27	27	27	100	
2020-062	135 Hybrido AU			001 001	001 001		001 - Negative	12	12	12	100	
2020-062	138 URS-10T			001 001	001 001		001 - Negative	10	10	10	100	
2020-062	140 MISSION A			001 001	001 001		001 - Negative	74	74	74	100	
2020-062	147 Mission120			001 001	001 001		001 - Negative	12	12	12	100	
184 URINALYSIS LEUKOCYTE												
2020-063				001 001	001 001		001 - Negative	47	47	47	100	
2020-063	011 AMES NM10S			001 001	001 001		001 - Negative	44	44	44	100	
2020-063	013 BAYER CLI			001 001	001 001		001 - Negative	36	36	36	100	
2020-063	067 URISCAN			001 001	001 001		001 - Negative	47	47	47	100	
2020-063	071 ATLAS			001 001	001 001		001 - Negative	20	20	20	100	
2020-063	072 URS-9-TECO			001 001	001 001		001 - Negative	94	94	94	100	
2020-063	080 ROCHE (CHE			001 001	001 001		001 - Negative	31	31	31	100	
2020-063	089 AIMSTRIP			001 001	001 001		001 - Negative	104	104	104	100	
2020-063	094 CLINITEK			001 001	001 001		001 - Negative	104	104	104	100	
2020-063	102 ACON LAB			001 001	001 001		001 - Negative	145	145	145	100	
2020-063	117 URS 10 A			001 001	001 001		001 - Negative	10	10	10	100	
2020-063	131 iCHEM VELO			001 001	001 001		001 - Negative	27	27	27	100	
2020-063	135 Hybrido AU			001 001	001 001		001 - Negative	12	12	12	100	
2020-063	138 URS-10T			001 001	001 001		001 - Negative	10	10	10	100	
2020-063	140 MISSION A			001 001	001 001		001 - Negative	74	73	73	100	
2020-063	140 MISSION A			001 001	001 001		003 - Small	74	1	73	0	
2020-063	147 Mission120			001 001	001 001		001 - Negative	12	12	12	100	
2020-064				001 001	001 001		001 - Negative	47	47	47	100	
2020-064	011 AMES NM10S			001 001	001 001		001 - Negative	44	44	44	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-064	013	BAYER CLI		001	001	001	001	001	- Negative	36	36	36	100
2020-064	067	URISCAN		001	001	001	001	001	- Negative	47	47	47	100
2020-064	071	ATLAS		001	001	001	001	001	- Negative	20	20	20	100
2020-064	072	URS-9-TECO		001	001	001	001	001	- Negative	94	94	94	100
2020-064	080	ROCHE (CHE		001	001	001	001	001	- Negative	31	31	31	100
2020-064	089	AIMSTRIP		001	001	001	001	001	- Negative	104	104	104	100
2020-064	094	CLINITEK		001	001	001	001	001	- Negative	104	104	104	100
2020-064	102	ACON LAB		001	001	001	001	001	- Negative	145	145	145	100
2020-064	117	URS 10 A		001	001	001	001	001	- Negative	10	10	10	100
2020-064	131	iCHEM VELO		001	001	001	001	001	- Negative	27	27	27	100
2020-064	135	Hybrido AU		001	001	001	001	001	- Negative	12	12	12	100
2020-064	138	URS-10T		001	001	001	001	001	- Negative	10	10	10	100
2020-064	140	MISSION A		001	001	001	001	001	- Negative	74	73	73	100
2020-064	140	MISSION A		001	001	001	001	001	002 - Traces	74	1	73	0
2020-064	147	Mission120		001	001	001	001	001	- Negative	12	12	12	100
184 URINALYSIS LEUKOCYTE													
2020-065				002	004	003	005		002 - Traces	47	1	47	100
2020-065				002	004	003	005		003 - Small	47	23	47	100
2020-065				002	004	003	005		004 - Moderate	47	16	47	100
2020-065				002	004	003	005		005 - Large	47	7	47	100
2020-065	011	AMES NM10S		002	004	003	005		001 - Negative	44	1	43	0
2020-065	011	AMES NM10S		002	004	003	005		002 - Traces	44	1	43	100
2020-065	011	AMES NM10S		002	004	003	005		003 - Small	44	23	43	100
2020-065	011	AMES NM10S		002	004	003	005		004 - Moderate	44	15	43	100
2020-065	011	AMES NM10S		002	004	003	005		005 - Large	44	4	43	100
2020-065	013	BAYER CLI		002	004	003	005		003 - Small	36	25	36	100
2020-065	013	BAYER CLI		002	004	003	005		004 - Moderate	36	8	36	100
2020-065	013	BAYER CLI		002	004	003	005		005 - Large	36	3	36	100
2020-065	067	URISCAN		004	005	003	005		003 - Small	47	4	47	100
2020-065	067	URISCAN		004	005	003	005		004 - Moderate	47	16	47	100

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-065	067	URISCAN		004 005	003 005		005 - Large	47	27	47	100
2020-065	071	ATLAS		003 005	003 005		004 - Moderate	20	17	20	100
2020-065	071	ATLAS		003 005	003 005		005 - Large	20	3	20	100
2020-065	072	URS-9-TECO		003 005	003 005		002 - Traces	94	2	92	0
2020-065	072	URS-9-TECO		003 005	003 005		003 - Small	94	30	92	100
2020-065	072	URS-9-TECO		003 005	003 005		004 - Moderate	94	58	92	100
2020-065	072	URS-9-TECO		003 005	003 005		005 - Large	94	4	92	100
2020-065	080	ROCHE (CHE		004 005	003 005		004 - Moderate	31	7	31	100
2020-065	080	ROCHE (CHE		004 005	003 005		005 - Large	31	23	31	100
2020-065	080	ROCHE (CHE		004 005	003 005		003 - Small	31	1	31	100
2020-065	089	AIMSTRIP		003 005	003 005		002 - Traces	104	1	103	0
2020-065	089	AIMSTRIP		003 005	003 005		003 - Small	104	46	103	100
2020-065	089	AIMSTRIP		003 005	003 005		004 - Moderate	104	51	103	100
2020-065	089	AIMSTRIP		003 005	003 005		005 - Large	104	6	103	100
2020-065	094	CLINITEK		002 004	003 005		003 - Small	104	71	104	100
2020-065	094	CLINITEK		002 004	003 005		004 - Moderate	104	33	104	100
2020-065	102	ACON LAB		003 005	003 005		002 - Traces	145	3	142	0
2020-065	102	ACON LAB		003 005	003 005		003 - Small	145	58	142	100
2020-065	102	ACON LAB		003 005	003 005		004 - Moderate	145	80	142	100
2020-065	102	ACON LAB		003 005	003 005		005 - Large	145	4	142	100
2020-065	117	URS 10 A		003 005	003 005		003 - Small	10	5	10	100
2020-065	117	URS 10 A		003 005	003 005		004 - Moderate	10	5	10	100
2020-065	131	iCHEM VELO		004 005	003 005		003 - Small	27	1	27	100
2020-065	131	iCHEM VELO		004 005	003 005		004 - Moderate	27	1	27	100
2020-065	131	iCHEM VELO		004 005	003 005		005 - Large	27	25	27	100
2020-065	135	Hybrido AU		004 005	003 005		005 - Large	12	12	12	100
2020-065	138	URS-10T		002 004	003 005		003 - Small	10	7	10	100
2020-065	138	URS-10T		002 004	003 005		004 - Moderate	10	3	10	100
2020-065	140	MISSION A		003 005	003 005		001 - Negative	74	1	70	0
2020-065	140	MISSION A		003 005	003 005		002 - Traces	74	3	70	0

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-065	140	MISSION A		003 005	003 005		003 - Small	74	31	70	100	
2020-065	140	MISSION A		003 005	003 005		004 - Moderate	74	37	70	100	
2020-065	140	MISSION A		003 005	003 005		005 - Large	74	2	70	100	
2020-065	147	Mission120		002 004	003 005		002 - Traces	12	1	12	100	
2020-065	147	Mission120		002 004	003 005		003 - Small	12	8	12	100	
2020-065	147	Mission120		002 004	003 005		004 - Moderate	12	3	12	100	

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2020-061	091	AIM STEP		002 002	002 002		002 - POSITIVE	36	36	36	100	
2020-061	101	HCG Alere		002 002	002 002		002 - POSITIVE	23	23	23	100	
2020-062	091	AIM STEP		001 001	001 001		001 - NEGATIVE	36	36	36	100	
2020-062	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2020-063	091	AIM STEP		001 001	001 001		001 - NEGATIVE	36	36	36	100	
2020-063	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2020-064	091	AIM STEP		002 002	002 002		002 - POSITIVE	36	36	36	100	
2020-064	101	HCG Alere		002 002	002 002		002 - POSITIVE	23	23	23	100	
2020-065	091	AIM STEP		001 001	001 001		001 - NEGATIVE	36	36	36	100	
2020-065	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	

Participant Summary

MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2020-071					002	002	002 - FATTY CAST	866	662	662	100
2020-071					002	002	003 - GRANULAR C	866	4	662	0
2020-071					002	002	005 - RED CELL C	866	190	662	0
2020-071					002	002	006 - WAXY CAST	866	3	662	0
2020-071					002	002	007 - WHITE CELL	866	7	662	0
2020-072					024	024	013 - AMORPHOUS	866	1	865	0
2020-072					024	024	024 - AMMONIUM B	866	865	865	100
2020-073					029	029	008 - ERYTHROCYT	866	1	862	0
2020-073					029	029	029 - BACTERIA	866	862	862	100
2020-073					029	029	032 - MUCOUS TH.	866	2	862	0
2020-073					029	029	038 - YEAST/FUN.	866	1	862	0
2020-074					040	040	039 - E. vermic.	866	1	865	0
2020-074					040	040	040 - S. haemat.	866	865	865	100
2020-075					011	011	001 - E.C. CAST	867	7	858	0
2020-075					011	011	011 - SQUAMOUS E	867	858	858	100
2020-075					011	011	012 - TRANSITIO.	867	2	858	0

Participant Summary
MONTH 2 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 98 OCCULT BLOOD							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
154 FECAL OCCULT BLOOD													
2020-081	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	470	470	470	100	
2020-081	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	676	1	675	0	
2020-081	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	676	675	675	100	
2020-082	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	470	470	470	100	
2020-082	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	676	675	675	100	
2020-082	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	676	1	675	0	
2020-083	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	470	2	468	0	
2020-083	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	470	468	468	100	
2020-083	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	676	2	674	0	
2020-083	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	676	674	674	100	
2020-084	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	470	470	470	100	
2020-084	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	676	2	674	0	
2020-084	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	676	674	674	100	
2020-085	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	470	468	468	100	
2020-085	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	470	2	468	0	
2020-085	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	676	671	671	100	
2020-085	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	676	5	671	0	
SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 90 SPECIAL TEST							

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-031	29 VITROS			11	14.064	0.355	2.52	13.000	15.100
2020-031	38 ROCHE E			22	13.705	0.750	5.47	11.500	16.000
2020-031	44 BECK.COUL.			18	15.661	0.945	6.04	12.800	18.500
2020-031	50 ARC			11	15.855	0.677	4.27	13.800	17.900
2020-032	29 VITROS			11	0.300	0.000	0.00	0.300	0.300
2020-032	38 ROCHE E			22	0.545	0.134	24.55	0.400	0.700
2020-032	44 BECK.COUL.			18	0.500	0.034	6.86	0.400	0.600
2020-032	50 ARC			11	0.964	0.360	37.36	0.700	1.200
2020-033	29 VITROS			11	5.573	0.129	2.31	5.200	6.000
2020-033	38 ROCHE E			22	5.795	0.336	5.80	4.800	6.800
2020-033	44 BECK.COUL.			18	6.094	0.352	5.77	5.000	7.200
2020-033	50 ARC			11	6.564	0.210	3.20	5.900	7.200
2020-034	29 VITROS			11	21.191	0.419	1.98	19.900	22.400
2020-034	38 ROCHE E			22	20.173	1.210	6.00	16.500	23.800
2020-034	44 BECK.COUL.			18	23.728	1.475	6.22	19.300	28.200
2020-034	50 ARC			11	23.782	0.949	3.99	20.900	26.600
2020-035	29 VITROS			11	24.609	0.446	1.81	23.300	25.900
2020-035	38 ROCHE E			22	23.050	1.422	6.17	18.800	27.300
2020-035	44 BECK.COUL.			18	27.135	1.284	4.73	23.300	31.000
2020-035	50 ARC			11	27.509	1.060	3.85	24.300	30.700
069	PROS. SPEC-ANTIG-PSA								
2020-031	29 VITROS			35	4.791	0.206	4.30	4.200	5.400
2020-031	36 TOSOH AIA			12	7.117	0.570	8.01	5.400	8.800
2020-031	38 ROCHE E			48	9.724	0.582	5.98	8.000	11.500
2020-031	39 DIMENS.RXL			15	8.840	0.367	4.15	7.700	9.900

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-031	44	BECK.COUL.		54	9.477	0.484	5.11	8.000	10.900
2020-031	47	D.D.XPAND		14	8.686	0.729	8.39	6.500	10.900
2020-031	50	ARC		17	7.771	0.392	5.05	6.600	8.900
2020-031	57	S. EXL 200		45	8.749	0.422	4.82	7.500	10.000
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2020-032	29	VITROS		35	0.103	0.017	16.41	0.100	0.100
2020-032	36	TOSOH AIA		12	0.208	0.064	30.72	0.200	0.300
2020-032	38	ROCHE E		48	0.200	0.000	0.00	0.200	0.200
2020-032	39	DIMENS.RXL		15	0.136	0.048	35.31	0.100	0.200
2020-032	44	BECK.COUL.		54	0.200	0.000	0.00	0.200	0.200
2020-032	47	D.D.XPAND		14	0.179	0.041	22.98	0.100	0.200
2020-032	50	ARC		17	0.124	0.042	34.34	0.100	0.200
2020-032	57	S. EXL 200		45	0.164	0.053	32.16	0.100	0.200
2020-033	29	VITROS		35	2.037	0.080	3.91	1.800	2.300
2020-033	36	TOSOH AIA		12	2.767	0.205	7.43	2.200	3.400
2020-033	38	ROCHE E		48	3.945	0.283	7.17	3.100	4.800
2020-033	39	DIMENS.RXL		15	3.440	0.136	3.94	3.000	3.800
2020-033	44	BECK.COUL.		54	3.766	0.188	5.00	3.200	4.300
2020-033	47	D.D.XPAND		14	3.386	0.270	7.96	2.600	4.200
2020-033	50	ARC		17	3.118	0.349	11.18	2.300	3.900
2020-033	57	S. EXL 200		45	3.391	0.170	5.01	2.900	3.900
2020-034	29	VITROS		35	6.971	0.355	5.09	5.900	8.000
2020-034	36	TOSOH AIA		12	10.117	0.736	7.27	7.900	12.300
2020-034	38	ROCHE E		48	14.209	0.891	6.27	11.500	16.900
2020-034	39	DIMENS.RXL		15	13.107	0.558	4.26	11.400	14.800
2020-034	44	BECK.COUL.		54	14.085	0.852	6.05	11.500	16.600
2020-034	47	D.D.XPAND		14	12.586	1.071	8.51	9.400	15.700
2020-034	50	ARC		17	11.441	0.790	6.90	9.100	13.800

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 90 SPECIAL TEST				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-034	57 S. EXL 200			45	12.984	0.674	5.19	11.000	15.000	
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}							
2020-035	29 VITROS			35	8.009	0.374	4.67	6.900	9.100	
2020-035	36 TOSOH AIA			12	11.692	0.981	8.39	8.800	14.600	
2020-035	38 ROCHE E			48	16.113	1.068	6.63	12.900	19.300	
2020-035	39 DIMENS.RXL			15	15.007	0.530	3.53	13.400	16.600	
2020-035	44 BECK.COUL.			54	16.209	1.158	7.14	12.700	19.700	
2020-035	47 D.D.XPAND			14	14.829	1.017	6.86	11.800	17.900	
2020-035	50 ARC			17	13.476	0.886	6.57	10.800	16.100	
2020-035	57 S. EXL 200			45	14.918	0.696	4.66	12.800	17.000	
071	VITAMIN B 12									
2020-031	29 VITROS			21	485.714	39.026	8.03	369.000	603.000	
2020-031	38 ROCHE E			24	525.542	33.732	6.42	424.000	627.000	
2020-031	44 BECK.COUL.			28	387.607	31.167	8.04	294.000	481.000	
2020-031	50 ARC			16	489.188	41.322	8.45	367.000	611.000	
2020-032	29 VITROS			21	240.381	9.959	4.14	211.000	270.000	
2020-032	38 ROCHE E			24	215.000	36.590	17.02	161.000	269.000	
2020-032	44 BECK.COUL.			28	169.464	14.495	8.55	127.000	212.000	
2020-032	50 ARC			16	226.400	16.066	7.10	178.000	275.000	
2020-033	29 VITROS			21	332.810	14.889	4.47	288.000	377.000	
2020-033	38 ROCHE E			24	330.792	37.611	11.37	248.000	413.000	
2020-033	44 BECK.COUL.			28	261.107	17.763	6.80	208.000	314.000	
2020-033	50 ARC			16	331.750	55.537	16.74	249.000	415.000	
2020-034	29 VITROS			21	603.524	74.005	12.26	453.000	754.000	
2020-034	38 ROCHE E			24	701.250	52.876	7.54	543.000	860.000	
2020-034	44 BECK.COUL.			28	505.643	35.811	7.08	398.000	613.000	
2020-034	50 ARC			16	631.625	60.284	9.54	474.000	790.000	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
071 VITAMIN B 12		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}							
2020-035	29 VITROS			21	642.952	80.199	12.47	482.000	804.000	
2020-035	38 ROCHE E			24	781.958	68.277	8.73	586.000	977.000	
2020-035	44 BECK.COUL.			28	558.500	42.147	7.55	432.000	685.000	
2020-035	50 ARC			16	708.000	68.661	9.70	531.000	885.000	
SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
146 GLYCOHEMOGLOBIN										
2020-041	2 DCA-2000			12	7.392	0.330	4.47	6.400	8.400	
2020-041	12 COBAS/INT.			29	7.039	0.170	2.41	6.500	7.500	
2020-041	22 VITROS			37	7.081	0.143	2.02	6.700	7.500	
2020-041	23 IMOLA/RX			12	7.817	0.618	7.90	6.000	9.700	
2020-041	29 DADE DIME			18	7.472	0.325	4.34	6.500	8.400	
2020-041	30 DAYTONA			16	7.500	0.570	7.60	5.800	9.200	
2020-041	31 IMMUNOTURB			31	7.077	0.341	4.81	6.100	8.100	
2020-041	40 SIEMENS DI			68	7.482	0.276	3.69	6.700	8.300	
2020-041	46 ARCHITECT			10	6.880	0.363	5.27	5.800	8.000	
2020-042	2 DCA-2000			12	6.075	0.252	4.15	5.300	6.800	
2020-042	12 COBAS/INT.			29	5.832	0.196	3.37	5.200	6.400	
2020-042	22 VITROS			37	5.776	0.126	2.18	5.400	6.200	
2020-042	23 IMOLA/RX			12	6.342	0.401	6.32	5.100	7.500	
2020-042	29 DADE DIME			18	6.117	0.246	4.01	5.400	6.900	
2020-042	30 DAYTONA			16	5.981	0.298	4.99	5.100	6.900	
2020-042	31 IMMUNOTURB			31	5.720	0.268	4.68	4.900	6.500	
2020-042	40 SIEMENS DI			68	6.070	0.267	4.40	5.300	6.900	
2020-042	46 ARCHITECT			10	5.610	0.295	5.25	4.700	6.500	

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-043	2	DCA-2000		12	11.200	0.620	5.54	9.300	13.100
2020-043	12	COBAS/INT.		29	11.218	0.220	1.97	10.600	11.900
2020-043	22	VITROS		37	10.606	0.230	2.17	9.900	11.300
2020-043	23	IMOLA/RX		12	12.442	0.551	4.43	10.800	14.100
2020-043	29	DADE DIME		18	11.344	0.310	2.73	10.400	12.300
2020-043	30	DAYTONA		16	12.119	0.762	6.29	9.800	14.400
2020-043	31	IMMUNOTURB		31	11.203	0.486	4.34	9.700	12.700
2020-043	40	SIEMENS DI		68	11.332	0.376	3.32	10.200	12.500
2020-043	46	ARCHITECT		10	10.900	0.316	2.90	10.000	11.800
2020-044	2	DCA-2000		12	7.358	0.312	4.24	6.400	8.300
2020-044	12	COBAS/INT.		29	7.093	0.156	2.19	6.600	7.600
2020-044	22	VITROS		37	7.076	0.130	1.84	6.700	7.500
2020-044	23	IMOLA/RX		12	7.692	0.463	6.02	6.300	9.100
2020-044	29	DADE DIME		18	7.544	0.243	3.22	6.800	8.300
2020-044	30	DAYTONA		16	7.625	0.511	6.70	6.100	9.200
2020-044	31	IMMUNOTURB		31	7.032	0.334	4.76	6.000	8.000
2020-044	40	SIEMENS DI		68	7.501	0.239	3.18	6.800	8.200
2020-044	46	ARCHITECT		10	6.860	0.338	4.93	5.800	7.900
2020-045	2	DCA-2000		12	6.050	0.229	3.79	5.400	6.700
2020-045	12	COBAS/INT.		29	5.864	0.149	2.55	5.400	6.300
2020-045	22	VITROS		37	5.784	0.140	2.43	5.400	6.200
2020-045	23	IMOLA/RX		12	6.283	0.513	8.16	4.700	7.800
2020-045	29	DADE DIME		18	6.178	0.199	3.22	5.600	6.800
2020-045	30	DAYTONA		16	6.027	0.326	5.40	5.100	7.000
2020-045	31	IMMUNOTURB		31	5.757	0.255	4.43	5.000	6.500
2020-045	40	SIEMENS DI		68	6.104	0.234	3.84	5.400	6.800
2020-045	46	ARCHITECT		10	5.600	0.303	5.42	4.700	6.500

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-261	3	VITROS		32	42.281	3.475	8.22	34.000	51.000		
2020-261	5	ROCHE E		29	36.448	7.412	20.34	29.000	44.000		
2020-261	9	BECK.COUL.		44	47.818	4.190	8.76	38.000	57.000		
2020-261	11	ARC		12	45.750	1.299	2.84	37.000	55.000		
2020-261	25	S. DIME		15	33.571	2.352	7.01	27.000	40.000		
2020-262	3	VITROS		32	23.903	2.085	8.72	19.000	29.000		
2020-262	5	ROCHE E		29	23.172	3.228	13.93	19.000	28.000		
2020-262	9	BECK.COUL.		44	29.455	2.808	9.53	24.000	35.000		
2020-262	11	ARC		12	27.917	0.954	3.42	22.000	34.000		
2020-262	25	S. DIME		15	22.000	1.558	7.08	18.000	26.000		
2020-263	3	VITROS		32	80.625	5.290	6.56	65.000	97.000		
2020-263	5	ROCHE E		29	54.793	6.789	12.39	44.000	66.000		
2020-263	9	BECK.COUL.		44	79.698	6.091	7.64	64.000	96.000		
2020-263	11	ARC		12	75.167	2.115	2.81	60.000	90.000		
2020-263	25	S. DIME		15	54.643	3.286	6.01	44.000	66.000		
2020-264	3	VITROS		32	42.844	4.032	9.41	34.000	51.000		
2020-264	5	ROCHE E		29	34.741	4.308	12.40	28.000	42.000		
2020-264	9	BECK.COUL.		44	47.614	4.119	8.65	38.000	57.000		
2020-264	11	ARC		12	46.083	1.801	3.91	37.000	55.000		
2020-264	25	S. DIME		15	34.000	2.390	7.03	27.000	41.000		
2020-265	3	VITROS		32	32.188	2.973	9.24	26.000	39.000		
2020-265	5	ROCHE E		29	26.483	4.561	17.22	21.000	32.000		
2020-265	9	BECK.COUL.		44	33.841	3.233	9.55	27.000	41.000		
2020-265	11	ARC		12	32.545	0.891	2.74	26.000	39.000		

4/9/2020

PROFICIENCY TESTING SERVICE
LABORATORY SERVICES PROGRAM
DEPARTMENT OF HEALTH OF PUERTO RICO

PAGE : 131 (131)
DATE : 4/9/2020
TIME : 09:15:16

Participant Summary

MONTH 2 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-265	25 S. DIME				15	25.643	1.836	7.16	21.000	31.000