

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 15 WHOLE BLOOD GLUCOSE

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
157 WHOLE BLOOD GLUCOSE				CLIA TEST ID #: 0						EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}	
2018-491	2	ACCU-CHEK		4	101.500	2.693	2.65	91.000	112.000		
2018-491	8	PRECISION		3	70.667	7.409	10.48	64.000	78.000		
2018-491	14	PREC.PLUS		10	66.900	5.594	8.36	60.000	74.000		
2018-491	16	ASCENSIA		6	77.500	7.182	9.27	70.000	85.000		
2018-491	23	ABBOTT PRE		57	67.127	3.927	5.85	60.000	74.000		
2018-491	27	STAT SRIP.		31	84.452	5.847	6.92	76.000	93.000		
2018-491	33	AC.CHEK II		22	99.524	1.789	1.80	90.000	109.000		
2018-491	34	FREE STYL		66	83.338	5.531	6.64	75.000	92.000		
2018-491	34	FREE STYL		66	101.500	2.693	2.65	91.000	112.000		
2018-492	2	ACCU-CHEK		4	232.000	28.390	12.24	209.000	255.000		
2018-492	8	PRECISION		3	161.667	20.039	12.40	146.000	178.000		
2018-492	14	PREC.PLUS		10	160.500	8.176	5.09	144.000	177.000		
2018-492	16	ASCENSIA		6	202.500	9.828	4.85	182.000	223.000		
2018-492	23	ABBOTT PRE		57	155.836	8.500	5.45	140.000	171.000		
2018-492	27	STAT SRIP.		31	187.000	11.121	5.95	168.000	206.000		
2018-492	33	AC.CHEK II		22	211.190	3.445	1.63	190.000	232.000		
2018-492	34	FREE STYL		66	232.000	28.390	12.24	209.000	255.000		
2018-492	34	FREE STYL		66	190.046	9.632	5.07	171.000	209.000		
2018-493	2	ACCU-CHEK		4	371.250	2.947	0.79	334.000	408.000		
2018-493	8	PRECISION		3	420.000	23.930	5.70	378.000	462.000		
2018-493	14	PREC.PLUS		10	425.000	19.616	4.62	383.000	468.000		
2018-493	16	ASCENSIA		6	337.167	17.525	5.20	303.000	371.000		
2018-493	23	ABBOTT PRE		57	409.218	16.896	4.13	368.000	450.000		
2018-493	27	STAT SRIP.		31	320.129	15.256	4.77	288.000	352.000		
2018-493	33	AC.CHEK II		22	378.045	9.222	2.44	340.000	416.000		
2018-493	34	FREE STYL		66	470.212	21.562	4.59	423.000	517.000		
2018-493	34	FREE STYL		66	371.250	2.947	0.79	334.000	408.000		

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 10 CHEMISTRY			SUB: 10 ROUTINE			GRP: 15 WHOLE BLOOD GLUCOSE					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
157	WHOLE BLOOD GLUCOSE	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}								
2018-494	2	ACCU-CHEK		4	285.750	4.437	1.55	257.000	314.000		
2018-494	8	PRECISION		3	217.333	23.228	10.69	196.000	239.000		
2018-494	14	PREC.PLUS		10	218.400	11.834	5.42	197.000	240.000		
2018-494	16	ASCENSIA		6	285.667	12.750	4.46	257.000	314.000		
2018-494	23	ABBOTT PRE		57	214.200	10.046	4.69	193.000	236.000		
2018-494	27	STAT SRIP.		31	251.290	11.504	4.58	226.000	276.000		
2018-494	33	AC.CHEK II		22	283.727	6.593	2.32	255.000	312.000		
2018-494	34	FREE STYL		66	285.750	4.437	1.55	257.000	314.000		
2018-494	34	FREE STYL		66	257.585	14.061	5.46	232.000	283.000		
2018-495	2	ACCU-CHEK		4	385.250	4.918	1.28	347.000	424.000		
2018-495	8	PRECISION		3	305.333	27.329	8.95	275.000	336.000		
2018-495	14	PREC.PLUS		10	300.300	17.344	5.78	270.000	330.000		
2018-495	16	ASCENSIA		6	369.833	39.138	10.58	333.000	407.000		
2018-495	23	ABBOTT PRE		57	296.196	12.679	4.28	267.000	326.000		
2018-495	27	STAT SRIP.		31	337.903	17.098	5.06	304.000	372.000		
2018-495	33	AC.CHEK II		22	383.500	9.114	2.38	345.000	422.000		
2018-495	34	FREE STYL		66	385.250	4.918	1.28	347.000	424.000		
2018-495	34	FREE STYL		66	351.500	17.231	4.90	316.000	387.000		

SPEC: 20 HEMATOLOGY			SUB: 70 GENERAL			GRP: 140 COAGULATION (HEMAT.)					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
080	FIBRINOGEN	CLIA TEST ID #: 825	EVALUATION CRITERIA: TARGET VALUE 20 %								
2018-551	60	DADE SYSME	67 DADE	16	132.438	10.799	8.15	106.000	159.000		
2018-551	69	DIAG.STAGO	85 STA FIBRI	3	158.333	5.907	3.73	127.000	190.000		

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2018-551	73	ACL 8 TO E	87	HEMOSIL-C	6	158.500	9.106	5.75	127.000	190.000	
2018-551	73	ACL 8 TO E	88	HEMO. PT-F	6	165.167	20.860	12.63	132.000	198.000	
2018-551	73	ACL 8 TO E	89	HEMOSIL 2G	7	136.857	7.259	5.30	109.000	164.000	
080	FIBRINOGEN		CLIA TEST ID #: 825		EVALUATION CRITERIA: TARGET VALUE 20 %						
2018-552	60	DADE SYSME	67	DADE	16	249.500	17.117	6.86	200.000	299.000	
2018-552	69	DIAG.STAGO	85	STA FIBRI	3	284.333	4.989	1.75	227.000	341.000	
2018-552	73	ACL 8 TO E	87	HEMOSIL-C	6	322.333	15.553	4.83	258.000	387.000	
2018-552	73	ACL 8 TO E	88	HEMO. PT-F	6	434.333	62.864	14.47	347.000	521.000	
2018-552	73	ACL 8 TO E	89	HEMOSIL 2G	7	533.000	42.352	7.95	426.000	640.000	
2018-553	60	DADE SYSME	67	DADE	16	261.750	19.514	7.46	209.000	314.000	
2018-553	69	DIAG.STAGO	85	STA FIBRI	3	295.667	10.403	3.52	237.000	355.000	
2018-553	73	ACL 8 TO E	87	HEMOSIL-C	6	337.667	20.475	6.06	270.000	405.000	
2018-553	73	ACL 8 TO E	88	HEMO. PT-F	6	336.500	39.995	11.89	269.000	404.000	
2018-553	73	ACL 8 TO E	89	HEMOSIL 2G	7	342.000	16.587	4.85	274.000	410.000	
2018-554	60	DADE SYSME	67	DADE	16	263.875	17.553	6.65	211.000	317.000	
2018-554	69	DIAG.STAGO	85	STA FIBRI	3	289.667	10.656	3.68	232.000	348.000	
2018-554	73	ACL 8 TO E	87	HEMOSIL-C	6	336.667	15.096	4.48	269.000	404.000	
2018-554	73	ACL 8 TO E	88	HEMO. PT-F	6	330.500	42.547	12.87	264.000	397.000	
2018-554	73	ACL 8 TO E	89	HEMOSIL 2G	7	337.000	17.873	5.30	270.000	404.000	
2018-555	60	DADE SYSME	67	DADE	16	264.250	20.547	7.78	211.000	317.000	
2018-555	69	DIAG.STAGO	85	STA FIBRI	3	292.667	15.326	5.24	234.000	351.000	
2018-555	73	ACL 8 TO E	87	HEMOSIL-C	6	356.000	19.261	5.41	285.000	427.000	
2018-555	73	ACL 8 TO E	88	HEMO. PT-F	6	446.000	54.562	12.23	357.000	535.000	
2018-555	73	ACL 8 TO E	89	HEMOSIL 2G	7	519.571	37.576	7.23	416.000	623.000	
081	PART.THROM. TIME-PTT		CLIA TEST ID #: 835		EVALUATION CRITERIA: TARGET VALUE 15 %						
2018-551	5	DADE SYS.	35	DADE ACTIN	3	31.200	0.374	1.20	26.500	35.900	
2018-551	33	MLA ELECT.	54	QUICK COAG	11	30.127	1.011	3.36	25.600	34.600	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2018-551	45	BIOMERIEUX	41 BIOMERIEUX	3	32.900	2.478	7.53	28.000	37.800	
2018-551	45	BIOMERIEUX	54 QUICK COAG	3	33.400	1.838	5.50	28.400	38.400	
2018-551	58	AMELUNG KC	33 DADE ACTIN	5	38.780	1.335	3.44	33.000	44.600	
2018-551	58	AMELUNG KC	42 BIOMERIEUX	10	36.690	8.393	22.88	31.200	42.200	
2018-551	58	AMELUNG KC	59 DADE ACTIN	2	38.780	1.335	3.44	33.000	44.600	
2018-551	60	DADE SYSME	35 DADE ACTIN	18	33.061	1.020	3.08	28.100	38.000	
2018-551	62	DADE SYS.	35 DADE ACTIN	65	31.284	0.785	2.51	26.600	36.000	
2018-551	65	DADE BFTII	33 DADE ACTIN	3	36.567	2.614	7.15	31.100	42.100	
2018-551	65	DADE BFTII	35 DADE ACTIN	14	36.079	1.922	5.33	30.700	41.500	
2018-551	65	DADE BFTII	35 DADE ACTIN	14	36.567	2.614	7.15	31.100	42.100	
2018-551	69	DIAG.STAGO	49 STA-PTT	11	39.973	1.501	3.75	34.000	46.000	
2018-551	72	PACI HEMOS	33 DADE ACTIN	1	38.780	1.335	3.44	33.000	44.600	
2018-551	72	PACI HEMOS	60 PACIFIC	3	42.200	0.735	1.74	35.900	48.500	
2018-551	73	ACL 8 TO E	38 IL TEST A.	3	39.167	0.464	1.19	33.300	45.000	
2018-551	73	ACL 8 TO E	48 SynthAsil	1	38.780	1.335	3.44	33.000	44.600	
2018-551	73	ACL 8 TO E	56 PTT-HEMO	35	40.041	1.432	3.58	34.000	46.000	
2018-551	76	STAGO STAR	49 STA-PTT	19	42.200	0.735	1.74	35.900	48.500	
2018-551	76	STAGO STAR	49 STA-PTT	19	39.574	2.972	7.51	33.600	45.500	
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %							
2018-552	5	DADE SYS.	35 DADE ACTIN	3	79.433	1.960	2.47	67.500	91.300	
2018-552	33	MLA ELECT.	54 QUICK COAG	11	71.027	1.683	2.37	60.400	81.700	
2018-552	45	BIOMERIEUX	41 BIOMERIEUX	3	74.367	7.633	10.26	63.200	85.500	
2018-552	45	BIOMERIEUX	54 QUICK COAG	3	75.300	4.777	6.34	64.000	86.600	
2018-552	58	AMELUNG KC	33 DADE ACTIN	5	116.700	13.888	11.90	99.200	134.200	
2018-552	58	AMELUNG KC	42 BIOMERIEUX	10	77.870	18.777	24.11	66.200	89.600	
2018-552	58	AMELUNG KC	59 DADE ACTIN	2	116.700	13.888	11.90	99.200	134.200	
2018-552	60	DADE SYSME	35 DADE ACTIN	18	81.789	2.727	3.33	69.500	94.100	
2018-552	62	DADE SYS.	35 DADE ACTIN	65	78.723	3.698	4.70	66.900	90.500	
2018-552	65	DADE BFTII	33 DADE ACTIN	3	117.867	9.955	8.45	100.200	135.500	
2018-552	65	DADE BFTII	35 DADE ACTIN	14	117.867	9.955	8.45	100.200	135.500	
2018-552	65	DADE BFTII	35 DADE ACTIN	14	92.336	7.261	7.86	78.500	106.200	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2018-552	69	DIAG.STAGO	49 STA-PTT	11	82.045	4.152	5.06	69.700	94.400	
2018-552	72	PACI HEMOS	33 DADE ACTIN	1	116.700	13.888	11.90	99.200	134.200	
2018-552	72	PACI HEMOS	60 PACIFIC	3	81.900	4.032	4.92	69.600	94.200	
2018-552	73	ACL 8 TO E	38 IL TEST A.	3	67.133	13.610	20.27	57.100	77.200	
2018-552	73	ACL 8 TO E	48 SynthASil	1	116.700	13.888	11.90	99.200	134.200	
2018-552	73	ACL 8 TO E	56 PTT-HEMO	35	102.776	5.089	4.95	87.400	118.200	
2018-552	76	STAGO STAR	49 STA-PTT	19	83.663	6.557	7.84	71.100	96.200	
2018-552	76	STAGO STAR	49 STA-PTT	19	81.900	4.032	4.92	69.600	94.200	
081	PART.THROM. TIME-PTT		CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %						
2018-553	5	DADE SYS.	35 DADE ACTIN	3	24.400	0.216	0.89	20.700	28.100	
2018-553	33	MLA ELECT.	54 QUICK COAG	11	23.455	0.847	3.61	19.900	27.000	
2018-553	45	BIOMERIEUX	41 BIOMERIEUX	3	24.200	1.023	4.23	20.600	27.800	
2018-553	45	BIOMERIEUX	54 QUICK COAG	3	24.967	1.907	7.64	21.200	28.700	
2018-553	58	AMELUNG KC	33 DADE ACTIN	5	27.820	1.736	6.24	23.600	32.000	
2018-553	58	AMELUNG KC	42 BIOMERIEUX	10	26.240	5.299	20.20	22.300	30.200	
2018-553	58	AMELUNG KC	59 DADE ACTIN	2	27.820	1.736	6.24	23.600	32.000	
2018-553	60	DADE SYSME	35 DADE ACTIN	18	26.211	0.769	2.94	22.300	30.100	
2018-553	62	DADE SYS.	35 DADE ACTIN	65	24.480	0.926	3.78	20.800	28.200	
2018-553	65	DADE BFTII	33 DADE ACTIN	3	25.567	2.223	8.70	21.700	29.400	
2018-553	65	DADE BFTII	35 DADE ACTIN	14	24.607	1.173	4.77	20.900	28.300	
2018-553	65	DADE BFTII	35 DADE ACTIN	14	25.567	2.223	8.70	21.700	29.400	
2018-553	69	DIAG.STAGO	49 STA-PTT	11	28.118	0.564	2.01	23.900	32.300	
2018-553	72	PACI HEMOS	33 DADE ACTIN	1	27.820	1.736	6.24	23.600	32.000	
2018-553	72	PACI HEMOS	60 PACIFIC	3	33.767	1.721	5.10	28.700	38.800	
2018-553	73	ACL 8 TO E	38 IL TEST A.	3	25.500	0.816	3.20	21.700	29.300	
2018-553	73	ACL 8 TO E	48 SynthASil	1	27.820	1.736	6.24	23.600	32.000	
2018-553	73	ACL 8 TO E	56 PTT-HEMO	35	27.194	1.005	3.69	23.100	31.300	
2018-553	76	STAGO STAR	49 STA-PTT	19	33.767	1.721	5.10	28.700	38.800	
2018-553	76	STAGO STAR	49 STA-PTT	19	27.595	1.437	5.21	23.500	31.700	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE	15	%					
2018-554	5	DADE SYS.	35	DADE ACTIN	3	24.267	0.125	0.51	20.600	27.900
2018-554	33	MLA ELECT.	54	QUICK COAG	11	23.445	0.669	2.85	19.900	27.000
2018-554	45	BIOMERIEUX	41	BIOMERIEUX	3	24.600	1.134	4.61	20.900	28.300
2018-554	45	BIOMERIEUX	54	QUICK COAG	3	25.967	1.725	6.64	22.100	29.900
2018-554	58	AMELUNG KC	33	DADE ACTIN	5	27.380	1.186	4.33	23.300	31.500
2018-554	58	AMELUNG KC	42	BIOMERIEUX	10	26.350	5.872	22.29	22.400	30.300
2018-554	58	AMELUNG KC	59	DADE ACTIN	2	27.380	1.186	4.33	23.300	31.500
2018-554	60	DADE SYSME	35	DADE ACTIN	18	26.128	0.831	3.18	22.200	30.000
2018-554	62	DADE SYS.	35	DADE ACTIN	65	24.425	0.863	3.53	20.800	28.100
2018-554	65	DADE BFTII	33	DADE ACTIN	3	26.267	2.182	8.31	22.300	30.200
2018-554	65	DADE BFTII	35	DADE ACTIN	14	26.267	2.182	8.31	22.300	30.200
2018-554	65	DADE BFTII	35	DADE ACTIN	14	24.636	1.362	5.53	20.900	28.300
2018-554	69	DIAG.STAGO	49	STA-PTT	11	28.227	0.819	2.90	24.000	32.500
2018-554	72	PACI HEMOS	33	DADE ACTIN	1	27.380	1.186	4.33	23.300	31.500
2018-554	72	PACI HEMOS	60	PACIFIC	3	33.300	2.061	6.19	28.300	38.300
2018-554	73	ACL 8 TO E	38	IL TEST A.	3	25.467	0.464	1.82	21.600	29.300
2018-554	73	ACL 8 TO E	48	SynthASil	1	27.380	1.186	4.33	23.300	31.500
2018-554	73	ACL 8 TO E	56	PTT-HEMO	35	27.437	0.934	3.40	23.300	31.600
2018-554	76	STAGO STAR	49	STA-PTT	19	33.300	2.061	6.19	28.300	38.300
2018-554	76	STAGO STAR	49	STA-PTT	19	27.106	1.044	3.85	23.000	31.200
2018-555	5	DADE SYS.	35	DADE ACTIN	3	54.667	0.330	0.60	46.500	62.900
2018-555	33	MLA ELECT.	54	QUICK COAG	11	51.036	1.538	3.01	43.400	58.700
2018-555	45	BIOMERIEUX	41	BIOMERIEUX	3	54.633	4.955	9.07	46.400	62.800
2018-555	45	BIOMERIEUX	54	QUICK COAG	3	52.900	1.233	2.33	45.000	60.800
2018-555	58	AMELUNG KC	33	DADE ACTIN	5	81.240	6.477	7.97	69.100	93.400
2018-555	58	AMELUNG KC	42	BIOMERIEUX	10	55.270	12.292	22.24	47.000	63.600
2018-555	58	AMELUNG KC	59	DADE ACTIN	2	81.240	6.477	7.97	69.100	93.400
2018-555	60	DADE SYSME	35	DADE ACTIN	18	57.561	1.583	2.75	48.900	66.200

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-555	62	DADE SYS.	35 DADE ACTIN	65	53.411	2.322	4.35	45.400	61.400
2018-555	65	DADE BFTII	33 DADE ACTIN	3	79.800	10.072	12.62	67.800	91.800
2018-555	65	DADE BFTII	35 DADE ACTIN	14	79.800	10.072	12.62	67.800	91.800
2018-555	65	DADE BFTII	35 DADE ACTIN	14	65.343	4.916	7.52	55.500	75.100
2018-555	69	DIAG.STAGO	49 STA-PTT	11	57.309	2.941	5.13	48.700	65.900
2018-555	72	PACI HEMOS	33 DADE ACTIN	1	81.240	6.477	7.97	69.100	93.400
2018-555	72	PACI HEMOS	60 PACIFIC	3	57.567	2.330	4.05	48.900	66.200
2018-555	73	ACL 8 TO E	38 IL TEST A.	3	50.233	8.796	17.51	42.700	57.800
2018-555	73	ACL 8 TO E	48 SynthASil	1	81.240	6.477	7.97	69.100	93.400
2018-555	73	ACL 8 TO E	56 PTT-HEMO	35	69.682	2.585	3.71	59.200	80.100
2018-555	76	STAGO STAR	49 STA-PTT	19	59.328	5.163	8.70	50.400	68.200
2018-555	76	STAGO STAR	49 STA-PTT	19	57.567	2.330	4.05	48.900	66.200
082 PROTHROMBIN TIME (PT)		CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %					
2018-551	5	DADE SYS.	33 DADE INNOV	3	12.567	0.094	0.75	10.700	14.500
2018-551	25	IL ACL 100	13 HEMOSIL PT	1	13.250	0.502	3.79	11.300	15.200
2018-551	33	MLA ELECT.	37 QUICK COAG	11	13.655	0.514	3.77	11.600	15.700
2018-551	45	BIOMERIEUX	8 Thromborel	5	13.160	0.344	2.61	11.200	15.100
2018-551	45	BIOMERIEUX	37 QUICK COAG	3	13.567	0.189	1.39	11.500	15.600
2018-551	57	ACL 2-6000	13 HEMOSIL PT	1	13.250	0.502	3.79	11.300	15.200
2018-551	58	AMELUNG KC	15 BIOMERIEUX	13	11.783	0.534	4.53	10.000	13.600
2018-551	58	AMELUNG KC	33 DADE INNOV	6	11.483	0.511	4.45	9.800	13.200
2018-551	60	DADE SYSME	33 DADE INNOV	19	12.600	0.222	1.77	10.700	14.500
2018-551	62	DADE SYS.	33 DADE INNOV	66	12.224	0.336	2.75	10.400	14.100
2018-551	65	DADE BFTII	8 Thromborel	1	13.160	0.344	2.61	11.200	15.100
2018-551	65	DADE BFTII	9 DADE TH. C	1	12.567	0.094	0.75	10.700	14.500
2018-551	65	DADE BFTII	33 DADE INNOV	16	11.506	0.543	4.72	9.800	13.200
2018-551	69	DIAG.STAGO	36 STA NEOPLA	3	13.900	0.283	2.03	11.800	16.000
2018-551	69	DIAG.STAGO	44 STA NEO CI	7	13.829	0.562	4.07	11.800	15.900
2018-551	72	PACI HEMOS	39 TROMBO.-D	4	13.250	0.502	3.79	11.300	15.200
2018-551	73	ACL 8 TO E	13 HEMOSIL PT	2	13.250	0.502	3.79	11.300	15.200
2018-551	73	ACL 8 TO E	41 HEMOSIL RE	25	12.380	0.524	4.23	10.500	14.200

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-551	73 ACL 8 TO E	43 HEMOSIL PT		12	14.567	0.320	2.19	12.400	16.800
2018-551	76 STAGO STAR	28 STA NEOPL.		3	14.000	0.408	2.92	11.900	16.100
2018-551	76 STAGO STAR	44 STA NEO CI		14	13.350	0.360	2.70	11.300	15.400
082 PROTHROMBIN TIME(PT)		CLIA TEST ID #: 845	EVALUATION CRITERIA: TARGET VALUE 15 %						
2018-552	5 DADE SYS.	33 DADE INNOV		3	42.933	2.106	4.91	36.500	49.400
2018-552	25 IL ACL 100	13 HEMOSIL PT		1	36.900	4.135	11.21	31.400	42.400
2018-552	33 MLA ELECT.	37 QUICK COAG		11	51.573	2.406	4.67	43.800	59.300
2018-552	45 BIOMERIEUX	8 Thromborel		5	57.040	4.007	7.03	48.500	65.600
2018-552	45 BIOMERIEUX	37 QUICK COAG		3	51.333	1.808	3.52	43.600	59.000
2018-552	57 ACL 2-6000	13 HEMOSIL PT		1	36.900	4.135	11.21	31.400	42.400
2018-552	58 AMELUNG KC	15 BIOMERIEUX		13	30.392	2.282	7.51	25.800	35.000
2018-552	58 AMELUNG KC	33 DADE INNOV		6	41.417	1.298	3.13	35.200	47.600
2018-552	60 DADE SYSME	33 DADE INNOV		19	45.353	2.012	4.44	38.500	52.200
2018-552	62 DADE SYS.	33 DADE INNOV		66	42.415	2.700	6.36	36.100	48.800
2018-552	65 DADE BFTII	8 Thromborel		1	57.040	4.007	7.03	48.500	65.600
2018-552	65 DADE BFTII	9 DADE TH. C		1	42.933	2.106	4.91	36.500	49.400
2018-552	65 DADE BFTII	33 DADE INNOV		16	43.675	2.960	6.78	37.100	50.200
2018-552	69 DIAG.STAGO	36 STA NEOPLA		3	58.167	0.850	1.46	49.400	66.900
2018-552	69 DIAG.STAGO	44 STA NEO CI		7	59.914	2.860	4.77	50.900	68.900
2018-552	72 PACI HEMOS	39 TROMBO.-D		4	36.900	4.135	11.21	31.400	42.400
2018-552	73 ACL 8 TO E	13 HEMOSIL PT		2	36.900	4.135	11.21	31.400	42.400
2018-552	73 ACL 8 TO E	41 HEMOSIL RE		25	56.204	3.110	5.53	47.800	64.600
2018-552	73 ACL 8 TO E	43 HEMOSIL PT		12	67.208	3.704	5.51	57.100	77.300
2018-552	76 STAGO STAR	28 STA NEOPL.		3	62.500	7.770	12.43	53.100	71.900
2018-552	76 STAGO STAR	44 STA NEO CI		14	64.371	2.877	4.47	54.700	74.000
2018-553	5 DADE SYS.	33 DADE INNOV		3	10.300	0.216	2.10	8.800	11.800
2018-553	25 IL ACL 100	13 HEMOSIL PT		1	12.000	0.524	4.37	10.200	13.800
2018-553	33 MLA ELECT.	37 QUICK COAG		11	12.618	0.262	2.08	10.700	14.500
2018-553	45 BIOMERIEUX	8 Thromborel		5	11.960	0.492	4.12	10.200	13.800



**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-553	45	BIOMERIEUX	37 QUICK COAG	3	12.067	0.170	1.41	10.300	13.900
2018-553	57	ACL 2-6000	13 HEMOSIL PT	1	12.000	0.524	4.37	10.200	13.800
2018-553	58	AMELUNG KC	15 BIOMERIEUX	13	10.950	0.687	6.28	9.300	12.600
2018-553	58	AMELUNG KC	33 DADE INNOV	6	9.450	0.525	5.56	8.000	10.900
2018-553	60	DADE SYSME	33 DADE INNOV	19	10.863	0.237	2.18	9.200	12.500
2018-553	62	DADE SYS.	33 DADE INNOV	66	10.342	0.314	3.04	8.800	11.900
2018-553	65	DADE BFTII	8 Thromborel	1	11.960	0.492	4.12	10.200	13.800
2018-553	65	DADE BFTII	9 DADE TH. C	1	10.300	0.216	2.10	8.800	11.800
2018-553	65	DADE BFTII	33 DADE INNOV	16	9.240	0.357	3.87	7.900	10.600
2018-553	69	DIAG.STAGO	36 STA NEOPLA	3	13.067	0.189	1.44	11.100	15.000
2018-553	69	DIAG.STAGO	44 STA NEO CI	7	12.800	0.513	4.01	10.900	14.700
2018-553	72	PACI HEMOS	39 TROMBO.-D	4	12.000	0.524	4.37	10.200	13.800
2018-553	73	ACL 8 TO E	13 HEMOSIL PT	2	12.000	0.524	4.37	10.200	13.800
2018-553	73	ACL 8 TO E	41 HEMOSIL RE	25	11.408	0.558	4.89	9.700	13.100
2018-553	73	ACL 8 TO E	43 HEMOSIL PT	12	13.150	0.407	3.10	11.200	15.100
2018-553	76	STAGO STAR	28 STA NEOPL.	3	12.900	0.455	3.52	11.000	14.800
2018-553	76	STAGO STAR	44 STA NEO CI	14	12.186	0.449	3.68	10.400	14.000
082 PROTHROMBIN TIME(PT)			CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %				
2018-554	5	DADE SYS.	33 DADE INNOV	3	10.400	0.082	0.79	8.800	12.000
2018-554	25	IL ACL 100	13 HEMOSIL PT	1	12.250	0.669	5.46	10.400	14.100
2018-554	33	MLA ELECT.	37 QUICK COAG	11	12.655	0.323	2.55	10.800	14.600
2018-554	45	BIOMERIEUX	8 Thromborel	5	12.160	0.383	3.15	10.300	14.000
2018-554	45	BIOMERIEUX	37 QUICK COAG	3	12.200	0.163	1.34	10.400	14.000
2018-554	57	ACL 2-6000	13 HEMOSIL PT	1	12.250	0.669	5.46	10.400	14.100
2018-554	58	AMELUNG KC	15 BIOMERIEUX	13	10.967	0.630	5.75	9.300	12.600
2018-554	58	AMELUNG KC	33 DADE INNOV	6	9.517	0.157	1.65	8.100	10.900
2018-554	60	DADE SYSME	33 DADE INNOV	19	10.868	0.200	1.84	9.200	12.500
2018-554	62	DADE SYS.	33 DADE INNOV	66	10.271	0.281	2.74	8.700	11.800
2018-554	65	DADE BFTII	8 Thromborel	1	12.160	0.383	3.15	10.300	14.000
2018-554	65	DADE BFTII	9 DADE TH. C	1	10.400	0.082	0.79	8.800	12.000
2018-554	65	DADE BFTII	33 DADE INNOV	16	9.569	0.612	6.40	8.100	11.000

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-554	69 DIAG.STAGO	36 STA NEOPLA		3	12.900	0.283	2.19	11.000	14.800		
2018-554	69 DIAG.STAGO	44 STA NEO CI		7	12.543	0.600	4.78	10.700	14.400		
2018-554	72 PACI HEMOS	39 TROMBO.-D		4	12.250	0.669	5.46	10.400	14.100		
2018-554	73 ACL 8 TO E	13 HEMOSIL PT		2	12.250	0.669	5.46	10.400	14.100		
2018-554	73 ACL 8 TO E	41 HEMOSIL RE		25	11.356	0.505	4.45	9.700	13.100		
2018-554	73 ACL 8 TO E	43 HEMOSIL PT		12	13.300	0.354	2.66	11.300	15.300		
2018-554	76 STAGO STAR	28 STA NEOPL.		3	12.667	0.205	1.62	10.800	14.600		
2018-554	76 STAGO STAR	44 STA NEO CI		14	12.229	0.306	2.50	10.400	14.100		
082 PROTHROMBIN TIME(PT)		CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %							
2018-555	5 DADE SYS.	33 DADE INNOV		3	25.300	0.787	3.11	21.500	29.100		
2018-555	25 IL ACL 100	13 HEMOSIL PT		1	23.125	2.022	8.74	19.700	26.600		
2018-555	33 MLA ELECT.	37 QUICK COAG		11	28.673	1.340	4.67	24.400	33.000		
2018-555	45 BIOMERIEUX	8 Thromborel		5	33.640	0.647	1.92	28.600	38.700		
2018-555	45 BIOMERIEUX	37 QUICK COAG		3	29.533	2.439	8.26	25.100	34.000		
2018-555	57 ACL 2-6000	13 HEMOSIL PT		1	23.125	2.022	8.74	19.700	26.600		
2018-555	58 AMELUNG KC	15 BIOMERIEUX		13	19.242	0.950	4.94	16.400	22.100		
2018-555	58 AMELUNG KC	33 DADE INNOV		6	23.800	1.179	4.95	20.200	27.400		
2018-555	60 DADE SYSME	33 DADE INNOV		19	26.389	0.681	2.58	22.400	30.300		
2018-555	62 DADE SYS.	33 DADE INNOV		66	24.970	1.271	5.09	21.200	28.700		
2018-555	65 DADE BFTII	8 Thromborel		1	33.640	0.647	1.92	28.600	38.700		
2018-555	65 DADE BFTII	9 DADE TH. C		1	25.300	0.787	3.11	21.500	29.100		
2018-555	65 DADE BFTII	33 DADE INNOV		16	25.169	1.408	5.60	21.400	28.900		
2018-555	69 DIAG.STAGO	36 STA NEOPLA		3	31.900	0.864	2.71	27.100	36.700		
2018-555	69 DIAG.STAGO	44 STA NEO CI		7	31.514	2.064	6.55	26.800	36.200		
2018-555	72 PACI HEMOS	39 TROMBO.-D		4	23.125	2.022	8.74	19.700	26.600		
2018-555	73 ACL 8 TO E	13 HEMOSIL PT		2	23.125	2.022	8.74	19.700	26.600		
2018-555	73 ACL 8 TO E	41 HEMOSIL RE		25	31.136	1.628	5.23	26.500	35.800		
2018-555	73 ACL 8 TO E	43 HEMOSIL PT		12	35.755	1.241	3.47	30.400	41.100		
2018-555	76 STAGO STAR	28 STA NEOPL.		3	33.433	3.297	9.86	28.400	38.400		
2018-555	76 STAGO STAR	44 STA NEO CI		14	34.786	1.921	5.52	29.600	40.000		

**Participant Summary**  
**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
075 RED CELL COUNT (RBC)			CLIA TEST ID #: 775      EVALUATION CRITERIA: TARGET VALUE 6 %					
2018-521	33	COULTER JT	13	3.362	0.101	3.00	3.160	3.560
2018-521	46	COULTER S	7	3.336	0.069	2.06	3.140	3.540
2018-521	84	ABBOTT C-D	9	3.248	0.299	9.20	3.050	3.440
2018-521	86	ABBOTT C-D	5	3.360	0.039	1.18	3.160	3.560
2018-521	98	SYSMEX K-	4	3.235	0.027	0.83	3.040	3.430
2018-521	115	ABBOTT C-D	149	3.338	0.066	1.96	3.140	3.540
2018-521	116	ADVIA-120	20	4.742	0.087	1.84	4.460	5.030
2018-521	120	COULTER	5	3.174	0.387	12.20	2.980	3.360
2018-521	123	EX.22/2280	14	3.336	0.042	1.24	3.140	3.540
2018-521	125	COULTER AC	151	2.327	0.042	1.79	2.190	2.470
2018-521	126	COULTER LH	16	3.331	0.047	1.42	3.130	3.530
2018-521	127	SYSMEX XT	24	3.359	0.026	0.79	3.160	3.560
2018-521	128	C. LH-500	17	3.426	0.085	2.49	3.220	3.630
2018-521	130	ABX PENTRA	21	2.318	0.041	1.78	2.180	2.460
2018-521	132	SYSMEX XS	136	3.294	0.041	1.25	3.100	3.490
2018-521	137	B.C DXH	40	3.307	0.041	1.24	3.110	3.510
2018-521	138	SY KX-21N	3	3.257	0.024	0.72	3.060	3.450
2018-521	145	BC-5390	17	3.280	0.053	1.61	3.080	3.480
2018-522	33	COULTER JT	13	4.493	0.100	2.22	4.220	4.760
2018-522	46	COULTER S	7	4.481	0.100	2.23	4.210	4.750
2018-522	84	ABBOTT C-D	9	4.557	0.114	2.49	4.280	4.830
2018-522	86	ABBOTT C-D	5	4.502	0.048	1.06	4.230	4.770
2018-522	98	SYSMEX K-	4	4.405	0.081	1.83	4.140	4.670
2018-522	115	ABBOTT C-D	149	4.527	0.089	1.97	4.260	4.800
2018-522	116	ADVIA-120	20	2.530	0.052	2.06	2.380	2.680
2018-522	120	COULTER	5	4.608	0.168	3.65	4.330	4.880
2018-522	123	EX.22/2280	14	4.508	0.053	1.18	4.240	4.780

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-522	125	COULTER AC		151	4.654	0.068	1.45	4.370	4.930
2018-522	126	COULTER LH		16	4.493	0.055	1.23	4.220	4.760
2018-522	127	SYSMEX XT		24	4.535	0.049	1.09	4.260	4.810
2018-522	128	C. LH-500		17	4.530	0.068	1.51	4.260	4.800
2018-522	130	ABX PENTRA		21	4.627	0.054	1.16	4.350	4.900
2018-522	132	SYSMEX XS		136	4.467	0.049	1.10	4.200	4.730
2018-522	137	B.C DXH		40	4.442	0.055	1.24	4.180	4.710
2018-522	138	SY KX-21N		3	4.390	0.028	0.64	4.130	4.650
2018-522	145	BC-5390		17	4.473	0.061	1.36	4.200	4.740
075 RED CELL COUNT (RBC)				CLIA TEST ID #: 775      EVALUATION CRITERIA: TARGET VALUE 6 %					
2018-523	33	COULTER JT		13	4.766	0.103	2.16	4.480	5.050
2018-523	46	COULTER S		7	4.743	0.113	2.38	4.460	5.030
2018-523	84	ABBOTT C-D		9	4.727	0.178	3.76	4.440	5.010
2018-523	86	ABBOTT C-D		5	4.742	0.076	1.60	4.460	5.030
2018-523	98	SYSMEX K-		4	4.638	0.035	0.75	4.360	4.920
2018-523	115	ABBOTT C-D		149	4.771	0.098	2.05	4.490	5.060
2018-523	116	ADVIA-120		20	5.488	0.093	1.70	5.160	5.820
2018-523	120	COULTER		5	4.850	0.247	5.10	4.560	5.140
2018-523	123	EX.22/2280		14	4.725	0.070	1.49	4.440	5.010
2018-523	125	COULTER AC		151	5.070	0.081	1.59	4.770	5.370
2018-523	126	COULTER LH		16	4.741	0.060	1.26	4.460	5.030
2018-523	127	SYSMEX XT		24	4.774	0.068	1.43	4.490	5.060
2018-523	128	C. LH-500		17	4.792	0.080	1.67	4.500	5.080
2018-523	130	ABX PENTRA		21	5.033	0.071	1.40	4.730	5.330
2018-523	132	SYSMEX XS		136	4.722	0.051	1.08	4.440	5.010
2018-523	137	B.C DXH		40	4.665	0.053	1.13	4.390	4.940
2018-523	138	SY KX-21N		3	4.610	0.043	0.94	4.330	4.890
2018-523	145	BC-5390		17	4.705	0.089	1.90	4.420	4.990
2018-524	33	COULTER JT		13	3.325	0.067	2.00	3.130	3.520

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-524	46	COULTER S		7	3.337	0.070	2.11	3.140	3.540
2018-524	84	ABBOTT C-D		9	3.537	0.442	12.50	3.320	3.750
2018-524	86	ABBOTT C-D		5	3.350	0.043	1.28	3.150	3.550
2018-524	98	SYSMEX K-		4	3.278	0.023	0.69	3.080	3.470
2018-524	115	ABBOTT C-D		149	3.352	0.060	1.79	3.150	3.550
2018-524	116	ADVIA-120		20	4.709	0.096	2.03	4.430	4.990
2018-524	120	COULTER		5	3.664	0.578	15.77	3.440	3.880
2018-524	123	EX.22/2280		14	3.354	0.040	1.19	3.150	3.550
2018-524	125	COULTER AC		151	4.651	0.077	1.65	4.370	4.930
2018-524	126	COULTER LH		16	3.345	0.046	1.37	3.140	3.550
2018-524	127	SYSMEX XT		24	3.373	0.044	1.30	3.170	3.580
2018-524	128	C. LH-500		17	3.391	0.074	2.20	3.190	3.590
2018-524	130	ABX PENTRA		21	4.622	0.066	1.43	4.350	4.900
2018-524	132	SYSMEX XS		136	3.289	0.037	1.13	3.090	3.490
2018-524	137	B.C DXH		40	3.309	0.043	1.29	3.110	3.510
2018-524	138	SY KX-21N		3	3.233	0.021	0.64	3.040	3.430
2018-524	145	BC-5390		17	3.283	0.058	1.76	3.090	3.480
075 RED CELL COUNT (RBC)		CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %					
2018-525	33	COULTER JT		13	5.207	0.106	2.04	4.890	5.520
2018-525	46	COULTER S		7	5.171	0.108	2.09	4.860	5.480
2018-525	84	ABBOTT C-D		9	4.953	0.915	18.48	4.660	5.250
2018-525	86	ABBOTT C-D		5	5.202	0.104	2.00	4.890	5.510
2018-525	98	SYSMEX K-		4	5.095	0.072	1.42	4.790	5.400
2018-525	115	ABBOTT C-D		149	5.284	0.108	2.05	4.970	5.600
2018-525	116	ADVIA-120		20	5.485	0.091	1.66	5.160	5.810
2018-525	120	COULTER		5	4.712	1.165	24.73	4.430	4.990
2018-525	123	EX.22/2280		14	5.228	0.053	1.02	4.910	5.540
2018-525	125	COULTER AC		151	2.321	0.043	1.87	2.180	2.460
2018-525	126	COULTER LH		16	5.165	0.066	1.27	4.860	5.470
2018-525	127	SYSMEX XT		24	5.201	0.083	1.60	4.890	5.510
2018-525	128	C. LH-500		17	5.232	0.113	2.15	4.920	5.550

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-525	130	ABX PENTRA		21	2.323	0.040	1.71	2.180	2.460
2018-525	132	SYSMEX XS		136	5.171	0.055	1.07	4.860	5.480
2018-525	137	B.C DXH		40	5.097	0.061	1.21	4.790	5.400
2018-525	138	SY KX-21N		3	5.037	0.029	0.57	4.730	5.340
2018-525	145	BC-5390		17	5.206	0.088	1.69	4.890	5.520
076 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %					
2018-521	33	COULTER JT		13	26.177	0.814	3.11	24.600	27.700
2018-521	46	COULTER S		7	25.514	0.464	1.82	24.000	27.000
2018-521	84	ABBOTT C-D		9	24.967	1.508	6.04	23.500	26.500
2018-521	86	ABBOTT C-D		6	26.233	1.115	4.25	24.700	27.800
2018-521	98	SYSMEX K-		4	23.850	1.132	4.75	22.400	25.300
2018-521	115	ABBOTT C-D		149	23.209	0.620	2.67	21.800	24.600
2018-521	116	ADVIA-120		20	35.900	0.875	2.44	33.700	38.100
2018-521	120	COULTER		5	24.840	3.009	12.11	23.300	26.300
2018-521	123	EX.22/2280		14	26.764	0.471	1.76	25.200	28.400
2018-521	125	COULTER AC		152	18.970	0.432	2.28	17.800	20.100
2018-521	126	COULTER LH		16	26.069	0.415	1.59	24.500	27.600
2018-521	127	SYSMEX XT		24	26.304	0.424	1.61	24.700	27.900
2018-521	128	C. LH-500		17	26.594	0.579	2.18	25.000	28.200
2018-521	130	ABX PENTRA		21	18.786	0.407	2.17	17.700	19.900
2018-521	132	SYSMEX XS		135	25.988	0.407	1.57	24.400	27.500
2018-521	137	B.C DXH		40	26.943	0.394	1.46	25.300	28.600
2018-521	138	SY KX-21N		3	24.233	0.262	1.08	22.800	25.700
2018-521	145	BC-5390		17	25.676	0.440	1.71	24.100	27.200
2018-522	33	COULTER JT		13	38.508	0.767	1.99	36.200	40.800
2018-522	46	COULTER S		7	37.514	0.874	2.33	35.300	39.800
2018-522	84	ABBOTT C-D		9	38.611	2.707	7.01	36.300	40.900
2018-522	86	ABBOTT C-D		6	38.633	2.189	5.67	36.300	41.000
2018-522	98	SYSMEX K-		4	35.550	2.254	6.34	33.400	37.700

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-522	115	ABBOTT C-D		149	34.017	0.908	2.67	32.000	36.100
2018-522	116	ADVIA-120		20	16.150	0.349	2.16	15.200	17.100
2018-522	120	COULTER		5	39.400	1.801	4.57	37.000	41.800
2018-522	123	EX.22/2280		14	39.700	0.859	2.16	37.300	42.100
2018-522	125	COULTER AC		152	38.364	0.701	1.83	36.100	40.700
2018-522	126	COULTER LH		16	38.625	0.547	1.42	36.300	40.900
2018-522	127	SYSMEX XT		24	38.713	0.703	1.82	36.400	41.000
2018-522	128	C. LH-500		17	38.606	0.533	1.38	36.300	40.900
2018-522	130	ABX PENTRA		21	38.243	0.546	1.43	35.900	40.500
2018-522	132	SYSMEX XS		135	38.697	0.593	1.53	36.400	41.000
2018-522	137	B.C DXH		40	39.393	0.453	1.15	37.000	41.800
2018-522	138	SY KX-21N		3	35.633	0.340	0.95	33.500	37.800
2018-522	145	BC-5390		17	38.212	0.571	1.49	35.900	40.500
076 HEMATOCRIT (HCT)			CLIA TEST ID #: 785				EVALUATION CRITERIA: TARGET VALUE 6 %		
2018-523	33	COULTER JT		13	41.062	0.947	2.31	38.600	43.500
2018-523	46	COULTER S		7	39.886	0.645	1.62	37.500	42.300
2018-523	84	ABBOTT C-D		9	40.700	3.882	9.54	38.300	43.100
2018-523	86	ABBOTT C-D		6	40.967	2.395	5.85	38.500	43.400
2018-523	98	SYSMEX K-		4	37.775	2.269	6.01	35.500	40.000
2018-523	115	ABBOTT C-D		149	36.139	0.930	2.57	34.000	38.300
2018-523	116	ADVIA-120		20	46.745	1.119	2.39	43.900	49.500
2018-523	120	COULTER		5	42.040	2.573	6.12	39.500	44.600
2018-523	123	EX.22/2280		14	41.929	1.135	2.71	39.400	44.400
2018-523	125	COULTER AC		152	43.447	0.868	2.00	40.800	46.100
2018-523	126	COULTER LH		16	40.888	0.533	1.30	38.400	43.300
2018-523	127	SYSMEX XT		24	40.988	0.704	1.72	38.500	43.400
2018-523	128	C. LH-500		17	41.053	0.665	1.62	38.600	43.500
2018-523	130	ABX PENTRA		21	43.267	0.688	1.59	40.700	45.900
2018-523	132	SYSMEX XS		135	41.010	0.630	1.54	38.500	43.500
2018-523	137	B.C DXH		40	41.535	0.486	1.17	39.000	44.000
2018-523	138	SY KX-21N		3	37.700	0.283	0.75	35.400	40.000

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 130 COMPL BLOOD COUNT

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-523	145 BC-5390				17	40.371	0.751	1.86	37.900	42.800
076 HEMATOCRIT (HCT)			CLIA TEST ID #: 785	EVALUATION CRITERIA: TARGET VALUE 6 %						
2018-524	33 COULTER JT				13	29.300	0.587	2.00	27.500	31.100
2018-524	46 COULTER S				7	28.757	0.507	1.76	27.000	30.500
2018-524	84 ABBOTT C-D				9	30.856	5.006	16.23	29.000	32.700
2018-524	86 ABBOTT C-D				6	29.383	1.623	5.52	27.600	31.100
2018-524	98 SYSMEX K-				4	27.200	1.595	5.87	25.600	28.800
2018-524	115 ABBOTT C-D				149	25.653	0.596	2.32	24.100	27.200
2018-524	116 ADVIA-120				20	35.610	0.831	2.33	33.500	37.700
2018-524	120 COULTER				5	32.100	4.072	12.68	30.200	34.000
2018-524	123 EX.22/2280				14	30.507	0.574	1.88	28.700	32.300
2018-524	125 COULTER AC				152	38.311	0.921	2.40	36.000	40.600
2018-524	126 COULTER LH				16	29.569	0.466	1.58	27.800	31.300
2018-524	127 SYSMEX XT				24	30.038	0.519	1.73	28.200	31.800
2018-524	128 C. LH-500				17	29.741	0.463	1.56	28.000	31.500
2018-524	130 ABX PENTRA				21	38.200	0.598	1.57	35.900	40.500
2018-524	132 SYSMEX XS				135	29.709	0.465	1.56	27.900	31.500
2018-524	137 B.C DXH				40	30.463	0.408	1.34	28.600	32.300
2018-524	138 SY KX-21N				3	26.867	0.309	1.15	25.300	28.500
2018-524	145 BC-5390				17	29.000	0.519	1.79	27.300	30.700
2018-525	33 COULTER JT				13	48.477	1.116	2.30	45.600	51.400
2018-525	46 COULTER S				7	47.771	0.587	1.23	44.900	50.600
2018-525	84 ABBOTT C-D				9	42.867	9.773	22.80	40.300	45.400
2018-525	86 ABBOTT C-D				6	48.667	3.060	6.29	45.700	51.600
2018-525	98 SYSMEX K-				4	44.650	2.729	6.11	42.000	47.300
2018-525	115 ABBOTT C-D				149	42.763	1.146	2.68	40.200	45.300
2018-525	116 ADVIA-120				20	47.058	0.876	1.86	44.200	49.900
2018-525	120 COULTER				5	43.740	12.461	28.49	41.100	46.400
2018-525	123 EX.22/2280				14	50.114	1.493	2.98	47.100	53.100



## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-525	125	COULTER AC		152	18.876	0.410	2.17	17.700	20.000
2018-525	126	COULTER LH		16	48.625	0.809	1.66	45.700	51.500
2018-525	127	SYSMEX XT		24	48.667	0.937	1.92	45.700	51.600
2018-525	128	C. LH-500		17	48.488	0.975	2.01	45.600	51.400
2018-525	130	ABX PENTRA		21	18.838	0.385	2.04	17.700	20.000
2018-525	132	SYSMEX XS		135	48.857	0.721	1.48	45.900	51.800
2018-525	137	B.C DXH		40	49.140	0.675	1.37	46.200	52.100
2018-525	138	SY KX-21N		3	44.467	0.471	1.06	41.800	47.100
2018-525	145	BC-5390		17	48.065	0.876	1.82	45.200	50.900
077	HEMOGLOBIN (HGB)		CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 %					
2018-521	33	COULTER JT		13	8.985	0.238	2.65	8.400	9.600
2018-521	46	COULTER S		7	8.757	0.118	1.35	8.100	9.400
2018-521	84	ABBOTT C-D		9	8.556	0.746	8.71	8.000	9.200
2018-521	86	ABBOTT C-D		6	8.983	0.134	1.50	8.400	9.600
2018-521	98	SYSMEX K-		5	8.760	0.162	1.85	8.100	9.400
2018-521	115	ABBOTT C-D		149	9.143	0.207	2.26	8.500	9.800
2018-521	116	ADVIA-120		20	13.200	0.235	1.78	12.300	14.100
2018-521	120	COULTER		5	8.500	0.846	9.95	7.900	9.100
2018-521	123	EX.22/2280		14	9.043	0.184	2.04	8.400	9.700
2018-521	125	COULTER AC		152	6.771	0.114	1.68	6.300	7.200
2018-521	126	COULTER LH		16	8.913	0.078	0.88	8.300	9.500
2018-521	127	SYSMEX XT		24	8.817	0.107	1.21	8.200	9.400
2018-521	128	C. LH-500		17	9.029	0.187	2.07	8.400	9.700
2018-521	130	ABX PENTRA		21	6.667	0.094	1.41	6.200	7.100
2018-521	132	SYSMEX XS		135	8.629	0.103	1.19	8.000	9.200
2018-521	136	HEMO-CUE		3	8.667	0.170	1.96	8.100	9.300
2018-521	137	B.C DXH		40	8.823	0.099	1.12	8.200	9.400
2018-521	138	SY KX-21N		3	8.833	0.047	0.53	8.200	9.500
2018-521	145	BC-5390		16	8.713	0.111	1.28	8.100	9.300

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
077	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 %						
2018-522	33	COULTER JT		13	13.231	0.297	2.25	12.300	14.200
2018-522	46	COULTER S		7	13.271	0.255	1.92	12.300	14.200
2018-522	84	ABBOTT C-D		9	13.133	0.611	4.65	12.200	14.100
2018-522	86	ABBOTT C-D		6	13.333	0.160	1.20	12.400	14.300
2018-522	98	SYSMEX K-		5	13.200	0.089	0.68	12.300	14.100
2018-522	115	ABBOTT C-D		149	13.387	0.252	1.89	12.400	14.300
2018-522	116	ADVIA-120		20	5.820	0.117	2.00	5.400	6.200
2018-522	120	COULTER		5	13.540	0.344	2.54	12.600	14.500
2018-522	123	EX.22/2280		14	13.421	0.245	1.83	12.500	14.400
2018-522	125	COULTER AC		152	13.879	0.185	1.33	12.900	14.900
2018-522	126	COULTER LH		16	13.313	0.122	0.92	12.400	14.200
2018-522	127	SYSMEX XT		24	13.246	0.104	0.79	12.300	14.200
2018-522	128	C. LH-500		17	13.312	0.153	1.15	12.400	14.200
2018-522	130	ABX PENTRA		21	13.995	0.162	1.16	13.000	15.000
2018-522	132	SYSMEX XS		135	13.157	0.128	0.97	12.200	14.100
2018-522	136	HEMO-CUE		3	12.933	0.403	3.11	12.000	13.800
2018-522	137	B.C DXH		40	13.193	0.121	0.92	12.300	14.100
2018-522	138	SY KX-21N		3	13.067	0.125	0.95	12.200	14.000
2018-522	145	BC-5390		16	13.044	0.137	1.05	12.100	14.000
2018-523	33	COULTER JT		13	13.954	0.325	2.33	13.000	14.900
2018-523	46	COULTER S		7	14.029	0.333	2.37	13.000	15.000
2018-523	84	ABBOTT C-D		9	13.822	0.868	6.28	12.900	14.800
2018-523	86	ABBOTT C-D		6	14.050	0.096	0.68	13.100	15.000
2018-523	98	SYSMEX K-		5	14.020	0.075	0.53	13.000	15.000
2018-523	115	ABBOTT C-D		149	14.077	0.250	1.78	13.100	15.100
2018-523	116	ADVIA-120		20	17.420	0.347	1.99	16.200	18.600
2018-523	120	COULTER		5	14.440	0.789	5.46	13.400	15.500
2018-523	123	EX.22/2280		14	14.154	0.169	1.20	13.200	15.100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-523	125	COULTER AC		152	15.715	0.218	1.39	14.600	16.800
2018-523	126	COULTER LH		16	14.069	0.140	1.00	13.100	15.100
2018-523	127	SYSMEX XT		24	14.008	0.108	0.77	13.000	15.000
2018-523	128	C. LH-500		17	14.082	0.238	1.69	13.100	15.100
2018-523	130	ABX PENTRA		21	15.914	0.196	1.23	14.800	17.000
2018-523	132	SYSMEX XS		135	13.950	0.123	0.88	13.000	14.900
2018-523	136	HEMO-CUE		3	13.800	0.294	2.13	12.800	14.800
2018-523	137	B.C DXH		40	13.969	0.114	0.81	13.000	14.900
2018-523	138	SY KX-21N		3	13.833	0.125	0.90	12.900	14.800
2018-523	145	BC-5390		16	13.744	0.146	1.06	12.800	14.700
077 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 %					
2018-524	33	COULTER JT		13	10.083	0.114	1.13	9.400	10.800
2018-524	46	COULTER S		7	10.043	0.213	2.12	9.300	10.700
2018-524	84	ABBOTT C-D		9	10.389	1.419	13.66	9.700	11.100
2018-524	86	ABBOTT C-D		6	10.050	0.189	1.88	9.300	10.800
2018-524	98	SYSMEX K-		5	10.080	0.133	1.32	9.400	10.800
2018-524	115	ABBOTT C-D		149	10.234	0.236	2.31	9.500	11.000
2018-524	116	ADVIA-120		20	13.250	0.269	2.03	12.300	14.200
2018-524	120	COULTER		5	10.900	1.563	14.34	10.100	11.700
2018-524	123	EX.22/2280		14	10.164	0.199	1.95	9.500	10.900
2018-524	125	COULTER AC		152	13.894	0.181	1.31	12.900	14.900
2018-524	126	COULTER LH		16	10.119	0.088	0.87	9.400	10.800
2018-524	127	SYSMEX XT		24	10.104	0.098	0.97	9.400	10.800
2018-524	128	C. LH-500		17	10.188	0.171	1.68	9.500	10.900
2018-524	130	ABX PENTRA		21	13.957	0.224	1.60	13.000	14.900
2018-524	132	SYSMEX XS		135	9.945	0.103	1.04	9.200	10.600
2018-524	136	HEMO-CUE		3	9.900	0.141	1.43	9.200	10.600
2018-524	137	B.C DXH		40	10.143	0.120	1.18	9.400	10.900
2018-524	138	SY KX-21N		3	10.100	0.082	0.81	9.400	10.800
2018-524	145	BC-5390		16	9.856	0.132	1.34	9.200	10.500

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
077	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE	7	%				
2018-525	33	COULTER JT		13	16.617	0.134	0.81	15.500	17.800
2018-525	46	COULTER S		7	16.800	0.428	2.55	15.600	18.000
2018-525	84	ABBOTT C-D		9	15.411	3.176	20.61	14.300	16.500
2018-525	86	ABBOTT C-D		6	16.933	0.047	0.28	15.700	18.100
2018-525	98	SYSMEX K-		5	16.860	0.206	1.22	15.700	18.000
2018-525	115	ABBOTT C-D		149	16.884	0.303	1.79	15.700	18.100
2018-525	116	ADVIA-120		20	17.450	0.369	2.12	16.200	18.700
2018-525	120	COULTER		5	14.940	4.042	27.06	13.900	16.000
2018-525	123	EX.22/2280		14	17.107	0.365	2.14	15.900	18.300
2018-525	125	COULTER AC		152	6.771	0.104	1.54	6.300	7.200
2018-525	126	COULTER LH		16	16.638	0.190	1.14	15.500	17.800
2018-525	127	SYSMEX XT		24	16.742	0.193	1.16	15.600	17.900
2018-525	128	C. LH-500		17	16.741	0.243	1.45	15.600	17.900
2018-525	130	ABX PENTRA		21	6.681	0.096	1.43	6.200	7.100
2018-525	132	SYSMEX XS		135	16.836	0.154	0.92	15.700	18.000
2018-525	136	HEMO-CUE		3	16.167	0.450	2.78	15.000	17.300
2018-525	137	B.C DXH		40	16.555	0.177	1.07	15.400	17.700
2018-525	138	SY KX-21N		3	16.733	0.125	0.75	15.600	17.900
2018-525	145	BC-5390		16	16.556	0.187	1.13	15.400	17.700
078	WHITE CELL COUNT-WBC	CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE	15	%				
2018-521	33	COULTER JT		13	12.200	0.643	5.27	10.400	14.000
2018-521	46	COULTER S		7	12.171	0.361	2.97	10.300	14.000
2018-521	84	ABBOTT C-D		11	11.455	3.010	26.28	9.700	13.200
2018-521	86	ABBOTT C-D		6	11.200	0.493	4.40	9.500	12.900
2018-521	98	SYSMEX K-		5	11.440	0.801	7.01	9.700	13.200
2018-521	115	ABBOTT C-D		148	11.176	0.360	3.22	9.500	12.900
2018-521	116	ADVIA-120		20	7.380	0.394	5.35	6.300	8.500
2018-521	120	COULTER		5	10.420	4.089	39.24	8.900	12.000
2018-521	123	EX.22/2280		14	11.571	0.219	1.89	9.800	13.300

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-521	125	COULTER AC		152	2.373	0.086	3.63	2.000	2.700
2018-521	126	COULTER LH		17	11.988	0.312	2.60	10.200	13.800
2018-521	127	SYSMEX XT		24	12.046	0.473	3.92	10.200	13.900
2018-521	128	C. LH-500		17	12.131	0.406	3.34	10.300	14.000
2018-521	130	ABX PENTRA		21	2.371	0.063	2.65	2.000	2.700
2018-521	132	SYSMEX XS		134	12.209	0.227	1.86	10.400	14.000
2018-521	137	B.C DXH		39	12.146	0.227	1.87	10.300	14.000
2018-521	138	SY KX-21N		3	11.233	0.249	2.22	9.500	12.900
2018-521	145	BC-5390		16	12.100	0.260	2.15	10.300	13.900
078 WHITE CELL COUNT-WBC		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %					
2018-522	33	COULTER JT		13	6.300	0.260	4.13	5.400	7.200
2018-522	46	COULTER S		7	6.457	0.272	4.21	5.500	7.400
2018-522	84	ABBOTT C-D		11	6.764	0.790	11.68	5.700	7.800
2018-522	86	ABBOTT C-D		6	6.033	0.275	4.56	5.100	6.900
2018-522	98	SYSMEX K-		5	6.060	0.314	5.18	5.200	7.000
2018-522	115	ABBOTT C-D		148	6.067	0.182	3.01	5.200	7.000
2018-522	116	ADVIA-120		20	2.950	0.124	4.22	2.500	3.400
2018-522	120	COULTER		5	6.660	0.356	5.34	5.700	7.700
2018-522	123	EX.22/2280		14	6.214	0.177	2.84	5.300	7.100
2018-522	125	COULTER AC		152	7.353	0.220	2.99	6.200	8.500
2018-522	126	COULTER LH		17	6.300	0.203	3.22	5.400	7.200
2018-522	127	SYSMEX XT		24	6.242	0.261	4.19	5.300	7.200
2018-522	128	C. LH-500		17	6.324	0.190	3.00	5.400	7.300
2018-522	130	ABX PENTRA		21	7.295	0.168	2.30	6.200	8.400
2018-522	132	SYSMEX XS		134	6.404	0.162	2.54	5.400	7.400
2018-522	137	B.C DXH		39	6.323	0.151	2.39	5.400	7.300
2018-522	138	SY KX-21N		3	6.067	0.125	2.06	5.200	7.000
2018-522	145	BC-5390		16	6.356	0.127	2.00	5.400	7.300
2018-523	33	COULTER JT		13	3.123	0.112	3.59	2.700	3.600

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-523	46	COULTER S		7	3.071	0.070	2.28	2.600	3.500
2018-523	84	ABBOTT C-D		11	3.420	0.861	25.18	2.900	3.900
2018-523	86	ABBOTT C-D		6	2.900	0.100	3.45	2.500	3.300
2018-523	98	SYSMEX K-		5	2.940	0.102	3.47	2.500	3.400
2018-523	115	ABBOTT C-D		148	2.880	0.101	3.51	2.400	3.300
2018-523	116	ADVIA-120		20	15.595	0.681	4.37	13.300	17.900
2018-523	120	COULTER		5	5.660	5.120	90.47	4.800	6.500
2018-523	123	EX.22/2280		14	3.079	0.094	3.05	2.600	3.500
2018-523	125	COULTER AC		152	16.090	0.485	3.01	13.700	18.500
2018-523	126	COULTER LH		17	2.971	0.118	3.96	2.500	3.400
2018-523	127	SYSMEX XT		24	2.883	0.160	5.54	2.500	3.300
2018-523	128	C. LH-500		17	3.138	0.141	4.49	2.700	3.600
2018-523	130	ABX PENTRA		21	16.014	0.333	2.08	13.600	18.400
2018-523	132	SYSMEX XS		134	2.974	0.085	2.85	2.500	3.400
2018-523	137	B.C DXH		39	2.890	0.106	3.66	2.500	3.300
2018-523	138	SY KX-21N		3	2.933	0.125	4.25	2.500	3.400
2018-523	145	BC-5390		16	3.031	0.092	3.02	2.600	3.500
078	WHITE CELL COUNT-WBC		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %				
2018-524	33	COULTER JT		13	5.177	0.236	4.56	4.400	6.000
2018-524	46	COULTER S		7	5.129	0.045	0.88	4.400	5.900
2018-524	84	ABBOTT C-D		11	5.564	0.874	15.71	4.700	6.400
2018-524	86	ABBOTT C-D		6	4.950	0.126	2.54	4.200	5.700
2018-524	98	SYSMEX K-		5	4.920	0.271	5.51	4.200	5.700
2018-524	115	ABBOTT C-D		148	4.996	0.145	2.91	4.200	5.700
2018-524	116	ADVIA-120		20	7.455	0.309	4.14	6.300	8.600
2018-524	120	COULTER		5	5.680	0.806	14.19	4.800	6.500
2018-524	123	EX.22/2280		14	5.064	0.111	2.19	4.300	5.800
2018-524	125	COULTER AC		152	7.340	0.205	2.79	6.200	8.400
2018-524	126	COULTER LH		17	5.071	0.140	2.77	4.300	5.800
2018-524	127	SYSMEX XT		24	4.917	0.140	2.86	4.200	5.700
2018-524	128	C. LH-500		17	5.200	0.209	4.01	4.400	6.000

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 130 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-524	130	ABX PENTRA		21	7.305	0.140	1.91	6.200	8.400
2018-524	132	SYSMEX XS		134	5.105	0.118	2.32	4.300	5.900
2018-524	137	B.C DXH		39	5.110	0.106	2.07	4.300	5.900
2018-524	138	SY KX-21N		3	4.933	0.047	0.96	4.200	5.700
2018-524	145	BC-5390		16	5.194	0.130	2.50	4.400	6.000
078 WHITE CELL COUNT-WBC		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %					
2018-525	33	COULTER JT		13	17.023	0.802	4.71	14.500	19.600
2018-525	46	COULTER S		7	17.057	0.817	4.79	14.500	19.600
2018-525	84	ABBOTT C-D		11	15.600	4.414	28.29	13.300	17.900
2018-525	86	ABBOTT C-D		6	15.367	0.704	4.58	13.100	17.700
2018-525	98	SYSMEX K-		5	16.000	1.049	6.56	13.600	18.400
2018-525	115	ABBOTT C-D		148	15.107	0.486	3.22	12.800	17.400
2018-525	116	ADVIA-120		20	15.655	0.460	2.94	13.300	18.000
2018-525	120	COULTER		5	14.300	6.016	42.07	12.200	16.400
2018-525	123	EX.22/2280		14	16.079	0.300	1.87	13.700	18.500
2018-525	125	COULTER AC		152	2.362	0.086	3.64	2.000	2.700
2018-525	126	COULTER LH		17	16.282	0.463	2.84	13.800	18.700
2018-525	127	SYSMEX XT		24	15.958	0.708	4.43	13.600	18.400
2018-525	128	C. LH-500		17	16.729	0.411	2.46	14.200	19.200
2018-525	130	ABX PENTRA		21	2.343	0.079	3.38	2.000	2.700
2018-525	132	SYSMEX XS		134	16.783	0.330	1.97	14.300	19.300
2018-525	137	B.C DXH		39	16.936	0.436	2.58	14.400	19.500
2018-525	138	SY KX-21N		3	16.200	0.082	0.50	13.800	18.600
2018-525	145	BC-5390		16	16.463	0.352	2.14	14.000	18.900

## Participant Summary

MONTH 7 OF YEAR 2018

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2018-501					008	008	008 - EOSINOPHIL	840	839	839	100
2018-501					008	008	007 - BASOPHIL,	840	1	839	0
2018-502					014	014	014 - LYMPHOCYTE	839	750	750	100
2018-502					014	014	017 - MONOCYTE	839	89	750	0
2018-503					003	003	002 - PROMYELOCY	839	9	820	0
2018-503					003	003	014 - LYMPHOCYTE	839	3	820	0
2018-503					003	003	004 - METAMYELOC	839	1	820	0
2018-503					003	003	001 - BLAST	839	5	820	0
2018-503					003	003	003 - MYELOCYTE	839	820	820	100
2018-503					003	003	017 - MONOCYTE	839	1	820	0
2018-504					020	020	019 - PLATELETS,	839	1	837	0
2018-504					020	020	020 - GIANT PLA	839	837	837	100
2018-504					020	020	033 - POLYCHROMA	839	1	837	0
2018-505							018 - PLATELETS,	839	1	839	0
2018-505							048 - ERYTHROCYT	839	2	839	0
2018-505							013 - LYMPHOCYTE	839	2	839	0
2018-505							046 - ERYTHROCYT	839	267	839	0
2018-505							044 - ERYTHROCYT	839	1	839	0
2018-505							030 - ERYTHROCYT	839	524	839	0
2018-505							052 - SPHEROCYTE	839	42	839	0



**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 135 PLATELET COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
079	PLATELET COUNT	CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %							
2018-521	16	COULTER		7	421.857	32.673	7.75	316.000	527.000	
2018-521	31	ABBOTT C-D		8	454.375	50.515	11.12	341.000	568.000	
2018-521	33	ABBOTT C-D		6	477.833	24.883	5.21	358.000	597.000	
2018-521	39	SYSMEX K-		6	496.500	59.225	11.93	372.000	621.000	
2018-521	50	COULT.JT,M		13	437.769	28.991	6.62	328.000	547.000	
2018-521	63	ABBOTT C-D		148	469.048	21.043	4.49	352.000	586.000	
2018-521	64	ADVIA-120		20	218.100	15.086	6.92	164.000	273.000	
2018-521	68	COULTERACT		5	376.000	154.134	40.99	282.000	470.000	
2018-521	71	EX.22/2280		14	433.143	40.084	9.25	325.000	541.000	
2018-521	73	COULTER AC		152	69.762	4.443	6.37	52.000	87.000	
2018-521	74	COULTER LH		16	433.750	21.067	4.86	325.000	542.000	
2018-521	75	SYSMEX XT		24	509.826	14.233	2.79	382.000	637.000	
2018-521	76	COUL.LH-50		17	465.353	25.354	5.45	349.000	582.000	
2018-521	78	ABX PENTRA		21	70.667	4.998	7.07	53.000	88.000	
2018-521	79	SYSMEX XS		134	470.328	47.790	10.16	353.000	588.000	
2018-521	83	B.C600,800		40	461.575	12.542	2.72	346.000	577.000	
2018-521	84	KX21,300		3	518.000	4.243	0.82	389.000	648.000	
2018-521	91	BC-5390		16	483.250	36.574	7.57	362.000	604.000	
2018-522	16	COULTER		7	238.571	8.192	3.43	179.000	298.000	
2018-522	31	ABBOTT C-D		8	254.750	15.180	5.96	191.000	318.000	
2018-522	33	ABBOTT C-D		6	267.500	13.562	5.07	201.000	334.000	
2018-522	39	SYSMEX K-		6	266.333	22.209	8.34	200.000	333.000	
2018-522	50	COULT.JT,M		13	230.154	14.239	6.19	173.000	288.000	
2018-522	63	ABBOTT C-D		148	267.822	12.500	4.67	201.000	335.000	
2018-522	64	ADVIA-120		20	55.100	5.108	9.27	41.000	69.000	
2018-522	68	COULTERACT		5	263.600	38.176	14.48	198.000	330.000	
2018-522	71	EX.22/2280		14	250.143	15.047	6.02	188.000	313.000	
2018-522	73	COULTER AC		152	246.517	10.941	4.44	185.000	308.000	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 135 PLATELET COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2018-522	74	COULTER LH		16	235.438	8.958	3.80	177.000	294.000	
2018-522	75	SYSMEX XT		24	256.458	9.005	3.51	192.000	321.000	
2018-522	76	COUL.LH-50		17	240.353	11.438	4.76	180.000	300.000	
2018-522	78	ABX PENTRA		21	250.619	9.988	3.99	188.000	313.000	
2018-522	79	SYSMEX XS		134	236.963	18.374	7.75	178.000	296.000	
2018-522	83	B.C600,800		40	246.500	6.874	2.79	185.000	308.000	
2018-522	84	KX21,300		3	274.000	4.320	1.58	206.000	343.000	
2018-522	91	BC-5390		16	268.875	15.128	5.63	202.000	336.000	
079	PLATELET COUNT		CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %						
2018-523	16	COULTER		7	307.143	8.476	2.76	230.000	384.000	
2018-523	31	ABBOTT C-D		8	329.000	25.773	7.83	247.000	411.000	
2018-523	33	ABBOTT C-D		6	324.667	19.729	6.08	244.000	406.000	
2018-523	39	SYSMEX K-		6	327.667	24.527	7.49	246.000	410.000	
2018-523	50	COULT.JT,M		13	296.000	15.089	5.10	222.000	370.000	
2018-523	63	ABBOTT C-D		148	320.469	15.646	4.88	240.000	401.000	
2018-523	64	ADVIA-120		20	441.550	18.288	4.14	331.000	552.000	
2018-523	68	COULTERAcT		5	349.000	77.115	22.10	262.000	436.000	
2018-523	71	EX.22/2280		14	314.500	15.065	4.79	236.000	393.000	
2018-523	73	COULTER AC		152	476.311	17.109	3.59	357.000	595.000	
2018-523	74	COULTER LH		16	295.438	10.828	3.67	222.000	369.000	
2018-523	75	SYSMEX XT		24	321.250	10.682	3.33	241.000	402.000	
2018-523	76	COUL.LH-50		17	307.412	11.030	3.59	231.000	384.000	
2018-523	78	ABX PENTRA		21	479.143	21.700	4.53	359.000	599.000	
2018-523	79	SYSMEX XS		134	306.933	23.187	7.55	230.000	384.000	
2018-523	83	B.C600,800		40	311.300	7.135	2.29	233.000	389.000	
2018-523	84	KX21,300		3	340.667	13.912	4.08	256.000	426.000	
2018-523	91	BC-5390		16	329.188	21.472	6.52	247.000	411.000	
2018-524	16	COULTER		7	108.714	5.496	5.06	82.000	136.000	
2018-524	31	ABBOTT C-D		8	111.500	11.424	10.25	84.000	139.000	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 135 PLATELET COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-524	33	ABBOTT C-D		6	123.833	9.651	7.79	93.000	155.000
2018-524	39	SYSMEX K-		6	127.500	18.500	14.51	96.000	159.000
2018-524	50	COULT.JT,M		13	99.154	6.781	6.84	74.000	124.000
2018-524	63	ABBOTT C-D		148	130.408	7.280	5.58	98.000	163.000
2018-524	64	ADVIA-120		20	214.737	6.965	3.24	161.000	268.000
2018-524	68	COULTERACT		5	134.200	61.454	45.79	101.000	168.000
2018-524	71	EX.22/2280		14	111.857	12.082	10.80	84.000	140.000
2018-524	73	COULTER AC		152	245.747	10.226	4.16	184.000	307.000
2018-524	74	COULTER LH		16	104.250	4.161	3.99	78.000	130.000
2018-524	75	SYSMEX XT		24	113.667	4.506	3.96	85.000	142.000
2018-524	76	COUL.LH-50		17	104.125	3.887	3.73	78.000	130.000
2018-524	78	ABX PENTRA		21	247.381	9.162	3.70	186.000	309.000
2018-524	79	SYSMEX XS		134	104.754	10.140	9.68	79.000	131.000
2018-524	83	B.C600,800		40	108.450	3.768	3.47	81.000	136.000
2018-524	84	KX21,300		3	133.000	2.160	1.62	100.000	166.000
2018-524	91	BC-5390		16	126.938	9.250	7.29	95.000	159.000
079 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %					
2018-525	16	COULTER		7	619.143	29.098	4.70	464.000	774.000
2018-525	31	ABBOTT C-D		8	695.875	67.442	9.69	522.000	870.000
2018-525	33	ABBOTT C-D		6	642.833	28.876	4.49	482.000	804.000
2018-525	39	SYSMEX K-		6	665.000	43.432	6.53	499.000	831.000
2018-525	50	COULT.JT,M		13	601.308	27.373	4.55	451.000	752.000
2018-525	63	ABBOTT C-D		148	630.231	29.481	4.68	473.000	788.000
2018-525	64	ADVIA-120		20	440.316	14.715	3.34	330.000	550.000
2018-525	68	COULTERACT		5	533.800	245.593	46.01	400.000	667.000
2018-525	71	EX.22/2280		14	643.929	26.340	4.09	483.000	805.000
2018-525	73	COULTER AC		152	69.500	4.182	6.02	52.000	87.000
2018-525	74	COULTER LH		16	576.813	18.742	3.25	433.000	721.000
2018-525	75	SYSMEX XT		24	640.333	31.271	4.88	480.000	800.000
2018-525	76	COUL.LH-50		17	628.412	35.277	5.61	471.000	786.000
2018-525	78	ABX PENTRA		21	71.333	3.496	4.90	54.000	89.000

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 135 PLATELET COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2018-525	79	SYSMEX XS		134	644.992	28.639	4.44	484.000	806.000	
2018-525	83	B.C600,800		40	623.333	14.190	2.28	468.000	779.000	
2018-525	84	KX21,300		3	712.000	10.198	1.43	534.000	890.000	
2018-525	91	BC-5390		16	691.813	27.747	4.01	519.000	865.000	

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
149 RETICULOCYTE COUNT		CLIA TEST ID #: 0			EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}					
2018-531	1	CELL-DYN		2	4.167	1.410	33.85	0	8.400	
2018-531	7	M.B.MANUAL		23	11.400	2.961	25.97	2.500	20.300	
2018-531	9	M.B. RICCA		35	9.637	2.120	22.00	3.300	16.000	
2018-531	17	COUL.LH		11	7.818	0.369	4.72	5.900	9.800	
2018-531	18	SYSMEX XT		10	7.510	1.070	14.25	4.300	10.700	
2018-531	23	ABBOTT RUB		6	4.167	1.410	33.85	0	8.400	
2018-531	24	Unicel DxH		11	8.209	0.434	5.28	6.200	10.300	
2018-531	25	SYSMEX XN		6	7.100	0.577	8.13	5.300	8.900	
2018-532	1	CELL-DYN		2	1.150	0.275	23.95	0.300	2.000	
2018-532	7	M.B.MANUAL		23	1.945	0.604	31.02	0.100	3.800	
2018-532	9	M.B. RICCA		35	1.740	0.500	28.76	0.200	3.200	
2018-532	17	COUL.LH		11	1.355	0.197	14.55	0.800	1.900	
2018-532	18	SYSMEX XT		10	1.170	0.249	21.30	0.400	1.900	
2018-532	23	ABBOTT RUB		6	1.150	0.275	23.95	0.300	2.000	
2018-532	24	Unicel DxH		11	1.127	0.086	7.65	0.800	1.400	
2018-532	25	SYSMEX XN		6	1.200	0.129	10.76	0.800	1.600	
2018-533	1	CELL-DYN		2	2.133	0.621	29.11	0.300	4.000	
2018-533	7	M.B.MANUAL		23	5.814	1.776	30.55	0.500	11.100	

**Participant Summary**  
**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 137 RETICULOCYTE COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2018-533	9	M.B. RICCA		35	5.094	1.059	20.78	1.900	8.300
2018-533	17	COUL.LH		11	3.682	0.134	3.63	2.800	4.600
2018-533	18	SYSMEX XT		10	3.230	0.510	15.79	1.700	4.800
2018-533	23	ABBOTT RUB		6	2.133	0.621	29.11	0.300	4.000
2018-533	24	Unicel DxH		11	3.800	0.259	6.83	2.900	4.800
2018-533	25	SYSMEX XN		6	3.250	0.377	11.62	2.100	4.400
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}						
2018-534	1	CELL-DYN		2	1.150	0.180	15.68	0.600	1.700
2018-534	7	M.B.MANUAL		23	2.009	0.735	36.58	0	4.200
2018-534	9	M.B. RICCA		35	1.789	0.571	31.93	0.100	3.500
2018-534	17	COUL.LH		11	1.309	0.188	14.37	0.700	1.900
2018-534	18	SYSMEX XT		10	1.180	0.227	19.25	0.500	1.900
2018-534	23	ABBOTT RUB		6	1.150	0.180	15.68	0.600	1.700
2018-534	24	Unicel DxH		11	1.118	0.111	9.96	0.800	1.500
2018-534	25	SYSMEX XN		6	1.167	0.075	6.39	0.900	1.500
2018-535	1	CELL-DYN		2	4.150	1.078	25.98	0.900	7.400
2018-535	7	M.B.MANUAL		23	11.423	3.008	26.33	2.400	20.400
2018-535	9	M.B. RICCA		35	10.069	1.981	19.67	4.100	16.000
2018-535	17	COUL.LH		11	7.882	0.333	4.22	5.900	9.900
2018-535	18	SYSMEX XT		10	7.560	1.265	16.73	3.800	11.400
2018-535	23	ABBOTT RUB		6	4.150	1.078	25.98	0.900	7.400
2018-535	24	Unicel DxH		11	8.127	0.512	6.30	6.100	10.200
2018-535	25	SYSMEX XN		6	6.950	0.665	9.57	5.000	8.900

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 138 SED. RATE (ESR)

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
153	SED. RATE (ESR)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{GREATER}			
2018-541	3	SEDIGREN		3	40.000	3.559	8.90	29.000	51.000
2018-541	4	SEDIPLAST		560	36.878	9.132	24.76	9.000	64.000
2018-541	10	SEDI RATE		124	42.098	10.059	23.89	12.000	72.000
2018-541	11	SEDIMAT-15		64	132.286	13.255	10.02	93.000	172.000
2018-541	14	GLOBAL SC.		9	45.889	11.939	26.02	10.000	82.000
2018-541	17	SEDI SYSTE		7	40.286	6.111	15.17	22.000	59.000
2018-541	19	MINI CUBE		8	59.125	11.879	20.09	23.000	95.000
2018-542	3	SEDIGREN		3	4.000	2.160	54.01	0	10.000
2018-542	4	SEDIPLAST		560	1.499	0.993	66.24	0	4.000
2018-542	10	SEDI RATE		124	1.992	0.746	37.45	0.000	4.000
2018-542	11	SEDIMAT-15		64	4.571	1.697	37.12	0	10.000
2018-542	14	GLOBAL SC.		9	2.111	0.875	41.44	0	5.000
2018-542	17	SEDI SYSTE		7	1.714	0.700	40.82	0.000	4.000
2018-542	19	MINI CUBE		8	7.625	1.495	19.60	3.000	12.000
2018-543	3	SEDIGREN		3	19.333	6.600	34.14	0.000	39.000
2018-543	4	SEDIPLAST		560	20.117	5.862	29.14	3.000	38.000
2018-543	10	SEDI RATE		124	24.407	8.529	34.95	0	50.000
2018-543	11	SEDIMAT-15		64	63.969	13.980	21.85	22.000	106.000
2018-543	14	GLOBAL SC.		9	24.778	11.679	47.13	0	60.000
2018-543	17	SEDI SYSTE		7	19.143	4.389	22.93	6.000	32.000
2018-543	19	MINI CUBE		8	30.000	6.000	20.00	12.000	48.000
2018-544	3	SEDIGREN		3	48.667	6.600	13.56	29.000	68.000
2018-544	4	SEDIPLAST		560	37.117	8.911	24.01	10.000	64.000
2018-544	10	SEDI RATE		124	42.097	10.932	25.97	9.000	75.000
2018-544	11	SEDIMAT-15		64	132.397	13.988	10.57	90.000	174.000
2018-544	14	GLOBAL SC.		9	40.778	11.660	28.59	6.000	76.000

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)									
ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
2018-544	17	SEDI SYSTE		7	41.000	9.055	22.09	14.000	68.000				
2018-544	19	MINI CUBE		8	57.750	12.296	21.29	21.000	95.000				
153 SED. RATE (ESR)		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}									
2018-545	3	SEDIGREN		3	3.000	1.414	47.14	0	7.000				
2018-545	4	SEDIPLAST		560	1.531	0.933	60.91	0	4.000				
2018-545	10	SEDI RATE		124	2.016	0.823	40.82	0.000	4.000				
2018-545	11	SEDIMAT-15		64	4.524	2.100	46.41	0	11.000				
2018-545	14	GLOBAL SC.		9	2.111	0.875	41.44	0	5.000				
2018-545	17	SEDI SYSTE		7	1.714	1.030	60.09	0	5.000				
2018-545	19	MINI CUBE		8	8.750	1.561	17.84	4.000	13.000				

  

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 125 DIFF. FOR 5 PAR NEW									
ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2018-521	1	CBC REG	1 NEUTROPHIL	1	COULTER M.	11	54.473	0.965	1.77	40.900	68.100		
2018-521	1	CBC REG	1 NEUTROPHIL	8	CELL-DYN	149	61.429	2.109	3.43	46.100	76.800		
2018-521	1	CBC REG	1 NEUTROPHIL	15	COUL ACT 5	14	58.000	2.533	4.37	43.500	72.500		
2018-521	1	CBC REG	1 NEUTROPHIL	16	EX.22-2280	13	64.169	2.495	3.89	48.100	80.200		
2018-521	1	CBC REG	1 NEUTROPHIL	17	CELL-DYN	4	62.200	0.424	0.68	46.700	77.800		
2018-521	1	CBC REG	1 NEUTROPHIL	18	COLT.LH	17	57.294	1.922	3.35	43.000	71.600		
2018-521	1	CBC REG	1 NEUTROPHIL	20	SYSMEX XT-	25	45.271	0.940	2.08	34.000	56.600		
2018-521	1	CBC REG	1 NEUTROPHIL	21	COUL LH500	17	54.581	1.437	2.63	40.900	68.200		
2018-521	1	CBC REG	1 NEUTROPHIL	25	XS SERIES	133	36.929	0.960	2.60	27.700	46.200		
2018-521	1	CBC REG	1 NEUTROPHIL	29	B.C600,800	40	52.978	0.802	1.51	39.700	66.200		
2018-521	1	CBC REG	1 NEUTROPHIL	32	BC-5390	17	61.029	0.980	1.61	45.800	76.300		
2018-521	1	CBC REG	2 LYMPHOCYTE	1	COULTER M.	11	25.955	1.314	5.06	19.500	32.400		
2018-521	1	CBC REG	2 LYMPHOCYTE	8	CELL-DYN	149	31.857	0.872	2.74	23.900	39.800		
2018-521	1	CBC REG	2 LYMPHOCYTE	15	COUL ACT 5	14	31.750	2.066	6.51	23.800	39.700		

## Participant Summary

### MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-521	1	CBC REG	2 LYMPHOCYTE	16	EX.22-2280	13	31.892	1.741	5.46	23.900	39.900
2018-521	1	CBC REG	2 LYMPHOCYTE	17	CELL-DYN	4	32.050	0.427	1.33	24.000	40.100
2018-521	1	CBC REG	2 LYMPHOCYTE	18	COLT.LH	17	25.147	1.442	5.74	18.900	31.400
2018-521	1	CBC REG	2 LYMPHOCYTE	20	SYSMEX XT-	25	33.204	0.964	2.90	24.900	41.500
2018-521	1	CBC REG	2 LYMPHOCYTE	21	COUL LH500	17	25.824	1.101	4.26	19.400	32.300
2018-521	1	CBC REG	2 LYMPHOCYTE	25	XS SERIES	133	34.804	0.685	1.97	26.100	43.500
2018-521	1	CBC REG	2 LYMPHOCYTE	29	B.C600,800	40	27.890	0.847	3.04	20.900	34.900
2018-521	1	CBC REG	2 LYMPHOCYTE	32	BC-5390	17	28.282	0.740	2.62	21.200	35.400
2018-521	1	CBC REG	3 MONOCYTE	1	COULTER M.	11	8.436	1.147	13.59	6.300	10.500
2018-521	1	CBC REG	3 MONOCYTE	8	CELL-DYN	149	1.498	0.963	64.27	0	3.500
2018-521	1	CBC REG	3 MONOCYTE	8	CELL-DYN	149	1.450	0.150	10.34	0	3.500
2018-521	1	CBC REG	3 MONOCYTE	15	COUL ACT 5	14	2.979	0.857	28.77	1.000	5.000
2018-521	1	CBC REG	3 MONOCYTE	16	EX.22-2280	12	2.575	0.825	32.03	0.600	4.600
2018-521	1	CBC REG	3 MONOCYTE	17	CELL-DYN	4	1.450	0.150	10.34	0	3.500
2018-521	1	CBC REG	3 MONOCYTE	18	COLT.LH	17	6.947	1.959	28.20	4.900	8.900
2018-521	1	CBC REG	3 MONOCYTE	20	SYSMEX XT-	25	12.196	0.709	5.82	9.100	15.200
2018-521	1	CBC REG	3 MONOCYTE	21	COUL LH500	17	8.135	2.434	29.92	6.100	10.200
2018-521	1	CBC REG	3 MONOCYTE	25	XS SERIES	133	14.007	1.000	7.14	10.500	17.500
2018-521	1	CBC REG	3 MONOCYTE	29	B.C600,800	40	8.233	0.594	7.21	6.200	10.300
2018-521	1	CBC REG	3 MONOCYTE	32	BC-5390	17	1.559	0.584	37.47	0	3.600
2018-521	1	CBC REG	4 EOSINOPHIL	1	COULTER M.	11	10.764	0.629	5.84	8.100	13.500
2018-521	1	CBC REG	4 EOSINOPHIL	8	CELL-DYN	150	4.033	1.228	30.45	2.000	6.000
2018-521	1	CBC REG	4 EOSINOPHIL	15	COUL ACT 5	14	7.357	0.481	6.53	5.400	9.400
2018-521	1	CBC REG	4 EOSINOPHIL	16	EX.22-2280	14	0.423	0.321	75.97	0	2.400
2018-521	1	CBC REG	4 EOSINOPHIL	17	CELL-DYN	4	3.350	0.166	4.95	1.400	5.400
2018-521	1	CBC REG	4 EOSINOPHIL	18	COLT.LH	17	10.376	0.594	5.72	7.800	13.000
2018-521	1	CBC REG	4 EOSINOPHIL	20	SYSMEX XT-	25	9.450	0.682	7.22	7.100	11.800
2018-521	1	CBC REG	4 EOSINOPHIL	21	COUL LH500	17	10.759	0.445	4.13	8.100	13.400
2018-521	1	CBC REG	4 EOSINOPHIL	25	XS SERIES	133	8.658	0.645	7.44	6.500	10.800
2018-521	1	CBC REG	4 EOSINOPHIL	29	B.C600,800	40	10.765	0.411	3.82	8.100	13.500
2018-521	1	CBC REG	4 EOSINOPHIL	32	BC-5390	17	9.129	0.421	4.61	6.800	11.400



## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-521	1	CBC REG	5 BASOPHIL	1	COULTER M.	11	0.373	0.263	70.61	0	2.400
2018-521	1	CBC REG	5 BASOPHIL	8	CELL-DYN	147	1.052	0.336	31.92	0	3.100
2018-521	1	CBC REG	5 BASOPHIL	11	CELL-DYN	3	0.700	0.374	53.45	0	2.700
2018-521	1	CBC REG	5 BASOPHIL	15	COUL ACT 5	15	66.967	1.046	1.56	50.200	83.700
2018-521	1	CBC REG	5 BASOPHIL	16	EX.22-2280	13	0.892	0.500	56.03	0	2.900
2018-521	1	CBC REG	5 BASOPHIL	17	CELL-DYN	4	0.950	0.087	9.12	0	3.000
2018-521	1	CBC REG	5 BASOPHIL	18	COLT.LH	17	0.294	0.267	90.69	0	2.300
2018-521	1	CBC REG	5 BASOPHIL	20	SYSMEX XT-	25	60.709	1.177	1.94	45.500	75.900
2018-521	1	CBC REG	5 BASOPHIL	21	COUL LH500	17	0.259	0.175	67.65	0	2.300
2018-521	1	CBC REG	5 BASOPHIL	25	XS SERIES	133	5.565	0.426	7.66	3.600	7.600
2018-521	1	CBC REG	5 BASOPHIL	29	B.C600,800	40	0.115	0.086	74.83	0	2.100
2018-521	1	CBC REG	5 BASOPHIL	32	BC-5390	17	26.641	0.188	0.71	20.000	33.300
2018-521	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	131	58.726	2.439	4.15	44.000	73.400
2018-521	2	ACT PENTRA	1 NEUTROPHIL	19	ACT.2 DIFF	7	58.500	3.412	5.83	43.900	73.100
2018-521	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	20	55.710	2.324	4.17	41.800	69.600
2018-521	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	131	31.065	2.357	7.59	23.300	38.800
2018-521	2	ACT PENTRA	2 LYMPHOCYTE	19	ACT.2 DIFF	7	31.971	3.646	11.41	24.000	40.000
2018-521	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	20	31.175	2.080	6.67	23.400	39.000
2018-521	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	131	2.982	0.672	22.53	1.000	5.000
2018-521	2	ACT PENTRA	3 MONOCYTE	19	ACT.2 DIFF	7	2.800	0.581	20.74	0.800	4.800
2018-521	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	20	3.345	0.781	23.36	1.300	5.300
2018-521	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	132	7.048	0.819	11.63	5.000	9.000
2018-521	2	ACT PENTRA	4 EOSINOPHIL	19	ACT.2 DIFF	6	6.367	0.471	7.40	4.400	8.400
2018-521	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	20	6.615	0.820	12.39	4.600	8.600
2018-521	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	130	66.229	1.450	2.19	49.700	82.800
2018-521	2	ACT PENTRA	5 BASOPHIL	19	ACT.2 DIFF	7	57.357	22.133	38.59	43.000	71.700
2018-521	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	20	3.337	0.058	1.74	1.300	5.300
2018-521	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	20	62.689	1.168	1.86	47.000	78.400
2018-521	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	20	22.905	0.992	4.33	17.200	28.600
2018-521	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	20	8.250	0.439	5.32	6.200	10.300
2018-521	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	20	2.480	0.695	28.04	0.500	4.500

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2018-521	3	ADVIA 120	5	BASOPHIL	9	ADVIA-120	20	1.715	0.550	32.06	0	3.700
2018-521	3	ADVIA 120	6	LUC	9	ADVIA-120	19	3.484	0.360	10.34	1.500	5.500
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2018-522	1	CBC REG	1	NEUTROPHIL	1	COULTER M.	11	57.445	0.581	1.01	43.100	71.800
2018-522	1	CBC REG	1	NEUTROPHIL	8	CELL-DYN	149	62.175	2.522	4.06	46.600	77.700
2018-522	1	CBC REG	1	NEUTROPHIL	15	COUL ACT 5	14	59.593	0.791	1.33	44.700	74.500
2018-522	1	CBC REG	1	NEUTROPHIL	16	EX.22-2280	13	63.092	2.967	4.70	47.300	78.900
2018-522	1	CBC REG	1	NEUTROPHIL	17	CELL-DYN	4	63.175	0.327	0.52	47.400	79.000
2018-522	1	CBC REG	1	NEUTROPHIL	18	COLT.LH	17	59.918	1.122	1.87	44.900	74.900
2018-522	1	CBC REG	1	NEUTROPHIL	20	SYSMEX XT-	25	46.583	1.466	3.15	34.900	58.200
2018-522	1	CBC REG	1	NEUTROPHIL	21	COUL LH500	17	57.375	1.229	2.14	43.000	71.700
2018-522	1	CBC REG	1	NEUTROPHIL	25	XS SERIES	133	37.811	1.174	3.10	28.400	47.300
2018-522	1	CBC REG	1	NEUTROPHIL	29	B.C600,800	40	57.453	0.893	1.55	43.100	71.800
2018-522	1	CBC REG	1	NEUTROPHIL	32	BC-5390	17	64.212	1.203	1.87	48.200	80.300
2018-522	1	CBC REG	2	LYMPHOCYTE	1	COULTER M.	11	28.027	0.833	2.97	21.000	35.000
2018-522	1	CBC REG	2	LYMPHOCYTE	8	CELL-DYN	149	32.639	0.850	2.60	24.500	40.800
2018-522	1	CBC REG	2	LYMPHOCYTE	15	COUL ACT 5	14	34.221	0.893	2.61	25.700	42.800
2018-522	1	CBC REG	2	LYMPHOCYTE	16	EX.22-2280	13	32.985	1.645	4.99	24.700	41.200
2018-522	1	CBC REG	2	LYMPHOCYTE	17	CELL-DYN	4	32.725	0.277	0.85	24.500	40.900
2018-522	1	CBC REG	2	LYMPHOCYTE	18	COLT.LH	17	26.735	1.072	4.01	20.100	33.400
2018-522	1	CBC REG	2	LYMPHOCYTE	20	SYSMEX XT-	24	35.952	1.181	3.28	27.000	44.900
2018-522	1	CBC REG	2	LYMPHOCYTE	21	COUL LH500	17	27.718	1.236	4.46	20.800	34.600
2018-522	1	CBC REG	2	LYMPHOCYTE	25	XS SERIES	133	37.490	0.954	2.54	28.100	46.900
2018-522	1	CBC REG	2	LYMPHOCYTE	29	B.C600,800	40	29.065	0.922	3.17	21.800	36.300
2018-522	1	CBC REG	2	LYMPHOCYTE	32	BC-5390	17	30.235	1.110	3.67	22.700	37.800
2018-522	1	CBC REG	3	MONOCYTE	1	COULTER M.	11	9.164	0.764	8.34	6.900	11.500
2018-522	1	CBC REG	3	MONOCYTE	8	CELL-DYN	150	1.831	1.280	69.91	0	3.800
2018-522	1	CBC REG	3	MONOCYTE	8	CELL-DYN	150	1.250	0.377	30.20	0	3.300
2018-522	1	CBC REG	3	MONOCYTE	15	COUL ACT 5	14	2.693	0.589	21.86	0.700	4.700
2018-522	1	CBC REG	3	MONOCYTE	16	EX.22-2280	13	2.638	1.119	42.43	0.600	4.600
2018-522	1	CBC REG	3	MONOCYTE	17	CELL-DYN	4	1.250	0.377	30.20	0	3.300

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-522	1	CBC REG	3 MONOCYTE	18	COLT.LH	17	8.247	0.982	11.91	6.200	10.300
2018-522	1	CBC REG	3 MONOCYTE	20	SYSMEX XT-	25	7.208	1.403	19.46	5.200	9.200
2018-522	1	CBC REG	3 MONOCYTE	21	COUL LH500	17	9.147	2.284	24.97	6.900	11.400
2018-522	1	CBC REG	3 MONOCYTE	25	XS SERIES	133	9.617	1.125	11.70	7.200	12.000
2018-522	1	CBC REG	3 MONOCYTE	29	B.C600,800	40	8.090	0.695	8.59	6.100	10.100
2018-522	1	CBC REG	3 MONOCYTE	32	BC-5390	17	1.729	0.631	36.50	0	3.700
2018-522	1	CBC REG	4 EOSINOPHIL	1	COULTER M.	11	4.945	0.287	5.81	2.900	6.900
2018-522	1	CBC REG	4 EOSINOPHIL	8	CELL-DYN	149	1.867	0.775	41.53	0	3.900
2018-522	1	CBC REG	4 EOSINOPHIL	15	COUL ACT 5	15	3.436	0.264	7.67	1.400	5.400
2018-522	1	CBC REG	4 EOSINOPHIL	16	EX.22-2280	13	0.454	0.310	68.39	0	2.500
2018-522	1	CBC REG	4 EOSINOPHIL	17	CELL-DYN	4	1.500	0.469	31.27	0	3.500
2018-522	1	CBC REG	4 EOSINOPHIL	18	COLT.LH	17	4.771	0.297	6.22	2.800	6.800
2018-522	1	CBC REG	4 EOSINOPHIL	20	SYSMEX XT-	24	10.317	0.877	8.50	7.700	12.900
2018-522	1	CBC REG	4 EOSINOPHIL	21	COUL LH500	17	4.965	0.303	6.10	3.000	7.000
2018-522	1	CBC REG	4 EOSINOPHIL	25	XS SERIES	133	9.245	0.740	8.00	6.900	11.600
2018-522	1	CBC REG	4 EOSINOPHIL	29	B.C600,800	40	5.028	0.257	5.11	3.000	7.000
2018-522	1	CBC REG	4 EOSINOPHIL	32	BC-5390	17	3.824	0.251	6.56	1.800	5.800
2018-522	1	CBC REG	5 BASOPHIL	1	COULTER M.	11	0.418	0.298	71.24	0	2.400
2018-522	1	CBC REG	5 BASOPHIL	8	CELL-DYN	149	1.299	0.599	46.13	0	3.300
2018-522	1	CBC REG	5 BASOPHIL	15	COUL ACT 5	14	69.079	2.278	3.30	51.800	86.300
2018-522	1	CBC REG	5 BASOPHIL	16	EX.22-2280	13	0.831	0.551	66.33	0	2.800
2018-522	1	CBC REG	5 BASOPHIL	17	CELL-DYN	4	1.275	0.476	37.36	0	3.300
2018-522	1	CBC REG	5 BASOPHIL	18	COLT.LH	17	0.294	0.083	28.15	0	2.300
2018-522	1	CBC REG	5 BASOPHIL	20	SYSMEX XT-	25	63.075	1.089	1.73	47.300	78.800
2018-522	1	CBC REG	5 BASOPHIL	21	COUL LH500	17	0.175	0.120	68.51	0	2.200
2018-522	1	CBC REG	5 BASOPHIL	25	XS SERIES	133	5.800	0.502	8.66	3.800	7.800
2018-522	1	CBC REG	5 BASOPHIL	29	B.C600,800	40	0.365	0.257	70.53	0	2.400
2018-522	1	CBC REG	5 BASOPHIL	32	BC-5390	17	25.676	0.389	1.51	19.300	32.100
2018-522	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	131	59.782	1.647	2.76	44.800	74.700
2018-522	2	ACT PENTRA	1 NEUTROPHIL	19	ACT.2 DIFF	7	59.100	1.343	2.27	44.300	73.900
2018-522	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	20	56.225	1.004	1.79	42.200	70.300

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2018-522	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	131	33.749	1.527	4.52	25.300	42.200			
2018-522	2 ACT PENTRA	2 LYMPHOCYTE	19 ACT.2 DIFF	7	34.100	1.517	4.45	25.600	42.600			
2018-522	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	20	33.310	0.765	2.30	25.000	41.600			
2018-522	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	131	2.891	0.713	24.67	0.900	4.900			
2018-522	2 ACT PENTRA	3 MONOCYTE	19 ACT.2 DIFF	7	2.871	0.707	24.61	0.900	4.900			
2018-522	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	20	3.485	0.604	17.32	1.500	5.500			
2018-522	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	131	3.557	0.366	10.28	1.600	5.600			
2018-522	2 ACT PENTRA	4 EOSINOPHIL	19 ACT.2 DIFF	7	3.400	0.283	8.32	1.400	5.400			
2018-522	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	20	3.495	0.315	9.02	1.500	5.500			
2018-522	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	130	68.314	2.932	4.29	51.200	85.400			
2018-522	2 ACT PENTRA	5 BASOPHIL	19 ACT.2 DIFF	7	59.100	22.849	38.66	44.300	73.900			
2018-522	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	20	3.500	0.210	5.99	1.500	5.500			
2018-522	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	20	45.774	2.035	4.45	34.300	57.200			
2018-522	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	20	34.316	2.982	8.69	25.700	42.900			
2018-522	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	20	13.635	1.002	7.35	10.200	17.000			
2018-522	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	20	1.045	0.468	44.83	0	3.000			
2018-522	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	20	0.900	0.298	33.15	0	2.900			
2018-522	3 ADVIA 120	6 LUC	9 ADVIA-120	19	5.121	0.930	18.16	3.100	7.100			
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2018-523	1 CBC REG	1 NEUTROPHIL	1 COULTER M.	11	67.009	4.280	6.39	50.300	83.800			
2018-523	1 CBC REG	1 NEUTROPHIL	8 CELL-DYN	148	69.097	2.133	3.09	51.800	86.400			
2018-523	1 CBC REG	1 NEUTROPHIL	15 COUL ACT 5	14	72.021	0.626	0.87	54.000	90.000			
2018-523	1 CBC REG	1 NEUTROPHIL	16 EX.22-2280	13	68.854	2.690	3.91	51.600	86.100			
2018-523	1 CBC REG	1 NEUTROPHIL	17 CELL-DYN	4	70.025	1.062	1.52	52.500	87.500			
2018-523	1 CBC REG	1 NEUTROPHIL	18 COLT.LH	17	68.376	3.613	5.28	51.300	85.500			
2018-523	1 CBC REG	1 NEUTROPHIL	20 SYSMEX XT-	25	56.038	2.983	5.32	42.000	70.000			
2018-523	1 CBC REG	1 NEUTROPHIL	21 COUL LH500	17	64.700	3.558	5.50	48.500	80.900			
2018-523	1 CBC REG	1 NEUTROPHIL	25 XS SERIES	132	44.474	1.897	4.26	33.400	55.600			
2018-523	1 CBC REG	1 NEUTROPHIL	29 B.C600,800	40	64.018	1.145	1.79	48.000	80.000			
2018-523	1 CBC REG	1 NEUTROPHIL	32 BC-5390	17	70.831	1.159	1.64	53.100	88.500			
2018-523	1 CBC REG	2 LYMPHOCYTE	1 COULTER M.	11	21.645	1.033	4.77	16.200	27.100			

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-523	1 CBC REG	2 LYMPHOCYTE	8 CELL-DYN	151	27.298	1.442	5.28	20.500	34.100		
2018-523	1 CBC REG	2 LYMPHOCYTE	15 COUL ACT 5	14	19.379	0.635	3.28	14.500	24.200		
2018-523	1 CBC REG	2 LYMPHOCYTE	16 EX.22-2280	13	27.285	1.794	6.58	20.500	34.100		
2018-523	1 CBC REG	2 LYMPHOCYTE	17 CELL-DYN	4	27.925	0.942	3.37	20.900	34.900		
2018-523	1 CBC REG	2 LYMPHOCYTE	18 COLT.LH	17	22.224	1.788	8.04	16.700	27.800		
2018-523	1 CBC REG	2 LYMPHOCYTE	20 SYSMEX XT-	25	30.017	2.987	9.95	22.500	37.500		
2018-523	1 CBC REG	2 LYMPHOCYTE	21 COUL LH500	17	22.541	0.825	3.66	16.900	28.200		
2018-523	1 CBC REG	2 LYMPHOCYTE	25 XS SERIES	133	31.739	1.491	4.70	23.800	39.700		
2018-523	1 CBC REG	2 LYMPHOCYTE	29 B.C600,800	40	24.651	0.742	3.01	18.500	30.800		
2018-523	1 CBC REG	2 LYMPHOCYTE	32 BC-5390	17	25.319	1.054	4.16	19.000	31.600		
2018-523	1 CBC REG	3 MONOCYTE	1 COULTER M.	11	8.909	3.969	44.55	6.700	11.100		
2018-523	1 CBC REG	3 MONOCYTE	8 CELL-DYN	148	1.171	0.466	39.80	0	3.200		
2018-523	1 CBC REG	3 MONOCYTE	8 CELL-DYN	148	0.550	0.328	59.61	0	2.600		
2018-523	1 CBC REG	3 MONOCYTE	15 COUL ACT 5	14	2.800	0.370	13.23	0.800	4.800		
2018-523	1 CBC REG	3 MONOCYTE	16 EX.22-2280	13	2.046	0.934	45.63	0.000	4.000		
2018-523	1 CBC REG	3 MONOCYTE	17 CELL-DYN	4	0.550	0.328	59.61	0	2.600		
2018-523	1 CBC REG	3 MONOCYTE	18 COLT.LH	17	7.571	2.413	31.88	5.600	9.600		
2018-523	1 CBC REG	3 MONOCYTE	20 SYSMEX XT-	25	2.004	1.335	66.59	0.000	4.000		
2018-523	1 CBC REG	3 MONOCYTE	21 COUL LH500	17	9.612	3.312	34.46	7.200	12.000		
2018-523	1 CBC REG	3 MONOCYTE	25 XS SERIES	133	6.201	1.860	29.99	4.200	8.200		
2018-523	1 CBC REG	3 MONOCYTE	29 B.C600,800	40	8.420	0.971	11.53	6.300	10.500		
2018-523	1 CBC REG	3 MONOCYTE	32 BC-5390	17	2.524	0.831	32.95	0.500	4.500		
2018-523	1 CBC REG	4 EOSINOPHIL	1 COULTER M.	11	2.164	0.629	29.05	0.200	4.200		
2018-523	1 CBC REG	4 EOSINOPHIL	8 CELL-DYN	149	1.001	0.816	81.56	0	3.000		
2018-523	1 CBC REG	4 EOSINOPHIL	15 COUL ACT 5	14	5.793	0.435	7.51	3.800	7.800		
2018-523	1 CBC REG	4 EOSINOPHIL	16 EX.22-2280	13	1.008	0.569	56.47	0	3.000		
2018-523	1 CBC REG	4 EOSINOPHIL	17 CELL-DYN	4	0.400	0.520	129.90	0	2.400		
2018-523	1 CBC REG	4 EOSINOPHIL	18 COLT.LH	17	1.994	0.353	17.69	0.000	4.000		
2018-523	1 CBC REG	4 EOSINOPHIL	20 SYSMEX XT-	25	11.942	0.921	7.72	9.000	14.900		
2018-523	1 CBC REG	4 EOSINOPHIL	21 COUL LH500	17	2.935	0.557	18.98	0.900	4.900		
2018-523	1 CBC REG	4 EOSINOPHIL	25 XS SERIES	133	10.853	0.898	8.28	8.100	13.600		

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-523	1	CBC REG	4 EOSINOPHIL	29	B.C600,800	40	2.708	0.345	12.75	0.700	4.700
2018-523	1	CBC REG	4 EOSINOPHIL	32	BC-5390	17	1.306	0.239	18.29	0	3.300
2018-523	1	CBC REG	5 BASOPHIL	1	COULTER M.	11	0.273	0.186	68.31	0	2.300
2018-523	1	CBC REG	5 BASOPHIL	8	CELL-DYN	149	1.349	0.518	38.40	0	3.300
2018-523	1	CBC REG	5 BASOPHIL	15	COUL ACT 5	14	84.729	1.937	2.29	63.500	105.900
2018-523	1	CBC REG	5 BASOPHIL	16	EX.22-2280	13	0.642	0.340	53.02	0	2.600
2018-523	1	CBC REG	5 BASOPHIL	17	CELL-DYN	4	1.175	1.176	100.05	0	3.200
2018-523	1	CBC REG	5 BASOPHIL	18	COLT.LH	17	0.165	0.153	92.99	0	2.200
2018-523	1	CBC REG	5 BASOPHIL	20	SYSMEX XT-	25	70.454	0.761	1.08	52.800	88.100
2018-523	1	CBC REG	5 BASOPHIL	21	COUL LH500	17	0.212	0.132	62.48	0	2.200
2018-523	1	CBC REG	5 BASOPHIL	25	XS SERIES	133	6.705	0.628	9.36	4.700	8.700
2018-523	1	CBC REG	5 BASOPHIL	29	B.C600,800	40	0.440	0.214	48.69	0	2.400
2018-523	1	CBC REG	5 BASOPHIL	32	BC-5390	17	25.482	0.391	1.54	19.100	31.900
2018-523	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	131	72.141	1.579	2.19	54.100	90.200
2018-523	2	ACT PENTRA	1 NEUTROPHIL	19	ACT.2 DIFF	7	71.843	2.141	2.98	53.900	89.800
2018-523	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	20	67.970	1.270	1.87	51.000	85.000
2018-523	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	131	19.037	1.145	6.02	14.300	23.800
2018-523	2	ACT PENTRA	2 LYMPHOCYTE	19	ACT.2 DIFF	7	19.214	1.521	7.91	14.400	24.000
2018-523	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	20	18.325	0.875	4.77	13.700	22.900
2018-523	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	131	3.060	0.704	23.00	1.100	5.100
2018-523	2	ACT PENTRA	3 MONOCYTE	19	ACT.2 DIFF	7	2.714	0.818	30.15	0.700	4.700
2018-523	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	20	3.725	0.433	11.64	1.700	5.700
2018-523	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	131	5.825	0.419	7.20	3.800	7.800
2018-523	2	ACT PENTRA	4 EOSINOPHIL	19	ACT.2 DIFF	7	5.714	0.203	3.55	3.700	7.700
2018-523	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	20	5.775	0.308	5.33	3.800	7.800
2018-523	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	131	84.078	1.971	2.34	63.100	105.100
2018-523	2	ACT PENTRA	5 BASOPHIL	19	ACT.2 DIFF	7	73.057	28.205	38.61	54.800	91.300
2018-523	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	20	4.240	0.128	3.02	2.200	6.200
2018-523	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	20	73.042	1.123	1.54	54.800	91.300
2018-523	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	20	14.526	1.548	10.66	10.900	18.200
2018-523	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	20	8.035	0.408	5.07	6.000	10.000

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2018-523	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	20	2.240	0.442	19.73	0.200	4.200	
2018-523	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	20	1.365	0.419	30.67	0	3.400	
2018-523	3	ADVIA 120	6 LUC	9	ADVIA-120	19	2.232	0.241	10.79	0.200	4.200	
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2018-524	1	CBC REG	1 NEUTROPHIL	1	COULTER M.	11	64.436	1.069	1.66	48.300	80.500	
2018-524	1	CBC REG	1 NEUTROPHIL	8	CELL-DYN	149	67.169	1.846	2.75	50.400	84.000	
2018-524	1	CBC REG	1 NEUTROPHIL	15	COUL ACT 5	14	60.021	0.832	1.39	45.000	75.000	
2018-524	1	CBC REG	1 NEUTROPHIL	16	EX.22-2280	13	66.562	1.876	2.82	49.900	83.200	
2018-524	1	CBC REG	1 NEUTROPHIL	17	CELL-DYN	4	68.000	0.212	0.31	51.000	85.000	
2018-524	1	CBC REG	1 NEUTROPHIL	18	COLT.LH	17	67.329	2.536	3.77	50.500	84.200	
2018-524	1	CBC REG	1 NEUTROPHIL	20	SYSMEX XT-	25	52.265	1.576	3.02	39.200	65.300	
2018-524	1	CBC REG	1 NEUTROPHIL	21	COUL LH500	17	65.794	2.543	3.87	49.300	82.200	
2018-524	1	CBC REG	1 NEUTROPHIL	25	XS SERIES	133	43.046	1.171	2.72	32.300	53.800	
2018-524	1	CBC REG	1 NEUTROPHIL	29	B.C600,800	40	64.485	1.034	1.60	48.400	80.600	
2018-524	1	CBC REG	1 NEUTROPHIL	32	BC-5390	17	70.788	1.964	2.77	53.100	88.500	
2018-524	1	CBC REG	2 LYMPHOCYTE	1	COULTER M.	11	25.618	1.114	4.35	19.200	32.000	
2018-524	1	CBC REG	2 LYMPHOCYTE	8	CELL-DYN	149	29.447	0.873	2.96	22.100	36.800	
2018-524	1	CBC REG	2 LYMPHOCYTE	15	COUL ACT 5	15	34.067	0.784	2.30	25.600	42.600	
2018-524	1	CBC REG	2 LYMPHOCYTE	16	EX.22-2280	13	29.677	0.962	3.24	22.300	37.100	
2018-524	1	CBC REG	2 LYMPHOCYTE	17	CELL-DYN	4	29.200	0.424	1.45	21.900	36.500	
2018-524	1	CBC REG	2 LYMPHOCYTE	18	COLT.LH	17	25.819	0.990	3.83	19.400	32.300	
2018-524	1	CBC REG	2 LYMPHOCYTE	20	SYSMEX XT-	24	32.735	1.328	4.06	24.600	40.900	
2018-524	1	CBC REG	2 LYMPHOCYTE	21	COUL LH500	17	25.124	1.390	5.53	18.800	31.400	
2018-524	1	CBC REG	2 LYMPHOCYTE	25	XS SERIES	133	33.761	0.941	2.79	25.300	42.200	
2018-524	1	CBC REG	2 LYMPHOCYTE	29	B.C600,800	40	26.674	0.589	2.21	20.000	33.300	
2018-524	1	CBC REG	2 LYMPHOCYTE	32	BC-5390	17	26.775	1.216	4.54	20.100	33.500	
2018-524	1	CBC REG	3 MONOCYTE	1	COULTER M.	11	7.755	1.178	15.19	5.800	9.800	
2018-524	1	CBC REG	3 MONOCYTE	8	CELL-DYN	150	1.050	0.409	38.98	0	3.100	
2018-524	1	CBC REG	3 MONOCYTE	8	CELL-DYN	150	1.672	0.983	58.79	0	3.700	
2018-524	1	CBC REG	3 MONOCYTE	15	COUL ACT 5	16	2.625	0.488	18.59	0.600	4.600	
2018-524	1	CBC REG	3 MONOCYTE	16	EX.22-2280	13	2.538	0.674	26.56	0.500	4.500	

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-524	1	CBC REG	3 MONOCYTE	17	CELL-DYN	4	1.050	0.409	38.98	0	3.100
2018-524	1	CBC REG	3 MONOCYTE	18	COLT.LH	17	5.394	2.324	43.08	3.400	7.400
2018-524	1	CBC REG	3 MONOCYTE	20	SYSMEX XT-	25	3.775	1.298	34.37	1.800	5.800
2018-524	1	CBC REG	3 MONOCYTE	21	COUL LH500	17	7.200	2.250	31.26	5.200	9.200
2018-524	1	CBC REG	3 MONOCYTE	25	XS SERIES	133	6.360	1.000	15.72	4.400	8.400
2018-524	1	CBC REG	3 MONOCYTE	29	B.C600,800	40	6.580	0.912	13.86	4.600	8.600
2018-524	1	CBC REG	3 MONOCYTE	32	BC-5390	17	1.894	0.847	44.71	0	3.900
2018-524	1	CBC REG	4 EOSINOPHIL	1	COULTER M.	11	1.855	0.281	15.14	0	3.900
2018-524	1	CBC REG	4 EOSINOPHIL	8	CELL-DYN	147	0.581	0.573	98.68	0	2.600
2018-524	1	CBC REG	4 EOSINOPHIL	15	COUL ACT 5	15	3.460	0.298	8.63	1.500	5.500
2018-524	1	CBC REG	4 EOSINOPHIL	16	EX.22-2280	13	0.500	0.402	80.38	0	2.500
2018-524	1	CBC REG	4 EOSINOPHIL	17	CELL-DYN	4	0.775	0.492	63.46	0	2.800
2018-524	1	CBC REG	4 EOSINOPHIL	18	COLT.LH	17	1.594	0.213	13.35	0	3.600
2018-524	1	CBC REG	4 EOSINOPHIL	20	SYSMEX XT-	25	11.325	0.720	6.36	8.500	14.200
2018-524	1	CBC REG	4 EOSINOPHIL	21	COUL LH500	16	1.706	0.351	20.56	0	3.700
2018-524	1	CBC REG	4 EOSINOPHIL	25	XS SERIES	133	10.363	0.866	8.35	7.800	13.000
2018-524	1	CBC REG	4 EOSINOPHIL	29	B.C600,800	40	1.780	0.182	10.22	0	3.800
2018-524	1	CBC REG	4 EOSINOPHIL	32	BC-5390	17	0.906	0.163	17.95	0	2.900
2018-524	1	CBC REG	5 BASOPHIL	1	COULTER M.	11	0.336	0.192	57.08	0	2.300
2018-524	1	CBC REG	5 BASOPHIL	8	CELL-DYN	148	1.078	0.518	48.02	0	3.100
2018-524	1	CBC REG	5 BASOPHIL	15	COUL ACT 5	14	68.264	2.000	2.93	51.200	85.300
2018-524	1	CBC REG	5 BASOPHIL	16	EX.22-2280	13	0.723	0.496	68.65	0	2.700
2018-524	1	CBC REG	5 BASOPHIL	17	CELL-DYN	4	0.950	0.862	90.70	0	3.000
2018-524	1	CBC REG	5 BASOPHIL	18	COLT.LH	17	0.218	0.201	92.21	0	2.200
2018-524	1	CBC REG	5 BASOPHIL	20	SYSMEX XT-	24	68.017	0.782	1.15	51.000	85.000
2018-524	1	CBC REG	5 BASOPHIL	21	COUL LH500	17	0.194	0.100	51.43	0	2.200
2018-524	1	CBC REG	5 BASOPHIL	25	XS SERIES	133	6.451	0.544	8.44	4.500	8.500
2018-524	1	CBC REG	5 BASOPHIL	29	B.C600,800	40	0.538	0.228	42.36	0	2.500
2018-524	1	CBC REG	5 BASOPHIL	32	BC-5390	17	25.982	0.378	1.45	19.500	32.500
2018-524	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	131	59.896	1.761	2.94	44.900	74.900
2018-524	2	ACT PENTRA	1 NEUTROPHIL	19	ACT.2 DIFF	7	59.557	1.450	2.43	44.700	74.400



**Participant Summary**  
**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2018-524	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	20	56.880	0.905	1.59	42.700	71.100			
2018-524	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	130	33.688	1.463	4.34	25.300	42.100			
2018-524	2 ACT PENTRA	2 LYMPHOCYTE	19 ACT.2 DIFF	7	33.914	1.484	4.37	25.400	42.400			
2018-524	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	20	32.760	0.593	1.81	24.600	41.000			
2018-524	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	129	2.910	0.706	24.26	0.900	4.900			
2018-524	2 ACT PENTRA	3 MONOCYTE	19 ACT.2 DIFF	7	2.729	0.694	25.45	0.700	4.700			
2018-524	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	20	3.535	0.488	13.81	1.500	5.500			
2018-524	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	129	3.576	0.364	10.17	1.600	5.600			
2018-524	2 ACT PENTRA	4 EOSINOPHIL	19 ACT.2 DIFF	7	3.343	0.232	6.94	1.300	5.300			
2018-524	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	20	3.355	0.319	9.49	1.400	5.400			
2018-524	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	130	67.771	2.346	3.46	50.800	84.700			
2018-524	2 ACT PENTRA	5 BASOPHIL	19 ACT.2 DIFF	7	58.914	22.772	38.65	44.200	73.600			
2018-524	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	20	3.485	0.174	4.99	1.500	5.500			
2018-524	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	20	62.458	0.898	1.44	46.800	78.100			
2018-524	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	20	23.053	0.880	3.82	17.300	28.800			
2018-524	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	19	8.342	0.451	5.41	6.300	10.400			
2018-524	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	20	2.450	0.625	25.51	0.500	4.500			
2018-524	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	20	1.680	0.493	29.32	0	3.700			
2018-524	3 ADVIA 120	6 LUC	9 ADVIA-120	19	3.505	0.456	13.01	1.500	5.500			
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2018-525	1 CBC REG	1 NEUTROPHIL	1 COULTER M.	11	61.964	4.262	6.88	46.500	77.500			
2018-525	1 CBC REG	1 NEUTROPHIL	8 CELL-DYN	149	66.719	1.680	2.52	50.000	83.400			
2018-525	1 CBC REG	1 NEUTROPHIL	15 COUL ACT 5	14	58.850	1.363	2.32	44.100	73.600			
2018-525	1 CBC REG	1 NEUTROPHIL	16 EX.22-2280	13	70.177	4.438	6.32	52.600	87.700			
2018-525	1 CBC REG	1 NEUTROPHIL	17 CELL-DYN	4	66.900	0.490	0.73	50.200	83.600			
2018-525	1 CBC REG	1 NEUTROPHIL	18 COLT.LH	17	63.065	2.462	3.90	47.300	78.800			
2018-525	1 CBC REG	1 NEUTROPHIL	20 SYSMEX XT-	25	54.079	1.437	2.66	40.600	67.600			
2018-525	1 CBC REG	1 NEUTROPHIL	21 COUL LH500	17	60.741	3.651	6.01	45.600	75.900			
2018-525	1 CBC REG	1 NEUTROPHIL	25 XS SERIES	133	42.578	1.778	4.18	31.900	53.200			
2018-525	1 CBC REG	1 NEUTROPHIL	29 B.C600,800	40	57.308	1.209	2.11	43.000	71.600			
2018-525	1 CBC REG	1 NEUTROPHIL	32 BC-5390	17	66.318	0.943	1.42	49.700	82.900			

## Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-525	1	CBC REG	2 LYMPHOCYTE	1	COULTER M.	11	20.136	1.576	7.83	15.100	25.200
2018-525	1	CBC REG	2 LYMPHOCYTE	8	CELL-DYN	149	26.958	0.804	2.98	20.200	33.700
2018-525	1	CBC REG	2 LYMPHOCYTE	15	COUL ACT 5	14	31.286	1.222	3.91	23.500	39.100
2018-525	1	CBC REG	2 LYMPHOCYTE	16	EX.22-2280	13	26.531	1.690	6.37	19.900	33.200
2018-525	1	CBC REG	2 LYMPHOCYTE	17	CELL-DYN	4	27.475	0.421	1.53	20.600	34.300
2018-525	1	CBC REG	2 LYMPHOCYTE	18	COLT.LH	17	19.571	1.215	6.21	14.700	24.500
2018-525	1	CBC REG	2 LYMPHOCYTE	20	SYSMEX XT-	25	27.329	1.592	5.82	20.500	34.200
2018-525	1	CBC REG	2 LYMPHOCYTE	21	COUL LH500	17	20.306	0.950	4.68	15.200	25.400
2018-525	1	CBC REG	2 LYMPHOCYTE	25	XS SERIES	133	30.366	1.140	3.75	22.800	38.000
2018-525	1	CBC REG	2 LYMPHOCYTE	29	B.C600,800	40	22.956	0.796	3.47	17.200	28.700
2018-525	1	CBC REG	2 LYMPHOCYTE	32	BC-5390	17	23.265	0.686	2.95	17.400	29.100
2018-525	1	CBC REG	3 MONOCYTE	1	COULTER M.	11	6.400	3.505	54.76	4.400	8.400
2018-525	1	CBC REG	3 MONOCYTE	8	CELL-DYN	149	0.979	0.454	46.43	0	3.000
2018-525	1	CBC REG	3 MONOCYTE	8	CELL-DYN	149	1.025	0.083	8.09	0	3.000
2018-525	1	CBC REG	3 MONOCYTE	15	COUL ACT 5	14	2.786	0.658	23.61	0.800	4.800
2018-525	1	CBC REG	3 MONOCYTE	16	EX.22-2280	13	1.654	1.328	80.27	0	3.700
2018-525	1	CBC REG	3 MONOCYTE	17	CELL-DYN	4	1.025	0.083	8.09	0	3.000
2018-525	1	CBC REG	3 MONOCYTE	18	COLT.LH	17	8.035	1.688	21.00	6.000	10.000
2018-525	1	CBC REG	3 MONOCYTE	20	SYSMEX XT-	26	7.325	0.849	11.59	5.300	9.300
2018-525	1	CBC REG	3 MONOCYTE	21	COUL LH500	16	8.669	3.443	39.71	6.500	10.800
2018-525	1	CBC REG	3 MONOCYTE	25	XS SERIES	134	10.038	1.562	15.56	7.500	12.500
2018-525	1	CBC REG	3 MONOCYTE	29	B.C600,800	39	8.805	0.419	4.76	6.600	11.000
2018-525	1	CBC REG	3 MONOCYTE	32	BC-5390	17	1.494	0.495	33.15	0	3.500
2018-525	1	CBC REG	4 EOSINOPHIL	1	COULTER M.	10	11.170	0.261	2.34	8.400	14.000
2018-525	1	CBC REG	4 EOSINOPHIL	8	CELL-DYN	149	4.151	0.981	23.64	2.200	6.200
2018-525	1	CBC REG	4 EOSINOPHIL	15	COUL ACT 5	14	7.071	0.544	7.70	5.100	9.100
2018-525	1	CBC REG	4 EOSINOPHIL	16	EX.22-2280	13	0.731	0.446	61.05	0	2.700
2018-525	1	CBC REG	4 EOSINOPHIL	17	CELL-DYN	4	3.775	0.536	14.19	1.800	5.800
2018-525	1	CBC REG	4 EOSINOPHIL	18	COLT.LH	17	9.159	1.225	13.37	6.900	11.400
2018-525	1	CBC REG	4 EOSINOPHIL	20	SYSMEX XT-	26	11.264	0.877	7.79	8.400	14.100
2018-525	1	CBC REG	4 EOSINOPHIL	21	COUL LH500	17	10.482	0.666	6.36	7.900	13.100

### Participant Summary

MONTH 7 OF YEAR 2018

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2018-525	1	CBC REG	4 EOSINOPHIL	25 XS SERIES	132	10.189	0.900	8.83	7.600	12.700	
2018-525	1	CBC REG	4 EOSINOPHIL	29 B.C600,800	40	10.865	0.784	7.22	8.100	13.600	
2018-525	1	CBC REG	4 EOSINOPHIL	32 BC-5390	17	8.924	0.310	3.47	6.700	11.200	
2018-525	1	CBC REG	5 BASOPHIL	1 COULTER M.	10	0.310	0.336	108.39	0	2.300	
2018-525	1	CBC REG	5 BASOPHIL	8 CELL-DYN	149	1.065	0.409	38.38	0	3.100	
2018-525	1	CBC REG	5 BASOPHIL	15 COUL ACT 5	15	66.820	1.130	1.69	50.100	83.500	
2018-525	1	CBC REG	5 BASOPHIL	16 EX.22-2280	13	0.433	0.266	61.30	0	2.400	
2018-525	1	CBC REG	5 BASOPHIL	17 CELL-DYN	4	0.800	0.173	21.65	0	2.800	
2018-525	1	CBC REG	5 BASOPHIL	18 COLT.LH	17	0.188	0.178	94.48	0	2.200	
2018-525	1	CBC REG	5 BASOPHIL	20 SYSMEX XT-	25	67.138	1.114	1.66	50.400	83.900	
2018-525	1	CBC REG	5 BASOPHIL	21 COUL LH500	17	0.182	0.189	103.43	0	2.200	
2018-525	1	CBC REG	5 BASOPHIL	25 XS SERIES	133	6.846	0.658	9.61	4.800	8.800	
2018-525	1	CBC REG	5 BASOPHIL	29 B.C600,800	40	0.110	0.067	60.91	0	2.100	
2018-525	1	CBC REG	5 BASOPHIL	32 BC-5390	17	27.412	0.211	0.77	20.600	34.300	
2018-525	2	ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	132	59.134	2.632	4.45	44.400	73.900	
2018-525	2	ACT PENTRA	1 NEUTROPHIL	19 ACT.2 DIFF	7	58.843	2.843	4.83	44.100	73.600	
2018-525	2	ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	20	55.970	2.173	3.88	42.000	70.000	
2018-525	2	ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	130	30.740	2.235	7.27	23.100	38.400	
2018-525	2	ACT PENTRA	2 LYMPHOCYTE	19 ACT.2 DIFF	7	31.971	3.103	9.71	24.000	40.000	
2018-525	2	ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	20	30.755	2.000	6.50	23.100	38.400	
2018-525	2	ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	131	2.879	0.704	24.45	0.900	4.900	
2018-525	2	ACT PENTRA	3 MONOCYTE	19 ACT.2 DIFF	7	2.571	0.645	25.09	0.600	4.600	
2018-525	2	ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	20	3.068	0.454	14.81	1.100	5.100	
2018-525	2	ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	131	7.088	0.799	11.28	5.100	9.100	
2018-525	2	ACT PENTRA	4 EOSINOPHIL	19 ACT.2 DIFF	7	6.257	0.232	3.71	4.300	8.300	
2018-525	2	ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	20	6.870	1.147	16.70	4.900	8.900	
2018-525	2	ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	130	66.219	1.350	2.04	49.700	82.800	
2018-525	2	ACT PENTRA	5 BASOPHIL	19 ACT.2 DIFF	7	57.514	22.194	38.59	43.100	71.900	
2018-525	2	ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	20	3.332	0.057	1.70	1.300	5.300	
2018-525	3	ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	20	73.285	0.735	1.00	55.000	91.600	
2018-525	3	ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	20	14.200	0.560	3.94	10.700	17.800	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
2018-525	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	20	8.000	0.322	4.03	6.000	10.000				
2018-525	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	20	2.170	0.564	25.99	0.200	4.200				
2018-525	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	20	1.230	0.427	34.69	0	3.200				
2018-525	3 ADVIA 120	6 LUC	9 ADVIA-120	19	2.163	0.454	21.01	0.200	4.200				

SPEC: 20 HEMATOLOGY      SUB: 75 BODY FLUID      GRP: 145 BODY FLUID CELL COU.

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}							
2018-561				50	101.694	8.413	8.27	76.000	127.000				
2018-562				50	10.400	1.697	16.32	8.000	13.000				
2018-563				50	102.122	6.980	6.83	81.000	123.000				
2018-564				50	10.408	1.384	13.30	8.000	13.000				
2018-565				50	9.653	1.636	16.95	7.000	12.000				
160	BF WHITE CELL COUNT												
2018-561				50	98.480	10.830	11.00	74.000	123.000				
2018-562				50	9.755	1.286	13.18	7.000	12.000				
2018-563				50	97.776	10.947	11.20	73.000	122.000				
2018-564				50	9.633	1.494	15.51	7.000	12.000				
2018-565				50	9.469	1.527	16.12	7.000	12.000				

**Participant Summary**  
**MONTH 7 OF YEAR 2018**

SPEC: 20 HEMATOLOGY		SUB: 75 BODY FLUID		GRP: 146 BF DIFF. CELL COUNT					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
161 BF DIFF. CELL COUNT		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2018-561	1	MONONUCL.		51	24.314	4.336	17.83	18.000	30.000
2018-561	2	POLYMORPH.		51	75.686	4.336	5.73	63.000	89.000
2018-562	1	MONONUCL.		51	25.100	3.612	14.39	19.000	31.000
2018-562	2	POLYMORPH.		51	75.235	3.959	5.26	63.000	87.000
2018-563	1	MONONUCL.		51	24.500	3.971	16.21	18.000	31.000
2018-563	2	POLYMORPH.		51	75.500	3.971	5.26	64.000	87.000
2018-564	1	MONONUCL.		51	24.571	3.057	12.44	18.000	31.000
2018-564	2	POLYMORPH.		51	75.429	3.057	4.05	66.000	85.000
2018-565	1	MONONUCL.		51	24.940	3.635	14.58	19.000	31.000
2018-565	2	POLYMORPH.		51	75.060	3.635	4.84	64.000	86.000

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
083 ABO GROUP													
2018-571				004 004	004 004		004 - B	100	99	99	100		
2018-571				004 004	004 004		005 - AB	100	1	99	0		
2018-572				004 004	004 004		004 - B	100	100	100	100		
2018-573				008 008	008 008		008 - O	100	100	100	100		
2018-574				001 003	001 003		002 - A1	100	3	100	100		
2018-574				001 003	001 003		001 - A	100	97	100	100		
2018-575				001 003	001 003		001 - A	100	97	100	100		
2018-575				001 003	001 003		002 - A1	100	3	100	100		
126 RH TYPE													
2018-571				001 003	001 003		001 - Rh+	100	100	100	100		
2018-572				001 003	001 003		001 - Rh+	100	100	100	100		
2018-573				001 003	001 003		001 - Rh+	100	100	100	100		
2018-574				003 004	003 004		003 - Rh-Du np	100	13	100	100		
2018-574				003 004	003 004		004 - Rh-	100	87	100	100		
2018-575				001 003	001 003		001 - Rh+	100	100	100	100		

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2018-581				002 002	002 002		002 - POSITIVE	69	68	68	100
2018-581				002 002	002 002		001 - NEGATIVE	69	1	68	0
2018-582				001 001	001 001		001 - NEGATIVE	69	69	69	100
2018-583				001 001	001 001		001 - NEGATIVE	69	69	69	100
2018-584				001 001	001 001		001 - NEGATIVE	69	69	69	100
2018-585				002 002	002 002		002 - POSITIVE	69	69	69	100

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2018-291	001 Gel	001 Poly AHG		002 004	002 004		003 - 3+	10	4	10	100
2018-291	001 Gel	001 Poly AHG		002 004	002 004		004 - 4+	10	6	10	100
2018-291	001 Gel	002 MnoAnt-IgG		002 004	002 004		003 - 3+	9	3	9	100
2018-291	001 Gel	002 MnoAnt-IgG		002 004	002 004		002 - 2+	9	1	9	100
2018-291	001 Gel	002 MnoAnt-IgG		002 004	002 004		004 - 4+	9	5	9	100
2018-291	001 Gel	003 Anti-C3b/d			002 004		001 - Negativo	1	1	0	0
2018-291	002 Tubo	001 Poly AHG		002 004	002 004		003 - 3+	28	8	28	100
2018-291	002 Tubo	001 Poly AHG		002 004		002 004	002 - 2+	28	19	28	100
2018-291	002 Tubo	001 Poly AHG			002 004		004 - 4+	28	1	28	100
2018-291	002 Tubo	002 MnoAnt-IgG		002 004			002 - 2+	13	9	13	100
2018-291	002 Tubo	002 MnoAnt-IgG		002 004	002 004		003 - 3+	13	4	13	100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-291	002 Tubo	003 Anti-C3b/d			002 004		001 - Negativo	4	4	0	0	
179 DIRECT ANTIG(COOMBS)												
2018-292	001 Gel	001 Poly AHG		001 001	001 001		001 - Negativo	10	10	10	100	
2018-292	001 Gel	002 MnoAnt-IgG			001 001		001 - Negativo	7	7	7	100	
2018-292	002 Tubo	001 Poly AHG		001 001	001 001		001 - Negativo	29	29	29	100	
2018-292	002 Tubo	002 MnoAnt-IgG		001 001	001 001		001 - Negativo	10	10	10	100	
2018-292	002 Tubo	003 Anti-C3b/d			001 001		001 - Negativo	2	2	2	100	
2018-293	001 Gel	001 Poly AHG			002 004		003 - 3+	10	5	10	100	
2018-293	001 Gel	001 Poly AHG		002 004			004 - 4+	10	5	10	100	
2018-293	001 Gel	002 MnoAnt-IgG		002 004	002 004		004 - 4+	9	4	9	100	
2018-293	001 Gel	002 MnoAnt-IgG		002 004	002 004		003 - 3+	9	4	9	100	
2018-293	001 Gel	002 MnoAnt-IgG		002 004	002 004		002 - 2+	9	1	9	100	
2018-293	001 Gel	003 Anti-C3b/d			002 004		001 - Negativo	1	1	0	0	
2018-293	002 Tubo	001 Poly AHG		002 004			002 - 2+	28	18	28	100	
2018-293	002 Tubo	001 Poly AHG		002 004	002 004		004 - 4+	28	1	28	100	
2018-293	002 Tubo	001 Poly AHG			002 004		003 - 3+	28	9	28	100	
2018-293	002 Tubo	002 MnoAnt-IgG		002 004			002 - 2+	13	10	13	100	
2018-293	002 Tubo	002 MnoAnt-IgG			002 004		003 - 3+	13	3	13	100	
2018-293	002 Tubo	003 Anti-C3b/d			002 004		001 - Negativo	4	3	1	0	
2018-293	002 Tubo	003 Anti-C3b/d			002 004		002 - 2+	4	1	1	100	
2018-294	001 Gel	001 Poly AHG			002 004		003 - 3+	10	5	10	100	
2018-294	001 Gel	001 Poly AHG		002 004			004 - 4+	10	5	10	100	
2018-294	001 Gel	002 MnoAnt-IgG		002 004			004 - 4+	9	5	9	100	
2018-294	001 Gel	002 MnoAnt-IgG			002 004		003 - 3+	9	4	9	100	
2018-294	001 Gel	003 Anti-C3b/d			002 004		001 - Negativo	1	1	0	0	
2018-294	002 Tubo	001 Poly AHG			002 004		003 - 3+	28	13	28	100	
2018-294	002 Tubo	001 Poly AHG		002 004	002 004		004 - 4+	28	1	28	100	



**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-294	002 Tubo	001 Poly AHG		002 004			002 - 2+	28	14	28	100	
2018-294	002 Tubo	002 MnoAnt-IgG			002 004		003 - 3+	13	4	13	100	
2018-294	002 Tubo	002 MnoAnt-IgG		002 004			002 - 2+	13	9	13	100	
2018-294	002 Tubo	003 Anti-C3b/d			002 004		001 - Negativo	4	3	1	0	
2018-294	002 Tubo	003 Anti-C3b/d			002 004		002 - 2+	4	1	1	100	
179 DIRECT ANTIG(COOMBS)												
2018-295	001 Gel	001 Poly AHG		001 001	001 001		001 - Negativo	10	10	10	100	
2018-295	001 Gel	002 MnoAnt-IgG			001 001		001 - Negativo	7	7	7	100	
2018-295	002 Tubo	001 Poly AHG		001 001	001 001		002 - 2+	29	1	28	0	
2018-295	002 Tubo	001 Poly AHG		001 001	001 001		001 - Negativo	29	28	28	100	
2018-295	002 Tubo	002 MnoAnt-IgG		001 001	001 001		001 - Negativo	10	10	10	100	
2018-295	002 Tubo	003 Anti-C3b/d			001 001		001 - Negativo	2	2	2	100	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY												SUB: 100 COMPATIBILITY TESTING												GRP: 170 COMPATIBILITY TEST.											
TEST	LEVEL 1		LEVEL 2		LEVEL 3		PART		REFEREE		COMMER		REPORTED		PART		# LABS		# "SF"		Score														
SAMPLE	RESULT CODE		RESULT CODE		RESULT CODE		LOW/HIGH		LOW/HIGH		LOW/HIGH		VALUE		LABS		ANSW		LABS																
085 COMPATIBIL. TESTING																																			
2018-591							001	001	001	001					001	-	COMPATIBLE	41	41	41	100														
2018-592							001	001	001	001					001	-	COMPATIBLE	41	41	41	100														
2018-593							001	001	001	001					001	-	COMPATIBLE	41	41	41	100														
2018-594							002	002	002	002					002	-	INCOMPATIB	41	41	41	100														
2018-595							001	001	001	001					001	-	COMPATIBLE	41	41	41	100														

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 110 ANTIBODY IDENTIFICATION				GRP: 180 ANTIBODY IDENT.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS LABS	# LABS ANSW	# "SF" LABS	Score
086 ANTIBODY IDENT.TEST.											
2018-591				007 007	007 007		007 - Anti E	11	11	11	100
2018-592				001 001	001 001		001 - UN. ANT.ND	11	11	11	100
2018-593				001 001	001 001		001 - UN. ANT.ND	11	11	11	100
2018-594				006 006	006 006		006 - Anti c	11	11	11	100
2018-595				001 001	001 001		001 - UN. ANT.ND	11	11	11	100
SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC, HIV			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS LABS	# LABS ANSW	# "SF" LABS	Score
093 ANTI-HUMAN IMM. (HIV)											
2018-601	002	ABBOTT			001 001		001 - NEG/NR	2	2	2	100
2018-601	006	BIOTECH			001 001		001 - NEG/NR	1	1	1	100
2018-601	023	ORTHO			001 001		001 - NEG/NR	2	2	2	100
2018-601	039	ECI VITROS			001 001		001 - NEG/NR	5	5	5	100
2018-601	042	UNI-GOLD		001 001	001 001		001 - NEG/NR	28	28	28	100
2018-601	044	ADVIA CENT			001 001		001 - NEG/NR	9	9	9	100
2018-601	048	ABBOTT			001 001		001 - NEG/NR	6	6	6	100
2018-601	050	ARCHITECT			001 001		001 - NEG/NR	2	2	2	100
2018-601	053	GS HIV AG/			001 001		001 - NEG/NR	2	2	2	100
2018-601	054	BIO-RAD 5G			001 001		001 - NEG/NR	1	1	1	100
2018-601	055	ALERE DETE			001 001		001 - NEG/NR	8	8	8	100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 230 HBsAG, ANTI HBC,HIV

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
093 ANTI-HUMAN IMM. (HIV)											
2018-602	002	ABBOTT			002	002	002 - POS/R	2	2	2	100
2018-602	006	BIOTECH			002	002	002 - POS/R	1	1	1	100
2018-602	023	ORTHO			002	002	002 - POS/R	2	2	2	100
2018-602	039	ECI VITROS			002	002	002 - POS/R	5	5	5	100
2018-602	042	UNI-GOLD		002	002	002	001 - NEG/NR	28	2	26	0
2018-602	042	UNI-GOLD		002	002	002	002 - POS/R	28	26	26	100
2018-602	044	ADVIA CENT			002	002	002 - POS/R	9	9	9	100
2018-602	048	ABBOTT			002	002	002 - POS/R	6	6	6	100
2018-602	050	ARCHITECT			002	002	002 - POS/R	2	2	2	100
2018-602	053	GS HIV AG/			002	002	002 - POS/R	2	2	2	100
2018-602	054	BIO-RAD 5G			002	002	002 - POS/R	1	1	1	100
2018-602	055	ALERE DETE			002	002	002 - POS/R	8	8	8	100
2018-603	002	ABBOTT			001	001	001 - NEG/NR	2	2	2	100
2018-603	006	BIOTECH			001	001	001 - NEG/NR	1	1	1	100
2018-603	023	ORTHO			001	001	001 - NEG/NR	2	2	2	100
2018-603	039	ECI VITROS			001	001	001 - NEG/NR	5	5	5	100
2018-603	042	UNI-GOLD		001	001	001	001 - NEG/NR	28	28	28	100
2018-603	044	ADVIA CENT			001	001	001 - NEG/NR	9	9	9	100
2018-603	048	ABBOTT			001	001	001 - NEG/NR	6	6	6	100
2018-603	050	ARCHITECT			001	001	001 - NEG/NR	2	2	2	100
2018-603	053	GS HIV AG/			001	001	001 - NEG/NR	2	2	2	100
2018-603	054	BIO-RAD 5G			001	001	001 - NEG/NR	1	1	1	100
2018-603	055	ALERE DETE			001	001	001 - NEG/NR	8	8	8	100
2018-604	002	ABBOTT			002	002	002 - POS/R	2	2	2	100
2018-604	006	BIOTECH			002	002	002 - POS/R	1	1	1	100
2018-604	023	ORTHO			002	002	002 - POS/R	2	2	2	100

## Participant Summary

MONTH 7 OF YEAR 2018

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 230 HBsAG, ANTI HBC, HIV

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2018-604	039 ECI VITROS				002 002		002 - POS/R	5	5	5	100
2018-604	042 UNI-GOLD			002 002	002 002		001 - NEG/NR	28	4	24	0
2018-604	042 UNI-GOLD			002 002	002 002		002 - POS/R	28	24	24	100
2018-604	044 ADVIA CENT				002 002		002 - POS/R	9	9	9	100
2018-604	048 ABBOTT				002 002		002 - POS/R	6	6	6	100
2018-604	050 ARCHITECT				002 002		002 - POS/R	2	2	2	100
2018-604	053 GS HIV AG/				002 002		002 - POS/R	2	2	2	100
2018-604	054 BIO-RAD 5G				002 002		002 - POS/R	1	1	1	100
2018-604	055 ALERE DETE				002 002		002 - POS/R	8	8	8	100
093 ANTI-HUMAN IMM. (HIV)											
2018-605	002 ABBOTT				001 001		001 - NEG/NR	2	2	2	100
2018-605	006 BIOTECH				001 001		001 - NEG/NR	1	1	1	100
2018-605	023 ORTHO				001 001		001 - NEG/NR	2	2	2	100
2018-605	039 ECI VITROS				001 001		001 - NEG/NR	5	5	5	100
2018-605	042 UNI-GOLD			001 001	001 001		001 - NEG/NR	28	28	28	100
2018-605	044 ADVIA CENT				001 001		001 - NEG/NR	9	9	9	100
2018-605	048 ABBOTT				001 001		001 - NEG/NR	6	6	6	100
2018-605	050 ARCHITECT				001 001		001 - NEG/NR	2	2	2	100
2018-605	053 GS HIV AG/				001 001		001 - NEG/NR	2	2	2	100
2018-605	054 BIO-RAD 5G				001 001		001 - NEG/NR	1	1	1	100
2018-605	055 ALERE DETE				001 001		001 - NEG/NR	8	8	8	100
096 HEPATITIS (HBsAG)											
2018-601	002 ABBOTT			002 002	002 002		002 - POS/ R	2	2	2	100
2018-601	023 ORTHO			002 002	002 002		002 - POS/ R	3	3	3	100
2018-601	038 ELEC. 2010			002 002	002 002		002 - POS/ R	2	2	2	100
2018-601	039 ECI VITROS			002 002	002 002		002 - POS/ R	4	4	4	100
2018-601	044 ADVIA CENT			002 002	002 002		002 - POS/ R	9	9	9	100
2018-601	048 ABBOTT			002 002	002 002		002 - POS/ R	6	6	6	100
2018-601	052 COBAS E 60			002 002	002 002		002 - POS/ R	1	1	1	100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
096 HEPATITIS (HBsAG)												
2018-602	002	ABBOTT		002	002	002	002	002 - POS/ R	2	2	2	100
2018-602	023	ORTHO		002	002	002	002	002 - POS/ R	3	3	3	100
2018-602	038	ELEC. 2010		002	002	002	002	002 - POS/ R	2	2	2	100
2018-602	039	ECI VITROS		002	002	002	002	002 - POS/ R	4	4	4	100
2018-602	044	ADVIA CENT		002	002	002	002	002 - POS/ R	9	9	9	100
2018-602	048	ABBOTT		002	002	002	002	002 - POS/ R	6	6	6	100
2018-602	052	COBAS E 60		002	002	002	002	002 - POS/ R	1	1	1	100
2018-603	002	ABBOTT		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-603	023	ORTHO		001	001	001	001	001 - NEG/ NR	3	3	3	100
2018-603	038	ELEC. 2010		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-603	039	ECI VITROS		001	001	001	001	001 - NEG/ NR	4	4	4	100
2018-603	044	ADVIA CENT		001	001	001	001	001 - NEG/ NR	9	9	9	100
2018-603	048	ABBOTT		001	001	001	001	001 - NEG/ NR	6	6	6	100
2018-603	052	COBAS E 60		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-604	002	ABBOTT		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-604	023	ORTHO		001	001	001	001	001 - NEG/ NR	3	3	3	100
2018-604	038	ELEC. 2010		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-604	039	ECI VITROS		001	001	001	001	001 - NEG/ NR	4	4	4	100
2018-604	044	ADVIA CENT		001	001	001	001	001 - NEG/ NR	9	9	9	100
2018-604	048	ABBOTT		001	001	001	001	001 - NEG/ NR	6	6	6	100
2018-604	052	COBAS E 60		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-605	002	ABBOTT		002	002	002	002	002 - POS/ R	2	2	2	100
2018-605	023	ORTHO		002	002	002	002	002 - POS/ R	3	3	3	100
2018-605	038	ELEC. 2010		002	002	002	002	002 - POS/ R	2	2	2	100
2018-605	039	ECI VITROS		002	002	002	002	002 - POS/ R	4	4	4	100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-605	044 ADVIA CENT			002 002	002 002		002 - POS/ R	9	9	9	100	
2018-605	048 ABBOTT			002 002	002 002		002 - POS/ R	6	6	6	100	
2018-605	052 COBAS E 60			002 002	002 002		002 - POS/ R	1	1	1	100	
	097 HEPATITIS (antiHBC)											
2018-601	002 ABBOTT			001 001	001 001		001 - NEG/ NR	2	2	2	100	
2018-601	023 ORTHO			001 001	001 001		001 - NEG/ NR	1	1	1	100	
2018-601	039 ECI VITROS			001 001	001 001		001 - NEG/ NR	1	1	1	100	
2018-601	044 ADVIA CENT			001 001	001 001		001 - NEG/ NR	8	8	8	100	
2018-601	048 ABBOTT			001 001	001 001		001 - NEG/ NR	3	3	3	100	
2018-602	002 ABBOTT			002 002	002 002		002 - POS/ R	2	2	2	100	
2018-602	023 ORTHO			002 002	002 002		002 - POS/ R	1	1	1	100	
2018-602	039 ECI VITROS			002 002	002 002		002 - POS/ R	1	1	1	100	
2018-602	044 ADVIA CENT			002 002	002 002		002 - POS/ R	8	8	8	100	
2018-602	048 ABBOTT			002 002	002 002		002 - POS/ R	3	3	3	100	
2018-603	002 ABBOTT				001 001		002 - POS/ R	2	2	0	0	
2018-603	023 ORTHO				001 001		001 - NEG/ NR	1	1	1	100	
2018-603	039 ECI VITROS				001 001		001 - NEG/ NR	1	1	1	100	
2018-603	044 ADVIA CENT				001 001		001 - NEG/ NR	8	8	8	100	
2018-603	048 ABBOTT				001 001		002 - POS/ R	3	2	1	0	
2018-603	048 ABBOTT				001 001		001 - NEG/ NR	3	1	1	100	
2018-604	002 ABBOTT			001 001	001 001		001 - NEG/ NR	2	2	2	100	
2018-604	023 ORTHO			001 001	001 001		001 - NEG/ NR	1	1	1	100	
2018-604	039 ECI VITROS			001 001	001 001		001 - NEG/ NR	1	1	1	100	
2018-604	044 ADVIA CENT			001 001	001 001		001 - NEG/ NR	8	8	8	100	
2018-604	048 ABBOTT			001 001	001 001		002 - POS/ R	3	1	2	0	
2018-604	048 ABBOTT			001 001	001 001		001 - NEG/ NR	3	2	2	100	

## Participant Summary

MONTH 7 OF YEAR 2018

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
097 HEPATITIS (antiHBC)												
2018-605	002	ABBOTT		002	002	002	002	002 - POS/ R	2	2	2	100
2018-605	023	ORTHO		002	002	002	002	002 - POS/ R	1	1	1	100
2018-605	039	ECI VITROS		002	002	002	002	002 - POS/ R	1	1	1	100
2018-605	044	ADVIA CENT		002	002	002	002	002 - POS/ R	8	8	8	100
2018-605	048	ABBOTT		002	002	002	002	002 - POS/ R	3	3	3	100
171 HEPATITIS C												
2018-601	002	ABBOTT		002	002			002 - POS/ R	2	2	2	100
2018-601	023	ORTHO		002	002			002 - POS/ R	2	2	2	100
2018-601	038	ELEC. 2010		002	002			002 - POS/ R	1	1	1	100
2018-601	039	ECI VITROS		002	002			002 - POS/ R	5	5	5	100
2018-601	044	ADVIA CENT		002	002			002 - POS/ R	8	8	8	100
2018-601	048	ABBOTT		002	002			002 - POS/ R	5	5	5	100
2018-602	002	ABBOTT		001	001			001 - NEG/ NR	2	2	2	100
2018-602	023	ORTHO		001	001			001 - NEG/ NR	2	2	2	100
2018-602	038	ELEC. 2010		001	001			001 - NEG/ NR	1	1	1	100
2018-602	039	ECI VITROS		001	001			001 - NEG/ NR	5	5	5	100
2018-602	044	ADVIA CENT		001	001			001 - NEG/ NR	8	6	6	100
2018-602	044	ADVIA CENT		001	001			002 - POS/ R	8	2	6	0
2018-602	048	ABBOTT		001	001			001 - NEG/ NR	5	5	5	100
2018-603	002	ABBOTT		001	001			001 - NEG/ NR	2	2	2	100
2018-603	023	ORTHO		001	001			001 - NEG/ NR	2	2	2	100
2018-603	038	ELEC. 2010		001	001			001 - NEG/ NR	1	1	1	100
2018-603	039	ECI VITROS		001	001			001 - NEG/ NR	5	5	5	100
2018-603	044	ADVIA CENT		001	001			001 - NEG/ NR	8	8	8	100
2018-603	048	ABBOTT		001	001			001 - NEG/ NR	5	5	5	100



**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL			GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
171 HEPATITIS C											
2018-604	002	ABBOTT		002	002		002 - POS/ R	2	2	2	100
2018-604	023	ORTHO		002	002		002 - POS/ R	2	2	2	100
2018-604	038	ELEC. 2010		002	002		002 - POS/ R	1	1	1	100
2018-604	039	ECI VITROS		002	002		002 - POS/ R	5	5	5	100
2018-604	044	ADVIA CENT		002	002		002 - POS/ R	8	8	8	100
2018-604	048	ABBOTT		002	002		002 - POS/ R	5	5	5	100
2018-605	002	ABBOTT		002	002		002 - POS/ R	2	2	2	100
2018-605	023	ORTHO		002	002		002 - POS/ R	2	2	2	100
2018-605	038	ELEC. 2010		002	002		002 - POS/ R	1	1	1	100
2018-605	039	ECI VITROS		002	002		002 - POS/ R	5	5	5	100
2018-605	044	ADVIA CENT		002	002		002 - POS/ R	8	8	8	100
2018-605	048	ABBOTT		002	002		002 - POS/ R	5	5	5	100
181 HAV Total											
2018-601	002	ABBOTT			002	002	002 - POS/ R	2	2	2	100
2018-601	023	ORTHO			002	002	002 - POS/ R	1	1	1	100
2018-601	044	ADVIA CENT			002	002	002 - POS/ R	4	4	4	100
2018-601	048	ABBOTT			002	002	002 - POS/ R	1	1	1	100
2018-602	002	ABBOTT			001	001	002 - POS/ R	2	2	0	0
2018-602	023	ORTHO			001	001	001 - NEG/ NR	1	1	1	100
2018-602	044	ADVIA CENT			001	001	001 - NEG/ NR	4	4	4	100
2018-602	048	ABBOTT			001	001	002 - POS/ R	1	1	0	0
2018-603	002	ABBOTT			002	002	002 - POS/ R	2	2	2	100
2018-603	023	ORTHO			002	002	002 - POS/ R	1	1	1	100
2018-603	044	ADVIA CENT			002	002	002 - POS/ R	4	4	4	100

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-603	048	ABBOTT			002	002	002 - POS/ R	1	1	1	100	
181 HAV Total												
2018-604	002	ABBOTT			001	001	002 - POS/ R	2	2	0	0	
2018-604	023	ORTHO			001	001	001 - NEG/ NR	1	1	1	100	
2018-604	044	ADVIA CENT			001	001	001 - NEG/ NR	4	4	4	100	
2018-604	048	ABBOTT			001	001	002 - POS/ R	1	1	0	0	
2018-605	002	ABBOTT			001	001	001 - NEG/ NR	2	2	2	100	
2018-605	023	ORTHO			001	001	001 - NEG/ NR	1	1	1	100	
2018-605	044	ADVIA CENT			001	001	001 - NEG/ NR	4	4	4	100	
2018-605	048	ABBOTT			001	001	001 - NEG/ NR	1	1	1	100	
182 HEPATITIS(Anti-HBs)												
2018-601	002	ABBOTT		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-601	023	ORTHO		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-601	039	ECI VITROS		001	001	001	001	001 - NEG/ NR	4	4	4	100
2018-601	044	ADVIA CENT		001	001	001	001	001 - NEG/ NR	3	3	3	100
2018-601	048	ABBOTT		001	001	001	001	001 - NEG/ NR	6	6	6	100
2018-601	050	ARCHITECT		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-602	002	ABBOTT		001	001	001	001	001 - NEG/ NR	2	2	2	100
2018-602	023	ORTHO		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-602	039	ECI VITROS		001	001	001	001	001 - NEG/ NR	4	4	4	100
2018-602	044	ADVIA CENT		001	001	001	001	001 - NEG/ NR	3	3	3	100
2018-602	048	ABBOTT		001	001	001	001	001 - NEG/ NR	6	6	6	100
2018-602	050	ARCHITECT		001	001	001	001	001 - NEG/ NR	1	1	1	100
2018-603	002	ABBOTT		002	002	002	002	002 - POS/ R	2	2	2	100
2018-603	023	ORTHO		002	002	002	002	002 - POS/ R	1	1	1	100

## Participant Summary

MONTH 7 OF YEAR 2018

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC, HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-603	039 ECI VITROS			002 002	002 002		002 - POS/ R	4	4	4	100	
2018-603	044 ADVIA CENT			002 002	002 002		002 - POS/ R	3	3	3	100	
2018-603	048 ABBOTT			002 002	002 002		002 - POS/ R	6	6	6	100	
2018-603	050 ARCHITECT			002 002	002 002		002 - POS/ R	1	1	1	100	
182 HEPATITIS(Anti-HBs)												
2018-604	002 ABBOTT			002 002	002 002		002 - POS/ R	2	2	2	100	
2018-604	023 ORTHO			002 002	002 002		002 - POS/ R	1	1	1	100	
2018-604	039 ECI VITROS			002 002	002 002		002 - POS/ R	4	4	4	100	
2018-604	044 ADVIA CENT			002 002	002 002		002 - POS/ R	3	3	3	100	
2018-604	048 ABBOTT			002 002	002 002		002 - POS/ R	6	6	6	100	
2018-604	050 ARCHITECT			002 002	002 002		002 - POS/ R	1	1	1	100	
2018-605	002 ABBOTT			001 001	001 001		001 - NEG/ NR	2	2	2	100	
2018-605	023 ORTHO			001 001	001 001		001 - NEG/ NR	1	1	1	100	
2018-605	039 ECI VITROS			001 001	001 001		001 - NEG/ NR	4	4	4	100	
2018-605	044 ADVIA CENT			001 001	001 001		001 - NEG/ NR	3	3	3	100	
2018-605	048 ABBOTT			001 001	001 001		001 - NEG/ NR	6	6	6	100	
2018-605	050 ARCHITECT			001 001	001 001		001 - NEG/ NR	1	1	1	100	
183 HAV IgM												
2018-601	002 ABBOTT			001 001	001 001		001 - Negative	2	2	2	100	
2018-601	023 ORTHO			001 001	001 001		001 - Negative	1	1	1	100	
2018-601	039 ECI VITROS			001 001	001 001		001 - Negative	4	4	4	100	
2018-601	044 ADVIA CENT			001 001	001 001		001 - Negative	6	6	6	100	
2018-601	048 ABBOTT			001 001	001 001		001 - Negative	4	4	4	100	
2018-602	002 ABBOTT			001 001	001 001		001 - Negative	2	2	2	100	
2018-602	023 ORTHO			001 001	001 001		001 - Negative	1	1	1	100	
2018-602	039 ECI VITROS			001 001	001 001		001 - Negative	4	4	4	100	
2018-602	044 ADVIA CENT			001 001	001 001		001 - Negative	6	6	6	100	

**Participant Summary**

**MONTH 7 OF YEAR 2018**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC,HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2018-602	048	ABBOTT		001 001	001 001		001 - Negative	4	4	4	100	
	183 HAV IgM											
2018-603	002	ABBOTT		001 001	001 001		001 - Negative	2	2	2	100	
2018-603	023	ORTHO		001 001	001 001		001 - Negative	1	1	1	100	
2018-603	039	ECI VITROS		001 001	001 001		001 - Negative	4	4	4	100	
2018-603	044	ADVIA CENT		001 001	001 001		001 - Negative	6	6	6	100	
2018-603	048	ABBOTT		001 001	001 001		001 - Negative	4	4	4	100	
2018-604	002	ABBOTT		001 001	001 001		001 - Negative	2	2	2	100	
2018-604	023	ORTHO		001 001	001 001		001 - Negative	1	1	1	100	
2018-604	039	ECI VITROS		001 001	001 001		001 - Negative	4	4	4	100	
2018-604	044	ADVIA CENT		001 001	001 001		001 - Negative	6	6	6	100	
2018-604	048	ABBOTT		001 001	001 001		001 - Negative	4	4	4	100	
2018-605	002	ABBOTT		001 001	001 001		001 - Negative	2	2	2	100	
2018-605	023	ORTHO		001 001	001 001		001 - Negative	1	1	1	100	
2018-605	039	ECI VITROS		001 001	001 001		001 - Negative	4	4	4	100	
2018-605	044	ADVIA CENT		001 001	001 001		001 - Negative	6	6	6	100	
2018-605	048	ABBOTT		001 001	001 001		001 - Negative	4	4	4	100	