

Participant Summary
MONTH 7 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 15 WHOLE BLOOD GLUCOSE

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
157 WHOLE BLOOD GLUCOSE			CLIA TEST ID #: 0						EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}
2020-491	20	ABBOT FREE	18	500.000	0.000	0.00	450.000	550.000	
2020-491	23	ABBOTT PRE	32	442.656	43.833	9.90	398.000	487.000	
2020-491	27	STAT SRIP.	50	431.000	32.279	7.49	388.000	474.000	
2020-491	33	AC.CHEK II	23	596.182	7.685	1.29	537.000	656.000	
2020-491	34	FREE STYL	36	500.000	0.000	0.00	450.000	550.000	
2020-492	20	ABBOT FREE	18	20.000	0.000	0.00	14.000	26.000	
2020-492	23	ABBOTT PRE	32	20.000	0.000	0.00	14.000	26.000	
2020-492	27	STAT SRIP.	50	10.745	2.301	21.42	5.000	17.000	
2020-492	33	AC.CHEK II	23	15.217	0.930	6.11	9.000	21.000	
2020-492	34	FREE STYL	36	20.000	0.000	0.00	14.000	26.000	
2020-493	20	ABBOT FREE	18	500.000	0.000	0.00	450.000	550.000	
2020-493	23	ABBOTT PRE	32	415.031	46.467	11.20	374.000	457.000	
2020-493	27	STAT SRIP.	50	349.160	36.564	10.47	314.000	384.000	
2020-493	33	AC.CHEK II	23	508.696	25.022	4.92	458.000	560.000	
2020-493	34	FREE STYL	36	500.000	0.000	0.00	450.000	550.000	
2020-494	20	ABBOT FREE	18	213.722	12.875	6.02	192.000	235.000	
2020-494	23	ABBOTT PRE	32	188.438	9.756	5.18	170.000	207.000	
2020-494	27	STAT SRIP.	50	162.612	8.412	5.17	146.000	179.000	
2020-494	33	AC.CHEK II	23	197.818	5.882	2.97	178.000	218.000	
2020-494	34	FREE STYL	36	212.444	8.197	3.86	191.000	234.000	
2020-495	20	ABBOT FREE	18	400.882	15.491	3.86	361.000	441.000	
2020-495	23	ABBOTT PRE	32	328.125	24.220	7.38	295.000	361.000	
2020-495	27	STAT SRIP.	50	278.000	22.717	8.17	250.000	306.000	
2020-495	33	AC.CHEK II	23	364.182	9.764	2.68	328.000	401.000	
2020-495	34	FREE STYL	36	422.139	17.232	4.08	380.000	464.000	

Participant Summary
MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 140 COAGULATION (HEMAT.)						
ANALYTE											
SAMPLE	METHOD	SYST/INST		REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
080	FIBRINOGEN	CLIA TEST ID #: 825		EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-551	60 DADE SYSME	67	DADE		12	113.500	7.826	6.90	91.000	136.000	
2020-552	60 DADE SYSME	67	DADE		12	389.583	26.094	6.70	312.000	468.000	
2020-553	60 DADE SYSME	67	DADE		12	287.750	19.984	6.94	230.000	345.000	
2020-554	60 DADE SYSME	67	DADE		12	330.333	22.969	6.95	264.000	396.000	
2020-555	60 DADE SYSME	67	DADE		12	290.250	17.650	6.08	232.000	348.000	
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835		EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-551	33 MLA ELECT.	54	QUICK COAG		12	37.842	2.930	7.74	32.200	43.500	
2020-551	58 AMELUNG KC	42	BIOMERIEUX		10	45.950	5.152	11.21	39.100	52.800	
2020-551	60 DADE SYSME	35	DADE ACTIN		14	34.493	0.985	2.86	29.300	39.700	
2020-551	62 DADE SYS.	35	DADE ACTIN		61	32.613	1.787	5.48	27.700	37.500	
2020-551	65 DADE BFTII	35	DADE ACTIN		13	36.331	2.563	7.06	30.900	41.800	
2020-551	69 DIAG.STAGO	49	STA-PTT		10	47.000	2.064	4.39	40.000	54.100	
2020-551	73 ACL 8 TO E	56	PTT-HEMO		37	35.884	1.045	2.91	30.500	41.300	
2020-551	76 STAGO STAR	49	STA-PTT		21	46.329	4.702	10.15	39.400	53.300	
2020-552	33 MLA ELECT.	54	QUICK COAG		12	29.342	1.628	5.55	24.900	33.700	
2020-552	58 AMELUNG KC	42	BIOMERIEUX		10	34.730	2.628	7.57	29.500	39.900	
2020-552	60 DADE SYSME	35	DADE ACTIN		14	32.071	1.140	3.55	27.300	36.900	
2020-552	62 DADE SYS.	35	DADE ACTIN		61	30.888	1.644	5.32	26.300	35.500	
2020-552	65 DADE BFTII	35	DADE ACTIN		13	30.323	1.843	6.08	25.800	34.900	
2020-552	69 DIAG.STAGO	49	STA-PTT		10	36.400	0.840	2.31	30.900	41.900	
2020-552	73 ACL 8 TO E	56	PTT-HEMO		37	33.619	1.019	3.03	28.600	38.700	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-552	76 STAGO STAR	49 STA-PTT		21	34.957	2.245	6.42	29.700	40.200		
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %								
2020-553	33 MLA ELECT.	54 QUICK COAG		12	64.875	1.672	2.58	55.100	74.600		
2020-553	58 AMELUNG KC	42 BIOMERIEUX		10	58.640	4.543	7.75	49.800	67.400		
2020-553	60 DADE SYSME	35 DADE ACTIN		14	59.607	2.475	4.15	50.700	68.500		
2020-553	62 DADE SYS.	35 DADE ACTIN		61	56.235	3.069	5.46	47.800	64.700		
2020-553	65 DADE BFTII	35 DADE ACTIN		13	65.715	2.544	3.87	55.900	75.600		
2020-553	69 DIAG.STAGO	49 STA-PTT		10	58.750	2.006	3.41	49.900	67.600		
2020-553	73 ACL 8 TO E	56 PTT-HEMO		37	69.874	2.730	3.91	59.400	80.400		
2020-553	76 STAGO STAR	49 STA-PTT		21	59.429	5.435	9.15	50.500	68.300		
2020-554	33 MLA ELECT.	54 QUICK COAG		12	27.500	1.760	6.40	23.400	31.600		
2020-554	58 AMELUNG KC	42 BIOMERIEUX		10	29.440	2.301	7.82	25.000	33.900		
2020-554	60 DADE SYSME	35 DADE ACTIN		14	27.257	0.714	2.62	23.200	31.300		
2020-554	62 DADE SYS.	35 DADE ACTIN		61	25.990	1.127	4.34	22.100	29.900		
2020-554	65 DADE BFTII	35 DADE ACTIN		13	25.431	1.687	6.63	21.600	29.200		
2020-554	69 DIAG.STAGO	49 STA-PTT		10	30.300	1.083	3.57	25.800	34.800		
2020-554	73 ACL 8 TO E	56 PTT-HEMO		37	29.714	1.076	3.62	25.300	34.200		
2020-554	76 STAGO STAR	49 STA-PTT		21	29.338	2.197	7.49	24.900	33.700		
2020-555	33 MLA ELECT.	54 QUICK COAG		12	93.100	4.550	4.89	79.100	107.100		
2020-555	58 AMELUNG KC	42 BIOMERIEUX		10	84.910	5.976	7.04	72.200	97.600		
2020-555	60 DADE SYSME	35 DADE ACTIN		14	85.086	3.182	3.74	72.300	97.800		
2020-555	62 DADE SYS.	35 DADE ACTIN		61	79.003	3.821	4.84	67.200	90.900		
2020-555	65 DADE BFTII	35 DADE ACTIN		13	93.377	6.484	6.94	79.400	107.400		
2020-555	69 DIAG.STAGO	49 STA-PTT		10	83.460	2.652	3.18	70.900	96.000		
2020-555	73 ACL 8 TO E	56 PTT-HEMO		37	98.814	3.817	3.86	84.000	113.600		
2020-555	76 STAGO STAR	49 STA-PTT		21	85.395	9.879	11.57	72.600	98.200		
082	PROTHROMBIN TIME(P.T)	CLIA TEST ID #: 845									
2020-551	33 MLA ELECT.	37 QUICK COAG		12	17.400	1.361	7.82	14.800	20.000		

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-551	58	AMELUNG KC	15 BIOMERIEUX	12	14.675	0.847	5.77	12.500	16.900	
2020-551	60	DADE SYSME	33 DADE INNOV	15	14.887	0.270	1.82	12.700	17.100	
2020-551	62	DADE SYS.	33 DADE INNOV	62	14.469	0.548	3.79	12.300	16.600	
2020-551	65	DADE BFTII	33 DADE INNOV	14	14.314	1.239	8.65	12.200	16.500	
2020-551	73	ACL 8 TO E	41 HEMOSIL RE	30	16.131	0.416	2.58	13.700	18.600	
2020-551	73	ACL 8 TO E	43 HEMOSIL PT	11	20.391	0.881	4.32	17.300	23.400	
2020-551	76	STAGO STAR	44 STA NEO CI	14	18.314	0.823	4.49	15.600	21.100	
082 PROTHROMBIN TIME(PT)					CLIA TEST ID #: 845 EVALUATION CRITERIA: TARGET VALUE 15 %					
2020-552	33	MLA ELECT.	37 QUICK COAG	12	13.208	0.780	5.91	11.200	15.200	
2020-552	58	AMELUNG KC	15 BIOMERIEUX	12	11.258	0.663	5.89	9.600	12.900	
2020-552	60	DADE SYSME	33 DADE INNOV	15	10.893	0.188	1.72	9.300	12.500	
2020-552	62	DADE SYS.	33 DADE INNOV	62	10.356	0.414	3.99	8.800	11.900	
2020-552	65	DADE BFTII	33 DADE INNOV	14	9.664	0.656	6.79	8.200	11.100	
2020-552	73	ACL 8 TO E	41 HEMOSIL RE	30	11.857	0.427	3.60	10.100	13.600	
2020-552	73	ACL 8 TO E	43 HEMOSIL PT	11	13.800	0.684	4.95	11.700	15.900	
2020-552	76	STAGO STAR	44 STA NEO CI	14	12.843	0.512	3.99	10.900	14.800	
2020-553	33	MLA ELECT.	37 QUICK COAG	12	31.367	3.352	10.69	26.700	36.100	
2020-553	58	AMELUNG KC	15 BIOMERIEUX	12	20.033	2.112	10.54	17.000	23.000	
2020-553	60	DADE SYSME	33 DADE INNOV	15	26.300	1.037	3.94	22.400	30.200	
2020-553	62	DADE SYS.	33 DADE INNOV	62	25.328	1.118	4.41	21.500	29.100	
2020-553	65	DADE BFTII	33 DADE INNOV	14	26.471	1.759	6.65	22.500	30.400	
2020-553	73	ACL 8 TO E	41 HEMOSIL RE	30	33.014	1.681	5.09	28.100	38.000	
2020-553	73	ACL 8 TO E	43 HEMOSIL PT	11	38.664	2.578	6.67	32.900	44.500	
2020-553	76	STAGO STAR	44 STA NEO CI	14	36.450	3.002	8.24	31.000	41.900	
2020-554	33	MLA ELECT.	37 QUICK COAG	12	13.217	0.768	5.81	11.200	15.200	
2020-554	58	AMELUNG KC	15 BIOMERIEUX	12	10.975	0.536	4.88	9.300	12.600	
2020-554	60	DADE SYSME	33 DADE INNOV	15	10.880	0.147	1.35	9.200	12.500	
2020-554	62	DADE SYS.	33 DADE INNOV	62	10.364	0.364	3.51	8.800	11.900	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-554	65	DADE BFTII	33	DADE INNOV	14	9.438	0.570	6.04	8.000	10.900	
2020-554	73	ACL 8 TO E	41	HEMOSIL RE	30	11.827	0.457	3.86	10.100	13.600	
2020-554	73	ACL 8 TO E	43	HEMOSIL PT	11	13.945	0.593	4.25	11.900	16.000	
2020-554	76	STAGO STAR	44	STA NEO CI	14	12.800	0.600	4.69	10.900	14.700	
082 PROTHROMBIN TIME(PT)		CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-555	33	MLA ELECT.	37	QUICK COAG	12	54.608	4.414	8.08	46.400	62.800	
2020-555	58	AMELUNG KC	15	BIOMERIEUX	12	30.009	1.884	6.28	25.500	34.500	
2020-555	60	DADE SYSME	33	DADE INNOV	15	45.427	1.817	4.00	38.600	52.200	
2020-555	62	DADE SYS.	33	DADE INNOV	62	43.235	2.263	5.23	36.700	49.700	
2020-555	65	DADE BFTII	33	DADE INNOV	14	47.736	3.423	7.17	40.600	54.900	
2020-555	73	ACL 8 TO E	41	HEMOSIL RE	30	59.563	3.460	5.81	50.600	68.500	
2020-555	73	ACL 8 TO E	43	HEMOSIL PT	11	72.491	6.666	9.20	61.600	83.400	
2020-555	76	STAGO STAR	44	STA NEO CI	14	67.593	7.263	10.74	57.500	77.700	

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 100 HEMA. CELL IDENTIFI.							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2020-501					018 018		071 - Poikilo.	859	1	855	0
2020-501					018 018		018 - PLATELETS,	859	855	855	100
2020-501					018 018		013 - LYMPHOCYTE	859	1	855	0
2020-501					018 018		020 - GIANT PLA	859	2	855	0
2020-502					025 025		024 - AUER ROD	860	6	851	0
2020-502					025 025		025 - DOHLE INCL	860	851	851	100
2020-502					025 025		018 - PLATELETS,	860	1	851	0
2020-502					025 025		006 - SEGMENTED	860	2	851	0
2020-503					071 071		053 - STOMATOCYT	1036	17	846	0

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-503					071 071		055 - TEAR-TAILE	1036	32	846	0
2020-503					071 071		049 - FRAGMENTED	1036	21	846	0
2020-503					071 071		036 - BASOPHILIC	1036	1	846	0
2020-503					071 071		018 - PLATELETS,	1036	5	846	0
2020-503					071 071		071 - Poikilo.	1036	846	846	100
2020-503					071 071		048 - ERYTHROCYT	1036	29	846	0
2020-503					071 071		050 - OVALOCYTE	1036	25	846	0
2020-503					071 071		033 - POLYCHROMA	1036	3	846	0
2020-503					071 071		044 - ERYTHROCYT	1036	28	846	0
2020-503					071 071		041 - ACANTHOCYT	1036	4	846	0
2020-503					071 071		020 - GIANT PLA	1036	3	846	0
2020-503					071 071		052 - SPHEROCYTE	1036	5	846	0
2020-503					071 071		030 - ERYTHROCYT	1036	2	846	0
2020-503					071 071		042 - BITE CELL	1036	1	846	0
2020-503					071 071		046 - ERYTHROCYT	1036	14	846	0
072 HEMA. CELL IDENTIFI.											
2020-504					001 001		014 - LYMPHOCYTE	859	14	841	0
2020-504					001 001		001 - BLAST	859	841	841	100
2020-504					001 001		013 - LYMPHOCYTE	859	2	841	0
2020-504					001 001		069 - Prolymphoc	859	2	841	0
2020-505					017 017		014 - LYMPHOCYTE	859	38	820	0
2020-505					017 017		004 - METAMYELOLOC	859	1	820	0
2020-505					017 017		017 - MONOCYTE	859	820	820	100

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{GREATER}			
2020-531	7	M.B.MANUAL		18	5.650	1.681	29.75	0.600	10.700
2020-531	9	M.B. RICCA		31	5.403	1.680	31.10	0.400	10.400
2020-531	24	Unicel DxH		12	3.558	0.278	7.82	2.700	4.400
2020-531	25	SYSMEX XN		25	2.972	0.266	8.95	2.200	3.800
2020-532	7	M.B.MANUAL		18	1.233	0.337	27.30	0.200	2.200
2020-532	9	M.B. RICCA		31	1.235	0.303	24.55	0.300	2.100
2020-532	24	Unicel DxH		12	0.700	0.082	11.66	0.500	0.900
2020-532	25	SYSMEX XN		25	0.992	0.116	11.73	0.600	1.300
2020-533	7	M.B.MANUAL		18	5.872	2.204	37.53	0	12.500
2020-533	9	M.B. RICCA		31	5.245	1.318	25.13	1.300	9.200
2020-533	24	Unicel DxH		12	3.600	0.436	12.11	2.300	4.900
2020-533	25	SYSMEX XN		25	2.920	0.232	7.93	2.200	3.700
2020-534	7	M.B.MANUAL		18	10.656	2.652	24.89	2.700	18.600
2020-534	9	M.B. RICCA		31	10.845	2.075	19.13	4.600	17.100
2020-534	24	Unicel DxH		12	8.308	0.366	4.41	6.200	10.400
2020-534	25	SYSMEX XN		25	5.864	0.701	11.96	3.800	8.000
2020-535	7	M.B.MANUAL		18	1.189	0.333	28.02	0.200	2.200
2020-535	9	M.B. RICCA		31	1.273	0.277	21.74	0.400	2.100
2020-535	24	Unicel DxH		12	0.725	0.164	22.61	0.200	1.200
2020-535	25	SYSMEX XN		25	0.976	0.134	13.68	0.600	1.400

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
153	SED. RATE (ESR)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{GREATER}				
2020-541	4	SEDIPLAST		559	40.327	7.919	19.64	17.000	64.000	
2020-541	10	SEDI RATE		118	40.462	8.120	20.07	16.000	65.000	
2020-541	11	SEDIMAT-15		68	131.076	11.666	8.90	96.000	166.000	
2020-541	14	GLOBAL SC.		11	42.091	6.431	15.28	23.000	61.000	
2020-541	19	MINI CUBE		16	34.563	10.482	30.33	3.000	66.000	
2020-542	4	SEDIPLAST		559	41.947	7.335	17.49	20.000	64.000	
2020-542	10	SEDI RATE		118	40.718	7.629	18.74	18.000	64.000	
2020-542	11	SEDIMAT-15		68	123.060	16.406	13.33	74.000	172.000	
2020-542	14	GLOBAL SC.		11	42.545	4.887	11.49	28.000	57.000	
2020-542	19	MINI CUBE		16	66.625	20.814	31.24	4.000	129.000	
2020-543	4	SEDIPLAST		559	42.063	7.269	17.28	20.000	64.000	
2020-543	10	SEDI RATE		118	42.157	7.635	18.11	19.000	65.000	
2020-543	11	SEDIMAT-15		68	124.269	13.401	10.78	84.000	164.000	
2020-543	14	GLOBAL SC.		11	42.000	7.628	18.16	19.000	65.000	
2020-543	19	MINI CUBE		16	62.875	18.422	29.30	8.000	118.000	
2020-544	4	SEDIPLAST		559	6.254	2.068	33.07	0.000	12.000	
2020-544	10	SEDI RATE		118	8.435	1.912	22.66	3.000	14.000	
2020-544	11	SEDIMAT-15		68	13.388	7.029	52.50	0	34.000	
2020-544	14	GLOBAL SC.		11	8.727	3.595	41.20	0	20.000	
2020-544	19	MINI CUBE		16	17.867	1.024	5.73	13.000	22.000	
2020-545	4	SEDIPLAST		559	41.833	7.561	18.07	19.000	65.000	
2020-545	10	SEDI RATE		118	41.632	8.311	19.96	17.000	67.000	
2020-545	11	SEDIMAT-15		68	124.818	14.388	11.53	82.000	168.000	
2020-545	14	GLOBAL SC.		11	42.455	8.489	20.00	17.000	68.000	
2020-545	19	MINI CUBE		16	67.500	16.248	24.07	19.000	116.000	

Participant Summary
MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2020521X	1 SYSMEX CBC	1 NEUTROPHIL	20 SYSMEX XT-	16	61.656	1.712	2.78	46.200	77.100				
2020521X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	146	58.483	1.389	2.37	43.900	73.100				
2020521X	1 SYSMEX CBC	2 LYMPHOCYTE	20 SYSMEX XT-	16	23.300	2.211	9.49	17.500	29.100				
2020521X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	145	16.494	1.400	8.49	12.400	20.600				
2020521X	1 SYSMEX CBC	3 MONOCYTE	20 SYSMEX XT-	16	1.519	0.432	28.44	0	3.500				
2020521X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	145	0.664	0.203	30.56	0	2.700				
2020521X	1 SYSMEX CBC	4 EOSINOPHIL	20 SYSMEX XT-	16	13.525	0.785	5.81	10.100	16.900				
2020521X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	143	14.767	0.679	4.60	11.100	18.500				
2020521X	1 SYSMEX CBC	5 BASOPHIL	20 SYSMEX XT-	16	86.644	1.194	1.38	65.000	108.300				
2020521X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	147	9.570	0.888	9.28	7.200	12.000				
2020521X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	154	71.074	1.439	2.02	53.300	88.800				
2020521X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	23	67.796	1.785	2.63	50.800	84.700				
2020521X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	154	18.695	1.221	6.53	14.000	23.400				
2020521X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	23	17.557	1.621	9.23	13.200	21.900				
2020521X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	154	4.831	0.777	16.08	2.800	6.800				
2020521X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	23	5.004	0.696	13.90	3.000	7.000				
2020521X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	154	5.332	0.422	7.92	3.300	7.300				
2020521X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	23	5.452	0.420	7.70	3.500	7.500				
2020521X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	154	84.016	2.476	2.95	63.000	105.000				
2020521X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	23	4.130	0.104	2.52	2.100	6.100				
2020521X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	19	72.884	0.756	1.04	54.700	91.100				
2020521X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	19	13.953	0.485	3.48	10.500	17.400				
2020521X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	19	7.553	0.353	4.67	5.600	9.600				
2020521X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	19	3.511	0.466	13.26	1.500	5.500				
2020521X	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	19	0.553	0.139	25.16	0	2.600				
2020521X	3 ADVIA 120	6 LUC	9 ADVIA-120	19	1.879	0.263	13.98	0	3.900				
2020521X	4 ABBOTT	1 NEUTROPHIL	8 CELL-DYN	130	77.764	0.710	0.91	58.300	97.200				
2020521X	4 ABBOTT	2 LYMPHOCYTE	8 CELL-DYN	129	14.888	1.055	7.09	11.200	18.600				

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2020521X	4	ABBOTT	3	MONOCYTE	8	CELL-DYN	129	3.594	0.583	16.23	1.600	5.600
2020521X	4	ABBOTT	4	EOSINOPHIL	8	CELL-DYN	129	3.443	0.209	6.06	1.400	5.400
2020521X	4	ABBOTT	5	BASOPHIL	8	CELL-DYN	129	0.409	0.250	61.19	0	2.400
2020521X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	46	67.298	0.761	1.13	50.500	84.100
2020521X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	46	21.493	0.852	3.97	16.100	26.900
2020521X	5	BECTON D	3	MONOCYTE	29	B.C600,800	46	6.809	0.772	11.34	4.800	8.800
2020521X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	45	4.389	0.401	9.13	2.400	6.400
2020521X	5	BECTON D	5	BASOPHIL	29	B.C600,800	46	0.067	0.047	69.56	0	2.100
2020521X	6	BC Mindray	1	NEUTROPHIL	32	BC-5390	28	61.385	0.783	1.28	46.000	76.700
2020521X	6	BC Mindray	1	NEUTROPHIL	36	BC DXH 500	17	83.425	1.412	1.69	62.600	104.300
2020521X	6	BC Mindray	2	LYMPHOCYTE	32	BC-5390	28	19.007	0.928	4.88	14.300	23.800
2020521X	6	BC Mindray	2	LYMPHOCYTE	36	BC DXH 500	17	7.800	1.530	19.61	5.800	9.800
2020521X	6	BC Mindray	3	MONOCYTE	32	BC-5390	28	6.850	0.656	9.58	4.900	8.900
2020521X	6	BC Mindray	3	MONOCYTE	36	BC DXH 500	17	0.044	0.061	139.24	0	2.000
2020521X	6	BC Mindray	4	EOSINOPHIL	32	BC-5390	28	12.800	0.484	3.78	9.600	16.000
2020521X	6	BC Mindray	4	EOSINOPHIL	36	BC DXH 500	17	8.656	0.827	9.55	6.500	10.800
2020521X	6	BC Mindray	5	BASOPHIL	32	BC-5390	28	31.961	0.152	0.48	24.000	40.000
2020521X	6	BC Mindray	5	BASOPHIL	36	BC DXH 500	17	0.031	0.077	245.76	0	2.000
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2020522X	1	SYSMEX CBC	1	NEUTROPHIL	20	SYSMEX XT-	16	68.450	0.965	1.41	51.300	85.600
2020522X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	145	59.257	1.054	1.78	44.400	74.100
2020522X	1	SYSMEX CBC	2	LYMPHOCYTE	20	SYSMEX XT-	16	14.406	0.804	5.58	10.800	18.000
2020522X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	146	15.335	0.654	4.26	11.500	19.200
2020522X	1	SYSMEX CBC	3	MONOCYTE	20	SYSMEX XT-	16	1.713	0.267	15.57	0	3.700
2020522X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	146	1.020	0.329	32.29	0	3.000
2020522X	1	SYSMEX CBC	4	EOSINOPHIL	20	SYSMEX XT-	16	15.431	0.539	3.50	11.600	19.300
2020522X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	144	15.275	0.530	3.47	11.500	19.100
2020522X	1	SYSMEX CBC	5	BASOPHIL	20	SYSMEX XT-	16	82.556	1.247	1.51	61.900	103.200
2020522X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	145	9.212	0.756	8.20	6.900	11.500
2020522X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	155	53.789	1.973	3.67	40.300	67.200
2020522X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	23	52.800	2.861	5.42	39.600	66.000

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020522X	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	153	37.093	1.980	5.34	27.800	46.400
2020522X	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	23	35.052	2.661	7.59	26.300	43.800
2020522X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	155	4.885	0.803	16.43	2.900	6.900
2020522X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	23	4.861	0.778	16.01	2.900	6.900
2020522X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	153	4.092	0.491	12.00	2.100	6.100
2020522X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	23	4.174	0.386	9.24	2.200	6.200
2020522X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	154	62.992	2.983	4.74	47.200	78.700
2020522X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	23	3.083	0.096	3.12	1.100	5.100
2020522X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	19	61.705	0.670	1.09	46.300	77.100
2020522X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	19	22.984	0.497	2.16	17.200	28.700
2020522X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	19	7.795	0.363	4.66	5.800	9.800
2020522X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	19	3.626	0.613	16.90	1.600	5.600
2020522X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	19	0.779	0.253	32.42	0	2.800
2020522X	3	ADVIA 120	6 LUC	9	ADVIA-120	19	3.626	0.452	12.46	1.600	5.600
2020522X	4	ABBOTT	1 NEUTROPHIL	8	CELL-DYN	130	52.586	1.440	2.74	39.400	65.700
2020522X	4	ABBOTT	2 LYMPHOCYTE	8	CELL-DYN	129	35.426	2.429	6.86	26.600	44.300
2020522X	4	ABBOTT	3 MONOCYTE	8	CELL-DYN	128	7.029	1.834	26.09	5.000	9.000
2020522X	4	ABBOTT	4 EOSINOPHIL	8	CELL-DYN	130	4.179	0.392	9.37	2.200	6.200
2020522X	4	ABBOTT	5 BASOPHIL	8	CELL-DYN	129	0.750	0.594	79.18	0	2.800
2020522X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	46	51.978	0.712	1.37	39.000	65.000
2020522X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	46	36.756	1.305	3.55	27.600	45.900
2020522X	5	BECTON D	3 MONOCYTE	29	B.C600,800	46	2.826	0.843	29.84	0.800	4.800
2020522X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	46	8.278	0.375	4.53	6.200	10.300
2020522X	5	BECTON D	5 BASOPHIL	29	B.C600,800	46	0.117	0.048	40.91	0	2.100
2020522X	6	BC Mindray	1 NEUTROPHIL	32	BC-5390	28	46.948	0.867	1.85	35.200	58.700
2020522X	6	BC Mindray	1 NEUTROPHIL	36	BC DXH 500	17	74.663	2.613	3.50	56.000	93.300
2020522X	6	BC Mindray	2 LYMPHOCYTE	32	BC-5390	28	38.963	1.488	3.82	29.200	48.700
2020522X	6	BC Mindray	2 LYMPHOCYTE	36	BC DXH 500	17	15.788	2.490	15.77	11.800	19.700
2020522X	6	BC Mindray	3 MONOCYTE	32	BC-5390	28	6.839	1.105	16.15	4.800	8.800
2020522X	6	BC Mindray	3 MONOCYTE	36	BC DXH 500	17	0.081	0.088	108.51	0	2.100
2020522X	6	BC Mindray	4 EOSINOPHIL	32	BC-5390	28	7.170	0.521	7.27	5.200	9.200

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020522X	6 BC Mindray	4 EOSINOPHIL	36 BC DXH 500	17	9.365	1.893	20.21	7.000	11.700		
2020522X	6 BC Mindray	5 BASOPHIL	32 BC-5390	27	29.367	0.456	1.55	22.000	36.700		
2020522X	6 BC Mindray	5 BASOPHIL	36 BC DXH 500	18	0.024	0.042	180.28	0	2.000		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020523X	1 SYSMEX CBC	1 NEUTROPHIL	20 SYSMEX XT-	16	54.825	2.661	4.85	41.100	68.500		
2020523X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	145	52.495	2.682	5.11	39.400	65.600		
2020523X	1 SYSMEX CBC	2 LYMPHOCYTE	20 SYSMEX XT-	16	30.844	3.024	9.80	23.100	38.600		
2020523X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	145	24.154	3.401	14.08	18.100	30.200		
2020523X	1 SYSMEX CBC	3 MONOCYTE	20 SYSMEX XT-	16	1.875	0.282	15.03	0	3.900		
2020523X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	146	0.944	0.241	25.52	0	2.900		
2020523X	1 SYSMEX CBC	4 EOSINOPHIL	20 SYSMEX XT-	16	12.456	0.395	3.17	9.300	15.600		
2020523X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	145	14.001	0.702	5.02	10.500	17.500		
2020523X	1 SYSMEX CBC	5 BASOPHIL	20 SYSMEX XT-	16	75.438	1.451	1.92	56.600	94.300		
2020523X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	145	8.349	0.858	10.28	6.300	10.400		
2020523X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	154	53.935	2.718	5.04	40.500	67.400		
2020523X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	23	53.574	2.805	5.24	40.200	67.000		
2020523X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	154	34.131	2.189	6.41	25.600	42.700		
2020523X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	23	31.800	2.638	8.29	23.900	39.800		
2020523X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	154	4.592	0.845	18.39	2.600	6.600		
2020523X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	23	4.587	0.772	16.83	2.600	6.600		
2020523X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	154	7.281	0.869	11.94	5.300	9.300		
2020523X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	23	7.091	0.697	9.83	5.100	9.100		
2020523X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	154	60.915	2.035	3.34	45.700	76.100		
2020523X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	23	3.000	0.052	1.74	1.000	5.000		
2020523X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	19	43.326	1.129	2.61	32.500	54.200		
2020523X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	19	34.989	1.239	3.54	26.200	43.700		
2020523X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	19	14.442	0.836	5.79	10.800	18.100		
2020523X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	19	1.358	0.221	16.26	0	3.400		
2020523X	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	19	1.050	0.342	32.57	0	3.100		
2020523X	3 ADVIA 120	6 LUC	9 ADVIA-120	19	5.484	0.692	12.62	3.500	7.500		
2020523X	4 ABBOTT	1 NEUTROPHIL	8 CELL-DYN	129	68.393	0.987	1.44	51.300	85.500		

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2020523X	4	ABBOTT	2	LYMPHOCYTE	8	CELL-DYN	129	21.189	1.550	7.31	15.900	26.500
2020523X	4	ABBOTT	3	MONOCYTE	8	CELL-DYN	129	4.931	0.902	18.30	2.900	6.900
2020523X	4	ABBOTT	4	EOSINOPHIL	8	CELL-DYN	130	5.002	0.305	6.09	3.000	7.000
2020523X	4	ABBOTT	5	BASOPHIL	8	CELL-DYN	128	0.569	0.407	71.51	0	2.600
2020523X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	46	59.933	0.723	1.21	44.900	74.900
2020523X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	46	30.891	0.848	2.74	23.200	38.600
2020523X	5	BECTON D	3	MONOCYTE	29	B.C600,800	46	3.948	0.438	11.10	1.900	5.900
2020523X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	46	5.217	0.251	4.82	3.200	7.200
2020523X	5	BECTON D	5	BASOPHIL	29	B.C600,800	46	0.011	0.031	286.36	0	2.000
2020523X	6	BC Mindray	1	NEUTROPHIL	32	BC-5390	28	54.111	0.776	1.43	40.600	67.600
2020523X	6	BC Mindray	1	NEUTROPHIL	36	BC DXH 500	18	79.941	5.260	6.58	60.000	99.900
2020523X	6	BC Mindray	2	LYMPHOCYTE	32	BC-5390	28	28.926	0.982	3.40	21.700	36.200
2020523X	6	BC Mindray	2	LYMPHOCYTE	36	BC DXH 500	16	8.987	1.869	20.79	6.700	11.200
2020523X	6	BC Mindray	3	MONOCYTE	32	BC-5390	28	7.382	1.023	13.86	5.400	9.400
2020523X	6	BC Mindray	3	MONOCYTE	36	BC DXH 500	17	0.069	0.077	111.71	0	2.100
2020523X	6	BC Mindray	4	EOSINOPHIL	32	BC-5390	28	9.715	0.311	3.20	7.300	12.100
2020523X	6	BC Mindray	4	EOSINOPHIL	36	BC DXH 500	17	9.318	1.596	17.13	7.000	11.600
2020523X	6	BC Mindray	5	BASOPHIL	32	BC-5390	28	30.136	0.308	1.02	22.600	37.700
2020523X	6	BC Mindray	5	BASOPHIL	36	BC DXH 500	17	0.012	0.032	273.86	0	2.000
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2020524X	1	SYSMEX CBC	1	NEUTROPHIL	20	SYSMEX XT-	16	67.738	1.267	1.87	50.800	84.700
2020524X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	145	59.326	1.218	2.05	44.500	74.200
2020524X	1	SYSMEX CBC	2	LYMPHOCYTE	20	SYSMEX XT-	16	14.913	0.902	6.05	11.200	18.600
2020524X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	145	15.298	0.967	6.32	11.500	19.100
2020524X	1	SYSMEX CBC	3	MONOCYTE	20	SYSMEX XT-	16	1.594	0.437	27.39	0	3.600
2020524X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	145	0.927	0.213	22.96	0	2.900
2020524X	1	SYSMEX CBC	4	EOSINOPHIL	20	SYSMEX XT-	16	15.431	0.418	2.71	11.600	19.300
2020524X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	145	15.205	0.576	3.79	11.400	19.000
2020524X	1	SYSMEX CBC	5	BASOPHIL	20	SYSMEX XT-	16	82.738	0.805	0.97	62.100	103.400
2020524X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	146	9.195	0.818	8.90	6.900	11.500
2020524X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	154	71.090	1.599	2.25	53.300	88.900

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020524X	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	24	67.883	1.683	2.48	50.900	84.900
2020524X	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	154	18.695	1.123	6.01	14.000	23.400
2020524X	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	23	17.417	1.542	8.85	13.100	21.800
2020524X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	154	4.854	0.783	16.12	2.900	6.900
2020524X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	23	5.074	0.705	13.90	3.100	7.100
2020524X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	154	5.370	0.429	7.98	3.400	7.400
2020524X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	23	5.522	0.409	7.40	3.500	7.500
2020524X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	154	83.633	2.162	2.58	62.700	104.500
2020524X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	22	4.118	0.094	2.27	2.100	6.100
2020524X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	19	72.795	0.520	0.71	54.600	91.000
2020524X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	19	13.947	0.387	2.78	10.500	17.400
2020524X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	19	7.611	0.354	4.65	5.600	9.600
2020524X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	19	3.542	0.513	14.49	1.500	5.500
2020524X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	19	0.595	0.167	28.07	0	2.600
2020524X	3	ADVIA 120	6 LUC	9	ADVIA-120	19	1.921	0.317	16.51	0	3.900
2020524X	4	ABBOTT	1 NEUTROPHIL	8	CELL-DYN	130	52.037	1.590	3.06	39.000	65.000
2020524X	4	ABBOTT	2 LYMPHOCYTE	8	CELL-DYN	129	35.847	2.621	7.31	26.900	44.800
2020524X	4	ABBOTT	3 MONOCYTE	8	CELL-DYN	129	7.068	2.182	30.87	5.100	9.100
2020524X	4	ABBOTT	4 EOSINOPHIL	8	CELL-DYN	129	4.236	0.394	9.29	2.200	6.200
2020524X	4	ABBOTT	5 BASOPHIL	8	CELL-DYN	130	0.728	0.542	74.42	0	2.700
2020524X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	46	51.749	0.880	1.70	38.800	64.700
2020524X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	46	37.454	1.582	4.22	28.100	46.800
2020524X	5	BECTON D	3 MONOCYTE	29	B.C600,800	46	2.539	0.770	30.33	0.500	4.500
2020524X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	46	8.209	0.364	4.43	6.200	10.300
2020524X	5	BECTON D	5 BASOPHIL	29	B.C600,800	46	0.113	0.045	39.60	0	2.100
2020524X	6	BC Mindray	1 NEUTROPHIL	32	BC-5390	28	46.970	1.140	2.43	35.200	58.700
2020524X	6	BC Mindray	1 NEUTROPHIL	36	BC DXH 500	17	75.725	2.373	3.13	56.800	94.700
2020524X	6	BC Mindray	2 LYMPHOCYTE	32	BC-5390	28	38.641	1.693	4.38	29.000	48.300
2020524X	6	BC Mindray	2 LYMPHOCYTE	36	BC DXH 500	17	15.238	2.351	15.43	11.400	19.000
2020524X	6	BC Mindray	3 MONOCYTE	32	BC-5390	28	7.346	1.073	14.60	5.300	9.300
2020524X	6	BC Mindray	3 MONOCYTE	36	BC DXH 500	17	0.056	0.070	125.22	0	2.100

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020524X	6 BC Mindray	4 EOSINOPHIL	32 BC-5390	28	7.114	0.655	9.21	5.100	9.100		
2020524X	6 BC Mindray	4 EOSINOPHIL	36 BC DXH 500	17	8.912	1.657	18.59	6.700	11.100		
2020524X	6 BC Mindray	5 BASOPHIL	32 BC-5390	28	29.189	0.438	1.50	21.900	36.500		
2020524X	6 BC Mindray	5 BASOPHIL	36 BC DXH 500	17	0.035	0.048	135.40	0	2.000		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2020525X	1 SYSMEX CBC	1 NEUTROPHIL	20 SYSMEX XT-	16	61.363	1.637	2.67	46.000	76.700		
2020525X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	145	58.655	1.418	2.42	44.000	73.300		
2020525X	1 SYSMEX CBC	2 LYMPHOCYTE	20 SYSMEX XT-	16	23.313	2.137	9.16	17.500	29.100		
2020525X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	145	16.250	1.363	8.39	12.200	20.300		
2020525X	1 SYSMEX CBC	3 MONOCYTE	20 SYSMEX XT-	16	1.531	0.247	16.12	0	3.500		
2020525X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	144	0.706	0.198	28.03	0	2.700		
2020525X	1 SYSMEX CBC	4 EOSINOPHIL	20 SYSMEX XT-	16	13.794	0.742	5.38	10.300	17.200		
2020525X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	147	14.752	0.835	5.66	11.100	18.400		
2020525X	1 SYSMEX CBC	5 BASOPHIL	20 SYSMEX XT-	16	86.525	1.370	1.58	64.900	108.200		
2020525X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	145	9.519	0.801	8.41	7.100	11.900		
2020525X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	154	54.390	2.721	5.00	40.800	68.000		
2020525X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	23	53.383	2.605	4.88	40.000	66.700		
2020525X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	154	33.690	2.485	7.38	25.300	42.100		
2020525X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	23	31.661	2.766	8.74	23.700	39.600		
2020525X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	154	4.586	0.808	17.62	2.600	6.600		
2020525X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	23	4.587	0.567	12.35	2.600	6.600		
2020525X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	154	7.261	0.701	9.65	5.300	9.300		
2020525X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	23	7.252	0.792	10.92	5.300	9.300		
2020525X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	154	60.922	1.800	2.96	45.700	76.200		
2020525X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	23	3.009	0.060	1.98	1.000	5.000		
2020525X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	19	43.905	1.118	2.55	32.900	54.900		
2020525X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	19	34.521	1.974	5.72	25.900	43.200		
2020525X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	19	14.537	0.838	5.77	10.900	18.200		
2020525X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	19	1.374	0.447	32.54	0	3.400		
2020525X	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	19	0.883	0.195	22.08	0	2.900		
2020525X	3 ADVIA 120	6 LUC	9 ADVIA-120	19	5.337	0.505	9.46	3.300	7.300		

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020525X	4 ABBOTT	1 NEUTROPHIL	8 CELL-DYN	131	77.571	1.830	2.36	58.200	97.000		
2020525X	4 ABBOTT	2 LYMPHOCYTE	8 CELL-DYN	128	14.843	1.051	7.08	11.100	18.600		
2020525X	4 ABBOTT	3 MONOCYTE	8 CELL-DYN	128	3.641	0.523	14.37	1.600	5.600		
2020525X	4 ABBOTT	4 EOSINOPHIL	8 CELL-DYN	130	3.405	0.240	7.06	1.400	5.400		
2020525X	4 ABBOTT	5 BASOPHIL	8 CELL-DYN	128	0.445	0.260	58.34	0	2.400		
2020525X	5 BECTON D	1 NEUTROPHIL	29 B.C600,800	46	67.185	0.808	1.20	50.400	84.000		
2020525X	5 BECTON D	2 LYMPHOCYTE	29 B.C600,800	46	21.413	0.742	3.47	16.100	26.800		
2020525X	5 BECTON D	3 MONOCYTE	29 B.C600,800	46	6.993	0.662	9.47	5.000	9.000		
2020525X	5 BECTON D	4 EOSINOPHIL	29 B.C600,800	46	4.335	0.367	8.47	2.300	6.300		
2020525X	5 BECTON D	5 BASOPHIL	29 B.C600,800	46	0.074	0.044	59.41	0	2.100		
2020525X	6 BC Mindray	1 NEUTROPHIL	32 BC-5390	28	61.493	1.003	1.63	46.100	76.900		
2020525X	6 BC Mindray	1 NEUTROPHIL	36 BC DXH 500	17	83.031	1.261	1.52	62.300	103.800		
2020525X	6 BC Mindray	2 LYMPHOCYTE	32 BC-5390	28	18.637	0.994	5.34	14.000	23.300		
2020525X	6 BC Mindray	2 LYMPHOCYTE	36 BC DXH 500	17	8.075	1.424	17.63	6.100	10.100		
2020525X	6 BC Mindray	3 MONOCYTE	32 BC-5390	28	7.418	1.022	13.77	5.400	9.400		
2020525X	6 BC Mindray	3 MONOCYTE	36 BC DXH 500	17	0.044	0.100	228.12	0	2.000		
2020525X	6 BC Mindray	4 EOSINOPHIL	32 BC-5390	28	12.500	0.508	4.06	9.400	15.600		
2020525X	6 BC Mindray	4 EOSINOPHIL	36 BC DXH 500	17	8.788	0.921	10.48	6.600	11.000		
2020525X	6 BC Mindray	5 BASOPHIL	32 BC-5390	27	32.007	0.174	0.54	24.000	40.000		
2020525X	6 BC Mindray	5 BASOPHIL	36 BC DXH 500	18	0.071	0.190	269.52	0	2.100		

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD		SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
185	Diff 6 Parameter Ins		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25 %	OR 2.00	{GREATER}					
2020521Y	1 6 PAR REG	1	NEUTROPHIL	31 SYSMEX XN	34	62.082	4.091	6.59	46.600	77.600		
2020521Y	1 6 PAR REG	1	NEUTROPHIL	35 XN350/450/	33	58.003	1.298	2.24	43.500	72.500		
2020521Y	1 6 PAR REG	2	LYMPHOCYTE	31 SYSMEX XN	34	16.594	1.228	7.40	12.400	20.700		
2020521Y	1 6 PAR REG	2	LYMPHOCYTE	35 XN350/450/	33	17.178	0.816	4.75	12.900	21.500		
2020521Y	1 6 PAR REG	3	MONOCYTE	31 SYSMEX XN	34	1.297	0.755	58.23	0	3.300		
2020521Y	1 6 PAR REG	3	MONOCYTE	35 XN350/450/	33	0.653	0.962	147.36	0	2.700		
2020521Y	1 6 PAR REG	4	EOSINOPHIL	31 SYSMEX XN	34	15.003	1.811	12.07	11.300	18.800		
2020521Y	1 6 PAR REG	4	EOSINOPHIL	35 XN350/450/	33	15.584	0.715	4.59	11.700	19.500		
2020521Y	1 6 PAR REG	5	BASOPHIL	31 SYSMEX XN	34	5.076	2.227	43.87	3.100	7.100		
2020521Y	1 6 PAR REG	5	BASOPHIL	35 XN350/450/	33	8.816	0.965	10.94	6.600	11.000		
2020521Y	1 6 PAR REG	6	IMMATURE G	31 SYSMEX XN	34	15.141	2.222	14.68	11.400	18.900		
2020521Y	1 6 PAR REG	6	IMMATURE G	35 XN350/450/	33	15.874	0.934	5.88	11.900	19.800		
2020521Y	1 6 PAR REG	7	NRBC	31 SYSMEX XN	21	0.719	0.385	53.54	0	2.700		
2020522Y	1 6 PAR REG	1	NEUTROPHIL	31 SYSMEX XN	34	60.552	1.962	3.24	45.400	75.700		
2020522Y	1 6 PAR REG	1	NEUTROPHIL	35 XN350/450/	33	56.997	1.582	2.78	42.700	71.200		
2020522Y	1 6 PAR REG	2	LYMPHOCYTE	31 SYSMEX XN	34	16.448	1.167	7.10	12.300	20.600		
2020522Y	1 6 PAR REG	2	LYMPHOCYTE	35 XN350/450/	33	17.613	1.413	8.02	13.200	22.000		
2020522Y	1 6 PAR REG	3	MONOCYTE	31 SYSMEX XN	34	1.653	0.524	31.68	0	3.700		
2020522Y	1 6 PAR REG	3	MONOCYTE	35 XN350/450/	33	1.003	0.357	35.64	0	3.000		
2020522Y	1 6 PAR REG	4	EOSINOPHIL	31 SYSMEX XN	34	15.830	0.556	3.51	11.900	19.800		
2020522Y	1 6 PAR REG	4	EOSINOPHIL	35 XN350/450/	33	15.429	0.625	4.05	11.600	19.300		
2020522Y	1 6 PAR REG	5	BASOPHIL	31 SYSMEX XN	34	5.168	1.692	32.75	3.200	7.200		
2020522Y	1 6 PAR REG	5	BASOPHIL	35 XN350/450/	33	8.871	1.123	12.66	6.700	11.100		
2020522Y	1 6 PAR REG	6	IMMATURE G	31 SYSMEX XN	34	15.703	1.012	6.45	11.800	19.600		
2020522Y	1 6 PAR REG	6	IMMATURE G	35 XN350/450/	33	16.042	1.267	7.90	12.000	20.100		
2020522Y	1 6 PAR REG	7	NRBC	31 SYSMEX XN	21	0.081	0.150	185.27	0	2.100		
2020523Y	1 6 PAR REG	1	NEUTROPHIL	31 SYSMEX XN	34	52.706	1.585	3.01	39.500	65.900		

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020523Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	33	50.358	1.292	2.57	37.800	62.900			
2020523Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	34	25.803	1.384	5.36	19.400	32.300			
2020523Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	33	26.990	1.638	6.07	20.200	33.700			
2020523Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	34	1.853	0.521	28.10	0	3.900			
2020523Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	33	0.906	0.587	64.75	0	2.900			
2020523Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	34	14.072	0.267	1.90	10.600	17.600			
2020523Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	33	14.068	0.403	2.86	10.600	17.600			
2020523Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	34	5.163	1.022	19.79	3.200	7.200			
2020523Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	33	7.677	0.815	10.61	5.700	9.700			
2020523Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	34	13.659	0.698	5.11	10.200	17.100			
2020523Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	33	13.435	0.600	4.47	10.100	16.800			
2020523Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.000	0.000	0.00	0	2.000			
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020524Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	34	61.091	2.861	4.68	45.800	76.400			
2020524Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	33	57.368	1.521	2.65	43.000	71.700			
2020524Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	34	16.106	1.404	8.72	12.100	20.100			
2020524Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	33	17.513	1.498	8.56	13.100	21.900			
2020524Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	34	1.679	0.448	26.70	0	3.700			
2020524Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	33	1.034	0.397	38.34	0	3.000			
2020524Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	34	15.700	0.954	6.08	11.800	19.600			
2020524Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	33	15.222	1.401	9.20	11.400	19.000			
2020524Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	34	4.997	2.195	43.93	3.000	7.000			
2020524Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	33	8.594	0.985	11.46	6.400	10.700			
2020524Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	34	15.835	1.101	6.95	11.900	19.800			
2020524Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	33	16.071	0.942	5.86	12.100	20.100			
2020524Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.000	0.000	0.00	0	2.000			
2020525Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	34	62.130	3.665	5.90	46.600	77.700			
2020525Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	33	58.497	1.484	2.54	43.900	73.100			
2020525Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	34	16.342	1.063	6.50	12.300	20.400			

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020525Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	33	17.069	0.613	3.59	12.800	21.300			
2020525Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	1.358	0.767	56.48	0	3.400			
2020525Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	33	0.691	1.047	151.66	0	2.700			
2020525Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	34	15.178	1.858	12.24	11.400	19.000			
2020525Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	33	14.987	0.576	3.85	11.200	18.700			
2020525Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	34	5.074	2.176	42.88	3.100	7.100			
2020525Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	33	9.110	1.053	11.56	6.800	11.400			
2020525Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	34	15.303	2.211	14.45	11.500	19.100			
2020525Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	33	15.977	0.774	4.85	12.000	20.000			
2020525Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.724	0.350	48.41	0	2.700			

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 128 DIFF 3 PARAMETER INS

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
186 DIFF 3 PARAMETER INS		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020521Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	37	61.546	1.004	1.63	46.200	76.900			
2020521Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	48	60.398	1.236	2.05	45.300	75.500			
2020521Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	37	13.839	0.814	5.88	10.400	17.300			
2020521Z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	48	9.920	1.135	11.44	7.400	12.400			
2020521Z	8 3ST	3 Neutrophil	1 KX 21 - Sy	37	24.532	0.824	3.36	18.400	30.700			
2020521Z	8 3ST	3 Neutrophil	3 ABBOTT C-D	48	29.678	1.643	5.54	22.300	37.100			
2020521Z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	14.056	0.759	5.40	10.500	17.600			
2020521Z	9 3D	1 Lymphocyte	4 JT - Coult	13	14.023	1.696	12.10	10.500	17.500			
2020521Z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	4.006	0.580	14.49	2.000	6.000			
2020521Z	9 3D	2 Mix M/E/Ba	4 JT - Coult	11	3.864	0.940	24.34	1.900	5.900			
2020521Z	9 3D	3 Neutrophil	2 AcT Diff 2	16	81.950	0.850	1.04	61.500	102.400			
2020521Z	9 3D	3 Neutrophil	4 JT - Coult	11	81.845	1.193	1.46	61.400	102.300			

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 128 DIFF 3 PARAMETER INS

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
186	DIFF 3 PARAMETER INS	CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25	%	OR 2.00	{GREATER}						
2020522Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	37	12.054	1.338	11.10	9.000	15.100				
2020522Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	48	14.411	1.585	11.00	10.800	18.000				
2020522Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	37	18.441	1.239	6.72	13.800	23.100				
2020522Z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	48	11.578	2.550	22.02	8.700	14.500				
2020522Z	8 3ST	3 Neutrophil	1 KX 21 - Sy	37	69.505	1.705	2.45	52.100	86.900				
2020522Z	8 3ST	3 Neutrophil	3 ABBOTT C-D	48	73.789	3.677	4.98	55.300	92.200				
2020522Z	9 3D	1 Lymphocyte	2 AcT Diff 2	17	58.488	2.774	4.74	43.900	73.100				
2020522Z	9 3D	1 Lymphocyte	4 JT - Coult	13	58.085	5.640	9.71	43.600	72.600				
2020522Z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	9.331	1.743	18.68	7.000	11.700				
2020522Z	9 3D	2 Mix M/E/Ba	4 JT - Coult	11	8.618	3.127	36.28	6.500	10.800				
2020522Z	9 3D	3 Neutrophil	2 AcT Diff 2	16	31.925	2.572	8.06	23.900	39.900				
2020522Z	9 3D	3 Neutrophil	4 JT - Coult	11	31.355	3.335	10.64	23.500	39.200				
2020523Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	38	32.141	0.923	2.87	24.100	40.200				
2020523Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	48	32.043	0.827	2.58	24.000	40.100				
2020523Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	36	15.147	0.966	6.38	11.400	18.900				
2020523Z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	48	9.824	1.622	16.51	7.400	12.300				
2020523Z	8 3ST	3 Neutrophil	1 KX 21 - Sy	37	52.700	1.280	2.43	39.500	65.900				
2020523Z	8 3ST	3 Neutrophil	3 ABBOTT C-D	48	58.133	1.953	3.36	43.600	72.700				
2020523Z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	31.394	0.924	2.94	23.500	39.200				
2020523Z	9 3D	1 Lymphocyte	4 JT - Coult	13	31.354	3.168	10.10	23.500	39.200				
2020523Z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	6.163	0.535	8.68	4.200	8.200				
2020523Z	9 3D	2 Mix M/E/Ba	4 JT - Coult	11	4.509	1.501	33.28	2.500	6.500				
2020523Z	9 3D	3 Neutrophil	2 AcT Diff 2	16	62.431	0.825	1.32	46.800	78.000				
2020523Z	9 3D	3 Neutrophil	4 JT - Coult	11	62.955	1.041	1.65	47.200	78.700				
2020524Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	37	61.573	0.800	1.30	46.200	77.000				
2020524Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	48	60.513	1.247	2.06	45.400	75.600				
2020524Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	37	14.019	0.721	5.14	10.500	17.500				

Participant Summary
MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 128 DIFF 3 PARAMETER INS									
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020524Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	48	9.933	1.207	12.15	7.400	12.400		
2020524Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	37	24.408	0.709	2.91	18.300	30.500		
2020524Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	48	29.550	1.727	5.84	22.200	36.900		
2020524Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	14.273	0.547	3.83	10.700	17.800		
2020524Z	9 3D	1	Lymphocyte	4	JT - Coult	13	14.331	1.376	9.60	10.700	17.900		
2020524Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	4.031	0.536	13.29	2.000	6.000		
2020524Z	9 3D	2	Mix M/E/Ba	4	JT - Coult	11	3.518	0.933	26.52	1.500	5.500		
2020524Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	81.669	0.659	0.81	61.300	102.100		
2020524Z	9 3D	3	Neutrophil	4	JT - Coult	11	81.818	0.874	1.07	61.400	102.300		
186 DIFF 3 PARAMETER INS		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2020525Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	36	32.231	0.752	2.33	24.200	40.300		
2020525Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	47	32.516	1.198	3.68	24.400	40.600		
2020525Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	36	15.017	0.888	5.91	11.300	18.800		
2020525Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	49	9.310	2.318	24.89	7.000	11.600		
2020525Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	36	52.753	0.857	1.62	39.600	65.900		
2020525Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	46	57.870	2.184	3.77	43.400	72.300		
2020525Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	32.394	0.885	2.73	24.300	40.500		
2020525Z	9 3D	1	Lymphocyte	4	JT - Coult	13	30.577	2.910	9.52	22.900	38.200		
2020525Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	5.881	0.979	16.64	3.900	7.900		
2020525Z	9 3D	2	Mix M/E/Ba	4	JT - Coult	11	5.727	0.969	16.91	3.700	7.700		
2020525Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	61.750	1.173	1.90	46.300	77.200		
2020525Z	9 3D	3	Neutrophil	4	JT - Coult	11	62.709	1.527	2.44	47.000	78.400		

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMPL BLOOD COUNT									
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
549 RED CELL COUNT (RBC)		CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %									
2020-521	1 Sysmex CBC	98	SYSMEX K-	12	5.729	0.043	0.76	5.390	6.070				
2020-521	1 Sysmex CBC	127	SYSMEX XT	16	5.652	0.097	1.72	5.310	5.990				

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-521	1	Sysmex CBC	132 SYSMEX XS	148	5.579	0.062	1.11	5.240	5.910
2020-521	1	Sysmex CBC	141 SYSMEX XN	69	5.614	0.063	1.13	5.280	5.950
2020-521	2	ACT PENTRA	125 COULTER AC	148	5.067	0.085	1.68	4.760	5.370
2020-521	2	ACT PENTRA	130 ABX PENTRA	23	5.013	0.073	1.46	4.710	5.310
2020-521	3	ADVIA 120	116 ADVIA-120	19	5.324	0.068	1.28	5.000	5.640
2020-521	4	CELL-DYN	115 ABBOTT C-D	119	5.031	0.135	2.68	4.730	5.330
2020-521	4	CELL-DYN	149 BC DXH 500	13	4.990	0.068	1.36	4.690	5.290
2020-521	5	BECTON D	137 B.C DXH	44	5.376	0.070	1.31	5.050	5.700
2020-521	6	BC Mindray	145 BC-5390	28	5.182	0.082	1.58	4.870	5.490
2020-521	8	3 ST	84 ABBOTT C-D	43	5.773	0.151	2.62	5.430	6.120
2020-521	8	3 ST	98 SYSMEX K-	16	5.758	0.068	1.17	5.410	6.100
2020-521	9	3 D	33 COULTER JT	12	5.756	0.087	1.52	5.410	6.100
2020-521	9	3 D	120 COULTER	16	5.682	0.148	2.60	5.340	6.020
549 RED CELL COUNT (RBC)		CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %					
2020-522	1	Sysmex CBC	98 SYSMEX K-	12	2.498	0.028	1.13	2.350	2.650
2020-522	1	Sysmex CBC	127 SYSMEX XT	16	2.435	0.031	1.27	2.290	2.580
2020-522	1	Sysmex CBC	132 SYSMEX XS	148	2.350	0.028	1.18	2.210	2.490
2020-522	1	Sysmex CBC	141 SYSMEX XN	69	2.323	0.035	1.52	2.180	2.460
2020-522	2	ACT PENTRA	125 COULTER AC	148	4.647	0.077	1.67	4.370	4.930
2020-522	2	ACT PENTRA	130 ABX PENTRA	23	4.606	0.066	1.43	4.330	4.880
2020-522	3	ADVIA 120	116 ADVIA-120	19	4.563	0.099	2.16	4.290	4.840
2020-522	4	CELL-DYN	115 ABBOTT C-D	119	2.239	0.061	2.73	2.100	2.370
2020-522	4	CELL-DYN	149 BC DXH 500	13	2.232	0.045	2.04	2.100	2.370
2020-522	5	BECTON D	137 B.C DXH	44	2.567	0.042	1.63	2.410	2.720
2020-522	6	BC Mindray	145 BC-5390	28	2.266	0.033	1.45	2.130	2.400
2020-522	8	3 ST	84 ABBOTT C-D	43	2.557	0.083	3.25	2.400	2.710
2020-522	8	3 ST	98 SYSMEX K-	16	2.510	0.032	1.29	2.360	2.660
2020-522	9	3 D	33 COULTER JT	12	2.359	0.053	2.26	2.220	2.500
2020-522	9	3 D	120 COULTER	16	2.390	0.087	3.65	2.250	2.530

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %							
2020-523	1 Sysmex CBC	98 SYSMEX K-		12	4.142	0.052	1.25	3.890	4.390	
2020-523	1 Sysmex CBC	127 SYSMEX XT		16	4.593	0.054	1.17	4.320	4.870	
2020-523	1 Sysmex CBC	132 SYSMEX XS		148	4.508	0.048	1.06	4.240	4.780	
2020-523	1 Sysmex CBC	141 SYSMEX XN		69	4.524	0.051	1.14	4.250	4.800	
2020-523	2 ACT PENTRA	125 COULTER AC		148	2.392	0.047	1.96	2.250	2.540	
2020-523	2 ACT PENTRA	130 ABX PENTRA		23	2.385	0.041	1.72	2.240	2.530	
2020-523	3 ADVIA 120	116 ADVIA-120		19	2.362	0.053	2.23	2.220	2.500	
2020-523	4 CELL-DYN	115 ABBOTT C-D		119	4.622	0.131	2.84	4.340	4.900	
2020-523	4 CELL-DYN	149 BC DXH 500		13	4.557	0.075	1.65	4.280	4.830	
2020-523	5 BECTON D	137 B.C DXH		44	4.107	0.065	1.57	3.860	4.350	
2020-523	6 BC Mindray	145 BC-5390		28	4.419	0.066	1.50	4.150	4.680	
2020-523	8 3 ST	84 ABBOTT C-D		43	4.255	0.101	2.36	4.000	4.510	
2020-523	8 3 ST	98 SYSMEX K-		16	4.169	0.056	1.34	3.920	4.420	
2020-523	9 3 D	33 COULTER JT		12	4.769	0.079	1.66	4.480	5.060	
2020-523	9 3 D	120 COULTER		16	4.699	0.090	1.92	4.420	4.980	
2020-524	1 Sysmex CBC	98 SYSMEX K-		12	5.687	0.035	0.62	5.350	6.030	
2020-524	1 Sysmex CBC	127 SYSMEX XT		16	2.434	0.027	1.11	2.290	2.580	
2020-524	1 Sysmex CBC	132 SYSMEX XS		148	2.349	0.030	1.26	2.210	2.490	
2020-524	1 Sysmex CBC	141 SYSMEX XN		69	2.318	0.031	1.33	2.180	2.460	
2020-524	2 ACT PENTRA	125 COULTER AC		148	5.053	0.087	1.72	4.750	5.360	
2020-524	2 ACT PENTRA	130 ABX PENTRA		23	5.005	0.058	1.16	4.700	5.310	
2020-524	3 ADVIA 120	116 ADVIA-120		19	5.252	0.098	1.87	4.940	5.570	
2020-524	4 CELL-DYN	115 ABBOTT C-D		119	2.237	0.058	2.58	2.100	2.370	
2020-524	4 CELL-DYN	149 BC DXH 500		13	2.222	0.035	1.58	2.090	2.360	
2020-524	5 BECTON D	137 B.C DXH		44	2.568	0.039	1.54	2.410	2.720	
2020-524	6 BC Mindray	145 BC-5390		28	2.266	0.039	1.74	2.130	2.400	
2020-524	8 3 ST	84 ABBOTT C-D		43	5.755	0.145	2.52	5.410	6.100	
2020-524	8 3 ST	98 SYSMEX K-		16	5.758	0.066	1.15	5.410	6.100	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD		SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-524	9	3 D	33	COULTER JT	12	5.727	0.084	1.47	5.380	6.070	
2020-524	9	3 D	120	COULTER	16	5.689	0.076	1.34	5.350	6.030	
549	RED CELL COUNT (RBC)		CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %							
2020-525	1	Sysmex CBC	98	SYSMEX K-	12	4.153	0.066	1.58	3.900	4.400	
2020-525	1	Sysmex CBC	127	SYSMEX XT	16	5.644	0.087	1.53	5.310	5.980	
2020-525	1	Sysmex CBC	132	SYSMEX XS	148	5.590	0.055	0.98	5.260	5.930	
2020-525	1	Sysmex CBC	141	SYSMEX XN	69	5.622	0.055	0.98	5.290	5.960	
2020-525	2	ACT PENTRA	125	COULTER AC	148	2.385	0.069	2.90	2.240	2.530	
2020-525	2	ACT PENTRA	130	ABX PENTRA	23	2.380	0.033	1.40	2.240	2.520	
2020-525	3	ADVIA 120	116	ADVIA-120	19	2.360	0.052	2.18	2.220	2.500	
2020-525	4	CELL-DYN	115	ABBOTT C-D	119	5.022	0.147	2.93	4.720	5.320	
2020-525	4	CELL-DYN	149	BC DXH 500	13	4.955	0.057	1.15	4.660	5.250	
2020-525	5	BECTON D	137	B.C DXH	44	5.344	0.063	1.18	5.020	5.660	
2020-525	6	BC Mindray	145	BC-5390	28	5.184	0.084	1.61	4.870	5.500	
2020-525	8	3 ST	84	ABBOTT C-D	43	4.271	0.122	2.86	4.020	4.530	
2020-525	8	3 ST	98	SYSMEX K-	16	4.169	0.041	0.99	3.920	4.420	
2020-525	9	3 D	33	COULTER JT	12	4.812	0.068	1.41	4.520	5.100	
2020-525	9	3 D	120	COULTER	16	4.730	0.085	1.79	4.450	5.010	
550	HEMATOCRIT (HCT)		CLIA TEST ID #: 785								
2020-521	1	Sysmex CBC	98	SYSMEX K-	12	49.258	0.734	1.49	46.300	52.200	
2020-521	1	Sysmex CBC	127	SYSMEX XT	16	52.213	1.148	2.20	49.100	55.300	
2020-521	1	Sysmex CBC	132	SYSMEX XS	147	51.846	1.037	2.00	48.700	55.000	
2020-521	1	Sysmex CBC	141	SYSMEX XN	70	51.039	1.025	2.01	48.000	54.100	
2020-521	2	ACT PENTRA	125	COULTER AC	147	44.375	0.793	1.79	41.700	47.000	
2020-521	2	ACT PENTRA	130	ABX PENTRA	23	44.448	0.731	1.65	41.800	47.100	
2020-521	3	ADVIA 120	116	ADVIA-120	19	46.878	0.729	1.56	44.100	49.700	
2020-521	4	CELL-DYN	115	ABBOTT C-D	119	43.218	1.116	2.58	40.600	45.800	
2020-521	4	CELL-DYN	149	BC DXH 500	12	47.673	1.028	2.16	44.800	50.500	
2020-521	5	BECTON D	137	B.C DXH	44	55.881	0.648	1.16	52.500	59.200	
2020-521	6	BC Mindray	145	BC-5390	28	48.482	1.158	2.39	45.600	51.400	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-521	8 3 ST	84 ABBOTT C-D			43	52.393	1.724	3.29	49.200	55.500	
2020-521	8 3 ST	98 SYSMEX K-			16	49.669	0.924	1.86	46.700	52.600	
2020-521	9 3 D	33 COULTER JT			12	53.233	0.918	1.72	50.000	56.400	
2020-521	9 3 D	120 COULTER			16	53.406	1.535	2.87	50.200	56.600	
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %							
2020-522	1 Sysmex CBC	98 SYSMEX K-			12	17.533	0.332	1.90	16.500	18.600	
2020-522	1 Sysmex CBC	127 SYSMEX XT			16	18.469	0.335	1.81	17.400	19.600	
2020-522	1 Sysmex CBC	132 SYSMEX XS			147	17.855	0.340	1.91	16.800	18.900	
2020-522	1 Sysmex CBC	141 SYSMEX XN			70	17.387	0.385	2.21	16.300	18.400	
2020-522	2 ACT PENTRA	125 COULTER AC			147	38.203	0.721	1.89	35.900	40.500	
2020-522	2 ACT PENTRA	130 ABX PENTRA			23	38.257	0.836	2.18	36.000	40.600	
2020-522	3 ADVIA 120	116 ADVIA-120			19	35.600	0.704	1.98	33.500	37.700	
2020-522	4 CELL-DYN	115 ABBOTT C-D			119	15.604	0.403	2.59	14.700	16.500	
2020-522	4 CELL-DYN	149 BC DXH 500			12	16.891	0.396	2.35	15.900	17.900	
2020-522	5 BECTON D	137 B.C DXH			44	20.337	0.333	1.64	19.100	21.600	
2020-522	6 BC Mindray	145 BC-5390			28	17.936	0.473	2.64	16.900	19.000	
2020-522	8 3 ST	84 ABBOTT C-D			43	18.440	0.664	3.60	17.300	19.500	
2020-522	8 3 ST	98 SYSMEX K-			16	17.656	0.320	1.81	16.600	18.700	
2020-522	9 3 D	33 COULTER JT			12	17.483	0.529	3.03	16.400	18.500	
2020-522	9 3 D	120 COULTER			16	17.975	0.750	4.17	16.900	19.100	
2020-523	1 Sysmex CBC	98 SYSMEX K-			12	32.255	0.577	1.79	30.300	34.200	
2020-523	1 Sysmex CBC	127 SYSMEX XT			16	38.956	0.857	2.20	36.600	41.300	
2020-523	1 Sysmex CBC	132 SYSMEX XS			147	38.452	0.699	1.82	36.100	40.800	
2020-523	1 Sysmex CBC	141 SYSMEX XN			70	38.093	0.777	2.04	35.800	40.400	
2020-523	2 ACT PENTRA	125 COULTER AC			147	19.018	0.403	2.12	17.900	20.200	
2020-523	2 ACT PENTRA	130 ABX PENTRA			23	19.013	0.445	2.34	17.900	20.200	
2020-523	3 ADVIA 120	116 ADVIA-120			19	16.184	0.351	2.17	15.200	17.200	
2020-523	4 CELL-DYN	115 ABBOTT C-D			119	37.195	1.025	2.75	35.000	39.400	
2020-523	4 CELL-DYN	149 BC DXH 500			12	39.264	0.796	2.03	36.900	41.600	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-523	5	BECTON D	137	B.C DXH	44	38.235	0.528	1.38	35.900	40.500	
2020-523	6	BC Mindray	145	BC-5390	28	38.379	0.901	2.35	36.100	40.700	
2020-523	8	3 ST	84	ABBOTT C-D	43	35.181	0.999	2.84	33.100	37.300	
2020-523	8	3 ST	98	SYSMEX K-	16	32.594	0.589	1.81	30.600	34.500	
2020-523	9	3 D	33	COULTER JT	12	39.875	0.788	1.98	37.500	42.300	
2020-523	9	3 D	120	COULTER	16	39.688	0.852	2.15	37.300	42.100	
550 HEMATOCRIT (HCT)			CLIA TEST ID #: 785			EVALUATION CRITERIA: TARGET VALUE 6 %					
2020-524	1	Sysmex CBC	98	SYSMEX K-	12	48.864	0.809	1.66	45.900	51.800	
2020-524	1	Sysmex CBC	127	SYSMEX XT	16	18.419	0.272	1.48	17.300	19.500	
2020-524	1	Sysmex CBC	132	SYSMEX XS	147	17.818	0.352	1.98	16.700	18.900	
2020-524	1	Sysmex CBC	141	SYSMEX XN	70	17.359	0.374	2.16	16.300	18.400	
2020-524	2	ACT PENTRA	125	COULTER AC	147	44.287	0.785	1.77	41.600	46.900	
2020-524	2	ACT PENTRA	130	ABX PENTRA	23	44.461	0.732	1.65	41.800	47.100	
2020-524	3	ADVIA 120	116	ADVIA-120	19	46.179	1.081	2.34	43.400	48.900	
2020-524	4	CELL-DYN	115	ABBOTT C-D	119	15.602	0.370	2.37	14.700	16.500	
2020-524	4	CELL-DYN	149	BC DXH 500	12	16.855	0.268	1.59	15.800	17.900	
2020-524	5	BECTON D	137	B.C DXH	44	20.351	0.329	1.62	19.100	21.600	
2020-524	6	BC Mindray	145	BC-5390	28	17.932	0.468	2.61	16.900	19.000	
2020-524	8	3 ST	84	ABBOTT C-D	43	52.414	1.552	2.96	49.300	55.600	
2020-524	8	3 ST	98	SYSMEX K-	16	49.594	0.754	1.52	46.600	52.600	
2020-524	9	3 D	33	COULTER JT	12	52.950	0.865	1.63	49.800	56.100	
2020-524	9	3 D	120	COULTER	16	53.563	1.083	2.02	50.300	56.800	
2020-525	1	Sysmex CBC	98	SYSMEX K-	12	32.318	0.539	1.67	30.400	34.300	
2020-525	1	Sysmex CBC	127	SYSMEX XT	16	52.056	1.019	1.96	48.900	55.200	
2020-525	1	Sysmex CBC	132	SYSMEX XS	147	51.863	0.979	1.89	48.800	55.000	
2020-525	1	Sysmex CBC	141	SYSMEX XN	70	51.156	0.961	1.88	48.100	54.200	
2020-525	2	ACT PENTRA	125	COULTER AC	147	18.988	0.529	2.79	17.800	20.100	
2020-525	2	ACT PENTRA	130	ABX PENTRA	23	19.035	0.394	2.07	17.900	20.200	
2020-525	3	ADVIA 120	116	ADVIA-120	19	16.200	0.383	2.36	15.200	17.200	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-525	4	CELL-DYN	115 ABBOTT C-D	119	43.344	1.250	2.88	40.700	45.900		
2020-525	4	CELL-DYN	149 BC DXH 500	12	47.327	0.930	1.97	44.500	50.200		
2020-525	5	BECTON D	137 B.C DXH	44	55.716	0.669	1.20	52.400	59.100		
2020-525	6	BC Mindray	145 BC-5390	28	48.400	1.165	2.41	45.500	51.300		
2020-525	8	3 ST	84 ABBOTT C-D	43	35.295	1.147	3.25	33.200	37.400		
2020-525	8	3 ST	98 SYSMEX K-	16	32.631	0.503	1.54	30.700	34.600		
2020-525	9	3 D	33 COULTER JT	12	40.092	0.649	1.62	37.700	42.500		
2020-525	9	3 D	120 COULTER	16	40.131	1.040	2.59	37.700	42.500		
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}								
2020-521	1	Sysmex CBC	98 SYSMEX K-	11	17.964	0.235	1.31	16.700	19.200		
2020-521	1	Sysmex CBC	127 SYSMEX XT	16	17.456	0.132	0.76	16.200	18.700		
2020-521	1	Sysmex CBC	132 SYSMEX XS	148	17.539	0.188	1.07	16.300	18.800		
2020-521	1	Sysmex CBC	141 SYSMEX XN	69	17.526	0.172	0.98	16.300	18.800		
2020-521	2	ACT PENTRA	125 COULTER AC	149	16.025	0.223	1.39	14.900	17.100		
2020-521	2	ACT PENTRA	130 ABX PENTRA	23	16.141	0.123	0.76	15.000	17.300		
2020-521	3	ADVIA 120	116 ADVIA-120	19	17.174	0.245	1.42	16.000	18.400		
2020-521	4	CELL-DYN	115 ABBOTT C-D	119	16.451	0.285	1.73	15.300	17.600		
2020-521	4	CELL-DYN	149 BC DXH 500	12	15.218	0.330	2.17	14.200	16.300		
2020-521	5	BECTON D	137 B.C DXH	44	17.130	0.179	1.04	15.900	18.300		
2020-521	6	BC Mindray	145 BC-5390	28	16.696	0.215	1.29	15.500	17.900		
2020-521	8	3 ST	84 ABBOTT C-D	43	18.143	0.379	2.09	16.900	19.400		
2020-521	8	3 ST	98 SYSMEX K-	16	18.025	0.273	1.51	16.800	19.300		
2020-521	9	3 D	33 COULTER JT	12	18.600	0.381	2.05	17.300	19.900		
2020-521	9	3 D	120 COULTER	16	18.100	0.442	2.44	16.800	19.400		
2020-522	1	Sysmex CBC	98 SYSMEX K-	11	6.145	0.123	2.01	5.700	6.600		
2020-522	1	Sysmex CBC	127 SYSMEX XT	16	5.988	0.060	1.00	5.600	6.400		
2020-522	1	Sysmex CBC	132 SYSMEX XS	148	5.860	0.076	1.29	5.400	6.300		
2020-522	1	Sysmex CBC	141 SYSMEX XN	69	5.947	0.067	1.13	5.500	6.400		
2020-522	2	ACT PENTRA	125 COULTER AC	149	13.838	0.208	1.51	12.900	14.800		

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-522	2	ACT PENTRA	130 ABX PENTRA	23	13.822	0.174	1.26	12.900	14.800
2020-522	3	ADVIA 120	116 ADVIA-120	19	13.122	0.178	1.36	12.200	14.000
2020-522	4	CELL-DYN	115 ABBOTT C-D	119	5.705	0.155	2.73	5.300	6.100
2020-522	4	CELL-DYN	149 BC DXH 500	12	4.627	0.114	2.45	4.300	5.000
2020-522	5	BECTON D	137 B.C DXH	44	6.344	0.082	1.29	5.900	6.800
2020-522	6	BC Mindray	145 BC-5390	28	6.000	0.076	1.26	5.600	6.400
2020-522	8	3 ST	84 ABBOTT C-D	43	6.421	0.156	2.44	6.000	6.900
2020-522	8	3 ST	98 SYSMEX K-	16	6.125	0.097	1.58	5.700	6.600
2020-522	9	3 D	33 COULTER JT	12	5.708	0.180	3.15	5.300	6.100
2020-522	9	3 D	120 COULTER	16	5.881	0.155	2.64	5.500	6.300
551 HEMOGLOBIN (HGB)			CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}				
2020-523	1	Sysmex CBC	98 SYSMEX K-	11	11.991	0.138	1.15	11.200	12.800
2020-523	1	Sysmex CBC	127 SYSMEX XT	16	13.144	0.141	1.07	12.200	14.100
2020-523	1	Sysmex CBC	132 SYSMEX XS	148	13.095	0.142	1.09	12.200	14.000
2020-523	1	Sysmex CBC	141 SYSMEX XN	69	13.181	0.138	1.04	12.300	14.100
2020-523	2	ACT PENTRA	125 COULTER AC	149	6.721	0.127	1.89	6.300	7.200
2020-523	2	ACT PENTRA	130 ABX PENTRA	23	6.587	0.095	1.44	6.100	7.000
2020-523	3	ADVIA 120	116 ADVIA-120	19	5.879	0.077	1.30	5.500	6.300
2020-523	4	CELL-DYN	115 ABBOTT C-D	119	13.252	0.224	1.69	12.300	14.200
2020-523	4	CELL-DYN	149 BC DXH 500	12	11.609	0.243	2.09	10.800	12.400
2020-523	5	BECTON D	137 B.C DXH	44	11.916	0.158	1.33	11.100	12.800
2020-523	6	BC Mindray	145 BC-5390	28	13.118	0.160	1.22	12.200	14.000
2020-523	8	3 ST	84 ABBOTT C-D	43	12.179	0.246	2.02	11.300	13.000
2020-523	8	3 ST	98 SYSMEX K-	16	11.963	0.141	1.18	11.100	12.800
2020-523	9	3 D	33 COULTER JT	12	13.792	0.287	2.08	12.800	14.800
2020-523	9	3 D	120 COULTER	16	13.569	0.208	1.54	12.600	14.500
2020-524	1	Sysmex CBC	98 SYSMEX K-	11	17.945	0.192	1.07	16.700	19.200
2020-524	1	Sysmex CBC	127 SYSMEX XT	16	5.994	0.066	1.10	5.600	6.400
2020-524	1	Sysmex CBC	132 SYSMEX XS	148	5.847	0.083	1.41	5.400	6.300

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-524	1	Sysmex CBC	141	SYSMEX XN	69	5.936	0.070	1.18	5.500	6.400	
2020-524	2	ACT PENTRA	125	COULTER AC	149	16.018	0.232	1.45	14.900	17.100	
2020-524	2	ACT PENTRA	130	ABX PENTRA	23	16.155	0.150	0.93	15.000	17.300	
2020-524	3	ADVIA 120	116	ADVIA-120	19	17.068	0.297	1.74	15.900	18.300	
2020-524	4	CELL-DYN	115	ABBOTT C-D	119	5.697	0.177	3.10	5.300	6.100	
2020-524	4	CELL-DYN	149	BC DXH 500	12	4.618	0.083	1.80	4.300	4.900	
2020-524	5	BECTON D	137	B.C DXH	44	6.347	0.087	1.37	5.900	6.800	
2020-524	6	BC Mindray	145	BC-5390	28	6.032	0.104	1.72	5.600	6.500	
2020-524	8	3 ST	84	ABBOTT C-D	43	18.169	0.382	2.10	16.900	19.400	
2020-524	8	3 ST	98	SYSMEX K-	16	17.981	0.178	0.99	16.700	19.200	
2020-524	9	3 D	33	COULTER JT	12	18.625	0.327	1.76	17.300	19.900	
2020-524	9	3 D	120	COULTER	16	18.250	0.296	1.62	17.000	19.500	
551	HEMOGLOBIN (HGB)		CLIA	TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}						
2020-525	1	Sysmex CBC	98	SYSMEX K-	11	11.982	0.175	1.46	11.100	12.800	
2020-525	1	Sysmex CBC	127	SYSMEX XT	16	17.413	0.136	0.78	16.200	18.600	
2020-525	1	Sysmex CBC	132	SYSMEX XS	148	17.557	0.167	0.95	16.300	18.800	
2020-525	1	Sysmex CBC	141	SYSMEX XN	69	17.559	0.161	0.92	16.300	18.800	
2020-525	2	ACT PENTRA	125	COULTER AC	149	6.706	0.161	2.40	6.200	7.200	
2020-525	2	ACT PENTRA	130	ABX PENTRA	23	6.609	0.102	1.54	6.100	7.100	
2020-525	3	ADVIA 120	116	ADVIA-120	19	5.911	0.099	1.68	5.500	6.300	
2020-525	4	CELL-DYN	115	ABBOTT C-D	119	16.331	0.299	1.83	15.200	17.500	
2020-525	4	CELL-DYN	149	BC DXH 500	12	15.055	0.246	1.64	14.000	16.100	
2020-525	5	BECTON D	137	B.C DXH	44	17.100	0.174	1.02	15.900	18.300	
2020-525	6	BC Mindray	145	BC-5390	28	16.644	0.164	0.99	15.500	17.800	
2020-525	8	3 ST	84	ABBOTT C-D	43	12.228	0.249	2.04	11.400	13.100	
2020-525	8	3 ST	98	SYSMEX K-	16	11.981	0.107	0.90	11.100	12.800	
2020-525	9	3 D	33	COULTER JT	12	13.758	0.266	1.93	12.800	14.700	
2020-525	9	3 D	120	COULTER	16	13.581	0.227	1.67	12.600	14.500	
552	WHITE CELL COUNT(WBC)		CLIA	TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %						
2020-521	1	Sysmex CBC	98	SYSMEX K-	12	19.933	0.352	1.77	16.900	22.900	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-521	1	Sysmex CBC	127	SYSMEX XT	16	19.113	0.765	4.00	16.200	22.000	
2020-521	1	Sysmex CBC	132	SYSMEX XS	146	18.894	0.448	2.37	16.100	21.700	
2020-521	1	Sysmex CBC	141	SYSMEX XN	70	19.087	0.515	2.70	16.200	22.000	
2020-521	2	ACT PENTRA	125	COULTER AC	149	18.411	0.475	2.58	15.600	21.200	
2020-521	2	ACT PENTRA	130	ABX PENTRA	23	18.100	0.696	3.85	15.400	20.800	
2020-521	3	ADVIA 120	116	ADVIA-120	19	16.868	0.891	5.28	14.300	19.400	
2020-521	4	CELL-DYN	115	ABBOTT C-D	119	19.637	0.532	2.71	16.700	22.600	
2020-521	4	CELL-DYN	149	BC DXH 500	12	18.236	0.352	1.93	15.500	21.000	
2020-521	5	BECTON D	137	B.C DXH	44	21.244	0.473	2.23	18.100	24.400	
2020-521	6	BC Mindray	145	BC-5390	28	18.850	0.507	2.69	16.000	21.700	
2020-521	8	3 ST	84	ABBOTT C-D	43	21.902	1.167	5.33	18.600	25.200	
2020-521	8	3 ST	98	SYSMEX K-	16	20.094	0.458	2.28	17.100	23.100	
2020-521	9	3 D	33	COULTER JT	12	22.892	0.924	4.04	19.500	26.300	
2020-521	9	3 D	120	COULTER	16	21.206	0.672	3.17	18.000	24.400	
552 WHITE CELL COUNT(WBC		CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %								
2020-522	1	Sysmex CBC	98	SYSMEX K-	12	2.733	0.075	2.73	2.300	3.100	
2020-522	1	Sysmex CBC	127	SYSMEX XT	16	3.800	0.122	3.22	3.200	4.400	
2020-522	1	Sysmex CBC	132	SYSMEX XS	146	3.770	0.098	2.61	3.200	4.300	
2020-522	1	Sysmex CBC	141	SYSMEX XN	70	3.278	0.414	12.62	2.800	3.800	
2020-522	2	ACT PENTRA	125	COULTER AC	149	7.742	0.240	3.09	6.600	8.900	
2020-522	2	ACT PENTRA	130	ABX PENTRA	23	7.678	0.304	3.95	6.500	8.800	
2020-522	3	ADVIA 120	116	ADVIA-120	19	7.758	0.396	5.11	6.600	8.900	
2020-522	4	CELL-DYN	115	ABBOTT C-D	119	2.802	0.135	4.83	2.400	3.200	
2020-522	4	CELL-DYN	149	BC DXH 500	12	2.164	0.077	3.57	1.800	2.500	
2020-522	5	BECTON D	137	B.C DXH	44	3.931	0.134	3.40	3.300	4.500	
2020-522	6	BC Mindray	145	BC-5390	28	3.529	0.096	2.72	3.000	4.100	
2020-522	8	3 ST	84	ABBOTT C-D	43	3.174	0.193	6.08	2.700	3.700	
2020-522	8	3 ST	98	SYSMEX K-	16	2.706	0.056	2.05	2.300	3.100	
2020-522	9	3 D	33	COULTER JT	12	2.083	0.191	9.16	1.800	2.400	
2020-522	9	3 D	120	COULTER	16	2.231	0.223	9.99	1.900	2.600	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
552	WHITE CELL COUNT(WBC	CLIA TEST ID #:	805	EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-523	1	Sysmex CBC	98	SYSMEX K-	12	8.025	0.159	1.98	6.800	9.200	
2020-523	1	Sysmex CBC	127	SYSMEX XT	16	7.431	0.259	3.49	6.300	8.500	
2020-523	1	Sysmex CBC	132	SYSMEX XS	146	7.082	0.504	7.12	6.000	8.100	
2020-523	1	Sysmex CBC	141	SYSMEX XN	70	6.841	0.469	6.86	5.800	7.900	
2020-523	2	ACT PENTRA	125	COULTER AC	149	2.435	0.125	5.14	2.100	2.800	
2020-523	2	ACT PENTRA	130	ABX PENTRA	23	2.439	0.113	4.64	2.100	2.800	
2020-523	3	ADVIA 120	116	ADVIA-120	19	3.132	0.138	4.40	2.700	3.600	
2020-523	4	CELL-DYN	115	ABBOTT C-D	119	7.094	0.177	2.50	6.000	8.200	
2020-523	4	CELL-DYN	149	BC DXH 500	12	6.436	0.155	2.41	5.500	7.400	
2020-523	5	BECTON D	137	B.C DXH	44	9.574	0.200	2.09	8.100	11.000	
2020-523	6	BC Mindray	145	BC-5390	28	8.200	0.248	3.02	7.000	9.400	
2020-523	8	3 ST	84	ABBOTT C-D	43	8.756	0.464	5.30	7.400	10.100	
2020-523	8	3 ST	98	SYSMEX K-	16	8.050	0.132	1.64	6.800	9.300	
2020-523	9	3 D	33	COULTER JT	12	8.467	0.361	4.27	7.200	9.700	
2020-523	9	3 D	120	COULTER	16	8.213	0.257	3.13	7.000	9.400	
2020-524	1	Sysmex CBC	98	SYSMEX K-	12	19.942	0.340	1.71	17.000	22.900	
2020-524	1	Sysmex CBC	127	SYSMEX XT	16	3.775	0.148	3.92	3.200	4.300	
2020-524	1	Sysmex CBC	132	SYSMEX XS	146	3.750	0.100	2.68	3.200	4.300	
2020-524	1	Sysmex CBC	141	SYSMEX XN	70	3.267	0.381	11.67	2.800	3.800	
2020-524	2	ACT PENTRA	125	COULTER AC	149	18.380	0.547	2.98	15.600	21.100	
2020-524	2	ACT PENTRA	130	ABX PENTRA	23	18.213	0.713	3.91	15.500	20.900	
2020-524	3	ADVIA 120	116	ADVIA-120	19	16.995	0.911	5.36	14.400	19.500	
2020-524	4	CELL-DYN	115	ABBOTT C-D	119	2.768	0.101	3.64	2.400	3.200	
2020-524	4	CELL-DYN	149	BC DXH 500	12	2.173	0.086	3.97	1.800	2.500	
2020-524	5	BECTON D	137	B.C DXH	44	3.921	0.141	3.59	3.300	4.500	
2020-524	6	BC Mindray	145	BC-5390	28	3.550	0.138	3.88	3.000	4.100	
2020-524	8	3 ST	84	ABBOTT C-D	43	21.819	1.197	5.49	18.500	25.100	
2020-524	8	3 ST	98	SYSMEX K-	16	20.100	0.294	1.46	17.100	23.100	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-524	9 3 D	33 COULTER JT		12	22.700	0.903	3.98	19.300	26.100
2020-524	9 3 D	120 COULTER		16	21.344	0.633	2.97	18.100	24.500
552 WHITE CELL COUNT(WBC)				CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %				
2020-525	1 Sysmex CBC	98 SYSMEX K-		12	8.092	0.138	1.71	6.900	9.300
2020-525	1 Sysmex CBC	127 SYSMEX XT		16	19.081	0.641	3.36	16.200	21.900
2020-525	1 Sysmex CBC	132 SYSMEX XS		146	18.863	0.422	2.24	16.000	21.700
2020-525	1 Sysmex CBC	141 SYSMEX XN		70	19.049	0.462	2.43	16.200	21.900
2020-525	2 ACT PENTRA	125 COULTER AC		149	2.438	0.122	4.99	2.100	2.800
2020-525	2 ACT PENTRA	130 ABX PENTRA		23	2.413	0.090	3.73	2.100	2.800
2020-525	3 ADVIA 120	116 ADVIA-120		19	3.105	0.115	3.69	2.600	3.600
2020-525	4 CELL-DYN	115 ABBOTT C-D		119	19.482	0.521	2.67	16.600	22.400
2020-525	4 CELL-DYN	149 BC DXH 500		12	18.045	0.309	1.71	15.300	20.800
2020-525	5 BECTON D	137 B.C DXH		44	21.240	0.444	2.09	18.100	24.400
2020-525	6 BC Mindray	145 BC-5390		28	18.814	0.499	2.65	16.000	21.600
2020-525	8 3 ST	84 ABBOTT C-D		43	8.947	0.490	5.48	7.600	10.300
2020-525	8 3 ST	98 SYSMEX K-		16	8.006	0.175	2.18	6.800	9.200
2020-525	9 3 D	33 COULTER JT		12	8.542	0.335	3.93	7.300	9.800
2020-525	9 3 D	120 COULTER		16	8.325	0.277	3.33	7.100	9.600

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
553 PLATELET COUNT				CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %				
2020-521	1 SYSMEX CBC	39 SYSMEX K-		12	378.083	28.831	7.63	284.000	473.000
2020-521	1 SYSMEX CBC	75 SYSMEX XT		16	434.438	37.069	8.53	326.000	543.000
2020-521	1 SYSMEX CBC	79 SYSMEX XS		147	418.007	27.105	6.48	314.000	523.000
2020-521	1 SYSMEX CBC	87 SYSMEX XN		69	455.164	14.455	3.18	341.000	569.000
2020-521	2 ACT5 PENTR	73 COULTER AC		145	488.245	20.310	4.16	366.000	610.000
2020-521	2 ACT5 PENTR	78 ABX PENTRA		23	485.652	9.111	1.88	364.000	607.000

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-521	3	ADVIA 120	64	ADVIA-120	19	433.105	29.424	6.79	325.000	541.000	
2020-521	4	CELLDYN 4K	63	ABBOTT C-D	119	477.795	20.566	4.30	358.000	597.000	
2020-521	4	CELLDYN 4K	94	BC DXH 500	13	486.692	16.895	3.47	365.000	608.000	
2020-521	5	BECTON D	83	B.C600,800	44	463.442	13.708	2.96	348.000	579.000	
2020-521	6	BC Mindray	91	BC-5390	28	534.852	15.518	2.90	401.000	669.000	
2020-521	8	3 ST	31	ABBOTT C-D	42	376.732	31.362	8.32	283.000	471.000	
2020-521	8	3 ST	39	SYSMEX K-	15	384.467	19.200	4.99	288.000	481.000	
2020-521	9	3 D	50	COULT.JT,M	10	577.000	48.452	8.40	433.000	721.000	
2020-521	9	3 D	68	COULTERACT	16	544.938	32.455	5.96	409.000	681.000	
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-522	1	SYSMEX CBC	39	SYSMEX K-	12	67.167	3.869	5.76	50.000	84.000	
2020-522	1	SYSMEX CBC	75	SYSMEX XT	16	70.375	4.442	6.31	53.000	88.000	
2020-522	1	SYSMEX CBC	79	SYSMEX XS	147	66.646	4.037	6.06	50.000	83.000	
2020-522	1	SYSMEX CBC	87	SYSMEX XN	69	60.551	3.778	6.24	45.000	76.000	
2020-522	2	ACT5 PENTR	73	COULTER AC	145	243.923	10.519	4.31	183.000	305.000	
2020-522	2	ACT5 PENTR	78	ABX PENTRA	23	246.636	9.088	3.68	185.000	308.000	
2020-522	3	ADVIA 120	64	ADVIA-120	19	221.316	13.483	6.09	166.000	277.000	
2020-522	4	CELLDYN 4K	63	ABBOTT C-D	119	74.957	4.991	6.66	56.000	94.000	
2020-522	4	CELLDYN 4K	94	BC DXH 500	13	84.308	6.169	7.32	63.000	105.000	
2020-522	5	BECTON D	83	B.C600,800	44	88.167	2.069	2.35	66.000	110.000	
2020-522	6	BC Mindray	91	BC-5390	28	51.821	3.230	6.23	39.000	65.000	
2020-522	8	3 ST	31	ABBOTT C-D	42	69.561	6.559	9.43	52.000	87.000	
2020-522	8	3 ST	39	SYSMEX K-	15	67.200	5.205	7.75	50.000	84.000	
2020-522	9	3 D	50	COULT.JT,M	10	72.700	6.724	9.25	55.000	91.000	
2020-522	9	3 D	68	COULTERACT	16	71.563	9.096	12.71	54.000	89.000	
2020-523	1	SYSMEX CBC	39	SYSMEX K-	12	195.417	11.101	5.68	147.000	244.000	
2020-523	1	SYSMEX CBC	75	SYSMEX XT	16	206.813	16.719	8.08	155.000	259.000	
2020-523	1	SYSMEX CBC	79	SYSMEX XS	147	200.264	13.851	6.92	150.000	250.000	
2020-523	1	SYSMEX CBC	87	SYSMEX XN	69	209.750	7.764	3.70	157.000	262.000	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 134 PLATELET COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-523	2	ACT5 PENTR	73 COULTER AC	145	66.021	4.273	6.47	50.000	83.000	
2020-523	2	ACT5 PENTR	78 ABX PENTRA	23	67.783	4.252	6.27	51.000	85.000	
2020-523	3	ADVIA 120	64 ADVIA-120	19	54.474	4.235	7.77	41.000	68.000	
2020-523	4	CELLDYN 4K	63 ABBOTT C-D	119	255.822	12.815	5.01	192.000	320.000	
2020-523	4	CELLDYN 4K	94 BC DXH 500	13	252.000	9.207	3.65	189.000	315.000	
2020-523	5	BECTON D	83 B.C600,800	44	277.814	12.137	4.37	208.000	347.000	
2020-523	6	BC Mindray	91 BC-5390	28	259.393	10.241	3.95	195.000	324.000	
2020-523	8	3 ST	31 ABBOTT C-D	42	196.390	16.622	8.46	147.000	245.000	
2020-523	8	3 ST	39 SYSMEX K-	15	197.533	7.347	3.72	148.000	247.000	
2020-523	9	3 D	50 COULT.JT,M	10	284.200	21.752	7.65	213.000	355.000	
2020-523	9	3 D	68 COULTERACT	16	273.313	15.386	5.63	205.000	342.000	
553 PLATELET COUNT		CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-524	1	SYSMEX CBC	39 SYSMEX K-	12	378.333	25.021	6.61	284.000	473.000	
2020-524	1	SYSMEX CBC	75 SYSMEX XT	16	69.563	3.774	5.43	52.000	87.000	
2020-524	1	SYSMEX CBC	79 SYSMEX XS	147	65.917	4.381	6.65	49.000	82.000	
2020-524	1	SYSMEX CBC	87 SYSMEX XN	69	59.456	3.332	5.60	45.000	74.000	
2020-524	2	ACT5 PENTR	73 COULTER AC	145	483.021	20.090	4.16	362.000	604.000	
2020-524	2	ACT5 PENTR	78 ABX PENTRA	23	481.739	18.066	3.75	361.000	602.000	
2020-524	3	ADVIA 120	64 ADVIA-120	19	426.158	25.860	6.07	320.000	533.000	
2020-524	4	CELLDYN 4K	63 ABBOTT C-D	119	74.437	5.458	7.33	56.000	93.000	
2020-524	4	CELLDYN 4K	94 BC DXH 500	13	86.231	10.009	11.61	65.000	108.000	
2020-524	5	BECTON D	83 B.C600,800	44	89.884	2.730	3.04	67.000	112.000	
2020-524	6	BC Mindray	91 BC-5390	28	51.607	2.858	5.54	39.000	65.000	
2020-524	8	3 ST	31 ABBOTT C-D	42	378.537	32.207	8.51	284.000	473.000	
2020-524	8	3 ST	39 SYSMEX K-	15	385.867	18.676	4.84	289.000	482.000	
2020-524	9	3 D	50 COULT.JT,M	10	571.200	44.575	7.80	428.000	714.000	
2020-524	9	3 D	68 COULTERACT	16	548.625	25.739	4.69	411.000	686.000	
2020-525	1	SYSMEX CBC	39 SYSMEX K-	12	191.667	12.795	6.68	144.000	240.000	
2020-525	1	SYSMEX CBC	75 SYSMEX XT	16	431.625	33.179	7.69	324.000	540.000	

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT

ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-525	1	SYSMEX CBC	79 SYSMEX XS	147	416.734	27.734	6.66	313.000	521.000	
2020-525	1	SYSMEX CBC	87 SYSMEX XN	69	451.309	14.289	3.17	338.000	564.000	
2020-525	2	ACT5 PENTR	73 COULTER AC	145	65.133	4.565	7.01	49.000	81.000	
2020-525	2	ACT5 PENTR	78 ABX PENTRA	23	66.636	3.310	4.97	50.000	83.000	
2020-525	3	ADVIA 120	64 ADVIA-120	19	55.316	4.846	8.76	41.000	69.000	
2020-525	4	CELLDYN 4K	63 ABBOTT C-D	119	477.559	22.408	4.69	358.000	597.000	
2020-525	4	CELLDYN 4K	94 BC DXH 500	13	484.385	16.850	3.48	363.000	605.000	
2020-525	5	BECTON D	83 B.C600,800	44	462.884	16.519	3.57	347.000	579.000	
2020-525	6	BC Mindray	91 BC-5390	28	536.333	14.981	2.79	402.000	670.000	
2020-525	8	3 ST	31 ABBOTT C-D	42	198.439	15.833	7.98	149.000	248.000	
2020-525	8	3 ST	39 SYSMEX K-	15	195.467	12.176	6.23	147.000	244.000	
2020-525	9	3 D	50 COULT.JT,M	10	285.600	20.641	7.23	214.000	357.000	
2020-525	9	3 D	68 COULTERACT	16	276.750	12.804	4.63	208.000	346.000	

SPEC: 20 HEMATOLOGY SUB: 75 BODY FLUID GRP: 145 BODY FLUID CELL COU.

ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}						
	2020-561			47	1044.000	119.853	11.48	684.000	1404.000
	2020-562			47	0.978	1.053	107.61	0	4.000
	2020-563			47	1035.870	104.082	10.05	724.000	1348.000
	2020-564			47	0.830	1.117	134.61	0	4.000
	2020-565			47	1020.043	115.943	11.37	672.000	1368.000
160	BF WHITE CELL COUNT								
	2020-561			46	314.239	37.343	11.88	202.000	426.000

Participant Summary

MONTH 7 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 75 BODY FLUID GRP: 145 BODY FLUID CELL COU.

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
160 BF WHITE CELL COUNT		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
2020-562					46	5.109	1.902	37.24	0	11.000
2020-563					46	309.244	32.744	10.59	211.000	407.000
2020-564					46	5.065	1.292	25.51	1.000	9.000
2020-565					46	308.222	28.895	9.37	222.000	395.000

SPE: 30 IMMUNOHEMATOLOGY SUB: 80 ABO-RH GRP: 150 ABO-RH

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
083 ABO GROUP											
2020-571				001 003	001 003		002 - A1	91	4	91	100
2020-571				001 003	001 003		001 - A	91	87	91	100
2020-572				008 008	008 008		008 - O	91	91	91	100
2020-573				005 007	005 007		006 - A1B	91	2	91	100
2020-573				005 007	005 007		005 - AB	91	89	91	100
2020-574				008 008	008 008		008 - O	91	91	91	100
2020-575				004 004	004 004		004 - B	91	91	91	100

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
126 RH TYPE													
2020-571				003 004	003 004		003 - Rh-Du np	91	9	90	100		
2020-571				003 004	003 004		004 - Rh-	91	81	90	100		
2020-571				003 004	003 004		001 - Rh+	91	1	90	0		
2020-572				001 003	001 003		001 - Rh+	91	91	91	100		
2020-573				001 003	001 003		003 - Rh-Du np	91	1	91	100		
2020-573				001 003	001 003		001 - Rh+	91	90	91	100		
2020-574				003 004	003 004		004 - Rh-	91	81	90	100		
2020-574				003 004	003 004		001 - Rh+	91	1	90	0		
2020-574				003 004	003 004		003 - Rh-Du np	91	9	90	100		
2020-575				001 003	001 003		001 - Rh+	91	91	91	100		

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2020-581				002 002	002 002		002 - POSITIVE	68	66	66	100
2020-581				002 002	002 002		001 - NEGATIVE	68	2	66	0
2020-582				001 001	001 001		001 - NEGATIVE	68	68	68	100
2020-583				001 001	001 001		001 - NEGATIVE	68	68	68	100
2020-584				002 002	002 002		002 - POSITIVE	68	68	68	100
2020-585				001 001	001 001		001 - NEGATIVE	68	68	68	100

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2020-291	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2020-291	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	33	33	33	100
2020-291	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	17	17	17	100
2020-292	001 Poly AHG	001 Gel		002 004			004 - 4+	12	11	12	100
2020-292	001 Poly AHG	001 Gel		002 004			003 - 3+	12	1	12	100
2020-292	001 Poly AHG	002 Tubo					002 - 2+	33	16	33	0
2020-292	001 Poly AHG	002 Tubo					003 - 3+	33	12	33	0
2020-292	001 Poly AHG	002 Tubo					004 - 4+	33	5	33	0
2020-292	002 MnoAnt-IgG	002 Tubo					002 - 2+	16	8	16	0

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-292	002 MnoAnt-IgG	002 Tubo					004 - 4+	16	1	16	0
2020-292	002 MnoAnt-IgG	002 Tubo					003 - 3+	16	7	16	0
179 DIRECT ANTIG(COOMBS)											
2020-293	001 Poly AHG	001 Gel		002 004			004 - 4+	12	11	12	100
2020-293	001 Poly AHG	001 Gel		002 004			003 - 3+	12	1	12	100
2020-293	001 Poly AHG	002 Tubo					002 - 2+	33	21	33	0
2020-293	001 Poly AHG	002 Tubo					004 - 4+	33	6	33	0
2020-293	001 Poly AHG	002 Tubo					003 - 3+	33	6	33	0
2020-293	002 MnoAnt-IgG	002 Tubo					003 - 3+	16	7	16	0
2020-293	002 MnoAnt-IgG	002 Tubo					002 - 2+	16	9	16	0
2020-294	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2020-294	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	33	33	33	100
2020-294	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	17	17	17	100
2020-295	001 Poly AHG	001 Gel		002 004			004 - 4+	12	11	12	100
2020-295	001 Poly AHG	001 Gel		002 004			003 - 3+	12	1	12	100
2020-295	001 Poly AHG	002 Tubo					002 - 2+	32	17	32	0
2020-295	001 Poly AHG	002 Tubo					003 - 3+	32	10	32	0
2020-295	001 Poly AHG	002 Tubo					004 - 4+	32	5	32	0
2020-295	002 MnoAnt-IgG	002 Tubo					004 - 4+	16	2	16	0
2020-295	002 MnoAnt-IgG	002 Tubo					003 - 3+	16	5	16	0
2020-295	002 MnoAnt-IgG	002 Tubo					002 - 2+	16	9	16	0

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY SUB: 100 COMPATIBILITY TESTING GRP: 170 COMPATIBILITY TEST.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
085 COMPATIBIL. TESTING											
2020-591				002 002	002 002		002 - INCOMPATIB	41	41	41	100
2020-592				001 001	001 001		001 - COMPATIBLE	41	41	41	100
2020-593				001 001	001 001		001 - COMPATIBLE	41	41	41	100
2020-594				002 002	002 002		002 - INCOMPATIB	41	41	41	100
2020-595				001 001	001 001		001 - COMPATIBLE	41	41	41	100

SPE: 30 IMMUNOHEMATOLOGY SUB: 110 ANTIBODY IDENTIFICATION GRP: 180 ANTIBODY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
086 ANTIBODY IDENT. TEST.											
2020-594					005 005		005 - Anti C	18	9	18	100
2020-594					016 016		016 - Anti K	18	9	18	100

Participant Summary

MONTH 7 OF YEAR 2020

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC, HIV						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
093 ANTI-HUMAN IMM. (HIV)												
2020-601	055 ALERE DETE			002 002	002 002		002 - POS/R	31	31	31	100	
2020-602	055 ALERE DETE			002 002	002 002		002 - POS/R	31	31	31	100	
2020-603	055 ALERE DETE			001 001	001 001		001 - NEG/NR	31	31	31	100	
2020-604	055 ALERE DETE			002 002	002 002		002 - POS/R	31	31	31	100	
2020-605	055 ALERE DETE			001 001	001 001		001 - NEG/NR	31	31	31	100	