

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE				SYST/INST		REAGENTS	NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD						LABS	VALUE	DEVIATION	CV	LOW	HIGH
001 ALAN AMINO	ALT/SGPT			CLIA TEST ID #: 255		EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-391	5	ENZ. 37 C	194	BIOLIS-24S	154	(IDG)	10	104.200	5.381	5.16	83.000	125.000
2020-391	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	16	103.000	3.588	3.48	82.000	124.000
2020-391	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	108.500	2.021	1.86	87.000	130.000
2020-391	11	MOD. IFCC	38	SIEMENS DI	143	SIEMENS	78	110.351	3.027	2.74	88.000	132.000
2020-391	11	MOD. IFCC	184	INTEGRA	93	ROCHE	31	97.548	3.876	3.97	78.000	117.000
2020-391	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	25	89.320	2.664	2.98	71.000	107.000
2020-391	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	32	113.188	8.995	7.95	91.000	136.000
2020-391	11	MOD. IFCC	203	ROCHE	93	ROCHE	26	98.920	2.667	2.70	79.000	119.000
2020-391	11	MOD. IFCC	212	S. D. EXL	143	SIEMENS	10	112.500	4.588	4.08	90.000	135.000
2020-391	11	MOD. IFCC	214	Envoy 500	150	ELITECH	10	107.800	4.468	4.14	86.000	129.000
2020-391	13	ALA &aKETO	165	VITROS	63	KODAK	11	95.909	2.065	2.15	77.000	115.000
2020-391	13	ALA &aKETO	165	VITROS	155	ALTV	119	95.410	2.450	2.57	76.000	114.000
2020-392	5	ENZ. 37 C	194	BIOLIS-24S	154	(IDG)	10	314.700	12.650	4.02	252.000	378.000
2020-392	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	16	336.313	12.251	3.64	269.000	404.000
2020-392	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	322.583	5.545	1.72	258.000	387.000
2020-392	11	MOD. IFCC	38	SIEMENS DI	143	SIEMENS	78	345.935	10.187	2.94	277.000	415.000
2020-392	11	MOD. IFCC	184	INTEGRA	93	ROCHE	31	312.774	13.104	4.19	250.000	375.000
2020-392	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	25	295.080	7.200	2.44	236.000	354.000
2020-392	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	32	354.781	27.258	7.68	284.000	426.000
2020-392	11	MOD. IFCC	203	ROCHE	93	ROCHE	26	325.160	5.704	1.75	260.000	390.000
2020-392	11	MOD. IFCC	212	S. D. EXL	143	SIEMENS	10	354.300	15.595	4.40	283.000	425.000
2020-392	11	MOD. IFCC	214	Envoy 500	150	ELITECH	10	341.700	15.402	4.51	273.000	410.000
2020-392	13	ALA &aKETO	165	VITROS	63	KODAK	11	304.182	9.143	3.01	243.000	365.000
2020-392	13	ALA &aKETO	165	VITROS	155	ALTV	119	298.863	7.725	2.58	239.000	359.000
2020-393	5	ENZ. 37 C	194	BIOLIS-24S	154	(IDG)	10	86.700	4.314	4.98	69.000	104.000
2020-393	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	16	81.875	2.891	3.53	66.000	98.000
2020-393	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	89.000	1.958	2.20	71.000	107.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	78	88.558	2.728	3.08	71.000	106.000		
2020-393	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	76.600	2.703	3.53	61.000	92.000		
2020-393	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	70.760	1.773	2.51	57.000	85.000		
2020-393	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	32	89.531	6.995	7.81	72.000	107.000		
2020-393	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	77.846	1.915	2.46	62.000	93.000		
2020-393	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	10	90.100	4.253	4.72	72.000	108.000		
2020-393	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	86.000	3.493	4.06	69.000	103.000		
2020-393	13 ALA &aKETO	165 VITROS	63 KODAK	11	76.909	2.875	3.74	62.000	92.000		
2020-393	13 ALA &aKETO	165 VITROS	155 ALTV	119	75.703	2.326	3.07	61.000	91.000		
001 ALAN AMINO ALT/SGPT		CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-394	5 ENZ. 37 C	194 BIOLIS-24S	154 (IDG)	10	31.400	5.817	18.53	25.000	38.000		
2020-394	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	16	20.938	1.298	6.20	17.000	25.000		
2020-394	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	35.583	1.441	4.05	28.000	43.000		
2020-394	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	78	25.597	1.804	7.05	20.000	31.000		
2020-394	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	19.129	1.100	5.75	15.000	23.000		
2020-394	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	17.480	0.700	4.00	14.000	21.000		
2020-394	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	32	23.563	2.549	10.82	19.000	28.000		
2020-394	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	19.115	0.800	4.19	15.000	23.000		
2020-394	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	10	26.000	1.789	6.88	21.000	31.000		
2020-394	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	22.800	0.980	4.30	18.000	27.000		
2020-394	13 ALA &aKETO	165 VITROS	63 KODAK	11	16.091	0.900	5.59	13.000	19.000		
2020-394	13 ALA &aKETO	165 VITROS	155 ALTV	119	16.145	0.850	5.26	13.000	19.000		
2020-395	5 ENZ. 37 C	194 BIOLIS-24S	154 (IDG)	10	214.600	9.233	4.30	172.000	258.000		
2020-395	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	16	221.688	7.380	3.33	177.000	266.000		
2020-395	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	218.750	4.639	2.12	175.000	263.000		
2020-395	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	78	232.218	6.450	2.78	186.000	279.000		
2020-395	11 MOD. IFCC	184 INTEGRA	93 ROCHE	31	209.839	8.069	3.85	168.000	252.000		
2020-395	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	25	192.800	4.472	2.32	154.000	231.000		
2020-395	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	32	238.719	15.979	6.69	191.000	286.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-395	11 MOD. IFCC	203 ROCHE	93 ROCHE	26	214.962	4.433	2.06	172.000	258.000		
2020-395	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	10	236.100	10.084	4.27	189.000	283.000		
2020-395	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	227.300	8.730	3.84	182.000	273.000		
2020-395	13 ALA &aKETO	165 VITROS	63 KODAK	11	202.909	4.719	2.33	162.000	243.000		
2020-395	13 ALA &aKETO	165 VITROS	155 ALTV	119	200.000	4.741	2.37	160.000	240.000		
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-391	3 DYE BI-B.G	165 VITROS	63 KODAK	142	2.472	0.065	2.65	2.200	2.700		
2020-391	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	2.584	0.117	4.55	2.300	2.800		
2020-391	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	25	2.476	0.086	3.48	2.200	2.700		
2020-391	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	33	2.625	0.061	2.33	2.400	2.900		
2020-391	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	19	2.547	0.075	2.95	2.300	2.800		
2020-391	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	12	2.708	0.086	3.18	2.400	3.000		
2020-391	3 DYE BI-B.G	203 ROCHE	93 ROCHE	31	2.642	0.124	4.69	2.400	2.900		
2020-391	3 DYE BI-B.G	214 Envoy 500	150 ELITECH	10	2.720	0.060	2.21	2.400	3.000		
2020-391	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	82	2.386	0.077	3.23	2.100	2.600		
2020-391	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	14	2.385	0.066	2.77	2.100	2.600		
2020-391	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	2.482	0.072	2.88	2.200	2.700		
2020-392	3 DYE BI-B.G	165 VITROS	63 KODAK	142	4.730	0.154	3.25	4.300	5.200		
2020-392	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	4.647	0.140	3.02	4.200	5.100		
2020-392	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	25	4.554	0.087	1.90	4.100	5.000		
2020-392	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	33	4.555	0.128	2.81	4.100	5.000		
2020-392	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	19	4.558	0.099	2.17	4.100	5.000		
2020-392	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	12	4.667	0.143	3.07	4.200	5.100		
2020-392	3 DYE BI-B.G	203 ROCHE	93 ROCHE	31	4.783	0.139	2.91	4.300	5.300		
2020-392	3 DYE BI-B.G	214 Envoy 500	150 ELITECH	10	4.710	0.104	2.22	4.200	5.200		
2020-392	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	82	4.465	0.157	3.52	4.000	4.900		
2020-392	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	14	4.443	0.112	2.51	4.000	4.900		
2020-392	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	4.573	0.075	1.64	4.100	5.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
002	ALBUMIN	CLIA TEST ID #: 265				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-393	3 DYE BI-B.G	165 VITROS	63 KODAK		142	2.273	0.064	2.83	2.000	2.500	
2020-393	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		37	2.381	0.111	4.67	2.100	2.600	
2020-393	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		25	2.275	0.066	2.91	2.000	2.500	
2020-393	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		33	2.425	0.050	2.06	2.200	2.700	
2020-393	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	2.342	0.082	3.48	2.100	2.600	
2020-393	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX		12	2.442	0.095	3.91	2.200	2.700	
2020-393	3 DYE BI-B.G	203 ROCHE	93 ROCHE		31	2.471	0.092	3.73	2.200	2.700	
2020-393	3 DYE BI-B.G	214 Envoy 500	150 ELITECH		10	2.550	0.081	3.16	2.300	2.800	
2020-393	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		82	2.183	0.084	3.86	2.000	2.400	
2020-393	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		14	2.185	0.053	2.44	2.000	2.400	
2020-393	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL		11	2.282	0.072	3.14	2.100	2.500	
2020-394	3 DYE BI-B.G	165 VITROS	63 KODAK		142	1.746	0.056	3.23	1.600	1.900	
2020-394	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		37	1.846	0.083	4.47	1.700	2.000	
2020-394	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		25	1.740	0.063	3.63	1.600	1.900	
2020-394	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		33	1.891	0.045	2.39	1.700	2.100	
2020-394	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	1.800	0.056	3.12	1.600	2.000	
2020-394	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX		12	1.933	0.075	3.86	1.700	2.100	
2020-394	3 DYE BI-B.G	203 ROCHE	93 ROCHE		31	1.910	0.082	4.28	1.700	2.100	
2020-394	3 DYE BI-B.G	214 Envoy 500	150 ELITECH		10	1.970	0.090	4.57	1.800	2.200	
2020-394	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		82	1.643	0.068	4.16	1.500	1.800	
2020-394	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		14	1.607	0.103	6.43	1.400	1.800	
2020-394	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL		11	1.718	0.083	4.85	1.500	1.900	
2020-395	3 DYE BI-B.G	165 VITROS	63 KODAK		142	3.633	0.101	2.77	3.300	4.000	
2020-395	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		37	3.678	0.142	3.85	3.300	4.000	
2020-395	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		25	3.529	0.061	1.73	3.200	3.900	
2020-395	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		33	3.622	0.074	2.04	3.300	4.000	
2020-395	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	3.584	0.081	2.27	3.200	3.900	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-395	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	12	3.683	0.090	2.44	3.300	4.100		
2020-395	3 DYE BI-B.G	203 ROCHE	93 ROCHE	31	3.790	0.117	3.10	3.400	4.200		
2020-395	3 DYE BI-B.G	214 Envoy 500	150 ELITECH	10	3.740	0.066	1.77	3.400	4.100		
2020-395	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	82	3.422	0.115	3.37	3.100	3.800		
2020-395	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	14	3.400	0.076	2.22	3.100	3.700		
2020-395	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	11	3.555	0.050	1.40	3.200	3.900		
003 ALKALINE PHOSPHATE		CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-391	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	86.033	4.246	4.94	60.000	112.000		
2020-391	3 COLORIMET.	203 ROCHE	93 ROCHE	22	81.762	2.671	3.27	57.000	106.000		
2020-391	6 KINETIC 37	165 VITROS	63 KODAK	142	97.324	5.069	5.21	68.000	127.000		
2020-391	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	83.700	4.026	4.81	59.000	109.000		
2020-391	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	73.250	3.585	4.89	51.000	95.000		
2020-391	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	26	74.962	6.334	8.45	52.000	97.000		
2020-391	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	26	81.000	6.226	7.69	57.000	105.000		
2020-391	11 Modified I	38 SIEMENS DI	143 SIEMENS	77	89.026	5.269	5.92	62.000	116.000		
2020-391	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	10	86.300	5.951	6.90	60.000	112.000		
2020-392	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	268.862	9.209	3.43	188.000	350.000		
2020-392	3 COLORIMET.	203 ROCHE	93 ROCHE	22	257.955	13.506	5.24	181.000	335.000		
2020-392	6 KINETIC 37	165 VITROS	63 KODAK	142	190.944	9.348	4.90	134.000	248.000		
2020-392	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	263.300	14.093	5.35	184.000	342.000		
2020-392	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	229.917	11.507	5.00	161.000	299.000		
2020-392	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	26	236.885	19.130	8.08	166.000	308.000		
2020-392	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	26	242.385	16.836	6.95	170.000	315.000		
2020-392	11 Modified I	38 SIEMENS DI	143 SIEMENS	77	285.880	11.139	3.90	200.000	372.000		
2020-392	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	10	267.300	16.835	6.30	187.000	347.000		
2020-393	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	69.600	3.422	4.92	49.000	90.000		
2020-393	3 COLORIMET.	203 ROCHE	93 ROCHE	22	67.409	3.271	4.85	47.000	88.000		
2020-393	6 KINETIC 37	165 VITROS	63 KODAK	142	83.837	4.759	5.68	59.000	109.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	68.100	3.113	4.57	48.000	89.000		
2020-393	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	59.250	2.742	4.63	41.000	77.000		
2020-393	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	26	60.808	5.069	8.34	43.000	79.000		
2020-393	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	26	65.423	5.408	8.27	46.000	85.000		
2020-393	11 Modified I	38 SIEMENS DI	143 SIEMENS	77	72.429	4.406	6.08	51.000	94.000		
2020-393	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	10	69.000	5.568	8.07	48.000	90.000		
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-394	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	25.333	1.578	6.23	18.000	33.000		
2020-394	3 COLORIMET.	203 ROCHE	93 ROCHE	22	24.182	1.825	7.55	17.000	31.000		
2020-394	6 KINETIC 37	165 VITROS	63 KODAK	142	39.873	2.780	6.97	28.000	52.000		
2020-394	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	23.200	2.315	9.98	16.000	30.000		
2020-394	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	21.167	1.344	6.35	15.000	28.000		
2020-394	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	26	21.192	3.000	14.16	15.000	28.000		
2020-394	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	26	23.923	2.630	10.99	17.000	31.000		
2020-394	11 Modified I	38 SIEMENS DI	143 SIEMENS	77	24.453	3.304	13.51	17.000	32.000		
2020-394	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	10	25.300	2.193	8.67	18.000	33.000		
2020-395	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	176.133	9.294	5.28	123.000	229.000		
2020-395	3 COLORIMET.	203 ROCHE	93 ROCHE	22	170.429	7.938	4.66	119.000	222.000		
2020-395	6 KINETIC 37	165 VITROS	63 KODAK	142	159.944	8.328	5.21	112.000	208.000		
2020-395	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	170.900	8.700	5.09	120.000	222.000		
2020-395	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	151.667	6.549	4.32	106.000	197.000		
2020-395	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	26	154.577	12.314	7.97	108.000	201.000		
2020-395	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	26	161.885	13.299	8.22	113.000	210.000		
2020-395	11 Modified I	38 SIEMENS DI	143 SIEMENS	77	187.987	8.176	4.35	132.000	244.000		
2020-395	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	10	175.000	12.434	7.11	123.000	228.000		
004	AMYLASE	CLIA TEST ID #: 285									
2020-391	3 ENZYM. 37C	165 VITROS	63 KODAK	79	49.532	7.165	14.47	35.000	64.000		
2020-391	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	81.083	2.753	3.39	57.000	105.000		
2020-391	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	27	76.074	1.923	2.53	53.000	99.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-391	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	76.500	2.291	3.00	54.000	99.000		
2020-391	9 ENZ/COL.37	203 ROCHE	93 ROCHE	21	73.900	1.221	1.65	52.000	96.000		
2020-391	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	58.529	3.310	5.65	41.000	76.000		
004 AMYLASE		CLIA TEST ID #: 285	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-392	3 ENZYM. 37C	165 VITROS	63 KODAK	79	182.013	8.036	4.41	127.000	237.000		
2020-392	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	245.417	4.609	1.88	172.000	319.000		
2020-392	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	27	229.259	4.222	1.84	160.000	298.000		
2020-392	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	227.917	7.029	3.08	160.000	296.000		
2020-392	9 ENZ/COL.37	203 ROCHE	93 ROCHE	21	226.190	4.382	1.94	158.000	294.000		
2020-392	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	180.471	10.466	5.80	126.000	235.000		
2020-393	3 ENZYM. 37C	165 VITROS	63 KODAK	79	38.468	5.468	14.21	27.000	50.000		
2020-393	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	66.417	2.060	3.10	46.000	86.000		
2020-393	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	27	62.222	1.595	2.56	44.000	81.000		
2020-393	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	62.333	2.014	3.23	44.000	81.000		
2020-393	9 ENZ/COL.37	203 ROCHE	93 ROCHE	21	60.762	1.411	2.32	43.000	79.000		
2020-393	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	48.412	2.498	5.16	34.000	63.000		
2020-394	3 ENZYM. 37C	165 VITROS	63 KODAK	79	30.000	0.000	0.00	21.000	39.000		
2020-394	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	25.167	1.213	4.82	18.000	33.000		
2020-394	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	27	24.926	0.813	3.26	17.000	32.000		
2020-394	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	24.917	0.640	2.57	17.000	32.000		
2020-394	9 ENZ/COL.37	203 ROCHE	93 ROCHE	21	23.571	0.660	2.80	17.000	31.000		
2020-394	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	17	18.353	1.453	7.92	13.000	24.000		
2020-395	3 ENZYM. 37C	165 VITROS	63 KODAK	79	121.823	6.635	5.45	85.000	158.000		
2020-395	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	12	164.750	3.982	2.42	115.000	214.000		
2020-395	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	27	153.444	3.166	2.06	107.000	199.000		
2020-395	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	152.583	4.192	2.75	107.000	198.000		
2020-395	9 ENZ/COL.37	203 ROCHE	93 ROCHE	21	150.000	3.222	2.15	105.000	195.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-395	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	120.059	6.864	5.72	84.000	156.000
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-391	1	ASPAR&KETO	165	VITROS	63	KODAK	135	132.082	4.161	3.15	106.000	158.000
2020-391	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	13	128.308	8.452	6.59	103.000	154.000
2020-391	5	ENZYMA.37C	38	SIEMENS DI	143	SIEMENS	83	123.819	4.396	3.55	99.000	149.000
2020-391	5	ENZYMA.37C	212	S. D. EXL	143	SIEMENS	14	124.000	4.158	3.35	99.000	149.000
2020-391	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	122.455	2.350	1.92	98.000	147.000
2020-391	11	MOD. IFCC	184	INTEGRA	93	ROCHE	32	124.094	6.621	5.34	99.000	149.000
2020-391	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	107.920	4.578	4.24	86.000	130.000
2020-391	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	32	134.484	6.676	4.96	108.000	161.000
2020-391	11	MOD. IFCC	203	ROCHE	93	ROCHE	24	122.458	3.808	3.11	98.000	147.000
2020-391	11	MOD. IFCC	214	Envoy 500	150	ELITECH	10	130.800	7.705	5.89	105.000	157.000
2020-391	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	10	121.300	3.132	2.58	97.000	146.000
2020-392	1	ASPAR&KETO	165	VITROS	63	KODAK	135	457.112	17.935	3.92	366.000	549.000
2020-392	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	13	357.462	22.339	6.25	286.000	429.000
2020-392	5	ENZYMA.37C	38	SIEMENS DI	143	SIEMENS	83	358.346	9.720	2.71	287.000	430.000
2020-392	5	ENZYMA.37C	212	S. D. EXL	143	SIEMENS	14	364.000	13.480	3.70	291.000	437.000
2020-392	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	342.818	5.149	1.50	274.000	411.000
2020-392	11	MOD. IFCC	184	INTEGRA	93	ROCHE	32	354.375	20.489	5.78	284.000	425.000
2020-392	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	312.400	12.897	4.13	250.000	375.000
2020-392	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	32	377.935	22.136	5.86	302.000	454.000
2020-392	11	MOD. IFCC	203	ROCHE	93	ROCHE	24	359.542	12.124	3.37	288.000	431.000
2020-392	11	MOD. IFCC	214	Envoy 500	150	ELITECH	10	367.400	18.073	4.92	294.000	441.000
2020-392	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	10	348.000	11.198	3.22	278.000	418.000
2020-393	1	ASPAR&KETO	165	VITROS	63	KODAK	135	107.925	3.630	3.36	86.000	130.000
2020-393	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	13	108.308	6.329	5.84	87.000	130.000
2020-393	5	ENZYMA.37C	38	SIEMENS DI	143	SIEMENS	83	102.651	3.938	3.84	82.000	123.000
2020-393	5	ENZYMA.37C	212	S. D. EXL	143	SIEMENS	14	103.071	3.654	3.55	82.000	124.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2020-393	10	KINETIC 37	215 B.C DXC	134	BECKM.COUL	11	103.455	2.426	2.34	83.000	124.000	
2020-393	11	MOD. IFCC	184 INTEGRA	93	ROCHE	32	102.281	4.692	4.59	82.000	123.000	
2020-393	11	MOD. IFCC	186 BECK. COUL	76	AU-BECKMAN	26	89.120	3.468	3.89	71.000	107.000	
2020-393	11	MOD. IFCC	191 RX-DAYTONA	126	RANDOX	32	111.063	5.408	4.87	89.000	133.000	
2020-393	11	MOD. IFCC	203 ROCHE	93	ROCHE	24	102.542	3.617	3.53	82.000	123.000	
2020-393	11	MOD. IFCC	214 Envoy 500	150	ELITECH	10	108.300	6.709	6.19	87.000	130.000	
2020-393	13	NADH wo P5	197 ARCHITECT	1	ABBOTT	10	99.400	4.200	4.23	80.000	119.000	
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE 20 %									
2020-394	1	ASPAR&KETO	165 VITROS	63	KODAK	135	44.887	1.583	3.53	36.000	54.000	
2020-394	1	ASPAR&KETO	194 BIOLIS-24S	154	(IDG)	13	40.923	2.464	6.02	33.000	49.000	
2020-394	5	ENZYM.37C	38 SIEMENS DI	143	SIEMENS	83	43.145	2.078	4.82	35.000	52.000	
2020-394	5	ENZYM.37C	212 S. D. EXL	143	SIEMENS	14	41.615	2.371	5.70	33.000	50.000	
2020-394	10	KINETIC 37	215 B.C DXC	134	BECKM.COUL	11	48.636	1.432	2.94	39.000	58.000	
2020-394	11	MOD. IFCC	184 INTEGRA	93	ROCHE	32	41.281	2.004	4.85	33.000	50.000	
2020-394	11	MOD. IFCC	186 BECK. COUL	76	AU-BECKMAN	26	35.720	1.401	3.92	29.000	43.000	
2020-394	11	MOD. IFCC	191 RX-DAYTONA	126	RANDOX	32	45.194	2.967	6.56	36.000	54.000	
2020-394	11	MOD. IFCC	203 ROCHE	93	ROCHE	24	40.833	1.491	3.65	33.000	49.000	
2020-394	11	MOD. IFCC	214 Envoy 500	150	ELITECH	10	43.900	2.300	5.24	35.000	53.000	
2020-394	13	NADH wo P5	197 ARCHITECT	1	ABBOTT	10	40.300	1.269	3.15	32.000	48.000	
2020-395	1	ASPAR&KETO	165 VITROS	63	KODAK	135	284.978	9.474	3.32	228.000	342.000	
2020-395	1	ASPAR&KETO	194 BIOLIS-24S	154	(IDG)	13	255.385	12.833	5.03	204.000	306.000	
2020-395	5	ENZYM.37C	38 SIEMENS DI	143	SIEMENS	83	243.988	7.791	3.19	195.000	293.000	
2020-395	5	ENZYM.37C	212 S. D. EXL	143	SIEMENS	14	244.071	8.606	3.53	195.000	293.000	
2020-395	10	KINETIC 37	215 B.C DXC	134	BECKM.COUL	11	236.091	3.942	1.67	189.000	283.000	
2020-395	11	MOD. IFCC	184 INTEGRA	93	ROCHE	32	244.063	11.454	4.69	195.000	293.000	
2020-395	11	MOD. IFCC	186 BECK. COUL	76	AU-BECKMAN	26	211.760	7.986	3.77	169.000	254.000	
2020-395	11	MOD. IFCC	191 RX-DAYTONA	126	RANDOX	32	260.968	12.963	4.97	209.000	313.000	
2020-395	11	MOD. IFCC	203 ROCHE	93	ROCHE	24	245.542	8.756	3.57	196.000	295.000	
2020-395	11	MOD. IFCC	214 Envoy 500	150	ELITECH	10	252.600	13.537	5.36	202.000	303.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-395	13	NADH w _o P5	197	ARCHITECT	1	ABBOTT	10	237.800	5.930	2.49	190.000	285.000
006	BILIRUBIN, TOTAL		CLIA TEST ID #: 305		EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-391	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	1.138	0.057	5.05	0.910	1.370
2020-391	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	81	1.289	0.064	4.99	1.030	1.550
2020-391	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	23	1.385	0.039	2.81	1.110	1.660
2020-391	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	1.411	0.054	3.84	1.130	1.690
2020-391	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	1.247	0.083	6.69	1.000	1.500
2020-391	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	1.438	0.157	10.94	1.150	1.730
2020-391	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	1.420	0.173	12.19	1.140	1.700
2020-391	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	1.171	0.097	8.29	0.940	1.400
2020-391	13	AZOBIL.	184	INTEGRA	93	ROCHE	18	1.146	0.060	5.26	0.920	1.370
2020-391	13	AZOBIL.	203	ROCHE	93	ROCHE	22	1.159	0.029	2.47	0.930	1.390
2020-391	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	1.259	0.130	10.29	1.010	1.510
2020-392	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	4.255	0.203	4.77	3.400	5.110
2020-392	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	81	4.949	0.226	4.57	3.960	5.940
2020-392	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	23	4.807	0.113	2.35	3.850	5.770
2020-392	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	5.162	0.294	5.70	4.130	6.190
2020-392	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	4.892	0.290	5.93	3.910	5.870
2020-392	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	5.094	0.453	8.89	4.080	6.110
2020-392	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	5.238	0.262	5.01	4.190	6.290
2020-392	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	4.614	0.225	4.87	3.690	5.540
2020-392	13	AZOBIL.	184	INTEGRA	93	ROCHE	18	4.253	0.202	4.76	3.400	5.100
2020-392	13	AZOBIL.	203	ROCHE	93	ROCHE	22	4.350	0.145	3.33	3.480	5.220
2020-392	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	4.656	0.327	7.03	3.720	5.590
2020-393	3	E-M NO BK.	184	INTEGRA	93	ROCHE	11	0.857	0.047	5.52	0.690	1.030
2020-393	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	81	0.952	0.060	6.34	0.760	1.140
2020-393	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	23	1.060	0.049	4.62	0.850	1.270
2020-393	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	24	1.045	0.066	6.31	0.840	1.250

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	10	0.914	0.091	9.99	0.730	1.100		
2020-393	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	1.153	0.144	12.45	0.920	1.380		
2020-393	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	1.102	0.085	7.75	0.880	1.320		
2020-393	7 DIAZO-DYPH	165 VITROS	63 KODAK	139	0.853	0.092	10.82	0.680	1.020		
2020-393	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	0.855	0.047	5.46	0.680	1.030		
2020-393	13 AZOBIL.	203 ROCHE	93 ROCHE	22	0.869	0.031	3.59	0.700	1.040		
2020-393	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	0.933	0.091	9.79	0.750	1.120		
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305				EVALUATION CRITERIA: TARGET VALUE 20 %					
2020-394	3 E-M NO BK.	184 INTEGRA	93 ROCHE	11	0.066	0.044	65.61	0.050	0.080		
2020-394	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	81	0.099	0.036	36.25	0.080	0.120		
2020-394	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	23	0.120	0.027	22.33	0.100	0.140		
2020-394	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	24	0.132	0.051	38.46	0.110	0.160		
2020-394	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	10	0.093	0.040	42.75	0.070	0.110		
2020-394	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	0.130	0.038	29.19	0.100	0.160		
2020-394	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	0.188	0.092	48.94	0.150	0.230		
2020-394	7 DIAZO-DYPH	165 VITROS	63 KODAK	139	0.103	0.021	20.16	0.080	0.120		
2020-394	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	0.076	0.038	50.35	0.060	0.090		
2020-394	13 AZOBIL.	203 ROCHE	93 ROCHE	22	0.136	0.039	28.98	0.110	0.160		
2020-394	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	0.099	0.003	3.03	0.080	0.120		
2020-395	3 E-M NO BK.	184 INTEGRA	93 ROCHE	11	2.775	0.078	2.79	2.220	3.330		
2020-395	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	81	3.097	0.144	4.65	2.480	3.720		
2020-395	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	23	3.140	0.066	2.10	2.510	3.770		
2020-395	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	24	3.293	0.190	5.78	2.630	3.950		
2020-395	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	10	3.074	0.171	5.55	2.460	3.690		
2020-395	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	3.252	0.335	10.31	2.600	3.900		
2020-395	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	3.363	0.214	6.35	2.690	4.040		
2020-395	7 DIAZO-DYPH	165 VITROS	63 KODAK	139	2.887	0.148	5.12	2.310	3.460		
2020-395	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	2.743	0.105	3.81	2.190	3.290		
2020-395	13 AZOBIL.	203 ROCHE	93 ROCHE	22	2.780	0.077	2.76	2.220	3.340		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-395	19 DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	2.955	0.241	8.14	2.360	3.550	
010	CALCIUM, TOTAL	CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE 1.00								
2020-391	3 ARSEN.-III	165	VITROS	63	KODAK	156	7.425	0.168	2.27	6.400	8.400	
2020-391	3 ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	28	7.474	0.167	2.23	6.500	8.500	
2020-391	3 ARSEN.-III	191	RX-DAYTONA	126	RANDOX	35	7.851	0.213	2.71	6.900	8.900	
2020-391	3 ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	7.320	0.240	3.28	6.300	8.300	
2020-391	3 ARSEN.-III	199	Rx-IMOLA	126	RANDOX	11	7.809	0.291	3.72	6.800	8.800	
2020-391	3 ARSEN.-III	214	Envoy 500	150	ELITECH	10	7.660	0.307	4.01	6.700	8.700	
2020-391	8 CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	84	7.371	0.251	3.41	6.400	8.400	
2020-391	8 CRESOLPHT.	184	INTEGRA	93	ROCHE	19	7.558	0.270	3.57	6.600	8.600	
2020-391	8 CRESOLPHT.	203	ROCHE	93	ROCHE	10	7.650	0.136	1.78	6.700	8.700	
2020-391	8 CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	7.150	0.407	5.69	6.200	8.200	
2020-391	14 ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	7.367	0.201	2.73	6.400	8.400	
2020-391	15 CA-NM-BAPT	184	INTEGRA	93	ROCHE	18	7.533	0.245	3.25	6.500	8.500	
2020-391	15 CA-NM-BAPT	203	ROCHE	93	ROCHE	20	7.632	0.166	2.17	6.600	8.600	
2020-392	3 ARSEN.-III	165	VITROS	63	KODAK	156	11.030	0.226	2.05	10.000	12.000	
2020-392	3 ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	28	11.311	0.278	2.46	10.300	12.300	
2020-392	3 ARSEN.-III	191	RX-DAYTONA	126	RANDOX	35	11.457	0.269	2.34	10.500	12.500	
2020-392	3 ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	11.300	0.398	3.53	10.300	12.300	
2020-392	3 ARSEN.-III	199	Rx-IMOLA	126	RANDOX	11	11.409	0.281	2.46	10.400	12.400	
2020-392	3 ARSEN.-III	214	Envoy 500	150	ELITECH	10	11.460	0.317	2.76	10.500	12.500	
2020-392	8 CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	84	10.930	0.294	2.69	9.900	11.900	
2020-392	8 CRESOLPHT.	184	INTEGRA	93	ROCHE	19	11.117	0.297	2.67	10.100	12.100	
2020-392	8 CRESOLPHT.	203	ROCHE	93	ROCHE	10	11.180	0.394	3.53	10.200	12.200	
2020-392	8 CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	10.807	0.442	4.09	9.800	11.800	
2020-392	14 ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	10.842	0.171	1.57	9.800	11.800	
2020-392	15 CA-NM-BAPT	184	INTEGRA	93	ROCHE	18	11.171	0.284	2.55	10.200	12.200	
2020-392	15 CA-NM-BAPT	203	ROCHE	93	ROCHE	20	11.279	0.196	1.74	10.300	12.300	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
010	CALCIUM, TOTAL	CLIA TEST ID #: 345	EVALUATION CRITERIA: TARGET VALUE			1.00					
2020-393	3 ARSEN.-III	165 VITROS	63 KODAK		156	7.094	0.184	2.60	6.100	8.100	
2020-393	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN		28	7.141	0.162	2.26	6.100	8.100	
2020-393	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX		35	7.471	0.243	3.26	6.500	8.500	
2020-393	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT		15	6.953	0.255	3.67	6.000	8.000	
2020-393	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX		11	7.418	0.262	3.53	6.400	8.400	
2020-393	3 ARSEN.-III	214 Envoy 500	150 ELITECH		10	7.150	0.186	2.60	6.200	8.200	
2020-393	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS		84	7.022	0.259	3.69	6.000	8.000	
2020-393	8 CRESOLPHT.	184 INTEGRA	93 ROCHE		19	7.205	0.261	3.62	6.200	8.200	
2020-393	8 CRESOLPHT.	203 ROCHE	93 ROCHE		10	7.350	0.150	2.04	6.400	8.400	
2020-393	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS		14	6.857	0.392	5.72	5.900	7.900	
2020-393	14 ELECTRODE	215 B.C DXC	134 BECKM.COUL		12	7.050	0.194	2.75	6.100	8.100	
2020-393	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE		18	7.128	0.258	3.62	6.100	8.100	
2020-393	15 CA-NM-BAPT	203 ROCHE	93 ROCHE		20	7.295	0.132	1.81	6.300	8.300	
2020-394	3 ARSEN.-III	165 VITROS	63 KODAK		156	6.179	0.165	2.67	5.200	7.200	
2020-394	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN		28	6.215	0.138	2.22	5.200	7.200	
2020-394	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX		35	6.597	0.247	3.74	5.600	7.600	
2020-394	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT		15	6.100	0.213	3.49	5.100	7.100	
2020-394	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX		11	6.582	0.341	5.18	5.600	7.600	
2020-394	3 ARSEN.-III	214 Envoy 500	150 ELITECH		10	6.210	0.281	4.52	5.200	7.200	
2020-394	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS		84	6.240	0.300	4.80	5.200	7.200	
2020-394	8 CRESOLPHT.	184 INTEGRA	93 ROCHE		19	6.342	0.248	3.91	5.300	7.300	
2020-394	8 CRESOLPHT.	203 ROCHE	93 ROCHE		10	6.460	0.136	2.10	5.500	7.500	
2020-394	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS		14	6.100	0.382	6.26	5.100	7.100	
2020-394	14 ELECTRODE	215 B.C DXC	134 BECKM.COUL		12	6.275	0.148	2.36	5.300	7.300	
2020-394	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE		18	6.306	0.196	3.10	5.300	7.300	
2020-394	15 CA-NM-BAPT	203 ROCHE	93 ROCHE		20	6.435	0.142	2.21	5.400	7.400	
2020-395	3 ARSEN.-III	165 VITROS	63 KODAK		156	9.334	0.198	2.12	8.300	10.300	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-395	3	ARSEN.-III	186 BECK. COUL	76	AU-BECKMAN	28	9.370	0.205	2.19	8.400	10.400
2020-395	3	ARSEN.-III	191 RX-DAYTONA	126	RANDOX	35	9.680	0.298	3.08	8.700	10.700
2020-395	3	ARSEN.-III	197 ARCHITECT	1	ABBOTT	15	9.273	0.340	3.66	8.300	10.300
2020-395	3	ARSEN.-III	199 Rx-IMOLA	126	RANDOX	11	9.600	0.273	2.84	8.600	10.600
2020-395	3	ARSEN.-III	214 Envoy 500	150	ELITECH	10	9.500	0.303	3.19	8.500	10.500
2020-395	8	CRESOLPHT.	38 SIEMENS DI	143	SIEMENS	84	9.096	0.282	3.10	8.100	10.100
2020-395	8	CRESOLPHT.	184 INTEGRA	93	ROCHE	19	9.363	0.318	3.40	8.400	10.400
2020-395	8	CRESOLPHT.	203 ROCHE	93	ROCHE	10	9.580	0.189	1.97	8.600	10.600
2020-395	8	CRESOLPHT.	212 S. D. EXL	143	SIEMENS	14	8.950	0.398	4.44	8.000	10.000
2020-395	14	ELECTRODE	215 B.C DXC	134	BECKM.COUL	12	9.108	0.193	2.12	8.100	10.100
2020-395	15	CA-NM-BAPT	184 INTEGRA	93	ROCHE	18	9.311	0.285	3.06	8.300	10.300
2020-395	15	CA-NM-BAPT	203 ROCHE	93	ROCHE	20	9.510	0.176	1.85	8.500	10.500
011	CHLORIDE		CLIA TEST ID #: 355		EVALUATION CRITERIA: TARGET VALUE 5 %						
2020-391	3	ISE DILUTE	184 INTEGRA	93	ROCHE	39	89.846	1.994	2.22	85.000	94.000
2020-391	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	28	88.714	1.508	1.70	84.000	93.000
2020-391	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	16	91.063	1.519	1.67	87.000	96.000
2020-391	3	ISE DILUTE	203 ROCHE	93	ROCHE	33	85.969	2.023	2.35	82.000	90.000
2020-391	3	ISE DILUTE	215 B.C DXC	134	BECKM.COUL	12	90.583	2.326	2.57	86.000	95.000
2020-391	4	ISE UND-DI	165 VITROS	63	KODAK	148	89.673	1.574	1.76	85.000	94.000
2020-391	4	ISE UND-DI	165 VITROS	63	KODAK	148	88.667	1.700	1.92	84.000	93.000
2020-391	4	ISE UND-DI	191 RX-DAYTONA	126	RANDOX	36	92.250	1.639	1.78	88.000	97.000
2020-391	4	ISE UND-DI	191 RX-DAYTONA	126	RANDOX	36	89.250	2.005	2.25	85.000	94.000
2020-391	4	ISE UND-DI	194 BIOLIS-24S	149	JAS	12	92.000	1.472	1.60	87.000	97.000
2020-391	4	ISE UND-DI	199 Rx-IMOLA	126	RANDOX	12	88.333	1.650	1.87	84.000	93.000
2020-391	4	ISE UND-DI	214 Envoy 500	150	ELITECH	10	89.600	1.625	1.81	85.000	94.000
2020-391	9	QUICK LYTE	38 SIEMENS DI	143	SIEMENS	82	88.738	1.773	2.00	84.000	93.000
2020-391	9	QUICK LYTE	212 S. D. EXL	143	SIEMENS	14	88.643	2.057	2.32	84.000	93.000
2020-392	3	ISE DILUTE	184 INTEGRA	93	ROCHE	39	104.641	2.465	2.36	99.000	110.000
2020-392	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	28	102.321	1.774	1.73	97.000	107.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-392	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	16	103.625	1.615	1.56	98.000	109.000
2020-392	3	ISE DILUTE	203	ROCHE	93	ROCHE	33	101.061	2.295	2.27	96.000	106.000
2020-392	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	103.833	2.075	2.00	99.000	109.000
2020-392	4	ISE UND-DI	165	VITROS	63	KODAK	148	105.333	2.494	2.37	100.000	111.000
2020-392	4	ISE UND-DI	165	VITROS	63	KODAK	148	103.445	1.945	1.88	98.000	109.000
2020-392	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	36	104.800	2.081	1.99	100.000	110.000
2020-392	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	36	106.000	1.871	1.76	101.000	111.000
2020-392	4	ISE UND-DI	194	BIOLIS-24S	149	JAS	12	105.917	2.060	1.94	101.000	111.000
2020-392	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	12	103.917	2.216	2.13	99.000	109.000
2020-392	4	ISE UND-DI	214	Envoy 500	150	ELITECH	10	107.600	3.007	2.79	102.000	113.000
2020-392	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	82	106.325	1.794	1.69	101.000	112.000
2020-392	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	106.071	2.120	2.00	101.000	111.000
011	CHLORIDE	CLIA TEST ID #: 355				EVALUATION CRITERIA: TARGET VALUE 5 %						
2020-393	3	ISE DILUTE	184	INTEGRA	93	ROCHE	39	88.237	1.677	1.90	84.000	93.000
2020-393	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	28	87.750	1.503	1.71	83.000	92.000
2020-393	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	16	89.438	2.150	2.40	85.000	94.000
2020-393	3	ISE DILUTE	203	ROCHE	93	ROCHE	33	84.750	1.732	2.04	81.000	89.000
2020-393	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	88.583	2.564	2.89	84.000	93.000
2020-393	4	ISE UND-DI	165	VITROS	63	KODAK	148	88.551	1.723	1.95	84.000	93.000
2020-393	4	ISE UND-DI	165	VITROS	63	KODAK	148	87.333	1.247	1.43	83.000	92.000
2020-393	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	36	89.750	1.090	1.21	85.000	94.000
2020-393	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	36	87.833	2.167	2.47	83.000	92.000
2020-393	4	ISE UND-DI	194	BIOLIS-24S	149	JAS	12	91.250	1.588	1.74	87.000	96.000
2020-393	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	12	86.500	1.708	1.97	82.000	91.000
2020-393	4	ISE UND-DI	214	Envoy 500	150	ELITECH	10	87.700	2.100	2.39	83.000	92.000
2020-393	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	82	86.550	1.717	1.98	82.000	91.000
2020-393	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	86.500	1.991	2.30	82.000	91.000
2020-394	3	ISE DILUTE	184	INTEGRA	93	ROCHE	39	85.308	1.666	1.95	81.000	90.000
2020-394	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	28	85.607	1.319	1.54	81.000	90.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-394	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	16	86.938	1.886	2.17	83.000	91.000		
2020-394	3 ISE DILUTE	203 ROCHE	93 ROCHE	33	82.182	2.124	2.58	78.000	86.000		
2020-394	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	87.083	1.935	2.22	83.000	91.000		
2020-394	4 ISE UND-DI	165 VITROS	63 KODAK	148	84.333	1.247	1.48	80.000	89.000		
2020-394	4 ISE UND-DI	165 VITROS	63 KODAK	148	86.411	1.629	1.88	82.000	91.000		
2020-394	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	36	85.000	2.014	2.37	81.000	89.000		
2020-394	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	36	88.750	2.487	2.80	84.000	93.000		
2020-394	4 ISE UND-DI	194 BIOLIS-24S	149 JAS	12	89.083	2.722	3.06	85.000	94.000		
2020-394	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	12	83.333	2.095	2.51	79.000	88.000		
2020-394	4 ISE UND-DI	214 Envoy 500	150 ELITECH	10	84.200	1.939	2.30	80.000	88.000		
2020-394	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	82	83.333	1.571	1.89	79.000	88.000		
2020-394	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	83.357	1.950	2.34	79.000	88.000		
011 CHLORIDE		CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2020-395	3 ISE DILUTE	184 INTEGRA	93 ROCHE	39	97.816	2.076	2.12	93.000	103.000		
2020-395	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	28	95.185	1.564	1.64	90.000	100.000		
2020-395	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	16	97.125	1.409	1.45	92.000	102.000		
2020-395	3 ISE DILUTE	203 ROCHE	93 ROCHE	33	93.970	2.139	2.28	89.000	99.000		
2020-395	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	97.083	2.139	2.20	92.000	102.000		
2020-395	4 ISE UND-DI	165 VITROS	63 KODAK	148	96.644	1.715	1.77	92.000	101.000		
2020-395	4 ISE UND-DI	165 VITROS	63 KODAK	148	96.333	2.055	2.13	92.000	101.000		
2020-395	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	36	98.250	0.433	0.44	93.000	103.000		
2020-395	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	36	97.056	2.738	2.82	92.000	102.000		
2020-395	4 ISE UND-DI	194 BIOLIS-24S	149 JAS	12	99.083	1.256	1.27	94.000	104.000		
2020-395	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	12	95.917	2.139	2.23	91.000	101.000		
2020-395	4 ISE UND-DI	214 Envoy 500	150 ELITECH	10	98.900	2.022	2.04	94.000	104.000		
2020-395	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	82	97.000	1.757	1.81	92.000	102.000		
2020-395	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	96.929	2.186	2.26	92.000	102.000		
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-391	1 ENZYMATIC	38 SIEMENS DI	143 SIEMENS	82	136.463	4.623	3.39	123.000	150.000		
2020-391	1 ENZYMATIC	165 VITROS	63 KODAK	146	135.806	3.737	2.75	122.000	149.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-391	1	ENZYMATIC	184	INTEGRA	93	ROCHE	23	143.478	3.855	2.69	129.000	158.000
2020-391	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	149.711	6.840	4.57	135.000	165.000
2020-391	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	146.667	7.409	5.05	132.000	161.000
2020-391	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	17	144.471	5.054	3.50	130.000	159.000
2020-391	1	ENZYMATIC	199	Rx-IMOLA	126	RANDOX	10	148.400	4.224	2.85	134.000	163.000
2020-391	1	ENZYMATIC	203	ROCHE	93	ROCHE	18	141.333	4.055	2.87	127.000	155.000
2020-391	1	ENZYMATIC	212	S. D. EXL	143	SIEMENS	14	136.143	5.139	3.77	123.000	150.000
2020-391	1	ENZYMATIC	214	Envoy 500	150	ELITECH	11	145.364	6.596	4.54	131.000	160.000
2020-391	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	12	143.667	4.308	3.00	129.000	158.000
2020-391	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	25	140.083	3.390	2.42	126.000	154.000
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365		EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-392	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	82	191.646	6.575	3.43	172.000	211.000
2020-392	1	ENZYMATIC	165	VITROS	63	KODAK	146	214.222	6.237	2.91	193.000	236.000
2020-392	1	ENZYMATIC	184	INTEGRA	93	ROCHE	23	202.783	8.267	4.08	183.000	223.000
2020-392	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	213.333	9.741	4.57	192.000	235.000
2020-392	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	216.564	10.290	4.75	195.000	238.000
2020-392	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	17	207.765	7.550	3.63	187.000	229.000
2020-392	1	ENZYMATIC	199	Rx-IMOLA	126	RANDOX	10	212.600	7.365	3.46	191.000	234.000
2020-392	1	ENZYMATIC	203	ROCHE	93	ROCHE	18	204.167	8.255	4.04	184.000	225.000
2020-392	1	ENZYMATIC	212	S. D. EXL	143	SIEMENS	14	191.214	7.311	3.82	172.000	210.000
2020-392	1	ENZYMATIC	214	Envoy 500	150	ELITECH	11	212.636	7.643	3.59	191.000	234.000
2020-392	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	12	208.000	3.894	1.87	187.000	229.000
2020-392	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	25	203.958	5.012	2.46	184.000	224.000
2020-393	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	82	130.659	4.829	3.70	118.000	144.000
2020-393	1	ENZYMATIC	165	VITROS	63	KODAK	146	129.407	3.655	2.82	116.000	142.000
2020-393	1	ENZYMATIC	184	INTEGRA	93	ROCHE	23	136.696	5.344	3.91	123.000	150.000
2020-393	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	143.667	6.944	4.83	129.000	158.000
2020-393	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	39	144.103	5.490	3.81	130.000	159.000
2020-393	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	17	138.235	5.440	3.94	124.000	152.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	10	141.700	4.406	3.11	128.000	156.000		
2020-393	1 ENZYMATI	203 ROCHE	93 ROCHE	18	135.222	3.867	2.86	122.000	149.000		
2020-393	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	130.786	2.883	2.20	118.000	144.000		
2020-393	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	140.545	8.564	6.09	126.000	155.000		
2020-393	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	136.417	4.152	3.04	123.000	150.000		
2020-393	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	25	134.917	3.278	2.43	121.000	148.000		
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-394	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	82	118.805	4.212	3.55	107.000	131.000		
2020-394	1 ENZYMATI	165 VITROS	63 KODAK	146	113.166	3.656	3.23	102.000	124.000		
2020-394	1 ENZYMATI	184 INTEGRA	93 ROCHE	23	123.478	3.775	3.06	111.000	136.000		
2020-394	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	39	129.769	5.274	4.06	117.000	143.000		
2020-394	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	39	127.000	3.559	2.80	114.000	140.000		
2020-394	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	125.235	4.697	3.75	113.000	138.000		
2020-394	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	10	126.200	5.455	4.32	114.000	139.000		
2020-394	1 ENZYMATI	203 ROCHE	93 ROCHE	18	120.389	4.704	3.91	108.000	132.000		
2020-394	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	119.000	3.117	2.62	107.000	131.000		
2020-394	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	126.182	7.272	5.76	114.000	139.000		
2020-394	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	121.417	2.060	1.70	109.000	134.000		
2020-394	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	25	120.375	2.884	2.40	108.000	132.000		
2020-395	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	82	162.494	5.698	3.51	146.000	179.000		
2020-395	1 ENZYMATI	165 VITROS	63 KODAK	146	173.821	5.450	3.14	156.000	191.000		
2020-395	1 ENZYMATI	184 INTEGRA	93 ROCHE	23	174.826	6.155	3.52	157.000	192.000		
2020-395	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	39	189.000	7.789	4.12	170.000	208.000		
2020-395	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	39	183.154	7.963	4.35	165.000	201.000		
2020-395	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	175.353	5.911	3.37	158.000	193.000		
2020-395	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	10	181.000	4.561	2.52	163.000	199.000		
2020-395	1 ENZYMATI	203 ROCHE	93 ROCHE	18	173.722	4.964	2.86	156.000	191.000		
2020-395	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	164.143	3.815	2.32	148.000	181.000		
2020-395	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	181.545	9.346	5.15	163.000	200.000		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-395	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	12	175.083	2.871	1.64	158.000	193.000
2020-395	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	25	170.875	4.014	2.35	154.000	188.000
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375				EVALUATION CRITERIA: TARGET VALUE 30 %						
2020-391	11	ENC.-COLO.	165	VITROS	63	KODAK	140	54.162	2.834	5.23	37.900	70.400
2020-391	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	33	62.659	2.982	4.76	43.900	81.500
2020-391	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	25	56.320	3.328	5.91	39.400	73.200
2020-391	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	37	50.797	6.011	11.83	35.600	66.000
2020-391	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	37	44.333	5.185	11.70	31.000	57.600
2020-391	11	ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	13	31.300	2.050	6.55	21.900	40.700
2020-391	11	ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	46.933	8.659	18.45	32.900	61.000
2020-391	11	ENC.-COLO.	203	ROCHE	93	ROCHE	28	61.881	2.888	4.67	43.300	80.400
2020-391	12	HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	16	55.793	1.956	3.51	39.100	72.500
2020-391	12	HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	58.900	2.343	3.98	41.200	76.600
2020-391	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	56.692	3.230	5.70	39.700	73.700
2020-391	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	81	55.129	2.043	3.71	38.600	71.700
2020-391	13	AMIN.W.E.Q	212	S. D. EXL	143	SIEMENS	14	53.571	2.770	5.17	37.500	69.600
2020-392	11	ENC.-COLO.	165	VITROS	63	KODAK	140	73.825	4.139	5.61	51.700	96.000
2020-392	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	33	77.075	3.325	4.31	54.000	100.200
2020-392	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	25	78.996	4.342	5.50	55.300	102.700
2020-392	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	37	62.000	4.082	6.58	43.400	80.600
2020-392	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	37	70.581	7.669	10.87	49.400	91.800
2020-392	11	ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	13	45.317	1.976	4.36	31.700	58.900
2020-392	11	ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	65.325	10.172	15.57	45.700	84.900
2020-392	11	ENC.-COLO.	203	ROCHE	93	ROCHE	28	76.764	3.463	4.51	53.700	99.800
2020-392	12	HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	16	77.480	3.196	4.12	54.200	100.700
2020-392	12	HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	82.600	3.904	4.73	57.800	107.400
2020-392	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	77.567	2.568	3.31	54.300	100.800
2020-392	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	81	61.698	2.796	4.53	43.200	80.200
2020-392	13	AMIN.W.E.Q	212	S. D. EXL	143	SIEMENS	14	59.714	3.194	5.35	41.800	77.600

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
013	CHOLESTEROL, (HDL)	CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE	30	%						
2020-393	11 ENC.-COLO.	165 VITROS	63 KODAK		140	52.583	2.766	5.26	36.800	68.400	
2020-393	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	60.997	3.369	5.52	42.700	79.300	
2020-393	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		25	53.316	3.573	6.70	37.300	69.300	
2020-393	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	49.192	5.845	11.88	34.400	63.900	
2020-393	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	42.667	4.784	11.21	29.900	55.500	
2020-393	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)		13	31.262	3.382	10.82	21.900	40.600	
2020-393	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX		12	45.783	8.621	18.83	32.000	59.500	
2020-393	11 ENC.-COLO.	203 ROCHE	93 ROCHE		28	59.675	3.043	5.10	41.800	77.600	
2020-393	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		16	53.767	1.917	3.57	37.600	69.900	
2020-393	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH		10	57.200	2.600	4.55	40.000	74.400	
2020-393	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL		12	53.917	2.113	3.92	37.700	70.100	
2020-393	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		81	54.280	2.384	4.39	38.000	70.600	
2020-393	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		14	52.214	2.980	5.71	36.600	67.900	
2020-394	11 ENC.-COLO.	165 VITROS	63 KODAK		140	49.949	2.523	5.05	35.000	64.900	
2020-394	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	56.969	3.770	6.62	39.900	74.100	
2020-394	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		25	47.300	3.148	6.65	33.100	61.500	
2020-394	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	36.000	2.944	8.18	25.200	46.800	
2020-394	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	43.625	5.633	12.91	30.500	56.700	
2020-394	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)		13	26.385	2.058	7.80	18.500	34.300	
2020-394	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX		12	42.608	6.772	15.89	29.800	55.400	
2020-394	11 ENC.-COLO.	203 ROCHE	93 ROCHE		28	55.389	2.843	5.13	38.800	72.000	
2020-394	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		16	48.400	3.317	6.85	33.900	62.900	
2020-394	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH		10	52.000	1.949	3.75	36.400	67.600	
2020-394	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL		12	49.475	2.201	4.45	34.600	64.300	
2020-394	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		81	52.019	2.189	4.21	36.400	67.600	
2020-394	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		14	50.929	2.658	5.22	35.700	66.200	
2020-395	11 ENC.-COLO.	165 VITROS	63 KODAK		140	63.293	3.336	5.27	44.300	82.300	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-395	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	70.838	3.990	5.63	49.600	92.100	
2020-395	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		25	66.820	3.909	5.85	46.800	86.900	
2020-395	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	53.000	3.559	6.72	37.100	68.900	
2020-395	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX		37	61.861	6.626	10.71	43.300	80.400	
2020-395	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)		13	38.408	3.050	7.94	26.900	49.900	
2020-395	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX		12	60.258	8.596	14.27	42.200	78.300	
2020-395	11 ENC.-COLO.	203 ROCHE	93 ROCHE		28	69.721	3.096	4.44	48.800	90.600	
2020-395	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		16	67.087	2.553	3.80	47.000	87.200	
2020-395	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH		10	72.900	1.868	2.56	51.000	94.800	
2020-395	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL		12	67.708	2.981	4.40	47.400	88.000	
2020-395	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		81	58.627	2.139	3.65	41.000	76.200	
2020-395	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		14	56.857	2.695	4.74	39.800	73.900	
014 CREATINE KINASE (CK)		CLIA TEST ID #: 385		EVALUATION CRITERIA: TARGET VALUE 30 %							
2020-391	3 KINETIC 37	165 VITROS	63 KODAK		72	166.681	10.612	6.37	117.000	217.000	
2020-391	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN		18	173.611	7.917	4.56	122.000	226.000	
2020-391	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)		13	151.462	7.078	4.67	106.000	197.000	
2020-391	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL		11	198.273	6.863	3.46	139.000	258.000	
2020-391	8 Hexokinase	184 INTEGRA	93 ROCHE		14	188.929	8.198	4.34	132.000	246.000	
2020-391	8 Hexokinase	203 ROCHE	93 ROCHE		12	188.750	4.585	2.43	132.000	245.000	
2020-391	9 SIEMENS	38 SIEMENS DI	143 SIEMENS		31	188.067	3.660	1.95	132.000	244.000	
2020-392	3 KINETIC 37	165 VITROS	63 KODAK		72	434.472	24.130	5.55	304.000	565.000	
2020-392	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN		18	558.611	29.569	5.29	391.000	726.000	
2020-392	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)		13	484.692	19.483	4.02	339.000	630.000	
2020-392	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL		11	620.182	21.234	3.42	434.000	806.000	
2020-392	8 Hexokinase	184 INTEGRA	93 ROCHE		14	580.714	21.492	3.70	407.000	755.000	
2020-392	8 Hexokinase	203 ROCHE	93 ROCHE		12	590.333	13.174	2.23	413.000	767.000	
2020-392	9 SIEMENS	38 SIEMENS DI	143 SIEMENS		31	587.452	20.216	3.44	411.000	764.000	
2020-393	3 KINETIC 37	165 VITROS	63 KODAK		72	133.239	9.158	6.87	93.000	173.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-393	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	137.471	4.146	3.02	96.000	179.000
2020-393	3	KINETIC 37	194	BIOLIS-24S	154	(IDG)	13	117.615	8.214	6.98	82.000	153.000
2020-393	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	159.091	5.946	3.74	111.000	207.000
2020-393	8	Hexokinase	184	INTEGRA	93	ROCHE	14	151.071	5.230	3.46	106.000	196.000
2020-393	8	Hexokinase	203	ROCHE	93	ROCHE	12	149.333	3.727	2.50	105.000	194.000
2020-393	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	31	150.065	4.127	2.75	105.000	195.000
014	CREATINE KINASE (CK)		CLIA	TEST ID #: 385		EVALUATION CRITERIA: TARGET VALUE 30 %						
2020-394	3	KINETIC 37	165	VITROS	63	KODAK	72	37.338	3.369	9.02	26.000	49.000
2020-394	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	37.722	2.231	5.91	26.000	49.000
2020-394	3	KINETIC 37	194	BIOLIS-24S	154	(IDG)	13	32.385	2.897	8.94	23.000	42.000
2020-394	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	44.818	2.124	4.74	31.000	58.000
2020-394	8	Hexokinase	184	INTEGRA	93	ROCHE	14	45.538	1.824	4.00	32.000	59.000
2020-394	8	Hexokinase	203	ROCHE	93	ROCHE	12	44.750	1.422	3.18	31.000	58.000
2020-394	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	31	43.129	1.896	4.40	30.000	56.000
2020-395	3	KINETIC 37	165	VITROS	63	KODAK	72	327.208	19.796	6.05	229.000	425.000
2020-395	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	368.611	17.873	4.85	258.000	479.000
2020-395	3	KINETIC 37	194	BIOLIS-24S	154	(IDG)	13	319.538	15.072	4.72	224.000	415.000
2020-395	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	416.455	13.459	3.23	292.000	541.000
2020-395	8	Hexokinase	184	INTEGRA	93	ROCHE	14	389.071	14.270	3.67	272.000	506.000
2020-395	8	Hexokinase	203	ROCHE	93	ROCHE	12	396.083	9.041	2.28	277.000	515.000
2020-395	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	31	394.452	10.361	2.63	276.000	513.000
016	CREATININE		CLIA	TEST ID #: 405		EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}						
2020-391	2	AK. PIC. N	191	RX-DAYTONA	126	RANDOX	10	3.533	0.356	10.07	3.000	4.060
2020-391	2	AK. PIC. N	194	BIOLIS-24S	154	(IDG)	13	4.268	0.067	1.57	3.630	4.910
2020-391	3	AK. PIC. L	191	RX-DAYTONA	126	RANDOX	24	3.493	0.214	6.14	2.970	4.020
2020-391	6	KIN. ALK P	184	INTEGRA	93	ROCHE	25	4.403	0.153	3.47	3.740	5.060
2020-391	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	28	4.186	0.103	2.47	3.560	4.810
2020-391	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	18	4.299	0.106	2.47	3.650	4.940
2020-391	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	4.330	0.137	3.17	3.680	4.980

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
	2020-391	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	4.302	0.111	2.58	3.660	4.950		
	2020-391	8 IDMS-CREA	165 VITROS	63 KODAK	152	4.062	0.104	2.56	3.450	4.670		
	2020-391	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	75	4.355	0.121	2.77	3.700	5.010		
	2020-391	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	10	4.410	0.146	3.31	3.750	5.070		
016 CREATININE			CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}								
	2020-392	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	5.013	0.339	6.76	4.260	5.770		
	2020-392	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	13	5.951	0.112	1.88	5.060	6.840		
	2020-392	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	24	4.882	0.299	6.13	4.150	5.610		
	2020-392	6 KIN. ALK P	184 INTEGRA	93 ROCHE	25	5.803	0.164	2.83	4.930	6.670		
	2020-392	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	28	5.882	0.154	2.62	5.000	6.760		
	2020-392	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	6.366	0.163	2.55	5.410	7.320		
	2020-392	6 KIN. ALK P	203 ROCHE	93 ROCHE	22	5.820	0.159	2.72	4.950	6.690		
	2020-392	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	6.001	0.169	2.81	5.100	6.900		
	2020-392	8 IDMS-CREA	165 VITROS	63 KODAK	152	6.018	0.155	2.58	5.120	6.920		
	2020-392	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	75	6.155	0.167	2.72	5.230	7.080		
	2020-392	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	10	6.290	0.211	3.35	5.350	7.230		
	2020-393	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	3.462	0.303	8.74	2.940	3.980		
	2020-393	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	13	4.075	0.063	1.54	3.460	4.690		
	2020-393	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	24	3.367	0.191	5.66	2.860	3.870		
	2020-393	6 KIN. ALK P	184 INTEGRA	93 ROCHE	25	4.193	0.142	3.38	3.560	4.820		
	2020-393	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	28	4.007	0.101	2.52	3.410	4.610		
	2020-393	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	4.052	0.157	3.87	3.440	4.660		
	2020-393	6 KIN. ALK P	203 ROCHE	93 ROCHE	22	4.176	0.122	2.93	3.550	4.800		
	2020-393	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	4.122	0.144	3.50	3.500	4.740		
	2020-393	8 IDMS-CREA	165 VITROS	63 KODAK	152	3.882	0.108	2.78	3.300	4.460		
	2020-393	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	75	4.125	0.105	2.55	3.510	4.740		
	2020-393	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	10	4.172	0.123	2.96	3.550	4.800		
	2020-394	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	3.151	0.240	7.61	2.680	3.620		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020-394	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	13	3.718	0.091	2.45	3.160	4.280			
2020-394	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	24	3.023	0.169	5.60	2.570	3.480			
2020-394	6 KIN. ALK P	184 INTEGRA	93 ROCHE	25	3.812	0.085	2.24	3.240	4.380			
2020-394	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	28	3.575	0.104	2.90	3.040	4.110			
2020-394	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	3.601	0.082	2.29	3.060	4.140			
2020-394	6 KIN. ALK P	203 ROCHE	93 ROCHE	22	3.776	0.111	2.95	3.210	4.340			
2020-394	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	3.723	0.104	2.79	3.160	4.280			
2020-394	8 IDMS-CREA	165 VITROS	63 KODAK	152	3.436	0.100	2.90	2.920	3.950			
2020-394	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	75	3.528	0.083	2.34	3.000	4.060			
2020-394	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	10	3.589	0.109	3.04	3.050	4.130			
016 CREATININE		CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE	15 % OR 0.30 {GREATER}								
2020-395	2 AK. PIC. N	191 RX-DAYTONA	126 RANDOX	10	4.374	0.303	6.94	3.720	5.030			
2020-395	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	13	5.098	0.111	2.17	4.330	5.860			
2020-395	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	24	4.202	0.238	5.67	3.570	4.830			
2020-395	6 KIN. ALK P	184 INTEGRA	93 ROCHE	25	5.118	0.080	1.57	4.350	5.890			
2020-395	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	28	5.022	0.102	2.03	4.270	5.780			
2020-395	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	5.303	0.150	2.83	4.510	6.100			
2020-395	6 KIN. ALK P	203 ROCHE	93 ROCHE	22	5.163	0.153	2.96	4.390	5.940			
2020-395	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	12	5.197	0.108	2.09	4.420	5.980			
2020-395	8 IDMS-CREA	165 VITROS	63 KODAK	152	5.065	0.132	2.61	4.310	5.830			
2020-395	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	75	5.219	0.136	2.60	4.440	6.000			
2020-395	11 IDMS Jaffe	212 S. D. EXL	143 SIEMENS	10	5.306	0.159	2.99	4.510	6.100			
017 GLUCOSE		CLIA TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE	10 % OR 6.00 {GREATER}								
2020-391	2 GL. OXI. E	215 B.C DXC	134 BECKM.COUL	12	72.333	1.650	2.28	65.000	80.000			
2020-391	3 GL. OXI. C	165 VITROS	63 KODAK	163	75.000	4.546	6.06	68.000	83.000			
2020-391	3 GL. OXI. C	165 VITROS	63 KODAK	163	71.759	2.030	2.83	65.000	79.000			
2020-391	3 GL. OXI. C	213 SELECTRA P	150 ELITECH	12	79.000	2.082	2.64	71.000	87.000			
2020-391	3 GL. OXI. C	214 Envoy 500	150 ELITECH	12	79.333	1.650	2.08	71.000	87.000			
2020-391	5 HEXOKINASE	38 SIEMENS DI	143 SIEMENS	84	78.940	2.441	3.09	71.000	87.000			
2020-391	5 HEXOKINASE	186 BECK. COUL	76 AU-BECKMAN	28	74.593	2.215	2.97	67.000	82.000			

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-391	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	28	80.679	3.129	3.88	73.000	89.000
2020-391	5	HEXOKINASE	203 ROCHE	93	ROCHE	17	74.882	2.847	3.80	67.000	82.000
2020-391	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	78.357	1.716	2.19	71.000	86.000
2020-391	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	79.600	2.458	3.09	72.000	88.000
2020-391	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	13	73.385	2.021	2.75	66.000	81.000
2020-391	12	HEXOKINASE	184 INTEGRA	93	ROCHE	29	75.931	2.876	3.79	68.000	84.000
2020-391	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	72.176	1.790	2.48	65.000	79.000
2020-391	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	75.500	2.372	3.14	68.000	83.000
017	GLUCOSE	CLIA	TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}							
2020-392	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	144.833	1.818	1.26	130.000	159.000
2020-392	3	GL. OXI. C	165 VITROS	63	KODAK	163	147.778	5.411	3.66	133.000	163.000
2020-392	3	GL. OXI. C	165 VITROS	63	KODAK	163	147.432	3.651	2.48	133.000	162.000
2020-392	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	151.750	7.224	4.76	137.000	167.000
2020-392	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	154.833	3.782	2.44	139.000	170.000
2020-392	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	84	153.357	4.278	2.79	138.000	169.000
2020-392	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	28	146.370	4.120	2.81	132.000	161.000
2020-392	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	28	154.000	5.763	3.74	139.000	169.000
2020-392	5	HEXOKINASE	203 ROCHE	93	ROCHE	17	147.941	6.881	4.65	133.000	163.000
2020-392	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	153.714	2.373	1.54	138.000	169.000
2020-392	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	149.100	4.969	3.33	134.000	164.000
2020-392	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	13	144.077	3.647	2.53	130.000	158.000
2020-392	12	HEXOKINASE	184 INTEGRA	93	ROCHE	29	149.111	3.705	2.48	134.000	164.000
2020-392	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	142.118	4.282	3.01	128.000	156.000
2020-392	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	148.375	5.611	3.78	134.000	163.000
2020-393	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	66.250	1.876	2.83	60.000	73.000
2020-393	3	GL. OXI. C	165 VITROS	63	KODAK	163	64.602	1.928	2.98	58.000	71.000
2020-393	3	GL. OXI. C	165 VITROS	63	KODAK	163	68.778	2.973	4.32	62.000	76.000
2020-393	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	72.167	3.262	4.52	65.000	79.000
2020-393	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	73.000	1.472	2.02	66.000	80.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	84	71.602	2.012	2.81	64.000	79.000
2020-393	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	28	67.750	2.681	3.96	61.000	75.000
2020-393	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	28	74.357	3.014	4.05	67.000	82.000
2020-393	5	HEXOKINASE	203 ROCHE	93	ROCHE	17	69.059	1.798	2.60	62.000	76.000
2020-393	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	70.857	1.301	1.84	64.000	78.000
2020-393	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	72.000	2.720	3.78	65.000	79.000
2020-393	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	13	68.154	2.070	3.04	61.000	75.000
2020-393	12	HEXOKINASE	184 INTEGRA	93	ROCHE	29	68.966	2.297	3.33	62.000	76.000
2020-393	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	65.235	2.016	3.09	59.000	72.000
2020-393	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	68.563	2.449	3.57	62.000	75.000
017	GLUCOSE	CLIA TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE	10	% OR 6.00 {GREATER}						
2020-394	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	48.167	1.404	2.92	42.000	54.000
2020-394	3	GL. OXI. C	165 VITROS	63	KODAK	163	50.444	2.455	4.87	44.000	56.000
2020-394	3	GL. OXI. C	165 VITROS	63	KODAK	163	44.748	1.336	2.99	39.000	51.000
2020-394	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	52.167	3.625	6.95	46.000	58.000
2020-394	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	54.167	2.154	3.98	48.000	60.000
2020-394	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	84	52.890	1.718	3.25	47.000	59.000
2020-394	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	28	49.926	1.654	3.31	44.000	56.000
2020-394	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	28	55.643	3.061	5.50	50.000	62.000
2020-394	5	HEXOKINASE	203 ROCHE	93	ROCHE	17	51.000	1.328	2.60	45.000	57.000
2020-394	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	52.357	1.288	2.46	46.000	58.000
2020-394	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	54.700	2.052	3.75	49.000	61.000
2020-394	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	13	50.846	1.610	3.17	45.000	57.000
2020-394	12	HEXOKINASE	184 INTEGRA	93	ROCHE	29	50.690	1.932	3.81	45.000	57.000
2020-394	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	46.824	1.543	3.30	41.000	53.000
2020-394	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	50.438	1.767	3.50	44.000	56.000
2020-395	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	108.583	1.552	1.43	98.000	119.000
2020-395	3	GL. OXI. C	165 VITROS	63	KODAK	163	112.444	3.624	3.22	101.000	124.000
2020-395	3	GL. OXI. C	165 VITROS	63	KODAK	163	109.654	2.535	2.31	99.000	121.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-395	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	115.917	5.392	4.65	104.000	128.000
2020-395	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	117.917	3.402	2.89	106.000	130.000
2020-395	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	84	116.159	3.078	2.65	105.000	128.000
2020-395	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	28	110.296	3.342	3.03	99.000	121.000
2020-395	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	28	117.571	4.709	4.01	106.000	129.000
2020-395	5	HEXOKINASE	203 ROCHE	93	ROCHE	17	113.118	3.066	2.71	102.000	124.000
2020-395	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	115.929	1.486	1.28	104.000	128.000
2020-395	6	HEXOKI. CO	191 RX-DAYTONA	126	RANDOX	10	114.800	2.088	1.82	103.000	126.000
2020-395	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	13	109.538	3.225	2.94	99.000	120.000
2020-395	12	HEXOKINASE	184 INTEGRA	93	ROCHE	29	113.103	4.155	3.67	102.000	124.000
2020-395	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	107.471	2.404	2.24	97.000	118.000
2020-395	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	113.188	4.447	3.93	102.000	125.000
018 IRON, TOTAL			CLIA TEST ID #: 425		EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-391	4	COULOMETR.	165 VITROS	63	KODAK	14	101.143	5.642	5.58	81.000	121.000
2020-392	4	COULOMETR.	165 VITROS	63	KODAK	14	292.714	12.003	4.10	234.000	351.000
2020-393	4	COULOMETR.	165 VITROS	63	KODAK	14	87.143	5.343	6.13	70.000	105.000
2020-394	4	COULOMETR.	165 VITROS	63	KODAK	14	49.500	4.289	8.66	40.000	59.000
2020-395	4	COULOMETR.	165 VITROS	63	KODAK	14	192.071	6.573	3.42	154.000	230.000
019 LACTASE DEHY. (LDH)			CLIA TEST ID #: 435								
2020-391	3	KINETIC-37	186 BECK. COUL	76	AU-BECKMAN	18	170.722	12.516	7.33	137.000	205.000
2020-391	3	KINETIC-37	194 BIOLIS-24S	154	(IDG)	12	187.083	9.819	5.25	150.000	225.000
2020-391	3	KINETIC-37	215 B.C DXC	134	BECKM.COUL	11	174.455	5.960	3.42	140.000	209.000
2020-391	6	NADH	184 INTEGRA	93	ROCHE	16	220.125	7.288	3.31	176.000	264.000
2020-391	6	NADH	203 ROCHE	93	ROCHE	22	211.500	8.606	4.07	169.000	254.000
2020-391	7	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	34	203.735	8.019	3.94	163.000	244.000
2020-391	9	LDH1	165 VITROS	63	KODAK	65	258.369	5.958	2.31	207.000	310.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
019	LACTASE DEHY. (LDH)			CLIA TEST ID #: 435		EVALUATION CRITERIA: TARGET VALUE 20 %					
2020-392	3 KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	494.889	39.523	7.99	396.000	594.000
2020-392	3 KINETIC-37	194	BIOLIS-24S	154	(IDG)	12	543.583	17.666	3.25	435.000	652.000
2020-392	3 KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	477.000	16.158	3.39	382.000	572.000
2020-392	6 NADH	184	INTEGRA	93	ROCHE	16	608.313	17.585	2.89	487.000	730.000
2020-392	6 NADH	203	ROCHE	93	ROCHE	22	604.682	26.326	4.35	484.000	726.000
2020-392	7 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	34	587.235	21.389	3.64	470.000	705.000
2020-392	9 LDH1	165	VITROS	63	KODAK	65	730.846	32.360	4.43	585.000	877.000
2020-393	3 KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	140.778	10.470	7.44	113.000	169.000
2020-393	3 KINETIC-37	194	BIOLIS-24S	154	(IDG)	12	155.917	3.989	2.56	125.000	187.000
2020-393	3 KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	142.909	4.522	3.16	114.000	171.000
2020-393	6 NADH	184	INTEGRA	93	ROCHE	16	180.563	5.612	3.11	144.000	217.000
2020-393	6 NADH	203	ROCHE	93	ROCHE	22	175.545	5.614	3.20	140.000	211.000
2020-393	7 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	34	168.912	6.312	3.74	135.000	203.000
2020-393	9 LDH1	165	VITROS	63	KODAK	65	215.262	4.996	2.32	172.000	258.000
2020-394	3 KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	56.611	3.759	6.64	45.000	68.000
2020-394	3 KINETIC-37	194	BIOLIS-24S	154	(IDG)	12	62.667	2.925	4.67	50.000	75.000
2020-394	3 KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	58.727	2.049	3.49	47.000	70.000
2020-394	6 NADH	184	INTEGRA	93	ROCHE	16	75.133	2.604	3.47	60.000	90.000
2020-394	6 NADH	203	ROCHE	93	ROCHE	22	72.182	2.639	3.66	58.000	87.000
2020-394	7 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	34	69.382	5.765	8.31	56.000	83.000
2020-394	9 LDH1	165	VITROS	63	KODAK	65	86.578	4.183	4.83	69.000	104.000
2020-395	3 KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	333.111	23.727	7.12	266.000	400.000
2020-395	3 KINETIC-37	194	BIOLIS-24S	154	(IDG)	12	372.750	20.080	5.39	298.000	447.000
2020-395	3 KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	331.091	13.035	3.94	265.000	397.000
2020-395	6 NADH	184	INTEGRA	93	ROCHE	16	417.438	23.680	5.67	334.000	501.000
2020-395	6 NADH	203	ROCHE	93	ROCHE	22	414.364	13.196	3.18	331.000	497.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-395	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	34	399.118	11.923	2.99	319.000	479.000	
2020-395	9	LDH1	165	VITROS	63	KODAK	65	494.688	11.147	2.25	396.000	594.000	
021 MAGNESIUM		CLIA TEST ID #: 455				EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-391	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	3.509	0.100	2.84	2.600	4.400	
2020-391	5	COLOR./K E	165	VITROS	63	KODAK	22	3.845	0.103	2.69	2.900	4.800	
2020-392	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	5.291	0.162	3.06	4.000	6.600	
2020-392	5	COLOR./K E	165	VITROS	63	KODAK	22	5.732	0.205	3.58	4.300	7.200	
2020-393	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	3.345	0.116	3.46	2.500	4.200	
2020-393	5	COLOR./K E	165	VITROS	63	KODAK	22	3.664	0.115	3.14	2.700	4.600	
2020-394	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	2.927	0.105	3.59	2.200	3.700	
2020-394	5	COLOR./K E	165	VITROS	63	KODAK	22	3.245	0.089	2.74	2.400	4.100	
2020-395	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	4.418	0.170	3.84	3.300	5.500	
2020-395	5	COLOR./K E	165	VITROS	63	KODAK	22	4.786	0.142	2.97	3.600	6.000	
022 POTASSIUM		CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00							
2020-391	3	ISE DILUT.	184	INTEGRA	93	ROCHE	37	4.070	0.087	2.13	3.100	5.100	
2020-391	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	28	3.993	0.070	1.76	3.000	5.000	
2020-391	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	17	3.976	0.139	3.51	3.000	5.000	
2020-391	3	ISE DILUT.	203	ROCHE	93	ROCHE	34	4.124	0.095	2.31	3.100	5.100	
2020-391	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	3.917	0.090	2.29	2.900	4.900	
2020-391	4	ISE UNDIL.	165	VITROS	63	KODAK	151	4.012	0.089	2.22	3.000	5.000	
2020-391	4	ISE UNDIL.	191	RX-DAYTONA	126	RANDOX	34	4.024	0.106	2.63	3.000	5.000	
2020-391	4	ISE UNDIL.	194	BIOLIS-24S	149	JAS	12	3.983	0.114	2.87	3.000	5.000	
2020-391	4	ISE UNDIL.	199	Rx-IMOLA	126	RANDOX	12	3.950	0.104	2.64	3.000	5.000	
2020-391	4	ISE UNDIL.	214	Envoy 500	150	ELITECH	10	4.000	0.045	1.12	3.000	5.000	
2020-391	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	82	4.024	0.075	1.87	3.000	5.000	
2020-391	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	4.000	0.093	2.31	3.000	5.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
022 POTASSIUM		CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE			1.00					
2020-392	3 ISE DILUT.	184 INTEGRA	93 ROCHE		37	5.762	0.160	2.78	4.800	6.800	
2020-392	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		28	5.725	0.095	1.66	4.700	6.700	
2020-392	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	5.741	0.124	2.16	4.700	6.700	
2020-392	3 ISE DILUT.	203 ROCHE	93 ROCHE		34	5.876	0.133	2.26	4.900	6.900	
2020-392	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	5.750	0.096	1.67	4.800	6.800	
2020-392	4 ISE UNDIL.	165 VITROS	63 KODAK		151	5.857	0.124	2.12	4.900	6.900	
2020-392	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		34	5.712	0.139	2.43	4.700	6.700	
2020-392	4 ISE UNDIL.	194 BIOLIS-24S	149 JAS		12	5.258	0.193	3.68	4.300	6.300	
2020-392	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	5.633	0.160	2.84	4.600	6.600	
2020-392	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	5.760	0.233	4.05	4.800	6.800	
2020-392	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		82	5.801	0.104	1.79	4.800	6.800	
2020-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	5.764	0.072	1.25	4.800	6.800	
2020-393	3 ISE DILUT.	184 INTEGRA	93 ROCHE		37	3.889	0.080	2.05	2.900	4.900	
2020-393	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		28	3.839	0.072	1.89	2.800	4.800	
2020-393	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	3.813	0.086	2.25	2.800	4.800	
2020-393	3 ISE DILUT.	203 ROCHE	93 ROCHE		34	3.942	0.074	1.88	2.900	4.900	
2020-393	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	3.758	0.111	2.97	2.800	4.800	
2020-393	4 ISE UNDIL.	165 VITROS	63 KODAK		151	3.861	0.090	2.33	2.900	4.900	
2020-393	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		34	3.871	0.118	3.04	2.900	4.900	
2020-393	4 ISE UNDIL.	194 BIOLIS-24S	149 JAS		12	3.900	0.141	3.63	2.900	4.900	
2020-393	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	3.792	0.111	2.94	2.800	4.800	
2020-393	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	3.830	0.127	3.31	2.800	4.800	
2020-393	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		82	3.844	0.082	2.13	2.800	4.800	
2020-393	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	3.843	0.098	2.55	2.800	4.800	
2020-394	3 ISE DILUT.	184 INTEGRA	93 ROCHE		37	3.492	0.067	1.93	2.500	4.500	
2020-394	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		28	3.457	0.056	1.63	2.500	4.500	
2020-394	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	3.394	0.135	3.97	2.400	4.400	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-394	3 ISE DILUT.	203 ROCHE	93 ROCHE		34	3.535	0.090	2.56	2.500	4.500	
2020-394	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	3.350	0.050	1.49	2.400	4.400	
2020-394	4 ISE UNDIL.	165 VITROS	63 KODAK		151	3.427	0.084	2.45	2.400	4.400	
2020-394	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		34	3.479	0.102	2.94	2.500	4.500	
2020-394	4 ISE UNDIL.	194 BIOLIS-24S	149 JAS		12	3.583	0.162	4.53	2.600	4.600	
2020-394	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	3.400	0.091	2.68	2.400	4.400	
2020-394	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	3.450	0.150	4.35	2.500	4.500	
2020-394	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		82	3.429	0.071	2.07	2.400	4.400	
2020-394	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	3.429	0.080	2.32	2.400	4.400	
022 POTASSIUM		CLIA TEST ID #: 465		EVALUATION CRITERIA: TARGET VALUE		1.00					
2020-395	3 ISE DILUT.	184 INTEGRA	93 ROCHE		37	4.928	0.080	1.63	3.900	5.900	
2020-395	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		28	4.854	0.068	1.40	3.900	5.900	
2020-395	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		17	4.847	0.119	2.46	3.800	5.800	
2020-395	3 ISE DILUT.	203 ROCHE	93 ROCHE		34	5.021	0.118	2.36	4.000	6.000	
2020-395	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		12	4.833	0.085	1.76	3.800	5.800	
2020-395	4 ISE UNDIL.	165 VITROS	63 KODAK		151	4.928	0.095	1.92	3.900	5.900	
2020-395	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		34	4.850	0.109	2.25	3.900	5.900	
2020-395	4 ISE UNDIL.	194 BIOLIS-24S	149 JAS		12	4.567	0.229	5.00	3.600	5.600	
2020-395	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	4.767	0.085	1.78	3.800	5.800	
2020-395	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	4.830	0.168	3.47	3.800	5.800	
2020-395	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		82	4.899	0.097	1.99	3.900	5.900	
2020-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	4.871	0.080	1.63	3.900	5.900	
023 SODIUM		CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE		4.00					
2020-391	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	130.939	2.707	2.07	127.000	135.000	
2020-391	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		28	130.357	2.057	1.58	126.000	134.000	
2020-391	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		12	130.000	1.354	1.04	126.000	134.000	
2020-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	128.429	2.060	1.60	124.000	132.000	
2020-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	131.625	2.368	1.80	128.000	136.000	
2020-391	3 ISE DILUTE	203 ROCHE	93 ROCHE		25	131.417	2.080	1.58	127.000	135.000	
2020-391	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	130.250	2.203	1.69	126.000	134.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-391	4 ISE UNDIL.	165 VITROS	63 KODAK		145	130.538	1.884	1.44	127.000	135.000	
2020-391	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		29	131.069	2.348	1.79	127.000	135.000	
2020-391	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	129.750	1.831	1.41	126.000	134.000	
2020-391	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	128.429	2.060	1.60	124.000	132.000	
2020-391	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	129.900	2.022	1.56	126.000	134.000	
2020-391	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		83	132.778	2.114	1.59	129.000	137.000	
2020-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	132.214	2.335	1.77	128.000	136.000	
2020-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	132.667	1.247	0.94	129.000	137.000	
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2020-392	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	153.219	2.678	1.75	149.000	157.000	
2020-392	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		28	154.286	2.281	1.48	150.000	158.000	
2020-392	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		12	145.083	1.935	1.33	141.000	149.000	
2020-392	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	155.250	2.926	1.88	151.000	159.000	
2020-392	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	150.143	2.231	1.49	146.000	154.000	
2020-392	3 ISE DILUTE	203 ROCHE	93 ROCHE		25	155.960	3.352	2.15	152.000	160.000	
2020-392	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	153.833	2.192	1.43	150.000	158.000	
2020-392	4 ISE UNDIL.	165 VITROS	63 KODAK		145	155.660	2.252	1.45	152.000	160.000	
2020-392	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		29	154.897	3.188	2.06	151.000	159.000	
2020-392	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	153.250	3.269	2.13	149.000	157.000	
2020-392	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	150.143	2.231	1.49	146.000	154.000	
2020-392	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	154.700	2.685	1.74	151.000	159.000	
2020-392	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		83	155.073	2.458	1.59	151.000	159.000	
2020-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	155.000	2.160	1.39	151.000	159.000	
2020-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	155.429	2.744	1.77	151.000	159.000	
2020-393	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	127.909	2.417	1.89	124.000	132.000	
2020-393	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		28	128.536	1.721	1.34	125.000	133.000	
2020-393	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		12	128.583	1.498	1.16	125.000	133.000	
2020-393	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	126.000	1.512	1.20	122.000	130.000	
2020-393	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	128.250	2.681	2.09	124.000	132.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-393	3 ISE DILUTE	203 ROCHE	93 ROCHE		25	129.280	1.800	1.39	125.000	133.000	
2020-393	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	128.000	2.944	2.30	124.000	132.000	
2020-393	4 ISE UNDIL.	165 VITROS	63 KODAK		145	128.438	1.917	1.49	124.000	132.000	
2020-393	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		29	128.586	2.205	1.72	125.000	133.000	
2020-393	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	127.083	2.139	1.68	123.000	131.000	
2020-393	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	128.100	2.343	1.83	124.000	132.000	
2020-393	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	126.000	1.512	1.20	122.000	130.000	
2020-393	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		83	129.988	1.856	1.43	126.000	134.000	
2020-393	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	129.786	2.305	1.78	126.000	134.000	
2020-393	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	130.667	1.247	0.95	127.000	135.000	
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2020-394	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	123.909	1.928	1.56	120.000	128.000	
2020-394	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		28	123.821	1.670	1.35	120.000	128.000	
2020-394	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		12	126.333	2.461	1.95	122.000	130.000	
2020-394	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	123.500	2.622	2.12	120.000	128.000	
2020-394	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	122.000	1.927	1.58	118.000	126.000	
2020-394	3 ISE DILUTE	203 ROCHE	93 ROCHE		25	124.560	1.722	1.38	121.000	129.000	
2020-394	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	123.750	1.479	1.20	120.000	128.000	
2020-394	4 ISE UNDIL.	165 VITROS	63 KODAK		145	123.703	1.894	1.53	120.000	128.000	
2020-394	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		29	123.483	2.284	1.85	119.000	127.000	
2020-394	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	121.667	2.687	2.21	118.000	126.000	
2020-394	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	122.000	1.927	1.58	118.000	126.000	
2020-394	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	122.900	2.773	2.26	119.000	127.000	
2020-394	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		83	125.793	1.974	1.57	122.000	130.000	
2020-394	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	126.333	0.471	0.37	122.000	130.000	
2020-394	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	125.286	1.868	1.49	121.000	129.000	
2020-395	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	141.758	2.257	1.59	138.000	146.000	
2020-395	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		28	142.179	1.794	1.26	138.000	146.000	
2020-395	3 ISE DILUTE	194 BIOLIS-24S	149 JAS		12	138.167	0.898	0.65	134.000	142.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-395	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	138.857	2.167	1.56	135.000	143.000	
2020-395	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		16	143.125	1.900	1.33	139.000	147.000	
2020-395	3 ISE DILUTE	203 ROCHE	93 ROCHE		25	144.120	2.438	1.69	140.000	148.000	
2020-395	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	141.750	1.689	1.19	138.000	146.000	
2020-395	4 ISE UNDIL.	165 VITROS	63 KODAK		145	142.725	1.773	1.24	139.000	147.000	
2020-395	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		29	142.448	2.908	2.04	138.000	146.000	
2020-395	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	141.083	1.847	1.31	137.000	145.000	
2020-395	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	141.400	2.154	1.52	137.000	145.000	
2020-395	4 ISE UNDIL.	214 Envoy 500	150 ELITECH		10	138.857	2.167	1.56	135.000	143.000	
2020-395	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		83	143.341	2.085	1.45	139.000	147.000	
2020-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	143.000	2.360	1.65	139.000	147.000	
2020-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	143.000	2.449	1.71	139.000	147.000	
024	TOTAL PROTEIN	CLIA TEST ID #: 485	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-391	1 BIURET	38 SIEMENS DI	143 SIEMENS		82	3.875	0.095	2.45	3.500	4.300	
2020-391	1 BIURET	38 SIEMENS DI	143 SIEMENS		82	3.833	0.149	3.89	3.500	4.200	
2020-391	1 BIURET	165 VITROS	63 KODAK		140	3.809	0.110	2.89	3.400	4.200	
2020-391	1 BIURET	184 INTEGRA	93 ROCHE		31	3.684	0.161	4.37	3.300	4.100	
2020-391	1 BIURET	186 BECK. COUL	76 AU-BECKMAN		24	3.667	0.094	2.57	3.300	4.000	
2020-391	1 BIURET	191 RX-DAYTONA	126 RANDOX		35	3.768	0.116	3.07	3.400	4.100	
2020-391	1 BIURET	197 ARCHITECT	1 ABBOTT		18	3.689	0.124	3.37	3.300	4.100	
2020-391	1 BIURET	199 Rx-IMOLA	126 RANDOX		11	3.709	0.090	2.43	3.300	4.100	
2020-391	1 BIURET	203 ROCHE	93 ROCHE		26	3.784	0.092	2.44	3.400	4.200	
2020-391	1 BIURET	212 S. D. EXL	143 SIEMENS		14	3.929	0.139	3.53	3.500	4.300	
2020-391	1 BIURET	214 Envoy 500	150 ELITECH		10	3.750	0.128	3.43	3.400	4.100	
2020-391	1 BIURET	215 B.C DXC	134 BECKM.COUL		12	3.733	0.155	4.14	3.400	4.100	
2020-392	1 BIURET	38 SIEMENS DI	143 SIEMENS		82	7.033	0.287	4.08	6.300	7.700	
2020-392	1 BIURET	38 SIEMENS DI	143 SIEMENS		82	7.156	0.197	2.75	6.400	7.900	
2020-392	1 BIURET	165 VITROS	63 KODAK		140	6.906	0.169	2.45	6.200	7.600	
2020-392	1 BIURET	184 INTEGRA	93 ROCHE		31	6.600	0.195	2.96	5.900	7.300	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-392	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	24	6.754	0.171	2.53	6.100	7.400
2020-392	1	BIURET	191	RX-DAYTONA	126	RANDOX	35	6.834	0.272	3.99	6.200	7.500
2020-392	1	BIURET	197	ARCHITECT	1	ABBOTT	18	6.944	0.171	2.46	6.300	7.600
2020-392	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	6.809	0.227	3.34	6.100	7.500
2020-392	1	BIURET	203	ROCHE	93	ROCHE	26	6.792	0.278	4.10	6.100	7.500
2020-392	1	BIURET	212	S. D. EXL	143	SIEMENS	14	7.243	0.311	4.29	6.500	8.000
2020-392	1	BIURET	214	Envoy 500	150	ELITECH	10	6.890	0.137	2.00	6.200	7.600
2020-392	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	6.925	0.205	2.96	6.200	7.600
024 TOTAL PROTEIN		CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %						
2020-393	1	BIURET	38	SIEMENS DI	143	SIEMENS	82	3.600	0.153	4.24	3.200	4.000
2020-393	1	BIURET	38	SIEMENS DI	143	SIEMENS	82	3.590	0.103	2.87	3.200	3.900
2020-393	1	BIURET	165	VITROS	63	KODAK	140	3.544	0.114	3.21	3.200	3.900
2020-393	1	BIURET	184	INTEGRA	93	ROCHE	31	3.390	0.173	5.10	3.100	3.700
2020-393	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	24	3.392	0.081	2.39	3.100	3.700
2020-393	1	BIURET	191	RX-DAYTONA	126	RANDOX	35	3.460	0.142	4.10	3.100	3.800
2020-393	1	BIURET	197	ARCHITECT	1	ABBOTT	18	3.378	0.118	3.50	3.000	3.700
2020-393	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	3.418	0.119	3.49	3.100	3.800
2020-393	1	BIURET	203	ROCHE	93	ROCHE	26	3.488	0.077	2.19	3.100	3.800
2020-393	1	BIURET	212	S. D. EXL	143	SIEMENS	14	3.621	0.137	3.79	3.300	4.000
2020-393	1	BIURET	214	Envoy 500	150	ELITECH	10	3.440	0.066	1.93	3.100	3.800
2020-393	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	3.433	0.131	3.82	3.100	3.800
2020-394	1	BIURET	38	SIEMENS DI	143	SIEMENS	82	2.760	0.105	3.80	2.500	3.000
2020-394	1	BIURET	38	SIEMENS DI	143	SIEMENS	82	2.767	0.094	3.41	2.500	3.000
2020-394	1	BIURET	165	VITROS	63	KODAK	140	2.746	0.089	3.24	2.500	3.000
2020-394	1	BIURET	184	INTEGRA	93	ROCHE	31	2.629	0.130	4.94	2.400	2.900
2020-394	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	24	2.638	0.075	2.86	2.400	2.900
2020-394	1	BIURET	191	RX-DAYTONA	126	RANDOX	35	2.683	0.116	4.32	2.400	3.000
2020-394	1	BIURET	197	ARCHITECT	1	ABBOTT	18	2.600	0.105	4.05	2.300	2.900
2020-394	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	2.636	0.098	3.71	2.400	2.900

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-394	1 BIURET	203	ROCHE	93	ROCHE	26	2.715	0.072	2.64	2.400	3.000
2020-394	1 BIURET	212	S. D. EXL	143	SIEMENS	14	2.807	0.116	4.14	2.500	3.100
2020-394	1 BIURET	214	Envoy 500	150	ELITECH	10	2.680	0.060	2.24	2.400	2.900
2020-394	1 BIURET	215	B.C DXC	134	BECKM.COUL	12	2.692	0.161	5.96	2.400	3.000
024	TOTAL PROTEIN	CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-395	1 BIURET	38	SIEMENS DI	143	SIEMENS	82	5.433	0.160	2.94	4.900	6.000
2020-395	1 BIURET	38	SIEMENS DI	143	SIEMENS	82	5.531	0.157	2.84	5.000	6.100
2020-395	1 BIURET	165	VITROS	63	KODAK	140	5.378	0.126	2.34	4.800	5.900
2020-395	1 BIURET	184	INTEGRA	93	ROCHE	31	5.252	0.285	5.43	4.700	5.800
2020-395	1 BIURET	186	BECK. COUL	76	AU-BECKMAN	24	5.213	0.105	2.02	4.700	5.700
2020-395	1 BIURET	191	RX-DAYTONA	126	RANDOX	35	5.331	0.214	4.00	4.800	5.900
2020-395	1 BIURET	197	ARCHITECT	1	ABBOTT	18	5.311	0.141	2.65	4.800	5.800
2020-395	1 BIURET	199	Rx-IMOLA	126	RANDOX	11	5.300	0.195	3.69	4.800	5.800
2020-395	1 BIURET	203	ROCHE	93	ROCHE	26	5.360	0.139	2.59	4.800	5.900
2020-395	1 BIURET	212	S. D. EXL	143	SIEMENS	14	5.543	0.164	2.95	5.000	6.100
2020-395	1 BIURET	214	Envoy 500	150	ELITECH	10	5.310	0.130	2.45	4.800	5.800
2020-395	1 BIURET	215	B.C DXC	134	BECKM.COUL	12	5.325	0.164	3.08	4.800	5.900
025	TRIGLYCERIDES	CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-391	2 ENZ. COLO.	165	VITROS	63	KODAK	143	128.664	4.960	3.85	96.000	161.000
2020-391	2 ENZ. COLO.	203	ROCHE	93	ROCHE	17	116.647	4.269	3.66	87.000	146.000
2020-391	4 ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	80	109.654	3.849	3.51	82.000	137.000
2020-391	4 ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	15	108.429	3.941	3.63	81.000	136.000
2020-391	9 GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	26	117.640	2.667	2.27	88.000	147.000
2020-391	9 GPO TRIND.	191	RX-DAYTONA	126	RANDOX	23	116.217	6.934	5.97	87.000	145.000
2020-391	9 GPO TRIND.	194	BIOLIS-24S	154	(IDG)	10	105.300	2.722	2.59	79.000	132.000
2020-391	9 GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	128.909	5.367	4.16	97.000	161.000
2020-391	10 ENZYM. GPO	184	INTEGRA	93	ROCHE	28	116.500	6.044	5.19	87.000	146.000
2020-391	10 ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	13	113.923	8.343	7.32	85.000	142.000
2020-391	10 ENZYM. GPO	203	ROCHE	93	ROCHE	16	117.875	4.196	3.56	88.000	147.000
2020-391	10 ENZYM. GPO	214	Envoy 500	150	ELITECH	11	114.909	6.171	5.37	86.000	144.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-391	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	10	112.500	2.419	2.15	84.000	141.000	
025	TRIGLYCERIDES	CLIA TEST ID #: 495		EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-392	2 ENZ. COLO.	165	VITROS	63	KODAK	143	232.085	6.738	2.90	174.000	290.000	
2020-392	2 ENZ. COLO.	203	ROCHE	93	ROCHE	17	193.647	6.361	3.28	145.000	242.000	
2020-392	4 ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	80	194.000	4.917	2.53	146.000	243.000	
2020-392	4 ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	15	192.286	4.861	2.53	144.000	240.000	
2020-392	9 GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	26	201.423	6.500	3.23	151.000	252.000	
2020-392	9 GPO TRIND.	191	RX-DAYTONA	126	RANDOX	23	194.043	7.805	4.02	146.000	243.000	
2020-392	9 GPO TRIND.	194	BIOLIS-24S	154	(IDG)	10	171.500	6.217	3.63	129.000	214.000	
2020-392	9 GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	213.091	7.465	3.50	160.000	266.000	
2020-392	10 ENZYM. GPO	184	INTEGRA	93	ROCHE	28	189.643	11.815	6.23	142.000	237.000	
2020-392	10 ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	13	189.692	9.400	4.96	142.000	237.000	
2020-392	10 ENZYM. GPO	203	ROCHE	93	ROCHE	16	191.250	6.750	3.53	143.000	239.000	
2020-392	10 ENZYM. GPO	214	Envoy 500	150	ELITECH	11	196.364	7.655	3.90	147.000	245.000	
2020-392	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	10	191.500	3.722	1.94	144.000	239.000	
2020-393	2 ENZ. COLO.	165	VITROS	63	KODAK	143	119.727	4.650	3.88	90.000	150.000	
2020-393	2 ENZ. COLO.	203	ROCHE	93	ROCHE	17	110.706	2.492	2.25	83.000	138.000	
2020-393	4 ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	80	101.692	3.436	3.38	76.000	127.000	
2020-393	4 ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	15	100.500	2.946	2.93	75.000	126.000	
2020-393	9 GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	26	109.308	3.439	3.15	82.000	137.000	
2020-393	9 GPO TRIND.	191	RX-DAYTONA	126	RANDOX	23	108.913	6.043	5.55	82.000	136.000	
2020-393	9 GPO TRIND.	194	BIOLIS-24S	154	(IDG)	10	97.800	1.661	1.70	73.000	122.000	
2020-393	9 GPO TRIND.	215	B.C DXC	134	BECKM.COUL	11	120.273	5.189	4.31	90.000	150.000	
2020-393	10 ENZYM. GPO	184	INTEGRA	93	ROCHE	28	109.179	5.645	5.17	82.000	136.000	
2020-393	10 ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	13	108.077	8.024	7.42	81.000	135.000	
2020-393	10 ENZYM. GPO	203	ROCHE	93	ROCHE	16	110.563	3.588	3.24	83.000	138.000	
2020-393	10 ENZYM. GPO	214	Envoy 500	150	ELITECH	11	107.455	4.639	4.32	81.000	134.000	
2020-393	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	10	103.400	2.871	2.78	78.000	129.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
025	TRIGLYCERIDES	CLIA TEST ID #: 495	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-394	2 ENZ. COLO.	165 VITROS	63 KODAK		143	93.720	3.582	3.82	70.000	117.000	
2020-394	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	90.412	2.568	2.84	68.000	113.000	
2020-394	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		80	79.885	3.101	3.88	60.000	100.000	
2020-394	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		15	78.286	3.554	4.54	59.000	98.000	
2020-394	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		26	88.280	2.050	2.32	66.000	110.000	
2020-394	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		23	88.000	5.779	6.57	66.000	110.000	
2020-394	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)		10	84.600	2.728	3.22	63.000	106.000	
2020-394	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		11	104.000	3.885	3.74	78.000	130.000	
2020-394	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		28	89.143	4.565	5.12	67.000	111.000	
2020-394	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		13	86.692	9.219	10.63	65.000	108.000	
2020-394	10 ENZYM. GPO	203 ROCHE	93 ROCHE		16	90.125	3.160	3.51	68.000	113.000	
2020-394	10 ENZYM. GPO	214 Envoy 500	150 ELITECH		11	87.364	5.381	6.16	66.000	109.000	
2020-394	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		10	83.300	1.552	1.86	62.000	104.000	
2020-395	2 ENZ. COLO.	165 VITROS	63 KODAK		143	179.741	6.201	3.45	135.000	225.000	
2020-395	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	157.176	3.222	2.05	118.000	196.000	
2020-395	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		80	151.731	3.761	2.48	114.000	190.000	
2020-395	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		15	150.571	3.353	2.23	113.000	188.000	
2020-395	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		26	159.280	3.726	2.34	119.000	199.000	
2020-395	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		23	156.391	6.787	4.34	117.000	195.000	
2020-395	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)		10	138.400	4.737	3.42	104.000	173.000	
2020-395	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		11	169.364	4.637	2.74	127.000	212.000	
2020-395	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		28	155.536	7.543	4.85	117.000	194.000	
2020-395	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		13	152.846	8.752	5.73	115.000	191.000	
2020-395	10 ENZYM. GPO	203 ROCHE	93 ROCHE		16	156.938	5.043	3.21	118.000	196.000	
2020-395	10 ENZYM. GPO	214 Envoy 500	150 ELITECH		11	157.273	6.383	4.06	118.000	197.000	
2020-395	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		10	151.700	3.407	2.25	114.000	190.000	
026	UREA NITROGEN (BUN)	CLIA TEST ID #: 505	EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}								
2020-391	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS		84	19.250	0.829	4.31	17.000	21.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-391	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	18.494	0.936	5.06	16.000	20.000
2020-391	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	40	18.175	0.771	4.24	16.000	20.000
2020-391	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	18.667	0.609	3.26	17.000	21.000
2020-391	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	18.143	0.833	4.59	16.000	20.000
2020-391	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	18.667	0.943	5.05	17.000	21.000
2020-391	2	UREAS.GLDH	203	ROCHE	93	ROCHE	34	18.147	0.733	4.04	16.000	20.000
2020-391	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	18.214	1.013	5.56	16.000	20.000
2020-391	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	18.200	0.748	4.11	16.000	20.000
2020-391	8	UREASE QUI	165	VITROS	63	KODAK	158	16.468	0.547	3.32	14.000	18.000
026 UREA NITROGEN (BUN)			CLIA TEST ID #: 505			EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}						
2020-392	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	36.867	1.471	3.99	34.000	40.000
2020-392	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	35.000	1.225	3.50	32.000	38.000
2020-392	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	40	36.000	1.533	4.26	33.000	39.000
2020-392	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	36.143	3.720	10.29	33.000	39.000
2020-392	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	36.926	1.086	2.94	34.000	40.000
2020-392	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	35.639	1.669	4.68	32.000	39.000
2020-392	2	UREAS.GLDH	203	ROCHE	93	ROCHE	34	35.971	1.403	3.90	33.000	39.000
2020-392	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	37.286	1.278	3.43	34.000	41.000
2020-392	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	35.700	1.269	3.55	32.000	39.000
2020-392	8	UREASE QUI	165	VITROS	63	KODAK	158	33.720	0.851	2.52	31.000	37.000
2020-393	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	17.250	0.829	4.81	15.000	19.000
2020-393	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	16.869	0.897	5.32	15.000	19.000
2020-393	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	40	16.350	0.726	4.44	14.000	18.000
2020-393	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	17.074	0.663	3.88	15.000	19.000
2020-393	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	16.286	0.700	4.30	14.000	18.000
2020-393	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	17.200	0.748	4.35	15.000	19.000
2020-393	2	UREAS.GLDH	203	ROCHE	93	ROCHE	34	16.735	0.656	3.92	15.000	19.000
2020-393	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	16.857	1.059	6.28	15.000	19.000
2020-393	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	16.400	0.663	4.04	14.000	18.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-393	8	UREASE QUI	165	VITROS	63	KODAK	158	15.000	0.490	3.27	13.000	17.000
026	UREA NITROGEN (BUN)		CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}							
2020-394	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	12.146	0.798	6.57	10.000	14.000
2020-394	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	12.500	0.866	6.93	11.000	15.000
2020-394	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	40	12.100	0.436	3.60	10.000	14.000
2020-394	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	12.429	0.728	5.86	10.000	14.000
2020-394	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	12.536	0.626	4.99	11.000	15.000
2020-394	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	12.528	0.833	6.65	11.000	15.000
2020-394	2	UREAS.GLDH	203	ROCHE	93	ROCHE	34	11.941	0.725	6.07	10.000	14.000
2020-394	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	12.385	0.836	6.75	10.000	14.000
2020-394	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	12.000	0.632	5.27	10.000	14.000
2020-394	8	UREASE QUI	165	VITROS	63	KODAK	158	10.892	0.512	4.70	9.000	13.000
2020-395	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	27.000	1.581	5.86	25.000	29.000
2020-395	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	84	27.738	1.292	4.66	25.000	30.000
2020-395	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	40	27.103	0.871	3.21	25.000	30.000
2020-395	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	27.815	0.772	2.77	25.000	30.000
2020-395	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	28	27.571	1.178	4.27	25.000	30.000
2020-395	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	36	27.528	1.343	4.88	25.000	30.000
2020-395	2	UREAS.GLDH	203	ROCHE	93	ROCHE	34	27.382	1.138	4.16	25.000	30.000
2020-395	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	27.429	1.178	4.29	25.000	30.000
2020-395	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	27.400	0.917	3.34	25.000	30.000
2020-395	8	UREASE QUI	165	VITROS	63	KODAK	158	24.853	0.658	2.65	23.000	27.000
027	URIC ACID		CLIA TEST ID #: 515		EVALUATION CRITERIA: TARGET VALUE 17 %							
2020-391	2	URICASE	38	SIEMENS DI	143	SIEMENS	48	4.543	0.125	2.75	3.800	5.300
2020-391	2	URICASE	165	VITROS	63	KODAK	94	4.481	0.106	2.36	3.700	5.200
2020-391	2	URICASE	184	INTEGRA	93	ROCHE	19	4.711	0.189	4.01	3.900	5.500
2020-391	2	URICASE	197	ARCHITECT	1	ABBOTT	10	4.670	0.142	3.04	3.900	5.500
2020-391	2	URICASE	203	ROCHE	93	ROCHE	23	4.445	0.127	2.86	3.700	5.200
2020-391	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	18	4.644	0.117	2.51	3.900	5.400

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-391	5 URICASE 37	194 BIOLIS-24S	154 (IDG)	10	5.240	0.128	2.44	4.300	6.100		
2020-391	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	11	4.364	0.077	1.77	3.600	5.100		
027 URIC ACID		CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %								
2020-392	2 URICASE	38 SIEMENS DI	143 SIEMENS	48	9.450	0.223	2.36	7.800	11.100		
2020-392	2 URICASE	165 VITROS	63 KODAK	94	9.350	0.218	2.33	7.800	10.900		
2020-392	2 URICASE	184 INTEGRA	93 ROCHE	19	9.721	0.286	2.94	8.100	11.400		
2020-392	2 URICASE	197 ARCHITECT	1 ABBOTT	10	9.660	0.201	2.08	8.000	11.300		
2020-392	2 URICASE	203 ROCHE	93 ROCHE	23	9.309	0.215	2.30	7.700	10.900		
2020-392	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	9.378	0.266	2.83	7.800	11.000		
2020-392	5 URICASE 37	194 BIOLIS-24S	154 (IDG)	10	10.250	0.269	2.63	8.500	12.000		
2020-392	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	11	8.718	0.119	1.37	7.200	10.200		
2020-393	2 URICASE	38 SIEMENS DI	143 SIEMENS	48	4.067	0.133	3.27	3.400	4.800		
2020-393	2 URICASE	165 VITROS	63 KODAK	94	4.031	0.099	2.46	3.300	4.700		
2020-393	2 URICASE	184 INTEGRA	93 ROCHE	19	4.216	0.153	3.63	3.500	4.900		
2020-393	2 URICASE	197 ARCHITECT	1 ABBOTT	10	4.180	0.140	3.35	3.500	4.900		
2020-393	2 URICASE	203 ROCHE	93 ROCHE	23	3.991	0.097	2.44	3.300	4.700		
2020-393	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	4.200	0.120	2.86	3.500	4.900		
2020-393	5 URICASE 37	194 BIOLIS-24S	154 (IDG)	10	4.840	0.092	1.89	4.000	5.700		
2020-393	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	11	3.927	0.114	2.89	3.300	4.600		
2020-394	2 URICASE	38 SIEMENS DI	143 SIEMENS	48	2.848	0.104	3.65	2.400	3.300		
2020-394	2 URICASE	165 VITROS	63 KODAK	94	2.749	0.092	3.36	2.300	3.200		
2020-394	2 URICASE	184 INTEGRA	93 ROCHE	19	2.937	0.113	3.86	2.400	3.400		
2020-394	2 URICASE	197 ARCHITECT	1 ABBOTT	10	2.960	0.102	3.45	2.500	3.500		
2020-394	2 URICASE	203 ROCHE	93 ROCHE	23	2.774	0.079	2.86	2.300	3.200		
2020-394	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	3.000	0.094	3.14	2.500	3.500		
2020-394	5 URICASE 37	194 BIOLIS-24S	154 (IDG)	10	3.480	0.117	3.35	2.900	4.100		
2020-394	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	11	2.700	0.043	1.58	2.200	3.200		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
027	URIC ACID	CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE	17	%						
2020-395	2 URICASE	38 SIEMENS DI	143 SIEMENS	48	6.994	0.160	2.29	5.800	8.200		
2020-395	2 URICASE	165 VITROS	63 KODAK	94	6.928	0.135	1.96	5.800	8.100		
2020-395	2 URICASE	184 INTEGRA	93 ROCHE	19	7.279	0.305	4.20	6.000	8.500		
2020-395	2 URICASE	197 ARCHITECT	1 ABBOTT	10	7.200	0.179	2.48	6.000	8.400		
2020-395	2 URICASE	203 ROCHE	93 ROCHE	23	6.909	0.198	2.86	5.700	8.100		
2020-395	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	7.011	0.188	2.68	5.800	8.200		
2020-395	5 URICASE 37	194 BIOLIS-24S	154 (IDG)	10	7.860	0.254	3.23	6.500	9.200		
2020-395	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	11	6.655	0.078	1.18	5.500	7.800		
135	GGT (GAMMA GLU. TRA)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	20	%						
2020-391	5 KINETIC 37	165 VITROS	63 KODAK	21	176.762	4.374	2.47	141.000	212.000		
2020-392	5 KINETIC 37	165 VITROS	63 KODAK	21	493.762	15.377	3.11	395.000	593.000		
2020-393	5 KINETIC 37	165 VITROS	63 KODAK	21	142.857	2.867	2.01	114.000	171.000		
2020-394	5 KINETIC 37	165 VITROS	63 KODAK	21	34.476	1.180	3.42	28.000	41.000		
2020-395	5 KINETIC 37	165 VITROS	63 KODAK	21	354.143	9.115	2.57	283.000	425.000		
136	LACTIC ACID										
2020-391	1 LACTATE-P.	165 VITROS	63 KODAK	18	1.822	0.063	3.45	1.500	2.200		
2020-392	1 LACTATE-P.	165 VITROS	63 KODAK	18	4.459	0.114	2.56	3.600	5.400		
2020-393	1 LACTATE-P.	165 VITROS	63 KODAK	18	1.600	0.058	3.61	1.300	1.900		
2020-394	1 LACTATE-P.	165 VITROS	63 KODAK	18	0.967	0.047	4.88	0.800	1.200		
2020-395	1 LACTATE-P.	165 VITROS	63 KODAK	18	3.256	0.117	3.58	2.600	3.900		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
137	LIPASE	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-391	1	COLORIMET.	203 ROCHE	93	ROCHE	14	29.643	3.497	11.80	24.000	36.000
2020-391	2	ENZIMATIC	165 VITROS	63	KODAK	78	298.487	11.825	3.96	239.000	358.000
2020-391	2	ENZIMATIC	215 B.C DXC	134	BECKM.COUL	12	38.167	2.703	7.08	31.000	46.000
2020-391	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	32.571	2.382	7.31	26.000	39.000
2020-391	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	32.375	0.992	3.06	26.000	39.000
2020-391	4	ENZ/COLOR	203 ROCHE	93	ROCHE	13	29.615	1.862	6.29	24.000	36.000
2020-391	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	20	121.600	11.155	9.17	97.000	146.000
2020-392	1	COLORIMET.	203 ROCHE	93	ROCHE	14	68.643	8.607	12.54	55.000	82.000
2020-392	2	ENZIMATIC	165 VITROS	63	KODAK	78	695.115	26.060	3.75	556.000	834.000
2020-392	2	ENZIMATIC	215 B.C DXC	134	BECKM.COUL	12	69.500	4.682	6.74	56.000	83.000
2020-392	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	79.286	4.495	5.67	63.000	95.000
2020-392	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	74.625	1.932	2.59	60.000	90.000
2020-392	4	ENZ/COLOR	203 ROCHE	93	ROCHE	13	65.923	6.776	10.28	53.000	79.000
2020-392	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	20	270.750	24.659	9.11	217.000	325.000
2020-393	1	COLORIMET.	203 ROCHE	93	ROCHE	14	26.643	3.061	11.49	21.000	32.000
2020-393	2	ENZIMATIC	165 VITROS	63	KODAK	78	260.154	10.486	4.03	208.000	312.000
2020-393	2	ENZIMATIC	215 B.C DXC	134	BECKM.COUL	12	34.667	2.560	7.39	28.000	42.000
2020-393	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	27.857	2.615	9.39	22.000	33.000
2020-393	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	28.125	1.053	3.74	23.000	34.000
2020-393	4	ENZ/COLOR	203 ROCHE	93	ROCHE	13	25.692	1.727	6.72	21.000	31.000
2020-393	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	20	109.150	10.036	9.19	87.000	131.000
2020-394	1	COLORIMET.	203 ROCHE	93	ROCHE	14	16.786	1.264	7.53	13.000	20.000
2020-394	2	ENZIMATIC	165 VITROS	63	KODAK	78	145.221	6.166	4.25	116.000	174.000
2020-394	2	ENZIMATIC	215 B.C DXC	134	BECKM.COUL	12	26.583	2.929	11.02	21.000	32.000
2020-394	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	15.929	2.120	13.31	13.000	19.000
2020-394	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	14	17.250	0.829	4.81	14.000	21.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-394	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	16.615	0.923	5.56	13.000	20.000
2020-394	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	20	72.300	9.435	13.05	58.000	87.000
137	LIPASE		CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %					
2020-395	1	COLORIMET.	203	ROCHE	93	ROCHE	14	49.286	6.192	12.56	39.000	59.000
2020-395	2	ENZIMATIC	165	VITROS	63	KODAK	78	512.077	19.144	3.74	410.000	614.000
2020-395	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	54.250	3.677	6.78	43.000	65.000
2020-395	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	53.375	1.654	3.10	43.000	64.000
2020-395	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	55.462	3.650	6.58	44.000	67.000
2020-395	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	47.615	4.412	9.27	38.000	57.000
2020-395	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	20	197.050	18.043	9.16	158.000	236.000
138	PHOSPHORUS											
2020-391	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	45	2.587	0.088	3.42	2.100	3.100
2020-391	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	2.667	0.107	4.00	2.100	3.200
2020-391	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	17	2.506	0.207	8.27	2.000	3.000
2020-391	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	2.367	0.135	5.70	1.900	2.800
2020-391	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	2.414	0.122	5.04	1.900	2.900
2020-391	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	2.517	0.099	3.92	2.000	3.000
2020-391	2	END POINT	184	INTEGRA	93	ROCHE	10	2.420	0.098	4.05	1.900	2.900
2020-392	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	45	5.340	0.125	2.35	4.300	6.400
2020-392	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	5.162	0.138	2.68	4.100	6.200
2020-392	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	17	5.135	0.274	5.34	4.100	6.200
2020-392	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	15	5.013	0.209	4.18	4.000	6.000
2020-392	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	29	5.155	0.251	4.88	4.100	6.200
2020-392	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	5.492	0.189	3.44	4.400	6.600
2020-392	2	END POINT	184	INTEGRA	93	ROCHE	10	5.030	0.119	2.36	4.000	6.000
2020-393	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	45	2.267	0.089	3.95	1.800	2.700
2020-393	1	PHOSPHOMO.	165	VITROS	63	KODAK	82	2.410	0.112	4.64	1.900	2.900
2020-393	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	17	2.241	0.257	11.46	1.800	2.700

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-393	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	15	2.120	0.138	6.49	1.700	2.500
2020-393	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	29	2.164	0.101	4.66	1.700	2.600
2020-393	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	12	2.242	0.111	4.97	1.800	2.700
2020-393	2	END POINT	184 INTEGRA	93	ROCHE	10	2.100	0.063	3.01	1.700	2.500
138	PHOSPHORUS	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %					
2020-394	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	45	1.452	0.084	5.78	1.200	1.700
2020-394	1	PHOSPHOMO.	165 VITROS	63	KODAK	82	1.593	0.112	7.03	1.300	1.900
2020-394	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	17	1.413	0.132	9.32	1.100	1.700
2020-394	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	15	1.327	0.100	7.52	1.100	1.600
2020-394	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	29	1.382	0.076	5.49	1.100	1.700
2020-394	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	12	1.375	0.123	8.97	1.100	1.700
2020-394	2	END POINT	184 INTEGRA	93	ROCHE	10	1.350	0.081	5.97	1.100	1.600
2020-395	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	45	3.978	0.103	2.59	3.200	4.800
2020-395	1	PHOSPHOMO.	165 VITROS	63	KODAK	82	3.985	0.117	2.94	3.200	4.800
2020-395	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	17	3.859	0.275	7.11	3.100	4.600
2020-395	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	15	3.727	0.169	4.54	3.000	4.500
2020-395	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	29	3.882	0.167	4.30	3.100	4.700
2020-395	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	12	4.067	0.149	3.67	3.300	4.900
2020-395	2	END POINT	184 INTEGRA	93	ROCHE	10	3.770	0.100	2.67	3.000	4.500
169	CO2					EVALUATION CRITERIA: TARGET VALUE 3 SD					
2020-391	1	Generic Me	165 VITROS	63	KODAK	32	12.188	1.704	13.98	7.000	17.000
2020-391	1	Generic Me	184 INTEGRA	93	ROCHE	15	12.267	0.998	8.13	9.000	15.000
2020-391	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	21	12.000	1.773	14.77	7.000	17.000
2020-391	1	Generic Me	203 ROCHE	93	ROCHE	15	11.600	1.665	14.36	7.000	17.000
2020-391	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	13.803	1.786	12.94	8.000	19.000
2020-391	2	ENZYMATIC	165 VITROS	63	KODAK	109	12.110	1.627	13.44	7.000	17.000
2020-391	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	12.455	1.373	11.02	8.000	17.000
2020-392	1	Generic Me	165 VITROS	63	KODAK	32	18.844	2.138	11.34	12.000	25.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-392	1	Generic Me	184 INTEGRA	93	ROCHE	15	19.200	2.561	13.34	12.000	27.000
2020-392	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	21	20.905	2.287	10.94	14.000	28.000
2020-392	1	Generic Me	203 ROCHE	93	ROCHE	15	20.200	1.327	6.57	16.000	24.000
2020-392	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	22.149	2.194	9.91	16.000	29.000
2020-392	2	ENZYMATIC	165 VITROS	63	KODAK	109	19.257	1.535	7.97	15.000	24.000
2020-392	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	20.364	2.267	11.13	14.000	27.000
169 CO2			CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-393	1	Generic Me	165 VITROS	63	KODAK	32	10.969	1.591	14.50	6.000	16.000
2020-393	1	Generic Me	184 INTEGRA	93	ROCHE	15	11.533	1.454	12.61	7.000	16.000
2020-393	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	21	11.400	1.068	9.37	8.000	15.000
2020-393	1	Generic Me	203 ROCHE	93	ROCHE	15	11.267	1.289	11.44	7.000	15.000
2020-393	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	13.030	1.736	13.32	8.000	18.000
2020-393	2	ENZYMATIC	165 VITROS	63	KODAK	109	11.303	1.536	13.59	7.000	16.000
2020-393	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	11.727	2.339	19.95	5.000	19.000
2020-394	1	Generic Me	165 VITROS	63	KODAK	32	10.063	1.435	14.26	6.000	14.000
2020-394	1	Generic Me	184 INTEGRA	93	ROCHE	15	10.267	1.236	12.04	7.000	14.000
2020-394	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	21	10.333	1.392	13.47	6.000	15.000
2020-394	1	Generic Me	203 ROCHE	93	ROCHE	15	9.533	1.784	18.71	4.000	15.000
2020-394	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	11.910	1.699	14.27	7.000	17.000
2020-394	2	ENZYMATIC	165 VITROS	63	KODAK	109	10.239	1.299	12.68	6.000	14.000
2020-394	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	10.455	1.671	15.99	5.000	15.000
2020-395	1	Generic Me	165 VITROS	63	KODAK	32	14.813	1.991	13.44	9.000	21.000
2020-395	1	Generic Me	184 INTEGRA	93	ROCHE	15	15.667	2.055	13.12	10.000	22.000
2020-395	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	21	16.000	1.976	12.35	10.000	22.000
2020-395	1	Generic Me	203 ROCHE	93	ROCHE	15	15.333	2.357	15.37	8.000	22.000
2020-395	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	17.672	2.003	11.33	12.000	24.000
2020-395	2	ENZYMATIC	165 VITROS	63	KODAK	109	15.239	1.489	9.77	11.000	20.000
2020-395	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	16.182	2.208	13.65	10.000	23.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
170	DIRECT BILIRUBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 20 %	{GREATER}					
2020-391	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.410	0.030	7.38	0.300	0.500		
2020-391	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.192	0.083	43.08	0	0.400		
2020-391	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.425	0.043	10.19	0.300	0.600		
2020-391	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	0.400	0.038	9.45	0.300	0.500		
2020-391	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	12	0.433	0.062	14.39	0.200	0.600		
2020-392	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	1.482	0.074	4.99	1.200	1.800		
2020-392	11 SPECT. WIT	165 VITROS	63 KODAK	55	1.475	0.130	8.81	1.100	1.900		
2020-392	11 SPECT. WIT	165 VITROS	63 KODAK	55	1.667	0.338	20.26	0.700	2.700		
2020-392	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	1.414	0.083	5.89	1.100	1.700		
2020-392	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	12	1.475	0.136	9.23	1.100	1.900		
2020-393	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.300	0.000	0.00	0.200	0.400		
2020-393	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.075	0.095	125.85	0	0.400		
2020-393	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.275	0.043	15.75	0.100	0.400		
2020-393	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	0.307	0.026	8.38	0.200	0.400		
2020-393	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	12	0.325	0.060	18.31	0.100	0.500		
2020-394	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.074	0.044	59.27	0	0.200		
2020-394	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.100	0.000	0.00	0.100	0.100		
2020-394	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.000	0.000	0.00	0.000	0.000		
2020-394	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	0.200	0.000	0.00	0.200	0.200		
2020-394	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	12	0.067	0.047	70.71	0	0.200		
2020-395	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.971	0.057	5.88	0.800	1.200		
2020-395	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.981	0.235	23.94	0.300	1.700		
2020-395	11 SPECT. WIT	165 VITROS	63 KODAK	55	0.975	0.083	8.50	0.700	1.200		
2020-395	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	0.960	0.061	6.36	0.800	1.200		
2020-395	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	12	0.975	0.083	8.50	0.700	1.200		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 20 BLOOD GASES

GRP: 30 BLOOD GASES

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
007	BLOOD GASES-pCO2	CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE	8 % OR 5.00 {GREATER}						
2020-441	260 SIEMENS				10	52.790	2.467	4.67	47.800	57.800
2020-441	260 SIEMENS				10	49.988	2.389	4.78	45.000	55.000
2020-441	265 ROCHE COBA				30	54.479	1.227	2.25	49.500	59.500
2020-442	260 SIEMENS				10	56.850	2.319	4.08	51.900	61.900
2020-442	260 SIEMENS				10	61.270	2.305	3.76	56.300	66.300
2020-442	265 ROCHE COBA				30	63.690	1.349	2.12	58.600	68.800
2020-443	260 SIEMENS				10	51.840	2.465	4.76	46.800	56.800
2020-443	260 SIEMENS				10	48.400	4.044	8.36	43.400	53.400
2020-443	265 ROCHE COBA				30	52.697	3.154	5.99	47.700	57.700
2020-444	260 SIEMENS				10	56.825	2.049	3.61	51.800	61.800
2020-444	260 SIEMENS				10	60.490	2.144	3.54	55.500	65.500
2020-444	265 ROCHE COBA				30	63.963	1.322	2.07	58.800	69.100
2020-445	260 SIEMENS				10	78.460	3.077	3.92	72.200	84.700
2020-445	260 SIEMENS				10	71.625	2.291	3.20	65.900	77.400
2020-445	265 ROCHE COBA				30	81.713	1.968	2.41	75.200	88.300
008	BLOOD GASES-pH	CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04						
2020-441	260 SIEMENS				10	7.365	0.020	0.27	7.325	7.405
2020-441	265 ROCHE COBA				30	7.326	0.017	0.24	7.286	7.366
2020-442	260 SIEMENS				10	7.214	0.005	0.07	7.174	7.254
2020-442	265 ROCHE COBA				30	7.194	0.007	0.09	7.154	7.234

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04					
2020-443	260	SIEMENS			10	7.365	0.019	0.26	7.325	7.405
2020-443	265	ROCHE COBA			30	7.333	0.016	0.22	7.293	7.373
2020-444	260	SIEMENS			10	7.215	0.004	0.05	7.175	7.255
2020-444	265	ROCHE COBA			30	7.193	0.007	0.10	7.153	7.233
2020-445	260	SIEMENS			10	7.116	0.005	0.08	7.076	7.156
2020-445	265	ROCHE COBA			30	7.099	0.006	0.09	7.059	7.139
009	BLOOD GASES-pO2		CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE	3 SD					
2020-441	260	SIEMENS			10	124.750	10.277	8.24	93.900	155.600
2020-441	265	ROCHE COBA			31	127.958	8.353	6.53	102.900	153.000
2020-442	260	SIEMENS			10	314.460	30.771	9.79	222.100	406.800
2020-442	265	ROCHE COBA			31	326.145	15.582	4.78	279.400	372.900
2020-443	260	SIEMENS			10	125.900	9.339	7.42	97.900	153.900
2020-443	265	ROCHE COBA			31	124.277	8.368	6.73	99.200	149.400
2020-444	260	SIEMENS			10	311.090	26.376	8.48	232.000	390.200
2020-444	265	ROCHE COBA			31	330.352	19.871	6.02	270.700	390.000
2020-445	260	SIEMENS			10	89.990	23.312	25.91	20.100	159.900
2020-445	265	ROCHE COBA			31	76.493	12.303	16.08	39.600	113.400
SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-421	116 ROCHE COBA			15	18.773	1.313	6.99	14.800	22.700	
2020-421	125 B.C.ACCESS			10	22.160	1.300	5.87	18.300	26.100	
2020-422	116 ROCHE COBA			15	9.627	0.748	7.77	7.400	11.900	
2020-422	125 B.C.ACCESS			10	11.060	0.655	5.92	9.100	13.000	
2020-423	116 ROCHE COBA			15	11.160	0.879	7.88	8.500	13.800	
2020-423	125 B.C.ACCESS			10	12.580	0.623	4.95	10.700	14.400	
2020-424	116 ROCHE COBA			15	26.113	1.860	7.12	20.500	31.700	
2020-424	125 B.C.ACCESS			10	30.940	1.456	4.71	26.600	35.300	
2020-425	116 ROCHE COBA			15	5.687	0.469	8.24	4.300	7.100	
2020-425	125 B.C.ACCESS			10	6.350	0.467	7.36	4.900	7.800	
030	FREE THYROXINE	CLIA TEST ID #: 545								
2020-421	115 VITROS ECI			32	5.281	0.153	2.90	4.800	5.700	
2020-421	115 VITROS ECI			32	1.967	0.047	2.40	1.800	2.100	
2020-421	116 ROCHE COBA			44	2.293	0.170	7.40	1.800	2.800	
2020-421	125 B.C.ACCESS			43	1.956	0.121	6.17	1.600	2.300	
2020-421	149 ARC i2000s			21	1.857	0.118	6.34	1.500	2.200	
2020-421	160 SIEMENS EX			43	2.302	0.147	6.39	1.900	2.700	
2020-421	160 SIEMENS EX			43	1.967	0.047	2.40	1.800	2.100	
2020-422	115 VITROS ECI			32	0.933	0.047	5.05	0.800	1.100	
2020-422	115 VITROS ECI			32	2.668	0.089	3.35	2.400	2.900	
2020-422	116 ROCHE COBA			44	1.288	0.110	8.50	1.000	1.600	
2020-422	125 B.C.ACCESS			43	1.116	0.080	7.21	0.900	1.400	
2020-422	149 ARC i2000s			21	0.957	0.058	6.09	0.800	1.100	
2020-422	160 SIEMENS EX			43	1.227	0.063	5.10	1.000	1.400	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-422	160	SIEMENS EX		43	0.933	0.047	5.05	0.800	1.100	
030	FREE THYROXINE		CLIA TEST ID #: 545							EVALUATION CRITERIA: TARGET VALUE 3 SD
2020-423	115	VITROS ECI		32	3.126	0.095	3.04	2.800	3.400	
2020-423	115	VITROS ECI		32	1.067	0.094	8.84	0.800	1.300	
2020-423	116	ROCHE COBA		44	1.445	0.116	8.02	1.100	1.800	
2020-423	125	B.C.ACCESS		43	1.253	0.095	7.57	1.000	1.500	
2020-423	149	ARC i2000s		21	1.085	0.048	4.40	0.900	1.200	
2020-423	160	SIEMENS EX		43	1.385	0.084	6.08	1.100	1.600	
2020-423	160	SIEMENS EX		43	1.067	0.094	8.84	0.800	1.300	
2020-424	115	VITROS ECI		32	3.033	0.094	3.11	2.800	3.300	
2020-424	115	VITROS ECI		32	6.674	0.134	2.01	6.300	7.100	
2020-424	116	ROCHE COBA		44	3.200	0.228	7.12	2.500	3.900	
2020-424	125	B.C.ACCESS		43	2.577	0.148	5.73	2.100	3.000	
2020-424	149	ARC i2000s		21	2.465	0.215	8.73	1.800	3.100	
2020-424	160	SIEMENS EX		43	3.272	0.207	6.33	2.700	3.900	
2020-424	160	SIEMENS EX		43	3.033	0.094	3.11	2.800	3.300	
2020-425	115	VITROS ECI		32	1.363	0.070	5.11	1.200	1.600	
2020-425	115	VITROS ECI		32	0.567	0.047	8.32	0.400	0.700	
2020-425	116	ROCHE COBA		44	0.783	0.072	9.20	0.600	1.000	
2020-425	125	B.C.ACCESS		43	0.742	0.062	8.34	0.600	0.900	
2020-425	149	ARC i2000s		21	0.645	0.050	7.71	0.500	0.800	
2020-425	160	SIEMENS EX		43	0.795	0.058	7.24	0.600	1.000	
2020-425	160	SIEMENS EX		43	0.567	0.047	8.32	0.400	0.700	
032	hCG-QUANTITATIVE		CLIA TEST ID #: 555							
2020-421	115	VITROS ECI		17	1194.112	68.268	5.72	989.300	1398.900	
2020-421	116	ROCHE COBA		20	666.210	35.259	5.29	560.400	772.000	
2020-421	125	B.C.ACCESS		11	455.973	21.803	4.78	390.600	521.400	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-422	115 VITROS ECI			17	398.871	21.242	5.33	335.100	462.600	
2020-422	116 ROCHE COBA			20	208.915	10.204	4.88	178.300	239.500	
2020-422	125 B.C.ACCESS			11	178.609	8.820	4.94	152.200	205.100	
2020-423	115 VITROS ECI			17	519.694	30.547	5.88	428.100	611.300	
2020-423	116 ROCHE COBA			20	279.360	15.962	5.71	231.500	327.200	
2020-423	125 B.C.ACCESS			11	232.491	13.017	5.60	193.400	271.500	
2020-424	115 VITROS ECI			17	1821.459	120.588	6.62	1459.700	2183.200	
2020-424	116 ROCHE COBA			20	1049.590	46.612	4.44	909.800	1189.400	
2020-424	125 B.C.ACCESS			11	650.545	31.946	4.91	554.700	746.400	
2020-425	115 VITROS ECI			17	47.271	3.125	6.61	37.900	56.600	
2020-425	116 ROCHE COBA			20	22.905	1.146	5.00	19.500	26.300	
2020-425	125 B.C.ACCESS			11	22.382	0.664	2.97	20.400	24.400	
034	PROGESTERONE	CLIA TEST ID #: 0								
2020-421	116 ROCHE COBA			10	18.270	1.138	6.23	14.900	21.700	
2020-422	116 ROCHE COBA			10	7.640	0.599	7.84	5.800	9.400	
2020-423	116 ROCHE COBA			10	9.600	0.564	5.87	7.900	11.300	
2020-424	116 ROCHE COBA			10	25.030	1.661	6.64	20.000	30.000	
2020-425	116 ROCHE COBA			10	2.100	0.290	13.80	1.200	3.000	
035	PROLACTIN									
2020-421	116 ROCHE COBA			17	19.006	1.238	6.52	15.300	22.700	
2020-421	125 B.C.ACCESS			10	17.900	0.829	4.63	15.400	20.400	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
035	PROLACTIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE	3	SD			
2020-422	116 ROCHE COBA			17	7.935	0.498	6.28	6.400	9.400	
2020-422	125 B.C.ACCESS			10	7.610	0.305	4.01	6.700	8.500	
2020-423	116 ROCHE COBA			17	9.600	0.644	6.70	7.700	11.500	
2020-423	125 B.C.ACCESS			10	9.290	0.333	3.58	8.300	10.300	
2020-424	116 ROCHE COBA			17	28.312	1.785	6.30	23.000	33.700	
2020-424	125 B.C.ACCESS			10	26.070	1.205	4.62	22.500	29.700	
2020-425	116 ROCHE COBA			17	3.306	0.231	7.00	2.600	4.000	
2020-425	125 B.C.ACCESS			10	2.940	0.215	7.33	2.300	3.600	
037	TESTOSTERONE									
2020-421	116 ROCHE COBA			10	9.956	0.854	8.58	7.400	12.500	
2020-422	116 ROCHE COBA			10	4.956	1.444	29.14	0.600	9.300	
2020-423	116 ROCHE COBA			10	5.233	0.386	7.37	4.100	6.400	
2020-424	116 ROCHE COBA			10	13.222	1.061	8.03	10.000	16.400	
2020-425	116 ROCHE COBA			10	1.389	0.145	10.43	1.000	1.800	
039	THYRO. STI. HOR.-TSH		CLIA TEST ID #: 585							
2020-421	32 DADE DIME.			10	3.320	0.316	9.51	2.400	4.300	
2020-421	106 TOSOH AIA			13	3.885	0.330	8.50	2.900	4.900	
2020-421	115 VITROS ECI			37	4.197	0.150	3.57	3.700	4.600	
2020-421	116 ROCHE COBA			54	3.549	0.153	4.30	3.100	4.000	
2020-421	125 B.C.ACCESS			60	3.333	0.188	5.63	2.800	3.900	
2020-421	149 ARC i2000s			22	3.255	0.159	4.88	2.800	3.700	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-421	160 SIEMENS EX			69	2.913	0.178	6.10	2.400	3.400
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-422	32 DADE DIME.			10	1.450	0.120	8.30	1.100	1.800
2020-422	106 TOSOH AIA			13	1.754	0.139	7.94	1.300	2.200
2020-422	115 VITROS ECI			37	1.862	0.094	5.05	1.600	2.100
2020-422	116 ROCHE COBA			54	1.740	0.098	5.62	1.400	2.000
2020-422	125 B.C.ACCESS			60	1.450	0.090	6.23	1.200	1.700
2020-422	149 ARC i2000s			22	1.432	0.082	5.72	1.200	1.700
2020-422	160 SIEMENS EX			69	1.355	0.096	7.06	1.100	1.600
2020-423	32 DADE DIME.			10	1.800	0.155	8.61	1.300	2.300
2020-423	106 TOSOH AIA			13	2.085	0.192	9.19	1.500	2.700
2020-423	115 VITROS ECI			37	2.246	0.100	4.46	1.900	2.500
2020-423	116 ROCHE COBA			54	2.030	0.109	5.38	1.700	2.400
2020-423	125 B.C.ACCESS			60	1.742	0.099	5.71	1.400	2.000
2020-423	149 ARC i2000s			22	1.695	0.098	5.76	1.400	2.000
2020-423	160 SIEMENS EX			69	1.610	0.101	6.27	1.300	1.900
2020-424	32 DADE DIME.			10	5.040	0.727	14.42	2.900	7.200
2020-424	106 TOSOH AIA			13	5.454	0.489	8.97	4.000	6.900
2020-424	115 VITROS ECI			37	6.014	0.186	3.09	5.500	6.600
2020-424	116 ROCHE COBA			54	4.932	0.255	5.16	4.200	5.700
2020-424	125 B.C.ACCESS			60	4.813	0.300	6.24	3.900	5.700
2020-424	149 ARC i2000s			22	4.623	0.281	6.08	3.800	5.500
2020-424	160 SIEMENS EX			69	4.113	0.248	6.03	3.400	4.900
2020-425	32 DADE DIME.			10	0.650	0.067	10.32	0.400	0.900
2020-425	106 TOSOH AIA			13	0.762	0.074	9.69	0.500	1.000
2020-425	115 VITROS ECI			37	0.732	0.047	6.39	0.600	0.900
2020-425	116 ROCHE COBA			54	0.883	0.066	7.47	0.700	1.100

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-425	125	B.C.ACCESS		60	0.667	0.047	7.07	0.500	0.800	
2020-425	149	ARC i2000s		22	0.627	0.045	7.10	0.500	0.800	
2020-425	160	SIEMENS EX		69	0.647	0.063	9.73	0.500	0.800	
040 THYROXINE (T4)			CLIA TEST ID #: 595				EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}			
2020-421	32	DADE DIME.		53	9.987	0.624	6.25	8.000	12.000	
2020-421	106	TOSOH AIA		10	8.350	0.813	9.73	6.700	10.000	
2020-421	115	VITROS ECI		32	8.313	0.414	4.98	6.700	10.000	
2020-421	116	ROCHE COBA		51	8.590	0.489	5.69	6.900	10.300	
2020-421	125	B.C.ACCESS		51	8.039	0.551	6.85	6.400	9.600	
2020-421	149	ARC i2000s		19	8.453	0.315	3.73	6.800	10.100	
2020-422	32	DADE DIME.		53	5.113	0.524	10.24	4.100	6.100	
2020-422	106	TOSOH AIA		10	4.350	0.396	9.09	3.400	5.400	
2020-422	115	VITROS ECI		32	4.397	0.207	4.71	3.400	5.400	
2020-422	116	ROCHE COBA		51	4.565	0.243	5.32	3.600	5.600	
2020-422	125	B.C.ACCESS		51	4.100	0.395	9.64	3.100	5.100	
2020-422	149	ARC i2000s		19	3.958	0.237	5.99	3.000	5.000	
2020-423	32	DADE DIME.		53	5.926	0.534	9.01	4.700	7.100	
2020-423	106	TOSOH AIA		10	4.870	0.316	6.50	3.900	5.900	
2020-423	115	VITROS ECI		32	5.003	0.252	5.04	4.000	6.000	
2020-423	116	ROCHE COBA		51	5.235	0.291	5.56	4.200	6.300	
2020-423	125	B.C.ACCESS		51	4.749	0.437	9.21	3.700	5.700	
2020-423	149	ARC i2000s		19	4.626	0.279	6.03	3.600	5.600	
2020-424	32	DADE DIME.		53	13.860	0.621	4.48	11.100	16.600	
2020-424	106	TOSOH AIA		10	10.930	1.022	9.35	8.700	13.100	
2020-424	115	VITROS ECI		32	11.269	0.624	5.54	9.000	13.500	
2020-424	116	ROCHE COBA		51	11.845	0.614	5.18	9.500	14.200	
2020-424	125	B.C.ACCESS		51	10.849	0.677	6.24	8.700	13.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-424	149	ARC i2000s		19	12.437	0.670	5.39	9.900	14.900	
040	THYROXINE (T4)		CLIA TEST ID #: 595		EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}					
2020-425	32	DADE DIME.		53	3.138	0.396	12.62	2.100	4.100	
2020-425	106	TOSOH AIA		10	2.750	0.216	7.84	1.800	3.800	
2020-425	115	VITROS ECI		32	2.665	0.103	3.88	1.700	3.700	
2020-425	116	ROCHE COBA		51	2.865	0.186	6.48	1.900	3.900	
2020-425	125	B.C.ACCESS		51	2.144	0.270	12.59	1.100	3.100	
2020-425	149	ARC i2000s		19	2.347	0.366	15.60	1.300	3.300	
141	FOLATE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %					
2020-421	115	VITROS ECI		12	6.425	0.849	13.21	4.800	8.000	
2020-421	116	ROCHE COBA		16	4.675	0.501	10.71	3.500	5.800	
2020-421	125	B.C.ACCESS		12	6.508	0.514	7.90	4.900	8.100	
2020-421	149	ARC i2000s		10	7.380	0.389	5.28	5.500	9.200	
2020-421	160	SIEMENS EX		10	3.980	0.489	12.30	3.000	5.000	
2020-422	115	VITROS ECI		12	2.733	0.359	13.13	2.100	3.400	
2020-422	116	ROCHE COBA		16	2.020	0.054	2.68	1.500	2.500	
2020-422	125	B.C.ACCESS		12	2.692	0.189	7.03	2.000	3.400	
2020-422	149	ARC i2000s		10	2.750	0.441	16.04	2.100	3.400	
2020-422	160	SIEMENS EX		10	1.870	0.261	13.96	1.400	2.300	
2020-423	115	VITROS ECI		12	3.200	0.342	10.67	2.400	4.000	
2020-423	116	ROCHE COBA		16	2.100	0.215	10.24	1.600	2.600	
2020-423	125	B.C.ACCESS		12	3.392	0.425	12.54	2.500	4.200	
2020-423	149	ARC i2000s		10	3.170	0.372	11.72	2.400	4.000	
2020-423	160	SIEMENS EX		10	2.220	0.319	14.36	1.700	2.800	
2020-424	115	VITROS ECI		12	10.967	1.537	14.01	8.200	13.700	
2020-424	116	ROCHE COBA		16	8.069	0.573	7.10	6.100	10.100	
2020-424	125	B.C.ACCESS		12	9.783	0.833	8.52	7.300	12.200	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-424	149	ARC i2000s		10	11.830	1.036	8.76	8.900	14.800	
2020-424	160	SIEMENS EX		10	5.690	0.653	11.48	4.300	7.100	
141	FOLATE			CLIA TEST ID #: 0 EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-425	115	VITROS ECI		12	1.642	0.275	16.77	1.200	2.100	
2020-425	116	ROCHE COBA		16	2.000	0.000	0.00	1.500	2.500	
2020-425	125	B.C.ACCESS		12	1.233	0.137	11.14	0.900	1.500	
2020-425	149	ARC i2000s		10	1.150	0.634	55.17	0.900	1.400	
2020-425	160	SIEMENS EX		10	1.060	0.215	20.32	0.800	1.300	
142	hLH (HUMAN LUT. HOR)									
2020-421	116	ROCHE COBA		15	30.960	1.781	5.75	23.200	38.700	
2020-421	125	B.C.ACCESS		10	23.870	1.385	5.80	17.900	29.800	
2020-422	116	ROCHE COBA		15	13.080	0.731	5.59	9.800	16.400	
2020-422	125	B.C.ACCESS		10	9.630	0.742	7.70	7.200	12.000	
2020-423	116	ROCHE COBA		15	16.093	0.866	5.38	12.100	20.100	
2020-423	125	B.C.ACCESS		10	11.850	0.856	7.22	8.900	14.800	
2020-424	116	ROCHE COBA		15	45.000	2.466	5.48	33.800	56.300	
2020-424	125	B.C.ACCESS		10	33.230	2.314	6.96	24.900	41.500	
2020-425	116	ROCHE COBA		15	4.807	0.345	7.19	3.600	6.000	
2020-425	125	B.C.ACCESS		10	3.380	0.244	7.22	2.500	4.200	

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)								
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART #	LABS #	"SF"	Score	
SAMPLE	RESULT	RESULT	RESULT	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
031 hCG-QUALITATIVE												
2020-421	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-421	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-421	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	84	84	100	
2020-421	164	HCG Alere		002 002	002 002		002 - POSITIVE	66	66	66	100	
2020-421	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-421	167	Tanner Sci		002 002	002 002		002 - POSITIVE	18	18	18	100	
	031	hCG-QUALITATIVE										
2020-422	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-422	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-422	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	84	84	100	
2020-422	164	HCG Alere		002 002	002 002		002 - POSITIVE	66	66	66	100	
2020-422	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-422	167	Tanner Sci		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-423	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-423	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-423	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	84	84	100	
2020-423	164	HCG Alere		002 002	002 002		002 - POSITIVE	66	66	66	100	
2020-423	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-423	167	Tanner Sci		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-424	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-424	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	48	48	100	
2020-424	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	84	84	100	
2020-424	164	HCG Alere		002 002	002 002		002 - POSITIVE	66	66	66	100	
2020-424	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-424	167	Tanner Sci		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-425	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-425	142	INSTANT VI		002 002	002 002		001 - NEGATIVE	48	1	47	0	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-425	142	INSTANT VI		002 002	002 002		002 - POSITIVE	48	47	47	100	
2020-425	150	AIM STEP		002 002	002 002		001 - NEGATIVE	84	2	82	0	
2020-425	150	AIM STEP		002 002	002 002		002 - POSITIVE	84	82	82	100	
2020-425	164	HCG Alere		002 002	002 002		002 - POSITIVE	66	66	66	100	
2020-425	165	Osom HCG C		002 002	002 002		002 - POSITIVE	18	18	18	100	
2020-425	167	Tanner Sci		002 002	002 002		002 - POSITIVE	18	17	17	100	
2020-425	167	Tanner Sci		002 002	002 002		001 - NEGATIVE	18	1	17	0	

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2020-411	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-411	083	STAMBIO		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-411	112	ONE STEP D		002 002	002 002		002 - POSITIVE	12	12	12	100	
2020-411	114	INSTANT VI		002 002	002 002		002 - POSITIVE	123	123	123	100	
2020-411	118	AIM STEP		002 002	002 002		002 - POSITIVE	143	143	143	100	
2020-411	130	ALERE HCG		002 002	002 002		002 - POSITIVE	88	88	88	100	
2020-411	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100	
2020-411	133	FaStep		002 002	002 002		002 - POSITIVE	17	17	17	100	
2020-411	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	67	67	67	100	
2020-412	018	BIO SIGN		001 001	001 001		001 - NEGATIVE	21	21	21	100	
2020-412	083	STAMBIO		001 001	001 001		001 - NEGATIVE	10	10	10	100	
2020-412	112	ONE STEP D		001 001	001 001		001 - NEGATIVE	12	12	12	100	
2020-412	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	123	123	123	100	
2020-412	118	AIM STEP		001 001	001 001		001 - NEGATIVE	143	143	143	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-412	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	88	87	87	100	
2020-412	130	ALERE HCG		001 001	001 001		002 - POSITIVE	88	1	87	0	
2020-412	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2020-412	133	FaStep		001 001	001 001		001 - NEGATIVE	17	17	17	100	
2020-412	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	67	67	67	100	
151 hCG ONLY QUALITATIVE												
2020-413	018	BIO SIGN		001 001	001 001		001 - NEGATIVE	21	21	21	100	
2020-413	083	STAMBIO		001 001	001 001		001 - NEGATIVE	10	10	10	100	
2020-413	112	ONE STEP D		001 001	001 001		001 - NEGATIVE	12	12	12	100	
2020-413	114	INSTANT VI		001 001	001 001		002 - POSITIVE	123	1	122	0	
2020-413	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	123	122	122	100	
2020-413	118	AIM STEP		001 001	001 001		001 - NEGATIVE	143	143	143	100	
2020-413	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	88	88	88	100	
2020-413	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2020-413	133	FaStep		001 001	001 001		001 - NEGATIVE	17	17	17	100	
2020-413	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	67	67	67	100	
2020-414	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-414	083	STAMBIO		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-414	112	ONE STEP D		002 002	002 002		002 - POSITIVE	12	12	12	100	
2020-414	114	INSTANT VI		002 002	002 002		002 - POSITIVE	123	122	122	100	
2020-414	114	INSTANT VI		002 002	002 002		001 - NEGATIVE	123	1	122	0	
2020-414	118	AIM STEP		002 002	002 002		001 - NEGATIVE	143	1	142	0	
2020-414	118	AIM STEP		002 002	002 002		002 - POSITIVE	143	142	142	100	
2020-414	130	ALERE HCG		002 002	002 002		002 - POSITIVE	88	88	88	100	
2020-414	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100	
2020-414	133	FaStep		002 002	002 002		002 - POSITIVE	17	17	17	100	
2020-414	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	67	67	67	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2020-415	018	BIO SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100
2020-415	083	STAMBIO		002	002	002	002	002 - POSITIVE	10	10	10	100
2020-415	112	ONE STEP D		002	002	002	002	002 - POSITIVE	12	12	12	100
2020-415	114	INSTANT VI		002	002	002	002	002 - POSITIVE	123	122	122	100
2020-415	114	INSTANT VI		002	002	002	002	001 - NEGATIVE	123	1	122	0
2020-415	118	AIM STEP		002	002	002	002	002 - POSITIVE	143	143	143	100
2020-415	130	ALERE HCG		002	002	002	002	002 - POSITIVE	88	88	88	100
2020-415	131	OSOM HCG		002	002	002	002	002 - POSITIVE	25	25	25	100
2020-415	133	FaStep		002	002	002	002	002 - POSITIVE	17	17	17	100
2020-415	134	Tanner Sc.		002	002	002	002	002 - POSITIVE	67	67	67	100

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)					
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. LABS	PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
173 T3 UPTAKE												
CLIA TEST ID #: 565				EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-421	1	%U	84 VITROS ECI	1	27	25-35	52.970	1.520	2.87	48.410	57.530	
2020-421	1	%U	91 B.C.ACCESS	2	26	35-45	47.327	2.174	4.59	40.810	53.850	
2020-421	2	TUV	95 Arc i2000s	3	13	0.7 - 1.2	0.824	0.077	9.35	0.590	1.050	
2020-421	4	%	32 DIMENSION	2	58	35-45	41.516	1.786	4.30	36.160	46.870	
2020-421	4	%	91 B.C.ACCESS	2	10	35-45	46.440	2.027	4.36	40.360	52.520	
2020-421	5	TBI	91 B.C.ACCESS	2	11	35-45	46.109	3.060	6.64	36.930	55.290	
2020-421	5	TBI	98 ROCHE COBA	4	47	0.8-1.3	0.697	0.063	9.00	0.510	0.890	
2020-422	1	%U	84 VITROS ECI	1	27	25-35	57.096	1.594	2.79	52.310	61.880	
2020-422	1	%U	91 B.C.ACCESS	2	26	35-45	50.619	2.991	5.91	41.650	59.590	
2020-422	2	TUV	95 Arc i2000s	3	13	0.7 - 1.2	0.620	0.059	9.57	0.440	0.800	
2020-422	4	%	32 DIMENSION	2	58	35-45	41.488	2.300	5.54	34.590	48.390	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 55 ENDO.(T3 UPTAKE)

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-422	4 %	91	B.C.ACCESS	2	35-45	10	49.960	1.407	2.82	45.740	54.180	
2020-422	5 TBI	91	B.C.ACCESS	2	35-45	11	49.345	2.713	5.50	41.210	57.480	
2020-422	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.841	0.061	7.26	0.660	1.020	
173	T3 UPTAKE	CLIA TEST ID #: 565		EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-423	1 %U	84	VITROS ECI	1	25-35	27	55.963	1.633	2.92	51.060	60.860	
2020-423	1 %U	91	B.C.ACCESS	2	35-45	26	50.173	2.240	4.47	43.450	56.890	
2020-423	2 TUV	95	Arc i2000s	3	0.7 - 1.2	13	0.660	0.057	8.69	0.490	0.830	
2020-423	4 %	32	DIMENSION	2	35-45	58	41.717	2.073	4.97	35.500	47.940	
2020-423	4 %	91	B.C.ACCESS	2	35-45	10	49.840	2.008	4.03	43.820	55.860	
2020-423	5 TBI	91	B.C.ACCESS	2	35-45	11	49.300	2.324	4.71	42.330	56.270	
2020-423	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.821	0.060	7.36	0.640	1.000	
2020-424	1 %U	84	VITROS ECI	1	25-35	27	50.556	1.450	2.87	46.200	54.910	
2020-424	1 %U	91	B.C.ACCESS	2	35-45	26	46.119	1.965	4.26	40.220	52.020	
2020-424	2 TUV	95	Arc i2000s	3	0.7 - 1.2	13	0.945	0.079	8.38	0.710	1.180	
2020-424	4 %	32	DIMENSION	2	35-45	58	41.650	1.852	4.45	36.090	47.210	
2020-424	4 %	91	B.C.ACCESS	2	35-45	10	45.090	1.159	2.57	41.610	48.570	
2020-424	5 TBI	91	B.C.ACCESS	2	35-45	11	44.600	2.858	6.41	36.030	53.170	
2020-424	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.589	0.069	11.70	0.380	0.800	
2020-425	1 %U	84	VITROS ECI	1	25-35	27	58.333	1.764	3.02	53.040	63.630	
2020-425	1 %U	91	B.C.ACCESS	2	35-45	26	52.500	2.591	4.94	44.730	60.270	
2020-425	2 TUV	95	Arc i2000s	3	0.7 - 1.2	13	0.520	0.049	9.45	0.370	0.670	
2020-425	4 %	32	DIMENSION	2	35-45	58	41.843	1.952	4.67	35.990	47.700	
2020-425	4 %	91	B.C.ACCESS	2	35-45	10	51.120	1.594	3.12	46.340	55.900	
2020-425	5 TBI	91	B.C.ACCESS	2	35-45	11	50.736	2.618	5.16	42.880	58.590	
2020-425	5 TBI	98	ROCHE COBA	4	0.8-1.3	47	0.923	0.060	6.49	0.740	1.100	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
043 DIGOXIN		CLIA TEST ID #: 645	EVALUATION CRITERIA: TARGET VALUE	20 % OR 0.20	{GREATER}					
2020-421	50 VITROS			14	1.179	0.067	5.72	0.900	1.400	
2020-422	50 VITROS			14	0.643	0.073	11.33	0.400	0.800	
2020-423	50 VITROS			14	0.729	0.080	10.92	0.500	0.900	
2020-424	50 VITROS			14	1.750	0.124	7.08	1.400	2.100	
2020-425	50 VITROS			14	0.493	0.059	12.04	0.300	0.700	
045 FERRITIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25 %						
2020-421	50 VITROS			12	85.242	4.477	5.25	63.900	106.600	
2020-421	90 ROCHE COBA			14	146.523	8.723	5.95	109.900	183.200	
2020-422	50 VITROS			12	36.517	1.628	4.46	27.400	45.600	
2020-422	90 ROCHE COBA			14	62.650	3.082	4.92	47.000	78.300	
2020-423	50 VITROS			12	44.008	2.636	5.99	33.000	55.000	
2020-423	90 ROCHE COBA			14	76.607	4.186	5.46	57.500	95.800	
2020-424	50 VITROS			12	129.083	8.411	6.52	96.800	161.400	
2020-424	90 ROCHE COBA			14	219.671	16.810	7.65	164.800	274.600	
2020-425	50 VITROS			12	17.133	0.961	5.61	12.900	21.400	
2020-425	90 ROCHE COBA			14	28.064	1.849	6.59	21.000	35.100	
046 GENTAMICIN		CLIA TEST ID #: 665								
2020-421	50 VITROS			13	7.100	0.349	4.91	5.300	8.900	
2020-422	50 VITROS			13	2.454	0.075	3.04	1.800	3.100	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
046 GENTAMICIN		CLIA TEST ID #: 665	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-423	50 VITROS			13	3.169	0.099	3.13	2.400	4.000		
2020-424	50 VITROS			13	11.446	1.083	9.46	8.600	14.300		
2020-425	50 VITROS			13	0.615	0.036	5.86	0.500	0.800		
050 PHENYTOIN		CLIA TEST ID #: 695									
2020-421	50 VITROS			16	14.481	0.502	3.46	10.900	18.100		
2020-421	92 ROCHE COBA			10	14.960	0.697	4.66	11.200	18.700		
2020-422	50 VITROS			16	4.444	0.221	4.97	3.300	5.600		
2020-422	92 ROCHE COBA			10	4.640	0.246	5.30	3.500	5.800		
2020-423	50 VITROS			16	5.963	0.255	4.27	4.500	7.500		
2020-423	92 ROCHE COBA			10	6.140	0.284	4.62	4.600	7.700		
2020-424	50 VITROS			16	22.225	1.135	5.11	16.700	27.800		
2020-424	92 ROCHE COBA			10	22.540	0.564	2.50	16.900	28.200		
2020-425	50 VITROS			16	3.000	0.000	0.00	2.300	3.800		
2020-425	92 ROCHE COBA			10	0.860	0.150	17.40	0.600	1.100		
054 THEOPHYLLINE		CLIA TEST ID #: 735									
2020-421	50 VITROS			13	32.046	1.070	3.34	24.000	40.100		
2020-422	50 VITROS			13	12.431	0.639	5.14	9.300	15.500		
2020-423	50 VITROS			13	15.600	0.826	5.29	11.700	19.500		
2020-424	50 VITROS			13	45.708	1.777	3.89	34.300	57.100		

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
054 THEOPHYLLINE		CLIA TEST ID #: 735	EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-425	50 VITROS			13	3.746	0.388	10.34	2.800	4.700	
127 VANCOMYCIN		CLIA TEST ID #: 0								
2020-421	50 VITROS			16	25.838	0.762	2.95	19.400	32.300	
2020-422	50 VITROS			16	8.419	0.420	4.99	6.300	10.500	
2020-423	50 VITROS			16	11.038	0.430	3.90	8.300	13.800	
2020-424	50 VITROS			16	41.531	1.854	4.46	31.100	51.900	
2020-425	50 VITROS			16	5.000	0.000	0.00	3.800	6.300	
SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
065 URINA-PROTEIN QUANT.			EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}							
2020-451	58 VITROS			38	9.108	0.863	9.48	4.100	14.100	
2020-451	62 DADE DIM.			52	22.808	2.178	9.55	17.100	28.500	
2020-451	79 BECK. COUL			12	5.583	1.665	29.83	0.600	10.600	
2020-451	93 ARCHITECT			15	6.780	0.098	1.45	1.800	11.800	
2020-451	95 INTEGRA			12	2.525	0.846	33.49	0	7.500	
2020-451	113 ROCHE COBA			26	3.860	0.385	9.97	0	8.900	
2020-451	139 S. EXL.200			11	23.236	3.002	12.92	17.400	29.000	
2020-452	58 VITROS			38	143.743	9.333	6.49	107.800	179.700	
2020-452	62 DADE DIM.			52	115.569	2.452	2.12	86.700	144.500	
2020-452	79 BECK. COUL			12	99.833	6.318	6.33	74.900	124.800	
2020-452	93 ARCHITECT			15	97.286	1.908	1.96	73.000	121.600	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-452	95	INTEGRA			12	91.483	7.492	8.19	68.600	114.400
2020-452	113	ROCHE COBA			26	94.227	2.524	2.68	70.700	117.800
2020-452	139	S. EXL.200			11	115.764	1.672	1.44	86.800	144.700
065 URINA-PROTEIN QUANT.			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-453	58	VITROS			38	131.189	6.970	5.31	98.400	164.000
2020-453	62	DADE DIM.			52	112.796	2.037	1.81	84.600	141.000
2020-453	79	BECK. COUL			12	100.708	5.137	5.10	75.500	125.900
2020-453	93	ARCHITECT			15	96.021	2.541	2.65	72.000	120.000
2020-453	95	INTEGRA			12	87.725	6.859	7.82	65.800	109.700
2020-453	113	ROCHE COBA			26	90.184	1.788	1.98	67.600	112.700
2020-453	139	S. EXL.200			11	112.982	2.047	1.81	84.700	141.200
2020-454	58	VITROS			38	7.595	0.491	6.46	2.600	12.600
2020-454	62	DADE DIM.			52	15.210	1.568	10.31	10.200	20.200
2020-454	79	BECK. COUL			12	5.008	0.658	13.13	0.000	10.000
2020-454	93	ARCHITECT			15	7.100	0.913	12.86	2.100	12.100
2020-454	95	INTEGRA			12	3.308	0.777	23.49	0	8.300
2020-454	113	ROCHE COBA			26	3.964	0.521	13.15	0	9.000
2020-454	139	S. EXL.200			11	15.518	2.098	13.52	10.500	20.500
2020-455	58	VITROS			38	128.108	8.395	6.55	96.100	160.100
2020-455	62	DADE DIM.			52	123.670	1.819	1.47	92.800	154.600
2020-455	79	BECK. COUL			12	116.375	5.631	4.84	87.300	145.500
2020-455	93	ARCHITECT			15	115.907	2.701	2.33	86.900	144.900
2020-455	95	INTEGRA			12	102.267	8.708	8.52	76.700	127.800
2020-455	113	ROCHE COBA			26	107.704	2.918	2.71	80.800	134.600
2020-455	139	S. EXL.200			11	122.936	2.726	2.22	92.200	153.700
067 URI-SPECIFIC GRAVITY				EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-451	11	AMES NM10S			49	1.013	0.004	0.40	1.001	1.026
2020-451	13	BAYER CLI			37	1.014	0.003	0.33	1.004	1.024

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-451	67	URISCAN	51	1.013	0.004	0.44	1.000	1.026
2020-451	71	ATLAS	19	1.030	0.001	0.07	1.028	1.032
2020-451	72	URS-9-TECO	106	1.014	0.004	0.43	1.001	1.027
2020-451	80	ROCHE (CHE	36	1.011	0.003	0.30	1.002	1.020
2020-451	89	AIMSTRIP	108	1.014	0.004	0.44	1.000	1.027
2020-451	94	CLINITEK	110	1.013	0.003	0.34	1.003	1.024
2020-451	102	ACON LAB	175	1.017	0.004	0.40	1.005	1.029
2020-451	117	URS 10 A	10	1.015	0.007	0.72	0.993	1.037
2020-451	131	iCHEM VELO	27	1.025	0.001	0.09	1.023	1.028
2020-451	135	Hybrido AU	12	1.025	0.001	0.12	1.021	1.029
2020-451	138	URS-10T	12	1.012	0.003	0.32	1.003	1.022
2020-451	140	MISSION A	63	1.016	0.004	0.43	1.003	1.029
2020-451	147	Mission120	16	1.019	0.002	0.16	1.014	1.024
2020-451	148	Mission500	14	1.019	0.002	0.17	1.014	1.025
067 URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0			EVALUATION CRITERIA: TARGET VALUE 3 SD		
2020-452	11	AMES NM10S	49	1.016	0.004	0.38	1.005	1.028
2020-452	13	BAYER CLI	37	1.016	0.003	0.31	1.006	1.025
2020-452	67	URISCAN	51	1.014	0.002	0.23	1.007	1.020
2020-452	71	ATLAS	19	1.030	0.000	0.04	1.029	1.031
2020-452	72	URS-9-TECO	106	1.016	0.004	0.41	1.004	1.029
2020-452	80	ROCHE (CHE	36	1.013	0.002	0.24	1.006	1.020
2020-452	89	AIMSTRIP	108	1.014	0.004	0.40	1.002	1.026
2020-452	94	CLINITEK	110	1.016	0.003	0.31	1.006	1.025
2020-452	102	ACON LAB	175	1.016	0.004	0.42	1.003	1.029
2020-452	117	URS 10 A	10	1.018	0.006	0.55	1.001	1.034
2020-452	131	iCHEM VELO	27	1.028	0.001	0.08	1.026	1.031
2020-452	135	Hybrido AU	12	1.028	0.001	0.13	1.024	1.032
2020-452	138	URS-10T	12	1.014	0.004	0.36	1.003	1.025
2020-452	140	MISSION A	63	1.016	0.004	0.43	1.003	1.029
2020-452	147	Mission120	16	1.019	0.002	0.16	1.014	1.024
2020-452	148	Mission500	14	1.019	0.002	0.17	1.014	1.025

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
067	URI-SPECIFIC GRAVITY	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-453	11 AMES NM10S			49	1.021	0.003	0.32	1.011	1.031
2020-453	13 BAYER CLI			37	1.020	0.003	0.25	1.012	1.028
2020-453	67 URISCAN			51	1.019	0.003	0.29	1.010	1.028
2020-453	71 ATLAS			19	1.026	0.001	0.13	1.022	1.030
2020-453	72 URS-9-TECO			106	1.020	0.002	0.23	1.013	1.027
2020-453	80 ROCHE (CHE			36	1.012	0.003	0.27	1.004	1.021
2020-453	89 AIMSTRIP			108	1.019	0.003	0.31	1.010	1.029
2020-453	94 CLINITEK			110	1.021	0.003	0.27	1.013	1.030
2020-453	102 ACON LAB			175	1.019	0.003	0.25	1.011	1.027
2020-453	117 URS 10 A			10	1.021	0.001	0.15	1.016	1.025
2020-453	131 iCHEM VELO			27	1.026	0.001	0.09	1.023	1.029
2020-453	135 Hybrido AU			12	1.025	0.001	0.13	1.021	1.029
2020-453	138 URS-10T			12	1.019	0.003	0.27	1.011	1.028
2020-453	140 MISSION A			63	1.019	0.003	0.28	1.010	1.027
2020-453	147 Mission120			16	1.021	0.002	0.21	1.015	1.028
2020-453	148 Mission500			14	1.020	0.000	0.00	1.020	1.020
2020-454	11 AMES NM10S			49	1.022	0.003	0.33	1.012	1.032
2020-454	13 BAYER CLI			37	1.021	0.004	0.36	1.010	1.032
2020-454	67 URISCAN			51	1.023	0.004	0.36	1.012	1.034
2020-454	71 ATLAS			19	1.031	0.001	0.11	1.027	1.034
2020-454	72 URS-9-TECO			106	1.020	0.004	0.44	1.006	1.033
2020-454	80 ROCHE (CHE			36	1.017	0.002	0.23	1.010	1.024
2020-454	89 AIMSTRIP			108	1.021	0.004	0.40	1.008	1.033
2020-454	94 CLINITEK			110	1.022	0.003	0.28	1.014	1.031
2020-454	102 ACON LAB			175	1.023	0.003	0.30	1.014	1.032
2020-454	117 URS 10 A			10	1.048	0.084	8.03	0.796	1.300
2020-454	131 iCHEM VELO			27	1.026	0.001	0.08	1.024	1.029
2020-454	135 Hybrido AU			12	1.026	0.001	0.08	1.024	1.029

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-454	138	URS-10T	12	1.018	0.005	0.46	1.004	1.033	
2020-454	140	MISSION A	63	1.022	0.003	0.26	1.014	1.030	
2020-454	147	Mission120	16	1.020	0.000	0.00	1.020	1.020	
2020-454	148	Mission500	14	1.023	0.003	0.30	1.014	1.032	
067 URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0						EVALUATION CRITERIA: TARGET VALUE 3 SD
2020-455	11	AMES NM10S	49	1.024	0.002	0.22	1.018	1.031	
2020-455	13	BAYER CLI	37	1.022	0.003	0.27	1.014	1.031	
2020-455	67	URISCAN	51	1.025	0.004	0.39	1.013	1.037	
2020-455	71	ATLAS	19	1.030	0.001	0.06	1.028	1.031	
2020-455	72	URS-9-TECO	106	1.023	0.003	0.30	1.013	1.032	
2020-455	80	ROCHE (CHE	36	1.016	0.002	0.19	1.010	1.022	
2020-455	89	AIMSTRIP	108	1.023	0.003	0.30	1.014	1.032	
2020-455	94	CLINITEK	110	1.024	0.002	0.22	1.017	1.031	
2020-455	102	ACON LAB	175	1.024	0.002	0.24	1.016	1.031	
2020-455	117	URS 10 A	10	1.024	0.004	0.35	1.013	1.035	
2020-455	131	iCHEM VELO	27	1.025	0.001	0.08	1.022	1.027	
2020-455	135	Hybrido AU	12	1.024	0.001	0.13	1.020	1.028	
2020-455	138	URS-10T	12	1.022	0.003	0.31	1.012	1.031	
2020-455	140	MISSION A	63	1.024	0.002	0.22	1.017	1.030	
2020-455	147	Mission120	16	1.023	0.003	0.29	1.014	1.032	
2020-455	148	Mission500	14	1.024	0.003	0.25	1.017	1.032	
166 MICROALBUMIN			EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-451	58	VITROS	32	6.000	0.000	0.00	1.000	11.000	
2020-451	62	DADE DIM.	71	1.275	2.208	173.21	0	6.300	
2020-451	62	DADE DIM.	71	2.343	0.989	42.21	0	7.300	
2020-451	79	BECK. COUL	17	1.275	2.208	173.21	0	6.300	
2020-451	79	BECK. COUL	17	2.194	2.705	123.27	0	7.200	
2020-451	85	Rx DAYTONA	19	1.632	2.102	128.84	0	6.600	
2020-451	93	ARCHITECT	18	4.988	0.032	0.65	0.000	10.000	
2020-451	95	INTEGRA	24	2.004	1.417	70.67	0	7.000	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-451	113	ROCHE COBA		24	10.625	3.672	34.56	5.600	15.600
2020-451	139	S. EXL.200		14	1.275	2.208	173.21	0	6.300
2020-451	139	S. EXL.200		14	2.271	1.077	47.43	0	7.300
166	MICROALBUMIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}				
2020-452	58	VITROS		32	850.847	34.689	4.08	638.100	1063.600
2020-452	62	DADE DIM.		71	840.887	39.241	4.67	630.700	1051.100
2020-452	62	DADE DIM.		71	819.775	70.638	8.62	614.800	1024.700
2020-452	79	BECK. COUL		17	794.488	70.786	8.91	595.900	993.100
2020-452	79	BECK. COUL		17	819.775	70.638	8.62	614.800	1024.700
2020-452	85	Rx DAYTONA		19	777.842	82.780	10.64	583.400	972.300
2020-452	93	ARCHITECT		18	823.176	39.332	4.78	617.400	1029.000
2020-452	95	INTEGRA		24	845.825	59.194	7.00	634.400	1057.300
2020-452	113	ROCHE COBA		24	787.663	32.549	4.13	590.700	984.600
2020-452	139	S. EXL.200		14	850.769	45.651	5.37	638.100	1063.500
2020-452	139	S. EXL.200		14	819.775	70.638	8.62	614.800	1024.700
2020-453	58	VITROS		32	854.578	36.323	4.25	640.900	1068.200
2020-453	62	DADE DIM.		71	819.425	73.557	8.98	614.600	1024.300
2020-453	62	DADE DIM.		71	848.721	44.504	5.24	636.500	1060.900
2020-453	79	BECK. COUL		17	819.425	73.557	8.98	614.600	1024.300
2020-453	79	BECK. COUL		17	813.331	67.616	8.31	610.000	1016.700
2020-453	85	Rx DAYTONA		19	798.095	89.433	11.21	598.600	997.600
2020-453	93	ARCHITECT		18	864.800	52.909	6.12	648.600	1081.000
2020-453	95	INTEGRA		24	848.404	50.140	5.91	636.300	1060.500
2020-453	113	ROCHE COBA		24	802.083	32.723	4.08	601.600	1002.600
2020-453	139	S. EXL.200		14	819.425	73.557	8.98	614.600	1024.300
2020-453	139	S. EXL.200		14	861.285	45.789	5.32	646.000	1076.600
2020-454	58	VITROS		32	6.032	0.177	2.93	1.000	11.000
2020-454	62	DADE DIM.		71	3.979	1.333	33.50	0	9.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-454	62	DADE DIM.		71	1.275	2.208	173.21	0	6.300
2020-454	79	BECK. COUL		17	2.282	2.660	116.57	0	7.300
2020-454	79	BECK. COUL		17	1.275	2.208	173.21	0	6.300
2020-454	85	Rx DAYTONA		19	1.795	2.113	117.75	0	6.800
2020-454	93	ARCHITECT		18	4.988	0.032	0.65	0.000	10.000
2020-454	95	INTEGRA		24	1.800	1.192	66.22	0	6.800
2020-454	113	ROCHE COBA		24	10.625	3.672	34.56	5.600	15.600
2020-454	139	S. EXL.200		14	4.100	1.711	41.73	0	9.100
2020-454	139	S. EXL.200		14	1.275	2.208	173.21	0	6.300
166	MICROALBUMIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-455	58	VITROS		32	1025.494	51.065	4.98	769.100	1281.900
2020-455	62	DADE DIM.		71	910.300	51.577	5.67	682.700	1137.900
2020-455	62	DADE DIM.		71	1013.894	72.725	7.17	760.400	1267.400
2020-455	79	BECK. COUL		17	910.300	51.577	5.67	682.700	1137.900
2020-455	79	BECK. COUL		17	945.244	115.742	12.24	708.900	1181.600
2020-455	85	Rx DAYTONA		19	935.437	64.399	6.88	701.600	1169.300
2020-455	93	ARCHITECT		18	1027.476	47.525	4.63	770.600	1284.300
2020-455	95	INTEGRA		24	1010.065	77.575	7.68	757.500	1262.600
2020-455	113	ROCHE COBA		24	954.288	37.081	3.89	715.700	1192.900
2020-455	139	S. EXL.200		14	910.300	51.577	5.67	682.700	1137.900
2020-455	139	S. EXL.200		14	1029.446	70.483	6.85	772.100	1286.800
167	CREATININE								
2020-451	58	VITROS		46	240.261	10.598	4.41	180.200	300.300
2020-451	62	DADE DIM.		62	235.719	9.578	4.06	176.800	294.600
2020-451	79	BECK. COUL		15	220.480	14.053	6.37	165.400	275.600
2020-451	85	Rx DAYTONA		10	222.830	22.170	9.95	167.100	278.500
2020-451	93	ARCHITECT		17	231.600	7.828	3.38	173.700	289.500
2020-451	95	INTEGRA		22	245.862	11.955	4.86	184.400	307.300
2020-451	113	ROCHE COBA		27	246.219	9.834	3.99	184.700	307.800
2020-451	139	S. EXL.200		18	242.341	6.164	2.54	181.800	302.900

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-451	139 S. EXL.200		18	246.783	9.884	4.01	185.100	308.500
167 CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-452	58 VITROS		46	231.343	8.077	3.49	173.500	289.200
2020-452	62 DADE DIM.		62	229.054	9.154	4.00	171.800	286.300
2020-452	79 BECK. COUL		15	206.927	15.831	7.65	155.200	258.700
2020-452	85 Rx DAYTONA		10	218.070	22.465	10.30	163.600	272.600
2020-452	93 ARCHITECT		17	226.871	9.840	4.34	170.200	283.600
2020-452	95 INTEGRA		22	238.710	11.118	4.66	179.000	298.400
2020-452	113 ROCHE COBA		27	240.619	7.728	3.21	180.500	300.800
2020-452	139 S. EXL.200		18	238.183	8.988	3.77	178.600	297.700
2020-452	139 S. EXL.200		18	236.065	6.337	2.68	177.000	295.100
2020-453	58 VITROS		46	223.252	7.618	3.41	167.400	279.100
2020-453	62 DADE DIM.		62	229.987	9.817	4.27	172.500	287.500
2020-453	79 BECK. COUL		15	215.187	13.824	6.42	161.400	269.000
2020-453	85 Rx DAYTONA		10	216.780	24.813	11.45	162.600	271.000
2020-453	93 ARCHITECT		17	224.994	8.237	3.66	168.700	281.200
2020-453	95 INTEGRA		22	235.976	11.504	4.88	177.000	295.000
2020-453	113 ROCHE COBA		27	240.593	7.799	3.24	180.400	300.700
2020-453	139 S. EXL.200		18	235.241	4.491	1.91	176.400	294.100
2020-453	139 S. EXL.200		18	239.383	7.891	3.30	179.500	299.200
2020-454	58 VITROS		46	3.229	0.207	6.42	0	8.200
2020-454	62 DADE DIM.		62	7.884	5.829	73.94	2.900	12.900
2020-454	79 BECK. COUL		15	0.364	0.475	130.26	0	5.400
2020-454	85 Rx DAYTONA		10	0.970	1.040	107.24	0	6.000
2020-454	93 ARCHITECT		17	4.925	0.241	4.90	0	9.900
2020-454	95 INTEGRA		22	0.538	1.116	207.38	0	5.500
2020-454	113 ROCHE COBA		27	3.835	0.809	21.10	0	8.800
2020-454	139 S. EXL.200		18	8.317	3.545	42.63	3.300	13.300

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-454	139 S. EXL.200			18	8.883	5.691	64.06	3.900	13.900	
167 CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}							
2020-455	58 VITROS			46	3.227	0.209	6.48	0	8.200	
2020-455	62 DADE DIM.			62	7.892	5.826	73.82	2.900	12.900	
2020-455	79 BECK. COUL			15	0.443	0.420	94.93	0	5.400	
2020-455	85 Rx DAYTONA			10	1.030	1.421	137.92	0	6.000	
2020-455	93 ARCHITECT			17	4.925	0.241	4.90	0	9.900	
2020-455	95 INTEGRA			22	0.710	1.289	181.69	0	5.700	
2020-455	113 ROCHE COBA			27	3.804	0.923	24.28	0	8.800	
2020-455	139 S. EXL.200			18	8.883	5.691	64.06	3.900	13.900	
2020-455	139 S. EXL.200			18	8.317	3.545	42.63	3.300	13.300	

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
057 URINALYSIS-BILIRUBIN											
2020-451	011 AMES NM10S			057 057	057 057		057 - NEGATIVE	49	49	49	100
2020-451	013 BAYER CLI			057 057	057 057		057 - NEGATIVE	38	38	38	100
2020-451	067 URISCAN			057 057	057 057		057 - NEGATIVE	50	50	50	100
2020-451	071 ATLAS			057 057	057 057		057 - NEGATIVE	19	19	19	100
2020-451	072 URS-9-TECO			057 057	057 057		057 - NEGATIVE	105	105	105	100
2020-451	080 ROCHE (CHE			057 057	057 057		057 - NEGATIVE	35	35	35	100
2020-451	089 AIMSTRIP			057 057	057 057		057 - NEGATIVE	107	107	107	100
2020-451	094 CLINITEK			057 057	057 057		057 - NEGATIVE	109	109	109	100
2020-451	102 ACON LAB			057 057	057 057		057 - NEGATIVE	170	170	170	100
2020-451	117 URS 10 A			057 057	057 057		057 - NEGATIVE	10	10	10	100
2020-451	131 iCHEM VELO			057 057	057 057		057 - NEGATIVE	25	25	25	100
2020-451	135 Hybrid AU			057 057	057 057		057 - NEGATIVE	12	12	12	100
2020-451	138 URS-10T			057 057	057 057		057 - NEGATIVE	12	12	12	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-451	140	MISSION A		057 057	057 057		057 - NEGATIVE	65	65	65	100	
2020-451	147	Mission120		057 057	057 057		057 - NEGATIVE	16	16	16	100	
2020-451	148	Mission500		057 057	057 057		057 - NEGATIVE	19	19	19	100	
057 URINALYSIS-BILIRUBIN												
2020-452	011	AMES NM10S		058 060	057 059		059 - 2+	49	36	49	100	
2020-452	011	AMES NM10S		058 060	057 059		058 - TRACES,1+	49	8	49	100	
2020-452	011	AMES NM10S		058 060	057 059		060 - 3+	49	5	49	100	
2020-452	013	BAYER CLI		058 060	057 059		059 - 2+	38	30	38	100	
2020-452	013	BAYER CLI		058 060	057 059		058 - TRACES,1+	38	7	38	100	
2020-452	013	BAYER CLI		058 060	057 059		060 - 3+	38	1	38	100	
2020-452	067	URISCAN		057 059	057 059		059 - 2+	50	11	50	100	
2020-452	067	URISCAN		057 059	057 059		058 - TRACES,1+	50	34	50	100	
2020-452	067	URISCAN		057 059	057 059		057 - NEGATIVE	50	5	50	100	
2020-452	071	ATLAS		057 059	057 059		059 - 2+	19	1	19	100	
2020-452	071	ATLAS		057 059	057 059		058 - TRACES,1+	19	18	19	100	
2020-452	072	URS-9-TECO		058 060	057 059		059 - 2+	105	58	105	100	
2020-452	072	URS-9-TECO		058 060	057 059		057 - NEGATIVE	105	5	105	100	
2020-452	072	URS-9-TECO		058 060	057 059		060 - 3+	105	16	105	100	
2020-452	072	URS-9-TECO		058 060	057 059		058 - TRACES,1+	105	26	105	100	
2020-452	080	ROCHE (CHE		057 059	057 059		057 - NEGATIVE	35	1	35	100	
2020-452	080	ROCHE (CHE		057 059	057 059		059 - 2+	35	3	35	100	
2020-452	080	ROCHE (CHE		057 059	057 059		058 - TRACES,1+	35	31	35	100	
2020-452	089	AIMSTRIP		058 060	057 059		057 - NEGATIVE	107	5	107	100	
2020-452	089	AIMSTRIP		058 060	057 059		058 - TRACES,1+	107	40	107	100	
2020-452	089	AIMSTRIP		058 060	057 059		059 - 2+	107	52	107	100	
2020-452	089	AIMSTRIP		058 060	057 059		060 - 3+	107	10	107	100	
2020-452	094	CLINITEK		058 060	057 059		058 - TRACES,1+	109	21	109	100	
2020-452	094	CLINITEK		058 060	057 059		059 - 2+	109	87	109	100	
2020-452	094	CLINITEK		058 060	057 059		060 - 3+	109	1	109	100	
2020-452	102	ACON LAB		057 059	057 059		060 - 3+	170	3	167	0	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-452	102	ACON LAB		057	059	057	059	058	- TRACES,1+	170	111	167	100
2020-452	102	ACON LAB		057	059	057	059	059	- 2+	170	48	167	100
2020-452	102	ACON LAB		057	059	057	059	057	- NEGATIVE	170	8	167	100
2020-452	117	URS 10 A		058	060	057	059	060	- 3+	10	1	10	100
2020-452	117	URS 10 A		058	060	057	059	058	- TRACES,1+	10	1	10	100
2020-452	117	URS 10 A		058	060	057	059	059	- 2+	10	8	10	100
2020-452	131	iCHEM VELO		057	057	057	059	057	- NEGATIVE	25	25	25	100
2020-452	135	Hybrido AU		058	060	057	059	060	- 3+	12	1	12	100
2020-452	135	Hybrido AU		058	060	057	059	059	- 2+	12	11	12	100
2020-452	138	URS-10T		058	060	057	059	058	- TRACES,1+	12	4	12	100
2020-452	138	URS-10T		058	060	057	059	059	- 2+	12	6	12	100
2020-452	138	URS-10T		058	060	057	059	060	- 3+	12	1	12	100
2020-452	138	URS-10T		058	060	057	059	057	- NEGATIVE	12	1	12	100
2020-452	140	MISSION A		057	059	057	059	059	- 2+	65	15	65	100
2020-452	140	MISSION A		057	059	057	059	058	- TRACES,1+	65	42	65	100
2020-452	140	MISSION A		057	059	057	059	057	- NEGATIVE	65	8	65	100
2020-452	147	Mission120		057	059	057	059	057	- NEGATIVE	16	1	16	100
2020-452	147	Mission120		057	059	057	059	059	- 2+	16	1	16	100
2020-452	147	Mission120		057	059	057	059	058	- TRACES,1+	16	14	16	100
2020-452	148	Mission500		057	059	057	059	058	- TRACES,1+	19	13	19	100
2020-452	148	Mission500		057	059	057	059	059	- 2+	19	5	19	100
2020-452	148	Mission500		057	059	057	059	057	- NEGATIVE	19	1	19	100
057 URINALYSIS-BILIRUBIN													
2020-453	011	AMES NM10S		057	057	057	057	057	- NEGATIVE	49	49	49	100
2020-453	013	BAYER CLI		057	057	057	057	057	- NEGATIVE	38	38	38	100
2020-453	067	URISCAN		057	057	057	057	057	- NEGATIVE	50	50	50	100
2020-453	071	ATLAS		057	057	057	057	057	- NEGATIVE	19	19	19	100
2020-453	072	URS-9-TECO		057	057	057	057	057	- NEGATIVE	105	105	105	100
2020-453	080	ROCHE (CHE		057	057	057	057	057	- NEGATIVE	35	35	35	100
2020-453	089	AIMSTRIP		057	057	057	057	057	- NEGATIVE	107	107	107	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-453	094	CLINITEK		057	057	057	057	057 - NEGATIVE	109	109	109	100
2020-453	102	ACON LAB		057	057	057	057	057 - NEGATIVE	170	170	170	100
2020-453	117	URS 10 A		057	057	057	057	057 - NEGATIVE	10	10	10	100
2020-453	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	25	25	25	100
2020-453	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	12	12	12	100
2020-453	138	URS-10T		057	057	057	057	057 - NEGATIVE	12	12	12	100
2020-453	140	MISSION A		057	057	057	057	057 - NEGATIVE	65	65	65	100
2020-453	147	Mission120		057	057	057	057	057 - NEGATIVE	16	16	16	100
2020-453	148	Mission500		057	057	057	057	057 - NEGATIVE	19	19	19	100
057 URINALYSIS-BILIRUBIN												
2020-454	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	49	49	49	100
2020-454	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	38	38	38	100
2020-454	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	50	50	100
2020-454	071	ATLAS		057	057	057	057	057 - NEGATIVE	19	19	19	100
2020-454	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	105	105	105	100
2020-454	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	35	35	35	100
2020-454	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	107	106	106	100
2020-454	089	AIMSTRIP		057	057	057	057	058 - TRACES, 1+	107	1	106	0
2020-454	094	CLINITEK		057	057	057	057	057 - NEGATIVE	109	108	108	100
2020-454	094	CLINITEK		057	057	057	057	059 - 2+	109	1	108	0
2020-454	102	ACON LAB		057	057	057	057	057 - NEGATIVE	170	170	170	100
2020-454	117	URS 10 A		057	057	057	057	057 - NEGATIVE	10	10	10	100
2020-454	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	25	25	25	100
2020-454	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	12	12	12	100
2020-454	138	URS-10T		057	057	057	057	057 - NEGATIVE	12	12	12	100
2020-454	140	MISSION A		057	057	057	057	057 - NEGATIVE	65	64	64	100
2020-454	140	MISSION A		057	057	057	057	060 - 3+	65	1	64	0
2020-454	147	Mission120		057	057	057	057	057 - NEGATIVE	16	16	16	100
2020-454	148	Mission500		057	057	057	057	057 - NEGATIVE	19	19	19	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
057 URINALYSIS-BILIRUBIN													
2020-455	011	AMES NM10S		057	057	057	057	059 - 2+	49	1	48	0	
2020-455	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	49	48	48	100	
2020-455	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	38	38	38	100	
2020-455	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	50	50	100	
2020-455	071	ATLAS		057	057	057	057	057 - NEGATIVE	19	19	19	100	
2020-455	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	105	105	105	100	
2020-455	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	35	35	35	100	
2020-455	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	107	106	106	100	
2020-455	089	AIMSTRIP		057	057	057	057	060 - 3+	107	1	106	0	
2020-455	094	CLINITEK		057	057	057	057	057 - NEGATIVE	109	109	109	100	
2020-455	102	ACON LAB		057	057	057	057	057 - NEGATIVE	170	170	170	100	
2020-455	117	URS 10 A		057	057	057	057	057 - NEGATIVE	10	10	10	100	
2020-455	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	25	25	25	100	
2020-455	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	12	12	12	100	
2020-455	138	URS-10T		057	057	057	057	057 - NEGATIVE	12	12	12	100	
2020-455	140	MISSION A		057	057	057	057	057 - NEGATIVE	65	65	65	100	
2020-455	147	Mission120		057	057	057	057	057 - NEGATIVE	16	16	16	100	
2020-455	148	Mission500		057	057	057	057	057 - NEGATIVE	19	19	19	100	
058 URINALYSIS-GLUCOSE													
2020-451	011	AMES NM10S		076	078	075	077	077 - 1000mg/dl	49	34	49	100	
2020-451	011	AMES NM10S		076	078	075	077	078 - 2000mg/dl	49	1	49	100	
2020-451	011	AMES NM10S		076	078	075	077	075 - 250mg/dl-+	49	1	49	100	
2020-451	011	AMES NM10S		076	078	075	077	076 - 500mg/dl2+	49	13	49	100	
2020-451	013	BAYER CLI		076	078	075	077	076 - 500mg/dl2+	37	14	37	100	
2020-451	013	BAYER CLI		076	078	075	077	077 - 1000mg/dl	37	20	37	100	
2020-451	013	BAYER CLI		076	078	075	077	075 - 250mg/dl-+	37	3	37	100	
2020-451	067	URISCAN		077	078	075	077	075 - 250mg/dl-+	50	1	50	100	
2020-451	067	URISCAN		077	078	075	077	078 - 2000mg/dl	50	30	50	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-451	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	50	4	50	100
2020-451	067	URISCAN		077 078	075 077		077 - 1000mg/dl	50	15	50	100
2020-451	071	ATLAS		075 077	075 077		076 - 500mg/dl2+	19	17	19	100
2020-451	071	ATLAS		075 077	075 077		077 - 1000mg/dl	19	2	19	100
2020-451	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	107	45	107	100
2020-451	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	107	50	107	100
2020-451	072	URS-9-TECO		076 078	075 077		075 - 250mg/dl-+	107	7	107	100
2020-451	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	107	5	107	100
2020-451	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	36	32	36	100
2020-451	080	ROCHE (CHE		076 078	075 077		075 - 250mg/dl-+	36	2	36	100
2020-451	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	36	1	36	100
2020-451	080	ROCHE (CHE		076 078	075 077		076 - 500mg/dl2+	36	1	36	100
2020-451	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	107	11	102	100
2020-451	089	AIMSTRIP		075 077	075 077		078 - 2000mg/dl	107	3	102	0
2020-451	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	107	62	102	100
2020-451	089	AIMSTRIP		075 077	075 077		073 - NEGATIVE	107	1	102	0
2020-451	089	AIMSTRIP		075 077	075 077		074 - TRACES	107	1	102	0
2020-451	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	107	29	102	100
2020-451	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	109	3	108	100
2020-451	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	109	3	108	100
2020-451	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	109	34	108	100
2020-451	094	CLINITEK		076 078	075 077		073 - NEGATIVE	109	1	108	0
2020-451	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	109	68	108	100
2020-451	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	174	105	170	100
2020-451	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	174	5	170	100
2020-451	102	ACON LAB		075 077	075 077		078 - 2000mg/dl	174	2	170	0
2020-451	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	174	60	170	100
2020-451	102	ACON LAB		075 077	075 077		073 - NEGATIVE	174	2	170	0
2020-451	117	URS 10 A		076 078	075 077		076 - 500mg/dl2+	10	4	10	100
2020-451	117	URS 10 A		076 078	075 077		075 - 250mg/dl-+	10	1	10	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-451	117	URS 10 A		076 078	075 077		077 - 1000mg/dl	10	5	10	100
2020-451	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	26	3	26	100
2020-451	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	26	23	26	100
2020-451	135	Hybrido AU		077 078	075 077		078 - 2000mg/dl	12	7	12	100
2020-451	135	Hybrido AU		077 078	075 077		077 - 1000mg/dl	12	5	12	100
2020-451	138	URS-10T		076 078	075 077		078 - 2000mg/dl	12	1	12	100
2020-451	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	12	4	12	100
2020-451	138	URS-10T		076 078	075 077		077 - 1000mg/dl	12	7	12	100
2020-451	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	63	23	63	100
2020-451	140	MISSION A		076 078	075 077		077 - 1000mg/dl	63	35	63	100
2020-451	140	MISSION A		076 078	075 077		078 - 2000mg/dl	63	4	63	100
2020-451	140	MISSION A		076 078	075 077		075 - 250mg/dl-+	63	1	63	100
2020-451	147	Mission120		076 078	075 077		078 - 2000mg/dl	16	3	16	100
2020-451	147	Mission120		076 078	075 077		077 - 1000mg/dl	16	8	16	100
2020-451	147	Mission120		076 078	075 077		076 - 500mg/dl2+	16	5	16	100
2020-451	148	Mission500		076 078	075 077		076 - 500mg/dl2+	18	3	18	100
2020-451	148	Mission500		076 078	075 077		077 - 1000mg/dl	18	14	18	100
2020-451	148	Mission500		076 078	075 077		078 - 2000mg/dl	18	1	18	100
058 URINALYSIS-GLUCOSE											
2020-452	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	49	49	49	100
2020-452	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	37	37	37	100
2020-452	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100
2020-452	071	ATLAS		073 073	073 073		073 - NEGATIVE	19	19	19	100
2020-452	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	107	106	106	100
2020-452	072	URS-9-TECO		073 073	073 073		078 - 2000mg/dl	107	1	106	0
2020-452	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	36	36	36	100
2020-452	089	AIMSTRIP		073 073	073 073		076 - 500mg/dl2+	107	1	106	0
2020-452	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	107	106	106	100
2020-452	094	CLINITEK		073 073	073 073		076 - 500mg/dl2+	109	1	108	0
2020-452	094	CLINITEK		073 073	073 073		073 - NEGATIVE	109	108	108	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	102	ACON LAB		073 073	073 073		073 - NEGATIVE	174	174	174	100	
2020-452	117	URS 10 A		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2020-452	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	26	26	26	100	
2020-452	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2020-452	138	URS-10T		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2020-452	140	MISSION A		073 073	073 073		073 - NEGATIVE	63	63	63	100	
2020-452	147	Mission120		073 073	073 073		073 - NEGATIVE	16	16	16	100	
2020-452	148	Mission500		073 073	073 073		073 - NEGATIVE	18	18	18	100	
058 URINALYSIS-GLUCOSE												
2020-453	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	49	49	49	100	
2020-453	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	37	37	37	100	
2020-453	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100	
2020-453	071	ATLAS		073 073	073 073		073 - NEGATIVE	19	19	19	100	
2020-453	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	107	107	107	100	
2020-453	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	36	36	36	100	
2020-453	089	AIMSTRIP		073 073	073 073		076 - 500mg/dl2+	107	1	105	0	
2020-453	089	AIMSTRIP		073 073	073 073		074 - TRACES	107	1	105	0	
2020-453	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	107	105	105	100	
2020-453	094	CLINITEK		073 073	073 073		073 - NEGATIVE	109	109	109	100	
2020-453	102	ACON LAB		073 073	073 073		073 - NEGATIVE	174	174	174	100	
2020-453	117	URS 10 A		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2020-453	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	26	26	26	100	
2020-453	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2020-453	138	URS-10T		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2020-453	140	MISSION A		073 073	073 073		073 - NEGATIVE	63	63	63	100	
2020-453	147	Mission120		073 073	073 073		073 - NEGATIVE	16	16	16	100	
2020-453	148	Mission500		073 073	073 073		073 - NEGATIVE	18	18	18	100	
2020-454	011	AMES NM10S		074 076	074 076		075 - 250mg/dl-+	49	25	48	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-454	011	AMES NM10S		074 076	074 076		074 - TRACES	49	23	48	100
2020-454	011	AMES NM10S		074 076	074 076		073 - NEGATIVE	49	1	48	0
2020-454	013	BAYER CLI		074 076	074 076		075 - 250mg/dl-+	37	22	36	100
2020-454	013	BAYER CLI		074 076	074 076		074 - TRACES	37	14	36	100
2020-454	013	BAYER CLI		074 076	074 076		073 - NEGATIVE	37	1	36	0
2020-454	067	URISCAN		074 076	074 076		076 - 500mg/dl2+	50	19	43	100
2020-454	067	URISCAN		074 076	074 076		074 - TRACES	50	2	43	100
2020-454	067	URISCAN		074 076	074 076		073 - NEGATIVE	50	1	43	0
2020-454	067	URISCAN		074 076	074 076		077 - 1000mg/dl	50	5	43	0
2020-454	067	URISCAN		074 076	074 076		078 - 2000mg/dl	50	1	43	0
2020-454	067	URISCAN		074 076	074 076		075 - 250mg/dl-+	50	22	43	100
2020-454	071	ATLAS		076 078	074 076		076 - 500mg/dl2+	19	3	19	100
2020-454	071	ATLAS		076 078	074 076		077 - 1000mg/dl	19	16	19	100
2020-454	072	URS-9-TECO		074 076	074 076		075 - 250mg/dl-+	107	63	101	100
2020-454	072	URS-9-TECO		074 076	074 076		076 - 500mg/dl2+	107	10	101	100
2020-454	072	URS-9-TECO		074 076	074 076		073 - NEGATIVE	107	6	101	0
2020-454	072	URS-9-TECO		074 076	074 076		074 - TRACES	107	28	101	100
2020-454	080	ROCHE (CHE		076 078	074 076		077 - 1000mg/dl	36	34	35	100
2020-454	080	ROCHE (CHE		076 078	074 076		073 - NEGATIVE	36	1	35	0
2020-454	080	ROCHE (CHE		076 078	074 076		078 - 2000mg/dl	36	1	35	100
2020-454	089	AIMSTRIP		074 076	074 076		074 - TRACES	107	26	103	100
2020-454	089	AIMSTRIP		074 076	074 076		076 - 500mg/dl2+	107	1	103	100
2020-454	089	AIMSTRIP		074 076	074 076		073 - NEGATIVE	107	4	103	0
2020-454	089	AIMSTRIP		074 076	074 076		075 - 250mg/dl-+	107	76	103	100
2020-454	094	CLINITEK		074 076	074 076		075 - 250mg/dl-+	109	57	106	100
2020-454	094	CLINITEK		074 076	074 076		076 - 500mg/dl2+	109	1	106	100
2020-454	094	CLINITEK		074 076	074 076		074 - TRACES	109	48	106	100
2020-454	094	CLINITEK		074 076	074 076		073 - NEGATIVE	109	3	106	0
2020-454	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	174	142	172	100
2020-454	102	ACON LAB		074 076	074 076		073 - NEGATIVE	174	2	172	0

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-454	102	ACON LAB		074	076	074	076	074	- TRACES	174	30	172	100
2020-454	117	URS 10 A		074	076	074	076	075	- 250mg/dl-+	10	7	8	100
2020-454	117	URS 10 A		074	076	074	076	076	- 500mg/dl2+	10	1	8	100
2020-454	117	URS 10 A		074	076	074	076	073	- NEGATIVE	10	2	8	0
2020-454	131	iCHEM VELO		075	077	074	076	077	- 1000mg/dl	26	3	26	100
2020-454	131	iCHEM VELO		075	077	074	076	076	- 500mg/dl2+	26	23	26	100
2020-454	135	Hybrido AU		077	078	074	076	078	- 2000mg/dl	12	8	12	100
2020-454	135	Hybrido AU		077	078	074	076	077	- 1000mg/dl	12	4	12	100
2020-454	138	URS-10T		074	076	074	076	075	- 250mg/dl-+	12	8	11	100
2020-454	138	URS-10T		074	076	074	076	076	- 500mg/dl2+	12	1	11	100
2020-454	138	URS-10T		074	076	074	076	074	- TRACES	12	2	11	100
2020-454	138	URS-10T		074	076	074	076	073	- NEGATIVE	12	1	11	0
2020-454	140	MISSION A		074	076	074	076	075	- 250mg/dl-+	63	54	62	100
2020-454	140	MISSION A		074	076	074	076	073	- NEGATIVE	63	1	62	0
2020-454	140	MISSION A		074	076	074	076	074	- TRACES	63	8	62	100
2020-454	147	Mission120		074	076	074	076	075	- 250mg/dl-+	16	13	16	100
2020-454	147	Mission120		074	076	074	076	074	- TRACES	16	3	16	100
2020-454	148	Mission500		074	076	074	076	073	- NEGATIVE	18	1	17	0
2020-454	148	Mission500		074	076	074	076	075	- 250mg/dl-+	18	16	17	100
2020-454	148	Mission500		074	076	074	076	074	- TRACES	18	1	17	100
058 URINALYSIS-GLUCOSE													
2020-455	011	AMES NM10S		076	078	075	077	077	- 1000mg/dl	49	27	49	100
2020-455	011	AMES NM10S		076	078	075	077	078	- 2000mg/dl	49	3	49	100
2020-455	011	AMES NM10S		076	078	075	077	076	- 500mg/dl2+	49	19	49	100
2020-455	013	BAYER CLI		076	078	075	077	077	- 1000mg/dl	37	22	37	100
2020-455	013	BAYER CLI		076	078	075	077	075	- 250mg/dl-+	37	2	37	100
2020-455	013	BAYER CLI		076	078	075	077	076	- 500mg/dl2+	37	13	37	100
2020-455	067	URISCAN		077	078	075	077	078	- 2000mg/dl	50	32	50	100
2020-455	067	URISCAN		077	078	075	077	075	- 250mg/dl-+	50	1	50	100
2020-455	067	URISCAN		077	078	075	077	077	- 1000mg/dl	50	14	50	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	50	3	50	100
2020-455	071	ATLAS		076 078	075 077		077 - 1000mg/dl	19	19	19	100
2020-455	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	107	54	107	100
2020-455	072	URS-9-TECO		076 078	075 077		075 - 250mg/dl-+	107	7	107	100
2020-455	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	107	11	107	100
2020-455	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	107	35	107	100
2020-455	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	36	1	36	100
2020-455	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	36	35	36	100
2020-455	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	107	26	104	100
2020-455	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	107	7	104	100
2020-455	089	AIMSTRIP		075 077	075 077		078 - 2000mg/dl	107	3	104	0
2020-455	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	107	71	104	100
2020-455	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	109	2	109	100
2020-455	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	109	2	109	100
2020-455	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	109	72	109	100
2020-455	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	109	33	109	100
2020-455	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	174	117	173	100
2020-455	102	ACON LAB		075 077	075 077		078 - 2000mg/dl	174	1	173	0
2020-455	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	174	1	173	100
2020-455	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	174	55	173	100
2020-455	117	URS 10 A		076 078	075 077		076 - 500mg/dl2+	10	4	10	100
2020-455	117	URS 10 A		076 078	075 077		075 - 250mg/dl-+	10	1	10	100
2020-455	117	URS 10 A		076 078	075 077		077 - 1000mg/dl	10	5	10	100
2020-455	131	icHEM VELO		075 077	075 077		077 - 1000mg/dl	26	3	26	100
2020-455	131	icHEM VELO		075 077	075 077		076 - 500mg/dl2+	26	23	26	100
2020-455	135	Hybrido AU		077 078	075 077		078 - 2000mg/dl	12	8	12	100
2020-455	135	Hybrido AU		077 078	075 077		077 - 1000mg/dl	12	4	12	100
2020-455	138	URS-10T		076 078	075 077		078 - 2000mg/dl	12	1	12	100
2020-455	138	URS-10T		076 078	075 077		077 - 1000mg/dl	12	8	12	100
2020-455	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	12	3	12	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-455	140	MISSION A		076 078	075 077		077 - 1000mg/dl	63	35	63	100	
2020-455	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	63	28	63	100	
2020-455	147	Mission120		076 078	075 077		078 - 2000mg/dl	16	2	16	100	
2020-455	147	Mission120		076 078	075 077		077 - 1000mg/dl	16	9	16	100	
2020-455	147	Mission120		076 078	075 077		076 - 500mg/dl2+	16	5	16	100	
2020-455	148	Mission500		076 078	075 077		076 - 500mg/dl2+	18	2	18	100	
2020-455	148	Mission500		076 078	075 077		078 - 2000mg/dl	18	1	18	100	
2020-455	148	Mission500		076 078	075 077		077 - 1000mg/dl	18	15	18	100	
059 URINALYS.HG.OR BLOOD												
2020-451	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	49	49	49	100	
2020-451	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	37	37	37	100	
2020-451	067	URISCAN		066 066	066 066		066 - NEGATIVE	50	50	50	100	
2020-451	071	ATLAS		066 066	066 066		066 - NEGATIVE	19	19	19	100	
2020-451	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	107	107	107	100	
2020-451	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	36	36	36	100	
2020-451	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	107	106	106	100	
2020-451	089	AIMSTRIP		066 066	066 066		069 - POS.(3+)	107	1	106	0	
2020-451	094	CLINITEK		066 066	066 066		066 - NEGATIVE	109	109	109	100	
2020-451	102	ACON LAB		066 066	066 066		066 - NEGATIVE	175	175	175	100	
2020-451	117	URS 10 A		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-451	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	27	27	27	100	
2020-451	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	12	12	12	100	
2020-451	138	URS-10T		066 066	066 066		066 - NEGATIVE	12	12	12	100	
2020-451	140	MISSION A		066 066	066 066		066 - NEGATIVE	63	63	63	100	
2020-451	147	Mission120		066 066	066 066		066 - NEGATIVE	16	16	16	100	
2020-451	148	Mission500		066 066	066 066		066 - NEGATIVE	16	16	16	100	
2020-452	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	49	43	47	100	
2020-452	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	49	4	47	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-452	011	AMES NM10S		068 069	068 069		067 - TRACES	49	2	47	0
2020-452	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	37	37	37	100
2020-452	067	URISCAN		068 069	068 069		069 - POS.(3+))	50	43	50	100
2020-452	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	50	7	50	100
2020-452	071	ATLAS		068 069	068 069		069 - POS.(3+))	19	19	19	100
2020-452	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	107	13	107	100
2020-452	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	107	94	107	100
2020-452	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	36	32	36	100
2020-452	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	36	4	36	100
2020-452	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	107	105	106	100
2020-452	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	107	1	106	100
2020-452	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	107	1	106	0
2020-452	094	CLINITEK		068 069	068 069		066 - NEGATIVE	109	1	108	0
2020-452	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	109	8	108	100
2020-452	094	CLINITEK		068 069	068 069		069 - POS.(3+))	109	100	108	100
2020-452	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	175	1	174	100
2020-452	102	ACON LAB		068 069	068 069		066 - NEGATIVE	175	1	174	0
2020-452	102	ACON LAB		068 069	068 069		069 - POS.(3+))	175	173	174	100
2020-452	117	URS 10 A		068 069	068 069		068 - POS.(1-2+)	10	1	10	100
2020-452	117	URS 10 A		068 069	068 069		069 - POS.(3+))	10	9	10	100
2020-452	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	27	8	27	100
2020-452	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	27	19	27	100
2020-452	135	Hybrido AU		067 069	068 069		069 - POS.(3+))	12	1	12	100
2020-452	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	11	12	100
2020-452	138	URS-10T		068 069	068 069		069 - POS.(3+))	12	11	12	100
2020-452	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	12	1	12	100
2020-452	140	MISSION A		068 069	068 069		069 - POS.(3+))	63	62	63	100
2020-452	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	63	1	63	100
2020-452	147	Mission120		068 069	068 069		069 - POS.(3+))	16	16	16	100
2020-452	148	Mission500		068 069	068 069		069 - POS.(3+))	16	16	16	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
059 URINALYS.HG.OR BLOOD													
2020-453	011	AMES NM10S		068	069	068	069	068 - POS.(1-2+)	49	3	49	100	
2020-453	011	AMES NM10S		068	069	068	069	069 - POS.(3+)	49	46	49	100	
2020-453	013	BAYER CLI		068	069	068	069	069 - POS.(3+)	37	37	37	100	
2020-453	067	URISCAN		068	069	068	069	069 - POS.(3+)	50	47	50	100	
2020-453	067	URISCAN		068	069	068	069	068 - POS.(1-2+)	50	3	50	100	
2020-453	071	ATLAS		068	069	068	069	069 - POS.(3+)	19	19	19	100	
2020-453	072	URS-9-TECO		068	069	068	069	069 - POS.(3+)	107	103	107	100	
2020-453	072	URS-9-TECO		068	069	068	069	068 - POS.(1-2+)	107	4	107	100	
2020-453	080	ROCHE (CHE		068	069	068	069	068 - POS.(1-2+)	36	4	36	100	
2020-453	080	ROCHE (CHE		068	069	068	069	069 - POS.(3+)	36	32	36	100	
2020-453	089	AIMSTRIP		068	069	068	069	069 - POS.(3+)	107	104	107	100	
2020-453	089	AIMSTRIP		068	069	068	069	068 - POS.(1-2+)	107	3	107	100	
2020-453	094	CLINITEK		068	069	068	069	069 - POS.(3+)	109	109	109	100	
2020-453	102	ACON LAB		068	069	068	069	068 - POS.(1-2+)	175	1	174	100	
2020-453	102	ACON LAB		068	069	068	069	066 - NEGATIVE	175	1	174	0	
2020-453	102	ACON LAB		068	069	068	069	069 - POS.(3+)	175	173	174	100	
2020-453	117	URS 10 A		068	069	068	069	069 - POS.(3+)	10	10	10	100	
2020-453	131	iCHEM VELO		068	069	068	069	069 - POS.(3+)	27	25	27	100	
2020-453	131	iCHEM VELO		068	069	068	069	068 - POS.(1-2+)	27	2	27	100	
2020-453	135	Hybrido AU		067	069	068	069	069 - POS.(3+)	12	3	12	100	
2020-453	135	Hybrido AU		067	069	068	069	068 - POS.(1-2+)	12	9	12	100	
2020-453	138	URS-10T		068	069	068	069	068 - POS.(1-2+)	12	1	12	100	
2020-453	138	URS-10T		068	069	068	069	069 - POS.(3+)	12	11	12	100	
2020-453	140	MISSION A		068	069	068	069	069 - POS.(3+)	63	62	63	100	
2020-453	140	MISSION A		068	069	068	069	068 - POS.(1-2+)	63	1	63	100	
2020-453	147	Mission120		068	069	068	069	069 - POS.(3+)	16	16	16	100	
2020-453	148	Mission500		068	069	068	069	069 - POS.(3+)	16	16	16	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
059 URINALYS.HG.OR BLOOD												
2020-454	011	AMES NM10S		066	066	066	066	066 - NEGATIVE	49	49	49	100
2020-454	013	BAYER CLI		066	066	066	066	066 - NEGATIVE	37	37	37	100
2020-454	067	URISCAN		066	066	066	066	066 - NEGATIVE	50	50	50	100
2020-454	071	ATLAS		066	066	066	066	066 - NEGATIVE	19	19	19	100
2020-454	072	URS-9-TECO		066	066	066	066	066 - NEGATIVE	107	107	107	100
2020-454	080	ROCHE (CHE		066	066	066	066	066 - NEGATIVE	36	36	36	100
2020-454	089	AIMSTRIP		066	066	066	066	069 - POS.(3+)	107	1	106	0
2020-454	089	AIMSTRIP		066	066	066	066	066 - NEGATIVE	107	106	106	100
2020-454	094	CLINITEK		066	066	066	066	066 - NEGATIVE	109	107	107	100
2020-454	094	CLINITEK		066	066	066	066	067 - TRACES	109	2	107	0
2020-454	102	ACON LAB		066	066	066	066	066 - NEGATIVE	175	174	174	100
2020-454	102	ACON LAB		066	066	066	066	069 - POS.(3+)	175	1	174	0
2020-454	117	URS 10 A		066	066	066	066	066 - NEGATIVE	10	10	10	100
2020-454	131	iCHEM VELO		066	066	066	066	069 - POS.(3+)	27	1	26	0
2020-454	131	iCHEM VELO		066	066	066	066	066 - NEGATIVE	27	26	26	100
2020-454	135	Hybrido AU		066	066	066	066	066 - NEGATIVE	12	12	12	100
2020-454	138	URS-10T		066	066	066	066	066 - NEGATIVE	12	12	12	100
2020-454	140	MISSION A		066	066	066	066	066 - NEGATIVE	63	63	63	100
2020-454	147	Mission120		066	066	066	066	066 - NEGATIVE	16	16	16	100
2020-454	148	Mission500		066	066	066	066	066 - NEGATIVE	16	16	16	100
2020-455	011	AMES NM10S		068	069	068	069	069 - POS.(3+)	49	45	49	100
2020-455	011	AMES NM10S		068	069	068	069	068 - POS.(1-2+)	49	4	49	100
2020-455	013	BAYER CLI		068	069	068	069	069 - POS.(3+)	37	37	37	100
2020-455	067	URISCAN		068	069	068	069	069 - POS.(3+)	50	46	50	100
2020-455	067	URISCAN		068	069	068	069	068 - POS.(1-2+)	50	4	50	100
2020-455	071	ATLAS		068	069	068	069	069 - POS.(3+)	19	19	19	100
2020-455	072	URS-9-TECO		068	069	068	069	068 - POS.(1-2+)	107	8	105	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	072	URS-9-TECO		068 069	068 069		067 - TRACES	107	1	105	0
2020-455	072	URS-9-TECO		068 069	068 069		066 - NEGATIVE	107	1	105	0
2020-455	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	107	97	105	100
2020-455	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	36	31	36	100
2020-455	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	36	5	36	100
2020-455	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	107	6	106	100
2020-455	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	107	1	106	0
2020-455	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	107	100	106	100
2020-455	094	CLINITEK		068 069	068 069		066 - NEGATIVE	109	1	108	0
2020-455	094	CLINITEK		068 069	068 069		069 - POS.(3+))	109	102	108	100
2020-455	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	109	6	108	100
2020-455	102	ACON LAB		068 069	068 069		066 - NEGATIVE	175	1	174	0
2020-455	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	175	5	174	100
2020-455	102	ACON LAB		068 069	068 069		069 - POS.(3+))	175	169	174	100
2020-455	117	URS 10 A		068 069	068 069		069 - POS.(3+))	10	10	10	100
2020-455	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	27	25	27	100
2020-455	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	27	2	27	100
2020-455	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	7	12	100
2020-455	135	Hybrido AU		067 069	068 069		069 - POS.(3+))	12	5	12	100
2020-455	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	12	2	12	100
2020-455	138	URS-10T		068 069	068 069		069 - POS.(3+))	12	10	12	100
2020-455	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	63	2	63	100
2020-455	140	MISSION A		068 069	068 069		069 - POS.(3+))	63	61	63	100
2020-455	147	Mission120		068 069	068 069		069 - POS.(3+))	16	16	16	100
2020-455	148	Mission500		068 069	068 069		068 - POS.(1-2+)	16	2	16	100
2020-455	148	Mission500		068 069	068 069		069 - POS.(3+))	16	14	16	100
060 URINALYSIS-KETONES											
2020-451	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	49	49	49	100
2020-451	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	37	37	37	100
2020-451	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-451	071	ATLAS		090 090	090 090		090 - NEGATIVE	19	19	19	100
2020-451	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	108	108	108	100
2020-451	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	36	36	36	100
2020-451	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	107	107	107	100
2020-451	094	CLINITEK		090 090	090 090		090 - NEGATIVE	109	109	109	100
2020-451	102	ACON LAB		090 090	090 090		090 - NEGATIVE	173	173	173	100
2020-451	117	URS 10 A		090 090	090 090		090 - NEGATIVE	10	10	10	100
2020-451	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100
2020-451	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-451	138	URS-10T		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-451	140	MISSION A		090 090	090 090		090 - NEGATIVE	63	62	62	100
2020-451	140	MISSION A		090 090	090 090		091 - TRACE	63	1	62	0
2020-451	147	Mission120		090 090	090 090		090 - NEGATIVE	16	16	16	100
2020-451	148	Mission500		090 090	090 090		090 - NEGATIVE	17	17	17	100
060 URINALYSIS-KETONES											
2020-452	011	AMES NM10S		093 094	093 094		094 - LARGE-3-4+	49	48	49	100
2020-452	011	AMES NM10S		093 094	093 094		093 - MOD. (2+)	49	1	49	100
2020-452	013	BAYER CLI		093 094	093 094		094 - LARGE-3-4+	37	35	37	100
2020-452	013	BAYER CLI		093 094	093 094		093 - MOD. (2+)	37	2	37	100
2020-452	067	URISCAN		093 094	093 094		092 - SMALL (1+)	50	1	49	0
2020-452	067	URISCAN		093 094	093 094		094 - LARGE-3-4+	50	37	49	100
2020-452	067	URISCAN		093 094	093 094		093 - MOD. (2+)	50	12	49	100
2020-452	071	ATLAS		093 094	093 094		094 - LARGE-3-4+	19	19	19	100
2020-452	072	URS-9-TECO		093 094	093 094		093 - MOD. (2+)	108	4	108	100
2020-452	072	URS-9-TECO		093 094	093 094		094 - LARGE-3-4+	108	104	108	100
2020-452	080	ROCHE (CHE		093 094	093 094		094 - LARGE-3-4+	36	29	36	100
2020-452	080	ROCHE (CHE		093 094	093 094		093 - MOD. (2+)	36	7	36	100
2020-452	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	107	8	106	100
2020-452	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	107	98	106	100
2020-452	089	AIMSTRIP		093 094	093 094		090 - NEGATIVE	107	1	106	0

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	094	CLINITEK		093 094	093 094		093 - MOD. (2+)	109	6	109	100	
2020-452	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	109	103	109	100	
2020-452	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	173	169	172	100	
2020-452	102	ACON LAB		093 094	093 094		090 - NEGATIVE	173	1	172	0	
2020-452	102	ACON LAB		093 094	093 094		093 - MOD. (2+)	173	3	172	100	
2020-452	117	URS 10 A		093 094	093 094		094 - LARGE-3-4+	10	8	10	100	
2020-452	117	URS 10 A		093 094	093 094		093 - MOD. (2+)	10	2	10	100	
2020-452	131	iCHEM VELO		092 094	093 094		093 - MOD. (2+)	27	16	27	100	
2020-452	131	iCHEM VELO		092 094	093 094		094 - LARGE-3-4+	27	11	27	100	
2020-452	135	Hybrido AU		093 094	093 094		094 - LARGE-3-4+	12	12	12	100	
2020-452	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	12	10	11	100	
2020-452	138	URS-10T		093 094	093 094		092 - SMALL (1+)	12	1	11	0	
2020-452	138	URS-10T		093 094	093 094		093 - MOD. (2+)	12	1	11	100	
2020-452	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	63	63	63	100	
2020-452	147	Mission120		093 094	093 094		094 - LARGE-3-4+	16	16	16	100	
2020-452	148	Mission500		093 094	093 094		094 - LARGE-3-4+	17	17	17	100	
060 URINALYSIS-KETONES												
2020-453	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	49	49	49	100	
2020-453	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	37	37	37	100	
2020-453	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100	
2020-453	071	ATLAS		090 090	090 090		090 - NEGATIVE	19	19	19	100	
2020-453	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	108	108	108	100	
2020-453	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	36	36	36	100	
2020-453	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	107	107	107	100	
2020-453	094	CLINITEK		090 090	090 090		090 - NEGATIVE	109	108	108	100	
2020-453	094	CLINITEK		090 090	090 090		094 - LARGE-3-4+	109	1	108	0	
2020-453	102	ACON LAB		090 090	090 090		090 - NEGATIVE	173	172	172	100	
2020-453	102	ACON LAB		090 090	090 090		091 - TRACE	173	1	172	0	
2020-453	117	URS 10 A		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-453	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-453	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-453	138	URS-10T		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-453	140	MISSION A		090 090	090 090		090 - NEGATIVE	63	63	63	100
2020-453	147	Mission120		090 090	090 090		090 - NEGATIVE	16	16	16	100
2020-453	148	Mission500		090 090	090 090		090 - NEGATIVE	17	17	17	100
060 URINALYSIS-KETONES											
2020-454	011	AMES NM10S		092 094	092 094		094 - LARGE-3-4+	49	17	49	100
2020-454	011	AMES NM10S		092 094	092 094		092 - SMALL (1+)	49	3	49	100
2020-454	011	AMES NM10S		092 094	092 094		093 - MOD. (2+)	49	29	49	100
2020-454	013	BAYER CLI		092 094	092 094		092 - SMALL (1+)	37	4	37	100
2020-454	013	BAYER CLI		092 094	092 094		093 - MOD. (2+)	37	29	37	100
2020-454	013	BAYER CLI		092 094	092 094		094 - LARGE-3-4+	37	4	37	100
2020-454	067	URISCAN		092 094	092 094		091 - TRACE	50	1	49	0
2020-454	067	URISCAN		092 094	092 094		093 - MOD. (2+)	50	24	49	100
2020-454	067	URISCAN		092 094	092 094		092 - SMALL (1+)	50	4	49	100
2020-454	067	URISCAN		092 094	092 094		094 - LARGE-3-4+	50	21	49	100
2020-454	071	ATLAS		092 094	092 094		092 - SMALL (1+)	19	1	18	100
2020-454	071	ATLAS		092 094	092 094		093 - MOD. (2+)	19	17	18	100
2020-454	071	ATLAS		092 094	092 094		091 - TRACE	19	1	18	0
2020-454	072	URS-9-TECO		092 094	092 094		094 - LARGE-3-4+	108	45	108	100
2020-454	072	URS-9-TECO		092 094	092 094		093 - MOD. (2+)	108	60	108	100
2020-454	072	URS-9-TECO		092 094	092 094		092 - SMALL (1+)	108	3	108	100
2020-454	080	ROCHE (CHE		092 094	092 094		093 - MOD. (2+)	36	20	36	100
2020-454	080	ROCHE (CHE		092 094	092 094		092 - SMALL (1+)	36	4	36	100
2020-454	080	ROCHE (CHE		092 094	092 094		094 - LARGE-3-4+	36	12	36	100
2020-454	089	AIMSTRIP		092 094	092 094		093 - MOD. (2+)	107	52	107	100
2020-454	089	AIMSTRIP		092 094	092 094		092 - SMALL (1+)	107	13	107	100
2020-454	089	AIMSTRIP		092 094	092 094		094 - LARGE-3-4+	107	42	107	100
2020-454	094	CLINITEK		092 094	092 094		094 - LARGE-3-4+	109	8	109	100
2020-454	094	CLINITEK		092 094	092 094		093 - MOD. (2+)	109	85	109	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-454	094	CLINITEK		092	094	092	094	092 - SMALL (1+)	109	16	109	100	
2020-454	102	ACON LAB		093	094	092	094	092 - SMALL (1+)	173	7	173	100	
2020-454	102	ACON LAB		093	094	092	094	093 - MOD. (2+)	173	61	173	100	
2020-454	102	ACON LAB		093	094	092	094	094 - LARGE-3-4+	173	105	173	100	
2020-454	117	URS 10 A		092	094	092	094	094 - LARGE-3-4+	10	4	10	100	
2020-454	117	URS 10 A		092	094	092	094	093 - MOD. (2+)	10	4	10	100	
2020-454	117	URS 10 A		092	094	092	094	092 - SMALL (1+)	10	2	10	100	
2020-454	131	iCHEM VELO		091	093	092	094	094 - LARGE-3-4+	27	2	27	100	
2020-454	131	iCHEM VELO		091	093	092	094	092 - SMALL (1+)	27	14	27	100	
2020-454	131	iCHEM VELO		091	093	092	094	093 - MOD. (2+)	27	11	27	100	
2020-454	135	Hybrido AU		092	094	092	094	093 - MOD. (2+)	12	11	12	100	
2020-454	135	Hybrido AU		092	094	092	094	094 - LARGE-3-4+	12	1	12	100	
2020-454	138	URS-10T		093	094	092	094	092 - SMALL (1+)	12	2	12	100	
2020-454	138	URS-10T		093	094	092	094	093 - MOD. (2+)	12	2	12	100	
2020-454	138	URS-10T		093	094	092	094	094 - LARGE-3-4+	12	8	12	100	
2020-454	140	MISSION A		093	094	092	094	093 - MOD. (2+)	63	10	63	100	
2020-454	140	MISSION A		093	094	092	094	094 - LARGE-3-4+	63	53	63	100	
2020-454	147	Mission120		093	094	092	094	094 - LARGE-3-4+	16	16	16	100	
2020-454	148	Mission500		093	094	092	094	093 - MOD. (2+)	17	6	17	100	
2020-454	148	Mission500		093	094	092	094	094 - LARGE-3-4+	17	11	17	100	
060 URINALYSIS-KETONES													
2020-455	011	AMES NM10S		090	090	090	090	090 - NEGATIVE	49	49	49	100	
2020-455	013	BAYER CLI		090	090	090	090	090 - NEGATIVE	37	37	37	100	
2020-455	067	URISCAN		090	090	090	090	090 - NEGATIVE	50	50	50	100	
2020-455	071	ATLAS		090	090	090	090	090 - NEGATIVE	19	19	19	100	
2020-455	072	URS-9-TECO		090	090	090	090	090 - NEGATIVE	108	108	108	100	
2020-455	080	ROCHE (CHE		090	090	090	090	090 - NEGATIVE	36	36	36	100	
2020-455	089	AIMSTRIP		090	090	090	090	094 - LARGE-3-4+	107	2	105	0	
2020-455	089	AIMSTRIP		090	090	090	090	090 - NEGATIVE	107	105	105	100	
2020-455	094	CLINITEK		090	090	090	090	090 - NEGATIVE	109	109	109	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	102	ACON LAB		090 090	090 090		094 - LARGE-3-4+	173	1	172	0
2020-455	102	ACON LAB		090 090	090 090		090 - NEGATIVE	173	172	172	100
2020-455	117	URS 10 A		090 090	090 090		090 - NEGATIVE	10	10	10	100
2020-455	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100
2020-455	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-455	138	URS-10T		090 090	090 090		090 - NEGATIVE	12	12	12	100
2020-455	140	MISSION A		090 090	090 090		090 - NEGATIVE	63	63	63	100
2020-455	147	Mission120		090 090	090 090		090 - NEGATIVE	16	16	16	100
2020-455	148	Mission500		090 090	090 090		090 - NEGATIVE	17	17	17	100
061 URINALYSIS-NITRATE											
2020-451	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	49	49	49	100
2020-451	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	38	38	38	100
2020-451	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100
2020-451	071	ATLAS		070 070	070 070		070 - NEGATIVE	19	19	19	100
2020-451	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	106	106	106	100
2020-451	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	36	36	36	100
2020-451	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	107	107	107	100
2020-451	094	CLINITEK		070 070	070 070		070 - NEGATIVE	109	109	109	100
2020-451	102	ACON LAB		070 070	070 070		070 - NEGATIVE	174	174	174	100
2020-451	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100
2020-451	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100
2020-451	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100
2020-451	138	URS-10T		070 070	070 070		070 - NEGATIVE	12	12	12	100
2020-451	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100
2020-451	147	Mission120		070 070	070 070		070 - NEGATIVE	16	16	16	100
2020-451	148	Mission500		070 070	070 070		070 - NEGATIVE	17	17	17	100
2020-452	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	49	49	49	100
2020-452	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	38	38	38	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score			
2020-452	067	URISCAN		070	070	070	070	070	-	NEGATIVE	50	50	50	100
2020-452	071	ATLAS		070	070	070	070	070	-	NEGATIVE	19	19	19	100
2020-452	072	URS-9-TECO		070	070	070	070	070	-	NEGATIVE	106	106	106	100
2020-452	080	ROCHE (CHE		070	070	070	070	070	-	NEGATIVE	36	36	36	100
2020-452	089	AIMSTRIP		070	070	070	070	070	-	NEGATIVE	107	107	107	100
2020-452	094	CLINITEK		070	070	070	070	070	-	NEGATIVE	109	109	109	100
2020-452	102	ACON LAB		070	070	070	070	070	-	NEGATIVE	174	174	174	100
2020-452	117	URS 10 A		070	070	070	070	070	-	NEGATIVE	10	10	10	100
2020-452	131	iCHEM VELO		070	070	070	070	070	-	NEGATIVE	27	27	27	100
2020-452	135	Hybrido AU		070	070	070	070	070	-	NEGATIVE	12	12	12	100
2020-452	138	URS-10T		070	070	070	070	070	-	NEGATIVE	12	12	12	100
2020-452	140	MISSION A		070	070	070	070	070	-	NEGATIVE	63	63	63	100
2020-452	147	Mission120		070	070	070	070	070	-	NEGATIVE	16	16	16	100
2020-452	148	Mission500		070	070	070	070	070	-	NEGATIVE	17	17	17	100
061 URINALYSIS-NITRATE														
2020-453	011	AMES NM10S		070	070	070	070	070	-	NEGATIVE	49	49	49	100
2020-453	013	BAYER CLI		070	070	070	070	070	-	NEGATIVE	38	38	38	100
2020-453	067	URISCAN		070	070	070	070	070	-	NEGATIVE	50	50	50	100
2020-453	071	ATLAS		070	070	070	070	070	-	NEGATIVE	19	19	19	100
2020-453	072	URS-9-TECO		070	070	070	070	070	-	NEGATIVE	106	106	106	100
2020-453	080	ROCHE (CHE		070	070	070	070	070	-	NEGATIVE	36	36	36	100
2020-453	089	AIMSTRIP		070	070	070	070	070	-	NEGATIVE	107	107	107	100
2020-453	094	CLINITEK		070	070	070	070	070	-	NEGATIVE	109	109	109	100
2020-453	102	ACON LAB		070	070	070	070	070	-	NEGATIVE	174	174	174	100
2020-453	117	URS 10 A		070	070	070	070	070	-	NEGATIVE	10	10	10	100
2020-453	131	iCHEM VELO		070	070	070	070	070	-	NEGATIVE	27	27	27	100
2020-453	135	Hybrido AU		070	070	070	070	070	-	NEGATIVE	12	12	12	100
2020-453	138	URS-10T		070	070	070	070	070	-	NEGATIVE	12	12	12	100
2020-453	140	MISSION A		070	070	070	070	070	-	NEGATIVE	63	63	63	100
2020-453	147	Mission120		070	070	070	070	070	-	NEGATIVE	16	16	16	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-453	148	Mission500		070 070	070 070		070 - NEGATIVE	17	17	17	100		
061 URINALYSIS-NITRATE													
2020-454	011	AMES NM10S		071 072	071 072		072 - POSITIVE	49	49	49	100		
2020-454	013	BAYER CLI		071 072	071 072		072 - POSITIVE	38	38	38	100		
2020-454	067	URISCAN		071 072	071 072		072 - POSITIVE	50	50	50	100		
2020-454	071	ATLAS		071 072	071 072		072 - POSITIVE	19	19	19	100		
2020-454	072	URS-9-TECO		071 072	071 072		072 - POSITIVE	106	105	106	100		
2020-454	072	URS-9-TECO		071 072	071 072		071 - WEAK POS.	106	1	106	100		
2020-454	080	ROCHE (CHE		071 072	071 072		072 - POSITIVE	36	36	36	100		
2020-454	089	AIMSTRIP		071 072	071 072		072 - POSITIVE	107	106	106	100		
2020-454	089	AIMSTRIP		071 072	071 072		070 - NEGATIVE	107	1	106	0		
2020-454	094	CLINITEK		071 072	071 072		070 - NEGATIVE	109	1	108	0		
2020-454	094	CLINITEK		071 072	071 072		072 - POSITIVE	109	108	108	100		
2020-454	102	ACON LAB		071 072	071 072		072 - POSITIVE	174	173	174	100		
2020-454	102	ACON LAB		071 072	071 072		071 - WEAK POS.	174	1	174	100		
2020-454	117	URS 10 A		071 072	071 072		072 - POSITIVE	10	10	10	100		
2020-454	131	iCHEM VELO		071 072	071 072		072 - POSITIVE	27	27	27	100		
2020-454	135	Hybrido AU		071 072	071 072		072 - POSITIVE	12	12	12	100		
2020-454	138	URS-10T		071 072	071 072		072 - POSITIVE	12	12	12	100		
2020-454	140	MISSION A		071 072	071 072		072 - POSITIVE	63	63	63	100		
2020-454	147	Mission120		071 072	071 072		072 - POSITIVE	16	16	16	100		
2020-454	148	Mission500		071 072	071 072		072 - POSITIVE	17	17	17	100		
2020-455	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	49	49	49	100		
2020-455	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	38	38	38	100		
2020-455	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100		
2020-455	071	ATLAS		070 070	070 070		070 - NEGATIVE	19	19	19	100		
2020-455	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	106	106	106	100		
2020-455	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	36	36	36	100		

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-455	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	107	107	107	100	
2020-455	094	CLINITEK		070 070	070 070		070 - NEGATIVE	109	109	109	100	
2020-455	102	ACON LAB		070 070	070 070		070 - NEGATIVE	174	174	174	100	
2020-455	117	URS 10 A		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2020-455	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100	
2020-455	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2020-455	138	URS-10T		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2020-455	140	MISSION A		070 070	070 070		070 - NEGATIVE	63	63	63	100	
2020-455	147	Mission120		070 070	070 070		070 - NEGATIVE	16	16	16	100	
2020-455	148	Mission500		070 070	070 070		070 - NEGATIVE	17	17	17	100	
062 URINALYSIS-pH												
2020-451	011	AMES NM10S		086 088	086 088		087 - 7.5	49	10	49	100	
2020-451	011	AMES NM10S		086 088	086 088		086 - 7.0	49	1	49	100	
2020-451	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	49	38	49	100	
2020-451	013	BAYER CLI		086 088	086 088		088 - 8.0OR MORE	38	25	38	100	
2020-451	013	BAYER CLI		086 088	086 088		087 - 7.5	38	13	38	100	
2020-451	067	URISCAN		085 088	086 088		086 - 7.0	50	3	50	100	
2020-451	067	URISCAN		085 088	086 088		087 - 7.5	50	37	50	100	
2020-451	067	URISCAN		085 088	086 088		088 - 8.0OR MORE	50	10	50	100	
2020-451	071	ATLAS		085 088	086 088		087 - 7.5	19	14	19	100	
2020-451	071	ATLAS		085 088	086 088		088 - 8.0OR MORE	19	5	19	100	
2020-451	072	URS-9-TECO		085 088	086 088		087 - 7.5	106	64	106	100	
2020-451	072	URS-9-TECO		085 088	086 088		086 - 7.0	106	21	106	100	
2020-451	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	106	19	106	100	
2020-451	072	URS-9-TECO		085 088	086 088		085 - 6.5	106	2	106	100	
2020-451	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	36	22	36	100	
2020-451	080	ROCHE (CHE		086 088	086 088		087 - 7.5	36	1	36	100	
2020-451	080	ROCHE (CHE		086 088	086 088		086 - 7.0	36	13	36	100	
2020-451	089	AIMSTRIP		086 088	086 088		086 - 7.0	107	3	107	100	
2020-451	089	AIMSTRIP		086 088	086 088		087 - 7.5	107	46	107	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-451	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	107	58	107	100	
2020-451	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	109	82	109	100	
2020-451	094	CLINITEK		086 088	086 088		087 - 7.5	109	25	109	100	
2020-451	094	CLINITEK		086 088	086 088		086 - 7.0	109	2	109	100	
2020-451	102	ACON LAB		086 088	086 088		086 - 7.0	174	1	173	100	
2020-451	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	174	123	173	100	
2020-451	102	ACON LAB		086 088	086 088		085 - 6.5	174	1	173	0	
2020-451	102	ACON LAB		086 088	086 088		087 - 7.5	174	49	173	100	
2020-451	117	URS 10 A		085 088	086 088		088 - 8.00R MORE	10	4	10	100	
2020-451	117	URS 10 A		085 088	086 088		085 - 6.5	10	1	10	100	
2020-451	117	URS 10 A		085 088	086 088		087 - 7.5	10	5	10	100	
2020-451	131	iCHEM VELO		086 088	086 088		086 - 7.0	27	12	27	100	
2020-451	131	iCHEM VELO		086 088	086 088		088 - 8.00R MORE	27	15	27	100	
2020-451	135	Hybrido AU		086 088	086 088		087 - 7.5	12	2	12	100	
2020-451	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	12	10	12	100	
2020-451	138	URS-10T		085 088	086 088		086 - 7.0	12	4	12	100	
2020-451	138	URS-10T		085 088	086 088		087 - 7.5	12	8	12	100	
2020-451	140	MISSION A		086 088	086 088		088 - 8.00R MORE	63	41	63	100	
2020-451	140	MISSION A		086 088	086 088		087 - 7.5	63	22	63	100	
2020-451	147	Mission120		086 088	086 088		087 - 7.5	16	7	16	100	
2020-451	147	Mission120		086 088	086 088		088 - 8.00R MORE	16	9	16	100	
2020-451	148	Mission500		086 088	086 088		088 - 8.00R MORE	17	14	17	100	
2020-451	148	Mission500		086 088	086 088		087 - 7.5	17	3	17	100	
062 URINALYSIS-pH												
2020-452	011	AMES NM10S		086 088	086 088		088 - 8.00R MORE	49	42	49	100	
2020-452	011	AMES NM10S		086 088	086 088		087 - 7.5	49	6	49	100	
2020-452	011	AMES NM10S		086 088	086 088		086 - 7.0	49	1	49	100	
2020-452	013	BAYER CLI		086 088	086 088		088 - 8.00R MORE	38	31	37	100	
2020-452	013	BAYER CLI		086 088	086 088		087 - 7.5	38	6	37	100	
2020-452	013	BAYER CLI		086 088	086 088		084 - 6.0	38	1	37	0	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	067	URISCAN		085 088	086 088		086 - 7.0	50	2	50	100	
2020-452	067	URISCAN		085 088	086 088		088 - 8.0OR MORE	50	12	50	100	
2020-452	067	URISCAN		085 088	086 088		087 - 7.5	50	36	50	100	
2020-452	071	ATLAS		085 088	086 088		087 - 7.5	19	10	19	100	
2020-452	071	ATLAS		085 088	086 088		088 - 8.0OR MORE	19	8	19	100	
2020-452	071	ATLAS		085 088	086 088		086 - 7.0	19	1	19	100	
2020-452	072	URS-9-TECO		085 088	086 088		087 - 7.5	106	70	106	100	
2020-452	072	URS-9-TECO		085 088	086 088		086 - 7.0	106	8	106	100	
2020-452	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	106	28	106	100	
2020-452	080	ROCHE (CHE		086 088	086 088		086 - 7.0	36	1	36	100	
2020-452	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	36	35	36	100	
2020-452	089	AIMSTRIP		086 088	086 088		087 - 7.5	107	17	107	100	
2020-452	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	107	90	107	100	
2020-452	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	109	92	109	100	
2020-452	094	CLINITEK		086 088	086 088		087 - 7.5	109	17	109	100	
2020-452	102	ACON LAB		086 088	086 088		087 - 7.5	174	23	174	100	
2020-452	102	ACON LAB		086 088	086 088		086 - 7.0	174	2	174	100	
2020-452	102	ACON LAB		086 088	086 088		088 - 8.0OR MORE	174	149	174	100	
2020-452	117	URS 10 A		086 088	086 088		087 - 7.5	10	2	10	100	
2020-452	117	URS 10 A		086 088	086 088		088 - 8.0OR MORE	10	8	10	100	
2020-452	131	iCHEM VELO		086 088	086 088		086 - 7.0	27	13	27	100	
2020-452	131	iCHEM VELO		086 088	086 088		088 - 8.0OR MORE	27	14	27	100	
2020-452	135	Hybrido AU		086 088	086 088		088 - 8.0OR MORE	12	12	12	100	
2020-452	138	URS-10T		085 088	086 088		088 - 8.0OR MORE	12	3	12	100	
2020-452	138	URS-10T		085 088	086 088		086 - 7.0	12	1	12	100	
2020-452	138	URS-10T		085 088	086 088		087 - 7.5	12	8	12	100	
2020-452	140	MISSION A		086 088	086 088		088 - 8.0OR MORE	63	57	63	100	
2020-452	140	MISSION A		086 088	086 088		087 - 7.5	63	6	63	100	
2020-452	147	Mission120		086 088	086 088		087 - 7.5	16	5	16	100	
2020-452	147	Mission120		086 088	086 088		088 - 8.0OR MORE	16	11	16	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	148	Mission500		086 088	086 088		087 - 7.5	17	3	17	100	
2020-452	148	Mission500		086 088	086 088		088 - 8.0OR MORE	17	14	17	100	
062 URINALYSIS-pH												
2020-453	011	AMES NM10S		084 088	084 088		086 - 7.0	49	28	49	100	
2020-453	011	AMES NM10S		084 088	084 088		087 - 7.5	49	21	49	100	
2020-453	013	BAYER CLI		085 088	084 088		087 - 7.5	38	19	38	100	
2020-453	013	BAYER CLI		085 088	084 088		086 - 7.0	38	19	38	100	
2020-453	067	URISCAN		084 088	084 088		085 - 6.5	50	22	50	100	
2020-453	067	URISCAN		084 088	084 088		087 - 7.5	50	1	50	100	
2020-453	067	URISCAN		084 088	084 088		086 - 7.0	50	27	50	100	
2020-453	071	ATLAS		084 088	084 088		086 - 7.0	19	19	19	100	
2020-453	072	URS-9-TECO		084 088	084 088		086 - 7.0	106	94	106	100	
2020-453	072	URS-9-TECO		084 088	084 088		085 - 6.5	106	8	106	100	
2020-453	072	URS-9-TECO		084 088	084 088		087 - 7.5	106	4	106	100	
2020-453	080	ROCHE (CHE		084 088	084 088		086 - 7.0	36	36	36	100	
2020-453	089	AIMSTRIP		084 088	084 088		084 - 6.0	107	1	107	100	
2020-453	089	AIMSTRIP		084 088	084 088		087 - 7.5	107	26	107	100	
2020-453	089	AIMSTRIP		084 088	084 088		086 - 7.0	107	79	107	100	
2020-453	089	AIMSTRIP		084 088	084 088		085 - 6.5	107	1	107	100	
2020-453	094	CLINITEK		084 088	084 088		087 - 7.5	109	54	109	100	
2020-453	094	CLINITEK		084 088	084 088		086 - 7.0	109	54	109	100	
2020-453	094	CLINITEK		084 088	084 088		085 - 6.5	109	1	109	100	
2020-453	102	ACON LAB		084 088	084 088		087 - 7.5	174	21	174	100	
2020-453	102	ACON LAB		084 088	084 088		086 - 7.0	174	148	174	100	
2020-453	102	ACON LAB		084 088	084 088		085 - 6.5	174	5	174	100	
2020-453	117	URS 10 A		084 088	084 088		086 - 7.0	10	8	10	100	
2020-453	117	URS 10 A		084 088	084 088		085 - 6.5	10	1	10	100	
2020-453	117	URS 10 A		084 088	084 088		087 - 7.5	10	1	10	100	
2020-453	131	iCHEM VELO		084 088	084 088		086 - 7.0	27	27	27	100	
2020-453	135	Hybrido AU		084 088	084 088		087 - 7.5	12	3	12	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-453	135	Hybrido AU		084 088	084 088		086 - 7.0	12	9	12	100	
2020-453	138	URS-10T		084 088	084 088		086 - 7.0	12	10	12	100	
2020-453	138	URS-10T		084 088	084 088		085 - 6.5	12	1	12	100	
2020-453	138	URS-10T		084 088	084 088		084 - 6.0	12	1	12	100	
2020-453	140	MISSION A		084 088	084 088		087 - 7.5	63	1	63	100	
2020-453	140	MISSION A		084 088	084 088		086 - 7.0	63	59	63	100	
2020-453	140	MISSION A		084 088	084 088		085 - 6.5	63	3	63	100	
2020-453	147	Mission120		084 088	084 088		085 - 6.5	16	1	16	100	
2020-453	147	Mission120		084 088	084 088		086 - 7.0	16	15	16	100	
2020-453	148	Mission500		084 088	084 088		086 - 7.0	17	17	17	100	
062 URINALYSIS-pH												
2020-454	011	AMES NM10S		081 085	080 084		083 - 5.5	49	22	49	100	
2020-454	011	AMES NM10S		081 085	080 084		082 - 5.0	49	18	49	100	
2020-454	011	AMES NM10S		081 085	080 084		084 - 6.0	49	9	49	100	
2020-454	013	BAYER CLI		081 085	080 084		082 - 5.0	38	16	38	100	
2020-454	013	BAYER CLI		081 085	080 084		084 - 6.0	38	5	38	100	
2020-454	013	BAYER CLI		081 085	080 084		083 - 5.5	38	17	38	100	
2020-454	067	URISCAN		080 084	080 084		082 - 5.0	50	42	50	100	
2020-454	067	URISCAN		080 084	080 084		083 - 5.5	50	8	50	100	
2020-454	071	ATLAS		080 084	080 084		082 - 5.0	19	18	19	100	
2020-454	071	ATLAS		080 084	080 084		083 - 5.5	19	1	19	100	
2020-454	072	URS-9-TECO		082 086	080 084		084 - 6.0	106	68	106	100	
2020-454	072	URS-9-TECO		082 086	080 084		083 - 5.5	106	20	106	100	
2020-454	072	URS-9-TECO		082 086	080 084		082 - 5.0	106	15	106	100	
2020-454	072	URS-9-TECO		082 086	080 084		085 - 6.5	106	3	106	100	
2020-454	080	ROCHE (CHE		080 084	080 084		082 - 5.0	36	36	36	100	
2020-454	089	AIMSTRIP		080 084	080 084		085 - 6.5	107	1	104	0	
2020-454	089	AIMSTRIP		080 084	080 084		084 - 6.0	107	16	104	100	
2020-454	089	AIMSTRIP		080 084	080 084		083 - 5.5	107	30	104	100	
2020-454	089	AIMSTRIP		080 084	080 084		087 - 7.5	107	2	104	0	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-454	089	AIMSTRIP		080 084	080 084		082 - 5.0	107	58	104	100
2020-454	094	CLINITEK		081 085	080 084		082 - 5.0	109	26	109	100
2020-454	094	CLINITEK		081 085	080 084		083 - 5.5	109	72	109	100
2020-454	094	CLINITEK		081 085	080 084		084 - 6.0	109	11	109	100
2020-454	102	ACON LAB		082 086	080 084		084 - 6.0	174	85	174	100
2020-454	102	ACON LAB		082 086	080 084		083 - 5.5	174	41	174	100
2020-454	102	ACON LAB		082 086	080 084		082 - 5.0	174	48	174	100
2020-454	117	URS 10 A		080 084	080 084		083 - 5.5	10	3	9	100
2020-454	117	URS 10 A		080 084	080 084		084 - 6.0	10	1	9	100
2020-454	117	URS 10 A		080 084	080 084		082 - 5.0	10	5	9	100
2020-454	117	URS 10 A		080 084	080 084		085 - 6.5	10	1	9	0
2020-454	131	iCHEM VELO		080 084	080 084		082 - 5.0	27	26	27	100
2020-454	131	iCHEM VELO		080 084	080 084		084 - 6.0	27	1	27	100
2020-454	135	Hybrido AU		080 084	080 084		083 - 5.5	12	1	12	100
2020-454	135	Hybrido AU		080 084	080 084		082 - 5.0	12	11	12	100
2020-454	138	URS-10T		082 086	080 084		082 - 5.0	12	4	12	100
2020-454	138	URS-10T		082 086	080 084		083 - 5.5	12	3	12	100
2020-454	138	URS-10T		082 086	080 084		084 - 6.0	12	5	12	100
2020-454	140	MISSION A		081 085	080 084		082 - 5.0	63	9	63	100
2020-454	140	MISSION A		081 085	080 084		084 - 6.0	63	17	63	100
2020-454	140	MISSION A		081 085	080 084		083 - 5.5	63	37	63	100
2020-454	147	Mission120		081 085	080 084		082 - 5.0	16	5	16	100
2020-454	147	Mission120		081 085	080 084		083 - 5.5	16	11	16	100
2020-454	148	Mission500		082 086	080 084		083 - 5.5	17	1	17	100
2020-454	148	Mission500		082 086	080 084		084 - 6.0	17	15	17	100
2020-454	148	Mission500		082 086	080 084		082 - 5.0	17	1	17	100
062 URINALYSIS-pH											
2020-455	011	AMES NM10S		082 086	082 086		085 - 6.5	49	2	49	100
2020-455	011	AMES NM10S		082 086	082 086		084 - 6.0	49	46	49	100
2020-455	011	AMES NM10S		082 086	082 086		083 - 5.5	49	1	49	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	013	BAYER CLI		082 086	082 086		083 - 5.5	38	2	38	100
2020-455	013	BAYER CLI		082 086	082 086		084 - 6.0	38	36	38	100
2020-455	067	URISCAN		081 085	082 086		084 - 6.0	50	9	50	100
2020-455	067	URISCAN		081 085	082 086		083 - 5.5	50	31	50	100
2020-455	067	URISCAN		081 085	082 086		082 - 5.0	50	10	50	100
2020-455	071	ATLAS		082 086	082 086		084 - 6.0	19	19	19	100
2020-455	072	URS-9-TECO		082 086	082 086		083 - 5.5	106	6	106	100
2020-455	072	URS-9-TECO		082 086	082 086		084 - 6.0	106	67	106	100
2020-455	072	URS-9-TECO		082 086	082 086		086 - 7.0	106	1	106	100
2020-455	072	URS-9-TECO		082 086	082 086		085 - 6.5	106	32	106	100
2020-455	080	ROCHE (CHE		082 086	082 086		084 - 6.0	36	36	36	100
2020-455	089	AIMSTRIP		082 086	082 086		084 - 6.0	107	93	107	100
2020-455	089	AIMSTRIP		082 086	082 086		082 - 5.0	107	5	107	100
2020-455	089	AIMSTRIP		082 086	082 086		085 - 6.5	107	9	107	100
2020-455	094	CLINITEK		082 086	082 086		085 - 6.5	109	1	109	100
2020-455	094	CLINITEK		082 086	082 086		083 - 5.5	109	1	109	100
2020-455	094	CLINITEK		082 086	082 086		084 - 6.0	109	107	109	100
2020-455	102	ACON LAB		082 086	082 086		084 - 6.0	174	167	174	100
2020-455	102	ACON LAB		082 086	082 086		083 - 5.5	174	2	174	100
2020-455	102	ACON LAB		082 086	082 086		085 - 6.5	174	3	174	100
2020-455	102	ACON LAB		082 086	082 086		082 - 5.0	174	2	174	100
2020-455	117	URS 10 A		082 086	082 086		082 - 5.0	10	1	10	100
2020-455	117	URS 10 A		082 086	082 086		085 - 6.5	10	1	10	100
2020-455	117	URS 10 A		082 086	082 086		084 - 6.0	10	7	10	100
2020-455	117	URS 10 A		082 086	082 086		083 - 5.5	10	1	10	100
2020-455	131	iCHEM VELO		082 086	082 086		082 - 5.0	27	1	27	100
2020-455	131	iCHEM VELO		082 086	082 086		084 - 6.0	27	26	27	100
2020-455	135	Hybrido AU		082 086	082 086		085 - 6.5	12	3	12	100
2020-455	135	Hybrido AU		082 086	082 086		084 - 6.0	12	9	12	100
2020-455	138	URS-10T		082 086	082 086		083 - 5.5	12	2	12	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-455	138	URS-10T		082 086	082 086		085 - 6.5	12	1	12	100	
2020-455	138	URS-10T		082 086	082 086		084 - 6.0	12	9	12	100	
2020-455	140	MISSION A		082 086	082 086		085 - 6.5	63	1	63	100	
2020-455	140	MISSION A		082 086	082 086		084 - 6.0	63	61	63	100	
2020-455	140	MISSION A		082 086	082 086		083 - 5.5	63	1	63	100	
2020-455	147	Mission120		082 086	082 086		084 - 6.0	16	16	16	100	
2020-455	148	Mission500		082 086	082 086		084 - 6.0	17	17	17	100	
064 URIN-PROTEIN QUALI.												
2020-451	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	50	50	50	100	
2020-451	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	37	36	36	100	
2020-451	013	BAYER CLI		051 051	051 051		052 - TRACES	37	1	36	0	
2020-451	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	50	50	100	
2020-451	071	ATLAS		051 051	051 051		051 - NEGATIVE	19	19	19	100	
2020-451	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	107	107	107	100	
2020-451	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	36	36	36	100	
2020-451	089	AIMSTRIP		051 051	051 051		054 - 2+	106	1	102	0	
2020-451	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	106	102	102	100	
2020-451	089	AIMSTRIP		051 051	051 051		052 - TRACES	106	3	102	0	
2020-451	094	CLINITEK		051 051	051 051		051 - NEGATIVE	109	109	109	100	
2020-451	102	ACON LAB		051 051	051 051		051 - NEGATIVE	173	165	165	100	
2020-451	102	ACON LAB		051 051	051 051		052 - TRACES	173	8	165	0	
2020-451	117	URS 10 A		051 051	051 051		051 - NEGATIVE	10	10	10	100	
2020-451	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	27	26	26	100	
2020-451	131	iCHEM VELO		051 051	051 051		052 - TRACES	27	1	26	0	
2020-451	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100	
2020-451	138	URS-10T		051 051	051 051		051 - NEGATIVE	12	12	12	100	
2020-451	140	MISSION A		051 051	051 051		051 - NEGATIVE	63	62	62	100	
2020-451	140	MISSION A		051 051	051 051		052 - TRACES	63	1	62	0	
2020-451	147	Mission120		051 051	051 051		051 - NEGATIVE	16	16	16	100	
2020-451	148	Mission500			051 051		051 - NEGATIVE	17	13	13	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-451	148	Mission500			051	051	052 - TRACES	17	4	13	0	
064 URIN-PROTEIN QUALI.												
2020-452	011	AMES NM10S		053	055	053	055	052 - TRACES	50	1	49	0
2020-452	011	AMES NM10S		053	055	053	055	055 - 3+	50	18	49	100
2020-452	011	AMES NM10S		053	055	053	055	053 - 1+	50	1	49	100
2020-452	011	AMES NM10S		053	055	053	055	054 - 2+	50	30	49	100
2020-452	013	BAYER CLI		053	055	053	055	054 - 2+	37	21	36	100
2020-452	013	BAYER CLI		053	055	053	055	056 - 4+	37	1	36	0
2020-452	013	BAYER CLI		053	055	053	055	055 - 3+	37	14	36	100
2020-452	013	BAYER CLI		053	055	053	055	053 - 1+	37	1	36	100
2020-452	067	URISCAN		053	055	053	055	055 - 3+	50	5	50	100
2020-452	067	URISCAN		053	055	053	055	054 - 2+	50	42	50	100
2020-452	067	URISCAN		053	055	053	055	053 - 1+	50	3	50	100
2020-452	071	ATLAS		053	055	053	055	055 - 3+	19	1	19	100
2020-452	071	ATLAS		053	055	053	055	053 - 1+	19	1	19	100
2020-452	071	ATLAS		053	055	053	055	054 - 2+	19	17	19	100
2020-452	072	URS-9-TECO		054	056	053	055	056 - 4+	107	1	107	100
2020-452	072	URS-9-TECO		054	056	053	055	055 - 3+	107	54	107	100
2020-452	072	URS-9-TECO		054	056	053	055	054 - 2+	107	50	107	100
2020-452	072	URS-9-TECO		054	056	053	055	053 - 1+	107	2	107	100
2020-452	080	ROCHE (CHE		053	055	053	055	053 - 1+	36	1	36	100
2020-452	080	ROCHE (CHE		053	055	053	055	054 - 2+	36	24	36	100
2020-452	080	ROCHE (CHE		053	055	053	055	055 - 3+	36	11	36	100
2020-452	089	AIMSTRIP		053	055	053	055	056 - 4+	106	3	102	0
2020-452	089	AIMSTRIP		053	055	053	055	053 - 1+	106	4	102	100
2020-452	089	AIMSTRIP		053	055	053	055	055 - 3+	106	36	102	100
2020-452	089	AIMSTRIP		053	055	053	055	051 - NEGATIVE	106	1	102	0
2020-452	089	AIMSTRIP		053	055	053	055	054 - 2+	106	62	102	100
2020-452	094	CLINITEK		053	055	053	055	054 - 2+	109	84	109	100
2020-452	094	CLINITEK		053	055	053	055	055 - 3+	109	20	109	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	094	CLINITEK		053 055	053 055		053 - 1+	109	5	109	100	
2020-452	102	ACON LAB		053 055	053 055		053 - 1+	173	2	172	100	
2020-452	102	ACON LAB		053 055	053 055		054 - 2+	173	94	172	100	
2020-452	102	ACON LAB		053 055	053 055		056 - 4+	173	1	172	0	
2020-452	102	ACON LAB		053 055	053 055		055 - 3+	173	76	172	100	
2020-452	117	URS 10 A		054 056	053 055		054 - 2+	10	5	10	100	
2020-452	117	URS 10 A		054 056	053 055		055 - 3+	10	5	10	100	
2020-452	131	iCHEM VELO		053 055	053 055		054 - 2+	27	24	27	100	
2020-452	131	iCHEM VELO		053 055	053 055		055 - 3+	27	3	27	100	
2020-452	135	Hybrido AU		053 055	053 055		054 - 2+	12	12	12	100	
2020-452	138	URS-10T		054 056	053 055		054 - 2+	12	3	12	100	
2020-452	138	URS-10T		054 056	053 055		055 - 3+	12	9	12	100	
2020-452	140	MISSION A		053 055	053 055		054 - 2+	63	46	63	100	
2020-452	140	MISSION A		053 055	053 055		053 - 1+	63	2	63	100	
2020-452	140	MISSION A		053 055	053 055		055 - 3+	63	15	63	100	
2020-452	147	Mission120		053 055	053 055		053 - 1+	16	3	16	100	
2020-452	147	Mission120		053 055	053 055		054 - 2+	16	12	16	100	
2020-452	147	Mission120		053 055	053 055		055 - 3+	16	1	16	100	
2020-452	148	Mission500		054 056	053 055		055 - 3+	17	12	17	100	
2020-452	148	Mission500		054 056	053 055		054 - 2+	17	4	17	100	
2020-452	148	Mission500		054 056	053 055		053 - 1+	17	1	17	100	
064 URIN-PROTEIN QUALI.												
2020-453	011	AMES NM10S		053 055	053 055		055 - 3+	50	14	50	100	
2020-453	011	AMES NM10S		053 055	053 055		054 - 2+	50	35	50	100	
2020-453	011	AMES NM10S		053 055	053 055		053 - 1+	50	1	50	100	
2020-453	013	BAYER CLI		053 055	053 055		055 - 3+	37	11	37	100	
2020-453	013	BAYER CLI		053 055	053 055		053 - 1+	37	1	37	100	
2020-453	013	BAYER CLI		053 055	053 055		054 - 2+	37	25	37	100	
2020-453	067	URISCAN		053 055	053 055		054 - 2+	50	45	50	100	
2020-453	067	URISCAN		053 055	053 055		053 - 1+	50	3	50	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-453	067	URISCAN		053 055	053 055		055 - 3+	50	2	50	100	
2020-453	071	ATLAS		053 055	053 055		054 - 2+	19	18	19	100	
2020-453	071	ATLAS		053 055	053 055		053 - 1+	19	1	19	100	
2020-453	072	URS-9-TECO		053 055	053 055		056 - 4+	107	2	105	0	
2020-453	072	URS-9-TECO		053 055	053 055		054 - 2+	107	65	105	100	
2020-453	072	URS-9-TECO		053 055	053 055		053 - 1+	107	5	105	100	
2020-453	072	URS-9-TECO		053 055	053 055		055 - 3+	107	35	105	100	
2020-453	080	ROCHE (CHE		053 055	053 055		054 - 2+	36	24	36	100	
2020-453	080	ROCHE (CHE		053 055	053 055		053 - 1+	36	1	36	100	
2020-453	080	ROCHE (CHE		053 055	053 055		055 - 3+	36	11	36	100	
2020-453	089	AIMSTRIP		053 055	053 055		056 - 4+	106	2	104	0	
2020-453	089	AIMSTRIP		053 055	053 055		054 - 2+	106	78	104	100	
2020-453	089	AIMSTRIP		053 055	053 055		055 - 3+	106	21	104	100	
2020-453	089	AIMSTRIP		053 055	053 055		053 - 1+	106	5	104	100	
2020-453	094	CLINITEK		053 055	053 055		054 - 2+	109	89	109	100	
2020-453	094	CLINITEK		053 055	053 055		053 - 1+	109	5	109	100	
2020-453	094	CLINITEK		053 055	053 055		055 - 3+	109	15	109	100	
2020-453	102	ACON LAB		053 055	053 055		055 - 3+	173	48	171	100	
2020-453	102	ACON LAB		053 055	053 055		051 - NEGATIVE	173	1	171	0	
2020-453	102	ACON LAB		053 055	053 055		054 - 2+	173	122	171	100	
2020-453	102	ACON LAB		053 055	053 055		053 - 1+	173	1	171	100	
2020-453	102	ACON LAB		053 055	053 055		056 - 4+	173	1	171	0	
2020-453	117	URS 10 A		053 055	053 055		055 - 3+	10	1	10	100	
2020-453	117	URS 10 A		053 055	053 055		054 - 2+	10	9	10	100	
2020-453	131	iCHEM VELO		053 055	053 055		054 - 2+	27	24	27	100	
2020-453	131	iCHEM VELO		053 055	053 055		055 - 3+	27	3	27	100	
2020-453	135	Hybrido AU		053 055	053 055		054 - 2+	12	12	12	100	
2020-453	138	URS-10T		054 056	053 055		055 - 3+	12	9	12	100	
2020-453	138	URS-10T		054 056	053 055		054 - 2+	12	2	12	100	
2020-453	138	URS-10T		054 056	053 055		053 - 1+	12	1	12	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-453	140	MISSION A		053 055	053 055		053 - 1+	63	3	62	100	
2020-453	140	MISSION A		053 055	053 055		054 - 2+	63	52	62	100	
2020-453	140	MISSION A		053 055	053 055		056 - 4+	63	1	62	0	
2020-453	140	MISSION A		053 055	053 055		055 - 3+	63	7	62	100	
2020-453	147	Mission120		053 055	053 055		054 - 2+	16	13	16	100	
2020-453	147	Mission120		053 055	053 055		053 - 1+	16	3	16	100	
2020-453	148	Mission500		053 055	053 055		055 - 3+	17	6	17	100	
2020-453	148	Mission500		053 055	053 055		053 - 1+	17	2	17	100	
2020-453	148	Mission500		053 055	053 055		054 - 2+	17	9	17	100	
064 URIN-PROTEIN QUALI.												
2020-454	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	50	50	50	100	
2020-454	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	37	37	37	100	
2020-454	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	50	50	100	
2020-454	071	ATLAS		051 051	051 051		051 - NEGATIVE	19	19	19	100	
2020-454	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	107	106	106	100	
2020-454	072	URS-9-TECO		051 051	051 051		052 - TRACES	107	1	106	0	
2020-454	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	36	36	36	100	
2020-454	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	106	105	105	100	
2020-454	089	AIMSTRIP		051 051	051 051		054 - 2+	106	1	105	0	
2020-454	094	CLINITEK		051 051	051 051		051 - NEGATIVE	109	109	109	100	
2020-454	102	ACON LAB		051 051	051 051		051 - NEGATIVE	173	173	173	100	
2020-454	117	URS 10 A		051 051	051 051		051 - NEGATIVE	10	10	10	100	
2020-454	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	27	27	27	100	
2020-454	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100	
2020-454	138	URS-10T		051 051	051 051		051 - NEGATIVE	12	11	11	100	
2020-454	138	URS-10T		051 051	051 051		053 - 1+	12	1	11	0	
2020-454	140	MISSION A		051 051	051 051		051 - NEGATIVE	63	63	63	100	
2020-454	147	Mission120		051 051	051 051		051 - NEGATIVE	16	16	16	100	
2020-454	148	Mission500		051 051	051 051		051 - NEGATIVE	17	17	17	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
064 URIN-PROTEIN QUALI.												
2020-455	011	AMES NM10S		053	055	053	055	054 - 2+	50	29	50	100
2020-455	011	AMES NM10S		053	055	053	055	053 - 1+	50	3	50	100
2020-455	011	AMES NM10S		053	055	053	055	055 - 3+	50	18	50	100
2020-455	013	BAYER CLI		053	055	053	055	054 - 2+	37	26	37	100
2020-455	013	BAYER CLI		053	055	053	055	053 - 1+	37	2	37	100
2020-455	013	BAYER CLI		053	055	053	055	055 - 3+	37	9	37	100
2020-455	067	URISCAN		053	055	053	055	053 - 1+	50	4	50	100
2020-455	067	URISCAN		053	055	053	055	054 - 2+	50	44	50	100
2020-455	067	URISCAN		053	055	053	055	055 - 3+	50	2	50	100
2020-455	071	ATLAS		053	055	053	055	055 - 3+	19	1	19	100
2020-455	071	ATLAS		053	055	053	055	053 - 1+	19	1	19	100
2020-455	071	ATLAS		053	055	053	055	054 - 2+	19	17	19	100
2020-455	072	URS-9-TECO		053	055	053	055	054 - 2+	107	62	106	100
2020-455	072	URS-9-TECO		053	055	053	055	056 - 4+	107	1	106	0
2020-455	072	URS-9-TECO		053	055	053	055	055 - 3+	107	36	106	100
2020-455	072	URS-9-TECO		053	055	053	055	053 - 1+	107	8	106	100
2020-455	080	ROCHE (CHE		053	055	053	055	053 - 1+	36	1	36	100
2020-455	080	ROCHE (CHE		053	055	053	055	055 - 3+	36	13	36	100
2020-455	080	ROCHE (CHE		053	055	053	055	054 - 2+	36	22	36	100
2020-455	089	AIMSTRIP		053	055	053	055	054 - 2+	106	81	104	100
2020-455	089	AIMSTRIP		053	055	053	055	053 - 1+	106	5	104	100
2020-455	089	AIMSTRIP		053	055	053	055	056 - 4+	106	1	104	0
2020-455	089	AIMSTRIP		053	055	053	055	051 - NEGATIVE	106	1	104	0
2020-455	089	AIMSTRIP		053	055	053	055	055 - 3+	106	18	104	100
2020-455	094	CLINITEK		053	055	053	055	053 - 1+	109	5	109	100
2020-455	094	CLINITEK		053	055	053	055	054 - 2+	109	87	109	100
2020-455	094	CLINITEK		053	055	053	055	055 - 3+	109	17	109	100
2020-455	102	ACON LAB		053	055	053	055	054 - 2+	173	130	171	100
2020-455	102	ACON LAB		053	055	053	055	052 - TRACES	173	1	171	0

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	102	ACON LAB		053 055	053 055		055 - 3+	173	40	171	100
2020-455	102	ACON LAB		053 055	053 055		051 - NEGATIVE	173	1	171	0
2020-455	102	ACON LAB		053 055	053 055		053 - 1+	173	1	171	100
2020-455	117	URS 10 A		053 055	053 055		054 - 2+	10	9	10	100
2020-455	117	URS 10 A		053 055	053 055		055 - 3+	10	1	10	100
2020-455	131	iCHEM VELO		053 055	053 055		055 - 3+	27	3	27	100
2020-455	131	iCHEM VELO		053 055	053 055		054 - 2+	27	24	27	100
2020-455	135	Hybrido AU		053 055	053 055		054 - 2+	12	12	12	100
2020-455	138	URS-10T		054 056	053 055		054 - 2+	12	5	12	100
2020-455	138	URS-10T		054 056	053 055		053 - 1+	12	1	12	100
2020-455	138	URS-10T		054 056	053 055		055 - 3+	12	6	12	100
2020-455	140	MISSION A		053 055	053 055		054 - 2+	63	53	62	100
2020-455	140	MISSION A		053 055	053 055		053 - 1+	63	2	62	100
2020-455	140	MISSION A		053 055	053 055		052 - TRACES	63	1	62	0
2020-455	140	MISSION A		053 055	053 055		055 - 3+	63	7	62	100
2020-455	147	Mission120		053 055	053 055		054 - 2+	16	13	16	100
2020-455	147	Mission120		053 055	053 055		053 - 1+	16	3	16	100
2020-455	148	Mission500		053 055	053 055		055 - 3+	17	4	17	100
2020-455	148	Mission500		053 055	053 055		054 - 2+	17	11	17	100
2020-455	148	Mission500		053 055	053 055		053 - 1+	17	2	17	100
066 URINAL.-UROBILINOGEN											
2020-451	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-451	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	37	37	37	100
2020-451	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-451	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	19	19	19	100
2020-451	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	107	107	107	100
2020-451	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	36	36	36	100
2020-451	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	106	106	106	100
2020-451	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	109	109	109	100
2020-451	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	172	172	172	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-451	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-451	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2020-451	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2020-451	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100	
2020-451	147	Mission120		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2020-451	148	Mission500		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
066 URINAL.-UROBILINOGEN												
2020-452	011	AMES NM10S		061 063	061 063		064 - 8	50	1	49	0	
2020-452	011	AMES NM10S		061 063	061 063		061 - NORMAL,NEG	50	5	49	100	
2020-452	011	AMES NM10S		061 063	061 063		063 - 4	50	18	49	100	
2020-452	011	AMES NM10S		061 063	061 063		062 - 2	50	26	49	100	
2020-452	013	BAYER CLI			061 063		062 - 2	37	14	37	100	
2020-452	013	BAYER CLI			061 063		063 - 4	37	8	37	100	
2020-452	013	BAYER CLI			061 063		061 - NORMAL,NEG	37	15	37	100	
2020-452	067	URISCAN		062 064	061 063		064 - 8	50	2	50	100	
2020-452	067	URISCAN		062 064	061 063		061 - NORMAL,NEG	50	9	50	100	
2020-452	067	URISCAN		062 064	061 063		062 - 2	50	10	50	100	
2020-452	067	URISCAN		062 064	061 063		063 - 4	50	29	50	100	
2020-452	071	ATLAS		061 063	061 063		063 - 4	19	8	19	100	
2020-452	071	ATLAS		061 063	061 063		062 - 2	19	11	19	100	
2020-452	072	URS-9-TECO		061 063	061 063		062 - 2	107	50	105	100	
2020-452	072	URS-9-TECO		061 063	061 063		063 - 4	107	35	105	100	
2020-452	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	107	20	105	100	
2020-452	072	URS-9-TECO		061 063	061 063		064 - 8	107	2	105	0	
2020-452	080	ROCHE (CHE		061 061	061 063		062 - 2	36	2	36	100	
2020-452	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	36	34	36	100	
2020-452	089	AIMSTRIP			061 063		061 - NORMAL,NEG	106	58	106	100	
2020-452	089	AIMSTRIP			061 063		063 - 4	106	12	106	100	
2020-452	089	AIMSTRIP			061 063		062 - 2	106	36	106	100	
2020-452	094	CLINITEK		061 063	061 063		063 - 4	109	31	109	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-452	094	CLINITEK		061 063	061 063		062 - 2	109	62	109	100	
2020-452	094	CLINITEK		061 063	061 063		061 - NORMAL,NEG	109	16	109	100	
2020-452	102	ACON LAB			061 063		061 - NORMAL,NEG	172	134	172	100	
2020-452	102	ACON LAB			061 063		063 - 4	172	6	172	100	
2020-452	102	ACON LAB			061 063		062 - 2	172	32	172	100	
2020-452	131	iCHEM VELO		062 064	061 063		061 - NORMAL,NEG	27	1	27	100	
2020-452	131	iCHEM VELO		062 064	061 063		062 - 2	27	1	27	100	
2020-452	131	iCHEM VELO		062 064	061 063		063 - 4	27	25	27	100	
2020-452	135	Hybrido AU		061 063	061 063		062 - 2	12	12	12	100	
2020-452	138	URS-10T		062 064	061 063		062 - 2	12	4	12	100	
2020-452	138	URS-10T		062 064	061 063		063 - 4	12	6	12	100	
2020-452	138	URS-10T		062 064	061 063		061 - NORMAL,NEG	12	2	12	100	
2020-452	140	MISSION A			061 063		062 - 2	63	13	63	100	
2020-452	140	MISSION A			061 063		061 - NORMAL,NEG	63	50	63	100	
2020-452	147	Mission120		061 061	061 063		061 - NORMAL,NEG	16	16	16	100	
2020-452	148	Mission500		061 061	061 063		063 - 4	17	1	17	100	
2020-452	148	Mission500		061 061	061 063		062 - 2	17	1	17	100	
2020-452	148	Mission500		061 061	061 063		061 - NORMAL,NEG	17	15	17	100	
066 URINAL.-UROBILINOGEN												
2020-453	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-453	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2020-453	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-453	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	19	19	19	100	
2020-453	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	107	107	107	100	
2020-453	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2020-453	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	106	106	106	100	
2020-453	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	109	109	109	100	
2020-453	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	172	172	172	100	
2020-453	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-453	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-453	138	URS-10T		061 061	061 061		062 - 2	12	1	11	0
2020-453	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	12	11	11	100
2020-453	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100
2020-453	147	Mission120		061 061	061 061		061 - NORMAL,NEG	16	16	16	100
2020-453	148	Mission500		061 061	061 061		061 - NORMAL,NEG	17	17	17	100
066 URINAL.-UROBILINOGEN											
2020-454	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-454	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	37	37	37	100
2020-454	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-454	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	19	19	19	100
2020-454	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	107	107	107	100
2020-454	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	36	36	36	100
2020-454	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	106	106	106	100
2020-454	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	109	109	109	100
2020-454	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	172	172	172	100
2020-454	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100
2020-454	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100
2020-454	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	12	12	12	100
2020-454	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100
2020-454	147	Mission120		061 061	061 061		061 - NORMAL,NEG	16	16	16	100
2020-454	148	Mission500		061 061	061 061		061 - NORMAL,NEG	17	17	17	100
2020-455	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-455	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	37	37	37	100
2020-455	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100
2020-455	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	19	19	19	100
2020-455	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	107	107	107	100
2020-455	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	36	36	36	100
2020-455	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	106	106	106	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-455	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	109	109	109	100	
2020-455	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	172	172	172	100	
2020-455	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-455	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2020-455	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2020-455	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	63	63	63	100	
2020-455	147	Mission120		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2020-455	148	Mission500		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
184 URINALYSIS LEUKOCYTE												
2020-451				001 001	001 001		001 - Negative	34	34	34	100	
2020-451	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100	
2020-451	013	BAYER CLI		001 001	001 001		001 - Negative	37	37	37	100	
2020-451	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100	
2020-451	071	ATLAS		001 001	001 001		001 - Negative	18	18	18	100	
2020-451	072	URS-9-TECO		001 001	001 001		001 - Negative	97	97	97	100	
2020-451	080	ROCHE (CHE		001 001	001 001		001 - Negative	32	32	32	100	
2020-451	089	AIMSTRIP		001 001	001 001		001 - Negative	100	100	100	100	
2020-451	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100	
2020-451	102	ACON LAB		001 001	001 001		001 - Negative	149	149	149	100	
2020-451	117	URS 10 A		001 001	001 001		001 - Negative	10	10	10	100	
2020-451	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100	
2020-451	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100	
2020-451	138	URS-10T		001 001	001 001		001 - Negative	12	12	12	100	
2020-451	140	MISSION A		001 001	001 001		001 - Negative	71	71	71	100	
2020-451	147	Mission120		001 001	001 001		001 - Negative	17	17	17	100	
2020-451	148	Mission500		001 001	001 001		001 - Negative	15	15	15	100	
2020-452				001 001	001 001		001 - Negative	34	34	34	100	
2020-452	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-452	013	BAYER CLI		001 001	001 001		001 - Negative	37	37	37	100
2020-452	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100
2020-452	071	ATLAS		001 001	001 001		001 - Negative	18	18	18	100
2020-452	072	URS-9-TECO		001 001	001 001		001 - Negative	97	97	97	100
2020-452	080	ROCHE (CHE		001 001	001 001		001 - Negative	32	32	32	100
2020-452	089	AIMSTRIP		001 001	001 001		005 - Large	100	1	99	0
2020-452	089	AIMSTRIP		001 001	001 001		001 - Negative	100	99	99	100
2020-452	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2020-452	102	ACON LAB		001 001	001 001		001 - Negative	149	149	149	100
2020-452	117	URS 10 A		001 001	001 001		001 - Negative	10	10	10	100
2020-452	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100
2020-452	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2020-452	138	URS-10T		001 001	001 001		001 - Negative	12	12	12	100
2020-452	140	MISSION A		001 001	001 001		001 - Negative	71	71	71	100
2020-452	147	Mission120		001 001	001 001		001 - Negative	17	17	17	100
2020-452	148	Mission500		001 001	001 001		001 - Negative	15	15	15	100
184 URINALYSIS LEUKOCYTE											
2020-453				001 001	001 001		001 - Negative	34	34	34	100
2020-453	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100
2020-453	013	BAYER CLI		001 001	001 001		001 - Negative	37	37	37	100
2020-453	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100
2020-453	071	ATLAS		001 001	001 001		001 - Negative	18	18	18	100
2020-453	072	URS-9-TECO		001 001	001 001		001 - Negative	97	97	97	100
2020-453	080	ROCHE (CHE		001 001	001 001		001 - Negative	32	32	32	100
2020-453	089	AIMSTRIP		001 001	001 001		001 - Negative	100	100	100	100
2020-453	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2020-453	102	ACON LAB		001 001	001 001		001 - Negative	149	148	148	100
2020-453	102	ACON LAB		001 001	001 001		002 - Traces	149	1	148	0
2020-453	117	URS 10 A		001 001	001 001		001 - Negative	10	10	10	100
2020-453	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-453	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2020-453	138	URS-10T		001 001	001 001		001 - Negative	12	12	12	100
2020-453	140	MISSION A		001 001	001 001		001 - Negative	71	71	71	100
2020-453	147	Mission120		001 001	001 001		001 - Negative	17	17	17	100
2020-453	148	Mission500		001 001	001 001		001 - Negative	15	15	15	100
184 URINALYSIS LEUKOCYTE											
2020-454				003 005	003 005		003 - Small	34	13	34	100
2020-454				003 005	003 005		005 - Large	34	4	34	100
2020-454				003 005	003 005		004 - Moderate	34	17	34	100
2020-454	011	AMES NM10S		002 004	003 005		003 - Small	44	32	44	100
2020-454	011	AMES NM10S		002 004	003 005		002 - Traces	44	1	44	100
2020-454	011	AMES NM10S		002 004	003 005		005 - Large	44	1	44	100
2020-454	011	AMES NM10S		002 004	003 005		004 - Moderate	44	10	44	100
2020-454	013	BAYER CLI		002 004	003 005		004 - Moderate	37	10	37	100
2020-454	013	BAYER CLI		002 004	003 005		003 - Small	37	25	37	100
2020-454	013	BAYER CLI		002 004	003 005		005 - Large	37	2	37	100
2020-454	067	URISCAN		004 005	003 005		004 - Moderate	49	18	49	100
2020-454	067	URISCAN		004 005	003 005		005 - Large	49	26	49	100
2020-454	067	URISCAN		004 005	003 005		003 - Small	49	5	49	100
2020-454	071	ATLAS		004 005	003 005		004 - Moderate	18	7	18	100
2020-454	071	ATLAS		004 005	003 005		005 - Large	18	11	18	100
2020-454	072	URS-9-TECO		003 005	003 005		005 - Large	97	3	96	100
2020-454	072	URS-9-TECO		003 005	003 005		002 - Traces	97	1	96	0
2020-454	072	URS-9-TECO		003 005	003 005		004 - Moderate	97	55	96	100
2020-454	072	URS-9-TECO		003 005	003 005		003 - Small	97	38	96	100
2020-454	080	ROCHE (CHE		004 005	003 005		004 - Moderate	32	4	31	100
2020-454	080	ROCHE (CHE		004 005	003 005		001 - Negative	32	1	31	0
2020-454	080	ROCHE (CHE		004 005	003 005		005 - Large	32	26	31	100
2020-454	080	ROCHE (CHE		004 005	003 005		003 - Small	32	1	31	100
2020-454	089	AIMSTRIP		002 004	003 005		003 - Small	100	48	99	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-454	089	AIMSTRIP		002 004	003 005		004 - Moderate	100	44	99	100	
2020-454	089	AIMSTRIP		002 004	003 005		001 - Negative	100	1	99	0	
2020-454	089	AIMSTRIP		002 004	003 005		005 - Large	100	4	99	100	
2020-454	089	AIMSTRIP		002 004	003 005		002 - Traces	100	3	99	100	
2020-454	094	CLINITEK		002 004	003 005		004 - Moderate	107	22	107	100	
2020-454	094	CLINITEK		002 004	003 005		003 - Small	107	85	107	100	
2020-454	102	ACON LAB		003 005	003 005		003 - Small	149	60	146	100	
2020-454	102	ACON LAB		003 005	003 005		004 - Moderate	149	83	146	100	
2020-454	102	ACON LAB		003 005	003 005		002 - Traces	149	3	146	0	
2020-454	102	ACON LAB		003 005	003 005		005 - Large	149	3	146	100	
2020-454	117	URS 10 A		002 004	003 005		003 - Small	10	6	10	100	
2020-454	117	URS 10 A		002 004	003 005		004 - Moderate	10	4	10	100	
2020-454	131	iCHEM VELO		004 005	003 005		005 - Large	26	23	26	100	
2020-454	131	iCHEM VELO		004 005	003 005		004 - Moderate	26	3	26	100	
2020-454	135	Hybrido AU		004 005	003 005		005 - Large	12	12	12	100	
2020-454	138	URS-10T		003 005	003 005		003 - Small	12	5	12	100	
2020-454	138	URS-10T		003 005	003 005		004 - Moderate	12	7	12	100	
2020-454	140	MISSION A		002 004	003 005		004 - Moderate	71	33	71	100	
2020-454	140	MISSION A		002 004	003 005		005 - Large	71	2	71	100	
2020-454	140	MISSION A		002 004	003 005		002 - Traces	71	2	71	100	
2020-454	140	MISSION A		002 004	003 005		003 - Small	71	34	71	100	
2020-454	147	Mission120		002 004	003 005		002 - Traces	17	2	17	100	
2020-454	147	Mission120		002 004	003 005		004 - Moderate	17	5	17	100	
2020-454	147	Mission120		002 004	003 005		003 - Small	17	10	17	100	
2020-454	148	Mission500		003 005	003 005		004 - Moderate	15	13	15	100	
2020-454	148	Mission500		003 005	003 005		003 - Small	15	2	15	100	
184 URINALYSIS LEUKOCYTE												
2020-455				001 001	001 001		003 - Small	34	1	33	0	
2020-455				001 001	001 001		001 - Negative	34	33	33	100	
2020-455	011	AMES NM10S		001 001	001 001		001 - Negative	44	44	44	100	

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-455	013	BAYER CLI		001 001	001 001		001 - Negative	37	37	37	100
2020-455	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100
2020-455	071	ATLAS		001 001	001 001		001 - Negative	18	18	18	100
2020-455	072	URS-9-TECO		001 001	001 001		001 - Negative	97	97	97	100
2020-455	080	ROCHE (CHE		001 001	001 001		001 - Negative	32	32	32	100
2020-455	089	AIMSTRIP		001 001	001 001		004 - Moderate	100	1	99	0
2020-455	089	AIMSTRIP		001 001	001 001		001 - Negative	100	99	99	100
2020-455	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2020-455	102	ACON LAB		001 001	001 001		001 - Negative	149	149	149	100
2020-455	117	URS 10 A		001 001	001 001		001 - Negative	10	10	10	100
2020-455	131	iCHEM VELO		001 001	001 001		001 - Negative	26	25	25	100
2020-455	131	iCHEM VELO		001 001	001 001		003 - Small	26	1	25	0
2020-455	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2020-455	138	URS-10T		001 001	001 001		001 - Negative	12	12	12	100
2020-455	140	MISSION A		001 001	001 001		001 - Negative	71	71	71	100
2020-455	147	Mission120		001 001	001 001		001 - Negative	17	17	17	100
2020-455	148	Mission500		001 001	001 001		001 - Negative	15	15	15	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2020-451	091	AIM STEP		001	001	001	001	001 - NEGATIVE	36	36	36	100
2020-451	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100
2020-452	091	AIM STEP		002	002	002	002	002 - POSITIVE	36	36	36	100
2020-452	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2020-453	091	AIM STEP		002	002	002	002	002 - POSITIVE	36	36	36	100
2020-453	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2020-454	091	AIM STEP		001	001	001	001	001 - NEGATIVE	36	36	36	100
2020-454	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100
2020-455	091	AIM STEP		001	001	001	001	001 - NEGATIVE	36	36	36	100
2020-455	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2020-461					023	023	018 - HIPPURIC	868	2	866	0
2020-461					023	023	023 - URIC ACID	868	866	866	100
2020-462					012	012	012 - TRANSITIO.	868	587	587	100
2020-462					012	012	042 - WBC in clu	868	1	587	0
2020-462					012	012	009 - LEUKOCYTE	868	2	587	0
2020-462					012	012	001 - E.C. CAST	868	1	587	0
2020-462					012	012	010 - RENAL TUB.	868	277	587	0
2020-463					042	042	042 - WBC in clu	869	729	729	100
2020-463					042	042	007 - WHITE CELL	869	3	729	0
2020-463					042	042	012 - TRANSITIO.	869	1	729	0
2020-463					042	042	009 - LEUKOCYTE	869	136	729	0
2020-464					006	006	030 - CYLINDROID	868	1	859	0
2020-464					006	006	006 - WAXY CAST	868	859	859	100
2020-464					006	006	016 - CHOLESTER.	868	2	859	0
2020-464					006	006	004 - HYALINE C.	868	6	859	0
2020-465					031	031	031 - FIBERS	868	868	868	100

Participant Summary

MONTH 6 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 98 OCCULT BLOOD							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
154 FECAL OCCULT BLOOD													
2020-471	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	452	451	451	100	
2020-471	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	452	1	451	0	
2020-471	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	689	688	688	100	
2020-471	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	689	1	688	0	
2020-472	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	452	451	451	100	
2020-472	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	452	1	451	0	
2020-472	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	689	687	687	100	
2020-472	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	689	2	687	0	
2020-473	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	452	452	452	100	
2020-473	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	689	687	687	100	
2020-473	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	689	2	687	0	
2020-474	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	452	452	452	100	
2020-474	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	689	687	687	100	
2020-474	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	689	2	687	0	
2020-475	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	452	452	452	100	
2020-475	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	689	688	688	100	
2020-475	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	689	1	688	0	
SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 90 SPECIAL TEST							

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-421	29 VITROS			11	14.427	0.486	3.37	13.000	15.900
2020-421	38 ROCHE E			21	14.286	0.786	5.50	11.900	16.600
2020-421	44 BECK.COUL.			18	16.035	0.662	4.13	14.000	18.000
2020-421	50 ARC			12	16.825	0.803	4.77	14.400	19.200
2020-422	29 VITROS			11	4.836	0.149	3.09	4.400	5.300
2020-422	38 ROCHE E			21	5.148	0.261	5.07	4.400	5.900
2020-422	44 BECK.COUL.			18	5.500	0.293	5.33	4.600	6.400
2020-422	50 ARC			12	5.825	0.209	3.58	5.200	6.500
2020-423	29 VITROS			11	6.382	0.225	3.52	5.700	7.100
2020-423	38 ROCHE E			21	6.638	0.363	5.47	5.500	7.700
2020-423	44 BECK.COUL.			18	7.059	0.309	4.37	6.100	8.000
2020-423	50 ARC			12	7.333	0.314	4.29	6.400	8.300
2020-424	29 VITROS			11	23.309	0.696	2.99	21.200	25.400
2020-424	38 ROCHE E			21	21.843	1.366	6.25	17.700	25.900
2020-424	44 BECK.COUL.			18	25.247	1.423	5.63	21.000	29.500
2020-424	50 ARC			12	25.464	0.856	3.36	22.900	28.000
2020-425	29 VITROS			11	0.773	0.062	7.98	0.600	1.000
2020-425	38 ROCHE E			21	1.086	0.108	9.96	0.800	1.400
2020-425	44 BECK.COUL.			18	1.111	0.105	9.43	0.800	1.400
2020-425	50 ARC			12	1.291	0.202	15.65	1.000	1.600
069	PROS. SPEC-ANTIG-PSA								
2020-421	29 VITROS			34	5.215	0.287	5.51	4.400	6.100
2020-421	36 TOSOH AIA			12	6.292	0.804	12.77	4.700	7.900
2020-421	38 ROCHE E			49	7.645	0.522	6.83	6.100	9.200
2020-421	39 DIMENS.RXL			57	7.004	0.436	6.23	5.700	8.300

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-421	44	BECK.COUL.		54	8.009	0.537	6.70	6.400	9.600
2020-421	50	ARC		17	5.788	0.353	6.10	4.700	6.800
2020-421	57	S. EXL 200		15	6.820	0.548	8.03	5.200	8.500
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-422	29	VITROS		34	2.021	0.111	5.47	1.700	2.400
2020-422	36	TOSOH AIA		12	2.217	0.316	14.25	1.700	2.800
2020-422	38	ROCHE E		49	2.682	0.176	6.55	2.200	3.200
2020-422	39	DIMENS.RXL		57	2.367	0.163	6.87	1.900	2.900
2020-422	44	BECK.COUL.		54	2.746	0.182	6.64	2.200	3.300
2020-422	50	ARC		17	1.924	0.148	7.68	1.500	2.400
2020-422	57	S. EXL 200		15	2.333	0.205	8.81	1.800	2.900
2020-423	29	VITROS		34	2.559	0.131	5.11	2.200	3.000
2020-423	36	TOSOH AIA		12	2.883	0.362	12.57	2.200	3.600
2020-423	38	ROCHE E		49	3.437	0.241	7.02	2.700	4.200
2020-423	39	DIMENS.RXL		57	3.067	0.197	6.42	2.500	3.700
2020-423	44	BECK.COUL.		54	3.563	0.231	6.49	2.900	4.300
2020-423	50	ARC		17	2.476	0.148	5.96	2.000	2.900
2020-423	57	S. EXL 200		15	3.020	0.264	8.74	2.300	3.800
2020-424	29	VITROS		34	7.868	0.486	6.17	6.400	9.300
2020-424	36	TOSOH AIA		12	9.275	1.359	14.65	7.000	11.600
2020-424	38	ROCHE E		49	11.398	0.799	7.01	9.000	13.800
2020-424	39	DIMENS.RXL		57	10.666	0.714	6.70	8.500	12.800
2020-424	44	BECK.COUL.		54	12.161	0.820	6.74	9.700	14.600
2020-424	50	ARC		17	8.776	0.576	6.56	7.000	10.500
2020-424	57	S. EXL 200		15	10.573	0.839	7.93	8.100	13.100
2020-425	29	VITROS		34	0.412	0.032	7.82	0.300	0.500
2020-425	36	TOSOH AIA		12	0.400	0.058	14.43	0.300	0.500

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-425	38	ROCHE E			49	0.455	0.050	10.93	0.300	0.600
2020-425	39	DIMENS.RXL			57	0.389	0.062	15.86	0.300	0.500
2020-425	44	BECK.COUL.			54	0.462	0.048	10.49	0.300	0.600
2020-425	50	ARC			17	0.300	0.000	0.00	0.300	0.300
2020-425	57	S. EXL 200			15	0.413	0.072	17.37	0.300	0.500
071 VITAMIN B 12			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2020-421	29	VITROS			21	604.857	54.175	8.96	454.000	756.000
2020-421	38	ROCHE E			23	583.591	32.674	5.60	486.000	682.000
2020-421	44	BECK.COUL.			29	427.857	23.904	5.59	356.000	500.000
2020-421	50	ARC			16	584.813	28.461	4.87	499.000	670.000
2020-421	57	S. EXL 200			12	618.250	38.958	6.30	501.000	735.000
2020-422	29	VITROS			21	346.286	17.707	5.11	293.000	399.000
2020-422	38	ROCHE E			23	304.348	54.039	17.76	228.000	380.000
2020-422	44	BECK.COUL.			29	243.345	18.640	7.66	187.000	299.000
2020-422	50	ARC			16	342.750	28.446	8.30	257.000	428.000
2020-422	57	S. EXL 200			12	422.333	34.822	8.25	318.000	527.000
2020-423	29	VITROS			21	381.667	20.499	5.37	320.000	443.000
2020-423	38	ROCHE E			23	334.087	59.062	17.68	251.000	418.000
2020-423	44	BECK.COUL.			29	269.414	19.049	7.07	212.000	327.000
2020-423	50	ARC			16	374.000	29.492	7.89	286.000	462.000
2020-423	57	S. EXL 200			12	450.833	24.956	5.54	376.000	526.000
2020-424	29	VITROS			21	763.238	92.153	12.07	572.000	954.000
2020-424	38	ROCHE E			23	828.609	48.916	5.90	682.000	975.000
2020-424	44	BECK.COUL.			29	593.964	38.377	6.46	479.000	709.000
2020-424	50	ARC			16	811.125	51.571	6.36	656.000	966.000
2020-424	57	S. EXL 200			12	800.167	39.012	4.88	683.000	917.000

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 90 SPECIAL TEST				
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
071 VITAMIN B 12		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}				
2020-425	29	VITROS		21	279.714	9.488	3.39	251.000	308.000	
2020-425	38	ROCHE E		23	210.478	46.769	22.22	158.000	263.000	
2020-425	44	BECK.COUL.		29	170.103	14.563	8.56	128.000	213.000	
2020-425	50	ARC		16	258.188	33.138	12.84	194.000	323.000	
2020-425	57	S. EXL 200		12	349.250	26.649	7.63	269.000	429.000	

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 95 GLYCOHEMOGLOBINE				
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
146 GLYCOHEMOGLOBIN										
2020-431	2	DCA-2000		12	7.650	0.194	2.53	7.100	8.200	
2020-431	12	COBAS/INT.		30	6.983	0.246	3.53	6.200	7.700	
2020-431	22	VITROS		36	7.136	0.160	2.24	6.700	7.600	
2020-431	30	DAYTONA		15	7.620	0.624	8.19	5.700	9.500	
2020-431	31	IMMUNOTURB		29	7.031	0.149	2.12	6.600	7.500	
2020-431	40	SIEMENS DI		85	7.384	0.287	3.88	6.500	8.200	
2020-431	40	SIEMENS DI		85	7.767	0.262	3.38	7.000	8.600	
2020-431	46	ARCHITECT		13	6.992	0.383	5.48	5.800	8.100	
2020-432	2	DCA-2000		12	6.392	0.222	3.47	5.700	7.100	
2020-432	12	COBAS/INT.		30	5.873	0.206	3.52	5.300	6.500	
2020-432	22	VITROS		36	5.880	0.098	1.67	5.600	6.200	
2020-432	30	DAYTONA		15	6.253	0.507	8.11	4.700	7.800	
2020-432	31	IMMUNOTURB		29	5.764	0.191	3.32	5.200	6.300	
2020-432	40	SIEMENS DI		85	6.033	0.125	2.07	5.700	6.400	
2020-432	40	SIEMENS DI		85	6.086	0.268	4.41	5.300	6.900	
2020-432	46	ARCHITECT		13	5.854	0.313	5.34	4.900	6.800	

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 95 GLYCOHEMOGLOBINE

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}				
2020-433	2	DCA-2000			12	6.383	0.177	2.78	5.900	6.900
2020-433	12	COBAS/INT.			30	5.887	0.205	3.47	5.300	6.500
2020-433	22	VITROS			36	5.867	0.118	2.01	5.500	6.200
2020-433	30	DAYTONA			15	6.227	0.427	6.85	4.900	7.500
2020-433	31	IMMUNOTURB			29	5.693	0.232	4.07	5.000	6.400
2020-433	40	SIEMENS DI			85	6.067	0.094	1.55	5.800	6.300
2020-433	40	SIEMENS DI			85	6.111	0.243	3.98	5.400	6.800
2020-433	46	ARCHITECT			13	5.831	0.297	5.10	4.900	6.700
2020-434	2	DCA-2000			12	11.042	0.390	3.54	9.900	12.200
2020-434	12	COBAS/INT.			30	10.643	0.337	3.17	9.600	11.700
2020-434	22	VITROS			36	10.186	0.231	2.26	9.500	10.900
2020-434	30	DAYTONA			15	11.793	0.894	7.58	9.100	14.500
2020-434	31	IMMUNOTURB			29	10.546	0.211	2.00	9.900	11.200
2020-434	40	SIEMENS DI			85	10.445	0.260	2.49	9.700	11.200
2020-434	40	SIEMENS DI			85	10.733	0.189	1.76	10.200	11.300
2020-434	46	ARCHITECT			13	10.885	0.704	6.46	8.800	13.000
2020-435	2	DCA-2000			12	7.458	0.263	3.52	6.700	8.200
2020-435	12	COBAS/INT.			30	7.030	0.213	3.03	6.400	7.700
2020-435	22	VITROS			36	7.164	0.169	2.35	6.700	7.700
2020-435	30	DAYTONA			15	7.940	1.000	12.59	6.000	9.900
2020-435	31	IMMUNOTURB			29	7.038	0.137	1.95	6.600	7.500
2020-435	40	SIEMENS DI			85	7.600	0.294	3.87	6.700	8.500
2020-435	40	SIEMENS DI			85	7.418	0.266	3.59	6.600	8.200
2020-435	46	ARCHITECT			13	6.992	0.405	5.79	5.800	8.200

Participant Summary

MONTH 6 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	20 %						
2020-281	3	VITROS		31	42.367	4.483	10.58	34.000	51.000	
2020-281	5	ROCHE E		30	33.750	2.641	7.82	27.000	41.000	
2020-281	9	BECK.COUL.		45	48.233	4.269	8.85	39.000	58.000	
2020-281	11	ARC		13	46.000	2.320	5.04	37.000	55.000	
2020-281	25	S. DIME		15	33.733	2.909	8.62	27.000	40.000	
2020-282	3	VITROS		31	32.667	3.048	9.33	26.000	39.000	
2020-282	5	ROCHE E		30	24.500	2.113	8.62	20.000	29.000	
2020-282	9	BECK.COUL.		45	34.488	3.553	10.30	28.000	41.000	
2020-282	11	ARC		13	32.083	1.256	3.91	26.000	39.000	
2020-282	25	S. DIME		15	25.333	1.700	6.71	20.000	30.000	
2020-283	3	VITROS		31	25.067	3.010	12.01	20.000	30.000	
2020-283	5	ROCHE E		30	22.655	3.187	14.07	18.000	27.000	
2020-283	9	BECK.COUL.		45	29.930	3.187	10.65	24.000	36.000	
2020-283	11	ARC		13	27.750	1.090	3.93	22.000	33.000	
2020-283	25	S. DIME		15	22.400	1.451	6.48	18.000	27.000	
2020-284	3	VITROS		31	79.900	6.513	8.15	64.000	96.000	
2020-284	5	ROCHE E		30	53.310	4.713	8.84	43.000	64.000	
2020-284	9	BECK.COUL.		45	80.000	6.731	8.41	64.000	96.000	
2020-284	11	ARC		13	74.846	3.592	4.80	60.000	90.000	
2020-284	25	S. DIME		15	55.333	2.981	5.39	44.000	66.000	
2020-285	3	VITROS		31	60.367	4.159	6.89	48.000	72.000	
2020-285	5	ROCHE E		30	44.552	5.424	12.17	36.000	53.000	
2020-285	9	BECK.COUL.		45	64.884	5.864	9.04	52.000	78.000	
2020-285	11	ARC		13	59.538	2.951	4.96	48.000	71.000	
2020-285	25	S. DIME		15	45.467	2.802	6.16	36.000	55.000	

8/20/2020

PROFICIENCY TESTING SERVICE
LABORATORY SERVICES PROGRAM
DEPARTMENT OF HEALTH OF PUERTO RICO

PAGE : 127 (127)
DATE : 8/20/2020
TIME : 08:19:41

Participant Summary

MONTH 6 OF YEAR 2020