

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
080 FIBRINOGEN				CLIA TEST ID #: 825						EVALUATION CRITERIA: TARGET VALUE 20 %	
2023-551	73 ACL 8 TO E	87 HEMOSIL-C		10	169.300	9.144	5.40	135.000	203.000		
2023-551	79 CS-2500	67 DADE		15	140.867	8.484	6.02	113.000	169.000		
2023-552	73 ACL 8 TO E	87 HEMOSIL-C		10	327.700	20.259	6.18	262.000	393.000		
2023-552	79 CS-2500	67 DADE		15	264.533	12.774	4.83	212.000	317.000		
2023-553	73 ACL 8 TO E	87 HEMOSIL-C		10	277.800	14.063	5.06	222.000	333.000		
2023-553	79 CS-2500	67 DADE		15	232.000	12.628	5.44	186.000	278.000		
2023-554	73 ACL 8 TO E	87 HEMOSIL-C		10	331.900	21.206	6.39	266.000	398.000		
2023-554	79 CS-2500	67 DADE		15	259.733	12.080	4.65	208.000	312.000		
2023-555	73 ACL 8 TO E	87 HEMOSIL-C		10	649.700	57.857	8.91	520.000	780.000		
2023-555	79 CS-2500	67 DADE		15	471.000	27.164	5.77	377.000	565.000		
081 PART.THROM. TIME-PTT				CLIA TEST ID #: 835						EVALUATION CRITERIA: TARGET VALUE 15 %	
2023-551	62 DADE SYS.	35 DADE ACTIN		57	30.237	1.161	3.84	25.700	34.800		
2023-551	65 DADE BFTII	35 DADE ACTIN		11	32.264	2.900	8.99	27.400	37.100		
2023-551	69 DIAG.STAGO	49 STA-PTT		17	36.335	1.490	4.10	30.900	41.800		
2023-551	73 ACL 8 TO E	56 PTT-HEMO		35	30.600	0.597	1.95	26.000	35.200		
2023-551	76 STAGO STAR	49 STA-PTT		23	35.430	1.903	5.37	30.100	40.700		
2023-551	79 CS-2500	35 DADE ACTIN		15	31.913	1.333	4.18	27.100	36.700		
2023-552	62 DADE SYS.	35 DADE ACTIN		57	45.686	1.895	4.15	38.800	52.500		
2023-552	65 DADE BFTII	35 DADE ACTIN		11	49.036	5.457	11.13	41.700	56.400		
2023-552	69 DIAG.STAGO	49 STA-PTT		17	51.456	1.520	2.95	43.700	59.200		
2023-552	73 ACL 8 TO E	56 PTT-HEMO		35	48.171	0.809	1.68	40.900	55.400		
2023-552	76 STAGO STAR	49 STA-PTT		23	50.774	3.539	6.97	43.200	58.400		
2023-552	79 CS-2500	35 DADE ACTIN		15	47.833	1.854	3.88	40.700	55.000		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS			LABS	VALUE	DEVIATION	CV	LOW	HIGH
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835									
			EVALUATION CRITERIA: TARGET VALUE	15	%						
2023-553	62 DADE SYS.	35 DADE ACTIN		57	26.160	0.683	2.61	22.200	30.100		
2023-553	65 DADE BFTII	35 DADE ACTIN		11	26.127	2.124	8.13	22.200	30.000		
2023-553	69 DIAG.STAGO	49 STA-PTT		17	29.106	0.667	2.29	24.700	33.500		
2023-553	73 ACL 8 TO E	56 PTT-HEMO		35	29.403	0.740	2.52	25.000	33.800		
2023-553	76 STAGO STAR	49 STA-PTT		23	28.861	1.189	4.12	24.500	33.200		
2023-553	79 CS-2500	35 DADE ACTIN		15	27.140	0.430	1.59	23.100	31.200		
2023-554	62 DADE SYS.	35 DADE ACTIN		57	36.870	1.337	3.63	31.300	42.400		
2023-554	65 DADE BFTII	35 DADE ACTIN		11	38.918	3.471	8.92	33.100	44.800		
2023-554	69 DIAG.STAGO	49 STA-PTT		17	41.588	1.301	3.13	35.400	47.800		
2023-554	73 ACL 8 TO E	56 PTT-HEMO		35	41.109	0.929	2.26	34.900	47.300		
2023-554	76 STAGO STAR	49 STA-PTT		23	40.670	2.474	6.08	34.600	46.800		
2023-554	79 CS-2500	35 DADE ACTIN		15	38.171	0.699	1.83	32.400	43.900		
2023-555	62 DADE SYS.	35 DADE ACTIN		57	26.091	0.586	2.24	22.200	30.000		
2023-555	65 DADE BFTII	35 DADE ACTIN		11	25.782	2.279	8.84	21.900	29.600		
2023-555	69 DIAG.STAGO	49 STA-PTT		17	28.518	1.047	3.67	24.200	32.800		
2023-555	73 ACL 8 TO E	56 PTT-HEMO		35	38.865	1.491	3.84	33.000	44.700		
2023-555	76 STAGO STAR	49 STA-PTT		23	28.165	1.550	5.50	23.900	32.400		
2023-555	79 CS-2500	35 DADE ACTIN		15	26.553	0.359	1.35	22.600	30.500		
082	PROTHROMBIN TIME(P.T)	CLIA TEST ID #: 845									
2023-551	62 DADE SYS.	33 DADE INNOV		58	12.726	0.449	3.53	10.800	14.600		
2023-551	65 DADE BFTII	33 DADE INNOV		12	12.342	0.514	4.16	10.500	14.200		
2023-551	73 ACL 8 TO E	41 HEMOSIL RE		30	14.193	0.383	2.70	12.100	16.300		
2023-551	76 STAGO STAR	44 STA NEO CI		15	15.440	0.647	4.19	13.100	17.800		
2023-551	79 CS-2500	33 DADE INNOV		16	13.063	0.162	1.24	11.100	15.000		
2023-552	62 DADE SYS.	33 DADE INNOV		58	26.282	1.274	4.85	22.300	30.200		
2023-552	65 DADE BFTII	33 DADE INNOV		12	28.700	1.493	5.20	24.400	33.000		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-552	73	ACL 8 TO E	41 HEMOSIL RE	30	33.003	1.386	4.20	28.100	38.000	
2023-552	76	STAGO STAR	44 STA NEO CI	15	34.120	1.877	5.50	29.000	39.200	
2023-552	79	CS-2500	33 DADE INNOV	16	25.969	0.611	2.35	22.100	29.900	
082 PROTHROMBIN TIME(PT)		CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %						
2023-553	62	DADE SYS.	33 DADE INNOV	58	10.255	0.287	2.80	8.700	11.800	
2023-553	65	DADE BFTII	33 DADE INNOV	12	9.742	0.661	6.79	8.300	11.200	
2023-553	73	ACL 8 TO E	41 HEMOSIL RE	30	10.750	0.191	1.78	9.100	12.400	
2023-553	76	STAGO STAR	44 STA NEO CI	15	12.473	0.347	2.78	10.600	14.300	
2023-553	79	CS-2500	33 DADE INNOV	16	10.719	0.113	1.05	9.100	12.300	
2023-554	62	DADE SYS.	33 DADE INNOV	58	17.424	0.756	4.34	14.800	20.000	
2023-554	65	DADE BFTII	33 DADE INNOV	12	18.633	1.097	5.89	15.800	21.400	
2023-554	73	ACL 8 TO E	41 HEMOSIL RE	30	20.637	0.718	3.48	17.500	23.700	
2023-554	76	STAGO STAR	44 STA NEO CI	15	21.420	0.798	3.73	18.200	24.600	
2023-554	79	CS-2500	33 DADE INNOV	16	17.463	0.306	1.75	14.800	20.100	
2023-555	62	DADE SYS.	33 DADE INNOV	58	10.156	0.370	3.64	8.600	11.700	
2023-555	65	DADE BFTII	33 DADE INNOV	12	9.708	0.596	6.14	8.300	11.200	
2023-555	73	ACL 8 TO E	41 HEMOSIL RE	30	10.657	0.316	2.96	9.100	12.300	
2023-555	76	STAGO STAR	44 STA NEO CI	15	12.493	0.446	3.57	10.600	14.400	
2023-555	79	CS-2500	33 DADE INNOV	16	10.525	0.214	2.03	8.900	12.100	

Participant Summary

MONTH 6 OF YEAR 2023

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2023-501					008 008		005 - BAND (STAB	903	2	899	0
2023-501					008 008		007 - BASOPHIL,	903	2	899	0
2023-501					008 008		008 - EOSINOPHIL	903	899	899	100
2023-502					020 020		017 - MONOCYTE	903	1	901	0
2023-502					020 020		019 - PLATELETS,	903	1	901	0
2023-502					020 020		020 - GIANT PLA	903	901	901	100
2023-503					011 011		011 - HYPERSEGME	904	896	896	100
2023-503					011 011		006 - SEGMENTED	904	8	896	0
2023-504					046 046		003 - MYELOCYTE	903	1	890	0
2023-504					046 046		033 - POLYCHROMA	903	2	890	0
2023-504					046 046		046 - ERYTHROCYT	903	890	890	100
2023-504					046 046		050 - OVALOCYTE	903	1	890	0
2023-504					046 046		052 - SPHEROCYTE	903	8	890	0
2023-504					046 046		067 - CABOT RING	903	1	890	0
2023-505					003 003		013 - LYMPHOCYTE	903	2	888	0
2023-505					003 003		001 - BLAST	903	5	888	0
2023-505					003 003		014 - LYMPHOCYTE	903	2	888	0
2023-505					003 003		002 - PROMYELOCY	903	1	888	0
2023-505					003 003		003 - MYELOCYTE	903	888	888	100
2023-505					003 003		004 - METAMYELOC	903	5	888	0

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{GREATER}			
2023-531	7	M.B.MANUAL		12	9.900	1.431	14.46	5.600	14.200
2023-531	9	M.B. RICCA		11	10.600	1.953	18.43	4.700	16.500
2023-531	24	Unicel DxH		19	8.000	0.614	7.67	6.000	10.000
2023-531	25	SYSMEX XN		31	6.513	1.020	15.67	3.500	9.600
2023-532	7	M.B.MANUAL		12	5.367	1.317	24.55	1.400	9.300
2023-532	9	M.B. RICCA		11	5.264	1.317	25.02	1.300	9.200
2023-532	24	Unicel DxH		19	3.647	0.344	9.43	2.600	4.700
2023-532	25	SYSMEX XN		31	2.833	0.432	15.23	1.500	4.100
2023-533	7	M.B.MANUAL		12	10.056	2.519	25.05	2.500	17.600
2023-533	9	M.B. RICCA		11	10.073	1.801	17.88	4.700	15.500
2023-533	24	Unicel DxH		19	7.911	0.671	8.48	5.900	9.900
2023-533	25	SYSMEX XN		31	6.642	1.024	15.41	3.600	9.700
2023-534	7	M.B.MANUAL		12	1.300	0.316	24.33	0.400	2.200
2023-534	9	M.B. RICCA		11	1.236	0.367	29.72	0.100	2.300
2023-534	24	Unicel DxH		19	0.632	0.086	13.64	0.400	0.900
2023-534	25	SYSMEX XN		31	0.960	0.088	9.16	0.700	1.200
2023-535	7	M.B.MANUAL		12	4.744	0.600	12.65	2.900	6.500
2023-535	9	M.B. RICCA		11	5.673	2.658	46.86	0	13.600
2023-535	24	Unicel DxH		19	3.668	0.355	9.69	2.600	4.700
2023-535	25	SYSMEX XN		31	2.865	0.431	15.04	1.600	4.200

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
153	SED. RATE (ESR)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 35 %	{GREATER}				
2023-541	4	SEDIPLAST		619	7.993	2.088	26.12	2.000	14.000	
2023-541	10	SEDI RATE		83	9.074	1.698	18.71	4.000	14.000	
2023-541	11	SEDIMAT-15		54	15.714	3.918	24.93	4.000	27.000	
2023-541	19	MINI CUBE		34	15.333	1.795	11.71	10.000	21.000	
2023-542	4	SEDIPLAST		619	49.158	8.773	17.85	23.000	75.000	
2023-542	10	SEDI RATE		83	49.073	9.032	18.41	22.000	76.000	
2023-542	11	SEDIMAT-15		54	137.214	6.951	5.07	89.000	185.000	
2023-542	19	MINI CUBE		34	59.000	9.978	16.91	29.000	89.000	
2023-543	4	SEDIPLAST		619	8.153	2.208	27.09	2.000	15.000	
2023-543	10	SEDI RATE		83	9.217	1.665	18.06	4.000	14.000	
2023-543	11	SEDIMAT-15		54	19.929	5.203	26.11	4.000	36.000	
2023-543	19	MINI CUBE		34	15.833	1.922	12.14	10.000	22.000	
2023-544	4	SEDIPLAST		619	8.028	2.136	26.61	2.000	14.000	
2023-544	10	SEDI RATE		83	9.476	1.508	15.91	5.000	14.000	
2023-544	11	SEDIMAT-15		54	16.429	5.778	35.17	0	34.000	
2023-544	19	MINI CUBE		34	16.000	1.563	9.77	10.000	22.000	
2023-545	4	SEDIPLAST		619	49.462	8.866	17.92	23.000	76.000	
2023-545	10	SEDI RATE		83	48.244	8.424	17.46	23.000	74.000	
2023-545	11	SEDIMAT-15		54	136.429	8.748	6.41	89.000	184.000	
2023-545	19	MINI CUBE		34	67.111	16.636	24.79	17.000	117.000	
SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 125 DIFF. FOR 5 PAR NEW						

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25 % OR 2.00	{GREATER}						
2023521x	1	SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	150	47.738	1.464	3.07	35.800	59.700	
2023521x	1	SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	150	30.544	1.699	5.56	22.900	38.200	
2023521x	1	SYSMEX CBC	3 MONOCYTE	25 XS SERIES	150	1.393	0.275	19.72	0	3.400	
2023521x	1	SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	150	12.743	0.482	3.78	9.600	15.900	
2023521x	1	SYSMEX CBC	5 BASOPHIL	25 XS SERIES	150	7.609	0.713	9.37	5.600	9.600	
2023521x	2	ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	123	61.505	1.780	2.89	46.100	76.900	
2023521x	2	ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	30	58.373	1.436	2.46	43.800	73.000	
2023521x	2	ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	123	32.928	1.720	5.22	24.700	41.200	
2023521x	2	ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	30	31.500	1.500	4.76	23.600	39.400	
2023521x	2	ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	122	1.794	0.520	28.98	0	3.800	
2023521x	2	ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	30	2.750	0.613	22.27	0.800	4.800	
2023521x	2	ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	121	3.761	0.617	16.41	1.800	5.800	
2023521x	2	ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	30	4.013	0.643	16.03	2.000	6.000	
2023521x	2	ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	121	69.038	1.901	2.75	51.800	86.300	
2023521x	2	ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	30	3.452	0.093	2.70	1.500	5.500	
2023521x	3	ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	17	60.788	0.916	1.51	45.600	76.000	
2023521x	3	ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	17	23.659	0.779	3.29	17.700	29.600	
2023521x	3	ADVIA 120	3 MONOCYTE	9 ADVIA-120	17	8.171	0.439	5.37	6.100	10.200	
2023521x	3	ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	17	2.282	0.285	12.50	0.300	4.300	
2023521x	3	ADVIA 120	5 BASOPHIL	9 ADVIA-120	17	0.665	0.242	36.47	0	2.700	
2023521x	3	ADVIA 120	6 LUC	9 ADVIA-120	16	4.738	0.414	8.73	2.700	6.700	
2023521x	4	AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	59	61.834	0.872	1.41	46.400	77.300	
2023521x	4	AB- CELL-D	1 NEUTROPHIL	37 ABB.CEL RU	51	60.422	1.063	1.76	45.300	75.500	
2023521x	4	AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	59	27.307	1.065	3.90	20.500	34.100	
2023521x	4	AB- CELL-D	2 LYMPHOCYTE	37 ABB.CEL RU	51	29.371	1.255	4.27	22.000	36.700	
2023521x	4	AB- CELL-D	3 MONOCYTE	8 CELL-DYN	59	5.981	0.638	10.67	4.000	8.000	
2023521x	4	AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	51	6.094	0.522	8.56	4.100	8.100	
2023521x	4	AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	59	3.879	0.222	5.72	1.900	5.900	
2023521x	4	AB- CELL-D	4 EOSINOPHIL	37 ABB.CEL RU	51	3.798	0.187	4.93	1.800	5.800	
2023521x	4	AB- CELL-D	5 BASOPHIL	8 CELL-DYN	60	0.900	0.341	37.84	0	2.900	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023521x	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	51	0.284	0.163	57.37	0	2.300
2023521x	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	60	61.807	1.226	1.98	46.400	77.300
2023521x	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	60	25.846	1.349	5.22	19.400	32.300
2023521x	5	BECTON D	3 MONOCYTE	29	B.C600,800	60	6.447	0.548	8.51	4.400	8.400
2023521x	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	60	5.864	0.347	5.91	3.900	7.900
2023521x	5	BECTON D	5 BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023521x	6	MR Mindray	1 NEUTROPHIL	32	BC-5390	52	55.575	0.693	1.25	41.700	69.500
2023521x	6	MR Mindray	2 LYMPHOCYTE	32	BC-5390	52	28.683	1.070	3.73	21.500	35.900
2023521x	6	MR Mindray	3 MONOCYTE	32	BC-5390	52	6.598	0.763	11.56	4.600	8.600
2023521x	6	MR Mindray	4 EOSINOPHIL	32	BC-5390	52	9.133	0.401	4.39	6.800	11.400
2023521x	6	MR Mindray	5 BASOPHIL	32	BC-5390	52	28.829	0.252	0.88	21.600	36.000
2023521x	7	DXH	1 NEUTROPHIL	36	BC DXH 500	58	71.693	1.807	2.52	53.800	89.600
2023521x	7	DXH	2 LYMPHOCYTE	36	BC DXH 500	58	21.921	1.424	6.49	16.400	27.400
2023521x	7	DXH	3 MONOCYTE	36	BC DXH 500	58	0.616	0.350	56.89	0	2.600
2023521x	7	DXH	4 EOSINOPHIL	36	BC DXH 500	58	5.344	0.506	9.48	3.300	7.300
2023521x	7	DXH	5 BASOPHIL	36	BC DXH 500	58	0.188	0.299	159.70	0	2.200
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}					
2023522x	1	SYSMEX CBC	1 NEUTROPHIL	25	XS SERIES	150	53.909	1.809	3.36	40.400	67.400
2023522x	1	SYSMEX CBC	2 LYMPHOCYTE	25	XS SERIES	150	22.410	1.662	7.42	16.800	28.000
2023522x	1	SYSMEX CBC	3 MONOCYTE	25	XS SERIES	150	0.912	0.170	18.62	0	2.900
2023522x	1	SYSMEX CBC	4 EOSINOPHIL	25	XS SERIES	150	14.138	0.657	4.65	10.600	17.700
2023522x	1	SYSMEX CBC	5 BASOPHIL	25	XS SERIES	150	8.818	0.770	8.73	6.600	11.000
2023522x	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	123	76.444	1.346	1.76	57.300	95.600
2023522x	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	30	72.027	1.856	2.58	54.000	90.000
2023522x	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	123	17.093	1.103	6.45	12.800	21.400
2023522x	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	30	16.250	1.081	6.65	12.200	20.300
2023522x	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	122	1.614	0.496	30.72	0	3.600
2023522x	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	30	2.640	0.685	25.94	0.600	4.600
2023522x	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	122	4.775	0.917	19.21	2.800	6.800
2023522x	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	30	4.957	1.026	20.69	3.000	7.000
2023522x	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	122	85.503	1.770	2.07	64.100	106.900

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023522x	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	30	4.272	0.087	2.03	2.300	6.300
2023522x	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	17	71.135	0.654	0.92	53.400	88.900
2023522x	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	17	15.529	0.325	2.09	11.600	19.400
2023522x	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	17	7.441	0.275	3.69	5.400	9.400
2023522x	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	17	2.371	0.325	13.71	0.400	4.400
2023522x	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	0.459	0.168	36.67	0	2.500
2023522x	3	ADVIA 120	6 LUC	9	ADVIA-120	16	3.319	0.270	8.13	1.300	5.300
2023522x	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	59	77.495	0.969	1.25	58.100	96.900
2023522x	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	51	76.300	0.973	1.28	57.200	95.400
2023522x	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	59	13.657	0.998	7.31	10.200	17.100
2023522x	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	51	14.575	1.134	7.78	10.900	18.200
2023522x	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	59	3.400	0.393	11.55	1.400	5.400
2023522x	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	51	3.641	0.382	10.49	1.600	5.600
2023522x	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	59	4.857	0.237	4.88	2.900	6.900
2023522x	4	AB- CELL-D	4 EOSINOPHIL	37	ABB.CEL RU	51	4.882	0.266	5.45	2.900	6.900
2023522x	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	60	0.697	0.299	42.90	0	2.700
2023522x	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	51	0.604	0.206	34.07	0	2.600
2023522x	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	61	68.342	1.134	1.66	51.300	85.400
2023522x	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	59	20.257	1.207	5.96	15.200	25.300
2023522x	5	BECTON D	3 MONOCYTE	29	B.C600,800	60	8.090	0.857	10.59	6.100	10.100
2023522x	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	60	3.241	0.252	7.77	1.200	5.200
2023522x	5	BECTON D	5 BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023522x	6	MR Mindray	1 NEUTROPHIL	32	BC-5390	52	61.335	0.620	1.01	46.000	76.700
2023522x	6	MR Mindray	2 LYMPHOCYTE	32	BC-5390	52	20.642	0.947	4.59	15.500	25.800
2023522x	6	MR Mindray	3 MONOCYTE	32	BC-5390	52	6.185	0.775	12.54	4.200	8.200
2023522x	6	MR Mindray	4 EOSINOPHIL	32	BC-5390	52	11.781	0.394	3.35	8.800	14.700
2023522x	6	MR Mindray	5 BASOPHIL	32	BC-5390	52	31.055	0.191	0.62	23.300	38.800
2023522x	7	DXH	1 NEUTROPHIL	36	BC DXH 500	58	79.865	0.937	1.17	59.900	99.800
2023522x	7	DXH	2 LYMPHOCYTE	36	BC DXH 500	58	10.825	0.976	9.02	8.100	13.500
2023522x	7	DXH	3 MONOCYTE	36	BC DXH 500	58	0.512	0.324	63.34	0	2.500
2023522x	7	DXH	4 EOSINOPHIL	36	BC DXH 500	58	8.644	0.559	6.47	6.500	10.800

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
2023522x	7 DXH	5 BASOPHIL	36 BC DXH 500	58	0.082	0.082	99.34	0	2.100				
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2023523x	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	150	54.382	1.866	3.43	40.800	68.000				
2023523x	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	150	21.781	2.119	9.73	16.300	27.200				
2023523x	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	150	0.853	0.174	20.44	0	2.900				
2023523x	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	150	14.230	0.661	4.64	10.700	17.800				
2023523x	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	151	8.886	0.793	8.93	6.700	11.100				
2023523x	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	123	76.230	1.640	2.15	57.200	95.300				
2023523x	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	30	71.921	1.236	1.72	53.900	89.900				
2023523x	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	123	17.157	1.351	7.87	12.900	21.400				
2023523x	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	30	16.093	0.893	5.55	12.100	20.100				
2023523x	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	122	1.611	0.489	30.36	0	3.600				
2023523x	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	30	2.663	0.643	24.13	0.700	4.700				
2023523x	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	122	4.856	0.950	19.57	2.900	6.900				
2023523x	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	30	4.914	0.792	16.12	2.900	6.900				
2023523x	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	122	84.983	2.357	2.77	63.700	106.200				
2023523x	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	29	4.254	0.082	1.93	2.300	6.300				
2023523x	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	17	71.000	0.562	0.79	53.300	88.800				
2023523x	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	17	15.576	0.478	3.07	11.700	19.500				
2023523x	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	17	7.400	0.340	4.59	5.400	9.400				
2023523x	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	17	2.465	0.270	10.95	0.500	4.500				
2023523x	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	17	0.447	0.142	31.74	0	2.400				
2023523x	3 ADVIA 120	6 LUC	9 ADVIA-120	16	3.281	0.357	10.89	1.300	5.300				
2023523x	4 AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	59	71.173	0.713	1.00	53.400	89.000				
2023523x	4 AB- CELL-D	1 NEUTROPHIL	37 ABB.CEL RU	51	69.908	1.047	1.50	52.400	87.400				
2023523x	4 AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	59	19.586	0.999	5.10	14.700	24.500				
2023523x	4 AB- CELL-D	2 LYMPHOCYTE	37 ABB.CEL RU	51	20.976	1.321	6.30	15.700	26.200				
2023523x	4 AB- CELL-D	3 MONOCYTE	8 CELL-DYN	59	4.068	0.550	13.52	2.100	6.100				
2023523x	4 AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	51	4.308	0.551	12.78	2.300	6.300				
2023523x	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	59	4.749	0.263	5.54	2.700	6.700				
2023523x	4 AB- CELL-D	4 EOSINOPHIL	37 ABB.CEL RU	51	4.690	0.191	4.08	2.700	6.700				

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023523x	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	60	0.456	0.225	49.34	0	2.500
2023523x	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	51	0.158	0.140	88.69	0	2.200
2023523x	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	66.656	1.079	1.62	50.000	83.300
2023523x	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	22.305	1.202	5.39	16.700	27.900
2023523x	5	BECTON D	3	MONOCYTE	29	B.C600,800	60	7.935	0.671	8.46	5.900	9.900
2023523x	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	60	3.083	0.239	7.77	1.100	5.100
2023523x	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023523x	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	51	61.434	0.777	1.26	46.100	76.800
2023523x	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	20.542	0.903	4.40	15.400	25.700
2023523x	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	6.250	0.781	12.50	4.300	8.300
2023523x	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	52	11.751	0.416	3.54	8.800	14.700
2023523x	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	31.108	0.179	0.58	23.300	38.900
2023523x	7	DXH	1	NEUTROPHIL	36	BC DXH 500	58	79.730	0.923	1.16	59.800	99.700
2023523x	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	58	11.004	1.015	9.22	8.300	13.800
2023523x	7	DXH	3	MONOCYTE	36	BC DXH 500	58	0.472	0.313	66.22	0	2.500
2023523x	7	DXH	4	EOSINOPHIL	36	BC DXH 500	58	8.661	0.427	4.93	6.500	10.800
2023523x	7	DXH	5	BASOPHIL	36	BC DXH 500	58	0.064	0.055	85.35	0	2.100
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2023524x	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	150	56.778	1.510	2.66	42.600	71.000
2023524x	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	151	18.531	1.508	8.14	13.900	23.200
2023524x	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	150	0.886	0.242	27.28	0	2.900
2023524x	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	150	14.875	0.491	3.30	11.200	18.600
2023524x	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	150	8.923	0.883	9.90	6.700	11.200
2023524x	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	123	59.478	2.843	4.78	44.600	74.300
2023524x	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	30	55.977	2.431	4.34	42.000	70.000
2023524x	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	123	30.462	2.079	6.83	22.800	38.100
2023524x	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	29	28.724	2.091	7.28	21.500	35.900
2023524x	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	122	3.334	1.191	35.72	1.300	5.300
2023524x	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	30	5.377	1.376	25.59	3.400	7.400
2023524x	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	122	6.508	1.115	17.13	4.500	8.500
2023524x	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	30	6.513	1.023	15.70	4.500	8.500

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023524x	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	122	66.541	1.829	2.75	49.900	83.200
2023524x	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	30	3.372	0.064	1.89	1.400	5.400
2023524x	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	17	42.159	1.034	2.45	31.600	52.700
2023524x	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	17	34.975	1.327	3.80	26.200	43.700
2023524x	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	17	14.782	0.704	4.76	11.100	18.500
2023524x	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	17	1.394	0.284	20.36	0	3.400
2023524x	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	0.594	0.170	28.56	0	2.600
2023524x	3	ADVIA 120	6 LUC	9	ADVIA-120	16	6.506	1.042	16.01	4.500	8.500
2023524x	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	59	50.531	0.895	1.77	37.900	63.200
2023524x	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	51	49.924	1.242	2.49	37.400	62.400
2023524x	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	59	38.561	1.303	3.38	28.900	48.200
2023524x	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	51	39.890	2.115	5.30	29.900	49.900
2023524x	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	59	6.495	0.876	13.49	4.500	8.500
2023524x	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	51	6.565	1.222	18.62	4.600	8.600
2023524x	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	59	3.457	0.335	9.70	1.500	5.500
2023524x	4	AB- CELL-D	4 EOSINOPHIL	37	ABB.CEL RU	51	3.406	0.357	10.47	1.400	5.400
2023524x	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	60	0.828	0.400	48.29	0	2.800
2023524x	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	51	0.057	0.120	209.17	0	2.100
2023524x	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	60	52.908	1.338	2.53	39.700	66.100
2023524x	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	60	33.490	1.855	5.54	25.100	41.900
2023524x	5	BECTON D	3 MONOCYTE	29	B.C600,800	60	7.122	0.720	10.11	5.100	9.100
2023524x	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	60	6.492	0.371	5.71	4.500	8.500
2023524x	5	BECTON D	5 BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023524x	6	MR Mindray	1 NEUTROPHIL	32	BC-5390	52	46.900	0.986	2.10	35.200	58.600
2023524x	6	MR Mindray	2 LYMPHOCYTE	32	BC-5390	52	39.855	1.746	4.38	29.900	49.800
2023524x	6	MR Mindray	3 MONOCYTE	32	BC-5390	52	6.471	1.081	16.71	4.500	8.500
2023524x	6	MR Mindray	4 EOSINOPHIL	32	BC-5390	51	6.678	1.255	18.79	4.700	8.700
2023524x	6	MR Mindray	5 BASOPHIL	32	BC-5390	52	27.867	0.539	1.93	20.900	34.800
2023524x	7	DXH	1 NEUTROPHIL	36	BC DXH 500	58	46.264	2.761	5.97	34.700	57.800
2023524x	7	DXH	2 LYMPHOCYTE	36	BC DXH 500	58	40.235	3.307	8.22	30.200	50.300
2023524x	7	DXH	3 MONOCYTE	36	BC DXH 500	58	0.331	0.431	130.14	0	2.300

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023524x	7 DXH	4 EOSINOPHIL	36 BC DXH 500	58	12.600	1.067	8.47	9.500	15.800		
2023524x	7 DXH	5 BASOPHIL	36 BC DXH 500	58	0.398	0.717	179.94	0	2.400		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765				EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}					
2023525x	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	150	48.136	1.879	3.90	36.100	60.200		
2023525x	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	150	30.028	2.263	7.54	22.500	37.500		
2023525x	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	152	1.420	0.321	22.61	0	3.400		
2023525x	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	149	12.843	0.554	4.31	9.600	16.100		
2023525x	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	150	7.581	0.701	9.24	5.600	9.600		
2023525x	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	123	61.657	1.821	2.95	46.200	77.100		
2023525x	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	30	58.767	1.794	3.05	44.100	73.500		
2023525x	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	123	32.638	1.882	5.76	24.500	40.800		
2023525x	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	30	31.047	1.754	5.65	23.300	38.800		
2023525x	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	122	1.744	0.519	29.76	0	3.700		
2023525x	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	29	2.786	0.642	23.02	0.800	4.800		
2023525x	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	123	3.846	0.660	17.16	1.800	5.800		
2023525x	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	30	4.083	0.621	15.22	2.100	6.100		
2023525x	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	121	68.887	2.167	3.15	51.700	86.100		
2023525x	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	30	3.448	0.081	2.36	1.400	5.400		
2023525x	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	17	60.747	0.629	1.04	45.600	75.900		
2023525x	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	17	23.859	0.879	3.68	17.900	29.800		
2023525x	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	17	8.176	0.422	5.16	6.100	10.200		
2023525x	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	17	2.247	0.243	10.81	0.200	4.200		
2023525x	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	17	0.524	0.183	34.99	0	2.500		
2023525x	3 ADVIA 120	6 LUC	9 ADVIA-120	16	4.750	0.477	10.04	2.800	6.800		
2023525x	4 AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	59	61.786	0.808	1.31	46.300	77.200		
2023525x	4 AB- CELL-D	1 NEUTROPHIL	37 ABB.CEL RU	51	60.162	0.943	1.57	45.100	75.200		
2023525x	4 AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	59	27.274	1.162	4.26	20.500	34.100		
2023525x	4 AB- CELL-D	2 LYMPHOCYTE	37 ABB.CEL RU	51	29.620	1.502	5.07	22.200	37.000		
2023525x	4 AB- CELL-D	3 MONOCYTE	8 CELL-DYN	59	6.021	0.572	9.51	4.000	8.000		
2023525x	4 AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	51	6.069	0.658	10.84	4.100	8.100		
2023525x	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	59	3.931	0.235	5.99	1.900	5.900		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 125 DIFF. FOR 5 PAR NEW								
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023525x	4	AB- CELL-D	4	EOSINOPHIL	37	ABB.CEL RU	51	3.716	0.231	6.22	1.700	5.700
2023525x	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	60	0.944	0.371	39.26	0	2.900
2023525x	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	51	0.308	0.207	67.12	0	2.300
2023525x	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	61.520	1.114	1.81	46.100	76.900
2023525x	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	25.959	1.224	4.72	19.500	32.400
2023525x	5	BECTON D	3	MONOCYTE	29	B.C600,800	61	6.564	0.647	9.86	4.600	8.600
2023525x	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	59	5.917	0.417	7.04	3.900	7.900
2023525x	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023525x	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	52	53.427	1.084	2.03	40.100	66.800
2023525x	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	31.329	1.591	5.08	23.500	39.200
2023525x	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	6.371	0.920	14.44	4.400	8.400
2023525x	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	52	8.892	0.477	5.37	6.700	11.100
2023525x	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	28.644	0.349	1.22	21.500	35.800
2023525x	7	DXH	1	NEUTROPHIL	36	BC DXH 500	58	72.266	1.821	2.52	54.200	90.300
2023525x	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	58	21.459	1.702	7.93	16.100	26.800
2023525x	7	DXH	3	MONOCYTE	36	BC DXH 500	58	0.579	0.378	65.28	0	2.600
2023525x	7	DXH	4	EOSINOPHIL	36	BC DXH 500	59	5.414	0.532	9.82	3.400	7.400
2023525x	7	DXH	5	BASOPHIL	36	BC DXH 500	57	0.175	0.298	170.73	0	2.200

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 127 DIFF 6 PARAM INSTRUM									
ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
185 Diff 6 Parameter Ins			CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE				25 % OR 2.00	{GREATER}				
2023521y	1	6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	47	50.315	1.236	2.46	37.700	62.900	
2023521y	1	6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	135	47.966	1.404	2.93	36.000	60.000	
2023521y	1	6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	48	29.427	0.905	3.08	22.100	36.800	
2023521y	1	6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	134	30.641	1.444	4.71	23.000	38.300	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
2023521y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	47	2.479	0.468	18.87	0.500	4.500				
2023521y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	135	1.292	0.301	23.30	0	3.300				
2023521y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	48	12.694	0.501	3.94	9.500	15.900				
2023521y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	134	12.613	0.608	4.82	9.500	15.800				
2023521y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	47	5.143	0.840	16.33	3.100	7.100				
2023521y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	135	7.542	0.623	8.26	5.500	9.500				
2023521y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	47	12.457	0.407	3.27	9.300	15.600				
2023521y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	135	12.782	0.696	5.45	9.600	16.000				
2023521y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.010	0.030	300.00	0	2.000				
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2023522y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	47	57.433	1.256	2.19	43.100	71.800				
2023522y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	135	54.346	1.204	2.21	40.800	67.900				
2023522y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	47	21.363	0.545	2.55	16.000	26.700				
2023522y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	135	22.263	1.055	4.74	16.700	27.800				
2023522y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	47	1.517	0.316	20.86	0	3.500				
2023522y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	135	0.777	0.320	41.18	0	2.800				
2023522y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	47	14.583	0.535	3.67	10.900	18.200				
2023522y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	135	14.486	0.625	4.31	10.900	18.100				
2023522y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	47	5.237	1.036	19.77	3.200	7.200				
2023522y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	135	8.302	0.690	8.31	6.200	10.400				
2023522y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	47	13.904	0.241	1.74	10.400	17.400				
2023522y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	135	14.533	0.677	4.66	10.900	18.200				
2023522y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.000	0.000	0.00	0	2.000				
2023523y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	48	57.310	1.532	2.67	43.000	71.600				
2023523y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	134	54.422	1.316	2.42	40.800	68.000				
2023523y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	47	21.135	0.504	2.39	15.900	26.400				
2023523y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	135	21.827	1.279	5.86	16.400	27.300				
2023523y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	47	1.464	0.303	20.73	0	3.500				
2023523y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	135	0.736	0.314	42.70	0	2.700				

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023523y	1 6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	47	14.772	0.512	3.47	11.100	18.500
2023523y	1 6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	135	14.669	0.578	3.94	11.000	18.300
2023523y	1 6 PAR REG	5	BASOPHIL	31	SYSMEX XN	47	5.059	0.660	13.05	3.100	7.100
2023523y	1 6 PAR REG	5	BASOPHIL	35	SYS XNL SR	135	8.467	0.675	7.97	6.400	10.600
2023523y	1 6 PAR REG	6	IMMATURE G	31	SYSMEX XN	47	13.989	0.381	2.73	10.500	17.500
2023523y	1 6 PAR REG	6	IMMATURE G	35	SYS XNL SR	135	14.505	0.639	4.41	10.900	18.100
2023523y	1 6 PAR REG	7	NRBC	31	SYSMEX XN	21	0.000	0.000	0.00	0	2.000
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2023524y	1 6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	47	59.117	1.121	1.90	44.300	73.900
2023524y	1 6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	135	56.002	1.222	2.18	42.000	70.000
2023524y	1 6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	47	19.023	0.789	4.15	14.300	23.800
2023524y	1 6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	135	19.819	1.196	6.03	14.900	24.800
2023524y	1 6 PAR REG	3	MONOCYTE	31	SYSMEX XN	47	1.660	0.438	26.39	0	3.700
2023524y	1 6 PAR REG	3	MONOCYTE	35	SYS XNL SR	136	0.873	0.484	55.44	0	2.900
2023524y	1 6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	47	15.006	0.584	3.89	11.300	18.800
2023524y	1 6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	134	14.730	0.718	4.87	11.000	18.400
2023524y	1 6 PAR REG	5	BASOPHIL	31	SYSMEX XN	47	5.193	1.029	19.82	3.200	7.200
2023524y	1 6 PAR REG	5	BASOPHIL	35	SYS XNL SR	135	8.587	0.760	8.85	6.400	10.700
2023524y	1 6 PAR REG	6	IMMATURE G	31	SYSMEX XN	47	14.213	0.380	2.68	10.700	17.800
2023524y	1 6 PAR REG	6	IMMATURE G	35	SYS XNL SR	135	14.878	0.656	4.41	11.200	18.600
2023524y	1 6 PAR REG	7	NRBC	31	SYSMEX XN	21	0.090	0.169	186.53	0	2.100
2023525y	1 6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	47	50.339	1.043	2.07	37.800	62.900
2023525y	1 6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	135	48.090	1.324	2.75	36.100	60.100
2023525y	1 6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	47	29.280	0.767	2.62	22.000	36.600
2023525y	1 6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	135	30.559	1.564	5.12	22.900	38.200
2023525y	1 6 PAR REG	3	MONOCYTE	31	SYSMEX XN	48	2.468	0.504	20.41	0.500	4.500
2023525y	1 6 PAR REG	3	MONOCYTE	35	SYS XNL SR	134	1.346	0.368	27.35	0	3.300
2023525y	1 6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	47	12.733	0.597	4.69	9.500	15.900
2023525y	1 6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	133	12.642	0.711	5.63	9.500	15.800

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023525y	1	6 PAR REG	5	BASOPHIL	31	SYSMEX XN	46	5.059	0.759	15.00	3.100	7.100
2023525y	1	6 PAR REG	5	BASOPHIL	35	SYS XNL SR	135	7.418	0.669	9.02	5.400	9.400
2023525y	1	6 PAR REG	6	IMMATURE G	31	SYSMEX XN	47	12.626	0.505	4.00	9.500	15.800
2023525y	1	6 PAR REG	6	IMMATURE G	35	SYS XNL SR	134	12.935	0.670	5.18	9.700	16.200
2023525y	1	6 PAR REG	7	NRBC	31	SYSMEX XN	21	0.000	0.000	0.00	0	2.000

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 128 DIFF 3 PARAMETER INS

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
186 DIFF 3 PARAMETER INS			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2023521z	8	3ST	1	Lymphocyte	1	KX 21 - Sy	20	30.925	1.174	3.80	23.200	38.700
2023521z	8	3ST	1	Lymphocyte	3	ABBOTT C-D	27	30.989	1.092	3.52	23.200	38.700
2023521z	8	3ST	2	Mix M/E/Ba	1	KX 21 - Sy	20	17.658	1.231	6.97	13.200	22.100
2023521z	8	3ST	2	Mix M/E/Ba	3	ABBOTT C-D	27	11.741	1.455	12.40	8.800	14.700
2023521z	8	3ST	3	Neutrophil	1	KX 21 - Sy	20	51.370	1.145	2.23	38.500	64.200
2023521z	8	3ST	3	Neutrophil	3	ABBOTT C-D	27	57.311	2.240	3.91	43.000	71.600
2023521z	9	3D	1	Lymphocyte	2	AcT Diff 2	16	31.994	0.936	2.92	24.000	40.000
2023521z	9	3D	2	Mix M/E/Ba	2	AcT Diff 2	16	6.388	0.703	11.00	4.400	8.400
2023521z	9	3D	3	Neutrophil	2	AcT Diff 2	16	61.619	1.100	1.78	46.200	77.000
2023522z	8	3ST	1	Lymphocyte	1	KX 21 - Sy	20	61.905	0.867	1.40	46.400	77.400
2023522z	8	3ST	1	Lymphocyte	3	ABBOTT C-D	27	59.959	1.340	2.23	45.000	74.900
2023522z	8	3ST	2	Mix M/E/Ba	1	KX 21 - Sy	20	13.025	0.977	7.50	9.800	16.300
2023522z	8	3ST	2	Mix M/E/Ba	3	ABBOTT C-D	27	10.067	0.599	5.95	7.600	12.600
2023522z	8	3ST	3	Neutrophil	1	KX 21 - Sy	20	25.070	0.654	2.61	18.800	31.300
2023522z	8	3ST	3	Neutrophil	3	ABBOTT C-D	27	30.000	1.400	4.67	22.500	37.500
2023522z	9	3D	1	Lymphocyte	2	AcT Diff 2	16	15.475	0.719	4.65	11.600	19.300
2023522z	9	3D	2	Mix M/E/Ba	2	AcT Diff 2	16	3.619	0.573	15.82	1.600	5.600
2023522z	9	3D	3	Neutrophil	2	AcT Diff 2	16	80.944	1.006	1.24	60.700	101.200

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 128 DIFF 3 PARAMETER INS									
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE				
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
186	DIFF 3 PARAMETER INS	CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25	% OR 2.00	{GREATER}							
2023523z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	20	61.720	0.946	1.53	46.300	77.200				
2023523z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	27	60.104	1.371	2.28	45.100	75.100				
2023523z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	20	13.290	0.775	5.83	10.000	16.600				
2023523z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	27	10.067	0.674	6.69	7.600	12.600				
2023523z	8 3ST	3 Neutrophil	1 KX 21 - Sy	20	24.960	0.718	2.88	18.700	31.200				
2023523z	8 3ST	3 Neutrophil	3 ABBOTT C-D	27	29.841	1.276	4.28	22.400	37.300				
2023523z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	15.338	0.741	4.83	11.500	19.200				
2023523z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	3.619	0.486	13.44	1.600	5.600				
2023523z	9 3D	3 Neutrophil	2 AcT Diff 2	16	81.113	0.960	1.18	60.800	101.400				
2023524z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	20	12.190	1.011	8.30	9.100	15.200				
2023524z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	27	14.381	1.794	12.47	10.800	18.000				
2023524z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	20	19.380	1.396	7.20	14.500	24.200				
2023524z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	27	12.663	1.393	11.00	9.500	15.800				
2023524z	8 3ST	3 Neutrophil	1 KX 21 - Sy	20	68.470	1.706	2.49	51.400	85.600				
2023524z	8 3ST	3 Neutrophil	3 ABBOTT C-D	27	73.022	2.627	3.60	54.800	91.300				
2023524z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	58.706	2.835	4.83	44.000	73.400				
2023524z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	8.263	0.988	11.96	6.200	10.300				
2023524z	9 3D	3 Neutrophil	2 AcT Diff 2	16	32.775	2.744	8.37	24.600	41.000				
2023525z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	20	31.005	1.011	3.26	23.300	38.800				
2023525z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	27	30.856	0.964	3.13	23.100	38.600				
2023525z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	20	17.495	0.719	4.11	13.100	21.900				
2023525z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	27	12.081	1.325	10.97	9.100	15.100				
2023525z	8 3ST	3 Neutrophil	1 KX 21 - Sy	20	51.540	1.115	2.16	38.700	64.400				
2023525z	8 3ST	3 Neutrophil	3 ABBOTT C-D	27	57.093	2.058	3.61	42.800	71.400				
2023525z	9 3D	1 Lymphocyte	2 AcT Diff 2	15	32.300	0.919	2.84	24.200	40.400				
2023525z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	6.456	0.628	9.73	4.500	8.500				
2023525z	9 3D	3 Neutrophil	2 AcT Diff 2	16	61.281	1.164	1.90	46.000	76.600				

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
549	RED CELL COUNT (RBC)	CLIA	TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %								
2023-521	1 Sysmex CBC	132	SYSMEX XS		151	4.634	0.058	1.24	4.360	4.910		
2023-521	1 Sysmex CBC	141	SYSMEX XN		47	4.625	0.062	1.34	4.350	4.900		
2023-521	1 Sysmex CBC	152	Sysmex XN		138	4.632	0.053	1.14	4.350	4.910		
2023-521	2 ACT PENTRA	125	COULTER AC		121	4.496	0.066	1.46	4.230	4.770		
2023-521	2 ACT PENTRA	130	ABX PENTRA		29	4.460	0.071	1.59	4.190	4.730		
2023-521	3 ADVIA 120	116	ADVIA-120		17	4.388	0.058	1.32	4.120	4.650		
2023-521	4 AB-CELL DY	115	ABBOTT C-D		58	4.548	0.082	1.80	4.280	4.820		
2023-521	4 AB-CELL DY	153	AB-CEL-RUB		52	4.694	0.091	1.94	4.410	4.980		
2023-521	5 BECTON D	137	B.C DXH		59	4.150	0.074	1.78	3.900	4.400		
2023-521	6 MR Mindray	145	BC-5390		52	4.498	0.063	1.41	4.230	4.770		
2023-521	7 DXH500 Ser	149	BC DXH 500		58	4.766	0.102	2.15	4.480	5.050		
2023-521	8 3 ST	84	ABBOTT C-D		27	4.131	0.082	1.98	3.880	4.380		
2023-521	8 3 ST	138	SY KX-21N		15	4.101	0.048	1.17	3.850	4.350		
2023-521	9 3 D	33	COULTER JT		10	4.619	0.072	1.56	4.340	4.900		
2023-521	9 3 D	120	COULTER		16	4.594	0.092	2.01	4.320	4.870		
2023-522	1 Sysmex CBC	132	SYSMEX XS		151	5.763	0.061	1.07	5.420	6.110		
2023-522	1 Sysmex CBC	141	SYSMEX XN		47	5.790	0.066	1.14	5.440	6.140		
2023-522	1 Sysmex CBC	152	Sysmex XN		138	5.810	0.062	1.07	5.460	6.160		
2023-522	2 ACT PENTRA	125	COULTER AC		121	4.985	0.069	1.39	4.690	5.280		
2023-522	2 ACT PENTRA	130	ABX PENTRA		29	4.952	0.052	1.06	4.660	5.250		
2023-522	3 ADVIA 120	116	ADVIA-120		17	5.234	0.054	1.02	4.920	5.550		
2023-522	4 AB-CELL DY	115	ABBOTT C-D		58	4.994	0.089	1.79	4.690	5.290		
2023-522	4 AB-CELL DY	153	AB-CEL-RUB		52	5.171	0.101	1.96	4.860	5.480		
2023-522	5 BECTON D	137	B.C DXH		59	5.487	0.080	1.45	5.160	5.820		
2023-522	6 MR Mindray	145	BC-5390		52	5.143	0.085	1.66	4.830	5.450		
2023-522	7 DXH500 Ser	149	BC DXH 500		58	5.232	0.073	1.40	4.920	5.550		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-522	8 3 ST	84 ABBOTT C-D		27	5.805	0.143	2.46	5.460	6.150
2023-522	8 3 ST	138 SY KX-21N		15	5.826	0.088	1.51	5.480	6.180
2023-522	9 3 D	33 COULTER JT		10	5.609	0.059	1.04	5.270	5.950
2023-522	9 3 D	120 COULTER		16	5.559	0.126	2.27	5.230	5.890
549 RED CELL COUNT (RBC)		CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %						
2023-523	1 Sysmex CBC	132 SYSMEX XS		151	6.451	0.078	1.21	6.060	6.840
2023-523	1 Sysmex CBC	141 SYSMEX XN		47	6.477	0.063	0.98	6.090	6.870
2023-523	1 Sysmex CBC	152 Sysmex XN		138	6.515	0.072	1.11	6.120	6.910
2023-523	2 ACT PENTRA	125 COULTER AC		121	4.977	0.070	1.41	4.680	5.280
2023-523	2 ACT PENTRA	130 ABX PENTRA		29	4.947	0.052	1.05	4.650	5.240
2023-523	3 ADVIA 120	116 ADVIA-120		17	5.229	0.100	1.91	4.920	5.540
2023-523	4 AB-CELL DY	115 ABBOTT C-D		58	6.085	0.098	1.61	5.720	6.450
2023-523	4 AB-CELL DY	153 AB-CEL-RUB		52	6.321	0.141	2.23	5.940	6.700
2023-523	5 BECTON D	137 B.C DXH		59	6.192	0.078	1.26	5.820	6.560
2023-523	6 MR Mindray	145 BC-5390		52	5.135	0.080	1.55	4.830	5.440
2023-523	7 DXH500 Ser	149 BC DXH 500		58	5.260	0.088	1.67	4.940	5.580
2023-523	8 3 ST	84 ABBOTT C-D		27	5.780	0.140	2.43	5.430	6.130
2023-523	8 3 ST	138 SY KX-21N		15	5.823	0.067	1.16	5.470	6.170
2023-523	9 3 D	33 COULTER JT		10	5.614	0.072	1.28	5.280	5.950
2023-523	9 3 D	120 COULTER		16	5.621	0.134	2.39	5.280	5.960
2023-524	1 Sysmex CBC	132 SYSMEX XS		151	2.304	0.034	1.46	2.170	2.440
2023-524	1 Sysmex CBC	141 SYSMEX XN		47	2.290	0.035	1.52	2.150	2.430
2023-524	1 Sysmex CBC	152 Sysmex XN		138	2.268	0.030	1.32	2.130	2.400
2023-524	2 ACT PENTRA	125 COULTER AC		121	2.327	0.047	2.01	2.190	2.470
2023-524	2 ACT PENTRA	130 ABX PENTRA		29	2.317	0.040	1.74	2.180	2.460
2023-524	3 ADVIA 120	116 ADVIA-120		17	2.168	0.031	1.45	2.040	2.300
2023-524	4 AB-CELL DY	115 ABBOTT C-D		58	2.176	0.053	2.42	2.050	2.310
2023-524	4 AB-CELL DY	153 AB-CEL-RUB		52	2.220	0.044	1.97	2.090	2.350
2023-524	5 BECTON D	137 B.C DXH		59	2.485	0.045	1.83	2.340	2.630

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-524	6 MR Mindray	145 BC-5390			52	2.210	0.034	1.56	2.080	2.340	
2023-524	7 DXH500 Ser	149 BC DXH 500			58	2.464	0.039	1.58	2.320	2.610	
2023-524	8 3 ST	84 ABBOTT C-D			27	2.520	0.059	2.33	2.370	2.670	
2023-524	8 3 ST	138 SY KX-21N			15	2.526	0.038	1.50	2.370	2.680	
2023-524	9 3 D	33 COULTER JT			10	2.352	0.042	1.80	2.210	2.490	
2023-524	9 3 D	120 COULTER			16	2.363	0.044	1.85	2.220	2.500	
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %								
2023-525	1 Sysmex CBC	132 SYSMEX XS			151	4.626	0.055	1.20	4.350	4.900	
2023-525	1 Sysmex CBC	141 SYSMEX XN			47	4.636	0.052	1.13	4.360	4.910	
2023-525	1 Sysmex CBC	152 Sysmex XN			138	4.633	0.052	1.13	4.350	4.910	
2023-525	2 ACT PENTRA	125 COULTER AC			121	4.481	0.074	1.65	4.210	4.750	
2023-525	2 ACT PENTRA	130 ABX PENTRA			29	4.464	0.051	1.15	4.200	4.730	
2023-525	3 ADVIA 120	116 ADVIA-120			17	4.364	0.097	2.23	4.100	4.630	
2023-525	4 AB-CELL DY	115 ABBOTT C-D			58	4.521	0.070	1.55	4.250	4.790	
2023-525	4 AB-CELL DY	153 AB-CEL-RUB			52	4.655	0.077	1.66	4.380	4.930	
2023-525	5 BECTON D	137 B.C DXH			59	4.139	0.065	1.58	3.890	4.390	
2023-525	6 MR Mindray	145 BC-5390			52	4.499	0.061	1.36	4.230	4.770	
2023-525	7 DXH500 Ser	149 BC DXH 500			58	4.722	0.075	1.60	4.440	5.010	
2023-525	8 3 ST	84 ABBOTT C-D			27	4.141	0.068	1.64	3.890	4.390	
2023-525	8 3 ST	138 SY KX-21N			15	4.118	0.058	1.42	3.870	4.370	
2023-525	9 3 D	33 COULTER JT			10	4.644	0.058	1.24	4.370	4.920	
2023-525	9 3 D	120 COULTER			16	4.589	0.076	1.66	4.310	4.860	
550	HEMATOCRIT (HCT)	CLIA TEST ID #: 785									
2023-521	1 Sysmex CBC	132 SYSMEX XS			152	40.450	0.708	1.75	38.000	42.900	
2023-521	1 Sysmex CBC	141 SYSMEX XN			47	39.453	0.674	1.71	37.100	41.800	
2023-521	1 Sysmex CBC	152 Sysmex XN			137	39.613	0.787	1.99	37.200	42.000	
2023-521	2 ACT PENTRA	125 COULTER AC			121	37.644	0.750	1.99	35.400	39.900	
2023-521	2 ACT PENTRA	130 ABX PENTRA			29	37.169	0.664	1.79	34.900	39.400	
2023-521	3 ADVIA 120	116 ADVIA-120			17	36.306	0.749	2.06	34.100	38.500	
2023-521	4 AB-CELL DY	115 ABBOTT C-D			59	37.407	0.943	2.52	35.200	39.700	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-521	4 AB-CELL DY	153 AB-CEL-RUB		51	38.120	1.144	3.00	35.800	40.400		
2023-521	5 BECTON D	137 B.C DXH		60	39.345	0.759	1.93	37.000	41.700		
2023-521	6 MR Mindray	145 BC-5390		52	38.469	0.883	2.29	36.200	40.800		
2023-521	7 DXH500 Ser	149 BC DXH 500		58	42.002	1.074	2.56	39.500	44.500		
2023-521	8 3 ST	84 ABBOTT C-D		27	34.389	1.003	2.92	32.300	36.500		
2023-521	8 3 ST	138 SY KX-21N		15	32.100	0.737	2.29	30.200	34.000		
2023-521	9 3 D	33 COULTER JT		10	39.050	0.880	2.25	36.700	41.400		
2023-521	9 3 D	120 COULTER		16	38.550	0.696	1.80	36.200	40.900		
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785	EVALUATION CRITERIA: TARGET VALUE 6 %								
2023-522	1 Sysmex CBC	132 SYSMEX XS		152	54.150	0.956	1.77	50.900	57.400		
2023-522	1 Sysmex CBC	141 SYSMEX XN		47	52.913	0.799	1.51	49.700	56.100		
2023-522	1 Sysmex CBC	152 Sysmex XN		137	53.164	1.026	1.93	50.000	56.400		
2023-522	2 ACT PENTRA	125 COULTER AC		121	44.110	0.776	1.76	41.500	46.800		
2023-522	2 ACT PENTRA	130 ABX PENTRA		29	43.779	0.617	1.41	41.200	46.400		
2023-522	3 ADVIA 120	116 ADVIA-120		17	47.082	1.200	2.55	44.300	49.900		
2023-522	4 AB-CELL DY	115 ABBOTT C-D		59	43.815	1.096	2.50	41.200	46.400		
2023-522	4 AB-CELL DY	153 AB-CEL-RUB		51	45.151	1.293	2.86	42.400	47.900		
2023-522	5 BECTON D	137 B.C DXH		60	57.971	0.912	1.57	54.500	61.400		
2023-522	6 MR Mindray	145 BC-5390		52	47.633	1.297	2.72	44.800	50.500		
2023-522	7 DXH500 Ser	149 BC DXH 500		58	50.256	1.082	2.15	47.200	53.300		
2023-522	8 3 ST	84 ABBOTT C-D		27	53.319	1.852	3.47	50.100	56.500		
2023-522	8 3 ST	138 SY KX-21N		15	50.307	1.309	2.60	47.300	53.300		
2023-522	9 3 D	33 COULTER JT		10	51.670	1.370	2.65	48.600	54.800		
2023-522	9 3 D	120 COULTER		16	50.988	0.926	1.82	47.900	54.000		
2023-523	1 Sysmex CBC	132 SYSMEX XS		152	60.381	1.090	1.80	56.800	64.000		
2023-523	1 Sysmex CBC	141 SYSMEX XN		47	59.466	0.846	1.42	55.900	63.000		
2023-523	1 Sysmex CBC	152 Sysmex XN		137	59.726	1.158	1.94	56.100	63.300		
2023-523	2 ACT PENTRA	125 COULTER AC		121	44.068	0.817	1.85	41.400	46.700		
2023-523	2 ACT PENTRA	130 ABX PENTRA		29	43.672	0.664	1.52	41.100	46.300		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-523	3	ADVIA 120	116 ADVIA-120	17	47.153	1.172	2.49	44.300	50.000	
2023-523	4	AB-CELL DY	115 ABBOTT C-D	59	51.629	1.295	2.51	48.500	54.700	
2023-523	4	AB-CELL DY	153 AB-CEL-RUB	51	53.296	1.358	2.55	50.100	56.500	
2023-523	5	BECTON D	137 B.C DXH	60	65.608	1.054	1.61	61.700	69.500	
2023-523	6	MR Mindray	145 BC-5390	52	47.653	1.098	2.30	44.800	50.500	
2023-523	7	DXH500 Ser	149 BC DXH 500	58	50.559	1.129	2.23	47.500	53.600	
2023-523	8	3 ST	84 ABBOTT C-D	27	53.063	1.763	3.32	49.900	56.200	
2023-523	8	3 ST	138 SY KX-21N	15	50.273	1.041	2.07	47.300	53.300	
2023-523	9	3 D	33 COULTER JT	10	51.500	1.007	1.96	48.400	54.600	
2023-523	9	3 D	120 COULTER	16	51.569	1.026	1.99	48.500	54.700	
550 HEMATOCRIT (HCT)				CLIA TEST ID #: 785	EVALUATION CRITERIA: TARGET VALUE 6 %					
2023-524	1	Sysmex CBC	132 SYSMEX XS	152	18.334	0.348	1.90	17.200	19.400	
2023-524	1	Sysmex CBC	141 SYSMEX XN	47	17.702	0.294	1.66	16.600	18.800	
2023-524	1	Sysmex CBC	152 Sysmex XN	137	17.711	0.348	1.97	16.600	18.800	
2023-524	2	ACT PENTRA	125 COULTER AC	121	18.211	0.354	1.94	17.100	19.300	
2023-524	2	ACT PENTRA	130 ABX PENTRA	29	17.993	0.327	1.82	16.900	19.100	
2023-524	3	ADVIA 120	116 ADVIA-120	17	15.847	0.403	2.54	14.900	16.800	
2023-524	4	AB-CELL DY	115 ABBOTT C-D	59	15.966	0.444	2.78	15.000	16.900	
2023-524	4	AB-CELL DY	153 AB-CEL-RUB	51	16.010	0.423	2.64	15.000	17.000	
2023-524	5	BECTON D	137 B.C DXH	60	20.407	0.387	1.90	19.200	21.600	
2023-524	6	MR Mindray	145 BC-5390	52	17.118	0.394	2.30	16.100	18.100	
2023-524	7	DXH500 Ser	149 BC DXH 500	58	19.126	0.487	2.55	18.000	20.300	
2023-524	8	3 ST	84 ABBOTT C-D	27	18.162	0.430	2.37	17.100	19.300	
2023-524	8	3 ST	138 SY KX-21N	15	17.440	0.486	2.78	16.400	18.500	
2023-524	9	3 D	33 COULTER JT	10	17.380	0.397	2.28	16.300	18.400	
2023-524	9	3 D	120 COULTER	16	17.569	0.306	1.74	16.500	18.600	
2023-525	1	Sysmex CBC	132 SYSMEX XS	152	40.343	0.700	1.74	37.900	42.800	
2023-525	1	Sysmex CBC	141 SYSMEX XN	47	39.577	0.589	1.49	37.200	42.000	
2023-525	1	Sysmex CBC	152 Sysmex XN	137	39.647	0.719	1.81	37.300	42.000	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-525	2	ACT PENTRA	125 COULTER AC	121	37.484	0.732	1.95	35.200	39.700	
2023-525	2	ACT PENTRA	130 ABX PENTRA	29	37.217	0.558	1.50	35.000	39.500	
2023-525	3	ADVIA 120	116 ADVIA-120	17	36.135	0.929	2.57	34.000	38.300	
2023-525	4	AB-CELL DY	115 ABBOTT C-D	59	37.208	0.990	2.66	35.000	39.400	
2023-525	4	AB-CELL DY	153 AB-CEL-RUB	51	37.975	0.863	2.27	35.700	40.300	
2023-525	5	BECTON D	137 B.C DXH	60	39.402	0.665	1.69	37.000	41.800	
2023-525	6	MR Mindray	145 BC-5390	52	38.338	0.985	2.57	36.000	40.600	
2023-525	7	DXH500 Ser	149 BC DXH 500	58	41.582	0.820	1.97	39.100	44.100	
2023-525	8	3 ST	84 ABBOTT C-D	27	34.467	0.944	2.74	32.400	36.500	
2023-525	8	3 ST	138 SY KX-21N	15	32.180	0.846	2.63	30.200	34.100	
2023-525	9	3 D	33 COULTER JT	10	39.060	1.037	2.66	36.700	41.400	
2023-525	9	3 D	120 COULTER	16	38.581	0.741	1.92	36.300	40.900	
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}						
2023-521	1	Sysmex CBC	132 SYSMEX XS	152	13.501	0.148	1.10	12.600	14.400	
2023-521	1	Sysmex CBC	141 SYSMEX XN	47	13.533	0.120	0.89	12.600	14.500	
2023-521	1	Sysmex CBC	152 Sysmex XN	137	13.582	0.122	0.90	12.600	14.500	
2023-521	2	ACT PENTRA	125 COULTER AC	121	13.526	0.165	1.22	12.600	14.500	
2023-521	2	ACT PENTRA	130 ABX PENTRA	29	13.503	0.197	1.46	12.600	14.400	
2023-521	3	ADVIA 120	116 ADVIA-120	17	13.276	0.190	1.43	12.300	14.200	
2023-521	4	AB-CELL DY	115 ABBOTT C-D	60	13.412	0.237	1.76	12.500	14.400	
2023-521	4	AB-CELL DY	153 AB-CEL-RUB	50	13.414	0.261	1.95	12.500	14.400	
2023-521	5	BECTON D	137 B.C DXH	60	12.368	0.182	1.47	11.500	13.200	
2023-521	6	MR Mindray	145 BC-5390	52	13.027	0.116	0.89	12.100	13.900	
2023-521	7	DXH500 Ser	149 BC DXH 500	58	14.845	0.253	1.70	13.800	15.900	
2023-521	8	3 ST	84 ABBOTT C-D	27	11.974	0.215	1.80	11.100	12.800	
2023-521	8	3 ST	138 SY KX-21N	15	11.853	0.171	1.44	11.000	12.700	
2023-521	9	3 D	33 COULTER JT	10	13.740	0.246	1.79	12.800	14.700	
2023-521	9	3 D	120 COULTER	16	13.456	0.294	2.18	12.500	14.400	
2023-522	1	Sysmex CBC	132 SYSMEX XS	152	18.060	0.193	1.07	16.800	19.300	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-522	1	Sysmex CBC	141 SYSMEX XN	47	18.007	0.167	0.93	16.700	19.300		
2023-522	1	Sysmex CBC	152 Sysmex XN	137	18.069	0.158	0.87	16.800	19.300		
2023-522	2	ACT PENTRA	125 COULTER AC	121	15.829	0.185	1.17	14.700	16.900		
2023-522	2	ACT PENTRA	130 ABX PENTRA	29	15.907	0.229	1.44	14.800	17.000		
2023-522	3	ADVIA 120	116 ADVIA-120	17	17.388	0.242	1.39	16.200	18.600		
2023-522	4	AB-CELL DY	115 ABBOTT C-D	60	16.632	0.264	1.59	15.500	17.800		
2023-522	4	AB-CELL DY	153 AB-CEL-RUB	50	16.698	0.303	1.81	15.500	17.900		
2023-522	5	BECTON D	137 B.C DXH	60	17.513	0.174	0.99	16.300	18.700		
2023-522	6	MR Mindray	145 BC-5390	52	16.260	0.164	1.01	15.100	17.400		
2023-522	7	DXH500 Ser	149 BC DXH 500	58	17.919	0.263	1.47	16.700	19.200		
2023-522	8	3 ST	84 ABBOTT C-D	27	18.433	0.437	2.37	17.100	19.700		
2023-522	8	3 ST	138 SY KX-21N	15	18.347	0.250	1.36	17.100	19.600		
2023-522	9	3 D	33 COULTER JT	10	18.370	0.372	2.02	17.100	19.700		
2023-522	9	3 D	120 COULTER	16	17.981	0.334	1.86	16.700	19.200		
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE	7 % OR 2.00 {LOWER}							
2023-523	1	Sysmex CBC	132 SYSMEX XS	152	20.201	0.220	1.09	18.800	21.600		
2023-523	1	Sysmex CBC	141 SYSMEX XN	47	20.157	0.171	0.85	18.700	21.600		
2023-523	1	Sysmex CBC	152 Sysmex XN	137	20.217	0.185	0.92	18.800	21.600		
2023-523	2	ACT PENTRA	125 COULTER AC	121	15.836	0.202	1.28	14.700	16.900		
2023-523	2	ACT PENTRA	130 ABX PENTRA	29	15.914	0.237	1.49	14.800	17.000		
2023-523	3	ADVIA 120	116 ADVIA-120	17	17.447	0.287	1.65	16.200	18.700		
2023-523	4	AB-CELL DY	115 ABBOTT C-D	60	18.018	0.296	1.64	16.800	19.300		
2023-523	4	AB-CELL DY	153 AB-CEL-RUB	50	18.108	0.389	2.15	16.800	19.400		
2023-523	5	BECTON D	137 B.C DXH	60	19.776	0.199	1.01	18.400	21.200		
2023-523	6	MR Mindray	145 BC-5390	52	16.265	0.186	1.14	15.100	17.400		
2023-523	7	DXH500 Ser	149 BC DXH 500	58	17.916	0.257	1.44	16.700	19.200		
2023-523	8	3 ST	84 ABBOTT C-D	27	18.400	0.488	2.65	17.100	19.700		
2023-523	8	3 ST	138 SY KX-21N	15	18.353	0.250	1.36	17.100	19.600		
2023-523	9	3 D	33 COULTER JT	10	18.400	0.392	2.13	17.100	19.700		
2023-523	9	3 D	120 COULTER	16	18.106	0.270	1.49	16.800	19.400		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
551	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}								
2023-524	1 Sysmex CBC	132 SYSMEX XS			152	5.891	0.090	1.53	5.500	6.300	
2023-524	1 Sysmex CBC	141 SYSMEX XN			47	5.972	0.089	1.49	5.600	6.400	
2023-524	1 Sysmex CBC	152 Sysmex XN			137	5.988	0.073	1.22	5.600	6.400	
2023-524	2 ACT PENTRA	125 COULTER AC			121	6.265	0.110	1.76	5.800	6.700	
2023-524	2 ACT PENTRA	130 ABX PENTRA			29	6.128	0.101	1.65	5.700	6.600	
2023-524	3 ADVIA 120	116 ADVIA-120			17	5.782	0.104	1.80	5.400	6.200	
2023-524	4 AB-CELL DY	115 ABBOTT C-D			60	5.867	0.096	1.64	5.500	6.300	
2023-524	4 AB-CELL DY	153 AB-CEL-RUB			50	5.690	0.130	2.28	5.300	6.100	
2023-524	5 BECTON D	137 B.C DXH			60	6.485	0.082	1.26	6.000	6.900	
2023-524	6 MR Mindray	145 BC-5390			52	5.590	0.069	1.24	5.200	6.000	
2023-524	7 DXH500 Ser	149 BC DXH 500			58	6.645	0.100	1.51	6.200	7.100	
2023-524	8 3 ST	84 ABBOTT C-D			27	6.452	0.191	2.96	6.000	6.900	
2023-524	8 3 ST	138 SY KX-21N			15	6.253	0.131	2.09	5.800	6.700	
2023-524	9 3 D	33 COULTER JT			10	5.780	0.244	4.22	5.400	6.200	
2023-524	9 3 D	120 COULTER			16	6.044	0.154	2.55	5.600	6.500	
2023-525	1 Sysmex CBC	132 SYSMEX XS			152	13.499	0.135	1.00	12.600	14.400	
2023-525	1 Sysmex CBC	141 SYSMEX XN			47	13.570	0.112	0.83	12.600	14.500	
2023-525	1 Sysmex CBC	152 Sysmex XN			137	13.573	0.129	0.95	12.600	14.500	
2023-525	2 ACT PENTRA	125 COULTER AC			121	13.526	0.186	1.37	12.600	14.500	
2023-525	2 ACT PENTRA	130 ABX PENTRA			29	13.531	0.190	1.40	12.600	14.500	
2023-525	3 ADVIA 120	116 ADVIA-120			17	13.235	0.178	1.35	12.300	14.200	
2023-525	4 AB-CELL DY	115 ABBOTT C-D			60	13.302	0.218	1.64	12.400	14.200	
2023-525	4 AB-CELL DY	153 AB-CEL-RUB			50	13.272	0.256	1.93	12.300	14.200	
2023-525	5 BECTON D	137 B.C DXH			60	12.339	0.152	1.23	11.500	13.200	
2023-525	6 MR Mindray	145 BC-5390			52	12.994	0.130	1.00	12.100	13.900	
2023-525	7 DXH500 Ser	149 BC DXH 500			58	14.719	0.173	1.17	13.700	15.700	
2023-525	8 3 ST	84 ABBOTT C-D			27	12.067	0.239	1.98	11.200	12.900	
2023-525	8 3 ST	138 SY KX-21N			15	11.887	0.131	1.10	11.100	12.700	

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-525	9 3 D	33 COULTER JT		10	13.650	0.358	2.63	12.700	14.600
2023-525	9 3 D	120 COULTER		16	13.481	0.201	1.49	12.500	14.400
552 WHITE CELL COUNT(WBC)				CLIA TEST ID #: 805 EVALUATION CRITERIA: TARGET VALUE 15 %					
2023-521	1 Sysmex CBC	132 SYSMEX XS		151	7.836	0.259	3.30	6.700	9.000
2023-521	1 Sysmex CBC	141 SYSMEX XN		48	7.755	0.180	2.32	6.600	8.900
2023-521	1 Sysmex CBC	152 Sysmex XN		137	7.442	0.203	2.72	6.300	8.600
2023-521	2 ACT PENTRA	125 COULTER AC		121	7.567	0.237	3.13	6.400	8.700
2023-521	2 ACT PENTRA	130 ABX PENTRA		29	7.610	0.325	4.27	6.500	8.800
2023-521	3 ADVIA 120	116 ADVIA-120		17	7.600	0.474	6.24	6.500	8.700
2023-521	4 AB-CELL DY	115 ABBOTT C-D		60	7.731	0.281	3.64	6.600	8.900
2023-521	4 AB-CELL DY	153 AB-CEL-RUB		50	7.753	0.259	3.34	6.600	8.900
2023-521	5 BECTON D	137 B.C DXH		60	9.700	0.238	2.45	8.200	11.200
2023-521	6 MR Mindray	145 BC-5390		52	8.381	0.194	2.32	7.100	9.600
2023-521	7 DXH500 Ser	149 BC DXH 500		58	7.368	0.194	2.63	6.300	8.500
2023-521	8 3 ST	84 ABBOTT C-D		27	7.996	0.405	5.07	6.800	9.200
2023-521	8 3 ST	138 SY KX-21N		15	7.907	0.202	2.55	6.700	9.100
2023-521	9 3 D	33 COULTER JT		10	8.160	0.269	3.30	6.900	9.400
2023-521	9 3 D	120 COULTER		16	8.180	0.180	2.20	7.000	9.400
2023-522	1 Sysmex CBC	132 SYSMEX XS		151	20.677	0.725	3.51	17.600	23.800
2023-522	1 Sysmex CBC	141 SYSMEX XN		48	21.335	0.592	2.78	18.100	24.500
2023-522	1 Sysmex CBC	152 Sysmex XN		137	20.629	0.518	2.51	17.500	23.700
2023-522	2 ACT PENTRA	125 COULTER AC		121	18.021	0.548	3.04	15.300	20.700
2023-522	2 ACT PENTRA	130 ABX PENTRA		29	18.200	0.756	4.15	15.500	20.900
2023-522	3 ADVIA 120	116 ADVIA-120		17	16.918	0.786	4.65	14.400	19.500
2023-522	4 AB-CELL DY	115 ABBOTT C-D		60	20.570	0.789	3.83	17.500	23.700
2023-522	4 AB-CELL DY	153 AB-CEL-RUB		50	20.632	0.624	3.02	17.500	23.700
2023-522	5 BECTON D	137 B.C DXH		60	21.741	0.488	2.24	18.500	25.000
2023-522	6 MR Mindray	145 BC-5390		52	19.002	0.429	2.26	16.200	21.900
2023-522	7 DXH500 Ser	149 BC DXH 500		58	18.278	0.493	2.70	15.500	21.000

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023-522	8 3 ST	84	ABBOTT C-D	27	20.122	1.110	5.51	17.100	23.100		
2023-522	8 3 ST	138	SY KX-21N	15	19.880	0.511	2.57	16.900	22.900		
2023-522	9 3 D	33	COULTER JT	10	21.570	0.723	3.35	18.300	24.800		
2023-522	9 3 D	120	COULTER	16	20.906	0.498	2.38	17.800	24.000		
552 WHITE CELL COUNT(WBC)				CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %						
2023-523	1 Sysmex CBC	132	SYSMEX XS	151	22.472	0.865	3.85	19.100	25.800		
2023-523	1 Sysmex CBC	141	SYSMEX XN	48	23.371	0.651	2.79	19.900	26.900		
2023-523	1 Sysmex CBC	152	Sysmex XN	137	22.441	0.687	3.06	19.100	25.800		
2023-523	2 ACT PENTRA	125	COULTER AC	121	18.085	0.709	3.92	15.400	20.800		
2023-523	2 ACT PENTRA	130	ABX PENTRA	29	18.155	0.705	3.88	15.400	20.900		
2023-523	3 ADVIA 120	116	ADVIA-120	17	17.076	0.678	3.97	14.500	19.600		
2023-523	4 AB-CELL DY	115	ABBOTT C-D	60	9.733	0.335	3.44	8.300	11.200		
2023-523	4 AB-CELL DY	153	AB-CEL-RUB	50	9.832	0.344	3.50	8.400	11.300		
2023-523	5 BECTON D	137	B.C DXH	60	25.103	0.392	1.56	21.300	28.900		
2023-523	6 MR Mindray	145	BC-5390	52	18.892	0.417	2.21	16.100	21.700		
2023-523	7 DXH500 Ser	149	BC DXH 500	58	18.274	0.486	2.66	15.500	21.000		
2023-523	8 3 ST	84	ABBOTT C-D	27	20.152	1.041	5.16	17.100	23.200		
2023-523	8 3 ST	138	SY KX-21N	15	19.760	0.512	2.59	16.800	22.700		
2023-523	9 3 D	33	COULTER JT	10	21.430	0.642	3.00	18.200	24.600		
2023-523	9 3 D	120	COULTER	16	21.113	0.487	2.31	17.900	24.300		
2023-524	1 Sysmex CBC	132	SYSMEX XS	151	4.057	0.128	3.17	3.400	4.700		
2023-524	1 Sysmex CBC	141	SYSMEX XN	48	3.948	0.121	3.06	3.400	4.500		
2023-524	1 Sysmex CBC	152	Sysmex XN	137	3.780	0.118	3.11	3.200	4.300		
2023-524	2 ACT PENTRA	125	COULTER AC	121	2.445	0.118	4.81	2.100	2.800		
2023-524	2 ACT PENTRA	130	ABX PENTRA	29	2.459	0.133	5.40	2.100	2.800		
2023-524	3 ADVIA 120	116	ADVIA-120	17	3.488	0.102	2.93	3.000	4.000		
2023-524	4 AB-CELL DY	115	ABBOTT C-D	60	3.040	0.132	4.34	2.600	3.500		
2023-524	4 AB-CELL DY	153	AB-CEL-RUB	50	3.108	0.125	4.01	2.600	3.600		
2023-524	5 BECTON D	137	B.C DXH	60	4.197	0.162	3.85	3.600	4.800		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-524	6 MR Mindray	145 BC-5390		52	3.553	0.096	2.69	3.000	4.100		
2023-524	7 DXH500 Ser	149 BC DXH 500		58	2.346	0.117	4.99	2.000	2.700		
2023-524	8 3 ST	84 ABBOTT C-D		27	2.922	0.193	6.61	2.500	3.400		
2023-524	8 3 ST	138 SY KX-21N		15	2.767	0.087	3.14	2.400	3.200		
2023-524	9 3 D	33 COULTER JT		10	2.090	0.094	4.51	1.800	2.400		
2023-524	9 3 D	120 COULTER		16	2.313	0.141	6.09	2.000	2.700		
552 WHITE CELL COUNT(WBC)		CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %								
2023-525	1 Sysmex CBC	132 SYSMEX XS		151	7.846	0.254	3.24	6.700	9.000		
2023-525	1 Sysmex CBC	141 SYSMEX XN		48	7.700	0.209	2.72	6.500	8.900		
2023-525	1 Sysmex CBC	152 Sysmex XN		137	7.457	0.181	2.42	6.300	8.600		
2023-525	2 ACT PENTRA	125 COULTER AC		121	7.598	0.324	4.26	6.500	8.700		
2023-525	2 ACT PENTRA	130 ABX PENTRA		29	7.607	0.324	4.26	6.500	8.700		
2023-525	3 ADVIA 120	116 ADVIA-120		17	7.659	0.422	5.50	6.500	8.800		
2023-525	4 AB-CELL DY	115 ABBOTT C-D		60	7.675	0.210	2.74	6.500	8.800		
2023-525	4 AB-CELL DY	153 AB-CEL-RUB		50	7.738	0.238	3.08	6.600	8.900		
2023-525	5 BECTON D	137 B.C DXH		60	9.710	0.194	1.99	8.300	11.200		
2023-525	6 MR Mindray	145 BC-5390		52	8.665	0.241	2.78	7.400	10.000		
2023-525	7 DXH500 Ser	149 BC DXH 500		58	7.272	0.201	2.76	6.200	8.400		
2023-525	8 3 ST	84 ABBOTT C-D		27	8.026	0.491	6.12	6.800	9.200		
2023-525	8 3 ST	138 SY KX-21N		15	8.000	0.159	1.99	6.800	9.200		
2023-525	9 3 D	33 COULTER JT		10	8.180	0.209	2.55	7.000	9.400		
2023-525	9 3 D	120 COULTER		16	8.206	0.233	2.84	7.000	9.400		

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
553 PLATELET COUNT		CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %								
2023-521	1 SYSMEX CBC	79 SYSMEX XS		151	199.149	9.228	4.63	149.000	249.000		
2023-521	1 SYSMEX CBC	87 SYSMEX XN		49	204.625	6.833	3.34	153.000	256.000		

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-521	1	SYSMEX CBC	98 SYSMEX 350	136	204.368	7.496	3.67	153.000	255.000
2023-521	2	ACT5 PENTR	73 COULTER AC	118	247.179	9.993	4.04	185.000	309.000
2023-521	2	ACT5 PENTR	78 ABX PENTRA	29	245.750	7.799	3.17	184.000	307.000
2023-521	3	ADVIA 120	64 ADVIA-120	17	208.294	11.385	5.47	156.000	260.000
2023-521	4	AB- CELL-D	63 ABBOTT C-D	58	269.684	14.086	5.22	202.000	337.000
2023-521	4	AB- CELL-D	100 ABB.RUBY	50	270.220	15.595	5.77	203.000	338.000
2023-521	5	BECTON D	83 B.C600,800	60	254.881	7.054	2.77	191.000	319.000
2023-521	6	MR Mindray	91 BC-5390	52	263.731	8.454	3.21	198.000	330.000
2023-521	8	3 ST	31 ABBOTT C-D	27	186.963	19.416	10.39	140.000	234.000
2023-521	8	3 ST	84 KX21,300	16	198.133	5.226	2.64	149.000	248.000
2023-521	9	3 D	68 COULTERAct	16	275.438	12.791	4.64	207.000	344.000
2023-521	10	DxH	94 BC DXH 500	56	266.709	15.611	5.85	200.000	333.000
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %					
2023-522	1	SYSMEX CBC	79 SYSMEX XS	151	387.122	19.093	4.93	290.000	484.000
2023-522	1	SYSMEX CBC	87 SYSMEX XN	49	414.429	14.217	3.43	311.000	518.000
2023-522	1	SYSMEX CBC	98 SYSMEX 350	136	414.716	14.935	3.60	311.000	518.000
2023-522	2	ACT5 PENTR	73 COULTER AC	118	479.179	14.715	3.07	359.000	599.000
2023-522	2	ACT5 PENTR	78 ABX PENTRA	29	471.250	15.068	3.20	353.000	589.000
2023-522	3	ADVIA 120	64 ADVIA-120	17	411.647	20.945	5.09	309.000	515.000
2023-522	4	AB- CELL-D	63 ABBOTT C-D	58	497.466	28.020	5.63	373.000	622.000
2023-522	4	AB- CELL-D	100 ABB.RUBY	50	504.560	22.743	4.51	378.000	631.000
2023-522	5	BECTON D	83 B.C600,800	60	461.068	12.918	2.80	346.000	576.000
2023-522	6	MR Mindray	91 BC-5390	52	509.712	14.628	2.87	382.000	637.000
2023-522	8	3 ST	31 ABBOTT C-D	27	367.259	35.872	9.77	275.000	459.000
2023-522	8	3 ST	84 KX21,300	16	396.188	10.973	2.77	297.000	495.000
2023-522	9	3 D	68 COULTERAct	16	587.813	26.003	4.42	441.000	735.000
2023-522	10	DxH	94 BC DXH 500	56	502.400	20.631	4.11	377.000	628.000
2023-523	1	SYSMEX CBC	79 SYSMEX XS	151	381.329	21.135	5.54	286.000	477.000
2023-523	1	SYSMEX CBC	87 SYSMEX XN	49	423.714	17.158	4.05	318.000	530.000

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-523	1	SYSMEX CBC	98	SYSMEX 350	136	416.955	14.585	3.50	313.000	521.000	
2023-523	2	ACT5 PENTR	73	COULTER AC	118	477.265	17.155	3.59	358.000	597.000	
2023-523	2	ACT5 PENTR	78	ABX PENTRA	29	470.000	13.813	2.94	353.000	588.000	
2023-523	3	ADVIA 120	64	ADVIA-120	17	412.941	25.466	6.17	310.000	516.000	
2023-523	4	AB- CELL-D	63	ABBOTT C-D	58	158.439	16.056	10.13	119.000	198.000	
2023-523	4	AB- CELL-D	100	ABB.RUBY	50	147.347	11.208	7.61	111.000	184.000	
2023-523	5	BECTON D	83	B.C600,800	60	474.932	16.422	3.46	356.000	594.000	
2023-523	6	MR Mindray	91	BC-5390	52	510.442	13.421	2.63	383.000	638.000	
2023-523	8	3 ST	31	ABBOTT C-D	27	365.556	35.149	9.62	274.000	457.000	
2023-523	8	3 ST	84	KX21,300	16	399.313	15.296	3.83	299.000	499.000	
2023-523	9	3 D	68	COULTERACT	16	593.000	36.120	6.09	445.000	741.000	
2023-523	10	DxH	94	BC DXH 500	56	498.255	22.925	4.60	374.000	623.000	
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2023-524	1	SYSMEX CBC	79	SYSMEX XS	151	52.133	3.486	6.69	39.000	65.000	
2023-524	1	SYSMEX CBC	87	SYSMEX XN	49	46.122	2.480	5.38	35.000	58.000	
2023-524	1	SYSMEX CBC	98	SYSMEX 350	136	47.356	3.140	6.63	36.000	59.000	
2023-524	2	ACT5 PENTR	73	COULTER AC	118	72.381	5.264	7.27	54.000	90.000	
2023-524	2	ACT5 PENTR	78	ABX PENTRA	29	72.345	5.319	7.35	54.000	90.000	
2023-524	3	ADVIA 120	64	ADVIA-120	17	49.588	4.339	8.75	37.000	62.000	
2023-524	4	AB- CELL-D	63	ABBOTT C-D	58	79.054	6.632	8.39	59.000	99.000	
2023-524	4	AB- CELL-D	100	ABB.RUBY	50	76.000	5.485	7.22	57.000	95.000	
2023-524	5	BECTON D	83	B.C600,800	60	82.136	2.534	3.09	62.000	103.000	
2023-524	6	MR Mindray	91	BC-5390	52	58.135	3.470	5.97	44.000	73.000	
2023-524	8	3 ST	31	ABBOTT C-D	27	65.481	6.919	10.57	49.000	82.000	
2023-524	8	3 ST	84	KX21,300	16	65.125	3.276	5.03	49.000	81.000	
2023-524	9	3 D	68	COULTERACT	16	74.813	6.054	8.09	56.000	94.000	
2023-524	10	DxH	94	BC DXH 500	56	74.727	4.912	6.57	56.000	93.000	
2023-525	1	SYSMEX CBC	79	SYSMEX XS	151	197.459	11.147	5.65	148.000	247.000	
2023-525	1	SYSMEX CBC	87	SYSMEX XN	49	204.104	7.695	3.77	153.000	255.000	

Participant Summary
MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-525	1	SYSMEX CBC	98 SYSMEX 350	136	202.910	7.855	3.87	152.000	254.000
2023-525	2	ACT5 PENTR	73 COULTER AC	118	245.222	9.753	3.98	184.000	307.000
2023-525	2	ACT5 PENTR	78 ABX PENTRA	29	243.207	9.827	4.04	182.000	304.000
2023-525	3	ADVIA 120	64 ADVIA-120	17	207.588	11.692	5.63	156.000	259.000
2023-525	4	AB- CELL-D	63 ABBOTT C-D	58	268.931	15.658	5.82	202.000	336.000
2023-525	4	AB- CELL-D	100 ABB.RUBY	50	265.720	14.050	5.29	199.000	332.000
2023-525	5	BECTON D	83 B.C600,800	60	253.780	7.045	2.78	190.000	317.000
2023-525	6	MR Mindray	91 BC-5390	52	265.596	8.244	3.10	199.000	332.000
2023-525	8	3 ST	31 ABBOTT C-D	27	186.148	17.033	9.15	140.000	233.000
2023-525	8	3 ST	84 KX21,300	16	194.938	7.049	3.62	146.000	244.000
2023-525	9	3 D	68 COULTERACT	16	276.813	12.861	4.65	208.000	346.000
2023-525	10	DxH	94 BC DXH 500	56	263.818	13.893	5.27	198.000	330.000

SPEC: 20 HEMATOLOGY SUB: 75 BODY FLUID GRP: 145 BODY FLUID CELL COU.

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}						
2023-561				45	0.477	0.621	130.15	0	2.000
2023-562				45	0.489	0.687	140.54	0	3.000
2023-563				45	1037.114	94.185	9.08	755.000	1320.000
2023-564				45	0.511	0.687	134.43	0	3.000
2023-565				45	1039.545	104.620	10.06	726.000	1353.000

Participant Summary

MONTH 6 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 75 BODY FLUID GRP: 145 BODY FLUID CELL COU.

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
160	BF WHITE CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3			SD OR 25 %	{GREATER}		
2023-561					44	4.409	1.813	41.12	0	10.000
2023-562					44	4.409	1.992	45.18	0	10.000
2023-563					44	285.357	27.634	9.68	202.000	368.000
2023-564					44	4.727	1.935	40.93	0	11.000
2023-565					44	293.860	34.254	11.66	191.000	397.000

Participant Summary
MONTH 6 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
083 ABO GROUP													
2023-571				008 008	008 008		008 - O	82	82	82	100		
2023-572				008 008	008 008		008 - O	82	82	82	100		
2023-573				004 004	004 004		004 - B	82	82	82	100		
2023-574				001 003	001 003		001 - A	82	81	82	100		
2023-574				001 003	001 003		002 - A1	82	1	82	100		
2023-575				008 008	008 008		008 - O	82	82	82	100		
126 RH TYPE													
2023-571				003 004	003 004		003 - Rh-Du np	82	7	82	100		
2023-571				003 004	003 004		004 - Rh-	82	75	82	100		
2023-572				001 003	001 003		001 - Rh+	82	82	82	100		
2023-573				003 004	003 004		003 - Rh-Du np	82	7	82	100		
2023-573				003 004	003 004		004 - Rh-	82	75	82	100		
2023-574				001 003	001 003		001 - Rh+	82	82	82	100		
2023-575				001 003	001 003		001 - Rh+	82	81	82	100		
2023-575				001 003	001 003		003 - Rh-Du np	82	1	82	100		

Participant Summary

MONTH 6 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2023-581				001 001	001 001		001 - NEGATIVE	66	64	64	100
2023-581				001 001	001 001		002 - POSITIVE	66	2	64	0
2023-582				001 001	001 001		001 - NEGATIVE	66	66	66	100
2023-583				002 002	002 002		001 - NEGATIVE	66	2	64	0
2023-583				002 002	002 002		002 - POSITIVE	66	64	64	100
2023-584				001 001	001 001		001 - NEGATIVE	66	66	66	100
2023-585				002 002	002 002		002 - POSITIVE	66	66	66	100
SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2023-291	001 Poly AHG	001 Gel				002 004	002 - 2+	13	1	13	100
2023-291	001 Poly AHG	001 Gel				002 004	003 - 3+	13	6	13	100
2023-291	001 Poly AHG	001 Gel				002 004	004 - 4+	13	6	13	100
2023-291	001 Poly AHG	002 Tubo				002 004	002 - 2+	32	19	32	100
2023-291	001 Poly AHG	002 Tubo				002 004	003 - 3+	32	13	32	100
2023-291	002 MnoAnt-IgG	001 Gel				002 004	002 - 2+	12	1	12	100
2023-291	002 MnoAnt-IgG	001 Gel				002 004	003 - 3+	12	7	12	100
2023-291	002 MnoAnt-IgG	001 Gel				002 004	004 - 4+	12	4	12	100
2023-291	002 MnoAnt-IgG	002 Tubo				002 004	002 - 2+	18	14	18	100
2023-291	002 MnoAnt-IgG	002 Tubo				002 004	003 - 3+	18	4	18	100

Participant Summary

MONTH 6 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2023-292	001 Poly AHG	001 Gel				002 004	003 - 3+	13	8	13	100
2023-292	001 Poly AHG	001 Gel				002 004	004 - 4+	13	5	13	100
2023-292	001 Poly AHG	002 Tubo				002 004	001 - Negativo	32	2	30	0
2023-292	001 Poly AHG	002 Tubo				002 004	002 - 2+	32	21	30	100
2023-292	001 Poly AHG	002 Tubo				002 004	003 - 3+	32	9	30	100
2023-292	002 MnoAnt-IgG	001 Gel				002 004	003 - 3+	12	8	12	100
2023-292	002 MnoAnt-IgG	001 Gel				002 004	004 - 4+	12	4	12	100
2023-292	002 MnoAnt-IgG	002 Tubo		002	004		002 - 2+	18	16	18	100
2023-292	002 MnoAnt-IgG	002 Tubo		002	004		003 - 3+	18	2	18	100
2023-293	001 Poly AHG	001 Gel				002 004	003 - 3+	13	8	13	100
2023-293	001 Poly AHG	001 Gel				002 004	004 - 4+	13	5	13	100
2023-293	001 Poly AHG	002 Tubo				002 004	001 - Negativo	32	1	31	0
2023-293	001 Poly AHG	002 Tubo				002 004	002 - 2+	32	21	31	100
2023-293	001 Poly AHG	002 Tubo				002 004	003 - 3+	32	9	31	100
2023-293	001 Poly AHG	002 Tubo				002 004	004 - 4+	32	1	31	100
2023-293	002 MnoAnt-IgG	001 Gel				002 004	003 - 3+	12	8	12	100
2023-293	002 MnoAnt-IgG	001 Gel				002 004	004 - 4+	12	4	12	100
2023-293	002 MnoAnt-IgG	002 Tubo				002 004	002 - 2+	18	11	18	100
2023-293	002 MnoAnt-IgG	002 Tubo				002 004	003 - 3+	18	6	18	100
2023-293	002 MnoAnt-IgG	002 Tubo				002 004	004 - 4+	18	1	18	100
2023-294	001 Poly AHG	001 Gel		001	001	001 001	001 - Negativo	14	14	14	100
2023-294	001 Poly AHG	002 Tubo		001	001	001 001	001 - Negativo	32	32	32	100
2023-294	002 MnoAnt-IgG	001 Gel		001	001	001 001	001 - Negativo	11	11	11	100
2023-294	002 MnoAnt-IgG	002 Tubo		001	001	001 001	001 - Negativo	18	18	18	100
2023-295	001 Poly AHG	001 Gel		001	001	001 001	001 - Negativo	14	14	14	100

Participant Summary
MONTH 6 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-295	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	32	32	32	100		
2023-295	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100		
2023-295	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	18	18	18	100		

SPE: 30 IMMUNOHEMATOLOGY				SUB: 100 COMPATIBILITY TESTING				GRP: 170 COMPATIBILITY TEST.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
085 COMPATIBIL. TESTING													
2023-591				002 002	002 002		002 - INCOMPATIB	41	41	41	100		
2023-592				001 001	001 001		001 - COMPATIBLE	41	41	41	100		
2023-593				001 001	001 001		001 - COMPATIBLE	41	41	41	100		
2023-594				001 001	001 001		001 - COMPATIBLE	41	41	41	100		
2023-595				001 001	001 001		001 - COMPATIBLE	41	41	41	100		