

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE				SYST/INST		REAGENTS		NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD						LABS	VALUE	DEVIATION	CV	LOW	HIGH	
001 ALAN AMINO	ALT/SGPT	CLIA TEST ID #: 255		EVALUATION CRITERIA: TARGET VALUE 20 %									
2023-781	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	23	297.783	9.127	3.06	238.000	357.000		
2023-781	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	285.700	5.533	1.94	229.000	343.000		
2023-781	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	92	307.220	10.410	3.39	246.000	369.000		
2023-781	11 MOD. IFCC	184	INTEGRA	93	ROCHE	19	286.632	8.468	2.95	229.000	344.000		
2023-781	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	29	266.464	7.129	2.68	213.000	320.000		
2023-781	11 MOD. IFCC	191	RX-SERIES	126	RANDOX	36	347.171	21.090	6.07	278.000	417.000		
2023-781	11 MOD. IFCC	203	ROCHE	93	ROCHE	38	295.000	7.683	2.60	236.000	354.000		
2023-781	11 MOD. IFCC	219	ATELLICA	143	SIEMENS	13	339.846	10.369	3.05	272.000	408.000		
2023-781	12 IFCC witho	203	ROCHE	93	ROCHE	12	292.750	7.350	2.51	234.000	351.000		
2023-781	13 ALA &aKETO	165	VITROS	155	ALTV	129	277.622	8.012	2.89	222.000	333.000		
2023-782	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	23	120.435	4.261	3.54	96.000	145.000		
2023-782	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	124.700	2.610	2.09	100.000	150.000		
2023-782	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	92	128.473	4.663	3.63	103.000	154.000		
2023-782	11 MOD. IFCC	184	INTEGRA	93	ROCHE	19	117.632	3.344	2.84	94.000	141.000		
2023-782	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	29	107.517	3.775	3.51	86.000	129.000		
2023-782	11 MOD. IFCC	191	RX-SERIES	126	RANDOX	36	141.771	8.829	6.23	113.000	170.000		
2023-782	11 MOD. IFCC	203	ROCHE	93	ROCHE	38	121.053	4.006	3.31	97.000	145.000		
2023-782	11 MOD. IFCC	219	ATELLICA	143	SIEMENS	13	139.308	3.770	2.71	111.000	167.000		
2023-782	12 IFCC witho	203	ROCHE	93	ROCHE	12	119.833	2.939	2.45	96.000	144.000		
2023-782	13 ALA &aKETO	165	VITROS	155	ALTV	129	117.535	3.731	3.17	94.000	141.000		
2023-783	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	23	79.174	2.973	3.76	63.000	95.000		
2023-783	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	86.500	1.432	1.66	69.000	104.000		
2023-783	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	92	85.385	3.676	4.30	68.000	102.000		
2023-783	11 MOD. IFCC	184	INTEGRA	93	ROCHE	19	76.842	2.498	3.25	61.000	92.000		
2023-783	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	29	70.586	2.540	3.60	56.000	85.000		
2023-783	11 MOD. IFCC	191	RX-SERIES	126	RANDOX	36	93.250	6.006	6.44	75.000	112.000		
2023-783	11 MOD. IFCC	203	ROCHE	93	ROCHE	38	79.081	1.992	2.52	63.000	95.000		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	11 MOD. IFCC	219	ATELLICA	143	SIEMENS	13	91.308	2.893	3.17	73.000	110.000	
2023-783	12 IFCC witho	203	ROCHE	93	ROCHE	12	78.500	2.255	2.87	63.000	94.000	
2023-783	13 ALA &aKETO	165	VITROS	155	ALTV	129	77.236	2.628	3.40	62.000	93.000	
001	ALAN AMINO ALT/SGPT	CLIA TEST ID #: 255				EVALUATION CRITERIA: TARGET VALUE 20 %						
2023-784	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	23	410.043	15.040	3.67	328.000	492.000	
2023-784	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	383.500	8.868	2.31	307.000	460.000	
2023-784	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	92	419.604	14.359	3.42	336.000	504.000	
2023-784	11 MOD. IFCC	184	INTEGRA	93	ROCHE	19	390.632	12.995	3.33	313.000	469.000	
2023-784	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	29	364.379	11.961	3.28	292.000	437.000	
2023-784	11 MOD. IFCC	191	RX-SERIES	126	RANDOX	36	465.167	41.718	8.97	372.000	558.000	
2023-784	11 MOD. IFCC	203	ROCHE	93	ROCHE	38	403.162	12.104	3.00	323.000	484.000	
2023-784	11 MOD. IFCC	219	ATELLICA	143	SIEMENS	13	452.615	13.981	3.09	362.000	543.000	
2023-784	12 IFCC witho	203	ROCHE	93	ROCHE	12	400.333	11.564	2.89	320.000	480.000	
2023-784	13 ALA &aKETO	165	VITROS	155	ALTV	129	381.250	15.088	3.96	305.000	458.000	
2023-785	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	23	222.826	8.041	3.61	178.000	267.000	
2023-785	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	220.600	3.720	1.69	176.000	265.000	
2023-785	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	92	232.681	7.318	3.15	186.000	279.000	
2023-785	11 MOD. IFCC	184	INTEGRA	93	ROCHE	19	217.105	6.439	2.97	174.000	261.000	
2023-785	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	29	200.321	7.196	3.59	160.000	240.000	
2023-785	11 MOD. IFCC	191	RX-SERIES	126	RANDOX	36	260.472	15.805	6.07	208.000	313.000	
2023-785	11 MOD. IFCC	203	ROCHE	93	ROCHE	38	222.108	6.900	3.11	178.000	267.000	
2023-785	11 MOD. IFCC	219	ATELLICA	143	SIEMENS	13	255.000	7.706	3.02	204.000	306.000	
2023-785	12 IFCC witho	203	ROCHE	93	ROCHE	12	219.917	5.361	2.44	176.000	264.000	
2023-785	13 ALA &aKETO	165	VITROS	155	ALTV	129	211.325	5.824	2.76	169.000	254.000	
002	ALBUMIN	CLIA TEST ID #: 265				EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-781	3 DYE BI-B.G	165	VITROS	63	KODAK	144	4.557	0.171	3.76	4.100	5.000	
2023-781	3 DYE BI-B.G	184	INTEGRA	93	ROCHE	29	4.417	0.095	2.15	4.000	4.900	
2023-781	3 DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	30	4.372	0.091	2.07	3.900	4.800	
2023-781	3 DYE BI-B.G	191	RX-SERIES	126	RANDOX	37	4.346	0.139	3.19	3.900	4.800	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-781	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	13	4.338	0.121	2.79	3.900	4.800
2023-781	3	DYE BI-B.G	203	ROCHE	93	ROCHE	49	4.584	0.138	3.00	4.100	5.000
2023-781	3	DYE BI-B.G	219	ATELLICA	143	SIEMENS	12	4.425	0.136	3.08	4.000	4.900
2023-781	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	96	4.128	0.125	3.02	3.700	4.500
2023-781	5	DYE B-HABA	197	ARCHITECT	1	ABBOTT	11	4.309	0.108	2.51	3.900	4.700
002	ALBUMIN	CLIA TEST ID #: 265				EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-782	3	DYE BI-B.G	165	VITROS	63	KODAK	144	2.798	0.091	3.24	2.500	3.100
2023-782	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	29	2.848	0.104	3.64	2.600	3.100
2023-782	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	30	2.823	0.062	2.18	2.500	3.100
2023-782	3	DYE BI-B.G	191	RX-SERIES	126	RANDOX	37	2.881	0.104	3.59	2.600	3.200
2023-782	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	13	2.762	0.074	2.67	2.500	3.000
2023-782	3	DYE BI-B.G	203	ROCHE	93	ROCHE	49	2.935	0.102	3.48	2.600	3.200
2023-782	3	DYE BI-B.G	219	ATELLICA	143	SIEMENS	12	2.833	0.085	3.00	2.600	3.100
2023-782	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	96	2.568	0.084	3.25	2.300	2.800
2023-782	5	DYE B-HABA	197	ARCHITECT	1	ABBOTT	11	2.736	0.064	2.35	2.500	3.000
2023-783	3	DYE BI-B.G	165	VITROS	63	KODAK	144	2.405	0.072	2.97	2.200	2.600
2023-783	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	29	2.462	0.081	3.27	2.200	2.700
2023-783	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	30	2.452	0.062	2.54	2.200	2.700
2023-783	3	DYE BI-B.G	191	RX-SERIES	126	RANDOX	37	2.542	0.083	3.26	2.300	2.800
2023-783	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	13	2.408	0.092	3.81	2.200	2.600
2023-783	3	DYE BI-B.G	203	ROCHE	93	ROCHE	49	2.545	0.095	3.73	2.300	2.800
2023-783	3	DYE BI-B.G	219	ATELLICA	143	SIEMENS	12	2.442	0.076	3.11	2.200	2.700
2023-783	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	96	2.194	0.072	3.28	2.000	2.400
2023-783	5	DYE B-HABA	197	ARCHITECT	1	ABBOTT	11	2.364	0.048	2.04	2.100	2.600
2023-784	3	DYE BI-B.G	165	VITROS	63	KODAK	144	5.700	0.204	3.58	5.100	6.300
2023-784	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	29	5.372	0.189	3.52	4.800	5.900
2023-784	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	30	5.320	0.133	2.49	4.800	5.900
2023-784	3	DYE BI-B.G	191	RX-SERIES	126	RANDOX	37	5.214	0.170	3.25	4.700	5.700

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-784	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	13	5.254	0.139	2.65	4.700	5.800
2023-784	3	DYE BI-B.G	203	ROCHE	93	ROCHE	49	5.563	0.175	3.14	5.000	6.100
2023-784	3	DYE BI-B.G	219	ATELLICA	143	SIEMENS	12	5.350	0.171	3.19	4.800	5.900
2023-784	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	96	5.106	0.153	3.00	4.600	5.600
2023-784	5	DYE B-HABA	197	ARCHITECT	1	ABBOTT	11	5.273	0.135	2.57	4.700	5.800
002	ALBUMIN	CLIA TEST ID #: 265				EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-785	3	DYE BI-B.G	165	VITROS	63	KODAK	144	3.861	0.121	3.13	3.500	4.200
2023-785	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	29	3.828	0.094	2.46	3.400	4.200
2023-785	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	30	3.757	0.096	2.54	3.400	4.100
2023-785	3	DYE BI-B.G	191	RX-SERIES	126	RANDOX	37	3.776	0.117	3.10	3.400	4.200
2023-785	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	13	3.715	0.110	2.96	3.300	4.100
2023-785	3	DYE BI-B.G	203	ROCHE	93	ROCHE	49	3.918	0.127	3.25	3.500	4.300
2023-785	3	DYE BI-B.G	219	ATELLICA	143	SIEMENS	12	3.800	0.100	2.63	3.400	4.200
2023-785	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	96	3.481	0.101	2.91	3.100	3.800
2023-785	5	DYE B-HABA	197	ARCHITECT	1	ABBOTT	11	3.673	0.096	2.62	3.300	4.000
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275				EVALUATION CRITERIA: TARGET VALUE 30 %						
2023-781	3	COLORIMET.	184	INTEGRA	93	ROCHE	24	280.708	10.757	3.83	196.000	365.000
2023-781	3	COLORIMET.	203	ROCHE	93	ROCHE	36	279.694	9.907	3.54	196.000	364.000
2023-781	5	KINETIC 30	197	ARCHITECT	1	ABBOTT	11	260.000	11.265	4.33	182.000	338.000
2023-781	6	KINETIC 37	165	VITROS	63	KODAK	144	210.830	13.460	6.38	148.000	274.000
2023-781	6	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	222.400	18.752	8.43	156.000	289.000
2023-781	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	30	231.931	23.931	10.32	162.000	302.000
2023-781	10	IFCC	219	ATELLICA	143	SIEMENS	13	271.077	11.997	4.43	190.000	352.000
2023-781	11	Modified I	38	SIEMENS DI	143	SIEMENS	89	292.239	10.559	3.61	205.000	380.000
2023-781	11	Modified I	191	RX-SERIES	126	RANDOX	30	238.967	24.501	10.25	167.000	311.000
2023-781	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	12	255.000	14.888	5.84	179.000	332.000
2023-782	3	COLORIMET.	184	INTEGRA	93	ROCHE	24	119.083	5.671	4.76	83.000	155.000
2023-782	3	COLORIMET.	203	ROCHE	93	ROCHE	36	118.750	3.868	3.26	83.000	154.000
2023-782	5	KINETIC 30	197	ARCHITECT	1	ABBOTT	11	111.182	5.605	5.04	78.000	145.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-782	6	KINETIC 37	165 VITROS	63	KODAK	144	130.148	9.177	7.05	91.000	169.000	
2023-782	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	10	94.700	7.938	8.38	66.000	123.000	
2023-782	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	30	97.483	10.274	10.54	68.000	127.000	
2023-782	10	IFCC	219 ATELLICA	143	SIEMENS	13	114.462	3.993	3.49	80.000	149.000	
2023-782	11	Modified I	38 SIEMENS DI	143	SIEMENS	89	123.000	5.356	4.35	86.000	160.000	
2023-782	11	Modified I	191 RX-SERIES	126	RANDOX	30	106.433	12.377	11.63	75.000	138.000	
2023-782	12	P NITRO PH	197 ARCHITECT	1	ABBOTT	12	107.750	6.043	5.61	75.000	140.000	
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %									
2023-783	3	COLORIMET.	184 INTEGRA	93	ROCHE	24	81.917	3.662	4.47	57.000	106.000	
2023-783	3	COLORIMET.	203 ROCHE	93	ROCHE	36	81.528	3.245	3.98	57.000	106.000	
2023-783	5	KINETIC 30	197 ARCHITECT	1	ABBOTT	11	76.182	2.949	3.87	53.000	99.000	
2023-783	6	KINETIC 37	165 VITROS	63	KODAK	144	97.669	6.517	6.67	68.000	127.000	
2023-783	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	10	65.600	5.834	8.89	46.000	85.000	
2023-783	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	30	66.448	7.238	10.89	47.000	86.000	
2023-783	10	IFCC	219 ATELLICA	143	SIEMENS	13	78.846	2.797	3.55	55.000	103.000	
2023-783	11	Modified I	38 SIEMENS DI	143	SIEMENS	89	84.409	4.738	5.61	59.000	110.000	
2023-783	11	Modified I	191 RX-SERIES	126	RANDOX	30	73.400	8.554	11.65	51.000	95.000	
2023-783	12	P NITRO PH	197 ARCHITECT	1	ABBOTT	12	73.750	4.361	5.91	52.000	96.000	
2023-784	3	COLORIMET.	184 INTEGRA	93	ROCHE	24	389.250	15.128	3.89	272.000	506.000	
2023-784	3	COLORIMET.	203 ROCHE	93	ROCHE	36	384.500	17.035	4.43	269.000	500.000	
2023-784	5	KINETIC 30	197 ARCHITECT	1	ABBOTT	11	357.818	15.741	4.40	250.000	465.000	
2023-784	6	KINETIC 37	165 VITROS	63	KODAK	144	225.835	13.858	6.14	158.000	294.000	
2023-784	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	10	309.200	29.363	9.50	216.000	402.000	
2023-784	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	30	317.690	34.117	10.74	222.000	413.000	
2023-784	10	IFCC	219 ATELLICA	143	SIEMENS	13	366.846	15.236	4.15	257.000	477.000	
2023-784	11	Modified I	38 SIEMENS DI	143	SIEMENS	89	405.356	12.940	3.19	284.000	527.000	
2023-784	11	Modified I	191 RX-SERIES	126	RANDOX	30	321.033	34.239	10.67	225.000	417.000	
2023-784	12	P NITRO PH	197 ARCHITECT	1	ABBOTT	12	346.500	19.530	5.64	243.000	450.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275		EVALUATION CRITERIA: TARGET VALUE 30 %								
2023-785	3	COLORIMET.	184 INTEGRA	93	ROCHE	24	214.250	8.447	3.94	150.000	279.000	
2023-785	3	COLORIMET.	203 ROCHE	93	ROCHE	36	211.500	8.578	4.06	148.000	275.000	
2023-785	5	KINETIC 30	197 ARCHITECT	1	ABBOTT	11	197.909	8.597	4.34	139.000	257.000	
2023-785	6	KINETIC 37	165 VITROS	63	KODAK	144	187.779	12.314	6.56	131.000	244.000	
2023-785	6	KINETIC 37	215 B.C DXC	134	BECKM.COUL	10	171.100	13.270	7.76	120.000	222.000	
2023-785	8	BOW. MC-37	186 BECK. COUL	76	AU-BECKMAN	30	174.533	18.931	10.85	122.000	227.000	
2023-785	10	IFCC	219 ATELLICA	143	SIEMENS	13	203.231	8.459	4.16	142.000	264.000	
2023-785	11	Modified I	38 SIEMENS DI	143	SIEMENS	89	221.511	7.401	3.34	155.000	288.000	
2023-785	11	Modified I	191 RX-SERIES	126	RANDOX	30	187.333	19.489	10.40	131.000	244.000	
2023-785	12	P NITRO PH	197 ARCHITECT	1	ABBOTT	12	191.917	10.594	5.52	134.000	249.000	
004	AMYLASE	CLIA TEST ID #: 285										
2023-781	3	ENZYM. 37C	165 VITROS	63	KODAK	85	141.353	5.208	3.68	99.000	184.000	
2023-781	3	ENZYM. 37C	215 B.C DXC	134	BECKM.COUL	10	212.500	5.084	2.39	149.000	276.000	
2023-781	4	HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	27	206.630	4.066	1.97	145.000	269.000	
2023-781	9	ENZ/COL.37	184 INTEGRA	93	ROCHE	10	202.100	5.873	2.91	141.000	263.000	
2023-781	9	ENZ/COL.37	203 ROCHE	93	ROCHE	25	198.800	4.569	2.30	139.000	258.000	
2023-781	10	CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	165.625	8.200	4.95	116.000	215.000	
2023-782	3	ENZYM. 37C	165 VITROS	63	KODAK	85	66.679	4.624	6.94	47.000	87.000	
2023-782	3	ENZYM. 37C	215 B.C DXC	134	BECKM.COUL	10	93.500	2.729	2.92	65.000	122.000	
2023-782	4	HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	27	88.704	1.717	1.94	62.000	115.000	
2023-782	9	ENZ/COL.37	184 INTEGRA	93	ROCHE	10	89.400	2.538	2.84	63.000	116.000	
2023-782	9	ENZ/COL.37	203 ROCHE	93	ROCHE	25	86.880	2.355	2.71	61.000	113.000	
2023-782	10	CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	71.750	3.614	5.04	50.000	93.000	
2023-783	3	ENZYM. 37C	165 VITROS	63	KODAK	85	50.188	3.912	7.80	35.000	65.000	
2023-783	3	ENZYM. 37C	215 B.C DXC	134	BECKM.COUL	10	65.800	1.400	2.13	46.000	86.000	
2023-783	4	HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	27	61.370	1.469	2.39	43.000	80.000	
2023-783	9	ENZ/COL.37	184 INTEGRA	93	ROCHE	10	63.200	2.891	4.57	44.000	82.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	9	ENZ/COL.37	203	ROCHE	93	ROCHE	25	61.160	1.666	2.72	43.000	80.000
2023-783	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	16	49.313	2.200	4.46	35.000	64.000
004	AMYLASE	CLIA TEST ID #: 285				EVALUATION CRITERIA: TARGET VALUE 30 %						
2023-784	3	ENZYM. 37C	165	VITROS	63	KODAK	85	185.071	7.550	4.08	130.000	241.000
2023-784	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	10	292.000	11.916	4.08	204.000	380.000
2023-784	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	27	283.852	5.140	1.81	199.000	369.000
2023-784	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	10	277.000	9.930	3.58	194.000	360.000
2023-784	9	ENZ/COL.37	203	ROCHE	93	ROCHE	25	273.120	9.271	3.39	191.000	355.000
2023-784	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	16	226.125	12.990	5.74	158.000	294.000
2023-785	3	ENZYM. 37C	165	VITROS	63	KODAK	85	109.541	5.304	4.84	77.000	142.000
2023-785	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	10	163.500	4.272	2.61	114.000	213.000
2023-785	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	27	157.037	3.636	2.32	110.000	204.000
2023-785	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	10	154.700	4.314	2.79	108.000	201.000
2023-785	9	ENZ/COL.37	203	ROCHE	93	ROCHE	25	151.040	3.538	2.34	106.000	196.000
2023-785	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	16	126.000	7.141	5.67	88.000	164.000
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295				EVALUATION CRITERIA: TARGET VALUE 20 %						
2023-781	1	ASPAR&KETO	165	VITROS	63	KODAK	141	402.935	20.927	5.19	322.000	484.000
2023-781	5	ENZYM. 37C	38	SIEMENS DI	143	SIEMENS	95	311.479	8.109	2.60	249.000	374.000
2023-781	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	297.500	6.561	2.21	238.000	357.000
2023-781	11	MOD. IFCC	184	INTEGRA	93	ROCHE	21	311.190	12.561	4.04	249.000	373.000
2023-781	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	30	272.828	11.945	4.38	218.000	327.000
2023-781	11	MOD. IFCC	191	RX-SERIES	126	RANDOX	37	365.459	24.191	6.62	292.000	439.000
2023-781	11	MOD. IFCC	203	ROCHE	93	ROCHE	36	317.750	14.215	4.47	254.000	381.000
2023-781	11	MOD. IFCC	219	ATELLICA	143	SIEMENS	13	350.538	10.825	3.09	280.000	421.000
2023-781	12	IFCC witho	203	ROCHE	93	ROCHE	12	319.500	10.452	3.27	256.000	383.000
2023-781	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	14	303.143	10.113	3.34	243.000	364.000
2023-782	1	ASPAR&KETO	165	VITROS	63	KODAK	141	160.928	6.996	4.35	129.000	193.000
2023-782	5	ENZYM. 37C	38	SIEMENS DI	143	SIEMENS	95	140.368	3.992	2.84	112.000	168.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	139.900	3.208	2.29	112.000	168.000
2023-782	11	MOD. IFCC	184	INTEGRA	93	ROCHE	21	142.571	5.395	3.78	114.000	171.000
2023-782	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	30	123.828	5.972	4.82	99.000	149.000
2023-782	11	MOD. IFCC	191	RX-SERIES	126	RANDOX	37	167.528	10.587	6.32	134.000	201.000
2023-782	11	MOD. IFCC	203	ROCHE	93	ROCHE	36	143.889	5.606	3.90	115.000	173.000
2023-782	11	MOD. IFCC	219	ATELLICA	143	SIEMENS	13	160.692	4.936	3.07	129.000	193.000
2023-782	12	IFCC witho	203	ROCHE	93	ROCHE	12	144.750	4.621	3.19	116.000	174.000
2023-782	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	14	136.643	4.700	3.44	109.000	164.000
005 ASP. AMINO AST/SGOT			CLIA TEST ID #: 295			EVALUATION CRITERIA: TARGET VALUE 20 %						
2023-783	1	ASPAR&KETO	165	VITROS	63	KODAK	141	112.144	4.540	4.05	90.000	135.000
2023-783	5	ENZYMA.37C	38	SIEMENS DI	143	SIEMENS	95	100.596	2.871	2.85	80.000	121.000
2023-783	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	101.800	3.219	3.16	81.000	122.000
2023-783	11	MOD. IFCC	184	INTEGRA	93	ROCHE	21	100.905	5.042	5.00	81.000	121.000
2023-783	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	30	88.931	3.999	4.50	71.000	107.000
2023-783	11	MOD. IFCC	191	RX-SERIES	126	RANDOX	37	119.757	6.740	5.63	96.000	144.000
2023-783	11	MOD. IFCC	203	ROCHE	93	ROCHE	36	102.722	4.100	3.99	82.000	123.000
2023-783	11	MOD. IFCC	219	ATELLICA	143	SIEMENS	13	114.846	4.258	3.71	92.000	138.000
2023-783	12	IFCC witho	203	ROCHE	93	ROCHE	12	104.000	3.082	2.96	83.000	125.000
2023-783	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	14	98.929	3.595	3.63	79.000	119.000
2023-784	1	ASPAR&KETO	165	VITROS	63	KODAK	141	585.297	30.950	5.29	468.000	702.000
2023-784	5	ENZYMA.37C	38	SIEMENS DI	143	SIEMENS	95	424.947	13.984	3.29	340.000	510.000
2023-784	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	397.500	17.333	4.36	318.000	477.000
2023-784	11	MOD. IFCC	184	INTEGRA	93	ROCHE	21	412.905	19.046	4.61	330.000	495.000
2023-784	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	30	363.483	16.128	4.44	291.000	436.000
2023-784	11	MOD. IFCC	191	RX-SERIES	126	RANDOX	37	478.892	41.211	8.61	383.000	575.000
2023-784	11	MOD. IFCC	203	ROCHE	93	ROCHE	36	421.056	17.026	4.04	337.000	505.000
2023-784	11	MOD. IFCC	219	ATELLICA	143	SIEMENS	13	454.231	16.154	3.56	363.000	545.000
2023-784	12	IFCC witho	203	ROCHE	93	ROCHE	12	426.167	14.346	3.37	341.000	511.000
2023-784	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	14	402.429	15.193	3.78	322.000	483.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY SUB: 10 ROUTINE GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
005 ASP. AMINO	AST/SGOT	CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE	20	%						
2023-785	1 ASPAR&KETO	165 VITROS	63 KODAK	141	292.971	13.634	4.65	234.000	352.000		
2023-785	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	95	240.147	7.217	3.01	192.000	288.000		
2023-785	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	10	235.500	6.469	2.75	188.000	283.000		
2023-785	11 MOD. IFCC	184 INTEGRA	93 ROCHE	21	241.286	10.398	4.31	193.000	290.000		
2023-785	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	30	212.667	12.429	5.84	170.000	255.000		
2023-785	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	282.243	16.355	5.79	226.000	339.000		
2023-785	11 MOD. IFCC	203 ROCHE	93 ROCHE	36	245.139	9.998	4.08	196.000	294.000		
2023-785	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	13	271.000	9.806	3.62	217.000	325.000		
2023-785	12 IFCC witho	203 ROCHE	93 ROCHE	12	245.917	8.321	3.38	197.000	295.000		
2023-785	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	14	233.357	8.482	3.63	187.000	280.000		
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305									
2023-781	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	94	4.949	0.210	4.25	3.960	5.940		
2023-781	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	28	4.700	0.118	2.51	3.760	5.640		
2023-781	4 J-G WITH B	191 RX-SERIES	126 RANDOX	27	5.192	0.412	7.94	4.150	6.230		
2023-781	4 J-G WITH B	191 RX-SERIES	126 RANDOX	27	4.760	0.501	10.53	3.810	5.710		
2023-781	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	10	5.113	0.226	4.41	4.090	6.140		
2023-781	7 DIAZO-DYPH	165 VITROS	63 KODAK	142	4.695	0.219	4.66	3.760	5.630		
2023-781	7 DIAZO-DYPH	165 VITROS	63 KODAK	142	4.378	0.076	1.74	3.500	5.250		
2023-781	13 AZOBIL.	184 INTEGRA	93 ROCHE	11	4.383	0.155	3.53	3.510	5.260		
2023-781	13 AZOBIL.	203 ROCHE	93 ROCHE	33	4.455	0.166	3.72	3.560	5.350		
2023-781	14 VANADATE	219 ATELLICA	143 SIEMENS	13	5.305	0.179	3.37	4.240	6.370		
2023-781	18 COLORIMETR	184 INTEGRA	93 ROCHE	11	4.497	0.237	5.27	3.600	5.400		
2023-781	18 COLORIMETR	203 ROCHE	93 ROCHE	14	4.424	0.173	3.91	3.540	5.310		
2023-781	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	4.891	0.438	8.95	3.910	5.870		
2023-782	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	94	1.835	0.094	5.11	1.470	2.200		
2023-782	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	28	1.821	0.060	3.32	1.460	2.190		
2023-782	4 J-G WITH B	191 RX-SERIES	126 RANDOX	27	1.752	0.237	13.55	1.400	2.100		
2023-782	4 J-G WITH B	191 RX-SERIES	126 RANDOX	27	1.942	0.191	9.84	1.550	2.330		

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	10	1.942	0.163	8.39	1.550	2.330
2023-782	7	DIAZO-DYPH	165	VITROS	63	KODAK	142	1.648	0.040	2.40	1.320	1.980
2023-782	7	DIAZO-DYPH	165	VITROS	63	KODAK	142	1.767	0.112	6.35	1.410	2.120
2023-782	13	AZOBIL.	184	INTEGRA	93	ROCHE	11	1.620	0.080	4.92	1.300	1.940
2023-782	13	AZOBIL.	203	ROCHE	93	ROCHE	33	1.626	0.054	3.35	1.300	1.950
2023-782	14	VANADATE	219	ATELLICA	143	SIEMENS	13	1.969	0.075	3.79	1.580	2.360
2023-782	18	COLORIMETR	184	INTEGRA	93	ROCHE	11	1.655	0.067	4.03	1.320	1.990
2023-782	18	COLORIMETR	203	ROCHE	93	ROCHE	14	1.613	0.092	5.70	1.290	1.940
2023-782	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	1.876	0.170	9.04	1.500	2.250
006 BILIRUBIN, TOTAL			CLIA TEST ID #: 305			EVALUATION CRITERIA: TARGET VALUE 20 %						
2023-783	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	94	1.089	0.068	6.26	0.870	1.310
2023-783	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	28	1.136	0.041	3.64	0.910	1.360
2023-783	4	J-G WITH B	191	RX-SERIES	126	RANDOX	27	1.199	0.089	7.40	0.960	1.440
2023-783	4	J-G WITH B	191	RX-SERIES	126	RANDOX	27	1.085	0.061	5.65	0.870	1.300
2023-783	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	10	1.306	0.110	8.41	1.040	1.570
2023-783	7	DIAZO-DYPH	165	VITROS	63	KODAK	142	1.079	0.082	7.59	0.860	1.300
2023-783	7	DIAZO-DYPH	165	VITROS	63	KODAK	142	1.030	0.043	4.18	0.820	1.240
2023-783	13	AZOBIL.	184	INTEGRA	93	ROCHE	11	0.967	0.052	5.36	0.770	1.160
2023-783	13	AZOBIL.	203	ROCHE	93	ROCHE	33	0.977	0.047	4.85	0.780	1.170
2023-783	14	VANADATE	219	ATELLICA	143	SIEMENS	13	1.203	0.030	2.51	0.960	1.440
2023-783	18	COLORIMETR	184	INTEGRA	93	ROCHE	11	0.995	0.058	5.80	0.800	1.190
2023-783	18	COLORIMETR	203	ROCHE	93	ROCHE	14	0.979	0.030	3.11	0.780	1.180
2023-783	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	1.155	0.103	8.92	0.920	1.390
2023-784	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	94	6.991	0.315	4.50	5.590	8.390
2023-784	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	28	6.436	0.228	3.54	5.150	7.720
2023-784	4	J-G WITH B	191	RX-SERIES	126	RANDOX	27	6.902	0.643	9.31	5.520	8.280
2023-784	4	J-G WITH B	191	RX-SERIES	126	RANDOX	27	7.278	0.554	7.61	5.820	8.730
2023-784	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	10	7.258	0.210	2.89	5.810	8.710
2023-784	7	DIAZO-DYPH	165	VITROS	63	KODAK	142	6.170	0.059	0.95	4.940	7.400

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-784	7	DIAZO-DYPH	165 VITROS	63 KODAK	142	6.596	0.298	4.52	5.280	7.910		
2023-784	13	AZOBIL.	184 INTEGRA	93 ROCHE	11	6.129	0.291	4.74	4.900	7.350		
2023-784	13	AZOBIL.	203 ROCHE	93 ROCHE	33	6.278	0.310	4.93	5.020	7.530		
2023-784	14	VANADATE	219 ATELLICA	143 SIEMENS	13	7.319	0.311	4.25	5.860	8.780		
2023-784	18	COLORIMETR	184 INTEGRA	93 ROCHE	11	6.244	0.379	6.07	4.990	7.490		
2023-784	18	COLORIMETR	203 ROCHE	93 ROCHE	14	6.231	0.206	3.30	4.980	7.480		
2023-784	19	DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	6.759	0.534	7.91	5.410	8.110		
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305		EVALUATION CRITERIA: TARGET VALUE 20 %								
2023-785	4	J-G WITH B	38 SIEMENS DI	143 SIEMENS	94	3.622	0.154	4.25	2.900	4.350		
2023-785	4	J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	28	3.515	0.122	3.48	2.810	4.220		
2023-785	4	J-G WITH B	191 RX-SERIES	126 RANDOX	27	3.793	0.333	8.77	3.030	4.550		
2023-785	4	J-G WITH B	191 RX-SERIES	126 RANDOX	27	3.563	0.419	11.77	2.850	4.280		
2023-785	5	J-G NO BK.	215 B.C DXC	134 BECKM.COUL	10	3.747	0.340	9.08	3.000	4.500		
2023-785	7	DIAZO-DYPH	165 VITROS	63 KODAK	142	3.441	0.162	4.72	2.750	4.130		
2023-785	7	DIAZO-DYPH	165 VITROS	63 KODAK	142	3.218	0.062	1.92	2.570	3.860		
2023-785	13	AZOBIL.	184 INTEGRA	93 ROCHE	11	3.198	0.138	4.32	2.560	3.840		
2023-785	13	AZOBIL.	203 ROCHE	93 ROCHE	33	3.265	0.148	4.55	2.610	3.920		
2023-785	14	VANADATE	219 ATELLICA	143 SIEMENS	13	3.855	0.111	2.88	3.080	4.630		
2023-785	18	COLORIMETR	184 INTEGRA	93 ROCHE	11	3.284	0.179	5.45	2.630	3.940		
2023-785	18	COLORIMETR	203 ROCHE	93 ROCHE	14	3.196	0.132	4.13	2.560	3.830		
2023-785	19	DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	3.557	0.320	9.00	2.850	4.270		
010 CALCIUM, TOTAL		CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE 1.00								
2023-781	3	ARSEN.-III	165 VITROS	63 KODAK	159	10.186	0.200	1.96	9.200	11.200		
2023-781	3	ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	32	10.616	0.235	2.21	9.600	11.600		
2023-781	3	ARSEN.-III	191 RX-SERIES	126 RANDOX	37	10.603	0.361	3.41	9.600	11.600		
2023-781	3	ARSEN.-III	197 ARCHITECT	1 ABBOTT	12	10.567	0.364	3.44	9.600	11.600		
2023-781	3	ARSEN.-III	220 Medica E.	99 STAMBIO	12	10.542	0.253	2.40	9.500	11.500		
2023-781	5	CALC-NONFL	197 ARCHITECT	1 ABBOTT	10	10.380	0.204	1.96	9.400	11.400		
2023-781	8	CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	97	10.280	0.280	2.72	9.300	11.300		
2023-781	8	CRESOLPHT.	184 INTEGRA	93 ROCHE	15	10.407	0.289	2.77	9.400	11.400		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-781	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	10.510	0.308	2.93	9.500	11.500
2023-781	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	13	10.585	0.196	1.85	9.600	11.600
2023-781	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	10	10.250	0.494	4.82	9.300	11.300
2023-781	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	10.425	0.275	2.64	9.400	11.400
2023-781	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	37	10.611	0.250	2.36	9.600	11.600
010	CALCIUM, TOTAL	CLIA TEST ID #: 345				EVALUATION CRITERIA: TARGET VALUE 1.00						
2023-782	3	ARSEN.-III	165	VITROS	63	KODAK	159	7.243	0.158	2.18	6.200	8.200
2023-782	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	32	7.503	0.174	2.32	6.500	8.500
2023-782	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	37	7.822	0.276	3.53	6.800	8.800
2023-782	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	7.383	0.186	2.52	6.400	8.400
2023-782	3	ARSEN.-III	220	Medica E.	99	STAMBIO	12	7.492	0.198	2.64	6.500	8.500
2023-782	5	CALC-NONFL	197	ARCHITECT	1	ABBOTT	10	7.340	0.156	2.13	6.300	8.300
2023-782	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	97	7.400	0.246	3.33	6.400	8.400
2023-782	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	15	7.480	0.180	2.40	6.500	8.500
2023-782	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	7.590	0.234	3.09	6.600	8.600
2023-782	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	13	7.415	0.135	1.82	6.400	8.400
2023-782	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	10	7.560	0.465	6.15	6.600	8.600
2023-782	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	7.469	0.196	2.62	6.500	8.500
2023-782	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	37	7.583	0.191	2.52	6.600	8.600
2023-783	3	ARSEN.-III	165	VITROS	63	KODAK	159	6.559	0.138	2.11	5.600	7.600
2023-783	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	32	6.781	0.157	2.32	5.800	7.800
2023-783	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	37	7.130	0.239	3.36	6.100	8.100
2023-783	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	6.725	0.169	2.51	5.700	7.700
2023-783	3	ARSEN.-III	220	Medica E.	99	STAMBIO	12	6.792	0.247	3.63	5.800	7.800
2023-783	5	CALC-NONFL	197	ARCHITECT	1	ABBOTT	10	6.690	0.221	3.31	5.700	7.700
2023-783	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	97	6.779	0.262	3.87	5.800	7.800
2023-783	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	15	6.793	0.195	2.87	5.800	7.800
2023-783	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	6.870	0.142	2.06	5.900	7.900
2023-783	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	13	6.669	0.173	2.59	5.700	7.700

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	10	6.970	0.458	6.58	6.000	8.000
2023-783	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	6.707	0.153	2.28	5.700	7.700
2023-783	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	37	6.917	0.219	3.17	5.900	7.900
010	CALCIUM, TOTAL		CLIA TEST ID #: 345			EVALUATION CRITERIA: TARGET VALUE 1.00						
2023-784	3	ARSEN.-III	165	VITROS	63	KODAK	159	11.877	0.232	1.95	10.900	12.900
2023-784	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	32	12.581	0.361	2.87	11.600	13.600
2023-784	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	37	12.527	0.368	2.94	11.500	13.500
2023-784	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	12.533	0.350	2.79	11.500	13.500
2023-784	3	ARSEN.-III	220	Medica E.	99	STAMBIO	12	12.408	0.284	2.29	11.400	13.400
2023-784	5	CALC-NONFL	197	ARCHITECT	1	ABBOTT	10	12.590	0.333	2.65	11.600	13.600
2023-784	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	97	12.236	0.321	2.62	11.200	13.200
2023-784	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	15	12.287	0.508	4.14	11.300	13.300
2023-784	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	12.370	0.424	3.43	11.400	13.400
2023-784	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	13	12.408	0.297	2.40	11.400	13.400
2023-784	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	10	12.070	0.551	4.57	11.100	13.100
2023-784	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	12.288	0.322	2.62	11.300	13.300
2023-784	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	37	12.530	0.405	3.23	11.500	13.500
2023-785	3	ARSEN.-III	165	VITROS	63	KODAK	159	9.056	0.181	2.00	8.100	10.100
2023-785	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	32	9.394	0.247	2.63	8.400	10.400
2023-785	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	37	9.574	0.192	2.00	8.600	10.600
2023-785	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	9.083	0.291	3.20	8.100	10.100
2023-785	3	ARSEN.-III	220	Medica E.	99	STAMBIO	12	9.233	0.301	3.26	8.200	10.200
2023-785	5	CALC-NONFL	197	ARCHITECT	1	ABBOTT	10	9.080	0.252	2.78	8.100	10.100
2023-785	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	97	9.084	0.248	2.74	8.100	10.100
2023-785	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	15	9.320	0.251	2.69	8.300	10.300
2023-785	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	9.410	0.202	2.15	8.400	10.400
2023-785	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	13	9.323	0.252	2.70	8.300	10.300
2023-785	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	10	9.250	0.461	4.98	8.300	10.300
2023-785	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	9.219	0.207	2.24	8.200	10.200

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023-785	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	37	9.389	0.223	2.38	8.400	10.400	
011	CHLORIDE	CLIA TEST ID #: 355		EVALUATION CRITERIA: TARGET VALUE 5 %									
2023-781	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	106.438	3.061	2.88	101.000	112.000	
2023-781	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	30	102.621	1.400	1.36	97.000	108.000	
2023-781	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	13	105.769	1.527	1.44	100.000	111.000	
2023-781	3	ISE DILUTE	203	ROCHE	93	ROCHE	51	103.216	2.108	2.04	98.000	108.000	
2023-781	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	105.400	1.200	1.14	100.000	111.000	
2023-781	4	ISE UND-DI	165	VITROS	63	KODAK	71	104.648	1.965	1.88	99.000	110.000	
2023-781	4	ISE UND-DI	165	VITROS	156	Kit 32	86	105.274	1.614	1.53	100.000	111.000	
2023-781	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	37	104.568	2.400	2.29	99.000	110.000	
2023-781	4	ISE UND-DI	220	Medica E.	130	MEDICA	14	108.286	2.962	2.74	103.000	114.000	
2023-781	5	MER NIT/DI	197	ARCHITECT	1	ABBOTT	11	106.182	1.527	1.44	101.000	111.000	
2023-781	8	IMT	219	ATELLICA	143	SIEMENS	13	107.538	1.824	1.70	102.000	113.000	
2023-781	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	105.957	1.883	1.78	101.000	111.000	
2023-782	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	93.094	2.777	2.98	88.000	98.000	
2023-782	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	30	91.433	1.647	1.80	87.000	96.000	
2023-782	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	13	94.692	1.435	1.52	90.000	99.000	
2023-782	3	ISE DILUTE	203	ROCHE	93	ROCHE	51	89.960	1.788	1.99	85.000	94.000	
2023-782	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	94.600	1.960	2.07	90.000	99.000	
2023-782	4	ISE UND-DI	165	VITROS	63	KODAK	71	93.620	1.755	1.87	89.000	98.000	
2023-782	4	ISE UND-DI	165	VITROS	156	Kit 32	86	93.687	1.456	1.55	89.000	98.000	
2023-782	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	37	91.730	2.511	2.74	87.000	96.000	
2023-782	4	ISE UND-DI	220	Medica E.	130	MEDICA	14	92.571	2.195	2.37	88.000	97.000	
2023-782	5	MER NIT/DI	197	ARCHITECT	1	ABBOTT	11	94.364	1.367	1.45	90.000	99.000	
2023-782	8	IMT	219	ATELLICA	143	SIEMENS	13	95.077	2.235	2.35	90.000	100.000	
2023-782	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	90.883	1.850	2.04	86.000	95.000	
2023-783	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	90.156	2.399	2.66	86.000	95.000	
2023-783	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	30	89.000	1.366	1.54	85.000	93.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	13	92.077	1.328	1.44	87.000	97.000
2023-783	3	ISE DILUTE	203	ROCHE	93	ROCHE	51	87.333	1.833	2.10	83.000	92.000
2023-783	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	92.500	1.432	1.55	88.000	97.000
2023-783	4	ISE UND-DI	165	VITROS	63	KODAK	71	91.174	1.372	1.50	87.000	96.000
2023-783	4	ISE UND-DI	165	VITROS	156	Kit 32	86	91.512	1.314	1.44	87.000	96.000
2023-783	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	37	88.865	2.418	2.72	84.000	93.000
2023-783	4	ISE UND-DI	220	Medica E.	130	MEDICA	14	89.429	1.801	2.01	85.000	94.000
2023-783	5	MER NIT/DI	197	ARCHITECT	1	ABBOTT	11	92.273	1.355	1.47	88.000	97.000
2023-783	8	IMT	219	ATELLICA	143	SIEMENS	13	92.538	2.098	2.27	88.000	97.000
2023-783	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	87.699	1.515	1.73	83.000	92.000
011	CHLORIDE	CLIA TEST ID #: 355		EVALUATION CRITERIA: TARGET VALUE 5 %								
2023-784	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	116.438	3.889	3.34	111.000	122.000
2023-784	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	30	110.067	2.032	1.85	105.000	116.000
2023-784	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	13	113.077	1.591	1.41	107.000	119.000
2023-784	3	ISE DILUTE	203	ROCHE	93	ROCHE	51	112.098	2.060	1.84	106.000	118.000
2023-784	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	113.700	4.496	3.95	108.000	119.000
2023-784	4	ISE UND-DI	165	VITROS	63	KODAK	71	113.286	1.806	1.59	108.000	119.000
2023-784	4	ISE UND-DI	165	VITROS	156	Kit 32	86	113.476	1.985	1.75	108.000	119.000
2023-784	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	37	113.973	3.349	2.94	108.000	120.000
2023-784	4	ISE UND-DI	220	Medica E.	130	MEDICA	14	118.929	2.789	2.35	113.000	125.000
2023-784	5	MER NIT/DI	197	ARCHITECT	1	ABBOTT	11	112.818	2.167	1.92	107.000	118.000
2023-784	8	IMT	219	ATELLICA	143	SIEMENS	13	113.538	1.737	1.53	108.000	119.000
2023-784	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	115.978	2.224	1.92	110.000	122.000
2023-785	3	ISE DILUTE	184	INTEGRA	93	ROCHE	32	102.813	2.888	2.81	98.000	108.000
2023-785	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	30	99.172	1.723	1.74	94.000	104.000
2023-785	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	13	102.385	1.862	1.82	97.000	108.000
2023-785	3	ISE DILUTE	203	ROCHE	93	ROCHE	51	98.745	1.979	2.00	94.000	104.000
2023-785	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	102.500	1.285	1.25	97.000	108.000
2023-785	4	ISE UND-DI	165	VITROS	63	KODAK	71	101.239	1.748	1.73	96.000	106.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-785	4	ISE UND-DI	165 VITROS	156	Kit 32	86	101.512	1.673	1.65	96.000	107.000	
2023-785	4	ISE UND-DI	191 RX-SERIES	126	RANDOX	37	100.784	3.041	3.02	96.000	106.000	
2023-785	4	ISE UND-DI	220 Medica E.	130	MEDICA	14	102.500	1.991	1.94	97.000	108.000	
2023-785	5	MER NIT/DI	197 ARCHITECT	1	ABBOTT	11	101.818	1.192	1.17	97.000	107.000	
2023-785	8	IMT	219 ATELLICA	143	SIEMENS	13	102.923	2.401	2.33	98.000	108.000	
2023-785	9	QUICK LYTE	38 SIEMENS DI	143	SIEMENS	94	100.376	1.579	1.57	95.000	105.000	
012	CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2023-781	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	171.344	5.667	3.31	154.000	188.000	
2023-781	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	170.500	2.598	1.52	153.000	188.000	
2023-781	1	ENZYMATIC	165 VITROS	63	KODAK	146	198.660	6.564	3.30	179.000	219.000	
2023-781	1	ENZYMATIC	184 INTEGRA	93	ROCHE	12	177.250	6.469	3.65	160.000	195.000	
2023-781	1	ENZYMATIC	191 RX-SERIES	126	RANDOX	43	197.093	7.220	3.66	177.000	217.000	
2023-781	1	ENZYMATIC	197 ARCHITECT	1	ABBOTT	11	187.727	3.107	1.66	169.000	207.000	
2023-781	1	ENZYMATIC	203 ROCHE	93	ROCHE	22	189.182	5.340	2.82	170.000	208.000	
2023-781	1	ENZYMATIC	215 B.C DXC	134	BECKM.COUL	10	187.300	3.926	2.10	169.000	206.000	
2023-781	1	ENZYMATIC	220 Medica E.	130	MEDICA	10	188.100	2.119	1.13	169.000	207.000	
2023-781	5	FE-SALTS-N	197 ARCHITECT	1	ABBOTT	12	185.667	4.230	2.28	167.000	204.000	
2023-781	8	ENZYMATIC	186 BECK. COUL	76	AU-BECKMAN	28	182.000	2.952	1.62	164.000	200.000	
2023-781	8	ENZYMATIC	203 ROCHE	93	ROCHE	13	189.500	2.255	1.19	171.000	208.000	
2023-781	9	ENZYMATIC	203 ROCHE	93	ROCHE	11	188.182	6.102	3.24	169.000	207.000	
2023-781	9	ENZYMATIC	219 ATELLICA	143	SIEMENS	13	192.308	4.513	2.35	173.000	212.000	
2023-782	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	126.250	1.090	0.86	114.000	139.000	
2023-782	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	127.286	4.565	3.59	115.000	140.000	
2023-782	1	ENZYMATIC	165 VITROS	63	KODAK	146	136.608	4.223	3.09	123.000	150.000	
2023-782	1	ENZYMATIC	184 INTEGRA	93	ROCHE	12	131.167	5.550	4.23	118.000	144.000	
2023-782	1	ENZYMATIC	191 RX-SERIES	126	RANDOX	43	145.214	5.040	3.47	131.000	160.000	
2023-782	1	ENZYMATIC	197 ARCHITECT	1	ABBOTT	11	138.091	2.539	1.84	124.000	152.000	
2023-782	1	ENZYMATIC	203 ROCHE	93	ROCHE	22	139.227	3.450	2.48	125.000	153.000	
2023-782	1	ENZYMATIC	215 B.C DXC	134	BECKM.COUL	10	137.200	2.713	1.98	123.000	151.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	1	ENZYMATIC	220	Medica E.	130	MEDICA	10	140.400	2.835	2.02	126.000	154.000
2023-782	5	FE-SALTS-N	197	ARCHITECT	1	ABBOTT	12	137.583	3.861	2.81	124.000	151.000
2023-782	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	28	133.714	2.889	2.16	120.000	147.000
2023-782	8	ENZYMATIC	203	ROCHE	93	ROCHE	13	139.538	4.668	3.35	126.000	153.000
2023-782	9	ENZYMATIC	203	ROCHE	93	ROCHE	11	137.727	3.544	2.57	124.000	152.000
2023-782	9	ENZYMATIC	219	ATELLICA	143	SIEMENS	13	139.308	3.930	2.82	125.000	153.000
012 CHOLESTEROL, TOTAL			CLIA TEST ID #: 365			EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-783	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	92	117.728	4.230	3.59	106.000	130.000
2023-783	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	92	117.250	4.815	4.11	106.000	129.000
2023-783	1	ENZYMATIC	165	VITROS	63	KODAK	146	123.465	4.264	3.45	111.000	136.000
2023-783	1	ENZYMATIC	184	INTEGRA	93	ROCHE	12	120.583	5.880	4.88	109.000	133.000
2023-783	1	ENZYMATIC	191	RX-SERIES	126	RANDOX	43	134.326	4.954	3.69	121.000	148.000
2023-783	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	11	127.909	2.429	1.90	115.000	141.000
2023-783	1	ENZYMATIC	203	ROCHE	93	ROCHE	22	127.773	3.605	2.82	115.000	141.000
2023-783	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	10	125.400	3.072	2.45	113.000	138.000
2023-783	1	ENZYMATIC	220	Medica E.	130	MEDICA	10	129.400	2.691	2.08	116.000	142.000
2023-783	5	FE-SALTS-N	197	ARCHITECT	1	ABBOTT	12	126.583	3.707	2.93	114.000	139.000
2023-783	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	28	122.821	2.578	2.10	111.000	135.000
2023-783	8	ENZYMATIC	203	ROCHE	93	ROCHE	13	127.250	2.046	1.61	115.000	140.000
2023-783	9	ENZYMATIC	203	ROCHE	93	ROCHE	11	127.000	3.693	2.91	114.000	140.000
2023-783	9	ENZYMATIC	219	ATELLICA	143	SIEMENS	13	127.077	4.305	3.39	114.000	140.000
2023-784	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	92	195.000	5.339	2.74	176.000	215.000
2023-784	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	92	198.769	7.591	3.82	179.000	219.000
2023-784	1	ENZYMATIC	165	VITROS	63	KODAK	146	237.333	8.976	3.78	214.000	261.000
2023-784	1	ENZYMATIC	184	INTEGRA	93	ROCHE	12	205.667	10.734	5.22	185.000	226.000
2023-784	1	ENZYMATIC	191	RX-SERIES	126	RANDOX	43	231.047	9.895	4.28	208.000	254.000
2023-784	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	11	219.364	4.715	2.15	197.000	241.000
2023-784	1	ENZYMATIC	203	ROCHE	93	ROCHE	22	220.952	4.695	2.13	199.000	243.000
2023-784	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	10	222.100	9.803	4.41	200.000	244.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-784	1	ENZYMATIC	220 Medica E.	130	MEDICA	10	222.400	4.499	2.02	200.000	245.000	
2023-784	5	FE-SALTS-N	197 ARCHITECT	1	ABBOTT	12	216.417	6.639	3.07	195.000	238.000	
2023-784	8	ENZYMATIC	186 BECK. COUL	76	AU-BECKMAN	28	212.357	3.667	1.73	191.000	234.000	
2023-784	8	ENZYMATIC	203 ROCHE	93	ROCHE	13	220.000	8.832	4.01	198.000	242.000	
2023-784	9	ENZYMATIC	203 ROCHE	93	ROCHE	11	221.364	8.003	3.62	199.000	244.000	
2023-784	9	ENZYMATIC	219 ATELLICA	143	SIEMENS	13	220.923	7.032	3.18	199.000	243.000	
012	CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2023-785	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	153.000	5.103	3.34	138.000	168.000	
2023-785	1	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	92	153.500	3.202	2.09	138.000	169.000	
2023-785	1	ENZYMATIC	165 VITROS	63	KODAK	146	173.902	5.956	3.42	157.000	191.000	
2023-785	1	ENZYMATIC	184 INTEGRA	93	ROCHE	12	159.750	7.462	4.67	144.000	176.000	
2023-785	1	ENZYMATIC	191 RX-SERIES	126	RANDOX	43	179.116	6.949	3.88	161.000	197.000	
2023-785	1	ENZYMATIC	197 ARCHITECT	1	ABBOTT	11	166.091	2.021	1.22	149.000	183.000	
2023-785	1	ENZYMATIC	203 ROCHE	93	ROCHE	22	169.455	4.765	2.81	153.000	186.000	
2023-785	1	ENZYMATIC	215 B.C DXC	134	BECKM.COUL	10	162.500	20.151	12.40	146.000	179.000	
2023-785	1	ENZYMATIC	220 Medica E.	130	MEDICA	10	170.900	3.176	1.86	154.000	188.000	
2023-785	5	FE-SALTS-N	197 ARCHITECT	1	ABBOTT	12	166.333	4.230	2.54	150.000	183.000	
2023-785	8	ENZYMATIC	186 BECK. COUL	76	AU-BECKMAN	28	163.889	4.306	2.63	148.000	180.000	
2023-785	8	ENZYMATIC	203 ROCHE	93	ROCHE	13	168.250	2.861	1.70	151.000	185.000	
2023-785	9	ENZYMATIC	203 ROCHE	93	ROCHE	11	168.091	5.435	3.23	151.000	185.000	
2023-785	9	ENZYMATIC	219 ATELLICA	143	SIEMENS	13	170.769	5.206	3.05	154.000	188.000	
013	CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2023-781	11	ENC.-COLO.	165 VITROS	63	KODAK	143	59.399	3.619	6.09	41.600	77.200	
2023-781	11	ENC.-COLO.	184 INTEGRA	93	ROCHE	30	54.776	2.777	5.07	38.300	71.200	
2023-781	11	ENC.-COLO.	186 BECK. COUL	76	AU-BECKMAN	29	69.569	4.360	6.27	48.700	90.400	
2023-781	11	ENC.-COLO.	191 RX-SERIES	126	RANDOX	43	58.800	6.542	11.13	41.200	76.400	
2023-781	11	ENC.-COLO.	203 ROCHE	93	ROCHE	40	55.088	2.917	5.30	38.600	71.600	
2023-781	11	ENC.-COLO.	219 ATELLICA	143	SIEMENS	13	72.938	2.482	3.40	51.100	94.800	
2023-781	11	ENC.-COLO.	220 Medica E.	130	MEDICA	10	62.700	2.100	3.35	43.900	81.500	
2023-781	12	HOM-LQ.S.D	197 ARCHITECT	1	ABBOTT	15	66.567	3.737	5.61	46.600	86.500	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-781	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	10	66.010	2.939	4.45	46.200	85.800
2023-781	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	89	44.045	2.996	6.80	30.800	57.300
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375				EVALUATION CRITERIA: TARGET VALUE 30 %						
2023-782	11	ENC.-COLO.	165	VITROS	63	KODAK	143	43.129	2.652	6.15	30.200	56.100
2023-782	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	30	47.797	2.624	5.49	33.500	62.100
2023-782	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	29	53.038	3.178	5.99	37.100	68.900
2023-782	11	ENC.-COLO.	191	RX-SERIES	126	RANDOX	43	46.691	4.781	10.24	32.700	60.700
2023-782	11	ENC.-COLO.	203	ROCHE	93	ROCHE	40	47.743	2.338	4.90	33.400	62.100
2023-782	11	ENC.-COLO.	219	APELLICA	143	SIEMENS	13	56.169	2.314	4.12	39.300	73.000
2023-782	11	ENC.-COLO.	220	Medica E.	130	MEDICA	10	48.000	1.949	4.06	33.600	62.400
2023-782	12	HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	15	51.907	2.631	5.07	36.300	67.500
2023-782	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	10	50.600	3.822	7.55	35.400	65.800
2023-782	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	89	42.872	2.061	4.81	30.000	55.700
2023-783	11	ENC.-COLO.	165	VITROS	63	KODAK	143	40.467	2.213	5.47	28.300	52.600
2023-783	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	30	45.713	2.637	5.77	32.000	59.400
2023-783	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	29	49.307	3.188	6.47	34.500	64.100
2023-783	11	ENC.-COLO.	191	RX-SERIES	126	RANDOX	43	43.470	4.962	11.41	30.400	56.500
2023-783	11	ENC.-COLO.	203	ROCHE	93	ROCHE	40	45.780	2.193	4.79	32.000	59.500
2023-783	11	ENC.-COLO.	219	APELLICA	143	SIEMENS	13	52.262	1.903	3.64	36.600	67.900
2023-783	11	ENC.-COLO.	220	Medica E.	130	MEDICA	10	44.500	1.565	3.52	31.200	57.900
2023-783	12	HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	15	48.347	2.263	4.68	33.800	62.900
2023-783	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	10	47.210	4.980	10.55	33.000	61.400
2023-783	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	89	42.357	2.107	4.97	29.700	55.100
2023-784	11	ENC.-COLO.	165	VITROS	63	KODAK	143	72.973	4.439	6.08	51.100	94.900
2023-784	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	30	58.431	2.946	5.04	40.900	76.000
2023-784	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	29	79.414	4.719	5.94	55.600	103.200
2023-784	11	ENC.-COLO.	191	RX-SERIES	126	RANDOX	43	66.623	6.714	10.08	46.600	86.600
2023-784	11	ENC.-COLO.	203	ROCHE	93	ROCHE	40	58.603	3.053	5.21	41.000	76.200

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-784	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS	13	82.485	3.392	4.11	57.700	107.200			
2023-784	11 ENC.-COLO.	220 Medica E.	130 MEDICA	10	73.900	2.343	3.17	51.700	96.100			
2023-784	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	15	75.620	4.478	5.92	52.900	98.300			
2023-784	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	10	74.520	6.314	8.47	52.200	96.900			
2023-784	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	89	44.486	2.887	6.49	31.100	57.800			
013	CHOLESTEROL, (HDL)	CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %									
2023-785	11 ENC.-COLO.	165 VITROS	63 KODAK	143	52.535	3.206	6.10	36.800	68.300			
2023-785	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	30	51.669	3.107	6.01	36.200	67.200			
2023-785	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	29	62.786	3.492	5.56	44.000	81.600			
2023-785	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	43	54.505	6.154	11.29	38.200	70.900			
2023-785	11 ENC.-COLO.	203 ROCHE	93 ROCHE	40	52.344	2.734	5.22	36.600	68.000			
2023-785	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS	13	66.138	2.551	3.86	46.300	86.000			
2023-785	11 ENC.-COLO.	220 Medica E.	130 MEDICA	10	56.500	2.579	4.56	39.600	73.500			
2023-785	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	15	60.347	2.406	3.99	42.200	78.500			
2023-785	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	10	61.130	2.631	4.30	42.800	79.500			
2023-785	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	89	44.469	2.267	5.10	31.100	57.800			
014	CREATINE KINASE (CK)	CLIA TEST ID #: 385										
2023-781	3 KINETIC 37	165 VITROS	63 KODAK	74	355.521	18.324	5.15	249.000	462.000			
2023-781	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	17	441.647	33.100	7.49	309.000	574.000			
2023-781	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	489.500	10.269	2.10	343.000	636.000			
2023-781	5 SZASZ-MOD.	219 ATELLICA	143 SIEMENS	11	499.727	15.886	3.18	350.000	650.000			
2023-781	8 Hexokinase	203 ROCHE	93 ROCHE	11	483.182	13.009	2.69	338.000	628.000			
2023-781	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	477.552	8.439	1.77	334.000	621.000			
2023-781	10 ENZYMATIC-	203 ROCHE	93 ROCHE	10	484.400	13.908	2.87	339.000	630.000			
2023-782	3 KINETIC 37	165 VITROS	63 KODAK	74	179.444	12.020	6.70	126.000	233.000			
2023-782	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	17	178.412	11.607	6.51	125.000	232.000			
2023-782	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	204.800	4.490	2.19	143.000	266.000			
2023-782	5 SZASZ-MOD.	219 ATELLICA	143 SIEMENS	11	203.636	6.139	3.01	143.000	265.000			
2023-782	8 Hexokinase	203 ROCHE	93 ROCHE	11	199.545	3.474	1.74	140.000	259.000			

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	29	194.931	4.719	2.42	136.000	253.000
2023-782	10	ENZYMATIC-	203	ROCHE	93	ROCHE	10	197.200	4.445	2.25	138.000	256.000
014 CREATINE KINASE (CK)		CLIA TEST ID #: 385				EVALUATION CRITERIA: TARGET VALUE 30 %						
2023-783	3	KINETIC 37	165	VITROS	63	KODAK	74	120.153	8.120	6.76	84.000	156.000
2023-783	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	17	115.647	6.919	5.98	81.000	150.000
2023-783	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	10	135.000	2.236	1.66	95.000	176.000
2023-783	5	SZASZ-MOD.	219	ATELLICA	143	SIEMENS	11	132.909	5.107	3.84	93.000	173.000
2023-783	8	Hexokinase	203	ROCHE	93	ROCHE	11	130.091	2.678	2.06	91.000	169.000
2023-783	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	29	127.241	3.115	2.45	89.000	165.000
2023-783	10	ENZYMATIC-	203	ROCHE	93	ROCHE	10	129.700	3.770	2.91	91.000	169.000
2023-784	3	KINETIC 37	165	VITROS	63	KODAK	74	425.676	23.301	5.47	298.000	553.000
2023-784	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	17	604.706	44.009	7.28	423.000	786.000
2023-784	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	10	668.900	17.813	2.66	468.000	870.000
2023-784	5	SZASZ-MOD.	219	ATELLICA	143	SIEMENS	11	671.091	15.353	2.29	470.000	872.000
2023-784	8	Hexokinase	203	ROCHE	93	ROCHE	11	665.818	24.165	3.63	466.000	866.000
2023-784	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	29	656.310	15.918	2.43	459.000	853.000
2023-784	10	ENZYMATIC-	203	ROCHE	93	ROCHE	10	669.700	20.421	3.05	469.000	871.000
2023-785	3	KINETIC 37	165	VITROS	63	KODAK	74	293.592	15.039	5.12	206.000	382.000
2023-785	3	KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	17	332.588	23.053	6.93	233.000	432.000
2023-785	4	OLIVER ROS	215	B.C DXC	134	BECKM.COUL	10	358.500	42.119	11.75	251.000	466.000
2023-785	5	SZASZ-MOD.	219	ATELLICA	143	SIEMENS	11	373.364	8.510	2.28	261.000	485.000
2023-785	8	Hexokinase	203	ROCHE	93	ROCHE	11	366.909	10.317	2.81	257.000	477.000
2023-785	9	SIEMENS	38	SIEMENS DI	143	SIEMENS	29	358.897	7.378	2.06	251.000	467.000
2023-785	10	ENZYMATIC-	203	ROCHE	93	ROCHE	10	360.700	7.695	2.13	252.000	469.000
016 CREATININE		CLIA TEST ID #: 405				EVALUATION CRITERIA: TARGET VALUE 10 % OR 0.30 {GREATER}						
2023-781	2	AK. PIC. N	191	RX-SERIES	126	RANDOX	30	4.447	0.202	4.54	4.000	4.890
2023-781	4	ENZYMATIC	203	ROCHE	93	ROCHE	10	5.330	0.206	3.87	4.800	5.860
2023-781	4	ENZYMATIC	220	Medica E.	130	MEDICA	13	5.597	0.180	3.22	5.040	6.160

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2023-781	6 KIN. ALK P	184 INTEGRA	93 ROCHE	16	5.101	0.267	5.23	4.590	5.610			
2023-781	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	32	5.343	0.185	3.47	4.810	5.880			
2023-781	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	5.869	0.191	3.25	5.280	6.460			
2023-781	6 KIN. ALK P	203 ROCHE	93 ROCHE	28	5.271	0.168	3.19	4.740	5.800			
2023-781	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	10	5.421	0.089	1.64	4.880	5.960			
2023-781	8 IDMS-CREA	165 VITROS	63 KODAK	157	4.953	0.158	3.20	4.460	5.450			
2023-781	9 MODIFIED	219 ATELLICA	143 SIEMENS	13	5.318	0.134	2.53	4.790	5.850			
2023-781	10 JAFFE COMP	184 INTEGRA	93 ROCHE	10	5.260	0.107	2.03	4.730	5.790			
2023-781	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	89	5.515	0.152	2.75	4.960	6.070			
016 CREATININE		CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 10 % OR 0.30 {GREATER}									
2023-782	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	2.234	0.146	6.53	1.930	2.530			
2023-782	4 ENZYMATIC	203 ROCHE	93 ROCHE	10	2.599	0.084	3.23	2.300	2.900			
2023-782	4 ENZYMATIC	220 Medica E.	130 MEDICA	13	2.630	0.051	1.94	2.330	2.930			
2023-782	6 KIN. ALK P	184 INTEGRA	93 ROCHE	16	2.559	0.122	4.78	2.260	2.860			
2023-782	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	32	2.540	0.085	3.36	2.240	2.840			
2023-782	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	2.646	0.071	2.70	2.350	2.950			
2023-782	6 KIN. ALK P	203 ROCHE	93 ROCHE	28	2.589	0.068	2.64	2.290	2.890			
2023-782	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	10	2.611	0.079	3.01	2.310	2.910			
2023-782	8 IDMS-CREA	165 VITROS	63 KODAK	157	2.493	0.088	3.54	2.190	2.790			
2023-782	9 MODIFIED	219 ATELLICA	143 SIEMENS	13	2.505	0.084	3.35	2.200	2.800			
2023-782	10 JAFFE COMP	184 INTEGRA	93 ROCHE	10	2.622	0.093	3.56	2.320	2.920			
2023-782	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	89	2.674	0.081	3.04	2.370	2.970			
2023-783	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	1.731	0.121	6.98	1.430	2.030			
2023-783	4 ENZYMATIC	203 ROCHE	93 ROCHE	10	1.936	0.069	3.55	1.640	2.240			
2023-783	4 ENZYMATIC	220 Medica E.	130 MEDICA	13	1.958	0.038	1.96	1.660	2.260			
2023-783	6 KIN. ALK P	184 INTEGRA	93 ROCHE	16	1.895	0.094	4.98	1.600	2.200			
2023-783	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	32	1.877	0.072	3.81	1.580	2.180			
2023-783	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	1.896	0.048	2.55	1.600	2.200			
2023-783	6 KIN. ALK P	203 ROCHE	93 ROCHE	28	1.932	0.060	3.11	1.630	2.230			

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	10	1.933	0.034	1.75	1.630	2.230
2023-783	8	IDMS-CREA	165	VITROS	63	KODAK	157	1.843	0.063	3.40	1.540	2.140
2023-783	9	MODIFIED	219	ATELLICA	143	SIEMENS	13	1.842	0.043	2.35	1.540	2.140
2023-783	10	JAFFE COMP	184	INTEGRA	93	ROCHE	10	1.957	0.053	2.72	1.660	2.260
2023-783	11	IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	89	1.950	0.071	3.66	1.650	2.250
016 CREATININE			CLIA TEST ID #: 405			EVALUATION CRITERIA: TARGET VALUE 10 % OR 0.30 {GREATER}						
2023-784	2	AK. PIC. N	191	RX-SERIES	126	RANDOX	30	5.799	0.268	4.63	5.220	6.380
2023-784	4	ENZYMATIC	203	ROCHE	93	ROCHE	10	6.877	0.285	4.15	6.190	7.560
2023-784	4	ENZYMATIC	220	Medica E.	130	MEDICA	13	7.508	0.186	2.48	6.760	8.260
2023-784	6	KIN. ALK P	184	INTEGRA	93	ROCHE	16	6.535	0.448	6.85	5.880	7.190
2023-784	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	32	7.025	0.301	4.29	6.320	7.730
2023-784	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	18	7.892	0.213	2.70	7.100	8.680
2023-784	6	KIN. ALK P	203	ROCHE	93	ROCHE	28	6.830	0.295	4.32	6.150	7.510
2023-784	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	10	6.673	1.543	23.12	6.010	7.340
2023-784	8	IDMS-CREA	165	VITROS	63	KODAK	157	6.304	0.192	3.05	5.670	6.930
2023-784	9	MODIFIED	219	ATELLICA	143	SIEMENS	13	6.985	0.173	2.47	6.290	7.680
2023-784	10	JAFFE COMP	184	INTEGRA	93	ROCHE	10	6.685	0.160	2.39	6.020	7.350
2023-784	11	IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	89	7.421	0.217	2.92	6.680	8.160
2023-785	2	AK. PIC. N	191	RX-SERIES	126	RANDOX	30	3.537	0.226	6.39	3.180	3.890
2023-785	4	ENZYMATIC	203	ROCHE	93	ROCHE	10	4.219	0.140	3.32	3.800	4.640
2023-785	4	ENZYMATIC	220	Medica E.	130	MEDICA	13	4.354	0.122	2.79	3.920	4.790
2023-785	6	KIN. ALK P	184	INTEGRA	93	ROCHE	16	4.099	0.219	5.35	3.690	4.510
2023-785	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	32	4.196	0.162	3.87	3.780	4.620
2023-785	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	18	4.489	0.085	1.88	4.040	4.940
2023-785	6	KIN. ALK P	203	ROCHE	93	ROCHE	28	4.174	0.166	3.98	3.760	4.590
2023-785	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	10	3.997	0.967	24.19	3.600	4.400
2023-785	8	IDMS-CREA	165	VITROS	63	KODAK	157	3.989	0.125	3.13	3.590	4.390
2023-785	9	MODIFIED	219	ATELLICA	143	SIEMENS	13	4.122	0.092	2.23	3.710	4.530
2023-785	10	JAFFE COMP	184	INTEGRA	93	ROCHE	10	4.173	0.097	2.33	3.760	4.590

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023-785	11	IDMS	Jaffe	38	SIEMENS DI	143	SIEMENS	89	4.300	0.111	2.58	3.870	4.730
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}							
2023-781	2	GL. OXI. E		215	B.C DXC	134	BECKM.COUL	10	128.200	2.676	2.09	115.000	141.000
2023-781	3	GL. OXI. C		165	VITROS	63	KODAK	160	134.652	2.942	2.19	121.000	148.000
2023-781	3	GL. OXI. C		165	VITROS	63	KODAK	160	130.556	3.304	2.53	118.000	144.000
2023-781	3	GL. OXI. C		213	SELECTRA P	150	ELITECH	10	126.000	3.376	2.68	113.000	139.000
2023-781	5	HEXOKINASE		38	SIEMENS DI	143	SIEMENS	97	140.340	3.152	2.25	126.000	154.000
2023-781	5	HEXOKINASE		184	INTEGRA	93	ROCHE	10	133.700	4.124	3.08	120.000	147.000
2023-781	5	HEXOKINASE		186	BECK. COUL	76	AU-BECKMAN	32	130.125	2.631	2.02	117.000	143.000
2023-781	5	HEXOKINASE		191	RX-SERIES	126	RANDOX	32	133.839	3.446	2.57	120.000	147.000
2023-781	5	HEXOKINASE		197	ARCHITECT	1	ABBOTT	10	128.900	2.844	2.21	116.000	142.000
2023-781	5	HEXOKINASE		203	ROCHE	93	ROCHE	28	134.071	3.251	2.42	121.000	147.000
2023-781	6	HEXOKI. CO		220	Medica E.	130	MEDICA	13	138.923	3.269	2.35	125.000	153.000
2023-781	12	HEXOKINASE		184	INTEGRA	93	ROCHE	21	131.429	4.614	3.51	118.000	145.000
2023-781	12	HEXOKINASE		197	ARCHITECT	1	ABBOTT	14	129.071	3.807	2.95	116.000	142.000
2023-781	12	HEXOKINASE		203	ROCHE	93	ROCHE	22	131.591	4.469	3.40	118.000	145.000
2023-781	12	HEXOKINASE		219	APELLICA	143	SIEMENS	13	131.250	1.233	0.94	118.000	144.000
2023-782	2	GL. OXI. E		215	B.C DXC	134	BECKM.COUL	10	70.400	2.375	3.37	63.000	77.000
2023-782	3	GL. OXI. C		165	VITROS	63	KODAK	160	72.222	1.988	2.75	65.000	79.000
2023-782	3	GL. OXI. C		165	VITROS	63	KODAK	160	71.525	2.454	3.43	64.000	79.000
2023-782	3	GL. OXI. C		213	SELECTRA P	150	ELITECH	10	70.700	3.068	4.34	64.000	78.000
2023-782	5	HEXOKINASE		38	SIEMENS DI	143	SIEMENS	97	77.588	2.024	2.61	70.000	85.000
2023-782	5	HEXOKINASE		184	INTEGRA	93	ROCHE	10	73.100	1.700	2.33	66.000	80.000
2023-782	5	HEXOKINASE		186	BECK. COUL	76	AU-BECKMAN	32	71.469	1.837	2.57	64.000	79.000
2023-782	5	HEXOKINASE		191	RX-SERIES	126	RANDOX	32	75.906	2.673	3.52	68.000	83.000
2023-782	5	HEXOKINASE		197	ARCHITECT	1	ABBOTT	10	70.400	1.685	2.39	63.000	77.000
2023-782	5	HEXOKINASE		203	ROCHE	93	ROCHE	28	72.607	1.589	2.19	65.000	80.000
2023-782	6	HEXOKI. CO		220	Medica E.	130	MEDICA	13	76.462	2.678	3.50	69.000	84.000
2023-782	12	HEXOKINASE		184	INTEGRA	93	ROCHE	21	72.048	1.963	2.73	65.000	79.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	14	70.143	1.807	2.58	63.000	77.000
2023-782	12	HEXOKINASE	203	ROCHE	93	ROCHE	22	72.045	3.022	4.19	65.000	79.000
2023-782	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	13	72.250	1.010	1.40	65.000	79.000
017	GLUCOSE		CLIA TEST ID #: 415			EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2023-783	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	10	57.200	1.249	2.18	51.000	63.000
2023-783	3	GL. OXI. C	165	VITROS	63	KODAK	160	56.369	1.424	2.53	50.000	62.000
2023-783	3	GL. OXI. C	165	VITROS	63	KODAK	160	58.222	2.250	3.86	52.000	64.000
2023-783	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	10	57.000	2.966	5.20	51.000	63.000
2023-783	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	97	62.979	1.839	2.92	57.000	69.000
2023-783	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	59.400	2.289	3.85	53.000	65.000
2023-783	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	32	57.813	1.570	2.72	52.000	64.000
2023-783	5	HEXOKINASE	191	RX-SERIES	126	RANDOX	32	62.750	2.761	4.40	56.000	69.000
2023-783	5	HEXOKINASE	197	ARCHITECT	1	ABBOTT	10	56.600	1.200	2.12	51.000	63.000
2023-783	5	HEXOKINASE	203	ROCHE	93	ROCHE	28	58.857	1.641	2.79	53.000	65.000
2023-783	6	HEXOKI. CO	220	Medica E.	130	MEDICA	13	61.462	1.646	2.68	55.000	68.000
2023-783	12	HEXOKINASE	184	INTEGRA	93	ROCHE	21	58.000	1.225	2.11	52.000	64.000
2023-783	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	14	56.214	1.423	2.53	50.000	62.000
2023-783	12	HEXOKINASE	203	ROCHE	93	ROCHE	22	58.143	1.726	2.97	52.000	64.000
2023-783	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	13	58.917	0.954	1.62	53.000	65.000
2023-784	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	10	166.000	6.066	3.65	149.000	183.000
2023-784	3	GL. OXI. C	165	VITROS	63	KODAK	160	168.889	3.446	2.04	152.000	186.000
2023-784	3	GL. OXI. C	165	VITROS	63	KODAK	160	177.120	4.497	2.54	159.000	195.000
2023-784	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	10	162.000	6.450	3.98	146.000	178.000
2023-784	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	97	180.813	4.938	2.73	163.000	199.000
2023-784	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	170.900	5.576	3.26	154.000	188.000
2023-784	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	32	167.313	3.754	2.24	151.000	184.000
2023-784	5	HEXOKINASE	191	RX-SERIES	126	RANDOX	32	171.000	6.315	3.69	154.000	188.000
2023-784	5	HEXOKINASE	197	ARCHITECT	1	ABBOTT	10	167.800	2.750	1.64	151.000	185.000
2023-784	5	HEXOKINASE	203	ROCHE	93	ROCHE	28	173.679	3.752	2.16	156.000	191.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-784	6	HEXOKI. CO	220	Medica E.	130	MEDICA	13	180.077	6.753	3.75	162.000	198.000
2023-784	12	HEXOKINASE	184	INTEGRA	93	ROCHE	21	169.286	6.467	3.82	152.000	186.000
2023-784	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	14	166.429	4.515	2.71	150.000	183.000
2023-784	12	HEXOKINASE	203	ROCHE	93	ROCHE	22	169.136	7.400	4.37	152.000	186.000
2023-784	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	13	167.231	5.659	3.38	151.000	184.000
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2023-785	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	10	104.300	2.369	2.27	94.000	115.000
2023-785	3	GL. OXI. C	165	VITROS	63	KODAK	160	108.405	2.397	2.21	98.000	119.000
2023-785	3	GL. OXI. C	165	VITROS	63	KODAK	160	106.333	2.749	2.59	96.000	117.000
2023-785	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	10	103.200	2.821	2.73	93.000	114.000
2023-785	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	97	113.979	2.585	2.27	103.000	125.000
2023-785	5	HEXOKINASE	184	INTEGRA	93	ROCHE	10	110.000	4.050	3.68	99.000	121.000
2023-785	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	32	105.774	2.524	2.39	95.000	116.000
2023-785	5	HEXOKINASE	191	RX-SERIES	126	RANDOX	32	110.813	4.592	4.14	100.000	122.000
2023-785	5	HEXOKINASE	197	ARCHITECT	1	ABBOTT	10	104.500	2.110	2.02	94.000	115.000
2023-785	5	HEXOKINASE	203	ROCHE	93	ROCHE	28	108.321	2.592	2.39	97.000	119.000
2023-785	6	HEXOKI. CO	220	Medica E.	130	MEDICA	13	114.000	4.472	3.92	103.000	125.000
2023-785	12	HEXOKINASE	184	INTEGRA	93	ROCHE	21	107.238	2.991	2.79	97.000	118.000
2023-785	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	14	104.000	2.268	2.18	94.000	114.000
2023-785	12	HEXOKINASE	203	ROCHE	93	ROCHE	22	106.000	3.177	3.00	95.000	117.000
2023-785	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	13	106.500	1.708	1.60	96.000	117.000
018	IRON, TOTAL	CLIA TEST ID #: 425				EVALUATION CRITERIA: TARGET VALUE 20 %						
2023-781	4	COULOMETR.	165	VITROS	63	KODAK	16	277.375	8.283	2.99	222.000	333.000
2023-782	4	COULOMETR.	165	VITROS	63	KODAK	16	119.813	7.038	5.87	96.000	144.000
2023-783	4	COULOMETR.	165	VITROS	63	KODAK	16	88.938	6.309	7.09	71.000	107.000
2023-784	4	COULOMETR.	165	VITROS	63	KODAK	16	404.313	11.389	2.82	323.000	485.000

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
018	IRON, TOTAL	CLIA TEST ID #: 425		EVALUATION CRITERIA: TARGET VALUE 20 %							
2023-785	4 COULOMETR.	165 VITROS	63 KODAK	16	207.438	8.710	4.20	166.000	249.000		
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435									
2023-781	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	19	456.368	27.612	6.05	365.000	548.000		
2023-781	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	10	444.900	9.914	2.23	356.000	534.000		
2023-781	6 NADH	203 ROCHE	93 ROCHE	25	573.160	23.291	4.06	459.000	688.000		
2023-781	6 NADH	219 ATELLICA	143 SIEMENS	10	554.600	11.732	2.12	444.000	666.000		
2023-781	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	37	538.162	16.232	3.02	431.000	646.000		
2023-781	9 LDH1	165 VITROS	63 KODAK	68	667.612	33.712	5.05	534.000	801.000		
2023-782	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	19	201.263	13.416	6.67	161.000	242.000		
2023-782	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	10	204.900	6.332	3.09	164.000	246.000		
2023-782	6 NADH	203 ROCHE	93 ROCHE	25	255.880	9.937	3.88	205.000	307.000		
2023-782	6 NADH	219 ATELLICA	143 SIEMENS	10	255.700	8.521	3.33	205.000	307.000		
2023-782	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	37	239.865	9.456	3.94	192.000	288.000		
2023-782	9 LDH1	165 VITROS	63 KODAK	68	303.318	6.631	2.19	243.000	364.000		
2023-783	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	19	141.211	9.742	6.90	113.000	169.000		
2023-783	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	10	145.500	4.225	2.90	116.000	175.000		
2023-783	6 NADH	203 ROCHE	93 ROCHE	25	180.800	7.940	4.39	145.000	217.000		
2023-783	6 NADH	219 ATELLICA	143 SIEMENS	10	181.000	5.495	3.04	145.000	217.000		
2023-783	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	37	168.667	6.642	3.94	135.000	202.000		
2023-783	9 LDH1	165 VITROS	63 KODAK	68	217.500	4.570	2.10	174.000	261.000		
2023-784	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	19	610.526	36.584	5.99	488.000	733.000		
2023-784	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	10	591.700	19.489	3.29	473.000	710.000		
2023-784	6 NADH	203 ROCHE	93 ROCHE	25	767.720	30.638	3.99	614.000	921.000		
2023-784	6 NADH	219 ATELLICA	143 SIEMENS	10	729.000	25.271	3.47	583.000	875.000		
2023-784	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	37	732.811	24.768	3.38	586.000	879.000		
2023-784	9 LDH1	165 VITROS	63 KODAK	68	866.164	47.755	5.51	693.000	1039.000		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435		EVALUATION CRITERIA: TARGET VALUE 20 %								
2023-785	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	19	347.895	23.158	6.66	278.000	417.000			
2023-785	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	10	337.300	30.328	8.99	270.000	405.000			
2023-785	6 NADH	203 ROCHE	93 ROCHE	25	438.760	17.909	4.08	351.000	527.000			
2023-785	6 NADH	219 ATELLICA	143 SIEMENS	10	430.000	10.517	2.45	344.000	516.000			
2023-785	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	37	412.919	13.413	3.25	330.000	496.000			
2023-785	9 LDH1	165 VITROS	63 KODAK	68	509.388	15.945	3.13	408.000	611.000			
021	MAGNESIUM	CLIA TEST ID #: 455		EVALUATION CRITERIA: TARGET VALUE 25 %								
2023-781	5 COLOR./K E	165 VITROS	63 KODAK	26	4.800	0.078	1.63	3.600	6.000			
2023-782	5 COLOR./K E	165 VITROS	63 KODAK	26	2.604	0.081	3.10	2.000	3.300			
2023-783	5 COLOR./K E	165 VITROS	63 KODAK	26	2.092	0.073	3.49	1.600	2.600			
2023-784	5 COLOR./K E	165 VITROS	63 KODAK	26	6.176	0.148	2.39	4.600	7.700			
2023-785	5 COLOR./K E	165 VITROS	63 KODAK	26	3.888	0.085	2.18	2.900	4.900			
022	POTASSIUM	CLIA TEST ID #: 465		EVALUATION CRITERIA: TARGET VALUE 1.00								
2023-781	3 ISE DILUT.	184 INTEGRA	93 ROCHE	30	4.933	0.087	1.76	3.900	5.900			
2023-781	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	31	4.867	0.075	1.53	3.900	5.900			
2023-781	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	16	5.013	0.105	2.10	4.000	6.000			
2023-781	3 ISE DILUT.	203 ROCHE	93 ROCHE	52	5.012	0.089	1.78	4.000	6.000			
2023-781	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	10	4.880	0.060	1.23	3.900	5.900			
2023-781	4 ISE UNDIL.	165 VITROS	63 KODAK	71	4.986	0.100	2.00	4.000	6.000			
2023-781	4 ISE UNDIL.	165 VITROS	156 Kit 32	87	4.994	0.086	1.72	4.000	6.000			
2023-781	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX	36	4.794	0.143	2.99	3.800	5.800			
2023-781	4 ISE UNDIL.	220 Medica E.	130 MEDICA	14	4.864	0.104	2.14	3.900	5.900			
2023-781	5 IMT	219 ATELLICA	143 SIEMENS	13	4.877	0.058	1.18	3.900	5.900			
2023-781	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	93	4.890	0.089	1.81	3.900	5.900			

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
022 POTASSIUM		CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE 1.00									
2023-782	3 ISE DILUT.	184 INTEGRA	93 ROCHE	30	3.073	0.057	1.87	2.100	4.100			
2023-782	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	31	3.077	0.055	1.79	2.100	4.100			
2023-782	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	16	3.075	0.056	1.82	2.100	4.100			
2023-782	3 ISE DILUT.	203 ROCHE	93 ROCHE	52	3.135	0.076	2.41	2.100	4.100			
2023-782	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	10	2.980	0.075	2.51	2.000	4.000			
2023-782	4 ISE UNLIL.	165 VITROS	63 KODAK	71	3.079	0.065	2.11	2.100	4.100			
2023-782	4 ISE UNLIL.	165 VITROS	156 Kit 32	87	3.058	0.095	3.11	2.100	4.100			
2023-782	4 ISE UNLIL.	191 RX-SERIES	126 RANDOX	36	3.106	0.101	3.26	2.100	4.100			
2023-782	4 ISE UNLIL.	220 Medica E.	130 MEDICA	14	2.986	0.074	2.49	2.000	4.000			
2023-782	5 IMT	219 ATELLICA	143 SIEMENS	13	3.031	0.061	2.00	2.000	4.000			
2023-782	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	93	3.002	0.061	2.02	2.000	4.000			
2023-783	3 ISE DILUT.	184 INTEGRA	93 ROCHE	30	2.660	0.066	2.49	1.700	3.700			
2023-783	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	31	2.661	0.055	2.06	1.700	3.700			
2023-783	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	16	2.656	0.050	1.87	1.700	3.700			
2023-783	3 ISE DILUT.	203 ROCHE	93 ROCHE	52	2.710	0.063	2.34	1.700	3.700			
2023-783	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	10	2.560	0.066	2.59	1.600	3.600			
2023-783	4 ISE UNLIL.	165 VITROS	63 KODAK	71	2.645	0.062	2.36	1.600	3.600			
2023-783	4 ISE UNLIL.	165 VITROS	156 Kit 32	87	2.665	0.055	2.05	1.700	3.700			
2023-783	4 ISE UNLIL.	191 RX-SERIES	126 RANDOX	36	2.737	0.120	4.37	1.700	3.700			
2023-783	4 ISE UNLIL.	220 Medica E.	130 MEDICA	14	2.571	0.059	2.29	1.600	3.600			
2023-783	5 IMT	219 ATELLICA	143 SIEMENS	13	2.631	0.046	1.75	1.600	3.600			
2023-783	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	93	2.576	0.054	2.09	1.600	3.600			
2023-784	3 ISE DILUT.	184 INTEGRA	93 ROCHE	30	6.150	0.092	1.50	5.200	7.200			
2023-784	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	31	6.081	0.133	2.19	5.100	7.100			
2023-784	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	16	6.256	0.141	2.26	5.300	7.300			
2023-784	3 ISE DILUT.	203 ROCHE	93 ROCHE	52	6.263	0.095	1.52	5.300	7.300			
2023-784	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	10	6.220	0.209	3.36	5.200	7.200			

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-784	4	ISE UNDIL.	165	VITROS	63	KODAK	71	6.301	0.114	1.80	5.300	7.300
2023-784	4	ISE UNDIL.	165	VITROS	156	Kit 32	87	6.287	0.133	2.11	5.300	7.300
2023-784	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	36	5.931	0.203	3.41	4.900	6.900
2023-784	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	6.171	0.148	2.41	5.200	7.200
2023-784	5	IMT	219	ATELLICA	143	SIEMENS	13	6.046	0.084	1.39	5.000	7.000
2023-784	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	93	6.120	0.103	1.69	5.100	7.100
022	POTASSIUM	CLIA TEST ID #: 465		EVALUATION CRITERIA: TARGET VALUE 1.00								
2023-785	3	ISE DILUT.	184	INTEGRA	93	ROCHE	30	4.167	0.065	1.56	3.200	5.200
2023-785	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	31	4.143	0.072	1.73	3.100	5.100
2023-785	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	16	4.175	0.066	1.58	3.200	5.200
2023-785	3	ISE DILUT.	203	ROCHE	93	ROCHE	52	4.231	0.093	2.20	3.200	5.200
2023-785	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	10	4.130	0.046	1.11	3.100	5.100
2023-785	4	ISE UNDIL.	165	VITROS	63	KODAK	71	4.208	0.083	1.98	3.200	5.200
2023-785	4	ISE UNDIL.	165	VITROS	156	Kit 32	87	4.198	0.083	1.97	3.200	5.200
2023-785	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	36	4.120	0.104	2.52	3.100	5.100
2023-785	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	4.064	0.097	2.39	3.100	5.100
2023-785	5	IMT	219	ATELLICA	143	SIEMENS	13	4.123	0.089	2.16	3.100	5.100
2023-785	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	93	4.100	0.071	1.72	3.100	5.100
023	SODIUM	CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00								
2023-781	3	ISE DILUTE	184	INTEGRA	93	ROCHE	27	151.852	2.305	1.52	148.000	156.000
2023-781	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	31	152.067	1.879	1.24	148.000	156.000
2023-781	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	12	156.000	2.082	1.33	152.000	160.000
2023-781	3	ISE DILUTE	203	ROCHE	93	ROCHE	46	153.489	2.518	1.64	149.000	157.000
2023-781	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	151.600	1.625	1.07	148.000	156.000
2023-781	4	ISE UNDIL.	165	VITROS	63	KODAK	66	155.909	2.353	1.51	152.000	160.000
2023-781	4	ISE UNDIL.	165	VITROS	156	Kit 32	92	155.967	2.263	1.45	152.000	160.000
2023-781	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	152.543	2.406	1.58	149.000	157.000
2023-781	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	150.000	0.816	0.54	146.000	154.000
2023-781	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	153.571	3.156	2.05	150.000	158.000
2023-781	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	154.714	2.864	1.85	151.000	159.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-781	5 IMT	197 ARCHITECT	1	ABBOTT	12	154.500	2.784	1.80	151.000	159.000		
2023-781	5 IMT	219 ATELLICA	143	SIEMENS	13	151.846	1.703	1.12	148.000	156.000		
2023-781	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	94	152.149	2.436	1.60	148.000	156.000		
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00									
2023-782	3 ISE DILUTE	184 INTEGRA	93	ROCHE	27	135.222	1.950	1.44	131.000	139.000		
2023-782	3 ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	31	135.774	2.059	1.52	132.000	140.000		
2023-782	3 ISE DILUTE	197 ARCHITECT	1	ABBOTT	12	136.833	2.075	1.52	133.000	141.000		
2023-782	3 ISE DILUTE	203 ROCHE	93	ROCHE	46	137.065	2.862	2.09	133.000	141.000		
2023-782	3 ISE DILUTE	215 B.C DXC	134	BECKM.COUL	10	135.500	2.617	1.93	132.000	140.000		
2023-782	4 ISE UNDIL.	165 VITROS	63	KODAK	66	136.030	2.202	1.62	132.000	140.000		
2023-782	4 ISE UNDIL.	165 VITROS	156	Kit 32	92	135.889	2.002	1.47	132.000	140.000		
2023-782	4 ISE UNDIL.	191 RX-SERIES	126	RANDOX	35	134.000	0.816	0.61	130.000	138.000		
2023-782	4 ISE UNDIL.	191 RX-SERIES	126	RANDOX	35	134.800	2.605	1.93	131.000	139.000		
2023-782	4 ISE UNDIL.	220 Medica E.	130	MEDICA	14	138.286	1.278	0.92	134.000	142.000		
2023-782	4 ISE UNDIL.	220 Medica E.	130	MEDICA	14	135.143	2.875	2.13	131.000	139.000		
2023-782	5 IMT	197 ARCHITECT	1	ABBOTT	12	136.417	2.100	1.54	132.000	140.000		
2023-782	5 IMT	219 ATELLICA	143	SIEMENS	13	136.385	2.435	1.79	132.000	140.000		
2023-782	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	94	136.634	2.431	1.78	133.000	141.000		
2023-783	3 ISE DILUTE	184 INTEGRA	93	ROCHE	27	131.423	1.498	1.14	127.000	135.000		
2023-783	3 ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	31	132.032	2.117	1.60	128.000	136.000		
2023-783	3 ISE DILUTE	197 ARCHITECT	1	ABBOTT	12	133.083	2.139	1.61	129.000	137.000		
2023-783	3 ISE DILUTE	203 ROCHE	93	ROCHE	46	133.244	1.899	1.43	129.000	137.000		
2023-783	3 ISE DILUTE	215 B.C DXC	134	BECKM.COUL	10	131.700	1.847	1.40	128.000	136.000		
2023-783	4 ISE UNDIL.	165 VITROS	63	KODAK	66	131.923	1.924	1.46	128.000	136.000		
2023-783	4 ISE UNDIL.	165 VITROS	156	Kit 32	92	132.200	1.765	1.34	128.000	136.000		
2023-783	4 ISE UNDIL.	191 RX-SERIES	126	RANDOX	35	131.086	2.019	1.54	127.000	135.000		
2023-783	4 ISE UNDIL.	191 RX-SERIES	126	RANDOX	35	129.333	0.471	0.36	125.000	133.000		
2023-783	4 ISE UNDIL.	220 Medica E.	130	MEDICA	14	131.857	2.263	1.72	128.000	136.000		
2023-783	4 ISE UNDIL.	220 Medica E.	130	MEDICA	14	132.857	2.474	1.86	129.000	137.000		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	5	IMT	197	ARCHITECT	1	ABBOTT	12	132.583	1.552	1.17	129.000	137.000
2023-783	5	IMT	219	ATELLICA	143	SIEMENS	13	133.154	2.143	1.61	129.000	137.000
2023-783	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	133.245	2.025	1.52	129.000	137.000
023	SODIUM		CLIA TEST ID #: 475			EVALUATION CRITERIA: TARGET VALUE 4.00						
2023-784	3	ISE DILUTE	184	INTEGRA	93	ROCHE	27	162.000	2.388	1.47	158.000	166.000
2023-784	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	31	163.645	2.707	1.65	160.000	168.000
2023-784	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	12	167.333	2.981	1.78	163.000	171.000
2023-784	3	ISE DILUTE	203	ROCHE	93	ROCHE	46	165.422	2.517	1.52	161.000	169.000
2023-784	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	163.400	4.673	2.86	159.000	167.000
2023-784	4	ISE UNDIL.	165	VITROS	63	KODAK	66	170.379	2.340	1.37	166.000	174.000
2023-784	4	ISE UNDIL.	165	VITROS	156	Kit 32	92	170.111	2.911	1.71	166.000	174.000
2023-784	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	161.333	1.700	1.05	157.000	165.000
2023-784	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	165.629	3.750	2.26	162.000	170.000
2023-784	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	165.857	3.943	2.38	162.000	170.000
2023-784	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	166.571	3.041	1.83	163.000	171.000
2023-784	5	IMT	197	ARCHITECT	1	ABBOTT	12	165.167	2.230	1.35	161.000	169.000
2023-784	5	IMT	219	ATELLICA	143	SIEMENS	13	160.154	2.507	1.57	156.000	164.000
2023-784	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	162.065	2.635	1.63	158.000	166.000
2023-785	3	ISE DILUTE	184	INTEGRA	93	ROCHE	27	146.074	1.676	1.15	142.000	150.000
2023-785	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	31	147.300	2.734	1.86	143.000	151.000
2023-785	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	12	148.083	1.801	1.22	144.000	152.000
2023-785	3	ISE DILUTE	203	ROCHE	93	ROCHE	46	148.432	1.888	1.27	144.000	152.000
2023-785	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	10	146.800	1.249	0.85	143.000	151.000
2023-785	4	ISE UNDIL.	165	VITROS	63	KODAK	66	148.864	2.289	1.54	145.000	153.000
2023-785	4	ISE UNDIL.	165	VITROS	156	Kit 32	92	148.944	2.391	1.61	145.000	153.000
2023-785	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	147.200	2.681	1.82	143.000	151.000
2023-785	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	35	144.000	0.816	0.57	140.000	148.000
2023-785	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	146.857	2.850	1.94	143.000	151.000
2023-785	4	ISE UNDIL.	220	Medica E.	130	MEDICA	14	149.429	3.017	2.02	145.000	153.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-785	5	IMT	197	ARCHITECT	1	ABBOTT	12	147.917	1.935	1.31	144.000	152.000
2023-785	5	IMT	219	ATELLICA	143	SIEMENS	13	146.462	2.561	1.75	142.000	150.000
2023-785	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	94	146.774	2.220	1.51	143.000	151.000
024 TOTAL PROTEIN			CLIA TEST ID #: 485			EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-781	1	BIURET	38	SIEMENS DI	143	SIEMENS	95	7.069	0.186	2.64	6.400	7.800
2023-781	1	BIURET	165	VITROS	63	KODAK	143	6.726	0.148	2.21	6.100	7.400
2023-781	1	BIURET	184	INTEGRA	93	ROCHE	19	6.437	0.175	2.72	5.800	7.100
2023-781	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	29	6.759	0.190	2.82	6.100	7.400
2023-781	1	BIURET	191	RX-SERIES	126	RANDOX	37	6.814	0.232	3.40	6.100	7.500
2023-781	1	BIURET	197	ARCHITECT	1	ABBOTT	24	6.925	0.148	2.14	6.200	7.600
2023-781	1	BIURET	203	ROCHE	93	ROCHE	42	6.766	0.151	2.23	6.100	7.400
2023-781	1	BIURET	215	B.C DXC	134	BECKM.COUL	10	6.850	0.169	2.46	6.200	7.500
2023-781	1	BIURET	219	ATELLICA	143	SIEMENS	13	6.892	0.149	2.16	6.200	7.600
2023-781	1	BIURET	220	Medica E.	130	MEDICA	10	6.520	0.178	2.73	5.900	7.200
2023-781	3	BIURET COL	203	ROCHE	93	ROCHE	10	6.860	0.201	2.93	6.200	7.500
2023-782	1	BIURET	38	SIEMENS DI	143	SIEMENS	95	4.483	0.114	2.53	4.000	4.900
2023-782	1	BIURET	165	VITROS	63	KODAK	143	4.407	0.129	2.92	4.000	4.800
2023-782	1	BIURET	184	INTEGRA	93	ROCHE	19	4.158	0.150	3.60	3.700	4.600
2023-782	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	29	4.290	0.124	2.89	3.900	4.700
2023-782	1	BIURET	191	RX-SERIES	126	RANDOX	37	4.368	0.190	4.36	3.900	4.800
2023-782	1	BIURET	197	ARCHITECT	1	ABBOTT	24	4.215	0.111	2.63	3.800	4.600
2023-782	1	BIURET	203	ROCHE	93	ROCHE	42	4.355	0.116	2.66	3.900	4.800
2023-782	1	BIURET	215	B.C DXC	134	BECKM.COUL	10	4.350	0.120	2.77	3.900	4.800
2023-782	1	BIURET	219	ATELLICA	143	SIEMENS	13	4.331	0.126	2.92	3.900	4.800
2023-782	1	BIURET	220	Medica E.	130	MEDICA	10	4.150	0.136	3.28	3.700	4.600
2023-782	3	BIURET COL	203	ROCHE	93	ROCHE	10	4.440	0.102	2.30	4.000	4.900
2023-783	1	BIURET	38	SIEMENS DI	143	SIEMENS	95	3.890	0.112	2.88	3.500	4.300
2023-783	1	BIURET	165	VITROS	63	KODAK	143	3.860	0.104	2.71	3.500	4.200

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-783	1	BIURET	184	INTEGRA	93	ROCHE	19	3.605	0.143	3.97	3.200	4.000
2023-783	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	29	3.721	0.119	3.19	3.300	4.100
2023-783	1	BIURET	191	RX-SERIES	126	RANDOX	37	3.806	0.133	3.50	3.400	4.200
2023-783	1	BIURET	197	ARCHITECT	1	ABBOTT	24	3.640	0.111	3.06	3.300	4.000
2023-783	1	BIURET	203	ROCHE	93	ROCHE	42	3.788	0.096	2.52	3.400	4.200
2023-783	1	BIURET	215	B.C DXC	134	BECKM.COUL	10	3.800	0.118	3.11	3.400	4.200
2023-783	1	BIURET	219	ATELLICA	143	SIEMENS	13	3.731	0.091	2.44	3.400	4.100
2023-783	1	BIURET	220	Medica E.	130	MEDICA	10	3.630	0.135	3.71	3.300	4.000
2023-783	3	BIURET COL	203	ROCHE	93	ROCHE	10	3.870	0.155	4.01	3.500	4.300
024 TOTAL PROTEIN			CLIA TEST ID #: 485			EVALUATION CRITERIA: TARGET VALUE 10 %						
2023-784	1	BIURET	38	SIEMENS DI	143	SIEMENS	95	8.780	0.217	2.47	7.900	9.700
2023-784	1	BIURET	165	VITROS	63	KODAK	143	8.278	0.221	2.67	7.500	9.100
2023-784	1	BIURET	184	INTEGRA	93	ROCHE	19	8.006	0.334	4.17	7.200	8.800
2023-784	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	29	8.328	0.232	2.78	7.500	9.200
2023-784	1	BIURET	191	RX-SERIES	126	RANDOX	37	8.384	0.350	4.17	7.500	9.200
2023-784	1	BIURET	197	ARCHITECT	1	ABBOTT	24	8.630	0.208	2.41	7.800	9.500
2023-784	1	BIURET	203	ROCHE	93	ROCHE	42	8.281	0.273	3.29	7.500	9.100
2023-784	1	BIURET	215	B.C DXC	134	BECKM.COUL	10	8.450	0.422	5.00	7.600	9.300
2023-784	1	BIURET	219	ATELLICA	143	SIEMENS	13	8.354	0.250	2.99	7.500	9.200
2023-784	1	BIURET	220	Medica E.	130	MEDICA	10	8.040	0.196	2.44	7.200	8.800
2023-784	3	BIURET COL	203	ROCHE	93	ROCHE	10	8.470	0.253	2.99	7.600	9.300
2023-785	1	BIURET	38	SIEMENS DI	143	SIEMENS	95	6.014	0.140	2.32	5.400	6.600
2023-785	1	BIURET	165	VITROS	63	KODAK	143	5.779	0.138	2.38	5.200	6.400
2023-785	1	BIURET	184	INTEGRA	93	ROCHE	19	5.553	0.182	3.27	5.000	6.100
2023-785	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	29	5.746	0.152	2.65	5.200	6.300
2023-785	1	BIURET	191	RX-SERIES	126	RANDOX	37	5.830	0.225	3.86	5.200	6.400
2023-785	1	BIURET	197	ARCHITECT	1	ABBOTT	24	5.795	0.112	1.93	5.200	6.400
2023-785	1	BIURET	203	ROCHE	93	ROCHE	42	5.790	0.151	2.61	5.200	6.400
2023-785	1	BIURET	215	B.C DXC	134	BECKM.COUL	10	5.900	0.155	2.63	5.300	6.500

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-785	1	BIURET	219	APELLICA	143	SIEMENS	13	5.831	0.120	2.06	5.200	6.400
2023-785	1	BIURET	220	Medica E.	130	MEDICA	10	5.560	0.180	3.24	5.000	6.100
2023-785	3	BIURET COL	203	ROCHE	93	ROCHE	10	5.780	0.268	4.63	5.200	6.400
025 TRIGLYCERIDES		CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %						
2023-781	2	ENZ. COLO.	165	VITROS	63	KODAK	146	210.846	7.550	3.58	158.000	264.000
2023-781	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	13	171.692	3.646	2.12	129.000	215.000
2023-781	2	ENZ. COLO.	203	ROCHE	93	ROCHE	27	176.111	3.966	2.25	132.000	220.000
2023-781	2	ENZ. COLO.	220	Medica E.	130	MEDICA	10	165.500	5.065	3.06	124.000	207.000
2023-781	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	92	173.022	3.718	2.15	130.000	216.000
2023-781	5	HANTZCH-C)	197	ARCHITECT	1	ABBOTT	12	171.583	2.532	1.48	129.000	214.000
2023-781	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	183.310	5.826	3.18	137.000	229.000
2023-781	9	GPO TRIND.	191	RX-SERIES	126	RANDOX	33	174.030	7.175	4.12	131.000	218.000
2023-781	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	189.400	2.615	1.38	142.000	237.000
2023-781	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	15	173.400	5.314	3.06	130.000	217.000
2023-781	10	ENZYM. GPO	203	ROCHE	93	ROCHE	23	174.043	4.349	2.50	131.000	218.000
2023-781	10	ENZYM. GPO	219	APELLICA	143	SIEMENS	13	180.846	4.849	2.68	136.000	226.000
2023-781	11	GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	170.818	4.764	2.79	128.000	214.000
2023-782	2	ENZ. COLO.	165	VITROS	63	KODAK	146	133.160	5.883	4.42	100.000	166.000
2023-782	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	13	116.231	2.547	2.19	87.000	145.000
2023-782	2	ENZ. COLO.	203	ROCHE	93	ROCHE	27	118.111	3.035	2.57	89.000	148.000
2023-782	2	ENZ. COLO.	220	Medica E.	130	MEDICA	10	112.700	4.473	3.97	85.000	141.000
2023-782	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	92	109.337	3.231	2.96	82.000	137.000
2023-782	5	HANTZCH-C)	197	ARCHITECT	1	ABBOTT	12	109.500	1.756	1.60	82.000	137.000
2023-782	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	118.759	3.848	3.24	89.000	148.000
2023-782	9	GPO TRIND.	191	RX-SERIES	126	RANDOX	33	116.500	4.981	4.28	87.000	146.000
2023-782	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	120.400	3.353	2.78	90.000	151.000
2023-782	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	15	117.267	3.907	3.33	88.000	147.000
2023-782	10	ENZYM. GPO	203	ROCHE	93	ROCHE	23	116.652	3.509	3.01	87.000	146.000
2023-782	10	ENZYM. GPO	219	APELLICA	143	SIEMENS	13	117.923	4.215	3.57	88.000	147.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	107.455	1.876	1.75	81.000	134.000	
025	TRIGLYCERIDES	CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %						
2023-783	2 ENZ. COLO.	165	VITROS	63	KODAK	146	116.021	4.925	4.25	87.000	145.000	
2023-783	2 ENZ. COLO.	184	INTEGRA	93	ROCHE	13	102.462	2.308	2.25	77.000	128.000	
2023-783	2 ENZ. COLO.	203	ROCHE	93	ROCHE	27	104.444	2.713	2.60	78.000	131.000	
2023-783	2 ENZ. COLO.	220	Medica E.	130	MEDICA	10	99.700	3.900	3.91	75.000	125.000	
2023-783	4 ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	92	93.457	3.167	3.39	70.000	117.000	
2023-783	5 HANTZCH-C)	197	ARCHITECT	1	ABBOTT	12	95.500	2.141	2.24	72.000	119.000	
2023-783	9 GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	103.207	3.585	3.47	77.000	129.000	
2023-783	9 GPO TRIND.	191	RX-SERIES	126	RANDOX	33	101.424	5.646	5.57	76.000	127.000	
2023-783	9 GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	108.200	1.939	1.79	81.000	135.000	
2023-783	10 ENZYM. GPO	184	INTEGRA	93	ROCHE	15	103.533	3.721	3.59	78.000	129.000	
2023-783	10 ENZYM. GPO	203	ROCHE	93	ROCHE	23	102.261	3.273	3.20	77.000	128.000	
2023-783	10 ENZYM. GPO	219	ATELLICA	143	SIEMENS	13	102.923	3.872	3.76	77.000	129.000	
2023-783	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	93.545	1.373	1.47	70.000	117.000	
2023-784	2 ENZ. COLO.	165	VITROS	63	KODAK	146	265.743	10.897	4.10	199.000	332.000	
2023-784	2 ENZ. COLO.	184	INTEGRA	93	ROCHE	13	207.615	4.954	2.39	156.000	260.000	
2023-784	2 ENZ. COLO.	203	ROCHE	93	ROCHE	27	211.370	6.838	3.23	159.000	264.000	
2023-784	2 ENZ. COLO.	220	Medica E.	130	MEDICA	10	199.100	9.534	4.79	149.000	249.000	
2023-784	4 ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	92	214.308	5.540	2.59	161.000	268.000	
2023-784	5 HANTZCH-C)	197	ARCHITECT	1	ABBOTT	12	212.500	4.805	2.26	159.000	266.000	
2023-784	9 GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	223.931	7.552	3.37	168.000	280.000	
2023-784	9 GPO TRIND.	191	RX-SERIES	126	RANDOX	33	210.606	9.582	4.55	158.000	263.000	
2023-784	9 GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	236.900	9.710	4.10	178.000	296.000	
2023-784	10 ENZYM. GPO	184	INTEGRA	93	ROCHE	15	207.333	8.919	4.30	156.000	259.000	
2023-784	10 ENZYM. GPO	203	ROCHE	93	ROCHE	23	207.217	5.687	2.74	155.000	259.000	
2023-784	10 ENZYM. GPO	219	ATELLICA	143	SIEMENS	13	217.231	6.518	3.00	163.000	272.000	
2023-784	11 GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	207.545	4.998	2.41	156.000	259.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
025 TRIGLYCERIDES		CLIA TEST ID #: 495		EVALUATION CRITERIA: TARGET VALUE 25 %							
2023-785	2 ENZ. COLO.	165 VITROS	63 KODAK	146	178.972	6.410	3.58	134.000	224.000		
2023-785	2 ENZ. COLO.	184 INTEGRA	93 ROCHE	13	150.385	2.816	1.87	113.000	188.000		
2023-785	2 ENZ. COLO.	203 ROCHE	93 ROCHE	27	152.481	3.891	2.55	114.000	191.000		
2023-785	2 ENZ. COLO.	220 Medica E.	130 MEDICA	10	144.700	5.984	4.14	109.000	181.000		
2023-785	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	92	146.978	3.393	2.31	110.000	184.000		
2023-785	5 HANTZCH-C)	197 ARCHITECT	1 ABBOTT	12	147.083	2.396	1.63	110.000	184.000		
2023-785	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN	29	157.379	5.974	3.80	118.000	197.000		
2023-785	9 GPO TRIND.	191 RX-SERIES	126 RANDOX	33	150.970	6.511	4.31	113.000	189.000		
2023-785	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL	10	163.200	5.269	3.23	122.000	204.000		
2023-785	10 ENZYM. GPO	184 INTEGRA	93 ROCHE	15	152.133	4.829	3.17	114.000	190.000		
2023-785	10 ENZYM. GPO	203 ROCHE	93 ROCHE	23	151.174	3.941	2.61	113.000	189.000		
2023-785	10 ENZYM. GPO	219 ATELLICA	143 SIEMENS	13	155.385	4.876	3.14	117.000	194.000		
2023-785	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT	11	144.636	2.057	1.42	108.000	181.000		
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}							
2023-781	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	97	33.811	1.340	3.96	31.000	37.000		
2023-781	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	97	34.000	1.225	3.60	31.000	37.000		
2023-781	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	33	32.727	1.309	4.00	30.000	36.000		
2023-781	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	31	32.935	0.914	2.77	30.000	36.000		
2023-781	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX	38	30.974	1.597	5.16	28.000	34.000		
2023-781	2 UREAS.GLDH	203 ROCHE	93 ROCHE	52	32.784	1.108	3.38	30.000	36.000		
2023-781	2 UREAS.GLDH	219 ATELLICA	143 SIEMENS	13	33.154	0.533	1.61	30.000	36.000		
2023-781	5 CHROMATOG.	197 ARCHITECT	1 ABBOTT	11	33.545	1.233	3.68	31.000	37.000		
2023-781	8 UREASE QUI	165 VITROS	63 KODAK	159	30.541	0.920	3.01	28.000	33.000		
2023-781	10 UREASE 37	220 Medica E.	130 MEDICA	13	32.000	1.109	3.47	29.000	35.000		
2023-781	11 UREASE	197 ARCHITECT	1 ABBOTT	10	33.000	0.894	2.71	30.000	36.000		
2023-782	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	97	20.250	1.090	5.38	18.000	22.000		
2023-782	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	97	19.454	0.942	4.84	17.000	21.000		
2023-782	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	33	18.970	0.904	4.77	17.000	21.000		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	31	19.290	0.727	3.77	17.000	21.000
2023-782	2	UREAS.GLDH	191	RX-SERIES	126	RANDOX	38	18.892	0.980	5.19	17.000	21.000
2023-782	2	UREAS.GLDH	203	ROCHE	93	ROCHE	52	19.000	0.792	4.17	17.000	21.000
2023-782	2	UREAS.GLDH	219	ATELLICA	143	SIEMENS	13	20.000	0.679	3.40	18.000	22.000
2023-782	5	CHROMATOG.	197	ARCHITECT	1	ABBOTT	11	19.182	0.936	4.88	17.000	21.000
2023-782	8	UREASE QUI	165	VITROS	63	KODAK	159	16.892	0.701	4.15	15.000	19.000
2023-782	10	UREASE 37	220	Medica E.	130	MEDICA	13	18.846	0.662	3.51	17.000	21.000
2023-782	11	UREASE	197	ARCHITECT	1	ABBOTT	10	18.400	0.490	2.66	16.000	20.000
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505				EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}						
2023-783	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	16.750	0.829	4.95	15.000	19.000
2023-783	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	16.021	0.968	6.04	14.000	18.000
2023-783	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	33	15.879	0.640	4.03	14.000	18.000
2023-783	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	31	16.032	0.474	2.96	14.000	18.000
2023-783	2	UREAS.GLDH	191	RX-SERIES	126	RANDOX	38	16.135	0.935	5.79	14.000	18.000
2023-783	2	UREAS.GLDH	203	ROCHE	93	ROCHE	52	15.922	0.621	3.90	14.000	18.000
2023-783	2	UREAS.GLDH	219	ATELLICA	143	SIEMENS	13	16.692	0.606	3.63	15.000	19.000
2023-783	5	CHROMATOG.	197	ARCHITECT	1	ABBOTT	11	16.364	0.979	5.98	14.000	18.000
2023-783	8	UREASE QUI	165	VITROS	63	KODAK	159	13.923	0.489	3.51	12.000	16.000
2023-783	10	UREASE 37	220	Medica E.	130	MEDICA	13	15.692	0.462	2.94	14.000	18.000
2023-783	11	UREASE	197	ARCHITECT	1	ABBOTT	10	15.300	0.458	3.00	13.000	17.000
2023-784	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	43.000	1.871	4.35	39.000	47.000
2023-784	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	43.284	1.633	3.77	39.000	47.000
2023-784	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	33	42.061	1.722	4.09	38.000	46.000
2023-784	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	31	42.032	1.332	3.17	38.000	46.000
2023-784	2	UREAS.GLDH	191	RX-SERIES	126	RANDOX	38	38.684	2.091	5.41	35.000	42.000
2023-784	2	UREAS.GLDH	203	ROCHE	93	ROCHE	52	41.745	1.506	3.61	38.000	46.000
2023-784	2	UREAS.GLDH	219	ATELLICA	143	SIEMENS	13	40.750	0.595	1.46	37.000	44.000
2023-784	5	CHROMATOG.	197	ARCHITECT	1	ABBOTT	11	43.818	1.527	3.48	40.000	48.000
2023-784	8	UREASE QUI	165	VITROS	63	KODAK	159	39.948	1.146	2.87	36.000	44.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-784	10	UREASE 37	220	Medica E.	130	MEDICA	13	40.077	0.730	1.82	36.000	44.000
2023-784	11	UREASE	197	ARCHITECT	1	ABBOTT	10	42.400	0.663	1.56	39.000	46.000
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505				EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}						
2023-785	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	27.792	1.274	4.58	25.000	30.000
2023-785	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	97	28.500	1.500	5.26	26.000	31.000
2023-785	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	33	27.152	1.019	3.75	25.000	30.000
2023-785	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	31	27.355	0.900	3.29	25.000	30.000
2023-785	2	UREAS.GLDH	191	RX-SERIES	126	RANDOX	38	26.263	1.427	5.43	24.000	29.000
2023-785	2	UREAS.GLDH	203	ROCHE	93	ROCHE	52	27.039	0.928	3.43	25.000	29.000
2023-785	2	UREAS.GLDH	219	ATELLICA	143	SIEMENS	13	27.538	0.929	3.38	25.000	30.000
2023-785	5	CHROMATOG.	197	ARCHITECT	1	ABBOTT	11	27.545	1.076	3.91	25.000	30.000
2023-785	8	UREASE QUI	165	VITROS	63	KODAK	159	24.535	0.778	3.17	22.000	27.000
2023-785	10	UREASE 37	220	Medica E.	130	MEDICA	13	26.615	0.923	3.47	24.000	29.000
2023-785	11	UREASE	197	ARCHITECT	1	ABBOTT	10	26.900	0.539	2.00	24.000	29.000
027 URIC ACID		CLIA TEST ID #: 515				EVALUATION CRITERIA: TARGET VALUE 17 %						
2023-781	2	URICASE	38	SIEMENS DI	143	SIEMENS	51	8.575	0.218	2.54	7.100	10.000
2023-781	2	URICASE	165	VITROS	63	KODAK	96	8.548	0.199	2.33	7.100	10.000
2023-781	2	URICASE	184	INTEGRA	93	ROCHE	11	8.582	0.252	2.93	7.100	10.000
2023-781	2	URICASE	203	ROCHE	93	ROCHE	31	8.433	0.269	3.19	7.000	9.900
2023-781	2	URICASE	219	ATELLICA	143	SIEMENS	11	8.873	0.267	3.00	7.400	10.400
2023-781	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	18	8.433	0.189	2.24	7.000	9.900
2023-781	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	10	8.060	0.111	1.38	6.700	9.400
2023-781	6	ENZY. COL	203	ROCHE	93	ROCHE	13	8.662	0.169	1.95	7.200	10.100
2023-782	2	URICASE	38	SIEMENS DI	143	SIEMENS	51	4.227	0.083	1.96	3.500	4.900
2023-782	2	URICASE	165	VITROS	63	KODAK	96	4.176	0.142	3.40	3.500	4.900
2023-782	2	URICASE	184	INTEGRA	93	ROCHE	11	4.282	0.159	3.70	3.600	5.000
2023-782	2	URICASE	203	ROCHE	93	ROCHE	31	4.167	0.111	2.65	3.500	4.900
2023-782	2	URICASE	219	ATELLICA	143	SIEMENS	11	4.464	0.177	3.97	3.700	5.200
2023-782	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	18	4.272	0.128	3.00	3.500	5.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-782	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	10	4.180	0.075	1.79	3.500	4.900		
2023-782	6 ENZY. COL	203 ROCHE	93 ROCHE	13	4.262	0.084	1.96	3.500	5.000		
027 URIC ACID		CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %								
2023-783	2 URICASE	38 SIEMENS DI	143 SIEMENS	51	3.231	0.094	2.91	2.700	3.800		
2023-783	2 URICASE	165 VITROS	63 KODAK	96	3.217	0.087	2.71	2.700	3.800		
2023-783	2 URICASE	184 INTEGRA	93 ROCHE	11	3.255	0.137	4.22	2.700	3.800		
2023-783	2 URICASE	203 ROCHE	93 ROCHE	31	3.200	0.113	3.52	2.700	3.700		
2023-783	2 URICASE	219 ATELLICA	143 SIEMENS	11	3.436	0.107	3.11	2.900	4.000		
2023-783	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	3.289	0.099	3.02	2.700	3.800		
2023-783	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	10	3.230	0.046	1.42	2.700	3.800		
2023-783	6 ENZY. COL	203 ROCHE	93 ROCHE	13	3.246	0.084	2.60	2.700	3.800		
2023-784	2 URICASE	38 SIEMENS DI	143 SIEMENS	51	11.331	0.293	2.59	9.400	13.300		
2023-784	2 URICASE	165 VITROS	63 KODAK	96	11.431	0.319	2.79	9.500	13.400		
2023-784	2 URICASE	184 INTEGRA	93 ROCHE	11	11.327	0.481	4.24	9.400	13.300		
2023-784	2 URICASE	203 ROCHE	93 ROCHE	31	11.158	0.441	3.95	9.300	13.100		
2023-784	2 URICASE	219 ATELLICA	143 SIEMENS	11	11.555	0.387	3.35	9.600	13.500		
2023-784	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	11.033	0.238	2.16	9.200	12.900		
2023-784	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	10	10.440	0.301	2.88	8.700	12.200		
2023-784	6 ENZY. COL	203 ROCHE	93 ROCHE	13	11.400	0.280	2.46	9.500	13.300		
2023-785	2 URICASE	38 SIEMENS DI	143 SIEMENS	51	6.737	0.166	2.46	5.600	7.900		
2023-785	2 URICASE	165 VITROS	63 KODAK	96	6.676	0.182	2.73	5.500	7.800		
2023-785	2 URICASE	184 INTEGRA	93 ROCHE	11	6.836	0.187	2.74	5.700	8.000		
2023-785	2 URICASE	203 ROCHE	93 ROCHE	31	6.661	0.257	3.86	5.500	7.800		
2023-785	2 URICASE	219 ATELLICA	143 SIEMENS	11	7.009	0.250	3.57	5.800	8.200		
2023-785	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	18	6.671	0.152	2.29	5.500	7.800		
2023-785	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	10	6.320	0.581	9.19	5.200	7.400		
2023-785	6 ENZY. COL	203 ROCHE	93 ROCHE	13	6.777	0.131	1.93	5.600	7.900		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
135	GGT (GAMMA GLU. TRA)		CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %					
2023-781	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	261.000	8.832	3.38	209.000	313.000
2023-781	5	KINETIC 37	165	VITROS	63	KODAK	24	452.833	13.406	2.96	362.000	543.000
2023-782	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	103.400	4.386	4.24	83.000	124.000
2023-782	5	KINETIC 37	165	VITROS	63	KODAK	24	206.042	4.730	2.30	165.000	247.000
2023-783	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	66.600	2.498	3.75	53.000	80.000
2023-783	5	KINETIC 37	165	VITROS	63	KODAK	24	135.250	2.727	2.02	108.000	162.000
2023-784	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	362.100	13.824	3.82	290.000	435.000
2023-784	5	KINETIC 37	165	VITROS	63	KODAK	24	582.957	17.266	2.96	466.000	700.000
2023-785	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	194.100	6.252	3.22	155.000	233.000
2023-785	5	KINETIC 37	165	VITROS	63	KODAK	24	357.708	10.085	2.82	286.000	429.000
136	LACTIC ACID											
2023-781	1	LACTATE-P.	165	VITROS	63	KODAK	22	4.541	0.111	2.45	3.600	5.400
2023-782	1	LACTATE-P.	165	VITROS	63	KODAK	22	1.677	0.052	3.08	1.300	2.000
2023-783	1	LACTATE-P.	165	VITROS	63	KODAK	22	1.032	0.047	4.51	0.800	1.200
2023-784	1	LACTATE-P.	165	VITROS	63	KODAK	22	6.214	0.189	3.04	5.000	7.500
2023-785	1	LACTATE-P.	165	VITROS	63	KODAK	22	3.368	0.082	2.43	2.700	4.000
137	LIPASE											
2023-781	1	COLORIMET.	203	ROCHE	93	ROCHE	17	63.529	3.696	5.82	51.000	76.000
2023-781	2	ENZIMATIC	165	VITROS	63	KODAK	84	527.494	17.899	3.39	422.000	633.000
2023-781	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	60.400	2.245	3.72	48.000	72.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-781	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	61.800	3.516	5.69	49.000	74.000
2023-781	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	69.867	3.631	5.20	56.000	84.000
2023-781	4	ENZ/COLOR	203	ROCHE	93	ROCHE	12	63.917	3.947	6.17	51.000	77.000
2023-781	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	17	68.765	1.986	2.89	55.000	83.000
137	LIPASE	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %								
2023-782	1	COLORIMET.	203	ROCHE	93	ROCHE	17	33.353	1.713	5.14	27.000	40.000
2023-782	2	ENZIMATIC	165	VITROS	63	KODAK	84	245.157	8.929	3.64	196.000	294.000
2023-782	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	37.800	1.536	4.06	30.000	45.000
2023-782	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	33.200	1.661	5.00	27.000	40.000
2023-782	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	35.133	3.030	8.62	28.000	42.000
2023-782	4	ENZ/COLOR	203	ROCHE	93	ROCHE	12	33.750	1.920	5.69	27.000	41.000
2023-782	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	17	36.765	1.516	4.12	29.000	44.000
2023-783	1	COLORIMET.	203	ROCHE	93	ROCHE	17	26.118	1.131	4.33	21.000	31.000
2023-783	2	ENZIMATIC	165	VITROS	63	KODAK	84	172.631	6.850	3.97	138.000	207.000
2023-783	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	33.100	1.513	4.57	26.000	40.000
2023-783	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	26.700	1.792	6.71	21.000	32.000
2023-783	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	26.333	2.357	8.95	21.000	32.000
2023-783	4	ENZ/COLOR	203	ROCHE	93	ROCHE	12	27.333	1.312	4.80	22.000	33.000
2023-783	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	17	29.353	1.234	4.20	23.000	35.000
2023-784	1	COLORIMET.	203	ROCHE	93	ROCHE	17	82.647	5.291	6.40	66.000	99.000
2023-784	2	ENZIMATIC	165	VITROS	63	KODAK	84	682.723	23.881	3.50	546.000	819.000
2023-784	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	74.300	2.238	3.01	59.000	89.000
2023-784	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	78.900	4.277	5.42	63.000	95.000
2023-784	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	92.467	4.470	4.83	74.000	111.000
2023-784	4	ENZ/COLOR	203	ROCHE	93	ROCHE	12	82.750	6.098	7.37	66.000	99.000
2023-784	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	17	89.706	2.492	2.78	72.000	108.000
2023-785	1	COLORIMET.	203	ROCHE	93	ROCHE	17	51.176	2.640	5.16	41.000	61.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-785	2	ENZIMATIC	165	VITROS	63	KODAK	84	414.361	14.054	3.39	331.000	497.000
2023-785	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	51.400	2.245	4.37	41.000	62.000
2023-785	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	49.800	2.272	4.56	40.000	60.000
2023-785	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	55.133	2.500	4.53	44.000	66.000
2023-785	4	ENZ/COLOR	203	ROCHE	93	ROCHE	12	51.833	3.184	6.14	41.000	62.000
2023-785	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	17	56.118	2.541	4.53	45.000	67.000
138	PHOSPHORUS	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %								
2023-781	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	53	5.412	0.152	2.81	4.300	6.500
2023-781	1	PHOSPHOMO.	165	VITROS	63	KODAK	93	5.022	0.173	3.45	4.000	6.000
2023-781	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	15	5.000	0.155	3.10	4.000	6.000
2023-781	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	19	5.100	0.149	2.92	4.100	6.100
2023-781	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	40	5.120	0.140	2.73	4.100	6.100
2023-781	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	10	5.440	0.233	4.29	4.400	6.500
2023-781	1	PHOSPHOMO.	219	ATELLICA	143	SIEMENS	11	5.382	0.083	1.55	4.300	6.500
2023-782	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	53	3.120	0.127	4.07	2.500	3.700
2023-782	1	PHOSPHOMO.	165	VITROS	63	KODAK	93	3.116	0.142	4.54	2.500	3.700
2023-782	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	15	2.967	0.140	4.71	2.400	3.600
2023-782	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	19	2.926	0.125	4.27	2.300	3.500
2023-782	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	40	2.920	0.093	3.18	2.300	3.500
2023-782	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	10	3.100	0.167	5.40	2.500	3.700
2023-782	1	PHOSPHOMO.	219	ATELLICA	143	SIEMENS	11	3.082	0.057	1.87	2.500	3.700
2023-783	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	53	2.564	0.109	4.26	2.100	3.100
2023-783	1	PHOSPHOMO.	165	VITROS	63	KODAK	93	2.676	0.115	4.30	2.100	3.200
2023-783	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	15	2.400	0.110	4.56	1.900	2.900
2023-783	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	19	2.383	0.083	3.50	1.900	2.900
2023-783	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	40	2.413	0.075	3.10	1.900	2.900
2023-783	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	10	2.520	0.117	4.63	2.000	3.000
2023-783	1	PHOSPHOMO.	219	ATELLICA	143	SIEMENS	11	2.509	0.051	2.05	2.000	3.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
138	PHOSPHORUS		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE	20 %						
2023-784	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	53	6.848	0.245	3.57	5.500	8.200	
2023-784	1	PHOSPHOMO.	165 VITROS	63	KODAK	93	6.311	0.208	3.30	5.000	7.600	
2023-784	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	15	6.380	0.259	4.06	5.100	7.700	
2023-784	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	19	6.547	0.239	3.65	5.200	7.900	
2023-784	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	40	6.595	0.134	2.03	5.300	7.900	
2023-784	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	10	7.040	0.317	4.50	5.600	8.400	
2023-784	1	PHOSPHOMO.	219 ATELLICA	143	SIEMENS	11	6.764	0.172	2.54	5.400	8.100	
2023-785	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	53	4.440	0.123	2.77	3.600	5.300	
2023-785	1	PHOSPHOMO.	165 VITROS	63	KODAK	93	4.243	0.151	3.57	3.400	5.100	
2023-785	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	15	4.187	0.189	4.52	3.300	5.000	
2023-785	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	19	4.226	0.220	5.20	3.400	5.100	
2023-785	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	40	4.228	0.114	2.70	3.400	5.100	
2023-785	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	10	4.550	0.180	3.96	3.600	5.500	
2023-785	1	PHOSPHOMO.	219 ATELLICA	143	SIEMENS	11	4.418	0.094	2.12	3.500	5.300	
169	CO2				EVALUATION CRITERIA: TARGET VALUE	3 SD						
2023-781	1	Generic Me	165 VITROS	63	KODAK	10	19.500	1.688	8.66	14.000	25.000	
2023-781	1	Generic Me	184 INTEGRA	93	ROCHE	10	21.500	2.872	13.36	13.000	30.000	
2023-781	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	26	21.615	3.127	14.46	12.000	31.000	
2023-781	1	Generic Me	203 ROCHE	93	ROCHE	23	20.435	2.356	11.53	13.000	28.000	
2023-781	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	78	22.923	2.111	9.21	17.000	29.000	
2023-781	2	ENZYMATIC	165 VITROS	63	KODAK	61	20.217	1.694	8.38	15.000	25.000	
2023-781	2	ENZYMATIC	165 VITROS	156	Kit 32	77	20.355	2.327	11.43	13.000	27.000	
2023-781	2	ENZYMATIC	203 ROCHE	93	ROCHE	11	20.636	2.422	11.74	13.000	28.000	
2023-781	2	ENZYMATIC	219 ATELLICA	143	SIEMENS	11	22.909	1.240	5.41	19.000	27.000	
2023-782	1	Generic Me	165 VITROS	63	KODAK	10	12.100	1.375	11.36	8.000	16.000	
2023-782	1	Generic Me	184 INTEGRA	93	ROCHE	10	13.600	2.289	16.83	7.000	20.000	
2023-782	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	26	13.077	1.817	13.90	8.000	19.000	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-782	1	Generic Me	203	ROCHE	93	ROCHE	23	12.826	1.307	10.19	9.000	17.000
2023-782	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	78	14.408	1.566	10.87	10.000	19.000
2023-782	2	ENZYMATIC	165	VITROS	63	KODAK	61	11.967	1.689	14.11	7.000	17.000
2023-782	2	ENZYMATIC	165	VITROS	156	Kit 32	77	12.171	1.534	12.60	8.000	17.000
2023-782	2	ENZYMATIC	203	ROCHE	93	ROCHE	11	12.364	1.367	11.05	8.000	16.000
2023-782	2	ENZYMATIC	219	ATELLICA	143	SIEMENS	11	14.364	1.772	12.34	9.000	20.000
169	CO2	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD								
2023-783	1	Generic Me	165	VITROS	63	KODAK	10	10.000	1.673	16.73	5.000	15.000
2023-783	1	Generic Me	184	INTEGRA	93	ROCHE	10	12.700	2.147	16.91	6.000	19.000
2023-783	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	26	11.769	2.189	18.60	5.000	18.000
2023-783	1	Generic Me	203	ROCHE	93	ROCHE	23	11.217	1.559	13.90	7.000	16.000
2023-783	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	78	13.455	1.575	11.71	9.000	18.000
2023-783	2	ENZYMATIC	165	VITROS	63	KODAK	61	10.443	1.420	13.60	6.000	15.000
2023-783	2	ENZYMATIC	165	VITROS	156	Kit 32	77	10.507	1.370	13.04	6.000	15.000
2023-783	2	ENZYMATIC	203	ROCHE	93	ROCHE	11	11.455	0.988	8.62	8.000	14.000
2023-783	2	ENZYMATIC	219	ATELLICA	143	SIEMENS	11	13.455	1.777	13.21	8.000	19.000
2023-784	1	Generic Me	165	VITROS	63	KODAK	10	25.100	1.578	6.29	20.000	30.000
2023-784	1	Generic Me	184	INTEGRA	93	ROCHE	10	25.200	2.182	8.66	19.000	32.000
2023-784	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	26	27.462	3.237	11.79	18.000	37.000
2023-784	1	Generic Me	203	ROCHE	93	ROCHE	23	24.522	2.517	10.27	17.000	32.000
2023-784	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	78	27.872	2.096	7.52	22.000	34.000
2023-784	2	ENZYMATIC	165	VITROS	63	KODAK	61	24.738	1.983	8.01	19.000	31.000
2023-784	2	ENZYMATIC	165	VITROS	156	Kit 32	77	25.320	2.322	9.17	18.000	32.000
2023-784	2	ENZYMATIC	203	ROCHE	93	ROCHE	11	26.364	1.823	6.91	21.000	32.000
2023-784	2	ENZYMATIC	219	ATELLICA	143	SIEMENS	11	27.545	1.616	5.87	23.000	32.000
2023-785	1	Generic Me	165	VITROS	63	KODAK	10	15.300	1.552	10.15	11.000	20.000
2023-785	1	Generic Me	184	INTEGRA	93	ROCHE	10	17.400	2.653	15.25	9.000	25.000
2023-785	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	26	17.423	2.734	15.69	9.000	26.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-785	1 Generic Me	203 ROCHE	93 ROCHE	23	16.870	2.091	12.40	11.000	23.000		
2023-785	2 ENZYMATIC	38 SIEMENS DI	143 SIEMENS	78	18.974	2.230	11.75	12.000	26.000		
2023-785	2 ENZYMATIC	165 VITROS	63 KODAK	61	15.852	1.745	11.01	11.000	21.000		
2023-785	2 ENZYMATIC	165 VITROS	156 Kit 32	77	16.187	1.958	12.09	10.000	22.000		
2023-785	2 ENZYMATIC	203 ROCHE	93 ROCHE	11	16.545	2.061	12.46	10.000	23.000		
2023-785	2 ENZYMATIC	219 ATELLICA	143 SIEMENS	11	18.273	1.420	7.77	14.000	23.000		
170 DIRECT BILIRUBIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}								
2023-781	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	1.152	0.070	6.07	0.900	1.400		
2023-781	5 JEND.-GROG	197 ARCHITECT	1 ABBOTT	10	2.020	0.282	13.97	1.200	2.900		
2023-781	11 SPECT. WIT	165 VITROS	63 KODAK	67	0.702	0.141	20.08	0.300	1.100		
2023-781	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	20	1.050	0.092	8.78	0.800	1.300		
2023-781	14 VANADATE O	219 ATELLICA	143 SIEMENS	12	2.133	0.131	6.15	1.700	2.600		
2023-781	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	10	1.120	0.060	5.36	0.900	1.300		
2023-781	18 DIAZ. SULF	191 RX-SERIES	126 RANDOX	11	1.745	0.167	9.58	1.200	2.200		
2023-782	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.476	0.043	9.04	0.300	0.600		
2023-782	5 JEND.-GROG	197 ARCHITECT	1 ABBOTT	10	0.830	0.100	12.11	0.500	1.100		
2023-782	11 SPECT. WIT	165 VITROS	63 KODAK	67	0.073	0.143	196.72	0	0.500		
2023-782	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	20	0.430	0.064	14.89	0.200	0.600		
2023-782	14 VANADATE O	219 ATELLICA	143 SIEMENS	12	0.883	0.055	6.26	0.700	1.100		
2023-782	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	10	0.480	0.040	8.33	0.400	0.600		
2023-782	18 DIAZ. SULF	191 RX-SERIES	126 RANDOX	11	0.664	0.064	9.69	0.500	0.900		
2023-783	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.288	0.032	11.28	0.200	0.400		
2023-783	5 JEND.-GROG	197 ARCHITECT	1 ABBOTT	10	0.530	0.064	12.08	0.300	0.700		
2023-783	11 SPECT. WIT	165 VITROS	63 KODAK	67	0.027	0.080	299.42	0	0.300		
2023-783	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	20	0.285	0.036	12.53	0.200	0.400		
2023-783	14 VANADATE O	219 ATELLICA	143 SIEMENS	12	0.558	0.064	11.46	0.400	0.800		
2023-783	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	10	0.310	0.030	9.68	0.200	0.400		
2023-783	18 DIAZ. SULF	191 RX-SERIES	126 RANDOX	11	0.418	0.039	9.22	0.300	0.500		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
170	DIRECT BILIRUBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}								
2023-784	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	1.508	0.057	3.75	1.200	1.800		
2023-784	5 JEND.-GROG	197 ARCHITECT	1 ABBOTT	10	2.560	0.375	14.64	1.400	3.700		
2023-784	11 SPECT. WIT	165 VITROS	63 KODAK	67	1.000	0.168	16.82	0.500	1.500		
2023-784	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	20	1.350	0.112	8.28	1.000	1.700		
2023-784	14 VANADATE O	219 ATELLICA	143 SIEMENS	12	2.775	0.159	5.72	2.200	3.300		
2023-784	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	10	1.390	0.170	12.23	0.900	1.900		
2023-784	18 DIAZ. SULF	191 RX-SERIES	126 RANDOX	11	2.327	0.341	14.67	1.300	3.400		
2023-785	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	50	0.882	0.059	6.68	0.700	1.100		
2023-785	5 JEND.-GROG	197 ARCHITECT	1 ABBOTT	10	1.580	0.223	14.10	0.900	2.200		
2023-785	11 SPECT. WIT	165 VITROS	63 KODAK	67	0.440	0.115	26.08	0.100	0.800		
2023-785	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	20	0.810	0.089	10.97	0.500	1.100		
2023-785	14 VANADATE O	219 ATELLICA	143 SIEMENS	12	1.650	0.087	5.25	1.300	2.000		
2023-785	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	10	0.840	0.120	14.29	0.500	1.200		
2023-785	18 DIAZ. SULF	191 RX-SERIES	126 RANDOX	11	1.318	0.153	11.58	0.900	1.800		

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 15 WHOLE BLOOD GLUCOSE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
157	WHOLE BLOOD GLUCOSE		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}								
2023-881	20 ABBOT FREE			19	500.000	0.000	0.00	450.000	550.000		
2023-881	23 ABBOTT PRE			27	421.192	14.536	3.45	379.000	463.000		
2023-881	27 STAT SRIP.			52	267.442	23.434	8.76	241.000	294.000		
2023-881	33 AC.CHEK II			33	382.625	23.860	6.24	344.000	421.000		
2023-881	34 FREE STYL			46	476.089	28.242	5.93	428.000	524.000		
2023-882	20 ABBOT FREE			19	193.947	5.995	3.09	175.000	213.000		
2023-882	23 ABBOTT PRE			27	168.667	8.242	4.89	152.000	186.000		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 10 ROUTINE			GRP: 15 WHOLE BLOOD GLUCOSE					
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-882	27	STAT SRIP.		52	126.808	6.745	5.32	114.000	139.000		
2023-882	33	AC.CHEK II		33	149.645	7.403	4.95	135.000	165.000		
2023-882	34	FREE STYL		46	192.578	11.613	6.03	173.000	212.000		
157 WHOLE BLOOD GLUCOSE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}							
2023-883	20	ABBOT FREE		19	199.263	7.496	3.76	179.000	219.000		
2023-883	23	ABBOTT PRE		27	170.444	7.671	4.50	153.000	187.000		
2023-883	27	STAT SRIP.		52	127.481	6.963	5.46	115.000	140.000		
2023-883	33	AC.CHEK II		33	149.125	8.894	5.96	134.000	164.000		
2023-883	34	FREE STYL		46	191.711	13.135	6.85	173.000	211.000		
2023-884	20	ABBOT FREE		19	127.211	4.200	3.30	114.000	140.000		
2023-884	23	ABBOTT PRE		27	115.444	6.285	5.44	104.000	127.000		
2023-884	27	STAT SRIP.		52	94.902	4.442	4.68	85.000	104.000		
2023-884	33	AC.CHEK II		33	104.188	4.766	4.57	94.000	115.000		
2023-884	34	FREE STYL		46	128.089	8.369	6.53	115.000	141.000		
2023-885	20	ABBOT FREE		19	128.368	4.195	3.27	116.000	141.000		
2023-885	23	ABBOTT PRE		27	116.741	5.635	4.83	105.000	128.000		
2023-885	27	STAT SRIP.		52	94.942	4.012	4.23	85.000	104.000		
2023-885	33	AC.CHEK II		33	105.156	4.338	4.13	95.000	116.000		
2023-885	34	FREE STYL		46	127.733	8.067	6.32	115.000	141.000		
SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES					

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE	8	% OR 5.00	{GREATER}			
2023-831	265	ROCHE COBA			32	65.415	1.055	1.61	60.200	70.600
2023-831	265	ROCHE COBA			32	66.250	2.165	3.27	61.000	71.600
2023-832	265	ROCHE COBA			32	46.000	0.000	0.00	41.000	51.000
2023-832	265	ROCHE COBA			32	47.930	0.667	1.39	42.900	52.900
2023-833	265	ROCHE COBA			32	21.470	0.602	2.80	16.500	26.500
2023-833	265	ROCHE COBA			32	19.500	0.500	2.56	14.500	24.500
2023-834	265	ROCHE COBA			32	47.750	1.090	2.28	42.800	52.800
2023-834	265	ROCHE COBA			32	47.567	0.557	1.17	42.600	52.600
2023-835	265	ROCHE COBA			32	54.889	1.949	3.55	49.900	59.900
2023-835	265	ROCHE COBA			32	53.250	0.433	0.81	48.300	58.300
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04					
2023-831	265	ROCHE COBA			32	7.230	0.006	0.08	7.191	7.271
2023-832	265	ROCHE COBA			32	7.372	0.003	0.04	7.332	7.412
2023-833	265	ROCHE COBA			32	7.677	0.006	0.08	7.637	7.717
2023-834	265	ROCHE COBA			32	7.374	0.005	0.07	7.334	7.414
2023-835	265	ROCHE COBA			32	7.231	0.008	0.11	7.191	7.271
009	BLOOD GASES-pO2		CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE	3	SD				
2023-831	265	ROCHE COBA			32	55.588	10.337	18.60	24.600	86.600
2023-831	265	ROCHE COBA			32	75.000	3.391	4.52	64.800	85.200

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
009	BLOOD GASES-pO2	CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2023-832	265 ROCHE COBA			32	112.000	4.637	4.14	98.100	125.900	
2023-832	265 ROCHE COBA			32	107.854	3.859	3.58	96.300	119.400	
2023-833	265 ROCHE COBA			32	153.796	8.127	5.28	129.400	178.200	
2023-833	265 ROCHE COBA			32	159.250	3.192	2.00	149.700	168.800	
2023-834	265 ROCHE COBA			32	110.750	3.700	3.34	99.700	121.800	
2023-834	265 ROCHE COBA			32	106.615	3.629	3.40	95.700	117.500	
2023-835	265 ROCHE COBA			32	281.774	29.259	10.38	194.000	369.600	
2023-835	265 ROCHE COBA			32	256.000	14.765	5.77	211.700	300.300	

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0								
2023-811	116 ROCHE COBA			15	30.433	1.296	4.26	26.500	34.300	
2023-811	125 B.C.ACCESS			10	32.300	2.136	6.61	25.900	38.700	
2023-811	149 ARC i2000s			11	30.209	1.943	6.43	24.400	36.000	
2023-812	116 ROCHE COBA			15	15.733	0.765	4.86	13.400	18.000	
2023-812	125 B.C.ACCESS			10	15.810	1.088	6.88	12.500	19.100	
2023-812	149 ARC i2000s			11	15.136	1.364	9.01	11.000	19.200	
2023-813	116 ROCHE COBA			15	12.340	0.598	4.84	10.500	14.100	
2023-813	125 B.C.ACCESS			10	12.330	1.027	8.33	9.200	15.400	
2023-813	149 ARC i2000s			11	11.809	0.921	7.80	9.000	14.600	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2023-814	116 ROCHE COBA			15	39.533	1.479	3.74	35.100	44.000	
2023-814	125 B.C.ACCESS			10	40.450	2.706	6.69	32.300	48.600	
2023-814	149 ARC i2000s			11	40.436	2.930	7.24	31.600	49.200	
2023-815	116 ROCHE COBA			15	24.000	1.100	4.58	20.700	27.300	
2023-815	125 B.C.ACCESS			10	25.220	0.842	3.34	22.700	27.700	
2023-815	149 ARC i2000s			11	23.618	1.890	8.00	17.900	29.300	
030	FREE THYROXINE	CLIA TEST ID #: 545								
2023-811	115 VITROS ECI			44	6.256	0.239	3.81	5.500	7.000	
2023-811	116 ROCHE COBA			46	2.850	0.133	4.67	2.500	3.200	
2023-811	125 B.C.ACCESS			45	2.261	0.071	3.16	2.000	2.500	
2023-811	149 ARC i2000s			23	2.113	0.108	5.09	1.800	2.400	
2023-811	160 SIEMENS EX			51	2.644	0.102	3.87	2.300	3.000	
2023-811	169 ATELLICA			10	2.280	0.098	4.30	2.000	2.600	
2023-812	115 VITROS ECI			44	3.763	0.199	5.28	3.200	4.400	
2023-812	116 ROCHE COBA			46	1.739	0.082	4.72	1.500	2.000	
2023-812	125 B.C.ACCESS			45	1.507	0.071	4.72	1.300	1.700	
2023-812	149 ARC i2000s			23	1.391	0.072	5.15	1.200	1.600	
2023-812	160 SIEMENS EX			51	1.612	0.076	4.70	1.400	1.800	
2023-812	169 ATELLICA			10	1.440	0.080	5.56	1.200	1.700	
2023-813	115 VITROS ECI			44	2.900	0.188	6.48	2.300	3.500	
2023-813	116 ROCHE COBA			46	1.446	0.068	4.72	1.200	1.700	
2023-813	125 B.C.ACCESS			45	1.286	0.055	4.25	1.100	1.500	
2023-813	149 ARC i2000s			23	1.196	0.062	5.22	1.000	1.400	
2023-813	160 SIEMENS EX			51	1.335	0.071	5.31	1.100	1.500	
2023-813	169 ATELLICA			10	1.210	0.054	4.45	1.000	1.400	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY SUB: 30 ENDOCRINOLOGY GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
030	FREE THYROXINE	CLIA TEST ID #: 545	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2023-814	115 VITROS ECI			44	6.963	0.057	0.82	6.800	7.100	
2023-814	116 ROCHE COBA			46	3.538	0.164	4.63	3.000	4.000	
2023-814	125 B.C.ACCESS			45	2.602	0.106	4.09	2.300	2.900	
2023-814	149 ARC i2000s			23	2.409	0.144	5.99	2.000	2.800	
2023-814	160 SIEMENS EX			51	3.137	0.133	4.23	2.700	3.500	
2023-814	169 ATELLICA			10	2.740	0.111	4.06	2.400	3.100	
2023-815	115 VITROS ECI			44	5.426	0.200	3.69	4.800	6.000	
2023-815	116 ROCHE COBA			46	2.432	0.126	5.17	2.100	2.800	
2023-815	125 B.C.ACCESS			45	2.009	0.079	3.94	1.800	2.200	
2023-815	149 ARC i2000s			23	1.883	0.092	4.87	1.600	2.200	
2023-815	160 SIEMENS EX			51	2.186	0.089	4.05	1.900	2.500	
2023-815	169 ATELLICA			10	1.940	0.066	3.42	1.700	2.100	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555								
2023-811	115 VITROS ECI			20	711.405	31.502	4.43	616.900	805.900	
2023-811	116 ROCHE COBA			17	582.576	29.772	5.11	493.300	671.900	
2023-811	125 B.C.ACCESS			11	570.536	21.058	3.69	507.400	633.700	
2023-812	115 VITROS ECI			20	259.545	10.174	3.92	229.000	290.100	
2023-812	116 ROCHE COBA			17	214.682	10.703	4.99	182.600	246.800	
2023-812	125 B.C.ACCESS			11	222.718	8.408	3.78	197.500	247.900	
2023-813	115 VITROS ECI			20	155.490	7.242	4.66	133.800	177.200	
2023-813	116 ROCHE COBA			17	125.812	5.946	4.73	108.000	143.600	
2023-813	125 B.C.ACCESS			11	133.191	4.375	3.28	120.100	146.300	
2023-814	115 VITROS ECI			20	980.760	45.888	4.68	843.100	1118.400	
2023-814	116 ROCHE COBA			17	827.571	32.873	3.97	729.000	926.200	
2023-814	125 B.C.ACCESS			11	782.045	26.097	3.34	703.800	860.300	

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
032	hCG-QUANTITATIVE		CLIA TEST ID #: 555	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2023-815	115	VITROS ECI		20	514.890	22.519	4.37	447.300	582.400	
2023-815	116	ROCHE COBA		17	423.406	20.008	4.73	363.400	483.400	
2023-815	125	B.C.ACCESS		11	423.073	12.961	3.06	384.200	462.000	
035	PROLACTIN		CLIA TEST ID #: 0							
2023-811	116	ROCHE COBA		16	30.463	1.530	5.02	25.900	35.100	
2023-811	125	B.C.ACCESS		11	27.345	1.786	6.53	22.000	32.700	
2023-812	116	ROCHE COBA		16	13.088	0.720	5.50	10.900	15.200	
2023-812	125	B.C.ACCESS		11	12.591	0.585	4.65	10.800	14.300	
2023-813	116	ROCHE COBA		16	9.275	0.502	5.41	7.800	10.800	
2023-813	125	B.C.ACCESS		11	8.845	0.509	5.75	7.300	10.400	
2023-814	116	ROCHE COBA		16	42.138	2.259	5.36	35.400	48.900	
2023-814	125	B.C.ACCESS		11	35.764	2.660	7.44	27.800	43.700	
2023-815	116	ROCHE COBA		16	22.856	1.309	5.73	18.900	26.800	
2023-815	125	B.C.ACCESS		11	20.773	1.024	4.93	17.700	23.800	
037	TESTOSTERONE									
2023-811	149	ARC i2000s		10	12.540	1.356	10.81	8.500	16.600	
2023-812	149	ARC i2000s		10	4.770	0.393	8.23	3.600	5.900	
2023-813	149	ARC i2000s		10	3.750	1.365	36.41	0	7.800	
2023-814	149	ARC i2000s		10	14.020	1.960	13.98	8.100	19.900	
2023-815	149	ARC i2000s		10	9.000	0.714	7.93	6.900	11.100	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY SUB: 30 ENDOCRINOLOGY GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2023-811	106 TOSOH AIA			11	6.736	0.891	13.22	4.100	9.400
2023-811	115 VITROS ECI			51	6.594	0.200	3.04	6.000	7.200
2023-811	116 ROCHE COBA			51	5.806	0.240	4.13	5.100	6.500
2023-811	125 B.C.ACCESS			57	5.474	0.279	5.09	4.600	6.300
2023-811	149 ARC i2000s			25	4.816	0.289	6.01	3.900	5.700
2023-811	160 SIEMENS EX			82	4.833	0.308	6.37	3.900	5.800
2023-811	169 ATELLICA			11	5.845	0.363	6.21	4.800	6.900
2023-812	106 TOSOH AIA			11	2.955	0.382	12.94	1.800	4.100
2023-812	115 VITROS ECI			51	3.038	0.106	3.48	2.700	3.400
2023-812	116 ROCHE COBA			51	2.812	0.145	5.16	2.400	3.200
2023-812	125 B.C.ACCESS			57	2.430	0.131	5.39	2.000	2.800
2023-812	149 ARC i2000s			25	2.183	0.125	5.71	1.800	2.600
2023-812	160 SIEMENS EX			82	2.227	0.138	6.20	1.800	2.600
2023-812	169 ATELLICA			11	2.673	0.148	5.55	2.200	3.100
2023-813	106 TOSOH AIA			11	2.155	0.215	9.97	1.500	2.800
2023-813	115 VITROS ECI			51	2.118	0.069	3.26	1.900	2.300
2023-813	116 ROCHE COBA			51	2.057	0.083	4.05	1.800	2.300
2023-813	125 B.C.ACCESS			57	1.735	0.098	5.66	1.400	2.000
2023-813	149 ARC i2000s			25	1.524	0.103	6.76	1.200	1.800
2023-813	160 SIEMENS EX			82	1.615	0.099	6.13	1.300	1.900
2023-813	169 ATELLICA			11	1.900	0.095	5.02	1.600	2.200
2023-814	106 TOSOH AIA			11	8.836	1.012	11.45	5.800	11.900
2023-814	115 VITROS ECI			51	8.690	0.315	3.63	7.700	9.600
2023-814	116 ROCHE COBA			51	7.414	0.297	4.00	6.500	8.300
2023-814	125 B.C.ACCESS			57	7.484	0.459	6.13	6.100	8.900
2023-814	149 ARC i2000s			25	6.488	0.350	5.40	5.400	7.500

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)			
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023-814	160	SIEMENS EX		82	6.329	0.334	5.28	5.300	7.300
2023-814	169	ATELLICA		11	7.609	0.423	5.56	6.300	8.900
039	THYRO. STI. HOR.-TSH		CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2023-815	106	TOSOH AIA		11	5.200	0.632	12.16	3.300	7.100
2023-815	115	VITROS ECI		51	5.138	0.162	3.16	4.700	5.600
2023-815	116	ROCHE COBA		51	4.549	0.176	3.87	4.000	5.100
2023-815	125	B.C.ACCESS		57	4.169	0.221	5.31	3.500	4.800
2023-815	149	ARC i2000s		25	3.688	0.217	5.87	3.000	4.300
2023-815	160	SIEMENS EX		82	3.774	0.210	5.56	3.100	4.400
2023-815	169	ATELLICA		11	4.473	0.200	4.48	3.900	5.100
040	THYROXINE (T4)		CLIA TEST ID #: 595	EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}					
2023-811	32	SIEM DIME.		51	13.061	0.669	5.12	10.400	15.700
2023-811	115	VITROS ECI		41	11.200	0.468	4.18	9.000	13.400
2023-811	116	ROCHE COBA		47	11.417	0.600	5.25	9.100	13.700
2023-811	125	B.C.ACCESS		46	11.835	0.824	6.96	9.500	14.200
2023-811	149	ARC i2000s		23	11.341	0.474	4.18	9.100	13.600
2023-811	169	ATELLICA		11	12.364	0.555	4.49	9.900	14.800
2023-812	32	SIEM DIME.		51	6.994	0.462	6.61	5.600	8.400
2023-812	115	VITROS ECI		41	6.315	0.282	4.47	5.100	7.600
2023-812	116	ROCHE COBA		47	6.539	0.372	5.69	5.200	7.800
2023-812	125	B.C.ACCESS		46	6.909	0.441	6.38	5.500	8.300
2023-812	149	ARC i2000s		23	5.965	0.287	4.81	4.800	7.200
2023-812	169	ATELLICA		11	7.345	0.380	5.17	5.900	8.800
2023-813	32	SIEM DIME.		51	5.680	0.474	8.35	4.500	6.800
2023-813	115	VITROS ECI		41	5.133	0.227	4.42	4.100	6.200
2023-813	116	ROCHE COBA		47	5.391	0.361	6.70	4.300	6.500
2023-813	125	B.C.ACCESS		46	5.554	0.419	7.55	4.400	6.700
2023-813	149	ARC i2000s		23	4.752	0.252	5.30	3.800	5.800

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-813	169	ATELLICA		11	6.000	0.393	6.55	4.800	7.200	
040	THYROXINE (T4)		CLIA TEST ID #: 595		EVALUATION CRITERIA: TARGET VALUE	20 %	OR 1.00	{GREATER}		
2023-814	32	SIEM DIME.		51	16.852	0.761	4.51	13.500	20.200	
2023-814	115	VITROS ECI		41	13.923	0.600	4.31	11.100	16.700	
2023-814	116	ROCHE COBA		47	14.628	0.771	5.27	11.700	17.600	
2023-814	125	B.C.ACCESS		46	14.384	0.987	6.86	11.500	17.300	
2023-814	149	ARC i2000s		23	14.657	0.975	6.65	11.700	17.600	
2023-814	169	ATELLICA		11	14.573	0.564	3.87	11.700	17.500	
2023-815	32	SIEM DIME.		51	10.365	0.493	4.76	8.300	12.400	
2023-815	115	VITROS ECI		41	9.205	0.411	4.47	7.400	11.000	
2023-815	116	ROCHE COBA		47	9.480	0.540	5.70	7.600	11.400	
2023-815	125	B.C.ACCESS		46	9.959	0.656	6.59	8.000	12.000	
2023-815	149	ARC i2000s		23	9.209	0.386	4.19	7.400	11.100	
2023-815	169	ATELLICA		11	10.455	0.677	6.48	8.400	12.500	
141	FOLATE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE	25 %				
2023-811	115	VITROS ECI		14	7.529	0.779	10.34	5.600	9.400	
2023-811	116	ROCHE COBA		20	6.974	0.647	9.28	5.200	8.700	
2023-811	125	B.C.ACCESS		13	9.046	0.589	6.51	6.800	11.300	
2023-811	149	ARC i2000s		13	9.592	0.815	8.49	7.200	12.000	
2023-812	115	VITROS ECI		14	2.757	0.176	6.39	2.100	3.400	
2023-812	116	ROCHE COBA		20	2.621	0.352	13.42	2.000	3.300	
2023-812	125	B.C.ACCESS		13	3.838	0.329	8.58	2.900	4.800	
2023-812	149	ARC i2000s		13	4.762	2.448	51.42	3.600	6.000	
2023-813	115	VITROS ECI		14	1.964	0.129	6.56	1.500	2.500	
2023-813	116	ROCHE COBA		20	2.042	0.067	3.30	1.500	2.600	
2023-813	125	B.C.ACCESS		13	2.585	0.268	10.39	1.900	3.200	
2023-813	149	ARC i2000s		13	2.554	0.282	11.03	1.900	3.200	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
141 FOLATE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %						
2023-814	115 VITROS ECI			14	11.043	1.190	10.77	8.300	13.800
2023-814	116 ROCHE COBA			20	10.953	1.851	16.90	8.200	13.700
2023-814	125 B.C.ACCESS			13	12.331	1.136	9.22	9.200	15.400
2023-814	149 ARC i2000s			13	13.638	0.913	6.69	10.200	17.000
2023-815	115 VITROS ECI			14	5.329	0.576	10.81	4.000	6.700
2023-815	116 ROCHE COBA			20	5.137	1.170	22.78	3.900	6.400
2023-815	125 B.C.ACCESS			13	6.777	0.519	7.66	5.100	8.500
2023-815	149 ARC i2000s			13	6.892	0.487	7.07	5.200	8.600
142 hLH (HUMAN LUT. HOR)									
2023-811	116 ROCHE COBA			13	31.469	1.196	3.80	23.600	39.300
2023-811	125 B.C.ACCESS			10	26.580	1.190	4.48	19.900	33.200
2023-811	149 ARC i2000s			10	20.890	1.507	7.21	15.700	26.100
2023-812	116 ROCHE COBA			13	15.269	0.486	3.18	11.500	19.100
2023-812	125 B.C.ACCESS			10	12.830	0.420	3.27	9.600	16.000
2023-812	149 ARC i2000s			10	9.490	0.575	6.06	7.100	11.900
2023-813	116 ROCHE COBA			13	11.346	0.397	3.50	8.500	14.200
2023-813	125 B.C.ACCESS			10	9.150	0.403	4.41	6.900	11.400
2023-813	149 ARC i2000s			10	6.990	0.543	7.77	5.200	8.700
2023-814	116 ROCHE COBA			13	40.877	1.661	4.06	30.700	51.100
2023-814	125 B.C.ACCESS			10	32.790	0.966	2.95	24.600	41.000
2023-814	149 ARC i2000s			10	27.610	2.883	10.44	20.700	34.500
2023-815	116 ROCHE COBA			13	24.615	1.031	4.19	18.500	30.800
2023-815	125 B.C.ACCESS			10	20.250	1.124	5.55	15.200	25.300
2023-815	149 ARC i2000s			10	16.080	1.339	8.32	12.100	20.100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
031 hCG-QUALITATIVE												
2023-811	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-811	142	INSTANT VI		002	002	002	002	002 - POSITIVE	38	38	38	100
2023-811	150	AIM STEP		002	002	002	002	002 - POSITIVE	96	96	96	100
2023-811	164	HCG Alere		002	002	002	002	001 - NEGATIVE	72	1	71	0
2023-811	164	HCG Alere		002	002	002	002	002 - POSITIVE	72	71	71	100
2023-811	165	Osom HCG C		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-811	166	FaStep		002	002	002	002	002 - POSITIVE	19	19	19	100
2023-811	168	STANBIO 20		002	002	002	002	002 - POSITIVE	14	14	14	100
2023-812	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-812	142	INSTANT VI		002	002	002	002	002 - POSITIVE	38	38	38	100
2023-812	150	AIM STEP		002	002	002	002	002 - POSITIVE	96	96	96	100
2023-812	164	HCG Alere		002	002	002	002	001 - NEGATIVE	72	1	71	0
2023-812	164	HCG Alere		002	002	002	002	002 - POSITIVE	72	71	71	100
2023-812	165	Osom HCG C		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-812	166	FaStep		002	002	002	002	002 - POSITIVE	19	19	19	100
2023-812	168	STANBIO 20		002	002	002	002	002 - POSITIVE	14	14	14	100
2023-813	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-813	142	INSTANT VI		002	002	002	002	002 - POSITIVE	38	38	38	100
2023-813	150	AIM STEP		002	002	002	002	002 - POSITIVE	96	96	96	100
2023-813	164	HCG Alere		002	002	002	002	001 - NEGATIVE	72	1	71	0
2023-813	164	HCG Alere		002	002	002	002	002 - POSITIVE	72	71	71	100
2023-813	165	Osom HCG C		002	002	002	002	002 - POSITIVE	23	23	23	100
2023-813	166	FaStep		002	002	002	002	002 - POSITIVE	19	19	19	100
2023-813	168	STANBIO 20		002	002	002	002	002 - POSITIVE	14	14	14	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
031 hCG-QUALITATIVE											
2023-814	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	23	23	23	100
2023-814	142	INSTANT VI		002 002	002 002		002 - POSITIVE	38	38	38	100
2023-814	150	AIM STEP		002 002	002 002		002 - POSITIVE	96	96	96	100
2023-814	164	HCG Alere		002 002	002 002		002 - POSITIVE	72	72	72	100
2023-814	165	Osom HCG C		002 002	002 002		002 - POSITIVE	23	23	23	100
2023-814	166	FaStep		002 002	002 002		002 - POSITIVE	19	19	19	100
2023-814	168	STANBIO 20		002 002	002 002		002 - POSITIVE	14	14	14	100
2023-815	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	23	23	23	100
2023-815	142	INSTANT VI		002 002	002 002		002 - POSITIVE	38	38	38	100
2023-815	150	AIM STEP		002 002	002 002		002 - POSITIVE	96	96	96	100
2023-815	164	HCG Alere		002 002	002 002		002 - POSITIVE	72	72	72	100
2023-815	165	Osom HCG C		002 002	002 002		002 - POSITIVE	23	23	23	100
2023-815	166	FaStep		002 002	002 002		002 - POSITIVE	19	19	19	100
2023-815	168	STANBIO 20		002 002	002 002		002 - POSITIVE	14	14	14	100
SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
151 hCG ONLY QUALITATIVE											
2023-801	018	BIO SIGN		001 001	001 001		001 - NEGATIVE	24	24	24	100
2023-801	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	80	80	80	100
2023-801	118	AIM STEP		001 001	001 001		001 - NEGATIVE	191	191	191	100
2023-801	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	92	92	92	100
2023-801	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	29	29	29	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-801	133	FaStep		001 001	001 001		001 - NEGATIVE	43	43	43	100	
2023-801	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	51	51	51	100	
2023-801	136	CLARITY		001 001	001 001		001 - NEGATIVE	16	16	16	100	
151 hCG ONLY QUALITATIVE												
2023-802	018	BIO SIGN		002 002	002 002		002 - POSITIVE	24	24	24	100	
2023-802	114	INSTANT VI		002 002	002 002		002 - POSITIVE	80	80	80	100	
2023-802	118	AIM STEP		002 002	002 002		002 - POSITIVE	191	191	191	100	
2023-802	130	ALERE HCG		002 002	002 002		002 - POSITIVE	92	92	92	100	
2023-802	131	OSOM HCG		002 002	002 002		002 - POSITIVE	29	29	29	100	
2023-802	133	FaStep		002 002	002 002		002 - POSITIVE	43	43	43	100	
2023-802	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	51	51	51	100	
2023-802	136	CLARITY		002 002	002 002		002 - POSITIVE	16	16	16	100	
2023-803	018	BIO SIGN		001 001	001 001		001 - NEGATIVE	24	24	24	100	
2023-803	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	80	80	80	100	
2023-803	118	AIM STEP		001 001	001 001		001 - NEGATIVE	191	191	191	100	
2023-803	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	92	92	92	100	
2023-803	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	29	29	29	100	
2023-803	133	FaStep		001 001	001 001		001 - NEGATIVE	43	43	43	100	
2023-803	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	51	51	51	100	
2023-803	136	CLARITY		001 001	001 001		001 - NEGATIVE	16	16	16	100	
2023-804	018	BIO SIGN		002 002	002 002		002 - POSITIVE	24	24	24	100	
2023-804	114	INSTANT VI		002 002	002 002		002 - POSITIVE	80	80	80	100	
2023-804	118	AIM STEP		002 002	002 002		002 - POSITIVE	191	191	191	100	
2023-804	130	ALERE HCG		002 002	002 002		001 - NEGATIVE	92	1	91	0	
2023-804	130	ALERE HCG		002 002	002 002		002 - POSITIVE	92	91	91	100	
2023-804	131	OSOM HCG		002 002	002 002		002 - POSITIVE	29	29	29	100	
2023-804	133	FaStep		002 002	002 002		002 - POSITIVE	43	43	43	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-804	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	51	51	51	100	
2023-804	136	CLARITY		002 002	002 002		002 - POSITIVE	16	16	16	100	
151 hCG ONLY QUALITATIVE												
2023-805	018	BIO SIGN		002 002	002 002		002 - POSITIVE	24	24	24	100	
2023-805	114	INSTANT VI		002 002	002 002		002 - POSITIVE	80	80	80	100	
2023-805	118	AIM STEP		002 002	002 002		002 - POSITIVE	191	191	191	100	
2023-805	130	ALERE HCG		002 002	002 002		001 - NEGATIVE	92	1	91	0	
2023-805	130	ALERE HCG		002 002	002 002		002 - POSITIVE	92	91	91	100	
2023-805	131	OSOM HCG		002 002	002 002		002 - POSITIVE	29	29	29	100	
2023-805	133	FaStep		002 002	002 002		002 - POSITIVE	43	43	43	100	
2023-805	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	51	51	51	100	
2023-805	136	CLARITY		002 002	002 002		002 - POSITIVE	16	16	16	100	

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
173	T3 UPTAKE	CLIA TEST ID #: 565	EVALUATION CRITERIA: TARGET VALUE 3 SD									
2023-811	1 %U	84 VITROS ECI	1 25-35	36	51.586	1.423	2.76	47.320	55.850			
2023-811	1 %U	91 B.C.ACCESS	2 35-45	29	46.493	2.807	6.04	38.070	54.920			
2023-811	2 TUV	95 Arc i2000s	3 0.7 - 1.2	17	0.943	0.076	8.02	0.720	1.170			
2023-811	4 %	32 DIMENSION	2 35-45	58	40.028	1.800	4.50	34.630	45.430			
2023-811	4 %	91 B.C.ACCESS	2 35-45	13	46.508	3.022	6.50	37.440	55.570			
2023-811	4 %	106 ATELLICA	1 25-35	10	36.995	1.905	5.15	31.280	42.710			
2023-811	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.667	0.065	9.73	0.470	0.860			
2023-812	1 %U	84 VITROS ECI	1 25-35	36	57.061	1.201	2.10	53.460	60.660			
2023-812	1 %U	91 B.C.ACCESS	2 35-45	29	49.603	2.634	5.31	41.700	57.510			
2023-812	2 TUV	95 Arc i2000s	3 0.7 - 1.2	17	0.716	0.070	9.81	0.510	0.930			

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023-812	4 %	32 DIMENSION	2 35-45	58	38.984	1.330	3.41	34.990	42.970			
2023-812	4 %	91 B.C.ACCESS	2 35-45	13	50.308	3.212	6.38	40.670	59.940			
2023-812	4 %	106 ATELLICA	1 25-35	10	45.334	2.947	6.50	36.490	54.180			
2023-812	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.783	0.067	8.59	0.580	0.980			
173 T3 UPTAKE		CLIA TEST ID #: 565	EVALUATION CRITERIA: TARGET VALUE 3 SD									
2023-813	1 %U	84 VITROS ECI	1 25-35	36	58.531	1.421	2.43	54.270	62.790			
2023-813	1 %U	91 B.C.ACCESS	2 35-45	29	49.879	2.899	5.81	41.180	58.580			
2023-813	2 TUV	95 Arc i2000s	3 0.7 - 1.2	17	0.663	0.064	9.59	0.470	0.850			
2023-813	4 %	32 DIMENSION	2 35-45	58	38.915	1.285	3.30	35.060	42.770			
2023-813	4 %	91 B.C.ACCESS	2 35-45	13	51.331	3.151	6.14	41.880	60.780			
2023-813	4 %	106 ATELLICA	1 25-35	10	47.247	2.851	6.03	38.700	55.800			
2023-813	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.801	0.074	9.28	0.580	1.020			
2023-814	1 %U	84 VITROS ECI	1 25-35	36	48.657	1.329	2.73	44.670	52.650			
2023-814	1 %U	91 B.C.ACCESS	2 35-45	29	45.162	2.473	5.47	37.740	52.580			
2023-814	2 TUV	95 Arc i2000s	3 0.7 - 1.2	17	1.065	0.088	8.30	0.800	1.330			
2023-814	4 %	32 DIMENSION	2 35-45	58	39.922	1.404	3.52	35.710	44.130			
2023-814	4 %	91 B.C.ACCESS	2 35-45	13	45.900	2.944	6.41	37.070	54.730			
2023-814	4 %	106 ATELLICA	1 25-35	10	33.284	1.872	5.63	27.670	38.900			
2023-814	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.589	0.090	15.32	0.320	0.860			
2023-815	1 %U	84 VITROS ECI	1 25-35	36	53.506	1.415	2.65	49.260	57.750			
2023-815	1 %U	91 B.C.ACCESS	2 35-45	29	47.745	2.761	5.78	39.460	56.030			
2023-815	2 TUV	95 Arc i2000s	3 0.7 - 1.2	17	0.871	0.075	8.63	0.650	1.100			
2023-815	4 %	32 DIMENSION	2 35-45	58	39.371	1.271	3.23	35.560	43.180			
2023-815	4 %	91 B.C.ACCESS	2 35-45	13	48.208	3.048	6.32	39.060	57.350			
2023-815	4 %	106 ATELLICA	1 25-35	10	39.461	1.946	4.93	33.620	45.300			
2023-815	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.711	0.079	11.16	0.470	0.950			

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
043 DIGOXIN		CLIA TEST ID #: 645	EVALUATION CRITERIA: TARGET VALUE	20	% OR 0.20	{GREATER}				
2023-811	50 VITROS			15	1.280	0.083	6.51	1.000	1.500	
2023-812	50 VITROS			15	0.620	0.147	23.70	0.400	0.800	
2023-813	50 VITROS			15	0.560	0.095	17.00	0.400	0.800	
2023-814	50 VITROS			15	1.707	0.106	6.23	1.400	2.000	
2023-815	50 VITROS			15	0.967	0.094	9.75	0.800	1.200	
045 FERRITIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%					
2023-811	50 VITROS			18	114.706	5.154	4.49	86.000	143.400	
2023-811	90 ROCHE COBA			14	193.450	12.345	6.38	145.100	241.800	
2023-811	98 ARC i200sr			10	183.340	15.839	8.64	137.500	229.200	
2023-811	105 ATELLICA			10	132.900	5.570	4.19	99.700	166.100	
2023-812	50 VITROS			18	49.694	2.429	4.89	37.300	62.100	
2023-812	90 ROCHE COBA			14	87.679	8.205	9.36	65.800	109.600	
2023-812	98 ARC i200sr			10	82.370	4.966	6.03	61.800	103.000	
2023-812	105 ATELLICA			10	55.850	2.732	4.89	41.900	69.800	
2023-813	50 VITROS			18	36.759	1.893	5.15	27.600	45.900	
2023-813	90 ROCHE COBA			14	64.579	5.590	8.66	48.400	80.700	
2023-813	98 ARC i200sr			10	58.870	4.004	6.80	44.200	73.600	
2023-813	105 ATELLICA			10	40.420	0.928	2.30	30.300	50.500	
2023-814	50 VITROS			18	158.167	8.362	5.29	118.600	197.700	
2023-814	90 ROCHE COBA			14	259.764	16.551	6.37	194.800	324.700	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-814	98	ARC i200sr		10	255.630	18.755	7.34	191.700	319.500	
2023-814	105	ATELLICA		10	179.140	14.481	8.08	134.400	223.900	
045	FERRITIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %					
2023-815	50	VITROS		18	87.711	3.998	4.56	65.800	109.600	
2023-815	90	ROCHE COBA		14	146.464	11.391	7.78	109.800	183.100	
2023-815	98	ARC i200sr		10	144.300	10.632	7.37	108.200	180.400	
2023-815	105	ATELLICA		10	97.900	2.207	2.25	73.400	122.400	
046	GENTAMICIN		CLIA TEST ID #: 665							
2023-811	50	VITROS		15	10.293	0.687	6.67	7.700	12.900	
2023-812	50	VITROS		15	3.787	0.131	3.46	2.800	4.700	
2023-813	50	VITROS		15	2.360	0.131	5.54	1.800	3.000	
2023-814	50	VITROS		15	14.607	1.698	11.62	11.000	18.300	
2023-815	50	VITROS		15	7.540	0.435	5.77	5.700	9.400	
050	PHENYTOIN		CLIA TEST ID #: 695							
2023-811	50	VITROS		15	20.933	0.812	3.88	15.700	26.200	
2023-811	92	ROCHE COBA		10	18.770	0.911	4.85	14.100	23.500	
2023-812	50	VITROS		15	6.747	0.403	5.97	5.100	8.400	
2023-812	92	ROCHE COBA		10	6.930	0.520	7.50	5.200	8.700	
2023-813	50	VITROS		15	3.900	0.341	8.73	2.900	4.900	
2023-813	92	ROCHE COBA		10	4.060	0.273	6.72	3.000	5.100	
2023-814	50	VITROS		15	32.013	3.008	9.40	24.000	40.000	
2023-814	92	ROCHE COBA		10	26.610	1.449	5.45	20.000	33.300	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
050	PHENYTOIN	CLIA TEST ID #: 695		EVALUATION CRITERIA: TARGET VALUE 25 %						
2023-815	50 VITROS			15	14.680	0.751	5.12	11.000	18.400	
2023-815	92 ROCHE COBA			10	14.090	0.797	5.66	10.600	17.600	
054	THEOPHYLLINE	CLIA TEST ID #: 735								
2023-811	50 VITROS			12	38.842	1.491	3.84	29.100	48.600	
2023-812	50 VITROS			12	16.508	0.928	5.62	12.400	20.600	
2023-813	50 VITROS			12	11.083	0.751	6.78	8.300	13.900	
2023-814	50 VITROS			12	51.458	2.760	5.36	38.600	64.300	
2023-815	50 VITROS			12	29.817	1.492	5.00	22.400	37.300	
127	VANCOMYCIN	CLIA TEST ID #: 0								
2023-811	50 VITROS			19	34.774	1.107	3.18	26.100	43.500	
2023-811	92 ROCHE COBA			11	35.082	3.124	8.91	26.300	43.900	
2023-812	50 VITROS			19	12.484	0.502	4.02	9.400	15.600	
2023-812	92 ROCHE COBA			11	13.164	0.707	5.37	9.900	16.500	
2023-813	50 VITROS			19	7.958	0.366	4.60	6.000	9.900	
2023-813	92 ROCHE COBA			11	7.982	0.732	9.17	6.000	10.000	
2023-814	50 VITROS			19	49.916	1.900	3.81	37.400	62.400	
2023-814	92 ROCHE COBA			11	49.764	2.625	5.28	37.300	62.200	
2023-815	50 VITROS			19	24.432	0.962	3.94	18.300	30.500	
2023-815	92 ROCHE COBA			11	25.845	1.043	4.04	19.400	32.300	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
065	URINA-PROTEIN QUANT.	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	% OR 5.00	{GREATER}				
2023-841	58	VITROS		40	4.974	0.160	3.22	0.000	10.000	
2023-841	62	SIEM DIM.		51	16.286	1.637	10.05	11.300	21.300	
2023-841	79	BECK. COUL		20	4.763	0.808	16.96	0	9.800	
2023-841	93	ARCHITECT		19	6.794	0.023	0.34	1.800	11.800	
2023-841	113	ROCHE COBA		34	4.012	0.279	6.96	0	9.000	
2023-841	146	ATELLICA		10	14.456	1.901	13.15	9.500	19.500	
2023-842	58	VITROS		40	5.846	1.051	17.98	0.800	10.800	
2023-842	62	SIEM DIM.		51	26.080	1.968	7.54	19.600	32.600	
2023-842	79	BECK. COUL		20	6.120	1.650	26.96	1.100	11.100	
2023-842	93	ARCHITECT		19	6.706	0.417	6.22	1.700	11.700	
2023-842	113	ROCHE COBA		34	3.788	0.591	15.59	0	8.800	
2023-842	146	ATELLICA		10	25.656	3.703	14.43	19.200	32.100	
2023-843	58	VITROS		40	157.051	9.212	5.87	117.800	196.300	
2023-843	62	SIEM DIM.		51	112.962	2.945	2.61	84.700	141.200	
2023-843	79	BECK. COUL		20	101.285	4.588	4.53	76.000	126.600	
2023-843	93	ARCHITECT		19	97.900	2.187	2.23	73.400	122.400	
2023-843	113	ROCHE COBA		34	90.647	1.688	1.86	68.000	113.300	
2023-843	146	ATELLICA		10	111.389	4.204	3.77	83.500	139.200	
2023-844	58	VITROS		40	158.769	13.371	8.42	119.100	198.500	
2023-844	62	SIEM DIM.		51	114.924	2.723	2.37	86.200	143.700	
2023-844	79	BECK. COUL		20	95.937	2.512	2.62	72.000	119.900	
2023-844	93	ARCHITECT		19	95.256	1.879	1.97	71.400	119.100	
2023-844	113	ROCHE COBA		34	89.791	2.271	2.53	67.300	112.200	
2023-844	146	ATELLICA		10	114.033	5.098	4.47	85.500	142.500	
2023-845	58	VITROS		40	153.846	12.263	7.97	115.400	192.300	

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS			
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023-845	62	SIEM DIM.		51	113.190	2.493	2.20	84.900	141.500
2023-845	79	BECK. COUL		20	99.100	4.810	4.85	74.300	123.900
2023-845	93	ARCHITECT		19	95.006	1.653	1.74	71.300	118.800
2023-845	113	ROCHE COBA		34	87.682	1.722	1.96	65.800	109.600
2023-845	146	ATELLICA		10	111.822	5.018	4.49	83.900	139.800
067 URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2023-841	11	AMES NM10S		41	1.022	0.003	0.31	1.013	1.032
2023-841	13	BAYER CLI		25	1.019	0.005	0.53	1.003	1.036
2023-841	67	URISCAN		58	1.021	0.003	0.32	1.011	1.031
2023-841	72	URS-9-TECO		67	1.022	0.004	0.35	1.011	1.033
2023-841	80	ROCHE (CHE		34	1.016	0.002	0.22	1.010	1.023
2023-841	89	AIMSTRIP		138	1.021	0.004	0.37	1.010	1.032
2023-841	94	CLINITEK		106	1.022	0.004	0.42	1.009	1.034
2023-841	102	ACON LAB		139	1.020	0.003	0.25	1.013	1.028
2023-841	131	iCHEM VELO		32	1.026	0.001	0.09	1.024	1.029
2023-841	140	MISSION A		46	1.021	0.002	0.21	1.014	1.027
2023-841	145	CLINITEK		34	1.033	0.001	0.09	1.030	1.036
2023-841	147	Mission120		40	1.021	0.002	0.23	1.014	1.028
2023-841	148	Mission500		101	1.020	0.002	0.20	1.014	1.027
2023-841	151	Clari. Uro		18	1.021	0.004	0.35	1.010	1.032
2023-842	11	AMES NM10S		41	1.011	0.004	0.37	1.000	1.023
2023-842	13	BAYER CLI		25	1.012	0.003	0.31	1.003	1.022
2023-842	67	URISCAN		58	1.008	0.004	0.41	0.996	1.021
2023-842	72	URS-9-TECO		67	1.014	0.005	0.45	1.000	1.028
2023-842	80	ROCHE (CHE		34	1.012	0.003	0.27	1.004	1.021
2023-842	89	AIMSTRIP		138	1.011	0.004	0.40	0.999	1.023
2023-842	94	CLINITEK		106	1.011	0.002	0.20	1.005	1.017
2023-842	102	ACON LAB		139	1.010	0.003	0.30	1.001	1.019
2023-842	131	iCHEM VELO		32	1.026	0.001	0.11	1.022	1.029

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-842	140	MISSION A	46	1.011	0.003	0.33	1.001	1.021
2023-842	145	CLINITEK	34	1.031	0.001	0.10	1.028	1.034
2023-842	147	Mission120	40	1.013	0.003	0.28	1.005	1.022
2023-842	148	Mission500	101	1.012	0.003	0.31	1.002	1.021
2023-842	151	Clari. Uro	18	1.009	0.002	0.25	1.002	1.017
067 URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0 EVALUATION CRITERIA: TARGET VALUE 3 SD					
2023-843	11	AMES NM10S	41	1.022	0.003	0.32	1.012	1.032
2023-843	13	BAYER CLI	25	1.021	0.002	0.23	1.014	1.028
2023-843	67	URISCAN	58	1.018	0.003	0.32	1.008	1.028
2023-843	72	URS-9-TECO	67	1.022	0.003	0.32	1.012	1.032
2023-843	80	ROCHE (CHE	34	1.013	0.002	0.25	1.005	1.020
2023-843	89	AIMSTRIP	138	1.019	0.002	0.23	1.012	1.026
2023-843	94	CLINITEK	106	1.021	0.003	0.25	1.013	1.029
2023-843	102	ACON LAB	139	1.019	0.003	0.26	1.011	1.027
2023-843	131	iCHEM VELO	32	1.025	0.001	0.11	1.021	1.028
2023-843	140	MISSION A	46	1.018	0.003	0.26	1.010	1.026
2023-843	145	CLINITEK	34	1.030	0.001	0.09	1.028	1.033
2023-843	147	Mission120	40	1.021	0.002	0.20	1.015	1.027
2023-843	148	Mission500	101	1.019	0.002	0.21	1.012	1.025
2023-843	151	Clari. Uro	18	1.017	0.002	0.24	1.010	1.024
2023-844	11	AMES NM10S	41	1.016	0.004	0.38	1.004	1.028
2023-844	13	BAYER CLI	25	1.016	0.003	0.30	1.007	1.025
2023-844	67	URISCAN	58	1.012	0.003	0.30	1.003	1.022
2023-844	72	URS-9-TECO	67	1.018	0.004	0.39	1.006	1.030
2023-844	80	ROCHE (CHE	34	1.012	0.002	0.25	1.005	1.020
2023-844	89	AIMSTRIP	138	1.016	0.004	0.38	1.004	1.027
2023-844	94	CLINITEK	106	1.016	0.003	0.28	1.007	1.024
2023-844	102	ACON LAB	139	1.016	0.004	0.38	1.004	1.027
2023-844	131	iCHEM VELO	32	1.028	0.001	0.11	1.025	1.031

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS			
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023-844	140	MISSION A		46	1.015	0.004	0.39	1.003	1.027
2023-844	145	CLINITEK		34	1.031	0.001	0.10	1.028	1.034
2023-844	147	Mission120		40	1.019	0.002	0.21	1.012	1.025
2023-844	148	Mission500		101	1.018	0.002	0.23	1.012	1.025
2023-844	151	Clari. Uro		18	1.012	0.003	0.30	1.003	1.021
067 URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD					
2023-845	11	AMES NM10S		41	1.022	0.003	0.32	1.012	1.031
2023-845	13	BAYER CLI		25	1.021	0.003	0.26	1.013	1.030
2023-845	67	URISCAN		58	1.019	0.003	0.31	1.009	1.028
2023-845	72	URS-9-TECO		67	1.022	0.003	0.34	1.012	1.033
2023-845	80	ROCHE (CHE		34	1.014	0.002	0.22	1.007	1.020
2023-845	89	AIMSTRIP		138	1.020	0.002	0.22	1.013	1.027
2023-845	94	CLINITEK		106	1.022	0.003	0.29	1.013	1.031
2023-845	102	ACON LAB		139	1.019	0.002	0.23	1.012	1.026
2023-845	131	iCHEM VELO		32	1.026	0.001	0.10	1.023	1.029
2023-845	140	MISSION A		46	1.019	0.003	0.29	1.010	1.028
2023-845	145	CLINITEK		34	1.027	0.001	0.10	1.025	1.030
2023-845	147	Mission120		40	1.021	0.002	0.23	1.014	1.028
2023-845	148	Mission500		101	1.020	0.000	0.00	1.020	1.020
2023-845	151	Clari. Uro		18	1.018	0.002	0.23	1.011	1.025
166 MICROALBUMIN		EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}							
2023-841	58	VITROS		50	6.032	0.160	2.65	1.000	11.000
2023-841	62	SIEM DIM.		81	2.913	1.132	38.88	0	7.900
2023-841	79	BECK. COUL		19	2.363	2.963	125.39	0	7.400
2023-841	85	Rx DAYTONA		17	1.682	2.521	149.88	0	6.700
2023-841	93	ARCHITECT		23	5.000	0.000	0.00	0.000	10.000
2023-841	95	INTEGRA		19	3.526	1.600	45.38	0	8.500
2023-841	113	ROCHE COBA		40	11.579	1.801	15.55	6.600	16.600
2023-841	146	ATELLICA		12	3.000	0.000	0.00	0	8.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
166	MICROALBUMIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	% OR 5.00	{GREATER}				
2023-842	58	VITROS		50	6.000	0.000	0.00	1.000	11.000	
2023-842	62	SIEM DIM.		81	2.468	1.158	46.90	0	7.500	
2023-842	79	BECK. COUL		19	2.205	2.707	122.74	0	7.200	
2023-842	85	Rx DAYTONA		17	1.412	2.054	145.47	0	6.400	
2023-842	93	ARCHITECT		23	5.000	0.000	0.00	0.000	10.000	
2023-842	95	INTEGRA		19	3.484	1.596	45.81	0	8.500	
2023-842	113	ROCHE COBA		40	11.308	2.493	22.04	6.300	16.300	
2023-842	146	ATELLICA		12	3.000	0.000	0.00	0	8.000	
2023-843	58	VITROS		50	870.442	46.439	5.34	652.800	1088.100	
2023-843	62	SIEM DIM.		81	838.609	45.626	5.44	629.000	1048.300	
2023-843	79	BECK. COUL		19	806.374	84.986	10.54	604.800	1008.000	
2023-843	85	Rx DAYTONA		17	814.250	80.015	9.83	610.700	1017.800	
2023-843	93	ARCHITECT		23	878.939	31.575	3.59	659.200	1098.700	
2023-843	95	INTEGRA		19	858.032	47.225	5.50	643.500	1072.500	
2023-843	113	ROCHE COBA		40	831.582	37.908	4.56	623.700	1039.500	
2023-843	146	ATELLICA		12	751.027	83.980	11.18	563.300	938.800	
2023-844	58	VITROS		50	801.649	44.583	5.56	601.200	1002.100	
2023-844	62	SIEM DIM.		81	793.667	47.126	5.94	595.300	992.100	
2023-844	79	BECK. COUL		19	775.553	78.461	10.12	581.700	969.400	
2023-844	85	Rx DAYTONA		17	766.729	70.314	9.17	575.000	958.400	
2023-844	93	ARCHITECT		23	813.570	31.715	3.90	610.200	1017.000	
2023-844	95	INTEGRA		19	810.226	55.747	6.88	607.700	1012.800	
2023-844	113	ROCHE COBA		40	780.995	36.959	4.73	585.700	976.200	
2023-844	146	ATELLICA		12	685.009	28.460	4.15	513.800	856.300	
2023-845	58	VITROS		50	840.543	45.569	5.42	630.400	1050.700	
2023-845	62	SIEM DIM.		81	818.363	52.700	6.44	613.800	1023.000	

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS			
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-845	79	BECK. COUL		19	786.458	74.074	9.42	589.800	983.100
2023-845	85	Rx DAYTONA		17	796.535	70.533	8.85	597.400	995.700
2023-845	93	ARCHITECT		23	845.059	28.855	3.41	633.800	1056.300
2023-845	95	INTEGRA		19	823.289	41.476	5.04	617.500	1029.100
2023-845	113	ROCHE COBA		40	804.905	39.063	4.85	603.700	1006.100
2023-845	146	APELLICA		12	699.209	28.725	4.11	524.400	874.000
167	CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2023-841	58	VITROS		52	3.204	0.064	1.99	0	8.200
2023-841	62	SIEM DIM.		77	9.171	5.525	60.24	4.200	14.200
2023-841	79	BECK. COUL		20	0.884	1.796	203.08	0	5.900
2023-841	85	Rx DAYTONA		10	1.160	1.543	133.00	0	6.200
2023-841	93	ARCHITECT		21	5.000	0.000	0.00	0.000	10.000
2023-841	95	INTEGRA		17	0.206	0.338	163.94	0	5.200
2023-841	113	ROCHE COBA		41	3.578	1.229	34.34	0	8.600
2023-841	146	APELLICA		12	3.000	0.000	0.00	0	8.000
2023-842	58	VITROS		52	241.629	13.928	5.76	181.200	302.000
2023-842	62	SIEM DIM.		77	239.632	7.625	3.18	179.700	299.500
2023-842	79	BECK. COUL		20	205.653	12.729	6.19	154.200	257.100
2023-842	85	Rx DAYTONA		10	218.430	19.021	8.71	163.800	273.000
2023-842	93	ARCHITECT		21	231.924	9.299	4.01	173.900	289.900
2023-842	95	INTEGRA		17	243.429	10.231	4.20	182.600	304.300
2023-842	113	ROCHE COBA		41	245.545	7.457	3.04	184.200	306.900
2023-842	146	APELLICA		12	224.433	5.703	2.54	168.300	280.500
2023-843	58	VITROS		52	3.202	0.051	1.59	0	8.200
2023-843	62	SIEM DIM.		77	9.095	5.568	61.22	4.100	14.100
2023-843	79	BECK. COUL		20	0.930	1.722	185.16	0	5.900
2023-843	85	Rx DAYTONA		10	1.300	2.251	173.17	0	6.300
2023-843	93	ARCHITECT		21	5.000	0.000	0.00	0.000	10.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-843	95	INTEGRA		17	0.425	0.293	68.85	0	5.400	
2023-843	113	ROCHE COBA		41	3.539	1.322	37.37	0	8.500	
2023-843	146	ATELLICA		12	3.000	0.000	0.00	0	8.000	
167	CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2023-844	58	VITROS		52	236.296	8.623	3.65	177.200	295.400	
2023-844	62	SIEM DIM.		77	233.075	7.417	3.18	174.800	291.300	
2023-844	79	BECK. COUL		20	195.570	16.990	8.69	146.700	244.500	
2023-844	85	Rx DAYTONA		10	213.960	19.647	9.18	160.500	267.500	
2023-844	93	ARCHITECT		21	226.895	8.717	3.84	170.200	283.600	
2023-844	95	INTEGRA		17	237.618	12.301	5.18	178.200	297.000	
2023-844	113	ROCHE COBA		41	238.650	7.949	3.33	179.000	298.300	
2023-844	146	ATELLICA		12	217.342	5.398	2.48	163.000	271.700	
2023-845	58	VITROS		52	236.636	8.422	3.56	177.500	295.800	
2023-845	62	SIEM DIM.		77	242.688	7.844	3.23	182.000	303.400	
2023-845	79	BECK. COUL		20	210.853	17.850	8.47	158.100	263.600	
2023-845	85	Rx DAYTONA		10	219.330	19.484	8.88	164.500	274.200	
2023-845	93	ARCHITECT		21	232.329	9.968	4.29	174.200	290.400	
2023-845	95	INTEGRA		17	244.859	11.317	4.62	183.600	306.100	
2023-845	113	ROCHE COBA		41	245.753	6.854	2.79	184.300	307.200	
2023-845	146	ATELLICA		12	223.733	5.568	2.49	167.800	279.700	

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
057 URINALYSIS-BILIRUBIN												
2023-841	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	42	42	42	100
2023-841	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	25	25	25	100
2023-841	067	URISCAN		057	057	057	057	057 - NEGATIVE	57	57	57	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score			
2023-841	072	URS-9-TECO		057	057	057	057	057	-	NEGATIVE	64	64	64	100
2023-841	080	ROCHE (CHE		057	057	057	057	057	-	NEGATIVE	33	33	33	100
2023-841	089	AIMSTRIP		057	057	057	057	057	-	NEGATIVE	137	137	137	100
2023-841	094	CLINITEK		057	057	057	057	057	-	NEGATIVE	106	106	106	100
2023-841	102	ACON LAB		057	057	057	057	057	-	NEGATIVE	135	135	135	100
2023-841	131	iCHEM VELO		057	057	057	057	057	-	NEGATIVE	32	32	32	100
2023-841	138	URS-10T		057	057	057	057	057	-	NEGATIVE	11	11	11	100
2023-841	140	MISSION A		057	057	057	057	057	-	NEGATIVE	45	45	45	100
2023-841	145	CLINITEK		057	057	057	057	057	-	NEGATIVE	34	34	34	100
2023-841	147	Mission120		057	057	057	057	057	-	NEGATIVE	40	40	40	100
2023-841	148	Mission500		057	057	057	057	057	-	NEGATIVE	104	104	104	100
2023-841	151	Clari. Uro		057	057	057	057	057	-	NEGATIVE	19	19	19	100
057 URINALYSIS-BILIRUBIN														
2023-842	011	AMES NM10S		057	057	057	057	057	-	NEGATIVE	42	42	42	100
2023-842	013	BAYER CLI		057	057	057	057	057	-	NEGATIVE	25	25	25	100
2023-842	067	URISCAN		057	057	057	057	057	-	NEGATIVE	57	57	57	100
2023-842	072	URS-9-TECO		057	057	057	057	057	-	NEGATIVE	64	64	64	100
2023-842	080	ROCHE (CHE		057	057	057	057	057	-	NEGATIVE	33	33	33	100
2023-842	089	AIMSTRIP		057	057	057	057	057	-	NEGATIVE	137	137	137	100
2023-842	094	CLINITEK		057	057	057	057	057	-	NEGATIVE	106	106	106	100
2023-842	102	ACON LAB		057	057	057	057	057	-	NEGATIVE	135	135	135	100
2023-842	131	iCHEM VELO		057	057	057	057	057	-	NEGATIVE	32	32	32	100
2023-842	138	URS-10T		057	057	057	057	057	-	NEGATIVE	11	11	11	100
2023-842	140	MISSION A		057	057	057	057	057	-	NEGATIVE	45	45	45	100
2023-842	145	CLINITEK		057	057	057	057	057	-	NEGATIVE	34	34	34	100
2023-842	147	Mission120		057	057	057	057	057	-	NEGATIVE	40	40	40	100
2023-842	148	Mission500		057	057	057	057	057	-	NEGATIVE	104	104	104	100
2023-842	151	Clari. Uro		057	057	057	057	057	-	NEGATIVE	19	19	19	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
057 URINALYSIS-BILIRUBIN													
2023-843	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	42	42	42	100	
2023-843	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	25	25	25	100	
2023-843	067	URISCAN		057	057	057	057	057 - NEGATIVE	57	57	57	100	
2023-843	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	64	64	64	100	
2023-843	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	33	33	33	100	
2023-843	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	137	137	137	100	
2023-843	094	CLINITEK		057	057	057	057	057 - NEGATIVE	106	106	106	100	
2023-843	102	ACON LAB		057	057	057	057	057 - NEGATIVE	135	135	135	100	
2023-843	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	32	32	32	100	
2023-843	138	URS-10T		057	057	057	057	057 - NEGATIVE	11	11	11	100	
2023-843	140	MISSION A		057	057	057	057	057 - NEGATIVE	45	45	45	100	
2023-843	145	CLINITEK		057	057	057	057	057 - NEGATIVE	34	34	34	100	
2023-843	147	Mission120		057	057	057	057	057 - NEGATIVE	40	40	40	100	
2023-843	148	Mission500		057	057	057	057	057 - NEGATIVE	104	103	103	100	
2023-843	148	Mission500		057	057	057	057	060 - 3+	104	1	103	0	
2023-843	151	Clari. Uro		057	057	057	057	057 - NEGATIVE	19	19	19	100	
2023-844	011	AMES NM10S		058	060	057	059	058 - TRACES, 1+	42	3	42	100	
2023-844	011	AMES NM10S		058	060	057	059	059 - 2+	42	33	42	100	
2023-844	011	AMES NM10S		058	060	057	059	060 - 3+	42	6	42	100	
2023-844	013	BAYER CLI		058	060	057	059	058 - TRACES, 1+	25	1	25	100	
2023-844	013	BAYER CLI		058	060	057	059	059 - 2+	25	24	25	100	
2023-844	067	URISCAN		057	059	057	059	057 - NEGATIVE	57	11	57	100	
2023-844	067	URISCAN		057	059	057	059	058 - TRACES, 1+	57	37	57	100	
2023-844	067	URISCAN		057	059	057	059	059 - 2+	57	9	57	100	
2023-844	072	URS-9-TECO		058	060	057	059	058 - TRACES, 1+	64	14	64	100	
2023-844	072	URS-9-TECO		058	060	057	059	059 - 2+	64	33	64	100	
2023-844	072	URS-9-TECO		058	060	057	059	060 - 3+	64	17	64	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-844	080	ROCHE (CHE		057	059	057	059	058 -	TRACES,1+	33	20	31	100
2023-844	080	ROCHE (CHE		057	059	057	059	059 -	2+	33	11	31	100
2023-844	080	ROCHE (CHE		057	059	057	059	060 -	3+	33	2	31	0
2023-844	089	AIMSTRIP		057	059	057	059	057 -	NEGATIVE	137	1	128	100
2023-844	089	AIMSTRIP		057	059	057	059	058 -	TRACES,1+	137	74	128	100
2023-844	089	AIMSTRIP		057	059	057	059	059 -	2+	137	53	128	100
2023-844	089	AIMSTRIP		057	059	057	059	060 -	3+	137	9	128	0
2023-844	094	CLINITEK		058	060	057	059	057 -	NEGATIVE	106	1	106	100
2023-844	094	CLINITEK		058	060	057	059	058 -	TRACES,1+	106	12	106	100
2023-844	094	CLINITEK		058	060	057	059	059 -	2+	106	89	106	100
2023-844	094	CLINITEK		058	060	057	059	060 -	3+	106	4	106	100
2023-844	102	ACON LAB		057	059	057	059	058 -	TRACES,1+	135	82	131	100
2023-844	102	ACON LAB		057	059	057	059	059 -	2+	135	49	131	100
2023-844	102	ACON LAB		057	059	057	059	060 -	3+	135	4	131	0
2023-844	131	iCHEM VELO		057	057	057	059	057 -	NEGATIVE	32	32	32	100
2023-844	138	URS-10T		058	060	057	059	058 -	TRACES,1+	11	1	11	100
2023-844	138	URS-10T		058	060	057	059	059 -	2+	11	9	11	100
2023-844	138	URS-10T		058	060	057	059	060 -	3+	11	1	11	100
2023-844	140	MISSION A		057	059	057	059	058 -	TRACES,1+	45	27	44	100
2023-844	140	MISSION A		057	059	057	059	059 -	2+	45	17	44	100
2023-844	140	MISSION A		057	059	057	059	060 -	3+	45	1	44	0
2023-844	145	CLINITEK		057	057	057	059	057 -	NEGATIVE	34	33	34	100
2023-844	145	CLINITEK		057	057	057	059	058 -	TRACES,1+	34	1	34	100
2023-844	147	Mission120		057	059	057	059	058 -	TRACES,1+	40	26	39	100
2023-844	147	Mission120		057	059	057	059	059 -	2+	40	13	39	100
2023-844	147	Mission120		057	059	057	059	060 -	3+	40	1	39	0
2023-844	148	Mission500		057	059	057	059	057 -	NEGATIVE	104	1	104	100
2023-844	148	Mission500		057	059	057	059	058 -	TRACES,1+	104	88	104	100
2023-844	148	Mission500		057	059	057	059	059 -	2+	104	15	104	100
2023-844	151	Clari. Uro		058	060	057	059	058 -	TRACES,1+	19	7	19	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-844	151	Clari. Uro		058 060	057 059		059 - 2+	19	10	19	100	
2023-844	151	Clari. Uro		058 060	057 059		060 - 3+	19	2	19	100	
057 URINALYSIS-BILIRUBIN												
2023-845	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	42	42	42	100	
2023-845	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	25	25	25	100	
2023-845	067	URISCAN		057 057	057 057		057 - NEGATIVE	57	57	57	100	
2023-845	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	64	64	64	100	
2023-845	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	33	33	33	100	
2023-845	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	137	137	137	100	
2023-845	094	CLINITEK		057 057	057 057		057 - NEGATIVE	106	106	106	100	
2023-845	102	ACON LAB		057 057	057 057		057 - NEGATIVE	135	135	135	100	
2023-845	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	32	32	32	100	
2023-845	138	URS-10T		057 057	057 057		057 - NEGATIVE	11	11	11	100	
2023-845	140	MISSION A		057 057	057 057		057 - NEGATIVE	45	45	45	100	
2023-845	145	CLINITEK		057 057	057 057		057 - NEGATIVE	34	34	34	100	
2023-845	147	Mission120		057 057	057 057		057 - NEGATIVE	40	40	40	100	
2023-845	148	Mission500		057 057	057 057		057 - NEGATIVE	104	104	104	100	
2023-845	151	Clari. Uro		057 057	057 057		057 - NEGATIVE	19	19	19	100	
058 URINALYSIS-GLUCOSE												
2023-841	011	AMES NM10S		074 076	074 076		073 - NEGATIVE	42	1	41	0	
2023-841	011	AMES NM10S		074 076	074 076		074 - TRACES	42	14	41	100	
2023-841	011	AMES NM10S		074 076	074 076		075 - 250mg/dl-+	42	26	41	100	
2023-841	011	AMES NM10S		074 076	074 076		076 - 500mg/dl2+	42	1	41	100	
2023-841	013	BAYER CLI		074 076	074 076		073 - NEGATIVE	25	1	24	0	
2023-841	013	BAYER CLI		074 076	074 076		074 - TRACES	25	4	24	100	
2023-841	013	BAYER CLI		074 076	074 076		075 - 250mg/dl-+	25	20	24	100	
2023-841	067	URISCAN		075 077	074 076		075 - 250mg/dl-+	57	15	56	100	
2023-841	067	URISCAN		075 077	074 076		076 - 500mg/dl2+	57	29	56	100	
2023-841	067	URISCAN		075 077	074 076		077 - 1000mg/dl	57	12	56	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-841	067	URISCAN		075 077	074 076		078 - 2000mg/dl	57	1	56	0
2023-841	072	URS-9-TECO		074 076	074 076		073 - NEGATIVE	66	4	60	0
2023-841	072	URS-9-TECO		074 076	074 076		074 - TRACES	66	23	60	100
2023-841	072	URS-9-TECO		074 076	074 076		075 - 250mg/dl-+	66	34	60	100
2023-841	072	URS-9-TECO		074 076	074 076		076 - 500mg/dl2+	66	3	60	100
2023-841	072	URS-9-TECO		074 076	074 076		077 - 1000mg/dl	66	2	60	0
2023-841	080	ROCHE (CHE		076 078	074 076		077 - 1000mg/dl	34	32	34	100
2023-841	080	ROCHE (CHE		076 078	074 076		078 - 2000mg/dl	34	2	34	100
2023-841	089	AIMSTRIP		074 076	074 076		073 - NEGATIVE	137	1	136	0
2023-841	089	AIMSTRIP		074 076	074 076		074 - TRACES	137	19	136	100
2023-841	089	AIMSTRIP		074 076	074 076		075 - 250mg/dl-+	137	116	136	100
2023-841	089	AIMSTRIP		074 076	074 076		076 - 500mg/dl2+	137	1	136	100
2023-841	094	CLINITEK		074 076	074 076		074 - TRACES	106	37	106	100
2023-841	094	CLINITEK		074 076	074 076		075 - 250mg/dl-+	106	69	106	100
2023-841	102	ACON LAB		074 076	074 076		074 - TRACES	138	13	138	100
2023-841	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	138	124	138	100
2023-841	102	ACON LAB		074 076	074 076		076 - 500mg/dl2+	138	1	138	100
2023-841	131	iCHEM VELO		075 077	074 076		076 - 500mg/dl2+	32	28	32	100
2023-841	131	iCHEM VELO		075 077	074 076		077 - 1000mg/dl	32	4	32	100
2023-841	138	URS-10T		073 075	074 076		074 - TRACES	10	7	10	100
2023-841	138	URS-10T		073 075	074 076		075 - 250mg/dl-+	10	3	10	100
2023-841	140	MISSION A		074 076	074 076		073 - NEGATIVE	46	1	45	0
2023-841	140	MISSION A		074 076	074 076		074 - TRACES	46	3	45	100
2023-841	140	MISSION A		074 076	074 076		075 - 250mg/dl-+	46	40	45	100
2023-841	140	MISSION A		074 076	074 076		076 - 500mg/dl2+	46	2	45	100
2023-841	145	CLINITEK		076 078	074 076		075 - 250mg/dl-+	34	2	34	100
2023-841	145	CLINITEK		076 078	074 076		076 - 500mg/dl2+	34	5	34	100
2023-841	145	CLINITEK		076 078	074 076		077 - 1000mg/dl	34	25	34	100
2023-841	145	CLINITEK		076 078	074 076		078 - 2000mg/dl	34	2	34	100
2023-841	147	Mission120		074 076	074 076		074 - TRACES	40	5	40	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-841	147	Mission120		074 076	074 076		075 - 250mg/dl-+	40	35	40	100	
2023-841	148	Mission500		074 076	074 076		074 - TRACES	103	1	103	100	
2023-841	148	Mission500		074 076	074 076		075 - 250mg/dl-+	103	102	103	100	
2023-841	151	Clari. Uro		074 076	074 076		075 - 250mg/dl-+	18	18	18	100	
058 URINALYSIS-GLUCOSE												
2023-842	011	AMES NM10S		076 078	075 077		074 - TRACES	42	1	41	0	
2023-842	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	42	13	41	100	
2023-842	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	42	26	41	100	
2023-842	011	AMES NM10S		076 078	075 077		078 - 2000mg/dl	42	2	41	100	
2023-842	013	BAYER CLI		075 077	075 077		076 - 500mg/dl2+	25	13	25	100	
2023-842	013	BAYER CLI		075 077	075 077		077 - 1000mg/dl	25	12	25	100	
2023-842	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	57	4	57	100	
2023-842	067	URISCAN		077 078	075 077		077 - 1000mg/dl	57	18	57	100	
2023-842	067	URISCAN		077 078	075 077		078 - 2000mg/dl	57	35	57	100	
2023-842	072	URS-9-TECO		076 078	075 077		073 - NEGATIVE	66	1	65	0	
2023-842	072	URS-9-TECO		076 078	075 077		075 - 250mg/dl-+	66	2	65	100	
2023-842	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	66	21	65	100	
2023-842	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	66	37	65	100	
2023-842	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	66	5	65	100	
2023-842	080	ROCHE (CHE		076 078	075 077		075 - 250mg/dl-+	34	2	34	100	
2023-842	080	ROCHE (CHE		076 078	075 077		076 - 500mg/dl2+	34	1	34	100	
2023-842	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	34	29	34	100	
2023-842	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	34	2	34	100	
2023-842	089	AIMSTRIP		076 078	075 077		073 - NEGATIVE	137	1	136	0	
2023-842	089	AIMSTRIP		076 078	075 077		075 - 250mg/dl-+	137	1	136	100	
2023-842	089	AIMSTRIP		076 078	075 077		076 - 500mg/dl2+	137	58	136	100	
2023-842	089	AIMSTRIP		076 078	075 077		077 - 1000mg/dl	137	76	136	100	
2023-842	089	AIMSTRIP		076 078	075 077		078 - 2000mg/dl	137	1	136	100	
2023-842	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	106	1	106	100	
2023-842	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	106	44	106	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-842	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	106	61	106	100		
2023-842	102	ACON LAB		076 078	075 077		075 - 250mg/dl-+	138	1	138	100		
2023-842	102	ACON LAB		076 078	075 077		076 - 500mg/dl2+	138	52	138	100		
2023-842	102	ACON LAB		076 078	075 077		077 - 1000mg/dl	138	79	138	100		
2023-842	102	ACON LAB		076 078	075 077		078 - 2000mg/dl	138	6	138	100		
2023-842	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	32	28	32	100		
2023-842	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	32	4	32	100		
2023-842	138	URS-10T		075 077	075 077		076 - 500mg/dl2+	10	5	9	100		
2023-842	138	URS-10T		075 077	075 077		077 - 1000mg/dl	10	4	9	100		
2023-842	138	URS-10T		075 077	075 077		078 - 2000mg/dl	10	1	9	0		
2023-842	140	MISSION A		076 078	075 077		073 - NEGATIVE	46	1	45	0		
2023-842	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	46	8	45	100		
2023-842	140	MISSION A		076 078	075 077		077 - 1000mg/dl	46	31	45	100		
2023-842	140	MISSION A		076 078	075 077		078 - 2000mg/dl	46	6	45	100		
2023-842	145	CLINITEK		075 077	075 077		076 - 500mg/dl2+	34	31	33	100		
2023-842	145	CLINITEK		075 077	075 077		077 - 1000mg/dl	34	2	33	100		
2023-842	145	CLINITEK		075 077	075 077		078 - 2000mg/dl	34	1	33	0		
2023-842	147	Mission120		076 078	075 077		076 - 500mg/dl2+	40	1	40	100		
2023-842	147	Mission120		076 078	075 077		077 - 1000mg/dl	40	27	40	100		
2023-842	147	Mission120		076 078	075 077		078 - 2000mg/dl	40	12	40	100		
2023-842	148	Mission500		076 078	075 077		073 - NEGATIVE	103	1	102	0		
2023-842	148	Mission500		076 078	075 077		076 - 500mg/dl2+	103	29	102	100		
2023-842	148	Mission500		076 078	075 077		077 - 1000mg/dl	103	53	102	100		
2023-842	148	Mission500		076 078	075 077		078 - 2000mg/dl	103	20	102	100		
2023-842	151	Clari. Uro		076 078	075 077		076 - 500mg/dl2+	18	3	18	100		
2023-842	151	Clari. Uro		076 078	075 077		077 - 1000mg/dl	18	15	18	100		
058 URINALYSIS-GLUCOSE													
2023-843	011	AMES NM10S		076 078	076 078		076 - 500mg/dl2+	42	12	42	100		
2023-843	011	AMES NM10S		076 078	076 078		077 - 1000mg/dl	42	27	42	100		
2023-843	011	AMES NM10S		076 078	076 078		078 - 2000mg/dl	42	3	42	100		

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-843	013	BAYER CLI		076 078	076 078		075 - 250mg/dl-+	25	3	22	0	
2023-843	013	BAYER CLI		076 078	076 078		076 - 500mg/dl2+	25	9	22	100	
2023-843	013	BAYER CLI		076 078	076 078		077 - 1000mg/dl	25	13	22	100	
2023-843	067	URISCAN		077 078	076 078		076 - 500mg/dl2+	57	2	57	100	
2023-843	067	URISCAN		077 078	076 078		077 - 1000mg/dl	57	13	57	100	
2023-843	067	URISCAN		077 078	076 078		078 - 2000mg/dl	57	42	57	100	
2023-843	072	URS-9-TECO		076 078	076 078		075 - 250mg/dl-+	66	1	65	0	
2023-843	072	URS-9-TECO		076 078	076 078		076 - 500mg/dl2+	66	16	65	100	
2023-843	072	URS-9-TECO		076 078	076 078		077 - 1000mg/dl	66	37	65	100	
2023-843	072	URS-9-TECO		076 078	076 078		078 - 2000mg/dl	66	12	65	100	
2023-843	080	ROCHE (CHE		076 078	076 078		077 - 1000mg/dl	34	32	34	100	
2023-843	080	ROCHE (CHE		076 078	076 078		078 - 2000mg/dl	34	2	34	100	
2023-843	089	AIMSTRIP		076 078	076 078		075 - 250mg/dl-+	137	1	136	0	
2023-843	089	AIMSTRIP		076 078	076 078		076 - 500mg/dl2+	137	64	136	100	
2023-843	089	AIMSTRIP		076 078	076 078		077 - 1000mg/dl	137	72	136	100	
2023-843	094	CLINITEK		076 078	076 078		075 - 250mg/dl-+	106	3	103	0	
2023-843	094	CLINITEK		076 078	076 078		076 - 500mg/dl2+	106	30	103	100	
2023-843	094	CLINITEK		076 078	076 078		077 - 1000mg/dl	106	73	103	100	
2023-843	102	ACON LAB		076 078	076 078		075 - 250mg/dl-+	138	1	137	0	
2023-843	102	ACON LAB		076 078	076 078		076 - 500mg/dl2+	138	62	137	100	
2023-843	102	ACON LAB		076 078	076 078		077 - 1000mg/dl	138	73	137	100	
2023-843	102	ACON LAB		076 078	076 078		078 - 2000mg/dl	138	2	137	100	
2023-843	131	iCHEM VELO		075 077	076 078		076 - 500mg/dl2+	32	28	32	100	
2023-843	131	iCHEM VELO		075 077	076 078		077 - 1000mg/dl	32	4	32	100	
2023-843	138	URS-10T		076 078	076 078		076 - 500mg/dl2+	10	2	10	100	
2023-843	138	URS-10T		076 078	076 078		077 - 1000mg/dl	10	5	10	100	
2023-843	138	URS-10T		076 078	076 078		078 - 2000mg/dl	10	3	10	100	
2023-843	140	MISSION A		076 078	076 078		076 - 500mg/dl2+	46	11	46	100	
2023-843	140	MISSION A		076 078	076 078		077 - 1000mg/dl	46	27	46	100	
2023-843	140	MISSION A		076 078	076 078		078 - 2000mg/dl	46	8	46	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-843	145	CLINITEK		076 078	076 078		076 - 500mg/dl2+	34	4	34	100		
2023-843	145	CLINITEK		076 078	076 078		077 - 1000mg/dl	34	27	34	100		
2023-843	145	CLINITEK		076 078	076 078		078 - 2000mg/dl	34	3	34	100		
2023-843	147	Mission120		076 078	076 078		076 - 500mg/dl2+	40	3	40	100		
2023-843	147	Mission120		076 078	076 078		077 - 1000mg/dl	40	26	40	100		
2023-843	147	Mission120		076 078	076 078		078 - 2000mg/dl	40	11	40	100		
2023-843	148	Mission500		076 078	076 078		076 - 500mg/dl2+	103	37	103	100		
2023-843	148	Mission500		076 078	076 078		077 - 1000mg/dl	103	57	103	100		
2023-843	148	Mission500		076 078	076 078		078 - 2000mg/dl	103	9	103	100		
2023-843	151	Clari. Uro		076 078	076 078		076 - 500mg/dl2+	18	1	18	100		
2023-843	151	Clari. Uro		076 078	076 078		077 - 1000mg/dl	18	17	18	100		
058 URINALYSIS-GLUCOSE													
2023-844	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	42	42	42	100		
2023-844	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	25	25	25	100		
2023-844	067	URISCAN		073 073	073 073		073 - NEGATIVE	57	57	57	100		
2023-844	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	66	65	65	100		
2023-844	072	URS-9-TECO		073 073	073 073		078 - 2000mg/dl	66	1	65	0		
2023-844	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	34	34	34	100		
2023-844	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	137	136	136	100		
2023-844	089	AIMSTRIP		073 073	073 073		074 - TRACES	137	1	136	0		
2023-844	094	CLINITEK		073 073	073 073		073 - NEGATIVE	106	106	106	100		
2023-844	102	ACON LAB		073 073	073 073		073 - NEGATIVE	138	138	138	100		
2023-844	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	32	32	32	100		
2023-844	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2023-844	140	MISSION A		073 073	073 073		073 - NEGATIVE	46	46	46	100		
2023-844	145	CLINITEK		073 073	073 073		073 - NEGATIVE	34	34	34	100		
2023-844	147	Mission120		073 073	073 073		073 - NEGATIVE	40	40	40	100		
2023-844	148	Mission500		073 073	073 073		073 - NEGATIVE	103	103	103	100		
2023-844	151	Clari. Uro		073 073	073 073		073 - NEGATIVE	18	18	18	100		

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
058 URINALYSIS-GLUCOSE													
2023-845	011	AMES NM10S		073	073	073	073	073 - NEGATIVE	42	42	42	100	
2023-845	013	BAYER CLI		073	073	073	073	073 - NEGATIVE	25	25	25	100	
2023-845	067	URISCAN		073	073	073	073	073 - NEGATIVE	57	57	57	100	
2023-845	072	URS-9-TECO		073	073	073	073	073 - NEGATIVE	66	66	66	100	
2023-845	080	ROCHE (CHE		073	073	073	073	073 - NEGATIVE	34	34	34	100	
2023-845	089	AIMSTRIP		073	073	073	073	073 - NEGATIVE	137	137	137	100	
2023-845	094	CLINITEK		073	073	073	073	073 - NEGATIVE	106	104	104	100	
2023-845	094	CLINITEK		073	073	073	073	074 - TRACES	106	2	104	0	
2023-845	102	ACON LAB		073	073	073	073	073 - NEGATIVE	138	138	138	100	
2023-845	131	iCHEM VELO		073	073	073	073	073 - NEGATIVE	32	32	32	100	
2023-845	138	URS-10T		073	073	073	073	073 - NEGATIVE	10	10	10	100	
2023-845	140	MISSION A		073	073	073	073	073 - NEGATIVE	46	46	46	100	
2023-845	145	CLINITEK		073	073	073	073	073 - NEGATIVE	34	34	34	100	
2023-845	147	Mission120		073	073	073	073	073 - NEGATIVE	40	40	40	100	
2023-845	148	Mission500		073	073	073	073	073 - NEGATIVE	103	103	103	100	
2023-845	151	Clari. Uro		073	073	073	073	073 - NEGATIVE	18	18	18	100	
059 URINALYS.HG.OR BLOOD													
2023-841	011	AMES NM10S		066	066	066	066	066 - NEGATIVE	42	41	41	100	
2023-841	011	AMES NM10S		066	066	066	066	068 - POS.(1-2+)	42	1	41	0	
2023-841	013	BAYER CLI		066	066	066	066	066 - NEGATIVE	25	25	25	100	
2023-841	067	URISCAN		066	066	066	066	066 - NEGATIVE	57	56	56	100	
2023-841	067	URISCAN		066	066	066	066	067 - TRACES	57	1	56	0	
2023-841	072	URS-9-TECO		066	066	066	066	066 - NEGATIVE	66	66	66	100	
2023-841	080	ROCHE (CHE		066	066	066	066	066 - NEGATIVE	34	34	34	100	
2023-841	089	AIMSTRIP		066	066	066	066	066 - NEGATIVE	137	137	137	100	
2023-841	094	CLINITEK		066	066	066	066	066 - NEGATIVE	106	106	106	100	
2023-841	102	ACON LAB		066	066	066	066	066 - NEGATIVE	139	139	139	100	
2023-841	131	iCHEM VELO		066	066	066	066	066 - NEGATIVE	32	31	31	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-841	131	icHEM VELO		066 066	066 066		068 - POS.(1-2+)	32	1	31	0	
2023-841	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2023-841	140	MISSION A		066 066	066 066		066 - NEGATIVE	44	44	44	100	
2023-841	145	CLINITEK		066 066	066 066		066 - NEGATIVE	34	34	34	100	
2023-841	147	Mission120		066 066	066 066		066 - NEGATIVE	41	41	41	100	
2023-841	148	Mission500		066 066	066 066		066 - NEGATIVE	103	103	103	100	
2023-841	151	Clari. Uro		066 066	066 066		066 - NEGATIVE	18	18	18	100	
059 URINALYS.HG.OR BLOOD												
2023-842	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	42	42	42	100	
2023-842	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	25	25	25	100	
2023-842	067	URISCAN		066 066	066 066		066 - NEGATIVE	57	55	55	100	
2023-842	067	URISCAN		066 066	066 066		067 - TRACES	57	1	55	0	
2023-842	067	URISCAN		066 066	066 066		068 - POS.(1-2+)	57	1	55	0	
2023-842	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	66	66	66	100	
2023-842	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	34	34	34	100	
2023-842	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	137	137	137	100	
2023-842	094	CLINITEK		066 066	066 066		066 - NEGATIVE	106	106	106	100	
2023-842	102	ACON LAB		066 066	066 066		066 - NEGATIVE	139	139	139	100	
2023-842	131	icHEM VELO		066 066	066 066		066 - NEGATIVE	32	32	32	100	
2023-842	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2023-842	140	MISSION A		066 066	066 066		066 - NEGATIVE	44	44	44	100	
2023-842	145	CLINITEK		066 066	066 066		066 - NEGATIVE	34	34	34	100	
2023-842	147	Mission120		066 066	066 066		066 - NEGATIVE	41	41	41	100	
2023-842	148	Mission500		066 066	066 066		066 - NEGATIVE	103	103	103	100	
2023-842	151	Clari. Uro		066 066	066 066		066 - NEGATIVE	18	18	18	100	
2023-843	011	AMES NM10S		068 069	068 069		067 - TRACES	42	1	41	0	
2023-843	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	42	5	41	100	
2023-843	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	42	36	41	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-843	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	25	25	25	100
2023-843	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	57	6	57	100
2023-843	067	URISCAN		068 069	068 069		069 - POS.(3+))	57	51	57	100
2023-843	072	URS-9-TECO		068 069	068 069		067 - TRACES	66	2	64	0
2023-843	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	66	3	64	100
2023-843	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	66	61	64	100
2023-843	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	34	1	34	100
2023-843	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	34	33	34	100
2023-843	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	137	5	137	100
2023-843	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	137	132	137	100
2023-843	094	CLINITEK		068 069	068 069		066 - NEGATIVE	106	1	105	0
2023-843	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	106	5	105	100
2023-843	094	CLINITEK		068 069	068 069		069 - POS.(3+))	106	100	105	100
2023-843	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	139	6	139	100
2023-843	102	ACON LAB		068 069	068 069		069 - POS.(3+))	139	133	139	100
2023-843	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	32	2	32	100
2023-843	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	32	30	32	100
2023-843	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	10	10	100
2023-843	140	MISSION A		068 069	068 069		066 - NEGATIVE	44	1	42	0
2023-843	140	MISSION A		068 069	068 069		067 - TRACES	44	1	42	0
2023-843	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	44	2	42	100
2023-843	140	MISSION A		068 069	068 069		069 - POS.(3+))	44	40	42	100
2023-843	145	CLINITEK		068 069	068 069		069 - POS.(3+))	34	34	34	100
2023-843	147	Mission120		068 069	068 069		069 - POS.(3+))	41	41	41	100
2023-843	148	Mission500		068 069	068 069		068 - POS.(1-2+)	103	2	103	100
2023-843	148	Mission500		068 069	068 069		069 - POS.(3+))	103	101	103	100
2023-843	151	Clari. Uro		068 069	068 069		069 - POS.(3+))	18	18	18	100
059 URINALYS.HG.OR BLOOD											
2023-844	011	AMES NM10S		068 069	067 069		068 - POS.(1-2+)	42	8	42	100
2023-844	011	AMES NM10S		068 069	067 069		069 - POS.(3+))	42	34	42	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-844	013	BAYER CLI		068 069	067 069		068 - POS.(1-2+)	25	4	25	100
2023-844	013	BAYER CLI		068 069	067 069		069 - POS.(3+)	25	21	25	100
2023-844	067	URISCAN		068 069	067 069		068 - POS.(1-2+)	57	12	57	100
2023-844	067	URISCAN		068 069	067 069		069 - POS.(3+)	57	45	57	100
2023-844	072	URS-9-TECO		068 069	067 069		066 - NEGATIVE	66	1	65	0
2023-844	072	URS-9-TECO		068 069	067 069		068 - POS.(1-2+)	66	17	65	100
2023-844	072	URS-9-TECO		068 069	067 069		069 - POS.(3+)	66	48	65	100
2023-844	080	ROCHE (CHE		068 069	067 069		068 - POS.(1-2+)	34	1	34	100
2023-844	080	ROCHE (CHE		068 069	067 069		069 - POS.(3+)	34	33	34	100
2023-844	089	AIMSTRIP		068 069	067 069		066 - NEGATIVE	137	1	136	0
2023-844	089	AIMSTRIP		068 069	067 069		068 - POS.(1-2+)	137	4	136	100
2023-844	089	AIMSTRIP		068 069	067 069		069 - POS.(3+)	137	132	136	100
2023-844	094	CLINITEK		068 069	067 069		068 - POS.(1-2+)	106	24	106	100
2023-844	094	CLINITEK		068 069	067 069		069 - POS.(3+)	106	82	106	100
2023-844	102	ACON LAB		068 069	067 069		068 - POS.(1-2+)	139	4	139	100
2023-844	102	ACON LAB		068 069	067 069		069 - POS.(3+)	139	135	139	100
2023-844	131	iCHEM VELO		067 069	067 069		067 - TRACES	32	1	32	100
2023-844	131	iCHEM VELO		067 069	067 069		068 - POS.(1-2+)	32	30	32	100
2023-844	131	iCHEM VELO		067 069	067 069		069 - POS.(3+)	32	1	32	100
2023-844	138	URS-10T		068 069	067 069		068 - POS.(1-2+)	10	1	10	100
2023-844	138	URS-10T		068 069	067 069		069 - POS.(3+)	10	9	10	100
2023-844	140	MISSION A		068 069	067 069		068 - POS.(1-2+)	44	2	44	100
2023-844	140	MISSION A		068 069	067 069		069 - POS.(3+)	44	42	44	100
2023-844	145	CLINITEK		067 069	067 069		068 - POS.(1-2+)	34	18	34	100
2023-844	145	CLINITEK		067 069	067 069		069 - POS.(3+)	34	16	34	100
2023-844	147	Mission120		068 069	067 069		069 - POS.(3+)	41	41	41	100
2023-844	148	Mission500		068 069	067 069		068 - POS.(1-2+)	103	1	103	100
2023-844	148	Mission500		068 069	067 069		069 - POS.(3+)	103	102	103	100
2023-844	151	Clari. Uro		068 069	067 069		069 - POS.(3+)	18	18	18	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
059 URINALYS.HG.OR BLOOD													
2023-845	011 AMES NM10S			068 069	068 069		068 - POS.(1-2+)	42	2	42	100		
2023-845	011 AMES NM10S			068 069	068 069		069 - POS.(3+))	42	40	42	100		
2023-845	013 BAYER CLI			068 069	068 069		069 - POS.(3+))	25	25	25	100		
2023-845	067 URISCAN			068 069	068 069		068 - POS.(1-2+)	57	6	57	100		
2023-845	067 URISCAN			068 069	068 069		069 - POS.(3+))	57	51	57	100		
2023-845	072 URS-9-TECO			068 069	068 069		068 - POS.(1-2+)	66	6	66	100		
2023-845	072 URS-9-TECO			068 069	068 069		069 - POS.(3+))	66	60	66	100		
2023-845	080 ROCHE (CHE			068 069	068 069		068 - POS.(1-2+)	34	1	34	100		
2023-845	080 ROCHE (CHE			068 069	068 069		069 - POS.(3+))	34	33	34	100		
2023-845	089 AIMSTRIP			068 069	068 069		068 - POS.(1-2+)	137	2	137	100		
2023-845	089 AIMSTRIP			068 069	068 069		069 - POS.(3+))	137	135	137	100		
2023-845	094 CLINITEK			068 069	068 069		068 - POS.(1-2+)	106	3	106	100		
2023-845	094 CLINITEK			068 069	068 069		069 - POS.(3+))	106	103	106	100		
2023-845	102 ACON LAB			068 069	068 069		068 - POS.(1-2+)	139	2	139	100		
2023-845	102 ACON LAB			068 069	068 069		069 - POS.(3+))	139	137	139	100		
2023-845	131 iCHEM VELO			068 069	068 069		068 - POS.(1-2+)	32	2	32	100		
2023-845	131 iCHEM VELO			068 069	068 069		069 - POS.(3+))	32	30	32	100		
2023-845	138 URS-10T			068 069	068 069		069 - POS.(3+))	10	10	10	100		
2023-845	140 MISSION A			068 069	068 069		068 - POS.(1-2+)	44	2	44	100		
2023-845	140 MISSION A			068 069	068 069		069 - POS.(3+))	44	42	44	100		
2023-845	145 CLINITEK			068 069	068 069		068 - POS.(1-2+)	34	1	34	100		
2023-845	145 CLINITEK			068 069	068 069		069 - POS.(3+))	34	33	34	100		
2023-845	147 Mission120			068 069	068 069		069 - POS.(3+))	41	41	41	100		
2023-845	148 Mission500			068 069	068 069		068 - POS.(1-2+)	103	1	103	100		
2023-845	148 Mission500			068 069	068 069		069 - POS.(3+))	103	102	103	100		
2023-845	151 Clari. Uro			068 069	068 069		069 - POS.(3+))	18	18	18	100		
060 URINALYSIS-KETONES													
2023-841	011 AMES NM10S			092 094	092 094		092 - SMALL (1+)	42	15	42	100		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-841	011	AMES NM10S		092 094	092 094		093 - MOD. (2+)	42	22	42	100
2023-841	011	AMES NM10S		092 094	092 094		094 - LARGE-3-4+	42	5	42	100
2023-841	013	BAYER CLI		092 094	092 094		092 - SMALL (1+)	25	11	25	100
2023-841	013	BAYER CLI		092 094	092 094		093 - MOD. (2+)	25	14	25	100
2023-841	067	URISCAN		092 094	092 094		092 - SMALL (1+)	57	15	57	100
2023-841	067	URISCAN		092 094	092 094		093 - MOD. (2+)	57	40	57	100
2023-841	067	URISCAN		092 094	092 094		094 - LARGE-3-4+	57	2	57	100
2023-841	072	URS-9-TECO		092 094	092 094		092 - SMALL (1+)	66	12	66	100
2023-841	072	URS-9-TECO		092 094	092 094		093 - MOD. (2+)	66	40	66	100
2023-841	072	URS-9-TECO		092 094	092 094		094 - LARGE-3-4+	66	14	66	100
2023-841	080	ROCHE (CHE		092 094	092 094		092 - SMALL (1+)	34	10	34	100
2023-841	080	ROCHE (CHE		092 094	092 094		093 - MOD. (2+)	34	20	34	100
2023-841	080	ROCHE (CHE		092 094	092 094		094 - LARGE-3-4+	34	4	34	100
2023-841	089	AIMSTRIP		092 094	092 094		090 - NEGATIVE	137	1	135	0
2023-841	089	AIMSTRIP		092 094	092 094		091 - TRACE	137	1	135	0
2023-841	089	AIMSTRIP		092 094	092 094		092 - SMALL (1+)	137	23	135	100
2023-841	089	AIMSTRIP		092 094	092 094		093 - MOD. (2+)	137	98	135	100
2023-841	089	AIMSTRIP		092 094	092 094		094 - LARGE-3-4+	137	14	135	100
2023-841	094	CLINITEK		091 093	092 094		091 - TRACE	106	1	106	100
2023-841	094	CLINITEK		091 093	092 094		092 - SMALL (1+)	106	57	106	100
2023-841	094	CLINITEK		091 093	092 094		093 - MOD. (2+)	106	48	106	100
2023-841	102	ACON LAB		092 094	092 094		092 - SMALL (1+)	138	23	138	100
2023-841	102	ACON LAB		092 094	092 094		093 - MOD. (2+)	138	104	138	100
2023-841	102	ACON LAB		092 094	092 094		094 - LARGE-3-4+	138	11	138	100
2023-841	131	iCHEM VELO		091 093	092 094		092 - SMALL (1+)	32	21	32	100
2023-841	131	iCHEM VELO		091 093	092 094		093 - MOD. (2+)	32	9	32	100
2023-841	131	iCHEM VELO		091 093	092 094		094 - LARGE-3-4+	32	2	32	100
2023-841	138	URS-10T		092 094	092 094		092 - SMALL (1+)	10	3	10	100
2023-841	138	URS-10T		092 094	092 094		093 - MOD. (2+)	10	7	10	100
2023-841	140	MISSION A		092 094	092 094		092 - SMALL (1+)	47	2	47	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-841	140	MISSION A		092 094	092 094		093 - MOD. (2+)	47	31	47	100	
2023-841	140	MISSION A		092 094	092 094		094 - LARGE-3-4+	47	14	47	100	
2023-841	145	CLINITEK		092 094	092 094		092 - SMALL (1+)	34	7	34	100	
2023-841	145	CLINITEK		092 094	092 094		093 - MOD. (2+)	34	27	34	100	
2023-841	147	Mission120		092 094	092 094		092 - SMALL (1+)	40	1	40	100	
2023-841	147	Mission120		092 094	092 094		093 - MOD. (2+)	40	22	40	100	
2023-841	147	Mission120		092 094	092 094		094 - LARGE-3-4+	40	17	40	100	
2023-841	148	Mission500		092 094	092 094		092 - SMALL (1+)	102	2	102	100	
2023-841	148	Mission500		092 094	092 094		093 - MOD. (2+)	102	96	102	100	
2023-841	148	Mission500		092 094	092 094		094 - LARGE-3-4+	102	4	102	100	
2023-841	151	Clari. Uro		092 094	092 094		093 - MOD. (2+)	18	12	18	100	
2023-841	151	Clari. Uro		092 094	092 094		094 - LARGE-3-4+	18	6	18	100	
060 URINALYSIS-KETONES												
2023-842	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	42	42	42	100	
2023-842	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	25	25	25	100	
2023-842	067	URISCAN		090 090	090 090		090 - NEGATIVE	57	57	57	100	
2023-842	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	66	66	66	100	
2023-842	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	34	34	34	100	
2023-842	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	137	137	137	100	
2023-842	094	CLINITEK		090 090	090 090		090 - NEGATIVE	106	106	106	100	
2023-842	102	ACON LAB		090 090	090 090		090 - NEGATIVE	138	138	138	100	
2023-842	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	32	32	32	100	
2023-842	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2023-842	140	MISSION A		090 090	090 090		090 - NEGATIVE	47	47	47	100	
2023-842	145	CLINITEK		090 090	090 090		090 - NEGATIVE	34	34	34	100	
2023-842	147	Mission120		090 090	090 090		090 - NEGATIVE	40	40	40	100	
2023-842	148	Mission500		090 090	090 090		090 - NEGATIVE	102	102	102	100	
2023-842	151	Clari. Uro		090 090	090 090		090 - NEGATIVE	18	18	18	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
060 URINALYSIS-KETONES													
2023-843	011 AMES NM10S			090 090	090 090		090 - NEGATIVE	42	42	42	100		
2023-843	013 BAYER CLI			090 090	090 090		090 - NEGATIVE	25	25	25	100		
2023-843	067 URISCAN			090 090	090 090		090 - NEGATIVE	57	57	57	100		
2023-843	072 URS-9-TECO			090 090	090 090		090 - NEGATIVE	66	66	66	100		
2023-843	080 ROCHE (CHE			090 090	090 090		090 - NEGATIVE	34	34	34	100		
2023-843	089 AIMSTRIP			090 090	090 090		090 - NEGATIVE	137	136	136	100		
2023-843	089 AIMSTRIP			090 090	090 090		094 - LARGE-3-4+	137	1	136	0		
2023-843	094 CLINITEK			090 090	090 090		090 - NEGATIVE	106	106	106	100		
2023-843	102 ACON LAB			090 090	090 090		090 - NEGATIVE	138	138	138	100		
2023-843	131 iCHEM VELO			090 090	090 090		090 - NEGATIVE	32	32	32	100		
2023-843	138 URS-10T			090 090	090 090		090 - NEGATIVE	10	10	10	100		
2023-843	140 MISSION A			090 090	090 090		090 - NEGATIVE	47	47	47	100		
2023-843	145 CLINITEK			090 090	090 090		090 - NEGATIVE	34	34	34	100		
2023-843	147 Mission120			090 090	090 090		090 - NEGATIVE	40	40	40	100		
2023-843	148 Mission500			090 090	090 090		090 - NEGATIVE	102	101	101	100		
2023-843	148 Mission500			090 090	090 090		091 - TRACE	102	1	101	0		
2023-843	151 Clari. Uro			090 090	090 090		090 - NEGATIVE	18	18	18	100		
2023-844	011 AMES NM10S			093 094	093 094		093 - MOD. (2+)	42	3	42	100		
2023-844	011 AMES NM10S			093 094	093 094		094 - LARGE-3-4+	42	39	42	100		
2023-844	013 BAYER CLI			093 094	093 094		093 - MOD. (2+)	25	2	25	100		
2023-844	013 BAYER CLI			093 094	093 094		094 - LARGE-3-4+	25	23	25	100		
2023-844	067 URISCAN			093 094	093 094		093 - MOD. (2+)	57	18	57	100		
2023-844	067 URISCAN			093 094	093 094		094 - LARGE-3-4+	57	39	57	100		
2023-844	072 URS-9-TECO			093 094	093 094		093 - MOD. (2+)	66	3	66	100		
2023-844	072 URS-9-TECO			093 094	093 094		094 - LARGE-3-4+	66	63	66	100		
2023-844	080 ROCHE (CHE			093 094	093 094		093 - MOD. (2+)	34	4	34	100		
2023-844	080 ROCHE (CHE			093 094	093 094		094 - LARGE-3-4+	34	30	34	100		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-844	089	AIMSTRIP		093 094	093 094		090 - NEGATIVE	137	1	135	0
2023-844	089	AIMSTRIP		093 094	093 094		092 - SMALL (1+)	137	1	135	0
2023-844	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	137	2	135	100
2023-844	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	137	133	135	100
2023-844	094	CLINITEK		093 094	093 094		093 - MOD. (2+)	106	4	106	100
2023-844	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	106	102	106	100
2023-844	102	ACON LAB		093 094	093 094		093 - MOD. (2+)	138	2	138	100
2023-844	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	138	136	138	100
2023-844	131	iCHEM VELO		093 094	093 094		092 - SMALL (1+)	32	1	31	0
2023-844	131	iCHEM VELO		093 094	093 094		093 - MOD. (2+)	32	15	31	100
2023-844	131	iCHEM VELO		093 094	093 094		094 - LARGE-3-4+	32	16	31	100
2023-844	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	10	10	10	100
2023-844	140	MISSION A		093 094	093 094		093 - MOD. (2+)	47	1	47	100
2023-844	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	47	46	47	100
2023-844	145	CLINITEK		093 094	093 094		094 - LARGE-3-4+	34	34	34	100
2023-844	147	Mission120		093 094	093 094		094 - LARGE-3-4+	40	40	40	100
2023-844	148	Mission500		093 094	093 094		093 - MOD. (2+)	102	1	102	100
2023-844	148	Mission500		093 094	093 094		094 - LARGE-3-4+	102	101	102	100
2023-844	151	Clari. Uro		093 094	093 094		094 - LARGE-3-4+	18	18	18	100
060 URINALYSIS-KETONES											
2023-845	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	42	42	42	100
2023-845	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	25	25	25	100
2023-845	067	URISCAN		090 090	090 090		090 - NEGATIVE	57	57	57	100
2023-845	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	66	66	66	100
2023-845	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	34	34	34	100
2023-845	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	137	137	137	100
2023-845	094	CLINITEK		090 090	090 090		090 - NEGATIVE	106	106	106	100
2023-845	102	ACON LAB		090 090	090 090		090 - NEGATIVE	138	138	138	100
2023-845	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	32	32	32	100
2023-845	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-845	140	MISSION A		090 090	090 090		090 - NEGATIVE	47	47	47	100	
2023-845	145	CLINITEK		090 090	090 090		090 - NEGATIVE	34	34	34	100	
2023-845	147	Mission120		090 090	090 090		090 - NEGATIVE	40	40	40	100	
2023-845	148	Mission500		090 090	090 090		090 - NEGATIVE	102	101	101	100	
2023-845	148	Mission500		090 090	090 090		094 - LARGE-3-4+	102	1	101	0	
2023-845	151	Clari. Uro		090 090	090 090		090 - NEGATIVE	18	18	18	100	
061 URINALYSIS-NITRATE												
2023-841	011	AMES NM10S		071 072	071 072		072 - POSITIVE	42	42	42	100	
2023-841	013	BAYER CLI		071 072	071 072		072 - POSITIVE	25	25	25	100	
2023-841	067	URISCAN		071 072	071 072		072 - POSITIVE	57	57	57	100	
2023-841	072	URS-9-TECO		071 072	071 072		072 - POSITIVE	66	66	66	100	
2023-841	080	ROCHE (CHE		071 072	071 072		072 - POSITIVE	34	34	34	100	
2023-841	089	AIMSTRIP		071 072	071 072		071 - WEAK POS.	137	1	137	100	
2023-841	089	AIMSTRIP		071 072	071 072		072 - POSITIVE	137	136	137	100	
2023-841	094	CLINITEK		071 072	071 072		072 - POSITIVE	106	106	106	100	
2023-841	102	ACON LAB		071 072	071 072		072 - POSITIVE	138	138	138	100	
2023-841	131	iCHEM VELO		071 072	071 072		072 - POSITIVE	32	32	32	100	
2023-841	138	URS-10T		071 072	071 072		072 - POSITIVE	10	10	10	100	
2023-841	140	MISSION A		071 072	071 072		072 - POSITIVE	46	46	46	100	
2023-841	145	CLINITEK		071 072	071 072		071 - WEAK POS.	34	1	34	100	
2023-841	145	CLINITEK		071 072	071 072		072 - POSITIVE	34	33	34	100	
2023-841	147	Mission120		071 072	071 072		072 - POSITIVE	41	41	41	100	
2023-841	148	Mission500		071 072	071 072		072 - POSITIVE	102	102	102	100	
2023-841	151	Clari. Uro		071 072	071 072		072 - POSITIVE	18	18	18	100	
2023-842	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	42	42	42	100	
2023-842	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	25	25	25	100	
2023-842	067	URISCAN		070 070	070 070		070 - NEGATIVE	57	57	57	100	
2023-842	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	66	66	66	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-842	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	34	33	33	100
2023-842	080	ROCHE (CHE		070 070	070 070		072 - POSITIVE	34	1	33	0
2023-842	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	137	137	137	100
2023-842	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100
2023-842	102	ACON LAB		070 070	070 070		070 - NEGATIVE	138	138	138	100
2023-842	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	32	32	32	100
2023-842	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100
2023-842	140	MISSION A		070 070	070 070		070 - NEGATIVE	46	46	46	100
2023-842	145	CLINITEK		070 070	070 070		070 - NEGATIVE	34	34	34	100
2023-842	147	Mission120		070 070	070 070		070 - NEGATIVE	41	41	41	100
2023-842	148	Mission500		070 070	070 070		070 - NEGATIVE	102	102	102	100
2023-842	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	18	18	18	100
061 URINALYSIS-NITRATE											
2023-843	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	42	42	42	100
2023-843	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	25	25	25	100
2023-843	067	URISCAN		070 070	070 070		070 - NEGATIVE	57	57	57	100
2023-843	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	66	66	66	100
2023-843	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	34	34	34	100
2023-843	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	137	137	137	100
2023-843	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100
2023-843	102	ACON LAB		070 070	070 070		070 - NEGATIVE	138	138	138	100
2023-843	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	32	32	32	100
2023-843	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100
2023-843	140	MISSION A		070 070	070 070		070 - NEGATIVE	46	46	46	100
2023-843	145	CLINITEK		070 070	070 070		070 - NEGATIVE	34	34	34	100
2023-843	147	Mission120		070 070	070 070		070 - NEGATIVE	41	41	41	100
2023-843	148	Mission500		070 070	070 070		070 - NEGATIVE	102	102	102	100
2023-843	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	18	18	18	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
061 URINALYSIS-NITRATE													
2023-844	011	AMES NMI0S		070 070	070 070		070 - NEGATIVE	42	42	42	100		
2023-844	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	25	25	25	100		
2023-844	067	URISCAN		070 070	070 070		070 - NEGATIVE	57	57	57	100		
2023-844	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	66	66	66	100		
2023-844	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	34	34	34	100		
2023-844	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	137	137	137	100		
2023-844	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100		
2023-844	102	ACON LAB		070 070	070 070		070 - NEGATIVE	138	138	138	100		
2023-844	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	32	32	32	100		
2023-844	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2023-844	140	MISSION A		070 070	070 070		070 - NEGATIVE	46	46	46	100		
2023-844	145	CLINITEK		070 070	070 070		070 - NEGATIVE	34	34	34	100		
2023-844	147	Mission120		070 070	070 070		070 - NEGATIVE	41	41	41	100		
2023-844	148	Mission500		070 070	070 070		070 - NEGATIVE	102	102	102	100		
2023-844	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	18	18	18	100		
2023-845	011	AMES NMI0S		070 070	070 070		070 - NEGATIVE	42	42	42	100		
2023-845	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	25	25	25	100		
2023-845	067	URISCAN		070 070	070 070		070 - NEGATIVE	57	57	57	100		
2023-845	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	66	66	66	100		
2023-845	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	34	34	34	100		
2023-845	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	137	137	137	100		
2023-845	094	CLINITEK		070 070	070 070		070 - NEGATIVE	106	106	106	100		
2023-845	102	ACON LAB		070 070	070 070		070 - NEGATIVE	138	138	138	100		
2023-845	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	32	32	32	100		
2023-845	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100		
2023-845	140	MISSION A		070 070	070 070		070 - NEGATIVE	46	46	46	100		
2023-845	145	CLINITEK		070 070	070 070		070 - NEGATIVE	34	34	34	100		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-845	147	Mission120		070 070	070 070		070 - NEGATIVE	41	41	41	100	
2023-845	148	Mission500		070 070	070 070		070 - NEGATIVE	102	102	102	100	
2023-845	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	18	18	18	100	
062 URINALYSIS-pH												
2023-841	011	AMES NM10S		081 085	080 084		082 - 5.0	42	13	42	100	
2023-841	011	AMES NM10S		081 085	080 084		083 - 5.5	42	20	42	100	
2023-841	011	AMES NM10S		081 085	080 084		084 - 6.0	42	8	42	100	
2023-841	011	AMES NM10S		081 085	080 084		085 - 6.5	42	1	42	100	
2023-841	013	BAYER CLI		080 084	080 084		082 - 5.0	26	13	26	100	
2023-841	013	BAYER CLI		080 084	080 084		083 - 5.5	26	8	26	100	
2023-841	013	BAYER CLI		080 084	080 084		084 - 6.0	26	5	26	100	
2023-841	067	URISCAN		080 084	080 084		082 - 5.0	57	51	57	100	
2023-841	067	URISCAN		080 084	080 084		083 - 5.5	57	6	57	100	
2023-841	072	URS-9-TECO		082 086	080 084		082 - 5.0	66	16	66	100	
2023-841	072	URS-9-TECO		082 086	080 084		083 - 5.5	66	11	66	100	
2023-841	072	URS-9-TECO		082 086	080 084		084 - 6.0	66	37	66	100	
2023-841	072	URS-9-TECO		082 086	080 084		085 - 6.5	66	2	66	100	
2023-841	080	ROCHE (CHE		080 084	080 084		082 - 5.0	34	31	34	100	
2023-841	080	ROCHE (CHE		080 084	080 084		084 - 6.0	34	3	34	100	
2023-841	089	AIMSTRIP		082 086	080 084		082 - 5.0	137	46	137	100	
2023-841	089	AIMSTRIP		082 086	080 084		083 - 5.5	137	28	137	100	
2023-841	089	AIMSTRIP		082 086	080 084		084 - 6.0	137	63	137	100	
2023-841	094	CLINITEK		082 086	080 084		082 - 5.0	106	17	106	100	
2023-841	094	CLINITEK		082 086	080 084		083 - 5.5	106	42	106	100	
2023-841	094	CLINITEK		082 086	080 084		084 - 6.0	106	47	106	100	
2023-841	102	ACON LAB		082 086	080 084		082 - 5.0	138	39	138	100	
2023-841	102	ACON LAB		082 086	080 084		083 - 5.5	138	18	138	100	
2023-841	102	ACON LAB		082 086	080 084		084 - 6.0	138	81	138	100	
2023-841	131	iCHEM VELO		080 084	080 084		082 - 5.0	32	32	32	100	
2023-841	138	URS-10T		080 084	080 084		082 - 5.0	10	8	10	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-841	138	URS-10T		080 084	080 084		084 - 6.0	10	2	10	100	
2023-841	140	MISSION A		081 085	080 084		082 - 5.0	46	10	46	100	
2023-841	140	MISSION A		081 085	080 084		083 - 5.5	46	18	46	100	
2023-841	140	MISSION A		081 085	080 084		084 - 6.0	46	17	46	100	
2023-841	140	MISSION A		081 085	080 084		085 - 6.5	46	1	46	100	
2023-841	145	CLINITEK		081 085	080 084		082 - 5.0	34	4	34	100	
2023-841	145	CLINITEK		081 085	080 084		083 - 5.5	34	29	34	100	
2023-841	145	CLINITEK		081 085	080 084		084 - 6.0	34	1	34	100	
2023-841	147	Mission120		081 085	080 084		082 - 5.0	39	15	39	100	
2023-841	147	Mission120		081 085	080 084		083 - 5.5	39	24	39	100	
2023-841	148	Mission500		082 086	080 084		082 - 5.0	103	3	103	100	
2023-841	148	Mission500		082 086	080 084		083 - 5.5	103	2	103	100	
2023-841	148	Mission500		082 086	080 084		084 - 6.0	103	98	103	100	
2023-841	151	Clari. Uro		081 085	080 084		082 - 5.0	18	6	18	100	
2023-841	151	Clari. Uro		081 085	080 084		083 - 5.5	18	7	18	100	
2023-841	151	Clari. Uro		081 085	080 084		084 - 6.0	18	5	18	100	
062 URINALYSIS-pH												
2023-842	011	AMES NM10S		086 088	086 088		087 - 7.5	42	3	42	100	
2023-842	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	42	39	42	100	
2023-842	013	BAYER CLI		086 088	086 088		084 - 6.0	26	1	25	0	
2023-842	013	BAYER CLI		086 088	086 088		087 - 7.5	26	4	25	100	
2023-842	013	BAYER CLI		086 088	086 088		088 - 8.0OR MORE	26	21	25	100	
2023-842	067	URISCAN		086 088	086 088		087 - 7.5	57	8	57	100	
2023-842	067	URISCAN		086 088	086 088		088 - 8.0OR MORE	57	49	57	100	
2023-842	072	URS-9-TECO		086 088	086 088		084 - 6.0	66	1	65	0	
2023-842	072	URS-9-TECO		086 088	086 088		086 - 7.0	66	5	65	100	
2023-842	072	URS-9-TECO		086 088	086 088		087 - 7.5	66	25	65	100	
2023-842	072	URS-9-TECO		086 088	086 088		088 - 8.0OR MORE	66	35	65	100	
2023-842	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	34	29	34	100	
2023-842	080	ROCHE (CHE		086 088	086 088		086 - 7.0	34	5	34	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-842	089	AIMSTRIP		086 088	086 088		087 - 7.5	137	11	137	100	
2023-842	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	137	126	137	100	
2023-842	094	CLINITEK		086 088	086 088		087 - 7.5	106	6	106	100	
2023-842	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	106	100	106	100	
2023-842	102	ACON LAB		086 088	086 088		087 - 7.5	138	15	138	100	
2023-842	102	ACON LAB		086 088	086 088		088 - 8.0OR MORE	138	123	138	100	
2023-842	131	iCHEM VELO		086 088	086 088		088 - 8.0OR MORE	32	32	32	100	
2023-842	138	URS-10T		086 088	086 088		087 - 7.5	10	4	10	100	
2023-842	138	URS-10T		086 088	086 088		088 - 8.0OR MORE	10	6	10	100	
2023-842	140	MISSION A		086 088	086 088		087 - 7.5	46	4	46	100	
2023-842	140	MISSION A		086 088	086 088		088 - 8.0OR MORE	46	42	46	100	
2023-842	145	CLINITEK		086 088	086 088		088 - 8.0OR MORE	34	34	34	100	
2023-842	147	Mission120		086 088	086 088		087 - 7.5	39	5	39	100	
2023-842	147	Mission120		086 088	086 088		088 - 8.0OR MORE	39	34	39	100	
2023-842	148	Mission500		086 088	086 088		087 - 7.5	103	3	103	100	
2023-842	148	Mission500		086 088	086 088		088 - 8.0OR MORE	103	100	103	100	
2023-842	151	Clari. Uro		086 088	086 088		087 - 7.5	18	2	18	100	
2023-842	151	Clari. Uro		086 088	086 088		088 - 8.0OR MORE	18	16	18	100	
062 URINALYSIS-pH												
2023-843	011	AMES NM10S		084 088	084 088		086 - 7.0	42	30	42	100	
2023-843	011	AMES NM10S		084 088	084 088		087 - 7.5	42	12	42	100	
2023-843	013	BAYER CLI		084 088	084 088		086 - 7.0	26	15	26	100	
2023-843	013	BAYER CLI		084 088	084 088		087 - 7.5	26	11	26	100	
2023-843	067	URISCAN		084 088	084 088		084 - 6.0	57	1	57	100	
2023-843	067	URISCAN		084 088	084 088		085 - 6.5	57	15	57	100	
2023-843	067	URISCAN		084 088	084 088		086 - 7.0	57	40	57	100	
2023-843	067	URISCAN		084 088	084 088		087 - 7.5	57	1	57	100	
2023-843	072	URS-9-TECO		084 088	084 088		085 - 6.5	66	3	66	100	
2023-843	072	URS-9-TECO		084 088	084 088		086 - 7.0	66	60	66	100	
2023-843	072	URS-9-TECO		084 088	084 088		087 - 7.5	66	3	66	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-843	080	ROCHE (CHE		084	088	084	088	086	- 7.0	34	34	34	100
2023-843	089	AIMSTRIP		084	088	084	088	085	- 6.5	137	2	137	100
2023-843	089	AIMSTRIP		084	088	084	088	086	- 7.0	137	116	137	100
2023-843	089	AIMSTRIP		084	088	084	088	087	- 7.5	137	19	137	100
2023-843	094	CLINITEK		084	088	084	088	085	- 6.5	106	1	106	100
2023-843	094	CLINITEK		084	088	084	088	086	- 7.0	106	66	106	100
2023-843	094	CLINITEK		084	088	084	088	087	- 7.5	106	39	106	100
2023-843	102	ACON LAB		084	088	084	088	085	- 6.5	138	5	138	100
2023-843	102	ACON LAB		084	088	084	088	086	- 7.0	138	119	138	100
2023-843	102	ACON LAB		084	088	084	088	087	- 7.5	138	14	138	100
2023-843	131	iCHEM VELO		084	088	084	088	086	- 7.0	32	32	32	100
2023-843	138	URS-10T		084	088	084	088	086	- 7.0	10	8	10	100
2023-843	138	URS-10T		084	088	084	088	087	- 7.5	10	2	10	100
2023-843	140	MISSION A		084	088	084	088	085	- 6.5	46	5	46	100
2023-843	140	MISSION A		084	088	084	088	086	- 7.0	46	40	46	100
2023-843	140	MISSION A		084	088	084	088	087	- 7.5	46	1	46	100
2023-843	145	CLINITEK		084	088	084	088	086	- 7.0	34	34	34	100
2023-843	147	Mission120		084	088	084	088	085	- 6.5	39	8	39	100
2023-843	147	Mission120		084	088	084	088	086	- 7.0	39	31	39	100
2023-843	148	Mission500		084	088	084	088	085	- 6.5	103	1	103	100
2023-843	148	Mission500		084	088	084	088	086	- 7.0	103	102	103	100
2023-843	151	Clari. Uro		084	088	084	088	086	- 7.0	18	17	18	100
2023-843	151	Clari. Uro		084	088	084	088	087	- 7.5	18	1	18	100
062 URINALYSIS-pH													
2023-844	011	AMES NM10S		086	088	086	088	088	- 8.0OR MORE	42	36	42	100
2023-844	011	AMES NM10S		086	088	086	088	087	- 7.5	42	6	42	100
2023-844	013	BAYER CLI		086	088	086	088	087	- 7.5	26	12	26	100
2023-844	013	BAYER CLI		086	088	086	088	088	- 8.0OR MORE	26	14	26	100
2023-844	067	URISCAN		085	088	086	088	087	- 7.5	57	38	57	100
2023-844	067	URISCAN		085	088	086	088	088	- 8.0OR MORE	57	19	57	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-844	072	URS-9-TECO		085 088	086 088		086 - 7.0	66	5	66	100	
2023-844	072	URS-9-TECO		085 088	086 088		087 - 7.5	66	37	66	100	
2023-844	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	66	24	66	100	
2023-844	080	ROCHE (CHE		086 088	086 088		086 - 7.0	34	13	34	100	
2023-844	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	34	21	34	100	
2023-844	089	AIMSTRIP		086 088	086 088		087 - 7.5	137	34	137	100	
2023-844	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	137	103	137	100	
2023-844	094	CLINITEK		086 088	086 088		086 - 7.0	106	1	106	100	
2023-844	094	CLINITEK		086 088	086 088		087 - 7.5	106	24	106	100	
2023-844	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	106	81	106	100	
2023-844	102	ACON LAB		086 088	086 088		086 - 7.0	138	1	138	100	
2023-844	102	ACON LAB		086 088	086 088		087 - 7.5	138	35	138	100	
2023-844	102	ACON LAB		086 088	086 088		088 - 8.0OR MORE	138	102	138	100	
2023-844	131	iCHEM VELO		084 088	086 088		084 - 6.0	32	1	32	100	
2023-844	131	iCHEM VELO		084 088	086 088		086 - 7.0	32	29	32	100	
2023-844	131	iCHEM VELO		084 088	086 088		088 - 8.0OR MORE	32	2	32	100	
2023-844	138	URS-10T		085 088	086 088		087 - 7.5	10	7	10	100	
2023-844	138	URS-10T		085 088	086 088		088 - 8.0OR MORE	10	3	10	100	
2023-844	140	MISSION A		086 088	086 088		086 - 7.0	46	1	46	100	
2023-844	140	MISSION A		086 088	086 088		087 - 7.5	46	10	46	100	
2023-844	140	MISSION A		086 088	086 088		088 - 8.0OR MORE	46	35	46	100	
2023-844	145	CLINITEK		085 088	086 088		087 - 7.5	34	24	34	100	
2023-844	145	CLINITEK		085 088	086 088		088 - 8.0OR MORE	34	10	34	100	
2023-844	147	Mission120		086 088	086 088		087 - 7.5	39	14	39	100	
2023-844	147	Mission120		086 088	086 088		088 - 8.0OR MORE	39	25	39	100	
2023-844	148	Mission500		086 088	086 088		087 - 7.5	103	15	103	100	
2023-844	148	Mission500		086 088	086 088		088 - 8.0OR MORE	103	88	103	100	
2023-844	151	Clari. Uro		086 088	086 088		087 - 7.5	18	5	18	100	
2023-844	151	Clari. Uro		086 088	086 088		088 - 8.0OR MORE	18	13	18	100	

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
062 URINALYSIS-pH											
2023-845	011	AMES NM10S		084 088	084 088		085 - 6.5	42	1	42	100
2023-845	011	AMES NM10S		084 088	084 088		086 - 7.0	42	41	42	100
2023-845	013	BAYER CLI		084 088	084 088		085 - 6.5	26	2	26	100
2023-845	013	BAYER CLI		084 088	084 088		086 - 7.0	26	24	26	100
2023-845	067	URISCAN		083 087	084 088		085 - 6.5	57	36	57	100
2023-845	067	URISCAN		083 087	084 088		086 - 7.0	57	20	57	100
2023-845	067	URISCAN		083 087	084 088		087 - 7.5	57	1	57	100
2023-845	072	URS-9-TECO		084 088	084 088		085 - 6.5	66	10	66	100
2023-845	072	URS-9-TECO		084 088	084 088		086 - 7.0	66	56	66	100
2023-845	080	ROCHE (CHE		084 088	084 088		086 - 7.0	34	34	34	100
2023-845	089	AIMSTRIP		084 088	084 088		085 - 6.5	137	24	137	100
2023-845	089	AIMSTRIP		084 088	084 088		086 - 7.0	137	113	137	100
2023-845	094	CLINITEK		084 088	084 088		085 - 6.5	106	1	106	100
2023-845	094	CLINITEK		084 088	084 088		086 - 7.0	106	105	106	100
2023-845	102	ACON LAB		084 088	084 088		085 - 6.5	138	31	138	100
2023-845	102	ACON LAB		084 088	084 088		086 - 7.0	138	107	138	100
2023-845	131	iCHEM VELO		084 088	084 088		086 - 7.0	32	32	32	100
2023-845	138	URS-10T		084 088	084 088		086 - 7.0	10	10	10	100
2023-845	140	MISSION A		084 088	084 088		085 - 6.5	46	21	46	100
2023-845	140	MISSION A		084 088	084 088		086 - 7.0	46	25	46	100
2023-845	145	CLINITEK		084 088	084 088		086 - 7.0	34	34	34	100
2023-845	147	Mission120		083 087	084 088		085 - 6.5	39	34	39	100
2023-845	147	Mission120		083 087	084 088		086 - 7.0	39	5	39	100
2023-845	148	Mission500		084 088	084 088		085 - 6.5	103	4	103	100
2023-845	148	Mission500		084 088	084 088		086 - 7.0	103	99	103	100
2023-845	151	Clari. Uro		084 088	084 088		085 - 6.5	18	4	18	100
2023-845	151	Clari. Uro		084 088	084 088		086 - 7.0	18	14	18	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
064 URIN-PROTEIN QUALI.													
2023-841	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	42	42	42	100		
2023-841	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	25	25	25	100		
2023-841	067	URISCAN		051 051	051 051		051 - NEGATIVE	57	57	57	100		
2023-841	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	67	67	67	100		
2023-841	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	34	34	34	100		
2023-841	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	136	136	136	100		
2023-841	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	106	106	100		
2023-841	102	ACON LAB		051 051	051 051		051 - NEGATIVE	139	139	139	100		
2023-841	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	32	32	32	100		
2023-841	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100		
2023-841	140	MISSION A		051 051	051 051		051 - NEGATIVE	45	45	45	100		
2023-841	145	CLINITEK		051 051	051 051		051 - NEGATIVE	34	34	34	100		
2023-841	147	Mission120		051 051	051 051		051 - NEGATIVE	41	40	40	100		
2023-841	147	Mission120		051 051	051 051		052 - TRACES	41	1	40	0		
2023-841	148	Mission500		051 051	051 051		051 - NEGATIVE	103	103	103	100		
2023-841	151	Clari. Uro		051 051	051 051		051 - NEGATIVE	17	17	17	100		
2023-842	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	42	41	41	100		
2023-842	011	AMES NM10S		051 051	051 051		052 - TRACES	42	1	41	0		
2023-842	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	25	24	24	100		
2023-842	013	BAYER CLI		051 051	051 051		052 - TRACES	25	1	24	0		
2023-842	067	URISCAN		051 051	051 051		051 - NEGATIVE	57	53	53	100		
2023-842	067	URISCAN		051 051	051 051		052 - TRACES	57	3	53	0		
2023-842	067	URISCAN		051 051	051 051		053 - 1+	57	1	53	0		
2023-842	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	67	67	67	100		
2023-842	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	34	30	30	100		
2023-842	080	ROCHE (CHE		051 051	051 051		052 - TRACES	34	4	30	0		
2023-842	089	AIMSTRIP		051 051	051 051		054 - 2+	136	1	128	0		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-842	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	136	128	128	100	
2023-842	089	AIMSTRIP		051 051	051 051		052 - TRACES	136	7	128	0	
2023-842	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	102	102	100	
2023-842	094	CLINITEK		051 051	051 051		052 - TRACES	106	4	102	0	
2023-842	102	ACON LAB		051 051	051 051		051 - NEGATIVE	139	120	120	100	
2023-842	102	ACON LAB		051 051	051 051		052 - TRACES	139	19	120	0	
2023-842	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	32	32	32	100	
2023-842	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100	
2023-842	140	MISSION A		051 051	051 051		051 - NEGATIVE	45	42	42	100	
2023-842	140	MISSION A		051 051	051 051		052 - TRACES	45	3	42	0	
2023-842	145	CLINITEK		051 051	051 051		051 - NEGATIVE	34	34	34	100	
2023-842	147	Mission120		051 051	051 051		051 - NEGATIVE	41	40	40	100	
2023-842	147	Mission120		051 051	051 051		052 - TRACES	41	1	40	0	
2023-842	148	Mission500		051 051	051 051		051 - NEGATIVE	103	85	85	100	
2023-842	148	Mission500		051 051	051 051		052 - TRACES	103	17	85	0	
2023-842	148	Mission500		051 051	051 051		053 - 1+	103	1	85	0	
2023-842	151	Clari. Uro		051 051	051 051		051 - NEGATIVE	17	17	17	100	
064 URIN-PROTEIN QUALI.												
2023-843	011	AMES NM10S		053 055	053 055		054 - 2+	42	25	41	100	
2023-843	011	AMES NM10S		053 055	053 055		053 - 1+	42	2	41	100	
2023-843	011	AMES NM10S		053 055	053 055		055 - 3+	42	14	41	100	
2023-843	011	AMES NM10S		053 055	053 055		056 - 4+	42	1	41	0	
2023-843	013	BAYER CLI		053 055	053 055		054 - 2+	25	16	24	100	
2023-843	013	BAYER CLI		053 055	053 055		053 - 1+	25	1	24	100	
2023-843	013	BAYER CLI		053 055	053 055		055 - 3+	25	7	24	100	
2023-843	013	BAYER CLI		053 055	053 055		056 - 4+	25	1	24	0	
2023-843	067	URISCAN		053 055	053 055		054 - 2+	57	45	57	100	
2023-843	067	URISCAN		053 055	053 055		053 - 1+	57	12	57	100	
2023-843	072	URS-9-TECO		053 055	053 055		054 - 2+	67	39	67	100	
2023-843	072	URS-9-TECO		053 055	053 055		053 - 1+	67	3	67	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-843	072	URS-9-TECO		053 055	053 055		055 - 3+	67	25	67	100		
2023-843	080	ROCHE (CHE		053 055	053 055		054 - 2+	34	24	33	100		
2023-843	080	ROCHE (CHE		053 055	053 055		053 - 1+	34	1	33	100		
2023-843	080	ROCHE (CHE		053 055	053 055		055 - 3+	34	8	33	100		
2023-843	080	ROCHE (CHE		053 055	053 055		056 - 4+	34	1	33	0		
2023-843	089	AIMSTRIP		053 055	053 055		054 - 2+	136	94	135	100		
2023-843	089	AIMSTRIP		053 055	053 055		053 - 1+	136	1	135	100		
2023-843	089	AIMSTRIP		053 055	053 055		055 - 3+	136	40	135	100		
2023-843	089	AIMSTRIP		053 055	053 055		056 - 4+	136	1	135	0		
2023-843	094	CLINITEK		053 055	053 055		054 - 2+	106	88	106	100		
2023-843	094	CLINITEK		053 055	053 055		053 - 1+	106	2	106	100		
2023-843	094	CLINITEK		053 055	053 055		055 - 3+	106	16	106	100		
2023-843	102	ACON LAB		053 055	053 055		054 - 2+	139	108	139	100		
2023-843	102	ACON LAB		053 055	053 055		053 - 1+	139	6	139	100		
2023-843	102	ACON LAB		053 055	053 055		055 - 3+	139	25	139	100		
2023-843	131	iCHEM VELO		053 055	053 055		054 - 2+	32	29	32	100		
2023-843	131	iCHEM VELO		053 055	053 055		055 - 3+	32	3	32	100		
2023-843	138	URS-10T		054 056	053 055		054 - 2+	10	4	10	100		
2023-843	138	URS-10T		054 056	053 055		055 - 3+	10	6	10	100		
2023-843	140	MISSION A		053 055	053 055		054 - 2+	45	40	45	100		
2023-843	140	MISSION A		053 055	053 055		055 - 3+	45	5	45	100		
2023-843	145	CLINITEK		053 055	053 055		054 - 2+	34	25	33	100		
2023-843	145	CLINITEK		053 055	053 055		053 - 1+	34	1	33	100		
2023-843	145	CLINITEK		053 055	053 055		055 - 3+	34	7	33	100		
2023-843	145	CLINITEK		053 055	053 055		056 - 4+	34	1	33	0		
2023-843	147	Mission120		053 055	053 055		054 - 2+	41	38	41	100		
2023-843	147	Mission120		053 055	053 055		053 - 1+	41	3	41	100		
2023-843	148	Mission500		053 055	053 055		054 - 2+	103	82	103	100		
2023-843	148	Mission500		053 055	053 055		053 - 1+	103	10	103	100		
2023-843	148	Mission500		053 055	053 055		055 - 3+	103	11	103	100		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-843	151	Clari. Uro		053 055	053 055		054 - 2+	17	14	17	100	
2023-843	151	Clari. Uro		053 055	053 055		053 - 1+	17	1	17	100	
2023-843	151	Clari. Uro		053 055	053 055		055 - 3+	17	2	17	100	
064 URIN-PROTEIN QUALI.												
2023-844	011	AMES NM10S		054 056	053 055		054 - 2+	42	18	42	100	
2023-844	011	AMES NM10S		054 056	053 055		053 - 1+	42	1	42	100	
2023-844	011	AMES NM10S		054 056	053 055		055 - 3+	42	22	42	100	
2023-844	011	AMES NM10S		054 056	053 055		056 - 4+	42	1	42	100	
2023-844	013	BAYER CLI		053 055	053 055		054 - 2+	25	13	24	100	
2023-844	013	BAYER CLI		053 055	053 055		055 - 3+	25	11	24	100	
2023-844	013	BAYER CLI		053 055	053 055		056 - 4+	25	1	24	0	
2023-844	067	URISCAN		053 055	053 055		054 - 2+	57	47	56	100	
2023-844	067	URISCAN		053 055	053 055		053 - 1+	57	4	56	100	
2023-844	067	URISCAN		053 055	053 055		055 - 3+	57	5	56	100	
2023-844	067	URISCAN		053 055	053 055		056 - 4+	57	1	56	0	
2023-844	072	URS-9-TECO		054 056	053 055		054 - 2+	67	28	67	100	
2023-844	072	URS-9-TECO		054 056	053 055		053 - 1+	67	1	67	100	
2023-844	072	URS-9-TECO		054 056	053 055		055 - 3+	67	38	67	100	
2023-844	080	ROCHE (CHE		053 055	053 055		054 - 2+	34	27	33	100	
2023-844	080	ROCHE (CHE		053 055	053 055		055 - 3+	34	6	33	100	
2023-844	080	ROCHE (CHE		053 055	053 055		056 - 4+	34	1	33	0	
2023-844	089	AIMSTRIP		054 056	053 055		054 - 2+	136	61	136	100	
2023-844	089	AIMSTRIP		054 056	053 055		053 - 1+	136	1	136	100	
2023-844	089	AIMSTRIP		054 056	053 055		055 - 3+	136	74	136	100	
2023-844	094	CLINITEK		053 055	053 055		054 - 2+	106	84	106	100	
2023-844	094	CLINITEK		053 055	053 055		053 - 1+	106	2	106	100	
2023-844	094	CLINITEK		053 055	053 055		055 - 3+	106	20	106	100	
2023-844	102	ACON LAB		053 055	053 055		054 - 2+	139	82	137	100	
2023-844	102	ACON LAB		053 055	053 055		053 - 1+	139	3	137	100	
2023-844	102	ACON LAB		053 055	053 055		055 - 3+	139	52	137	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-844	102	ACON LAB		053 055	053 055		056 - 4+	139	2	137	0	
2023-844	131	iCHEM VELO		053 055	053 055		054 - 2+	32	19	30	100	
2023-844	131	iCHEM VELO		053 055	053 055		051 - NEGATIVE	32	1	30	0	
2023-844	131	iCHEM VELO		053 055	053 055		055 - 3+	32	11	30	100	
2023-844	131	iCHEM VELO		053 055	053 055		056 - 4+	32	1	30	0	
2023-844	138	URS-10T		054 056	053 055		054 - 2+	10	3	10	100	
2023-844	138	URS-10T		054 056	053 055		055 - 3+	10	6	10	100	
2023-844	138	URS-10T		054 056	053 055		056 - 4+	10	1	10	100	
2023-844	140	MISSION A		053 055	053 055		054 - 2+	45	34	45	100	
2023-844	140	MISSION A		053 055	053 055		055 - 3+	45	11	45	100	
2023-844	145	CLINITEK		053 055	053 055		054 - 2+	34	18	34	100	
2023-844	145	CLINITEK		053 055	053 055		053 - 1+	34	1	34	100	
2023-844	145	CLINITEK		053 055	053 055		055 - 3+	34	15	34	100	
2023-844	147	Mission120		053 055	053 055		054 - 2+	41	36	41	100	
2023-844	147	Mission120		053 055	053 055		053 - 1+	41	4	41	100	
2023-844	147	Mission120		053 055	053 055		055 - 3+	41	1	41	100	
2023-844	148	Mission500		053 055	053 055		054 - 2+	103	49	103	100	
2023-844	148	Mission500		053 055	053 055		053 - 1+	103	6	103	100	
2023-844	148	Mission500		053 055	053 055		055 - 3+	103	48	103	100	
2023-844	151	Clari. Uro		053 055	053 055		054 - 2+	17	11	17	100	
2023-844	151	Clari. Uro		053 055	053 055		055 - 3+	17	6	17	100	
064 URIN-PROTEIN QUALI.												
2023-845	011	AMES NM10S		053 055	053 055		054 - 2+	42	21	41	100	
2023-845	011	AMES NM10S		053 055	053 055		053 - 1+	42	1	41	100	
2023-845	011	AMES NM10S		053 055	053 055		055 - 3+	42	19	41	100	
2023-845	011	AMES NM10S		053 055	053 055		056 - 4+	42	1	41	0	
2023-845	013	BAYER CLI		053 055	053 055		054 - 2+	25	14	24	100	
2023-845	013	BAYER CLI		053 055	053 055		053 - 1+	25	1	24	100	
2023-845	013	BAYER CLI		053 055	053 055		055 - 3+	25	9	24	100	
2023-845	013	BAYER CLI		053 055	053 055		056 - 4+	25	1	24	0	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-845	067	URISCAN		053 055	053 055		054 - 2+	57	46	57	100
2023-845	067	URISCAN		053 055	053 055		053 - 1+	57	10	57	100
2023-845	067	URISCAN		053 055	053 055		055 - 3+	57	1	57	100
2023-845	072	URS-9-TECO		053 055	053 055		054 - 2+	67	33	67	100
2023-845	072	URS-9-TECO		053 055	053 055		053 - 1+	67	3	67	100
2023-845	072	URS-9-TECO		053 055	053 055		055 - 3+	67	31	67	100
2023-845	080	ROCHE (CHE		053 055	053 055		054 - 2+	34	25	33	100
2023-845	080	ROCHE (CHE		053 055	053 055		055 - 3+	34	8	33	100
2023-845	080	ROCHE (CHE		053 055	053 055		056 - 4+	34	1	33	0
2023-845	089	AIMSTRIP		053 055	053 055		054 - 2+	136	80	136	100
2023-845	089	AIMSTRIP		053 055	053 055		053 - 1+	136	2	136	100
2023-845	089	AIMSTRIP		053 055	053 055		055 - 3+	136	54	136	100
2023-845	094	CLINITEK		053 055	053 055		054 - 2+	106	86	106	100
2023-845	094	CLINITEK		053 055	053 055		053 - 1+	106	2	106	100
2023-845	094	CLINITEK		053 055	053 055		055 - 3+	106	18	106	100
2023-845	102	ACON LAB		053 055	053 055		054 - 2+	139	96	136	100
2023-845	102	ACON LAB		053 055	053 055		051 - NEGATIVE	139	1	136	0
2023-845	102	ACON LAB		053 055	053 055		053 - 1+	139	6	136	100
2023-845	102	ACON LAB		053 055	053 055		055 - 3+	139	34	136	100
2023-845	102	ACON LAB		053 055	053 055		056 - 4+	139	2	136	0
2023-845	131	iCHEM VELO		053 055	053 055		054 - 2+	32	25	32	100
2023-845	131	iCHEM VELO		053 055	053 055		055 - 3+	32	7	32	100
2023-845	138	URS-10T		054 056	053 055		054 - 2+	10	4	10	100
2023-845	138	URS-10T		054 056	053 055		055 - 3+	10	5	10	100
2023-845	138	URS-10T		054 056	053 055		056 - 4+	10	1	10	100
2023-845	140	MISSION A		053 055	053 055		054 - 2+	45	37	45	100
2023-845	140	MISSION A		053 055	053 055		055 - 3+	45	8	45	100
2023-845	145	CLINITEK		054 056	053 055		054 - 2+	34	11	34	100
2023-845	145	CLINITEK		054 056	053 055		053 - 1+	34	1	34	100
2023-845	145	CLINITEK		054 056	053 055		055 - 3+	34	21	34	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-845	145	CLINITEK		054 056	053 055		056 - 4+	34	1	34	100	
2023-845	147	Mission120		053 055	053 055		054 - 2+	41	37	41	100	
2023-845	147	Mission120		053 055	053 055		053 - 1+	41	4	41	100	
2023-845	148	Mission500		053 055	053 055		054 - 2+	103	71	103	100	
2023-845	148	Mission500		053 055	053 055		053 - 1+	103	8	103	100	
2023-845	148	Mission500		053 055	053 055		055 - 3+	103	24	103	100	
2023-845	151	Clari. Uro		053 055	053 055		055 - 3+	17	5	17	100	
2023-845	151	Clari. Uro		053 055	053 055		054 - 2+	17	12	17	100	
066 URINAL.-UROBILINOGEN												
2023-841	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	42	42	42	100	
2023-841	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	25	25	25	100	
2023-841	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2023-841	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	67	67	67	100	
2023-841	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-841	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	137	137	137	100	
2023-841	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	107	107	100	
2023-841	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	136	136	136	100	
2023-841	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	32	32	32	100	
2023-841	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2023-841	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	47	46	46	100	
2023-841	140	MISSION A		061 061	061 061		062 - 2	47	1	46	0	
2023-841	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-841	147	Mission120		061 061	061 061		061 - NORMAL,NEG	41	41	41	100	
2023-841	148	Mission500		061 061	061 061		061 - NORMAL,NEG	101	100	100	100	
2023-841	148	Mission500		061 061	061 061		062 - 2	101	1	100	0	
2023-841	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2023-842	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	42	42	42	100	
2023-842	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	25	25	25	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-842	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2023-842	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	67	67	67	100	
2023-842	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-842	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	137	137	137	100	
2023-842	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	107	107	100	
2023-842	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	136	136	136	100	
2023-842	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	32	32	32	100	
2023-842	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2023-842	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	47	46	46	100	
2023-842	140	MISSION A		061 061	061 061		062 - 2	47	1	46	0	
2023-842	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-842	147	Mission120		061 061	061 061		061 - NORMAL,NEG	41	41	41	100	
2023-842	148	Mission500		061 061	061 061		061 - NORMAL,NEG	101	100	100	100	
2023-842	148	Mission500		061 061	061 061		062 - 2	101	1	100	0	
2023-842	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
066 URINAL.-UROBILINOGEN												
2023-843	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	42	42	42	100	
2023-843	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	25	25	25	100	
2023-843	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2023-843	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	67	67	67	100	
2023-843	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-843	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	137	137	137	100	
2023-843	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	107	107	100	
2023-843	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	136	136	136	100	
2023-843	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	32	32	32	100	
2023-843	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2023-843	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	47	46	46	100	
2023-843	140	MISSION A		061 061	061 061		062 - 2	47	1	46	0	
2023-843	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-843	147	Mission120		061 061	061 061		061 - NORMAL,NEG	41	41	41	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-843	148	Mission500		061 061	061 061		061 - NORMAL,NEG	101	100	100	100		
2023-843	148	Mission500		061 061	061 061		062 - 2	101	1	100	0		
2023-843	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	16	16	16	100		
066 URINAL.-UROBILINOGEN													
2023-844	011	AMES NM10S		061 063	061 063		061 - NORMAL,NEG	42	5	38	100		
2023-844	011	AMES NM10S		061 063	061 063		062 - 2	42	18	38	100		
2023-844	011	AMES NM10S		061 063	061 063		063 - 4	42	15	38	100		
2023-844	011	AMES NM10S		061 063	061 063		064 - 8	42	4	38	0		
2023-844	013	BAYER CLI		061 063	061 063		061 - NORMAL,NEG	25	8	25	100		
2023-844	013	BAYER CLI		061 063	061 063		062 - 2	25	14	25	100		
2023-844	013	BAYER CLI		061 063	061 063		063 - 4	25	3	25	100		
2023-844	067	URISCAN			061 063		061 - NORMAL,NEG	57	26	55	100		
2023-844	067	URISCAN			061 063		062 - 2	57	10	55	100		
2023-844	067	URISCAN			061 063		063 - 4	57	19	55	100		
2023-844	067	URISCAN			061 063		064 - 8	57	2	55	0		
2023-844	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	67	9	62	100		
2023-844	072	URS-9-TECO		061 063	061 063		062 - 2	67	40	62	100		
2023-844	072	URS-9-TECO		061 063	061 063		063 - 4	67	13	62	100		
2023-844	072	URS-9-TECO		061 063	061 063		064 - 8	67	5	62	0		
2023-844	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	34	29	34	100		
2023-844	080	ROCHE (CHE		061 061	061 063		062 - 2	34	5	34	100		
2023-844	089	AIMSTRIP			061 063		061 - NORMAL,NEG	137	94	137	100		
2023-844	089	AIMSTRIP			061 063		062 - 2	137	35	137	100		
2023-844	089	AIMSTRIP			061 063		063 - 4	137	8	137	100		
2023-844	094	CLINITEK		062 064	061 063		061 - NORMAL,NEG	107	10	107	100		
2023-844	094	CLINITEK		062 064	061 063		062 - 2	107	41	107	100		
2023-844	094	CLINITEK		062 064	061 063		063 - 4	107	52	107	100		
2023-844	094	CLINITEK		062 064	061 063		064 - 8	107	4	107	100		
2023-844	102	ACON LAB			061 063		061 - NORMAL,NEG	136	107	136	100		
2023-844	102	ACON LAB			061 063		062 - 2	136	25	136	100		

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-844	102	ACON LAB			061	063	063 - 4	136	4	136	100	
2023-844	131	iCHEM VELO		062	064	061	063	061 - NORMAL,NEG	32	1	32	100
2023-844	131	iCHEM VELO		062	064	061	063	063 - 4	32	31	32	100
2023-844	138	URS-10T		062	064	061	063	061 - NORMAL,NEG	10	1	10	100
2023-844	138	URS-10T		062	064	061	063	062 - 2	10	3	10	100
2023-844	138	URS-10T		062	064	061	063	063 - 4	10	6	10	100
2023-844	140	MISSION A			061	063	062 - 2	47	7	47	100	
2023-844	140	MISSION A			061	063	063 - 4	47	3	47	100	
2023-844	140	MISSION A			061	063	061 - NORMAL,NEG	47	37	47	100	
2023-844	145	CLINITEK		061	063	061	063	061 - NORMAL,NEG	34	3	34	100
2023-844	145	CLINITEK		061	063	061	063	062 - 2	34	30	34	100
2023-844	145	CLINITEK		061	063	061	063	063 - 4	34	1	34	100
2023-844	147	Mission120		061	061	061	063	061 - NORMAL,NEG	41	38	41	100
2023-844	147	Mission120		061	061	061	063	062 - 2	41	3	41	100
2023-844	148	Mission500		061	061	061	063	061 - NORMAL,NEG	101	95	101	100
2023-844	148	Mission500		061	061	061	063	062 - 2	101	6	101	100
2023-844	151	Clari. Uro			061	063	061 - NORMAL,NEG	16	9	16	100	
2023-844	151	Clari. Uro			061	063	062 - 2	16	5	16	100	
2023-844	151	Clari. Uro			061	063	063 - 4	16	2	16	100	
066 URINAL.-UROBILINOGEN												
2023-845	011	AMES NM10S		061	061	061	061	061 - NORMAL,NEG	42	42	42	100
2023-845	013	BAYER CLI		061	061	061	061	061 - NORMAL,NEG	25	25	25	100
2023-845	067	URISCAN		061	061	061	061	061 - NORMAL,NEG	57	57	57	100
2023-845	072	URS-9-TECO		061	061	061	061	061 - NORMAL,NEG	67	67	67	100
2023-845	080	ROCHE (CHE		061	061	061	061	061 - NORMAL,NEG	34	34	34	100
2023-845	089	AIMSTRIP		061	061	061	061	061 - NORMAL,NEG	137	137	137	100
2023-845	094	CLINITEK		061	061	061	061	061 - NORMAL,NEG	107	107	107	100
2023-845	102	ACON LAB		061	061	061	061	061 - NORMAL,NEG	136	136	136	100
2023-845	131	iCHEM VELO		061	061	061	061	061 - NORMAL,NEG	32	32	32	100
2023-845	138	URS-10T		061	061	061	061	061 - NORMAL,NEG	10	10	10	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-845	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	47	46	46	100	
2023-845	140	MISSION A		061 061	061 061		062 - 2	47	1	46	0	
2023-845	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2023-845	147	Mission120		061 061	061 061		061 - NORMAL,NEG	41	41	41	100	
2023-845	148	Mission500		061 061	061 061		061 - NORMAL,NEG	101	100	100	100	
2023-845	148	Mission500		061 061	061 061		062 - 2	101	1	100	0	
2023-845	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
184 URINALYSIS LEUKOCYTE												
2023-841	011	AMES NM10S		002 004	003 005		002 - Traces	42	2	42	100	
2023-841	011	AMES NM10S		002 004	003 005		003 - Small	42	32	42	100	
2023-841	011	AMES NM10S		002 004	003 005		004 - Moderate	42	7	42	100	
2023-841	011	AMES NM10S		002 004	003 005		005 - Large	42	1	42	100	
2023-841	013	BAYER CLI		002 004	003 005		003 - Small	26	23	26	100	
2023-841	013	BAYER CLI		002 004	003 005		004 - Moderate	26	3	26	100	
2023-841	067	URISCAN		003 005	003 005		003 - Small	57	5	57	100	
2023-841	067	URISCAN		003 005	003 005		004 - Moderate	57	26	57	100	
2023-841	067	URISCAN		003 005	003 005		005 - Large	57	26	57	100	
2023-841	072	URS-9-TECO		003 005	003 005		002 - Traces	64	1	63	0	
2023-841	072	URS-9-TECO		003 005	003 005		003 - Small	64	21	63	100	
2023-841	072	URS-9-TECO		003 005	003 005		004 - Moderate	64	40	63	100	
2023-841	072	URS-9-TECO		003 005	003 005		005 - Large	64	2	63	100	
2023-841	080	ROCHE (CHE		004 005	003 005		004 - Moderate	31	1	31	100	
2023-841	080	ROCHE (CHE		004 005	003 005		005 - Large	31	30	31	100	
2023-841	089	AIMSTRIP		002 004	003 005		002 - Traces	136	6	136	100	
2023-841	089	AIMSTRIP		002 004	003 005		003 - Small	136	97	136	100	
2023-841	089	AIMSTRIP		002 004	003 005		004 - Moderate	136	31	136	100	
2023-841	089	AIMSTRIP		002 004	003 005		005 - Large	136	2	136	100	
2023-841	094	CLINITEK		002 004	003 005		003 - Small	105	97	105	100	
2023-841	094	CLINITEK		002 004	003 005		004 - Moderate	105	7	105	100	
2023-841	094	CLINITEK		002 004	003 005		005 - Large	105	1	105	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-841	102	ACON LAB		003 005	003 005		001 - Negative	128	1	124	0	
2023-841	102	ACON LAB		003 005	003 005		002 - Traces	128	3	124	0	
2023-841	102	ACON LAB		003 005	003 005		003 - Small	128	58	124	100	
2023-841	102	ACON LAB		003 005	003 005		004 - Moderate	128	65	124	100	
2023-841	102	ACON LAB		003 005	003 005		005 - Large	128	1	124	100	
2023-841	131	iCHEM VELO		004 005	003 005		004 - Moderate	32	3	32	100	
2023-841	131	iCHEM VELO		004 005	003 005		005 - Large	32	29	32	100	
2023-841	140	MISSION A		002 004	003 005		002 - Traces	50	3	50	100	
2023-841	140	MISSION A		002 004	003 005		003 - Small	50	27	50	100	
2023-841	140	MISSION A		002 004	003 005		004 - Moderate	50	19	50	100	
2023-841	140	MISSION A		002 004	003 005		005 - Large	50	1	50	100	
2023-841	145	CLINITEK		003 005	003 005		003 - Small	34	1	34	100	
2023-841	145	CLINITEK		003 005	003 005		004 - Moderate	34	33	34	100	
2023-841	147	Mission120		002 004	003 005		002 - Traces	39	5	39	100	
2023-841	147	Mission120		002 004	003 005		003 - Small	39	30	39	100	
2023-841	147	Mission120		002 004	003 005		004 - Moderate	39	4	39	100	
2023-841	148	Mission500		003 005	003 005		001 - Negative	104	1	98	0	
2023-841	148	Mission500		003 005	003 005		002 - Traces	104	5	98	0	
2023-841	148	Mission500		003 005	003 005		003 - Small	104	45	98	100	
2023-841	148	Mission500		003 005	003 005		004 - Moderate	104	51	98	100	
2023-841	148	Mission500		003 005	003 005		005 - Large	104	2	98	100	
2023-841	151	Clari. Uro		002 004	003 005		002 - Traces	18	1	18	100	
2023-841	151	Clari. Uro		002 004	003 005		003 - Small	18	9	18	100	
2023-841	151	Clari. Uro		002 004	003 005		004 - Moderate	18	7	18	100	
2023-841	151	Clari. Uro		002 004	003 005		005 - Large	18	1	18	100	
184 URINALYSIS LEUKOCYTE												
2023-842	011	AMES NM10S		001 001	001 001		001 - Negative	42	42	42	100	
2023-842	013	BAYER CLI		001 001	001 001		001 - Negative	26	26	26	100	
2023-842	067	URISCAN		001 001	001 001		001 - Negative	57	57	57	100	
2023-842	072	URS-9-TECO		001 001	001 001		001 - Negative	64	64	64	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2023-842	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100	
2023-842	089	AIMSTRIP		001 001	001 001		001 - Negative	136	136	136	100	
2023-842	094	CLINITEK		001 001	001 001		001 - Negative	105	105	105	100	
2023-842	102	ACON LAB		001 001	001 001		001 - Negative	128	128	128	100	
2023-842	131	iCHEM VELO		001 001	001 001		001 - Negative	32	32	32	100	
2023-842	140	MISSION A		001 001	001 001		001 - Negative	50	50	50	100	
2023-842	145	CLINITEK		001 001	001 001		001 - Negative	34	34	34	100	
2023-842	147	Mission120		001 001	001 001		001 - Negative	39	39	39	100	
2023-842	148	Mission500		001 001	001 001		001 - Negative	104	104	104	100	
2023-842	151	Clari. Uro		001 001	001 001		001 - Negative	18	18	18	100	
184 URINALYSIS LEUKOCYTE												
2023-843	011	AMES NM10S		001 001	001 001		001 - Negative	42	42	42	100	
2023-843	013	BAYER CLI		001 001	001 001		001 - Negative	26	26	26	100	
2023-843	067	URISCAN		001 001	001 001		001 - Negative	57	57	57	100	
2023-843	072	URS-9-TECO		001 001	001 001		001 - Negative	64	64	64	100	
2023-843	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100	
2023-843	089	AIMSTRIP		001 001	001 001		001 - Negative	136	136	136	100	
2023-843	094	CLINITEK		001 001	001 001		001 - Negative	105	105	105	100	
2023-843	102	ACON LAB		001 001	001 001		001 - Negative	128	128	128	100	
2023-843	131	iCHEM VELO		001 001	001 001		001 - Negative	32	32	32	100	
2023-843	140	MISSION A		001 001	001 001		001 - Negative	50	50	50	100	
2023-843	145	CLINITEK		001 001	001 001		001 - Negative	34	34	34	100	
2023-843	147	Mission120		001 001	001 001		001 - Negative	39	39	39	100	
2023-843	148	Mission500		001 001	001 001		001 - Negative	104	104	104	100	
2023-843	151	Clari. Uro		001 001	001 001		001 - Negative	18	18	18	100	
2023-844	011	AMES NM10S		001 001	001 001		001 - Negative	42	42	42	100	
2023-844	013	BAYER CLI		001 001	001 001		001 - Negative	26	26	26	100	
2023-844	067	URISCAN		001 001	001 001		001 - Negative	57	57	57	100	

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2023-844	072	URS-9-TECO		001 001	001 001		001 - Negative	64	64	64	100
2023-844	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100
2023-844	089	AIMSTRIP		001 001	001 001		001 - Negative	136	135	135	100
2023-844	089	AIMSTRIP		001 001	001 001		005 - Large	136	1	135	0
2023-844	094	CLINITEK		001 001	001 001		001 - Negative	105	105	105	100
2023-844	102	ACON LAB		001 001	001 001		001 - Negative	128	128	128	100
2023-844	131	iCHEM VELO		001 001	001 001		001 - Negative	32	32	32	100
2023-844	140	MISSION A		001 001	001 001		001 - Negative	50	49	49	100
2023-844	140	MISSION A		001 001	001 001		002 - Traces	50	1	49	0
2023-844	145	CLINITEK		001 001	001 001		001 - Negative	34	34	34	100
2023-844	147	Mission120		001 001	001 001		001 - Negative	39	39	39	100
2023-844	148	Mission500		001 001	001 001		001 - Negative	104	104	104	100
2023-844	151	Clari. Uro		001 001	001 001		001 - Negative	18	18	18	100
184 URINALYSIS LEUKOCYTE											
2023-845	011	AMES NM10S		001 001	001 001		001 - Negative	42	42	42	100
2023-845	013	BAYER CLI		001 001	001 001		001 - Negative	26	26	26	100
2023-845	067	URISCAN		001 001	001 001		001 - Negative	57	57	57	100
2023-845	072	URS-9-TECO		001 001	001 001		001 - Negative	64	64	64	100
2023-845	080	ROCHE (CHE		001 001	001 001		001 - Negative	31	31	31	100
2023-845	089	AIMSTRIP		001 001	001 001		001 - Negative	136	135	135	100
2023-845	089	AIMSTRIP		001 001	001 001		005 - Large	136	1	135	0
2023-845	094	CLINITEK		001 001	001 001		001 - Negative	105	105	105	100
2023-845	102	ACON LAB		001 001	001 001		001 - Negative	128	128	128	100
2023-845	131	iCHEM VELO		001 001	001 001		001 - Negative	32	32	32	100
2023-845	140	MISSION A		001 001	001 001		001 - Negative	50	49	49	100
2023-845	140	MISSION A		001 001	001 001		002 - Traces	50	1	49	0
2023-845	145	CLINITEK		001 001	001 001		001 - Negative	34	34	34	100
2023-845	147	Mission120		001 001	001 001		001 - Negative	39	39	39	100
2023-845	148	Mission500		001 001	001 001		001 - Negative	104	104	104	100
2023-845	151	Clari. Uro		001 001	001 001		001 - Negative	18	18	18	100

Participant Summary
MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2023-841	091	AIM STEP		001	001	001	001	001 - NEGATIVE	33	33	33	100
2023-841	101	HCG Alere		001	001	001	001	001 - NEGATIVE	25	25	25	100
2023-842	091	AIM STEP		001	001	001	001	001 - NEGATIVE	33	33	33	100
2023-842	101	HCG Alere		001	001	001	001	001 - NEGATIVE	25	25	25	100
2023-843	091	AIM STEP		001	001	001	001	001 - NEGATIVE	33	33	33	100
2023-843	101	HCG Alere		001	001	001	001	001 - NEGATIVE	25	25	25	100
2023-844	091	AIM STEP		002	002	002	002	002 - POSITIVE	33	33	33	100
2023-844	101	HCG Alere		002	002	002	002	002 - POSITIVE	25	25	25	100
2023-845	091	AIM STEP		002	002	002	002	002 - POSITIVE	33	33	33	100
2023-845	101	HCG Alere		002	002	002	002	002 - POSITIVE	25	25	25	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2023-851					028	028	014 - CA. OXALAT	904	25	878	0
2023-851					028	028	028 - TRIPLE PH.	904	878	878	100
2023-851					028	028	037 - SPERMAT. OZ.	904	1	878	0
2023-852					031	031	031 - FIBERS	904	884	884	100
2023-852					031	031	032 - MUCOUS TH.	904	19	884	0
2023-852					031	031	040 - S. haemat.	904	1	884	0
2023-853					037	037	028 - TRIPLE PH.	904	1	902	0
2023-853					037	037	037 - SPERMAT. OZ.	904	902	902	100
2023-853					037	037	039 - E. vermic.	904	1	902	0
2023-854					040	040	024 - AMMONIUM B	905	1	903	0
2023-854					040	040	031 - FIBERS	905	1	903	0
2023-854					040	040	040 - S. haemat.	905	903	903	100
2023-855					024	024	013 - AMORPHOUS	904	2	902	0
2023-855					024	024	024 - AMMONIUM B	904	902	902	100

Participant Summary

MONTH 10 OF YEAR 2023

SPE: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 98 OCCULT BLOOD						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
154 FECAL OCCULT BLOOD												
2023-861	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	401	401	401	100
2023-861	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	732	732	732	100
2023-862	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	401	401	401	100
2023-862	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	732	732	732	100
2023-863	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	401	401	401	100
2023-863	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	732	731	731	100
2023-863	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	732	1	731	0
2023-864	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	401	401	401	100
2023-864	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	732	732	732	100
2023-865	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	401	400	400	100
2023-865	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	401	1	400	0
2023-865	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	732	728	728	100
2023-865	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	732	4	728	0

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 90 SPECIAL TEST					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
068 CARCINOE. ANTIG. CEA				CLIA TEST ID #: 0							
				EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}							
2023-811	29	VITROS		12	33.208	0.872	2.63	30.600	35.800		
2023-811	38	ROCHE E		19	27.458	1.342	4.89	23.400	31.500		
2023-811	44	BECK.COUL.		19	23.837	1.105	4.63	20.500	27.200		
2023-811	50	ARC		10	35.180	2.078	5.91	28.900	41.400		

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2023-812	29 VITROS			12	12.067	0.256	2.12	11.300	12.800
2023-812	38 ROCHE E			19	10.679	0.432	4.05	9.400	12.000
2023-812	44 BECK.COUL.			19	8.995	0.432	4.81	7.700	10.300
2023-812	50 ARC			10	12.670	0.535	4.22	11.100	14.300
2023-813	29 VITROS			12	7.318	0.199	2.72	6.700	7.900
2023-813	38 ROCHE E			19	6.768	0.292	4.32	5.900	7.600
2023-813	44 BECK.COUL.			19	5.500	0.258	4.68	4.700	6.300
2023-813	50 ARC			10	8.310	0.386	4.64	7.200	9.500
2023-814	29 VITROS			12	47.700	1.734	3.64	42.500	52.900
2023-814	38 ROCHE E			19	37.963	1.585	4.18	33.200	42.700
2023-814	44 BECK.COUL.			19	34.663	1.705	4.92	29.500	39.800
2023-814	50 ARC			10	50.050	1.622	3.24	45.200	54.900
2023-815	29 VITROS			12	24.092	0.637	2.64	22.200	26.000
2023-815	38 ROCHE E			19	20.353	0.804	3.95	17.900	22.800
2023-815	44 BECK.COUL.			19	17.379	0.767	4.42	15.100	19.700
2023-815	50 ARC			10	25.540	0.825	3.23	23.100	28.000
069	PROS. SPEC-ANTIG-PSA								
2023-811	29 VITROS			44	11.193	0.398	3.55	10.000	12.400
2023-811	36 TOSOH AIA			10	14.320	0.690	4.82	12.300	16.400
2023-811	38 ROCHE E			49	21.622	1.173	5.43	18.100	25.100
2023-811	44 BECK.COUL.			53	21.242	1.220	5.74	17.600	24.900
2023-811	50 ARC			22	17.382	1.070	6.15	14.200	20.600
2023-811	57 S. EXL 200			66	22.354	1.216	5.44	18.700	26.000
2023-811	58 ATELLICA			11	18.309	0.875	4.78	15.700	20.900
2023-812	29 VITROS			44	4.328	0.200	4.63	3.700	4.900

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 90 SPECIAL TEST			
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-812	36	TOSOH AIA		10	5.460	0.452	8.28	4.100	6.800
2023-812	38	ROCHE E		49	8.102	0.429	5.30	6.800	9.400
2023-812	44	BECK.COUL.		53	7.839	0.448	5.71	6.500	9.200
2023-812	50	ARC		22	6.427	0.378	5.88	5.300	7.600
2023-812	57	S. EXL 200		66	8.097	0.440	5.43	6.800	9.400
2023-812	58	ATELLICA		11	6.900	0.484	7.02	5.400	8.400
069	PROS. SPEC-ANTIG-PSA		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}					
2023-813	29	VITROS		44	2.764	0.119	4.31	2.400	3.100
2023-813	36	TOSOH AIA		10	3.280	0.183	5.59	2.700	3.800
2023-813	38	ROCHE E		49	4.963	0.284	5.72	4.100	5.800
2023-813	44	BECK.COUL.		53	4.813	0.310	6.45	3.900	5.700
2023-813	50	ARC		22	3.886	0.247	6.36	3.100	4.600
2023-813	57	S. EXL 200		66	4.877	0.310	6.36	3.900	5.800
2023-813	58	ATELLICA		11	4.245	0.239	5.62	3.500	5.000
2023-814	29	VITROS		44	15.726	0.805	5.12	13.300	18.100
2023-814	36	TOSOH AIA		10	19.060	1.183	6.21	15.500	22.600
2023-814	38	ROCHE E		49	30.010	1.519	5.06	25.500	34.600
2023-814	44	BECK.COUL.		53	29.760	1.745	5.86	24.500	35.000
2023-814	50	ARC		22	24.809	1.437	5.79	20.500	29.100
2023-814	57	S. EXL 200		66	31.030	1.696	5.46	25.900	36.100
2023-814	58	ATELLICA		11	25.318	1.616	6.38	20.500	30.200
2023-815	29	VITROS		44	8.198	0.369	4.51	7.100	9.300
2023-815	36	TOSOH AIA		10	10.690	0.716	6.70	8.500	12.800
2023-815	38	ROCHE E		49	15.868	0.803	5.06	13.500	18.300
2023-815	44	BECK.COUL.		53	15.557	0.986	6.34	12.600	18.500
2023-815	50	ARC		22	12.673	0.650	5.13	10.700	14.600
2023-815	57	S. EXL 200		66	16.163	0.882	5.46	13.500	18.800
2023-815	58	ATELLICA		11	13.727	1.033	7.53	10.600	16.800

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
071 VITAMIN B 12		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}					
2023-811	29 VITROS			27	795.222	28.146	3.54	711.000	880.000		
2023-811	38 ROCHE E			29	832.138	35.475	4.26	726.000	939.000		
2023-811	44 BECK.COUL.			27	603.593	52.381	8.68	453.000	754.000		
2023-811	50 ARC			21	766.571	61.877	8.07	581.000	952.000		
2023-811	57 S. EXL 200			21	756.667	36.167	4.78	648.000	865.000		
2023-812	29 VITROS			27	374.185	7.884	2.11	351.000	398.000		
2023-812	38 ROCHE E			29	408.241	45.082	11.04	306.000	510.000		
2023-812	44 BECK.COUL.			27	307.519	28.354	9.22	231.000	384.000		
2023-812	50 ARC			21	374.429	27.384	7.31	292.000	457.000		
2023-812	57 S. EXL 200			21	398.143	20.571	5.17	336.000	460.000		
2023-813	29 VITROS			27	292.231	7.423	2.54	270.000	314.000		
2023-813	38 ROCHE E			29	311.759	44.407	14.24	234.000	390.000		
2023-813	44 BECK.COUL.			27	235.778	16.012	6.79	188.000	284.000		
2023-813	50 ARC			21	298.100	15.485	5.19	252.000	345.000		
2023-813	57 S. EXL 200			21	316.619	19.172	6.06	259.000	374.000		
2023-814	29 VITROS			27	832.231	52.096	6.26	676.000	989.000		
2023-814	38 ROCHE E			29	1110.107	69.521	6.26	902.000	1319.000		
2023-814	44 BECK.COUL.			27	784.778	59.562	7.59	606.000	963.000		
2023-814	50 ARC			21	1008.286	88.875	8.81	756.000	1260.000		
2023-814	57 S. EXL 200			21	990.143	35.897	3.63	882.000	1098.000		
2023-815	29 VITROS			27	645.556	30.365	4.70	554.000	737.000		
2023-815	38 ROCHE E			29	651.148	38.052	5.84	537.000	765.000		
2023-815	44 BECK.COUL.			27	497.259	32.418	6.52	400.000	595.000		
2023-815	50 ARC			21	598.571	54.053	9.03	449.000	748.000		
2023-815	57 S. EXL 200			21	609.476	27.605	4.53	527.000	692.000		

Participant Summary
MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY SUB: 60 SPECIAL TESTS GRP: 95 GLYCOHEMOGLOBINE

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD	OR	25	% {LOWER}		
2023-821	12	COBAS/INT.			27	6.248	0.285	4.56	5.400	7.100
2023-821	14	IMMUNOTUR.			15	6.500	0.292	4.49	5.600	7.400
2023-821	22	VITROS			51	6.136	0.123	2.00	5.800	6.500
2023-821	31	IMMUNOTURB			41	6.103	0.245	4.02	5.400	6.800
2023-821	40	SIEMENS DI			85	6.323	0.198	3.14	5.700	6.900
2023-821	46	ARCHITECT			19	5.926	0.102	1.72	5.600	6.200
2023-821	48	ATELLICA			11	5.845	0.123	2.11	5.500	6.200
2023-822	12	COBAS/INT.			27	8.627	0.212	2.46	8.000	9.300
2023-822	14	IMMUNOTUR.			15	9.079	0.421	4.64	7.800	10.300
2023-822	22	VITROS			51	8.574	0.202	2.35	8.000	9.200
2023-822	31	IMMUNOTURB			41	8.748	0.268	3.07	7.900	9.600
2023-822	40	SIEMENS DI			85	8.951	0.221	2.47	8.300	9.600
2023-822	46	ARCHITECT			19	8.342	0.203	2.44	7.700	9.000
2023-822	48	ATELLICA			11	8.200	0.128	1.56	7.800	8.600
2023-823	12	COBAS/INT.			27	5.593	0.228	4.07	4.900	6.300
2023-823	14	IMMUNOTUR.			15	5.764	0.172	2.98	5.200	6.300
2023-823	22	VITROS			51	5.500	0.094	1.71	5.200	5.800
2023-823	31	IMMUNOTURB			41	5.438	0.250	4.59	4.700	6.200
2023-823	40	SIEMENS DI			85	5.677	0.228	4.02	5.000	6.400
2023-823	46	ARCHITECT			19	5.437	0.113	2.09	5.100	5.800
2023-823	48	ATELLICA			11	5.400	0.113	2.09	5.100	5.700
2023-824	12	COBAS/INT.			27	6.285	0.249	3.96	5.500	7.000
2023-824	14	IMMUNOTUR.			15	6.380	0.281	4.40	5.500	7.200
2023-824	22	VITROS			51	6.096	0.109	1.78	5.800	6.400
2023-824	31	IMMUNOTURB			41	6.143	0.182	2.95	5.600	6.700

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE				
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023-824	40	SIEMENS DI		85	6.369	0.194	3.04	5.800	7.000
2023-824	46	ARCHITECT		19	5.942	0.099	1.67	5.600	6.200
2023-824	48	ATELLICA		11	5.873	0.142	2.42	5.400	6.300
146 GLYCOHEMOGLOBIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}					
2023-825	12	COBAS/INT.		27	5.652	0.279	4.94	4.800	6.500
2023-825	14	IMMUNOTUR.		15	5.800	0.285	4.92	4.900	6.700
2023-825	22	VITROS		51	5.492	0.105	1.91	5.200	5.800
2023-825	31	IMMUNOTURB		41	5.461	0.237	4.33	4.800	6.200
2023-825	40	SIEMENS DI		85	5.669	0.229	4.04	5.000	6.400
2023-825	46	ARCHITECT		19	5.447	0.123	2.26	5.100	5.800
2023-825	48	ATELLICA		11	5.382	0.111	2.07	5.000	5.700

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS		GRP: 99 Vitamin D				
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
178 Vitamin D				EVALUATION CRITERIA: TARGET VALUE 20 %					
2023-301	3	VITROS		44	125.405	1.415	1.13	100.000	150.000
2023-301	5	ROCHE E		33	113.303	8.170	7.21	91.000	136.000
2023-301	9	BECK.COUL.		50	110.667	6.018	5.44	89.000	133.000
2023-301	9	BECK.COUL.		50	108.408	8.239	7.60	87.000	130.000
2023-301	11	ARC		18	105.222	5.940	5.65	84.000	126.000
2023-301	25	S. DIME		32	81.438	4.322	5.31	65.000	98.000
2023-301	27	ATELLICA		10	135.800	10.703	7.88	109.000	163.000
2023-302	3	VITROS		44	54.929	3.712	6.76	44.000	66.000
2023-302	5	ROCHE E		33	55.333	5.341	9.65	44.000	66.000
2023-302	9	BECK.COUL.		50	55.755	5.101	9.15	45.000	67.000
2023-302	9	BECK.COUL.		50	37.667	2.055	5.46	30.000	45.000
2023-302	11	ARC		18	53.833	2.949	5.48	43.000	65.000

Participant Summary

MONTH 10 OF YEAR 2023

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 99 Vitamin D					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-302	25	S. DIME		32	42.438	2.839	6.69	34.000	51.000
2023-302	27	APELLICA		10	105.800	4.238	4.01	85.000	127.000
178	Vitamin D		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %				
2023-303	3	VITROS		44	41.595	3.928	9.44	33.000	50.000
2023-303	5	ROCHE E		33	43.848	4.335	9.89	35.000	53.000
2023-303	9	BECK.COUL.		50	30.000	1.633	5.44	24.000	36.000
2023-303	9	BECK.COUL.		50	47.542	4.138	8.70	38.000	57.000
2023-303	11	ARC		18	45.722	2.305	5.04	37.000	55.000
2023-303	25	S. DIME		32	34.438	2.318	6.73	28.000	41.000
2023-303	27	APELLICA		10	94.400	4.477	4.74	76.000	113.000
2023-304	3	VITROS		44	19.048	2.329	12.23	15.000	23.000
2023-304	5	ROCHE E		33	9.061	1.999	22.06	7.000	11.000
2023-304	9	BECK.COUL.		50	21.000	0.816	3.89	17.000	25.000
2023-304	9	BECK.COUL.		50	16.375	1.740	10.62	13.000	20.000
2023-304	11	ARC		18	15.500	0.957	6.18	12.000	19.000
2023-304	25	S. DIME		32	18.188	1.261	6.93	15.000	22.000
2023-304	27	APELLICA		10	35.200	2.993	8.50	28.000	42.000
2023-305	3	VITROS		44	48.000	3.952	8.23	38.000	58.000
2023-305	5	ROCHE E		33	32.576	3.877	11.90	26.000	39.000
2023-305	9	BECK.COUL.		50	43.000	1.414	3.29	34.000	52.000
2023-305	9	BECK.COUL.		50	40.854	2.843	6.96	33.000	49.000
2023-305	11	ARC		18	38.500	1.572	4.08	31.000	46.000
2023-305	25	S. DIME		32	34.844	2.464	7.07	28.000	42.000
2023-305	27	APELLICA		10	75.900	5.890	7.76	61.000	91.000