

**Participant Summary**  
**MONTH 3 OF YEAR 2022**

SPEC: 20 HEMATOLOGY      SUB: 75 BODY FLUID      GRP: 145 BODY FLUID CELL COU.

ANALYTE		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}						
	2022-171			46	0.477	0.690	144.67	0	3.000
	2022-172			46	1002.400	89.471	8.93	734.000	1271.000
	2022-173			46	0.488	0.544	111.47	0	2.000
	2022-174			46	0.614	0.681	111.05	0	3.000
	2022-175			46	1009.089	97.291	9.64	717.000	1301.000
160	BF WHITE CELL COUNT								
	2022-171			45	4.295	1.816	42.29	0	10.000
	2022-172			45	311.581	29.611	9.50	223.000	400.000
	2022-173			45	4.500	1.672	37.15	0	10.000
	2022-174			45	4.455	1.685	37.82	0	10.000
	2022-175			45	310.953	27.737	8.92	228.000	394.000

SPE: 30 IMMUNOHEMATOLOGY      SUB: 80 ABO-RH      GRP: 150 ABO-RH

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS	# LABS ANSW	# "SF" LABS	Score
083 ABO GROUP											
2022-181				008 008	008 008		008 - O	83	83	83	100
2022-182				001 003	001 003		001 - A	83	81	83	100

**Participant Summary**

**MONTH 3 OF YEAR 2022**

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-182				001 003	001 003		002 - A1	83	2	83	100		
	083 ABO GROUP												
2022-183				004 004	004 004		004 - B	83	83	83	100		
2022-184				008 008	008 008		008 - O	83	83	83	100		
2022-185				005 007	005 007		006 - A1B	83	1	83	100		
2022-185				005 007	005 007		005 - AB	83	82	83	100		
	126 RH TYPE												
2022-181				001 003	001 003		001 - Rh+	83	83	83	100		
2022-182				001 003	001 003		001 - Rh+	83	83	83	100		
2022-183				003 004	003 004		003 - Rh-Du np	83	11	83	100		
2022-183				003 004	003 004		004 - Rh-	83	72	83	100		
2022-184				003 004	003 004		003 - Rh-Du np	83	11	83	100		
2022-184				003 004	003 004		004 - Rh-	83	72	83	100		
2022-185				001 003	001 003		001 - Rh+	83	82	83	100		
2022-185				001 003	001 003		002 - Rh+ Du va.	83	1	83	100		

**Participant Summary**

**MONTH 3 OF YEAR 2022**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2022-191				002 002	002 002		002 - POSITIVE	65	65	65	100
2022-192				001 001	001 001		001 - NEGATIVE	65	64	64	100
2022-192				001 001	001 001		002 - POSITIVE	65	1	64	0
2022-193				002 002	002 002		002 - POSITIVE	65	65	65	100
2022-194				001 001	001 001		001 - NEGATIVE	65	65	65	100
2022-195				001 001	001 001		001 - NEGATIVE	65	65	65	100
SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2022-271	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	13	13	13	100
2022-271	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	32	32	32	100
2022-271	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	10	10	10	100
2022-271	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	16	16	16	100
2022-272	001 Poly AHG	001 Gel			002 004		003 - 3+	13	9	13	100
2022-272	001 Poly AHG	001 Gel			002 004		002 - 2+	13	1	13	100
2022-272	001 Poly AHG	001 Gel			002 004		004 - 4+	13	3	13	100
2022-272	001 Poly AHG	002 Tubo			002 004		004 - 4+	32	1	32	100
2022-272	001 Poly AHG	002 Tubo			002 004		002 - 2+	32	23	32	100

**Participant Summary**

**MONTH 3 OF YEAR 2022**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-272	001 Poly AHG	002 Tubo			002 004		003 - 3+	32	8	32	100	
2022-272	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	15	15	15	100	
179 DIRECT ANTIG(COOMBS)												
2022-273	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	13	13	13	100	
2022-273	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	32	32	32	100	
2022-273	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	10	10	10	100	
2022-273	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	16	16	16	100	
2022-274	001 Poly AHG	001 Gel			002 004		003 - 3+	13	10	13	100	
2022-274	001 Poly AHG	001 Gel			002 004		002 - 2+	13	1	13	100	
2022-274	001 Poly AHG	001 Gel			002 004		004 - 4+	13	2	13	100	
2022-274	001 Poly AHG	002 Tubo			002 004		003 - 3+	32	9	32	100	
2022-274	001 Poly AHG	002 Tubo			002 004		002 - 2+	32	21	32	100	
2022-274	001 Poly AHG	002 Tubo			002 004		004 - 4+	32	2	32	100	
2022-274	002 MnoAnt-IgG	002 Tubo			002 004		003 - 3+	15	3	15	100	
2022-274	002 MnoAnt-IgG	002 Tubo			002 004		004 - 4+	15	1	15	100	
2022-274	002 MnoAnt-IgG	002 Tubo			002 004		002 - 2+	15	11	15	100	
2022-275	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	13	13	13	100	
2022-275	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	32	32	32	100	
2022-275	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	10	10	10	100	
2022-275	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	16	16	16	100	

**Participant Summary**

**MONTH 3 OF YEAR 2022**

SPE: 30 IMMUNOHEMATOLOGY												SUB: 100 COMPATIBILITY TESTING												GRP: 170 COMPATIBILITY TEST.											
TEST	LEVEL 1		LEVEL 2		LEVEL 3		PART		REFEREE		COMMER		REPORTED		PART		# LABS		# "SF"		Score														
SAMPLE	RESULT CODE		RESULT CODE		RESULT CODE		LOW/HIGH		LOW/HIGH		LOW/HIGH		VALUE		LABS		ANSW		LABS																
085 COMPATIBIL. TESTING																																			
2022-201							001	001	001	001					001	-	COMPATIBLE	40	40	40	100														
2022-202							002	002	002	002					002	-	INCOMPATIB	40	40	40	100														
2022-203							001	001	001	001					001	-	COMPATIBLE	40	40	40	100														
2022-204							001	001	001	001					001	-	COMPATIBLE	40	40	40	100														
2022-205							001	001	001	001					001	-	COMPATIBLE	40	40	40	100														