

Participant Summary
MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
080 FIBRINOGEN				CLIA TEST ID #: 825						EVALUATION CRITERIA: TARGET VALUE 20 %	
2023-161	73 ACL 8 TO E	87 HEMOSIL-C		10	640.200	74.310	11.61	512.000	768.000		
2023-161	79 CS-2500	67 DADE		14	483.571	29.193	6.04	387.000	580.000		
2023-162	73 ACL 8 TO E	87 HEMOSIL-C		10	343.800	23.464	6.82	275.000	413.000		
2023-162	79 CS-2500	67 DADE		14	261.429	11.350	4.34	209.000	314.000		
2023-163	73 ACL 8 TO E	87 HEMOSIL-C		10	338.500	23.762	7.02	271.000	406.000		
2023-163	79 CS-2500	67 DADE		14	256.429	11.494	4.48	205.000	308.000		
2023-164	73 ACL 8 TO E	87 HEMOSIL-C		10	176.100	11.282	6.41	141.000	211.000		
2023-164	79 CS-2500	67 DADE		14	141.857	10.763	7.59	113.000	170.000		
2025-165	73 ACL 8 TO E	87 HEMOSIL-C		10	285.700	22.799	7.98	229.000	343.000		
2025-165	79 CS-2500	67 DADE		14	225.929	10.680	4.73	181.000	271.000		
081 PART.THROM. TIME-PTT				CLIA TEST ID #: 835						EVALUATION CRITERIA: TARGET VALUE 15 %	
2023-161	62 DADE SYS.	35 DADE ACTIN		56	25.978	0.862	3.32	22.100	29.900		
2023-161	65 DADE BFTII	35 DADE ACTIN		11	26.345	1.873	7.11	22.400	30.300		
2023-161	69 DIAG.STAGO	49 STA-PTT		17	28.759	0.618	2.15	24.400	33.100		
2023-161	73 ACL 8 TO E	56 PTT-HEMO		35	38.300	1.373	3.58	32.600	44.000		
2023-161	76 STAGO STAR	49 STA-PTT		23	28.191	1.234	4.38	24.000	32.400		
2023-161	79 CS-2500	35 DADE ACTIN		13	26.346	0.625	2.37	22.400	30.300		
2023-162	62 DADE SYS.	35 DADE ACTIN		56	37.089	1.668	4.50	31.500	42.700		
2023-162	65 DADE BFTII	35 DADE ACTIN		11	39.345	3.131	7.96	33.400	45.200		
2023-162	69 DIAG.STAGO	49 STA-PTT		17	41.635	1.257	3.02	35.400	47.900		
2023-162	73 ACL 8 TO E	56 PTT-HEMO		35	40.643	0.786	1.93	34.500	46.700		
2023-162	76 STAGO STAR	49 STA-PTT		23	41.035	2.539	6.19	34.900	47.200		
2023-162	79 CS-2500	35 DADE ACTIN		13	38.485	1.490	3.87	32.700	44.300		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %						
2023-163	62 DADE SYS.	35 DADE ACTIN		56	44.538	4.319	9.70	37.900	51.200
2023-163	65 DADE BFTII	35 DADE ACTIN		11	49.391	5.670	11.48	42.000	56.800
2023-163	69 DIAG.STAGO	49 STA-PTT		17	51.694	2.118	4.10	43.900	59.400
2023-163	73 ACL 8 TO E	56 PTT-HEMO		35	48.597	0.742	1.53	41.300	55.900
2023-163	76 STAGO STAR	49 STA-PTT		23	50.435	3.269	6.48	42.900	58.000
2023-163	79 CS-2500	35 DADE ACTIN		13	47.723	2.092	4.38	40.600	54.900
2023-164	62 DADE SYS.	35 DADE ACTIN		56	30.956	1.393	4.50	26.300	35.600
2023-164	65 DADE BFTII	35 DADE ACTIN		11	31.073	2.766	8.90	26.400	35.700
2023-164	69 DIAG.STAGO	49 STA-PTT		17	36.259	1.451	4.00	30.800	41.700
2023-164	73 ACL 8 TO E	56 PTT-HEMO		35	31.314	0.667	2.13	26.600	36.000
2023-164	76 STAGO STAR	49 STA-PTT		23	35.678	2.584	7.24	30.300	41.000
2023-164	79 CS-2500	35 DADE ACTIN		13	32.085	1.162	3.62	27.300	36.900
2025-165	62 DADE SYS.	35 DADE ACTIN		56	25.918	0.798	3.08	22.000	29.800
2025-165	65 DADE BFTII	35 DADE ACTIN		11	25.127	2.747	10.93	21.400	28.900
2025-165	69 DIAG.STAGO	49 STA-PTT		17	29.259	0.815	2.79	24.900	33.600
2025-165	73 ACL 8 TO E	56 PTT-HEMO		35	30.166	0.536	1.78	25.600	34.700
2025-165	76 STAGO STAR	49 STA-PTT		23	28.474	1.790	6.29	24.200	32.700
2025-165	79 CS-2500	35 DADE ACTIN		13	26.892	0.537	2.00	22.900	30.900
082	PROTHROMBIN TIME (PT)	CLIA TEST ID #: 845							
2023-161	62 DADE SYS.	33 DADE INNOV		57	10.275	0.366	3.56	8.700	11.800
2023-161	65 DADE BFTII	33 DADE INNOV		12	9.575	0.469	4.90	8.100	11.000
2023-161	73 ACL 8 TO E	41 HEMOSIL RE		30	10.255	0.311	3.04	8.700	11.800
2023-161	76 STAGO STAR	44 STA NEO CI		14	12.257	0.329	2.68	10.400	14.100
2023-161	79 CS-2500	33 DADE INNOV		14	10.571	0.294	2.78	9.000	12.200
2023-162	62 DADE SYS.	33 DADE INNOV		57	17.355	0.698	4.02	14.800	20.000
2023-162	65 DADE BFTII	33 DADE INNOV		12	18.692	1.127	6.03	15.900	21.500

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-162	73	ACL 8 TO E	41 HEMOSIL RE	30	20.563	0.623	3.03	17.500	23.600	
2023-162	76	STAGO STAR	44 STA NEO CI	14	21.371	0.757	3.54	18.200	24.600	
2023-162	79	CS-2500	33 DADE INNOV	14	17.450	0.275	1.57	14.800	20.100	
082 PROTHROMBIN TIME(PT)		CLIA TEST ID #: 845		EVALUATION CRITERIA: TARGET VALUE 15 %						
2023-163	62	DADE SYS.	33 DADE INNOV	57	26.349	1.361	5.17	22.400	30.300	
2023-163	65	DADE BFTII	33 DADE INNOV	12	29.667	1.148	3.87	25.200	34.100	
2023-163	73	ACL 8 TO E	41 HEMOSIL RE	30	33.193	1.163	3.50	28.200	38.200	
2023-163	76	STAGO STAR	44 STA NEO CI	14	33.607	2.472	7.36	28.600	38.600	
2023-163	79	CS-2500	33 DADE INNOV	14	25.979	0.516	1.99	22.100	29.900	
2023-164	62	DADE SYS.	33 DADE INNOV	57	12.668	0.437	3.45	10.800	14.600	
2023-164	65	DADE BFTII	33 DADE INNOV	12	12.408	0.442	3.57	10.500	14.300	
2023-164	73	ACL 8 TO E	41 HEMOSIL RE	30	14.521	0.270	1.86	12.300	16.700	
2023-164	76	STAGO STAR	44 STA NEO CI	14	15.379	0.521	3.39	13.100	17.700	
2023-164	79	CS-2500	33 DADE INNOV	14	13.114	0.173	1.32	11.100	15.100	
2025-165	62	DADE SYS.	33 DADE INNOV	57	10.223	0.316	3.09	8.700	11.800	
2025-165	65	DADE BFTII	33 DADE INNOV	12	9.683	0.593	6.12	8.200	11.100	
2025-165	73	ACL 8 TO E	41 HEMOSIL RE	30	11.037	0.287	2.60	9.400	12.700	
2025-165	76	STAGO STAR	44 STA NEO CI	14	12.550	0.672	5.35	10.700	14.400	
2025-165	79	CS-2500	33 DADE INNOV	14	10.707	0.128	1.20	9.100	12.300	

Participant Summary

MONTH 3 OF YEAR 2023

SPE: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 100 HEMA. CELL IDENTIFI.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2023-111					014 014		014 - LYMPHOCYTE	897	894	894	100
2023-111					014 014		017 - MONOCYTE	897	2	894	0
2023-111					014 014		025 - DOHLE INCL	897	1	894	0
2023-112					025 025		023 - ALDER ANOM	898	1	888	0
2023-112					025 025		025 - DOHLE INCL	898	888	888	100
2023-112					025 025		026 - LEUKOCYTE	898	1	888	0
2023-112					025 025		029 - TOXIC GRAN	898	2	888	0
2023-112					025 025		038 - HOWELL-JOL	898	1	888	0
2023-112					025 025		057 - CHEDIAK -	898	2	888	0
2023-112					025 025		074 - BLUE-GREEN	898	3	888	0
2023-113					054 054		054 - TARGET CEL	897	897	897	100
2023-114					033 033		033 - POLYCHROMA	897	871	871	100
2023-114					033 033		034 - PRORUBRICY	897	1	871	0
2023-114					033 033		035 - RUBRICYTE	897	8	871	0
2023-114					033 033		036 - BASOPHILIC	897	8	871	0
2023-114					033 033		073 - RETICULOCY	897	9	871	0
2023-115					002 003		001 - BLAST	897	57	836	0
2023-115					002 003		002 - PROMYELOCY	897	835	836	100
2023-115					002 003		003 - MYELOCYTE	897	1	836	100
2023-115					002 003		069 - Prolymphoc	897	4	836	0

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}						
2023-141	7	M.B.MANUAL		14	1.729	0.562	32.54	0.000	3.400
2023-141	9	M.B. RICCA		10	1.530	0.542	35.45	0	3.200
2023-141	24	Unicel DxH		19	0.768	0.145	18.91	0.300	1.200
2023-141	25	SYSMEX XN		28	0.889	0.094	10.56	0.600	1.200
2023-142	7	M.B.MANUAL		14	1.650	0.543	32.94	0.000	3.300
2023-142	9	M.B. RICCA		10	1.950	0.625	32.05	0.100	3.800
2023-142	24	Unicel DxH		19	0.779	0.132	16.96	0.400	1.200
2023-142	25	SYSMEX XN		28	0.886	0.087	9.88	0.600	1.100
2023-143	7	M.B.MANUAL		14	11.771	3.395	28.84	1.600	22.000
2023-143	9	M.B. RICCA		10	10.770	1.546	14.35	6.100	15.400
2023-143	24	Unicel DxH		19	7.832	0.806	10.29	5.400	10.200
2023-143	25	SYSMEX XN		28	5.371	1.264	23.53	1.600	9.200
2023-144	7	M.B.MANUAL		14	5.979	1.696	28.37	0.900	11.100
2023-144	9	M.B. RICCA		10	4.970	1.166	23.47	1.500	8.500
2023-144	24	Unicel DxH		19	3.674	0.448	12.20	2.300	5.000
2023-144	25	SYSMEX XN		28	2.564	0.539	21.03	0.900	4.200
2023-145	7	M.B.MANUAL		14	11.914	2.636	22.13	4.000	19.800
2023-145	9	M.B. RICCA		10	10.980	2.206	20.09	4.400	17.600
2023-145	24	Unicel DxH		19	8.061	0.424	5.27	6.000	10.100
2023-145	25	SYSMEX XN		28	5.296	1.276	24.09	1.500	9.100

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
153	SED. RATE (ESR)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 35 %	{GREATER}				
2023-151	4	SEDIPLAST		602	7.912	2.190	27.68	1.000	14.000	
2023-151	10	SEDI RATE		96	9.674	1.986	20.54	4.000	16.000	
2023-151	11	SEDIMAT-15		53	11.635	8.691	74.70	0	38.000	
2023-151	14	GLOBAL SC.		10	10.800	1.661	15.38	6.000	16.000	
2023-151	19	MINI CUBE		31	16.367	1.871	11.43	11.000	22.000	
2023-152	4	SEDIPLAST		602	48.224	8.316	17.24	23.000	73.000	
2023-152	10	SEDI RATE		96	47.883	7.300	15.25	26.000	70.000	
2023-152	11	SEDIMAT-15		53	128.577	22.316	17.36	62.000	196.000	
2023-152	14	GLOBAL SC.		10	52.800	11.505	21.79	18.000	87.000	
2023-152	19	MINI CUBE		31	77.935	20.199	25.92	17.000	139.000	
2023-153	4	SEDIPLAST		602	8.392	2.240	26.69	2.000	15.000	
2023-153	10	SEDI RATE		96	9.819	1.523	15.51	5.000	14.000	
2023-153	11	SEDIMAT-15		53	12.755	8.637	67.72	0	39.000	
2023-153	14	GLOBAL SC.		10	11.100	3.113	28.04	2.000	20.000	
2023-153	19	MINI CUBE		31	16.452	2.367	14.39	9.000	24.000	
2023-154	4	SEDIPLAST		602	8.228	2.226	27.05	2.000	15.000	
2023-154	10	SEDI RATE		96	9.553	1.568	16.42	5.000	14.000	
2023-154	11	SEDIMAT-15		53	11.887	7.752	65.22	0	35.000	
2023-154	14	GLOBAL SC.		10	10.700	1.735	16.21	5.000	16.000	
2023-154	19	MINI CUBE		31	16.367	1.888	11.54	11.000	22.000	
2023-155	4	SEDIPLAST		602	47.955	8.118	16.93	24.000	72.000	
2023-155	10	SEDI RATE		96	46.774	6.560	14.03	27.000	66.000	
2023-155	11	SEDIMAT-15		53	127.176	21.041	16.54	64.000	190.000	
2023-155	14	GLOBAL SC.		10	49.200	9.347	19.00	21.000	77.000	
2023-155	19	MINI CUBE		31	80.226	19.633	24.47	21.000	139.000	

Participant Summary
MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2023131X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	148	50.691	1.176	2.32	38.000	63.400				
2023131X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	148	26.654	1.029	3.86	20.000	33.300				
2023131X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	148	1.116	0.213	19.12	0	3.100				
2023131X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	148	13.201	0.520	3.94	9.900	16.500				
2023131X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	148	8.351	0.766	9.17	6.300	10.400				
2023131X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	123	61.326	1.596	2.60	46.000	76.700				
2023131X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	29	59.517	1.502	2.52	44.600	74.400				
2023131X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	123	33.033	1.469	4.45	24.800	41.300				
2023131X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	29	31.048	1.557	5.02	23.300	38.800				
2023131X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	123	1.727	0.354	20.48	0	3.700				
2023131X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	29	1.962	0.457	23.28	0.000	4.000				
2023131X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	123	3.934	0.345	8.78	1.900	5.900				
2023131X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	29	4.064	0.320	7.87	2.100	6.100				
2023131X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	123	67.981	1.402	2.06	51.000	85.000				
2023131X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	29	3.396	0.074	2.19	1.400	5.400				
2023131X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	17	64.335	0.718	1.12	48.300	80.400				
2023131X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	17	21.671	0.897	4.14	16.300	27.100				
2023131X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	17	8.735	0.363	4.15	6.600	10.900				
2023131X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	17	2.153	0.261	12.15	0.200	4.200				
2023131X	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	17	1.665	0.452	27.17	0	3.700				
2023131X	3 ADVIA 120	6 LUC	9 ADVIA-120	17	2.582	0.333	12.89	0.600	4.600				
2023131X	4 AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	69	74.287	0.768	1.03	55.700	92.900				
2023131X	4 AB- CELL-D	1 NEUTROPHIL	37 ABB.CEL RU	46	73.860	0.620	0.84	55.400	92.300				
2023131X	4 AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	69	16.745	1.246	7.44	12.600	20.900				
2023131X	4 AB- CELL-D	2 LYMPHOCYTE	37 ABB.CEL RU	46	16.538	0.786	4.75	12.400	20.700				
2023131X	4 AB- CELL-D	3 MONOCYTE	8 CELL-DYN	69	4.903	0.610	12.44	2.900	6.900				
2023131X	4 AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	46	5.440	0.420	7.73	3.400	7.400				
2023131X	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	69	3.806	0.181	4.77	1.800	5.800				

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023131X	4	AB- CELL-D	4	EOSINOPHIL	37	ABB.CEL RU	46	3.898	0.252	6.45	1.900	5.900
2023131X	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	69	0.340	0.205	60.40	0	2.300
2023131X	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	46	0.226	0.126	55.67	0	2.200
2023131X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	65.307	0.947	1.45	49.000	81.600
2023131X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	21.569	0.670	3.11	16.200	27.000
2023131X	5	BECTON D	3	MONOCYTE	29	B.C600,800	60	8.646	0.855	9.89	6.500	10.800
2023131X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	60	4.473	0.287	6.43	2.500	6.500
2023131X	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.012	0.032	275.16	0	2.000
2023131X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	52	63.882	1.274	1.99	47.900	79.900
2023131X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	19.998	1.033	5.17	15.000	25.000
2023131X	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	7.514	0.841	11.19	5.500	9.500
2023131X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	52	8.745	0.477	5.45	6.600	10.900
2023131X	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	31.184	0.163	0.52	23.400	39.000
2023131X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	53	68.779	1.812	2.63	51.600	86.000
2023131X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	53	26.402	1.738	6.58	19.800	33.000
2023131X	7	DXH	3	MONOCYTE	36	BC DXH 500	53	0.441	0.399	90.46	0	2.400
2023131X	7	DXH	4	EOSINOPHIL	36	BC DXH 500	53	4.085	0.492	12.05	2.100	6.100
2023131X	7	DXH	5	BASOPHIL	36	BC DXH 500	53	0.199	0.289	145.14	0	2.200
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2023132X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	148	48.834	1.131	2.32	36.600	61.000
2023132X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	148	29.408	1.122	3.81	22.100	36.800
2023132X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	148	1.520	0.282	18.52	0	3.500
2023132X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	148	12.596	0.658	5.22	9.400	15.700
2023132X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	148	7.649	0.696	9.10	5.600	9.600
2023132X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	123	61.125	1.990	3.26	45.800	76.400
2023132X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	29	59.055	2.254	3.82	44.300	73.800
2023132X	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	123	30.554	1.672	5.47	22.900	38.200
2023132X	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	29	28.441	1.802	6.34	21.300	35.600
2023132X	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	123	1.811	0.545	30.10	0	3.800
2023132X	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	29	2.321	0.538	23.20	0.300	4.300
2023132X	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	123	6.514	0.628	9.63	4.500	8.500

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2023132X	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	29	6.704	0.940	14.02	4.700	8.700
2023132X	2	ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	123	66.028	1.195	1.81	49.500	82.500
2023132X	2	ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	29	3.282	0.085	2.58	1.300	5.300
2023132X	3	ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	17	46.488	1.038	2.23	34.900	58.100
2023132X	3	ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	17	31.729	1.381	4.35	23.800	39.700
2023132X	3	ADVIA 120	3	MONOCYTE	9	ADVIA-120	17	15.688	0.927	5.91	11.800	19.600
2023132X	3	ADVIA 120	4	EOSINOPHIL	9	ADVIA-120	17	1.571	0.375	23.90	0	3.600
2023132X	3	ADVIA 120	5	BASOPHIL	9	ADVIA-120	17	2.124	0.641	30.20	0.100	4.100
2023132X	3	ADVIA 120	6	LUC	9	ADVIA-120	17	3.676	0.625	17.01	1.700	5.700
2023132X	4	AB- CELL-D	1	NEUTROPHIL	8	CELL-DYN	69	67.894	0.868	1.28	50.900	84.900
2023132X	4	AB- CELL-D	1	NEUTROPHIL	37	ABB.CEL RU	45	66.044	1.078	1.63	49.500	82.600
2023132X	4	AB- CELL-D	2	LYMPHOCYTE	8	CELL-DYN	69	22.470	1.457	6.48	16.900	28.100
2023132X	4	AB- CELL-D	2	LYMPHOCYTE	37	ABB.CEL RU	46	24.302	1.984	8.16	18.200	30.400
2023132X	4	AB- CELL-D	3	MONOCYTE	8	CELL-DYN	69	4.984	0.783	15.71	3.000	7.000
2023132X	4	AB- CELL-D	3	MONOCYTE	37	ABB.CEL RU	46	5.533	1.442	26.07	3.500	7.500
2023132X	4	AB- CELL-D	4	EOSINOPHIL	8	CELL-DYN	69	4.162	0.231	5.55	2.200	6.200
2023132X	4	AB- CELL-D	4	EOSINOPHIL	37	ABB.CEL RU	46	4.072	0.245	6.01	2.100	6.100
2023132X	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	69	0.415	0.273	65.81	0	2.400
2023132X	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	46	0.189	0.413	218.77	0	2.200
2023132X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	60.785	1.028	1.69	45.600	76.000
2023132X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	25.167	1.507	5.99	18.900	31.500
2023132X	5	BECTON D	3	MONOCYTE	29	B.C600,800	59	8.846	0.791	8.94	6.600	11.100
2023132X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	60	5.217	0.260	4.98	3.200	7.200
2023132X	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023132X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	52	55.241	1.403	2.54	41.400	69.100
2023132X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	30.125	1.632	5.42	22.600	37.700
2023132X	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	8.085	1.047	12.95	6.100	10.100
2023132X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	52	6.475	0.413	6.39	4.500	8.500
2023132X	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	28.537	0.309	1.08	21.400	35.700
2023132X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	53	48.860	3.963	8.11	36.600	61.100
2023132X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	53	38.377	4.473	11.66	28.800	48.000

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023132X	7 DXH	3 MONOCYTE	36 BC DXH 500	53	0.219	0.114	52.19	0	2.200		
2023132X	7 DXH	4 EOSINOPHIL	36 BC DXH 500	53	12.400	1.184	9.54	9.300	15.500		
2023132X	7 DXH	5 BASOPHIL	36 BC DXH 500	53	0.050	0.075	150.12	0	2.000		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2023133X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	148	59.594	1.472	2.47	44.700	74.500		
2023133X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	148	14.112	1.361	9.64	10.600	17.600		
2023133X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	148	1.015	0.251	24.77	0	3.000		
2023133X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	147	15.997	0.318	1.99	12.000	20.000		
2023133X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	149	9.315	0.925	9.93	7.000	11.600		
2023133X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	123	61.629	1.692	2.75	46.200	77.000		
2023133X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	29	59.621	1.679	2.82	44.700	74.500		
2023133X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	123	32.739	1.603	4.90	24.600	40.900		
2023133X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	29	30.848	1.404	4.55	23.100	38.600		
2023133X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	123	1.684	0.388	23.06	0	3.700		
2023133X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	29	2.066	0.463	22.40	0.100	4.100		
2023133X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	122	3.947	0.368	9.32	1.900	5.900		
2023133X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	29	4.139	0.303	7.31	2.100	6.100		
2023133X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	124	67.704	1.554	2.29	50.800	84.600		
2023133X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	29	3.393	0.054	1.59	1.400	5.400		
2023133X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	17	64.581	0.679	1.05	48.400	80.700		
2023133X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	17	21.613	1.021	4.72	16.200	27.000		
2023133X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	17	8.694	0.534	6.14	6.500	10.900		
2023133X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	17	2.135	0.299	14.00	0.100	4.100		
2023133X	3 ADVIA 120	5 BASOPHIL	9 ADVIA-120	17	1.618	0.340	21.02	0	3.600		
2023133X	3 ADVIA 120	6 LUC	9 ADVIA-120	17	2.471	0.542	21.94	0.500	4.500		
2023133X	4 AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	69	50.234	0.931	1.85	37.700	62.800		
2023133X	4 AB- CELL-D	1 NEUTROPHIL	37 ABB.CEL RU	46	49.230	1.230	2.50	36.900	61.500		
2023133X	4 AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	69	39.354	1.692	4.30	29.500	49.200		
2023133X	4 AB- CELL-D	2 LYMPHOCYTE	37 ABB.CEL RU	46	40.002	2.109	5.27	30.000	50.000		
2023133X	4 AB- CELL-D	3 MONOCYTE	8 CELL-DYN	69	5.912	1.323	22.38	3.900	7.900		
2023133X	4 AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	46	6.936	1.624	23.41	4.900	8.900		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2023133X	4	AB- CELL-D	4	EOSINOPHIL	8	CELL-DYN	69	3.672	0.346	9.42	1.700	5.700
2023133X	4	AB- CELL-D	4	EOSINOPHIL	37	ABB.CEL RU	46	3.676	0.408	11.10	1.700	5.700
2023133X	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	69	0.761	0.517	67.91	0	2.800
2023133X	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	46	0.166	0.155	93.50	0	2.200
2023133X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	58.428	1.353	2.32	43.800	73.000
2023133X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	25.710	2.466	9.59	19.300	32.100
2023133X	5	BECTON D	3	MONOCYTE	29	B.C600,800	60	7.817	1.237	15.82	5.800	9.800
2023133X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	60	8.045	0.378	4.70	6.000	10.100
2023133X	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.000	0.000	0.00	0	2.000
2023133X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	52	47.684	1.358	2.85	35.800	59.600
2023133X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	40.390	1.501	3.72	30.300	50.500
2023133X	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	7.190	0.947	13.17	5.200	9.200
2023133X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	52	4.751	0.914	19.24	2.800	6.800
2023133X	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	26.994	0.447	1.66	20.200	33.700
2023133X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	53	69.333	1.506	2.17	52.000	86.700
2023133X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	53	26.036	1.640	6.30	19.500	32.500
2023133X	7	DXH	3	MONOCYTE	36	BC DXH 500	53	0.296	0.305	102.96	0	2.300
2023133X	7	DXH	4	EOSINOPHIL	36	BC DXH 500	53	4.178	0.499	11.95	2.200	6.200
2023133X	7	DXH	5	BASOPHIL	36	BC DXH 500	53	0.144	0.226	156.47	0	2.100
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2023134X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	148	50.648	1.300	2.57	38.000	63.300
2023134X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	148	26.562	1.103	4.15	19.900	33.200
2023134X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	147	1.134	0.214	18.89	0	3.100
2023134X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	147	13.269	0.503	3.79	10.000	16.600
2023134X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	147	8.419	0.745	8.85	6.300	10.500
2023134X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	123	74.883	1.572	2.10	56.200	93.600
2023134X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	29	71.641	1.404	1.96	53.700	89.600
2023134X	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	123	18.145	1.368	7.54	13.600	22.700
2023134X	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	29	16.552	1.248	7.54	12.400	20.700
2023134X	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	124	2.311	0.471	20.37	0.300	4.300
2023134X	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	29	2.797	0.604	21.61	0.800	4.800

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2023134X	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	122	4.637	0.346	7.47	2.600	6.600
2023134X	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	29	4.817	0.336	6.98	2.800	6.800
2023134X	2	ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	123	84.153	1.192	1.42	63.100	105.200
2023134X	2	ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	29	4.182	0.076	1.81	2.200	6.200
2023134X	3	ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	17	74.212	0.479	0.65	55.700	92.800
2023134X	3	ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	17	12.535	0.744	5.94	9.400	15.700
2023134X	3	ADVIA 120	3	MONOCYTE	9	ADVIA-120	17	8.476	0.319	3.76	6.400	10.600
2023134X	3	ADVIA 120	4	EOSINOPHIL	9	ADVIA-120	17	2.506	0.226	9.02	0.500	4.500
2023134X	3	ADVIA 120	5	BASOPHIL	9	ADVIA-120	18	1.389	0.373	26.82	0	3.400
2023134X	3	ADVIA 120	6	LUC	9	ADVIA-120	17	1.888	0.343	18.15	0	3.900
2023134X	4	AB- CELL-D	1	NEUTROPHIL	8	CELL-DYN	71	74.306	0.791	1.06	55.700	92.900
2023134X	4	AB- CELL-D	1	NEUTROPHIL	37	ABB.CEL RU	46	73.887	0.603	0.82	55.400	92.400
2023134X	4	AB- CELL-D	2	LYMPHOCYTE	8	CELL-DYN	68	16.857	1.402	8.31	12.600	21.100
2023134X	4	AB- CELL-D	2	LYMPHOCYTE	37	ABB.CEL RU	45	16.641	0.684	4.11	12.500	20.800
2023134X	4	AB- CELL-D	3	MONOCYTE	8	CELL-DYN	68	4.894	0.523	10.69	2.900	6.900
2023134X	4	AB- CELL-D	3	MONOCYTE	37	ABB.CEL RU	46	5.448	0.372	6.82	3.400	7.400
2023134X	4	AB- CELL-D	4	EOSINOPHIL	8	CELL-DYN	69	3.781	0.210	5.55	1.800	5.800
2023134X	4	AB- CELL-D	4	EOSINOPHIL	37	ABB.CEL RU	46	3.827	0.227	5.94	1.800	5.800
2023134X	4	AB- CELL-D	5	BASOPHIL	8	CELL-DYN	69	0.318	0.175	55.04	0	2.300
2023134X	4	AB- CELL-D	5	BASOPHIL	37	ABB.CEL RU	46	0.187	0.117	62.70	0	2.200
2023134X	5	BECTON D	1	NEUTROPHIL	29	B.C600,800	60	65.508	1.055	1.61	49.100	81.900
2023134X	5	BECTON D	2	LYMPHOCYTE	29	B.C600,800	60	21.445	0.689	3.21	16.100	26.800
2023134X	5	BECTON D	3	MONOCYTE	29	B.C600,800	60	8.652	0.891	10.30	6.500	10.800
2023134X	5	BECTON D	4	EOSINOPHIL	29	B.C600,800	60	4.383	0.252	5.75	2.400	6.400
2023134X	5	BECTON D	5	BASOPHIL	29	B.C600,800	60	0.012	0.032	275.16	0	2.000
2023134X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	52	63.827	1.302	2.04	47.900	79.800
2023134X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	52	19.614	1.035	5.27	14.700	24.500
2023134X	6	MR Mindray	3	MONOCYTE	32	BC-5390	52	7.825	0.848	10.83	5.800	9.800
2023134X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	51	8.816	0.511	5.80	6.600	11.000
2023134X	6	MR Mindray	5	BASOPHIL	32	BC-5390	52	31.135	0.162	0.52	23.400	38.900
2023134X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	53	80.865	0.877	1.08	60.600	101.100

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023134X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	53	10.338	0.876	8.47	7.800	12.900
2023134X	7	DXH	3	MONOCYTE	36	BC DXH 500	53	0.213	0.188	88.14	0	2.200
2023134X	7	DXH	4	EOSINOPHIL	36	BC DXH 500	53	8.344	0.593	7.11	6.300	10.400
2023134X	7	DXH	5	BASOPHIL	36	BC DXH 500	53	0.087	0.181	207.86	0	2.100
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2023135X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	146	59.577	1.405	2.36	44.700	74.500
2023135X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	149	14.083	1.495	10.62	10.600	17.600
2023135X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	147	1.003	0.186	18.54	0	3.000
2023135X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	147	15.989	0.327	2.04	12.000	20.000
2023135X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	147	9.310	0.899	9.66	7.000	11.600
2023135X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	123	61.542	1.902	3.09	46.200	76.900
2023135X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	29	59.496	1.534	2.58	44.600	74.400
2023135X	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	123	30.104	1.714	5.69	22.600	37.600
2023135X	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	29	28.014	1.589	5.67	21.000	35.000
2023135X	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	123	1.815	0.552	30.43	0	3.800
2023135X	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	29	2.352	0.640	27.23	0.400	4.400
2023135X	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	124	6.546	0.977	14.92	4.500	8.500
2023135X	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	29	6.648	0.668	10.04	4.600	8.600
2023135X	2	ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	122	66.089	1.161	1.76	49.600	82.600
2023135X	2	ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	29	3.293	0.156	4.73	1.300	5.300
2023135X	3	ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	17	46.659	1.161	2.49	35.000	58.300
2023135X	3	ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	17	32.000	1.597	4.99	24.000	40.000
2023135X	3	ADVIA 120	3	MONOCYTE	9	ADVIA-120	17	15.524	0.861	5.55	11.600	19.400
2023135X	3	ADVIA 120	4	EOSINOPHIL	9	ADVIA-120	17	1.594	0.306	19.18	0	3.600
2023135X	3	ADVIA 120	5	BASOPHIL	9	ADVIA-120	17	2.041	0.668	32.72	0.000	4.000
2023135X	3	ADVIA 120	6	LUC	9	ADVIA-120	17	3.771	0.736	19.53	1.800	5.800
2023135X	4	AB- CELL-D	1	NEUTROPHIL	8	CELL-DYN	69	50.425	0.994	1.97	37.800	63.000
2023135X	4	AB- CELL-D	1	NEUTROPHIL	37	ABB.CEL RU	46	50.402	1.473	2.92	37.800	63.000
2023135X	4	AB- CELL-D	2	LYMPHOCYTE	8	CELL-DYN	69	38.900	1.700	4.37	29.200	48.600
2023135X	4	AB- CELL-D	2	LYMPHOCYTE	37	ABB.CEL RU	45	39.041	2.116	5.42	29.300	48.800
2023135X	4	AB- CELL-D	3	MONOCYTE	8	CELL-DYN	69	6.038	0.865	14.32	4.000	8.000

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2023135X	4 AB- CELL-D	3 MONOCYTE	37 ABB.CEL RU	46	6.822	1.210	17.73	4.800	8.800			
2023135X	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	69	3.699	0.265	7.18	1.700	5.700			
2023135X	4 AB- CELL-D	4 EOSINOPHIL	37 ABB.CEL RU	46	3.624	0.332	9.16	1.600	5.600			
2023135X	4 AB- CELL-D	5 BASOPHIL	8 CELL-DYN	69	0.772	0.406	52.56	0	2.800			
2023135X	4 AB- CELL-D	5 BASOPHIL	37 ABB.CEL RU	45	0.211	0.281	133.14	0	2.200			
2023135X	5 BECTON D	1 NEUTROPHIL	29 B.C600,800	60	58.435	1.514	2.59	43.800	73.000			
2023135X	5 BECTON D	2 LYMPHOCYTE	29 B.C600,800	59	25.583	2.617	10.23	19.200	32.000			
2023135X	5 BECTON D	3 MONOCYTE	29 B.C600,800	59	7.891	1.113	14.10	5.900	9.900			
2023135X	5 BECTON D	4 EOSINOPHIL	29 B.C600,800	61	8.108	0.378	4.66	6.100	10.100			
2023135X	5 BECTON D	5 BASOPHIL	29 B.C600,800	60	0.000	0.000	0.00	0	2.000			
2023135X	6 MR Mindray	1 NEUTROPHIL	32 BC-5390	52	47.729	1.600	3.35	35.800	59.700			
2023135X	6 MR Mindray	2 LYMPHOCYTE	32 BC-5390	52	40.155	2.018	5.03	30.100	50.200			
2023135X	6 MR Mindray	3 MONOCYTE	32 BC-5390	52	7.021	1.257	17.90	5.000	9.000			
2023135X	6 MR Mindray	4 EOSINOPHIL	32 BC-5390	51	4.938	0.777	15.73	2.900	6.900			
2023135X	6 MR Mindray	5 BASOPHIL	32 BC-5390	52	26.963	0.559	2.07	20.200	33.700			
2023135X	7 DXH	1 NEUTROPHIL	36 BC DXH 500	53	48.626	3.918	8.06	36.500	60.800			
2023135X	7 DXH	2 LYMPHOCYTE	36 BC DXH 500	53	38.619	4.152	10.75	29.000	48.300			
2023135X	7 DXH	3 MONOCYTE	36 BC DXH 500	53	0.266	0.239	89.79	0	2.300			
2023135X	7 DXH	4 EOSINOPHIL	36 BC DXH 500	53	12.388	1.220	9.85	9.300	15.500			
2023135X	7 DXH	5 BASOPHIL	36 BC DXH 500	53	0.155	0.313	202.52	0	2.200			

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25 % OR 2.00 {GREATER}								
2023131Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	40	54.621	0.582	1.07	41.000	68.300			
2023131Y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	127	51.396	1.262	2.45	38.500	64.200			
2023131Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	40	24.515	0.386	1.57	18.400	30.600			

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023131Y	1	6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	127	26.155	0.897	3.43	19.600	32.700
2023131Y	1	6 PAR REG	3	MONOCYTE	31	SYSMEX XN	40	2.308	0.202	8.74	0.300	4.300
2023131Y	1	6 PAR REG	3	MONOCYTE	35	SYS XNL SR	127	0.835	0.280	33.56	0	2.800
2023131Y	1	6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	40	13.783	0.517	3.75	10.300	17.200
2023131Y	1	6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	126	13.866	0.607	4.38	10.400	17.300
2023131Y	1	6 PAR REG	5	BASOPHIL	31	SYSMEX XN	40	4.821	0.092	1.91	2.800	6.800
2023131Y	1	6 PAR REG	5	BASOPHIL	35	SYS XNL SR	128	7.929	0.619	7.81	5.900	9.900
2023131Y	1	6 PAR REG	6	IMMATURE G	31	SYSMEX XN	40	13.638	0.468	3.43	10.200	17.000
2023131Y	1	6 PAR REG	6	IMMATURE G	35	SYS XNL SR	127	13.698	0.690	5.04	10.300	17.100
2023131Y	1	6 PAR REG	7	NRBC	31	SYSMEX XN	22	0.095	0.111	115.96	0	2.100
185 Diff 6 Parameter Ins			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2023132Y	1	6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	40	52.153	0.860	1.65	39.100	65.200
2023132Y	1	6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	127	49.757	1.311	2.64	37.300	62.200
2023132Y	1	6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	40	26.590	0.615	2.31	19.900	33.200
2023132Y	1	6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	127	28.661	0.950	3.31	21.500	35.800
2023132Y	1	6 PAR REG	3	MONOCYTE	31	SYSMEX XN	40	3.116	0.318	10.22	1.100	5.100
2023132Y	1	6 PAR REG	3	MONOCYTE	35	SYS XNL SR	127	1.184	0.284	24.00	0	3.200
2023132Y	1	6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	40	13.183	0.555	4.21	9.900	16.500
2023132Y	1	6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	128	13.084	0.629	4.80	9.800	16.400
2023132Y	1	6 PAR REG	5	BASOPHIL	31	SYSMEX XN	40	4.863	0.096	1.97	2.900	6.900
2023132Y	1	6 PAR REG	5	BASOPHIL	35	SYS XNL SR	126	7.506	0.600	8.00	5.500	9.500
2023132Y	1	6 PAR REG	6	IMMATURE G	31	SYSMEX XN	40	12.521	0.213	1.70	9.400	15.700
2023132Y	1	6 PAR REG	6	IMMATURE G	35	SYS XNL SR	127	13.135	0.512	3.90	9.900	16.400
2023132Y	1	6 PAR REG	7	NRBC	31	SYSMEX XN	23	0.032	0.055	174.38	0	2.000
2023133Y	1	6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	40	62.974	0.782	1.24	47.200	78.700
2023133Y	1	6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	127	59.140	1.075	1.82	44.400	73.900
2023133Y	1	6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	40	14.198	0.804	5.66	10.600	17.700
2023133Y	1	6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	127	14.959	0.945	6.32	11.200	18.700
2023133Y	1	6 PAR REG	3	MONOCYTE	31	SYSMEX XN	40	1.735	0.357	20.60	0	3.700

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023133Y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	127	0.689	0.234	33.88	0	2.700		
2023133Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	40	16.385	0.295	1.80	12.300	20.500		
2023133Y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	127	16.125	0.505	3.13	12.100	20.200		
2023133Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	40	4.734	0.163	3.44	2.700	6.700		
2023133Y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	127	9.092	0.770	8.46	6.800	11.400		
2023133Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	40	15.863	0.834	5.26	11.900	19.800		
2023133Y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	127	16.289	0.991	6.08	12.200	20.400		
2023133Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	22	0.091	0.153	168.82	0	2.100		
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2023134Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	40	54.545	0.480	0.88	40.900	68.200		
2023134Y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	127	51.480	1.257	2.44	38.600	64.300		
2023134Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	40	24.615	0.435	1.77	18.500	30.800		
2023134Y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	127	26.152	0.763	2.92	19.600	32.700		
2023134Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	40	2.216	0.194	8.76	0.200	4.200		
2023134Y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	127	0.835	0.179	21.44	0	2.800		
2023134Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	40	13.825	0.548	3.96	10.400	17.300		
2023134Y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	127	13.802	0.559	4.05	10.400	17.300		
2023134Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	40	4.813	0.108	2.24	2.800	6.800		
2023134Y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	126	7.933	0.602	7.59	5.900	9.900		
2023134Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	40	13.560	0.420	3.10	10.200	17.000		
2023134Y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	126	14.011	0.745	5.32	10.500	17.500		
2023134Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	21	0.070	0.095	136.28	0	2.100		
2023135Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	40	63.192	0.631	1.00	47.400	79.000		
2023135Y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	126	59.085	1.161	1.97	44.300	73.900		
2023135Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	40	14.021	0.714	5.09	10.500	17.500		
2023135Y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	126	14.927	1.064	7.13	11.200	18.700		
2023135Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	40	1.721	0.261	15.15	0	3.700		
2023135Y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	126	0.734	0.338	46.08	0	2.700		
2023135Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	40	16.355	0.354	2.17	12.300	20.400		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023135Y	1	6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	126	16.180	0.511	3.16	12.100	20.200
2023135Y	1	6 PAR REG	5	BASOPHIL	31	SYSMEX XN	40	4.761	0.158	3.32	2.800	6.800
2023135Y	1	6 PAR REG	5	BASOPHIL	35	SYS XNL SR	126	9.083	0.771	8.49	6.800	11.400
2023135Y	1	6 PAR REG	6	IMMATURE G	31	SYSMEX XN	40	15.825	0.811	5.13	11.900	19.800
2023135Y	1	6 PAR REG	6	IMMATURE G	35	SYS XNL SR	126	16.233	0.956	5.89	12.200	20.300
2023135Y	1	6 PAR REG	7	NRBC	31	SYSMEX XN	22	0.077	0.128	165.23	0	2.100

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 128 DIFF 3 PARAMETER INS

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
186 DIFF 3 PARAMETER INS			CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2023131Z	8	3ST	1	Lymphocyte	1	KX 21 - Sy	22	29.262	0.868	2.97	21.900	36.600
2023131Z	8	3ST	1	Lymphocyte	3	ABBOTT C-D	32	29.778	1.266	4.25	22.300	37.200
2023131Z	8	3ST	2	Mix M/E/Ba	1	KX 21 - Sy	22	16.800	0.658	3.91	12.600	21.000
2023131Z	8	3ST	2	Mix M/E/Ba	3	ABBOTT C-D	32	11.516	1.377	11.96	8.600	14.400
2023131Z	8	3ST	3	Neutrophil	1	KX 21 - Sy	22	53.914	0.639	1.18	40.400	67.400
2023131Z	8	3ST	3	Neutrophil	3	ABBOTT C-D	32	58.747	2.107	3.59	44.100	73.400
2023131Z	9	3D	1	Lymphocyte	2	AcT Diff 2	16	31.625	1.167	3.69	23.700	39.500
2023131Z	9	3D	1	Lymphocyte	4	JT - Coult	10	31.570	1.793	5.68	23.700	39.500
2023131Z	9	3D	2	Mix M/E/Ba	2	AcT Diff 2	16	5.175	1.105	21.35	3.200	7.200
2023131Z	9	3D	3	Neutrophil	2	AcT Diff 2	16	63.200	1.772	2.80	47.400	79.000
2023132Z	8	3ST	1	Lymphocyte	1	KX 21 - Sy	22	10.748	1.089	10.13	8.100	13.400
2023132Z	8	3ST	1	Lymphocyte	3	ABBOTT C-D	32	12.713	1.552	12.21	9.500	15.900
2023132Z	8	3ST	2	Mix M/E/Ba	1	KX 21 - Sy	22	17.152	0.983	5.73	12.900	21.400
2023132Z	8	3ST	2	Mix M/E/Ba	3	ABBOTT C-D	33	12.116	1.626	13.42	9.100	15.100
2023132Z	8	3ST	3	Neutrophil	1	KX 21 - Sy	22	71.957	1.030	1.43	54.000	89.900
2023132Z	8	3ST	3	Neutrophil	3	ABBOTT C-D	31	74.906	2.414	3.22	56.200	93.600
2023132Z	9	3D	1	Lymphocyte	2	AcT Diff 2	16	61.250	2.073	3.38	45.900	76.600

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 128 DIFF 3 PARAMETER INS

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023132Z	9 3D	1	Lymphocyte	4	JT - Coult	10	58.750	3.196	5.44	44.100	73.400		
2023132Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	7.688	1.580	20.55	5.700	9.700		
2023132Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	31.156	2.135	6.85	23.400	38.900		
186 DIFF 3 PARAMETER INS		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2023133Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	22	29.176	0.615	2.11	21.900	36.500		
2023133Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	32	29.529	1.118	3.79	22.100	36.900		
2023133Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	22	16.671	0.787	4.72	12.500	20.800		
2023133Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	32	11.481	1.422	12.39	8.600	14.400		
2023133Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	22	54.152	0.899	1.66	40.600	67.700		
2023133Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	32	58.875	1.696	2.88	44.200	73.600		
2023133Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	31.806	1.071	3.37	23.900	39.800		
2023133Z	9 3D	1	Lymphocyte	4	JT - Coult	10	31.590	2.249	7.12	23.700	39.500		
2023133Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	5.231	0.872	16.67	3.200	7.200		
2023133Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	63.325	2.240	3.54	47.500	79.200		
2023134Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	22	60.662	0.999	1.65	45.500	75.800		
2023134Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	32	59.665	1.478	2.48	44.700	74.600		
2023134Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	22	13.862	0.952	6.87	10.400	17.300		
2023134Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	32	10.494	0.544	5.18	7.900	13.100		
2023134Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	22	25.476	0.525	2.06	19.100	31.800		
2023134Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	32	29.909	1.331	4.45	22.400	37.400		
2023134Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	14.250	0.752	5.28	10.700	17.800		
2023134Z	9 3D	1	Lymphocyte	4	JT - Coult	10	13.870	1.151	8.30	10.400	17.300		
2023134Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	4.494	0.615	13.68	2.500	6.500		
2023134Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	81.263	0.841	1.03	60.900	101.600		
2023135Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	22	10.710	1.466	13.69	8.000	13.400		
2023135Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	32	14.116	1.868	13.23	10.600	17.600		
2023135Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	22	17.081	1.218	7.13	12.800	21.400		
2023135Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	32	11.634	1.842	15.84	8.700	14.500		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 128 DIFF 3 PARAMETER INS

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2023135Z	8	3ST	3	Neutrophil	1	KX 21 - Sy	22	72.185	1.403	1.94	54.100	90.200
2023135Z	8	3ST	3	Neutrophil	3	ABBOTT C-D	32	73.881	2.942	3.98	55.400	92.400
2023135Z	9	3D	1	Lymphocyte	2	AcT Diff 2	16	61.250	1.809	2.95	45.900	76.600
2023135Z	9	3D	1	Lymphocyte	4	JT - Coult	10	59.950	3.781	6.31	45.000	74.900
2023135Z	9	3D	2	Mix M/E/Ba	2	AcT Diff 2	16	6.800	1.007	14.81	4.800	8.800
2023135Z	9	3D	3	Neutrophil	2	AcT Diff 2	16	32.088	1.490	4.64	24.100	40.100

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMP BLOOD COUNT NEW

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
549	RED CELL COUNT (RBC)		CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %							
2023-131	1	Sysmex CBC	132	SYSMEX XS			151	5.809	0.071	1.22	5.460	6.160
2023-131	1	Sysmex CBC	141	SYSMEX XN			45	5.838	0.045	0.77	5.490	6.190
2023-131	1	Sysmex CBC	152	Sysmex XN			125	5.870	0.064	1.10	5.520	6.220
2023-131	2	ACT PENTRA	125	COULTER AC			123	4.396	0.084	1.91	4.130	4.660
2023-131	2	ACT PENTRA	130	ABX PENTRA			29	4.358	0.064	1.47	4.100	4.620
2023-131	3	ADVIA 120	116	ADVIA-120			17	4.444	0.092	2.06	4.180	4.710
2023-131	4	AB-CELL DY	115	ABBOTT C-D			66	4.949	0.117	2.36	4.650	5.250
2023-131	4	AB-CELL DY	153	AB-CEL-RUB			47	5.174	0.090	1.74	4.860	5.480
2023-131	5	BECTON D	137	B.C DXH			60	5.426	0.069	1.27	5.100	5.750
2023-131	6	MR Mindray	145	BC-5390			52	5.337	0.097	1.82	5.020	5.660
2023-131	7	DXH500 Ser	149	BC DXH 500			53	4.682	0.085	1.81	4.400	4.960
2023-131	8	3 ST	84	ABBOTT C-D			32	4.010	0.100	2.49	3.770	4.250
2023-131	8	3 ST	138	SY KX-21N			16	3.986	0.068	1.70	3.750	4.220
2023-131	9	3 D	33	COULTER JT			10	4.682	0.077	1.65	4.400	4.960
2023-131	9	3 D	120	COULTER			16	4.589	0.081	1.77	4.310	4.860
2023-132	1	Sysmex CBC	132	SYSMEX XS			151	4.634	0.051	1.11	4.360	4.910

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2023-132	1	Sysmex CBC	141 SYSMEX XN	45	4.650	0.043	0.93	4.370	4.930	
2023-132	1	Sysmex CBC	152 Sysmex XN	125	4.650	0.049	1.05	4.370	4.930	
2023-132	2	ACT PENTRA	125 COULTER AC	123	2.409	0.041	1.72	2.260	2.550	
2023-132	2	ACT PENTRA	130 ABX PENTRA	29	2.382	0.042	1.75	2.240	2.520	
2023-132	3	ADVIA 120	116 ADVIA-120	17	2.257	0.023	1.04	2.120	2.390	
2023-132	4	AB-CELL DY	115 ABBOTT C-D	66	4.432	0.088	1.98	4.170	4.700	
2023-132	4	AB-CELL DY	153 AB-CEL-RUB	47	4.645	0.090	1.93	4.370	4.920	
2023-132	5	BECTON D	137 B.C DXH	60	4.128	0.060	1.44	3.880	4.380	
2023-132	6	MR Mindray	145 BC-5390	52	4.429	0.053	1.19	4.160	4.700	
2023-132	7	DXH500 Ser	149 BC DXH 500	53	2.459	0.050	2.04	2.310	2.610	
2023-132	8	3 ST	84 ABBOTT C-D	32	2.469	0.083	3.37	2.320	2.620	
2023-132	8	3 ST	138 SY KX-21N	16	2.455	0.039	1.59	2.310	2.600	
2023-132	9	3 D	33 COULTER JT	10	2.352	0.048	2.04	2.210	2.490	
2023-132	9	3 D	120 COULTER	16	2.408	0.041	1.69	2.260	2.550	
549 RED CELL COUNT (RBC)			CLIA TEST ID #: 775				EVALUATION CRITERIA: TARGET VALUE 6 %			
2023-133	1	Sysmex CBC	132 SYSMEX XS	151	2.320	0.031	1.33	2.180	2.460	
2023-133	1	Sysmex CBC	141 SYSMEX XN	45	2.303	0.029	1.26	2.160	2.440	
2023-133	1	Sysmex CBC	152 Sysmex XN	125	2.296	0.027	1.19	2.160	2.430	
2023-133	2	ACT PENTRA	125 COULTER AC	123	4.382	0.067	1.53	4.120	4.640	
2023-133	2	ACT PENTRA	130 ABX PENTRA	29	4.344	0.070	1.61	4.080	4.600	
2023-133	3	ADVIA 120	116 ADVIA-120	17	4.421	0.082	1.86	4.160	4.690	
2023-133	4	AB-CELL DY	115 ABBOTT C-D	66	2.218	0.044	1.99	2.090	2.350	
2023-133	4	AB-CELL DY	153 AB-CEL-RUB	47	2.279	0.050	2.18	2.140	2.420	
2023-133	5	BECTON D	137 B.C DXH	60	2.515	0.039	1.54	2.360	2.670	
2023-133	6	MR Mindray	145 BC-5390	52	2.202	0.033	1.50	2.070	2.330	
2023-133	7	DXH500 Ser	149 BC DXH 500	53	4.655	0.082	1.77	4.380	4.930	
2023-133	8	3 ST	84 ABBOTT C-D	32	4.018	0.090	2.24	3.780	4.260	
2023-133	8	3 ST	138 SY KX-21N	16	3.979	0.066	1.67	3.740	4.220	
2023-133	9	3 D	33 COULTER JT	10	4.664	0.105	2.26	4.380	4.940	
2023-133	9	3 D	120 COULTER	16	4.583	0.061	1.34	4.310	4.860	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %						
2023-134	1 Sysmex CBC	132 SYSMEX XS		151	5.815	0.071	1.22	5.470	6.160
2023-134	1 Sysmex CBC	141 SYSMEX XN		45	5.841	0.049	0.83	5.490	6.190
2023-134	1 Sysmex CBC	152 Sysmex XN		125	5.866	0.067	1.15	5.510	6.220
2023-134	2 ACT PENTRA	125 COULTER AC		123	5.064	0.098	1.94	4.760	5.370
2023-134	2 ACT PENTRA	130 ABX PENTRA		29	5.018	0.074	1.46	4.720	5.320
2023-134	3 ADVIA 120	116 ADVIA-120		17	5.314	0.100	1.88	4.990	5.630
2023-134	4 AB-CELL DY	115 ABBOTT C-D		66	4.954	0.101	2.05	4.660	5.250
2023-134	4 AB-CELL DY	153 AB-CEL-RUB		47	5.185	0.108	2.08	4.870	5.500
2023-134	5 BECTON D	137 B.C DXH		60	5.410	0.072	1.33	5.090	5.740
2023-134	6 MR Mindray	145 BC-5390		52	5.323	0.094	1.77	5.000	5.640
2023-134	7 DXH500 Ser	149 BC DXH 500		53	5.173	0.098	1.90	4.860	5.480
2023-134	8 3 ST	84 ABBOTT C-D		32	5.562	0.138	2.48	5.230	5.900
2023-134	8 3 ST	138 SY KX-21N		16	5.596	0.081	1.44	5.260	5.930
2023-134	9 3 D	33 COULTER JT		10	5.734	0.058	1.01	5.390	6.080
2023-134	9 3 D	120 COULTER		16	5.719	0.100	1.75	5.380	6.060
2023-135	1 Sysmex CBC	132 SYSMEX XS		151	2.322	0.030	1.29	2.180	2.460
2023-135	1 Sysmex CBC	141 SYSMEX XN		45	2.308	0.031	1.34	2.170	2.450
2023-135	1 Sysmex CBC	152 Sysmex XN		125	2.294	0.027	1.19	2.160	2.430
2023-135	2 ACT PENTRA	125 COULTER AC		123	2.400	0.041	1.73	2.260	2.540
2023-135	2 ACT PENTRA	130 ABX PENTRA		29	2.390	0.042	1.75	2.250	2.530
2023-135	3 ADVIA 120	116 ADVIA-120		17	2.272	0.028	1.22	2.140	2.410
2023-135	4 AB-CELL DY	115 ABBOTT C-D		66	2.221	0.049	2.21	2.090	2.350
2023-135	4 AB-CELL DY	153 AB-CEL-RUB		47	2.284	0.044	1.93	2.150	2.420
2023-135	5 BECTON D	137 B.C DXH		60	2.511	0.038	1.53	2.360	2.660
2023-135	6 MR Mindray	145 BC-5390		52	2.204	0.031	1.40	2.070	2.340
2023-135	7 DXH500 Ser	149 BC DXH 500		53	2.473	0.057	2.29	2.320	2.620
2023-135	8 3 ST	84 ABBOTT C-D		32	2.469	0.061	2.46	2.320	2.620
2023-135	8 3 ST	138 SY KX-21N		16	2.458	0.033	1.34	2.310	2.610

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW							
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-135	9 3 D	33 COULTER JT		10	2.393	0.059	2.46	2.250	2.540		
2023-135	9 3 D	120 COULTER		16	2.400	0.064	2.67	2.260	2.540		
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785	EVALUATION CRITERIA: TARGET VALUE 6 %								
2023-131	1 Sysmex CBC	132 SYSMEX XS		152	53.584	1.011	1.89	50.400	56.800		
2023-131	1 Sysmex CBC	141 SYSMEX XN		45	52.668	0.837	1.59	49.500	55.800		
2023-131	1 Sysmex CBC	152 Sysmex XN		124	52.643	1.066	2.03	49.500	55.800		
2023-131	2 ACT PENTRA	125 COULTER AC		122	37.541	0.769	2.05	35.300	39.800		
2023-131	2 ACT PENTRA	130 ABX PENTRA		29	37.038	0.542	1.46	34.800	39.300		
2023-131	3 ADVIA 120	116 ADVIA-120		17	36.029	0.828	2.30	33.900	38.200		
2023-131	4 AB-CELL DY	115 ABBOTT C-D		69	44.028	1.215	2.76	41.400	46.700		
2023-131	4 AB-CELL DY	153 AB-CEL-RUB		44	45.698	1.208	2.64	43.000	48.400		
2023-131	5 BECTON D	137 B.C DXH		60	56.977	0.917	1.61	53.600	60.400		
2023-131	6 MR Mindray	145 BC-5390		52	47.398	1.159	2.44	44.600	50.200		
2023-131	7 DXH500 Ser	149 BC DXH 500		53	40.912	1.076	2.63	38.500	43.400		
2023-131	8 3 ST	84 ABBOTT C-D		32	34.406	0.952	2.77	32.300	36.500		
2023-131	8 3 ST	98 SYSMEX K-		10	31.533	0.683	2.17	29.600	33.400		
2023-131	8 3 ST	138 SY KX-21N		16	31.869	0.618	1.94	30.000	33.800		
2023-131	9 3 D	33 COULTER JT		10	39.480	1.136	2.88	37.100	41.800		
2023-131	9 3 D	120 COULTER		16	39.269	1.325	3.37	36.900	41.600		
2023-132	1 Sysmex CBC	132 SYSMEX XS		152	40.207	0.709	1.76	37.800	42.600		
2023-132	1 Sysmex CBC	141 SYSMEX XN		45	39.607	0.626	1.58	37.200	42.000		
2023-132	1 Sysmex CBC	152 Sysmex XN		124	39.574	0.714	1.80	37.200	41.900		
2023-132	2 ACT PENTRA	125 COULTER AC		122	18.893	0.426	2.25	17.800	20.000		
2023-132	2 ACT PENTRA	130 ABX PENTRA		29	18.469	0.398	2.16	17.400	19.600		
2023-132	3 ADVIA 120	116 ADVIA-120		17	15.782	0.179	1.13	14.800	16.700		
2023-132	4 AB-CELL DY	115 ABBOTT C-D		69	37.406	0.892	2.39	35.200	39.700		
2023-132	4 AB-CELL DY	153 AB-CEL-RUB		44	38.932	1.008	2.59	36.600	41.300		
2023-132	5 BECTON D	137 B.C DXH		60	38.400	0.658	1.71	36.100	40.700		
2023-132	6 MR Mindray	145 BC-5390		52	36.863	0.915	2.48	34.700	39.100		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-132	7 DXH500 Ser	149 BC DXH 500			53	19.545	0.499	2.55	18.400	20.700	
2023-132	8 3 ST	84 ABBOTT C-D			32	17.584	0.722	4.11	16.500	18.600	
2023-132	8 3 ST	98 SYSMEX K-			10	16.756	0.271	1.62	15.800	17.800	
2023-132	8 3 ST	138 SY KX-21N			16	16.819	0.433	2.58	15.800	17.800	
2023-132	9 3 D	33 COULTER JT			10	17.190	0.359	2.09	16.200	18.200	
2023-132	9 3 D	120 COULTER			16	17.669	0.498	2.82	16.600	18.700	
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %							
2023-133	1 Sysmex CBC	132 SYSMEX XS			152	18.222	0.343	1.88	17.100	19.300	
2023-133	1 Sysmex CBC	141 SYSMEX XN			45	17.656	0.291	1.65	16.600	18.700	
2023-133	1 Sysmex CBC	152 Sysmex XN			124	17.672	0.361	2.04	16.600	18.700	
2023-133	2 ACT PENTRA	125 COULTER AC			122	37.432	0.763	2.04	35.200	39.700	
2023-133	2 ACT PENTRA	130 ABX PENTRA			29	36.876	0.634	1.72	34.700	39.100	
2023-133	3 ADVIA 120	116 ADVIA-120			17	35.850	0.660	1.84	33.700	38.000	
2023-133	4 AB-CELL DY	115 ABBOTT C-D			69	15.788	0.327	2.07	14.800	16.700	
2023-133	4 AB-CELL DY	153 AB-CEL-RUB			44	15.914	0.403	2.53	15.000	16.900	
2023-133	5 BECTON D	137 B.C DXH			60	20.302	0.350	1.73	19.100	21.500	
2023-133	6 MR Mindray	145 BC-5390			52	16.465	0.356	2.16	15.500	17.500	
2023-133	7 DXH500 Ser	149 BC DXH 500			53	40.794	0.802	1.96	38.300	43.200	
2023-133	8 3 ST	84 ABBOTT C-D			32	34.559	0.822	2.38	32.500	36.600	
2023-133	8 3 ST	98 SYSMEX K-			10	31.544	0.587	1.86	29.700	33.400	
2023-133	8 3 ST	138 SY KX-21N			16	31.781	0.655	2.06	29.900	33.700	
2023-133	9 3 D	33 COULTER JT			10	39.290	1.147	2.92	36.900	41.600	
2023-133	9 3 D	120 COULTER			16	38.731	0.874	2.26	36.400	41.100	
2023-134	1 Sysmex CBC	132 SYSMEX XS			152	53.623	1.012	1.89	50.400	56.800	
2023-134	1 Sysmex CBC	141 SYSMEX XN			45	52.714	0.884	1.68	49.600	55.900	
2023-134	1 Sysmex CBC	152 Sysmex XN			124	52.751	1.085	2.06	49.600	55.900	
2023-134	2 ACT PENTRA	125 COULTER AC			122	45.053	0.843	1.87	42.400	47.800	
2023-134	2 ACT PENTRA	130 ABX PENTRA			29	44.418	0.619	1.39	41.800	47.100	
2023-134	3 ADVIA 120	116 ADVIA-120			17	47.265	0.911	1.93	44.400	50.100	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2023-134	4 AB-CELL DY	115 ABBOTT C-D		69	44.039	1.109	2.52	41.400	46.700		
2023-134	4 AB-CELL DY	153 AB-CEL-RUB		44	45.759	1.302	2.84	43.000	48.500		
2023-134	5 BECTON D	137 B.C DXH		60	56.783	0.949	1.67	53.400	60.200		
2023-134	6 MR Mindray	145 BC-5390		52	47.235	1.185	2.51	44.400	50.100		
2023-134	7 DXH500 Ser	149 BC DXH 500		53	48.851	1.121	2.29	45.900	51.800		
2023-134	8 3 ST	84 ABBOTT C-D		32	52.032	1.146	2.20	48.900	55.200		
2023-134	8 3 ST	98 SYSMEX K-		10	48.533	0.913	1.88	45.600	51.400		
2023-134	8 3 ST	138 SY KX-21N		16	48.788	0.905	1.85	45.900	51.700		
2023-134	9 3 D	33 COULTER JT		10	52.310	0.850	1.63	49.200	55.400		
2023-134	9 3 D	120 COULTER		16	52.594	1.484	2.82	49.400	55.700		
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %							
2023-135	1 Sysmex CBC	132 SYSMEX XS		152	18.219	0.332	1.82	17.100	19.300		
2023-135	1 Sysmex CBC	141 SYSMEX XN		45	17.676	0.285	1.61	16.600	18.700		
2023-135	1 Sysmex CBC	152 Sysmex XN		124	17.645	0.340	1.93	16.600	18.700		
2023-135	2 ACT PENTRA	125 COULTER AC		122	19.012	1.737	9.14	17.900	20.200		
2023-135	2 ACT PENTRA	130 ABX PENTRA		29	18.521	0.396	2.14	17.400	19.600		
2023-135	3 ADVIA 120	116 ADVIA-120		17	15.888	0.211	1.33	14.900	16.800		
2023-135	4 AB-CELL DY	115 ABBOTT C-D		69	15.803	0.316	2.00	14.900	16.800		
2023-135	4 AB-CELL DY	153 AB-CEL-RUB		44	15.975	0.398	2.49	15.000	16.900		
2023-135	5 BECTON D	137 B.C DXH		60	20.272	0.368	1.81	19.100	21.500		
2023-135	6 MR Mindray	145 BC-5390		52	16.467	0.412	2.50	15.500	17.500		
2023-135	7 DXH500 Ser	149 BC DXH 500		53	19.704	0.522	2.65	18.500	20.900		
2023-135	8 3 ST	84 ABBOTT C-D		32	17.631	0.523	2.97	16.600	18.700		
2023-135	8 3 ST	98 SYSMEX K-		10	16.744	0.267	1.60	15.700	17.700		
2023-135	8 3 ST	138 SY KX-21N		16	16.800	0.335	2.00	15.800	17.800		
2023-135	9 3 D	33 COULTER JT		10	17.350	0.338	1.95	16.300	18.400		
2023-135	9 3 D	120 COULTER		16	17.581	0.521	2.96	16.500	18.600		
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}							
2023-131	1 Sysmex CBC	132 SYSMEX XS		152	18.047	0.194	1.07	16.800	19.300		
2023-131	1 Sysmex CBC	141 SYSMEX XN		45	17.964	0.169	0.94	16.700	19.200		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2023-131	1	Sysmex CBC	152 Sysmex XN	124	18.035	0.151	0.84	16.800	19.300		
2023-131	2	ACT PENTRA	125 COULTER AC	123	13.473	0.192	1.43	12.500	14.400		
2023-131	2	ACT PENTRA	130 ABX PENTRA	29	13.424	0.179	1.34	12.500	14.400		
2023-131	3	ADVIA 120	116 ADVIA-120	17	13.229	0.214	1.61	12.300	14.200		
2023-131	4	AB-CELL DY	115 ABBOTT C-D	69	16.935	0.288	1.70	15.700	18.100		
2023-131	4	AB-CELL DY	153 AB-CEL-RUB	44	16.920	0.303	1.79	15.700	18.100		
2023-131	5	BECTON D	137 B.C DXH	60	17.375	0.203	1.17	16.200	18.600		
2023-131	6	MR Mindray	145 BC-5390	52	16.602	0.159	0.96	15.400	17.800		
2023-131	7	DXH500 Ser	149 BC DXH 500	53	14.538	0.207	1.42	13.500	15.600		
2023-131	8	3 ST	84 ABBOTT C-D	32	12.087	0.243	2.01	11.200	12.900		
2023-131	8	3 ST	98 SYSMEX K-	10	11.967	0.133	1.11	11.100	12.800		
2023-131	8	3 ST	138 SY KX-21N	16	11.975	0.139	1.16	11.100	12.800		
2023-131	9	3 D	33 COULTER JT	10	13.750	0.246	1.79	12.800	14.700		
2023-131	9	3 D	120 COULTER	16	13.619	0.419	3.07	12.700	14.600		
551 HEMOGLOBIN (HGB)			CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}						
2023-132	1	Sysmex CBC	132 SYSMEX XS	152	13.532	0.143	1.06	12.600	14.500		
2023-132	1	Sysmex CBC	141 SYSMEX XN	45	13.600	0.141	1.04	12.600	14.600		
2023-132	1	Sysmex CBC	152 Sysmex XN	124	13.618	0.110	0.81	12.700	14.600		
2023-132	2	ACT PENTRA	125 COULTER AC	123	6.684	0.111	1.66	6.200	7.200		
2023-132	2	ACT PENTRA	130 ABX PENTRA	29	6.497	0.096	1.48	6.000	7.000		
2023-132	3	ADVIA 120	116 ADVIA-120	17	5.859	0.133	2.27	5.400	6.300		
2023-132	4	AB-CELL DY	115 ABBOTT C-D	69	13.849	0.192	1.39	12.900	14.800		
2023-132	4	AB-CELL DY	153 AB-CEL-RUB	44	13.780	0.226	1.64	12.800	14.700		
2023-132	5	BECTON D	137 B.C DXH	60	11.863	0.163	1.38	11.000	12.700		
2023-132	6	MR Mindray	145 BC-5390	52	12.929	0.130	1.01	12.000	13.800		
2023-132	7	DXH500 Ser	149 BC DXH 500	53	6.832	0.099	1.44	6.400	7.300		
2023-132	8	3 ST	84 ABBOTT C-D	32	6.348	0.160	2.53	5.900	6.800		
2023-132	8	3 ST	98 SYSMEX K-	10	6.078	0.123	2.02	5.700	6.500		
2023-132	8	3 ST	138 SY KX-21N	16	6.100	0.073	1.20	5.700	6.500		
2023-132	9	3 D	33 COULTER JT	10	5.700	0.141	2.48	5.300	6.100		
2023-132	9	3 D	120 COULTER	16	5.981	0.207	3.46	5.600	6.400		

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
551	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE	7	% OR 2.00	{LOWER}			
2023-133	1	Sysmex CBC	132 SYSMEX XS	152	5.915	0.079	1.33	5.500	6.300
2023-133	1	Sysmex CBC	141 SYSMEX XN	45	6.018	0.102	1.69	5.600	6.400
2023-133	1	Sysmex CBC	152 Sysmex XN	124	6.014	0.064	1.06	5.600	6.400
2023-133	2	ACT PENTRA	125 COULTER AC	123	13.460	0.181	1.35	12.500	14.400
2023-133	2	ACT PENTRA	130 ABX PENTRA	29	13.369	0.170	1.27	12.400	14.300
2023-133	3	ADVIA 120	116 ADVIA-120	17	13.175	0.214	1.62	12.300	14.100
2023-133	4	AB-CELL DY	115 ABBOTT C-D	69	5.964	0.125	2.10	5.500	6.400
2023-133	4	AB-CELL DY	153 AB-CEL-RUB	44	5.705	0.138	2.42	5.300	6.100
2023-133	5	BECTON D	137 B.C DXH	60	6.292	0.089	1.41	5.900	6.700
2023-133	6	MR Mindray	145 BC-5390	52	5.639	0.069	1.22	5.200	6.000
2023-133	7	DXH500 Ser	149 BC DXH 500	53	14.477	0.191	1.32	13.500	15.500
2023-133	8	3 ST	84 ABBOTT C-D	32	12.168	0.189	1.55	11.300	13.000
2023-133	8	3 ST	98 SYSMEX K-	10	11.944	0.171	1.43	11.100	12.800
2023-133	8	3 ST	138 SY KX-21N	16	11.994	0.164	1.37	11.200	12.800
2023-133	9	3 D	33 COULTER JT	10	13.750	0.287	2.09	12.800	14.700
2023-133	9	3 D	120 COULTER	16	13.675	0.390	2.85	12.700	14.600
2023-134	1	Sysmex CBC	132 SYSMEX XS	152	18.030	0.198	1.10	16.800	19.300
2023-134	1	Sysmex CBC	141 SYSMEX XN	45	17.991	0.173	0.96	16.700	19.300
2023-134	1	Sysmex CBC	152 Sysmex XN	124	18.029	0.147	0.81	16.800	19.300
2023-134	2	ACT PENTRA	125 COULTER AC	123	16.281	0.202	1.24	15.100	17.400
2023-134	2	ACT PENTRA	130 ABX PENTRA	29	16.276	0.225	1.38	15.100	17.400
2023-134	3	ADVIA 120	116 ADVIA-120	17	17.512	0.242	1.38	16.300	18.700
2023-134	4	AB-CELL DY	115 ABBOTT C-D	69	16.891	0.228	1.35	15.700	18.100
2023-134	4	AB-CELL DY	153 AB-CEL-RUB	44	16.845	0.331	1.97	15.700	18.000
2023-134	5	BECTON D	137 B.C DXH	60	17.340	0.189	1.09	16.100	18.600
2023-134	6	MR Mindray	145 BC-5390	52	16.556	0.197	1.19	15.400	17.700
2023-134	7	DXH500 Ser	149 BC DXH 500	53	17.198	0.252	1.46	16.000	18.400
2023-134	8	3 ST	84 ABBOTT C-D	32	17.959	0.427	2.38	16.700	19.200

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-134	8 3 ST	98 SYSMEX K-			10	17.978	0.204	1.14	16.700	19.200	
2023-134	8 3 ST	138 SY KX-21N			16	17.988	0.289	1.61	16.700	19.200	
2023-134	9 3 D	33 COULTER JT			10	18.670	0.300	1.61	17.400	20.000	
2023-134	9 3 D	120 COULTER			16	18.400	0.472	2.56	17.100	19.700	
551	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}							
2023-135	1 Sysmex CBC	132 SYSMEX XS			152	5.934	0.082	1.38	5.500	6.300	
2023-135	1 Sysmex CBC	141 SYSMEX XN			45	6.027	0.095	1.58	5.600	6.400	
2023-135	1 Sysmex CBC	152 Sysmex XN			124	6.024	0.067	1.11	5.600	6.400	
2023-135	2 ACT PENTRA	125 COULTER AC			123	6.679	0.110	1.64	6.200	7.100	
2023-135	2 ACT PENTRA	130 ABX PENTRA			29	6.503	0.103	1.59	6.000	7.000	
2023-135	3 ADVIA 120	116 ADVIA-120			17	5.894	0.126	2.14	5.500	6.300	
2023-135	4 AB-CELL DY	115 ABBOTT C-D			69	5.963	0.108	1.82	5.500	6.400	
2023-135	4 AB-CELL DY	153 AB-CEL-RUB			44	5.735	0.126	2.19	5.300	6.100	
2023-135	5 BECTON D	137 B.C DXH			60	6.290	0.091	1.44	5.800	6.700	
2023-135	6 MR Mindray	145 BC-5390			52	5.631	0.072	1.28	5.200	6.000	
2023-135	7 DXH500 Ser	149 BC DXH 500			53	6.815	0.126	1.86	6.300	7.300	
2023-135	8 3 ST	84 ABBOTT C-D			32	6.355	0.134	2.11	5.900	6.800	
2023-135	8 3 ST	98 SYSMEX K-			10	6.089	0.137	2.25	5.700	6.500	
2023-135	8 3 ST	138 SY KX-21N			16	6.060	0.136	2.24	5.600	6.500	
2023-135	9 3 D	33 COULTER JT			10	5.680	0.147	2.59	5.300	6.100	
2023-135	9 3 D	120 COULTER			16	5.956	0.169	2.84	5.500	6.400	
552	WHITE CELL COUNT(WBC)	CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %							
2023-131	1 Sysmex CBC	132 SYSMEX XS			151	19.497	0.484	2.48	16.600	22.400	
2023-131	1 Sysmex CBC	141 SYSMEX XN			46	20.429	0.383	1.88	17.400	23.500	
2023-131	1 Sysmex CBC	152 Sysmex XN			124	19.549	0.393	2.01	16.600	22.500	
2023-131	2 ACT PENTRA	125 COULTER AC			123	7.557	0.227	3.01	6.400	8.700	
2023-131	2 ACT PENTRA	130 ABX PENTRA			29	7.576	0.225	2.97	6.400	8.700	
2023-131	3 ADVIA 120	116 ADVIA-120			17	7.176	0.300	4.18	6.100	8.300	
2023-131	4 AB-CELL DY	115 ABBOTT C-D			70	21.019	0.602	2.86	17.900	24.200	
2023-131	4 AB-CELL DY	153 AB-CEL-RUB			43	20.407	0.603	2.96	17.300	23.500	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-131	5	BECTON D	137 B.C DXH	60	20.502	0.421	2.05	17.400	23.600
2023-131	6	MR Mindray	145 BC-5390	52	18.018	0.392	2.18	15.300	20.700
2023-131	7	DXH500 Ser	149 BC DXH 500	53	7.187	0.227	3.16	6.100	8.300
2023-131	8	3 ST	84 ABBOTT C-D	32	8.356	0.485	5.81	7.100	9.600
2023-131	8	3 ST	98 SYSMEX K-	10	8.444	0.275	3.26	7.200	9.700
2023-131	8	3 ST	138 SY KX-21N	16	8.319	0.230	2.76	7.100	9.600
2023-131	9	3 D	33 COULTER JT	10	8.260	0.388	4.70	7.000	9.500
2023-131	9	3 D	120 COULTER	16	8.163	0.302	3.70	6.900	9.400
552 WHITE CELL COUNT(WBC)				CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %				
2023-132	1	Sysmex CBC	132 SYSMEX XS	151	7.454	0.293	3.93	6.300	8.600
2023-132	1	Sysmex CBC	141 SYSMEX XN	46	7.618	0.180	2.37	6.500	8.800
2023-132	1	Sysmex CBC	152 Sysmex XN	124	7.329	0.151	2.06	6.200	8.400
2023-132	2	ACT PENTRA	125 COULTER AC	123	2.408	0.092	3.82	2.000	2.800
2023-132	2	ACT PENTRA	130 ABX PENTRA	29	2.404	0.094	3.93	2.000	2.800
2023-132	3	ADVIA 120	116 ADVIA-120	17	2.953	0.104	3.51	2.500	3.400
2023-132	4	AB-CELL DY	115 ABBOTT C-D	70	8.169	0.262	3.21	6.900	9.400
2023-132	4	AB-CELL DY	153 AB-CEL-RUB	43	8.245	0.261	3.17	7.000	9.500
2023-132	5	BECTON D	137 B.C DXH	60	8.577	0.186	2.16	7.300	9.900
2023-132	6	MR Mindray	145 BC-5390	52	7.942	0.170	2.14	6.800	9.100
2023-132	7	DXH500 Ser	149 BC DXH 500	53	2.840	0.132	4.66	2.400	3.300
2023-132	8	3 ST	84 ABBOTT C-D	32	2.731	0.239	8.75	2.300	3.100
2023-132	8	3 ST	98 SYSMEX K-	10	2.733	0.176	6.45	2.300	3.100
2023-132	8	3 ST	138 SY KX-21N	16	2.619	0.088	3.37	2.200	3.000
2023-132	9	3 D	33 COULTER JT	10	2.130	0.078	3.67	1.800	2.400
2023-132	9	3 D	120 COULTER	16	2.294	0.130	5.66	1.900	2.600
2023-133	1	Sysmex CBC	132 SYSMEX XS	151	3.885	0.114	2.93	3.300	4.500
2023-133	1	Sysmex CBC	141 SYSMEX XN	46	3.778	0.109	2.89	3.200	4.300
2023-133	1	Sysmex CBC	152 Sysmex XN	124	3.595	0.100	2.79	3.100	4.100
2023-133	2	ACT PENTRA	125 COULTER AC	123	7.534	0.242	3.21	6.400	8.700

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2023-133	2	ACT PENTRA	130 ABX PENTRA	29	7.590	0.216	2.84	6.500	8.700
2023-133	3	ADVIA 120	116 ADVIA-120	17	7.144	0.355	4.97	6.100	8.200
2023-133	4	AB-CELL DY	115 ABBOTT C-D	70	3.261	0.123	3.78	2.800	3.800
2023-133	4	AB-CELL DY	153 AB-CEL-RUB	43	3.205	0.122	3.80	2.700	3.700
2023-133	5	BECTON D	137 B.C DXH	60	4.073	0.177	4.34	3.500	4.700
2023-133	6	MR Mindray	145 BC-5390	52	3.356	0.091	2.70	2.900	3.900
2023-133	7	DXH500 Ser	149 BC DXH 500	53	7.081	0.203	2.86	6.000	8.100
2023-133	8	3 ST	84 ABBOTT C-D	32	8.388	0.485	5.78	7.100	9.600
2023-133	8	3 ST	98 SYSMEX K-	10	8.400	0.262	3.12	7.100	9.700
2023-133	8	3 ST	138 SY KX-21N	16	8.369	0.247	2.95	7.100	9.600
2023-133	9	3 D	33 COULTER JT	10	8.230	0.415	5.04	7.000	9.500
2023-133	9	3 D	120 COULTER	16	8.150	0.306	3.76	6.900	9.400
552 WHITE CELL COUNT(WBC		CLIA	TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %					
2023-134	1	Sysmex CBC	132 SYSMEX XS	151	19.551	0.517	2.65	16.600	22.500
2023-134	1	Sysmex CBC	141 SYSMEX XN	46	20.338	0.495	2.43	17.300	23.400
2023-134	1	Sysmex CBC	152 Sysmex XN	124	19.491	0.371	1.91	16.600	22.400
2023-134	2	ACT PENTRA	125 COULTER AC	123	18.214	0.549	3.01	15.500	20.900
2023-134	2	ACT PENTRA	130 ABX PENTRA	29	18.234	0.639	3.50	15.500	21.000
2023-134	3	ADVIA 120	116 ADVIA-120	17	16.806	0.598	3.56	14.300	19.300
2023-134	4	AB-CELL DY	115 ABBOTT C-D	70	21.003	0.650	3.09	17.900	24.200
2023-134	4	AB-CELL DY	153 AB-CEL-RUB	43	20.321	0.543	2.67	17.300	23.400
2023-134	5	BECTON D	137 B.C DXH	60	20.432	0.407	1.99	17.400	23.500
2023-134	6	MR Mindray	145 BC-5390	52	17.996	0.361	2.00	15.300	20.700
2023-134	7	DXH500 Ser	149 BC DXH 500	53	18.433	0.396	2.15	15.700	21.200
2023-134	8	3 ST	84 ABBOTT C-D	32	20.291	1.229	6.06	17.200	23.300
2023-134	8	3 ST	98 SYSMEX K-	10	20.367	0.636	3.12	17.300	23.400
2023-134	8	3 ST	138 SY KX-21N	16	20.338	0.665	3.27	17.300	23.400
2023-134	9	3 D	33 COULTER JT	10	22.370	0.800	3.58	19.000	25.700
2023-134	9	3 D	120 COULTER	16	21.269	0.692	3.25	18.100	24.500

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
552	WHITE CELL COUNT(WBC	CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE	15	%					
2023-135	1	Sysmex CBC	132 SYSMEX XS	151	3.888	0.112	2.89	3.300	4.500	
2023-135	1	Sysmex CBC	141 SYSMEX XN	46	3.778	0.107	2.84	3.200	4.300	
2023-135	1	Sysmex CBC	152 Sysmex XN	124	3.602	0.104	2.90	3.100	4.100	
2023-135	2	ACT PENTRA	125 COULTER AC	123	2.416	0.086	3.57	2.100	2.800	
2023-135	2	ACT PENTRA	130 ABX PENTRA	29	2.431	0.095	3.91	2.100	2.800	
2023-135	3	ADVIA 120	116 ADVIA-120	17	3.029	0.118	3.88	2.600	3.500	
2023-135	4	AB-CELL DY	115 ABBOTT C-D	70	3.281	0.116	3.53	2.800	3.800	
2023-135	4	AB-CELL DY	153 AB-CEL-RUB	43	3.291	0.178	5.40	2.800	3.800	
2023-135	5	BECTON D	137 B.C DXH	60	4.080	0.178	4.36	3.500	4.700	
2023-135	6	MR Mindray	145 BC-5390	52	3.344	0.091	2.71	2.800	3.800	
2023-135	7	DXH500 Ser	149 BC DXH 500	53	2.894	0.150	5.17	2.500	3.300	
2023-135	8	3 ST	84 ABBOTT C-D	32	2.894	0.226	7.82	2.500	3.300	
2023-135	8	3 ST	98 SYSMEX K-	10	2.733	0.205	7.52	2.300	3.100	
2023-135	8	3 ST	138 SY KX-21N	16	2.631	0.085	3.21	2.200	3.000	
2023-135	9	3 D	33 COULTER JT	10	2.170	0.127	5.85	1.800	2.500	
2023-135	9	3 D	120 COULTER	16	2.319	0.155	6.68	2.000	2.700	

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 134 PLATELET COUNT NEW						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
553	PLATELET COUNT	CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE	25	%					
2023-131	1	SYSMEX CBC	79 SYSMEX XS	151	377.223	17.489	4.64	283.000	472.000	
2023-131	1	SYSMEX CBC	87 SYSMEX XN	46	404.178	12.879	3.19	303.000	505.000	
2023-131	1	SYSMEX CBC	98 SYSMEX 350	124	401.724	14.492	3.61	301.000	502.000	
2023-131	2	ACT5 PENTR	73 COULTER AC	117	242.043	10.020	4.14	182.000	303.000	
2023-131	2	ACT5 PENTR	78 ABX PENTRA	29	240.607	7.399	3.07	180.000	301.000	
2023-131	3	ADVIA 120	64 ADVIA-120	17	207.000	13.890	6.71	155.000	259.000	
2023-131	4	AB- CELL-D	63 ABBOTT C-D	65	484.462	22.339	4.61	363.000	606.000	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-131	4 AB- CELL-D	100 ABB.RUBY			46	463.523	16.261	3.51	348.000	579.000	
2023-131	5 BECTON D	83 B.C600,800			60	449.356	11.828	2.63	337.000	562.000	
2023-131	6 MR Mindray	91 BC-5390			52	562.769	17.108	3.04	422.000	703.000	
2023-131	8 3 ST	31 ABBOTT C-D			32	192.750	21.073	10.93	145.000	241.000	
2023-131	8 3 ST	84 KX21,300			17	207.588	13.360	6.44	156.000	259.000	
2023-131	9 3 D	68 COULTERAcT			16	287.867	10.776	3.74	216.000	360.000	
2023-131	10 DxH	94 BC DXH 500			52	251.824	12.871	5.11	189.000	315.000	
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2023-132	1 SYSMEX CBC	79 SYSMEX XS			151	191.142	9.084	4.75	143.000	239.000	
2023-132	1 SYSMEX CBC	87 SYSMEX XN			46	194.935	6.951	3.57	146.000	244.000	
2023-132	1 SYSMEX CBC	98 SYSMEX 350			124	193.934	7.343	3.79	145.000	242.000	
2023-132	2 ACT5 PENTR	73 COULTER AC			117	69.560	4.995	7.18	52.000	87.000	
2023-132	2 ACT5 PENTR	78 ABX PENTRA			29	70.621	5.314	7.52	53.000	88.000	
2023-132	3 ADVIA 120	64 ADVIA-120			17	51.412	4.088	7.95	39.000	64.000	
2023-132	4 AB- CELL-D	63 ABBOTT C-D			65	253.954	12.245	4.82	190.000	317.000	
2023-132	4 AB- CELL-D	100 ABB.RUBY			46	246.200	11.752	4.77	185.000	308.000	
2023-132	5 BECTON D	83 B.C600,800			60	257.183	6.559	2.55	193.000	321.000	
2023-132	6 MR Mindray	91 BC-5390			52	265.843	7.915	2.98	199.000	332.000	
2023-132	8 3 ST	31 ABBOTT C-D			32	65.063	7.084	10.89	49.000	81.000	
2023-132	8 3 ST	84 KX21,300			17	63.882	3.046	4.77	48.000	80.000	
2023-132	9 3 D	68 COULTERAcT			16	71.875	5.946	8.27	54.000	90.000	
2023-132	10 DxH	94 BC DXH 500			52	77.769	5.645	7.26	58.000	97.000	
2023-133	1 SYSMEX CBC	79 SYSMEX XS			151	54.987	3.339	6.07	41.000	69.000	
2023-133	1 SYSMEX CBC	87 SYSMEX XN			46	49.435	2.939	5.94	37.000	62.000	
2023-133	1 SYSMEX CBC	98 SYSMEX 350			124	50.106	3.041	6.07	38.000	63.000	
2023-133	2 ACT5 PENTR	73 COULTER AC			117	241.724	10.115	4.18	181.000	302.000	
2023-133	2 ACT5 PENTR	78 ABX PENTRA			29	238.071	6.590	2.77	179.000	298.000	
2023-133	3 ADVIA 120	64 ADVIA-120			17	207.500	13.555	6.53	156.000	259.000	
2023-133	4 AB- CELL-D	63 ABBOTT C-D			65	78.738	4.987	6.33	59.000	98.000	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-133	4 AB- CELL-D	100 ABB.RUBY			46	74.622	4.513	6.05	56.000	93.000	
2023-133	5 BECTON D	83 B.C600,800			60	84.508	2.424	2.87	63.000	106.000	
2023-133	6 MR Mindray	91 BC-5390			52	44.059	2.920	6.63	33.000	55.000	
2023-133	8 3 ST	31 ABBOTT C-D			32	190.625	17.175	9.01	143.000	238.000	
2023-133	8 3 ST	84 KX21,300			17	203.706	12.685	6.23	153.000	255.000	
2023-133	9 3 D	68 COULTERACT			16	284.188	14.427	5.08	213.000	355.000	
2023-133	10 DxH	94 BC DXH 500			52	249.942	15.354	6.14	187.000	312.000	
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2023-134	1 SYSMEX CBC	79 SYSMEX XS			151	375.946	18.186	4.84	282.000	470.000	
2023-134	1 SYSMEX CBC	87 SYSMEX XN			46	404.867	13.783	3.40	304.000	506.000	
2023-134	1 SYSMEX CBC	98 SYSMEX 350			124	401.664	13.873	3.45	301.000	502.000	
2023-134	2 ACT5 PENTR	73 COULTER AC			117	486.009	16.985	3.49	365.000	608.000	
2023-134	2 ACT5 PENTR	78 ABX PENTRA			29	482.536	14.836	3.07	362.000	603.000	
2023-134	3 ADVIA 120	64 ADVIA-120			17	429.176	25.585	5.96	322.000	536.000	
2023-134	4 AB- CELL-D	63 ABBOTT C-D			65	480.062	20.299	4.23	360.000	600.000	
2023-134	4 AB- CELL-D	100 ABB.RUBY			46	466.533	19.082	4.09	350.000	583.000	
2023-134	5 BECTON D	83 B.C600,800			60	449.000	12.766	2.84	337.000	561.000	
2023-134	6 MR Mindray	91 BC-5390			52	563.373	18.864	3.35	423.000	704.000	
2023-134	8 3 ST	31 ABBOTT C-D			32	371.750	36.194	9.74	279.000	465.000	
2023-134	8 3 ST	84 KX21,300			17	402.588	21.973	5.46	302.000	503.000	
2023-134	9 3 D	68 COULTERACT			16	606.563	35.572	5.86	455.000	758.000	
2023-134	10 DxH	94 BC DXH 500			52	487.118	23.038	4.73	365.000	609.000	
2023-135	1 SYSMEX CBC	79 SYSMEX XS			151	54.497	3.323	6.10	41.000	68.000	
2023-135	1 SYSMEX CBC	87 SYSMEX XN			46	49.413	2.901	5.87	37.000	62.000	
2023-135	1 SYSMEX CBC	98 SYSMEX 350			124	49.886	3.013	6.04	37.000	62.000	
2023-135	2 ACT5 PENTR	73 COULTER AC			117	69.574	6.427	9.24	52.000	87.000	
2023-135	2 ACT5 PENTR	78 ABX PENTRA			29	70.069	5.065	7.23	53.000	88.000	
2023-135	3 ADVIA 120	64 ADVIA-120			17	51.529	4.031	7.82	39.000	64.000	
2023-135	4 AB- CELL-D	63 ABBOTT C-D			65	79.169	4.941	6.24	59.000	99.000	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT NEW

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2023-135	4 AB- CELL-D	100 ABB.RUBY		46	74.889	4.508	6.02	56.000	94.000	
2023-135	5 BECTON D	83 B.C600,800		60	84.650	2.414	2.85	63.000	106.000	
2023-135	6 MR Mindray	91 BC-5390		52	43.627	2.883	6.61	33.000	55.000	
2023-135	8 3 ST	31 ABBOTT C-D		32	66.606	6.844	10.28	50.000	83.000	
2023-135	8 3 ST	84 KX21,300		17	63.294	2.824	4.46	47.000	79.000	
2023-135	9 3 D	68 COULTERAcT		16	73.563	5.385	7.32	55.000	92.000	
2023-135	10 DxH	94 BC DXH 500		52	78.231	5.823	7.44	59.000	98.000	

SPEC: 20 HEMATOLOGY SUB: 75 BODY FLUID GRP: 145 BODY FLUID CELL COU.

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
	2023-171			45	1097.093	114.419	10.43	754.000	1440.000	
	2023-172			45	0.390	0.488	125.00	0	2.000	
	2023-173			45	1102.286	101.614	9.22	797.000	1407.000	
	2023-174			45	0.465	0.659	141.77	0	2.000	
	2023-175			45	1083.488	104.256	9.62	771.000	1396.000	
160	BF WHITE CELL COUNT									
	2023-171			44	286.325	23.865	8.33	215.000	358.000	
	2023-172			44	4.524	1.829	40.43	0	10.000	

Participant Summary

MONTH 3 OF YEAR 2023

SPEC: 20 HEMATOLOGY		SUB: 75 BODY FLUID		GRP: 145 BODY FLUID CELL COU.						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
160 BF WHITE CELL COUNT		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
2023-173				44	290.415	32.045	11.03	194.000	387.000	
2023-174				44	4.881	1.592	32.61	0.000	10.000	
2023-175				44	289.643	30.975	10.69	197.000	383.000	

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFeree LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
083 ABO GROUP												
2023-181				004 004	004 004		004 - B	81	81	81	100	
2023-182				005 007	005 007		005 - AB	81	79	81	100	
2023-182				005 007	005 007		006 - A1B	81	2	81	100	
2023-183				001 003	001 003		001 - A	81	79	81	100	
2023-183				001 003	001 003		002 - A1	81	2	81	100	
2023-184				001 003	001 003		001 - A	81	80	81	100	
2023-184				001 003	001 003		002 - A1	81	1	81	100	
2023-185				008 008	008 008		008 - O	81	81	81	100	
126 RH TYPE												
2023-181							001 - Rh+	81	36	81	0	
2023-181							002 - Rh+ Du va.	81	4	81	0	
2023-181							003 - Rh-Du np	81	5	81	0	

Participant Summary

MONTH 3 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2023-181							004 - Rh-	81	36	81	0		
	126 RH TYPE												
2023-182				001 003	001 003		001 - Rh+	81	81	81	100		
2023-183				001 003	001 003		001 - Rh+	81	81	81	100		
2023-184				003 004	003 004		003 - Rh-Du np	81	9	81	100		
2023-184				003 004	003 004		004 - Rh-	81	72	81	100		
2023-185				003 004	003 004		003 - Rh-Du np	81	9	81	100		
2023-185				003 004	003 004		004 - Rh-	81	72	81	100		

Participant Summary

MONTH 3 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2023-191				002 002	002 002		001 - NEGATIVE	65	2	63	0
2023-191				002 002	002 002		002 - POSITIVE	65	63	63	100
2023-192				001 001	001 001		001 - NEGATIVE	65	65	65	100
2023-193				001 001	001 001		001 - NEGATIVE	65	65	65	100
2023-194				002 002	002 002		002 - POSITIVE	65	65	65	100
2023-195				001 001	001 001		001 - NEGATIVE	65	65	65	100

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2023-271	001 Poly AHG	001 Gel			002 004		003 - 3+	11	8	11	100
2023-271	001 Poly AHG	001 Gel			002 004		004 - 4+	11	3	11	100
2023-271	001 Poly AHG	002 Tubo			002 004		002 - 2+	33	25	33	100
2023-271	001 Poly AHG	002 Tubo			002 004		003 - 3+	33	7	33	100
2023-271	001 Poly AHG	002 Tubo			002 004		004 - 4+	33	1	33	100
2023-271	002 MnoAnt-IgG	001 Gel				002 004	003 - 3+	12	9	12	100
2023-271	002 MnoAnt-IgG	001 Gel				002 004	004 - 4+	12	3	12	100
2023-271	002 MnoAnt-IgG	002 Tubo				002 004	002 - 2+	16	12	16	100
2023-271	002 MnoAnt-IgG	002 Tubo				002 004	003 - 3+	16	3	16	100
2023-271	002 MnoAnt-IgG	002 Tubo				002 004	004 - 4+	16	1	16	100

Participant Summary

MONTH 3 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2023-272	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100
2023-272	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	33	33	33	100
2023-272	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2023-272	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	14	14	14	100
2023-273	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100
2023-273	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	33	33	33	100
2023-273	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2023-273	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	14	14	14	100
2023-274	001 Poly AHG	001 Gel			002 004		003 - 3+	11	8	11	100
2023-274	001 Poly AHG	001 Gel			002 004		004 - 4+	11	3	11	100
2023-274	001 Poly AHG	002 Tubo			002 004		002 - 2+	33	23	33	100
2023-274	001 Poly AHG	002 Tubo			002 004		003 - 3+	33	8	33	100
2023-274	001 Poly AHG	002 Tubo			002 004		004 - 4+	33	2	33	100
2023-274	002 MnoAnt-IgG	001 Gel				002 004	003 - 3+	12	9	12	100
2023-274	002 MnoAnt-IgG	001 Gel				002 004	004 - 4+	12	3	12	100
2023-274	002 MnoAnt-IgG	002 Tubo		002 004			002 - 2+	16	13	16	100
2023-274	002 MnoAnt-IgG	002 Tubo		002 004			003 - 3+	16	2	16	100
2023-274	002 MnoAnt-IgG	002 Tubo		002 004			004 - 4+	16	1	16	100
2023-275	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100
2023-275	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	33	33	33	100
2023-275	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2023-275	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	14	14	14	100

Participant Summary
MONTH 3 OF YEAR 2023

SPE: 30 IMMUNOHEMATOLOGY												SUB: 100 COMPATIBILITY TESTING				GRP: 170 COMPATIBILITY TEST.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score								
085 COMPATIBIL. TESTING																			
2023-201				002	002	002	002	002 - INCOMPATIB	39	39	39	100							
2023-202				001	001	001	001	001 - COMPATIBLE	39	39	39	100							
2023-203				001	001	001	001	001 - COMPATIBLE	39	39	39	100							
2023-204				001	001	001	001	001 - COMPATIBLE	39	39	39	100							
2023-205				002	002	002	002	002 - INCOMPATIB	39	39	39	100							