

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE				SYST/INST		REAGENTS		NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD						LABS	VALUE	DEVIATION	CV	LOW	HIGH	
001 ALAN AMINO	ALT/SGPT			CLIA TEST ID #: 255		EVALUATION CRITERIA: TARGET VALUE 20 %							
2021-391	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	19	18.316	1.379	7.53	15.000	22.000	
2021-391	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	31.364	0.979	3.12	25.000	38.000	
2021-391	11	MOD. IFCC	38	SIEMENS DI	143	SIEMENS	77	24.066	2.148	8.93	19.000	29.000	
2021-391	11	MOD. IFCC	184	INTEGRA	93	ROCHE	29	17.241	1.005	5.83	14.000	21.000	
2021-391	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	15.640	0.933	5.97	13.000	19.000	
2021-391	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	19.786	2.093	10.58	16.000	24.000	
2021-391	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	17.250	3.152	18.27	14.000	21.000	
2021-391	11	MOD. IFCC	203	ROCHE	93	ROCHE	28	17.259	0.843	4.88	14.000	21.000	
2021-391	13	ALA &aKETO	165	VITROS	155	ALTV	129	14.444	0.956	6.62	12.000	17.000	
2021-391	13	ALA &aKETO	165	VITROS	155	ALTV	129	15.039	0.913	6.07	12.000	18.000	
2021-392	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	19	261.789	8.141	3.11	209.000	314.000	
2021-392	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	251.545	3.775	1.50	201.000	302.000	
2021-392	11	MOD. IFCC	38	SIEMENS DI	143	SIEMENS	77	275.184	5.988	2.18	220.000	330.000	
2021-392	11	MOD. IFCC	184	INTEGRA	93	ROCHE	29	254.552	8.101	3.18	204.000	305.000	
2021-392	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	229.731	8.188	3.56	184.000	276.000	
2021-392	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	286.630	19.435	6.78	229.000	344.000	
2021-392	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	294.375	6.744	2.29	236.000	353.000	
2021-392	11	MOD. IFCC	203	ROCHE	93	ROCHE	28	253.000	7.562	2.99	202.000	304.000	
2021-392	13	ALA &aKETO	165	VITROS	155	ALTV	129	249.441	6.327	2.54	200.000	299.000	
2021-392	13	ALA &aKETO	165	VITROS	155	ALTV	129	246.889	6.279	2.54	198.000	296.000	
2021-393	5	ENZ. 37 C	197	ARCHITECT	1	ABBOTT	19	121.632	3.542	2.91	97.000	146.000	
2021-393	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	126.091	1.676	1.33	101.000	151.000	
2021-393	11	MOD. IFCC	38	SIEMENS DI	143	SIEMENS	77	131.733	3.043	2.31	105.000	158.000	
2021-393	11	MOD. IFCC	184	INTEGRA	93	ROCHE	29	118.828	2.913	2.45	95.000	143.000	
2021-393	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	107.385	3.498	3.26	86.000	129.000	
2021-393	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	134.679	8.233	6.11	108.000	162.000	
2021-393	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	133.500	2.915	2.18	107.000	160.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2021-393	11 MOD. IFCC	203	ROCHE	93	ROCHE	28	117.667	3.662	3.11	94.000	141.000		
2021-393	13 ALA &aKETO	165	VITROS	155	ALTV	129	116.889	2.961	2.53	94.000	140.000		
2021-393	13 ALA &aKETO	165	VITROS	155	ALTV	129	119.047	3.271	2.75	95.000	143.000		
001	ALAN AMINO ALT/SGPT	CLIA TEST ID #: 255		EVALUATION CRITERIA: TARGET VALUE 20 %									
2021-394	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	19	296.105	7.873	2.66	237.000	355.000		
2021-394	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	284.636	4.457	1.57	228.000	342.000		
2021-394	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	77	310.760	7.447	2.40	249.000	373.000		
2021-394	11 MOD. IFCC	184	INTEGRA	93	ROCHE	29	288.621	9.268	3.21	231.000	346.000		
2021-394	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	261.615	7.391	2.82	209.000	314.000		
2021-394	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	328.179	28.268	8.61	263.000	394.000		
2021-394	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	336.250	9.270	2.76	269.000	404.000		
2021-394	11 MOD. IFCC	203	ROCHE	93	ROCHE	28	288.148	6.381	2.21	231.000	346.000		
2021-394	13 ALA &aKETO	165	VITROS	155	ALTV	129	281.914	7.158	2.54	226.000	338.000		
2021-394	13 ALA &aKETO	165	VITROS	155	ALTV	129	277.556	7.890	2.84	222.000	333.000		
2021-395	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	19	404.947	13.205	3.26	324.000	486.000		
2021-395	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	11	382.636	4.773	1.25	306.000	459.000		
2021-395	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	77	421.408	10.425	2.47	337.000	506.000		
2021-395	11 MOD. IFCC	184	INTEGRA	93	ROCHE	29	392.069	11.614	2.96	314.000	470.000		
2021-395	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	360.038	11.257	3.13	288.000	432.000		
2021-395	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	444.321	39.453	8.88	355.000	533.000		
2021-395	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	455.500	27.120	5.95	364.000	547.000		
2021-395	11 MOD. IFCC	203	ROCHE	93	ROCHE	28	394.889	9.931	2.51	316.000	474.000		
2021-395	13 ALA &aKETO	165	VITROS	155	ALTV	129	377.667	10.583	2.80	302.000	453.000		
2021-395	13 ALA &aKETO	165	VITROS	155	ALTV	129	382.875	13.710	3.58	306.000	459.000		
002	ALBUMIN	CLIA TEST ID #: 265		EVALUATION CRITERIA: TARGET VALUE 10 %									
2021-391	3 DYE BI-B.G	165	VITROS	63	KODAK	148	1.854	0.060	3.22	1.700	2.000		
2021-391	3 DYE BI-B.G	184	INTEGRA	93	ROCHE	37	1.954	0.079	4.05	1.800	2.100		
2021-391	3 DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	26	1.854	0.063	3.42	1.700	2.000		
2021-391	3 DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	1.867	0.047	2.53	1.700	2.100		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-391	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	1.977	0.096	4.83	1.800	2.200		
2021-391	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	21	1.810	0.043	2.35	1.600	2.000		
2021-391	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	11	2.027	0.075	3.70	1.800	2.200		
2021-391	3 DYE BI-B.G	203 ROCHE	93 ROCHE	35	1.968	0.063	3.20	1.800	2.200		
2021-391	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	1.686	0.065	3.86	1.500	1.900		
2021-391	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	12	1.692	0.064	3.78	1.500	1.900		
2021-391	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	10	1.750	0.102	5.86	1.600	1.900		
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2021-392	3 DYE BI-B.G	165 VITROS	63 KODAK	148	4.188	0.130	3.10	3.800	4.600		
2021-392	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	4.161	0.098	2.35	3.700	4.600		
2021-392	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	26	3.940	0.069	1.76	3.500	4.300		
2021-392	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	4.032	0.135	3.36	3.600	4.400		
2021-392	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	4.300	0.082	1.90	3.900	4.700		
2021-392	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	21	3.895	0.065	1.68	3.500	4.300		
2021-392	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	11	4.100	0.085	2.08	3.700	4.500		
2021-392	3 DYE BI-B.G	203 ROCHE	93 ROCHE	35	4.135	0.087	2.11	3.700	4.500		
2021-392	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	3.808	0.110	2.88	3.400	4.200		
2021-392	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	12	3.858	0.138	3.58	3.500	4.200		
2021-392	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	10	3.880	0.075	1.93	3.500	4.300		
2021-393	3 DYE BI-B.G	165 VITROS	63 KODAK	148	2.797	0.081	2.89	2.500	3.100		
2021-393	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	2.897	0.083	2.87	2.600	3.200		
2021-393	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	26	2.750	0.069	2.52	2.500	3.000		
2021-393	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	2.967	0.236	7.95	2.700	3.300		
2021-393	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	2.845	0.119	4.17	2.600	3.100		
2021-393	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	21	2.710	0.053	1.94	2.400	3.000		
2021-393	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	11	2.900	0.085	2.94	2.600	3.200		
2021-393	3 DYE BI-B.G	203 ROCHE	93 ROCHE	35	2.903	0.079	2.71	2.600	3.200		
2021-393	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	2.574	0.074	2.88	2.300	2.800		
2021-393	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	12	2.600	0.091	3.51	2.300	2.900		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-393	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	10	2.710	0.070	2.58	2.400	3.000		
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2021-394	3 DYE BI-B.G	165 VITROS	63 KODAK	148	4.564	0.138	3.02	4.100	5.000		
2021-394	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	4.428	0.126	2.85	4.000	4.900		
2021-394	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	26	4.200	0.083	1.98	3.800	4.600		
2021-394	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	4.313	0.179	4.15	3.900	4.700		
2021-394	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	4.400	0.141	3.21	4.000	4.800		
2021-394	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	21	4.152	0.085	2.05	3.700	4.600		
2021-394	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	11	4.300	0.165	3.84	3.900	4.700		
2021-394	3 DYE BI-B.G	203 ROCHE	93 ROCHE	35	4.391	0.113	2.57	4.000	4.800		
2021-394	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	4.072	0.149	3.67	3.700	4.500		
2021-394	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	12	4.100	0.147	3.59	3.700	4.500		
2021-394	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	10	4.150	0.157	3.77	3.700	4.600		
2021-395	3 DYE BI-B.G	165 VITROS	63 KODAK	148	5.728	0.202	3.53	5.200	6.300		
2021-395	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	37	5.294	0.131	2.48	4.800	5.800		
2021-395	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	26	5.092	0.127	2.49	4.600	5.600		
2021-395	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	5.400	0.141	2.62	4.900	5.900		
2021-395	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX	31	5.161	0.230	4.45	4.600	5.700		
2021-395	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	21	5.010	0.092	1.84	4.500	5.500		
2021-395	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX	11	5.155	0.227	4.41	4.600	5.700		
2021-395	3 DYE BI-B.G	203 ROCHE	93 ROCHE	35	5.329	0.099	1.85	4.800	5.900		
2021-395	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	4.996	0.181	3.63	4.500	5.500		
2021-395	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	12	5.025	0.242	4.82	4.500	5.500		
2021-395	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL	10	5.130	0.078	1.52	4.600	5.600		
003 ALKALINE PHOSPHATE		CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-391	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	26.097	0.995	3.81	18.000	34.000		
2021-391	3 COLORIMET.	203 ROCHE	93 ROCHE	24	25.304	1.428	5.64	18.000	33.000		
2021-391	6 KINETIC 37	165 VITROS	63 KODAK	148	39.582	2.053	5.19	28.000	51.000		
2021-391	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	22.500	2.419	10.75	16.000	29.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-391	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	21.000	1.265	6.02	15.000	27.000		
2021-391	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	23.000	1.414	6.15	16.000	30.000		
2021-391	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	21.519	2.846	13.23	15.000	28.000		
2021-391	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	23	23.304	2.628	11.28	16.000	30.000		
2021-391	11 Modified I	38 SIEMENS DI	143 SIEMENS	76	26.880	3.570	13.28	19.000	35.000		
2021-391	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	12	24.333	1.700	6.98	17.000	32.000		
003 ALKALINE PHOSPHATE		CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-392	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	218.065	7.107	3.26	153.000	283.000		
2021-392	3 COLORIMET.	203 ROCHE	93 ROCHE	24	212.042	10.922	5.15	148.000	276.000		
2021-392	6 KINETIC 37	165 VITROS	63 KODAK	148	170.386	7.592	4.46	119.000	222.000		
2021-392	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	209.900	9.093	4.33	147.000	273.000		
2021-392	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	192.000	11.314	5.89	134.000	250.000		
2021-392	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	181.300	7.072	3.90	127.000	236.000		
2021-392	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	182.185	18.533	10.17	128.000	237.000		
2021-392	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	23	190.478	18.512	9.72	133.000	248.000		
2021-392	11 Modified I	38 SIEMENS DI	143 SIEMENS	76	229.173	8.475	3.70	160.000	298.000		
2021-392	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	12	209.083	10.412	4.98	146.000	272.000		
2021-393	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	107.033	4.004	3.74	75.000	139.000		
2021-393	3 COLORIMET.	203 ROCHE	93 ROCHE	24	104.208	5.802	5.57	73.000	135.000		
2021-393	6 KINETIC 37	165 VITROS	63 KODAK	148	113.441	5.084	4.48	79.000	147.000		
2021-393	6 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	102.100	5.319	5.21	71.000	133.000		
2021-393	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	89.400	3.555	3.98	63.000	116.000		
2021-393	6 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	95.333	7.587	7.96	67.000	124.000		
2021-393	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	89.000	9.745	10.95	62.000	116.000		
2021-393	9 ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	23	95.739	8.994	9.39	67.000	124.000		
2021-393	11 Modified I	38 SIEMENS DI	143 SIEMENS	76	111.827	5.686	5.08	78.000	145.000		
2021-393	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	12	101.917	4.924	4.83	71.000	132.000		
2021-394	3 COLORIMET.	184 INTEGRA	93 ROCHE	31	248.871	9.591	3.85	174.000	324.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-394	3	COLORIMET.	203 ROCHE	93 ROCHE	24	242.583	12.890	5.31	170.000	315.000	
2021-394	6	KINETIC 37	165 VITROS	63 KODAK	148	180.582	8.073	4.47	126.000	235.000	
2021-394	6	KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	243.500	29.904	12.28	170.000	317.000	
2021-394	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	221.667	13.888	6.27	155.000	288.000	
2021-394	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	208.100	8.360	4.02	146.000	271.000	
2021-394	8	BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	207.778	21.023	10.12	145.000	270.000	
2021-394	9	ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	23	216.043	20.750	9.60	151.000	281.000	
2021-394	11	Modified I	38 SIEMENS DI	143 SIEMENS	76	262.000	10.182	3.89	183.000	341.000	
2021-394	12	P NITRO PH	197 ARCHITECT	1 ABBOTT	12	237.083	11.800	4.98	166.000	308.000	
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-395	3	COLORIMET.	184 INTEGRA	93 ROCHE	31	340.032	11.128	3.27	238.000	442.000	
2021-395	3	COLORIMET.	203 ROCHE	93 ROCHE	24	328.333	15.537	4.73	230.000	427.000	
2021-395	6	KINETIC 37	165 VITROS	63 KODAK	148	195.367	10.280	5.26	137.000	254.000	
2021-395	6	KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	316.600	27.958	8.83	222.000	412.000	
2021-395	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	282.000	14.156	5.02	197.000	367.000	
2021-395	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	298.667	18.696	6.26	209.000	388.000	
2021-395	8	BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	284.444	31.540	11.09	199.000	370.000	
2021-395	9	ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	23	291.652	28.170	9.66	204.000	379.000	
2021-395	11	Modified I	38 SIEMENS DI	143 SIEMENS	76	357.716	12.352	3.45	250.000	465.000	
2021-395	12	P NITRO PH	197 ARCHITECT	1 ABBOTT	12	322.500	16.383	5.08	226.000	419.000	
004	AMYLASE	CLIA TEST ID #: 285									
2021-391	3	ENZYM. 37C	165 VITROS	63 KODAK	85	30.000	0.000	0.00	21.000	39.000	
2021-391	3	ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	11	25.909	1.164	4.49	18.000	34.000	
2021-391	4	HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	25	23.720	1.040	4.38	17.000	31.000	
2021-391	9	ENZ/COL.37	184 INTEGRA	93 ROCHE	10	26.800	0.748	2.79	19.000	35.000	
2021-391	9	ENZ/COL.37	203 ROCHE	93 ROCHE	23	25.273	1.135	4.49	18.000	33.000	
2021-391	10	CNPG3	186 BECK. COUL	76 AU-BECKMAN	18	17.889	1.487	8.31	13.000	23.000	
2021-392	3	ENZYM. 37C	165 VITROS	63 KODAK	85	143.071	5.444	3.81	100.000	186.000	
2021-392	3	ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	11	195.182	5.670	2.90	137.000	254.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-392	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	25	189.520	4.355	2.30	133.000	246.000		
2021-392	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	10	194.800	4.261	2.19	136.000	253.000		
2021-392	9 ENZ/COL.37	203 ROCHE	93 ROCHE	23	188.826	4.300	2.28	132.000	245.000		
2021-392	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	18	148.333	7.557	5.09	104.000	193.000		
004	AMYLASE	CLIA TEST ID #: 285	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-393	3 ENZYM. 37C	165 VITROS	63 KODAK	85	69.294	7.273	10.50	49.000	90.000		
2021-393	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	11	98.273	1.483	1.51	69.000	128.000		
2021-393	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	25	92.880	2.438	2.63	65.000	121.000		
2021-393	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	10	96.200	1.661	1.73	67.000	125.000		
2021-393	9 ENZ/COL.37	203 ROCHE	93 ROCHE	23	93.227	2.762	2.96	65.000	121.000		
2021-393	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	18	73.111	4.605	6.30	51.000	95.000		
2021-394	3 ENZYM. 37C	165 VITROS	63 KODAK	85	159.541	6.129	3.84	112.000	207.000		
2021-394	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	11	223.000	5.205	2.33	156.000	290.000		
2021-394	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	25	214.360	4.638	2.16	150.000	279.000		
2021-394	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	10	221.400	5.314	2.40	155.000	288.000		
2021-394	9 ENZ/COL.37	203 ROCHE	93 ROCHE	23	214.739	4.067	1.89	150.000	279.000		
2021-394	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	18	167.444	10.468	6.25	117.000	218.000		
2021-395	3 ENZYM. 37C	165 VITROS	63 KODAK	85	208.624	9.097	4.36	146.000	271.000		
2021-395	3 ENZYM. 37C	215 B.C DXC	134 BECKM.COUL	11	308.000	7.198	2.34	216.000	400.000		
2021-395	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	25	291.840	6.739	2.31	204.000	379.000		
2021-395	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	10	299.900	5.049	1.68	210.000	390.000		
2021-395	9 ENZ/COL.37	203 ROCHE	93 ROCHE	23	294.043	5.393	1.83	206.000	382.000		
2021-395	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	18	228.056	11.801	5.17	160.000	296.000		
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-391	1 ASPAR&KETO	165 VITROS	63 KODAK	144	45.818	1.387	3.03	37.000	55.000		
2021-391	1 ASPAR&KETO	165 VITROS	63 KODAK	144	43.000	4.637	10.78	34.000	52.000		
2021-391	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	10	38.200	2.441	6.39	31.000	46.000		
2021-391	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	81	41.222	1.721	4.18	33.000	49.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-391	5	ENZYM.37C	212	S. D. EXL	143	SIEMENS	11	41.182	1.336	3.24	33.000	49.000
2021-391	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	46.300	1.345	2.91	37.000	56.000
2021-391	11	MOD. IFCC	184	INTEGRA	93	ROCHE	31	41.226	1.861	4.52	33.000	49.000
2021-391	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	27	35.296	2.033	5.76	28.000	42.000
2021-391	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	43.643	3.254	7.46	35.000	52.000
2021-391	11	MOD. IFCC	203	ROCHE	93	ROCHE	26	40.040	1.777	4.44	32.000	48.000
2021-391	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	13	39.538	1.151	2.91	32.000	47.000
005	ASP. AMINO AST/SGOT	CLIA	TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-392	1	ASPAR&KETO	165	VITROS	63	KODAK	144	331.500	39.740	11.99	265.000	398.000
2021-392	1	ASPAR&KETO	165	VITROS	63	KODAK	144	359.901	10.683	2.97	288.000	432.000
2021-392	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	10	279.000	17.527	6.28	223.000	335.000
2021-392	5	ENZYM.37C	38	SIEMENS DI	143	SIEMENS	81	281.575	7.762	2.76	225.000	338.000
2021-392	5	ENZYM.37C	212	S. D. EXL	143	SIEMENS	11	280.818	6.847	2.44	225.000	337.000
2021-392	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	267.900	5.665	2.11	214.000	321.000
2021-392	11	MOD. IFCC	184	INTEGRA	93	ROCHE	31	289.935	9.211	3.18	232.000	348.000
2021-392	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	27	247.296	13.805	5.58	198.000	297.000
2021-392	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	308.679	18.363	5.95	247.000	370.000
2021-392	11	MOD. IFCC	203	ROCHE	93	ROCHE	26	284.577	9.881	3.47	228.000	341.000
2021-392	13	NADH wo P5	197	ARCHITECT	1	ABBOTT	13	273.538	5.500	2.01	219.000	328.000
2021-393	1	ASPAR&KETO	165	VITROS	63	KODAK	144	164.901	4.961	3.01	132.000	198.000
2021-393	1	ASPAR&KETO	165	VITROS	63	KODAK	144	162.500	5.025	3.09	130.000	195.000
2021-393	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	10	145.700	5.442	3.73	117.000	175.000
2021-393	5	ENZYM.37C	38	SIEMENS DI	143	SIEMENS	81	142.413	4.774	3.35	114.000	171.000
2021-393	5	ENZYM.37C	212	S. D. EXL	143	SIEMENS	11	141.818	4.914	3.47	113.000	170.000
2021-393	10	KINETIC 37	215	B.C DXC	134	BECKM.COUL	10	140.900	2.809	1.99	113.000	169.000
2021-393	11	MOD. IFCC	184	INTEGRA	93	ROCHE	31	147.258	4.355	2.96	118.000	177.000
2021-393	11	MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	27	125.346	4.945	3.95	100.000	150.000
2021-393	11	MOD. IFCC	191	RX-DAYTONA	126	RANDOX	28	155.429	7.277	4.68	124.000	187.000
2021-393	11	MOD. IFCC	203	ROCHE	93	ROCHE	26	144.480	4.768	3.30	116.000	173.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2021-393	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	13	139.154	2.957	2.13	111.000	167.000
005	ASP. AMINO AST/SGOT		CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-394	1	ASPAR&KETO		165	VITROS	63	KODAK	144	411.000	13.748	3.34	329.000	493.000
2021-394	1	ASPAR&KETO		165	VITROS	63	KODAK	144	414.986	16.001	3.86	332.000	498.000
2021-394	1	ASPAR&KETO		194	BIOLIS-24S	154	(IDG)	10	308.200	20.400	6.62	247.000	370.000
2021-394	5	ENZYMA.37C		38	SIEMENS DI	143	SIEMENS	81	317.734	10.559	3.32	254.000	381.000
2021-394	5	ENZYMA.37C		212	S. D. EXL	143	SIEMENS	11	320.909	7.179	2.24	257.000	385.000
2021-394	10	KINETIC 37		215	B.C DXC	134	BECKM.COUL	10	303.000	6.325	2.09	242.000	364.000
2021-394	11	MOD. IFCC		184	INTEGRA	93	ROCHE	31	326.700	10.090	3.09	261.000	392.000
2021-394	11	MOD. IFCC		186	BECK. COUL	76	AU-BECKMAN	27	277.556	14.143	5.10	222.000	333.000
2021-394	11	MOD. IFCC		191	RX-DAYTONA	126	RANDOX	28	342.429	17.441	5.09	274.000	411.000
2021-394	11	MOD. IFCC		203	ROCHE	93	ROCHE	26	318.115	10.906	3.43	254.000	382.000
2021-394	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	13	307.846	6.024	1.96	246.000	369.000
2021-395	1	ASPAR&KETO		165	VITROS	63	KODAK	144	602.648	29.403	4.88	482.000	723.000
2021-395	1	ASPAR&KETO		165	VITROS	63	KODAK	144	605.000	11.424	1.89	484.000	726.000
2021-395	1	ASPAR&KETO		194	BIOLIS-24S	154	(IDG)	10	390.500	27.746	7.11	312.000	469.000
2021-395	5	ENZYMA.37C		38	SIEMENS DI	143	SIEMENS	81	428.450	12.843	3.00	343.000	514.000
2021-395	5	ENZYMA.37C		212	S. D. EXL	143	SIEMENS	11	431.545	14.481	3.36	345.000	518.000
2021-395	10	KINETIC 37		215	B.C DXC	134	BECKM.COUL	10	406.500	17.101	4.21	325.000	488.000
2021-395	11	MOD. IFCC		184	INTEGRA	93	ROCHE	31	433.839	15.390	3.55	347.000	521.000
2021-395	11	MOD. IFCC		186	BECK. COUL	76	AU-BECKMAN	27	371.808	15.799	4.25	297.000	446.000
2021-395	11	MOD. IFCC		191	RX-DAYTONA	126	RANDOX	28	457.286	20.933	4.58	366.000	549.000
2021-395	11	MOD. IFCC		203	ROCHE	93	ROCHE	26	428.885	14.672	3.42	343.000	515.000
2021-395	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	13	414.538	7.929	1.91	332.000	497.000
006	BILIRUBIN, TOTAL		CLIA TEST ID #: 305										
2021-391	4	J-G WITH B		38	SIEMENS DI	143	SIEMENS	79	0.100	0.018	18.00	0.080	0.120
2021-391	4	J-G WITH B		186	BECK. COUL	76	AU-BECKMAN	24	0.145	0.037	25.49	0.120	0.170
2021-391	4	J-G WITH B		191	RX-DAYTONA	126	RANDOX	21	0.149	0.061	41.19	0.120	0.180
2021-391	4	J-G WITH B		212	S. D. EXL	143	SIEMENS	10	0.127	0.067	52.71	0.100	0.150

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-391	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	11	0.269	0.064	23.85	0.220	0.320		
2021-391	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	0.108	0.028	25.72	0.090	0.130		
2021-391	13 AZOBIL.	184 INTEGRA	93 ROCHE	20	0.115	0.121	104.96	0.090	0.140		
2021-391	13 AZOBIL.	203 ROCHE	93 ROCHE	24	0.135	0.038	28.29	0.110	0.160		
2021-391	18 COLORIMETR	203 ROCHE	93 ROCHE	10	0.147	0.030	20.19	0.120	0.180		
2021-391	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	12	0.102	0.004	3.67	0.080	0.120		
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305	EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-392	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	3.303	0.138	4.17	2.640	3.960		
2021-392	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	24	3.288	0.096	2.91	2.630	3.950		
2021-392	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	21	3.390	0.257	7.58	2.710	4.070		
2021-392	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	10	3.408	0.326	9.57	2.730	4.090		
2021-392	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	11	3.630	0.219	6.04	2.900	4.360		
2021-392	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	3.095	0.161	5.21	2.480	3.710		
2021-392	13 AZOBIL.	184 INTEGRA	93 ROCHE	20	2.943	0.088	2.98	2.350	3.530		
2021-392	13 AZOBIL.	203 ROCHE	93 ROCHE	24	2.855	0.058	2.04	2.280	3.430		
2021-392	18 COLORIMETR	203 ROCHE	93 ROCHE	10	2.975	0.152	5.10	2.380	3.570		
2021-392	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	12	3.107	0.142	4.58	2.490	3.730		
2021-393	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	1.414	0.088	6.21	1.130	1.700		
2021-393	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	24	1.515	0.046	3.01	1.210	1.820		
2021-393	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	21	1.482	0.125	8.47	1.190	1.780		
2021-393	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	10	1.469	0.141	9.62	1.180	1.760		
2021-393	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	11	1.688	0.106	6.26	1.350	2.030		
2021-393	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	1.324	0.079	6.00	1.060	1.590		
2021-393	13 AZOBIL.	184 INTEGRA	93 ROCHE	20	1.254	0.066	5.23	1.000	1.500		
2021-393	13 AZOBIL.	203 ROCHE	93 ROCHE	24	1.222	0.045	3.69	0.980	1.470		
2021-393	18 COLORIMETR	203 ROCHE	93 ROCHE	10	1.272	0.094	7.38	1.020	1.530		
2021-393	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	12	1.393	0.108	7.78	1.110	1.670		
2021-394	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	3.771	0.186	4.92	3.020	4.520		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-394	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN		24	3.728	0.088	2.37	2.980	4.470	
2021-394	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX		21	3.897	0.310	7.94	3.120	4.680	
2021-394	4 J-G WITH B	212 S. D. EXL	143 SIEMENS		10	3.899	0.358	9.18	3.120	4.680	
2021-394	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL		11	4.184	0.179	4.28	3.350	5.020	
2021-394	7 DIAZO-DYPH	165 VITROS	63 KODAK		143	3.514	0.176	5.00	2.810	4.220	
2021-394	13 AZOBIL.	184 INTEGRA	93 ROCHE		20	3.391	0.134	3.95	2.710	4.070	
2021-394	13 AZOBIL.	203 ROCHE	93 ROCHE		24	3.306	0.080	2.42	2.640	3.970	
2021-394	18 COLORIMETR	203 ROCHE	93 ROCHE		10	3.394	0.177	5.22	2.720	4.070	
2021-394	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT		12	3.563	0.148	4.17	2.850	4.280	
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305	EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-395	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS		79	5.281	0.222	4.19	4.220	6.340	
2021-395	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN		24	5.168	0.152	2.94	4.130	6.200	
2021-395	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX		21	5.601	0.422	7.53	4.480	6.720	
2021-395	4 J-G WITH B	212 S. D. EXL	143 SIEMENS		10	5.433	0.527	9.70	4.350	6.520	
2021-395	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL		11	5.841	0.255	4.37	4.670	7.010	
2021-395	7 DIAZO-DYPH	165 VITROS	63 KODAK		143	4.858	0.248	5.11	3.890	5.830	
2021-395	13 AZOBIL.	184 INTEGRA	93 ROCHE		20	4.683	0.161	3.43	3.750	5.620	
2021-395	13 AZOBIL.	203 ROCHE	93 ROCHE		24	4.718	0.147	3.12	3.770	5.660	
2021-395	18 COLORIMETR	203 ROCHE	93 ROCHE		10	4.775	0.246	5.16	3.820	5.730	
2021-395	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT		12	4.978	0.211	4.23	3.980	5.970	
010 CALCIUM, TOTAL		CLIA TEST ID #: 345	EVALUATION CRITERIA: TARGET VALUE 1.00								
2021-391	3 ARSEN.-III	165 VITROS	63 KODAK		162	5.696	0.147	2.58	4.700	6.700	
2021-391	3 ARSEN.-III	165 VITROS	63 KODAK		162	6.157	0.206	3.35	5.200	7.200	
2021-391	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN		29	5.710	0.121	2.12	4.700	6.700	
2021-391	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX		33	6.079	0.213	3.50	5.100	7.100	
2021-391	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT		14	5.621	0.161	2.87	4.600	6.600	
2021-391	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX		11	6.136	0.284	4.62	5.100	7.100	
2021-391	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS		82	5.743	0.237	4.12	4.700	6.700	
2021-391	8 CRESOLPHT.	184 INTEGRA	93 ROCHE		18	5.794	0.118	2.03	4.800	6.800	
2021-391	8 CRESOLPHT.	203 ROCHE	93 ROCHE		10	5.790	0.284	4.91	4.800	6.800	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-391	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	12	5.692	0.335	5.89	4.700	6.700
2021-391	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	11	5.764	0.123	2.13	4.800	6.800
2021-391	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	18	5.839	0.177	3.03	4.800	6.800
2021-391	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	24	5.896	0.137	2.32	4.900	6.900
010	CALCIUM, TOTAL		CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE 1.00							
2021-392	3	ARSEN.-III	165	VITROS	63	KODAK	162	10.300	0.278	2.70	9.300	11.300
2021-392	3	ARSEN.-III	165	VITROS	63	KODAK	162	9.830	0.210	2.13	8.800	10.800
2021-392	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	10.034	0.251	2.50	9.000	11.000
2021-392	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	33	10.267	0.329	3.21	9.300	11.300
2021-392	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	14	9.650	0.329	3.41	8.700	10.700
2021-392	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	11	10.300	0.416	4.03	9.300	11.300
2021-392	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	82	9.662	0.288	2.98	8.700	10.700
2021-392	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	9.961	0.189	1.90	9.000	11.000
2021-392	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	10.050	0.304	3.03	9.100	11.100
2021-392	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	12	9.625	0.295	3.06	8.600	10.600
2021-392	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	11	9.564	0.227	2.37	8.600	10.600
2021-392	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	18	9.956	0.214	2.15	9.000	11.000
2021-392	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	24	10.050	0.187	1.86	9.100	11.100
2021-393	3	ARSEN.-III	165	VITROS	63	KODAK	162	7.453	0.181	2.43	6.500	8.500
2021-393	3	ARSEN.-III	165	VITROS	63	KODAK	162	8.014	0.181	2.25	7.000	9.000
2021-393	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	7.514	0.164	2.18	6.500	8.500
2021-393	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	33	7.822	0.247	3.16	6.800	8.800
2021-393	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	14	7.257	0.238	3.28	6.300	8.300
2021-393	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	11	7.755	0.331	4.27	6.800	8.800
2021-393	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	82	7.383	0.225	3.05	6.400	8.400
2021-393	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	7.482	0.129	1.73	6.500	8.500
2021-393	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	7.570	0.241	3.18	6.600	8.600
2021-393	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	12	7.233	0.295	4.08	6.200	8.200
2021-393	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	11	7.436	0.137	1.84	6.400	8.400

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
2021-393	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	7.578	0.204	2.70	6.600	8.600				
2021-393	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	24	7.604	0.186	2.44	6.600	8.600				
010	CALCIUM, TOTAL	CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE 1.00									
2021-394	3 ARSEN.-III	165 VITROS	63 KODAK	162	10.671	0.249	2.33	9.700	11.700				
2021-394	3 ARSEN.-III	165 VITROS	63 KODAK	162	10.334	0.226	2.19	9.300	11.300				
2021-394	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	29	10.590	0.229	2.17	9.600	11.600				
2021-394	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX	33	10.847	0.313	2.89	9.800	11.800				
2021-394	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	14	10.100	0.215	2.13	9.100	11.100				
2021-394	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX	11	10.655	0.434	4.07	9.700	11.700				
2021-394	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	82	10.244	0.273	2.67	9.200	11.200				
2021-394	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	18	10.517	0.265	2.52	9.500	11.500				
2021-394	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	10.630	0.190	1.79	9.600	11.600				
2021-394	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	12	10.175	0.268	2.63	9.200	11.200				
2021-394	14 ELECTRODE	215 B.C DXC	134 BECKM.COUL	11	10.155	0.239	2.35	9.200	11.200				
2021-394	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	10.500	0.306	2.91	9.500	11.500				
2021-394	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	24	10.554	0.204	1.93	9.600	11.600				
2021-395	3 ARSEN.-III	165 VITROS	63 KODAK	162	11.935	0.268	2.25	10.900	12.900				
2021-395	3 ARSEN.-III	165 VITROS	63 KODAK	162	12.200	0.346	2.84	11.200	13.200				
2021-395	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	29	12.648	0.311	2.46	11.600	13.600				
2021-395	3 ARSEN.-III	191 RX-DAYTONA	126 RANDOX	33	12.669	0.396	3.12	11.700	13.700				
2021-395	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	14	12.071	0.276	2.29	11.100	13.100				
2021-395	3 ARSEN.-III	199 Rx-IMOLA	126 RANDOX	11	12.273	0.594	4.84	11.300	13.300				
2021-395	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	82	12.005	0.354	2.95	11.000	13.000				
2021-395	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	18	12.294	0.248	2.02	11.300	13.300				
2021-395	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	12.650	0.311	2.46	11.700	13.700				
2021-395	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	12	11.892	0.454	3.81	10.900	12.900				
2021-395	14 ELECTRODE	215 B.C DXC	134 BECKM.COUL	11	11.945	0.243	2.03	10.900	12.900				
2021-395	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	18	12.328	0.300	2.43	11.300	13.300				
2021-395	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	24	12.471	0.239	1.92	11.500	13.500				

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
011	CHLORIDE	CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2021-391	3 ISE DILUTE	184 INTEGRA	93 ROCHE		38	84.658	1.675	1.98	80.000	89.000	
2021-391	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	84.379	1.064	1.26	80.000	89.000	
2021-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		18	86.889	1.197	1.38	83.000	91.000	
2021-391	3 ISE DILUTE	203 ROCHE	93 ROCHE		37	81.838	2.020	2.47	78.000	86.000	
2021-391	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		11	84.727	1.355	1.60	80.000	89.000	
2021-391	4 ISE UND-DI	165 VITROS	63 KODAK		93	85.264	1.511	1.77	81.000	90.000	
2021-391	4 ISE UND-DI	165 VITROS	156 Kit 32		67	85.106	1.426	1.68	81.000	89.000	
2021-391	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX		31	83.710	1.938	2.31	80.000	88.000	
2021-391	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX		12	83.083	1.935	2.33	79.000	87.000	
2021-391	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	84.025	1.542	1.84	80.000	88.000	
2021-391	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	83.000	2.121	2.56	79.000	87.000	
2021-391	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	82.750	1.299	1.57	79.000	87.000	
2021-392	3 ISE DILUTE	184 INTEGRA	93 ROCHE		38	104.316	2.153	2.06	99.000	110.000	
2021-392	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	101.464	1.569	1.55	96.000	107.000	
2021-392	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		18	103.222	1.227	1.19	98.000	108.000	
2021-392	3 ISE DILUTE	203 ROCHE	93 ROCHE		37	100.838	1.498	1.49	96.000	106.000	
2021-392	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		11	100.636	1.666	1.66	96.000	106.000	
2021-392	4 ISE UND-DI	165 VITROS	63 KODAK		93	102.130	1.771	1.73	97.000	107.000	
2021-392	4 ISE UND-DI	165 VITROS	156 Kit 32		67	101.879	1.600	1.57	97.000	107.000	
2021-392	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX		31	102.258	2.382	2.33	97.000	107.000	
2021-392	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX		12	101.833	2.075	2.04	97.000	107.000	
2021-392	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	107.115	1.881	1.76	102.000	112.000	
2021-392	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	105.917	2.985	2.82	101.000	111.000	
2021-392	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	102.250	1.920	1.88	97.000	107.000	
2021-393	3 ISE DILUTE	184 INTEGRA	93 ROCHE		38	92.158	1.478	1.60	88.000	97.000	
2021-393	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	91.241	1.330	1.46	87.000	96.000	
2021-393	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		18	93.611	1.568	1.68	89.000	98.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-393	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	89.324	1.875	2.10	85.000	94.000
2021-393	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	11	92.091	1.311	1.42	87.000	97.000
2021-393	4	ISE UND-DI	165	VITROS	63	KODAK	93	91.772	1.596	1.74	87.000	96.000
2021-393	4	ISE UND-DI	165	VITROS	156	Kit 32	67	91.621	1.334	1.46	87.000	96.000
2021-393	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	31	90.677	1.574	1.74	86.000	95.000
2021-393	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	12	90.083	0.954	1.06	86.000	95.000
2021-393	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	93.026	1.432	1.54	88.000	98.000
2021-393	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	91.833	2.609	2.84	87.000	96.000
2021-393	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	91.000	2.121	2.33	86.000	96.000
011	CHLORIDE	CLIA TEST ID #: 355		EVALUATION CRITERIA: TARGET VALUE 5 %								
2021-394	3	ISE DILUTE	184	INTEGRA	93	ROCHE	38	106.079	2.321	2.19	101.000	111.000
2021-394	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	29	101.966	1.671	1.64	97.000	107.000
2021-394	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	18	104.222	1.548	1.48	99.000	109.000
2021-394	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	102.324	1.693	1.65	97.000	107.000
2021-394	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	11	102.636	1.720	1.68	98.000	108.000
2021-394	4	ISE UND-DI	165	VITROS	63	KODAK	93	103.598	2.187	2.11	98.000	109.000
2021-394	4	ISE UND-DI	165	VITROS	156	Kit 32	67	103.433	1.871	1.81	98.000	109.000
2021-394	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	31	103.806	2.520	2.43	99.000	109.000
2021-394	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	12	103.833	2.409	2.32	99.000	109.000
2021-394	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	108.810	2.273	2.09	103.000	114.000
2021-394	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	107.583	2.691	2.50	102.000	113.000
2021-394	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	103.750	2.861	2.76	99.000	109.000
2021-395	3	ISE DILUTE	184	INTEGRA	93	ROCHE	38	115.158	3.183	2.76	109.000	121.000
2021-395	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	29	109.621	2.296	2.09	104.000	115.000
2021-395	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	18	111.333	1.633	1.47	106.000	117.000
2021-395	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	110.472	1.936	1.75	105.000	116.000
2021-395	3	ISE DILUTE	215	B.C DXC	134	BECKM.COUL	11	109.273	2.799	2.56	104.000	115.000
2021-395	4	ISE UND-DI	165	VITROS	63	KODAK	93	111.478	2.478	2.22	106.000	117.000
2021-395	4	ISE UND-DI	165	VITROS	156	Kit 32	67	110.985	1.822	1.64	105.000	117.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-395	4	ISE UND-DI	191	RX-DAYTONA	126	RANDOX	31	113.097	2.998	2.65	107.000	119.000
2021-395	4	ISE UND-DI	199	Rx-IMOLA	126	RANDOX	12	112.417	3.593	3.20	107.000	118.000
2021-395	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	118.557	2.326	1.96	113.000	124.000
2021-395	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	117.250	3.218	2.74	111.000	123.000
2021-395	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	12	112.000	2.236	2.00	106.000	118.000
012	CHOLESTEROL, TOTAL		CLIA	TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %							
2021-391	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	80	114.848	4.487	3.91	103.000	126.000
2021-391	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	80	118.125	2.571	2.18	106.000	130.000
2021-391	1	ENZYMATIC	165	VITROS	63	KODAK	152	110.373	3.080	2.79	99.000	121.000
2021-391	1	ENZYMATIC	184	INTEGRA	93	ROCHE	21	116.000	3.281	2.83	104.000	128.000
2021-391	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	36	121.111	7.070	5.84	109.000	133.000
2021-391	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	16	118.500	2.264	1.91	107.000	130.000
2021-391	1	ENZYMATIC	199	Rx-IMOLA	126	RANDOX	10	118.700	3.100	2.61	107.000	131.000
2021-391	1	ENZYMATIC	203	ROCHE	93	ROCHE	21	117.381	3.078	2.62	106.000	129.000
2021-391	1	ENZYMATIC	212	S. D. EXL	143	SIEMENS	12	114.750	5.293	4.61	103.000	126.000
2021-391	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	11	117.545	3.846	3.27	106.000	129.000
2021-391	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	26	114.692	2.700	2.35	103.000	126.000
2021-392	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	80	189.375	4.470	2.36	170.000	208.000
2021-392	1	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	80	179.924	6.330	3.52	162.000	198.000
2021-392	1	ENZYMATIC	165	VITROS	63	KODAK	152	191.447	5.803	3.03	172.000	211.000
2021-392	1	ENZYMATIC	184	INTEGRA	93	ROCHE	21	185.476	5.197	2.80	167.000	204.000
2021-392	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	36	194.389	10.660	5.48	175.000	214.000
2021-392	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	16	185.438	3.278	1.77	167.000	204.000
2021-392	1	ENZYMATIC	199	Rx-IMOLA	126	RANDOX	10	195.500	5.679	2.90	176.000	215.000
2021-392	1	ENZYMATIC	203	ROCHE	93	ROCHE	21	187.048	4.715	2.52	168.000	206.000
2021-392	1	ENZYMATIC	212	S. D. EXL	143	SIEMENS	12	178.917	6.551	3.66	161.000	197.000
2021-392	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	11	187.545	8.360	4.46	169.000	206.000
2021-392	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	26	184.640	3.576	1.94	166.000	203.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
012	CHOLESTEROL, TOTAL	CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE	10 %							
2021-393	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	140.705	5.400	3.84	127.000	155.000	
2021-393	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	146.875	2.759	1.88	132.000	162.000	
2021-393	1 ENZYMATI	165 VITROS	63 KODAK		152	142.795	4.043	2.83	129.000	157.000	
2021-393	1 ENZYMATI	184 INTEGRA	93 ROCHE		21	144.286	3.326	2.31	130.000	159.000	
2021-393	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX		36	151.750	8.591	5.66	137.000	167.000	
2021-393	1 ENZYMATI	197 ARCHITECT	1 ABBOTT		16	146.500	3.873	2.64	132.000	161.000	
2021-393	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX		10	149.700	4.605	3.08	135.000	165.000	
2021-393	1 ENZYMATI	203 ROCHE	93 ROCHE		21	145.238	3.939	2.71	131.000	160.000	
2021-393	1 ENZYMATI	212 S. D. EXL	143 SIEMENS		12	140.167	5.336	3.81	126.000	154.000	
2021-393	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL		11	146.545	5.758	3.93	132.000	161.000	
2021-393	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN		26	143.115	3.434	2.40	129.000	157.000	
2021-394	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	198.375	2.781	1.40	179.000	218.000	
2021-394	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	187.169	7.000	3.74	168.000	206.000	
2021-394	1 ENZYMATI	165 VITROS	63 KODAK		152	202.801	5.602	2.76	183.000	223.000	
2021-394	1 ENZYMATI	184 INTEGRA	93 ROCHE		21	193.905	5.740	2.96	175.000	213.000	
2021-394	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX		36	203.333	10.690	5.26	183.000	224.000	
2021-394	1 ENZYMATI	197 ARCHITECT	1 ABBOTT		16	193.250	3.192	1.65	174.000	213.000	
2021-394	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX		10	201.000	5.158	2.57	181.000	221.000	
2021-394	1 ENZYMATI	203 ROCHE	93 ROCHE		21	195.700	3.689	1.89	176.000	215.000	
2021-394	1 ENZYMATI	212 S. D. EXL	143 SIEMENS		12	186.667	7.087	3.80	168.000	205.000	
2021-394	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL		11	198.727	7.840	3.95	179.000	219.000	
2021-394	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN		26	192.615	4.592	2.38	173.000	212.000	
2021-395	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	216.494	8.546	3.95	195.000	238.000	
2021-395	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		80	233.000	6.124	2.63	210.000	256.000	
2021-395	1 ENZYMATI	165 VITROS	63 KODAK		152	241.013	7.473	3.10	217.000	265.000	
2021-395	1 ENZYMATI	184 INTEGRA	93 ROCHE		21	224.190	5.077	2.26	202.000	247.000	
2021-395	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX		36	237.667	11.849	4.99	214.000	261.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-395	1 ENZYMATIC	197 ARCHITECT	1 ABBOTT	16	223.313	4.818	2.16	201.000	246.000		
2021-395	1 ENZYMATIC	199 Rx-IMOLA	126 RANDOX	10	234.600	8.958	3.82	211.000	258.000		
2021-395	1 ENZYMATIC	203 ROCHE	93 ROCHE	21	228.238	4.994	2.19	205.000	251.000		
2021-395	1 ENZYMATIC	212 S. D. EXL	143 SIEMENS	12	216.917	11.615	5.35	195.000	239.000		
2021-395	1 ENZYMATIC	215 B.C DXC	134 BECKM.COUL	11	232.182	9.485	4.09	209.000	255.000		
2021-395	8 ENZYMATIC	186 BECK. COUL	76 AU-BECKMAN	26	224.423	8.186	3.65	202.000	247.000		
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-391	11 ENC.-COLO.	165 VITROS	63 KODAK	147	42.333	0.471	1.11	29.600	55.000		
2021-391	11 ENC.-COLO.	165 VITROS	63 KODAK	147	42.599	2.254	5.29	29.800	55.400		
2021-391	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	45.830	2.691	5.87	32.100	59.600		
2021-391	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	43.296	2.673	6.17	30.300	56.300		
2021-391	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	33	40.933	3.411	8.33	28.700	53.200		
2021-391	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	10	27.870	1.418	5.09	19.500	36.200		
2021-391	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	11	40.691	3.982	9.79	28.500	52.900		
2021-391	11 ENC.-COLO.	203 ROCHE	93 ROCHE	31	45.257	1.975	4.36	31.700	58.800		
2021-391	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	18	44.383	3.114	7.02	31.100	57.700		
2021-391	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	11	44.582	1.605	3.60	31.200	58.000		
2021-391	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	48.409	1.962	4.05	33.900	62.900		
2021-391	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	12	48.417	1.977	4.08	33.900	62.900		
2021-392	11 ENC.-COLO.	165 VITROS	63 KODAK	147	62.000	1.633	2.63	43.400	80.600		
2021-392	11 ENC.-COLO.	165 VITROS	63 KODAK	147	62.692	3.550	5.66	43.900	81.500		
2021-392	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	61.467	3.006	4.89	43.000	79.900		
2021-392	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	69.677	4.539	6.51	48.800	90.600		
2021-392	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	33	63.176	5.275	8.35	44.200	82.100		
2021-392	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	10	45.660	3.731	8.17	32.000	59.400		
2021-392	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	11	64.300	5.288	8.22	45.000	83.600		
2021-392	11 ENC.-COLO.	203 ROCHE	93 ROCHE	31	60.848	3.086	5.07	42.600	79.100		
2021-392	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	18	68.767	4.535	6.60	48.100	89.400		
2021-392	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	11	69.736	3.361	4.82	48.800	90.700		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-392	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	56.476	2.204	3.90	39.500	73.400		
2021-392	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	12	56.750	3.345	5.89	39.700	73.800		
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-393	11 ENC.-COLO.	165 VITROS	63 KODAK	147	50.148	2.252	4.49	35.100	65.200		
2021-393	11 ENC.-COLO.	165 VITROS	63 KODAK	147	50.000	0.816	1.63	35.000	65.000		
2021-393	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	52.782	2.903	5.50	36.900	68.600		
2021-393	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	54.019	3.349	6.20	37.800	70.200		
2021-393	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	33	49.806	4.165	8.36	34.900	64.700		
2021-393	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	10	35.600	2.417	6.79	24.900	46.300		
2021-393	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	11	52.018	3.693	7.10	36.400	67.600		
2021-393	11 ENC.-COLO.	203 ROCHE	93 ROCHE	31	51.542	2.849	5.53	36.100	67.000		
2021-393	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	18	53.978	3.423	6.34	37.800	70.200		
2021-393	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	11	55.209	2.107	3.82	38.600	71.800		
2021-393	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	52.266	1.989	3.81	36.600	67.900		
2021-393	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	12	51.667	2.560	4.96	36.200	67.200		
2021-394	11 ENC.-COLO.	165 VITROS	63 KODAK	147	66.333	1.247	1.88	46.400	86.200		
2021-394	11 ENC.-COLO.	165 VITROS	63 KODAK	147	65.990	3.239	4.91	46.200	85.800		
2021-394	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	62.842	3.161	5.03	44.000	81.700		
2021-394	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	73.042	4.750	6.50	51.100	95.000		
2021-394	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	33	65.258	5.322	8.16	45.700	84.800		
2021-394	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	10	47.210	2.738	5.80	33.000	61.400		
2021-394	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	11	67.645	6.155	9.10	47.400	87.900		
2021-394	11 ENC.-COLO.	203 ROCHE	93 ROCHE	31	62.432	2.903	4.65	43.700	81.200		
2021-394	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	18	71.672	4.444	6.20	50.200	93.200		
2021-394	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	11	73.291	2.415	3.29	51.300	95.300		
2021-394	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	57.279	2.790	4.87	40.100	74.500		
2021-394	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	12	56.750	2.278	4.01	39.700	73.800		
2021-395	11 ENC.-COLO.	165 VITROS	63 KODAK	147	78.971	4.283	5.42	55.300	102.700		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-395	11 ENC.-COLO.	165 VITROS	63 KODAK	147	78.000	1.633	2.09	54.600	101.400		
2021-395	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	68.347	3.171	4.64	47.800	88.900		
2021-395	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	85.204	3.811	4.47	59.600	110.800		
2021-395	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	33	74.545	6.042	8.11	52.200	96.900		
2021-395	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	10	54.510	4.290	7.87	38.200	70.900		
2021-395	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	11	75.682	5.327	7.04	53.000	98.400		
2021-395	11 ENC.-COLO.	203 ROCHE	93 ROCHE	31	68.390	3.202	4.68	47.900	88.900		
2021-395	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	18	82.867	5.559	6.71	58.000	107.700		
2021-395	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	11	84.700	4.317	5.10	59.300	110.100		
2021-395	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	59.412	2.997	5.05	41.600	77.200		
2021-395	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	12	59.083	2.361	4.00	41.400	76.800		
014 CREATINE KINASE (CK)		CLIA TEST ID #: 385	EVALUATION CRITERIA: TARGET VALUE 30 %								
2021-391	3 KINETIC 37	165 VITROS	63 KODAK	76	36.237	3.268	9.02	25.000	47.000		
2021-391	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	19	33.158	1.814	5.47	23.000	43.000		
2021-391	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	42.200	1.470	3.48	30.000	55.000		
2021-391	8 Hexokinase	184 INTEGRA	93 ROCHE	12	40.333	1.374	3.41	28.000	52.000		
2021-391	8 Hexokinase	203 ROCHE	93 ROCHE	12	40.750	1.422	3.49	29.000	53.000		
2021-391	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	38.536	1.918	4.98	27.000	50.000		
2021-392	3 KINETIC 37	165 VITROS	63 KODAK	76	310.293	16.711	5.39	217.000	403.000		
2021-392	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	19	414.526	18.409	4.44	290.000	539.000		
2021-392	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	466.100	15.909	3.41	326.000	606.000		
2021-392	8 Hexokinase	184 INTEGRA	93 ROCHE	12	441.417	14.459	3.28	309.000	574.000		
2021-392	8 Hexokinase	203 ROCHE	93 ROCHE	12	442.333	4.570	1.03	310.000	575.000		
2021-392	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	444.690	12.717	2.86	311.000	578.000		
2021-393	3 KINETIC 37	165 VITROS	63 KODAK	76	173.263	10.291	5.94	121.000	225.000		
2021-393	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	19	195.684	8.880	4.54	137.000	254.000		
2021-393	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	228.200	7.054	3.09	160.000	297.000		
2021-393	8 Hexokinase	184 INTEGRA	93 ROCHE	12	215.000	5.132	2.39	151.000	280.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-393	8 Hexokinase	203 ROCHE	93 ROCHE	12	212.583	4.716	2.22	149.000	276.000		
2021-393	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	213.036	4.452	2.09	149.000	277.000		
014 CREATINE KINASE (CK)		CLIA TEST ID #: 385				EVALUATION CRITERIA: TARGET VALUE 30 %					
2021-394	3 KINETIC 37	165 VITROS	63 KODAK	76	338.500	17.938	5.30	237.000	440.000		
2021-394	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	19	469.579	20.783	4.43	329.000	610.000		
2021-394	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	533.600	15.914	2.98	374.000	694.000		
2021-394	8 Hexokinase	184 INTEGRA	93 ROCHE	12	496.667	13.810	2.78	348.000	646.000		
2021-394	8 Hexokinase	203 ROCHE	93 ROCHE	12	496.167	8.952	1.80	347.000	645.000		
2021-394	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	508.897	13.430	2.64	356.000	662.000		
2021-395	3 KINETIC 37	165 VITROS	63 KODAK	76	398.026	19.997	5.02	279.000	517.000		
2021-395	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	19	646.105	28.819	4.46	452.000	840.000		
2021-395	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	10	722.700	19.935	2.76	506.000	940.000		
2021-395	8 Hexokinase	184 INTEGRA	93 ROCHE	12	675.917	15.735	2.33	473.000	879.000		
2021-395	8 Hexokinase	203 ROCHE	93 ROCHE	12	686.083	11.821	1.72	480.000	892.000		
2021-395	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	688.517	19.963	2.90	482.000	895.000		
016 CREATININE		CLIA TEST ID #: 405				EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}					
2021-391	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	10	0.967	0.076	7.81	0.670	1.270		
2021-391	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	23	0.916	0.073	7.99	0.620	1.220		
2021-391	6 KIN. ALK P	184 INTEGRA	93 ROCHE	24	0.941	0.051	5.38	0.640	1.240		
2021-391	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	29	0.885	0.039	4.45	0.580	1.180		
2021-391	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	0.786	0.022	2.85	0.490	1.090		
2021-391	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	0.861	0.034	3.93	0.560	1.160		
2021-391	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	11	0.924	0.041	4.41	0.620	1.220		
2021-391	8 IDMS-CREA	165 VITROS	63 KODAK	159	0.838	0.037	4.46	0.540	1.140		
2021-391	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	73	0.885	0.046	5.20	0.590	1.190		
2021-392	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	10	4.122	0.195	4.72	3.500	4.740		
2021-392	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	23	3.772	0.203	5.39	3.210	4.340		
2021-392	6 KIN. ALK P	184 INTEGRA	93 ROCHE	24	4.407	0.160	3.63	3.750	5.070		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
	2021-392	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	29	4.428	0.111	2.50	3.760	5.090	
	2021-392	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	4.730	0.111	2.35	4.020	5.440	
	2021-392	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	4.340	0.126	2.90	3.690	4.990	
	2021-392	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	11	4.620	0.073	1.59	3.930	5.310	
	2021-392	8 IDMS-CREA	165 VITROS	63 KODAK	159	3.719	0.113	3.05	3.160	4.280	
	2021-392	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	73	4.687	0.136	2.90	3.980	5.390	
016 CREATININE			CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}							
	2021-393	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	10	2.268	0.120	5.31	1.930	2.610	
	2021-393	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	23	2.170	0.159	7.31	1.840	2.500	
	2021-393	6 KIN. ALK P	184 INTEGRA	93 ROCHE	24	2.474	0.098	3.95	2.100	2.840	
	2021-393	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	29	2.408	0.071	2.95	2.050	2.770	
	2021-393	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	2.422	0.066	2.73	2.060	2.790	
	2021-393	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	2.403	0.061	2.54	2.040	2.760	
	2021-393	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	11	2.516	0.055	2.19	2.140	2.890	
	2021-393	8 IDMS-CREA	165 VITROS	63 KODAK	159	2.182	0.070	3.19	1.850	2.510	
	2021-393	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	73	2.586	0.085	3.28	2.200	2.970	
	2021-394	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	10	4.565	0.220	4.83	3.880	5.250	
	2021-394	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	23	4.222	0.239	5.65	3.590	4.860	
	2021-394	6 KIN. ALK P	184 INTEGRA	93 ROCHE	24	4.843	0.227	4.69	4.120	5.570	
	2021-394	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	29	4.911	0.119	2.43	4.170	5.650	
	2021-394	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	5.307	0.137	2.58	4.510	6.100	
	2021-394	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	4.828	0.145	2.99	4.100	5.550	
	2021-394	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	11	5.190	0.074	1.43	4.410	5.970	
	2021-394	8 IDMS-CREA	165 VITROS	63 KODAK	159	4.062	0.121	2.98	3.450	4.670	
	2021-394	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	73	5.233	0.178	3.40	4.450	6.020	
	2021-395	2 AK. PIC. N	194 BIOLIS-24S	154 (IDG)	10	6.037	0.347	5.75	5.130	6.940	
	2021-395	3 AK. PIC. L	191 RX-DAYTONA	126 RANDOX	23	5.463	0.297	5.43	4.640	6.280	
	2021-395	6 KIN. ALK P	184 INTEGRA	93 ROCHE	24	6.185	0.252	4.08	5.260	7.110	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2021-395	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	29	6.451	0.194	3.01	5.480	7.420			
2021-395	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	18	7.094	0.183	2.58	6.030	8.160			
2021-395	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	6.235	0.126	2.03	5.300	7.170			
2021-395	6 KIN. ALK P	215 B.C DXC	134 BECKM.COUL	11	6.869	0.174	2.53	5.840	7.900			
2021-395	8 IDMS-CREA	165 VITROS	63 KODAK	159	4.924	0.153	3.11	4.190	5.660			
2021-395	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	73	6.947	0.215	3.09	5.900	7.990			
017 GLUCOSE		CLIA TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}									
2021-391	2 GL. OXI. E	215 B.C DXC	134 BECKM.COUL	11	39.727	1.052	2.65	34.000	46.000			
2021-391	3 GL. OXI. C	165 VITROS	63 KODAK	167	38.558	1.177	3.05	33.000	45.000			
2021-391	3 GL. OXI. C	213 SELECTRA P	150 ELITECH	10	44.300	1.269	2.86	38.000	50.000			
2021-391	3 GL. OXI. C	214 Envoy 500	150 ELITECH	10	44.500	1.285	2.89	39.000	51.000			
2021-391	5 HEXOKINASE	38 SIEMENS DI	143 SIEMENS	82	44.012	1.753	3.98	38.000	50.000			
2021-391	5 HEXOKINASE	186 BECK. COUL	76 AU-BECKMAN	29	42.241	1.164	2.76	36.000	48.000			
2021-391	5 HEXOKINASE	191 RX-DAYTONA	126 RANDOX	25	46.080	1.978	4.29	40.000	52.000			
2021-391	5 HEXOKINASE	203 ROCHE	93 ROCHE	20	41.400	1.158	2.80	35.000	47.000			
2021-391	5 HEXOKINASE	212 S. D. EXL	143 SIEMENS	12	44.250	2.487	5.62	38.000	50.000			
2021-391	6 HEXOKI. CO	194 BIOLIS-24S	154 (IDG)	10	42.700	1.418	3.32	37.000	49.000			
2021-391	12 HEXOKINASE	184 INTEGRA	93 ROCHE	27	43.111	1.100	2.55	37.000	49.000			
2021-391	12 HEXOKINASE	197 ARCHITECT	1 ABBOTT	18	39.611	0.891	2.25	34.000	46.000			
2021-391	12 HEXOKINASE	203 ROCHE	93 ROCHE	16	41.375	0.927	2.24	35.000	47.000			
2021-392	2 GL. OXI. E	215 B.C DXC	134 BECKM.COUL	11	115.364	1.772	1.54	104.000	127.000			
2021-392	3 GL. OXI. C	165 VITROS	63 KODAK	167	122.594	2.496	2.04	110.000	135.000			
2021-392	3 GL. OXI. C	213 SELECTRA P	150 ELITECH	10	124.300	3.068	2.47	112.000	137.000			
2021-392	3 GL. OXI. C	214 Envoy 500	150 ELITECH	10	124.700	3.926	3.15	112.000	137.000			
2021-392	5 HEXOKINASE	38 SIEMENS DI	143 SIEMENS	82	124.630	3.543	2.84	112.000	137.000			
2021-392	5 HEXOKINASE	186 BECK. COUL	76 AU-BECKMAN	29	120.286	2.462	2.05	108.000	132.000			
2021-392	5 HEXOKINASE	191 RX-DAYTONA	126 RANDOX	25	124.840	4.575	3.67	112.000	137.000			
2021-392	5 HEXOKINASE	203 ROCHE	93 ROCHE	20	119.000	2.828	2.38	107.000	131.000			
2021-392	5 HEXOKINASE	212 S. D. EXL	143 SIEMENS	12	124.667	4.625	3.71	112.000	137.000			

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-392	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	10	117.200	2.676	2.28	105.000	129.000
2021-392	12	HEXOKINASE	184 INTEGRA	93	ROCHE	27	123.000	3.333	2.71	111.000	135.000
2021-392	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	18	115.941	1.434	1.24	104.000	128.000
2021-392	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	119.875	3.295	2.75	108.000	132.000
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}					
2021-393	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	11	72.182	1.192	1.65	65.000	79.000
2021-393	3	GL. OXI. C	165 VITROS	63	KODAK	167	74.641	1.874	2.51	67.000	82.000
2021-393	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	10	78.800	2.750	3.49	71.000	87.000
2021-393	3	GL. OXI. C	214 Envoy 500	150	ELITECH	10	79.000	2.366	3.00	71.000	87.000
2021-393	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	77.863	2.370	3.04	70.000	86.000
2021-393	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	29	75.379	1.919	2.55	68.000	83.000
2021-393	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	25	79.560	2.981	3.75	72.000	88.000
2021-393	5	HEXOKINASE	203 ROCHE	93	ROCHE	20	73.500	1.746	2.38	66.000	81.000
2021-393	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	12	77.917	2.985	3.83	70.000	86.000
2021-393	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	10	74.200	2.088	2.81	67.000	82.000
2021-393	12	HEXOKINASE	184 INTEGRA	93	ROCHE	27	76.778	1.950	2.54	69.000	84.000
2021-393	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	18	72.444	2.386	3.29	65.000	80.000
2021-393	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	74.563	1.368	1.83	67.000	82.000
2021-394	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	11	126.818	2.289	1.80	114.000	140.000
2021-394	3	GL. OXI. C	165 VITROS	63	KODAK	167	134.313	2.963	2.21	121.000	148.000
2021-394	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	10	137.500	4.738	3.45	124.000	151.000
2021-394	3	GL. OXI. C	214 Envoy 500	150	ELITECH	10	137.500	4.696	3.42	124.000	151.000
2021-394	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	135.850	4.196	3.09	122.000	149.000
2021-394	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	29	131.138	3.758	2.87	118.000	144.000
2021-394	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	25	136.720	5.724	4.19	123.000	150.000
2021-394	5	HEXOKINASE	203 ROCHE	93	ROCHE	20	129.200	3.124	2.42	116.000	142.000
2021-394	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	12	135.417	4.462	3.30	122.000	149.000
2021-394	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	10	127.600	3.137	2.46	115.000	140.000
2021-394	12	HEXOKINASE	184 INTEGRA	93	ROCHE	27	134.704	3.547	2.63	121.000	148.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-394	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	127.222	2.417	1.90	115.000	140.000
2021-394	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	131.000	2.646	2.02	118.000	144.000
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2021-395	2	GL. OXI. E	215	B.C DXC	134	BECKM.COUL	11	162.545	2.641	1.62	146.000	179.000
2021-395	3	GL. OXI. C	165	VITROS	63	KODAK	167	174.200	3.996	2.29	157.000	192.000
2021-395	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	10	174.000	4.595	2.64	157.000	191.000
2021-395	3	GL. OXI. C	214	Envoy 500	150	ELITECH	10	173.400	5.295	3.05	156.000	191.000
2021-395	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	82	171.654	5.087	2.96	154.000	189.000
2021-395	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	29	167.429	4.632	2.77	151.000	184.000
2021-395	5	HEXOKINASE	191	RX-DAYTONA	126	RANDOX	25	171.625	5.736	3.34	154.000	189.000
2021-395	5	HEXOKINASE	203	ROCHE	93	ROCHE	20	165.650	4.163	2.51	149.000	182.000
2021-395	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	12	170.083	5.575	3.28	153.000	187.000
2021-395	6	HEXOKI. CO	194	BIOLIS-24S	154	(IDG)	10	162.000	4.690	2.90	146.000	178.000
2021-395	12	HEXOKINASE	184	INTEGRA	93	ROCHE	27	170.407	4.449	2.61	153.000	187.000
2021-395	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	161.500	4.787	2.96	145.000	178.000
2021-395	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	168.000	2.151	1.28	151.000	185.000
018	IRON, TOTAL	CLIA TEST ID #: 425				EVALUATION CRITERIA: TARGET VALUE 20 %						
2021-391	4	COULOMETR.	165	VITROS	63	KODAK	16	41.063	4.892	11.91	33.000	49.000
2021-391	4	COULOMETR.	165	VITROS	63	KODAK	16	46.833	1.344	2.87	37.000	56.000
2021-392	4	COULOMETR.	165	VITROS	63	KODAK	16	222.667	3.350	1.50	178.000	267.000
2021-392	4	COULOMETR.	165	VITROS	63	KODAK	16	240.533	9.251	3.85	192.000	289.000
2021-393	4	COULOMETR.	165	VITROS	63	KODAK	16	120.438	5.111	4.24	96.000	145.000
2021-393	4	COULOMETR.	165	VITROS	63	KODAK	16	121.833	2.672	2.19	97.000	146.000
2021-394	4	COULOMETR.	165	VITROS	63	KODAK	16	248.000	2.769	1.12	198.000	298.000
2021-394	4	COULOMETR.	165	VITROS	63	KODAK	16	273.875	12.062	4.40	219.000	329.000
2021-395	4	COULOMETR.	165	VITROS	63	KODAK	16	390.438	14.517	3.72	312.000	469.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-395	4	COULOMETR.	165	VITROS	63	KODAK	16	327.000	5.538	1.69	262.000	392.000
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435				EVALUATION CRITERIA: TARGET VALUE 20 %						
2021-391	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	19	61.526	4.604	7.48	49.000	74.000
2021-391	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	10	64.000	1.844	2.88	51.000	77.000
2021-391	6	NADH	184	INTEGRA	93	ROCHE	14	81.714	3.194	3.91	65.000	98.000
2021-391	6	NADH	203	ROCHE	93	ROCHE	22	76.524	2.038	2.66	61.000	92.000
2021-391	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	33	71.212	5.688	7.99	57.000	85.000
2021-391	9	LDH1	165	VITROS	63	KODAK	68	90.269	4.931	5.46	72.000	108.000
2021-392	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	19	416.105	28.961	6.96	333.000	499.000
2021-392	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	10	389.700	13.176	3.38	312.000	468.000
2021-392	6	NADH	184	INTEGRA	93	ROCHE	14	509.929	18.285	3.59	408.000	612.000
2021-392	6	NADH	203	ROCHE	93	ROCHE	22	499.000	11.897	2.38	399.000	599.000
2021-392	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	33	476.394	10.307	2.16	381.000	572.000
2021-392	9	LDH1	165	VITROS	63	KODAK	68	581.179	17.404	2.99	465.000	697.000
2021-393	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	19	211.000	15.379	7.29	169.000	253.000
2021-393	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	10	208.100	6.284	3.02	166.000	250.000
2021-393	6	NADH	184	INTEGRA	93	ROCHE	14	267.429	9.295	3.48	214.000	321.000
2021-393	6	NADH	203	ROCHE	93	ROCHE	22	257.286	5.444	2.12	206.000	309.000
2021-393	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	33	243.250	7.194	2.96	195.000	292.000
2021-393	9	LDH1	165	VITROS	63	KODAK	68	308.118	8.376	2.72	246.000	370.000
2021-394	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	19	464.842	32.145	6.92	372.000	558.000
2021-394	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	10	440.500	14.935	3.39	352.000	529.000
2021-394	6	NADH	184	INTEGRA	93	ROCHE	14	568.214	15.662	2.76	455.000	682.000
2021-394	6	NADH	203	ROCHE	93	ROCHE	22	562.500	15.631	2.78	450.000	675.000
2021-394	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	33	537.688	15.689	2.92	430.000	645.000
2021-394	9	LDH1	165	VITROS	63	KODAK	68	647.779	25.696	3.97	518.000	777.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435				EVALUATION CRITERIA: TARGET VALUE 20 %					
2021-395	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN		19	625.105	41.393	6.62	500.000	750.000	
2021-395	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL		10	580.600	17.333	2.99	464.000	697.000	
2021-395	6 NADH	184 INTEGRA	93 ROCHE		14	751.500	23.700	3.15	601.000	902.000	
2021-395	6 NADH	203 ROCHE	93 ROCHE		22	751.091	17.666	2.35	601.000	901.000	
2021-395	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS		33	722.152	20.688	2.86	578.000	867.000	
2021-395	9 LDH1	165 VITROS	63 KODAK		68	834.090	38.687	4.64	667.000	1001.000	
021	MAGNESIUM	CLIA TEST ID #: 455				EVALUATION CRITERIA: TARGET VALUE 25 %					
2021-391	5 COLOR./K E	165 VITROS	63 KODAK		25	1.338	0.063	4.73	1.000	1.700	
2021-392	5 COLOR./K E	165 VITROS	63 KODAK		25	4.408	0.150	3.40	3.300	5.500	
2021-393	5 COLOR./K E	165 VITROS	63 KODAK		25	2.660	0.069	2.60	2.000	3.300	
2021-394	5 COLOR./K E	165 VITROS	63 KODAK		25	4.880	0.141	2.90	3.700	6.100	
2021-395	5 COLOR./K E	165 VITROS	63 KODAK		25	6.204	0.201	3.24	4.700	7.800	
022	POTASSIUM	CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00					
2021-391	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	2.115	0.035	1.67	1.100	3.100	
2021-391	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		29	2.179	0.049	2.25	1.200	3.200	
2021-391	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		16	2.150	0.061	2.85	1.200	3.200	
2021-391	3 ISE DILUT.	203 ROCHE	93 ROCHE		38	2.192	0.087	3.97	1.200	3.200	
2021-391	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		11	2.036	0.048	2.36	1.000	3.000	
2021-391	4 ISE UNDIL.	165 VITROS	63 KODAK		95	2.112	0.041	1.94	1.100	3.100	
2021-391	4 ISE UNDIL.	165 VITROS	156 Kit 32		67	2.114	0.052	2.46	1.100	3.100	
2021-391	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		31	2.194	0.141	6.44	1.200	3.200	
2021-391	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	2.150	0.104	4.84	1.200	3.200	
2021-391	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	2.044	0.052	2.55	1.000	3.000	
2021-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	2.042	0.076	3.72	1.000	3.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
022 POTASSIUM		CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE			1.00					
2021-392	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	4.683	0.073	1.55	3.700	5.700	
2021-392	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		29	4.610	0.088	1.92	3.600	5.600	
2021-392	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		16	4.706	0.090	1.91	3.700	5.700	
2021-392	3 ISE DILUT.	203 ROCHE	93 ROCHE		38	4.711	0.085	1.81	3.700	5.700	
2021-392	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		11	4.600	0.060	1.31	3.600	5.600	
2021-392	4 ISE UNDIL.	165 VITROS	63 KODAK		95	4.695	0.090	1.91	3.700	5.700	
2021-392	4 ISE UNDIL.	165 VITROS	156 Kit 32		67	4.685	0.084	1.79	3.700	5.700	
2021-392	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		31	4.603	0.109	2.37	3.600	5.600	
2021-392	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	4.583	0.090	1.96	3.600	5.600	
2021-392	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	4.633	0.069	1.49	3.600	5.600	
2021-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	4.625	0.092	2.00	3.600	5.600	
2021-393	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	3.186	0.058	1.84	2.200	4.200	
2021-393	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		29	3.190	0.048	1.51	2.200	4.200	
2021-393	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		16	3.206	0.066	2.05	2.200	4.200	
2021-393	3 ISE DILUT.	203 ROCHE	93 ROCHE		38	3.237	0.074	2.29	2.200	4.200	
2021-393	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		11	3.127	0.045	1.42	2.100	4.100	
2021-393	4 ISE UNDIL.	165 VITROS	63 KODAK		95	3.185	0.064	2.00	2.200	4.200	
2021-393	4 ISE UNDIL.	165 VITROS	156 Kit 32		67	3.175	0.063	1.99	2.200	4.200	
2021-393	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		31	3.173	0.100	3.14	2.200	4.200	
2021-393	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		12	3.125	0.101	3.23	2.100	4.100	
2021-393	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	3.120	0.060	1.93	2.100	4.100	
2021-393	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	3.125	0.060	1.90	2.100	4.100	
2021-394	3 ISE DILUT.	184 INTEGRA	93 ROCHE		36	5.039	0.089	1.77	4.000	6.000	
2021-394	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN		29	4.955	0.085	1.72	4.000	6.000	
2021-394	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT		16	5.069	0.092	1.81	4.100	6.100	
2021-394	3 ISE DILUT.	203 ROCHE	93 ROCHE		38	5.089	0.082	1.61	4.100	6.100	
2021-394	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL		11	5.009	0.079	1.58	4.000	6.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
	2021-394	4 ISE UNDIL.	165 VITROS	63 KODAK	95	5.081	0.091	1.79	4.100	6.100
	2021-394	4 ISE UNDIL.	165 VITROS	156 Kit 32	67	5.058	0.093	1.84	4.100	6.100
	2021-394	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX	31	4.955	0.127	2.55	4.000	6.000
	2021-394	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX	12	4.925	0.109	2.21	3.900	5.900
	2021-394	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	79	4.999	0.081	1.63	4.000	6.000
	2021-394	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	12	4.967	0.075	1.50	4.000	6.000
022 POTASSIUM			CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE 1.00						
	2021-395	3 ISE DILUT.	184 INTEGRA	93 ROCHE	36	6.233	0.115	1.85	5.200	7.200
	2021-395	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	29	6.152	0.130	2.12	5.200	7.200
	2021-395	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	16	6.313	0.117	1.85	5.300	7.300
	2021-395	3 ISE DILUT.	203 ROCHE	93 ROCHE	38	6.270	0.121	1.92	5.300	7.300
	2021-395	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	11	6.264	0.115	1.84	5.300	7.300
	2021-395	4 ISE UNDIL.	165 VITROS	63 KODAK	95	6.304	0.097	1.54	5.300	7.300
	2021-395	4 ISE UNDIL.	165 VITROS	156 Kit 32	67	6.283	0.079	1.26	5.300	7.300
	2021-395	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX	31	6.152	0.183	2.97	5.200	7.200
	2021-395	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX	12	6.117	0.157	2.57	5.100	7.100
	2021-395	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	79	6.194	0.106	1.71	5.200	7.200
	2021-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	12	6.133	0.137	2.24	5.100	7.100
023 SODIUM			CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00						
	2021-391	3 ISE DILUTE	184 INTEGRA	93 ROCHE	33	120.833	2.339	1.94	117.000	125.000
	2021-391	3 ISE DILUTE	184 INTEGRA	93 ROCHE	33	121.848	1.925	1.58	118.000	126.000
	2021-391	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	121.483	1.523	1.25	117.000	125.000
	2021-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	123.125	1.691	1.37	119.000	127.000
	2021-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	122.667	1.921	1.57	119.000	127.000
	2021-391	3 ISE DILUTE	203 ROCHE	93 ROCHE	31	121.645	2.208	1.81	118.000	126.000
	2021-391	3 ISE DILUTE	203 ROCHE	93 ROCHE	31	121.750	2.107	1.73	118.000	126.000
	2021-391	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	11	120.000	2.000	1.67	116.000	124.000
	2021-391	4 ISE UNDIL.	165 VITROS	63 KODAK	88	119.500	1.500	1.26	116.000	124.000
	2021-391	4 ISE UNDIL.	165 VITROS	63 KODAK	88	120.151	1.688	1.40	116.000	124.000
	2021-391	4 ISE UNDIL.	165 VITROS	156 Kit 32	68	119.500	1.500	1.26	116.000	124.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-391	4 ISE UNDIL.	165 VITROS	156 Kit 32		68	119.851	1.677	1.40	116.000	124.000	
2021-391	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		26	122.462	2.561	2.09	118.000	126.000	
2021-391	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		26	121.500	1.500	1.23	118.000	126.000	
2021-391	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		11	121.091	1.928	1.59	117.000	125.000	
2021-391	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		80	123.975	1.835	1.48	120.000	128.000	
2021-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	122.583	2.178	1.78	119.000	127.000	
2021-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	122.500	2.062	1.68	119.000	127.000	
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2021-392	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	146.303	2.181	1.49	142.000	150.000	
2021-392	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	146.667	3.145	2.14	143.000	151.000	
2021-392	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	147.207	2.784	1.89	143.000	151.000	
2021-392	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	149.133	2.093	1.40	145.000	153.000	
2021-392	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	149.750	2.634	1.76	146.000	154.000	
2021-392	3 ISE DILUTE	203 ROCHE	93 ROCHE		31	148.125	1.364	0.92	144.000	152.000	
2021-392	3 ISE DILUTE	203 ROCHE	93 ROCHE		31	147.806	2.131	1.44	144.000	152.000	
2021-392	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		11	145.727	1.420	0.97	142.000	150.000	
2021-392	4 ISE UNDIL.	165 VITROS	63 KODAK		88	148.092	2.778	1.88	144.000	152.000	
2021-392	4 ISE UNDIL.	165 VITROS	63 KODAK		88	148.750	2.278	1.53	145.000	153.000	
2021-392	4 ISE UNDIL.	165 VITROS	156 Kit 32		68	147.940	2.298	1.55	144.000	152.000	
2021-392	4 ISE UNDIL.	165 VITROS	156 Kit 32		68	148.750	2.278	1.53	145.000	153.000	
2021-392	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		26	146.667	3.636	2.48	143.000	151.000	
2021-392	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		26	146.846	1.854	1.26	143.000	151.000	
2021-392	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		11	146.182	1.402	0.96	142.000	150.000	
2021-392	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		80	147.962	2.173	1.47	144.000	152.000	
2021-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	145.250	1.920	1.32	141.000	149.000	
2021-392	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		12	147.083	2.722	1.85	143.000	151.000	
2021-393	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	130.167	2.115	1.62	126.000	134.000	
2021-393	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	131.091	1.975	1.51	127.000	135.000	
2021-393	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	132.172	1.577	1.19	128.000	136.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2021-393	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	133.375	1.654	1.24	129.000	137.000	
2021-393	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	133.333	2.211	1.66	129.000	137.000	
2021-393	3	ISE DILUTE	203 ROCHE	93	ROCHE	31	132.129	1.791	1.36	128.000	136.000	
2021-393	3	ISE DILUTE	203 ROCHE	93	ROCHE	31	131.750	1.639	1.24	128.000	136.000	
2021-393	3	ISE DILUTE	215 B.C DXC	134	BECKM.COUL	11	131.273	1.213	0.92	127.000	135.000	
2021-393	4	ISE UNDIL.	165 VITROS	63	KODAK	88	131.000	1.871	1.43	127.000	135.000	
2021-393	4	ISE UNDIL.	165 VITROS	63	KODAK	88	131.011	1.974	1.51	127.000	135.000	
2021-393	4	ISE UNDIL.	165 VITROS	156	Kit 32	68	131.000	1.871	1.43	127.000	135.000	
2021-393	4	ISE UNDIL.	165 VITROS	156	Kit 32	68	130.765	1.637	1.25	127.000	135.000	
2021-393	4	ISE UNDIL.	191 RX-DAYTONA	126	RANDOX	26	131.154	1.634	1.25	127.000	135.000	
2021-393	4	ISE UNDIL.	191 RX-DAYTONA	126	RANDOX	26	130.667	2.427	1.86	127.000	135.000	
2021-393	4	ISE UNDIL.	199 Rx-IMOLA	126	RANDOX	11	130.182	1.029	0.79	126.000	134.000	
2021-393	6	QUICK LYTE	38 SIEMENS DI	143	SIEMENS	80	133.405	1.959	1.47	129.000	137.000	
2021-393	6	QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	132.583	2.564	1.93	129.000	137.000	
2021-393	6	QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	130.875	1.452	1.11	127.000	135.000	
023 SODIUM			CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00							
2021-394	3	ISE DILUTE	184 INTEGRA	93	ROCHE	33	148.212	2.434	1.64	144.000	152.000	
2021-394	3	ISE DILUTE	184 INTEGRA	93	ROCHE	33	147.667	2.285	1.55	144.000	152.000	
2021-394	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	29	148.690	2.245	1.51	145.000	153.000	
2021-394	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	151.200	1.939	1.28	147.000	155.000	
2021-394	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	151.125	2.260	1.50	147.000	155.000	
2021-394	3	ISE DILUTE	203 ROCHE	93	ROCHE	31	149.375	2.342	1.57	145.000	153.000	
2021-394	3	ISE DILUTE	203 ROCHE	93	ROCHE	31	149.839	2.002	1.34	146.000	154.000	
2021-394	3	ISE DILUTE	215 B.C DXC	134	BECKM.COUL	11	148.636	2.227	1.50	145.000	153.000	
2021-394	4	ISE UNDIL.	165 VITROS	63	KODAK	88	150.826	2.677	1.78	147.000	155.000	
2021-394	4	ISE UNDIL.	165 VITROS	63	KODAK	88	151.250	3.031	2.00	147.000	155.000	
2021-394	4	ISE UNDIL.	165 VITROS	156	Kit 32	68	150.632	2.578	1.71	147.000	155.000	
2021-394	4	ISE UNDIL.	165 VITROS	156	Kit 32	68	151.250	3.031	2.00	147.000	155.000	
2021-394	4	ISE UNDIL.	191 RX-DAYTONA	126	RANDOX	26	150.167	3.484	2.32	146.000	154.000	
2021-394	4	ISE UNDIL.	191 RX-DAYTONA	126	RANDOX	26	149.385	2.288	1.53	145.000	153.000	
2021-394	4	ISE UNDIL.	199 Rx-IMOLA	126	RANDOX	11	148.000	2.523	1.70	144.000	152.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-394	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	80	149.709	2.307	1.54	146.000	154.000	
2021-394	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	147.250	2.107	1.43	143.000	151.000	
2021-394	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	148.167	2.640	1.78	144.000	152.000	
023	SODIUM	CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2021-395	3 ISE DILUTE	184 INTEGRA	93	ROCHE	33	159.000	2.582	1.62	155.000	163.000	
2021-395	3 ISE DILUTE	184 INTEGRA	93	ROCHE	33	158.242	2.686	1.70	154.000	162.000	
2021-395	3 ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	29	160.241	2.661	1.66	156.000	164.000	
2021-395	3 ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	163.375	2.955	1.81	159.000	167.000	
2021-395	3 ISE DILUTE	197 ARCHITECT	1	ABBOTT	15	162.000	2.733	1.69	158.000	166.000	
2021-395	3 ISE DILUTE	203 ROCHE	93	ROCHE	31	161.065	3.482	2.16	157.000	165.000	
2021-395	3 ISE DILUTE	203 ROCHE	93	ROCHE	31	162.125	2.759	1.70	158.000	166.000	
2021-395	3 ISE DILUTE	215 B.C DXC	134	BECKM.COUL	11	160.364	2.346	1.46	156.000	164.000	
2021-395	4 ISE UNLIL.	165 VITROS	63	KODAK	88	164.750	1.920	1.17	161.000	169.000	
2021-395	4 ISE UNLIL.	165 VITROS	63	KODAK	88	163.453	3.052	1.87	159.000	167.000	
2021-395	4 ISE UNLIL.	165 VITROS	156	Kit 32	68	164.750	1.920	1.17	161.000	169.000	
2021-395	4 ISE UNLIL.	165 VITROS	156	Kit 32	68	163.463	2.488	1.52	159.000	167.000	
2021-395	4 ISE UNLIL.	191 RX-DAYTONA	126	RANDOX	26	161.000	3.431	2.13	157.000	165.000	
2021-395	4 ISE UNLIL.	191 RX-DAYTONA	126	RANDOX	26	160.500	4.153	2.59	157.000	165.000	
2021-395	4 ISE UNLIL.	199 Rx-IMOLA	126	RANDOX	11	160.000	2.697	1.69	156.000	164.000	
2021-395	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	80	159.550	2.549	1.60	156.000	164.000	
2021-395	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	158.417	3.095	1.95	154.000	162.000	
2021-395	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	12	157.375	1.728	1.10	153.000	161.000	
024	TOTAL PROTEIN	CLIA TEST ID #: 485	EVALUATION CRITERIA: TARGET VALUE 10 %								
2021-391	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	3.090	0.098	3.16	2.800	3.400	
2021-391	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	3.075	0.083	2.70	2.800	3.400	
2021-391	1 BIURET	165 VITROS	63	KODAK	148	2.978	0.086	2.87	2.700	3.300	
2021-391	1 BIURET	165 VITROS	63	KODAK	148	2.900	0.122	4.22	2.600	3.200	
2021-391	1 BIURET	184 INTEGRA	93	ROCHE	29	2.866	0.099	3.46	2.600	3.200	
2021-391	1 BIURET	186 BECK. COUL	76	AU-BECKMAN	25	2.938	0.081	2.75	2.600	3.200	
2021-391	1 BIURET	191 RX-DAYTONA	126	RANDOX	33	3.024	0.085	2.82	2.700	3.300	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-391	1 BIURET	197 ARCHITECT	1	ABBOTT	20	2.890	0.077	2.66	2.600	3.200		
2021-391	1 BIURET	199 Rx-IMOLA	126	RANDOX	10	3.010	0.094	3.13	2.700	3.300		
2021-391	1 BIURET	203 ROCHE	93	ROCHE	31	3.010	0.070	2.33	2.700	3.300		
2021-391	1 BIURET	212 S. D. EXL	143	SIEMENS	12	3.117	0.080	2.56	2.800	3.400		
2021-391	1 BIURET	215 B.C DXC	134	BECKM.COUL	11	2.955	0.156	5.28	2.700	3.300		
024	TOTAL PROTEIN	CLIA TEST ID #: 485	EVALUATION CRITERIA: TARGET VALUE 10 %									
2021-392	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	6.450	0.112	1.73	5.800	7.100		
2021-392	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	6.525	0.203	3.12	5.900	7.200		
2021-392	1 BIURET	165 VITROS	63	KODAK	148	6.050	0.269	4.45	5.400	6.700		
2021-392	1 BIURET	165 VITROS	63	KODAK	148	5.954	0.139	2.33	5.400	6.500		
2021-392	1 BIURET	184 INTEGRA	93	ROCHE	29	6.052	0.175	2.90	5.400	6.700		
2021-392	1 BIURET	186 BECK. COUL	76	AU-BECKMAN	25	6.133	0.170	2.77	5.500	6.700		
2021-392	1 BIURET	191 RX-DAYTONA	126	RANDOX	33	6.264	0.157	2.51	5.600	6.900		
2021-392	1 BIURET	197 ARCHITECT	1	ABBOTT	20	6.275	0.122	1.94	5.600	6.900		
2021-392	1 BIURET	199 Rx-IMOLA	126	RANDOX	10	6.270	0.241	3.84	5.600	6.900		
2021-392	1 BIURET	203 ROCHE	93	ROCHE	31	6.239	0.150	2.40	5.600	6.900		
2021-392	1 BIURET	212 S. D. EXL	143	SIEMENS	12	6.500	0.196	3.01	5.900	7.200		
2021-392	1 BIURET	215 B.C DXC	134	BECKM.COUL	11	6.191	0.227	3.67	5.600	6.800		
2021-393	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	4.529	0.134	2.96	4.100	5.000		
2021-393	1 BIURET	38 SIEMENS DI	143	SIEMENS	80	4.550	0.112	2.46	4.100	5.000		
2021-393	1 BIURET	165 VITROS	63	KODAK	148	4.240	0.109	2.57	3.800	4.700		
2021-393	1 BIURET	165 VITROS	63	KODAK	148	4.200	0.212	5.05	3.800	4.600		
2021-393	1 BIURET	184 INTEGRA	93	ROCHE	29	4.197	0.100	2.38	3.800	4.600		
2021-393	1 BIURET	186 BECK. COUL	76	AU-BECKMAN	25	4.271	0.117	2.74	3.800	4.700		
2021-393	1 BIURET	191 RX-DAYTONA	126	RANDOX	33	4.385	0.102	2.32	3.900	4.800		
2021-393	1 BIURET	197 ARCHITECT	1	ABBOTT	20	4.275	0.118	2.76	3.800	4.700		
2021-393	1 BIURET	199 Rx-IMOLA	126	RANDOX	10	4.350	0.169	3.88	3.900	4.800		
2021-393	1 BIURET	203 ROCHE	93	ROCHE	31	4.390	0.098	2.23	4.000	4.800		
2021-393	1 BIURET	212 S. D. EXL	143	SIEMENS	12	4.492	0.126	2.80	4.000	4.900		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-393	1 BIURET	215	B.C DXC	134	BECKM.COUL	11	4.327	0.181	4.19	3.900	4.800
024	TOTAL PROTEIN	CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2021-394	1 BIURET	38	SIEMENS DI	143	SIEMENS	80	6.950	0.112	1.61	6.300	7.600
2021-394	1 BIURET	38	SIEMENS DI	143	SIEMENS	80	6.985	0.258	3.69	6.300	7.700
2021-394	1 BIURET	165	VITROS	63	KODAK	148	6.425	0.286	4.45	5.800	7.100
2021-394	1 BIURET	165	VITROS	63	KODAK	148	6.372	0.140	2.19	5.700	7.000
2021-394	1 BIURET	184	INTEGRA	93	ROCHE	29	6.497	0.177	2.73	5.800	7.100
2021-394	1 BIURET	186	BECK. COUL	76	AU-BECKMAN	25	6.567	0.175	2.66	5.900	7.200
2021-394	1 BIURET	191	RX-DAYTONA	126	RANDOX	33	6.727	0.140	2.08	6.100	7.400
2021-394	1 BIURET	197	ARCHITECT	1	ABBOTT	20	6.700	0.121	1.81	6.000	7.400
2021-394	1 BIURET	199	Rx-IMOLA	126	RANDOX	10	6.680	0.214	3.20	6.000	7.300
2021-394	1 BIURET	203	ROCHE	93	ROCHE	31	6.681	0.142	2.13	6.000	7.300
2021-394	1 BIURET	212	S. D. EXL	143	SIEMENS	12	6.933	0.209	3.02	6.200	7.600
2021-394	1 BIURET	215	B.C DXC	134	BECKM.COUL	11	6.718	0.241	3.58	6.000	7.400
2021-395	1 BIURET	38	SIEMENS DI	143	SIEMENS	80	8.544	0.299	3.50	7.700	9.400
2021-395	1 BIURET	38	SIEMENS DI	143	SIEMENS	80	8.500	0.141	1.66	7.700	9.400
2021-395	1 BIURET	165	VITROS	63	KODAK	148	7.761	0.190	2.44	7.000	8.500
2021-395	1 BIURET	165	VITROS	63	KODAK	148	7.700	0.332	4.31	6.900	8.500
2021-395	1 BIURET	184	INTEGRA	93	ROCHE	29	7.900	0.172	2.18	7.100	8.700
2021-395	1 BIURET	186	BECK. COUL	76	AU-BECKMAN	25	8.083	0.302	3.74	7.300	8.900
2021-395	1 BIURET	191	RX-DAYTONA	126	RANDOX	33	8.194	0.216	2.64	7.400	9.000
2021-395	1 BIURET	197	ARCHITECT	1	ABBOTT	20	8.250	0.204	2.47	7.400	9.100
2021-395	1 BIURET	199	Rx-IMOLA	126	RANDOX	10	8.120	0.376	4.63	7.300	8.900
2021-395	1 BIURET	203	ROCHE	93	ROCHE	31	8.168	0.189	2.31	7.400	9.000
2021-395	1 BIURET	212	S. D. EXL	143	SIEMENS	12	8.425	0.289	3.43	7.600	9.300
2021-395	1 BIURET	215	B.C DXC	134	BECKM.COUL	11	8.227	0.218	2.65	7.400	9.100
025	TRIGLYCERIDES	CLIA TEST ID #: 495				EVALUATION CRITERIA: TARGET VALUE 25 %					
2021-391	2 ENZ. COLO.	165	VITROS	63	KODAK	150	88.993	3.341	3.75	67.000	111.000
2021-391	2 ENZ. COLO.	184	INTEGRA	93	ROCHE	10	86.800	1.887	2.17	65.000	109.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-391	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	85.353	2.140	2.51	64.000	107.000	
2021-391	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		78	76.500	3.046	3.98	57.000	96.000	
2021-391	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		13	74.615	2.923	3.92	56.000	93.000	
2021-391	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		27	82.704	2.942	3.56	62.000	103.000	
2021-391	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		24	82.958	4.430	5.34	62.000	104.000	
2021-391	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		10	96.600	4.716	4.88	72.000	121.000	
2021-391	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		26	87.231	2.873	3.29	65.000	109.000	
2021-391	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		10	79.900	5.281	6.61	60.000	100.000	
2021-391	10 ENZYM. GPO	203 ROCHE	93 ROCHE		20	85.900	3.375	3.93	64.000	107.000	
2021-391	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		12	80.250	1.422	1.77	60.000	100.000	
025 TRIGLYCERIDES		CLIA TEST ID #: 495		EVALUATION CRITERIA: TARGET VALUE 25 %							
2021-392	2 ENZ. COLO.	165 VITROS	63 KODAK		150	194.588	6.590	3.39	146.000	243.000	
2021-392	2 ENZ. COLO.	184 INTEGRA	93 ROCHE		10	170.600	3.980	2.33	128.000	213.000	
2021-392	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	170.375	2.497	1.47	128.000	213.000	
2021-392	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		78	166.949	3.869	2.32	125.000	209.000	
2021-392	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		13	164.615	5.400	3.28	123.000	206.000	
2021-392	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		27	172.185	6.902	4.01	129.000	215.000	
2021-392	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		24	167.250	7.166	4.28	125.000	209.000	
2021-392	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		10	182.600	6.530	3.58	137.000	228.000	
2021-392	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		26	172.500	4.281	2.48	129.000	216.000	
2021-392	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		10	163.200	7.277	4.46	122.000	204.000	
2021-392	10 ENZYM. GPO	203 ROCHE	93 ROCHE		20	169.895	3.655	2.15	127.000	212.000	
2021-392	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		12	166.500	2.958	1.78	125.000	208.000	
2021-393	2 ENZ. COLO.	165 VITROS	63 KODAK		150	133.664	4.733	3.54	100.000	167.000	
2021-393	2 ENZ. COLO.	184 INTEGRA	93 ROCHE		10	123.100	2.508	2.04	92.000	154.000	
2021-393	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	122.412	3.219	2.63	92.000	153.000	
2021-393	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		78	114.714	3.636	3.17	86.000	143.000	
2021-393	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		13	112.231	3.423	3.05	84.000	140.000	
2021-393	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		27	120.704	5.304	4.39	91.000	151.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-393	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	24	118.875	5.191	4.37	89.000	149.000
2021-393	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	132.100	4.323	3.27	99.000	165.000
2021-393	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	26	123.538	3.342	2.71	93.000	154.000
2021-393	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	10	115.100	6.715	5.83	86.000	144.000
2021-393	10	ENZYM. GPO	203	ROCHE	93	ROCHE	20	122.316	3.010	2.46	92.000	153.000
2021-393	11	GLY PO4 OX	197	ARCHITECT	1	ABBOTT	12	116.917	2.216	1.90	88.000	146.000
025 TRIGLYCERIDES			CLIA TEST ID #: 495			EVALUATION CRITERIA: TARGET VALUE 25 %						
2021-394	2	ENZ. COLO.	165	VITROS	63	KODAK	150	209.000	6.529	3.12	157.000	261.000
2021-394	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	10	181.000	5.000	2.76	136.000	226.000
2021-394	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	182.235	4.413	2.42	137.000	228.000
2021-394	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	78	178.390	4.710	2.64	134.000	223.000
2021-394	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	13	175.615	5.541	3.15	132.000	220.000
2021-394	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	27	184.074	5.881	3.19	138.000	230.000
2021-394	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	24	177.500	7.124	4.01	133.000	222.000
2021-394	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	197.300	7.695	3.90	148.000	247.000
2021-394	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	26	183.615	5.077	2.76	138.000	230.000
2021-394	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	10	173.500	11.741	6.77	130.000	217.000
2021-394	10	ENZYM. GPO	203	ROCHE	93	ROCHE	20	180.579	3.951	2.19	135.000	226.000
2021-394	11	GLY PO4 OX	197	ARCHITECT	1	ABBOTT	12	178.083	3.148	1.77	134.000	223.000
2021-395	2	ENZ. COLO.	165	VITROS	63	KODAK	150	258.839	7.852	3.03	194.000	324.000
2021-395	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	10	216.100	4.527	2.09	162.000	270.000
2021-395	2	ENZ. COLO.	203	ROCHE	93	ROCHE	17	217.882	4.600	2.11	163.000	272.000
2021-395	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	78	218.115	6.220	2.85	164.000	273.000
2021-395	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	13	213.923	8.499	3.97	160.000	267.000
2021-395	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	27	226.185	10.100	4.47	170.000	283.000
2021-395	9	GPO TRIND.	191	RX-DAYTONA	126	RANDOX	24	215.792	7.858	3.64	162.000	270.000
2021-395	9	GPO TRIND.	215	B.C DXC	134	BECKM.COUL	10	238.500	8.237	3.45	179.000	298.000
2021-395	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	26	219.423	5.108	2.33	165.000	274.000
2021-395	10	ENZYM. GPO	191	RX-DAYTONA	126	RANDOX	10	209.000	11.489	5.50	157.000	261.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-395	10	ENZYM.	GPO	203	ROCHE	93	ROCHE	20	218.789	5.952	2.72	164.000	273.000
2021-395	11	GLY PO4	OX	197	ARCHITECT	1	ABBOTT	12	216.917	3.040	1.40	163.000	271.000
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}									
2021-391	2	UREAS.GLDH		38	SIEMENS DI	143	SIEMENS	82	11.444	0.685	5.98	9.000	13.000
2021-391	2	UREAS.GLDH		184	INTEGRA	93	ROCHE	39	11.077	0.525	4.74	9.000	13.000
2021-391	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	11.759	0.625	5.31	10.000	14.000
2021-391	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	12.500	0.500	4.00	11.000	15.000
2021-391	2	UREAS.GLDH		191	RX-DAYTONA	126	RANDEX	32	11.645	0.863	7.41	10.000	14.000
2021-391	2	UREAS.GLDH		191	RX-DAYTONA	126	RANDEX	32	11.778	0.916	7.78	10.000	14.000
2021-391	2	UREAS.GLDH		203	ROCHE	93	ROCHE	39	11.026	0.530	4.81	9.000	13.000
2021-391	2	UREAS.GLDH		212	S. D. EXL	143	SIEMENS	12	11.333	0.943	8.32	9.000	13.000
2021-391	8	UREASE QUI		165	VITROS	63	KODAK	163	10.321	0.480	4.65	8.000	12.000
2021-391	11	UREASE		197	ARCHITECT	1	ABBOTT	11	11.364	0.881	7.76	9.000	13.000
2021-392	2	UREAS.GLDH		38	SIEMENS DI	143	SIEMENS	82	30.695	1.112	3.62	28.000	33.000
2021-392	2	UREAS.GLDH		184	INTEGRA	93	ROCHE	39	30.103	0.778	2.58	27.000	33.000
2021-392	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	31.333	0.943	3.01	29.000	34.000
2021-392	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	30.862	1.432	4.64	28.000	34.000
2021-392	2	UREAS.GLDH		191	RX-DAYTONA	126	RANDEX	32	29.556	1.707	5.78	27.000	32.000
2021-392	2	UREAS.GLDH		191	RX-DAYTONA	126	RANDEX	32	29.516	1.456	4.93	27.000	32.000
2021-392	2	UREAS.GLDH		203	ROCHE	93	ROCHE	39	29.795	1.264	4.24	27.000	32.000
2021-392	2	UREAS.GLDH		212	S. D. EXL	143	SIEMENS	12	30.167	0.898	2.98	27.000	33.000
2021-392	8	UREASE QUI		165	VITROS	63	KODAK	163	28.420	0.735	2.59	26.000	31.000
2021-392	11	UREASE		197	ARCHITECT	1	ABBOTT	11	30.636	1.494	4.88	28.000	33.000
2021-393	2	UREAS.GLDH		38	SIEMENS DI	143	SIEMENS	82	19.439	0.898	4.62	17.000	21.000
2021-393	2	UREAS.GLDH		184	INTEGRA	93	ROCHE	39	18.821	0.635	3.38	17.000	21.000
2021-393	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	19.679	0.889	4.52	18.000	22.000
2021-393	2	UREAS.GLDH		186	BECK. COUL	76	AU-BECKMAN	29	21.000	1.000	4.76	19.000	23.000
2021-393	2	UREAS.GLDH		191	RX-DAYTONA	126	RANDEX	32	19.250	1.275	6.62	17.000	21.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-393	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	32	19.111	1.100	5.76	17.000	21.000		
2021-393	2 UREAS.GLDH	203 ROCHE	93 ROCHE	39	18.949	0.677	3.58	17.000	21.000		
2021-393	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	12	19.500	0.957	4.91	18.000	22.000		
2021-393	8 UREASE QUI	165 VITROS	63 KODAK	163	17.770	0.514	2.89	16.000	20.000		
2021-393	11 UREASE	197 ARCHITECT	1 ABBOTT	11	19.000	1.044	5.50	17.000	21.000		
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505	EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}								
2021-394	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	82	33.432	1.217	3.64	30.000	36.000		
2021-394	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	39	32.769	1.165	3.55	30.000	36.000		
2021-394	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	35.167	1.067	3.03	32.000	38.000		
2021-394	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	33.414	1.352	4.05	30.000	36.000		
2021-394	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	32	32.222	1.872	5.81	29.000	35.000		
2021-394	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	32	31.844	2.123	6.67	29.000	35.000		
2021-394	2 UREAS.GLDH	203 ROCHE	93 ROCHE	39	32.385	0.923	2.85	29.000	35.000		
2021-394	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	12	32.833	1.213	3.70	30.000	36.000		
2021-394	8 UREASE QUI	165 VITROS	63 KODAK	163	31.227	0.770	2.47	28.000	34.000		
2021-394	11 UREASE	197 ARCHITECT	1 ABBOTT	11	33.636	1.720	5.11	31.000	37.000		
2021-395	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	82	42.378	1.487	3.51	39.000	46.000		
2021-395	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	39	41.744	1.103	2.64	38.000	46.000		
2021-395	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	44.833	0.898	2.00	41.000	49.000		
2021-395	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	42.724	2.318	5.43	39.000	47.000		
2021-395	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	32	38.778	2.250	5.80	35.000	42.000		
2021-395	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	32	39.344	2.975	7.56	36.000	43.000		
2021-395	2 UREAS.GLDH	203 ROCHE	93 ROCHE	39	41.359	1.209	2.92	38.000	45.000		
2021-395	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	12	41.333	1.650	3.99	38.000	45.000		
2021-395	8 UREASE QUI	165 VITROS	63 KODAK	163	40.099	0.947	2.36	36.000	44.000		
2021-395	11 UREASE	197 ARCHITECT	1 ABBOTT	11	42.636	1.920	4.50	39.000	46.000		
027 URIC ACID		CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %								
2021-391	2 URICASE	38 SIEMENS DI	143 SIEMENS	46	1.737	0.092	5.28	1.400	2.000		
2021-391	2 URICASE	165 VITROS	63 KODAK	100	1.715	0.064	3.74	1.400	2.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-391	2	URICASE	184	INTEGRA	93	ROCHE	17	1.765	0.076	4.32	1.500	2.100
2021-391	2	URICASE	197	ARCHITECT	1	ABBOTT	10	1.780	0.060	3.37	1.500	2.100
2021-391	2	URICASE	203	ROCHE	93	ROCHE	25	1.664	0.074	4.46	1.400	1.900
2021-391	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	1.858	0.075	4.03	1.500	2.200
2021-391	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	10	1.580	0.040	2.53	1.300	1.800
027 URIC ACID			CLIA TEST ID #: 515		EVALUATION CRITERIA: TARGET VALUE 17 %							
2021-392	2	URICASE	38	SIEMENS DI	143	SIEMENS	46	7.172	0.188	2.63	6.000	8.400
2021-392	2	URICASE	165	VITROS	63	KODAK	100	7.309	0.152	2.09	6.100	8.600
2021-392	2	URICASE	184	INTEGRA	93	ROCHE	17	7.541	0.277	3.67	6.300	8.800
2021-392	2	URICASE	197	ARCHITECT	1	ABBOTT	10	7.480	0.140	1.87	6.200	8.800
2021-392	2	URICASE	203	ROCHE	93	ROCHE	25	7.200	0.185	2.58	6.000	8.400
2021-392	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	7.342	0.118	1.61	6.100	8.600
2021-392	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	10	6.810	0.104	1.53	5.700	8.000
2021-393	2	URICASE	38	SIEMENS DI	143	SIEMENS	46	4.002	0.129	3.23	3.300	4.700
2021-393	2	URICASE	165	VITROS	63	KODAK	100	4.058	0.096	2.38	3.400	4.700
2021-393	2	URICASE	184	INTEGRA	93	ROCHE	17	4.206	0.139	3.31	3.500	4.900
2021-393	2	URICASE	197	ARCHITECT	1	ABBOTT	10	4.130	0.078	1.89	3.400	4.800
2021-393	2	URICASE	203	ROCHE	93	ROCHE	25	4.004	0.096	2.39	3.300	4.700
2021-393	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	4.163	0.087	2.09	3.500	4.900
2021-393	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	10	3.920	0.087	2.22	3.300	4.600
2021-394	2	URICASE	38	SIEMENS DI	143	SIEMENS	46	7.989	0.192	2.41	6.600	9.300
2021-394	2	URICASE	165	VITROS	63	KODAK	100	8.136	0.153	1.88	6.800	9.500
2021-394	2	URICASE	184	INTEGRA	93	ROCHE	17	8.382	0.307	3.67	7.000	9.800
2021-394	2	URICASE	197	ARCHITECT	1	ABBOTT	10	8.350	0.136	1.63	6.900	9.800
2021-394	2	URICASE	203	ROCHE	93	ROCHE	25	8.016	0.159	1.99	6.700	9.400
2021-394	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	8.100	0.152	1.88	6.700	9.500
2021-394	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	10	7.570	0.168	2.21	6.300	8.900

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
027	URIC ACID	CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %									
2021-395	2 URICASE	38 SIEMENS DI	143 SIEMENS	46	10.483	0.319	3.04	8.700	12.300			
2021-395	2 URICASE	165 VITROS	63 KODAK	100	10.774	0.219	2.03	8.900	12.600			
2021-395	2 URICASE	184 INTEGRA	93 ROCHE	17	10.971	0.375	3.42	9.100	12.800			
2021-395	2 URICASE	197 ARCHITECT	1 ABBOTT	10	11.000	0.205	1.86	9.100	12.900			
2021-395	2 URICASE	203 ROCHE	93 ROCHE	25	10.532	0.241	2.29	8.700	12.300			
2021-395	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN	19	10.679	0.248	2.33	8.900	12.500			
2021-395	5 URICASE 37	215 B.C DXC	134 BECKM.COUL	10	9.810	0.164	1.67	8.100	11.500			
135	GGT (GAMMA GLU. TRA)	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %									
2021-391	5 KINETIC 37	165 VITROS	63 KODAK	23	20.565	1.690	8.22	16.000	25.000			
2021-392	5 KINETIC 37	165 VITROS	63 KODAK	23	393.130	9.023	2.30	315.000	472.000			
2021-393	5 KINETIC 37	165 VITROS	63 KODAK	23	200.391	4.094	2.04	160.000	240.000			
2021-394	5 KINETIC 37	165 VITROS	63 KODAK	23	437.391	13.177	3.01	350.000	525.000			
2021-395	5 KINETIC 37	165 VITROS	63 KODAK	23	564.652	21.186	3.75	452.000	678.000			
136	LACTIC ACID											
2021-391	1 LACTATE-P.	165 VITROS	63 KODAK	19	0.500	0.000	0.00	0.400	0.600			
2021-392	1 LACTATE-P.	165 VITROS	63 KODAK	19	3.658	0.082	2.23	2.900	4.400			
2021-393	1 LACTATE-P.	165 VITROS	63 KODAK	19	1.495	0.051	3.41	1.200	1.800			
2021-394	1 LACTATE-P.	165 VITROS	63 KODAK	19	4.221	0.061	1.45	3.400	5.100			
2021-395	1 LACTATE-P.	165 VITROS	63 KODAK	19	5.711	0.102	1.79	4.600	6.900			

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
137	LIPASE	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-391	1	COLORIMET.	203	ROCHE	93	ROCHE	14	16.071	0.799	4.97	13.000	19.000
2021-391	2	ENZIMATIC	165	VITROS	63	KODAK	84	59.585	3.582	6.01	48.000	72.000
2021-391	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	20.100	1.446	7.19	16.000	24.000
2021-391	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	16	14.625	1.833	12.53	12.000	18.000
2021-391	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	16.000	0.784	4.90	13.000	19.000
2021-391	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	50.333	8.158	16.21	40.000	60.000
2021-391	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	43.500	10.210	23.47	35.000	52.000
2021-392	1	COLORIMET.	203	ROCHE	93	ROCHE	14	60.643	5.204	8.58	49.000	73.000
2021-392	2	ENZIMATIC	165	VITROS	63	KODAK	84	599.663	22.883	3.82	480.000	720.000
2021-392	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	57.400	2.458	4.28	46.000	69.000
2021-392	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	16	68.500	2.958	4.32	55.000	82.000
2021-392	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	57.923	5.121	8.84	46.000	70.000
2021-392	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	232.250	13.881	5.98	186.000	279.000
2021-392	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	204.389	18.524	9.06	164.000	245.000
2021-393	1	COLORIMET.	203	ROCHE	93	ROCHE	14	35.214	2.076	5.90	28.000	42.000
2021-393	2	ENZIMATIC	165	VITROS	63	KODAK	84	299.711	10.692	3.57	240.000	360.000
2021-393	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	36.700	1.735	4.73	29.000	44.000
2021-393	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	16	37.563	2.150	5.72	30.000	45.000
2021-393	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	34.154	1.994	5.84	27.000	41.000
2021-393	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	116.333	10.290	8.85	93.000	140.000
2021-393	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	125.000	8.367	6.69	100.000	150.000
2021-394	1	COLORIMET.	203	ROCHE	93	ROCHE	14	67.071	5.599	8.35	54.000	80.000
2021-394	2	ENZIMATIC	165	VITROS	63	KODAK	84	669.929	26.153	3.90	536.000	804.000
2021-394	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	10	62.600	2.728	4.36	50.000	75.000
2021-394	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	16	75.313	2.973	3.95	60.000	90.000
2021-394	4	ENZ/COLOR	203	ROCHE	93	ROCHE	13	63.615	5.772	9.07	51.000	76.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-394	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	18	256.000	14.731	5.75	205.000	307.000
2021-394	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	18	227.722	16.124	7.08	182.000	273.000
137	LIPASE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %						
2021-395	1	COLORIMET.	203 ROCHE	93	ROCHE	14	87.143	8.140	9.34	70.000	105.000
2021-395	2	ENZIMATIC	165 VITROS	63	KODAK	84	924.916	66.657	7.21	740.000	1110.000
2021-395	2	ENZIMATIC	215 B.C DXC	134	BECKM.COUL	10	78.300	3.466	4.43	63.000	94.000
2021-395	4	ENZ/COLOR	186 BECK. COUL	76	AU-BECKMAN	16	100.688	6.282	6.24	81.000	121.000
2021-395	4	ENZ/COLOR	203 ROCHE	93	ROCHE	13	81.692	8.361	10.23	65.000	98.000
2021-395	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	18	295.333	20.235	6.85	236.000	354.000
2021-395	5	SIEMENS 37	38 SIEMENS DI	143	SIEMENS	18	345.750	18.005	5.21	277.000	415.000
138	PHOSPHORUS										
2021-391	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	45	1.493	0.099	6.60	1.200	1.800
2021-391	1	PHOSPHOMO.	165 VITROS	63	KODAK	87	1.633	0.084	5.15	1.300	2.000
2021-391	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	1.413	0.093	6.56	1.100	1.700
2021-391	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	16	1.340	0.071	5.31	1.100	1.600
2021-391	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	31	1.360	0.084	6.18	1.100	1.600
2021-391	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	11	1.364	0.115	8.43	1.100	1.600
2021-391	2	END POINT	184 INTEGRA	93	ROCHE	10	1.340	0.049	3.66	1.100	1.600
2021-392	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	45	4.442	0.127	2.87	3.600	5.300
2021-392	1	PHOSPHOMO.	165 VITROS	63	KODAK	87	4.187	0.110	2.62	3.300	5.000
2021-392	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	4.294	0.134	3.13	3.400	5.200
2021-392	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	16	4.175	0.103	2.47	3.300	5.000
2021-392	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	31	4.220	0.101	2.40	3.400	5.100
2021-392	1	PHOSPHOMO.	215 B.C DXC	134	BECKM.COUL	11	4.382	0.180	4.11	3.500	5.300
2021-392	2	END POINT	184 INTEGRA	93	ROCHE	10	4.190	0.130	3.10	3.400	5.000
2021-393	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	45	2.800	0.090	3.23	2.200	3.400
2021-393	1	PHOSPHOMO.	165 VITROS	63	KODAK	87	2.855	0.086	3.01	2.300	3.400
2021-393	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	2.706	0.114	4.23	2.200	3.200

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-393	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	16	2.593	0.077	2.98	2.100	3.100
2021-393	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	31	2.650	0.099	3.74	2.100	3.200
2021-393	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	11	2.727	0.114	4.16	2.200	3.300
2021-393	2	END POINT	184	INTEGRA	93	ROCHE	10	2.610	0.030	1.15	2.100	3.100
138	PHOSPHORUS	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2021-394	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	45	4.824	0.134	2.77	3.900	5.800
2021-394	1	PHOSPHOMO.	165	VITROS	63	KODAK	87	4.517	0.114	2.52	3.600	5.400
2021-394	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	16	4.669	0.179	3.84	3.700	5.600
2021-394	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	16	4.538	0.136	3.01	3.600	5.400
2021-394	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	31	4.648	0.150	3.23	3.700	5.600
2021-394	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	11	4.873	0.242	4.96	3.900	5.800
2021-394	2	END POINT	184	INTEGRA	93	ROCHE	10	4.540	0.092	2.02	3.600	5.400
2021-395	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	45	6.055	0.160	2.65	4.800	7.300
2021-395	1	PHOSPHOMO.	165	VITROS	63	KODAK	87	5.619	0.142	2.52	4.500	6.700
2021-395	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	16	5.825	0.219	3.77	4.700	7.000
2021-395	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	16	5.775	0.195	3.38	4.600	6.900
2021-395	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	31	5.887	0.179	3.04	4.700	7.100
2021-395	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	11	6.273	0.328	5.23	5.000	7.500
2021-395	2	END POINT	184	INTEGRA	93	ROCHE	10	5.700	0.100	1.75	4.600	6.800
169	CO2					EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-391	1	Generic Me	165	VITROS	63	KODAK	17	14.412	1.942	13.48	9.000	20.000
2021-391	1	Generic Me	184	INTEGRA	93	ROCHE	13	13.462	1.737	12.90	8.000	19.000
2021-391	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	22	12.864	1.290	10.03	9.000	17.000
2021-391	1	Generic Me	203	ROCHE	93	ROCHE	16	12.938	1.676	12.95	8.000	18.000
2021-391	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	67	15.773	2.159	13.69	9.000	22.000
2021-391	2	ENZYMATIC	165	VITROS	63	KODAK	71	13.648	1.602	11.74	9.000	18.000
2021-391	2	ENZYMATIC	165	VITROS	156	Kit 32	61	12.900	1.650	12.79	8.000	18.000
2021-391	2	ENZYMATIC	203	ROCHE	93	ROCHE	14	12.500	1.402	11.21	8.000	17.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
169	CO2	CLIA TEST ID #: 0			EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-392	1	Generic Me	165 VITROS	63	KODAK	17	24.294	1.563	6.43	20.000	29.000
2021-392	1	Generic Me	184 INTEGRA	93	ROCHE	13	22.308	2.266	10.16	16.000	29.000
2021-392	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	22.818	1.992	8.73	17.000	29.000
2021-392	1	Generic Me	203 ROCHE	93	ROCHE	16	23.000	1.751	7.61	18.000	28.000
2021-392	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	25.258	2.554	10.11	18.000	33.000
2021-392	2	ENZYMATIC	165 VITROS	63	KODAK	71	23.268	1.482	6.37	19.000	28.000
2021-392	2	ENZYMATIC	165 VITROS	156	Kit 32	61	23.115	1.719	7.44	18.000	28.000
2021-392	2	ENZYMATIC	203 ROCHE	93	ROCHE	14	21.429	1.841	8.59	16.000	27.000
2021-393	1	Generic Me	165 VITROS	63	KODAK	17	17.588	1.942	11.04	12.000	23.000
2021-393	1	Generic Me	184 INTEGRA	93	ROCHE	13	16.692	1.538	9.22	12.000	21.000
2021-393	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	16.727	1.452	8.68	12.000	21.000
2021-393	1	Generic Me	203 ROCHE	93	ROCHE	16	16.125	2.288	14.19	9.000	23.000
2021-393	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	18.985	2.513	13.23	11.000	27.000
2021-393	2	ENZYMATIC	165 VITROS	63	KODAK	71	16.971	1.373	8.09	13.000	21.000
2021-393	2	ENZYMATIC	165 VITROS	156	Kit 32	61	16.367	1.341	8.19	12.000	20.000
2021-393	2	ENZYMATIC	203 ROCHE	93	ROCHE	14	15.929	1.751	10.99	11.000	21.000
2021-394	1	Generic Me	165 VITROS	63	KODAK	17	24.941	1.954	7.84	19.000	31.000
2021-394	1	Generic Me	184 INTEGRA	93	ROCHE	13	23.231	1.671	7.19	18.000	28.000
2021-394	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	24.091	1.240	5.15	20.000	28.000
2021-394	1	Generic Me	203 ROCHE	93	ROCHE	16	23.625	3.140	13.29	14.000	33.000
2021-394	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	25.985	2.762	10.63	18.000	34.000
2021-394	2	ENZYMATIC	165 VITROS	63	KODAK	71	24.408	1.866	7.64	19.000	30.000
2021-394	2	ENZYMATIC	165 VITROS	156	Kit 32	61	24.148	1.745	7.22	19.000	29.000
2021-394	2	ENZYMATIC	203 ROCHE	93	ROCHE	14	23.071	1.907	8.27	17.000	29.000
2021-395	1	Generic Me	165 VITROS	63	KODAK	17	29.000	2.376	8.19	22.000	36.000
2021-395	1	Generic Me	184 INTEGRA	93	ROCHE	13	26.923	1.542	5.73	22.000	32.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-395	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	29.136	1.841	6.32	24.000	35.000
2021-395	1	Generic Me	203 ROCHE	93	ROCHE	16	27.875	3.480	12.48	17.000	38.000
2021-395	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	67	30.507	2.651	8.69	23.000	38.000
2021-395	2	ENZYMATIC	165 VITROS	63	KODAK	71	28.521	1.686	5.91	23.000	34.000
2021-395	2	ENZYMATIC	165 VITROS	156	Kit 32	61	28.820	1.954	6.78	23.000	35.000
2021-395	2	ENZYMATIC	203 ROCHE	93	ROCHE	14	27.643	2.158	7.81	21.000	34.000
170	DIRECT BILIRUBIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}							
2021-391	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	48	0.067	0.051	77.06	0	0.200
2021-391	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.022	0.042	187.08	0	0.100
2021-391	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.000	0.000	0.00	0.000	0.000
2021-391	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.200	0.000	0.00	0.200	0.200
2021-391	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	11	0.100	0.000	0.00	0.100	0.100
2021-392	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	48	0.981	0.073	7.40	0.800	1.200
2021-392	11	SPECT. WIT	165 VITROS	63	KODAK	63	1.489	0.057	3.81	1.200	1.800
2021-392	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.829	0.163	19.66	0.300	1.300
2021-392	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.900	0.085	9.39	0.600	1.200
2021-392	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	11	0.982	0.039	3.93	0.800	1.200
2021-393	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	48	0.396	0.041	10.26	0.300	0.500
2021-393	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.152	0.140	92.33	0	0.600
2021-393	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.689	0.031	4.56	0.600	0.800
2021-393	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.367	0.070	19.07	0.200	0.600
2021-393	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	11	0.400	0.074	18.46	0.200	0.600
2021-394	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	48	1.102	0.076	6.87	0.900	1.300
2021-394	11	SPECT. WIT	165 VITROS	63	KODAK	63	1.689	0.057	3.35	1.400	2.000
2021-394	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.974	0.152	15.64	0.500	1.400
2021-394	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	1.029	0.088	8.56	0.800	1.300
2021-394	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	11	1.118	0.057	5.14	0.900	1.300

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
170	DIRECT BILIRUBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 20 %	{GREATER}						
2021-395	4 JEND.-GROG	38 SIEMENS DI	143 SIEMENS	48	1.532	0.109	7.14	1.200	1.900			
2021-395	11 SPECT. WIT	165 VITROS	63 KODAK	63	1.439	0.184	12.78	0.900	2.000			
2021-395	11 SPECT. WIT	165 VITROS	63 KODAK	63	2.278	0.113	4.97	1.800	2.700			
2021-395	13 AZOBIL.W/D	203 ROCHE	93 ROCHE	15	1.400	0.085	6.04	1.100	1.700			
2021-395	16 DIAZO ENDP	215 B.C DXC	134 BECKM.COUL	11	1.509	0.067	4.43	1.200	1.800			
<hr/>												
SPEC: 10 CHEMISTRY		SUB: 20 BLOOD GASES		GRP: 30 BLOOD GASES								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
007	BLOOD GASES-pCO2	CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE	8	% OR 5.00	{GREATER}						
2021-441	265 ROCHE COBA			30	95.903	3.122	3.26	88.200	103.600			
2021-442	265 ROCHE COBA			30	52.570	1.309	2.49	47.600	57.600			
2021-443	265 ROCHE COBA			30	27.462	0.407	1.48	22.500	32.500			
2021-444	265 ROCHE COBA			30	63.427	1.235	1.95	58.400	68.500			
2021-445	265 ROCHE COBA			30	52.803	1.090	2.06	47.800	57.800			
008	BLOOD GASES-pH	CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04								
2021-441	265 ROCHE COBA			30	7.068	0.008	0.12	7.029	7.109			
2021-442	265 ROCHE COBA			30	7.336	0.006	0.08	7.296	7.376			
2021-443	265 ROCHE COBA			30	7.577	0.005	0.07	7.537	7.617			

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE		0.04				
2021-444	265 ROCHE COBA				30	7.196	0.005	0.07	7.156	7.236
2021-445	265 ROCHE COBA				30	7.334	0.005	0.07	7.294	7.374
009	BLOOD GASES-pO2		CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE		3 SD				
2021-441	265 ROCHE COBA				30	86.043	12.689	14.75	48.000	124.100
2021-442	265 ROCHE COBA				30	123.141	4.992	4.05	108.200	138.100
2021-443	265 ROCHE COBA				30	157.050	5.196	3.31	141.500	172.600
2021-444	265 ROCHE COBA				30	328.747	16.080	4.89	280.500	377.000
2021-445	265 ROCHE COBA				30	119.459	5.170	4.33	103.900	135.000
SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
029	FOLL. STIM. HOR.-FSH		CLIA TEST ID #: 0							
2021-421	116 ROCHE COBA				16	15.675	1.440	9.19	11.400	20.000
2021-421	149 ARC i2000s				10	14.520	0.914	6.30	11.800	17.300
2021-422	116 ROCHE COBA				16	23.969	2.845	11.87	15.400	32.500
2021-422	149 ARC i2000s				10	23.030	1.228	5.33	19.300	26.700
2021-423	116 ROCHE COBA				16	33.125	3.451	10.42	22.800	43.500
2021-423	149 ARC i2000s				10	32.330	2.084	6.45	26.100	38.600

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2021-424	116 ROCHE COBA			16	9.006	0.822	9.13	6.500	11.500	
2021-424	149 ARC i2000s			10	8.190	0.487	5.94	6.700	9.700	
2021-425	116 ROCHE COBA			16	26.953	2.385	8.85	19.800	34.100	
2021-425	149 ARC i2000s			10	25.220	1.442	5.72	20.900	29.500	
030	FREE THYROXINE	CLIA TEST ID #: 545								
2021-421	115 VITROS ECI			38	3.239	0.141	4.34	2.800	3.700	
2021-421	116 ROCHE COBA			44	1.550	0.106	6.81	1.200	1.900	
2021-421	125 B.C.ACCESS			41	1.393	0.075	5.35	1.200	1.600	
2021-421	149 ARC i2000s			24	1.246	0.076	6.12	1.000	1.500	
2021-421	160 SIEMENS EX			45	1.458	0.095	6.55	1.200	1.700	
2021-422	115 VITROS ECI			38	5.226	0.152	2.90	4.800	5.700	
2021-422	116 ROCHE COBA			44	2.370	0.155	6.52	1.900	2.800	
2021-422	125 B.C.ACCESS			41	2.023	0.085	4.21	1.800	2.300	
2021-422	149 ARC i2000s			24	1.954	0.119	6.09	1.600	2.300	
2021-422	160 SIEMENS EX			45	2.158	0.134	6.22	1.800	2.600	
2021-423	115 VITROS ECI			38	6.542	0.157	2.39	6.100	7.000	
2021-423	116 ROCHE COBA			44	3.193	0.243	7.62	2.500	3.900	
2021-423	125 B.C.ACCESS			41	2.528	0.095	3.75	2.200	2.800	
2021-423	149 ARC i2000s			24	2.500	0.161	6.43	2.000	3.000	
2021-423	160 SIEMENS EX			45	2.824	0.165	5.84	2.300	3.300	
2021-424	115 VITROS ECI			38	1.459	0.063	4.35	1.300	1.700	
2021-424	116 ROCHE COBA			44	0.859	0.075	8.71	0.600	1.100	
2021-424	125 B.C.ACCESS			41	0.863	0.076	8.77	0.600	1.100	
2021-424	149 ARC i2000s			24	0.717	0.037	5.20	0.600	0.800	
2021-424	160 SIEMENS EX			45	0.864	0.057	6.57	0.700	1.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
030	FREE THYROXINE	CLIA TEST ID #: 545	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-425	115 VITROS ECI			38	5.629	0.178	3.15	5.100	6.200
2021-425	116 ROCHE COBA			44	2.577	0.173	6.71	2.100	3.100
2021-425	125 B.C.ACCESS			41	2.163	0.112	5.18	1.800	2.500
2021-425	149 ARC i2000s			24	2.079	0.104	5.00	1.800	2.400
2021-425	160 SIEMENS EX			45	2.327	0.139	5.97	1.900	2.700
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555							
2021-421	115 VITROS ECI			18	757.789	32.052	4.23	661.600	853.900
2021-421	116 ROCHE COBA			20	437.115	20.961	4.80	374.200	500.000
2021-421	125 B.C.ACCESS			12	311.367	17.493	5.62	258.900	363.800
2021-422	115 VITROS ECI			18	1781.644	66.295	3.72	1582.800	1980.500
2021-422	116 ROCHE COBA			20	1055.070	56.165	5.32	886.600	1223.600
2021-422	125 B.C.ACCESS			12	613.550	39.085	6.37	496.300	730.800
2021-423	115 VITROS ECI			18	2924.344	114.654	3.92	2580.400	3268.300
2021-423	116 ROCHE COBA			20	1706.650	100.735	5.90	1404.400	2008.900
2021-423	125 B.C.ACCESS			12	853.575	45.034	5.28	718.500	988.700
2021-424	115 VITROS ECI			18	3.644	0.414	11.36	2.400	4.900
2021-424	116 ROCHE COBA			20	1.580	0.175	11.07	1.100	2.100
2021-424	125 B.C.ACCESS			12	1.855	0.099	5.32	1.600	2.200
2021-425	115 VITROS ECI			18	2067.506	76.168	3.68	1839.000	2296.000
2021-425	116 ROCHE COBA			20	1225.800	56.663	4.62	1055.800	1395.800
2021-425	125 B.C.ACCESS			12	686.650	50.934	7.42	533.800	839.500
034	PROGESTERONE	CLIA TEST ID #: 0							
2021-421	116 ROCHE COBA			10	13.360	0.612	4.58	11.500	15.200
2021-422	116 ROCHE COBA			10	27.480	1.116	4.06	24.100	30.800

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
034	PROGESTERONE	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2021-423	116 ROCHE COBA			10	38.920	1.785	4.59	33.600	44.300	
2021-424	116 ROCHE COBA			10	0.490	0.104	21.31	0.200	0.800	
2021-425	116 ROCHE COBA			10	30.900	1.348	4.36	26.900	34.900	
035	PROLACTIN									
2021-421	116 ROCHE COBA			18	12.976	0.763	5.88	10.700	15.300	
2021-422	116 ROCHE COBA			18	25.788	1.433	5.56	21.500	30.100	
2021-423	116 ROCHE COBA			18	40.435	2.607	6.45	32.600	48.300	
2021-424	116 ROCHE COBA			18	3.583	0.295	8.23	2.700	4.500	
2021-425	116 ROCHE COBA			18	29.494	1.601	5.43	24.700	34.300	
037	TESTOSTERONE									
2021-421	116 ROCHE COBA			10	6.280	0.690	10.98	4.200	8.300	
2021-422	116 ROCHE COBA			10	11.920	1.317	11.05	8.000	15.900	
2021-423	116 ROCHE COBA			10	15.120	1.144	7.56	11.700	18.600	
2021-424	116 ROCHE COBA			10	1.220	0.209	17.12	0.600	1.800	
2021-425	116 ROCHE COBA			10	13.340	1.565	11.73	8.600	18.000	
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585								
2021-421	106 TOSOH AIA			14	2.829	0.299	10.56	1.900	3.700	
2021-421	115 VITROS ECI			44	3.020	0.110	3.64	2.700	3.400	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-421	116	ROCHE COBA		49	2.825	0.107	3.79	2.500	3.100
2021-421	116	ROCHE COBA		49	2.389	0.191	8.00	1.800	3.000
2021-421	125	B.C.ACCESS		56	2.364	0.129	5.45	2.000	2.800
2021-421	149	ARC i2000s		24	2.392	0.104	4.34	2.100	2.700
2021-421	160	SIEMENS EX		71	2.269	0.134	5.89	1.900	2.700
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-422	106	TOSOH AIA		14	5.264	0.429	8.14	4.000	6.600
2021-422	115	VITROS ECI		44	5.753	0.188	3.28	5.200	6.300
2021-422	116	ROCHE COBA		49	4.556	0.347	7.61	3.500	5.600
2021-422	116	ROCHE COBA		49	5.006	0.189	3.77	4.400	5.600
2021-422	125	B.C.ACCESS		56	4.575	0.244	5.34	3.800	5.300
2021-422	149	ARC i2000s		24	4.550	0.212	4.66	3.900	5.200
2021-422	160	SIEMENS EX		71	4.248	0.241	5.68	3.500	5.000
2021-423	106	TOSOH AIA		14	7.750	0.498	6.43	6.300	9.200
2021-423	115	VITROS ECI		44	8.507	0.302	3.55	7.600	9.400
2021-423	116	ROCHE COBA		49	6.856	0.452	6.60	5.500	8.200
2021-423	116	ROCHE COBA		49	7.073	0.286	4.04	6.200	7.900
2021-423	125	B.C.ACCESS		56	6.889	0.394	5.72	5.700	8.100
2021-423	149	ARC i2000s		24	6.748	0.306	4.54	5.800	7.700
2021-423	160	SIEMENS EX		71	6.220	0.361	5.80	5.100	7.300
2021-424	106	TOSOH AIA		14	0.850	0.063	7.37	0.700	1.000
2021-424	115	VITROS ECI		44	0.784	0.042	5.40	0.700	0.900
2021-424	116	ROCHE COBA		49	0.994	0.052	5.20	0.800	1.100
2021-424	116	ROCHE COBA		49	0.711	0.110	15.47	0.400	1.000
2021-424	125	B.C.ACCESS		56	0.755	0.050	6.58	0.600	0.900
2021-424	149	ARC i2000s		24	0.688	0.033	4.81	0.600	0.800
2021-424	160	SIEMENS EX		71	0.751	0.058	7.70	0.600	0.900

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-425	106 TOSOH AIA			14	5.886	0.597	10.15	4.100	7.700
2021-425	115 VITROS ECI			44	6.528	0.233	3.56	5.800	7.200
2021-425	116 ROCHE COBA			49	5.211	0.420	8.06	4.000	6.500
2021-425	116 ROCHE COBA			49	5.557	0.192	3.46	5.000	6.100
2021-425	125 B.C.ACCESS			56	5.166	0.223	4.32	4.500	5.800
2021-425	149 ARC i2000s			24	5.075	0.226	4.45	4.400	5.800
2021-425	160 SIEMENS EX			71	4.728	0.312	6.60	3.800	5.700
040	THYROXINE (T4)	CLIA TEST ID #: 595	EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}						
2021-421	32 DADE DIME.			47	7.065	0.553	7.83	5.700	8.500
2021-421	106 TOSOH AIA			10	6.290	0.905	14.39	5.000	7.500
2021-421	115 VITROS ECI			37	6.258	0.329	5.25	5.000	7.500
2021-421	116 ROCHE COBA			46	6.353	0.419	6.60	5.100	7.600
2021-421	125 B.C.ACCESS			46	5.954	0.473	7.94	4.800	7.100
2021-421	149 ARC i2000s			21	5.671	0.357	6.30	4.500	6.800
2021-421	160 SIEMENS EX			11	7.282	0.587	8.07	5.800	8.700
2021-422	32 DADE DIME.			47	11.272	0.744	6.60	9.000	13.500
2021-422	106 TOSOH AIA			10	9.790	1.194	12.19	7.800	11.700
2021-422	115 VITROS ECI			37	10.024	0.426	4.25	8.000	12.000
2021-422	116 ROCHE COBA			46	10.107	0.706	6.99	8.100	12.100
2021-422	125 B.C.ACCESS			46	9.280	0.731	7.88	7.400	11.100
2021-422	149 ARC i2000s			21	9.857	0.640	6.50	7.900	11.800
2021-422	160 SIEMENS EX			11	11.482	0.635	5.53	9.200	13.800
2021-423	32 DADE DIME.			47	15.939	1.063	6.67	12.800	19.100
2021-423	106 TOSOH AIA			10	13.020	1.383	10.62	10.400	15.600
2021-423	115 VITROS ECI			37	13.595	0.557	4.10	10.900	16.300
2021-423	116 ROCHE COBA			46	13.971	0.894	6.40	11.200	16.800
2021-423	125 B.C.ACCESS			46	11.941	0.936	7.84	9.600	14.300

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-423	149	ARC i2000s		21	13.971	0.932	6.67	11.200	16.800	
2021-423	160	SIEMENS EX		11	15.955	0.847	5.31	12.800	19.100	
040	THYROXINE (T4)		CLIA TEST ID #: 595		EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}					
2021-424	32	DADE DIME.		47	3.815	0.464	12.17	2.800	4.800	
2021-424	106	TOSOH AIA		10	3.530	0.540	15.31	2.500	4.500	
2021-424	115	VITROS ECI		37	3.330	0.159	4.78	2.300	4.300	
2021-424	116	ROCHE COBA		46	3.451	0.256	7.42	2.500	4.500	
2021-424	125	B.C.ACCESS		46	2.948	0.353	11.98	1.900	3.900	
2021-424	149	ARC i2000s		21	2.795	0.250	8.93	1.800	3.800	
2021-424	160	SIEMENS EX		11	4.027	0.471	11.70	3.000	5.000	
2021-425	32	DADE DIME.		47	12.415	0.873	7.03	9.900	14.900	
2021-425	106	TOSOH AIA		10	10.700	1.435	13.41	8.600	12.800	
2021-425	115	VITROS ECI		37	10.938	0.543	4.97	8.800	13.100	
2021-425	116	ROCHE COBA		46	11.109	0.773	6.96	8.900	13.300	
2021-425	125	B.C.ACCESS		46	10.041	0.809	8.05	8.000	12.000	
2021-425	149	ARC i2000s		21	10.700	0.496	4.64	8.600	12.800	
2021-425	160	SIEMENS EX		11	12.455	0.845	6.78	10.000	14.900	
141	FOLATE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %					
2021-421	115	VITROS ECI		15	2.787	0.424	15.22	2.100	3.500	
2021-421	115	VITROS ECI		15	3.560	0.273	7.66	2.700	4.500	
2021-421	116	ROCHE COBA		18	2.233	0.236	10.55	1.700	2.800	
2021-421	116	ROCHE COBA		18	3.600	0.000	0.00	2.700	4.500	
2021-421	125	B.C.ACCESS		11	3.655	0.311	8.52	2.700	4.600	
2021-421	149	ARC i2000s		11	4.045	0.523	12.92	3.000	5.100	
2021-422	115	VITROS ECI		15	7.600	0.319	4.19	5.700	9.500	
2021-422	115	VITROS ECI		15	6.413	0.813	12.68	4.800	8.000	
2021-422	116	ROCHE COBA		18	7.600	1.000	13.16	5.700	9.500	
2021-422	116	ROCHE COBA		18	5.572	1.081	19.40	4.200	7.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-422	125	B.C.ACCESS		11	7.700	0.497	6.46	5.800	9.600
2021-422	149	ARC i2000s		11	9.882	0.811	8.21	7.400	12.400
141	FOLATE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %				
2021-423	115	VITROS ECI		15	10.779	1.001	9.29	8.100	13.500
2021-423	115	VITROS ECI		15	11.600	0.447	3.86	8.700	14.500
2021-423	116	ROCHE COBA		18	9.978	1.416	14.20	7.500	12.500
2021-423	116	ROCHE COBA		18	11.600	2.000	17.24	8.700	14.500
2021-423	125	B.C.ACCESS		11	11.636	0.882	7.58	8.700	14.500
2021-423	149	ARC i2000s		11	15.700	1.408	8.97	11.800	19.600
2021-424	115	VITROS ECI		15	0.800	0.388	48.48	0.600	0.900
2021-424	115	VITROS ECI		15	1.093	0.139	12.70	0.800	1.400
2021-424	116	ROCHE COBA		18	0.800	0.000	0.00	0.600	0.900
2021-424	116	ROCHE COBA		18	2.000	0.000	0.00	1.500	2.500
2021-424	125	B.C.ACCESS		11	0.764	0.382	50.03	0.600	1.000
2021-424	149	ARC i2000s		11	1.255	0.576	45.90	0.900	1.600
2021-425	115	VITROS ECI		15	7.853	0.950	12.10	5.900	9.800
2021-425	115	VITROS ECI		15	8.700	0.204	2.34	6.500	10.900
2021-425	116	ROCHE COBA		18	6.900	0.593	8.60	5.200	8.600
2021-425	116	ROCHE COBA		18	8.700	1.000	11.49	6.500	10.900
2021-425	125	B.C.ACCESS		11	8.627	0.548	6.35	6.500	10.800
2021-425	149	ARC i2000s		11	11.082	1.002	9.04	8.300	13.900
142	hLH (HUMAN LUT. HOR)								
2021-421	116	ROCHE COBA		15	17.713	1.314	7.42	13.300	22.100
2021-422	116	ROCHE COBA		15	32.547	2.651	8.15	24.400	40.700
2021-423	116	ROCHE COBA		15	46.993	4.354	9.27	35.200	58.700

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
142 hLH (HUMAN LUT. HOR)			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %						
	2021-424	116 ROCHE COBA			15	5.813	0.481	8.28	4.400	7.300
	2021-425	116 ROCHE COBA			15	36.247	2.395	6.61	27.200	45.300

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS	# LABS ANSW	# "SF" LABS	Score
031 hCG-QUALITATIVE											
2021-421	113 BIO-SIGN			002 002	002 002		002 - POSITIVE	20	20	20	100
2021-421	142 INSTANT VI			002 002	002 002		002 - POSITIVE	45	45	45	100
2021-421	150 AIM STEP			002 002	002 002		002 - POSITIVE	82	82	82	100
2021-421	164 HCG Alere			002 002	002 002		002 - POSITIVE	69	69	69	100
2021-421	165 Osom HCG C			002 002	002 002		002 - POSITIVE	21	21	21	100
2021-421	166 FaStep			002 002	002 002		002 - POSITIVE	14	14	14	100
2021-421	167 Tanner Sci			002 002	002 002		002 - POSITIVE	12	12	12	100
2021-422	113 BIO-SIGN			002 002	002 002		002 - POSITIVE	20	20	20	100
2021-422	142 INSTANT VI			002 002	002 002		002 - POSITIVE	45	45	45	100
2021-422	150 AIM STEP			002 002	002 002		002 - POSITIVE	82	82	82	100
2021-422	164 HCG Alere			002 002	002 002		002 - POSITIVE	69	69	69	100
2021-422	165 Osom HCG C			002 002	002 002		002 - POSITIVE	21	21	21	100
2021-422	166 FaStep			002 002	002 002		002 - POSITIVE	14	14	14	100
2021-422	167 Tanner Sci			002 002	002 002		002 - POSITIVE	12	12	12	100
2021-423	113 BIO-SIGN			002 002	002 002		002 - POSITIVE	20	20	20	100
2021-423	142 INSTANT VI			002 002	002 002		002 - POSITIVE	45	45	45	100
2021-423	150 AIM STEP			002 002	002 002		002 - POSITIVE	82	82	82	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-423	164	HCG Alere		002 002	002 002		002 - POSITIVE	69	69	69	100	
2021-423	165	Osom HCG C		002 002	002 002		002 - POSITIVE	21	21	21	100	
2021-423	166	FaStep		002 002	002 002		001 - NEGATIVE	14	1	13	0	
2021-423	166	FaStep		002 002	002 002		002 - POSITIVE	14	13	13	100	
2021-423	167	Tanner Sci		002 002	002 002		002 - POSITIVE	12	12	12	100	
		031 hCG-QUALITATIVE										
2021-424	113	BIO-SIGN		001 001	001 001		001 - NEGATIVE	20	19	19	100	
2021-424	113	BIO-SIGN		001 001	001 001		002 - POSITIVE	20	1	19	0	
2021-424	142	INSTANT VI		001 001	001 001		001 - NEGATIVE	45	44	44	100	
2021-424	142	INSTANT VI		001 001	001 001		002 - POSITIVE	45	1	44	0	
2021-424	150	AIM STEP		001 001	001 001		001 - NEGATIVE	82	80	80	100	
2021-424	150	AIM STEP		001 001	001 001		002 - POSITIVE	82	2	80	0	
2021-424	164	HCG Alere		001 001	001 001		001 - NEGATIVE	69	69	69	100	
2021-424	165	Osom HCG C		001 001	001 001		001 - NEGATIVE	21	20	20	100	
2021-424	165	Osom HCG C		001 001	001 001		002 - POSITIVE	21	1	20	0	
2021-424	166	FaStep		001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-424	167	Tanner Sci		001 001	001 001		001 - NEGATIVE	12	12	12	100	
2021-425	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	20	20	20	100	
2021-425	142	INSTANT VI		002 002	002 002		002 - POSITIVE	45	45	45	100	
2021-425	150	AIM STEP		002 002	002 002		002 - POSITIVE	82	82	82	100	
2021-425	164	HCG Alere		002 002	002 002		002 - POSITIVE	69	69	69	100	
2021-425	165	Osom HCG C		002 002	002 002		002 - POSITIVE	21	21	21	100	
2021-425	166	FaStep		002 002	002 002		001 - NEGATIVE	14	1	13	0	
2021-425	166	FaStep		002 002	002 002		002 - POSITIVE	14	13	13	100	
2021-425	167	Tanner Sci		002 002	002 002		002 - POSITIVE	12	12	12	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2021-411	018	BIO SIGN		002	002	002	002	002 - POSITIVE	24	24	24	100
2021-411	114	INSTANT VI		002	002	002	002	002 - POSITIVE	120	120	120	100
2021-411	118	AIM STEP		002	002	002	002	002 - POSITIVE	147	147	147	100
2021-411	130	ALERE HCG		002	002	002	002	002 - POSITIVE	91	91	91	100
2021-411	131	OSOM HCG		002	002	002	002	002 - POSITIVE	26	26	26	100
2021-411	133	FaStep		002	002	002	002	002 - POSITIVE	25	25	25	100
2021-411	134	Tanner Sc.		002	002	002	002	002 - POSITIVE	64	64	64	100
2021-412	018	BIO SIGN		001	001	001	001	001 - NEGATIVE	24	23	23	100
2021-412	018	BIO SIGN		001	001	001	001	002 - POSITIVE	24	1	23	0
2021-412	114	INSTANT VI		001	001	001	001	001 - NEGATIVE	120	119	119	100
2021-412	114	INSTANT VI		001	001	001	001	002 - POSITIVE	120	1	119	0
2021-412	118	AIM STEP		001	001	001	001	001 - NEGATIVE	147	146	146	100
2021-412	118	AIM STEP		001	001	001	001	002 - POSITIVE	147	1	146	0
2021-412	130	ALERE HCG		001	001	001	001	001 - NEGATIVE	91	91	91	100
2021-412	131	OSOM HCG		001	001	001	001	001 - NEGATIVE	26	26	26	100
2021-412	133	FaStep		001	001	001	001	001 - NEGATIVE	25	25	25	100
2021-412	134	Tanner Sc.		001	001	001	001	001 - NEGATIVE	64	64	64	100
2021-413	018	BIO SIGN		002	002	002	002	002 - POSITIVE	24	24	24	100
2021-413	114	INSTANT VI		002	002	002	002	002 - POSITIVE	120	120	120	100
2021-413	118	AIM STEP		002	002	002	002	002 - POSITIVE	147	147	147	100
2021-413	130	ALERE HCG		002	002	002	002	002 - POSITIVE	91	91	91	100
2021-413	131	OSOM HCG		002	002	002	002	002 - POSITIVE	26	26	26	100
2021-413	133	FaStep		002	002	002	002	002 - POSITIVE	25	25	25	100
2021-413	134	Tanner Sc.		002	002	002	002	002 - POSITIVE	64	64	64	100
2021-414	018	BIO SIGN		001	001	001	001	001 - NEGATIVE	24	24	24	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-414	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	120	120	120	100	
2021-414	118	AIM STEP		001 001	001 001		001 - NEGATIVE	147	147	147	100	
2021-414	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	91	91	91	100	
2021-414	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	26	26	26	100	
2021-414	133	FaStep		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2021-414	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	64	64	64	100	
151 hCG ONLY QUALITATIVE												
2021-415	018	BIO SIGN		002 002	002 002		002 - POSITIVE	24	24	24	100	
2021-415	114	INSTANT VI		002 002	002 002		002 - POSITIVE	120	120	120	100	
2021-415	118	AIM STEP		002 002	002 002		002 - POSITIVE	147	147	147	100	
2021-415	130	ALERE HCG		002 002	002 002		002 - POSITIVE	91	91	91	100	
2021-415	131	OSOM HCG		002 002	002 002		002 - POSITIVE	26	26	26	100	
2021-415	133	FaStep		002 002	002 002		002 - POSITIVE	25	25	25	100	
2021-415	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	64	64	64	100	

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
173 T3 UPTAKE												
CLIA TEST ID #: 565			EVALUATION CRITERIA: TARGET VALUE 3 SD									
2021-421	1 %U	84 VITROS ECI	1 25-35	27	54.356	1.577	2.90	49.620	59.090			
2021-421	1 %U	91 B.C.ACCESS	2 35-45	29	47.572	2.496	5.25	40.080	55.060			
2021-421	2 TUV	95 Arc i2000s	3 0.7 - 1.2	16	0.824	0.037	4.54	0.710	0.940			
2021-421	4 %	32 DIMENSION	2 35-45	58	40.173	1.859	4.63	34.600	45.750			
2021-421	4 %	91 B.C.ACCESS	2 35-45	13	47.200	2.507	5.31	39.680	54.720			
2021-421	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.828	0.076	9.18	0.600	1.060			
2021-421	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.813	0.041	5.01	0.690	0.940			
2021-422	1 %U	84 VITROS ECI	1 25-35	27	49.922	1.430	2.86	45.630	54.210			

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 55 ENDO.(T3 UPTAKE)

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2021-422	1 %U	91 B.C.ACCESS	2 35-45	29	45.441	2.553	5.62	37.780	53.100			
2021-422	2 TUV	95 Arc i2000s	3 0.7 - 1.2	16	1.021	0.051	4.98	0.870	1.170			
2021-422	4 %	32 DIMENSION	2 35-45	58	39.569	1.795	4.54	34.180	44.950			
2021-422	4 %	91 B.C.ACCESS	2 35-45	13	44.869	1.607	3.58	40.050	49.690			
2021-422	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.715	0.052	7.23	0.560	0.870			
2021-422	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.758	0.048	6.36	0.610	0.900			
173 T3 UPTAKE		CLIA TEST ID #: 565	EVALUATION CRITERIA: TARGET VALUE 3 SD									
2021-423	1 %U	84 VITROS ECI	1 25-35	27	46.922	1.561	3.33	42.240	51.610			
2021-423	1 %U	91 B.C.ACCESS	2 35-45	29	42.914	2.159	5.03	36.440	49.390			
2021-423	2 TUV	95 Arc i2000s	3 0.7 - 1.2	16	1.189	0.069	5.83	0.980	1.400			
2021-423	4 %	32 DIMENSION	2 35-45	58	39.436	1.614	4.09	34.590	44.280			
2021-423	4 %	91 B.C.ACCESS	2 35-45	13	42.900	1.888	4.40	37.240	48.560			
2021-423	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.690	0.074	10.80	0.470	0.910			
2021-423	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.606	0.058	9.55	0.430	0.780			
2021-424	1 %U	84 VITROS ECI	1 25-35	27	58.185	1.675	2.88	53.160	63.210			
2021-424	1 %U	91 B.C.ACCESS	2 35-45	29	50.907	2.206	4.33	44.290	57.530			
2021-424	2 TUV	95 Arc i2000s	3 0.7 - 1.2	16	0.636	0.034	5.27	0.540	0.740			
2021-424	4 %	32 DIMENSION	2 35-45	58	40.258	1.805	4.48	34.840	45.670			
2021-424	4 %	91 B.C.ACCESS	2 35-45	13	50.723	2.326	4.59	43.740	57.700			
2021-424	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.933	0.044	4.67	0.800	1.060			
2021-424	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.903	0.182	20.13	0.360	1.450			
2021-425	1 %U	84 VITROS ECI	1 25-35	27	49.222	1.381	2.81	45.080	53.370			
2021-425	1 %U	91 B.C.ACCESS	2 35-45	29	45.107	1.966	4.36	39.210	51.010			
2021-425	2 TUV	95 Arc i2000s	3 0.7 - 1.2	16	1.069	0.061	5.68	0.890	1.250			
2021-425	4 %	32 DIMENSION	2 35-45	58	39.614	1.666	4.21	34.620	44.610			
2021-425	4 %	91 B.C.ACCESS	2 35-45	13	44.808	2.196	4.90	38.220	51.400			
2021-425	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.778	0.080	10.26	0.540	1.020			
2021-425	5 TBI	98 ROCHE COBA	4 0.8-1.3	43	0.689	0.057	8.21	0.520	0.860			

Participant Summary

MONTH 5 OF YEAR 2021

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
SPEC: 10 CHEMISTRY			SUB: 40 TOXICOLOGY			GRP: 60 TOXICOLOGY		
043 DIGOXIN		CLIA TEST ID #: 645	EVALUATION CRITERIA: TARGET VALUE 20 % OR 0.20 {GREATER}					
2021-421	50 VITROS		14	0.771	0.059	7.64	0.600	1.000
2021-422	50 VITROS		14	1.364	0.061	4.47	1.100	1.600
2021-423	50 VITROS		14	2.085	0.095	4.55	1.700	2.500
2021-424	50 VITROS		14	0.450	0.050	11.11	0.300	0.700
2021-425	50 VITROS		14	1.564	0.072	4.59	1.300	1.900
045 FERRITIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %					
2021-421	50 VITROS		17	57.525	2.206	3.83	43.100	71.900
2021-421	50 VITROS		17	86.700	9.000	10.38	65.000	108.400
2021-421	90 ROCHE COBA		13	96.692	5.069	5.24	72.500	120.900
2021-422	50 VITROS		17	169.000	15.000	8.88	126.800	211.200
2021-422	50 VITROS		17	115.688	5.382	4.65	86.800	144.600
2021-422	90 ROCHE COBA		13	191.377	13.044	6.82	143.500	239.200
2021-423	50 VITROS		17	181.625	10.005	5.51	136.200	227.000
2021-423	50 VITROS		17	267.200	25.000	9.36	200.400	334.000
2021-423	90 ROCHE COBA		13	290.354	20.587	7.09	217.800	362.900

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
045	FERRITIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %							
2021-424	50 VITROS			17	20.500	2.000	9.76	15.400	25.600	
2021-424	50 VITROS			17	18.025	0.796	4.42	13.500	22.500	
2021-424	90 ROCHE COBA			13	30.854	2.115	6.85	23.100	38.600	
2021-425	50 VITROS			17	131.438	4.677	3.56	98.600	164.300	
2021-425	50 VITROS			17	197.000	20.000	10.15	147.800	246.200	
2021-425	90 ROCHE COBA			13	218.162	12.218	5.60	163.600	272.700	
046	GENTAMICIN	CLIA TEST ID #: 665								
2021-421	50 VITROS			13	3.523	0.119	3.37	2.600	4.400	
2021-422	50 VITROS			13	8.383	0.412	4.91	6.300	10.500	
2021-423	50 VITROS			13	13.869	0.765	5.52	10.400	17.300	
2021-424	50 VITROS			13	0.600	0.000	0.00	0.500	0.800	
2021-425	50 VITROS			13	9.592	0.218	2.27	7.200	12.000	
050	PHENYTOIN	CLIA TEST ID #: 695								
2021-421	50 VITROS			16	6.825	0.353	5.17	5.100	8.500	
2021-421	92 ROCHE COBA			10	6.710	0.339	5.05	5.000	8.400	
2021-422	50 VITROS			16	16.956	0.708	4.18	12.700	21.200	
2021-422	92 ROCHE COBA			10	16.240	0.839	5.17	12.200	20.300	
2021-423	50 VITROS			16	28.369	0.924	3.26	21.300	35.500	
2021-423	92 ROCHE COBA			10	26.250	1.694	6.45	19.700	32.800	
2021-424	50 VITROS			16	2.600	2.154	82.85	2.000	3.300	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-424	92	ROCHE COBA		10	0.800	0.000	0.00	0.600	1.000	
050	PHENYTOIN		CLIA TEST ID #: 695		EVALUATION CRITERIA: TARGET VALUE 25 %					
2021-425	50	VITROS		16	19.738	0.844	4.28	14.800	24.700	
2021-425	92	ROCHE COBA		10	18.390	0.719	3.91	13.800	23.000	
054	THEOPHYLLINE		CLIA TEST ID #: 735							
2021-421	50	VITROS		13	14.931	0.693	4.64	11.200	18.700	
2021-422	50	VITROS		13	31.754	0.933	2.94	23.800	39.700	
2021-423	50	VITROS		13	48.754	2.111	4.33	36.600	60.900	
2021-424	50	VITROS		13	1.677	0.317	18.88	1.300	2.100	
2021-425	50	VITROS		13	36.085	1.228	3.40	27.100	45.100	
127	VANCOMYCIN		CLIA TEST ID #: 0							
2021-421	50	VITROS		18	12.306	0.454	3.69	9.200	15.400	
2021-422	50	VITROS		18	30.344	1.397	4.60	22.800	37.900	
2021-423	50	VITROS		18	50.667	3.256	6.43	38.000	63.300	
2021-424	50	VITROS		18	5.000	0.000	0.00	3.800	6.300	
2021-425	50	VITROS		18	36.518	1.893	5.18	27.400	45.600	
SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
065	URINA-PROTEIN QUANT.	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%	OR 5.00	{GREATER}		
2021-451	58	VITROS		41	155.195	10.848	6.99	116.400	194.000
2021-451	62	DADE DIM.		42	121.529	2.865	2.36	91.100	151.900
2021-451	79	BECK. COUL		13	104.192	4.971	4.77	78.100	130.200
2021-451	93	ARCHITECT		20	100.610	2.962	2.94	75.500	125.800
2021-451	95	INTEGRA		10	93.780	6.412	6.84	70.300	117.200
2021-451	113	ROCHE COBA		28	97.415	2.327	2.39	73.100	121.800
2021-451	139	S. EXL.200		15	119.973	1.914	1.60	90.000	150.000
2021-452	58	VITROS		41	143.561	11.591	8.07	107.700	179.500
2021-452	62	DADE DIM.		42	118.054	2.063	1.75	88.500	147.600
2021-452	79	BECK. COUL		13	105.323	4.393	4.17	79.000	131.700
2021-452	93	ARCHITECT		20	96.655	3.148	3.26	72.500	120.800
2021-452	95	INTEGRA		10	87.950	6.786	7.72	66.000	109.900
2021-452	113	ROCHE COBA		28	90.979	2.475	2.72	68.200	113.700
2021-452	139	S. EXL.200		15	116.780	2.236	1.91	87.600	146.000
2021-453	58	VITROS		41	9.075	0.565	6.23	4.100	14.100
2021-453	62	DADE DIM.		42	24.571	2.899	11.80	18.400	30.700
2021-453	79	BECK. COUL		13	5.525	1.526	27.62	0.500	10.500
2021-453	93	ARCHITECT		20	6.980	1.248	17.88	2.000	12.000
2021-453	95	INTEGRA		10	2.420	0.921	38.04	0	7.400
2021-453	113	ROCHE COBA		28	3.748	0.672	17.92	0	8.700
2021-453	139	S. EXL.200		15	24.060	1.287	5.35	18.000	30.100
2021-454	58	VITROS		41	127.000	9.860	7.76	95.300	158.800
2021-454	62	DADE DIM.		42	112.100	2.567	2.29	84.100	140.100
2021-454	79	BECK. COUL		13	102.433	3.805	3.71	76.800	128.000
2021-454	93	ARCHITECT		20	98.000	3.554	3.63	73.500	122.500
2021-454	95	INTEGRA		10	87.550	6.374	7.28	65.700	109.400

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-454	113	ROCHE COBA			28	91.261	1.933	2.12	68.400	114.100
2021-454	139	S. EXL.200			15	111.180	1.524	1.37	83.400	139.000
065 URINA-PROTEIN QUANT.			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2021-455	58	VITROS			41	7.390	0.488	6.60	2.400	12.400
2021-455	62	DADE DIM.			42	14.478	2.078	14.35	9.500	19.500
2021-455	79	BECK. COUL			13	4.867	0.896	18.40	0	9.900
2021-455	93	ARCHITECT			20	7.205	1.685	23.39	2.200	12.200
2021-455	95	INTEGRA			10	3.310	0.797	24.07	0	8.300
2021-455	113	ROCHE COBA			28	3.911	0.508	12.99	0	8.900
2021-455	139	S. EXL.200			15	14.053	0.999	7.11	9.100	19.100
067 URI-SPECIFIC GRAVITY				EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-451	11	AMES NM10S			50	1.017	0.003	0.34	1.007	1.028
2021-451	13	BAYER CLI			36	1.018	0.003	0.27	1.010	1.026
2021-451	67	URISCAN			51	1.013	0.003	0.28	1.005	1.022
2021-451	71	ATLAS			14	1.030	0.001	0.06	1.028	1.032
2021-451	72	URS-9-TECO			101	1.018	0.003	0.32	1.008	1.028
2021-451	80	ROCHE (CHE			37	1.013	0.002	0.24	1.006	1.020
2021-451	89	AIMSTRIP			108	1.017	0.003	0.33	1.006	1.027
2021-451	94	CLINITEK			107	1.018	0.003	0.29	1.009	1.027
2021-451	102	ACON LAB			157	1.018	0.003	0.34	1.007	1.028
2021-451	131	iCHEM VELO			27	1.029	0.001	0.09	1.026	1.032
2021-451	135	Hybrido AU			13	1.029	0.001	0.11	1.026	1.033
2021-451	138	URS-10T			10	1.018	0.003	0.25	1.010	1.025
2021-451	140	MISSION A			58	1.017	0.003	0.34	1.007	1.028
2021-451	145	CLINITEK			12	1.032	0.001	0.06	1.030	1.034
2021-451	147	Mission120			30	1.020	0.000	0.00	1.020	1.020
2021-451	148	Mission500			30	1.020	0.002	0.18	1.014	1.025
2021-452	11	AMES NM10S			50	1.020	0.003	0.29	1.011	1.029
2021-452	13	BAYER CLI			36	1.020	0.001	0.14	1.015	1.024

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-452	67	URISCAN	51	1.016	0.002	0.23	1.009	1.023
2021-452	71	ATLAS	14	1.028	0.002	0.15	1.023	1.033
2021-452	72	URS-9-TECO	101	1.019	0.003	0.29	1.010	1.027
2021-452	80	ROCHE (CHE	37	1.011	0.002	0.19	1.005	1.016
2021-452	89	AIMSTRIP	108	1.018	0.003	0.29	1.009	1.027
2021-452	94	CLINITEK	107	1.020	0.002	0.20	1.014	1.026
2021-452	102	ACON LAB	157	1.018	0.002	0.25	1.011	1.026
2021-452	131	iCHEM VELO	27	1.026	0.001	0.09	1.023	1.028
2021-452	135	Hybrido AU	13	1.025	0.002	0.16	1.020	1.030
2021-452	138	URS-10T	10	1.019	0.002	0.20	1.013	1.025
2021-452	140	MISSION A	58	1.018	0.003	0.29	1.009	1.027
2021-452	145	CLINITEK	12	1.027	0.001	0.07	1.025	1.029
2021-452	147	Mission120	30	1.020	0.000	0.00	1.020	1.020
2021-452	148	Mission500	30	1.019	0.002	0.19	1.014	1.025
067	URI-SPECIFIC GRAVITY	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2021-453	11	AMES NM10S	50	1.012	0.003	0.31	1.003	1.022
2021-453	13	BAYER CLI	36	1.013	0.003	0.29	1.004	1.022
2021-453	67	URISCAN	51	1.011	0.006	0.60	0.993	1.029
2021-453	71	ATLAS	14	1.030	0.001	0.09	1.027	1.033
2021-453	72	URS-9-TECO	101	1.012	0.004	0.36	1.001	1.023
2021-453	80	ROCHE (CHE	37	1.013	0.003	0.30	1.004	1.022
2021-453	89	AIMSTRIP	108	1.015	0.005	0.49	1.000	1.030
2021-453	94	CLINITEK	107	1.012	0.003	0.28	1.003	1.021
2021-453	102	ACON LAB	157	1.016	0.004	0.39	1.004	1.028
2021-453	131	iCHEM VELO	27	1.026	0.001	0.09	1.023	1.028
2021-453	135	Hybrido AU	13	1.025	0.001	0.11	1.022	1.028
2021-453	138	URS-10T	10	1.018	0.012	1.15	0.983	1.053
2021-453	140	MISSION A	58	1.014	0.004	0.41	1.001	1.026
2021-453	145	CLINITEK	12	1.031	0.001	0.07	1.029	1.033
2021-453	147	Mission120	30	1.017	0.003	0.27	1.009	1.025
2021-453	148	Mission500	30	1.019	0.002	0.17	1.014	1.025

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
067	URI-SPECIFIC GRAVITY	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2021-454	11	AMES NM10S		50	1.024	0.002	0.24	1.016	1.031
2021-454	13	BAYER CLI		36	1.023	0.002	0.24	1.016	1.030
2021-454	67	URISCAN		51	1.021	0.005	0.47	1.007	1.035
2021-454	71	ATLAS		14	1.030	0.000	0.04	1.028	1.031
2021-454	72	URS-9-TECO		101	1.022	0.003	0.29	1.013	1.030
2021-454	80	ROCHE (CHE		37	1.014	0.002	0.17	1.009	1.020
2021-454	89	AIMSTRIP		108	1.022	0.003	0.25	1.015	1.030
2021-454	94	CLINITEK		107	1.024	0.002	0.23	1.017	1.031
2021-454	102	ACON LAB		157	1.021	0.002	0.23	1.014	1.028
2021-454	131	iCHEM VELO		27	1.025	0.001	0.09	1.022	1.027
2021-454	135	Hybrido AU		13	1.024	0.001	0.12	1.020	1.028
2021-454	138	URS-10T		10	1.022	0.002	0.24	1.015	1.029
2021-454	140	MISSION A		58	1.021	0.003	0.27	1.013	1.030
2021-454	145	CLINITEK		12	1.030	0.001	0.05	1.029	1.032
2021-454	147	Mission120		30	1.023	0.002	0.24	1.015	1.030
2021-454	148	Mission500		30	1.021	0.002	0.17	1.016	1.026
2021-455	11	AMES NM10S		50	1.023	0.003	0.25	1.016	1.031
2021-455	13	BAYER CLI		36	1.022	0.003	0.32	1.013	1.032
2021-455	67	URISCAN		51	1.021	0.004	0.36	1.010	1.032
2021-455	71	ATLAS		14	1.031	0.002	0.17	1.026	1.037
2021-455	72	URS-9-TECO		101	1.020	0.004	0.41	1.008	1.033
2021-455	80	ROCHE (CHE		37	1.017	0.002	0.24	1.010	1.025
2021-455	89	AIMSTRIP		108	1.023	0.003	0.33	1.013	1.034
2021-455	94	CLINITEK		107	1.023	0.003	0.29	1.015	1.032
2021-455	102	ACON LAB		157	1.023	0.003	0.26	1.015	1.031
2021-455	131	iCHEM VELO		27	1.027	0.001	0.09	1.024	1.030
2021-455	135	Hybrido AU		13	1.027	0.001	0.13	1.023	1.031
2021-455	138	URS-10T		10	1.018	0.005	0.50	1.003	1.033

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-455	140	MISSION A		58	1.023	0.003	0.28	1.014	1.032
2021-455	145	CLINITEK		12	1.033	0.001	0.10	1.030	1.036
2021-455	147	Mission120		30	1.021	0.002	0.22	1.015	1.028
2021-455	148	Mission500		30	1.023	0.002	0.23	1.017	1.030
166	MICROALBUMIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}				
2021-451	58	VITROS		41	882.828	41.132	4.66	662.100	1103.500
2021-451	62	DADE DIM.		59	945.667	43.053	4.55	709.300	1182.100
2021-451	62	DADE DIM.		59	878.116	51.818	5.90	658.600	1097.600
2021-451	79	BECK. COUL		19	923.600	57.413	6.22	692.700	1154.500
2021-451	79	BECK. COUL		19	862.212	101.422	11.76	646.700	1077.800
2021-451	85	Rx DAYTONA		17	813.250	80.966	9.96	609.900	1016.600
2021-451	93	ARCHITECT		23	860.961	34.597	4.02	645.700	1076.200
2021-451	95	INTEGRA		24	879.909	32.256	3.67	659.900	1099.900
2021-451	113	ROCHE COBA		27	821.881	41.912	5.10	616.400	1027.400
2021-451	139	S. EXL.200		24	894.138	89.677	10.03	670.600	1117.700
2021-451	139	S. EXL.200		24	923.600	57.413	6.22	692.700	1154.500
2021-452	58	VITROS		41	864.318	42.542	4.92	648.200	1080.400
2021-452	62	DADE DIM.		59	871.072	50.774	5.83	653.300	1088.800
2021-452	62	DADE DIM.		59	949.333	39.886	4.20	712.000	1186.700
2021-452	79	BECK. COUL		19	848.894	98.374	11.59	636.700	1061.100
2021-452	79	BECK. COUL		19	895.800	60.031	6.70	671.900	1119.800
2021-452	85	Rx DAYTONA		17	789.300	79.452	10.07	592.000	986.600
2021-452	93	ARCHITECT		23	846.136	25.902	3.06	634.600	1057.700
2021-452	95	INTEGRA		24	857.891	39.191	4.57	643.400	1072.400
2021-452	113	ROCHE COBA		27	809.204	40.537	5.01	606.900	1011.500
2021-452	139	S. EXL.200		24	875.070	67.057	7.66	656.300	1093.800
2021-452	139	S. EXL.200		24	895.800	60.031	6.70	671.900	1119.800
2021-453	58	VITROS		41	6.026	0.158	2.62	1.000	11.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-453	62	DADE DIM.		59	5.000	0.000	0.00	0.000	10.000	
2021-453	62	DADE DIM.		59	3.164	1.343	42.46	0	8.200	
2021-453	79	BECK. COUL		19	2.200	0.980	44.54	0	7.200	
2021-453	79	BECK. COUL		19	2.144	2.814	131.24	0	7.100	
2021-453	85	Rx DAYTONA		17	1.888	2.799	148.24	0	6.900	
2021-453	93	ARCHITECT		23	4.945	0.208	4.21	0	9.900	
2021-453	95	INTEGRA		24	3.650	1.408	38.58	0	8.700	
2021-453	113	ROCHE COBA		27	10.526	3.543	33.66	5.500	15.500	
2021-453	139	S. EXL.200		24	2.200	0.980	44.54	0	7.200	
2021-453	139	S. EXL.200		24	2.563	1.351	52.74	0	7.600	
166 MICROALBUMIN		CLIA TEST ID #: 0					EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}			
2021-454	58	VITROS		41	863.553	54.285	6.29	647.700	1079.400	
2021-454	62	DADE DIM.		59	868.859	51.357	5.91	651.600	1086.100	
2021-454	62	DADE DIM.		59	948.000	40.357	4.26	711.000	1185.000	
2021-454	79	BECK. COUL		19	910.800	70.505	7.74	683.100	1138.500	
2021-454	79	BECK. COUL		19	847.259	86.942	10.26	635.400	1059.100	
2021-454	85	Rx DAYTONA		17	806.582	92.112	11.42	604.900	1008.200	
2021-454	93	ARCHITECT		23	852.423	23.151	2.72	639.300	1065.500	
2021-454	95	INTEGRA		24	866.604	37.795	4.36	650.000	1083.300	
2021-454	113	ROCHE COBA		27	809.177	39.047	4.83	606.900	1011.500	
2021-454	139	S. EXL.200		24	888.567	80.066	9.01	666.400	1110.700	
2021-454	139	S. EXL.200		24	910.800	70.505	7.74	683.100	1138.500	
2021-455	58	VITROS		41	6.026	0.158	2.62	1.000	11.000	
2021-455	62	DADE DIM.		59	5.000	0.000	0.00	0.000	10.000	
2021-455	62	DADE DIM.		59	4.388	1.431	32.60	0	9.400	
2021-455	79	BECK. COUL		19	2.128	2.825	132.76	0	7.100	
2021-455	79	BECK. COUL		19	2.600	1.200	46.15	0	7.600	
2021-455	85	Rx DAYTONA		17	1.453	2.288	157.48	0	6.500	
2021-455	93	ARCHITECT		23	4.945	0.475	9.60	0	9.900	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2021-455	95	INTEGRA		24	3.646	1.307	35.85	0	8.600	
2021-455	113	ROCHE COBA		27	10.526	3.543	33.66	5.500	15.500	
2021-455	139	S. EXL.200		24	4.250	1.672	39.35	0	9.300	
2021-455	139	S. EXL.200		24	2.600	1.200	46.15	0	7.600	
167 CREATININE			CLIA TEST ID #: 0						EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}	
2021-451	58	VITROS		48	244.900	6.510	2.66	183.700	306.100	
2021-451	58	VITROS		48	238.917	8.572	3.59	179.200	298.600	
2021-451	62	DADE DIM.		52	244.900	6.510	2.66	183.700	306.100	
2021-451	62	DADE DIM.		52	231.663	17.752	7.66	173.700	289.600	
2021-451	79	BECK. COUL		16	244.900	6.510	2.66	183.700	306.100	
2021-451	79	BECK. COUL		16	205.750	13.335	6.48	154.300	257.200	
2021-451	85	Rx DAYTONA		11	229.591	27.170	11.83	172.200	287.000	
2021-451	93	ARCHITECT		22	237.057	7.135	3.01	177.800	296.300	
2021-451	95	INTEGRA		21	248.505	8.881	3.57	186.400	310.600	
2021-451	113	ROCHE COBA		30	245.283	11.758	4.79	184.000	306.600	
2021-451	139	S. EXL.200		23	238.352	6.459	2.71	178.800	297.900	
2021-451	139	S. EXL.200		23	244.900	6.510	2.66	183.700	306.100	
2021-452	58	VITROS		48	232.102	7.225	3.11	174.100	290.100	
2021-452	58	VITROS		48	241.433	10.419	4.32	181.100	301.800	
2021-452	62	DADE DIM.		52	236.031	12.081	5.12	177.000	295.000	
2021-452	62	DADE DIM.		52	241.433	10.419	4.32	181.100	301.800	
2021-452	79	BECK. COUL		16	217.769	14.628	6.72	163.300	272.200	
2021-452	79	BECK. COUL		16	241.433	10.419	4.32	181.100	301.800	
2021-452	85	Rx DAYTONA		11	236.373	22.764	9.63	177.300	295.500	
2021-452	93	ARCHITECT		22	236.014	7.338	3.11	177.000	295.000	
2021-452	95	INTEGRA		21	249.267	9.842	3.95	187.000	311.600	
2021-452	113	ROCHE COBA		30	245.403	10.970	4.47	184.100	306.800	
2021-452	139	S. EXL.200		23	241.433	10.419	4.32	181.100	301.800	
2021-452	139	S. EXL.200		23	240.822	6.571	2.73	180.600	301.000	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
167	CREATININE	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%	OR 5.00	{GREATER}		
2021-453	58	VITROS		48	247.900	8.862	3.57	185.900	309.900
2021-453	58	VITROS		48	245.439	11.327	4.62	184.100	306.800
2021-453	62	DADE DIM.		52	247.900	8.862	3.57	185.900	309.900
2021-453	62	DADE DIM.		52	238.862	12.384	5.18	179.100	298.600
2021-453	79	BECK. COUL		16	247.900	8.862	3.57	185.900	309.900
2021-453	79	BECK. COUL		16	221.638	12.988	5.86	166.200	277.000
2021-453	85	Rx DAYTONA		11	228.291	20.816	9.12	171.200	285.400
2021-453	93	ARCHITECT		22	240.124	7.212	3.00	180.100	300.200
2021-453	95	INTEGRA		21	252.110	8.839	3.51	189.100	315.100
2021-453	113	ROCHE COBA		30	248.140	10.528	4.24	186.100	310.200
2021-453	139	S. EXL.200		23	245.559	5.426	2.21	184.200	306.900
2021-453	139	S. EXL.200		23	247.900	8.862	3.57	185.900	309.900
2021-454	58	VITROS		48	3.221	0.207	6.43	0	8.200
2021-454	58	VITROS		48	15.000	0.000	0.00	10.000	20.000
2021-454	62	DADE DIM.		52	7.731	5.600	72.44	2.700	12.700
2021-454	62	DADE DIM.		52	15.000	0.000	0.00	10.000	20.000
2021-454	79	BECK. COUL		16	0.767	1.449	189.00	0	5.800
2021-454	79	BECK. COUL		16	15.000	0.000	0.00	10.000	20.000
2021-454	85	Rx DAYTONA		11	1.964	2.654	135.16	0	7.000
2021-454	93	ARCHITECT		22	4.943	0.213	4.31	0	9.900
2021-454	95	INTEGRA		21	0.414	0.318	76.80	0	5.400
2021-454	113	ROCHE COBA		30	3.859	0.770	19.95	0	8.900
2021-454	139	S. EXL.200		23	15.000	0.000	0.00	10.000	20.000
2021-454	139	S. EXL.200		23	6.791	5.922	87.19	1.800	11.800
2021-455	58	VITROS		48	15.000	0.000	0.00	10.000	20.000
2021-455	58	VITROS		48	3.223	0.205	6.37	0	8.200
2021-455	62	DADE DIM.		52	15.000	0.000	0.00	10.000	20.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-455	62	DADE DIM.		52	7.875	5.813	73.82	2.900	12.900
2021-455	79	BECK. COUL		16	15.000	0.000	0.00	10.000	20.000
2021-455	79	BECK. COUL		16	0.707	1.474	208.58	0	5.700
2021-455	85	Rx DAYTONA		11	1.718	1.335	77.72	0	6.700
2021-455	93	ARCHITECT		22	4.895	0.425	8.68	0	9.900
2021-455	95	INTEGRA		21	0.110	0.158	143.45	0	5.100
2021-455	113	ROCHE COBA		30	3.859	0.770	19.95	0	8.900
2021-455	139	S. EXL.200		23	6.830	5.968	87.37	1.800	11.800
2021-455	139	S. EXL.200		23	15.000	0.000	0.00	10.000	20.000

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART #	LABS #	# "SF"	Score	
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
057 URINALYSIS-BILIRUBIN												
2021-451	011	AMES NM10S		058	060	057	059	057 - NEGATIVE	50	1	50	100
2021-451	011	AMES NM10S		058	060	057	059	058 - TRACES, 1+	50	7	50	100
2021-451	011	AMES NM10S		058	060	057	059	059 - 2+	50	36	50	100
2021-451	011	AMES NM10S		058	060	057	059	060 - 3+	50	6	50	100
2021-451	013	BAYER CLI		058	060	057	059	058 - TRACES, 1+	37	6	37	100
2021-451	013	BAYER CLI		058	060	057	059	059 - 2+	37	24	37	100
2021-451	013	BAYER CLI		058	060	057	059	060 - 3+	37	7	37	100
2021-451	067	URISCAN		057	059	057	059	057 - NEGATIVE	50	1	50	100
2021-451	067	URISCAN		057	059	057	059	058 - TRACES, 1+	50	29	50	100
2021-451	067	URISCAN		057	059	057	059	059 - 2+	50	20	50	100
2021-451	071	ATLAS		057	059	057	059	058 - TRACES, 1+	15	14	15	100
2021-451	071	ATLAS		057	059	057	059	059 - 2+	15	1	15	100
2021-451	072	URS-9-TECO		058	060	057	059	057 - NEGATIVE	100	6	100	100
2021-451	072	URS-9-TECO		058	060	057	059	058 - TRACES, 1+	100	14	100	100
2021-451	072	URS-9-TECO		058	060	057	059	059 - 2+	100	55	100	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-451	072	URS-9-TECO		058 060	057 059		060 - 3+	100	25	100	100
2021-451	080	ROCHE (CHE		057 059	057 059		057 - NEGATIVE	35	1	35	100
2021-451	080	ROCHE (CHE		057 059	057 059		058 - TRACES,1+	35	30	35	100
2021-451	080	ROCHE (CHE		057 059	057 059		059 - 2+	35	4	35	100
2021-451	089	AIMSTRIP		057 059	057 059		057 - NEGATIVE	109	4	96	100
2021-451	089	AIMSTRIP		057 059	057 059		058 - TRACES,1+	109	48	96	100
2021-451	089	AIMSTRIP		057 059	057 059		059 - 2+	109	44	96	100
2021-451	089	AIMSTRIP		057 059	057 059		060 - 3+	109	13	96	0
2021-451	094	CLINITEK		058 060	057 059		057 - NEGATIVE	105	1	105	100
2021-451	094	CLINITEK		058 060	057 059		058 - TRACES,1+	105	11	105	100
2021-451	094	CLINITEK		058 060	057 059		059 - 2+	105	92	105	100
2021-451	094	CLINITEK		058 060	057 059		060 - 3+	105	1	105	100
2021-451	102	ACON LAB		057 059	057 059		057 - NEGATIVE	154	10	141	100
2021-451	102	ACON LAB		057 059	057 059		058 - TRACES,1+	154	96	141	100
2021-451	102	ACON LAB		057 059	057 059		059 - 2+	154	35	141	100
2021-451	102	ACON LAB		057 059	057 059		060 - 3+	154	13	141	0
2021-451	131	iCHEM VELO		057 057	057 059		057 - NEGATIVE	26	26	26	100
2021-451	135	Hybrido AU		058 060	057 059		059 - 2+	13	13	13	100
2021-451	140	MISSION A		057 059	057 059		057 - NEGATIVE	58	5	58	100
2021-451	140	MISSION A		057 059	057 059		058 - TRACES,1+	58	41	58	100
2021-451	140	MISSION A		057 059	057 059		059 - 2+	58	12	58	100
2021-451	145	CLINITEK		057 057	057 059		057 - NEGATIVE	12	12	12	100
2021-451	147	Mission120		057 059	057 059		057 - NEGATIVE	31	2	31	100
2021-451	147	Mission120		057 059	057 059		058 - TRACES,1+	31	26	31	100
2021-451	147	Mission120		057 059	057 059		059 - 2+	31	3	31	100
2021-451	148	Mission500		057 059	057 059		058 - TRACES,1+	31	26	31	100
2021-451	148	Mission500		057 059	057 059		059 - 2+	31	5	31	100
057 URINALYSIS-BILIRUBIN											
2021-452	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	50	50	50	100
2021-452	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	37	37	37	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-452	067	URISCAN		057 057	057 057		057 - NEGATIVE	50	50	50	100
2021-452	071	ATLAS		057 057	057 057		057 - NEGATIVE	15	15	15	100
2021-452	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	100	100	100	100
2021-452	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	35	35	35	100
2021-452	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	109	109	109	100
2021-452	094	CLINITEK		057 057	057 057		057 - NEGATIVE	105	105	105	100
2021-452	102	ACON LAB		057 057	057 057		057 - NEGATIVE	154	154	154	100
2021-452	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	26	26	26	100
2021-452	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	13	13	13	100
2021-452	140	MISSION A		057 057	057 057		057 - NEGATIVE	58	58	58	100
2021-452	145	CLINITEK		057 057	057 057		057 - NEGATIVE	12	12	12	100
2021-452	147	Mission120		057 057	057 057		057 - NEGATIVE	31	31	31	100
2021-452	148	Mission500		057 057	057 057		057 - NEGATIVE	31	31	31	100
057 URINALYSIS-BILIRUBIN											
2021-453	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	50	50	50	100
2021-453	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	37	37	37	100
2021-453	067	URISCAN		057 057	057 057		057 - NEGATIVE	50	50	50	100
2021-453	071	ATLAS		057 057	057 057		057 - NEGATIVE	15	15	15	100
2021-453	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	100	100	100	100
2021-453	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	35	35	35	100
2021-453	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	109	109	109	100
2021-453	094	CLINITEK		057 057	057 057		057 - NEGATIVE	105	105	105	100
2021-453	102	ACON LAB		057 057	057 057		057 - NEGATIVE	154	154	154	100
2021-453	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	26	26	26	100
2021-453	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	13	13	13	100
2021-453	140	MISSION A		057 057	057 057		057 - NEGATIVE	58	58	58	100
2021-453	145	CLINITEK		057 057	057 057		057 - NEGATIVE	12	12	12	100
2021-453	147	Mission120		057 057	057 057		057 - NEGATIVE	31	31	31	100
2021-453	148	Mission500		057 057	057 057		057 - NEGATIVE	31	31	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
057 URINALYSIS-BILIRUBIN												
2021-454	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	50	50	50	100
2021-454	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	37	37	37	100
2021-454	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	50	50	100
2021-454	071	ATLAS		057	057	057	057	057 - NEGATIVE	15	15	15	100
2021-454	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	100	100	100	100
2021-454	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	35	35	35	100
2021-454	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	109	109	109	100
2021-454	094	CLINITEK		057	057	057	057	057 - NEGATIVE	105	105	105	100
2021-454	102	ACON LAB		057	057	057	057	057 - NEGATIVE	154	154	154	100
2021-454	131	ICHEM VELO		057	057	057	057	057 - NEGATIVE	26	26	26	100
2021-454	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	13	13	13	100
2021-454	140	MISSION A		057	057	057	057	057 - NEGATIVE	58	58	58	100
2021-454	145	CLINITEK		057	057	057	057	057 - NEGATIVE	12	12	12	100
2021-454	147	Mission120		057	057	057	057	057 - NEGATIVE	31	31	31	100
2021-454	148	Mission500		057	057	057	057	057 - NEGATIVE	31	31	31	100
2021-455	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	50	50	50	100
2021-455	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	37	37	37	100
2021-455	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	49	49	100
2021-455	067	URISCAN		057	057	057	057	058 - TRACES, 1+	50	1	49	0
2021-455	071	ATLAS		057	057	057	057	057 - NEGATIVE	15	15	15	100
2021-455	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	100	100	100	100
2021-455	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	35	35	35	100
2021-455	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	109	109	109	100
2021-455	094	CLINITEK		057	057	057	057	057 - NEGATIVE	105	105	105	100
2021-455	102	ACON LAB		057	057	057	057	057 - NEGATIVE	154	154	154	100
2021-455	131	ICHEM VELO		057	057	057	057	057 - NEGATIVE	26	26	26	100
2021-455	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	13	13	13	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-455	140	MISSION A		057 057	057 057		057 - NEGATIVE	58	58	58	100	
2021-455	145	CLINITEK		057 057	057 057		057 - NEGATIVE	12	12	12	100	
2021-455	147	Mission120		057 057	057 057		057 - NEGATIVE	31	31	31	100	
2021-455	148	Mission500		057 057	057 057		057 - NEGATIVE	31	31	31	100	
058 URINALYSIS-GLUCOSE												
2021-451	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	50	50	50	100	
2021-451	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	35	35	35	100	
2021-451	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100	
2021-451	071	ATLAS		073 073	073 073		073 - NEGATIVE	15	15	15	100	
2021-451	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	101	101	101	100	
2021-451	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	37	37	37	100	
2021-451	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	108	107	107	100	
2021-451	089	AIMSTRIP		073 073	073 073		075 - 250mg/dl-+	108	1	107	0	
2021-451	094	CLINITEK		073 073	073 073		073 - NEGATIVE	105	105	105	100	
2021-451	102	ACON LAB		073 073	073 073		073 - NEGATIVE	156	156	156	100	
2021-451	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	28	28	28	100	
2021-451	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	13	13	13	100	
2021-451	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2021-451	140	MISSION A		073 073	073 073		073 - NEGATIVE	58	58	58	100	
2021-451	145	CLINITEK		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2021-451	147	Mission120		073 073	073 073		073 - NEGATIVE	30	30	30	100	
2021-451	148	Mission500		073 073	073 073		073 - NEGATIVE	31	31	31	100	
2021-452	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	50	50	50	100	
2021-452	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	35	35	35	100	
2021-452	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100	
2021-452	071	ATLAS		073 073	073 073		073 - NEGATIVE	15	15	15	100	
2021-452	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	101	101	101	100	
2021-452	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	37	37	37	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-452	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	108	108	108	100	
2021-452	094	CLINITEK		073 073	073 073		073 - NEGATIVE	105	105	105	100	
2021-452	102	ACON LAB		073 073	073 073		073 - NEGATIVE	156	156	156	100	
2021-452	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	28	28	28	100	
2021-452	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	13	13	13	100	
2021-452	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2021-452	140	MISSION A		073 073	073 073		073 - NEGATIVE	58	58	58	100	
2021-452	145	CLINITEK		073 073	073 073		073 - NEGATIVE	12	12	12	100	
2021-452	147	Mission120		073 073	073 073		073 - NEGATIVE	30	30	30	100	
2021-452	148	Mission500		073 073	073 073		073 - NEGATIVE	31	31	31	100	
058 URINALYSIS-GLUCOSE												
2021-453	011	AMES NM10S		076 078	075 077		074 - TRACES	50	1	49	0	
2021-453	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	50	13	49	100	
2021-453	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	50	33	49	100	
2021-453	011	AMES NM10S		076 078	075 077		078 - 2000mg/dl	50	3	49	100	
2021-453	013	BAYER CLI		076 078	075 077		076 - 500mg/dl2+	35	13	35	100	
2021-453	013	BAYER CLI		076 078	075 077		077 - 1000mg/dl	35	21	35	100	
2021-453	013	BAYER CLI		076 078	075 077		078 - 2000mg/dl	35	1	35	100	
2021-453	067	URISCAN		077 078	075 077		075 - 250mg/dl-+	50	2	50	100	
2021-453	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	50	4	50	100	
2021-453	067	URISCAN		077 078	075 077		077 - 1000mg/dl	50	12	50	100	
2021-453	067	URISCAN		077 078	075 077		078 - 2000mg/dl	50	32	50	100	
2021-453	071	ATLAS		075 077	075 077		076 - 500mg/dl2+	15	15	15	100	
2021-453	072	URS-9-TECO		075 077	075 077		073 - NEGATIVE	101	1	96	0	
2021-453	072	URS-9-TECO		075 077	075 077		074 - TRACES	101	1	96	0	
2021-453	072	URS-9-TECO		075 077	075 077		075 - 250mg/dl-+	101	2	96	100	
2021-453	072	URS-9-TECO		075 077	075 077		076 - 500mg/dl2+	101	48	96	100	
2021-453	072	URS-9-TECO		075 077	075 077		077 - 1000mg/dl	101	46	96	100	
2021-453	072	URS-9-TECO		075 077	075 077		078 - 2000mg/dl	101	3	96	0	
2021-453	080	ROCHE (CHE		076 078	075 077		076 - 500mg/dl2+	37	1	37	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-453	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	37	35	37	100
2021-453	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	37	1	37	100
2021-453	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	108	2	106	100
2021-453	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	108	64	106	100
2021-453	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	108	40	106	100
2021-453	089	AIMSTRIP		075 077	075 077		078 - 2000mg/dl	108	2	106	0
2021-453	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	105	1	105	100
2021-453	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	105	39	105	100
2021-453	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	105	63	105	100
2021-453	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	105	2	105	100
2021-453	102	ACON LAB		076 078	075 077		075 - 250mg/dl-+	156	2	156	100
2021-453	102	ACON LAB		076 078	075 077		076 - 500mg/dl2+	156	75	156	100
2021-453	102	ACON LAB		076 078	075 077		077 - 1000mg/dl	156	76	156	100
2021-453	102	ACON LAB		076 078	075 077		078 - 2000mg/dl	156	3	156	100
2021-453	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	28	23	28	100
2021-453	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	28	5	28	100
2021-453	135	Hybrido AU		076 078	075 077		077 - 1000mg/dl	13	8	13	100
2021-453	135	Hybrido AU		076 078	075 077		078 - 2000mg/dl	13	5	13	100
2021-453	138	URS-10T		075 077	075 077		075 - 250mg/dl-+	10	1	9	100
2021-453	138	URS-10T		075 077	075 077		076 - 500mg/dl2+	10	5	9	100
2021-453	138	URS-10T		075 077	075 077		077 - 1000mg/dl	10	3	9	100
2021-453	138	URS-10T		075 077	075 077		078 - 2000mg/dl	10	1	9	0
2021-453	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	58	14	58	100
2021-453	140	MISSION A		076 078	075 077		077 - 1000mg/dl	58	37	58	100
2021-453	140	MISSION A		076 078	075 077		078 - 2000mg/dl	58	7	58	100
2021-453	145	CLINITEK		075 077	075 077		076 - 500mg/dl2+	12	12	12	100
2021-453	147	Mission120		076 078	075 077		077 - 1000mg/dl	30	21	30	100
2021-453	147	Mission120		076 078	075 077		078 - 2000mg/dl	30	9	30	100
2021-453	148	Mission500		076 078	075 077		075 - 250mg/dl-+	31	1	31	100
2021-453	148	Mission500		076 078	075 077		076 - 500mg/dl2+	31	8	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-453	148	Mission500		076 078	075 077		077 - 1000mg/dl	31	19	31	100		
2021-453	148	Mission500		076 078	075 077		078 - 2000mg/dl	31	3	31	100		
058 URINALYSIS-GLUCOSE													
2021-454	011	AMES NM10S		076 078	076 078		075 - 250mg/dl-+	50	1	49	0		
2021-454	011	AMES NM10S		076 078	076 078		076 - 500mg/dl2+	50	12	49	100		
2021-454	011	AMES NM10S		076 078	076 078		077 - 1000mg/dl	50	36	49	100		
2021-454	011	AMES NM10S		076 078	076 078		078 - 2000mg/dl	50	1	49	100		
2021-454	013	BAYER CLI		076 078	076 078		076 - 500mg/dl2+	35	15	35	100		
2021-454	013	BAYER CLI		076 078	076 078		077 - 1000mg/dl	35	19	35	100		
2021-454	013	BAYER CLI		076 078	076 078		078 - 2000mg/dl	35	1	35	100		
2021-454	067	URISCAN		077 078	076 078		075 - 250mg/dl-+	50	1	49	0		
2021-454	067	URISCAN		077 078	076 078		076 - 500mg/dl2+	50	5	49	100		
2021-454	067	URISCAN		077 078	076 078		077 - 1000mg/dl	50	13	49	100		
2021-454	067	URISCAN		077 078	076 078		078 - 2000mg/dl	50	31	49	100		
2021-454	071	ATLAS		076 078	076 078		077 - 1000mg/dl	15	14	15	100		
2021-454	071	ATLAS		076 078	076 078		078 - 2000mg/dl	15	1	15	100		
2021-454	072	URS-9-TECO		075 077	076 078		075 - 250mg/dl-+	101	1	101	100		
2021-454	072	URS-9-TECO		075 077	076 078		076 - 500mg/dl2+	101	51	101	100		
2021-454	072	URS-9-TECO		075 077	076 078		077 - 1000mg/dl	101	44	101	100		
2021-454	072	URS-9-TECO		075 077	076 078		078 - 2000mg/dl	101	5	101	100		
2021-454	080	ROCHE (CHE		076 078	076 078		076 - 500mg/dl2+	37	1	37	100		
2021-454	080	ROCHE (CHE		076 078	076 078		077 - 1000mg/dl	37	35	37	100		
2021-454	080	ROCHE (CHE		076 078	076 078		078 - 2000mg/dl	37	1	37	100		
2021-454	089	AIMSTRIP		075 077	076 078		075 - 250mg/dl-+	108	2	108	100		
2021-454	089	AIMSTRIP		075 077	076 078		076 - 500mg/dl2+	108	79	108	100		
2021-454	089	AIMSTRIP		075 077	076 078		077 - 1000mg/dl	108	25	108	100		
2021-454	089	AIMSTRIP		075 077	076 078		078 - 2000mg/dl	108	2	108	100		
2021-454	094	CLINITEK		076 078	076 078		076 - 500mg/dl2+	105	37	105	100		
2021-454	094	CLINITEK		076 078	076 078		077 - 1000mg/dl	105	66	105	100		
2021-454	094	CLINITEK		076 078	076 078		078 - 2000mg/dl	105	2	105	100		

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-454	102	ACON LAB		075 077	076 078		075 - 250mg/dl-+	156	3	156	100
2021-454	102	ACON LAB		075 077	076 078		076 - 500mg/dl2+	156	87	156	100
2021-454	102	ACON LAB		075 077	076 078		077 - 1000mg/dl	156	65	156	100
2021-454	102	ACON LAB		075 077	076 078		078 - 2000mg/dl	156	1	156	100
2021-454	131	iCHEM VELO		075 077	076 078		076 - 500mg/dl2+	28	23	28	100
2021-454	131	iCHEM VELO		075 077	076 078		077 - 1000mg/dl	28	5	28	100
2021-454	135	Hybrido AU		077 078	076 078		077 - 1000mg/dl	13	4	13	100
2021-454	135	Hybrido AU		077 078	076 078		078 - 2000mg/dl	13	9	13	100
2021-454	138	URS-10T		076 078	076 078		074 - TRACES	10	1	9	0
2021-454	138	URS-10T		076 078	076 078		076 - 500mg/dl2+	10	3	9	100
2021-454	138	URS-10T		076 078	076 078		077 - 1000mg/dl	10	6	9	100
2021-454	140	MISSION A		076 078	076 078		075 - 250mg/dl-+	58	3	55	0
2021-454	140	MISSION A		076 078	076 078		076 - 500mg/dl2+	58	10	55	100
2021-454	140	MISSION A		076 078	076 078		077 - 1000mg/dl	58	42	55	100
2021-454	140	MISSION A		076 078	076 078		078 - 2000mg/dl	58	3	55	100
2021-454	145	CLINITEK		076 078	076 078		077 - 1000mg/dl	12	11	12	100
2021-454	145	CLINITEK		076 078	076 078		078 - 2000mg/dl	12	1	12	100
2021-454	147	Mission120		076 078	076 078		076 - 500mg/dl2+	30	1	30	100
2021-454	147	Mission120		076 078	076 078		077 - 1000mg/dl	30	25	30	100
2021-454	147	Mission120		076 078	076 078		078 - 2000mg/dl	30	4	30	100
2021-454	148	Mission500		076 078	076 078		076 - 500mg/dl2+	31	11	31	100
2021-454	148	Mission500		076 078	076 078		077 - 1000mg/dl	31	20	31	100
058 URINALYSIS-GLUCOSE											
2021-455	011	AMES NM10S		074 076	074 076		073 - NEGATIVE	50	1	49	0
2021-455	011	AMES NM10S		074 076	074 076		074 - TRACES	50	22	49	100
2021-455	011	AMES NM10S		074 076	074 076		075 - 250mg/dl-+	50	25	49	100
2021-455	011	AMES NM10S		074 076	074 076		076 - 500mg/dl2+	50	2	49	100
2021-455	013	BAYER CLI		074 076	074 076		074 - TRACES	35	13	35	100
2021-455	013	BAYER CLI		074 076	074 076		075 - 250mg/dl-+	35	22	35	100
2021-455	067	URISCAN		074 076	074 076		075 - 250mg/dl-+	50	22	43	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	067	URISCAN		074 076	074 076		076 - 500mg/dl2+	50	21	43	100
2021-455	067	URISCAN		074 076	074 076		077 - 1000mg/dl	50	7	43	0
2021-455	071	ATLAS		076 078	074 076		076 - 500mg/dl2+	15	3	15	100
2021-455	071	ATLAS		076 078	074 076		077 - 1000mg/dl	15	12	15	100
2021-455	072	URS-9-TECO		074 076	074 076		073 - NEGATIVE	101	9	91	0
2021-455	072	URS-9-TECO		074 076	074 076		074 - TRACES	101	41	91	100
2021-455	072	URS-9-TECO		074 076	074 076		075 - 250mg/dl-+	101	46	91	100
2021-455	072	URS-9-TECO		074 076	074 076		076 - 500mg/dl2+	101	4	91	100
2021-455	072	URS-9-TECO		074 076	074 076		077 - 1000mg/dl	101	1	91	0
2021-455	080	ROCHE (CHE		076 078	074 076		075 - 250mg/dl-+	37	1	37	100
2021-455	080	ROCHE (CHE		076 078	074 076		076 - 500mg/dl2+	37	2	37	100
2021-455	080	ROCHE (CHE		076 078	074 076		077 - 1000mg/dl	37	33	37	100
2021-455	080	ROCHE (CHE		076 078	074 076		078 - 2000mg/dl	37	1	37	100
2021-455	089	AIMSTRIP		074 076	074 076		073 - NEGATIVE	108	2	106	0
2021-455	089	AIMSTRIP		074 076	074 076		074 - TRACES	108	23	106	100
2021-455	089	AIMSTRIP		074 076	074 076		075 - 250mg/dl-+	108	83	106	100
2021-455	094	CLINITEK		074 076	074 076		073 - NEGATIVE	105	1	102	0
2021-455	094	CLINITEK		074 076	074 076		074 - TRACES	105	41	102	100
2021-455	094	CLINITEK		074 076	074 076		075 - 250mg/dl-+	105	61	102	100
2021-455	094	CLINITEK		074 076	074 076		077 - 1000mg/dl	105	2	102	0
2021-455	102	ACON LAB		074 076	074 076		073 - NEGATIVE	156	1	155	0
2021-455	102	ACON LAB		074 076	074 076		074 - TRACES	156	27	155	100
2021-455	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	156	126	155	100
2021-455	102	ACON LAB		074 076	074 076		076 - 500mg/dl2+	156	2	155	100
2021-455	131	iCHEM VELO		075 077	074 076		074 - TRACES	28	1	28	100
2021-455	131	iCHEM VELO		075 077	074 076		076 - 500mg/dl2+	28	23	28	100
2021-455	131	iCHEM VELO		075 077	074 076		077 - 1000mg/dl	28	4	28	100
2021-455	135	Hybrido AU		077 078	074 076		077 - 1000mg/dl	13	5	13	100
2021-455	135	Hybrido AU		077 078	074 076		078 - 2000mg/dl	13	8	13	100
2021-455	138	URS-10T		073 075	074 076		073 - NEGATIVE	10	3	10	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-455	138	URS-10T		073 075	074 076		074 - TRACES	10	3	10	100	
2021-455	138	URS-10T		073 075	074 076		075 - 250mg/dl-+	10	3	10	100	
2021-455	138	URS-10T		073 075	074 076		076 - 500mg/dl2+	10	1	10	100	
2021-455	140	MISSION A		074 076	074 076		074 - TRACES	58	2	57	100	
2021-455	140	MISSION A		074 076	074 076		075 - 250mg/dl-+	58	54	57	100	
2021-455	140	MISSION A		074 076	074 076		076 - 500mg/dl2+	58	1	57	100	
2021-455	140	MISSION A		074 076	074 076		077 - 1000mg/dl	58	1	57	0	
2021-455	145	CLINITEK		076 078	074 076		077 - 1000mg/dl	12	11	12	100	
2021-455	145	CLINITEK		076 078	074 076		078 - 2000mg/dl	12	1	12	100	
2021-455	147	Mission120		074 076	074 076		074 - TRACES	30	1	30	100	
2021-455	147	Mission120		074 076	074 076		075 - 250mg/dl-+	30	29	30	100	
2021-455	148	Mission500		074 076	074 076		074 - TRACES	31	1	31	100	
2021-455	148	Mission500		074 076	074 076		075 - 250mg/dl-+	31	30	31	100	
059 URINALYS.HG.OR BLOOD												
2021-451	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	4	50	100	
2021-451	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	50	46	50	100	
2021-451	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	36	3	36	100	
2021-451	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	36	33	36	100	
2021-451	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	50	4	50	100	
2021-451	067	URISCAN		068 069	068 069		069 - POS.(3+))	50	46	50	100	
2021-451	071	ATLAS		068 069	068 069		069 - POS.(3+))	15	15	15	100	
2021-451	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	101	25	101	100	
2021-451	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	101	76	101	100	
2021-451	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	37	2	37	100	
2021-451	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	37	35	37	100	
2021-451	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	109	1	108	0	
2021-451	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	109	1	108	100	
2021-451	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	109	107	108	100	
2021-451	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	105	4	105	100	
2021-451	094	CLINITEK		068 069	068 069		069 - POS.(3+))	105	101	105	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-451	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	157	3	157	100		
2021-451	102	ACON LAB		068 069	068 069		069 - POS.(3+))	157	154	157	100		
2021-451	131	iCHEM VELO		067 069	068 069		067 - TRACES	27	1	27	100		
2021-451	131	iCHEM VELO		067 069	068 069		068 - POS.(1-2+)	27	20	27	100		
2021-451	131	iCHEM VELO		067 069	068 069		069 - POS.(3+))	27	6	27	100		
2021-451	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	13	12	13	100		
2021-451	135	Hybrido AU		067 069	068 069		069 - POS.(3+))	13	1	13	100		
2021-451	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	10	1	10	100		
2021-451	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	9	10	100		
2021-451	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	56	2	56	100		
2021-451	140	MISSION A		068 069	068 069		069 - POS.(3+))	56	54	56	100		
2021-451	145	CLINITEK		068 069	068 069		068 - POS.(1-2+)	12	3	12	100		
2021-451	145	CLINITEK		068 069	068 069		069 - POS.(3+))	12	9	12	100		
2021-451	147	Mission120		068 069	068 069		069 - POS.(3+))	31	31	31	100		
2021-451	148	Mission500		068 069	068 069		068 - POS.(1-2+)	31	1	31	100		
2021-451	148	Mission500		068 069	068 069		069 - POS.(3+))	31	30	31	100		
059 URINALYS.HG.OR BLOOD													
2021-452	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	1	50	100		
2021-452	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	50	49	50	100		
2021-452	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	36	1	36	100		
2021-452	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	36	35	36	100		
2021-452	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	50	3	50	100		
2021-452	067	URISCAN		068 069	068 069		069 - POS.(3+))	50	47	50	100		
2021-452	071	ATLAS		068 069	068 069		069 - POS.(3+))	15	15	15	100		
2021-452	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	101	7	101	100		
2021-452	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	101	94	101	100		
2021-452	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	37	1	37	100		
2021-452	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	37	36	37	100		
2021-452	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	109	1	109	100		
2021-452	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	109	108	109	100		

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-452	094	CLINITEK		068 069	068 069		066 - NEGATIVE	105	1	104	0	
2021-452	094	CLINITEK		068 069	068 069		069 - POS.(3+)	105	104	104	100	
2021-452	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	157	1	157	100	
2021-452	102	ACON LAB		068 069	068 069		069 - POS.(3+)	157	156	157	100	
2021-452	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	27	2	27	100	
2021-452	131	iCHEM VELO		068 069	068 069		069 - POS.(3+)	27	25	27	100	
2021-452	135	Hybrido AU		068 069	068 069		068 - POS.(1-2+)	13	3	13	100	
2021-452	135	Hybrido AU		068 069	068 069		069 - POS.(3+)	13	10	13	100	
2021-452	138	URS-10T		068 069	068 069		066 - NEGATIVE	10	1	9	0	
2021-452	138	URS-10T		068 069	068 069		069 - POS.(3+)	10	9	9	100	
2021-452	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	56	1	56	100	
2021-452	140	MISSION A		068 069	068 069		069 - POS.(3+)	56	55	56	100	
2021-452	145	CLINITEK		068 069	068 069		069 - POS.(3+)	12	12	12	100	
2021-452	147	Mission120		068 069	068 069		069 - POS.(3+)	31	31	31	100	
2021-452	148	Mission500		068 069	068 069		068 - POS.(1-2+)	31	1	31	100	
2021-452	148	Mission500		068 069	068 069		069 - POS.(3+)	31	30	31	100	
059 URINALYS.HG.OR BLOOD												
2021-453	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	50	50	50	100	
2021-453	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	36	36	36	100	
2021-453	067	URISCAN		066 066	066 066		066 - NEGATIVE	50	50	50	100	
2021-453	071	ATLAS		066 066	066 066		066 - NEGATIVE	15	15	15	100	
2021-453	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	101	101	101	100	
2021-453	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	37	37	37	100	
2021-453	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	109	109	109	100	
2021-453	094	CLINITEK		066 066	066 066		066 - NEGATIVE	105	104	104	100	
2021-453	094	CLINITEK		066 066	066 066		067 - TRACES	105	1	104	0	
2021-453	102	ACON LAB		066 066	066 066		066 - NEGATIVE	157	157	157	100	
2021-453	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	27	27	27	100	
2021-453	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	13	13	13	100	
2021-453	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-453	140	MISSION A		066 066	066 066		066 - NEGATIVE	56	56	56	100	
2021-453	145	CLINITEK		066 066	066 066		066 - NEGATIVE	12	12	12	100	
2021-453	147	Mission120		066 066	066 066		066 - NEGATIVE	31	31	31	100	
2021-453	148	Mission500		066 066	066 066		066 - NEGATIVE	31	31	31	100	
059 URINALYS.HG.OR BLOOD												
2021-454	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	50	4	50	100	
2021-454	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	50	46	50	100	
2021-454	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	36	3	36	100	
2021-454	013	BAYER CLI		068 069	068 069		069 - POS.(3+)	36	33	36	100	
2021-454	067	URISCAN		068 069	068 069		067 - TRACES	50	1	49	0	
2021-454	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	50	3	49	100	
2021-454	067	URISCAN		068 069	068 069		069 - POS.(3+)	50	46	49	100	
2021-454	071	ATLAS		068 069	068 069		069 - POS.(3+)	15	15	15	100	
2021-454	072	URS-9-TECO		068 069	068 069		066 - NEGATIVE	101	1	100	0	
2021-454	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	101	12	100	100	
2021-454	072	URS-9-TECO		068 069	068 069		069 - POS.(3+)	101	88	100	100	
2021-454	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	37	2	37	100	
2021-454	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+)	37	35	37	100	
2021-454	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	109	1	109	100	
2021-454	089	AIMSTRIP		068 069	068 069		069 - POS.(3+)	109	108	109	100	
2021-454	094	CLINITEK		068 069	068 069		066 - NEGATIVE	105	1	104	0	
2021-454	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	105	2	104	100	
2021-454	094	CLINITEK		068 069	068 069		069 - POS.(3+)	105	102	104	100	
2021-454	102	ACON LAB		068 069	068 069		066 - NEGATIVE	157	1	156	0	
2021-454	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	157	3	156	100	
2021-454	102	ACON LAB		068 069	068 069		069 - POS.(3+)	157	153	156	100	
2021-454	131	ICHEM VELO		068 069	068 069		068 - POS.(1-2+)	27	2	27	100	
2021-454	131	ICHEM VELO		068 069	068 069		069 - POS.(3+)	27	25	27	100	
2021-454	135	Hybrido AU		068 069	068 069		068 - POS.(1-2+)	13	1	13	100	
2021-454	135	Hybrido AU		068 069	068 069		069 - POS.(3+)	13	12	13	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-454	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	10	2	10	100	
2021-454	138	URS-10T		068 069	068 069		069 - POS.(3+)	10	8	10	100	
2021-454	140	MISSION A		068 069	068 069		066 - NEGATIVE	56	1	55	0	
2021-454	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	56	3	55	100	
2021-454	140	MISSION A		068 069	068 069		069 - POS.(3+)	56	52	55	100	
2021-454	145	CLINITEK		068 069	068 069		069 - POS.(3+)	12	12	12	100	
2021-454	147	Mission120		068 069	068 069		069 - POS.(3+)	31	31	31	100	
2021-454	148	Mission500		068 069	068 069		068 - POS.(1-2+)	31	3	31	100	
2021-454	148	Mission500		068 069	068 069		069 - POS.(3+)	31	28	31	100	
059 URINALYS.HG.OR BLOOD												
2021-455	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	50	50	50	100	
2021-455	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	36	35	35	100	
2021-455	013	BAYER CLI		066 066	066 066		067 - TRACES	36	1	35	0	
2021-455	067	URISCAN		066 066	066 066		066 - NEGATIVE	50	49	49	100	
2021-455	067	URISCAN		066 066	066 066		068 - POS.(1-2+)	50	1	49	0	
2021-455	071	ATLAS		066 066	066 066		066 - NEGATIVE	15	15	15	100	
2021-455	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	101	101	101	100	
2021-455	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	37	37	37	100	
2021-455	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	109	109	109	100	
2021-455	094	CLINITEK		066 066	066 066		066 - NEGATIVE	105	103	103	100	
2021-455	094	CLINITEK		066 066	066 066		067 - TRACES	105	2	103	0	
2021-455	102	ACON LAB		066 066	066 066		066 - NEGATIVE	157	157	157	100	
2021-455	131	ICHEM VELO		066 066	066 066		066 - NEGATIVE	27	27	27	100	
2021-455	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	13	13	13	100	
2021-455	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2021-455	140	MISSION A		066 066	066 066		066 - NEGATIVE	56	55	55	100	
2021-455	140	MISSION A		066 066	066 066		069 - POS.(3+)	56	1	55	0	
2021-455	145	CLINITEK		066 066	066 066		066 - NEGATIVE	12	12	12	100	
2021-455	147	Mission120		066 066	066 066		066 - NEGATIVE	31	31	31	100	
2021-455	148	Mission500		066 066	066 066		066 - NEGATIVE	31	31	31	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
060 URINALYSIS-KETONES												
2021-451	011	AMES NM10S		093 094	093 094		093 - MOD. (2+)	50	1	50	100	
2021-451	011	AMES NM10S		093 094	093 094		094 - LARGE-3-4+	50	49	50	100	
2021-451	013	BAYER CLI		093 094	093 094		094 - LARGE-3-4+	36	36	36	100	
2021-451	067	URISCAN		093 094	093 094		092 - SMALL (1+)	50	2	48	0	
2021-451	067	URISCAN		093 094	093 094		093 - MOD. (2+)	50	12	48	100	
2021-451	067	URISCAN		093 094	093 094		094 - LARGE-3-4+	50	36	48	100	
2021-451	071	ATLAS		093 094	093 094		094 - LARGE-3-4+	15	15	15	100	
2021-451	072	URS-9-TECO		093 094	093 094		093 - MOD. (2+)	101	4	101	100	
2021-451	072	URS-9-TECO		093 094	093 094		094 - LARGE-3-4+	101	97	101	100	
2021-451	080	ROCHE (CHE		093 094	093 094		093 - MOD. (2+)	37	1	37	100	
2021-451	080	ROCHE (CHE		093 094	093 094		094 - LARGE-3-4+	37	36	37	100	
2021-451	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	109	2	109	100	
2021-451	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	109	107	109	100	
2021-451	094	CLINITEK		093 094	093 094		093 - MOD. (2+)	105	1	105	100	
2021-451	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	105	104	105	100	
2021-451	102	ACON LAB		093 094	093 094		093 - MOD. (2+)	156	3	156	100	
2021-451	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	156	153	156	100	
2021-451	131	iCHEM VELO		092 094	093 094		093 - MOD. (2+)	27	16	27	100	
2021-451	131	iCHEM VELO		092 094	093 094		094 - LARGE-3-4+	27	11	27	100	
2021-451	135	Hybrido AU		093 094	093 094		094 - LARGE-3-4+	13	13	13	100	
2021-451	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	10	10	10	100	
2021-451	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	58	58	58	100	
2021-451	145	CLINITEK		093 094	093 094		094 - LARGE-3-4+	12	12	12	100	
2021-451	147	Mission120		093 094	093 094		094 - LARGE-3-4+	30	30	30	100	
2021-451	148	Mission500		093 094	093 094		093 - MOD. (2+)	31	2	31	100	
2021-451	148	Mission500		093 094	093 094		094 - LARGE-3-4+	31	29	31	100	
2021-452	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	50	50	50	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-452	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	36	36	36	100
2021-452	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100
2021-452	071	ATLAS		090 090	090 090		090 - NEGATIVE	15	15	15	100
2021-452	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	101	101	101	100
2021-452	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	37	37	37	100
2021-452	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	109	109	109	100
2021-452	094	CLINITEK		090 090	090 090		090 - NEGATIVE	105	105	105	100
2021-452	102	ACON LAB		090 090	090 090		090 - NEGATIVE	156	156	156	100
2021-452	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100
2021-452	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	13	13	13	100
2021-452	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100
2021-452	140	MISSION A		090 090	090 090		090 - NEGATIVE	58	58	58	100
2021-452	145	CLINITEK		090 090	090 090		090 - NEGATIVE	12	12	12	100
2021-452	147	Mission120		090 090	090 090		090 - NEGATIVE	30	30	30	100
2021-452	148	Mission500		090 090	090 090		090 - NEGATIVE	31	31	31	100
060 URINALYSIS-KETONES											
2021-453	011	AMES NMI0S		090 090	090 090		090 - NEGATIVE	50	50	50	100
2021-453	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	36	36	36	100
2021-453	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100
2021-453	071	ATLAS		090 090	090 090		090 - NEGATIVE	15	15	15	100
2021-453	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	101	101	101	100
2021-453	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	37	37	37	100
2021-453	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	109	109	109	100
2021-453	094	CLINITEK		090 090	090 090		090 - NEGATIVE	105	105	105	100
2021-453	102	ACON LAB		090 090	090 090		090 - NEGATIVE	156	156	156	100
2021-453	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100
2021-453	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	13	13	13	100
2021-453	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100
2021-453	140	MISSION A		090 090	090 090		090 - NEGATIVE	58	58	58	100
2021-453	145	CLINITEK		090 090	090 090		090 - NEGATIVE	12	12	12	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-453	147	Mission120		090	090	090	090	090 - NEGATIVE	30	30	30	100
2021-453	148	Mission500		090	090	090	090	090 - NEGATIVE	31	31	31	100
060 URINALYSIS-KETONES												
2021-454	011	AMES NM10S		090	090	090	090	090 - NEGATIVE	50	50	50	100
2021-454	013	BAYER CLI		090	090	090	090	090 - NEGATIVE	36	36	36	100
2021-454	067	URISCAN		090	090	090	090	090 - NEGATIVE	50	50	50	100
2021-454	071	ATLAS		090	090	090	090	090 - NEGATIVE	15	15	15	100
2021-454	072	URS-9-TECO		090	090	090	090	090 - NEGATIVE	101	100	100	100
2021-454	072	URS-9-TECO		090	090	090	090	093 - MOD. (2+)	101	1	100	0
2021-454	080	ROCHE (CHE		090	090	090	090	090 - NEGATIVE	37	37	37	100
2021-454	089	AIMSTRIP		090	090	090	090	090 - NEGATIVE	109	109	109	100
2021-454	094	CLINITEK		090	090	090	090	090 - NEGATIVE	105	105	105	100
2021-454	102	ACON LAB		090	090	090	090	090 - NEGATIVE	156	156	156	100
2021-454	131	iCHEM VELO		090	090	090	090	090 - NEGATIVE	27	27	27	100
2021-454	135	Hybrido AU		090	090	090	090	090 - NEGATIVE	13	13	13	100
2021-454	138	URS-10T		090	090	090	090	090 - NEGATIVE	10	10	10	100
2021-454	140	MISSION A		090	090	090	090	090 - NEGATIVE	58	57	57	100
2021-454	140	MISSION A		090	090	090	090	094 - LARGE-3-4+	58	1	57	0
2021-454	145	CLINITEK		090	090	090	090	090 - NEGATIVE	12	12	12	100
2021-454	147	Mission120		090	090	090	090	090 - NEGATIVE	30	30	30	100
2021-454	148	Mission500		090	090	090	090	090 - NEGATIVE	31	31	31	100
2021-455	011	AMES NM10S		092	094	092	094	091 - TRACE	50	1	49	0
2021-455	011	AMES NM10S		092	094	092	094	092 - SMALL (1+)	50	4	49	100
2021-455	011	AMES NM10S		092	094	092	094	093 - MOD. (2+)	50	36	49	100
2021-455	011	AMES NM10S		092	094	092	094	094 - LARGE-3-4+	50	9	49	100
2021-455	013	BAYER CLI		092	094	092	094	092 - SMALL (1+)	36	2	36	100
2021-455	013	BAYER CLI		092	094	092	094	093 - MOD. (2+)	36	30	36	100
2021-455	013	BAYER CLI		092	094	092	094	094 - LARGE-3-4+	36	4	36	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	067	URISCAN		092 094	092 094		091 - TRACE	50	1	49	0
2021-455	067	URISCAN		092 094	092 094		092 - SMALL (1+)	50	4	49	100
2021-455	067	URISCAN		092 094	092 094		093 - MOD. (2+)	50	37	49	100
2021-455	067	URISCAN		092 094	092 094		094 - LARGE-3-4+	50	8	49	100
2021-455	071	ATLAS		092 094	092 094		093 - MOD. (2+)	15	15	15	100
2021-455	072	URS-9-TECO		092 094	092 094		090 - NEGATIVE	101	4	96	0
2021-455	072	URS-9-TECO		092 094	092 094		091 - TRACE	101	1	96	0
2021-455	072	URS-9-TECO		092 094	092 094		092 - SMALL (1+)	101	5	96	100
2021-455	072	URS-9-TECO		092 094	092 094		093 - MOD. (2+)	101	69	96	100
2021-455	072	URS-9-TECO		092 094	092 094		094 - LARGE-3-4+	101	22	96	100
2021-455	080	ROCHE (CHE		092 094	092 094		091 - TRACE	37	1	36	0
2021-455	080	ROCHE (CHE		092 094	092 094		092 - SMALL (1+)	37	7	36	100
2021-455	080	ROCHE (CHE		092 094	092 094		093 - MOD. (2+)	37	22	36	100
2021-455	080	ROCHE (CHE		092 094	092 094		094 - LARGE-3-4+	37	7	36	100
2021-455	089	AIMSTRIP		092 094	092 094		090 - NEGATIVE	109	3	104	0
2021-455	089	AIMSTRIP		092 094	092 094		091 - TRACE	109	2	104	0
2021-455	089	AIMSTRIP		092 094	092 094		092 - SMALL (1+)	109	10	104	100
2021-455	089	AIMSTRIP		092 094	092 094		093 - MOD. (2+)	109	54	104	100
2021-455	089	AIMSTRIP		092 094	092 094		094 - LARGE-3-4+	109	40	104	100
2021-455	094	CLINITEK		092 094	092 094		090 - NEGATIVE	105	1	103	0
2021-455	094	CLINITEK		092 094	092 094		091 - TRACE	105	1	103	0
2021-455	094	CLINITEK		092 094	092 094		092 - SMALL (1+)	105	10	103	100
2021-455	094	CLINITEK		092 094	092 094		093 - MOD. (2+)	105	87	103	100
2021-455	094	CLINITEK		092 094	092 094		094 - LARGE-3-4+	105	6	103	100
2021-455	102	ACON LAB		092 094	092 094		092 - SMALL (1+)	156	10	156	100
2021-455	102	ACON LAB		092 094	092 094		093 - MOD. (2+)	156	86	156	100
2021-455	102	ACON LAB		092 094	092 094		094 - LARGE-3-4+	156	60	156	100
2021-455	131	iCHEM VELO		091 093	092 094		091 - TRACE	27	1	27	100
2021-455	131	iCHEM VELO		091 093	092 094		092 - SMALL (1+)	27	20	27	100
2021-455	131	iCHEM VELO		091 093	092 094		093 - MOD. (2+)	27	6	27	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-455	135	Hybrido AU		092 094	092 094		093 - MOD. (2+)	13	13	13	100	
2021-455	138	URS-10T		092 094	092 094		092 - SMALL (1+)	10	1	10	100	
2021-455	138	URS-10T		092 094	092 094		093 - MOD. (2+)	10	7	10	100	
2021-455	138	URS-10T		092 094	092 094		094 - LARGE-3-4+	10	2	10	100	
2021-455	140	MISSION A		093 094	092 094		090 - NEGATIVE	58	1	57	0	
2021-455	140	MISSION A		093 094	092 094		093 - MOD. (2+)	58	14	57	100	
2021-455	140	MISSION A		093 094	092 094		094 - LARGE-3-4+	58	43	57	100	
2021-455	145	CLINITEK		092 094	092 094		092 - SMALL (1+)	12	1	12	100	
2021-455	145	CLINITEK		092 094	092 094		093 - MOD. (2+)	12	10	12	100	
2021-455	145	CLINITEK		092 094	092 094		094 - LARGE-3-4+	12	1	12	100	
2021-455	147	Mission120		093 094	092 094		093 - MOD. (2+)	30	2	30	100	
2021-455	147	Mission120		093 094	092 094		094 - LARGE-3-4+	30	28	30	100	
2021-455	148	Mission500		092 094	092 094		092 - SMALL (1+)	31	1	31	100	
2021-455	148	Mission500		092 094	092 094		093 - MOD. (2+)	31	19	31	100	
2021-455	148	Mission500		092 094	092 094		094 - LARGE-3-4+	31	11	31	100	
061 URINALYSIS-NITRATE												
2021-451	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2021-451	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	36	36	36	100	
2021-451	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2021-451	071	ATLAS		070 070	070 070		070 - NEGATIVE	15	15	15	100	
2021-451	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	101	101	101	100	
2021-451	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	37	37	37	100	
2021-451	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	109	109	109	100	
2021-451	094	CLINITEK		070 070	070 070		070 - NEGATIVE	105	105	105	100	
2021-451	102	ACON LAB		070 070	070 070		070 - NEGATIVE	156	156	156	100	
2021-451	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100	
2021-451	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	13	13	13	100	
2021-451	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2021-451	140	MISSION A		070 070	070 070		070 - NEGATIVE	57	57	57	100	
2021-451	145	CLINITEK		070 070	070 070		070 - NEGATIVE	12	12	12	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-451	147	Mission120		070	070	070	070	070 - NEGATIVE	31	31	31	100	
2021-451	148	Mission500		070	070	070	070	070 - NEGATIVE	31	31	31	100	
061 URINALYSIS-NITRATE													
2021-452	011	AMES NM10S		070	070	070	070	070 - NEGATIVE	50	50	50	100	
2021-452	013	BAYER CLI		070	070	070	070	070 - NEGATIVE	36	36	36	100	
2021-452	067	URISCAN		070	070	070	070	070 - NEGATIVE	50	50	50	100	
2021-452	071	ATLAS		070	070	070	070	070 - NEGATIVE	15	15	15	100	
2021-452	072	URS-9-TECO		070	070	070	070	070 - NEGATIVE	101	101	101	100	
2021-452	080	ROCHE (CHE		070	070	070	070	070 - NEGATIVE	37	37	37	100	
2021-452	089	AIMSTRIP		070	070	070	070	070 - NEGATIVE	109	109	109	100	
2021-452	094	CLINITEK		070	070	070	070	070 - NEGATIVE	105	105	105	100	
2021-452	102	ACON LAB		070	070	070	070	070 - NEGATIVE	156	156	156	100	
2021-452	131	iCHEM VELO		070	070	070	070	070 - NEGATIVE	27	27	27	100	
2021-452	135	Hybrido AU		070	070	070	070	070 - NEGATIVE	13	13	13	100	
2021-452	138	URS-10T		070	070	070	070	070 - NEGATIVE	10	10	10	100	
2021-452	140	MISSION A		070	070	070	070	070 - NEGATIVE	57	57	57	100	
2021-452	145	CLINITEK		070	070	070	070	070 - NEGATIVE	12	12	12	100	
2021-452	147	Mission120		070	070	070	070	070 - NEGATIVE	31	31	31	100	
2021-452	148	Mission500		070	070	070	070	070 - NEGATIVE	31	31	31	100	
2021-453	011	AMES NM10S		070	070	070	070	070 - NEGATIVE	50	50	50	100	
2021-453	013	BAYER CLI		070	070	070	070	070 - NEGATIVE	36	36	36	100	
2021-453	067	URISCAN		070	070	070	070	070 - NEGATIVE	50	50	50	100	
2021-453	071	ATLAS		070	070	070	070	070 - NEGATIVE	15	15	15	100	
2021-453	072	URS-9-TECO		070	070	070	070	070 - NEGATIVE	101	101	101	100	
2021-453	080	ROCHE (CHE		070	070	070	070	070 - NEGATIVE	37	37	37	100	
2021-453	089	AIMSTRIP		070	070	070	070	070 - NEGATIVE	109	109	109	100	
2021-453	094	CLINITEK		070	070	070	070	070 - NEGATIVE	105	105	105	100	
2021-453	102	ACON LAB		070	070	070	070	070 - NEGATIVE	156	156	156	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-453	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100	
2021-453	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	13	13	13	100	
2021-453	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2021-453	140	MISSION A		070 070	070 070		070 - NEGATIVE	57	57	57	100	
2021-453	145	CLINITEK		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2021-453	147	Mission120		070 070	070 070		070 - NEGATIVE	31	31	31	100	
2021-453	148	Mission500		070 070	070 070		070 - NEGATIVE	31	31	31	100	
061 URINALYSIS-NITRATE												
2021-454	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2021-454	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	36	36	36	100	
2021-454	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2021-454	071	ATLAS		070 070	070 070		070 - NEGATIVE	15	15	15	100	
2021-454	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	101	101	101	100	
2021-454	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	37	37	37	100	
2021-454	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	109	109	109	100	
2021-454	094	CLINITEK		070 070	070 070		070 - NEGATIVE	105	105	105	100	
2021-454	102	ACON LAB		070 070	070 070		070 - NEGATIVE	156	156	156	100	
2021-454	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100	
2021-454	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	13	13	13	100	
2021-454	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2021-454	140	MISSION A		070 070	070 070		070 - NEGATIVE	57	56	56	100	
2021-454	140	MISSION A		070 070	070 070		072 - POSITIVE	57	1	56	0	
2021-454	145	CLINITEK		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2021-454	147	Mission120		070 070	070 070		070 - NEGATIVE	31	31	31	100	
2021-454	148	Mission500		070 070	070 070		070 - NEGATIVE	31	31	31	100	
2021-455	011	AMES NM10S		071 072	071 072		072 - POSITIVE	50	50	50	100	
2021-455	013	BAYER CLI		071 072	071 072		072 - POSITIVE	36	36	36	100	
2021-455	067	URISCAN		071 072	071 072		072 - POSITIVE	50	50	50	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-455	071	ATLAS		071	072	071	072	072 - POSITIVE	15	15	15	100
2021-455	072	URS-9-TECO		071	072	071	072	072 - POSITIVE	101	101	101	100
2021-455	080	ROCHE (CHE		071	072	071	072	072 - POSITIVE	37	37	37	100
2021-455	089	AIMSTRIP		071	072	071	072	071 - WEAK POS.	109	1	109	100
2021-455	089	AIMSTRIP		071	072	071	072	072 - POSITIVE	109	108	109	100
2021-455	094	CLINITEK		071	072	071	072	072 - POSITIVE	105	105	105	100
2021-455	102	ACON LAB		071	072	071	072	070 - NEGATIVE	156	1	155	0
2021-455	102	ACON LAB		071	072	071	072	072 - POSITIVE	156	155	155	100
2021-455	131	iCHEM VELO		071	072	071	072	072 - POSITIVE	27	27	27	100
2021-455	135	Hybrido AU		071	072	071	072	072 - POSITIVE	13	13	13	100
2021-455	138	URS-10T		071	072	071	072	072 - POSITIVE	10	10	10	100
2021-455	140	MISSION A		071	072	071	072	070 - NEGATIVE	57	1	56	0
2021-455	140	MISSION A		071	072	071	072	072 - POSITIVE	57	56	56	100
2021-455	145	CLINITEK		071	072	071	072	072 - POSITIVE	12	12	12	100
2021-455	147	Mission120		071	072	071	072	072 - POSITIVE	31	31	31	100
2021-455	148	Mission500		071	072	071	072	072 - POSITIVE	31	31	31	100
062 URINALYSIS-pH												
2021-451	011	AMES NM10S		086	088	086	088	087 - 7.5	50	11	50	100
2021-451	011	AMES NM10S		086	088	086	088	088 - 8.0OR MORE	50	39	50	100
2021-451	013	BAYER CLI		086	088	086	088	087 - 7.5	37	10	37	100
2021-451	013	BAYER CLI		086	088	086	088	088 - 8.0OR MORE	37	27	37	100
2021-451	067	URISCAN		085	088	086	088	086 - 7.0	50	2	50	100
2021-451	067	URISCAN		085	088	086	088	087 - 7.5	50	38	50	100
2021-451	067	URISCAN		085	088	086	088	088 - 8.0OR MORE	50	10	50	100
2021-451	071	ATLAS		085	088	086	088	087 - 7.5	15	15	15	100
2021-451	072	URS-9-TECO		085	088	086	088	085 - 6.5	100	1	100	100
2021-451	072	URS-9-TECO		085	088	086	088	086 - 7.0	100	14	100	100
2021-451	072	URS-9-TECO		085	088	086	088	087 - 7.5	100	71	100	100
2021-451	072	URS-9-TECO		085	088	086	088	088 - 8.0OR MORE	100	14	100	100
2021-451	080	ROCHE (CHE		086	088	086	088	086 - 7.0	37	4	37	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-451	080	ROCHE (CHE		086 088	086 088		088 - 8.00R MORE	37	33	37	100		
2021-451	089	AIMSTRIP		086 088	086 088		086 - 7.0	109	2	109	100		
2021-451	089	AIMSTRIP		086 088	086 088		087 - 7.5	109	44	109	100		
2021-451	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	109	63	109	100		
2021-451	094	CLINITEK		086 088	086 088		087 - 7.5	105	25	105	100		
2021-451	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	105	80	105	100		
2021-451	102	ACON LAB		086 088	086 088		087 - 7.5	156	42	156	100		
2021-451	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	156	114	156	100		
2021-451	131	iCHEM VELO		084 088	086 088		086 - 7.0	27	15	27	100		
2021-451	131	iCHEM VELO		084 088	086 088		088 - 8.00R MORE	27	12	27	100		
2021-451	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	13	13	13	100		
2021-451	138	URS-10T		085 088	086 088		086 - 7.0	10	1	10	100		
2021-451	138	URS-10T		085 088	086 088		087 - 7.5	10	8	10	100		
2021-451	138	URS-10T		085 088	086 088		088 - 8.00R MORE	10	1	10	100		
2021-451	140	MISSION A		086 088	086 088		087 - 7.5	58	24	58	100		
2021-451	140	MISSION A		086 088	086 088		088 - 8.00R MORE	58	34	58	100		
2021-451	145	CLINITEK		086 088	086 088		087 - 7.5	12	5	12	100		
2021-451	145	CLINITEK		086 088	086 088		088 - 8.00R MORE	12	7	12	100		
2021-451	147	Mission120		086 088	086 088		087 - 7.5	30	9	30	100		
2021-451	147	Mission120		086 088	086 088		088 - 8.00R MORE	30	21	30	100		
2021-451	148	Mission500		086 088	086 088		087 - 7.5	31	3	31	100		
2021-451	148	Mission500		086 088	086 088		088 - 8.00R MORE	31	28	31	100		
062 URINALYSIS-pH													
2021-452	011	AMES NM10S		085 088	085 088		086 - 7.0	50	4	50	100		
2021-452	011	AMES NM10S		085 088	085 088		087 - 7.5	50	45	50	100		
2021-452	011	AMES NM10S		085 088	085 088		088 - 8.00R MORE	50	1	50	100		
2021-452	013	BAYER CLI		085 088	085 088		087 - 7.5	37	33	37	100		
2021-452	013	BAYER CLI		085 088	085 088		086 - 7.0	37	4	37	100		
2021-452	067	URISCAN		084 088	085 088		085 - 6.5	50	1	50	100		
2021-452	067	URISCAN		084 088	085 088		086 - 7.0	50	34	50	100		

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-452	067	URISCAN		084 088	085 088		087 - 7.5	50	15	50	100	
2021-452	071	ATLAS		085 088	085 088		087 - 7.5	15	15	15	100	
2021-452	072	URS-9-TECO		084 088	085 088		085 - 6.5	100	1	100	100	
2021-452	072	URS-9-TECO		084 088	085 088		086 - 7.0	100	72	100	100	
2021-452	072	URS-9-TECO		084 088	085 088		087 - 7.5	100	27	100	100	
2021-452	080	ROCHE (CHE		084 088	085 088		086 - 7.0	37	36	37	100	
2021-452	080	ROCHE (CHE		084 088	085 088		088 - 8.0OR MORE	37	1	37	100	
2021-452	089	AIMSTRIP		085 088	085 088		086 - 7.0	109	28	109	100	
2021-452	089	AIMSTRIP		085 088	085 088		087 - 7.5	109	81	109	100	
2021-452	094	CLINITEK		085 088	085 088		086 - 7.0	105	10	105	100	
2021-452	094	CLINITEK		085 088	085 088		087 - 7.5	105	95	105	100	
2021-452	102	ACON LAB		085 088	085 088		086 - 7.0	156	71	156	100	
2021-452	102	ACON LAB		085 088	085 088		087 - 7.5	156	85	156	100	
2021-452	131	ICHEM VELO		084 088	085 088		086 - 7.0	27	27	27	100	
2021-452	135	Hybrido AU		085 088	085 088		087 - 7.5	13	13	13	100	
2021-452	138	URS-10T		084 088	085 088		086 - 7.0	10	8	10	100	
2021-452	138	URS-10T		084 088	085 088		087 - 7.5	10	2	10	100	
2021-452	140	MISSION A		085 088	085 088		086 - 7.0	58	19	58	100	
2021-452	140	MISSION A		085 088	085 088		087 - 7.5	58	39	58	100	
2021-452	145	CLINITEK		085 088	085 088		086 - 7.0	12	1	12	100	
2021-452	145	CLINITEK		085 088	085 088		087 - 7.5	12	11	12	100	
2021-452	147	Mission120		085 088	085 088		086 - 7.0	30	10	30	100	
2021-452	147	Mission120		085 088	085 088		087 - 7.5	30	20	30	100	
2021-452	148	Mission500		084 088	085 088		086 - 7.0	31	16	31	100	
2021-452	148	Mission500		084 088	085 088		087 - 7.5	31	15	31	100	
062 URINALYSIS-pH												
2021-453	011	AMES NM10S		086 088	086 088		086 - 7.0	50	1	50	100	
2021-453	011	AMES NM10S		086 088	086 088		087 - 7.5	50	6	50	100	
2021-453	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	50	43	50	100	
2021-453	013	BAYER CLI		086 088	086 088		087 - 7.5	37	4	37	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-453	013	BAYER CLI		086 088	086 088		088 - 8.00R MORE	37	33	37	100		
2021-453	067	URISCAN		086 088	086 088		087 - 7.5	50	13	50	100		
2021-453	067	URISCAN		086 088	086 088		088 - 8.00R MORE	50	37	50	100		
2021-453	071	ATLAS		086 088	086 088		088 - 8.00R MORE	15	15	15	100		
2021-453	072	URS-9-TECO		085 088	086 088		085 - 6.5	100	2	100	100		
2021-453	072	URS-9-TECO		085 088	086 088		086 - 7.0	100	5	100	100		
2021-453	072	URS-9-TECO		085 088	086 088		087 - 7.5	100	59	100	100		
2021-453	072	URS-9-TECO		085 088	086 088		088 - 8.00R MORE	100	34	100	100		
2021-453	080	ROCHE (CHE		086 088	086 088		086 - 7.0	37	2	37	100		
2021-453	080	ROCHE (CHE		086 088	086 088		088 - 8.00R MORE	37	35	37	100		
2021-453	089	AIMSTRIP		086 088	086 088		087 - 7.5	109	6	109	100		
2021-453	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	109	103	109	100		
2021-453	094	CLINITEK		086 088	086 088		087 - 7.5	105	12	105	100		
2021-453	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	105	93	105	100		
2021-453	102	ACON LAB		086 088	086 088		087 - 7.5	156	10	156	100		
2021-453	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	156	146	156	100		
2021-453	131	iCHEM VELO		086 088	086 088		086 - 7.0	27	1	27	100		
2021-453	131	iCHEM VELO		086 088	086 088		088 - 8.00R MORE	27	26	27	100		
2021-453	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	13	13	13	100		
2021-453	138	URS-10T		085 088	086 088		087 - 7.5	10	8	10	100		
2021-453	138	URS-10T		085 088	086 088		088 - 8.00R MORE	10	2	10	100		
2021-453	140	MISSION A		086 088	086 088		087 - 7.5	58	4	58	100		
2021-453	140	MISSION A		086 088	086 088		088 - 8.00R MORE	58	54	58	100		
2021-453	145	CLINITEK		086 088	086 088		088 - 8.00R MORE	12	12	12	100		
2021-453	147	Mission120		086 088	086 088		087 - 7.5	30	1	30	100		
2021-453	147	Mission120		086 088	086 088		088 - 8.00R MORE	30	29	30	100		
2021-453	148	Mission500		086 088	086 088		087 - 7.5	31	1	31	100		
2021-453	148	Mission500		086 088	086 088		088 - 8.00R MORE	31	30	31	100		
062 URINALYSIS-pH													
2021-454	011	AMES NM10S		084 088	084 088		085 - 6.5	50	2	50	100		

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-454	011	AMES NM10S		084 088	084 088		086 - 7.0	50	48	50	100
2021-454	013	BAYER CLI		084 088	084 088		086 - 7.0	37	37	37	100
2021-454	067	URISCAN		083 087	084 088		084 - 6.0	50	2	50	100
2021-454	067	URISCAN		083 087	084 088		085 - 6.5	50	38	50	100
2021-454	067	URISCAN		083 087	084 088		086 - 7.0	50	10	50	100
2021-454	071	ATLAS		084 088	084 088		086 - 7.0	15	15	15	100
2021-454	072	URS-9-TECO		084 088	084 088		084 - 6.0	100	2	100	100
2021-454	072	URS-9-TECO		084 088	084 088		085 - 6.5	100	21	100	100
2021-454	072	URS-9-TECO		084 088	084 088		086 - 7.0	100	77	100	100
2021-454	080	ROCHE (CHE		084 088	084 088		085 - 6.5	37	2	37	100
2021-454	080	ROCHE (CHE		084 088	084 088		086 - 7.0	37	35	37	100
2021-454	089	AIMSTRIP		084 088	084 088		085 - 6.5	109	40	109	100
2021-454	089	AIMSTRIP		084 088	084 088		086 - 7.0	109	69	109	100
2021-454	094	CLINITEK		084 088	084 088		085 - 6.5	105	5	105	100
2021-454	094	CLINITEK		084 088	084 088		086 - 7.0	105	100	105	100
2021-454	102	ACON LAB		084 088	084 088		085 - 6.5	156	48	156	100
2021-454	102	ACON LAB		084 088	084 088		086 - 7.0	156	108	156	100
2021-454	131	ICHEM VELO		084 088	084 088		086 - 7.0	27	27	27	100
2021-454	135	Hybrido AU		084 088	084 088		086 - 7.0	13	13	13	100
2021-454	138	URS-10T		084 088	084 088		085 - 6.5	10	4	10	100
2021-454	138	URS-10T		084 088	084 088		086 - 7.0	10	6	10	100
2021-454	140	MISSION A		083 087	084 088		083 - 5.5	58	1	58	100
2021-454	140	MISSION A		083 087	084 088		085 - 6.5	58	38	58	100
2021-454	140	MISSION A		083 087	084 088		086 - 7.0	58	19	58	100
2021-454	145	CLINITEK		084 088	084 088		082 - 5.0	12	1	11	0
2021-454	145	CLINITEK		084 088	084 088		086 - 7.0	12	11	11	100
2021-454	147	Mission120		083 087	084 088		085 - 6.5	30	29	30	100
2021-454	147	Mission120		083 087	084 088		086 - 7.0	30	1	30	100
2021-454	148	Mission500		084 088	084 088		085 - 6.5	31	2	31	100
2021-454	148	Mission500		084 088	084 088		086 - 7.0	31	29	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
062 URINALYSIS-pH											
2021-455	011	AMES NM10S		080 084	081 085		083 - 5.5	50	18	50	100
2021-455	011	AMES NM10S		080 084	081 085		084 - 6.0	50	6	50	100
2021-455	011	AMES NM10S		080 084	081 085		085 - 6.5	50	2	50	100
2021-455	011	AMES NM10S		080 084	081 085		082 - 5.0	50	24	50	100
2021-455	013	BAYER CLI		081 085	081 085		082 - 5.0	37	14	37	100
2021-455	013	BAYER CLI		081 085	081 085		083 - 5.5	37	17	37	100
2021-455	013	BAYER CLI		081 085	081 085		084 - 6.0	37	6	37	100
2021-455	067	URISCAN		080 084	081 085		082 - 5.0	50	38	50	100
2021-455	067	URISCAN		080 084	081 085		083 - 5.5	50	11	50	100
2021-455	067	URISCAN		080 084	081 085		084 - 6.0	50	1	50	100
2021-455	071	ATLAS		081 085	081 085		082 - 5.0	15	5	15	100
2021-455	071	ATLAS		081 085	081 085		083 - 5.5	15	10	15	100
2021-455	072	URS-9-TECO		082 086	081 085		082 - 5.0	100	16	100	100
2021-455	072	URS-9-TECO		082 086	081 085		083 - 5.5	100	17	100	100
2021-455	072	URS-9-TECO		082 086	081 085		084 - 6.0	100	61	100	100
2021-455	072	URS-9-TECO		082 086	081 085		085 - 6.5	100	6	100	100
2021-455	080	ROCHE (CHE		080 084	081 085		082 - 5.0	37	36	37	100
2021-455	080	ROCHE (CHE		080 084	081 085		084 - 6.0	37	1	37	100
2021-455	089	AIMSTRIP		081 085	081 085		082 - 5.0	109	39	109	100
2021-455	089	AIMSTRIP		081 085	081 085		083 - 5.5	109	40	109	100
2021-455	089	AIMSTRIP		081 085	081 085		084 - 6.0	109	30	109	100
2021-455	094	CLINITEK		081 085	081 085		082 - 5.0	105	24	105	100
2021-455	094	CLINITEK		081 085	081 085		083 - 5.5	105	71	105	100
2021-455	094	CLINITEK		081 085	081 085		084 - 6.0	105	10	105	100
2021-455	102	ACON LAB		082 086	081 085		082 - 5.0	156	37	156	100
2021-455	102	ACON LAB		082 086	081 085		083 - 5.5	156	27	156	100
2021-455	102	ACON LAB		082 086	081 085		084 - 6.0	156	92	156	100
2021-455	131	ICHEM VELO		080 084	081 085		082 - 5.0	27	27	27	100
2021-455	135	Hybrido AU		081 085	081 085		082 - 5.0	13	2	13	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	135	Hybrido AU		081 085	081 085		083 - 5.5	13	11	13	100
2021-455	138	URS-10T		081 085	081 085		083 - 5.5	10	3	10	100
2021-455	138	URS-10T		081 085	081 085		084 - 6.0	10	2	10	100
2021-455	138	URS-10T		081 085	081 085		085 - 6.5	10	2	10	100
2021-455	138	URS-10T		081 085	081 085		082 - 5.0	10	3	10	100
2021-455	140	MISSION A		081 085	081 085		085 - 6.5	58	1	58	100
2021-455	140	MISSION A		081 085	081 085		082 - 5.0	58	8	58	100
2021-455	140	MISSION A		081 085	081 085		083 - 5.5	58	26	58	100
2021-455	140	MISSION A		081 085	081 085		084 - 6.0	58	23	58	100
2021-455	145	CLINITEK		080 084	081 085		082 - 5.0	12	11	11	100
2021-455	145	CLINITEK		080 084	081 085		087 - 7.5	12	1	11	0
2021-455	147	Mission120		081 085	081 085		082 - 5.0	30	4	30	100
2021-455	147	Mission120		081 085	081 085		083 - 5.5	30	25	30	100
2021-455	147	Mission120		081 085	081 085		084 - 6.0	30	1	30	100
2021-455	148	Mission500		082 086	081 085		082 - 5.0	31	1	30	100
2021-455	148	Mission500		082 086	081 085		083 - 5.5	31	1	30	100
2021-455	148	Mission500		082 086	081 085		084 - 6.0	31	28	30	100
2021-455	148	Mission500		082 086	081 085		088 - 8.0OR MORE	31	1	30	0
064 URIN-PROTEIN QUALI.											
2021-451	011	AMES NM10S		054 056	053 055		054 - 2+	51	24	51	100
2021-451	011	AMES NM10S		054 056	053 055		053 - 1+	51	1	51	100
2021-451	011	AMES NM10S		054 056	053 055		055 - 3+	51	26	51	100
2021-451	013	BAYER CLI		053 055	053 055		054 - 2+	36	20	36	100
2021-451	013	BAYER CLI		053 055	053 055		055 - 3+	36	16	36	100
2021-451	067	URISCAN		053 055	053 055		054 - 2+	50	38	50	100
2021-451	067	URISCAN		053 055	053 055		053 - 1+	50	6	50	100
2021-451	067	URISCAN		053 055	053 055		055 - 3+	50	6	50	100
2021-451	071	ATLAS		053 055	053 055		054 - 2+	15	14	15	100
2021-451	071	ATLAS		053 055	053 055		053 - 1+	15	1	15	100
2021-451	072	URS-9-TECO		054 056	053 055		054 - 2+	101	48	101	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-451	072	URS-9-TECO		054 056	053 055		053 - 1+	101	3	101	100
2021-451	072	URS-9-TECO		054 056	053 055		055 - 3+	101	49	101	100
2021-451	072	URS-9-TECO		054 056	053 055		056 - 4+	101	1	101	100
2021-451	080	ROCHE (CHE		053 055	053 055		054 - 2+	37	29	37	100
2021-451	080	ROCHE (CHE		053 055	053 055		053 - 1+	37	1	37	100
2021-451	080	ROCHE (CHE		053 055	053 055		055 - 3+	37	7	37	100
2021-451	089	AIMSTRIP		053 055	053 055		054 - 2+	108	56	102	100
2021-451	089	AIMSTRIP		053 055	053 055		053 - 1+	108	3	102	100
2021-451	089	AIMSTRIP		053 055	053 055		055 - 3+	108	43	102	100
2021-451	089	AIMSTRIP		053 055	053 055		056 - 4+	108	6	102	0
2021-451	094	CLINITEK		053 055	053 055		054 - 2+	105	79	104	100
2021-451	094	CLINITEK		053 055	053 055		053 - 1+	105	3	104	100
2021-451	094	CLINITEK		053 055	053 055		055 - 3+	105	22	104	100
2021-451	094	CLINITEK		053 055	053 055		056 - 4+	105	1	104	0
2021-451	102	ACON LAB		054 056	053 055		054 - 2+	155	71	155	100
2021-451	102	ACON LAB		054 056	053 055		055 - 3+	155	82	155	100
2021-451	102	ACON LAB		054 056	053 055		056 - 4+	155	2	155	100
2021-451	131	ICHEM VELO		053 055	053 055		054 - 2+	27	23	26	100
2021-451	131	ICHEM VELO		053 055	053 055		055 - 3+	27	3	26	100
2021-451	131	ICHEM VELO		053 055	053 055		056 - 4+	27	1	26	0
2021-451	135	Hybrido AU		053 055	053 055		054 - 2+	13	13	13	100
2021-451	138	URS-10T		054 056	053 055		054 - 2+	10	2	10	100
2021-451	138	URS-10T		054 056	053 055		055 - 3+	10	7	10	100
2021-451	138	URS-10T		054 056	053 055		056 - 4+	10	1	10	100
2021-451	140	MISSION A		053 055	053 055		054 - 2+	57	40	57	100
2021-451	140	MISSION A		053 055	053 055		055 - 3+	57	17	57	100
2021-451	145	CLINITEK		053 055	053 055		054 - 2+	12	8	12	100
2021-451	145	CLINITEK		053 055	053 055		055 - 3+	12	4	12	100
2021-451	147	Mission120		053 055	053 055		054 - 2+	31	28	31	100
2021-451	147	Mission120		053 055	053 055		053 - 1+	31	2	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-451	147	Mission120		053 055	053 055		055 - 3+	31	1	31	100	
2021-451	148	Mission500		053 055	053 055		054 - 2+	31	17	31	100	
2021-451	148	Mission500		053 055	053 055		053 - 1+	31	1	31	100	
2021-451	148	Mission500		053 055	053 055		055 - 3+	31	13	31	100	
064 URIN-PROTEIN QUALI.												
2021-452	011	AMES NM10S		053 055	053 055		054 - 2+	51	27	50	100	
2021-452	011	AMES NM10S		053 055	053 055		055 - 3+	51	23	50	100	
2021-452	011	AMES NM10S		053 055	053 055		056 - 4+	51	1	50	0	
2021-452	013	BAYER CLI		053 055	053 055		054 - 2+	36	20	36	100	
2021-452	013	BAYER CLI		053 055	053 055		055 - 3+	36	16	36	100	
2021-452	067	URISCAN		053 055	053 055		054 - 2+	50	41	49	100	
2021-452	067	URISCAN		053 055	053 055		052 - TRACES	50	1	49	0	
2021-452	067	URISCAN		053 055	053 055		053 - 1+	50	7	49	100	
2021-452	067	URISCAN		053 055	053 055		055 - 3+	50	1	49	100	
2021-452	071	ATLAS		053 055	053 055		054 - 2+	15	14	15	100	
2021-452	071	ATLAS		053 055	053 055		053 - 1+	15	1	15	100	
2021-452	072	URS-9-TECO		053 055	053 055		054 - 2+	101	53	101	100	
2021-452	072	URS-9-TECO		053 055	053 055		053 - 1+	101	2	101	100	
2021-452	072	URS-9-TECO		053 055	053 055		055 - 3+	101	46	101	100	
2021-452	080	ROCHE (CHE		053 055	053 055		054 - 2+	37	29	37	100	
2021-452	080	ROCHE (CHE		053 055	053 055		055 - 3+	37	8	37	100	
2021-452	089	AIMSTRIP		053 055	053 055		056 - 4+	108	3	105	0	
2021-452	089	AIMSTRIP		053 055	053 055		054 - 2+	108	64	105	100	
2021-452	089	AIMSTRIP		053 055	053 055		053 - 1+	108	5	105	100	
2021-452	089	AIMSTRIP		053 055	053 055		055 - 3+	108	36	105	100	
2021-452	094	CLINITEK		053 055	053 055		054 - 2+	105	78	103	100	
2021-452	094	CLINITEK		053 055	053 055		053 - 1+	105	3	103	100	
2021-452	094	CLINITEK		053 055	053 055		055 - 3+	105	22	103	100	
2021-452	094	CLINITEK		053 055	053 055		056 - 4+	105	2	103	0	
2021-452	102	ACON LAB		053 055	053 055		054 - 2+	155	83	154	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-452	102	ACON LAB		053 055	053 055		053 - 1+	155	1	154	100
2021-452	102	ACON LAB		053 055	053 055		055 - 3+	155	70	154	100
2021-452	102	ACON LAB		053 055	053 055		056 - 4+	155	1	154	0
2021-452	131	iCHEM VELO		053 055	053 055		056 - 4+	27	2	25	0
2021-452	131	iCHEM VELO		053 055	053 055		054 - 2+	27	20	25	100
2021-452	131	iCHEM VELO		053 055	053 055		055 - 3+	27	5	25	100
2021-452	135	Hybrido AU		053 055	053 055		054 - 2+	13	13	13	100
2021-452	138	URS-10T		054 056	053 055		054 - 2+	10	2	10	100
2021-452	138	URS-10T		054 056	053 055		055 - 3+	10	7	10	100
2021-452	138	URS-10T		054 056	053 055		056 - 4+	10	1	10	100
2021-452	140	MISSION A		053 055	053 055		054 - 2+	57	45	57	100
2021-452	140	MISSION A		053 055	053 055		055 - 3+	57	12	57	100
2021-452	145	CLINITEK		053 055	053 055		054 - 2+	12	9	12	100
2021-452	145	CLINITEK		053 055	053 055		055 - 3+	12	3	12	100
2021-452	147	Mission120		053 055	053 055		054 - 2+	31	28	31	100
2021-452	147	Mission120		053 055	053 055		053 - 1+	31	2	31	100
2021-452	147	Mission120		053 055	053 055		055 - 3+	31	1	31	100
2021-452	148	Mission500		053 055	053 055		054 - 2+	31	18	30	100
2021-452	148	Mission500		053 055	053 055		053 - 1+	31	2	30	100
2021-452	148	Mission500		053 055	053 055		055 - 3+	31	10	30	100
2021-452	148	Mission500		053 055	053 055		056 - 4+	31	1	30	0
064 URIN-PROTEIN QUALI.											
2021-453	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	51	51	51	100
2021-453	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	36	36	36	100
2021-453	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	50	50	100
2021-453	071	ATLAS		051 051	051 051		051 - NEGATIVE	15	15	15	100
2021-453	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	101	100	100	100
2021-453	072	URS-9-TECO		051 051	051 051		052 - TRACES	101	1	100	0
2021-453	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	37	36	36	100
2021-453	080	ROCHE (CHE		051 051	051 051		052 - TRACES	37	1	36	0

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-453	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	108	107	107	100
2021-453	089	AIMSTRIP		051 051	051 051		052 - TRACES	108	1	107	0
2021-453	094	CLINITEK		051 051	051 051		051 - NEGATIVE	105	104	104	100
2021-453	094	CLINITEK		051 051	051 051		052 - TRACES	105	1	104	0
2021-453	102	ACON LAB		051 051	051 051		051 - NEGATIVE	155	146	146	100
2021-453	102	ACON LAB		051 051	051 051		052 - TRACES	155	9	146	0
2021-453	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	27	27	27	100
2021-453	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	13	13	13	100
2021-453	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100
2021-453	140	MISSION A		051 051	051 051		051 - NEGATIVE	57	53	53	100
2021-453	140	MISSION A		051 051	051 051		052 - TRACES	57	4	53	0
2021-453	145	CLINITEK		051 051	051 051		051 - NEGATIVE	12	12	12	100
2021-453	147	Mission120		051 051	051 051		051 - NEGATIVE	31	31	31	100
2021-453	148	Mission500		051 051	051 051		051 - NEGATIVE	31	28	28	100
2021-453	148	Mission500		051 051	051 051		052 - TRACES	31	3	28	0
064 URIN-PROTEIN QUALI.											
2021-454	011	AMES NM10S		053 055	053 055		054 - 2+	51	29	51	100
2021-454	011	AMES NM10S		053 055	053 055		053 - 1+	51	2	51	100
2021-454	011	AMES NM10S		053 055	053 055		055 - 3+	51	20	51	100
2021-454	013	BAYER CLI		053 055	053 055		054 - 2+	36	23	36	100
2021-454	013	BAYER CLI		053 055	053 055		053 - 1+	36	2	36	100
2021-454	013	BAYER CLI		053 055	053 055		055 - 3+	36	11	36	100
2021-454	067	URISCAN		053 055	053 055		054 - 2+	50	35	49	100
2021-454	067	URISCAN		053 055	053 055		052 - TRACES	50	1	49	0
2021-454	067	URISCAN		053 055	053 055		053 - 1+	50	14	49	100
2021-454	071	ATLAS		053 055	053 055		054 - 2+	15	14	15	100
2021-454	071	ATLAS		053 055	053 055		053 - 1+	15	1	15	100
2021-454	072	URS-9-TECO		053 055	053 055		054 - 2+	101	62	101	100
2021-454	072	URS-9-TECO		053 055	053 055		053 - 1+	101	7	101	100
2021-454	072	URS-9-TECO		053 055	053 055		055 - 3+	101	32	101	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-454	080	ROCHE (CHE		053	055	053	055	054 - 2+	37	31	37	100
2021-454	080	ROCHE (CHE		053	055	053	055	055 - 3+	37	6	37	100
2021-454	089	AIMSTRIP		053	055	053	055	054 - 2+	108	83	105	100
2021-454	089	AIMSTRIP		053	055	053	055	053 - 1+	108	6	105	100
2021-454	089	AIMSTRIP		053	055	053	055	055 - 3+	108	16	105	100
2021-454	089	AIMSTRIP		053	055	053	055	056 - 4+	108	3	105	0
2021-454	094	CLINITEK		053	055	053	055	054 - 2+	105	90	105	100
2021-454	094	CLINITEK		053	055	053	055	053 - 1+	105	4	105	100
2021-454	094	CLINITEK		053	055	053	055	055 - 3+	105	11	105	100
2021-454	102	ACON LAB		053	055	053	055	054 - 2+	155	121	155	100
2021-454	102	ACON LAB		053	055	053	055	053 - 1+	155	6	155	100
2021-454	102	ACON LAB		053	055	053	055	055 - 3+	155	28	155	100
2021-454	131	iCHEM VELO		053	055	053	055	054 - 2+	27	26	27	100
2021-454	131	iCHEM VELO		053	055	053	055	055 - 3+	27	1	27	100
2021-454	135	Hybrido AU		053	055	053	055	054 - 2+	13	12	13	100
2021-454	135	Hybrido AU		053	055	053	055	053 - 1+	13	1	13	100
2021-454	138	URS-10T		053	055	053	055	054 - 2+	10	6	10	100
2021-454	138	URS-10T		053	055	053	055	055 - 3+	10	4	10	100
2021-454	140	MISSION A		053	055	053	055	054 - 2+	57	49	56	100
2021-454	140	MISSION A		053	055	053	055	051 - NEGATIVE	57	1	56	0
2021-454	140	MISSION A		053	055	053	055	053 - 1+	57	2	56	100
2021-454	140	MISSION A		053	055	053	055	055 - 3+	57	5	56	100
2021-454	145	CLINITEK		053	055	053	055	054 - 2+	12	12	12	100
2021-454	147	Mission120		053	055	053	055	054 - 2+	31	28	31	100
2021-454	147	Mission120		053	055	053	055	053 - 1+	31	3	31	100
2021-454	148	Mission500		053	055	053	055	054 - 2+	31	25	31	100
2021-454	148	Mission500		053	055	053	055	053 - 1+	31	2	31	100
2021-454	148	Mission500		053	055	053	055	055 - 3+	31	4	31	100
064 URIN-PROTEIN QUALI.												
2021-455	011	AMES NM10S		051	051	051	051	051 - NEGATIVE	51	51	51	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	36	36	36	100
2021-455	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	50	50	100
2021-455	071	ATLAS		051 051	051 051		051 - NEGATIVE	15	15	15	100
2021-455	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	101	101	101	100
2021-455	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	37	37	37	100
2021-455	089	AIMSTRIP		051 051	051 051		054 - 2+	108	1	107	0
2021-455	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	108	107	107	100
2021-455	094	CLINITEK		051 051	051 051		051 - NEGATIVE	105	105	105	100
2021-455	102	ACON LAB		051 051	051 051		051 - NEGATIVE	155	155	155	100
2021-455	131	ICHEM VELO		051 051	051 051		051 - NEGATIVE	27	27	27	100
2021-455	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	13	13	13	100
2021-455	138	URS-10T		051 051	051 051		051 - NEGATIVE	10	10	10	100
2021-455	140	MISSION A		051 051	051 051		054 - 2+	57	1	56	0
2021-455	140	MISSION A		051 051	051 051		051 - NEGATIVE	57	56	56	100
2021-455	145	CLINITEK		051 051	051 051		051 - NEGATIVE	12	12	12	100
2021-455	147	Mission120		051 051	051 051		051 - NEGATIVE	31	31	31	100
2021-455	148	Mission500		051 051	051 051		051 - NEGATIVE	31	31	31	100
066 URINAL.-UROBILINOGEN											
2021-451	011	AMES NM10S		061 063	061 063		061 - NORMAL,NEG	51	6	51	100
2021-451	011	AMES NM10S		061 063	061 063		062 - 2	51	26	51	100
2021-451	011	AMES NM10S		061 063	061 063		063 - 4	51	19	51	100
2021-451	013	BAYER CLI		061 063	061 063		061 - NORMAL,NEG	36	14	36	100
2021-451	013	BAYER CLI		061 063	061 063		062 - 2	36	19	36	100
2021-451	013	BAYER CLI		061 063	061 063		063 - 4	36	3	36	100
2021-451	067	URISCAN			061 063		061 - NORMAL,NEG	50	21	50	100
2021-451	067	URISCAN			061 063		062 - 2	50	10	50	100
2021-451	067	URISCAN			061 063		063 - 4	50	19	50	100
2021-451	071	ATLAS		061 063	061 063		061 - NORMAL,NEG	16	1	16	100
2021-451	071	ATLAS		061 063	061 063		062 - 2	16	8	16	100
2021-451	071	ATLAS		061 063	061 063		063 - 4	16	7	16	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-451	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	101	29	101	100	
2021-451	072	URS-9-TECO		061 063	061 063		062 - 2	101	62	101	100	
2021-451	072	URS-9-TECO		061 063	061 063		063 - 4	101	10	101	100	
2021-451	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	37	34	37	100	
2021-451	080	ROCHE (CHE		061 061	061 063		062 - 2	37	3	37	100	
2021-451	089	AIMSTRIP			061 063		061 - NORMAL,NEG	108	77	108	100	
2021-451	089	AIMSTRIP			061 063		062 - 2	108	27	108	100	
2021-451	089	AIMSTRIP			061 063		063 - 4	108	4	108	100	
2021-451	094	CLINITEK		061 063	061 063		061 - NORMAL,NEG	104	19	104	100	
2021-451	094	CLINITEK		061 063	061 063		062 - 2	104	54	104	100	
2021-451	094	CLINITEK		061 063	061 063		063 - 4	104	31	104	100	
2021-451	102	ACON LAB			061 063		061 - NORMAL,NEG	154	123	154	100	
2021-451	102	ACON LAB			061 063		062 - 2	154	31	154	100	
2021-451	131	ICHEM VELO		062 064	061 063		062 - 2	27	5	27	100	
2021-451	131	ICHEM VELO		062 064	061 063		063 - 4	27	22	27	100	
2021-451	135	Hybrido AU		061 063	061 063		061 - NORMAL,NEG	13	3	13	100	
2021-451	135	Hybrido AU		061 063	061 063		062 - 2	13	10	13	100	
2021-451	138	URS-10T		061 063	061 063		061 - NORMAL,NEG	10	5	10	100	
2021-451	138	URS-10T		061 063	061 063		062 - 2	10	5	10	100	
2021-451	140	MISSION A			061 063		061 - NORMAL,NEG	57	44	57	100	
2021-451	140	MISSION A			061 063		062 - 2	57	11	57	100	
2021-451	140	MISSION A			061 063		063 - 4	57	2	57	100	
2021-451	145	CLINITEK		061 063	061 063		062 - 2	12	12	12	100	
2021-451	147	Mission120		061 061	061 063		061 - NORMAL,NEG	31	27	31	100	
2021-451	147	Mission120		061 061	061 063		062 - 2	31	4	31	100	
2021-451	148	Mission500		061 061	061 063		061 - NORMAL,NEG	31	29	31	100	
2021-451	148	Mission500		061 061	061 063		062 - 2	31	1	31	100	
2021-451	148	Mission500		061 061	061 063		063 - 4	31	1	31	100	
066 URINAL.-UROBILINOGEN												
2021-452	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-452	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2021-452	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2021-452	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2021-452	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	101	101	101	100	
2021-452	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2021-452	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	108	108	108	100	
2021-452	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	103	103	100	
2021-452	094	CLINITEK		061 061	061 061		062 - 2	104	1	103	0	
2021-452	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	154	154	154	100	
2021-452	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2021-452	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2021-452	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2021-452	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2021-452	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2021-452	147	Mission120		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
2021-452	148	Mission500		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
066 URINAL.-UROBILINOGEN												
2021-453	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2021-453	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2021-453	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2021-453	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2021-453	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	101	101	101	100	
2021-453	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2021-453	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	108	108	108	100	
2021-453	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	103	103	100	
2021-453	094	CLINITEK		061 061	061 061		062 - 2	104	1	103	0	
2021-453	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	154	154	154	100	
2021-453	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2021-453	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2021-453	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-453	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2021-453	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2021-453	147	Mission120		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
2021-453	148	Mission500		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
066 URINAL.-UROBILINOGEN												
2021-454	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2021-454	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2021-454	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2021-454	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2021-454	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	101	101	101	100	
2021-454	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2021-454	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	108	108	108	100	
2021-454	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	103	103	100	
2021-454	094	CLINITEK		061 061	061 061		062 - 2	104	1	103	0	
2021-454	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	154	153	153	100	
2021-454	102	ACON LAB		061 061	061 061		062 - 2	154	1	153	0	
2021-454	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2021-454	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2021-454	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2021-454	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	57	57	57	100	
2021-454	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2021-454	147	Mission120		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
2021-454	148	Mission500		061 061	061 061		061 - NORMAL,NEG	31	31	31	100	
2021-455	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2021-455	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2021-455	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2021-455	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2021-455	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	101	101	101	100	

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100
2021-455	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	108	108	108	100
2021-455	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	103	103	100
2021-455	094	CLINITEK		061 061	061 061		062 - 2	104	1	103	0
2021-455	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	154	154	154	100
2021-455	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100
2021-455	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100
2021-455	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	10	10	10	100
2021-455	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	57	57	57	100
2021-455	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	12	12	12	100
2021-455	147	Mission120		061 061	061 061		061 - NORMAL,NEG	31	31	31	100
2021-455	148	Mission500		061 061	061 061		061 - NORMAL,NEG	31	31	31	100
184 URINALYSIS LEUKOCYTE											
2021-451				001 001	001 001		001 - Negative	13	13	13	100
2021-451	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100
2021-451	013	BAYER CLI		001 001	001 001		001 - Negative	35	35	35	100
2021-451	067	URISCAN		001 001	001 001		001 - Negative	50	50	50	100
2021-451	071	ATLAS		001 001	001 001		001 - Negative	15	15	15	100
2021-451	072	URS-9-TECO		001 001	001 001		001 - Negative	93	93	93	100
2021-451	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100
2021-451	089	AIMSTRIP		001 001	001 001		001 - Negative	106	106	106	100
2021-451	094	CLINITEK		001 001	001 001		001 - Negative	104	104	104	100
2021-451	102	ACON LAB		001 001	001 001		001 - Negative	139	139	139	100
2021-451	131	iCHEM VELO		001 001	001 001		001 - Negative	28	28	28	100
2021-451	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100
2021-451	138	URS-10T		001 001	001 001		001 - Negative	10	10	10	100
2021-451	140	MISSION A		001 001	001 001		001 - Negative	65	65	65	100
2021-451	145	CLINITEK		001 001	001 001		001 - Negative	12	12	12	100
2021-451	147	Mission120		001 001	001 001		001 - Negative	30	30	30	100
2021-451	148	Mission500		001 001	001 001		001 - Negative	31	31	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
184 URINALYSIS LEUKOCYTE											
2021-452				001 001	001 001		001 - Negative	13	13	13	100
2021-452	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100
2021-452	013	BAYER CLI		001 001	001 001		001 - Negative	35	35	35	100
2021-452	067	URISCAN		001 001	001 001		001 - Negative	50	50	50	100
2021-452	071	ATLAS		001 001	001 001		001 - Negative	15	15	15	100
2021-452	072	URS-9-TECO		001 001	001 001		001 - Negative	93	93	93	100
2021-452	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100
2021-452	089	AIMSTRIP		001 001	001 001		001 - Negative	106	106	106	100
2021-452	094	CLINITEK		001 001	001 001		001 - Negative	104	104	104	100
2021-452	102	ACON LAB		001 001	001 001		001 - Negative	139	139	139	100
2021-452	131	iCHEM VELO		001 001	001 001		001 - Negative	28	28	28	100
2021-452	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100
2021-452	138	URS-10T		001 001	001 001		001 - Negative	10	10	10	100
2021-452	140	MISSION A		001 001	001 001		001 - Negative	65	65	65	100
2021-452	145	CLINITEK		001 001	001 001		001 - Negative	12	12	12	100
2021-452	147	Mission120		001 001	001 001		001 - Negative	30	30	30	100
2021-452	148	Mission500		001 001	001 001		001 - Negative	31	31	31	100
2021-453				001 001	001 001		001 - Negative	13	13	13	100
2021-453	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100
2021-453	013	BAYER CLI		001 001	001 001		001 - Negative	35	35	35	100
2021-453	067	URISCAN		001 001	001 001		001 - Negative	50	50	50	100
2021-453	071	ATLAS		001 001	001 001		001 - Negative	15	15	15	100
2021-453	072	URS-9-TECO		001 001	001 001		001 - Negative	93	92	92	100
2021-453	072	URS-9-TECO		001 001	001 001		002 - Traces	93	1	92	0
2021-453	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100
2021-453	089	AIMSTRIP		001 001	001 001		001 - Negative	106	106	106	100
2021-453	094	CLINITEK		001 001	001 001		001 - Negative	104	104	104	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-453	102	ACON LAB		001 001	001 001		001 - Negative	139	139	139	100
2021-453	131	iCHEM VELO		001 001	001 001		001 - Negative	28	28	28	100
2021-453	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100
2021-453	138	URS-10T		001 001	001 001		001 - Negative	10	10	10	100
2021-453	140	MISSION A		001 001	001 001		001 - Negative	65	64	64	100
2021-453	140	MISSION A		001 001	001 001		003 - Small	65	1	64	0
2021-453	145	CLINITEK		001 001	001 001		001 - Negative	12	12	12	100
2021-453	147	Mission120		001 001	001 001		001 - Negative	30	30	30	100
2021-453	148	Mission500		001 001	001 001		001 - Negative	31	31	31	100
184 URINALYSIS LEUKOCYTE											
2021-454				001 001	001 001		001 - Negative	13	13	13	100
2021-454	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100
2021-454	013	BAYER CLI		001 001	001 001		001 - Negative	35	35	35	100
2021-454	067	URISCAN		001 001	001 001		001 - Negative	50	50	50	100
2021-454	071	ATLAS		001 001	001 001		001 - Negative	15	15	15	100
2021-454	072	URS-9-TECO		001 001	001 001		001 - Negative	93	93	93	100
2021-454	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100
2021-454	089	AIMSTRIP		001 001	001 001		001 - Negative	106	106	106	100
2021-454	094	CLINITEK		001 001	001 001		001 - Negative	104	104	104	100
2021-454	102	ACON LAB		001 001	001 001		001 - Negative	139	139	139	100
2021-454	131	iCHEM VELO		001 001	001 001		001 - Negative	28	28	28	100
2021-454	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100
2021-454	138	URS-10T		001 001	001 001		001 - Negative	10	10	10	100
2021-454	140	MISSION A		001 001	001 001		001 - Negative	65	64	64	100
2021-454	140	MISSION A		001 001	001 001		004 - Moderate	65	1	64	0
2021-454	145	CLINITEK		001 001	001 001		001 - Negative	12	12	12	100
2021-454	147	Mission120		001 001	001 001		001 - Negative	30	30	30	100
2021-454	148	Mission500		001 001	001 001		001 - Negative	31	31	31	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
184 URINALYSIS LEUKOCYTE											
2021-455				002 004	003 005		003 - Small	13	8	13	100
2021-455				002 004	003 005		004 - Moderate	13	4	13	100
2021-455				002 004	003 005		005 - Large	13	1	13	100
2021-455	011	AMES NM10S		002 004	003 005		002 - Traces	47	2	47	100
2021-455	011	AMES NM10S		002 004	003 005		003 - Small	47	39	47	100
2021-455	011	AMES NM10S		002 004	003 005		004 - Moderate	47	5	47	100
2021-455	011	AMES NM10S		002 004	003 005		005 - Large	47	1	47	100
2021-455	013	BAYER CLI		002 004	003 005		003 - Small	35	25	35	100
2021-455	013	BAYER CLI		002 004	003 005		004 - Moderate	35	10	35	100
2021-455	067	URISCAN		004 005	003 005		001 - Negative	50	1	49	0
2021-455	067	URISCAN		004 005	003 005		003 - Small	50	5	49	100
2021-455	067	URISCAN		004 005	003 005		004 - Moderate	50	18	49	100
2021-455	067	URISCAN		004 005	003 005		005 - Large	50	26	49	100
2021-455	071	ATLAS		003 005	003 005		004 - Moderate	15	12	15	100
2021-455	071	ATLAS		003 005	003 005		005 - Large	15	3	15	100
2021-455	072	URS-9-TECO		003 005	003 005		003 - Small	93	46	93	100
2021-455	072	URS-9-TECO		003 005	003 005		004 - Moderate	93	46	93	100
2021-455	072	URS-9-TECO		003 005	003 005		005 - Large	93	1	93	100
2021-455	080	ROCHE (CHE		004 005	003 005		003 - Small	34	2	34	100
2021-455	080	ROCHE (CHE		004 005	003 005		004 - Moderate	34	5	34	100
2021-455	080	ROCHE (CHE		004 005	003 005		005 - Large	34	27	34	100
2021-455	089	AIMSTRIP		002 004	003 005		003 - Small	106	69	106	100
2021-455	089	AIMSTRIP		002 004	003 005		004 - Moderate	106	35	106	100
2021-455	089	AIMSTRIP		002 004	003 005		005 - Large	106	2	106	100
2021-455	094	CLINITEK		002 004	003 005		001 - Negative	104	1	103	0
2021-455	094	CLINITEK		002 004	003 005		002 - Traces	104	1	103	100
2021-455	094	CLINITEK		002 004	003 005		003 - Small	104	90	103	100
2021-455	094	CLINITEK		002 004	003 005		004 - Moderate	104	12	103	100
2021-455	102	ACON LAB		002 004	003 005		003 - Small	139	86	139	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-455	102	ACON LAB		002 004	003 005		004 - Moderate	139	52	139	100
2021-455	102	ACON LAB		002 004	003 005		005 - Large	139	1	139	100
2021-455	131	iCHEM VELO		004 005	003 005		003 - Small	28	1	28	100
2021-455	131	iCHEM VELO		004 005	003 005		005 - Large	28	27	28	100
2021-455	135	Hybrido AU		004 005	003 005		005 - Large	13	13	13	100
2021-455	138	URS-10T		003 005	003 005		003 - Small	10	5	10	100
2021-455	138	URS-10T		003 005	003 005		004 - Moderate	10	5	10	100
2021-455	140	MISSION A		002 004	003 005		002 - Traces	65	7	64	100
2021-455	140	MISSION A		002 004	003 005		003 - Small	65	35	64	100
2021-455	140	MISSION A		002 004	003 005		004 - Moderate	65	18	64	100
2021-455	140	MISSION A		002 004	003 005		005 - Large	65	4	64	100
2021-455	140	MISSION A		002 004	003 005		001 - Negative	65	1	64	0
2021-455	145	CLINITEK		003 005	003 005		004 - Moderate	12	12	12	100
2021-455	147	Mission120		002 004	003 005		002 - Traces	30	4	30	100
2021-455	147	Mission120		002 004	003 005		003 - Small	30	25	30	100
2021-455	147	Mission120		002 004	003 005		005 - Large	30	1	30	100
2021-455	148	Mission500		002 004	003 005		002 - Traces	31	1	31	100
2021-455	148	Mission500		002 004	003 005		003 - Small	31	15	31	100
2021-455	148	Mission500		002 004	003 005		004 - Moderate	31	14	31	100
2021-455	148	Mission500		002 004	003 005		005 - Large	31	1	31	100

Participant Summary
MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2021-451	091	AIM STEP		002	002	002	002	002 - POSITIVE	35	35	35	100
2021-451	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2021-452	091	AIM STEP		002	002	002	002	002 - POSITIVE	35	35	35	100
2021-452	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2021-453	091	AIM STEP		001	001	001	001	001 - NEGATIVE	35	35	35	100
2021-453	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100
2021-454	091	AIM STEP		001	001	001	001	001 - NEGATIVE	35	35	35	100
2021-454	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100
2021-455	091	AIM STEP		001	001	001	001	001 - NEGATIVE	35	35	35	100
2021-455	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2021-461					014	014	014 - CA. OXALAT	866	865	865	100
2021-461					014	014	015 - CA. SULFAT	866	1	865	0
2021-462					041	041	041 - T. vagina.	866	863	863	100
2021-462					041	041	043 - Giardia la	866	3	863	0
2021-463					007	007	001 - E.C. CAST	866	1	745	0
2021-463					007	007	002 - FATTY CAST	866	97	745	0
2021-463					007	007	003 - GRANULAR C	866	1	745	0
2021-463					007	007	005 - RED CELL C	866	1	745	0
2021-463					007	007	006 - WAXY CAST	866	1	745	0
2021-463					007	007	007 - WHITE CELL	866	745	745	100
2021-463					007	007	042 - WBC in clu	866	20	745	0
2021-464					008	008	005 - RED CELL C	867	3	863	0
2021-464					008	008	008 - ERYTHROCYT	867	863	863	100
2021-464					008	008	009 - LEUKOCYTE	867	1	863	0
2021-465					011	011	001 - E.C. CAST	866	3	854	0
2021-465					011	011	011 - SQUAMOUS E	866	854	854	100
2021-465					011	011	012 - TRANSITIO.	866	3	854	0
2021-465					011	011	029 - BACTERIA	866	1	854	0
2021-465					011	011	038 - YEAST/FUN.	866	5	854	0

Participant Summary

MONTH 5 OF YEAR 2021

SPE: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 98 OCCULT BLOOD						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
154 FECAL OCCULT BLOOD												
2021-471	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	441	440	440	100
2021-471	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	441	1	440	0
2021-471	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	692	692	692	100
2021-472	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	441	1	440	0
2021-472	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	441	440	440	100
2021-472	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	692	692	692	100
2021-473	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	441	441	441	100
2021-473	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	692	692	692	100
2021-474	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	441	441	441	100
2021-474	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	692	692	692	100
2021-475	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	441	440	440	100
2021-475	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	441	1	440	0
2021-475	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	692	691	691	100
2021-475	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	692	1	691	0

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 90 SPECIAL TEST					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. LABS	PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}								
2021-421	29 VITROS			13		8.577	0.299	3.49	7.700	9.500	
2021-421	38 ROCHE E			21		8.267	0.439	5.31	6.900	9.600	
2021-421	44 BECK.COUL.			15		9.040	0.460	5.09	7.700	10.400	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-421	50 ARC		11	9.609	0.571	5.94	7.900	11.300
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}					
2021-422	29 VITROS		13	20.000	0.605	3.03	18.200	21.800
2021-422	38 ROCHE E		21	18.000	0.970	5.39	15.100	20.900
2021-422	44 BECK.COUL.		15	21.193	1.191	5.62	17.600	24.800
2021-422	50 ARC		11	22.000	0.716	3.25	19.900	24.100
2021-423	29 VITROS		13	32.969	0.811	2.46	30.500	35.400
2021-423	38 ROCHE E		21	28.348	1.424	5.03	24.100	32.600
2021-423	44 BECK.COUL.		15	34.540	1.511	4.37	30.000	39.100
2021-423	50 ARC		11	35.355	1.972	5.58	29.400	41.300
2021-424	29 VITROS		13	0.431	0.046	10.71	0.300	0.500
2021-424	38 ROCHE E		21	0.671	0.139	20.63	0.500	0.800
2021-424	44 BECK.COUL.		15	0.567	0.047	8.32	0.400	0.700
2021-424	50 ARC		11	1.009	0.525	52.00	0.800	1.300
2021-425	29 VITROS		13	23.262	0.722	3.10	21.100	25.400
2021-425	38 ROCHE E		21	20.633	0.926	4.49	17.900	23.400
2021-425	44 BECK.COUL.		15	24.353	1.253	5.14	20.600	28.100
2021-425	50 ARC		11	25.200	0.765	3.04	22.900	27.500
069	PROS. SPEC-ANTIG-PSA							
2021-421	29 VITROS		40	3.267	0.135	4.12	2.900	3.700
2021-421	36 TOSOH AIA		13	4.015	0.308	7.68	3.100	4.900
2021-421	38 ROCHE E		48	5.679	0.444	7.82	4.300	7.000
2021-421	39 DIMENS.RXL		46	4.715	0.282	5.98	3.900	5.600
2021-421	44 BECK.COUL.		50	5.372	0.277	5.16	4.500	6.200
2021-421	50 ARC		19	4.147	0.360	8.69	3.100	5.200
2021-421	57 S. EXL 200		22	4.668	0.305	6.54	3.800	5.600

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}				
2021-422	29	VITROS		40	7.273	0.344	4.73	6.200	8.300	
2021-422	36	TOSOH AIA		13	9.308	0.692	7.44	7.200	11.400	
2021-422	38	ROCHE E		48	13.187	1.066	8.09	10.000	16.400	
2021-422	39	DIMENS.RXL		46	11.235	0.633	5.63	9.300	13.100	
2021-422	44	BECK.COUL.		50	12.756	0.776	6.08	10.400	15.100	
2021-422	50	ARC		19	10.058	0.748	7.44	7.800	12.300	
2021-422	57	S. EXL 200		22	10.995	0.742	6.75	8.800	13.200	
2021-423	29	VITROS		40	11.673	0.526	4.51	10.100	13.300	
2021-423	36	TOSOH AIA		13	14.877	1.104	7.42	11.600	18.200	
2021-423	38	ROCHE E		48	20.926	1.816	8.68	15.700	26.200	
2021-423	39	DIMENS.RXL		46	18.062	1.034	5.72	15.000	21.200	
2021-423	44	BECK.COUL.		50	20.620	1.191	5.78	17.000	24.200	
2021-423	50	ARC		19	16.284	1.372	8.43	12.200	20.400	
2021-423	57	S. EXL 200		22	18.150	1.090	6.01	14.900	21.400	
2021-424	29	VITROS		40	0.100	0.000	0.00	0.100	0.100	
2021-424	36	TOSOH AIA		13	0.058	0.049	84.52	0.000	0.100	
2021-424	38	ROCHE E		48	0.027	0.044	164.08	0.000	0.000	
2021-424	39	DIMENS.RXL		46	0.074	0.044	59.41	0.100	0.100	
2021-424	44	BECK.COUL.		50	0.008	0.027	335.41	0.000	0.000	
2021-424	50	ARC		19	0.000	0.000	0.00	0.000	0.000	
2021-424	57	S. EXL 200		22	0.050	0.050	100.00	0.000	0.100	
2021-425	29	VITROS		40	8.348	0.409	4.90	7.100	9.600	
2021-425	36	TOSOH AIA		13	10.608	0.802	7.56	8.200	13.000	
2021-425	38	ROCHE E		48	15.017	1.501	10.00	11.300	18.800	
2021-425	39	DIMENS.RXL		46	12.985	0.804	6.19	10.600	15.400	
2021-425	44	BECK.COUL.		50	14.612	0.838	5.73	12.100	17.100	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-425	50	ARC			19	11.750	0.658	5.60	9.800	13.700
2021-425	57	S. EXL 200			22	12.855	0.752	5.85	10.600	15.100
071 VITAMIN B 12			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2021-421	29	VITROS			25	388.440	20.453	5.27	327.000	450.000
2021-421	38	ROCHE E			28	376.643	61.203	16.25	282.000	471.000
2021-421	44	BECK.COUL.			25	299.760	15.114	5.04	254.000	345.000
2021-421	50	ARC			19	376.316	26.390	7.01	297.000	455.000
2021-421	57	S. EXL 200			13	421.846	31.142	7.38	328.000	515.000
2021-422	29	VITROS			25	674.760	52.775	7.82	516.000	833.000
2021-422	38	ROCHE E			28	676.893	54.807	8.10	512.000	841.000
2021-422	44	BECK.COUL.			25	499.560	32.126	6.43	403.000	596.000
2021-422	50	ARC			19	639.789	45.285	7.08	504.000	776.000
2021-422	57	S. EXL 200			13	663.154	32.870	4.96	565.000	762.000
2021-423	29	VITROS			25	781.520	82.217	10.52	586.000	977.000
2021-423	38	ROCHE E			28	998.036	73.644	7.38	777.000	1219.000
2021-423	44	BECK.COUL.			25	680.920	69.710	10.24	511.000	851.000
2021-423	50	ARC			19	925.889	51.461	5.56	772.000	1080.000
2021-423	57	S. EXL 200			13	925.077	46.539	5.03	785.000	1065.000
2021-424	29	VITROS			25	228.208	8.216	3.60	204.000	253.000
2021-424	38	ROCHE E			28	199.107	43.154	21.67	149.000	249.000
2021-424	44	BECK.COUL.			25	162.560	11.737	7.22	127.000	198.000
2021-424	50	ARC			19	203.421	26.660	13.11	153.000	254.000
2021-424	57	S. EXL 200			13	250.615	30.032	11.98	188.000	313.000
2021-425	29	VITROS			25	719.920	61.064	8.48	540.000	900.000
2021-425	38	ROCHE E			28	763.786	57.857	7.58	590.000	937.000
2021-425	44	BECK.COUL.			25	558.880	41.935	7.50	433.000	685.000

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 90 SPECIAL TEST				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2021-425	50	ARC		19	689.737	49.471	7.17	541.000	838.000	
2021-425	57	S. EXL 200		13	729.846	39.156	5.36	612.000	847.000	

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 95 GLYCOHEMOGLOBINE				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}							
2021-431	2	DCA-2000		11	5.236	0.342	6.53	4.200	6.300	
2021-431	12	COBAS/INT.		29	4.700	0.486	10.34	3.500	5.900	
2021-431	14	IMMUNOTUR.		10	4.600	0.605	13.15	3.500	5.800	
2021-431	22	VITROS		43	4.645	0.124	2.67	4.300	5.000	
2021-431	30	DAYTONA		14	5.607	0.489	8.72	4.200	7.000	
2021-431	31	IMMUNOTURB		30	4.411	0.238	5.40	3.700	5.100	
2021-431	40	SIEMENS DI		79	4.496	0.251	5.59	3.700	5.300	
2021-431	46	ARCHITECT		16	4.569	0.320	7.00	3.600	5.500	
2021-432	2	DCA-2000		11	3.000	0.000	0.00	3.000	3.000	
2021-432	12	COBAS/INT.		29	3.000	0.000	0.00	3.000	3.000	
2021-432	14	IMMUNOTUR.		10	3.000	0.000	0.00	3.000	3.000	
2021-432	22	VITROS		43	3.000	0.000	0.00	3.000	3.000	
2021-432	30	DAYTONA		14	3.000	0.000	0.00	3.000	3.000	
2021-432	31	IMMUNOTURB		30	3.000	0.000	0.00	3.000	3.000	
2021-432	40	SIEMENS DI		79	3.000	0.000	0.00	3.000	3.000	
2021-432	46	ARCHITECT		16	3.000	0.000	0.00	3.000	3.000	
2021-433	2	DCA-2000		11	5.255	0.306	5.82	4.300	6.200	
2021-433	12	COBAS/INT.		29	4.712	0.493	10.47	3.500	5.900	
2021-433	14	IMMUNOTUR.		10	4.490	0.444	9.88	3.400	5.600	
2021-433	22	VITROS		43	4.667	0.128	2.75	4.300	5.100	

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2021-433	30	DAYTONA		14	5.486	0.331	6.04	4.500	6.500		
2021-433	31	IMMUNOTURB		30	4.410	0.459	10.40	3.300	5.500		
2021-433	40	SIEMENS DI		79	4.501	0.235	5.23	3.800	5.200		
2021-433	46	ARCHITECT		16	4.638	0.367	7.92	3.500	5.700		
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}								
2021-434	2	DCA-2000		11	9.109	0.198	2.17	8.500	9.700		
2021-434	12	COBAS/INT.		29	8.664	0.420	4.85	7.400	9.900		
2021-434	14	IMMUNOTUR.		10	8.870	0.610	6.88	7.000	10.700		
2021-434	22	VITROS		43	8.726	0.227	2.61	8.000	9.400		
2021-434	30	DAYTONA		14	9.671	0.416	4.31	8.400	10.900		
2021-434	31	IMMUNOTURB		30	8.360	0.214	2.56	7.700	9.000		
2021-434	40	SIEMENS DI		79	8.852	0.279	3.15	8.000	9.700		
2021-434	46	ARCHITECT		16	8.607	0.430	4.99	7.300	9.900		
2021-435	2	DCA-2000		11	3.000	0.000	0.00	3.000	3.000		
2021-435	12	COBAS/INT.		29	3.000	0.000	0.00	3.000	3.000		
2021-435	14	IMMUNOTUR.		10	3.000	0.000	0.00	3.000	3.000		
2021-435	22	VITROS		43	3.000	0.000	0.00	3.000	3.000		
2021-435	30	DAYTONA		14	3.000	0.000	0.00	3.000	3.000		
2021-435	31	IMMUNOTURB		30	3.000	0.000	0.00	3.000	3.000		
2021-435	40	SIEMENS DI		79	3.000	0.000	0.00	3.000	3.000		
2021-435	46	ARCHITECT		16	3.000	0.000	0.00	3.000	3.000		

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 99 Vitamin D							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
178	Vitamin D		EVALUATION CRITERIA: TARGET VALUE 20 %								
2021-281	3	VITROS		38	79.081	5.543	7.01	63.000	95.000		
2021-281	5	ROCHE E		33	51.030	4.896	9.59	41.000	61.000		

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-281	9	BECK.COUL.		43	83.405	6.623	7.94	67.000	100.000
2021-281	9	BECK.COUL.		43	101.750	12.153	11.94	81.000	122.000
2021-281	11	ARC		18	72.833	3.609	4.96	58.000	87.000
2021-281	25	S. DIME		21	55.000	3.795	6.90	44.000	66.000
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %						
2021-282	3	VITROS		38	58.081	6.470	11.14	46.000	70.000
2021-282	5	ROCHE E		33	40.906	2.983	7.29	33.000	49.000
2021-282	9	BECK.COUL.		43	80.250	10.825	13.49	64.000	96.000
2021-282	9	BECK.COUL.		43	67.714	5.161	7.62	54.000	81.000
2021-282	11	ARC		18	58.389	2.811	4.82	47.000	70.000
2021-282	25	S. DIME		21	45.400	3.169	6.98	36.000	54.000
2021-283	3	VITROS		38	24.421	3.924	16.07	20.000	29.000
2021-283	5	ROCHE E		33	20.576	2.161	10.50	16.000	25.000
2021-283	9	BECK.COUL.		43	30.952	3.039	9.82	25.000	37.000
2021-283	9	BECK.COUL.		43	34.750	2.586	7.44	28.000	42.000
2021-283	11	ARC		18	27.278	1.096	4.02	22.000	33.000
2021-283	25	S. DIME		21	23.100	1.609	6.97	18.000	28.000
2021-284	3	VITROS		38	43.351	4.467	10.31	35.000	52.000
2021-284	5	ROCHE E		33	32.188	2.518	7.82	26.000	39.000
2021-284	9	BECK.COUL.		43	56.500	7.399	13.10	45.000	68.000
2021-284	9	BECK.COUL.		43	50.619	3.994	7.89	40.000	61.000
2021-284	11	ARC		18	45.556	1.641	3.60	36.000	55.000
2021-284	25	S. DIME		21	34.900	2.385	6.83	28.000	42.000
2021-285	3	VITROS		38	33.730	4.091	12.13	27.000	40.000
2021-285	5	ROCHE E		33	23.156	2.181	9.42	19.000	28.000
2021-285	9	BECK.COUL.		43	40.750	5.761	14.14	33.000	49.000
2021-285	9	BECK.COUL.		43	36.143	3.152	8.72	29.000	43.000

8/17/2021

PROFICIENCY TESTING SERVICE
LABORATORY SERVICES PROGRAM
DEPARTMENT OF HEALTH OF PUERTO RICO

PAGE : 123 (123)
DATE : 8/17/2021
TIME : 13:16:10

Participant Summary

MONTH 5 OF YEAR 2021

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2021-285	11 ARC			18	31.111	1.663	5.35	25.000	37.000
2021-285	25 S. DIME			21	26.100	1.814	6.95	21.000	31.000