

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
001 ALAN AMINO	ALT/SGPT	CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %										
2022-391	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	21	298.095	9.937	3.33	238.000	358.000				
2022-391	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	77	315.519	8.485	2.69	252.000	379.000				
2022-391	11 MOD. IFCC	184 INTEGRA	93 ROCHE	25	291.880	10.164	3.48	234.000	350.000				
2022-391	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	29	263.621	8.219	3.12	211.000	316.000				
2022-391	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	316.500	9.014	2.85	253.000	380.000				
2022-391	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	334.229	22.557	6.75	267.000	401.000				
2022-391	11 MOD. IFCC	203 ROCHE	93 ROCHE	30	293.517	5.793	1.97	235.000	352.000				
2022-391	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	341.700	12.133	3.55	273.000	410.000				
2022-391	13 ALA &aKETO	165 VITROS	155 ALTV	127	265.664	6.390	2.41	213.000	319.000				
2022-392	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	21	17.048	1.914	11.23	14.000	20.000				
2022-392	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	77	23.182	2.118	9.14	19.000	28.000				
2022-392	11 MOD. IFCC	184 INTEGRA	93 ROCHE	25	16.720	1.001	5.99	13.000	20.000				
2022-392	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	29	15.429	0.942	6.11	12.000	19.000				
2022-392	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	19.400	1.438	7.41	16.000	23.000				
2022-392	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	23.250	2.165	9.31	19.000	28.000				
2022-392	11 MOD. IFCC	203 ROCHE	93 ROCHE	30	16.862	0.937	5.56	13.000	20.000				
2022-392	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	20.800	3.458	16.63	17.000	25.000				
2022-392	13 ALA &aKETO	165 VITROS	155 ALTV	127	14.248	0.665	4.67	11.000	17.000				
2022-393	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	21	410.286	12.743	3.11	328.000	492.000				
2022-393	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	77	427.727	13.806	3.23	342.000	513.000				
2022-393	11 MOD. IFCC	184 INTEGRA	93 ROCHE	25	393.240	14.426	3.67	315.000	472.000				
2022-393	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	29	361.207	10.704	2.96	289.000	433.000				
2022-393	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	430.750	13.899	3.23	345.000	517.000				
2022-393	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	444.371	37.762	8.50	355.000	533.000				
2022-393	11 MOD. IFCC	203 ROCHE	93 ROCHE	30	398.200	8.689	2.18	319.000	478.000				
2022-393	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	462.300	15.697	3.40	370.000	555.000				
2022-393	13 ALA &aKETO	165 VITROS	155 ALTV	127	363.240	10.464	2.88	291.000	436.000				

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
001	ALAN AMINO ALT/SGPT	CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %										
2022-394	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	21	120.476	4.584	3.80	96.000	145.000				
2022-394	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	77	133.118	4.386	3.29	106.000	160.000				
2022-394	11 MOD. IFCC	184 INTEGRA	93 ROCHE	25	118.840	3.663	3.08	95.000	143.000				
2022-394	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	29	107.586	2.966	2.76	86.000	129.000				
2022-394	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	134.750	4.710	3.50	108.000	162.000				
2022-394	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	136.229	7.395	5.43	109.000	163.000				
2022-394	11 MOD. IFCC	203 ROCHE	93 ROCHE	30	118.793	2.657	2.24	95.000	143.000				
2022-394	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	140.400	5.295	3.77	112.000	168.000				
2022-394	13 ALA &aKETO	165 VITROS	155 ALTV	127	111.552	3.061	2.74	89.000	134.000				
2022-395	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	21	77.333	3.329	4.30	62.000	93.000				
2022-395	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	77	87.675	3.340	3.81	70.000	105.000				
2022-395	11 MOD. IFCC	184 INTEGRA	93 ROCHE	25	75.875	1.641	2.16	61.000	91.000				
2022-395	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	29	68.586	2.043	2.98	55.000	82.000				
2022-395	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	87.257	5.552	6.36	70.000	105.000				
2022-395	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	36	88.750	3.112	3.51	71.000	107.000				
2022-395	11 MOD. IFCC	203 ROCHE	93 ROCHE	30	76.267	1.526	2.00	61.000	92.000				
2022-395	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	92.700	7.772	8.38	74.000	111.000				
2022-395	13 ALA &aKETO	165 VITROS	155 ALTV	127	71.508	1.999	2.80	57.000	86.000				
002	ALBUMIN	CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %										
2022-391	3 DYE BI-B.G	165 VITROS	63 KODAK	144	4.538	0.128	2.82	4.100	5.000				
2022-391	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	35	4.509	0.156	3.45	4.100	5.000				
2022-391	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	28	4.268	0.097	2.26	3.800	4.700				
2022-391	3 DYE BI-B.G	191 RX-SERIES	126 RANDOX	39	4.387	0.181	4.14	3.900	4.800				
2022-391	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	19	4.211	0.125	2.97	3.800	4.600				
2022-391	3 DYE BI-B.G	203 ROCHE	93 ROCHE	39	4.547	0.139	3.06	4.100	5.000				
2022-391	3 DYE BI-B.G	219 ATELLICA	143 SIEMENS	10	4.420	0.117	2.64	4.000	4.900				
2022-391	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	4.144	0.119	2.87	3.700	4.600				
2022-391	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	13	4.138	0.133	3.22	3.700	4.600				

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE	10 %							
2022-392	3 DYE BI-B.G	165 VITROS	63 KODAK		144	1.887	0.050	2.66	1.700	2.100	
2022-392	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		35	1.994	0.068	3.43	1.800	2.200	
2022-392	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		28	1.889	0.042	2.20	1.700	2.100	
2022-392	3 DYE BI-B.G	191 RX-SERIES	126 RANDOX		39	2.005	0.085	4.22	1.800	2.200	
2022-392	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	1.847	0.094	5.08	1.700	2.000	
2022-392	3 DYE BI-B.G	203 ROCHE	93 ROCHE		39	2.050	0.123	6.00	1.800	2.300	
2022-392	3 DYE BI-B.G	219 ATELLICA	143 SIEMENS		10	1.950	0.050	2.56	1.800	2.100	
2022-392	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		80	1.720	0.066	3.86	1.500	1.900	
2022-392	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		13	1.723	0.105	6.09	1.600	1.900	
2022-393	3 DYE BI-B.G	165 VITROS	63 KODAK		144	5.727	0.163	2.84	5.200	6.300	
2022-393	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		35	5.385	0.133	2.47	4.800	5.900	
2022-393	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		28	5.167	0.119	2.30	4.700	5.700	
2022-393	3 DYE BI-B.G	191 RX-SERIES	126 RANDOX		39	5.256	0.215	4.08	4.700	5.800	
2022-393	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	5.142	0.166	3.24	4.600	5.700	
2022-393	3 DYE BI-B.G	203 ROCHE	93 ROCHE		39	5.513	0.186	3.37	5.000	6.100	
2022-393	3 DYE BI-B.G	219 ATELLICA	143 SIEMENS		10	5.360	0.143	2.66	4.800	5.900	
2022-393	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		80	5.106	0.161	3.15	4.600	5.600	
2022-393	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		13	5.138	0.150	2.91	4.600	5.700	
2022-394	3 DYE BI-B.G	165 VITROS	63 KODAK		144	2.866	0.082	2.86	2.600	3.200	
2022-394	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		35	2.982	0.078	2.63	2.700	3.300	
2022-394	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		28	2.821	0.094	3.33	2.500	3.100	
2022-394	3 DYE BI-B.G	191 RX-SERIES	126 RANDOX		39	2.932	0.098	3.33	2.600	3.200	
2022-394	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		19	2.774	0.102	3.67	2.500	3.100	
2022-394	3 DYE BI-B.G	203 ROCHE	93 ROCHE		39	3.021	0.092	3.05	2.700	3.300	
2022-394	3 DYE BI-B.G	219 ATELLICA	143 SIEMENS		10	2.920	0.075	2.56	2.600	3.200	
2022-394	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		80	2.639	0.080	3.04	2.400	2.900	
2022-394	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		13	2.638	0.133	5.05	2.400	2.900	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
002	ALBUMIN	CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE	10	%						
2022-395	3 DYE BI-B.G	165 VITROS	63 KODAK	144	2.446	0.072	2.93	2.200	2.700		
2022-395	3 DYE BI-B.G	184 INTEGRA	93 ROCHE	35	2.556	0.060	2.36	2.300	2.800		
2022-395	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN	28	2.429	0.065	2.66	2.200	2.700		
2022-395	3 DYE BI-B.G	191 RX-SERIES	126 RANDOX	39	2.553	0.104	4.09	2.300	2.800		
2022-395	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT	19	2.389	0.085	3.57	2.200	2.600		
2022-395	3 DYE BI-B.G	203 ROCHE	93 ROCHE	39	2.600	0.122	4.69	2.300	2.900		
2022-395	3 DYE BI-B.G	219 ATELLICA	143 SIEMENS	10	2.610	0.234	8.98	2.300	2.900		
2022-395	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS	80	2.249	0.073	3.23	2.000	2.500		
2022-395	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS	13	2.269	0.114	5.01	2.000	2.500		
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE	30	%						
2022-391	3 COLORIMET.	184 INTEGRA	93 ROCHE	28	239.037	8.771	3.67	167.000	311.000		
2022-391	3 COLORIMET.	203 ROCHE	93 ROCHE	26	235.600	8.275	3.51	165.000	306.000		
2022-391	5 KINETIC 30	197 ARCHITECT	1 ABBOTT	10	210.300	14.043	6.68	147.000	273.000		
2022-391	6 KINETIC 37	165 VITROS	63 KODAK	144	158.514	8.480	5.35	111.000	206.000		
2022-391	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	29	192.241	14.112	7.34	135.000	250.000		
2022-391	10 IFCC	219 ATELLICA	143 SIEMENS	10	238.100	8.573	3.60	167.000	310.000		
2022-391	11 Modified I	38 SIEMENS DI	143 SIEMENS	76	238.293	8.513	3.57	167.000	310.000		
2022-391	11 Modified I	191 RX-SERIES	126 RANDOX	29	193.138	23.424	12.13	135.000	251.000		
2022-391	11 Modified I	191 RX-SERIES	126 RANDOX	29	203.286	37.331	18.36	142.000	264.000		
2022-391	12 P NITRO PH	197 ARCHITECT	1 ABBOTT	11	210.455	14.093	6.70	147.000	274.000		
2022-392	3 COLORIMET.	184 INTEGRA	93 ROCHE	28	22.741	1.205	5.30	16.000	30.000		
2022-392	3 COLORIMET.	203 ROCHE	93 ROCHE	26	22.040	1.183	5.37	15.000	29.000		
2022-392	5 KINETIC 30	197 ARCHITECT	1 ABBOTT	10	20.000	1.844	9.22	14.000	26.000		
2022-392	6 KINETIC 37	165 VITROS	63 KODAK	144	36.608	2.559	6.99	26.000	48.000		
2022-392	8 BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	29	17.036	1.546	9.08	12.000	22.000		
2022-392	10 IFCC	219 ATELLICA	143 SIEMENS	10	23.400	1.114	4.76	16.000	30.000		
2022-392	11 Modified I	38 SIEMENS DI	143 SIEMENS	76	22.400	2.708	12.09	16.000	29.000		
2022-392	11 Modified I	191 RX-SERIES	126 RANDOX	29	20.586	3.317	16.11	14.000	27.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-392	11	Modified I	191	RX-SERIES	126	RANDOX	29	23.143	6.833	29.53	16.000	30.000
2022-392	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	11	20.455	2.189	10.70	14.000	27.000
003	ALKALINE PHOSPHATE		CLIA TEST ID #: 275				EVALUATION CRITERIA: TARGET VALUE 30 %					
2022-393	3	COLORIMET.	184	INTEGRA	93	ROCHE	28	328.536	12.999	3.96	230.000	427.000
2022-393	3	COLORIMET.	203	ROCHE	93	ROCHE	26	327.308	9.526	2.91	229.000	426.000
2022-393	5	KINETIC 30	197	ARCHITECT	1	ABBOTT	10	284.500	21.481	7.55	199.000	370.000
2022-393	6	KINETIC 37	165	VITROS	63	KODAK	144	172.853	10.668	6.17	121.000	225.000
2022-393	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	29	265.483	19.760	7.44	186.000	345.000
2022-393	10	IFCC	219	ATELLICA	143	SIEMENS	10	324.800	9.152	2.82	227.000	422.000
2022-393	11	Modified I	38	SIEMENS DI	143	SIEMENS	76	330.813	13.315	4.03	232.000	430.000
2022-393	11	Modified I	191	RX-SERIES	126	RANDOX	29	278.429	55.229	19.84	195.000	362.000
2022-393	11	Modified I	191	RX-SERIES	126	RANDOX	29	261.828	31.053	11.86	183.000	340.000
2022-393	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	11	287.727	18.518	6.44	201.000	374.000
2022-394	3	COLORIMET.	184	INTEGRA	93	ROCHE	28	100.821	4.009	3.98	71.000	131.000
2022-394	3	COLORIMET.	203	ROCHE	93	ROCHE	26	100.040	2.676	2.67	70.000	130.000
2022-394	5	KINETIC 30	197	ARCHITECT	1	ABBOTT	10	88.900	7.035	7.91	62.000	116.000
2022-394	6	KINETIC 37	165	VITROS	63	KODAK	144	101.496	6.135	6.04	71.000	132.000
2022-394	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	29	80.517	6.532	8.11	56.000	105.000
2022-394	10	IFCC	219	ATELLICA	143	SIEMENS	10	100.100	3.833	3.83	70.000	130.000
2022-394	11	Modified I	38	SIEMENS DI	143	SIEMENS	76	101.240	4.378	4.32	71.000	132.000
2022-394	11	Modified I	191	RX-SERIES	126	RANDOX	29	83.357	8.595	10.31	58.000	108.000
2022-394	11	Modified I	191	RX-SERIES	126	RANDOX	29	93.286	16.731	17.93	65.000	121.000
2022-394	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	11	89.091	5.869	6.59	62.000	116.000
2022-395	3	COLORIMET.	184	INTEGRA	93	ROCHE	28	67.750	2.558	3.78	47.000	88.000
2022-395	3	COLORIMET.	203	ROCHE	93	ROCHE	26	67.346	1.980	2.94	47.000	88.000
2022-395	5	KINETIC 30	197	ARCHITECT	1	ABBOTT	10	60.400	4.630	7.67	42.000	79.000
2022-395	6	KINETIC 37	165	VITROS	63	KODAK	144	77.660	5.196	6.69	54.000	101.000
2022-395	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	29	54.000	4.594	8.51	38.000	70.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-395	10 IFCC	219 ATELLICA	143	SIEMENS	10	69.900	7.661	10.96	49.000	91.000	
2022-395	11 Modified I	38 SIEMENS DI	143	SIEMENS	76	68.307	3.398	4.97	48.000	89.000	
2022-395	11 Modified I	191 RX-SERIES	126	RANDOX	29	57.857	5.125	8.86	41.000	75.000	
2022-395	11 Modified I	191 RX-SERIES	126	RANDOX	29	65.000	12.236	18.82	46.000	85.000	
2022-395	12 P NITRO PH	197 ARCHITECT	1	ABBOTT	11	60.091	4.055	6.75	42.000	78.000	
004 AMYLASE		CLIA TEST ID #: 285	EVALUATION CRITERIA: TARGET VALUE 30 %								
2022-391	3 ENZYM. 37C	165 VITROS	63	KODAK	84	162.107	5.878	3.63	113.000	211.000	
2022-391	4 HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	24	219.750	3.929	1.79	154.000	286.000	
2022-391	9 ENZ/COL.37	184 INTEGRA	93	ROCHE	12	229.250	7.108	3.10	160.000	298.000	
2022-391	9 ENZ/COL.37	203 ROCHE	93	ROCHE	23	221.478	7.762	3.50	155.000	288.000	
2022-391	10 CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	172.000	8.860	5.15	120.000	224.000	
2022-392	3 ENZYM. 37C	165 VITROS	63	KODAK	84	30.120	0.782	2.60	21.000	39.000	
2022-392	4 HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	24	23.261	0.606	2.60	16.000	30.000	
2022-392	9 ENZ/COL.37	184 INTEGRA	93	ROCHE	12	27.250	0.829	3.04	19.000	35.000	
2022-392	9 ENZ/COL.37	203 ROCHE	93	ROCHE	23	26.174	0.916	3.50	18.000	34.000	
2022-392	10 CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	17.867	1.087	6.09	13.000	23.000	
2022-393	3 ENZYM. 37C	165 VITROS	63	KODAK	84	216.892	7.619	3.51	152.000	282.000	
2022-393	4 HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	24	300.458	5.723	1.90	210.000	391.000	
2022-393	9 ENZ/COL.37	184 INTEGRA	93	ROCHE	12	309.083	7.041	2.28	216.000	402.000	
2022-393	9 ENZ/COL.37	203 ROCHE	93	ROCHE	23	302.087	11.209	3.71	211.000	393.000	
2022-393	10 CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	234.250	13.339	5.69	164.000	305.000	
2022-394	3 ENZYM. 37C	165 VITROS	63	KODAK	84	72.036	4.255	5.91	50.000	94.000	
2022-394	4 HYDRO ACT.	38 SIEMENS DI	143	SIEMENS	24	95.043	1.681	1.77	67.000	124.000	
2022-394	9 ENZ/COL.37	184 INTEGRA	93	ROCHE	12	99.750	2.165	2.17	70.000	130.000	
2022-394	9 ENZ/COL.37	203 ROCHE	93	ROCHE	23	96.091	2.021	2.10	67.000	125.000	
2022-394	10 CNPG3	186 BECK. COUL	76	AU-BECKMAN	16	73.938	4.380	5.92	52.000	96.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
004	AMYLASE	CLIA TEST ID #: 285		EVALUATION CRITERIA: TARGET VALUE 30 %							
2022-395	3 ENZYM. 37C	165 VITROS	63 KODAK	84	50.131	4.122	8.22	35.000	65.000		
2022-395	4 HYDRO ACT.	38 SIEMENS DI	143 SIEMENS	24	64.667	0.943	1.46	45.000	84.000		
2022-395	9 ENZ/COL.37	184 INTEGRA	93 ROCHE	12	68.667	1.795	2.61	48.000	89.000		
2022-395	9 ENZ/COL.37	203 ROCHE	93 ROCHE	23	66.455	1.233	1.86	47.000	86.000		
2022-395	10 CNPG3	186 BECK. COUL	76 AU-BECKMAN	16	49.938	2.904	5.82	35.000	65.000		
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %							
2022-391	1 ASPAR&KETO	165 VITROS	63 KODAK	139	408.832	13.317	3.26	327.000	491.000		
2022-391	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	318.190	9.019	2.83	255.000	382.000		
2022-391	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	12	316.750	7.108	2.24	253.000	380.000		
2022-391	11 MOD. IFCC	184 INTEGRA	93 ROCHE	27	326.000	10.958	3.36	261.000	391.000		
2022-391	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	28	279.607	14.686	5.25	224.000	336.000		
2022-391	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	362.919	36.569	10.08	290.000	436.000		
2022-391	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	360.000	30.243	8.40	288.000	432.000		
2022-391	11 MOD. IFCC	203 ROCHE	93 ROCHE	28	324.607	11.803	3.64	260.000	390.000		
2022-391	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	364.300	10.771	2.96	291.000	437.000		
2022-391	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	15	308.867	9.885	3.20	247.000	371.000		
2022-392	1 ASPAR&KETO	165 VITROS	63 KODAK	139	46.101	1.515	3.29	37.000	55.000		
2022-392	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	42.750	2.395	5.60	34.000	51.000		
2022-392	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	12	41.417	2.465	5.95	33.000	50.000		
2022-392	11 MOD. IFCC	184 INTEGRA	93 ROCHE	27	42.630	1.281	3.00	34.000	51.000		
2022-392	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	28	36.821	2.122	5.76	29.000	44.000		
2022-392	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	46.333	4.110	8.87	37.000	56.000		
2022-392	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	48.108	4.385	9.12	38.000	58.000		
2022-392	11 MOD. IFCC	203 ROCHE	93 ROCHE	28	42.519	1.772	4.17	34.000	51.000		
2022-392	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	49.800	2.561	5.14	40.000	60.000		
2022-392	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	15	40.267	1.181	2.93	32.000	48.000		
2022-393	1 ASPAR&KETO	165 VITROS	63 KODAK	139	597.688	26.253	4.39	478.000	717.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2022-393	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	430.888	13.330	3.09	345.000	517.000			
2022-393	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	12	431.167	11.767	2.73	345.000	517.000			
2022-393	11 MOD. IFCC	184 INTEGRA	93 ROCHE	27	429.111	14.700	3.43	343.000	515.000			
2022-393	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	28	371.964	20.058	5.39	298.000	446.000			
2022-393	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	466.278	44.725	9.59	373.000	560.000			
2022-393	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	474.667	52.690	11.10	380.000	570.000			
2022-393	11 MOD. IFCC	203 ROCHE	93 ROCHE	28	433.821	15.055	3.47	347.000	521.000			
2022-393	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	481.400	18.747	3.89	385.000	578.000			
2022-393	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	15	412.667	11.937	2.89	330.000	495.000			
005 ASP. AMINO AST/SGOT		CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %								
2022-394	1 ASPAR&KETO	165 VITROS	63 KODAK	139	164.533	5.101	3.10	132.000	197.000			
2022-394	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	144.346	4.260	2.95	115.000	173.000			
2022-394	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	12	142.917	3.040	2.13	114.000	172.000			
2022-394	11 MOD. IFCC	184 INTEGRA	93 ROCHE	27	147.815	4.046	2.74	118.000	177.000			
2022-394	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	28	128.111	6.309	4.92	102.000	154.000			
2022-394	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	160.000	13.140	8.21	128.000	192.000			
2022-394	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	163.917	11.964	7.30	131.000	197.000			
2022-394	11 MOD. IFCC	203 ROCHE	93 ROCHE	28	147.429	6.068	4.12	118.000	177.000			
2022-394	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	166.900	5.873	3.52	134.000	200.000			
2022-394	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	15	141.267	4.074	2.88	113.000	170.000			
2022-395	1 ASPAR&KETO	165 VITROS	63 KODAK	139	112.686	3.588	3.18	90.000	135.000			
2022-395	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	101.925	3.282	3.22	82.000	122.000			
2022-395	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	12	99.833	3.105	3.11	80.000	120.000			
2022-395	11 MOD. IFCC	184 INTEGRA	93 ROCHE	27	103.556	3.047	2.94	83.000	124.000			
2022-395	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	28	88.741	2.823	3.18	71.000	106.000			
2022-395	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	115.568	9.531	8.25	92.000	139.000			
2022-395	11 MOD. IFCC	191 RX-SERIES	126 RANDOX	37	112.000	11.518	10.28	90.000	134.000			
2022-395	11 MOD. IFCC	203 ROCHE	93 ROCHE	28	102.821	3.392	3.30	82.000	123.000			
2022-395	11 MOD. IFCC	219 ATELLICA	143 SIEMENS	10	121.800	13.504	11.09	97.000	146.000			

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-395	13	NADH w _o P5	197	ARCHITECT	1	ABBOTT	15	99.400	3.382	3.40	80.000	119.000
006	BILIRUBIN, TOTAL		CLIA TEST ID #: 305		EVALUATION CRITERIA: TARGET VALUE 20 %							
2022-391	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	3.914	0.148	3.78	3.130	4.700
2022-391	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	27	3.883	0.085	2.18	3.110	4.660
2022-391	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	3.924	0.318	8.12	3.140	4.710
2022-391	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	3.630	0.570	15.69	2.900	4.360
2022-391	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	3.900	0.192	4.92	3.120	4.680
2022-391	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	3.686	0.199	5.40	2.950	4.420
2022-391	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	3.600	0.082	2.27	2.880	4.320
2022-391	13	AZOBIL.	184	INTEGRA	93	ROCHE	15	3.502	0.155	4.43	2.800	4.200
2022-391	13	AZOBIL.	203	ROCHE	93	ROCHE	26	3.449	0.155	4.50	2.760	4.140
2022-391	14	VANADATE	219	ATELLICA	143	SIEMENS	10	4.206	0.213	5.07	3.360	5.050
2022-391	18	COLORIMETR	203	ROCHE	93	ROCHE	13	3.455	0.086	2.48	2.760	4.150
2022-391	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	3.873	0.208	5.36	3.100	4.650
2022-392	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	0.104	0.023	22.28	0.080	0.130
2022-392	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	27	0.179	0.020	11.22	0.140	0.210
2022-392	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	0.132	0.027	20.55	0.110	0.160
2022-392	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	0.148	0.052	35.00	0.120	0.180
2022-392	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	0.106	0.010	9.62	0.080	0.130
2022-392	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	0.267	0.047	17.68	0.210	0.320
2022-392	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	0.167	0.062	37.08	0.130	0.200
2022-392	13	AZOBIL.	184	INTEGRA	93	ROCHE	15	0.098	0.041	41.36	0.080	0.120
2022-392	13	AZOBIL.	203	ROCHE	93	ROCHE	26	0.145	0.018	12.69	0.120	0.170
2022-392	14	VANADATE	219	ATELLICA	143	SIEMENS	10	0.182	0.023	12.48	0.150	0.220
2022-392	18	COLORIMETR	203	ROCHE	93	ROCHE	13	0.153	0.026	16.68	0.120	0.180
2022-392	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	0.110	0.012	10.76	0.090	0.130
2022-393	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	5.468	0.251	4.59	4.370	6.560
2022-393	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	27	5.295	0.153	2.90	4.240	6.350

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-393	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	5.607	0.329	5.87	4.490	6.730
2022-393	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	5.196	0.722	13.89	4.160	6.240
2022-393	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	5.433	0.367	6.75	4.350	6.520
2022-393	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	5.041	0.290	5.75	4.030	6.050
2022-393	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	4.600	0.294	6.40	3.680	5.520
2022-393	13	AZOBIL.	184	INTEGRA	93	ROCHE	15	4.764	0.223	4.68	3.810	5.720
2022-393	13	AZOBIL.	203	ROCHE	93	ROCHE	26	4.778	0.252	5.27	3.820	5.730
2022-393	14	VANADATE	219	ATELLICA	143	SIEMENS	10	5.879	0.272	4.63	4.700	7.050
2022-393	18	COLORIMETR	203	ROCHE	93	ROCHE	13	4.884	0.098	2.00	3.910	5.860
2022-393	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	5.411	0.248	4.59	4.330	6.490
006 BILIRUBIN, TOTAL			CLIA TEST ID #: 305			EVALUATION CRITERIA: TARGET VALUE 20 %						
2022-394	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	1.461	0.075	5.16	1.170	1.750
2022-394	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	27	1.552	0.059	3.78	1.240	1.860
2022-394	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	1.532	0.109	7.08	1.230	1.840
2022-394	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	1.499	0.127	8.48	1.200	1.800
2022-394	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	1.462	0.084	5.73	1.170	1.750
2022-394	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	1.467	0.170	11.59	1.170	1.760
2022-394	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	1.466	0.099	6.75	1.170	1.760
2022-394	13	AZOBIL.	184	INTEGRA	93	ROCHE	15	1.295	0.062	4.83	1.040	1.550
2022-394	13	AZOBIL.	203	ROCHE	93	ROCHE	26	1.268	0.070	5.49	1.010	1.520
2022-394	14	VANADATE	219	ATELLICA	143	SIEMENS	10	1.629	0.068	4.20	1.300	1.950
2022-394	18	COLORIMETR	203	ROCHE	93	ROCHE	13	1.295	0.037	2.85	1.040	1.550
2022-394	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	10	1.520	0.116	7.65	1.220	1.820
2022-395	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	0.863	0.049	5.72	0.690	1.040
2022-395	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	27	0.961	0.047	4.84	0.770	1.150
2022-395	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	0.908	0.119	13.11	0.730	1.090
2022-395	4	J-G WITH B	191	RX-SERIES	126	RANDOX	31	0.850	0.100	11.81	0.680	1.020
2022-395	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	10	0.892	0.073	8.24	0.710	1.070
2022-395	7	DIAZO-DYPH	165	VITROS	63	KODAK	139	0.897	0.083	9.28	0.720	1.080

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-395	7 DIAZO-DYPH	165 VITROS	63 KODAK	139	0.933	0.047	5.05	0.750	1.120		
2022-395	13 AZOBIL.	184 INTEGRA	93 ROCHE	15	0.784	0.048	6.16	0.630	0.940		
2022-395	13 AZOBIL.	203 ROCHE	93 ROCHE	26	0.757	0.050	6.56	0.610	0.910		
2022-395	14 VANADATE	219 ATELLICA	143 SIEMENS	10	1.022	0.113	11.04	0.820	1.230		
2022-395	18 COLORIMETR	203 ROCHE	93 ROCHE	13	0.770	0.031	4.04	0.620	0.920		
2022-395	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	10	0.937	0.069	7.33	0.750	1.120		
010 CALCIUM, TOTAL		CLIA TEST ID #: 345	EVALUATION CRITERIA: TARGET VALUE 1.00								
2022-391	3 ARSEN.-III	165 VITROS	63 KODAK	157	10.421	0.202	1.94	9.400	11.400		
2022-391	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	31	10.703	0.212	1.98	9.700	11.700		
2022-391	3 ARSEN.-III	191 RX-SERIES	126 RANDOX	38	10.895	0.325	2.99	9.900	11.900		
2022-391	3 ARSEN.-III	191 RX-SERIES	126 RANDOX	38	10.750	0.391	3.63	9.800	11.800		
2022-391	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	12	10.483	0.258	2.46	9.500	11.500		
2022-391	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	81	10.263	0.295	2.87	9.300	11.300		
2022-391	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	16	10.669	0.280	2.62	9.700	11.700		
2022-391	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	10.660	0.215	2.02	9.700	11.700		
2022-391	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	13	10.177	0.454	4.46	9.200	11.200		
2022-391	8 CRESOLPHT.	219 ATELLICA	143 SIEMENS	10	10.520	0.189	1.79	9.500	11.500		
2022-391	15 CA-NM-BAPT	184 INTEGRA	93 ROCHE	20	10.755	0.220	2.05	9.800	11.800		
2022-391	15 CA-NM-BAPT	203 ROCHE	93 ROCHE	29	10.750	0.199	1.85	9.800	11.800		
2022-392	3 ARSEN.-III	165 VITROS	63 KODAK	157	5.728	0.154	2.69	4.700	6.700		
2022-392	3 ARSEN.-III	186 BECK. COUL	76 AU-BECKMAN	31	5.819	0.197	3.39	4.800	6.800		
2022-392	3 ARSEN.-III	191 RX-SERIES	126 RANDOX	38	6.463	0.433	6.70	5.500	7.500		
2022-392	3 ARSEN.-III	191 RX-SERIES	126 RANDOX	38	6.268	0.288	4.59	5.300	7.300		
2022-392	3 ARSEN.-III	197 ARCHITECT	1 ABBOTT	12	5.825	0.116	2.00	4.800	6.800		
2022-392	8 CRESOLPHT.	38 SIEMENS DI	143 SIEMENS	81	5.822	0.200	3.43	4.800	6.800		
2022-392	8 CRESOLPHT.	184 INTEGRA	93 ROCHE	16	5.956	0.106	1.78	5.000	7.000		
2022-392	8 CRESOLPHT.	203 ROCHE	93 ROCHE	10	6.010	0.122	2.03	5.000	7.000		
2022-392	8 CRESOLPHT.	212 S. D. EXL	143 SIEMENS	13	5.692	0.407	7.14	4.700	6.700		
2022-392	8 CRESOLPHT.	219 ATELLICA	143 SIEMENS	10	5.570	0.173	3.11	4.600	6.600		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-392	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	20	5.980	0.189	3.16	5.000	7.000
2022-392	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	29	6.036	0.163	2.70	5.000	7.000
010	CALCIUM, TOTAL		CLIA TEST ID #: 345			EVALUATION CRITERIA: TARGET VALUE 1.00						
2022-393	3	ARSEN.-III	165	VITROS	63	KODAK	157	12.159	0.213	1.75	11.200	13.200
2022-393	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	31	12.700	0.352	2.77	11.700	13.700
2022-393	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	11.888	0.314	2.64	10.900	12.900
2022-393	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	12.658	0.568	4.49	11.700	13.700
2022-393	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	12.467	0.579	4.65	11.500	13.500
2022-393	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	12.237	0.379	3.09	11.200	13.200
2022-393	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	16	12.475	0.256	2.05	11.500	13.500
2022-393	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	12.580	0.204	1.62	11.600	13.600
2022-393	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	13	12.323	0.377	3.06	11.300	13.300
2022-393	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	10	12.420	0.232	1.86	11.400	13.400
2022-393	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	20	12.665	0.317	2.50	11.700	13.700
2022-393	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	29	12.600	0.260	2.07	11.600	13.600
2022-394	3	ARSEN.-III	165	VITROS	63	KODAK	157	7.554	0.164	2.17	6.600	8.600
2022-394	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	31	7.629	0.269	3.53	6.600	8.600
2022-394	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	8.076	0.377	4.67	7.100	9.100
2022-394	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	8.075	0.556	6.89	7.100	9.100
2022-394	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	7.500	0.153	2.04	6.500	8.500
2022-394	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	7.453	0.206	2.77	6.500	8.500
2022-394	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	16	7.694	0.125	1.62	6.700	8.700
2022-394	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	7.710	0.158	2.05	6.700	8.700
2022-394	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	13	7.392	0.408	5.53	6.400	8.400
2022-394	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	10	7.460	0.196	2.63	6.500	8.500
2022-394	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	20	7.760	0.150	1.93	6.800	8.800
2022-394	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	29	7.728	0.200	2.59	6.700	8.700
2022-395	3	ARSEN.-III	165	VITROS	63	KODAK	157	6.760	0.157	2.32	5.800	7.800

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-395	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	31	6.816	0.223	3.27	5.800	7.800
2022-395	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	7.450	0.469	6.30	6.500	8.500
2022-395	3	ARSEN.-III	191	RX-SERIES	126	RANDOX	38	7.278	0.316	4.35	6.300	8.300
2022-395	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	12	6.750	0.126	1.86	5.800	7.800
2022-395	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	6.699	0.217	3.23	5.700	7.700
2022-395	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	16	6.906	0.164	2.37	5.900	7.900
2022-395	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	6.950	0.180	2.59	6.000	8.000
2022-395	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	13	6.600	0.384	5.82	5.600	7.600
2022-395	8	CRESOLPHT.	219	ATELLICA	143	SIEMENS	10	6.870	0.795	11.57	5.900	7.900
2022-395	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	20	6.921	0.124	1.79	5.900	7.900
2022-395	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	29	6.954	0.124	1.78	6.000	8.000
011	CHLORIDE	CLIA TEST ID #: 355				EVALUATION CRITERIA: TARGET VALUE 5 %						
2022-391	3	ISE DILUTE	184	INTEGRA	93	ROCHE	38	109.194	1.898	1.74	104.000	115.000
2022-391	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	29	106.276	1.506	1.42	101.000	112.000
2022-391	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	17	108.882	1.490	1.37	103.000	114.000
2022-391	3	ISE DILUTE	203	ROCHE	93	ROCHE	42	106.190	1.880	1.77	101.000	112.000
2022-391	4	ISE UND-DI	165	VITROS	63	KODAK	83	108.193	1.556	1.44	103.000	114.000
2022-391	4	ISE UND-DI	165	VITROS	156	Kit 32	72	108.775	1.396	1.28	103.000	114.000
2022-391	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	39	107.711	2.199	2.04	102.000	113.000
2022-391	4	ISE UND-DI	191	RX-SERIES	126	RANDOX	39	109.000	2.062	1.89	104.000	114.000
2022-391	8	IMT	219	ATELLICA	143	SIEMENS	10	110.700	3.551	3.21	105.000	116.000
2022-391	9	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	110.169	2.016	1.83	105.000	116.000
2022-391	9	QUICK LYTE	212	S. D. EXL	143	SIEMENS	13	109.231	1.761	1.61	104.000	115.000
2022-392	3	ISE DILUTE	184	INTEGRA	93	ROCHE	38	90.842	1.899	2.09	86.000	95.000
2022-392	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	29	92.310	1.600	1.73	88.000	97.000
2022-392	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	17	94.588	1.497	1.58	90.000	99.000
2022-392	3	ISE DILUTE	203	ROCHE	93	ROCHE	42	90.146	1.920	2.13	86.000	95.000
2022-392	4	ISE UND-DI	165	VITROS	63	KODAK	83	94.012	1.517	1.61	89.000	99.000
2022-392	4	ISE UND-DI	165	VITROS	156	Kit 32	72	94.278	1.397	1.48	90.000	99.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-392	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	90.750	3.345	3.69	86.000	95.000		
2022-392	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	89.821	2.024	2.25	85.000	94.000		
2022-392	8 IMT	219 ATELLICA	143 SIEMENS	10	92.600	1.855	2.00	88.000	97.000		
2022-392	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	89.679	1.573	1.75	85.000	94.000		
2022-392	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	13	88.231	1.527	1.73	84.000	93.000		
011 CHLORIDE		CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2022-393	3 ISE DILUTE	184 INTEGRA	93 ROCHE	38	116.447	2.672	2.29	111.000	122.000		
2022-393	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	112.069	2.132	1.90	106.000	118.000		
2022-393	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	17	114.588	1.647	1.44	109.000	120.000		
2022-393	3 ISE DILUTE	203 ROCHE	93 ROCHE	42	113.548	1.917	1.69	108.000	119.000		
2022-393	4 ISE UND-DI	165 VITROS	63 KODAK	83	115.659	1.843	1.59	110.000	121.000		
2022-393	4 ISE UND-DI	165 VITROS	156 Kit 32	72	116.000	1.846	1.59	110.000	122.000		
2022-393	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	115.605	2.870	2.48	110.000	121.000		
2022-393	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	115.500	1.871	1.62	110.000	121.000		
2022-393	8 IMT	219 ATELLICA	143 SIEMENS	10	117.800	4.308	3.66	112.000	124.000		
2022-393	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	118.782	2.341	1.97	113.000	125.000		
2022-393	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	13	118.385	2.371	2.00	112.000	124.000		
2022-394	3 ISE DILUTE	184 INTEGRA	93 ROCHE	38	98.289	1.904	1.94	93.000	103.000		
2022-394	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	98.000	1.254	1.28	93.000	103.000		
2022-394	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	17	100.471	1.684	1.68	95.000	105.000		
2022-394	3 ISE DILUTE	203 ROCHE	93 ROCHE	42	96.429	2.002	2.08	92.000	101.000		
2022-394	4 ISE UND-DI	165 VITROS	63 KODAK	83	99.952	1.728	1.73	95.000	105.000		
2022-394	4 ISE UND-DI	165 VITROS	156 Kit 32	72	100.319	1.623	1.62	95.000	105.000		
2022-394	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	98.250	2.947	3.00	93.000	103.000		
2022-394	4 ISE UND-DI	191 RX-SERIES	126 RANDOX	39	96.342	2.068	2.15	92.000	101.000		
2022-394	8 IMT	219 ATELLICA	143 SIEMENS	10	100.100	2.343	2.34	95.000	105.000		
2022-394	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	98.026	1.736	1.77	93.000	103.000		
2022-394	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	13	97.000	1.840	1.90	92.000	102.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
011	CHLORIDE	CLIA TEST ID #: 355		EVALUATION CRITERIA: TARGET VALUE 5 %							
2022-395	3 ISE DILUTE	184 INTEGRA	93 ROCHE		38	93.868	1.922	2.05	89.000	99.000	
2022-395	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	94.690	1.289	1.36	90.000	99.000	
2022-395	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		17	97.588	1.611	1.65	93.000	102.000	
2022-395	3 ISE DILUTE	203 ROCHE	93 ROCHE		42	92.857	1.794	1.93	88.000	98.000	
2022-395	4 ISE UND-DI	165 VITROS	63 KODAK		83	97.000	1.667	1.72	92.000	102.000	
2022-395	4 ISE UND-DI	165 VITROS	156 Kit 32		72	97.167	1.374	1.41	92.000	102.000	
2022-395	4 ISE UND-DI	191 RX-SERIES	126 RANDOX		39	92.553	1.901	2.05	88.000	97.000	
2022-395	4 ISE UND-DI	191 RX-SERIES	126 RANDOX		39	94.125	2.571	2.73	89.000	99.000	
2022-395	8 IMT	219 ATELLICA	143 SIEMENS		10	99.900	11.614	11.63	95.000	105.000	
2022-395	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		78	93.532	1.420	1.52	89.000	98.000	
2022-395	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		13	92.538	1.278	1.38	88.000	97.000	
012	CHOLESTEROL, TOTAL	CLIA TEST ID #: 365		EVALUATION CRITERIA: TARGET VALUE 10 %							
2022-391	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		78	200.143	6.312	3.15	180.000	220.000	
2022-391	1 ENZYMATI	165 VITROS	63 KODAK		147	215.753	5.997	2.78	194.000	237.000	
2022-391	1 ENZYMATI	184 INTEGRA	93 ROCHE		17	207.059	7.742	3.74	186.000	228.000	
2022-391	1 ENZYMATI	191 RX-SERIES	126 RANDOX		45	220.467	12.256	5.56	198.000	243.000	
2022-391	1 ENZYMATI	191 RX-SERIES	126 RANDOX		45	217.000	4.517	2.08	195.000	239.000	
2022-391	1 ENZYMATI	197 ARCHITECT	1 ABBOTT		13	208.154	5.842	2.81	187.000	229.000	
2022-391	1 ENZYMATI	203 ROCHE	93 ROCHE		21	214.600	3.484	1.62	193.000	236.000	
2022-391	1 ENZYMATI	212 S. D. EXL	143 SIEMENS		13	196.846	5.021	2.55	177.000	217.000	
2022-391	1 ENZYMATI	220 Medica E.	130 MEDICA		10	212.900	9.751	4.58	192.000	234.000	
2022-391	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN		28	203.679	4.645	2.28	183.000	224.000	
2022-391	9 ENZYMATI	219 ATELLICA	143 SIEMENS		10	212.900	5.166	2.43	192.000	234.000	
2022-392	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS		78	120.013	4.015	3.35	108.000	132.000	
2022-392	1 ENZYMATI	165 VITROS	63 KODAK		147	116.303	3.685	3.17	105.000	128.000	
2022-392	1 ENZYMATI	184 INTEGRA	93 ROCHE		17	122.059	3.733	3.06	110.000	134.000	
2022-392	1 ENZYMATI	191 RX-SERIES	126 RANDOX		45	128.000	2.608	2.04	115.000	141.000	
2022-392	1 ENZYMATI	191 RX-SERIES	126 RANDOX		45	127.600	9.217	7.22	115.000	140.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-392	1 ENZYMATI	197 ARCHITECT	1	ABBOTT	13	123.923	3.100	2.50	112.000	136.000		
2022-392	1 ENZYMATI	203 ROCHE	93	ROCHE	21	125.450	2.539	2.02	113.000	138.000		
2022-392	1 ENZYMATI	212 S. D. EXL	143	SIEMENS	13	117.154	3.759	3.21	105.000	129.000		
2022-392	1 ENZYMATI	220 Medica E.	130	MEDICA	10	126.600	5.800	4.58	114.000	139.000		
2022-392	8 ENZYMATI	186 BECK. COUL	76	AU-BECKMAN	28	118.964	3.190	2.68	107.000	131.000		
2022-392	9 ENZYMATI	219 ATELLICA	143	SIEMENS	10	121.700	2.100	1.73	110.000	134.000		
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %									
2022-393	1 ENZYMATI	38 SIEMENS DI	143	SIEMENS	78	233.115	8.386	3.60	210.000	256.000		
2022-393	1 ENZYMATI	165 VITROS	63	KODAK	147	257.041	7.176	2.79	231.000	283.000		
2022-393	1 ENZYMATI	184 INTEGRA	93	ROCHE	17	240.294	6.875	2.86	216.000	264.000		
2022-393	1 ENZYMATI	191 RX-SERIES	126	RANDOX	45	255.977	12.527	4.89	230.000	282.000		
2022-393	1 ENZYMATI	191 RX-SERIES	126	RANDOX	45	256.800	12.937	5.04	231.000	282.000		
2022-393	1 ENZYMATI	197 ARCHITECT	1	ABBOTT	13	245.385	5.705	2.32	221.000	270.000		
2022-393	1 ENZYMATI	203 ROCHE	93	ROCHE	21	248.429	5.395	2.17	224.000	273.000		
2022-393	1 ENZYMATI	212 S. D. EXL	143	SIEMENS	13	231.077	4.463	1.93	208.000	254.000		
2022-393	1 ENZYMATI	220 Medica E.	130	MEDICA	10	249.000	3.578	1.44	224.000	274.000		
2022-393	8 ENZYMATI	186 BECK. COUL	76	AU-BECKMAN	28	236.964	7.033	2.97	213.000	261.000		
2022-393	9 ENZYMATI	219 ATELLICA	143	SIEMENS	10	247.700	6.214	2.51	223.000	272.000		
2022-394	1 ENZYMATI	38 SIEMENS DI	143	SIEMENS	78	149.553	4.913	3.29	135.000	165.000		
2022-394	1 ENZYMATI	165 VITROS	63	KODAK	147	152.699	4.670	3.06	137.000	168.000		
2022-394	1 ENZYMATI	184 INTEGRA	93	ROCHE	17	154.471	3.867	2.50	139.000	170.000		
2022-394	1 ENZYMATI	191 RX-SERIES	126	RANDOX	45	163.200	2.482	1.52	147.000	180.000		
2022-394	1 ENZYMATI	191 RX-SERIES	126	RANDOX	45	160.409	8.483	5.29	144.000	176.000		
2022-394	1 ENZYMATI	197 ARCHITECT	1	ABBOTT	13	155.154	3.939	2.54	140.000	171.000		
2022-394	1 ENZYMATI	203 ROCHE	93	ROCHE	21	157.900	3.048	1.93	142.000	174.000		
2022-394	1 ENZYMATI	212 S. D. EXL	143	SIEMENS	13	147.000	3.397	2.31	132.000	162.000		
2022-394	1 ENZYMATI	220 Medica E.	130	MEDICA	10	159.800	5.653	3.54	144.000	176.000		
2022-394	8 ENZYMATI	186 BECK. COUL	76	AU-BECKMAN	28	150.929	4.191	2.78	136.000	166.000		
2022-394	9 ENZYMATI	219 ATELLICA	143	SIEMENS	10	155.900	3.727	2.39	140.000	171.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH				
012	CHOLESTEROL, TOTAL		CLIA TEST ID #: 365		EVALUATION CRITERIA: TARGET VALUE	10 %							
2022-395	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	78	135.935	4.265	3.14	122.000	150.000				
2022-395	1 ENZYMATI	165 VITROS	63 KODAK	147	136.179	4.039	2.97	123.000	150.000				
2022-395	1 ENZYMATI	184 INTEGRA	93 ROCHE	17	139.353	3.709	2.66	125.000	153.000				
2022-395	1 ENZYMATI	191 RX-SERIES	126 RANDOX	45	146.023	7.944	5.44	131.000	161.000				
2022-395	1 ENZYMATI	191 RX-SERIES	126 RANDOX	45	144.600	6.216	4.30	130.000	159.000				
2022-395	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	13	141.769	3.166	2.23	128.000	156.000				
2022-395	1 ENZYMATI	203 ROCHE	93 ROCHE	21	143.600	3.308	2.30	129.000	158.000				
2022-395	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	13	133.154	4.016	3.02	120.000	146.000				
2022-395	1 ENZYMATI	220 Medica E.	130 MEDICA	10	143.500	5.971	4.16	129.000	158.000				
2022-395	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	28	136.556	3.247	2.38	123.000	150.000				
2022-395	9 ENZYMATI	219 ATELLICA	143 SIEMENS	10	145.600	15.711	10.79	131.000	160.000				
013	CHOLESTEROL, (HDL)		CLIA TEST ID #: 375		EVALUATION CRITERIA: TARGET VALUE	30 %							
2022-391	11 ENC.-COLO.	165 VITROS	63 KODAK	142	70.756	3.034	4.29	49.500	92.000				
2022-391	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	66.767	3.299	4.94	46.700	86.800				
2022-391	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	28	78.296	4.797	6.13	54.800	101.800				
2022-391	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	67.298	6.282	9.33	47.100	87.500				
2022-391	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	76.500	1.118	1.46	53.600	99.500				
2022-391	11 ENC.-COLO.	203 ROCHE	93 ROCHE	32	66.452	2.205	3.32	46.500	86.400				
2022-391	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS	10	65.160	3.496	5.37	45.600	84.700				
2022-391	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	15	74.273	2.851	3.84	52.000	96.600				
2022-391	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	76	64.045	2.737	4.27	44.800	83.300				
2022-391	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	13	64.923	2.401	3.70	45.400	84.400				
2022-392	11 ENC.-COLO.	165 VITROS	63 KODAK	142	44.080	2.791	6.33	30.900	57.300				
2022-392	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	33	48.422	2.516	5.20	33.900	62.900				
2022-392	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	28	45.793	3.475	7.59	32.100	59.500				
2022-392	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	47.250	2.487	5.26	33.100	61.400				
2022-392	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	41.339	4.938	11.95	28.900	53.700				
2022-392	11 ENC.-COLO.	203 ROCHE	93 ROCHE	32	48.065	1.840	3.83	33.600	62.500				

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-392	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS		10	37.910	2.309	6.09	26.500	49.300	
2022-392	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		15	45.113	1.917	4.25	31.600	58.600	
2022-392	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		76	52.231	2.214	4.24	36.600	67.900	
2022-392	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		13	52.692	2.671	5.07	36.900	68.500	
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2022-393	11 ENC.-COLO.	165 VITROS	63 KODAK		142	85.758	4.082	4.76	60.000	111.500	
2022-393	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	72.678	2.712	3.73	50.900	94.500	
2022-393	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		28	91.571	5.875	6.42	64.100	119.000	
2022-393	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX		42	77.502	7.454	9.62	54.300	100.800	
2022-393	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX		42	85.750	2.278	2.66	60.000	111.500	
2022-393	11 ENC.-COLO.	203 ROCHE	93 ROCHE		32	73.810	2.739	3.71	51.700	96.000	
2022-393	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS		10	75.890	4.014	5.29	53.100	98.700	
2022-393	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		15	86.980	3.851	4.43	60.900	113.100	
2022-393	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		76	68.380	3.016	4.41	47.900	88.900	
2022-393	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		13	70.000	3.419	4.88	49.000	91.000	
2022-394	11 ENC.-COLO.	165 VITROS	63 KODAK		142	53.279	2.880	5.40	37.300	69.300	
2022-394	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	55.516	2.146	3.86	38.900	72.200	
2022-394	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		28	57.325	3.376	5.89	40.100	74.500	
2022-394	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX		42	59.250	3.031	5.12	41.500	77.000	
2022-394	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX		42	50.893	4.833	9.50	35.600	66.200	
2022-394	11 ENC.-COLO.	203 ROCHE	93 ROCHE		32	55.166	2.520	4.57	38.600	71.700	
2022-394	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS		10	47.840	2.457	5.14	33.500	62.200	
2022-394	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT		15	55.573	2.656	4.78	38.900	72.200	
2022-394	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS		76	57.061	2.402	4.21	39.900	74.200	
2022-394	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS		13	57.846	2.713	4.69	40.500	75.200	
2022-395	11 ENC.-COLO.	165 VITROS	63 KODAK		142	49.031	2.798	5.71	34.300	63.700	
2022-395	11 ENC.-COLO.	184 INTEGRA	93 ROCHE		33	51.538	2.521	4.89	36.100	67.000	
2022-395	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN		28	51.300	3.382	6.59	35.900	66.700	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-395	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	46.032	4.127	8.97	32.200	59.800		
2022-395	11 ENC.-COLO.	191 RX-SERIES	126 RANDOX	42	54.250	3.192	5.88	38.000	70.500		
2022-395	11 ENC.-COLO.	203 ROCHE	93 ROCHE	32	51.747	1.929	3.73	36.200	67.300		
2022-395	11 ENC.-COLO.	219 ATELLICA	143 SIEMENS	10	45.340	6.231	13.74	31.700	58.900		
2022-395	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	15	50.900	2.162	4.25	35.600	66.200		
2022-395	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	76	54.872	2.423	4.42	38.400	71.300		
2022-395	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	13	55.538	2.590	4.66	38.900	72.200		
014	CREATINE KINASE (CK)	CLIA TEST ID #: 385	EVALUATION CRITERIA: TARGET VALUE 30 %								
2022-391	3 KINETIC 37	165 VITROS	63 KODAK	72	378.042	20.788	5.50	265.000	491.000		
2022-391	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	20	480.789	16.533	3.44	337.000	625.000		
2022-391	8 Hexokinase	184 INTEGRA	93 ROCHE	11	529.909	21.614	4.08	371.000	689.000		
2022-391	8 Hexokinase	203 ROCHE	93 ROCHE	11	529.818	14.572	2.75	371.000	689.000		
2022-391	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	28	524.179	11.436	2.18	367.000	681.000		
2022-392	3 KINETIC 37	165 VITROS	63 KODAK	72	34.775	3.229	9.29	24.000	45.000		
2022-392	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	20	30.737	1.742	5.67	22.000	40.000		
2022-392	8 Hexokinase	184 INTEGRA	93 ROCHE	11	38.455	2.105	5.47	27.000	50.000		
2022-392	8 Hexokinase	203 ROCHE	93 ROCHE	11	37.818	1.696	4.48	26.000	49.000		
2022-392	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	28	36.179	2.036	5.63	25.000	47.000		
2022-393	3 KINETIC 37	165 VITROS	63 KODAK	72	451.083	26.062	5.78	316.000	586.000		
2022-393	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	20	667.150	31.631	4.74	467.000	867.000		
2022-393	8 Hexokinase	184 INTEGRA	93 ROCHE	11	713.455	27.533	3.86	499.000	927.000		
2022-393	8 Hexokinase	203 ROCHE	93 ROCHE	11	720.455	25.994	3.61	504.000	937.000		
2022-393	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	28	719.429	14.908	2.07	504.000	935.000		
2022-394	3 KINETIC 37	165 VITROS	63 KODAK	72	191.563	11.013	5.75	134.000	249.000		
2022-394	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	20	199.368	6.627	3.32	140.000	259.000		
2022-394	8 Hexokinase	184 INTEGRA	93 ROCHE	11	220.545	6.228	2.82	154.000	287.000		
2022-394	8 Hexokinase	203 ROCHE	93 ROCHE	11	221.636	7.401	3.34	155.000	288.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-394	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	28	220.111	4.709	2.14	154.000	286.000		
014	CREATINE KINASE (CK)	CLIA TEST ID #: 385	EVALUATION CRITERIA: TARGET VALUE 30 %								
2022-395	3 KINETIC 37	165 VITROS	63 KODAK	72	128.181	7.947	6.20	90.000	167.000		
2022-395	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	20	128.842	4.603	3.57	90.000	167.000		
2022-395	8 Hexokinase	184 INTEGRA	93 ROCHE	11	144.909	5.071	3.50	101.000	188.000		
2022-395	8 Hexokinase	203 ROCHE	93 ROCHE	11	142.636	5.347	3.75	100.000	185.000		
2022-395	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	28	143.679	3.836	2.67	101.000	187.000		
016	CREATININE	CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}								
2022-391	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	4.486	0.390	8.70	3.810	5.160		
2022-391	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	4.496	0.205	4.57	3.820	5.170		
2022-391	4 ENZYMATIC	220 Medica E.	130 MEDICA	10	5.551	0.080	1.44	4.720	6.380		
2022-391	6 KIN. ALK P	184 INTEGRA	93 ROCHE	22	5.168	0.192	3.72	4.390	5.940		
2022-391	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	31	5.251	0.197	3.76	4.460	6.040		
2022-391	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	19	5.637	0.204	3.62	4.790	6.480		
2022-391	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	5.163	0.164	3.17	4.390	5.940		
2022-391	8 IDMS-CREA	165 VITROS	63 KODAK	153	4.590	0.169	3.67	3.900	5.280		
2022-391	8 IDMS-CREA	165 VITROS	63 KODAK	153	4.634	0.108	2.33	3.940	5.330		
2022-391	9 MODIFIED	219 ATELLICA	143 SIEMENS	10	5.151	0.207	4.01	4.380	5.920		
2022-391	11 IDMS Jaffe	38 SIEMENS DI	143 SIEMENS	74	5.493	0.178	3.24	4.670	6.320		
2022-392	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	0.949	0.085	8.96	0.650	1.250		
2022-392	2 AK. PIC. N	191 RX-SERIES	126 RANDOX	30	0.978	0.084	8.56	0.680	1.280		
2022-392	4 ENZYMATIC	220 Medica E.	130 MEDICA	10	0.992	0.037	3.74	0.690	1.290		
2022-392	6 KIN. ALK P	184 INTEGRA	93 ROCHE	22	0.965	0.044	4.57	0.660	1.260		
2022-392	6 KIN. ALK P	186 BECK. COUL	76 AU-BECKMAN	31	0.906	0.038	4.25	0.610	1.210		
2022-392	6 KIN. ALK P	197 ARCHITECT	1 ABBOTT	19	0.806	0.028	3.49	0.510	1.110		
2022-392	6 KIN. ALK P	203 ROCHE	93 ROCHE	23	0.902	0.046	5.15	0.600	1.200		
2022-392	8 IDMS-CREA	165 VITROS	63 KODAK	153	0.819	0.030	3.71	0.520	1.120		
2022-392	8 IDMS-CREA	165 VITROS	63 KODAK	153	0.820	0.040	4.88	0.520	1.120		
2022-392	9 MODIFIED	219 ATELLICA	143 SIEMENS	10	0.848	0.063	7.45	0.550	1.150		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-392	11 IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	74	0.878	0.059	6.71	0.580	1.180	
016 CREATININE		CLIA TEST ID #: 405				EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}						
2022-393	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	5.931	0.448	7.56	5.040	6.820	
2022-393	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	5.844	0.300	5.13	4.970	6.720	
2022-393	4 ENZYMATIC	220	Medica E.	130	MEDICA	10	7.517	0.420	5.59	6.390	8.640	
2022-393	6 KIN. ALK P	184	INTEGRA	93	ROCHE	22	6.598	0.318	4.82	5.610	7.590	
2022-393	6 KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	31	6.859	0.345	5.02	5.830	7.890	
2022-393	6 KIN. ALK P	197	ARCHITECT	1	ABBOTT	19	7.625	0.227	2.98	6.480	8.770	
2022-393	6 KIN. ALK P	203	ROCHE	93	ROCHE	23	6.741	0.202	3.00	5.730	7.750	
2022-393	8 IDMS-CREA	165	VITROS	63	KODAK	153	5.916	0.253	4.28	5.030	6.800	
2022-393	8 IDMS-CREA	165	VITROS	63	KODAK	153	5.949	0.146	2.46	5.060	6.840	
2022-393	9 MODIFIED	219	ATELLICA	143	SIEMENS	10	6.765	0.228	3.37	5.750	7.780	
2022-393	11 IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	74	7.353	0.220	3.00	6.250	8.460	
2022-394	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	2.260	0.162	7.19	1.920	2.600	
2022-394	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	2.383	0.286	12.00	2.030	2.740	
2022-394	4 ENZYMATIC	220	Medica E.	130	MEDICA	10	2.691	0.135	5.02	2.290	3.090	
2022-394	6 KIN. ALK P	184	INTEGRA	93	ROCHE	22	2.587	0.099	3.83	2.200	2.970	
2022-394	6 KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	31	2.518	0.111	4.42	2.140	2.900	
2022-394	6 KIN. ALK P	197	ARCHITECT	1	ABBOTT	19	2.581	0.098	3.79	2.190	2.970	
2022-394	6 KIN. ALK P	203	ROCHE	93	ROCHE	23	2.549	0.081	3.18	2.170	2.930	
2022-394	8 IDMS-CREA	165	VITROS	63	KODAK	153	2.288	0.061	2.69	1.940	2.630	
2022-394	8 IDMS-CREA	165	VITROS	63	KODAK	153	2.268	0.085	3.77	1.930	2.610	
2022-394	9 MODIFIED	219	ATELLICA	143	SIEMENS	10	2.424	0.102	4.23	2.060	2.790	
2022-394	11 IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	74	2.688	0.107	3.96	2.290	3.090	
2022-395	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	1.726	0.125	7.26	1.430	2.030	
2022-395	2 AK. PIC. N	191	RX-SERIES	126	RANDOX	30	1.726	0.127	7.35	1.430	2.030	
2022-395	4 ENZYMATIC	220	Medica E.	130	MEDICA	10	1.955	0.058	2.95	1.660	2.260	
2022-395	6 KIN. ALK P	184	INTEGRA	93	ROCHE	22	1.894	0.077	4.04	1.590	2.190	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-395	6 KIN. ALK P	186 BECK. COUL	76	AU-BECKMAN	31	1.832	0.068	3.73	1.530	2.130		
2022-395	6 KIN. ALK P	197 ARCHITECT	1	ABBOTT	19	1.829	0.084	4.58	1.530	2.130		
2022-395	6 KIN. ALK P	203 ROCHE	93	ROCHE	23	1.872	0.074	3.93	1.570	2.170		
2022-395	8 IDMS-CREA	165 VITROS	63	KODAK	153	1.664	0.082	4.95	1.360	1.960		
2022-395	8 IDMS-CREA	165 VITROS	63	KODAK	153	1.665	0.049	2.95	1.360	1.960		
2022-395	9 MODIFIED	219 ATELLICA	143	SIEMENS	10	1.827	0.269	14.74	1.530	2.130		
2022-395	11 IDMS Jaffe	38 SIEMENS DI	143	SIEMENS	74	1.945	0.089	4.55	1.650	2.250		
017 GLUCOSE		CLIA TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}									
2022-391	3 GL. OXI. C	165 VITROS	63	KODAK	163	134.086	3.236	2.41	121.000	147.000		
2022-391	3 GL. OXI. C	213 SELECTRA P	150	ELITECH	10	132.800	3.544	2.67	120.000	146.000		
2022-391	5 HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	135.390	3.498	2.58	122.000	149.000		
2022-391	5 HEXOKINASE	184 INTEGRA	93	ROCHE	11	131.273	4.371	3.33	118.000	144.000		
2022-391	5 HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	31	125.567	2.765	2.20	113.000	138.000		
2022-391	5 HEXOKINASE	191 RX-SERIES	126	RANDOX	36	135.371	5.571	4.12	122.000	149.000		
2022-391	5 HEXOKINASE	191 RX-SERIES	126	RANDOX	36	133.800	5.492	4.10	120.000	147.000		
2022-391	5 HEXOKINASE	203 ROCHE	93	ROCHE	25	130.400	3.124	2.40	117.000	143.000		
2022-391	5 HEXOKINASE	212 S. D. EXL	143	SIEMENS	13	134.154	3.371	2.51	121.000	148.000		
2022-391	12 HEXOKINASE	184 INTEGRA	93	ROCHE	25	132.720	4.565	3.44	119.000	146.000		
2022-391	12 HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	126.235	3.843	3.04	114.000	139.000		
2022-391	12 HEXOKINASE	203 ROCHE	93	ROCHE	16	131.625	2.421	1.84	118.000	145.000		
2022-391	12 HEXOKINASE	219 ATELLICA	143	SIEMENS	10	128.500	2.540	1.98	116.000	141.000		
2022-392	3 GL. OXI. C	165 VITROS	63	KODAK	163	32.253	1.085	3.36	26.000	38.000		
2022-392	3 GL. OXI. C	213 SELECTRA P	150	ELITECH	10	36.200	1.720	4.75	30.000	42.000		
2022-392	5 HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	37.074	1.624	4.38	31.000	43.000		
2022-392	5 HEXOKINASE	184 INTEGRA	93	ROCHE	11	34.545	1.233	3.57	29.000	41.000		
2022-392	5 HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	31	33.533	1.118	3.33	28.000	40.000		
2022-392	5 HEXOKINASE	191 RX-SERIES	126	RANDOX	36	40.800	3.763	9.22	35.000	47.000		
2022-392	5 HEXOKINASE	191 RX-SERIES	126	RANDOX	36	39.000	2.138	5.48	33.000	45.000		
2022-392	5 HEXOKINASE	203 ROCHE	93	ROCHE	25	34.708	0.978	2.82	29.000	41.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-392	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	13	35.923	3.407	9.48	30.000	42.000
2022-392	12	HEXOKINASE	184 INTEGRA	93	ROCHE	25	35.120	1.032	2.94	29.000	41.000
2022-392	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	32.471	1.242	3.83	26.000	38.000
2022-392	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	34.563	0.864	2.50	29.000	41.000
2022-392	12	HEXOKINASE	219 ATELLICA	143	SIEMENS	10	34.600	0.663	1.92	29.000	41.000
017	GLUCOSE	CLIA	TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}							
2022-393	3	GL. OXI. C	165 VITROS	63	KODAK	163	177.167	4.583	2.59	159.000	195.000
2022-393	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	10	169.000	8.012	4.74	152.000	186.000
2022-393	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	174.720	4.607	2.64	157.000	192.000
2022-393	5	HEXOKINASE	184 INTEGRA	93	ROCHE	11	168.818	5.557	3.29	152.000	186.000
2022-393	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	31	163.194	5.409	3.31	147.000	180.000
2022-393	5	HEXOKINASE	191 RX-SERIES	126	RANDOX	36	172.639	8.083	4.68	155.000	190.000
2022-393	5	HEXOKINASE	191 RX-SERIES	126	RANDOX	36	171.000	8.532	4.99	154.000	188.000
2022-393	5	HEXOKINASE	203 ROCHE	93	ROCHE	25	170.040	4.737	2.79	153.000	187.000
2022-393	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	13	175.538	4.668	2.66	158.000	193.000
2022-393	12	HEXOKINASE	184 INTEGRA	93	ROCHE	25	171.080	4.890	2.86	154.000	188.000
2022-393	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	17	164.412	3.866	2.35	148.000	181.000
2022-393	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	170.250	4.465	2.62	153.000	187.000
2022-393	12	HEXOKINASE	219 ATELLICA	143	SIEMENS	10	165.300	4.001	2.42	149.000	182.000
2022-394	3	GL. OXI. C	165 VITROS	63	KODAK	163	69.407	2.101	3.03	62.000	76.000
2022-394	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	10	72.900	2.737	3.75	66.000	80.000
2022-394	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	82	73.593	2.023	2.75	66.000	81.000
2022-394	5	HEXOKINASE	184 INTEGRA	93	ROCHE	11	70.091	2.712	3.87	63.000	77.000
2022-394	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	31	67.867	2.141	3.15	61.000	75.000
2022-394	5	HEXOKINASE	191 RX-SERIES	126	RANDOX	36	74.400	3.262	4.38	67.000	82.000
2022-394	5	HEXOKINASE	191 RX-SERIES	126	RANDOX	36	74.500	3.140	4.22	67.000	82.000
2022-394	5	HEXOKINASE	203 ROCHE	93	ROCHE	25	69.920	1.647	2.36	63.000	77.000
2022-394	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	13	72.385	3.175	4.39	65.000	80.000
2022-394	12	HEXOKINASE	184 INTEGRA	93	ROCHE	25	71.200	1.960	2.75	64.000	78.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-394	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	17	67.471	2.278	3.38	61.000	74.000
2022-394	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	69.563	1.694	2.44	63.000	77.000
2022-394	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	10	69.500	1.360	1.96	63.000	76.000
017	GLUCOSE		CLIA TEST ID #: 415			EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}						
2022-395	3	GL. OXI. C	165	VITROS	63	KODAK	163	51.733	1.740	3.36	46.000	58.000
2022-395	3	GL. OXI. C	213	SELECTRA P	150	ELITECH	10	57.700	2.759	4.78	52.000	64.000
2022-395	5	HEXOKINASE	38	SIEMENS DI	143	SIEMENS	82	57.829	1.820	3.15	52.000	64.000
2022-395	5	HEXOKINASE	184	INTEGRA	93	ROCHE	11	54.636	2.496	4.57	49.000	61.000
2022-395	5	HEXOKINASE	186	BECK. COUL	76	AU-BECKMAN	31	53.419	1.964	3.68	47.000	59.000
2022-395	5	HEXOKINASE	191	RX-SERIES	126	RANDOX	36	59.486	2.623	4.41	53.000	65.000
2022-395	5	HEXOKINASE	191	RX-SERIES	126	RANDOX	36	57.200	2.713	4.74	51.000	63.000
2022-395	5	HEXOKINASE	203	ROCHE	93	ROCHE	25	54.720	1.484	2.71	49.000	61.000
2022-395	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	13	57.583	1.441	2.50	52.000	64.000
2022-395	12	HEXOKINASE	184	INTEGRA	93	ROCHE	25	55.680	1.542	2.77	50.000	62.000
2022-395	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	17	52.294	1.563	2.99	46.000	58.000
2022-395	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	54.813	1.184	2.16	49.000	61.000
2022-395	12	HEXOKINASE	219	ATELLICA	143	SIEMENS	10	56.800	6.493	11.43	51.000	63.000
018	IRON, TOTAL		CLIA TEST ID #: 425			EVALUATION CRITERIA: TARGET VALUE 20 %						
2022-391	4	COULOMETR.	165	VITROS	63	KODAK	15	282.933	8.346	2.95	226.000	340.000
2022-392	4	COULOMETR.	165	VITROS	63	KODAK	15	41.600	2.245	5.40	33.000	50.000
2022-393	4	COULOMETR.	165	VITROS	63	KODAK	15	407.786	10.811	2.65	326.000	489.000
2022-394	4	COULOMETR.	165	VITROS	63	KODAK	15	122.267	3.820	3.12	98.000	147.000
2022-395	4	COULOMETR.	165	VITROS	63	KODAK	15	86.267	5.144	5.96	69.000	104.000
019	LACTASE DEHY. (LDH)		CLIA TEST ID #: 435									
2022-391	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	20	457.850	37.470	8.18	366.000	549.000
2022-391	6	NADH	184	INTEGRA	93	ROCHE	12	556.250	31.105	5.59	445.000	668.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2022-391	6	NADH	203	ROCHE	93	ROCHE	21	556.143	18.586	3.34	445.000	667.000	
2022-391	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	31	533.935	15.641	2.93	427.000	641.000	
2022-391	9	LDH1	165	VITROS	63	KODAK	67	642.851	31.553	4.91	514.000	771.000	
019 LACTASE DEHY. (LDH)			CLIA TEST ID #: 435			EVALUATION CRITERIA: TARGET VALUE 20 %							
2022-392	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	20	54.300	5.283	9.73	43.000	65.000	
2022-392	6	NADH	184	INTEGRA	93	ROCHE	12	69.000	2.483	3.60	55.000	83.000	
2022-392	6	NADH	203	ROCHE	93	ROCHE	21	67.143	2.883	4.29	54.000	81.000	
2022-392	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	31	61.516	5.254	8.54	49.000	74.000	
2022-392	9	LDH1	165	VITROS	63	KODAK	67	76.076	3.859	5.07	61.000	91.000	
2022-393	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	20	620.000	52.042	8.39	496.000	744.000	
2022-393	6	NADH	184	INTEGRA	93	ROCHE	12	725.583	35.561	4.90	580.000	871.000	
2022-393	6	NADH	203	ROCHE	93	ROCHE	21	745.714	21.673	2.91	597.000	895.000	
2022-393	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	31	726.516	24.644	3.39	581.000	872.000	
2022-393	9	LDH1	165	VITROS	63	KODAK	67	829.591	40.889	4.93	664.000	996.000	
2022-394	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	20	201.700	16.662	8.26	161.000	242.000	
2022-394	6	NADH	184	INTEGRA	93	ROCHE	12	252.417	12.305	4.87	202.000	303.000	
2022-394	6	NADH	203	ROCHE	93	ROCHE	21	248.524	7.287	2.93	199.000	298.000	
2022-394	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	31	233.774	9.970	4.26	187.000	281.000	
2022-394	9	LDH1	165	VITROS	63	KODAK	67	291.955	7.343	2.52	234.000	350.000	
2022-395	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	20	139.100	11.755	8.45	111.000	167.000	
2022-395	6	NADH	184	INTEGRA	93	ROCHE	12	175.750	6.584	3.75	141.000	211.000	
2022-395	6	NADH	203	ROCHE	93	ROCHE	21	172.429	5.020	2.91	138.000	207.000	
2022-395	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	31	162.258	7.304	4.50	130.000	195.000	
2022-395	9	LDH1	165	VITROS	63	KODAK	67	204.273	5.470	2.68	163.000	245.000	
021 MAGNESIUM			CLIA TEST ID #: 455			EVALUATION CRITERIA: TARGET VALUE 25 %							
2022-391	5	COLOR./K E	165	VITROS	63	KODAK	25	4.832	0.101	2.09	3.600	6.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
021	MAGNESIUM	CLIA TEST ID #: 455				EVALUATION CRITERIA: TARGET VALUE 25 %						
2022-392	5 COLOR./K E	165	VITROS	63	KODAK	25	1.296	0.072	5.56	1.000	1.600	
2022-393	5 COLOR./K E	165	VITROS	63	KODAK	25	6.100	0.129	2.12	4.600	7.600	
2022-394	5 COLOR./K E	165	VITROS	63	KODAK	25	2.628	0.072	2.75	2.000	3.300	
2022-395	5 COLOR./K E	165	VITROS	63	KODAK	25	2.060	0.069	3.36	1.500	2.600	
022	POTASSIUM	CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00						
2022-391	3 ISE DILUT.	184	INTEGRA	93	ROCHE	36	5.217	0.103	1.97	4.200	6.200	
2022-391	3 ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	32	5.081	0.098	1.93	4.100	6.100	
2022-391	3 ISE DILUT.	197	ARCHITECT	1	ABBOTT	17	5.171	0.123	2.37	4.200	6.200	
2022-391	3 ISE DILUT.	203	ROCHE	93	ROCHE	43	5.230	0.095	1.82	4.200	6.200	
2022-391	4 ISE UNDIL.	165	VITROS	63	KODAK	85	5.195	0.094	1.82	4.200	6.200	
2022-391	4 ISE UNDIL.	165	VITROS	156	Kit 32	72	5.194	0.091	1.75	4.200	6.200	
2022-391	4 ISE UNDIL.	191	RX-SERIES	126	RANDOX	38	5.121	0.138	2.69	4.100	6.100	
2022-391	4 ISE UNDIL.	220	Medica E.	130	MEDICA	10	5.070	0.078	1.54	4.100	6.100	
2022-391	5 IMT	219	ATELLICA	143	SIEMENS	10	5.020	0.087	1.74	4.000	6.000	
2022-391	6 QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	5.144	0.090	1.76	4.100	6.100	
2022-391	6 QUICK LYTE	212	S. D. EXL	143	SIEMENS	13	5.131	0.091	1.77	4.100	6.100	
2022-392	3 ISE DILUT.	184	INTEGRA	93	ROCHE	36	2.100	0.053	2.51	1.100	3.100	
2022-392	3 ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	32	2.148	0.067	3.10	1.100	3.100	
2022-392	3 ISE DILUT.	197	ARCHITECT	1	ABBOTT	17	2.094	0.054	2.57	1.100	3.100	
2022-392	3 ISE DILUT.	203	ROCHE	93	ROCHE	43	2.151	0.080	3.72	1.200	3.200	
2022-392	4 ISE UNDIL.	165	VITROS	63	KODAK	85	2.089	0.041	1.96	1.100	3.100	
2022-392	4 ISE UNDIL.	165	VITROS	156	Kit 32	72	2.085	0.043	2.06	1.100	3.100	
2022-392	4 ISE UNDIL.	191	RX-SERIES	126	RANDOX	38	2.149	0.100	4.67	1.100	3.100	
2022-392	4 ISE UNDIL.	220	Medica E.	130	MEDICA	10	2.010	0.054	2.68	1.000	3.000	
2022-392	5 IMT	219	ATELLICA	143	SIEMENS	10	2.030	0.046	2.26	1.000	3.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-392	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	78	2.001	0.047	2.33	1.000	3.000	
2022-392	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	13	1.992	0.073	3.66	1.000	3.000	
022 POTASSIUM		CLIA TEST ID #: 465	EVALUATION CRITERIA: TARGET VALUE 1.00								
2022-393	3 ISE DILUT.	184 INTEGRA	93	ROCHE	36	6.464	0.113	1.75	5.500	7.500	
2022-393	3 ISE DILUT.	186 BECK. COUL	76	AU-BECKMAN	32	6.335	0.115	1.82	5.300	7.300	
2022-393	3 ISE DILUT.	197 ARCHITECT	1	ABBOTT	17	6.494	0.121	1.87	5.500	7.500	
2022-393	3 ISE DILUT.	203 ROCHE	93	ROCHE	43	6.488	0.103	1.58	5.500	7.500	
2022-393	4 ISE UNLUT.	165 VITROS	63	KODAK	85	6.546	0.123	1.88	5.500	7.500	
2022-393	4 ISE UNLUT.	165 VITROS	156	Kit 32	72	6.514	0.131	2.02	5.500	7.500	
2022-393	4 ISE UNLUT.	191 RX-SERIES	126	RANDOX	38	6.455	0.226	3.49	5.500	7.500	
2022-393	4 ISE UNLUT.	220 Medica E.	130	MEDICA	10	6.450	0.150	2.33	5.500	7.500	
2022-393	5 IMT	219 ATELLICA	143	SIEMENS	10	6.260	0.156	2.50	5.300	7.300	
2022-393	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	78	6.433	0.097	1.50	5.400	7.400	
2022-393	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	13	6.431	0.126	1.97	5.400	7.400	
2022-394	3 ISE DILUT.	184 INTEGRA	93	ROCHE	36	3.231	0.066	2.04	2.200	4.200	
2022-394	3 ISE DILUT.	186 BECK. COUL	76	AU-BECKMAN	32	3.222	0.074	2.29	2.200	4.200	
2022-394	3 ISE DILUT.	197 ARCHITECT	1	ABBOTT	17	3.224	0.073	2.26	2.200	4.200	
2022-394	3 ISE DILUT.	203 ROCHE	93	ROCHE	43	3.260	0.076	2.32	2.300	4.300	
2022-394	4 ISE UNLUT.	165 VITROS	63	KODAK	85	3.217	0.064	1.98	2.200	4.200	
2022-394	4 ISE UNLUT.	165 VITROS	156	Kit 32	72	3.225	0.062	1.93	2.200	4.200	
2022-394	4 ISE UNLUT.	191 RX-SERIES	126	RANDOX	38	3.230	0.087	2.68	2.200	4.200	
2022-394	4 ISE UNLUT.	220 Medica E.	130	MEDICA	10	3.140	0.049	1.56	2.100	4.100	
2022-394	5 IMT	219 ATELLICA	143	SIEMENS	10	3.140	0.066	2.11	2.100	4.100	
2022-394	6 QUICK LYTE	38 SIEMENS DI	143	SIEMENS	78	3.169	0.063	1.99	2.200	4.200	
2022-394	6 QUICK LYTE	212 S. D. EXL	143	SIEMENS	13	3.154	0.084	2.67	2.200	4.200	
2022-395	3 ISE DILUT.	184 INTEGRA	93	ROCHE	36	2.731	0.062	2.25	1.700	3.700	
2022-395	3 ISE DILUT.	186 BECK. COUL	76	AU-BECKMAN	32	2.731	0.068	2.50	1.700	3.700	
2022-395	3 ISE DILUT.	197 ARCHITECT	1	ABBOTT	17	2.735	0.068	2.49	1.700	3.700	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-395	3 ISE DILUT.	203 ROCHE	93 ROCHE		43	2.788	0.085	3.05	1.800	3.800	
2022-395	4 ISE UNDIL.	165 VITROS	63 KODAK		85	2.735	0.061	2.22	1.700	3.700	
2022-395	4 ISE UNDIL.	165 VITROS	156 Kit 32		72	2.721	0.053	1.94	1.700	3.700	
2022-395	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX		38	2.732	0.107	3.90	1.700	3.700	
2022-395	4 ISE UNDIL.	220 Medica E.	130 MEDICA		10	2.660	0.049	1.84	1.700	3.700	
2022-395	5 IMT	219 ATELLICA	143 SIEMENS		10	2.730	0.300	11.00	1.700	3.700	
2022-395	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		78	2.659	0.063	2.37	1.700	3.700	
2022-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		13	2.646	0.063	2.40	1.600	3.600	
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2022-391	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	151.813	2.214	1.46	148.000	156.000	
2022-391	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		32	146.600	1.497	1.02	143.000	151.000	
2022-391	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		32	151.156	2.399	1.59	147.000	155.000	
2022-391	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		11	154.182	2.037	1.32	150.000	158.000	
2022-391	3 ISE DILUTE	203 ROCHE	93 ROCHE		37	152.143	2.997	1.97	148.000	156.000	
2022-391	3 ISE DILUTE	203 ROCHE	93 ROCHE		37	153.000	2.325	1.52	149.000	157.000	
2022-391	4 ISE UNDIL.	165 VITROS	63 KODAK		79	154.308	2.108	1.37	150.000	158.000	
2022-391	4 ISE UNDIL.	165 VITROS	63 KODAK		79	154.000	15.000	9.74	150.000	158.000	
2022-391	4 ISE UNDIL.	165 VITROS	156 Kit 32		73	154.861	1.734	1.12	151.000	159.000	
2022-391	4 ISE UNDIL.	165 VITROS	156 Kit 32		73	154.000	15.000	9.74	150.000	158.000	
2022-391	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX		37	153.946	2.977	1.93	150.000	158.000	
2022-391	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX		37	155.667	2.625	1.69	152.000	160.000	
2022-391	4 ISE UNDIL.	220 Medica E.	130 MEDICA		10	152.700	1.187	0.78	149.000	157.000	
2022-391	5 IMT	197 ARCHITECT	1 ABBOTT		12	152.333	3.171	2.08	148.000	156.000	
2022-391	5 IMT	219 ATELLICA	143 SIEMENS		10	150.700	3.607	2.39	147.000	155.000	
2022-391	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	151.897	2.152	1.42	148.000	156.000	
2022-391	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		13	151.462	2.406	1.59	147.000	155.000	
2022-392	3 ISE DILUTE	184 INTEGRA	93 ROCHE		33	122.788	2.026	1.65	119.000	127.000	
2022-392	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		32	123.250	2.031	1.65	119.000	127.000	
2022-392	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		32	127.400	1.020	0.80	123.000	131.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-392	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	11	123.727	1.543	1.25	120.000	128.000
2022-392	3	ISE DILUTE	203 ROCHE	93	ROCHE	37	124.333	2.449	1.97	120.000	128.000
2022-392	3	ISE DILUTE	203 ROCHE	93	ROCHE	37	122.571	1.498	1.22	119.000	127.000
2022-392	4	ISE UNDIL.	165 VITROS	63	KODAK	79	121.000	12.000	9.92	117.000	125.000
2022-392	4	ISE UNDIL.	165 VITROS	63	KODAK	79	121.744	1.870	1.54	118.000	126.000
2022-392	4	ISE UNDIL.	165 VITROS	156	Kit 32	73	121.000	12.000	9.92	117.000	125.000
2022-392	4	ISE UNDIL.	165 VITROS	156	Kit 32	73	122.082	1.686	1.38	118.000	126.000
2022-392	4	ISE UNDIL.	191 RX-SERIES	126	RANDOX	37	123.389	2.264	1.84	119.000	127.000
2022-392	4	ISE UNDIL.	191 RX-SERIES	126	RANDOX	37	124.333	0.471	0.38	120.000	128.000
2022-392	4	ISE UNDIL.	220 Medica E.	130	MEDICA	10	122.800	1.400	1.14	119.000	127.000
2022-392	5	IMT	197 ARCHITECT	1	ABBOTT	12	122.917	2.253	1.83	119.000	127.000
2022-392	5	IMT	219 ATELLICA	143	SIEMENS	10	123.600	1.685	1.36	120.000	128.000
2022-392	6	QUICK LYTE	38 SIEMENS DI	143	SIEMENS	79	124.885	1.769	1.42	121.000	129.000
2022-392	6	QUICK LYTE	212 S. D. EXL	143	SIEMENS	13	123.846	2.178	1.76	120.000	128.000
023 SODIUM			CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00						
2022-393	3	ISE DILUTE	184 INTEGRA	93	ROCHE	33	162.909	2.353	1.44	159.000	167.000
2022-393	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	32	157.000	1.095	0.70	153.000	161.000
2022-393	3	ISE DILUTE	186 BECK. COUL	76	AU-BECKMAN	32	163.469	2.947	1.80	159.000	167.000
2022-393	3	ISE DILUTE	197 ARCHITECT	1	ABBOTT	11	166.545	1.777	1.07	163.000	171.000
2022-393	3	ISE DILUTE	203 ROCHE	93	ROCHE	37	165.189	2.513	1.52	161.000	169.000
2022-393	3	ISE DILUTE	203 ROCHE	93	ROCHE	37	165.286	1.666	1.01	161.000	169.000
2022-393	4	ISE UNDIL.	165 VITROS	63	KODAK	79	163.000	16.000	9.82	159.000	167.000
2022-393	4	ISE UNDIL.	165 VITROS	63	KODAK	79	170.013	2.310	1.36	166.000	174.000
2022-393	4	ISE UNDIL.	165 VITROS	156	Kit 32	73	169.472	2.357	1.39	165.000	173.000
2022-393	4	ISE UNDIL.	165 VITROS	156	Kit 32	73	163.000	16.000	9.82	159.000	167.000
2022-393	4	ISE UNDIL.	191 RX-SERIES	126	RANDOX	37	166.459	4.214	2.53	162.000	170.000
2022-393	4	ISE UNDIL.	191 RX-SERIES	126	RANDOX	37	169.000	0.816	0.48	165.000	173.000
2022-393	4	ISE UNDIL.	220 Medica E.	130	MEDICA	10	167.200	2.227	1.33	163.000	171.000
2022-393	5	IMT	197 ARCHITECT	1	ABBOTT	12	166.167	2.075	1.25	162.000	170.000
2022-393	5	IMT	219 ATELLICA	143	SIEMENS	10	160.600	3.747	2.33	157.000	165.000
2022-393	6	QUICK LYTE	38 SIEMENS DI	143	SIEMENS	79	163.367	2.522	1.54	159.000	167.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-393	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	13	163.615	2.897	1.77	160.000	168.000
023	SODIUM		CLIA TEST ID #: 475			EVALUATION CRITERIA: TARGET VALUE 4.00						
2022-394	3	ISE DILUTE	184	INTEGRA	93	ROCHE	33	134.182	2.153	1.60	130.000	138.000
2022-394	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	32	134.656	2.300	1.71	131.000	139.000
2022-394	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	32	134.600	0.800	0.59	131.000	139.000
2022-394	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	11	135.727	1.710	1.26	132.000	140.000
2022-394	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	134.429	2.441	1.82	130.000	138.000
2022-394	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	135.162	2.034	1.50	131.000	139.000
2022-394	4	ISE UNDIL.	165	VITROS	63	KODAK	79	134.000	13.000	9.70	130.000	138.000
2022-394	4	ISE UNDIL.	165	VITROS	63	KODAK	79	134.615	2.260	1.68	131.000	139.000
2022-394	4	ISE UNDIL.	165	VITROS	156	Kit 32	73	134.000	13.000	9.70	130.000	138.000
2022-394	4	ISE UNDIL.	165	VITROS	156	Kit 32	73	134.699	1.811	1.34	131.000	139.000
2022-394	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	37	134.472	2.115	1.57	130.000	138.000
2022-394	4	ISE UNDIL.	191	RX-SERIES	126	RANDOX	37	136.000	0.816	0.60	132.000	140.000
2022-394	4	ISE UNDIL.	220	Medica E.	130	MEDICA	10	135.200	1.720	1.27	131.000	139.000
2022-394	5	IMT	197	ARCHITECT	1	ABBOTT	12	134.917	1.605	1.19	131.000	139.000
2022-394	5	IMT	219	APELLICA	143	SIEMENS	10	134.700	3.035	2.25	131.000	139.000
2022-394	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	136.077	2.117	1.56	132.000	140.000
2022-394	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	13	135.154	2.032	1.50	131.000	139.000
2022-395	3	ISE DILUTE	184	INTEGRA	93	ROCHE	33	128.182	1.866	1.46	124.000	132.000
2022-395	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	32	130.200	0.980	0.75	126.000	134.000
2022-395	3	ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	32	128.688	1.810	1.41	125.000	133.000
2022-395	3	ISE DILUTE	197	ARCHITECT	1	ABBOTT	11	129.727	1.420	1.09	126.000	134.000
2022-395	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	128.571	1.294	1.01	125.000	133.000
2022-395	3	ISE DILUTE	203	ROCHE	93	ROCHE	37	129.703	2.012	1.55	126.000	134.000
2022-395	4	ISE UNDIL.	165	VITROS	63	KODAK	79	128.551	2.055	1.60	125.000	133.000
2022-395	4	ISE UNDIL.	165	VITROS	63	KODAK	79	129.000	12.000	9.30	125.000	133.000
2022-395	4	ISE UNDIL.	165	VITROS	156	Kit 32	73	128.781	1.988	1.54	125.000	133.000
2022-395	4	ISE UNDIL.	165	VITROS	156	Kit 32	73	129.000	12.000	9.30	125.000	133.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-395	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX		37	128.622	2.685	2.09	125.000	133.000	
2022-395	4 ISE UNDIL.	191 RX-SERIES	126 RANDOX		37	129.333	0.471	0.36	125.000	133.000	
2022-395	4 ISE UNDIL.	220 Medica E.	130 MEDICA		10	129.600	2.375	1.83	126.000	134.000	
2022-395	5 IMT	197 ARCHITECT	1 ABBOTT		12	129.833	1.908	1.47	126.000	134.000	
2022-395	5 IMT	219 ATELLICA	143 SIEMENS		10	134.000	13.183	9.84	130.000	138.000	
2022-395	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	130.308	1.705	1.31	126.000	134.000	
2022-395	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		13	129.231	1.887	1.46	125.000	133.000	
024	TOTAL PROTEIN	CLIA TEST ID #: 485	EVALUATION CRITERIA: TARGET VALUE 10 %								
2022-391	1 BIURET	38 SIEMENS DI	143 SIEMENS		80	7.120	0.177	2.49	6.400	7.800	
2022-391	1 BIURET	165 VITROS	63 KODAK		143	6.605	0.160	2.42	5.900	7.300	
2022-391	1 BIURET	165 VITROS	63 KODAK		143	6.563	0.245	3.73	5.900	7.200	
2022-391	1 BIURET	184 INTEGRA	93 ROCHE		25	6.648	0.276	4.15	6.000	7.300	
2022-391	1 BIURET	186 BECK. COUL	76 AU-BECKMAN		27	6.752	0.162	2.40	6.100	7.400	
2022-391	1 BIURET	191 RX-SERIES	126 RANDOX		38	6.942	0.229	3.30	6.200	7.600	
2022-391	1 BIURET	197 ARCHITECT	1 ABBOTT		20	6.875	0.137	2.00	6.200	7.600	
2022-391	1 BIURET	203 ROCHE	93 ROCHE		33	6.784	0.168	2.47	6.100	7.500	
2022-391	1 BIURET	212 S. D. EXL	143 SIEMENS		13	7.115	0.151	2.12	6.400	7.800	
2022-391	1 BIURET	219 ATELLICA	143 SIEMENS		10	6.880	0.223	3.24	6.200	7.600	
2022-392	1 BIURET	38 SIEMENS DI	143 SIEMENS		80	3.204	0.101	3.16	2.900	3.500	
2022-392	1 BIURET	165 VITROS	63 KODAK		143	2.963	0.111	3.75	2.700	3.300	
2022-392	1 BIURET	165 VITROS	63 KODAK		143	3.139	0.099	3.16	2.800	3.500	
2022-392	1 BIURET	184 INTEGRA	93 ROCHE		25	2.948	0.106	3.61	2.700	3.200	
2022-392	1 BIURET	186 BECK. COUL	76 AU-BECKMAN		27	3.026	0.084	2.79	2.700	3.300	
2022-392	1 BIURET	191 RX-SERIES	126 RANDOX		38	3.121	0.130	4.17	2.800	3.400	
2022-392	1 BIURET	197 ARCHITECT	1 ABBOTT		20	2.955	0.074	2.50	2.700	3.300	
2022-392	1 BIURET	203 ROCHE	93 ROCHE		33	3.097	0.116	3.74	2.800	3.400	
2022-392	1 BIURET	212 S. D. EXL	143 SIEMENS		13	3.162	0.115	3.63	2.800	3.500	
2022-392	1 BIURET	219 ATELLICA	143 SIEMENS		10	3.000	0.100	3.33	2.700	3.300	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
024 TOTAL PROTEIN		CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2022-393	1	BIURET	38 SIEMENS DI	143 SIEMENS	80	8.734	0.249	2.86	7.900	9.600	
2022-393	1	BIURET	165 VITROS	63 KODAK	143	8.156	0.219	2.68	7.300	9.000	
2022-393	1	BIURET	165 VITROS	63 KODAK	143	7.925	0.228	2.87	7.100	8.700	
2022-393	1	BIURET	184 INTEGRA	93 ROCHE	25	8.016	0.260	3.25	7.200	8.800	
2022-393	1	BIURET	186 BECK. COUL	76 AU-BECKMAN	27	8.244	0.248	3.01	7.400	9.100	
2022-393	1	BIURET	191 RX-SERIES	126 RANDOX	38	8.482	0.376	4.44	7.600	9.300	
2022-393	1	BIURET	197 ARCHITECT	1 ABBOTT	20	8.530	0.119	1.39	7.700	9.400	
2022-393	1	BIURET	203 ROCHE	93 ROCHE	33	8.248	0.254	3.07	7.400	9.100	
2022-393	1	BIURET	212 S. D. EXL	143 SIEMENS	13	8.831	0.201	2.28	7.900	9.700	
2022-393	1	BIURET	219 ATELLICA	143 SIEMENS	10	8.340	0.250	3.00	7.500	9.200	
2022-394	1	BIURET	38 SIEMENS DI	143 SIEMENS	80	4.673	0.134	2.86	4.200	5.100	
2022-394	1	BIURET	165 VITROS	63 KODAK	143	4.300	0.132	3.08	3.900	4.700	
2022-394	1	BIURET	165 VITROS	63 KODAK	143	4.466	0.117	2.63	4.000	4.900	
2022-394	1	BIURET	184 INTEGRA	93 ROCHE	25	4.336	0.135	3.12	3.900	4.800	
2022-394	1	BIURET	186 BECK. COUL	76 AU-BECKMAN	27	4.412	0.097	2.21	4.000	4.900	
2022-394	1	BIURET	191 RX-SERIES	126 RANDOX	38	4.535	0.123	2.72	4.100	5.000	
2022-394	1	BIURET	197 ARCHITECT	1 ABBOTT	20	4.385	0.085	1.95	3.900	4.800	
2022-394	1	BIURET	203 ROCHE	93 ROCHE	33	4.455	0.154	3.46	4.000	4.900	
2022-394	1	BIURET	212 S. D. EXL	143 SIEMENS	13	4.662	0.100	2.15	4.200	5.100	
2022-394	1	BIURET	219 ATELLICA	143 SIEMENS	10	4.430	0.142	3.20	4.000	4.900	
2022-395	1	BIURET	38 SIEMENS DI	143 SIEMENS	80	4.032	0.116	2.89	3.600	4.400	
2022-395	1	BIURET	165 VITROS	63 KODAK	143	3.908	0.108	2.76	3.500	4.300	
2022-395	1	BIURET	165 VITROS	63 KODAK	143	3.700	0.150	4.05	3.300	4.100	
2022-395	1	BIURET	184 INTEGRA	93 ROCHE	25	3.728	0.122	3.27	3.400	4.100	
2022-395	1	BIURET	186 BECK. COUL	76 AU-BECKMAN	27	3.812	0.070	1.83	3.400	4.200	
2022-395	1	BIURET	191 RX-SERIES	126 RANDOX	38	3.947	0.139	3.52	3.600	4.300	
2022-395	1	BIURET	197 ARCHITECT	1 ABBOTT	20	3.775	0.077	2.03	3.400	4.200	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2022-395	1	BIURET	203	ROCHE	93	ROCHE	33	3.852	0.105	2.72	3.500	4.200	
2022-395	1	BIURET	212	S. D. EXL	143	SIEMENS	13	4.000	0.104	2.59	3.600	4.400	
2022-395	1	BIURET	219	ATELLICA	143	SIEMENS	10	3.940	0.474	12.02	3.500	4.300	
025 TRIGLYCERIDES		CLIA TEST ID #: 495		EVALUATION CRITERIA: TARGET VALUE 25 %									
2022-391	2	ENZ. COLO.	165	VITROS	63	KODAK	147	232.884	8.870	3.81	175.000	291.000	
2022-391	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	12	201.333	4.625	2.30	151.000	252.000	
2022-391	2	ENZ. COLO.	203	ROCHE	93	ROCHE	20	197.300	4.766	2.42	148.000	247.000	
2022-391	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	76	194.093	4.173	2.15	146.000	243.000	
2022-391	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	14	192.714	3.514	1.82	145.000	241.000	
2022-391	5	HANTZCH-C)	197	ARCHITECT	1	ABBOTT	10	191.300	6.649	3.48	143.000	239.000	
2022-391	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	202.929	6.475	3.19	152.000	254.000	
2022-391	9	GPO TRIND.	191	RX-SERIES	126	RANDOX	32	198.161	9.426	4.76	149.000	248.000	
2022-391	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	22	200.773	4.327	2.16	151.000	251.000	
2022-391	10	ENZYM. GPO	203	ROCHE	93	ROCHE	21	197.900	3.491	1.76	148.000	247.000	
2022-391	10	ENZYM. GPO	219	ATELLICA	143	SIEMENS	10	204.300	7.616	3.73	153.000	255.000	
2022-391	11	GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	191.364	7.690	4.02	144.000	239.000	
2022-392	2	ENZ. COLO.	165	VITROS	63	KODAK	147	96.041	3.418	3.56	72.000	120.000	
2022-392	2	ENZ. COLO.	184	INTEGRA	93	ROCHE	12	92.250	1.831	1.99	69.000	115.000	
2022-392	2	ENZ. COLO.	203	ROCHE	93	ROCHE	20	90.789	1.852	2.04	68.000	113.000	
2022-392	4	ENZ. (GPD)	38	SIEMENS DI	143	SIEMENS	76	80.338	2.632	3.28	60.000	100.000	
2022-392	4	ENZ. (GPD)	212	S. D. EXL	143	SIEMENS	14	79.071	2.658	3.36	59.000	99.000	
2022-392	5	HANTZCH-C)	197	ARCHITECT	1	ABBOTT	10	82.600	2.154	2.61	62.000	103.000	
2022-392	9	GPO TRIND.	186	BECK. COUL	76	AU-BECKMAN	29	88.536	3.168	3.58	66.000	111.000	
2022-392	9	GPO TRIND.	191	RX-SERIES	126	RANDOX	32	88.452	4.507	5.10	66.000	111.000	
2022-392	10	ENZYM. GPO	184	INTEGRA	93	ROCHE	22	91.762	1.849	2.02	69.000	115.000	
2022-392	10	ENZYM. GPO	203	ROCHE	93	ROCHE	21	90.450	2.559	2.83	68.000	113.000	
2022-392	10	ENZYM. GPO	219	ATELLICA	143	SIEMENS	10	89.000	5.586	6.28	67.000	111.000	
2022-392	11	GLY PO4 OX	197	ARCHITECT	1	ABBOTT	11	82.000	3.464	4.22	62.000	103.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
025 TRIGLYCERIDES		CLIA TEST ID #: 495	EVALUATION CRITERIA: TARGET VALUE 25 %								
2022-393	2 ENZ. COLO.	165 VITROS	63 KODAK		147	291.228	11.283	3.87	218.000	364.000	
2022-393	2 ENZ. COLO.	184 INTEGRA	93 ROCHE		12	240.833	3.869	1.61	181.000	301.000	
2022-393	2 ENZ. COLO.	203 ROCHE	93 ROCHE		20	237.950	6.265	2.63	178.000	297.000	
2022-393	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		76	239.413	5.926	2.48	180.000	299.000	
2022-393	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		14	240.857	4.703	1.95	181.000	301.000	
2022-393	5 HANTZCH-C)	197 ARCHITECT	1 ABBOTT		10	236.300	5.622	2.38	177.000	295.000	
2022-393	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		29	248.571	9.649	3.88	186.000	311.000	
2022-393	9 GPO TRIND.	191 RX-SERIES	126 RANDOX		32	237.677	8.322	3.50	178.000	297.000	
2022-393	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		22	238.333	5.008	2.10	179.000	298.000	
2022-393	10 ENZYM. GPO	203 ROCHE	93 ROCHE		21	238.381	3.823	1.60	179.000	298.000	
2022-393	10 ENZYM. GPO	219 ATELLICA	143 SIEMENS		10	249.200	8.183	3.28	187.000	312.000	
2022-393	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		11	232.818	10.564	4.54	175.000	291.000	
2022-394	2 ENZ. COLO.	165 VITROS	63 KODAK		147	146.925	4.689	3.19	110.000	184.000	
2022-394	2 ENZ. COLO.	184 INTEGRA	93 ROCHE		12	134.583	2.290	1.70	101.000	168.000	
2022-394	2 ENZ. COLO.	203 ROCHE	93 ROCHE		20	131.800	2.839	2.15	99.000	165.000	
2022-394	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		76	123.716	3.078	2.49	93.000	155.000	
2022-394	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		14	121.786	3.004	2.47	91.000	152.000	
2022-394	5 HANTZCH-C)	197 ARCHITECT	1 ABBOTT		10	123.900	3.113	2.51	93.000	155.000	
2022-394	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		29	131.250	4.680	3.57	98.000	164.000	
2022-394	9 GPO TRIND.	191 RX-SERIES	126 RANDOX		32	129.097	6.808	5.27	97.000	161.000	
2022-394	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		22	133.909	2.391	1.79	100.000	167.000	
2022-394	10 ENZYM. GPO	203 ROCHE	93 ROCHE		21	131.750	2.527	1.92	99.000	165.000	
2022-394	10 ENZYM. GPO	219 ATELLICA	143 SIEMENS		10	133.300	7.430	5.57	100.000	167.000	
2022-394	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		11	122.727	4.266	3.48	92.000	153.000	
2022-395	2 ENZ. COLO.	165 VITROS	63 KODAK		147	125.231	4.014	3.21	94.000	157.000	
2022-395	2 ENZ. COLO.	184 INTEGRA	93 ROCHE		12	116.167	1.863	1.60	87.000	145.000	
2022-395	2 ENZ. COLO.	203 ROCHE	93 ROCHE		20	114.500	2.729	2.38	86.000	143.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2022-395	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	76	104.608	2.604	2.49	78.000	131.000			
2022-395	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS	14	103.714	2.711	2.61	78.000	130.000			
2022-395	5 HANTZCH-C)	197 ARCHITECT	1 ABBOTT	10	106.800	2.600	2.43	80.000	134.000			
2022-395	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN	29	113.036	4.516	3.99	85.000	141.000			
2022-395	9 GPO TRIND.	191 RX-SERIES	126 RANDOX	32	111.281	4.875	4.38	83.000	139.000			
2022-395	10 ENZYM. GPO	184 INTEGRA	93 ROCHE	22	115.727	2.157	1.86	87.000	145.000			
2022-395	10 ENZYM. GPO	203 ROCHE	93 ROCHE	21	113.900	1.868	1.64	85.000	142.000			
2022-395	10 ENZYM. GPO	219 ATELLICA	143 SIEMENS	10	118.400	14.305	12.08	89.000	148.000			
2022-395	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT	11	105.727	3.645	3.45	79.000	132.000			
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505	EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}									
2022-391	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	82	33.159	1.435	4.33	30.000	36.000			
2022-391	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	39	32.718	1.319	4.03	30.000	36.000			
2022-391	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	31	33.581	1.009	3.00	31.000	37.000			
2022-391	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX	38	32.405	1.618	4.99	29.000	35.000			
2022-391	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX	38	34.600	0.800	2.31	31.000	38.000			
2022-391	2 UREAS.GLDH	203 ROCHE	93 ROCHE	44	32.884	1.061	3.23	30.000	36.000			
2022-391	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	13	33.615	1.389	4.13	31.000	37.000			
2022-391	2 UREAS.GLDH	219 ATELLICA	143 SIEMENS	10	34.200	1.077	3.15	31.000	37.000			
2022-391	5 CHROMATOG.	197 ARCHITECT	1 ABBOTT	11	33.091	0.996	3.01	30.000	36.000			
2022-391	8 UREASE QUI	165 VITROS	63 KODAK	159	29.777	0.834	2.80	27.000	32.000			
2022-391	10 UREASE 37	220 Medica E.	130 MEDICA	10	31.100	0.700	2.25	28.000	34.000			
2022-392	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	82	11.309	0.748	6.61	9.000	13.000			
2022-392	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	39	11.154	0.662	5.93	9.000	13.000			
2022-392	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	31	11.613	0.549	4.73	10.000	14.000			
2022-392	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX	38	12.000	0.000	0.00	10.000	14.000			
2022-392	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX	38	11.757	0.713	6.06	10.000	14.000			
2022-392	2 UREAS.GLDH	203 ROCHE	93 ROCHE	44	11.163	0.713	6.39	9.000	13.000			
2022-392	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	13	11.417	0.493	4.32	9.000	13.000			
2022-392	2 UREAS.GLDH	219 ATELLICA	143 SIEMENS	10	12.100	0.539	4.45	10.000	14.000			

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-392	5	CHROMATOG.	197 ARCHITECT	1	ABBOTT	11	11.000	0.739	6.71	9.000	13.000
2022-392	8	UREASE QUI	165 VITROS	63	KODAK	159	9.912	0.468	4.72	8.000	12.000
2022-392	10	UREASE 37	220 Medica E.	130	MEDICA	10	11.000	0.447	4.07	9.000	13.000
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505		EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}							
2022-393	2	UREAS.GLDH	38 SIEMENS DI	143	SIEMENS	82	42.427	1.601	3.77	39.000	46.000
2022-393	2	UREAS.GLDH	184 INTEGRA	93	ROCHE	39	41.842	1.328	3.17	38.000	46.000
2022-393	2	UREAS.GLDH	186 BECK. COUL	76	AU-BECKMAN	31	42.355	1.258	2.97	39.000	46.000
2022-393	2	UREAS.GLDH	191 RX-SERIES	126	RANDOX	38	40.368	2.006	4.97	37.000	44.000
2022-393	2	UREAS.GLDH	191 RX-SERIES	126	RANDOX	38	44.400	0.490	1.10	40.000	48.000
2022-393	2	UREAS.GLDH	203 ROCHE	93	ROCHE	44	41.674	1.426	3.42	38.000	45.000
2022-393	2	UREAS.GLDH	212 S. D. EXL	143	SIEMENS	13	42.769	1.625	3.80	39.000	47.000
2022-393	2	UREAS.GLDH	219 ATELLICA	143	SIEMENS	10	42.100	0.831	1.97	38.000	46.000
2022-393	5	CHROMATOG.	197 ARCHITECT	1	ABBOTT	11	43.636	0.979	2.24	40.000	48.000
2022-393	8	UREASE QUI	165 VITROS	63	KODAK	159	38.785	1.051	2.71	35.000	42.000
2022-393	10	UREASE 37	220 Medica E.	130	MEDICA	10	40.100	2.468	6.15	36.000	44.000
2022-394	2	UREAS.GLDH	38 SIEMENS DI	143	SIEMENS	82	19.198	0.999	5.20	17.000	21.000
2022-394	2	UREAS.GLDH	184 INTEGRA	93	ROCHE	39	19.051	0.749	3.93	17.000	21.000
2022-394	2	UREAS.GLDH	186 BECK. COUL	76	AU-BECKMAN	31	19.613	0.748	3.81	18.000	22.000
2022-394	2	UREAS.GLDH	191 RX-SERIES	126	RANDOX	38	20.200	0.400	1.98	18.000	22.000
2022-394	2	UREAS.GLDH	191 RX-SERIES	126	RANDOX	38	19.946	1.064	5.33	18.000	22.000
2022-394	2	UREAS.GLDH	203 ROCHE	93	ROCHE	44	19.233	0.710	3.69	17.000	21.000
2022-394	2	UREAS.GLDH	212 S. D. EXL	143	SIEMENS	13	19.462	1.278	6.57	17.000	21.000
2022-394	2	UREAS.GLDH	219 ATELLICA	143	SIEMENS	10	20.400	0.800	3.92	18.000	22.000
2022-394	5	CHROMATOG.	197 ARCHITECT	1	ABBOTT	11	19.455	0.498	2.56	17.000	21.000
2022-394	8	UREASE QUI	165 VITROS	63	KODAK	159	17.120	0.630	3.68	15.000	19.000
2022-394	10	UREASE 37	220 Medica E.	130	MEDICA	10	18.900	0.831	4.40	17.000	21.000
2022-395	2	UREAS.GLDH	38 SIEMENS DI	143	SIEMENS	82	15.975	0.860	5.39	14.000	18.000
2022-395	2	UREAS.GLDH	184 INTEGRA	93	ROCHE	39	15.564	0.590	3.79	14.000	18.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-395	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN		31	16.161	0.573	3.55	14.000	18.000	
2022-395	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX		38	16.500	1.094	6.63	15.000	19.000	
2022-395	2 UREAS.GLDH	191 RX-SERIES	126 RANDOX		38	16.800	0.400	2.38	15.000	19.000	
2022-395	2 UREAS.GLDH	203 ROCHE	93 ROCHE		44	15.791	0.631	4.00	14.000	18.000	
2022-395	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS		13	15.769	1.187	7.53	14.000	18.000	
2022-395	2 UREAS.GLDH	219 ATELLICA	143 SIEMENS		10	17.600	1.562	8.88	16.000	20.000	
2022-395	5 CHROMATOG.	197 ARCHITECT	1 ABBOTT		11	16.091	0.668	4.15	14.000	18.000	
2022-395	8 UREASE QUI	165 VITROS	63 KODAK		159	14.025	0.527	3.76	12.000	16.000	
2022-395	10 UREASE 37	220 Medica E.	130 MEDICA		10	15.500	0.671	4.33	14.000	18.000	
027 URIC ACID		CLIA TEST ID #: 515	EVALUATION CRITERIA: TARGET VALUE 17 %								
2022-391	2 URICASE	38 SIEMENS DI	143 SIEMENS		42	8.315	0.168	2.02	6.900	9.700	
2022-391	2 URICASE	165 VITROS	63 KODAK		99	8.459	0.204	2.41	7.000	9.900	
2022-391	2 URICASE	184 INTEGRA	93 ROCHE		16	8.840	0.270	3.06	7.300	10.300	
2022-391	2 URICASE	203 ROCHE	93 ROCHE		25	8.296	0.224	2.69	6.900	9.700	
2022-391	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		21	8.452	0.105	1.24	7.000	9.900	
2022-391	6 ENZY. COL	203 ROCHE	93 ROCHE		11	8.445	0.167	1.98	7.000	9.900	
2022-392	2 URICASE	38 SIEMENS DI	143 SIEMENS		42	1.771	0.092	5.18	1.500	2.100	
2022-392	2 URICASE	165 VITROS	63 KODAK		99	1.836	0.066	3.59	1.500	2.100	
2022-392	2 URICASE	184 INTEGRA	93 ROCHE		16	1.833	0.047	2.57	1.500	2.100	
2022-392	2 URICASE	203 ROCHE	93 ROCHE		25	1.729	0.068	3.91	1.400	2.000	
2022-392	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		21	1.890	0.081	4.29	1.600	2.200	
2022-392	6 ENZY. COL	203 ROCHE	93 ROCHE		11	1.773	0.045	2.51	1.500	2.100	
2022-393	2 URICASE	38 SIEMENS DI	143 SIEMENS		42	10.929	0.268	2.45	9.100	12.800	
2022-393	2 URICASE	165 VITROS	63 KODAK		99	11.236	0.260	2.31	9.300	13.100	
2022-393	2 URICASE	184 INTEGRA	93 ROCHE		16	11.394	0.305	2.68	9.500	13.300	
2022-393	2 URICASE	203 ROCHE	93 ROCHE		25	10.948	0.316	2.89	9.100	12.800	
2022-393	5 URICASE 37	186 BECK. COUL	76 AU-BECKMAN		21	11.019	0.240	2.18	9.100	12.900	
2022-393	6 ENZY. COL	203 ROCHE	93 ROCHE		11	11.045	0.202	1.83	9.200	12.900	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
027	URIC ACID	CLIA TEST ID #: 515				EVALUATION CRITERIA: TARGET VALUE 17 %						
2022-394	2	URICASE	38	SIEMENS DI	143	SIEMENS	42	4.141	0.145	3.50	3.400	4.800
2022-394	2	URICASE	165	VITROS	63	KODAK	99	4.226	0.119	2.83	3.500	4.900
2022-394	2	URICASE	184	INTEGRA	93	ROCHE	16	4.331	0.116	2.67	3.600	5.100
2022-394	2	URICASE	203	ROCHE	93	ROCHE	25	4.152	0.110	2.65	3.400	4.900
2022-394	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	21	4.319	0.105	2.44	3.600	5.100
2022-394	6	ENZY. COL	203	ROCHE	93	ROCHE	11	4.209	0.090	2.14	3.500	4.900
2022-395	2	URICASE	38	SIEMENS DI	143	SIEMENS	42	3.107	0.109	3.51	2.600	3.600
2022-395	2	URICASE	165	VITROS	63	KODAK	99	3.204	0.093	2.91	2.700	3.700
2022-395	2	URICASE	184	INTEGRA	93	ROCHE	16	3.256	0.100	3.07	2.700	3.800
2022-395	2	URICASE	203	ROCHE	93	ROCHE	25	3.132	0.088	2.82	2.600	3.700
2022-395	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	21	3.276	0.068	2.09	2.700	3.800
2022-395	6	ENZY. COL	203	ROCHE	93	ROCHE	11	3.173	0.062	1.94	2.600	3.700
135	GGT (GAMMA GLU. TRA)	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2022-391	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	273.300	7.001	2.56	219.000	328.000
2022-391	5	KINETIC 37	165	VITROS	63	KODAK	22	468.545	12.572	2.68	375.000	562.000
2022-392	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	12.100	0.300	2.48	10.000	15.000
2022-392	5	KINETIC 37	165	VITROS	63	KODAK	22	21.864	1.391	6.36	17.000	26.000
2022-393	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	387.000	30.318	7.83	310.000	464.000
2022-393	5	KINETIC 37	165	VITROS	63	KODAK	22	608.190	15.924	2.62	487.000	730.000
2022-394	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	108.000	3.688	3.41	86.000	130.000
2022-394	5	KINETIC 37	165	VITROS	63	KODAK	22	211.455	5.566	2.63	169.000	254.000
2022-395	3	ALPHA GGP	186	BECK. COUL	76	AU-BECKMAN	10	68.100	1.814	2.66	54.000	82.000
2022-395	5	KINETIC 37	165	VITROS	63	KODAK	22	136.318	3.521	2.58	109.000	164.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
136	LACTIC ACID	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2022-391	1 LACTATE-P.	165	VITROS	63	KODAK	18	4.800	0.125	2.60	3.800	5.800	
2022-392	1 LACTATE-P.	165	VITROS	63	KODAK	18	0.500	0.000	0.00	0.400	0.600	
2022-393	1 LACTATE-P.	165	VITROS	63	KODAK	18	6.417	0.186	2.90	5.100	7.700	
2022-394	1 LACTATE-P.	165	VITROS	63	KODAK	18	2.033	0.058	2.84	1.600	2.400	
2022-395	1 LACTATE-P.	165	VITROS	63	KODAK	18	1.361	0.049	3.58	1.100	1.600	
137	LIPASE											
2022-391	1 COLORIMET.	203	ROCHE	93	ROCHE	14	71.077	5.967	8.40	57.000	85.000	
2022-391	2 ENZIMATIC	165	VITROS	63	KODAK	82	584.778	16.113	2.76	468.000	702.000	
2022-391	4 ENZ/COLOR	184	INTEGRA	93	ROCHE	10	67.500	3.775	5.59	54.000	81.000	
2022-391	4 ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	75.200	2.535	3.37	60.000	90.000	
2022-391	4 ENZ/COLOR	203	ROCHE	93	ROCHE	14	68.286	4.479	6.56	55.000	82.000	
2022-391	5 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	212.000	11.411	5.38	170.000	254.000	
2022-391	5 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	265.250	13.103	4.94	212.000	318.000	
2022-392	1 COLORIMET.	203	ROCHE	93	ROCHE	14	16.615	0.625	3.76	13.000	20.000	
2022-392	2 ENZIMATIC	165	VITROS	63	KODAK	82	63.098	3.274	5.19	50.000	76.000	
2022-392	4 ENZ/COLOR	184	INTEGRA	93	ROCHE	10	17.200	0.872	5.07	14.000	21.000	
2022-392	4 ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	15.571	0.904	5.80	12.000	19.000	
2022-392	4 ENZ/COLOR	203	ROCHE	93	ROCHE	14	16.500	0.500	3.03	13.000	20.000	
2022-392	5 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	56.000	6.164	11.01	45.000	67.000	
2022-392	5 SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	50.000	4.796	9.59	40.000	60.000	
2022-393	1 COLORIMET.	203	ROCHE	93	ROCHE	14	91.846	7.705	8.39	73.000	110.000	
2022-393	2 ENZIMATIC	165	VITROS	63	KODAK	82	789.667	50.892	6.44	632.000	948.000	
2022-393	4 ENZ/COLOR	184	INTEGRA	93	ROCHE	10	86.100	6.007	6.98	69.000	103.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-393	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	98.467	4.031	4.09	79.000	118.000
2022-393	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	88.286	6.088	6.90	71.000	106.000
2022-393	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	274.059	10.608	3.87	219.000	329.000
2022-393	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	346.250	21.229	6.13	277.000	416.000
137	LIPASE	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2022-394	1	COLORIMET.	203	ROCHE	93	ROCHE	14	36.462	2.308	6.33	29.000	44.000
2022-394	2	ENZIMATIC	165	VITROS	63	KODAK	82	269.975	7.948	2.94	216.000	324.000
2022-394	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	35.900	1.814	5.05	29.000	43.000
2022-394	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	37.200	2.135	5.74	30.000	45.000
2022-394	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	35.929	2.219	6.18	29.000	43.000
2022-394	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	135.000	10.247	7.59	108.000	162.000
2022-394	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	109.056	10.347	9.49	87.000	131.000
2022-395	1	COLORIMET.	203	ROCHE	93	ROCHE	14	28.154	1.875	6.66	23.000	34.000
2022-395	2	ENZIMATIC	165	VITROS	63	KODAK	82	186.646	6.564	3.52	149.000	224.000
2022-395	4	ENZ/COLOR	184	INTEGRA	93	ROCHE	10	27.900	1.446	5.18	22.000	33.000
2022-395	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	15	27.643	2.057	7.44	22.000	33.000
2022-395	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	27.786	1.206	4.34	22.000	33.000
2022-395	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	86.222	6.545	7.59	69.000	103.000
2022-395	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	102.500	8.078	7.88	82.000	123.000
138	PHOSPHORUS											
2022-391	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	42	5.367	0.121	2.25	4.300	6.400
2022-391	1	PHOSPHOMO.	165	VITROS	63	KODAK	90	4.977	0.141	2.84	4.000	6.000
2022-391	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	16	5.113	0.141	2.76	4.100	6.100
2022-391	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	4.900	0.141	2.89	3.900	5.900
2022-391	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	5.080	0.075	1.47	4.100	6.100
2022-391	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	4.986	0.247	4.96	4.000	6.000
2022-391	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	4.965	0.113	2.29	4.000	6.000
2022-391	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	34	5.094	0.147	2.89	4.100	6.100

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
138	PHOSPHORUS	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 20 %								
2022-392	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	42	1.707	0.087	5.07	1.400	2.000	
2022-392	1	PHOSPHOMO.	165 VITROS	63	KODAK	90	1.833	0.101	5.51	1.500	2.200	
2022-392	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	1.600	0.094	5.85	1.300	1.900	
2022-392	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	1.556	0.101	6.51	1.200	1.900	
2022-392	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	1.467	0.125	8.50	1.200	1.800	
2022-392	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	1.560	0.080	5.13	1.200	1.900	
2022-392	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	1.514	0.064	4.22	1.200	1.800	
2022-392	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	34	1.548	0.099	6.38	1.200	1.900	
2022-393	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	42	6.781	0.145	2.14	5.400	8.100	
2022-393	1	PHOSPHOMO.	165 VITROS	63	KODAK	90	6.267	0.177	2.82	5.000	7.500	
2022-393	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	6.375	0.148	2.32	5.100	7.700	
2022-393	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	6.333	0.125	1.97	5.100	7.600	
2022-393	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	6.520	0.160	2.45	5.200	7.800	
2022-393	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	6.314	0.236	3.73	5.100	7.600	
2022-393	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	6.353	0.133	2.10	5.100	7.600	
2022-393	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	34	6.474	0.225	3.48	5.200	7.800	
2022-394	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	42	3.168	0.097	3.07	2.500	3.800	
2022-394	1	PHOSPHOMO.	165 VITROS	63	KODAK	90	3.162	0.090	2.85	2.500	3.800	
2022-394	1	PHOSPHOMO.	184 INTEGRA	93	ROCHE	16	2.950	0.087	2.94	2.400	3.500	
2022-394	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	2.912	0.118	4.06	2.300	3.500	
2022-394	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	2.833	0.125	4.40	2.300	3.400	
2022-394	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	2.980	0.075	2.51	2.400	3.600	
2022-394	1	PHOSPHOMO.	186 BECK. COUL	76	AU-BECKMAN	18	2.871	0.128	4.45	2.300	3.400	
2022-394	1	PHOSPHOMO.	203 ROCHE	93	ROCHE	34	2.938	0.121	4.13	2.400	3.500	
2022-395	1	PHOSPHOMO.	38 SIEMENS DI	143	SIEMENS	42	2.559	0.083	3.23	2.000	3.100	
2022-395	1	PHOSPHOMO.	165 VITROS	63	KODAK	90	2.647	0.093	3.53	2.100	3.200	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-395	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	16	2.369	0.077	3.24	1.900	2.800
2022-395	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	2.267	0.094	4.16	1.800	2.700
2022-395	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	2.380	0.117	4.90	1.900	2.900
2022-395	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	2.314	0.064	2.76	1.900	2.800
2022-395	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	18	2.318	0.098	4.25	1.900	2.800
2022-395	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	34	2.374	0.088	3.73	1.900	2.800
169	CO2			CLIA TEST ID #: 0								
2022-391	1	Generic Me	165	VITROS	63	KODAK	15	21.333	1.989	9.32	15.000	27.000
2022-391	1	Generic Me	184	INTEGRA	93	ROCHE	14	19.929	2.549	12.79	12.000	28.000
2022-391	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	23	22.826	2.761	12.09	15.000	31.000
2022-391	1	Generic Me	203	ROCHE	93	ROCHE	18	21.444	1.863	8.69	16.000	27.000
2022-391	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	65	23.923	2.685	11.22	16.000	32.000
2022-391	2	ENZYMATIC	165	VITROS	63	KODAK	64	21.688	1.935	8.92	16.000	27.000
2022-391	2	ENZYMATIC	165	VITROS	156	Kit 32	65	21.600	2.037	9.43	15.000	28.000
2022-391	2	ENZYMATIC	203	ROCHE	93	ROCHE	12	20.833	1.344	6.45	17.000	25.000
2022-392	1	Generic Me	165	VITROS	63	KODAK	15	9.133	1.543	16.90	5.000	14.000
2022-392	1	Generic Me	184	INTEGRA	93	ROCHE	14	9.786	1.264	12.91	6.000	14.000
2022-392	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	23	10.545	1.117	10.59	7.000	14.000
2022-392	1	Generic Me	203	ROCHE	93	ROCHE	18	9.444	1.117	11.82	6.000	13.000
2022-392	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	65	11.906	1.809	15.19	6.000	17.000
2022-392	2	ENZYMATIC	165	VITROS	63	KODAK	64	9.873	1.279	12.95	6.000	14.000
2022-392	2	ENZYMATIC	165	VITROS	156	Kit 32	65	9.538	1.229	12.88	6.000	13.000
2022-392	2	ENZYMATIC	203	ROCHE	93	ROCHE	12	9.833	1.404	14.28	6.000	14.000
2022-393	1	Generic Me	165	VITROS	63	KODAK	15	25.200	1.514	6.01	21.000	30.000
2022-393	1	Generic Me	184	INTEGRA	93	ROCHE	14	23.786	2.704	11.37	16.000	32.000
2022-393	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	23	27.609	2.583	9.36	20.000	35.000
2022-393	1	Generic Me	203	ROCHE	93	ROCHE	18	25.353	1.532	6.04	21.000	30.000
2022-393	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	65	27.785	2.533	9.12	20.000	35.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2022-393	2	ENZYMATIC	165 VITROS	63	KODAK	64	25.688	2.290	8.92	19.000	33.000	
2022-393	2	ENZYMATIC	165 VITROS	156	Kit 32	65	25.923	2.114	8.16	20.000	32.000	
2022-393	2	ENZYMATIC	203 ROCHE	93	ROCHE	12	24.583	2.216	9.01	18.000	31.000	
169	CO2		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE	3	SD					
2022-394	1	Generic Me	165 VITROS	63	KODAK	15	13.333	1.776	13.32	8.000	19.000	
2022-394	1	Generic Me	184 INTEGRA	93	ROCHE	14	13.143	2.199	16.73	7.000	20.000	
2022-394	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	23	14.783	1.793	12.13	9.000	20.000	
2022-394	1	Generic Me	203 ROCHE	93	ROCHE	18	12.944	1.840	14.22	7.000	18.000	
2022-394	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	65	15.785	2.243	14.21	9.000	23.000	
2022-394	2	ENZYMATIC	165 VITROS	63	KODAK	64	13.444	1.735	12.90	8.000	19.000	
2022-394	2	ENZYMATIC	165 VITROS	156	Kit 32	65	12.985	1.678	12.92	8.000	18.000	
2022-394	2	ENZYMATIC	203 ROCHE	93	ROCHE	12	13.667	1.700	12.44	9.000	19.000	
2022-395	1	Generic Me	165 VITROS	63	KODAK	15	11.267	1.652	14.66	6.000	16.000	
2022-395	1	Generic Me	184 INTEGRA	93	ROCHE	14	11.071	1.791	16.18	6.000	16.000	
2022-395	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	23	12.696	1.516	11.94	8.000	17.000	
2022-395	1	Generic Me	203 ROCHE	93	ROCHE	18	11.111	1.410	12.69	7.000	15.000	
2022-395	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	65	14.047	1.956	13.92	8.000	20.000	
2022-395	2	ENZYMATIC	165 VITROS	63	KODAK	64	11.422	1.378	12.07	7.000	16.000	
2022-395	2	ENZYMATIC	165 VITROS	156	Kit 32	65	11.246	1.253	11.14	7.000	15.000	
2022-395	2	ENZYMATIC	203 ROCHE	93	ROCHE	12	11.500	1.756	15.27	6.000	17.000	
170	DIRECT BILIRUBIN				EVALUATION CRITERIA: TARGET VALUE	3	SD OR 20 % {GREATER}					
2022-391	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	45	1.038	0.068	6.52	0.800	1.200	
2022-391	5	JEND.-GROG	197 ARCHITECT	1	ABBOTT	11	1.755	0.144	8.19	1.300	2.200	
2022-391	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.595	0.110	18.52	0.300	0.900	
2022-391	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	16	0.906	0.125	13.78	0.500	1.300	
2022-391	18	DIAZ. SULF	191 RX-SERIES	126	RANDOX	12	1.533	0.201	13.13	0.900	2.100	
2022-392	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	45	0.071	0.045	63.74	0	0.200	
2022-392	5	JEND.-GROG	197 ARCHITECT	1	ABBOTT	11	0.100	0.000	0.00	0.100	0.100	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2022-392	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.008	0.027	334.66	0	0.100	
2022-392	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	16	0.200	0.000	0.00	0.200	0.200	
2022-392	18	DIAZ. SULF	191 RX-SERIES	126	RANDOX	12	0.075	0.083	110.55	0	0.300	
170	DIRECT BILIRUBIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}								
2022-393	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	45	1.371	0.083	6.08	1.100	1.600	
2022-393	5	JEND.-GROG	197 ARCHITECT	1	ABBOTT	11	2.291	0.162	7.08	1.800	2.800	
2022-393	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.923	0.172	18.61	0.400	1.400	
2022-393	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	16	1.163	0.145	12.49	0.700	1.600	
2022-393	18	DIAZ. SULF	191 RX-SERIES	126	RANDOX	12	2.200	0.308	14.01	1.300	3.100	
2022-394	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	45	0.391	0.046	11.84	0.300	0.500	
2022-394	5	JEND.-GROG	197 ARCHITECT	1	ABBOTT	11	0.745	0.066	8.79	0.500	0.900	
2022-394	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.047	0.123	263.80	0	0.400	
2022-394	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	16	0.369	0.058	15.81	0.200	0.500	
2022-394	18	DIAZ. SULF	191 RX-SERIES	126	RANDOX	12	0.558	0.104	18.58	0.200	0.900	
2022-395	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	45	0.218	0.038	17.56	0.100	0.300	
2022-395	5	JEND.-GROG	197 ARCHITECT	1	ABBOTT	11	0.482	0.094	19.43	0.200	0.800	
2022-395	11	SPECT. WIT	165 VITROS	63	KODAK	63	0.031	0.092	293.80	0	0.300	
2022-395	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	16	0.200	0.000	0.00	0.200	0.200	
2022-395	18	DIAZ. SULF	191 RX-SERIES	126	RANDOX	12	0.342	0.049	14.43	0.200	0.500	

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 15 WHOLE BLOOD GLUCOSE								
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
157	WHOLE BLOOD GLUCOSE		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}									
2022-491	20	ABBOT FREE		21	157.095	3.490	2.22	141.000	173.000			
2022-491	23	ABBOTT PRE		27	136.038	6.460	4.75	122.000	150.000			
2022-491	27	STAT SRIP.		45	117.022	7.120	6.08	105.000	129.000			

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 15 WHOLE BLOOD GLUCOSE					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-491	33	AC.CHEK II		30	128.897	3.100	2.41	116.000	142.000
2022-491	34	FREE STYL		46	155.333	10.283	6.62	140.000	171.000
157 WHOLE BLOOD GLUCOSE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}					
2022-492	20	ABBOT FREE		21	92.000	2.619	2.85	83.000	101.000
2022-492	23	ABBOTT PRE		27	80.259	4.888	6.09	72.000	88.000
2022-492	27	STAT SRIP.		45	72.733	4.484	6.17	65.000	80.000
2022-492	33	AC.CHEK II		30	81.793	2.325	2.84	74.000	90.000
2022-492	34	FREE STYL		46	92.087	5.090	5.53	83.000	101.000
2022-493	20	ABBOT FREE		21	154.381	4.445	2.88	139.000	170.000
2022-493	23	ABBOTT PRE		27	136.704	7.678	5.62	123.000	150.000
2022-493	27	STAT SRIP.		45	115.556	6.933	6.00	104.000	127.000
2022-493	33	AC.CHEK II		30	130.069	3.991	3.07	117.000	143.000
2022-493	34	FREE STYL		46	154.087	8.680	5.63	139.000	169.000
2022-494	20	ABBOT FREE		21	406.714	12.918	3.18	366.000	447.000
2022-494	23	ABBOTT PRE		27	369.111	17.777	4.82	332.000	406.000
2022-494	27	STAT SRIP.		45	249.533	22.129	8.87	225.000	274.000
2022-494	33	AC.CHEK II		30	332.276	11.289	3.40	299.000	366.000
2022-494	34	FREE STYL		46	405.043	31.098	7.68	365.000	446.000
2022-495	20	ABBOT FREE		21	91.619	3.748	4.09	82.000	101.000
2022-495	23	ABBOTT PRE		27	80.000	4.706	5.88	72.000	88.000
2022-495	27	STAT SRIP.		45	71.822	5.364	7.47	65.000	79.000
2022-495	33	AC.CHEK II		30	82.103	2.618	3.19	74.000	90.000
2022-495	34	FREE STYL		46	91.935	6.381	6.94	83.000	101.000
SPEC: 10 CHEMISTRY		SUB: 20 BLOOD GASES		GRP: 30 BLOOD GASES					

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE	8	% OR 5.00	{GREATER}			
2022-441	265 ROCHE COBA				30	66.917	1.560	2.33	61.600	72.300
2022-442	265 ROCHE COBA				30	43.990	0.682	1.55	39.000	49.000
2022-443	265 ROCHE COBA				30	22.227	0.579	2.61	17.200	27.200
2022-444	265 ROCHE COBA				30	64.560	1.097	1.70	59.400	69.700
2022-445	265 ROCHE COBA				30	43.873	0.658	1.50	38.900	48.900
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04					
2022-441	265 ROCHE COBA				30	7.221	0.005	0.08	7.181	7.261
2022-442	265 ROCHE COBA				30	7.404	0.004	0.06	7.365	7.445
2022-443	265 ROCHE COBA				30	7.534	0.005	0.06	7.494	7.574
2022-444	265 ROCHE COBA				30	7.193	0.005	0.07	7.153	7.233
2022-445	265 ROCHE COBA				30	7.404	0.005	0.07	7.364	7.444
009	BLOOD GASES-pO2		CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE	3	SD				
2022-441	265 ROCHE COBA				30	49.555	3.400	6.86	39.400	59.800
2022-442	265 ROCHE COBA				30	114.845	3.701	3.22	103.700	125.900
2022-443	265 ROCHE COBA				30	187.220	6.155	3.29	168.800	205.700
2022-444	265 ROCHE COBA				30	336.313	20.975	6.24	273.400	399.200

Participant Summary

MONTH 5 OF YEAR 2022

ANALYTE			NO. PART		TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES		GRP: 30 BLOOD GASES				
009	BLOOD GASES-pO2	CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2022-445	265 ROCHE COBA			30	114.753	4.655	4.06	100.800	128.700
SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE			NO. PART		TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0							
2022-421	116 ROCHE COBA			16	13.825	1.669	12.07	8.800	18.800
2022-421	149 ARC i2000s			10	13.310	1.120	8.42	9.900	16.700
2022-422	116 ROCHE COBA			16	9.781	1.231	12.58	6.100	13.500
2022-422	149 ARC i2000s			10	9.140	0.846	9.26	6.600	11.700
2022-423	116 ROCHE COBA			16	33.294	3.345	10.05	23.300	43.300
2022-423	149 ARC i2000s			10	33.440	3.209	9.60	23.800	43.100
2022-424	116 ROCHE COBA			16	16.956	1.976	11.65	11.000	22.900
2022-424	149 ARC i2000s			10	16.460	1.231	7.48	12.800	20.200
2022-425	116 ROCHE COBA			16	28.881	2.796	9.68	20.500	37.300
2022-425	149 ARC i2000s			10	28.770	2.368	8.23	21.700	35.900
030	FREE THYROXINE	CLIA TEST ID #: 545							
2022-421	115 VITROS ECI			42	3.007	0.161	5.35	2.500	3.500
2022-421	116 ROCHE COBA			50	1.400	0.075	5.35	1.200	1.600
2022-421	125 B.C.ACCESS			46	1.269	0.072	5.71	1.100	1.500
2022-421	149 ARC i2000s			23	1.086	0.034	3.16	1.000	1.200
2022-421	160 SIEMENS EX			49	1.324	0.080	6.01	1.100	1.600
2022-422	115 VITROS ECI			42	1.569	0.083	5.29	1.300	1.800

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-422	116	ROCHE COBA		50	0.861	0.060	6.97	0.700	1.000
2022-422	125	B.C.ACCESS		46	0.859	0.053	6.23	0.700	1.000
2022-422	149	ARC i2000s		23	0.718	0.049	6.82	0.600	0.900
2022-422	160	SIEMENS EX		49	0.867	0.062	7.14	0.700	1.100
030	FREE THYROXINE		CLIA TEST ID #: 545						EVALUATION CRITERIA: TARGET VALUE 3 SD
2022-423	115	VITROS ECI		42	1.793	0.101	5.63	1.500	2.100
2022-423	116	ROCHE COBA		50	1.504	0.078	5.20	1.300	1.700
2022-423	125	B.C.ACCESS		46	0.983	0.064	6.47	0.800	1.200
2022-423	149	ARC i2000s		23	1.313	0.061	4.66	1.100	1.500
2022-423	160	SIEMENS EX		49	1.167	0.062	5.30	1.000	1.400
2022-424	115	VITROS ECI		42	4.056	0.212	5.22	3.400	4.700
2022-424	116	ROCHE COBA		50	1.756	0.085	4.85	1.500	2.000
2022-424	125	B.C.ACCESS		46	1.543	0.088	5.68	1.300	1.800
2022-424	149	ARC i2000s		23	1.357	0.071	5.25	1.100	1.600
2022-424	160	SIEMENS EX		49	1.631	0.091	5.57	1.400	1.900
2022-425	115	VITROS ECI		42	6.821	0.144	2.11	6.400	7.300
2022-425	116	ROCHE COBA		50	3.168	0.183	5.77	2.600	3.700
2022-425	125	B.C.ACCESS		46	2.502	0.155	6.21	2.000	3.000
2022-425	149	ARC i2000s		23	2.483	0.146	5.90	2.000	2.900
2022-425	160	SIEMENS EX		49	2.906	0.146	5.04	2.500	3.300
032	hCG-QUANTITATIVE		CLIA TEST ID #: 555						
2022-421	115	VITROS ECI		19	169.942	6.422	3.78	150.700	189.200
2022-421	116	ROCHE COBA		19	130.168	6.396	4.91	111.000	149.400
2022-421	125	B.C.ACCESS		11	130.345	4.698	3.60	116.300	144.400
2022-422	115	VITROS ECI		19	2.261	0.549	24.27	0.600	3.900
2022-422	116	ROCHE COBA		19	0.721	0.188	26.07	0.200	1.300
2022-422	125	B.C.ACCESS		11	1.264	0.123	9.71	0.900	1.600

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2022-423	115 VITROS ECI			19	2.974	0.335	11.28	2.000	4.000	
2022-423	116 ROCHE COBA			19	2.526	0.288	11.40	1.700	3.400	
2022-423	125 B.C.ACCESS			11	3.473	0.200	5.77	2.900	4.100	
2022-424	115 VITROS ECI			19	290.368	11.185	3.85	256.800	323.900	
2022-424	116 ROCHE COBA			19	228.037	13.025	5.71	189.000	267.100	
2022-424	125 B.C.ACCESS			11	223.745	8.869	3.96	197.100	250.400	
2022-425	115 VITROS ECI			19	809.321	29.760	3.68	720.000	898.600	
2022-425	116 ROCHE COBA			19	643.321	33.309	5.18	543.400	743.200	
2022-425	125 B.C.ACCESS			11	574.445	22.555	3.93	506.800	642.100	
035	PROLACTIN	CLIA TEST ID #: 0								
2022-421	116 ROCHE COBA			18	8.261	0.245	2.97	7.500	9.000	
2022-422	116 ROCHE COBA			18	3.683	0.201	5.45	3.100	4.300	
2022-423	116 ROCHE COBA			18	11.735	0.556	4.74	10.100	13.400	
2022-424	116 ROCHE COBA			18	11.906	0.660	5.54	9.900	13.900	
2022-425	116 ROCHE COBA			18	26.850	1.405	5.23	22.600	31.100	
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585								
2022-421	106 TOSOH AIA			14	2.150	0.282	13.13	1.300	3.000	
2022-421	115 VITROS ECI			46	2.237	0.096	4.31	1.900	2.500	
2022-421	116 ROCHE COBA			55	2.231	0.114	5.09	1.900	2.600	
2022-421	125 B.C.ACCESS			59	1.890	0.092	4.88	1.600	2.200	
2022-421	149 ARC i2000s			25	1.684	0.116	6.86	1.300	2.000	
2022-421	160 SIEMENS EX			78	1.741	0.136	7.83	1.300	2.100	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2022-422	106 TOSOH AIA			14	0.892	0.107	12.01	0.600	1.200
2022-422	115 VITROS ECI			46	0.822	0.046	5.62	0.700	1.000
2022-422	116 ROCHE COBA			55	1.091	0.062	5.71	0.900	1.300
2022-422	125 B.C.ACCESS			59	0.829	0.058	7.04	0.700	1.000
2022-422	149 ARC i2000s			25	0.713	0.033	4.64	0.600	0.800
2022-422	160 SIEMENS EX			78	0.809	0.065	8.01	0.600	1.000
2022-423	106 TOSOH AIA			14	3.064	0.388	12.68	1.900	4.200
2022-423	115 VITROS ECI			46	2.780	0.096	3.44	2.500	3.100
2022-423	116 ROCHE COBA			55	2.996	0.150	5.01	2.500	3.400
2022-423	125 B.C.ACCESS			59	3.093	0.176	5.70	2.600	3.600
2022-423	149 ARC i2000s			25	2.568	0.162	6.30	2.100	3.100
2022-423	160 SIEMENS EX			78	2.792	0.206	7.36	2.200	3.400
2022-424	106 TOSOH AIA			14	2.921	0.421	14.42	1.700	4.200
2022-424	115 VITROS ECI			46	3.187	0.119	3.74	2.800	3.500
2022-424	116 ROCHE COBA			55	3.030	0.195	6.44	2.400	3.600
2022-424	125 B.C.ACCESS			59	2.627	0.133	5.05	2.200	3.000
2022-424	149 ARC i2000s			25	2.380	0.152	6.40	1.900	2.800
2022-424	160 SIEMENS EX			78	2.408	0.182	7.55	1.900	3.000
2022-425	106 TOSOH AIA			14	6.286	0.807	12.84	3.900	8.700
2022-425	115 VITROS ECI			46	6.887	0.234	3.40	6.200	7.600
2022-425	116 ROCHE COBA			55	5.978	0.261	4.36	5.200	6.800
2022-425	125 B.C.ACCESS			59	5.783	0.271	4.69	5.000	6.600
2022-425	149 ARC i2000s			25	5.184	0.365	7.04	4.100	6.300
2022-425	160 SIEMENS EX			78	5.032	0.317	6.30	4.100	6.000
040	THYROXINE (T4)	CLIA TEST ID #: 595	EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}						
2022-421	32 DADE DIME.			44	6.423	0.461	7.17	5.100	7.700

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-421	106	TOSOH AIA		11	5.600	1.000	17.86	4.400	6.600
2022-421	106	TOSOH AIA		11	5.373	0.484	9.02	4.300	6.400
2022-421	115	VITROS ECI		38	5.405	0.203	3.75	4.300	6.500
2022-421	116	ROCHE COBA		51	5.318	0.315	5.93	4.300	6.400
2022-421	125	B.C.ACCESS		46	5.271	0.431	8.18	4.200	6.300
2022-421	125	B.C.ACCESS		46	5.600	1.000	17.86	4.400	6.600
2022-421	149	ARC i2000s		22	4.886	0.275	5.63	3.900	5.900
2022-421	160	SIEMENS EX		13	6.346	0.696	10.96	5.100	7.600
040 THYROXINE (T4)			CLIA TEST ID #: 595				EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}		
2022-422	32	DADE DIME.		44	4.064	0.493	12.12	3.100	5.100
2022-422	106	TOSOH AIA		11	3.382	0.285	8.44	2.400	4.400
2022-422	106	TOSOH AIA		11	3.300	0.000	0.00	2.200	4.200
2022-422	115	VITROS ECI		38	3.450	0.141	4.09	2.500	4.500
2022-422	116	ROCHE COBA		51	3.418	0.237	6.94	2.400	4.400
2022-422	125	B.C.ACCESS		46	3.011	0.343	11.38	2.000	4.000
2022-422	125	B.C.ACCESS		46	3.300	0.000	0.00	2.200	4.200
2022-422	149	ARC i2000s		22	2.827	0.226	7.99	1.800	3.800
2022-422	160	SIEMENS EX		13	4.008	0.470	11.72	3.000	5.000
2022-423	32	DADE DIME.		44	11.532	0.642	5.57	9.200	13.800
2022-423	106	TOSOH AIA		11	8.900	1.000	11.24	7.100	10.600
2022-423	106	TOSOH AIA		11	8.918	0.743	8.33	7.100	10.700
2022-423	115	VITROS ECI		38	8.621	0.395	4.58	6.900	10.300
2022-423	116	ROCHE COBA		51	9.553	0.483	5.05	7.600	11.500
2022-423	125	B.C.ACCESS		46	8.900	1.000	11.24	7.100	10.600
2022-423	125	B.C.ACCESS		46	7.882	0.757	9.60	6.300	9.500
2022-423	149	ARC i2000s		22	10.018	0.521	5.20	8.000	12.000
2022-423	160	SIEMENS EX		13	11.400	0.790	6.93	9.100	13.700
2022-424	32	DADE DIME.		44	7.909	0.618	7.82	6.300	9.500

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-424	106	TOSOH AIA		11	7.009	0.553	7.90	5.600	8.400
2022-424	106	TOSOH AIA		11	6.800	1.000	14.71	5.400	8.200
2022-424	115	VITROS ECI		38	6.892	0.266	3.87	5.500	8.300
2022-424	116	ROCHE COBA		51	6.744	0.414	6.14	5.400	8.100
2022-424	125	B.C.ACCESS		46	6.711	0.648	9.66	5.400	8.100
2022-424	125	B.C.ACCESS		46	6.800	1.000	14.71	5.400	8.200
2022-424	149	ARC i2000s		22	6.477	0.396	6.12	5.200	7.800
2022-424	160	SIEMENS EX		13	7.723	0.578	7.48	6.200	9.300
040	THYROXINE (T4)		CLIA TEST ID #: 595		EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}				
2022-425	32	DADE DIME.		44	14.527	0.669	4.60	11.600	17.400
2022-425	106	TOSOH AIA		11	11.800	1.000	8.47	9.400	14.200
2022-425	106	TOSOH AIA		11	11.673	0.862	7.39	9.300	14.000
2022-425	115	VITROS ECI		38	12.726	0.612	4.81	10.200	15.300
2022-425	116	ROCHE COBA		51	12.380	0.612	4.95	9.900	14.900
2022-425	125	B.C.ACCESS		46	11.800	1.000	8.47	9.400	14.200
2022-425	125	B.C.ACCESS		46	11.800	0.838	7.10	9.400	14.200
2022-425	149	ARC i2000s		22	12.986	0.740	5.70	10.400	15.600
2022-425	160	SIEMENS EX		13	14.177	0.895	6.31	11.300	17.000
141	FOLATE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %				
2022-421	115	VITROS ECI		14	1.779	0.286	16.07	1.300	2.200
2022-421	116	ROCHE COBA		18	2.000	0.000	0.00	1.500	2.500
2022-421	125	B.C.ACCESS		11	2.064	0.192	9.30	1.500	2.600
2022-421	149	ARC i2000s		12	2.092	0.335	16.03	1.600	2.600
2022-422	115	VITROS ECI		14	0.964	0.229	23.72	0.700	1.200
2022-422	116	ROCHE COBA		18	2.000	0.000	0.00	1.500	2.500
2022-422	125	B.C.ACCESS		11	0.545	0.215	39.37	0.400	0.700
2022-422	149	ARC i2000s		12	1.125	0.499	44.31	0.800	1.400
2022-423	115	VITROS ECI		14	1.125	2.976	264.58	0.800	1.400

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)				
ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
	2022-423	116	ROCHE COBA		18	2.024	0.094	4.65	1.500	2.500
	2022-423	125	B.C.ACCESS		11	1.091	0.131	12.02	0.800	1.400
	2022-423	149	ARC i2000s		12	1.392	0.375	26.96	1.000	1.700
141 FOLATE			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %						
	2022-424	115	VITROS ECI		14	2.614	0.352	13.47	2.000	3.300
	2022-424	116	ROCHE COBA		18	2.256	0.310	13.72	1.700	2.800
	2022-424	125	B.C.ACCESS		11	3.218	0.307	9.54	2.400	4.000
	2022-424	149	ARC i2000s		12	3.308	0.463	13.99	2.500	4.100
	2022-425	115	VITROS ECI		14	7.336	0.949	12.94	5.500	9.200
	2022-425	116	ROCHE COBA		18	6.217	0.704	11.33	4.700	7.800
	2022-425	125	B.C.ACCESS		11	8.136	0.624	7.67	6.100	10.200
	2022-425	149	ARC i2000s		12	8.717	0.737	8.45	6.500	10.900
142 hLH (HUMAN LUT. HOR)										
	2022-421	116	ROCHE COBA		14	14.146	1.037	7.33	10.600	17.700
	2022-422	116	ROCHE COBA		14	6.271	0.422	6.72	4.700	7.800
	2022-423	116	ROCHE COBA		14	19.543	1.653	8.46	14.700	24.400
	2022-424	116	ROCHE COBA		14	19.529	1.168	5.98	14.600	24.400
	2022-425	116	ROCHE COBA		14	39.957	2.410	6.03	30.000	49.900

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY			GRP: 40 ENDOCRINOLOGY (GEN.)					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
031 hCG-QUALITATIVE											
2022-421	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	22	22	22	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-421	142	INSTANT VI		002 002	002 002		002 - POSITIVE	47	47	47	100	
2022-421	150	AIM STEP		002 002	002 002		001 - NEGATIVE	89	1	88	0	
2022-421	150	AIM STEP		002 002	002 002		002 - POSITIVE	89	88	88	100	
2022-421	164	HCG Alere		002 002	002 002		001 - NEGATIVE	71	1	70	0	
2022-421	164	HCG Alere		002 002	002 002		002 - POSITIVE	71	70	70	100	
2022-421	165	Osom HCG C		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022-421	166	FaStep		002 002	002 002		002 - POSITIVE	18	18	18	100	
2022-421	167	Tanner Sci		002 002	002 002		002 - POSITIVE	11	11	11	100	
2022-421	168	STANBIO 20		002 002	002 002		002 - POSITIVE	10	10	10	100	
031 hCG-QUALITATIVE												
2022-422	113	BIO-SIGN		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022-422	142	INSTANT VI		001 001	001 001		001 - NEGATIVE	47	47	47	100	
2022-422	150	AIM STEP		001 001	001 001		001 - NEGATIVE	89	87	87	100	
2022-422	150	AIM STEP		001 001	001 001		002 - POSITIVE	89	2	87	0	
2022-422	164	HCG Alere		001 001	001 001		001 - NEGATIVE	71	68	68	100	
2022-422	164	HCG Alere		001 001	001 001		002 - POSITIVE	71	3	68	0	
2022-422	165	Osom HCG C		001 001	001 001		001 - NEGATIVE	24	24	24	100	
2022-422	166	FaStep		001 001	001 001		001 - NEGATIVE	18	18	18	100	
2022-422	167	Tanner Sci		001 001	001 001		001 - NEGATIVE	11	11	11	100	
2022-422	168	STANBIO 20		001 001	001 001		001 - NEGATIVE	10	10	10	100	
2022-423	113	BIO-SIGN		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2022-423	142	INSTANT VI		001 001	001 001		002 - POSITIVE	47	5	42	0	
2022-423	142	INSTANT VI		001 001	001 001		001 - NEGATIVE	47	42	42	100	
2022-423	150	AIM STEP		001 001	001 001		002 - POSITIVE	89	17	72	0	
2022-423	150	AIM STEP		001 001	001 001		001 - NEGATIVE	89	72	72	100	
2022-423	164	HCG Alere		001 001	001 001		002 - POSITIVE	71	6	65	0	
2022-423	164	HCG Alere		001 001	001 001		001 - NEGATIVE	71	65	65	100	
2022-423	165	Osom HCG C		001 001	001 001		001 - NEGATIVE	24	24	24	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-423	166	FaStep		001 001	001 001		001 - NEGATIVE	18	18	18	100	
2022-423	167	Tanner Sci		001 001	001 001		002 - POSITIVE	11	1	10	0	
2022-423	167	Tanner Sci		001 001	001 001		001 - NEGATIVE	11	10	10	100	
2022-423	168	STANBIO 20		001 001	001 001		001 - NEGATIVE	10	8	8	100	
2022-423	168	STANBIO 20		001 001	001 001		002 - POSITIVE	10	2	8	0	
031 hCG-QUALITATIVE												
2022-424	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	22	22	22	100	
2022-424	142	INSTANT VI		002 002	002 002		002 - POSITIVE	47	47	47	100	
2022-424	150	AIM STEP		002 002	002 002		002 - POSITIVE	89	89	89	100	
2022-424	164	HCG Alere		002 002	002 002		001 - NEGATIVE	71	1	70	0	
2022-424	164	HCG Alere		002 002	002 002		002 - POSITIVE	71	70	70	100	
2022-424	165	Osom HCG C		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022-424	166	FaStep		002 002	002 002		002 - POSITIVE	18	18	18	100	
2022-424	167	Tanner Sci		002 002	002 002		002 - POSITIVE	11	11	11	100	
2022-424	168	STANBIO 20		002 002	002 002		002 - POSITIVE	10	10	10	100	
2022-425	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	22	22	22	100	
2022-425	142	INSTANT VI		002 002	002 002		002 - POSITIVE	47	47	47	100	
2022-425	150	AIM STEP		002 002	002 002		002 - POSITIVE	89	89	89	100	
2022-425	164	HCG Alere		002 002	002 002		002 - POSITIVE	71	71	71	100	
2022-425	165	Osom HCG C		002 002	002 002		002 - POSITIVE	24	24	24	100	
2022-425	166	FaStep		002 002	002 002		002 - POSITIVE	18	18	18	100	
2022-425	167	Tanner Sci		002 002	002 002		002 - POSITIVE	11	11	11	100	
2022-425	168	STANBIO 20		002 002	002 002		002 - POSITIVE	10	10	10	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2022-411	018	BIO SIGN		001	001	001	001	001 - NEGATIVE	23	23	23	100
2022-411	112	ONE STEP D		001	001	001	001	001 - NEGATIVE	10	10	10	100
2022-411	114	INSTANT VI		001	001	001	001	001 - NEGATIVE	110	110	110	100
2022-411	118	AIM STEP		001	001	001	001	001 - NEGATIVE	148	148	148	100
2022-411	130	ALERE HCG		001	001	001	001	001 - NEGATIVE	84	84	84	100
2022-411	131	OSOM HCG		001	001	001	001	001 - NEGATIVE	30	30	30	100
2022-411	133	FaStep		001	001	001	001	001 - NEGATIVE	32	32	32	100
2022-411	134	Tanner Sc.		001	001	001	001	001 - NEGATIVE	59	59	59	100
2022-411	135	STANBIO 20		001	001	001	001	001 - NEGATIVE	11	11	11	100
2022-411	136	CLARITY		001	001	001	001	001 - NEGATIVE	15	15	15	100
2022-412	018	BIO SIGN		002	002	002	002	002 - POSITIVE	23	23	23	100
2022-412	112	ONE STEP D		002	002	002	002	002 - POSITIVE	10	10	10	100
2022-412	114	INSTANT VI		002	002	002	002	002 - POSITIVE	110	110	110	100
2022-412	118	AIM STEP		002	002	002	002	002 - POSITIVE	148	148	148	100
2022-412	130	ALERE HCG		002	002	002	002	002 - POSITIVE	84	84	84	100
2022-412	131	OSOM HCG		002	002	002	002	002 - POSITIVE	30	30	30	100
2022-412	133	FaStep		002	002	002	002	002 - POSITIVE	32	32	32	100
2022-412	134	Tanner Sc.		002	002	002	002	002 - POSITIVE	59	59	59	100
2022-412	135	STANBIO 20		002	002	002	002	002 - POSITIVE	11	11	11	100
2022-412	136	CLARITY		002	002	002	002	002 - POSITIVE	15	15	15	100
2022-413	018	BIO SIGN		001	001	001	001	001 - NEGATIVE	23	23	23	100
2022-413	112	ONE STEP D		001	001	001	001	001 - NEGATIVE	10	10	10	100
2022-413	114	INSTANT VI		001	001	001	001	001 - NEGATIVE	110	110	110	100
2022-413	118	AIM STEP		001	001	001	001	001 - NEGATIVE	148	148	148	100
2022-413	130	ALERE HCG		001	001	001	001	001 - NEGATIVE	84	84	84	100
2022-413	131	OSOM HCG		001	001	001	001	001 - NEGATIVE	30	30	30	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 45 ENDOCRINOL.-hCG ONLY

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-413	133	FaStep		001 001	001 001		001 - NEGATIVE	32	32	32	100
2022-413	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	59	59	59	100
2022-413	135	STANBIO 20		001 001	001 001		001 - NEGATIVE	11	11	11	100
2022-413	136	CLARITY		001 001	001 001		001 - NEGATIVE	15	15	15	100
151 hCG ONLY QUALITATIVE											
2022-414	018	BIO SIGN		002 002	002 002		002 - POSITIVE	23	23	23	100
2022-414	112	ONE STEP D		002 002	002 002		002 - POSITIVE	10	10	10	100
2022-414	114	INSTANT VI		002 002	002 002		002 - POSITIVE	110	110	110	100
2022-414	118	AIM STEP		002 002	002 002		002 - POSITIVE	148	148	148	100
2022-414	130	ALERE HCG		002 002	002 002		002 - POSITIVE	84	84	84	100
2022-414	131	OSOM HCG		002 002	002 002		002 - POSITIVE	30	30	30	100
2022-414	133	FaStep		002 002	002 002		002 - POSITIVE	32	32	32	100
2022-414	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	59	59	59	100
2022-414	135	STANBIO 20		002 002	002 002		002 - POSITIVE	11	11	11	100
2022-414	136	CLARITY		002 002	002 002		002 - POSITIVE	15	15	15	100
2022-415	018	BIO SIGN		002 002	002 002		002 - POSITIVE	23	23	23	100
2022-415	112	ONE STEP D		002 002	002 002		002 - POSITIVE	10	10	10	100
2022-415	114	INSTANT VI		002 002	002 002		002 - POSITIVE	110	110	110	100
2022-415	118	AIM STEP		002 002	002 002		002 - POSITIVE	148	148	148	100
2022-415	130	ALERE HCG		002 002	002 002		002 - POSITIVE	84	84	84	100
2022-415	131	OSOM HCG		002 002	002 002		002 - POSITIVE	30	30	30	100
2022-415	133	FaStep		002 002	002 002		002 - POSITIVE	32	32	32	100
2022-415	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	59	59	59	100
2022-415	135	STANBIO 20		002 002	002 002		002 - POSITIVE	11	11	11	100
2022-415	136	CLARITY		002 002	002 002		002 - POSITIVE	15	15	15	100

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 55 ENDO.(T3 UPTAKE)

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 55 ENDO.(T3 UPTAKE)

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
173 T3 UPTAKE		CLIA TEST ID #: 565	EVALUATION CRITERIA: TARGET VALUE 3 SD								
2022-421	1 %U	84 VITROS ECI	1 25-35		31	55.119	1.323	2.40	51.150	59.090	
2022-421	1 %U	91 B.C.ACCESS	2 35-45		30	50.003	3.143	6.29	40.570	59.430	
2022-421	2 TUV	95 Arc i2000s	3 0.7 - 1.2		17	0.732	0.051	6.96	0.580	0.890	
2022-421	4 %	32 DIMENSION	2 35-45		49	40.533	2.322	5.73	33.570	47.500	
2022-421	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.840	0.045	5.38	0.700	0.980	
2022-421	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.778	0.075	9.64	0.550	1.000	
2022-422	1 %U	84 VITROS ECI	1 25-35		31	57.561	1.325	2.30	53.590	61.540	
2022-422	1 %U	91 B.C.ACCESS	2 35-45		30	52.359	3.194	6.10	42.780	61.940	
2022-422	2 TUV	95 Arc i2000s	3 0.7 - 1.2		17	0.624	0.040	6.39	0.500	0.740	
2022-422	4 %	32 DIMENSION	2 35-45		49	39.804	2.089	5.25	33.540	46.070	
2022-422	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.843	0.102	12.12	0.540	1.150	
2022-422	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.926	0.047	5.05	0.790	1.070	
2022-423	1 %U	84 VITROS ECI	1 25-35		31	29.574	1.050	3.55	26.430	32.720	
2022-423	1 %U	91 B.C.ACCESS	2 35-45		30	34.033	1.685	4.95	28.980	39.090	
2022-423	2 TUV	95 Arc i2000s	3 0.7 - 1.2		17	1.345	0.068	5.04	1.140	1.550	
2022-423	4 %	32 DIMENSION	2 35-45		49	28.842	1.866	6.47	23.240	34.440	
2022-423	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	1.003	0.176	17.51	0.480	1.530	
2022-423	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	1.128	0.037	3.25	1.020	1.240	
2022-424	1 %U	84 VITROS ECI	1 25-35		31	53.503	1.061	1.98	50.320	56.690	
2022-424	1 %U	91 B.C.ACCESS	2 35-45		30	50.377	2.901	5.76	41.670	59.080	
2022-424	2 TUV	95 Arc i2000s	3 0.7 - 1.2		17	0.802	0.048	5.98	0.660	0.950	
2022-424	4 %	32 DIMENSION	2 35-45		49	40.165	2.332	5.81	33.170	47.160	
2022-424	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.803	0.052	6.51	0.650	0.960	
2022-424	5 TBI	98 ROCHE COBA	4 0.8-1.3		48	0.750	0.067	8.99	0.550	0.950	
2022-425	1 %U	84 VITROS ECI	1 25-35		31	48.403	1.225	2.53	44.730	52.080	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY SUB: 30 ENDOCRINOLOGY GRP: 55 ENDO.(T3 UPTAKE)

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2022-425	1 %U	91	B.C.ACCESS	2	35-45	30	47.234	1.867	3.95	41.630	52.840		
2022-425	2 TUV	95	Arc i2000s	3	0.7 - 1.2	17	1.026	0.052	5.10	0.870	1.180		
2022-425	4 %	32	DIMENSION	2	35-45	49	40.655	2.488	6.12	33.190	48.120		
2022-425	5 TBI	98	ROCHE COBA	4	0.8-1.3	48	0.605	0.105	17.36	0.290	0.920		
2022-425	5 TBI	98	ROCHE COBA	4	0.8-1.3	48	0.629	0.059	9.39	0.450	0.810		

SPEC: 10 CHEMISTRY SUB: 40 TOXICOLOGY GRP: 60 TOXICOLOGY

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
043 DIGOXIN		CLIA TEST ID #: 645		EVALUATION CRITERIA: TARGET VALUE				20 %	OR 0.20	{GREATER}			
2022-421	50 VITROS					14	0.715	0.123	17.20	0.500	0.900		
2022-422	50 VITROS					14	0.538	0.084	15.52	0.300	0.700		
2022-423	50 VITROS					14	0.400	0.000	0.00	0.200	0.600		
2022-424	50 VITROS					14	0.792	0.092	11.57	0.600	1.000		
2022-425	50 VITROS					14	1.500	0.096	6.41	1.200	1.800		
045 FERRITIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE				25 %					
2022-421	50 VITROS					18	37.428	1.918	5.12	28.100	46.800		
2022-421	90 ROCHE COBA					14	61.214	2.132	3.48	45.900	76.500		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
045	FERRITIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %							
2022-422	50 VITROS			18	15.794	0.859	5.44	11.800	19.700	
2022-422	90 ROCHE COBA			14	24.793	1.040	4.20	18.600	31.000	
2022-423	50 VITROS			18	48.911	2.478	5.07	36.700	61.100	
2022-423	90 ROCHE COBA			14	79.754	2.849	3.57	59.800	99.700	
2022-424	50 VITROS			18	53.811	2.706	5.03	40.400	67.300	
2022-424	90 ROCHE COBA			14	88.121	3.878	4.40	66.100	110.200	
2022-425	50 VITROS			18	127.222	7.284	5.73	95.400	159.000	
2022-425	90 ROCHE COBA			14	204.586	8.357	4.08	153.400	255.700	
046	GENTAMICIN	CLIA TEST ID #: 665								
2022-421	50 VITROS			14	1.864	0.089	4.80	1.400	2.300	
2022-421	50 VITROS			14	1.513	0.232	15.31	1.100	1.900	
2022-422	50 VITROS			14	0.288	0.060	20.85	0.200	0.400	
2022-422	50 VITROS			14	0.600	0.000	0.00	0.500	0.800	
2022-423	50 VITROS			14	0.662	0.062	9.45	0.500	0.800	
2022-423	50 VITROS			14	0.450	0.187	41.57	0.300	0.600	
2022-424	50 VITROS			14	2.613	0.162	6.18	2.000	3.300	
2022-424	50 VITROS			14	3.086	0.125	4.04	2.300	3.900	
2022-425	50 VITROS			14	8.493	0.477	5.62	6.400	10.600	
2022-425	50 VITROS			14	7.350	0.158	2.15	5.500	9.200	
050	PHENYTOIN	CLIA TEST ID #: 695								
2022-421	50 VITROS			15	4.453	0.301	6.75	3.300	5.600	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 40 TOXICOLOGY

GRP: 60 TOXICOLOGY

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-421	92	ROCHE COBA		10	4.560	0.450	9.87	3.400	5.700
2022-421	92	ROCHE COBA		10	4.510	0.207	4.59	3.400	5.600
050 PHENYTOIN		CLIA TEST ID #: 695	EVALUATION CRITERIA: TARGET VALUE 25 %						
2022-422	50	VITROS		15	2.600	1.020	39.22	2.000	3.300
2022-422	92	ROCHE COBA		10	0.800	0.000	0.00	0.600	1.000
2022-422	92	ROCHE COBA		10	0.500	0.000	0.00	0.400	0.600
2022-423	50	VITROS		15	8.753	0.880	10.05	6.600	10.900
2022-423	92	ROCHE COBA		10	8.940	0.287	3.21	6.700	11.200
2022-423	92	ROCHE COBA		10	7.430	0.814	10.95	5.600	9.300
2022-424	50	VITROS		15	7.793	0.428	5.49	5.800	9.700
2022-424	92	ROCHE COBA		10	7.750	0.347	4.48	5.800	9.700
2022-424	92	ROCHE COBA		10	8.360	0.531	6.36	6.300	10.500
2022-425	50	VITROS		15	23.187	1.020	4.40	17.400	29.000
2022-425	92	ROCHE COBA		10	25.060	0.739	2.95	18.800	31.300
2022-425	92	ROCHE COBA		10	19.890	5.496	27.63	14.900	24.900
054 THEOPHYLLINE		CLIA TEST ID #: 735							
2022-421	50	VITROS		12	10.925	0.458	4.20	8.200	13.700
2022-422	50	VITROS		12	2.250	0.263	11.69	1.700	2.800
2022-423	50	VITROS		12	15.958	0.628	3.93	12.000	19.900
2022-424	50	VITROS		12	17.108	1.016	5.94	12.800	21.400
2022-425	50	VITROS		12	39.717	1.575	3.97	29.800	49.600

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
127	VANCOMYCIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%					
2022-421	50 VITROS			19	7.926	0.497	6.27	5.900	9.900	
2022-421	50 VITROS			19	8.500	0.383	4.51	6.400	10.600	
2022-421	92 ROCHE COBA			10	8.080	0.669	8.28	6.100	10.100	
2022-422	50 VITROS			19	2.933	1.267	43.20	2.200	3.700	
2022-422	50 VITROS			19	5.000	0.000	0.00	3.800	6.300	
2022-422	92 ROCHE COBA			10	4.000	0.000	0.00	3.000	5.000	
2022-423	50 VITROS			19	5.000	0.000	0.00	3.800	6.300	
2022-423	50 VITROS			19	3.217	0.634	19.70	2.400	4.000	
2022-423	92 ROCHE COBA			10	4.000	0.000	0.00	3.000	5.000	
2022-424	50 VITROS			19	14.667	0.692	4.72	11.000	18.300	
2022-424	50 VITROS			19	13.011	0.589	4.53	9.800	16.300	
2022-424	92 ROCHE COBA			10	13.330	0.642	4.82	10.000	16.700	
2022-425	50 VITROS			19	38.005	1.539	4.05	28.500	47.500	
2022-425	50 VITROS			19	44.017	2.154	4.89	33.000	55.000	
2022-425	92 ROCHE COBA			10	37.620	0.940	2.50	28.200	47.000	

SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
065	URINA-PROTEIN QUANT.		EVALUATION CRITERIA: TARGET VALUE	25	% OR 5.00	{GREATER}				
2022-451	58 VITROS			43	5.476	0.879	16.06	0.500	10.500	
2022-451	62 DADE DIM.			39	22.429	1.427	6.36	16.800	28.000	
2022-451	79 BECK. COUL			17	4.900	1.026	20.94	0	9.900	
2022-451	93 ARCHITECT			19	6.794	0.023	0.34	1.800	11.800	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-451	95	INTEGRA		10	2.500	0.901	36.04	0	7.500
2022-451	113	ROCHE COBA		32	4.063	0.400	9.85	0	9.100
2022-451	139	S. EXL.200		17	22.581	2.035	9.01	16.900	28.200
065 URINA-PROTEIN QUANT.		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2022-452	58	VITROS		43	5.000	0.000	0.00	0.000	10.000
2022-452	62	DADE DIM.		39	14.008	0.836	5.97	9.000	19.000
2022-452	79	BECK. COUL		17	4.365	0.683	15.64	0	9.400
2022-452	93	ARCHITECT		19	7.505	1.673	22.28	2.500	12.500
2022-452	95	INTEGRA		10	2.760	1.002	36.31	0	7.800
2022-452	113	ROCHE COBA		32	3.810	0.505	13.26	0	8.800
2022-452	139	S. EXL.200		17	15.019	1.103	7.34	10.000	20.000
2022-453	58	VITROS		43	160.905	7.006	4.35	120.700	201.100
2022-453	62	DADE DIM.		39	119.559	2.036	1.70	89.700	149.400
2022-453	79	BECK. COUL		17	109.306	6.533	5.98	82.000	136.600
2022-453	93	ARCHITECT		19	99.274	3.604	3.63	74.500	124.100
2022-453	95	INTEGRA		10	89.850	6.212	6.91	67.400	112.300
2022-453	113	ROCHE COBA		32	92.910	2.266	2.44	69.700	116.100
2022-453	139	S. EXL.200		17	119.650	1.871	1.56	89.700	149.600
2022-454	58	VITROS		43	167.048	7.244	4.34	125.300	208.800
2022-454	62	DADE DIM.		39	118.869	2.090	1.76	89.200	148.600
2022-454	79	BECK. COUL		17	103.453	7.108	6.87	77.600	129.300
2022-454	93	ARCHITECT		19	99.532	2.860	2.87	74.600	124.400
2022-454	95	INTEGRA		10	91.760	5.852	6.38	68.800	114.700
2022-454	113	ROCHE COBA		32	95.735	2.787	2.91	71.800	119.700
2022-454	139	S. EXL.200		17	118.982	1.560	1.31	89.200	148.700
2022-455	58	VITROS		43	134.286	9.632	7.17	100.700	167.900
2022-455	62	DADE DIM.		39	110.941	1.992	1.80	83.200	138.700

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-455	79	BECK. COUL		17	103.031	3.968	3.85	77.300	128.800
2022-455	93	ARCHITECT		19	97.179	3.648	3.75	72.900	121.500
2022-455	95	INTEGRA		10	87.820	5.808	6.61	65.900	109.800
2022-455	113	ROCHE COBA		32	91.155	2.429	2.66	68.400	113.900
2022-455	139	S. EXL.200		17	111.565	3.276	2.94	83.700	139.500
067	URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2022-451	11	AMES NM10S		46	1.013	0.004	0.38	1.002	1.025
2022-451	13	BAYER CLI		34	1.015	0.004	0.38	1.003	1.026
2022-451	67	URISCAN		55	1.011	0.004	0.41	0.999	1.024
2022-451	72	URS-9-TECO		91	1.016	0.004	0.44	1.003	1.029
2022-451	80	ROCHE (CHE		38	1.009	0.002	0.23	1.002	1.016
2022-451	89	AIMSTRIP		117	1.015	0.004	0.37	1.004	1.026
2022-451	94	CLINITEK		107	1.013	0.003	0.29	1.005	1.022
2022-451	102	ACON LAB		146	1.014	0.004	0.38	1.002	1.025
2022-451	131	iCHEM VELO		29	1.025	0.001	0.08	1.023	1.028
2022-451	135	Hybrido AU		12	1.024	0.002	0.16	1.019	1.028
2022-451	140	MISSION A		52	1.014	0.004	0.37	1.003	1.025
2022-451	145	CLINITEK		16	1.031	0.001	0.10	1.028	1.034
2022-451	147	Mission120		39	1.017	0.003	0.27	1.009	1.026
2022-451	148	Mission500		56	1.015	0.004	0.39	1.003	1.027
2022-452	11	AMES NM10S		46	1.022	0.003	0.28	1.014	1.031
2022-452	13	BAYER CLI		34	1.021	0.004	0.37	1.010	1.033
2022-452	67	URISCAN		55	1.022	0.003	0.34	1.011	1.032
2022-452	72	URS-9-TECO		91	1.022	0.003	0.29	1.013	1.031
2022-452	80	ROCHE (CHE		38	1.018	0.002	0.25	1.010	1.025
2022-452	89	AIMSTRIP		117	1.022	0.003	0.31	1.012	1.031
2022-452	94	CLINITEK		107	1.023	0.003	0.32	1.013	1.032
2022-452	102	ACON LAB		146	1.022	0.003	0.31	1.013	1.032
2022-452	131	iCHEM VELO		29	1.027	0.001	0.09	1.024	1.030

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2022-452	135	Hybrido AU		12	1.025	0.001	0.13	1.022	1.029
2022-452	140	MISSION A		52	1.022	0.003	0.28	1.014	1.031
2022-452	145	CLINITEK		16	1.033	0.001	0.11	1.030	1.036
2022-452	147	Mission120		39	1.022	0.002	0.23	1.015	1.029
2022-452	148	Mission500		56	1.022	0.003	0.26	1.014	1.030
067	URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2022-453	11	AMES NM10S		46	1.022	0.004	0.34	1.011	1.032
2022-453	13	BAYER CLI		34	1.022	0.003	0.29	1.013	1.031
2022-453	67	URISCAN		55	1.019	0.003	0.28	1.010	1.027
2022-453	72	URS-9-TECO		91	1.022	0.003	0.32	1.012	1.032
2022-453	80	ROCHE (CHE		38	1.013	0.003	0.27	1.005	1.021
2022-453	89	AIMSTRIP		117	1.019	0.002	0.24	1.012	1.027
2022-453	94	CLINITEK		107	1.022	0.003	0.25	1.014	1.030
2022-453	102	ACON LAB		146	1.020	0.003	0.27	1.012	1.028
2022-453	131	iCHEM VELO		29	1.026	0.001	0.09	1.023	1.029
2022-453	135	Hybrido AU		12	1.024	0.001	0.14	1.020	1.028
2022-453	140	MISSION A		52	1.020	0.003	0.28	1.011	1.029
2022-453	145	CLINITEK		16	1.027	0.001	0.11	1.024	1.031
2022-453	147	Mission120		39	1.022	0.003	0.30	1.013	1.031
2022-453	148	Mission500		56	1.020	0.000	0.00	1.020	1.020
2022-454	11	AMES NM10S		46	1.016	0.004	0.36	1.005	1.027
2022-454	13	BAYER CLI		34	1.016	0.003	0.31	1.007	1.026
2022-454	67	URISCAN		55	1.013	0.003	0.29	1.004	1.022
2022-454	72	URS-9-TECO		91	1.018	0.004	0.39	1.006	1.030
2022-454	80	ROCHE (CHE		38	1.011	0.003	0.27	1.003	1.020
2022-454	89	AIMSTRIP		117	1.014	0.004	0.39	1.003	1.026
2022-454	94	CLINITEK		107	1.016	0.003	0.29	1.007	1.024
2022-454	102	ACON LAB		146	1.016	0.004	0.38	1.004	1.028
2022-454	131	iCHEM VELO		29	1.028	0.001	0.08	1.026	1.031

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-454	135	Hybrido AU	12	1.027	0.001	0.12	1.024	1.031
2022-454	140	MISSION A	52	1.016	0.004	0.42	1.003	1.029
2022-454	145	CLINITEK	16	1.031	0.000	0.05	1.029	1.032
2022-454	147	Mission120	39	1.019	0.002	0.19	1.013	1.025
2022-454	148	Mission500	56	1.019	0.002	0.20	1.013	1.025
067 URI-SPECIFIC GRAVITY			CLIA TEST ID #: 0 EVALUATION CRITERIA: TARGET VALUE 3 SD					
2022-455	11	AMES NM10S	46	1.024	0.003	0.30	1.015	1.034
2022-455	13	BAYER CLI	34	1.025	0.002	0.22	1.018	1.031
2022-455	67	URISCAN	55	1.023	0.004	0.39	1.011	1.035
2022-455	72	URS-9-TECO	91	1.024	0.002	0.22	1.017	1.031
2022-455	80	ROCHE (CHE	38	1.015	0.000	0.00	1.015	1.015
2022-455	89	AIMSTRIP	117	1.022	0.002	0.24	1.014	1.029
2022-455	94	CLINITEK	107	1.024	0.002	0.22	1.018	1.031
2022-455	102	ACON LAB	146	1.023	0.003	0.26	1.015	1.031
2022-455	131	iCHEM VELO	29	1.025	0.001	0.09	1.022	1.028
2022-455	135	Hybrido AU	12	1.022	0.002	0.15	1.018	1.027
2022-455	140	MISSION A	52	1.024	0.003	0.31	1.014	1.033
2022-455	145	CLINITEK	16	1.030	0.001	0.09	1.027	1.033
2022-455	147	Mission120	39	1.024	0.003	0.30	1.015	1.034
2022-455	148	Mission500	56	1.021	0.002	0.22	1.014	1.028
166 MICROALBUMIN			EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2022-451	58	VITROS	43	6.050	0.320	5.29	1.100	11.100
2022-451	62	DADE DIM.	55	2.600	1.200	46.15	0	7.600
2022-451	62	DADE DIM.	55	2.411	1.303	54.05	0	7.400
2022-451	79	BECK. COUL	20	2.220	2.665	120.06	0	7.200
2022-451	85	Rx DAYTONA	16	0.838	1.671	199.48	0	5.800
2022-451	85	Rx DAYTONA	16	2.600	1.200	46.15	0	7.600
2022-451	93	ARCHITECT	22	4.995	0.021	0.43	0.000	10.000
2022-451	95	INTEGRA	22	2.718	1.177	43.31	0	7.700
2022-451	113	ROCHE COBA	33	11.710	1.590	13.58	6.700	16.700

Participant Summary
MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-451	113	ROCHE COBA		33	2.600	1.200	46.15	0	7.600	
2022-451	139	S. EXL.200		30	2.600	1.200	46.15	0	7.600	
2022-451	139	S. EXL.200		30	2.377	1.255	52.82	0	7.400	
166 MICROALBUMIN			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2022-452	58	VITROS		43	6.024	0.152	2.53	1.000	11.000	
2022-452	62	DADE DIM.		55	3.127	1.355	43.31	0	8.100	
2022-452	62	DADE DIM.		55	2.200	0.980	44.54	0	7.200	
2022-452	79	BECK. COUL		20	2.230	2.686	120.43	0	7.200	
2022-452	85	Rx DAYTONA		16	2.200	0.980	44.54	0	7.200	
2022-452	85	Rx DAYTONA		16	1.094	1.770	161.83	0	6.100	
2022-452	93	ARCHITECT		22	4.995	0.021	0.43	0.000	10.000	
2022-452	95	INTEGRA		22	2.750	1.104	40.15	0	7.800	
2022-452	113	ROCHE COBA		33	2.200	0.980	44.54	0	7.200	
2022-452	113	ROCHE COBA		33	11.710	1.590	13.58	6.700	16.700	
2022-452	139	S. EXL.200		30	2.200	0.980	44.54	0	7.200	
2022-452	139	S. EXL.200		30	3.123	1.306	41.81	0	8.100	
2022-453	58	VITROS		43	893.755	30.244	3.38	670.300	1117.200	
2022-453	62	DADE DIM.		55	893.600	7.684	0.86	670.200	1117.000	
2022-453	62	DADE DIM.		55	877.432	49.049	5.59	658.100	1096.800	
2022-453	79	BECK. COUL		20	847.465	67.272	7.94	635.600	1059.300	
2022-453	85	Rx DAYTONA		16	893.600	7.684	0.86	670.200	1117.000	
2022-453	85	Rx DAYTONA		16	760.600	109.344	14.38	570.500	950.800	
2022-453	93	ARCHITECT		22	877.995	34.544	3.93	658.500	1097.500	
2022-453	95	INTEGRA		22	897.805	25.409	2.83	673.400	1122.300	
2022-453	113	ROCHE COBA		33	846.282	32.422	3.83	634.700	1057.900	
2022-453	113	ROCHE COBA		33	893.600	7.684	0.86	670.200	1117.000	
2022-453	139	S. EXL.200		30	861.786	56.739	6.58	646.300	1077.200	
2022-453	139	S. EXL.200		30	893.600	7.684	0.86	670.200	1117.000	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
166	MICROALBUMIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25 % OR 5.00	{GREATER}					
2022-454	58	VITROS			43	849.293	36.358	4.28	637.000	1061.600
2022-454	62	DADE DIM.			55	858.498	44.125	5.14	643.900	1073.100
2022-454	62	DADE DIM.			55	858.600	21.951	2.56	644.000	1073.300
2022-454	79	BECK. COUL			20	810.960	67.314	8.30	608.200	1013.700
2022-454	85	Rx DAYTONA			16	751.773	94.406	12.56	563.800	939.700
2022-454	85	Rx DAYTONA			16	858.600	21.951	2.56	644.000	1073.300
2022-454	93	ARCHITECT			22	839.795	26.772	3.19	629.800	1049.700
2022-454	95	INTEGRA			22	852.364	31.025	3.64	639.300	1065.500
2022-454	113	ROCHE COBA			33	858.600	21.951	2.56	644.000	1073.300
2022-454	113	ROCHE COBA			33	808.985	37.418	4.63	606.700	1011.200
2022-454	139	S. EXL.200			30	858.600	21.951	2.56	644.000	1073.300
2022-454	139	S. EXL.200			30	867.290	84.160	9.70	650.500	1084.100
2022-455	58	VITROS			43	883.802	39.410	4.46	662.900	1104.800
2022-455	62	DADE DIM.			55	864.400	18.650	2.16	648.300	1080.500
2022-455	62	DADE DIM.			55	868.257	47.333	5.45	651.200	1085.300
2022-455	79	BECK. COUL			20	813.505	62.241	7.65	610.100	1016.900
2022-455	85	Rx DAYTONA			16	864.400	18.650	2.16	648.300	1080.500
2022-455	85	Rx DAYTONA			16	764.533	81.441	10.65	573.400	955.700
2022-455	93	ARCHITECT			22	856.505	27.826	3.25	642.400	1070.600
2022-455	95	INTEGRA			22	860.236	30.620	3.56	645.200	1075.300
2022-455	113	ROCHE COBA			33	816.672	36.771	4.50	612.500	1020.800
2022-455	113	ROCHE COBA			33	864.400	18.650	2.16	648.300	1080.500
2022-455	139	S. EXL.200			30	865.852	59.402	6.86	649.400	1082.300
2022-455	139	S. EXL.200			30	864.400	18.650	2.16	648.300	1080.500
167	CREATININE									
2022-451	58	VITROS			50	250.414	10.321	4.12	187.800	313.000
2022-451	62	DADE DIM.			51	240.998	11.159	4.63	180.700	301.200
2022-451	79	BECK. COUL			18	219.556	10.678	4.86	164.700	274.400

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-451	93	ARCHITECT		21	233.833	6.225	2.66	175.400	292.300
2022-451	95	INTEGRA		19	257.411	10.440	4.06	193.100	321.800
2022-451	113	ROCHE COBA		34	249.670	7.691	3.08	187.300	312.100
2022-451	139	S. EXL.200		26	244.258	7.137	2.92	183.200	305.300
167 CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2022-452	58	VITROS		50	3.200	0.000	0.00	0	8.200
2022-452	62	DADE DIM.		51	9.473	5.194	54.83	4.500	14.500
2022-452	79	BECK. COUL		18	0.618	1.406	227.57	0	5.600
2022-452	93	ARCHITECT		21	5.000	0.000	0.00	0.000	10.000
2022-452	95	INTEGRA		19	0.184	0.292	158.77	0	5.200
2022-452	113	ROCHE COBA		34	3.476	1.285	36.96	0	8.500
2022-452	139	S. EXL.200		26	8.112	6.188	76.29	3.100	13.100
2022-453	58	VITROS		50	239.071	7.125	2.98	179.300	298.800
2022-453	62	DADE DIM.		51	243.248	11.202	4.61	182.400	304.100
2022-453	79	BECK. COUL		18	221.883	11.422	5.15	166.400	277.400
2022-453	93	ARCHITECT		21	235.200	6.032	2.56	176.400	294.000
2022-453	95	INTEGRA		19	258.574	10.966	4.24	193.900	323.200
2022-453	113	ROCHE COBA		34	252.273	8.866	3.51	189.200	315.300
2022-453	139	S. EXL.200		26	247.248	6.848	2.77	185.400	309.100
2022-454	58	VITROS		50	234.139	7.235	3.09	175.600	292.700
2022-454	62	DADE DIM.		51	230.500	9.526	4.13	172.900	288.100
2022-454	79	BECK. COUL		18	200.783	12.328	6.14	150.600	251.000
2022-454	93	ARCHITECT		21	224.286	5.687	2.54	168.200	280.400
2022-454	95	INTEGRA		19	243.516	10.559	4.34	182.600	304.400
2022-454	113	ROCHE COBA		34	239.609	7.827	3.27	179.700	299.500
2022-454	139	S. EXL.200		26	233.832	6.021	2.57	175.400	292.300
2022-455	58	VITROS		50	3.200	0.000	0.00	0	8.200

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS			
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2022-455	62	DADE DIM.		51	9.476	5.191	54.77	4.500	14.500		
2022-455	79	BECK. COUL		18	0.771	1.359	176.35	0	5.800		
2022-455	93	ARCHITECT		21	5.000	0.000	0.00	0.000	10.000		
2022-455	95	INTEGRA		19	0.411	0.348	84.71	0	5.400		
2022-455	113	ROCHE COBA		34	3.447	1.349	39.13	0	8.400		
2022-455	139	S. EXL.200		26	8.369	6.529	78.01	3.400	13.400		

SPE: 10 CHEMISTRY				SUB: 50 URINALYSIS				GRP: 70 URINALYSIS				
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART #	LABS #	"SF"	Score	
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
057 URINALYSIS-BILIRUBIN												
2022-451	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	46	46	46	100
2022-451	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	34	34	34	100
2022-451	067	URISCAN		057	057	057	057	057 - NEGATIVE	54	54	54	100
2022-451	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	89	89	89	100
2022-451	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	36	36	36	100
2022-451	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	119	119	119	100
2022-451	094	CLINITEK		057	057	057	057	057 - NEGATIVE	106	106	106	100
2022-451	102	ACON LAB		057	057	057	057	057 - NEGATIVE	142	142	142	100
2022-451	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	28	28	28	100
2022-451	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	12	12	12	100
2022-451	140	MISSION A		057	057	057	057	057 - NEGATIVE	50	50	50	100
2022-451	145	CLINITEK		057	057	057	057	057 - NEGATIVE	16	16	16	100
2022-451	147	Mission120		057	057	057	057	057 - NEGATIVE	41	41	41	100
2022-451	148	Mission500		057	057	057	057	057 - NEGATIVE	59	59	59	100
2022-451	151	Clari. Uro		057	057	057	057	057 - NEGATIVE	11	11	11	100
2022-452	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	46	46	46	100
2022-452	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	34	34	34	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	067	URISCAN		057 057	057 057		057 - NEGATIVE	54	54	54	100
2022-452	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	89	89	89	100
2022-452	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	36	36	36	100
2022-452	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	119	119	119	100
2022-452	094	CLINITEK		057 057	057 057		057 - NEGATIVE	106	106	106	100
2022-452	102	ACON LAB		057 057	057 057		057 - NEGATIVE	142	142	142	100
2022-452	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	28	28	28	100
2022-452	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	12	12	12	100
2022-452	140	MISSION A		057 057	057 057		057 - NEGATIVE	50	50	50	100
2022-452	145	CLINITEK		057 057	057 057		057 - NEGATIVE	16	16	16	100
2022-452	147	Mission120		057 057	057 057		057 - NEGATIVE	41	41	41	100
2022-452	148	Mission500		057 057	057 057		057 - NEGATIVE	59	59	59	100
2022-452	151	Clari. Uro		057 057	057 057		057 - NEGATIVE	11	11	11	100
057 URINALYSIS-BILIRUBIN											
2022-453	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	46	46	46	100
2022-453	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	34	34	34	100
2022-453	067	URISCAN		057 057	057 057		057 - NEGATIVE	54	54	54	100
2022-453	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	89	89	89	100
2022-453	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	36	36	36	100
2022-453	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	119	119	119	100
2022-453	094	CLINITEK		057 057	057 057		057 - NEGATIVE	106	106	106	100
2022-453	102	ACON LAB		057 057	057 057		057 - NEGATIVE	142	142	142	100
2022-453	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	28	28	28	100
2022-453	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	12	12	12	100
2022-453	140	MISSION A		057 057	057 057		057 - NEGATIVE	50	50	50	100
2022-453	145	CLINITEK		057 057	057 057		057 - NEGATIVE	16	16	16	100
2022-453	147	Mission120		057 057	057 057		057 - NEGATIVE	41	41	41	100
2022-453	148	Mission500		057 057	057 057		057 - NEGATIVE	59	59	59	100
2022-453	151	Clari. Uro		057 057	057 057		057 - NEGATIVE	11	11	11	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
057 URINALYSIS-BILIRUBIN											
2022-454	011	AMES NM10S		058 060	058 060		060 - 3+	46	5	46	100
2022-454	011	AMES NM10S		058 060	058 060		059 - 2+	46	31	46	100
2022-454	011	AMES NM10S		058 060	058 060		058 - TRACES, 1+	46	10	46	100
2022-454	013	BAYER CLI		058 060	058 060		058 - TRACES, 1+	34	7	34	100
2022-454	013	BAYER CLI		058 060	058 060		059 - 2+	34	26	34	100
2022-454	013	BAYER CLI		058 060	058 060		060 - 3+	34	1	34	100
2022-454	067	URISCAN		057 059	058 060		059 - 2+	54	8	54	100
2022-454	067	URISCAN		057 059	058 060		058 - TRACES, 1+	54	42	54	100
2022-454	067	URISCAN		057 059	058 060		057 - NEGATIVE	54	4	54	100
2022-454	072	URS-9-TECO		058 060	058 060		060 - 3+	89	22	89	100
2022-454	072	URS-9-TECO		058 060	058 060		059 - 2+	89	43	89	100
2022-454	072	URS-9-TECO		058 060	058 060		058 - TRACES, 1+	89	24	89	100
2022-454	080	ROCHE (CHE		057 059	058 060		058 - TRACES, 1+	36	34	36	100
2022-454	080	ROCHE (CHE		057 059	058 060		059 - 2+	36	2	36	100
2022-454	089	AIMSTRIP		057 059	058 060		059 - 2+	119	38	119	100
2022-454	089	AIMSTRIP		057 059	058 060		058 - TRACES, 1+	119	70	119	100
2022-454	089	AIMSTRIP		057 059	058 060		060 - 3+	119	5	119	100
2022-454	089	AIMSTRIP		057 059	058 060		057 - NEGATIVE	119	6	119	100
2022-454	094	CLINITEK		058 060	058 060		058 - TRACES, 1+	106	27	106	100
2022-454	094	CLINITEK		058 060	058 060		059 - 2+	106	76	106	100
2022-454	094	CLINITEK		058 060	058 060		060 - 3+	106	3	106	100
2022-454	102	ACON LAB		057 059	058 060		060 - 3+	142	4	142	100
2022-454	102	ACON LAB		057 059	058 060		057 - NEGATIVE	142	1	142	100
2022-454	102	ACON LAB		057 059	058 060		058 - TRACES, 1+	142	104	142	100
2022-454	102	ACON LAB		057 059	058 060		059 - 2+	142	33	142	100
2022-454	131	ICHEM VELO		057 057	058 060		057 - NEGATIVE	28	28	28	100
2022-454	135	Hybrido AU		058 060	058 060		059 - 2+	12	12	12	100
2022-454	140	MISSION A		057 059	058 060		059 - 2+	50	15	50	100
2022-454	140	MISSION A		057 059	058 060		058 - TRACES, 1+	50	34	50	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	140	MISSION A		057 059	058 060		057 - NEGATIVE	50	1	50	100
2022-454	145	CLINITEK		057 057	058 060		057 - NEGATIVE	16	16	16	100
2022-454	147	Mission120		057 059	058 060		059 - 2+	41	6	41	100
2022-454	147	Mission120		057 059	058 060		058 - TRACES,1+	41	35	41	100
2022-454	148	Mission500		057 059	058 060		058 - TRACES,1+	59	50	59	100
2022-454	148	Mission500		057 059	058 060		059 - 2+	59	9	59	100
2022-454	151	Clari. Uro		058 060	058 060		059 - 2+	11	6	11	100
2022-454	151	Clari. Uro		058 060	058 060		058 - TRACES,1+	11	5	11	100
057 URINALYSIS-BILIRUBIN											
2022-455	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	46	46	46	100
2022-455	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	34	34	34	100
2022-455	067	URISCAN		057 057	057 057		057 - NEGATIVE	54	54	54	100
2022-455	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	89	89	89	100
2022-455	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	36	36	36	100
2022-455	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	119	119	119	100
2022-455	094	CLINITEK		057 057	057 057		057 - NEGATIVE	106	106	106	100
2022-455	102	ACON LAB		057 057	057 057		057 - NEGATIVE	142	142	142	100
2022-455	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	28	28	28	100
2022-455	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	12	12	12	100
2022-455	140	MISSION A		057 057	057 057		057 - NEGATIVE	50	50	50	100
2022-455	145	CLINITEK		057 057	057 057		057 - NEGATIVE	16	16	16	100
2022-455	147	Mission120		057 057	057 057		057 - NEGATIVE	41	41	41	100
2022-455	148	Mission500		057 057	057 057		057 - NEGATIVE	59	59	59	100
2022-455	151	Clari. Uro		057 057	057 057		057 - NEGATIVE	11	11	11	100
058 URINALYSIS-GLUCOSE											
2022-451	011	AMES NM10S		076 078	075 077		078 - 2000mg/dl	46	1	45	100
2022-451	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	46	33	45	100
2022-451	011	AMES NM10S		076 078	075 077		074 - TRACES	46	1	45	0
2022-451	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	46	11	45	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-451	013	BAYER CLI		076 078	075 077		077 - 1000mg/dl	33	18	33	100		
2022-451	013	BAYER CLI		076 078	075 077		078 - 2000mg/dl	33	1	33	100		
2022-451	013	BAYER CLI		076 078	075 077		076 - 500mg/dl2+	33	14	33	100		
2022-451	067	URISCAN		077 078	075 077		078 - 2000mg/dl	54	29	54	100		
2022-451	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	54	6	54	100		
2022-451	067	URISCAN		077 078	075 077		077 - 1000mg/dl	54	19	54	100		
2022-451	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	91	45	91	100		
2022-451	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	91	35	91	100		
2022-451	072	URS-9-TECO		076 078	075 077		075 - 250mg/dl-+	91	1	91	100		
2022-451	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	91	10	91	100		
2022-451	080	ROCHE (CHE		076 078	075 077		075 - 250mg/dl-+	38	1	38	100		
2022-451	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	38	3	38	100		
2022-451	080	ROCHE (CHE		076 078	075 077		076 - 500mg/dl2+	38	1	38	100		
2022-451	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	38	33	38	100		
2022-451	089	AIMSTRIP		076 078	075 077		076 - 500mg/dl2+	119	54	118	100		
2022-451	089	AIMSTRIP		076 078	075 077		073 - NEGATIVE	119	1	118	0		
2022-451	089	AIMSTRIP		076 078	075 077		077 - 1000mg/dl	119	60	118	100		
2022-451	089	AIMSTRIP		076 078	075 077		078 - 2000mg/dl	119	1	118	100		
2022-451	089	AIMSTRIP		076 078	075 077		075 - 250mg/dl-+	119	3	118	100		
2022-451	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	107	2	107	100		
2022-451	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	107	35	107	100		
2022-451	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	107	69	107	100		
2022-451	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	107	1	107	100		
2022-451	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	145	1	145	100		
2022-451	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	145	84	145	100		
2022-451	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	145	60	145	100		
2022-451	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	29	2	29	100		
2022-451	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	29	27	29	100		
2022-451	135	Hybrido AU		077 078	075 077		078 - 2000mg/dl	12	7	12	100		
2022-451	135	Hybrido AU		077 078	075 077		077 - 1000mg/dl	12	5	12	100		

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-451	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	10	4	10	100
2022-451	138	URS-10T		076 078	075 077		077 - 1000mg/dl	10	6	10	100
2022-451	140	MISSION A		076 078	075 077		078 - 2000mg/dl	51	8	51	100
2022-451	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	51	12	51	100
2022-451	140	MISSION A		076 078	075 077		077 - 1000mg/dl	51	31	51	100
2022-451	145	CLINITEK		075 077	075 077		074 - TRACES	16	1	15	0
2022-451	145	CLINITEK		075 077	075 077		077 - 1000mg/dl	16	5	15	100
2022-451	145	CLINITEK		075 077	075 077		076 - 500mg/dl2+	16	10	15	100
2022-451	147	Mission120		076 078	075 077		076 - 500mg/dl2+	39	3	39	100
2022-451	147	Mission120		076 078	075 077		078 - 2000mg/dl	39	9	39	100
2022-451	147	Mission120		076 078	075 077		077 - 1000mg/dl	39	27	39	100
2022-451	148	Mission500		076 078	075 077		076 - 500mg/dl2+	58	22	58	100
2022-451	148	Mission500		076 078	075 077		077 - 1000mg/dl	58	26	58	100
2022-451	148	Mission500		076 078	075 077		078 - 2000mg/dl	58	9	58	100
2022-451	148	Mission500		076 078	075 077		075 - 250mg/dl-+	58	1	58	100
2022-451	151	Clari. Uro		076 078	075 077		076 - 500mg/dl2+	10	3	10	100
2022-451	151	Clari. Uro		076 078	075 077		077 - 1000mg/dl	10	7	10	100
058 URINALYSIS-GLUCOSE											
2022-452	011	AMES NM10S		074 076	074 076		074 - TRACES	46	19	46	100
2022-452	011	AMES NM10S		074 076	074 076		075 - 250mg/dl-+	46	27	46	100
2022-452	013	BAYER CLI		073 075	074 076		074 - TRACES	33	19	33	100
2022-452	013	BAYER CLI		073 075	074 076		075 - 250mg/dl-+	33	13	33	100
2022-452	013	BAYER CLI		073 075	074 076		076 - 500mg/dl2+	33	1	33	100
2022-452	067	URISCAN		074 076	074 076		076 - 500mg/dl2+	54	14	49	100
2022-452	067	URISCAN		074 076	074 076		078 - 2000mg/dl	54	1	49	0
2022-452	067	URISCAN		074 076	074 076		075 - 250mg/dl-+	54	31	49	100
2022-452	067	URISCAN		074 076	074 076		074 - TRACES	54	4	49	100
2022-452	067	URISCAN		074 076	074 076		077 - 1000mg/dl	54	4	49	0
2022-452	072	URS-9-TECO		074 076	074 076		073 - NEGATIVE	91	5	86	0
2022-452	072	URS-9-TECO		074 076	074 076		074 - TRACES	91	27	86	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	072	URS-9-TECO		074 076	074 076		075 - 250mg/dl-+	91	56	86	100
2022-452	072	URS-9-TECO		074 076	074 076		076 - 500mg/dl2+	91	3	86	100
2022-452	080	ROCHE (CHE		076 078	074 076		077 - 1000mg/dl	38	35	38	100
2022-452	080	ROCHE (CHE		076 078	074 076		078 - 2000mg/dl	38	3	38	100
2022-452	089	AIMSTRIP		074 076	074 076		075 - 250mg/dl-+	119	97	116	100
2022-452	089	AIMSTRIP		074 076	074 076		073 - NEGATIVE	119	3	116	0
2022-452	089	AIMSTRIP		074 076	074 076		074 - TRACES	119	19	116	100
2022-452	094	CLINITEK		074 076	074 076		076 - 500mg/dl2+	107	1	107	100
2022-452	094	CLINITEK		074 076	074 076		074 - TRACES	107	48	107	100
2022-452	094	CLINITEK		074 076	074 076		075 - 250mg/dl-+	107	58	107	100
2022-452	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	145	122	141	100
2022-452	102	ACON LAB		074 076	074 076		073 - NEGATIVE	145	4	141	0
2022-452	102	ACON LAB		074 076	074 076		074 - TRACES	145	19	141	100
2022-452	131	ICHEM VELO		075 077	074 076		076 - 500mg/dl2+	29	26	29	100
2022-452	131	ICHEM VELO		075 077	074 076		077 - 1000mg/dl	29	2	29	100
2022-452	131	ICHEM VELO		075 077	074 076		075 - 250mg/dl-+	29	1	29	100
2022-452	135	Hybrido AU		077 078	074 076		078 - 2000mg/dl	12	7	11	100
2022-452	135	Hybrido AU		077 078	074 076		073 - NEGATIVE	12	1	11	0
2022-452	135	Hybrido AU		077 078	074 076		077 - 1000mg/dl	12	4	11	100
2022-452	138	URS-10T		073 075	074 076		075 - 250mg/dl-+	10	2	10	100
2022-452	138	URS-10T		073 075	074 076		074 - TRACES	10	6	10	100
2022-452	138	URS-10T		073 075	074 076		073 - NEGATIVE	10	1	10	100
2022-452	138	URS-10T		073 075	074 076		076 - 500mg/dl2+	10	1	10	100
2022-452	140	MISSION A		074 076	074 076		074 - TRACES	51	2	51	100
2022-452	140	MISSION A		074 076	074 076		075 - 250mg/dl-+	51	49	51	100
2022-452	145	CLINITEK		076 078	074 076		076 - 500mg/dl2+	16	3	16	100
2022-452	145	CLINITEK		076 078	074 076		077 - 1000mg/dl	16	10	16	100
2022-452	145	CLINITEK		076 078	074 076		078 - 2000mg/dl	16	3	16	100
2022-452	147	Mission120		074 076	074 076		075 - 250mg/dl-+	39	39	39	100
2022-452	148	Mission500		074 076	074 076		075 - 250mg/dl-+	58	58	58	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-452	151	Clari. Uro		074 076	074 076		074 - TRACES	10	1	10	100	
2022-452	151	Clari. Uro		074 076	074 076		075 - 250mg/dl-+	10	9	10	100	
058 URINALYSIS-GLUCOSE												
2022-453	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	46	46	46	100	
2022-453	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	33	33	33	100	
2022-453	067	URISCAN		073 073	073 073		073 - NEGATIVE	54	53	53	100	
2022-453	067	URISCAN		073 073	073 073		074 - TRACES	54	1	53	0	
2022-453	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	91	91	91	100	
2022-453	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	38	38	38	100	
2022-453	089	AIMSTRIP		073 073	073 073		074 - TRACES	119	1	118	0	
2022-453	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	119	118	118	100	
2022-453	094	CLINITEK		073 073	073 073		073 - NEGATIVE	107	107	107	100	
2022-453	102	ACON LAB		073 073	073 073		073 - NEGATIVE	145	145	145	100	
2022-453	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	29	29	29	100	
2022-453	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	11	11	100	
2022-453	135	Hybrido AU		073 073	073 073		077 - 1000mg/dl	12	1	11	0	
2022-453	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2022-453	140	MISSION A		073 073	073 073		073 - NEGATIVE	51	51	51	100	
2022-453	145	CLINITEK		073 073	073 073		073 - NEGATIVE	16	16	16	100	
2022-453	147	Mission120		073 073	073 073		073 - NEGATIVE	39	39	39	100	
2022-453	148	Mission500		073 073	073 073		073 - NEGATIVE	58	58	58	100	
2022-453	151	Clari. Uro		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2022-454	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	46	46	46	100	
2022-454	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	33	33	33	100	
2022-454	067	URISCAN		073 073	073 073		073 - NEGATIVE	54	54	54	100	
2022-454	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	91	91	91	100	
2022-454	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	38	38	38	100	
2022-454	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	119	119	119	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	094	CLINITEK		073 073	073 073		073 - NEGATIVE	107	107	107	100
2022-454	102	ACON LAB		073 073	073 073		073 - NEGATIVE	145	145	145	100
2022-454	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	29	28	28	100
2022-454	131	iCHEM VELO		073 073	073 073		074 - TRACES	29	1	28	0
2022-454	135	Hybrido AU		073 073	073 073		077 - 1000mg/dl	12	1	11	0
2022-454	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	12	11	11	100
2022-454	138	URS-10T		073 073	073 073		073 - NEGATIVE	10	10	10	100
2022-454	140	MISSION A		073 073	073 073		073 - NEGATIVE	51	51	51	100
2022-454	145	CLINITEK		073 073	073 073		073 - NEGATIVE	16	16	16	100
2022-454	147	Mission120		073 073	073 073		073 - NEGATIVE	39	39	39	100
2022-454	148	Mission500		073 073	073 073		073 - NEGATIVE	58	58	58	100
2022-454	151	Clari. Uro		073 073	073 073		073 - NEGATIVE	10	10	10	100
058 URINALYSIS-GLUCOSE											
2022-455	011	AMES NM10S		076 078	075 077		077 - 1000mg/dl	46	34	46	100
2022-455	011	AMES NM10S		076 078	075 077		075 - 250mg/dl-+	46	1	46	100
2022-455	011	AMES NM10S		076 078	075 077		078 - 2000mg/dl	46	1	46	100
2022-455	011	AMES NM10S		076 078	075 077		076 - 500mg/dl2+	46	10	46	100
2022-455	013	BAYER CLI		076 078	075 077		077 - 1000mg/dl	33	19	33	100
2022-455	013	BAYER CLI		076 078	075 077		078 - 2000mg/dl	33	1	33	100
2022-455	013	BAYER CLI		076 078	075 077		076 - 500mg/dl2+	33	13	33	100
2022-455	067	URISCAN		077 078	075 077		078 - 2000mg/dl	54	27	54	100
2022-455	067	URISCAN		077 078	075 077		076 - 500mg/dl2+	54	6	54	100
2022-455	067	URISCAN		077 078	075 077		077 - 1000mg/dl	54	21	54	100
2022-455	072	URS-9-TECO		076 078	075 077		077 - 1000mg/dl	91	54	91	100
2022-455	072	URS-9-TECO		076 078	075 077		078 - 2000mg/dl	91	11	91	100
2022-455	072	URS-9-TECO		076 078	075 077		076 - 500mg/dl2+	91	26	91	100
2022-455	080	ROCHE (CHE		076 078	075 077		078 - 2000mg/dl	38	3	38	100
2022-455	080	ROCHE (CHE		076 078	075 077		077 - 1000mg/dl	38	35	38	100
2022-455	089	AIMSTRIP		075 077	075 077		077 - 1000mg/dl	119	43	119	100
2022-455	089	AIMSTRIP		075 077	075 077		075 - 250mg/dl-+	119	2	119	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-455	089	AIMSTRIP		075 077	075 077		076 - 500mg/dl2+	119	74	119	100
2022-455	094	CLINITEK		076 078	075 077		077 - 1000mg/dl	107	69	107	100
2022-455	094	CLINITEK		076 078	075 077		078 - 2000mg/dl	107	2	107	100
2022-455	094	CLINITEK		076 078	075 077		075 - 250mg/dl-+	107	2	107	100
2022-455	094	CLINITEK		076 078	075 077		076 - 500mg/dl2+	107	34	107	100
2022-455	102	ACON LAB		075 077	075 077		076 - 500mg/dl2+	145	93	145	100
2022-455	102	ACON LAB		075 077	075 077		077 - 1000mg/dl	145	48	145	100
2022-455	102	ACON LAB		075 077	075 077		075 - 250mg/dl-+	145	4	145	100
2022-455	131	iCHEM VELO		075 077	075 077		077 - 1000mg/dl	29	2	29	100
2022-455	131	iCHEM VELO		075 077	075 077		076 - 500mg/dl2+	29	27	29	100
2022-455	135	Hybrido AU		077 078	075 077		078 - 2000mg/dl	12	6	11	100
2022-455	135	Hybrido AU		077 078	075 077		073 - NEGATIVE	12	1	11	0
2022-455	135	Hybrido AU		077 078	075 077		077 - 1000mg/dl	12	5	11	100
2022-455	138	URS-10T		076 078	075 077		076 - 500mg/dl2+	10	4	10	100
2022-455	138	URS-10T		076 078	075 077		077 - 1000mg/dl	10	6	10	100
2022-455	140	MISSION A		076 078	075 077		078 - 2000mg/dl	51	4	51	100
2022-455	140	MISSION A		076 078	075 077		076 - 500mg/dl2+	51	19	51	100
2022-455	140	MISSION A		076 078	075 077		077 - 1000mg/dl	51	28	51	100
2022-455	145	CLINITEK		076 078	075 077		076 - 500mg/dl2+	16	1	16	100
2022-455	145	CLINITEK		076 078	075 077		077 - 1000mg/dl	16	11	16	100
2022-455	145	CLINITEK		076 078	075 077		075 - 250mg/dl-+	16	1	16	100
2022-455	145	CLINITEK		076 078	075 077		078 - 2000mg/dl	16	3	16	100
2022-455	147	Mission120		076 078	075 077		077 - 1000mg/dl	39	30	39	100
2022-455	147	Mission120		076 078	075 077		078 - 2000mg/dl	39	4	39	100
2022-455	147	Mission120		076 078	075 077		076 - 500mg/dl2+	39	5	39	100
2022-455	148	Mission500		075 077	075 077		076 - 500mg/dl2+	58	34	56	100
2022-455	148	Mission500		075 077	075 077		074 - TRACES	58	1	56	0
2022-455	148	Mission500		075 077	075 077		077 - 1000mg/dl	58	22	56	100
2022-455	148	Mission500		075 077	075 077		078 - 2000mg/dl	58	1	56	0
2022-455	151	Clari. Uro		076 078	075 077		076 - 500mg/dl2+	10	3	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-455	151	Clari. Uro		076 078	075 077		077 - 1000mg/dl	10	7	10	100		
059 URINALYS.HG.OR BLOOD													
2022-451	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	46	46	46	100		
2022-451	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	33	33	33	100		
2022-451	067	URISCAN		066 066	066 066		066 - NEGATIVE	53	53	53	100		
2022-451	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	91	91	91	100		
2022-451	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	38	38	38	100		
2022-451	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	119	119	119	100		
2022-451	094	CLINITEK		066 066	066 066		066 - NEGATIVE	107	107	107	100		
2022-451	102	ACON LAB		066 066	066 066		066 - NEGATIVE	146	146	146	100		
2022-451	131	ICHEM VELO		066 066	066 066		066 - NEGATIVE	29	29	29	100		
2022-451	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	12	12	12	100		
2022-451	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100		
2022-451	140	MISSION A		066 066	066 066		067 - TRACES	50	1	49	0		
2022-451	140	MISSION A		066 066	066 066		066 - NEGATIVE	50	49	49	100		
2022-451	145	CLINITEK		066 066	066 066		066 - NEGATIVE	16	16	16	100		
2022-451	147	Mission120		066 066	066 066		066 - NEGATIVE	40	40	40	100		
2022-451	148	Mission500		066 066	066 066		066 - NEGATIVE	57	57	57	100		
2022-451	151	Clari. Uro		066 066	066 066		066 - NEGATIVE	10	10	10	100		
2022-452	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	46	46	46	100		
2022-452	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	33	33	33	100		
2022-452	067	URISCAN		066 066	066 066		066 - NEGATIVE	53	53	53	100		
2022-452	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	91	91	91	100		
2022-452	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	38	38	38	100		
2022-452	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	119	118	118	100		
2022-452	089	AIMSTRIP		066 066	066 066		068 - POS. (1-2+)	119	1	118	0		
2022-452	094	CLINITEK		066 066	066 066		068 - POS. (1-2+)	107	1	106	0		
2022-452	094	CLINITEK		066 066	066 066		066 - NEGATIVE	107	106	106	100		

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	102	ACON LAB		066 066	066 066		069 - POS.(3+)	146	1	145	0
2022-452	102	ACON LAB		066 066	066 066		066 - NEGATIVE	146	145	145	100
2022-452	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	29	29	29	100
2022-452	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	12	12	12	100
2022-452	138	URS-10T		066 066	066 066		066 - NEGATIVE	10	10	10	100
2022-452	140	MISSION A		066 066	066 066		066 - NEGATIVE	50	50	50	100
2022-452	145	CLINITEK		066 066	066 066		066 - NEGATIVE	16	16	16	100
2022-452	147	Mission120		066 066	066 066		066 - NEGATIVE	40	40	40	100
2022-452	148	Mission500		066 066	066 066		066 - NEGATIVE	57	57	57	100
2022-452	151	Clari. Uro		066 066	066 066		066 - NEGATIVE	10	10	10	100
059 URINALYS.HG.OR BLOOD											
2022-453	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	46	3	46	100
2022-453	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	46	43	46	100
2022-453	013	BAYER CLI		068 069	068 069		069 - POS.(3+)	33	33	33	100
2022-453	067	URISCAN		068 069	068 069		069 - POS.(3+)	53	48	53	100
2022-453	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	53	5	53	100
2022-453	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	91	7	91	100
2022-453	072	URS-9-TECO		068 069	068 069		069 - POS.(3+)	91	84	91	100
2022-453	080	ROCHE (CHE		068 069	068 069		066 - NEGATIVE	38	1	37	0
2022-453	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+)	38	35	37	100
2022-453	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	38	2	37	100
2022-453	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	119	3	119	100
2022-453	089	AIMSTRIP		068 069	068 069		069 - POS.(3+)	119	116	119	100
2022-453	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	107	1	107	100
2022-453	094	CLINITEK		068 069	068 069		069 - POS.(3+)	107	106	107	100
2022-453	102	ACON LAB		068 069	068 069		069 - POS.(3+)	146	144	146	100
2022-453	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	146	2	146	100
2022-453	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	29	2	29	100
2022-453	131	iCHEM VELO		068 069	068 069		069 - POS.(3+)	29	27	29	100
2022-453	135	Hybrido AU		068 069	068 069		069 - POS.(3+)	12	12	12	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-453	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	10	10	100
2022-453	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	50	1	50	100
2022-453	140	MISSION A		068 069	068 069		069 - POS.(3+))	50	49	50	100
2022-453	145	CLINITEK		068 069	068 069		069 - POS.(3+))	16	16	16	100
2022-453	147	Mission120		068 069	068 069		069 - POS.(3+))	40	40	40	100
2022-453	148	Mission500		068 069	068 069		068 - POS.(1-2+)	57	1	57	100
2022-453	148	Mission500		068 069	068 069		069 - POS.(3+))	57	56	57	100
2022-453	151	Clari. Uro		068 069	068 069		069 - POS.(3+))	10	10	10	100
059 URINALYS.HG.OR BLOOD											
2022-454	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	46	11	46	100
2022-454	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	46	35	46	100
2022-454	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	33	3	33	100
2022-454	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	33	30	33	100
2022-454	067	URISCAN		068 069	068 069		069 - POS.(3+))	53	42	53	100
2022-454	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	53	11	53	100
2022-454	072	URS-9-TECO		068 069	068 069		067 - TRACES	91	1	90	0
2022-454	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	91	24	90	100
2022-454	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	91	66	90	100
2022-454	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	38	3	38	100
2022-454	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	38	35	38	100
2022-454	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	119	4	119	100
2022-454	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	119	115	119	100
2022-454	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	107	9	107	100
2022-454	094	CLINITEK		068 069	068 069		069 - POS.(3+))	107	98	107	100
2022-454	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	146	4	146	100
2022-454	102	ACON LAB		068 069	068 069		069 - POS.(3+))	146	142	146	100
2022-454	131	iCHEM VELO		067 069	068 069		068 - POS.(1-2+)	29	25	29	100
2022-454	131	iCHEM VELO		067 069	068 069		067 - TRACES	29	1	29	100
2022-454	131	iCHEM VELO		067 069	068 069		069 - POS.(3+))	29	3	29	100
2022-454	135	Hybrido AU		067 069	068 069		068 - POS.(1-2+)	12	12	12	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	7	10	100
2022-454	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	10	3	10	100
2022-454	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	50	1	50	100
2022-454	140	MISSION A		068 069	068 069		069 - POS.(3+))	50	49	50	100
2022-454	145	CLINITEK		068 069	068 069		069 - POS.(3+))	16	8	16	100
2022-454	145	CLINITEK		068 069	068 069		068 - POS.(1-2+)	16	8	16	100
2022-454	147	Mission120		068 069	068 069		069 - POS.(3+))	40	40	40	100
2022-454	148	Mission500		068 069	068 069		068 - POS.(1-2+)	57	1	57	100
2022-454	148	Mission500		068 069	068 069		069 - POS.(3+))	57	56	57	100
2022-454	151	Clari. Uro		068 069	068 069		069 - POS.(3+))	10	10	10	100
059 URINALYS.HG.OR BLOOD											
2022-455	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	46	45	46	100
2022-455	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	46	1	46	100
2022-455	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	33	33	33	100
2022-455	067	URISCAN		068 069	068 069		069 - POS.(3+))	53	44	53	100
2022-455	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	53	9	53	100
2022-455	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	91	81	91	100
2022-455	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	91	10	91	100
2022-455	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	38	35	37	100
2022-455	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	38	2	37	100
2022-455	080	ROCHE (CHE		068 069	068 069		066 - NEGATIVE	38	1	37	0
2022-455	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	119	112	119	100
2022-455	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	119	7	119	100
2022-455	094	CLINITEK		068 069	068 069		069 - POS.(3+))	107	104	106	100
2022-455	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	107	2	106	100
2022-455	094	CLINITEK		068 069	068 069		066 - NEGATIVE	107	1	106	0
2022-455	102	ACON LAB		068 069	068 069		069 - POS.(3+))	146	139	146	100
2022-455	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	146	7	146	100
2022-455	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	29	2	29	100
2022-455	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	29	27	29	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-455	135	Hybrido AU		068 069	068 069		069 - POS.(3+))	12	11	12	100		
2022-455	135	Hybrido AU		068 069	068 069		068 - POS.(1-2+)	12	1	12	100		
2022-455	138	URS-10T		068 069	068 069		069 - POS.(3+))	10	10	10	100		
2022-455	140	MISSION A		068 069	068 069		069 - POS.(3+))	50	48	50	100		
2022-455	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	50	2	50	100		
2022-455	145	CLINITEK		068 069	068 069		069 - POS.(3+))	16	16	16	100		
2022-455	147	Mission120		068 069	068 069		069 - POS.(3+))	40	40	40	100		
2022-455	148	Mission500		068 069	068 069		069 - POS.(3+))	57	56	57	100		
2022-455	148	Mission500		068 069	068 069		068 - POS.(1-2+)	57	1	57	100		
2022-455	151	Clari. Uro		068 069	068 069		069 - POS.(3+))	10	10	10	100		
060 URINALYSIS-KETONES													
2022-451	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	46	46	46	100		
2022-451	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	33	33	33	100		
2022-451	067	URISCAN		090 090	090 090		090 - NEGATIVE	54	54	54	100		
2022-451	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	91	91	91	100		
2022-451	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	38	38	38	100		
2022-451	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	119	119	119	100		
2022-451	094	CLINITEK		090 090	090 090		090 - NEGATIVE	107	107	107	100		
2022-451	102	ACON LAB		090 090	090 090		090 - NEGATIVE	145	145	145	100		
2022-451	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	29	29	29	100		
2022-451	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100		
2022-451	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100		
2022-451	140	MISSION A		090 090	090 090		090 - NEGATIVE	52	52	52	100		
2022-451	145	CLINITEK		090 090	090 090		090 - NEGATIVE	16	16	16	100		
2022-451	147	Mission120		090 090	090 090		090 - NEGATIVE	39	39	39	100		
2022-451	148	Mission500		090 090	090 090		090 - NEGATIVE	57	57	57	100		
2022-451	151	Clari. Uro		090 090	090 090		090 - NEGATIVE	10	10	10	100		
2022-452	011	AMES NM10S		092 094	092 094		092 - SMALL (1+)	46	9	46	100		

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	011	AMES NM10S		092 094	092 094		093 - MOD. (2+)	46	33	46	100
2022-452	011	AMES NM10S		092 094	092 094		094 - LARGE-3-4+	46	4	46	100
2022-452	013	BAYER CLI		092 094	092 094		093 - MOD. (2+)	33	22	33	100
2022-452	013	BAYER CLI		092 094	092 094		092 - SMALL (1+)	33	11	33	100
2022-452	067	URISCAN		092 094	092 094		092 - SMALL (1+)	54	15	54	100
2022-452	067	URISCAN		092 094	092 094		093 - MOD. (2+)	54	32	54	100
2022-452	067	URISCAN		092 094	092 094		094 - LARGE-3-4+	54	7	54	100
2022-452	072	URS-9-TECO		092 094	092 094		092 - SMALL (1+)	91	5	91	100
2022-452	072	URS-9-TECO		092 094	092 094		093 - MOD. (2+)	91	64	91	100
2022-452	072	URS-9-TECO		092 094	092 094		094 - LARGE-3-4+	91	22	91	100
2022-452	080	ROCHE (CHE		092 094	092 094		092 - SMALL (1+)	38	14	37	100
2022-452	080	ROCHE (CHE		092 094	092 094		093 - MOD. (2+)	38	18	37	100
2022-452	080	ROCHE (CHE		092 094	092 094		091 - TRACE	38	1	37	0
2022-452	080	ROCHE (CHE		092 094	092 094		094 - LARGE-3-4+	38	5	37	100
2022-452	089	AIMSTRIP		092 094	092 094		091 - TRACE	119	1	118	0
2022-452	089	AIMSTRIP		092 094	092 094		092 - SMALL (1+)	119	28	118	100
2022-452	089	AIMSTRIP		092 094	092 094		093 - MOD. (2+)	119	77	118	100
2022-452	089	AIMSTRIP		092 094	092 094		094 - LARGE-3-4+	119	13	118	100
2022-452	094	CLINITEK		092 094	092 094		090 - NEGATIVE	107	1	105	0
2022-452	094	CLINITEK		092 094	092 094		091 - TRACE	107	1	105	0
2022-452	094	CLINITEK		092 094	092 094		094 - LARGE-3-4+	107	2	105	100
2022-452	094	CLINITEK		092 094	092 094		093 - MOD. (2+)	107	73	105	100
2022-452	094	CLINITEK		092 094	092 094		092 - SMALL (1+)	107	30	105	100
2022-452	102	ACON LAB		092 094	092 094		090 - NEGATIVE	145	2	141	0
2022-452	102	ACON LAB		092 094	092 094		093 - MOD. (2+)	145	93	141	100
2022-452	102	ACON LAB		092 094	092 094		092 - SMALL (1+)	145	24	141	100
2022-452	102	ACON LAB		092 094	092 094		094 - LARGE-3-4+	145	24	141	100
2022-452	102	ACON LAB		092 094	092 094		091 - TRACE	145	2	141	0
2022-452	131	iCHEM VELO		091 093	092 094		093 - MOD. (2+)	29	9	29	100
2022-452	131	iCHEM VELO		091 093	092 094		092 - SMALL (1+)	29	20	29	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	135	Hybrido AU		092 094	092 094		093 - MOD. (2+)	12	11	12	100
2022-452	135	Hybrido AU		092 094	092 094		092 - SMALL (1+)	12	1	12	100
2022-452	138	URS-10T		092 094	092 094		094 - LARGE-3-4+	10	1	10	100
2022-452	138	URS-10T		092 094	092 094		092 - SMALL (1+)	10	3	10	100
2022-452	138	URS-10T		092 094	092 094		093 - MOD. (2+)	10	6	10	100
2022-452	140	MISSION A		093 094	092 094		094 - LARGE-3-4+	52	27	52	100
2022-452	140	MISSION A		093 094	092 094		093 - MOD. (2+)	52	22	52	100
2022-452	140	MISSION A		093 094	092 094		092 - SMALL (1+)	52	3	52	100
2022-452	145	CLINITEK		092 094	092 094		092 - SMALL (1+)	16	2	16	100
2022-452	145	CLINITEK		092 094	092 094		093 - MOD. (2+)	16	14	16	100
2022-452	147	Mission120		093 094	092 094		094 - LARGE-3-4+	39	25	39	100
2022-452	147	Mission120		093 094	092 094		093 - MOD. (2+)	39	13	39	100
2022-452	147	Mission120		093 094	092 094		092 - SMALL (1+)	39	1	39	100
2022-452	148	Mission500		092 094	092 094		093 - MOD. (2+)	57	49	57	100
2022-452	148	Mission500		092 094	092 094		092 - SMALL (1+)	57	1	57	100
2022-452	148	Mission500		092 094	092 094		094 - LARGE-3-4+	57	7	57	100
2022-452	151	Clari. Uro		092 094	092 094		093 - MOD. (2+)	10	7	10	100
2022-452	151	Clari. Uro		092 094	092 094		092 - SMALL (1+)	10	1	10	100
2022-452	151	Clari. Uro		092 094	092 094		094 - LARGE-3-4+	10	2	10	100
060 URINALYSIS-KETONES											
2022-453	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	46	46	46	100
2022-453	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	33	33	33	100
2022-453	067	URISCAN		090 090	090 090		090 - NEGATIVE	54	54	54	100
2022-453	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	91	91	91	100
2022-453	080	ROCHE (CHE		090 090	090 090		094 - LARGE-3-4+	38	2	36	0
2022-453	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	38	36	36	100
2022-453	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	119	119	119	100
2022-453	094	CLINITEK		090 090	090 090		090 - NEGATIVE	107	107	107	100
2022-453	102	ACON LAB		090 090	090 090		090 - NEGATIVE	145	144	144	100
2022-453	102	ACON LAB		090 090	090 090		092 - SMALL (1+)	145	1	144	0

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-453	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	29	29	29	100
2022-453	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100
2022-453	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100
2022-453	140	MISSION A		090 090	090 090		090 - NEGATIVE	52	52	52	100
2022-453	145	CLINITEK		090 090	090 090		090 - NEGATIVE	16	16	16	100
2022-453	147	Mission120		090 090	090 090		090 - NEGATIVE	39	39	39	100
2022-453	148	Mission500		090 090	090 090		090 - NEGATIVE	57	57	57	100
2022-453	151	Clari. Uro		090 090	090 090		090 - NEGATIVE	10	10	10	100
060 URINALYSIS-KETONES											
2022-454	011	AMES NM10S		093 094	093 094		094 - LARGE-3-4+	46	44	46	100
2022-454	011	AMES NM10S		093 094	093 094		093 - MOD. (2+)	46	2	46	100
2022-454	013	BAYER CLI		093 094	093 094		093 - MOD. (2+)	33	2	33	100
2022-454	013	BAYER CLI		093 094	093 094		094 - LARGE-3-4+	33	31	33	100
2022-454	067	URISCAN		093 094	093 094		093 - MOD. (2+)	54	18	54	100
2022-454	067	URISCAN		093 094	093 094		094 - LARGE-3-4+	54	36	54	100
2022-454	072	URS-9-TECO		093 094	093 094		093 - MOD. (2+)	91	3	91	100
2022-454	072	URS-9-TECO		093 094	093 094		094 - LARGE-3-4+	91	88	91	100
2022-454	080	ROCHE (CHE		093 094	093 094		093 - MOD. (2+)	38	7	38	100
2022-454	080	ROCHE (CHE		093 094	093 094		094 - LARGE-3-4+	38	31	38	100
2022-454	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	119	5	119	100
2022-454	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	119	114	119	100
2022-454	094	CLINITEK		093 094	093 094		093 - MOD. (2+)	107	4	107	100
2022-454	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	107	103	107	100
2022-454	102	ACON LAB		093 094	093 094		090 - NEGATIVE	145	1	144	0
2022-454	102	ACON LAB		093 094	093 094		093 - MOD. (2+)	145	6	144	100
2022-454	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	145	138	144	100
2022-454	131	iCHEM VELO		092 094	093 094		093 - MOD. (2+)	29	20	29	100
2022-454	131	iCHEM VELO		092 094	093 094		094 - LARGE-3-4+	29	9	29	100
2022-454	135	Hybrido AU		093 094	093 094		094 - LARGE-3-4+	12	12	12	100
2022-454	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	10	10	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-454	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	52	52	52	100		
2022-454	145	CLINITEK		093 094	093 094		094 - LARGE-3-4+	16	16	16	100		
2022-454	147	Mission120		093 094	093 094		094 - LARGE-3-4+	39	39	39	100		
2022-454	148	Mission500		093 094	093 094		093 - MOD. (2+)	57	1	57	100		
2022-454	148	Mission500		093 094	093 094		094 - LARGE-3-4+	57	56	57	100		
2022-454	151	Clari. Uro		093 094	093 094		094 - LARGE-3-4+	10	10	10	100		
060 URINALYSIS-KETONES													
2022-455	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	46	46	46	100		
2022-455	013	BAYER CLI		090 090	090 090		094 - LARGE-3-4+	33	1	32	0		
2022-455	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	33	32	32	100		
2022-455	067	URISCAN		090 090	090 090		090 - NEGATIVE	54	54	54	100		
2022-455	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	91	91	91	100		
2022-455	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	38	38	38	100		
2022-455	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	119	119	119	100		
2022-455	094	CLINITEK		090 090	090 090		090 - NEGATIVE	107	107	107	100		
2022-455	102	ACON LAB		090 090	090 090		090 - NEGATIVE	145	142	142	100		
2022-455	102	ACON LAB		090 090	090 090		094 - LARGE-3-4+	145	3	142	0		
2022-455	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	29	29	29	100		
2022-455	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	12	12	12	100		
2022-455	138	URS-10T		090 090	090 090		090 - NEGATIVE	10	10	10	100		
2022-455	140	MISSION A		090 090	090 090		090 - NEGATIVE	52	52	52	100		
2022-455	145	CLINITEK		090 090	090 090		090 - NEGATIVE	16	16	16	100		
2022-455	147	Mission120		090 090	090 090		094 - LARGE-3-4+	39	1	38	0		
2022-455	147	Mission120		090 090	090 090		090 - NEGATIVE	39	38	38	100		
2022-455	148	Mission500		090 090	090 090		090 - NEGATIVE	57	57	57	100		
2022-455	151	Clari. Uro		090 090	090 090		090 - NEGATIVE	10	10	10	100		
061 URINALYSIS-NITRATE													
2022-451	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	46	46	46	100		
2022-451	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	34	34	34	100		

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-451	067	URISCAN		070 070	070 070		070 - NEGATIVE	54	54	54	100	
2022-451	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	91	91	91	100	
2022-451	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	38	38	38	100	
2022-451	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	119	119	119	100	
2022-451	094	CLINITEK		070 070	070 070		070 - NEGATIVE	107	107	107	100	
2022-451	102	ACON LAB		070 070	070 070		070 - NEGATIVE	145	145	145	100	
2022-451	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	29	29	29	100	
2022-451	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2022-451	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2022-451	140	MISSION A		070 070	070 070		070 - NEGATIVE	51	51	51	100	
2022-451	145	CLINITEK		070 070	070 070		070 - NEGATIVE	16	16	16	100	
2022-451	147	Mission120		070 070	070 070		070 - NEGATIVE	40	40	40	100	
2022-451	148	Mission500		070 070	070 070		070 - NEGATIVE	57	57	57	100	
2022-451	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	10	10	10	100	
061 URINALYSIS-NITRATE												
2022-452	011	AMES NM10S		071 072	071 072		072 - POSITIVE	46	46	46	100	
2022-452	013	BAYER CLI		071 072	071 072		072 - POSITIVE	34	34	34	100	
2022-452	067	URISCAN		071 072	071 072		072 - POSITIVE	54	54	54	100	
2022-452	072	URS-9-TECO		071 072	071 072		072 - POSITIVE	91	91	91	100	
2022-452	080	ROCHE (CHE		071 072	071 072		072 - POSITIVE	38	38	38	100	
2022-452	089	AIMSTRIP		071 072	071 072		072 - POSITIVE	119	118	119	100	
2022-452	089	AIMSTRIP		071 072	071 072		071 - WEAK POS.	119	1	119	100	
2022-452	094	CLINITEK		071 072	071 072		072 - POSITIVE	107	106	107	100	
2022-452	094	CLINITEK		071 072	071 072		071 - WEAK POS.	107	1	107	100	
2022-452	102	ACON LAB		071 072	071 072		072 - POSITIVE	145	143	145	100	
2022-452	102	ACON LAB		071 072	071 072		071 - WEAK POS.	145	2	145	100	
2022-452	131	iCHEM VELO		071 072	071 072		072 - POSITIVE	29	29	29	100	
2022-452	135	Hybrido AU		071 072	071 072		072 - POSITIVE	12	12	12	100	
2022-452	138	URS-10T		071 072	071 072		072 - POSITIVE	10	10	10	100	
2022-452	140	MISSION A		071 072	071 072		072 - POSITIVE	51	51	51	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-452	145	CLINITEK		071 072	071 072		072 - POSITIVE	16	16	16	100	
2022-452	147	Mission120		071 072	071 072		072 - POSITIVE	40	40	40	100	
2022-452	148	Mission500		071 072	071 072		072 - POSITIVE	57	57	57	100	
2022-452	151	Clari. Uro		071 072	071 072		072 - POSITIVE	10	10	10	100	
061 URINALYSIS-NITRATE												
2022-453	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	46	46	46	100	
2022-453	013	BAYER CLI		070 070	070 070		071 - WEAK POS.	34	1	33	0	
2022-453	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	34	33	33	100	
2022-453	067	URISCAN		070 070	070 070		070 - NEGATIVE	54	54	54	100	
2022-453	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	91	91	91	100	
2022-453	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	38	38	38	100	
2022-453	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	119	119	119	100	
2022-453	094	CLINITEK		070 070	070 070		070 - NEGATIVE	107	107	107	100	
2022-453	102	ACON LAB		070 070	070 070		070 - NEGATIVE	145	145	145	100	
2022-453	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	29	29	29	100	
2022-453	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100	
2022-453	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2022-453	140	MISSION A		070 070	070 070		070 - NEGATIVE	51	51	51	100	
2022-453	145	CLINITEK		070 070	070 070		070 - NEGATIVE	16	16	16	100	
2022-453	147	Mission120		070 070	070 070		070 - NEGATIVE	40	40	40	100	
2022-453	148	Mission500		070 070	070 070		070 - NEGATIVE	57	57	57	100	
2022-453	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2022-454	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	46	46	46	100	
2022-454	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	34	34	34	100	
2022-454	067	URISCAN		070 070	070 070		070 - NEGATIVE	54	54	54	100	
2022-454	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	91	90	90	100	
2022-454	072	URS-9-TECO		070 070	070 070		072 - POSITIVE	91	1	90	0	
2022-454	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	38	38	38	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	119	119	119	100
2022-454	094	CLINITEK		070 070	070 070		070 - NEGATIVE	107	107	107	100
2022-454	102	ACON LAB		070 070	070 070		070 - NEGATIVE	145	144	144	100
2022-454	102	ACON LAB		070 070	070 070		072 - POSITIVE	145	1	144	0
2022-454	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	29	29	29	100
2022-454	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100
2022-454	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100
2022-454	140	MISSION A		070 070	070 070		070 - NEGATIVE	51	51	51	100
2022-454	145	CLINITEK		070 070	070 070		070 - NEGATIVE	16	16	16	100
2022-454	147	Mission120		070 070	070 070		070 - NEGATIVE	40	40	40	100
2022-454	148	Mission500		070 070	070 070		070 - NEGATIVE	57	57	57	100
2022-454	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	10	10	10	100
061 URINALYSIS-NITRATE											
2022-455	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	46	46	46	100
2022-455	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	34	34	34	100
2022-455	067	URISCAN		070 070	070 070		070 - NEGATIVE	54	54	54	100
2022-455	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	91	91	91	100
2022-455	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	38	38	38	100
2022-455	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	119	119	119	100
2022-455	094	CLINITEK		070 070	070 070		070 - NEGATIVE	107	107	107	100
2022-455	102	ACON LAB		070 070	070 070		070 - NEGATIVE	145	145	145	100
2022-455	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	29	29	29	100
2022-455	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	12	12	12	100
2022-455	138	URS-10T		070 070	070 070		070 - NEGATIVE	10	10	10	100
2022-455	140	MISSION A		070 070	070 070		070 - NEGATIVE	51	51	51	100
2022-455	145	CLINITEK		070 070	070 070		070 - NEGATIVE	16	16	16	100
2022-455	147	Mission120		070 070	070 070		070 - NEGATIVE	40	40	40	100
2022-455	148	Mission500		070 070	070 070		070 - NEGATIVE	57	57	57	100
2022-455	151	Clari. Uro		070 070	070 070		070 - NEGATIVE	10	10	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
062 URINALYSIS-pH											
2022-451	011	AMES NM10S		086 088	086 088		088 - 8.00R MORE	46	35	46	100
2022-451	011	AMES NM10S		086 088	086 088		087 - 7.5	46	11	46	100
2022-451	013	BAYER CLI		086 088	086 088		087 - 7.5	34	12	34	100
2022-451	013	BAYER CLI		086 088	086 088		088 - 8.00R MORE	34	22	34	100
2022-451	067	URISCAN		085 088	086 088		087 - 7.5	54	43	54	100
2022-451	067	URISCAN		085 088	086 088		086 - 7.0	54	1	54	100
2022-451	067	URISCAN		085 088	086 088		088 - 8.00R MORE	54	10	54	100
2022-451	072	URS-9-TECO		085 088	086 088		087 - 7.5	90	60	90	100
2022-451	072	URS-9-TECO		085 088	086 088		085 - 6.5	90	2	90	100
2022-451	072	URS-9-TECO		085 088	086 088		086 - 7.0	90	16	90	100
2022-451	072	URS-9-TECO		085 088	086 088		088 - 8.00R MORE	90	12	90	100
2022-451	080	ROCHE (CHE		084 088	086 088		088 - 8.00R MORE	38	8	38	100
2022-451	080	ROCHE (CHE		084 088	086 088		086 - 7.0	38	30	38	100
2022-451	089	AIMSTRIP		086 088	086 088		087 - 7.5	119	58	119	100
2022-451	089	AIMSTRIP		086 088	086 088		086 - 7.0	119	1	119	100
2022-451	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	119	60	119	100
2022-451	094	CLINITEK		086 088	086 088		086 - 7.0	107	1	107	100
2022-451	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	107	76	107	100
2022-451	094	CLINITEK		086 088	086 088		087 - 7.5	107	30	107	100
2022-451	102	ACON LAB		086 088	086 088		086 - 7.0	145	1	145	100
2022-451	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	145	73	145	100
2022-451	102	ACON LAB		086 088	086 088		087 - 7.5	145	71	145	100
2022-451	131	iCHEM VELO		086 088	086 088		088 - 8.00R MORE	29	24	29	100
2022-451	131	iCHEM VELO		086 088	086 088		086 - 7.0	29	5	29	100
2022-451	135	Hybrido AU		086 088	086 088		082 - 5.0	12	1	11	0
2022-451	135	Hybrido AU		086 088	086 088		087 - 7.5	12	1	11	100
2022-451	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	12	10	11	100
2022-451	138	URS-10T		085 088	086 088		085 - 6.5	10	1	10	100
2022-451	138	URS-10T		085 088	086 088		087 - 7.5	10	9	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-451	140	MISSION A		085 088	086 088		088 - 8.00R MORE	52	20	52	100
2022-451	140	MISSION A		085 088	086 088		087 - 7.5	52	32	52	100
2022-451	145	CLINITEK		086 088	086 088		087 - 7.5	16	2	16	100
2022-451	145	CLINITEK		086 088	086 088		088 - 8.00R MORE	16	14	16	100
2022-451	147	Mission120		085 088	086 088		087 - 7.5	39	27	39	100
2022-451	147	Mission120		085 088	086 088		088 - 8.00R MORE	39	12	39	100
2022-451	148	Mission500		086 088	086 088		087 - 7.5	57	18	57	100
2022-451	148	Mission500		086 088	086 088		088 - 8.00R MORE	57	39	57	100
2022-451	151	Clari. Uro		085 088	086 088		088 - 8.00R MORE	10	3	10	100
2022-451	151	Clari. Uro		085 088	086 088		087 - 7.5	10	7	10	100
062 URINALYSIS-pH											
2022-452	011	AMES NM10S		080 084	080 084		082 - 5.0	46	22	46	100
2022-452	011	AMES NM10S		080 084	080 084		083 - 5.5	46	16	46	100
2022-452	011	AMES NM10S		080 084	080 084		084 - 6.0	46	8	46	100
2022-452	013	BAYER CLI		081 085	080 084		082 - 5.0	34	16	34	100
2022-452	013	BAYER CLI		081 085	080 084		084 - 6.0	34	1	34	100
2022-452	013	BAYER CLI		081 085	080 084		083 - 5.5	34	17	34	100
2022-452	067	URISCAN		080 084	080 084		083 - 5.5	54	10	54	100
2022-452	067	URISCAN		080 084	080 084		084 - 6.0	54	1	54	100
2022-452	067	URISCAN		080 084	080 084		082 - 5.0	54	43	54	100
2022-452	072	URS-9-TECO		082 086	080 084		083 - 5.5	90	13	90	100
2022-452	072	URS-9-TECO		082 086	080 084		084 - 6.0	90	58	90	100
2022-452	072	URS-9-TECO		082 086	080 084		082 - 5.0	90	16	90	100
2022-452	072	URS-9-TECO		082 086	080 084		085 - 6.5	90	3	90	100
2022-452	080	ROCHE (CHE		080 084	080 084		082 - 5.0	38	31	38	100
2022-452	080	ROCHE (CHE		080 084	080 084		084 - 6.0	38	7	38	100
2022-452	089	AIMSTRIP		080 084	080 084		082 - 5.0	119	59	119	100
2022-452	089	AIMSTRIP		080 084	080 084		084 - 6.0	119	24	119	100
2022-452	089	AIMSTRIP		080 084	080 084		083 - 5.5	119	36	119	100
2022-452	094	CLINITEK		081 085	080 084		083 - 5.5	107	66	107	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-452	094	CLINITEK		081 085	080 084		082 - 5.0	107	29	107	100
2022-452	094	CLINITEK		081 085	080 084		084 - 6.0	107	12	107	100
2022-452	102	ACON LAB		082 086	080 084		084 - 6.0	145	69	145	100
2022-452	102	ACON LAB		082 086	080 084		082 - 5.0	145	43	145	100
2022-452	102	ACON LAB		082 086	080 084		083 - 5.5	145	33	145	100
2022-452	131	ICHEM VELO		080 084	080 084		082 - 5.0	29	29	29	100
2022-452	135	Hybrido AU		080 084	080 084		083 - 5.5	12	1	12	100
2022-452	135	Hybrido AU		080 084	080 084		082 - 5.0	12	11	12	100
2022-452	138	URS-10T		080 084	080 084		082 - 5.0	10	5	10	100
2022-452	138	URS-10T		080 084	080 084		083 - 5.5	10	3	10	100
2022-452	138	URS-10T		080 084	080 084		084 - 6.0	10	2	10	100
2022-452	140	MISSION A		081 085	080 084		082 - 5.0	52	19	52	100
2022-452	140	MISSION A		081 085	080 084		083 - 5.5	52	20	52	100
2022-452	140	MISSION A		081 085	080 084		084 - 6.0	52	13	52	100
2022-452	145	CLINITEK		080 084	080 084		082 - 5.0	16	15	16	100
2022-452	145	CLINITEK		080 084	080 084		083 - 5.5	16	1	16	100
2022-452	147	Mission120		081 085	080 084		083 - 5.5	39	26	39	100
2022-452	147	Mission120		081 085	080 084		082 - 5.0	39	13	39	100
2022-452	148	Mission500		082 086	080 084		082 - 5.0	57	1	57	100
2022-452	148	Mission500		082 086	080 084		083 - 5.5	57	8	57	100
2022-452	148	Mission500		082 086	080 084		084 - 6.0	57	48	57	100
2022-452	151	Clari. Uro		081 085	080 084		083 - 5.5	10	4	10	100
2022-452	151	Clari. Uro		081 085	080 084		084 - 6.0	10	3	10	100
2022-452	151	Clari. Uro		081 085	080 084		082 - 5.0	10	3	10	100
062 URINALYSIS-pH											
2022-453	011	AMES NM10S		084 088	084 088		087 - 7.5	46	16	46	100
2022-453	011	AMES NM10S		084 088	084 088		086 - 7.0	46	30	46	100
2022-453	013	BAYER CLI		084 088	084 088		087 - 7.5	34	10	34	100
2022-453	013	BAYER CLI		084 088	084 088		086 - 7.0	34	24	34	100
2022-453	067	URISCAN		084 088	084 088		085 - 6.5	54	22	54	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-453	067	URISCAN		084 088	084 088		086 - 7.0	54	30	54	100
2022-453	067	URISCAN		084 088	084 088		087 - 7.5	54	2	54	100
2022-453	072	URS-9-TECO		084 088	084 088		085 - 6.5	90	3	90	100
2022-453	072	URS-9-TECO		084 088	084 088		087 - 7.5	90	5	90	100
2022-453	072	URS-9-TECO		084 088	084 088		086 - 7.0	90	82	90	100
2022-453	080	ROCHE (CHE		084 088	084 088		086 - 7.0	38	37	38	100
2022-453	080	ROCHE (CHE		084 088	084 088		085 - 6.5	38	1	38	100
2022-453	089	AIMSTRIP		084 088	084 088		086 - 7.0	119	97	119	100
2022-453	089	AIMSTRIP		084 088	084 088		085 - 6.5	119	1	119	100
2022-453	089	AIMSTRIP		084 088	084 088		087 - 7.5	119	21	119	100
2022-453	094	CLINITEK		084 088	084 088		087 - 7.5	107	23	107	100
2022-453	094	CLINITEK		084 088	084 088		086 - 7.0	107	84	107	100
2022-453	102	ACON LAB		084 088	084 088		086 - 7.0	145	122	145	100
2022-453	102	ACON LAB		084 088	084 088		085 - 6.5	145	6	145	100
2022-453	102	ACON LAB		084 088	084 088		087 - 7.5	145	17	145	100
2022-453	131	icHEM VELO		084 088	084 088		086 - 7.0	29	29	29	100
2022-453	135	Hybrido AU		084 088	084 088		086 - 7.0	12	10	12	100
2022-453	135	Hybrido AU		084 088	084 088		087 - 7.5	12	2	12	100
2022-453	138	URS-10T		084 088	084 088		086 - 7.0	10	10	10	100
2022-453	140	MISSION A		084 088	084 088		087 - 7.5	52	2	52	100
2022-453	140	MISSION A		084 088	084 088		086 - 7.0	52	45	52	100
2022-453	140	MISSION A		084 088	084 088		085 - 6.5	52	5	52	100
2022-453	145	CLINITEK		084 088	084 088		087 - 7.5	16	1	16	100
2022-453	145	CLINITEK		084 088	084 088		086 - 7.0	16	15	16	100
2022-453	147	Mission120		084 088	084 088		085 - 6.5	39	7	39	100
2022-453	147	Mission120		084 088	084 088		086 - 7.0	39	31	39	100
2022-453	147	Mission120		084 088	084 088		084 - 6.0	39	1	39	100
2022-453	148	Mission500		084 088	084 088		086 - 7.0	57	57	57	100
2022-453	151	Clari. Uro		084 088	084 088		086 - 7.0	10	10	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
062 URINALYSIS-pH											
2022-454	011	AMES NM10S		086 088	086 088		087 - 7.5	46	6	46	100
2022-454	011	AMES NM10S		086 088	086 088		088 - 8.00R MORE	46	40	46	100
2022-454	013	BAYER CLI		086 088	086 088		088 - 8.00R MORE	34	31	34	100
2022-454	013	BAYER CLI		086 088	086 088		087 - 7.5	34	3	34	100
2022-454	067	URISCAN		085 088	086 088		087 - 7.5	54	31	54	100
2022-454	067	URISCAN		085 088	086 088		088 - 8.00R MORE	54	23	54	100
2022-454	072	URS-9-TECO		085 088	086 088		087 - 7.5	90	62	90	100
2022-454	072	URS-9-TECO		085 088	086 088		086 - 7.0	90	6	90	100
2022-454	072	URS-9-TECO		085 088	086 088		088 - 8.00R MORE	90	22	90	100
2022-454	080	ROCHE (CHE		086 088	086 088		086 - 7.0	38	19	38	100
2022-454	080	ROCHE (CHE		086 088	086 088		088 - 8.00R MORE	38	19	38	100
2022-454	089	AIMSTRIP		086 088	086 088		086 - 7.0	119	1	119	100
2022-454	089	AIMSTRIP		086 088	086 088		088 - 8.00R MORE	119	107	119	100
2022-454	089	AIMSTRIP		086 088	086 088		087 - 7.5	119	11	119	100
2022-454	094	CLINITEK		086 088	086 088		088 - 8.00R MORE	107	97	107	100
2022-454	094	CLINITEK		086 088	086 088		087 - 7.5	107	10	107	100
2022-454	102	ACON LAB		086 088	086 088		087 - 7.5	145	12	145	100
2022-454	102	ACON LAB		086 088	086 088		088 - 8.00R MORE	145	133	145	100
2022-454	131	iCHEM VELO		086 088	086 088		086 - 7.0	29	1	29	100
2022-454	131	iCHEM VELO		086 088	086 088		088 - 8.00R MORE	29	28	29	100
2022-454	135	Hybrido AU		086 088	086 088		088 - 8.00R MORE	12	12	12	100
2022-454	138	URS-10T		085 088	086 088		087 - 7.5	10	6	10	100
2022-454	138	URS-10T		085 088	086 088		088 - 8.00R MORE	10	3	10	100
2022-454	138	URS-10T		085 088	086 088		086 - 7.0	10	1	10	100
2022-454	140	MISSION A		086 088	086 088		088 - 8.00R MORE	52	46	52	100
2022-454	140	MISSION A		086 088	086 088		087 - 7.5	52	6	52	100
2022-454	145	CLINITEK		086 088	086 088		087 - 7.5	16	2	16	100
2022-454	145	CLINITEK		086 088	086 088		088 - 8.00R MORE	16	14	16	100
2022-454	147	Mission120		086 088	086 088		088 - 8.00R MORE	39	38	39	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-454	147	Mission120		086 088	086 088		087 - 7.5	39	1	39	100	
2022-454	148	Mission500		086 088	086 088		087 - 7.5	57	4	57	100	
2022-454	148	Mission500		086 088	086 088		088 - 8.0OR MORE	57	53	57	100	
2022-454	151	Clari. Uro		086 088	086 088		087 - 7.5	10	2	10	100	
2022-454	151	Clari. Uro		086 088	086 088		088 - 8.0OR MORE	10	8	10	100	
062 URINALYSIS-pH												
2022-455	011	AMES NM10S		084 088	083 087		086 - 7.0	46	24	46	100	
2022-455	011	AMES NM10S		084 088	083 087		085 - 6.5	46	22	46	100	
2022-455	013	BAYER CLI		084 088	083 087		085 - 6.5	34	16	34	100	
2022-455	013	BAYER CLI		084 088	083 087		086 - 7.0	34	18	34	100	
2022-455	067	URISCAN		083 087	083 087		086 - 7.0	54	1	54	100	
2022-455	067	URISCAN		083 087	083 087		085 - 6.5	54	30	54	100	
2022-455	067	URISCAN		083 087	083 087		084 - 6.0	54	23	54	100	
2022-455	072	URS-9-TECO		083 087	083 087		085 - 6.5	90	66	89	100	
2022-455	072	URS-9-TECO		083 087	083 087		082 - 5.0	90	1	89	0	
2022-455	072	URS-9-TECO		083 087	083 087		084 - 6.0	90	12	89	100	
2022-455	072	URS-9-TECO		083 087	083 087		086 - 7.0	90	11	89	100	
2022-455	080	ROCHE (CHE		083 087	083 087		085 - 6.5	38	32	38	100	
2022-455	080	ROCHE (CHE		083 087	083 087		084 - 6.0	38	6	38	100	
2022-455	089	AIMSTRIP		083 087	083 087		086 - 7.0	119	34	119	100	
2022-455	089	AIMSTRIP		083 087	083 087		085 - 6.5	119	61	119	100	
2022-455	089	AIMSTRIP		083 087	083 087		084 - 6.0	119	24	119	100	
2022-455	094	CLINITEK		084 088	083 087		084 - 6.0	107	3	107	100	
2022-455	094	CLINITEK		084 088	083 087		086 - 7.0	107	60	107	100	
2022-455	094	CLINITEK		084 088	083 087		085 - 6.5	107	44	107	100	
2022-455	102	ACON LAB		083 087	083 087		085 - 6.5	145	98	145	100	
2022-455	102	ACON LAB		083 087	083 087		086 - 7.0	145	22	145	100	
2022-455	102	ACON LAB		083 087	083 087		084 - 6.0	145	25	145	100	
2022-455	131	iCHEM VELO		082 086	083 087		084 - 6.0	29	26	29	100	
2022-455	131	iCHEM VELO		082 086	083 087		086 - 7.0	29	3	29	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-455	135	Hybrido AU		083 087	083 087		086 - 7.0	12	3	12	100	
2022-455	135	Hybrido AU		083 087	083 087		085 - 6.5	12	9	12	100	
2022-455	138	URS-10T		083 087	083 087		084 - 6.0	10	1	10	100	
2022-455	138	URS-10T		083 087	083 087		085 - 6.5	10	7	10	100	
2022-455	138	URS-10T		083 087	083 087		086 - 7.0	10	2	10	100	
2022-455	140	MISSION A		083 087	083 087		083 - 5.5	52	2	52	100	
2022-455	140	MISSION A		083 087	083 087		085 - 6.5	52	38	52	100	
2022-455	140	MISSION A		083 087	083 087		086 - 7.0	52	4	52	100	
2022-455	140	MISSION A		083 087	083 087		084 - 6.0	52	8	52	100	
2022-455	145	CLINITEK		083 087	083 087		085 - 6.5	16	16	16	100	
2022-455	147	Mission120		083 087	083 087		084 - 6.0	39	2	39	100	
2022-455	147	Mission120		083 087	083 087		085 - 6.5	39	37	39	100	
2022-455	148	Mission500		083 087	083 087		085 - 6.5	57	57	57	100	
2022-455	151	Clari. Uro		082 086	083 087		085 - 6.5	10	4	10	100	
2022-455	151	Clari. Uro		082 086	083 087		084 - 6.0	10	6	10	100	
064 URIN-PROTEIN QUALI.												
2022-451	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	47	45	45	100	
2022-451	011	AMES NM10S		051 051	051 051		052 - TRACES	47	2	45	0	
2022-451	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	33	33	33	100	
2022-451	067	URISCAN		051 051	051 051		052 - TRACES	54	1	53	0	
2022-451	067	URISCAN		051 051	051 051		051 - NEGATIVE	54	53	53	100	
2022-451	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	91	91	91	100	
2022-451	080	ROCHE (CHE		051 051	051 051		052 - TRACES	38	2	36	0	
2022-451	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	38	36	36	100	
2022-451	089	AIMSTRIP		051 051	051 051		052 - TRACES	119	5	114	0	
2022-451	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	119	114	114	100	
2022-451	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	106	106	100	
2022-451	102	ACON LAB		051 051	051 051		052 - TRACES	144	1	143	0	
2022-451	102	ACON LAB		051 051	051 051		051 - NEGATIVE	144	143	143	100	
2022-451	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	29	29	29	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2022-451	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100		
2022-451	140	MISSION A		051 051	051 051		052 - TRACES	51	1	50	0		
2022-451	140	MISSION A		051 051	051 051		051 - NEGATIVE	51	50	50	100		
2022-451	145	CLINITEK		051 051	051 051		051 - NEGATIVE	16	16	16	100		
2022-451	147	Mission120		051 051	051 051		051 - NEGATIVE	40	40	40	100		
2022-451	148	Mission500		051 051	051 051		052 - TRACES	57	5	52	0		
2022-451	148	Mission500		051 051	051 051		051 - NEGATIVE	57	52	52	100		
2022-451	151	Clari. Uro		051 051	051 051		051 - NEGATIVE	10	10	10	100		
064 URIN-PROTEIN QUALI.													
2022-452	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	47	47	47	100		
2022-452	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	33	33	33	100		
2022-452	067	URISCAN		051 051	051 051		052 - TRACES	54	1	53	0		
2022-452	067	URISCAN		051 051	051 051		051 - NEGATIVE	54	53	53	100		
2022-452	072	URS-9-TECO		051 051	051 051		052 - TRACES	91	3	88	0		
2022-452	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	91	88	88	100		
2022-452	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	38	38	38	100		
2022-452	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	119	119	119	100		
2022-452	094	CLINITEK		051 051	051 051		051 - NEGATIVE	106	106	106	100		
2022-452	102	ACON LAB		051 051	051 051		051 - NEGATIVE	144	144	144	100		
2022-452	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	29	29	29	100		
2022-452	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	12	12	12	100		
2022-452	140	MISSION A		051 051	051 051		051 - NEGATIVE	51	51	51	100		
2022-452	145	CLINITEK		051 051	051 051		051 - NEGATIVE	16	16	16	100		
2022-452	147	Mission120		051 051	051 051		051 - NEGATIVE	40	40	40	100		
2022-452	148	Mission500		051 051	051 051		051 - NEGATIVE	57	57	57	100		
2022-452	151	Clari. Uro		051 051	051 051		051 - NEGATIVE	10	10	10	100		
2022-453	011	AMES NM10S		053 055	053 055		055 - 3+	47	18	47	100		
2022-453	011	AMES NM10S		053 055	053 055		053 - 1+	47	3	47	100		

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-453	011	AMES NM10S		053 055	053 055		054 - 2+	47	26	47	100
2022-453	013	BAYER CLI		053 055	053 055		054 - 2+	33	21	33	100
2022-453	013	BAYER CLI		053 055	053 055		055 - 3+	33	12	33	100
2022-453	067	URISCAN		053 055	053 055		053 - 1+	54	13	54	100
2022-453	067	URISCAN		053 055	053 055		054 - 2+	54	40	54	100
2022-453	067	URISCAN		053 055	053 055		055 - 3+	54	1	54	100
2022-453	072	URS-9-TECO		054 056	053 055		055 - 3+	91	53	90	100
2022-453	072	URS-9-TECO		054 056	053 055		051 - NEGATIVE	91	1	90	0
2022-453	072	URS-9-TECO		054 056	053 055		054 - 2+	91	37	90	100
2022-453	080	ROCHE (CHE		053 055	053 055		054 - 2+	38	27	38	100
2022-453	080	ROCHE (CHE		053 055	053 055		055 - 3+	38	11	38	100
2022-453	089	AIMSTRIP		053 055	053 055		053 - 1+	119	2	117	100
2022-453	089	AIMSTRIP		053 055	053 055		054 - 2+	119	62	117	100
2022-453	089	AIMSTRIP		053 055	053 055		056 - 4+	119	1	117	0
2022-453	089	AIMSTRIP		053 055	053 055		051 - NEGATIVE	119	1	117	0
2022-453	089	AIMSTRIP		053 055	053 055		055 - 3+	119	53	117	100
2022-453	094	CLINITEK		053 055	053 055		053 - 1+	106	4	106	100
2022-453	094	CLINITEK		053 055	053 055		054 - 2+	106	83	106	100
2022-453	094	CLINITEK		053 055	053 055		055 - 3+	106	19	106	100
2022-453	102	ACON LAB		054 056	053 055		055 - 3+	144	91	144	100
2022-453	102	ACON LAB		054 056	053 055		056 - 4+	144	2	144	100
2022-453	102	ACON LAB		054 056	053 055		054 - 2+	144	50	144	100
2022-453	102	ACON LAB		054 056	053 055		053 - 1+	144	1	144	100
2022-453	131	ICHEM VELO		053 055	053 055		054 - 2+	29	26	28	100
2022-453	131	ICHEM VELO		053 055	053 055		056 - 4+	29	1	28	0
2022-453	131	ICHEM VELO		053 055	053 055		055 - 3+	29	2	28	100
2022-453	135	Hybrido AU		053 055	053 055		054 - 2+	12	11	12	100
2022-453	135	Hybrido AU		053 055	053 055		055 - 3+	12	1	12	100
2022-453	140	MISSION A		053 055	053 055		053 - 1+	51	1	51	100
2022-453	140	MISSION A		053 055	053 055		054 - 2+	51	28	51	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-453	140	MISSION A		053 055	053 055		055 - 3+	51	22	51	100	
2022-453	145	CLINITEK		054 056	053 055		055 - 3+	16	8	16	100	
2022-453	145	CLINITEK		054 056	053 055		054 - 2+	16	8	16	100	
2022-453	147	Mission120		053 055	053 055		053 - 1+	40	1	40	100	
2022-453	147	Mission120		053 055	053 055		054 - 2+	40	36	40	100	
2022-453	147	Mission120		053 055	053 055		055 - 3+	40	3	40	100	
2022-453	148	Mission500		054 056	053 055		053 - 1+	57	2	57	100	
2022-453	148	Mission500		054 056	053 055		054 - 2+	57	9	57	100	
2022-453	148	Mission500		054 056	053 055		055 - 3+	57	46	57	100	
2022-453	151	Clari. Uro		054 056	053 055		055 - 3+	10	5	10	100	
2022-453	151	Clari. Uro		054 056	053 055		054 - 2+	10	5	10	100	
064 URIN-PROTEIN QUALI.												
2022-454	011	AMES NM10S		053 055	053 055		053 - 1+	47	2	47	100	
2022-454	011	AMES NM10S		053 055	053 055		054 - 2+	47	30	47	100	
2022-454	011	AMES NM10S		053 055	053 055		055 - 3+	47	15	47	100	
2022-454	013	BAYER CLI		053 055	053 055		055 - 3+	33	13	33	100	
2022-454	013	BAYER CLI		053 055	053 055		054 - 2+	33	20	33	100	
2022-454	067	URISCAN		053 055	053 055		055 - 3+	54	7	54	100	
2022-454	067	URISCAN		053 055	053 055		054 - 2+	54	41	54	100	
2022-454	067	URISCAN		053 055	053 055		053 - 1+	54	6	54	100	
2022-454	072	URS-9-TECO		054 056	053 055		056 - 4+	91	1	91	100	
2022-454	072	URS-9-TECO		054 056	053 055		055 - 3+	91	47	91	100	
2022-454	072	URS-9-TECO		054 056	053 055		054 - 2+	91	43	91	100	
2022-454	080	ROCHE (CHE		053 055	053 055		055 - 3+	38	10	38	100	
2022-454	080	ROCHE (CHE		053 055	053 055		054 - 2+	38	28	38	100	
2022-454	089	AIMSTRIP		054 056	053 055		056 - 4+	119	1	119	100	
2022-454	089	AIMSTRIP		054 056	053 055		055 - 3+	119	73	119	100	
2022-454	089	AIMSTRIP		054 056	053 055		053 - 1+	119	2	119	100	
2022-454	089	AIMSTRIP		054 056	053 055		054 - 2+	119	43	119	100	
2022-454	094	CLINITEK		053 055	053 055		055 - 3+	106	20	106	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	094	CLINITEK		053 055	053 055		054 - 2+	106	82	106	100
2022-454	094	CLINITEK		053 055	053 055		053 - 1+	106	4	106	100
2022-454	102	ACON LAB		054 056	053 055		055 - 3+	144	105	144	100
2022-454	102	ACON LAB		054 056	053 055		054 - 2+	144	35	144	100
2022-454	102	ACON LAB		054 056	053 055		056 - 4+	144	4	144	100
2022-454	131	iCHEM VELO		053 055	053 055		056 - 4+	29	1	28	0
2022-454	131	iCHEM VELO		053 055	053 055		054 - 2+	29	26	28	100
2022-454	131	iCHEM VELO		053 055	053 055		055 - 3+	29	2	28	100
2022-454	135	Hybrido AU		053 055	053 055		055 - 3+	12	1	12	100
2022-454	135	Hybrido AU		053 055	053 055		054 - 2+	12	11	12	100
2022-454	140	MISSION A		053 055	053 055		055 - 3+	51	24	51	100
2022-454	140	MISSION A		053 055	053 055		053 - 1+	51	1	51	100
2022-454	140	MISSION A		053 055	053 055		054 - 2+	51	26	51	100
2022-454	145	CLINITEK		053 055	053 055		055 - 3+	16	6	16	100
2022-454	145	CLINITEK		053 055	053 055		053 - 1+	16	1	16	100
2022-454	145	CLINITEK		053 055	053 055		054 - 2+	16	9	16	100
2022-454	147	Mission120		053 055	053 055		055 - 3+	40	13	40	100
2022-454	147	Mission120		053 055	053 055		053 - 1+	40	1	40	100
2022-454	147	Mission120		053 055	053 055		054 - 2+	40	26	40	100
2022-454	148	Mission500		054 056	053 055		055 - 3+	57	52	57	100
2022-454	148	Mission500		054 056	053 055		054 - 2+	57	4	57	100
2022-454	148	Mission500		054 056	053 055		053 - 1+	57	1	57	100
2022-454	151	Clari. Uro		054 056	053 055		055 - 3+	10	6	10	100
2022-454	151	Clari. Uro		054 056	053 055		054 - 2+	10	4	10	100
064 URIN-PROTEIN QUALI.											
2022-455	011	AMES NM10S		053 055	053 055		055 - 3+	47	9	47	100
2022-455	011	AMES NM10S		053 055	053 055		053 - 1+	47	3	47	100
2022-455	011	AMES NM10S		053 055	053 055		054 - 2+	47	35	47	100
2022-455	013	BAYER CLI		053 055	053 055		054 - 2+	33	24	33	100
2022-455	013	BAYER CLI		053 055	053 055		055 - 3+	33	9	33	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-455	067	URISCAN		053 055	053 055		055 - 3+	54	4	54	100
2022-455	067	URISCAN		053 055	053 055		053 - 1+	54	23	54	100
2022-455	067	URISCAN		053 055	053 055		054 - 2+	54	27	54	100
2022-455	072	URS-9-TECO		053 055	053 055		055 - 3+	91	30	91	100
2022-455	072	URS-9-TECO		053 055	053 055		053 - 1+	91	3	91	100
2022-455	072	URS-9-TECO		053 055	053 055		054 - 2+	91	58	91	100
2022-455	080	ROCHE (CHE		053 055	053 055		055 - 3+	38	9	38	100
2022-455	080	ROCHE (CHE		053 055	053 055		054 - 2+	38	29	38	100
2022-455	089	AIMSTRIP		053 055	053 055		056 - 4+	119	1	118	0
2022-455	089	AIMSTRIP		053 055	053 055		055 - 3+	119	38	118	100
2022-455	089	AIMSTRIP		053 055	053 055		053 - 1+	119	3	118	100
2022-455	089	AIMSTRIP		053 055	053 055		054 - 2+	119	77	118	100
2022-455	094	CLINITEK		053 055	053 055		053 - 1+	106	4	106	100
2022-455	094	CLINITEK		053 055	053 055		054 - 2+	106	91	106	100
2022-455	094	CLINITEK		053 055	053 055		055 - 3+	106	11	106	100
2022-455	102	ACON LAB		053 055	053 055		054 - 2+	144	90	141	100
2022-455	102	ACON LAB		053 055	053 055		053 - 1+	144	2	141	100
2022-455	102	ACON LAB		053 055	053 055		055 - 3+	144	49	141	100
2022-455	102	ACON LAB		053 055	053 055		056 - 4+	144	3	141	0
2022-455	131	iCHEM VELO		053 055	053 055		055 - 3+	29	2	29	100
2022-455	131	iCHEM VELO		053 055	053 055		054 - 2+	29	27	29	100
2022-455	135	Hybrido AU		053 055	053 055		054 - 2+	12	11	12	100
2022-455	135	Hybrido AU		053 055	053 055		055 - 3+	12	1	12	100
2022-455	140	MISSION A		053 055	053 055		052 - TRACES	51	1	50	0
2022-455	140	MISSION A		053 055	053 055		053 - 1+	51	2	50	100
2022-455	140	MISSION A		053 055	053 055		054 - 2+	51	40	50	100
2022-455	140	MISSION A		053 055	053 055		055 - 3+	51	8	50	100
2022-455	145	CLINITEK		053 055	053 055		053 - 1+	16	1	16	100
2022-455	145	CLINITEK		053 055	053 055		054 - 2+	16	15	16	100
2022-455	147	Mission120		053 055	053 055		053 - 1+	40	2	40	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-455	147	Mission120		053 055	053 055		054 - 2+	40	37	40	100	
2022-455	147	Mission120		053 055	053 055		055 - 3+	40	1	40	100	
2022-455	148	Mission500		053 055	053 055		055 - 3+	57	21	57	100	
2022-455	148	Mission500		053 055	053 055		054 - 2+	57	34	57	100	
2022-455	148	Mission500		053 055	053 055		053 - 1+	57	2	57	100	
2022-455	151	Clari. Uro		053 055	053 055		054 - 2+	10	8	10	100	
2022-455	151	Clari. Uro		053 055	053 055		055 - 3+	10	2	10	100	
066 URINAL.-UROBILINOGEN												
2022-451	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	47	47	47	100	
2022-451	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2022-451	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	54	54	54	100	
2022-451	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	91	91	91	100	
2022-451	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	38	38	38	100	
2022-451	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	119	119	119	100	
2022-451	094	CLINITEK		061 061	061 061		062 - 2	107	1	106	0	
2022-451	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	106	106	100	
2022-451	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	143	143	143	100	
2022-451	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	29	29	29	100	
2022-451	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2022-451	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2022-451	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2022-451	147	Mission120		061 061	061 061		061 - NORMAL,NEG	40	40	40	100	
2022-451	148	Mission500		061 061	061 061		061 - NORMAL,NEG	56	56	56	100	
2022-451	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
2022-452	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	47	47	47	100	
2022-452	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2022-452	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	54	54	54	100	
2022-452	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	91	91	91	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-452	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	38	38	38	100	
2022-452	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	119	119	119	100	
2022-452	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	106	106	100	
2022-452	094	CLINITEK		061 061	061 061		062 - 2	107	1	106	0	
2022-452	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	143	143	143	100	
2022-452	131	ICHEM VELO		061 061	061 061		061 - NORMAL,NEG	29	29	29	100	
2022-452	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2022-452	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2022-452	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2022-452	147	Mission120		061 061	061 061		061 - NORMAL,NEG	40	40	40	100	
2022-452	148	Mission500		061 061	061 061		061 - NORMAL,NEG	56	56	56	100	
2022-452	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
066 URINAL.-UROBILINOGEN												
2022-453	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	47	47	47	100	
2022-453	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2022-453	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	54	54	54	100	
2022-453	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	91	91	91	100	
2022-453	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	38	38	38	100	
2022-453	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	119	119	119	100	
2022-453	094	CLINITEK		061 061	061 061		062 - 2	107	1	106	0	
2022-453	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	106	106	100	
2022-453	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	143	143	143	100	
2022-453	131	ICHEM VELO		061 061	061 061		061 - NORMAL,NEG	29	29	29	100	
2022-453	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2022-453	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2022-453	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2022-453	147	Mission120		061 061	061 061		061 - NORMAL,NEG	40	40	40	100	
2022-453	148	Mission500		061 061	061 061		061 - NORMAL,NEG	56	56	56	100	
2022-453	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
066 URINAL.-UROBILINOGEN											
2022-454	011	AMES NM10S		061 063	061 063		063 - 4	47	13	47	100
2022-454	011	AMES NM10S		061 063	061 063		062 - 2	47	28	47	100
2022-454	011	AMES NM10S		061 063	061 063		061 - NORMAL,NEG	47	6	47	100
2022-454	013	BAYER CLI		061 063	061 063		062 - 2	34	20	34	100
2022-454	013	BAYER CLI		061 063	061 063		063 - 4	34	2	34	100
2022-454	013	BAYER CLI		061 063	061 063		061 - NORMAL,NEG	34	12	34	100
2022-454	067	URISCAN			061 063		061 - NORMAL,NEG	54	18	54	100
2022-454	067	URISCAN			061 063		062 - 2	54	7	54	100
2022-454	067	URISCAN			061 063		063 - 4	54	29	54	100
2022-454	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	91	19	89	100
2022-454	072	URS-9-TECO		061 063	061 063		064 - 8	91	2	89	0
2022-454	072	URS-9-TECO		061 063	061 063		063 - 4	91	22	89	100
2022-454	072	URS-9-TECO		061 063	061 063		062 - 2	91	48	89	100
2022-454	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	38	32	38	100
2022-454	080	ROCHE (CHE		061 061	061 063		062 - 2	38	6	38	100
2022-454	089	AIMSTRIP			061 063		061 - NORMAL,NEG	119	79	119	100
2022-454	089	AIMSTRIP			061 063		063 - 4	119	11	119	100
2022-454	089	AIMSTRIP			061 063		062 - 2	119	29	119	100
2022-454	094	CLINITEK		061 063	061 063		061 - NORMAL,NEG	107	22	107	100
2022-454	094	CLINITEK		061 063	061 063		063 - 4	107	34	107	100
2022-454	094	CLINITEK		061 063	061 063		062 - 2	107	51	107	100
2022-454	102	ACON LAB			061 063		063 - 4	143	3	143	100
2022-454	102	ACON LAB			061 063		062 - 2	143	27	143	100
2022-454	102	ACON LAB			061 063		061 - NORMAL,NEG	143	113	143	100
2022-454	131	iCHEM VELO		062 064	061 063		063 - 4	29	29	29	100
2022-454	135	Hybrido AU		061 063	061 063		061 - NORMAL,NEG	12	4	12	100
2022-454	135	Hybrido AU		061 063	061 063		062 - 2	12	8	12	100
2022-454	140	MISSION A		061 061	061 063		061 - NORMAL,NEG	51	42	51	100
2022-454	140	MISSION A		061 061	061 063		063 - 4	51	1	51	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-454	140	MISSION A		061 061	061 063		062 - 2	51	8	51	100	
2022-454	145	CLINITEK		061 063	061 063		062 - 2	16	16	16	100	
2022-454	147	Mission120		061 061	061 063		061 - NORMAL,NEG	40	39	40	100	
2022-454	147	Mission120		061 061	061 063		062 - 2	40	1	40	100	
2022-454	148	Mission500		061 061	061 063		063 - 4	56	1	56	100	
2022-454	148	Mission500		061 061	061 063		062 - 2	56	3	56	100	
2022-454	148	Mission500		061 061	061 063		061 - NORMAL,NEG	56	52	56	100	
2022-454	151	Clari. Uro			061 063		061 - NORMAL,NEG	10	6	10	100	
2022-454	151	Clari. Uro			061 063		062 - 2	10	4	10	100	
066 URINAL.-UROBILINOGEN												
2022-455	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	47	47	47	100	
2022-455	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	34	34	34	100	
2022-455	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	54	54	54	100	
2022-455	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	91	91	91	100	
2022-455	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	38	38	38	100	
2022-455	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	119	119	119	100	
2022-455	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	107	106	106	100	
2022-455	094	CLINITEK		061 061	061 061		062 - 2	107	1	106	0	
2022-455	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	143	143	143	100	
2022-455	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	29	29	29	100	
2022-455	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	12	12	12	100	
2022-455	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	51	51	51	100	
2022-455	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	16	16	16	100	
2022-455	147	Mission120		061 061	061 061		061 - NORMAL,NEG	40	40	40	100	
2022-455	148	Mission500		061 061	061 061		061 - NORMAL,NEG	56	56	56	100	
2022-455	151	Clari. Uro		061 061	061 061		061 - NORMAL,NEG	10	10	10	100	
184 URINALYSIS LEUKOCYTE												
2022-451	011	AMES NM10S		001 001	001 001		001 - Negative	46	46	46	100	
2022-451	013	BAYER CLI		001 001	001 001		001 - Negative	34	34	34	100	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-451	067	URISCAN		001 001	001 001		001 - Negative	54	54	54	100
2022-451	072	URS-9-TECO		001 001	001 001		001 - Negative	85	84	84	100
2022-451	072	URS-9-TECO		001 001	001 001		003 - Small	85	1	84	0
2022-451	080	ROCHE (CHE		001 001	001 001		001 - Negative	35	35	35	100
2022-451	089	AIMSTRIP		001 001	001 001		001 - Negative	116	116	116	100
2022-451	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2022-451	102	ACON LAB		001 001	001 001		001 - Negative	130	130	130	100
2022-451	131	iCHEM VELO		001 001	001 001		001 - Negative	29	29	29	100
2022-451	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2022-451	140	MISSION A		001 001	001 001		001 - Negative	58	58	58	100
2022-451	145	CLINITEK		001 001	001 001		001 - Negative	16	16	16	100
2022-451	147	Mission120		001 001	001 001		001 - Negative	38	38	38	100
2022-451	148	Mission500		001 001	001 001		001 - Negative	56	56	56	100
2022-451	151	Clari. Uro		001 001	001 001		001 - Negative	10	10	10	100
184 URINALYSIS LEUKOCYTE											
2022-452	011	AMES NM10S		002 004	003 005		003 - Small	46	37	46	100
2022-452	011	AMES NM10S		002 004	003 005		005 - Large	46	1	46	100
2022-452	011	AMES NM10S		002 004	003 005		002 - Traces	46	1	46	100
2022-452	011	AMES NM10S		002 004	003 005		004 - Moderate	46	7	46	100
2022-452	013	BAYER CLI		002 004	003 005		003 - Small	34	27	34	100
2022-452	013	BAYER CLI		002 004	003 005		002 - Traces	34	1	34	100
2022-452	013	BAYER CLI		002 004	003 005		005 - Large	34	1	34	100
2022-452	013	BAYER CLI		002 004	003 005		004 - Moderate	34	5	34	100
2022-452	067	URISCAN		003 005	003 005		005 - Large	54	26	54	100
2022-452	067	URISCAN		003 005	003 005		004 - Moderate	54	26	54	100
2022-452	067	URISCAN		003 005	003 005		003 - Small	54	2	54	100
2022-452	072	URS-9-TECO		003 005	003 005		003 - Small	85	33	84	100
2022-452	072	URS-9-TECO		003 005	003 005		005 - Large	85	1	84	100
2022-452	072	URS-9-TECO		003 005	003 005		002 - Traces	85	1	84	0
2022-452	072	URS-9-TECO		003 005	003 005		004 - Moderate	85	50	84	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-452	080	ROCHE (CHE		004	005	003	005	004 - Moderate	35	3	35	100
2022-452	080	ROCHE (CHE		004	005	003	005	005 - Large	35	32	35	100
2022-452	089	AIMSTRIP		002	004	003	005	002 - Traces	116	4	116	100
2022-452	089	AIMSTRIP		002	004	003	005	004 - Moderate	116	33	116	100
2022-452	089	AIMSTRIP		002	004	003	005	003 - Small	116	79	116	100
2022-452	094	CLINITEK		002	004	003	005	004 - Moderate	107	4	107	100
2022-452	094	CLINITEK		002	004	003	005	003 - Small	107	101	107	100
2022-452	094	CLINITEK		002	004	003	005	002 - Traces	107	2	107	100
2022-452	102	ACON LAB		002	004	003	005	005 - Large	130	1	129	100
2022-452	102	ACON LAB		002	004	003	005	002 - Traces	130	6	129	100
2022-452	102	ACON LAB		002	004	003	005	001 - Negative	130	1	129	0
2022-452	102	ACON LAB		002	004	003	005	003 - Small	130	73	129	100
2022-452	102	ACON LAB		002	004	003	005	004 - Moderate	130	49	129	100
2022-452	131	iCHEM VELO		004	005	003	005	004 - Moderate	29	2	29	100
2022-452	131	iCHEM VELO		004	005	003	005	005 - Large	29	27	29	100
2022-452	135	Hybrido AU		004	005	003	005	005 - Large	12	12	12	100
2022-452	140	MISSION A		002	004	003	005	004 - Moderate	58	24	58	100
2022-452	140	MISSION A		002	004	003	005	003 - Small	58	28	58	100
2022-452	140	MISSION A		002	004	003	005	005 - Large	58	3	58	100
2022-452	140	MISSION A		002	004	003	005	002 - Traces	58	3	58	100
2022-452	145	CLINITEK		003	005	003	005	003 - Small	16	1	16	100
2022-452	145	CLINITEK		003	005	003	005	004 - Moderate	16	15	16	100
2022-452	147	Mission120		002	004	003	005	002 - Traces	38	8	37	100
2022-452	147	Mission120		002	004	003	005	003 - Small	38	22	37	100
2022-452	147	Mission120		002	004	003	005	004 - Moderate	38	7	37	100
2022-452	147	Mission120		002	004	003	005	001 - Negative	38	1	37	0
2022-452	148	Mission500		003	005	003	005	003 - Small	56	14	56	100
2022-452	148	Mission500		003	005	003	005	005 - Large	56	2	56	100
2022-452	148	Mission500		003	005	003	005	004 - Moderate	56	40	56	100
2022-452	151	Clari. Uro		002	004	003	005	005 - Large	10	1	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2022-452	151	Clari. Uro		002 004	003 005		004 - Moderate	10	3	10	100	
2022-452	151	Clari. Uro		002 004	003 005		003 - Small	10	6	10	100	
184 URINALYSIS LEUKOCYTE												
2022-453	011	AMES NM10S		001 001	001 001		001 - Negative	46	46	46	100	
2022-453	013	BAYER CLI		001 001	001 001		001 - Negative	34	34	34	100	
2022-453	067	URISCAN		001 001	001 001		001 - Negative	54	54	54	100	
2022-453	072	URS-9-TECO		001 001	001 001		003 - Small	85	1	84	0	
2022-453	072	URS-9-TECO		001 001	001 001		001 - Negative	85	84	84	100	
2022-453	080	ROCHE (CHE		001 001	001 001		001 - Negative	35	35	35	100	
2022-453	089	AIMSTRIP		001 001	001 001		004 - Moderate	116	1	115	0	
2022-453	089	AIMSTRIP		001 001	001 001		001 - Negative	116	115	115	100	
2022-453	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100	
2022-453	102	ACON LAB		001 001	001 001		002 - Traces	130	1	129	0	
2022-453	102	ACON LAB		001 001	001 001		001 - Negative	130	129	129	100	
2022-453	131	iCHEM VELO		001 001	001 001		001 - Negative	29	29	29	100	
2022-453	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100	
2022-453	140	MISSION A		001 001	001 001		001 - Negative	58	57	57	100	
2022-453	140	MISSION A		001 001	001 001		003 - Small	58	1	57	0	
2022-453	145	CLINITEK		001 001	001 001		001 - Negative	16	16	16	100	
2022-453	147	Mission120		001 001	001 001		001 - Negative	38	38	38	100	
2022-453	148	Mission500		001 001	001 001		001 - Negative	56	56	56	100	
2022-453	151	Clari. Uro		001 001	001 001		001 - Negative	10	10	10	100	
2022-454	011	AMES NM10S		001 001	001 001		001 - Negative	46	46	46	100	
2022-454	013	BAYER CLI		001 001	001 001		001 - Negative	34	34	34	100	
2022-454	067	URISCAN		001 001	001 001		001 - Negative	54	54	54	100	
2022-454	072	URS-9-TECO		001 001	001 001		001 - Negative	85	85	85	100	
2022-454	080	ROCHE (CHE		001 001	001 001		001 - Negative	35	35	35	100	
2022-454	089	AIMSTRIP		001 001	001 001		005 - Large	116	1	113	0	

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2022-454	089	AIMSTRIP		001 001	001 001		002 - Traces	116	1	113	0
2022-454	089	AIMSTRIP		001 001	001 001		001 - Negative	116	113	113	100
2022-454	089	AIMSTRIP		001 001	001 001		003 - Small	116	1	113	0
2022-454	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2022-454	102	ACON LAB		001 001	001 001		001 - Negative	130	130	130	100
2022-454	131	iCHEM VELO		001 001	001 001		001 - Negative	29	29	29	100
2022-454	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2022-454	140	MISSION A		001 001	001 001		001 - Negative	58	58	58	100
2022-454	145	CLINITEK		001 001	001 001		001 - Negative	16	16	16	100
2022-454	147	Mission120		001 001	001 001		001 - Negative	38	38	38	100
2022-454	148	Mission500		001 001	001 001		001 - Negative	56	56	56	100
2022-454	151	Clari. Uro		001 001	001 001		001 - Negative	10	10	10	100
184 URINALYSIS LEUKOCYTE											
2022-455	011	AMES NM10S		001 001	001 001		001 - Negative	46	46	46	100
2022-455	013	BAYER CLI		001 001	001 001		001 - Negative	34	34	34	100
2022-455	067	URISCAN		001 001	001 001		001 - Negative	54	54	54	100
2022-455	072	URS-9-TECO		001 001	001 001		001 - Negative	85	85	85	100
2022-455	080	ROCHE (CHE		001 001	001 001		001 - Negative	35	35	35	100
2022-455	089	AIMSTRIP		001 001	001 001		001 - Negative	116	116	116	100
2022-455	094	CLINITEK		001 001	001 001		001 - Negative	107	107	107	100
2022-455	102	ACON LAB		001 001	001 001		001 - Negative	130	130	130	100
2022-455	131	iCHEM VELO		001 001	001 001		001 - Negative	29	29	29	100
2022-455	135	Hybrido AU		001 001	001 001		001 - Negative	12	12	12	100
2022-455	140	MISSION A		001 001	001 001		001 - Negative	58	58	58	100
2022-455	145	CLINITEK		001 001	001 001		001 - Negative	16	16	16	100
2022-455	147	Mission120		001 001	001 001		001 - Negative	38	38	38	100
2022-455	148	Mission500		001 001	001 001		001 - Negative	56	56	56	100
2022-455	151	Clari. Uro		001 001	001 001		001 - Negative	10	10	10	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2022-451	091	AIM STEP		001	001	001	001	001 - NEGATIVE	33	33	33	100
2022-451	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100
2022-452	091	AIM STEP		002	002	002	002	001 - NEGATIVE	33	1	32	0
2022-452	091	AIM STEP		002	002	002	002	002 - POSITIVE	33	32	32	100
2022-452	101	HCG Alere			002	002		001 - NEGATIVE	23	14	9	0
2022-452	101	HCG Alere			002	002		002 - POSITIVE	23	9	9	100
2022-453	091	AIM STEP		002	002	002	002	002 - POSITIVE	33	33	33	100
2022-453	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2022-454	091	AIM STEP		002	002	002	002	002 - POSITIVE	33	33	33	100
2022-454	101	HCG Alere		002	002	002	002	002 - POSITIVE	23	23	23	100
2022-455	091	AIM STEP		001	001	001	001	001 - NEGATIVE	33	33	33	100
2022-455	101	HCG Alere		001	001	001	001	001 - NEGATIVE	23	23	23	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2022-461					006	006	006 - WAXY CAST	880	879	879	100
2022-461					006	006	004 - HYALINE C.	880	1	879	0
2022-462					027	027	022 - TYROSINE	879	20	830	0
2022-462					027	027	026 - CA. CARBO.	879	1	830	0
2022-462					027	027	014 - CA. OXALAT	879	3	830	0
2022-462					027	027	021 - SULFANAM.	879	1	830	0
2022-462					027	027	027 - CA. PHOSP.	879	830	830	100
2022-462					027	027	025 - AMORPHOUS	879	2	830	0
2022-462					027	027	015 - CA. SULFAT	879	19	830	0
2022-462					027	027	028 - TRIPLE PH.	879	3	830	0
2022-463					032	032	031 - FIBERS	879	35	843	0
2022-463					032	032	012 - TRANSITIO.	879	1	843	0
2022-463					032	032	032 - MUCOUS TH.	879	843	843	100
2022-464					008	008	008 - ERYTHROCYT	879	877	877	100
2022-464					008	008	039 - E. vermic.	879	1	877	0
2022-464					008	008	005 - RED CELL C	879	1	877	0
2022-465					039	039	039 - E. vermic.	879	879	879	100

Participant Summary

MONTH 5 OF YEAR 2022

SPE: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 98 OCCULT BLOOD							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
154 FECAL OCCULT BLOOD													
2022-471	001	PEROXIDASE		001	001	001	001	002 - POSITIVE	426	2	424	0	
2022-471	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	426	424	424	100	
2022-471	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	706	703	703	100	
2022-471	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	706	3	703	0	
2022-472	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	426	424	424	100	
2022-472	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	426	2	424	0	
2022-472	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	706	703	703	100	
2022-472	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	706	3	703	0	
2022-473	001	PEROXIDASE		001	001	001	001	001 - NEGATIVE	426	426	426	100	
2022-473	002	INMMUASSAY		001	001	001	001	001 - NEGATIVE	706	705	705	100	
2022-473	002	INMMUASSAY		001	001	001	001	002 - POSITIVE	706	1	705	0	
2022-474	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	426	425	425	100	
2022-474	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	426	1	425	0	
2022-474	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	706	706	706	100	
2022-475	001	PEROXIDASE		002	002	002	002	002 - POSITIVE	426	425	425	100	
2022-475	001	PEROXIDASE		002	002	002	002	001 - NEGATIVE	426	1	425	0	
2022-475	002	INMMUASSAY		002	002	002	002	001 - NEGATIVE	706	3	703	0	
2022-475	002	INMMUASSAY		002	002	002	002	002 - POSITIVE	706	703	703	100	
SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS				GRP: 90 SPECIAL TEST							

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2022-421	38 ROCHE E			21	5.538	0.336	6.07	4.500	6.500
2022-421	44 BECK.COUL.			15	5.140	0.346	6.73	4.100	6.200
2022-421	50 ARC			10	6.570	0.335	5.10	5.600	7.600
2022-422	38 ROCHE E			21	0.790	0.134	16.94	0.600	1.000
2022-422	44 BECK.COUL.			15	0.620	0.083	13.43	0.500	0.800
2022-422	50 ARC			10	0.840	0.215	25.64	0.600	1.100
2022-423	38 ROCHE E			21	2.719	0.238	8.76	2.000	3.400
2022-423	44 BECK.COUL.			15	2.487	0.268	10.78	1.900	3.100
2022-423	50 ARC			10	2.590	0.114	4.39	2.200	2.900
2022-424	38 ROCHE E			21	8.724	0.436	5.00	7.400	10.000
2022-424	44 BECK.COUL.			15	8.420	0.446	5.30	7.100	9.800
2022-424	50 ARC			10	10.870	0.480	4.41	9.400	12.300
2022-425	38 ROCHE E			21	22.057	0.926	4.20	19.300	24.800
2022-425	44 BECK.COUL.			15	22.840	1.319	5.78	18.900	26.800
2022-425	50 ARC			10	28.890	0.394	1.36	27.700	30.100
069	PROS. SPEC-ANTIG-PSA								
2022-421	29 VITROS			41	2.710	0.156	5.75	2.200	3.200
2022-421	36 TOSOH AIA			13	3.423	0.255	7.44	2.700	4.200
2022-421	36 TOSOH AIA			13	4.033	0.340	8.43	3.000	5.000
2022-421	38 ROCHE E			53	5.264	0.367	6.97	4.200	6.400
2022-421	39 DIMENS.RXL			41	4.808	0.189	3.93	4.200	5.400
2022-421	44 BECK.COUL.			55	4.930	0.265	5.38	4.100	5.700
2022-421	44 BECK.COUL.			55	4.033	0.340	8.43	3.000	5.000
2022-421	50 ARC			22	4.245	0.292	6.88	3.400	5.100
2022-421	57 S. EXL 200			29	4.811	0.335	6.97	3.800	5.800

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY			SUB: 60 SPECIAL TESTS			GRP: 90 SPECIAL TEST				
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2022-421	57 S. EXL 200			29	4.033	0.340	8.43	3.000	5.000	
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}							
2022-422	29 VITROS			41	0.100	0.000	0.00	0.100	0.100	
2022-422	36 TOSOH AIA			13	0.133	0.047	35.36	0.100	0.200	
2022-422	36 TOSOH AIA			13	0.162	0.173	107.33	0.100	0.200	
2022-422	38 ROCHE E			53	0.096	0.019	20.00	0.100	0.100	
2022-422	39 DIMENS.RXL			41	0.088	0.033	37.27	0.100	0.100	
2022-422	44 BECK.COUL.			55	0.133	0.047	35.36	0.100	0.200	
2022-422	44 BECK.COUL.			55	0.096	0.019	19.61	0.100	0.100	
2022-422	50 ARC			22	0.014	0.034	251.66	0.000	0.000	
2022-422	57 S. EXL 200			29	0.133	0.047	35.36	0.100	0.200	
2022-422	57 S. EXL 200			29	0.076	0.043	56.41	0.100	0.100	
2022-423	29 VITROS			41	0.200	0.000	0.00	0.200	0.200	
2022-423	36 TOSOH AIA			13	0.100	0.000	0.00	0.100	0.100	
2022-423	36 TOSOH AIA			13	0.133	0.047	35.36	0.100	0.200	
2022-423	38 ROCHE E			53	0.110	0.029	26.89	0.100	0.100	
2022-423	39 DIMENS.RXL			41	0.100	0.000	0.00	0.100	0.100	
2022-423	44 BECK.COUL.			55	0.100	0.000	0.00	0.100	0.100	
2022-423	44 BECK.COUL.			55	0.133	0.047	35.36	0.100	0.200	
2022-423	50 ARC			22	0.100	0.000	0.00	0.100	0.100	
2022-423	57 S. EXL 200			29	0.100	0.000	0.00	0.100	0.100	
2022-423	57 S. EXL 200			29	0.133	0.047	35.36	0.100	0.200	
2022-424	29 VITROS			41	4.420	0.244	5.52	3.700	5.200	
2022-424	36 TOSOH AIA			13	7.033	0.403	5.73	5.800	8.200	
2022-424	36 TOSOH AIA			13	5.854	0.605	10.33	4.400	7.300	
2022-424	38 ROCHE E			53	9.106	0.563	6.19	7.400	10.800	
2022-424	39 DIMENS.RXL			41	8.359	0.377	4.52	7.200	9.500	
2022-424	44 BECK.COUL.			55	7.033	0.403	5.73	5.800	8.200	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-424	44	BECK.COUL.		55	8.476	0.513	6.05	6.900	10.000
2022-424	50	ARC		22	7.332	0.408	5.57	6.100	8.600
2022-424	57	S. EXL 200		29	7.033	0.403	5.73	5.800	8.200
2022-424	57	S. EXL 200		29	8.332	0.529	6.35	6.700	9.900
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2022-425	29	VITROS		41	11.739	0.725	6.18	9.600	13.900
2022-425	36	TOSOH AIA		13	15.538	0.986	6.34	12.600	18.500
2022-425	36	TOSOH AIA		13	20.100	1.306	6.50	16.200	24.000
2022-425	38	ROCHE E		53	24.268	1.382	5.70	20.100	28.400
2022-425	39	DIMENS.RXL		41	23.154	1.196	5.16	19.600	26.700
2022-425	44	BECK.COUL.		55	23.115	1.295	5.60	19.200	27.000
2022-425	44	BECK.COUL.		55	20.100	1.306	6.50	16.200	24.000
2022-425	50	ARC		22	20.005	1.086	5.43	16.700	23.300
2022-425	57	S. EXL 200		29	20.100	1.306	6.50	16.200	24.000
2022-425	57	S. EXL 200		29	23.355	1.564	6.70	18.700	28.000
071	VITAMIN B 12								
2022-421	29	VITROS		23	348.087	15.981	4.59	300.000	396.000
2022-421	38	ROCHE E		30	330.233	69.516	21.05	248.000	413.000
2022-421	44	BECK.COUL.		27	266.846	27.704	10.38	200.000	334.000
2022-421	50	ARC		19	346.000	20.688	5.98	284.000	408.000
2022-421	57	S. EXL 200		15	373.467	46.104	12.34	280.000	467.000
2022-422	29	VITROS		23	233.727	9.803	4.19	204.000	263.000
2022-422	38	ROCHE E		30	205.700	46.623	22.67	154.000	257.000
2022-422	44	BECK.COUL.		27	155.500	14.238	9.16	117.000	194.000
2022-422	50	ARC		19	218.278	20.904	9.58	164.000	273.000
2022-422	57	S. EXL 200		15	237.667	43.005	18.09	178.000	297.000
2022-423	29	VITROS		23	0.000	0.000	0.00	0.000	0.000
2022-423	38	ROCHE E		30	612.533	60.855	9.93	459.000	766.000

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2022-423	44	BECK.COUL.		27	352.385	38.650	10.97	264.000	440.000	
2022-423	50	ARC		19	545.579	37.169	6.81	434.000	657.000	
2022-423	57	S. EXL 200		15	511.933	67.246	13.14	384.000	640.000	
071 VITAMIN B 12			CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}					
2022-424	29	VITROS		23	466.217	23.778	5.10	395.000	538.000	
2022-424	38	ROCHE E		30	440.533	64.807	14.71	330.000	551.000	
2022-424	44	BECK.COUL.		27	346.370	29.058	8.39	260.000	433.000	
2022-424	50	ARC		19	433.444	17.183	3.96	382.000	485.000	
2022-424	57	S. EXL 200		15	483.200	41.579	8.60	362.000	604.000	
2022-425	29	VITROS		23	856.783	54.985	6.42	692.000	1022.000	
2022-425	38	ROCHE E		30	948.767	80.154	8.45	712.000	1186.000	
2022-425	44	BECK.COUL.		27	660.222	52.436	7.94	503.000	818.000	
2022-425	50	ARC		19	856.222	53.642	6.27	695.000	1017.000	
2022-425	57	S. EXL 200		15	904.133	63.261	7.00	714.000	1094.000	

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE						
ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
146 GLYCOHEMOGLOBIN										
2022-431	12	COBAS/INT.		30	4.967	0.281	5.66	4.100	5.800	
2022-431	14	IMMUNOTUR.		12	4.925	0.396	8.04	3.700	6.100	
2022-431	22	VITROS		45	4.784	0.092	1.92	4.500	5.100	
2022-431	30	DAYTONA		12	5.650	0.527	9.32	4.200	7.100	
2022-431	31	IMMUNOTURB		32	4.544	0.173	3.81	4.000	5.100	
2022-431	40	SIEMENS DI		82	5.061	0.176	3.47	4.500	5.600	
2022-431	46	ARCHITECT		17	4.681	0.095	2.03	4.400	5.000	
2022-432	12	COBAS/INT.		30	6.690	0.217	3.24	6.000	7.300	

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 95 GLYCOHEMOGLOBINE

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2022-432	14	IMMUNOTUR.		12	6.875	0.613	8.91	5.200	8.600
2022-432	22	VITROS		45	6.729	0.138	2.05	6.300	7.100
2022-432	30	DAYTONA		12	7.850	1.031	13.13	5.900	9.800
2022-432	31	IMMUNOTURB		32	6.541	0.210	3.22	5.900	7.200
2022-432	40	SIEMENS DI		82	6.890	0.181	2.63	6.300	7.400
2022-432	46	ARCHITECT		17	6.565	0.194	2.95	6.000	7.100
146	GLYCOHEMOGLOBIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2022-433	12	COBAS/INT.		30	10.810	0.291	2.70	9.900	11.700
2022-433	14	IMMUNOTUR.		12	11.025	0.902	8.18	8.300	13.700
2022-433	22	VITROS		45	10.723	0.275	2.57	9.900	11.500
2022-433	30	DAYTONA		12	12.142	0.938	7.72	9.300	15.000
2022-433	31	IMMUNOTURB		32	10.728	0.367	3.42	9.600	11.800
2022-433	40	SIEMENS DI		82	10.431	0.231	2.21	9.700	11.100
2022-433	46	ARCHITECT		17	10.388	0.331	3.18	9.400	11.400
2022-434	12	COBAS/INT.		30	4.983	0.249	5.00	4.200	5.700
2022-434	14	IMMUNOTUR.		12	4.917	0.329	6.69	3.900	5.900
2022-434	22	VITROS		45	4.807	0.083	1.72	4.600	5.100
2022-434	30	DAYTONA		12	5.567	0.464	8.34	4.200	7.000
2022-434	31	IMMUNOTURB		32	4.550	0.175	3.85	4.000	5.100
2022-434	40	SIEMENS DI		82	5.090	0.169	3.31	4.600	5.600
2022-434	46	ARCHITECT		17	4.647	0.158	3.39	4.200	5.100
2022-435	12	COBAS/INT.		30	6.717	0.240	3.57	6.000	7.400
2022-435	14	IMMUNOTUR.		12	6.992	0.633	9.05	5.200	8.700
2022-435	22	VITROS		45	6.750	0.122	1.80	6.400	7.100
2022-435	30	DAYTONA		12	7.517	0.395	5.26	6.300	8.700
2022-435	31	IMMUNOTURB		32	6.600	0.214	3.24	6.000	7.200
2022-435	40	SIEMENS DI		82	6.895	0.168	2.43	6.400	7.400
2022-435	46	ARCHITECT		17	6.565	0.227	3.46	5.900	7.200

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %								
2022-281	3	VITROS		41	44.488	3.736	8.40	36.000	53.000		
2022-281	5	ROCHE E		40	34.395	2.787	8.10	28.000	41.000		
2022-281	9	BECK.COUL.		47	47.340	3.460	7.31	38.000	57.000		
2022-281	9	BECK.COUL.		47	51.000	4.817	9.44	41.000	61.000		
2022-281	11	ARC		19	45.474	2.325	5.11	36.000	55.000		
2022-281	25	S. DIME		29	34.107	3.143	9.22	27.000	41.000		
2022-282	3	VITROS		41	26.024	3.189	12.25	21.000	31.000		
2022-282	5	ROCHE E		40	22.816	2.882	12.63	18.000	27.000		
2022-282	9	BECK.COUL.		47	31.600	3.007	9.51	25.000	38.000		
2022-282	9	BECK.COUL.		47	29.565	2.864	9.69	24.000	35.000		
2022-282	11	ARC		19	27.211	1.196	4.39	22.000	33.000		
2022-282	25	S. DIME		29	23.214	1.520	6.55	19.000	28.000		
2022-283	3	VITROS		41	32.732	3.092	9.45	26.000	39.000		
2022-283	5	ROCHE E		40	25.263	1.845	7.30	20.000	30.000		
2022-283	9	BECK.COUL.		47	34.043	3.007	8.83	27.000	41.000		
2022-283	9	BECK.COUL.		47	35.400	4.224	11.93	28.000	42.000		
2022-283	11	ARC		19	31.778	1.272	4.00	25.000	38.000		
2022-283	25	S. DIME		29	25.857	1.977	7.65	21.000	31.000		
2022-284	3	VITROS		41	61.585	4.819	7.82	49.000	74.000		
2022-284	5	ROCHE E		40	44.051	5.519	12.53	35.000	53.000		
2022-284	9	BECK.COUL.		47	73.600	5.004	6.80	59.000	88.000		
2022-284	9	BECK.COUL.		47	63.435	4.514	7.12	51.000	76.000		
2022-284	11	ARC		19	58.263	2.653	4.55	47.000	70.000		
2022-284	25	S. DIME		29	45.143	3.125	6.92	36.000	54.000		

Participant Summary

MONTH 5 OF YEAR 2022

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	20	%				
2022-285	3	VITROS		41	82.878	5.008	6.04	66.000	99.000
2022-285	5	ROCHE E		40	53.897	4.235	7.86	43.000	65.000
2022-285	9	BECK.COUL.		47	97.200	7.386	7.60	78.000	117.000
2022-285	9	BECK.COUL.		47	79.426	5.031	6.33	64.000	95.000
2022-285	11	ARC		19	73.000	3.920	5.37	58.000	88.000
2022-285	25	S. DIME		29	54.621	2.953	5.41	44.000	66.000