

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 15 WHOLE BLOOD GLUCOSE

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
157 WHOLE BLOOD GLUCOSE			CLIA TEST ID #: 0						EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}
2020-881	20	ABBOT FREE	18	422.111	12.853	3.05	380.000	464.000	
2020-881	23	ABBOTT PRE	31	355.677	24.565	6.91	320.000	391.000	
2020-881	27	STAT SRIP.	50	351.250	6.685	1.90	316.000	386.000	
2020-881	27	STAT SRIP.	50	279.740	19.710	7.05	252.000	308.000	
2020-881	33	AC.CHEK II	23	357.130	8.077	2.26	321.000	393.000	
2020-881	34	FREE STYL	33	404.939	25.961	6.41	364.000	445.000	
2020-882	20	ABBOT FREE	18	500.000	0.000	0.00	450.000	550.000	
2020-882	23	ABBOTT PRE	31	464.367	25.483	5.49	418.000	511.000	
2020-882	27	STAT SRIP.	50	349.265	28.945	8.29	314.000	384.000	
2020-882	27	STAT SRIP.	50	486.250	8.318	1.71	438.000	535.000	
2020-882	33	AC.CHEK II	23	495.217	15.604	3.15	446.000	545.000	
2020-882	34	FREE STYL	33	499.875	0.696	0.14	450.000	550.000	
2020-883	20	ABBOT FREE	18	500.000	0.000	0.00	450.000	550.000	
2020-883	23	ABBOTT PRE	31	486.032	33.662	6.93	437.000	535.000	
2020-883	27	STAT SRIP.	50	591.250	4.969	0.84	532.000	650.000	
2020-883	27	STAT SRIP.	50	428.080	32.796	7.66	385.000	471.000	
2020-883	33	AC.CHEK II	23	591.500	10.845	1.83	532.000	651.000	
2020-883	34	FREE STYL	33	500.000	0.000	0.00	450.000	550.000	
2020-884	20	ABBOT FREE	18	214.722	9.562	4.45	193.000	236.000	
2020-884	23	ABBOTT PRE	31	188.667	8.859	4.70	170.000	208.000	
2020-884	27	STAT SRIP.	50	164.959	6.833	4.14	148.000	181.000	
2020-884	27	STAT SRIP.	50	192.750	4.969	2.58	173.000	212.000	
2020-884	33	AC.CHEK II	23	192.957	5.238	2.71	174.000	212.000	
2020-884	34	FREE STYL	33	209.750	8.288	3.95	189.000	231.000	
2020-885	20	ABBOT FREE	18	20.000	0.000	0.00	14.000	26.000	

### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 15 WHOLE BLOOD GLUCOSE						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-885	23	ABBOTT PRE		31	19.900	0.300	1.51	14.000	26.000	
2020-885	27	STAT SRIP.		50	15.250	0.433	2.84	9.000	21.000	
2020-885	27	STAT SRIP.		50	10.213	1.009	9.88	4.000	16.000	
2020-885	33	AC.CHEK II		23	15.136	1.324	8.75	9.000	21.000	
2020-885	34	FREE STYL		33	20.000	0.000	0.00	14.000	26.000	
SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
080	FIBRINOGEN	CLIA TEST ID #: 825	EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-941	60	DADE SYSME	67 DADE	11	269.455	15.042	5.58	216.000	323.000	
2020-942	60	DADE SYSME	67 DADE	11	436.091	28.811	6.61	349.000	523.000	
2020-943	60	DADE SYSME	67 DADE	11	345.200	12.528	3.63	276.000	414.000	
2020-944	60	DADE SYSME	67 DADE	11	290.182	11.574	3.99	232.000	348.000	
2020-945	60	DADE SYSME	67 DADE	11	109.000	6.396	5.87	87.000	131.000	
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-941	33	MLA ELECT.	54 QUICK COAG	10	69.680	1.759	2.52	59.200	80.100	
2020-941	58	AMELUNG KC	42 BIOMERIEUX	10	58.510	4.072	6.96	49.700	67.300	
2020-941	60	DADE SYSME	35 DADE ACTIN	13	58.923	1.224	2.08	50.100	67.800	
2020-941	62	DADE SYS.	35 DADE ACTIN	59	55.170	2.997	5.43	46.900	63.400	
2020-941	65	DADE BFTII	35 DADE ACTIN	13	66.738	6.121	9.17	56.700	76.700	
2020-941	69	DIAG.STAGO	49 STA-PTT	12	56.558	2.290	4.05	48.100	65.000	
2020-941	73	ACL 8 TO E	56 PTT-HEMO	37	69.189	2.056	2.97	58.800	79.600	
2020-941	73	ACL 8 TO E	56 PTT-HEMO	37	67.350	1.011	1.50	57.200	77.500	
2020-941	76	STAGO STAR	49 STA-PTT	20	54.230	5.231	9.65	46.100	62.400	

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-942	33 MLA ELECT.	54 QUICK COAG		10	27.620	1.057	3.83	23.500	31.800	
2020-942	58 AMELUNG KC	42 BIOMERIEUX		10	32.390	1.862	5.75	27.500	37.200	
2020-942	60 DADE SYSME	35 DADE ACTIN		13	30.377	1.014	3.34	25.800	34.900	
2020-942	62 DADE SYS.	35 DADE ACTIN		59	29.804	2.047	6.87	25.300	34.300	
2020-942	65 DADE BFTII	35 DADE ACTIN		13	29.354	2.122	7.23	25.000	33.800	
2020-942	69 DIAG.STAGO	49 STA-PTT		12	33.267	0.972	2.92	28.300	38.300	
2020-942	73 ACL 8 TO E	56 PTT-HEMO		37	30.325	0.850	2.80	25.800	34.900	
2020-942	73 ACL 8 TO E	56 PTT-HEMO		37	27.908	0.772	2.76	23.700	32.100	
2020-942	76 STAGO STAR	49 STA-PTT		20	32.265	1.366	4.23	27.400	37.100	
2020-943	33 MLA ELECT.	54 QUICK COAG		10	28.760	1.146	3.98	24.400	33.100	
2020-943	58 AMELUNG KC	42 BIOMERIEUX		10	29.140	1.806	6.20	24.800	33.500	
2020-943	60 DADE SYSME	35 DADE ACTIN		13	27.300	1.575	5.77	23.200	31.400	
2020-943	62 DADE SYS.	35 DADE ACTIN		59	26.354	1.115	4.23	22.400	30.300	
2020-943	65 DADE BFTII	35 DADE ACTIN		13	25.338	2.160	8.52	21.500	29.100	
2020-943	69 DIAG.STAGO	49 STA-PTT		12	30.983	1.041	3.36	26.300	35.600	
2020-943	73 ACL 8 TO E	56 PTT-HEMO		37	30.059	0.915	3.04	25.600	34.600	
2020-943	73 ACL 8 TO E	56 PTT-HEMO		37	31.950	0.427	1.34	27.200	36.700	
2020-943	76 STAGO STAR	49 STA-PTT		20	29.195	1.436	4.92	24.800	33.600	
2020-944	33 MLA ELECT.	54 QUICK COAG		10	99.470	3.550	3.57	84.500	114.400	
2020-944	58 AMELUNG KC	42 BIOMERIEUX		10	88.520	7.229	8.17	75.200	101.800	
2020-944	60 DADE SYSME	35 DADE ACTIN		13	86.469	2.259	2.61	73.500	99.400	
2020-944	62 DADE SYS.	35 DADE ACTIN		59	81.244	5.158	6.35	69.100	93.400	
2020-944	65 DADE BFTII	35 DADE ACTIN		13	96.185	12.017	12.49	81.800	110.600	
2020-944	69 DIAG.STAGO	49 STA-PTT		12	83.125	3.200	3.85	70.700	95.600	
2020-944	73 ACL 8 TO E	56 PTT-HEMO		37	97.675	2.625	2.69	83.000	112.300	
2020-944	73 ACL 8 TO E	56 PTT-HEMO		37	99.803	3.230	3.24	84.800	114.800	
2020-944	76 STAGO STAR	49 STA-PTT		20	83.480	9.198	11.02	71.000	96.000	

### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE	15	%				
2020-945	33 MLA ELECT.	54 QUICK COAG		10	42.460	2.489	5.86	36.100	48.800
2020-945	58 AMELUNG KC	42 BIOMERIEUX		10	42.610	4.270	10.02	36.200	49.000
2020-945	60 DADE SYSME	35 DADE ACTIN		13	34.669	0.925	2.67	29.500	39.900
2020-945	62 DADE SYS.	35 DADE ACTIN		59	32.358	1.808	5.59	27.500	37.200
2020-945	65 DADE BFTII	35 DADE ACTIN		13	35.685	4.203	11.78	30.300	41.000
2020-945	69 DIAG.STAGO	49 STA-PTT		12	46.958	2.183	4.65	39.900	54.000
2020-945	73 ACL 8 TO E	56 PTT-HEMO		37	36.125	0.798	2.21	30.700	41.500
2020-945	73 ACL 8 TO E	56 PTT-HEMO		37	36.516	1.045	2.86	31.000	42.000
2020-945	76 STAGO STAR	49 STA-PTT		20	43.045	3.304	7.68	36.600	49.500
082	PROTHROMBIN TIME (PT)	CLIA TEST ID #: 845							
2020-941	33 MLA ELECT.	37 QUICK COAG		10	26.890	2.875	10.69	22.900	30.900
2020-941	58 AMELUNG KC	15 BIOMERIEUX		12	23.867	0.976	4.09	20.300	27.400
2020-941	58 AMELUNG KC	15 BIOMERIEUX		12	19.367	1.502	7.76	16.500	22.300
2020-941	60 DADE SYSME	33 DADE INNOV		14	24.371	0.476	1.95	20.700	28.000
2020-941	62 DADE SYS.	33 DADE INNOV		60	24.202	0.933	3.85	20.600	27.800
2020-941	65 DADE BFTII	33 DADE INNOV		14	24.679	1.477	5.99	21.000	28.400
2020-941	73 ACL 8 TO E	41 HEMOSIL RE		32	30.763	2.013	6.54	26.100	35.400
2020-941	76 STAGO STAR	44 STA NEO CI		14	33.793	1.809	5.35	28.700	38.900
2020-942	33 MLA ELECT.	37 QUICK COAG		10	12.160	0.794	6.53	10.300	14.000
2020-942	58 AMELUNG KC	15 BIOMERIEUX		12	10.792	0.787	7.29	9.200	12.400
2020-942	58 AMELUNG KC	15 BIOMERIEUX		12	9.617	0.701	7.29	8.200	11.100
2020-942	60 DADE SYSME	33 DADE INNOV		14	10.323	0.212	2.05	8.800	11.900
2020-942	62 DADE SYS.	33 DADE INNOV		60	9.805	0.256	2.61	8.300	11.300
2020-942	65 DADE BFTII	33 DADE INNOV		14	8.993	0.397	4.42	7.600	10.300
2020-942	73 ACL 8 TO E	41 HEMOSIL RE		32	11.069	0.326	2.95	9.400	12.700
2020-942	76 STAGO STAR	44 STA NEO CI		14	12.271	0.459	3.74	10.400	14.100
2020-943	33 MLA ELECT.	37 QUICK COAG		10	12.330	1.041	8.44	10.500	14.200

### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-943	58	AMELUNG KC	15 BIOMERIEUX	12	10.283	0.631	6.14	8.700	11.800	
2020-943	58	AMELUNG KC	15 BIOMERIEUX	12	10.508	0.733	6.98	8.900	12.100	
2020-943	60	DADE SYSME	33 DADE INNOV	14	10.823	0.105	0.97	9.200	12.400	
2020-943	62	DADE SYS.	33 DADE INNOV	60	10.390	0.352	3.39	8.800	11.900	
2020-943	65	DADE BFTII	33 DADE INNOV	14	9.564	0.364	3.80	8.100	11.000	
2020-943	73	ACL 8 TO E	41 HEMOSIL RE	32	11.710	0.411	3.51	10.000	13.500	
2020-943	76	STAGO STAR	44 STA NEO CI	14	12.500	0.457	3.65	10.600	14.400	
082 PROTHROMBIN TIME(PT)					CLIA TEST ID #: 845 EVALUATION CRITERIA: TARGET VALUE 15 %					
2020-944	33	MLA ELECT.	37 QUICK COAG	10	46.370	8.754	18.88	39.400	53.300	
2020-944	58	AMELUNG KC	15 BIOMERIEUX	12	30.558	3.023	9.89	26.000	35.100	
2020-944	58	AMELUNG KC	15 BIOMERIEUX	12	43.783	2.701	6.17	37.200	50.400	
2020-944	60	DADE SYSME	33 DADE INNOV	14	44.586	1.000	2.24	37.900	51.300	
2020-944	62	DADE SYS.	33 DADE INNOV	60	43.583	2.135	4.90	37.000	50.100	
2020-944	65	DADE BFTII	33 DADE INNOV	14	45.943	4.044	8.80	39.100	52.800	
2020-944	73	ACL 8 TO E	41 HEMOSIL RE	32	58.516	3.358	5.74	49.700	67.300	
2020-944	76	STAGO STAR	44 STA NEO CI	14	66.650	4.267	6.40	56.700	76.600	
2020-945	33	MLA ELECT.	37 QUICK COAG	10	16.380	1.311	8.01	13.900	18.800	
2020-945	58	AMELUNG KC	15 BIOMERIEUX	12	14.000	0.443	3.17	11.900	16.100	
2020-945	58	AMELUNG KC	15 BIOMERIEUX	12	14.375	1.068	7.43	12.200	16.500	
2020-945	60	DADE SYSME	33 DADE INNOV	14	15.086	0.304	2.02	12.800	17.300	
2020-945	62	DADE SYS.	33 DADE INNOV	60	14.695	0.587	3.99	12.500	16.900	
2020-945	65	DADE BFTII	33 DADE INNOV	14	15.157	1.102	7.27	12.900	17.400	
2020-945	73	ACL 8 TO E	41 HEMOSIL RE	32	16.638	0.706	4.24	14.100	19.100	
2020-945	76	STAGO STAR	44 STA NEO CI	14	17.943	0.715	3.98	15.300	20.600	

SPE: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 100 HEMA. CELL IDENTIFI.

## Participant Summary

MONTH 11 OF YEAR 2020

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2020-891					030 030		052 - SPHEROCYTE	862	2	731	0
2020-891					030 030		030 - ERYTHROCYT	862	731	731	100
2020-891					030 030		044 - ERYTHROCYT	862	21	731	0
2020-891					030 030		048 - ERYTHROCYT	862	107	731	0
2020-891					030 030		032 - NUCLEATED	862	1	731	0
2020-892					024 024		025 - DOHLE INCL	861	3	856	0
2020-892					024 024		038 - HOWELL-JOL	861	1	856	0
2020-892					024 024		001 - BLAST	861	1	856	0
2020-892					024 024		024 - AUER ROD	861	856	856	100
2020-893					017 017		020 - GIANT PLA	861	2	853	0
2020-893					017 017		014 - LYMPHOCYTE	861	3	853	0
2020-893					017 017		001 - BLAST	861	1	853	0
2020-893					017 017		030 - ERYTHROCYT	861	1	853	0
2020-893					017 017		017 - MONOCYTE	861	853	853	100
2020-893					017 017		029 - TOXIC GRAN	861	1	853	0
2020-894					008 008		007 - BASOPHIL,	861	2	855	0
2020-894					008 008		008 - EOSINOPHIL	861	855	855	100
2020-894					008 008		044 - ERYTHROCYT	861	1	855	0
2020-894					008 008		017 - MONOCYTE	861	1	855	0
2020-894					008 008		057 - CHEDIAK -	861	1	855	0
2020-894					008 008		006 - SEGMENTED	861	1	855	0
2020-895					057 057		048 - ERYTHROCYT	861	1	839	0
2020-895					057 057		057 - CHEDIAK -	861	839	839	100

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-895					057	057	023 - ALDER ANOM	861	6	839	0
2020-895					057	057	044 - ERYTHROCYT	861	1	839	0
2020-895					057	057	008 - EOSINOPHIL	861	1	839	0
2020-895					057	057	012 - MAST CELL,	861	1	839	0
2020-895					057	057	029 - TOXIC GRAN	861	7	839	0
2020-895					057	057	027 - PSEUDO-PEL	861	4	839	0
2020-895					057	057	071 - Poikilo.	861	1	839	0

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 137 RETICULOCYTE COUNT						
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}								
2020-921	7	M.B.MANUAL		18	11.550	3.745	32.43	0.300	22.800		
2020-921	9	M.B. RICCA		31	10.542	1.849	17.54	5.000	16.100		
2020-921	24	Unicel DxH		12	8.050	0.517	6.42	6.000	10.100		
2020-921	25	SYSMEX XN		25	6.024	0.762	12.64	3.700	8.300		
2020-922	7	M.B.MANUAL		18	1.700	0.412	24.25	0.500	2.900		
2020-922	9	M.B. RICCA		31	1.670	0.442	26.47	0.300	3.000		
2020-922	24	Unicel DxH		12	1.008	0.138	13.71	0.600	1.400		
2020-922	25	SYSMEX XN		25	1.124	0.103	9.17	0.800	1.400		
2020-923	7	M.B.MANUAL		18	11.272	3.438	30.50	1.000	21.600		
2020-923	9	M.B. RICCA		31	10.723	1.984	18.51	4.800	16.700		
2020-923	24	Unicel DxH		12	7.892	0.441	5.58	5.900	9.900		
2020-923	25	SYSMEX XN		25	6.016	0.826	13.74	3.500	8.500		

### Participant Summary

#### MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
2020-924	7	M.B.MANUAL		18	5.644	1.788	31.68	0.300	11.000	
2020-924	9	M.B. RICCA		31	5.548	1.324	23.86	1.600	9.500	
2020-924	24	Unicel DxH		12	4.067	0.425	10.45	2.800	5.300	
2020-924	25	SYSMEX XN		25	3.088	0.358	11.60	2.000	4.200	
2020-925	7	M.B.MANUAL		18	5.400	1.633	30.25	0.500	10.300	
2020-925	9	M.B. RICCA		31	5.687	1.219	21.44	2.000	9.300	
2020-925	24	Unicel DxH		12	4.042	0.352	8.72	3.000	5.100	
2020-925	25	SYSMEX XN		25	3.092	0.383	12.37	1.900	4.200	
SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
153	SED. RATE (ESR)									
2020-931	4	SEDIPLAST		557	50.538	8.674	17.16	25.000	77.000	
2020-931	10	SEDI RATE		120	50.839	9.826	19.33	21.000	80.000	
2020-931	11	SEDIMAT-15		71	135.014	9.589	7.10	101.000	169.000	
2020-931	14	GLOBAL SC.		12	54.083	7.751	14.33	31.000	77.000	
2020-931	19	MINI CUBE		17	71.588	27.444	38.34	0	154.000	
2020-932	4	SEDIPLAST		557	50.038	8.042	16.07	26.000	74.000	
2020-932	10	SEDI RATE		120	49.720	9.421	18.95	21.000	78.000	
2020-932	11	SEDIMAT-15		71	134.118	11.900	8.87	98.000	170.000	
2020-932	14	GLOBAL SC.		12	51.417	10.484	20.39	20.000	83.000	
2020-932	19	MINI CUBE		17	74.412	30.704	41.26	0	167.000	
2020-933	4	SEDIPLAST		557	9.548	2.332	24.43	3.000	17.000	
2020-933	10	SEDI RATE		120	11.863	2.663	22.45	4.000	20.000	



**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY			SUB: 70 GENERAL			GRP: 138 SED. RATE (ESR)					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-933	11	SEDIMAT-15		71	15.451	12.954	83.84	0	54.000		
2020-933	14	GLOBAL SC.		12	13.750	2.618	19.04	6.000	22.000		
2020-933	19	MINI CUBE		17	19.235	4.166	21.66	7.000	32.000		
153 SED. RATE (ESR)		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}							
2020-934	4	SEDIPLAST		557	50.258	8.282	16.48	25.000	75.000		
2020-934	10	SEDI RATE		120	48.496	9.393	19.37	20.000	77.000		
2020-934	11	SEDIMAT-15		71	130.104	19.671	15.12	71.000	189.000		
2020-934	14	GLOBAL SC.		12	55.750	13.033	23.38	17.000	95.000		
2020-934	19	MINI CUBE		17	73.000	25.370	34.75	0	149.000		
2020-935	4	SEDIPLAST		557	9.867	2.544	25.78	2.000	17.000		
2020-935	10	SEDI RATE		120	11.855	2.254	19.02	5.000	19.000		
2020-935	11	SEDIMAT-15		71	14.718	11.395	77.42	0	49.000		
2020-935	14	GLOBAL SC.		12	12.417	3.378	27.20	2.000	23.000		
2020-935	19	MINI CUBE		17	19.412	4.045	20.84	7.000	32.000		

SPEC: 20 HEMATOLOGY			SUB: 70 GENERAL			GRP: 125 DIFF. FOR 5 PAR NEW					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2020911X	1	SYSMEX CBC	1 NEUTROPHIL	20	57.435	1.011	1.76	43.100	71.800		
2020911X	1	SYSMEX CBC	1 NEUTROPHIL	25	47.635	1.945	4.08	35.700	59.500		
2020911X	1	SYSMEX CBC	2 LYMPHOCYTE	20	26.582	1.263	4.75	19.900	33.200		
2020911X	1	SYSMEX CBC	2 LYMPHOCYTE	25	30.721	2.475	8.06	23.000	38.400		
2020911X	1	SYSMEX CBC	3 MONOCYTE	20	2.559	0.240	9.39	0.600	4.600		
2020911X	1	SYSMEX CBC	3 MONOCYTE	25	1.722	0.437	25.39	0	3.700		
2020911X	1	SYSMEX CBC	4 EOSINOPHIL	20	13.424	0.449	3.35	10.100	16.800		
2020911X	1	SYSMEX CBC	4 EOSINOPHIL	25	12.261	0.790	6.44	9.200	15.300		
2020911X	1	SYSMEX CBC	5 BASOPHIL	20	72.771	0.968	1.33	54.600	91.000		

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020911X	1	SYSMEX CBC	5 BASOPHIL	25 XS SERIES	152	7.689	0.825	10.73	5.700	9.700	
2020911X	2	ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	160	53.356	3.254	6.10	40.000	66.700	
2020911X	2	ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	24	50.913	2.584	5.08	38.200	63.600	
2020911X	2	ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	154	33.078	2.053	6.21	24.800	41.300	
2020911X	2	ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	24	31.300	2.170	6.93	23.500	39.100	
2020911X	2	ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	155	8.141	2.184	26.83	6.100	10.200	
2020911X	2	ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	24	9.096	1.920	21.10	6.800	11.400	
2020911X	2	ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	155	5.222	0.486	9.31	3.200	7.200	
2020911X	2	ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	24	5.004	0.347	6.93	3.000	7.000	
2020911X	2	ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	155	71.265	2.381	3.34	53.400	89.100	
2020911X	2	ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	24	3.508	0.171	4.86	1.500	5.500	
2020911X	3	ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	18	62.928	0.969	1.54	47.200	78.700	
2020911X	3	ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	18	23.753	0.783	3.30	17.800	29.700	
2020911X	3	ADVIA 120	3 MONOCYTE	9 ADVIA-120	18	8.706	0.650	7.47	6.500	10.900	
2020911X	3	ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	18	2.072	0.352	17.01	0.100	4.100	
2020911X	3	ADVIA 120	5 BASOPHIL	9 ADVIA-120	18	0.572	0.179	31.26	0	2.600	
2020911X	3	ADVIA 120	6 LUC	9 ADVIA-120	17	2.553	0.519	20.34	0.600	4.600	
2020911X	4	AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	133	63.947	0.904	1.41	48.000	79.900	
2020911X	4	AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	128	24.291	1.467	6.04	18.200	30.400	
2020911X	4	AB- CELL-D	3 MONOCYTE	8 CELL-DYN	128	5.387	0.769	14.28	3.400	7.400	
2020911X	4	AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	126	5.994	0.326	5.43	4.000	8.000	
2020911X	4	AB- CELL-D	5 BASOPHIL	8 CELL-DYN	130	0.484	0.407	84.11	0	2.500	
2020911X	5	BECTON D	1 NEUTROPHIL	29 B.C600,800	49	60.541	0.798	1.32	45.400	75.700	
2020911X	5	BECTON D	2 LYMPHOCYTE	29 B.C600,800	49	28.263	0.721	2.55	21.200	35.300	
2020911X	5	BECTON D	3 MONOCYTE	29 B.C600,800	49	5.118	0.517	10.09	3.100	7.100	
2020911X	5	BECTON D	4 EOSINOPHIL	29 B.C600,800	49	6.041	0.310	5.14	4.000	8.000	
2020911X	5	BECTON D	5 BASOPHIL	29 B.C600,800	49	0.096	0.040	41.91	0	2.100	
2020911X	6	BC Mindray	1 NEUTROPHIL	32 BC-5390	28	54.170	0.796	1.47	40.600	67.700	
2020911X	6	BC Mindray	1 NEUTROPHIL	36 BC DXH 500	18	78.247	1.651	2.11	58.700	97.800	
2020911X	6	BC Mindray	2 LYMPHOCYTE	32 BC-5390	28	28.341	1.275	4.50	21.300	35.400	
2020911X	6	BC Mindray	2 LYMPHOCYTE	36 BC DXH 500	18	10.194	3.067	30.09	7.600	12.700	

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020911X	6 BC Mindray	3 MONOCYTE	32 BC-5390	28	7.864	0.998	12.68	5.900	9.900		
2020911X	6 BC Mindray	3 MONOCYTE	36 BC DXH 500	18	0.028	0.045	161.25	0	2.000		
2020911X	6 BC Mindray	4 EOSINOPHIL	32 BC-5390	28	9.550	0.399	4.17	7.200	11.900		
2020911X	6 BC Mindray	4 EOSINOPHIL	36 BC DXH 500	18	10.994	1.686	15.34	8.200	13.700		
2020911X	6 BC Mindray	5 BASOPHIL	32 BC-5390	28	28.629	0.233	0.81	21.500	35.800		
2020911X	6 BC Mindray	5 BASOPHIL	36 BC DXH 500	18	0.011	0.031	282.84	0	2.000		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020912X	1 SYSMEX CBC	1 NEUTROPHIL	20 SYSMEX XT-	17	67.518	1.477	2.19	50.600	84.400		
2020912X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	152	59.568	1.441	2.42	44.700	74.500		
2020912X	1 SYSMEX CBC	2 LYMPHOCYTE	20 SYSMEX XT-	17	15.794	1.646	10.42	11.800	19.700		
2020912X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	152	14.595	1.133	7.76	10.900	18.200		
2020912X	1 SYSMEX CBC	3 MONOCYTE	20 SYSMEX XT-	17	1.369	0.247	18.03	0	3.400		
2020912X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	152	0.886	0.294	33.25	0	2.900		
2020912X	1 SYSMEX CBC	4 EOSINOPHIL	20 SYSMEX XT-	17	15.224	0.380	2.49	11.400	19.000		
2020912X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	152	15.380	0.578	3.76	11.500	19.200		
2020912X	1 SYSMEX CBC	5 BASOPHIL	20 SYSMEX XT-	17	84.571	1.281	1.51	63.400	105.700		
2020912X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	152	9.421	0.922	9.79	7.100	11.800		
2020912X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	155	53.704	3.230	6.01	40.300	67.100		
2020912X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	24	50.821	2.845	5.60	38.100	63.500		
2020912X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	160	32.857	2.196	6.68	24.600	41.100		
2020912X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	24	30.933	2.258	7.30	23.200	38.700		
2020912X	2 ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	154	8.127	2.009	24.72	6.100	10.200		
2020912X	2 ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	24	9.375	2.159	23.03	7.000	11.700		
2020912X	2 ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	156	5.259	0.499	9.48	3.300	7.300		
2020912X	2 ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	24	5.267	0.539	10.23	3.300	7.300		
2020912X	2 ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	155	71.381	2.535	3.55	53.500	89.200		
2020912X	2 ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	24	3.483	0.175	5.02	1.500	5.500		
2020912X	3 ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	18	62.756	0.951	1.52	47.100	78.400		
2020912X	3 ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	18	24.024	0.597	2.48	18.000	30.000		
2020912X	3 ADVIA 120	3 MONOCYTE	9 ADVIA-120	18	8.518	0.457	5.36	6.400	10.600		
2020912X	3 ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	18	2.083	0.342	16.41	0.100	4.100		

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020912X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	18	0.572	0.137	23.88	0	2.600
2020912X	3	ADVIA 120	6 LUC	9	ADVIA-120	17	2.541	0.428	16.86	0.500	4.500
2020912X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	128	46.401	1.872	4.03	34.800	58.000
2020912X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	133	40.834	2.531	6.20	30.600	51.000
2020912X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	128	6.486	1.549	23.88	4.500	8.500
2020912X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	128	5.620	0.480	8.53	3.600	7.600
2020912X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	128	0.735	0.585	79.50	0	2.700
2020912X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	49	52.006	1.022	1.96	39.000	65.000
2020912X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	49	36.010	1.591	4.42	27.000	45.000
2020912X	5	BECTON D	3 MONOCYTE	29	B.C600,800	49	3.943	0.937	23.77	1.900	5.900
2020912X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	49	7.935	0.377	4.75	5.900	9.900
2020912X	5	BECTON D	5 BASOPHIL	29	B.C600,800	49	0.106	0.055	51.82	0	2.100
2020912X	6	BC Mindray	1 NEUTROPHIL	32	BC-5390	28	48.896	1.104	2.26	36.700	61.100
2020912X	6	BC Mindray	1 NEUTROPHIL	36	BC DXH 500	18	75.167	2.218	2.95	56.400	94.000
2020912X	6	BC Mindray	2 LYMPHOCYTE	32	BC-5390	28	36.556	1.709	4.68	27.400	45.700
2020912X	6	BC Mindray	2 LYMPHOCYTE	36	BC DXH 500	18	13.389	2.455	18.34	10.000	16.700
2020912X	6	BC Mindray	3 MONOCYTE	32	BC-5390	28	7.807	0.853	10.93	5.800	9.800
2020912X	6	BC Mindray	3 MONOCYTE	36	BC DXH 500	18	0.056	0.050	89.44	0	2.100
2020912X	6	BC Mindray	4 EOSINOPHIL	32	BC-5390	28	6.581	0.646	9.82	4.600	8.600
2020912X	6	BC Mindray	4 EOSINOPHIL	36	BC DXH 500	18	11.450	1.392	12.16	8.600	14.300
2020912X	6	BC Mindray	5 BASOPHIL	32	BC-5390	28	26.832	0.317	1.18	20.100	33.500
2020912X	6	BC Mindray	5 BASOPHIL	36	BC DXH 500	18	0.017	0.037	223.61	0	2.000
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE	25	% OR 2.00	{GREATER}			
2020913X	1	SYSMEX CBC	1 NEUTROPHIL	20	SYSMEX XT-	17	61.476	1.626	2.65	46.100	76.800
2020913X	1	SYSMEX CBC	1 NEUTROPHIL	25	XS SERIES	152	55.283	1.460	2.64	41.500	69.100
2020913X	1	SYSMEX CBC	2 LYMPHOCYTE	20	SYSMEX XT-	17	23.065	2.099	9.10	17.300	28.800
2020913X	1	SYSMEX CBC	2 LYMPHOCYTE	25	XS SERIES	152	20.552	0.788	3.83	15.400	25.700
2020913X	1	SYSMEX CBC	3 MONOCYTE	20	SYSMEX XT-	17	1.912	0.532	27.85	0	3.900
2020913X	1	SYSMEX CBC	3 MONOCYTE	25	XS SERIES	152	0.980	0.276	28.18	0	3.000
2020913X	1	SYSMEX CBC	4 EOSINOPHIL	20	SYSMEX XT-	17	13.547	0.702	5.18	10.200	16.900
2020913X	1	SYSMEX CBC	4 EOSINOPHIL	25	XS SERIES	152	14.001	0.696	4.97	10.500	17.500

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020913X	1	SYSMEX CBC	5 BASOPHIL	20	SYSMEX XT-	17	83.156	0.544	0.65	62.400	103.900
2020913X	1	SYSMEX CBC	5 BASOPHIL	25	XS SERIES	152	9.201	0.763	8.29	6.900	11.500
2020913X	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	155	50.716	2.594	5.11	38.000	63.400
2020913X	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	24	48.800	2.457	5.03	36.600	61.000
2020913X	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	155	34.577	2.100	6.07	25.900	43.200
2020913X	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	24	32.783	2.683	8.18	24.600	41.000
2020913X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	160	6.709	1.807	26.94	4.700	8.700
2020913X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	24	6.963	1.687	24.24	5.000	9.000
2020913X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	155	8.018	0.828	10.33	6.000	10.000
2020913X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	24	7.913	0.680	8.60	5.900	9.900
2020913X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	155	62.761	1.532	2.44	47.100	78.500
2020913X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	24	3.100	0.083	2.69	1.100	5.100
2020913X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	18	45.212	1.271	2.81	33.900	56.500
2020913X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	18	34.582	1.750	5.06	25.900	43.200
2020913X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	18	14.567	1.404	9.64	10.900	18.200
2020913X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	18	1.394	0.409	29.32	0	3.400
2020913X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	18	0.539	0.198	36.67	0	2.500
2020913X	3	ADVIA 120	6 LUC	9	ADVIA-120	17	3.676	0.695	18.90	1.700	5.700
2020913X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	128	73.430	0.698	0.95	55.100	91.800
2020913X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	128	13.667	1.157	8.46	10.300	17.100
2020913X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	133	3.460	0.623	18.01	1.500	5.500
2020913X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	127	9.164	0.336	3.67	6.900	11.500
2020913X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	129	0.424	0.282	66.59	0	2.400
2020913X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	49	65.829	0.804	1.22	49.400	82.300
2020913X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	49	21.765	1.081	4.97	16.300	27.200
2020913X	5	BECTON D	3 MONOCYTE	29	B.C600,800	49	8.912	1.173	13.16	6.700	11.100
2020913X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	49	3.437	0.251	7.31	1.400	5.400
2020913X	5	BECTON D	5 BASOPHIL	29	B.C600,800	49	0.073	0.049	66.09	0	2.100
2020913X	6	BC Mindray	1 NEUTROPHIL	32	BC-5390	28	63.170	0.898	1.42	47.400	79.000
2020913X	6	BC Mindray	1 NEUTROPHIL	36	BC DXH 500	18	79.911	2.431	3.04	59.900	99.900
2020913X	6	BC Mindray	2 LYMPHOCYTE	32	BC-5390	28	17.237	1.245	7.22	12.900	21.500

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020913X	6 BC Mindray	2	LYMPHOCYTE	36	BC DXH 500	18	5.278	2.438	46.19	3.300	7.300	
2020913X	6 BC Mindray	3	MONOCYTE	32	BC-5390	28	7.729	0.820	10.61	5.700	9.700	
2020913X	6 BC Mindray	3	MONOCYTE	36	BC DXH 500	18	0.024	0.042	180.28	0	2.000	
2020913X	6 BC Mindray	4	EOSINOPHIL	32	BC-5390	28	11.881	0.583	4.91	8.900	14.900	
2020913X	6 BC Mindray	4	EOSINOPHIL	36	BC DXH 500	18	14.435	1.203	8.33	10.800	18.000	
2020913X	6 BC Mindray	5	BASOPHIL	32	BC-5390	28	31.486	0.208	0.66	23.600	39.400	
2020913X	6 BC Mindray	5	BASOPHIL	36	BC DXH 500	18	0.006	0.024	400.00	0	2.000	
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020914X	1 SYSMEX CBC	1	NEUTROPHIL	20	SYSMEX XT-	17	57.812	1.579	2.73	43.400	72.300	
2020914X	1 SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	152	48.222	2.302	4.77	36.200	60.300	
2020914X	1 SYSMEX CBC	2	LYMPHOCYTE	20	SYSMEX XT-	17	26.124	1.490	5.70	19.600	32.700	
2020914X	1 SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	152	29.938	3.003	10.03	22.500	37.400	
2020914X	1 SYSMEX CBC	3	MONOCYTE	20	SYSMEX XT-	17	2.671	0.536	20.05	0.700	4.700	
2020914X	1 SYSMEX CBC	3	MONOCYTE	25	XS SERIES	152	1.683	0.410	24.36	0	3.700	
2020914X	1 SYSMEX CBC	4	EOSINOPHIL	20	SYSMEX XT-	17	13.394	0.326	2.44	10.000	16.700	
2020914X	1 SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	152	12.457	0.768	6.16	9.300	15.600	
2020914X	1 SYSMEX CBC	5	BASOPHIL	20	SYSMEX XT-	17	72.453	0.755	1.04	54.300	90.600	
2020914X	1 SYSMEX CBC	5	BASOPHIL	25	XS SERIES	152	7.648	0.767	10.03	5.600	9.600	
2020914X	2 ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	155	70.571	2.452	3.48	52.900	88.200	
2020914X	2 ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	24	65.979	2.042	3.09	49.500	82.500	
2020914X	2 ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	153	16.638	1.257	7.56	12.500	20.800	
2020914X	2 ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	24	15.283	1.381	9.03	11.500	19.100	
2020914X	2 ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	155	7.575	2.254	29.76	5.600	9.600	
2020914X	2 ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	24	9.254	2.043	22.08	6.900	11.600	
2020914X	2 ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	159	5.016	0.450	8.97	3.000	7.000	
2020914X	2 ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	25	5.036	0.453	9.00	3.000	7.000	
2020914X	2 ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	156	89.029	2.453	2.76	66.800	111.300	
2020914X	2 ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	23	4.352	0.125	2.86	2.400	6.400	
2020914X	3 ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	19	72.339	0.656	0.91	54.300	90.400	
2020914X	3 ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	18	14.878	0.670	4.50	11.200	18.600	
2020914X	3 ADVIA 120	3	MONOCYTE	9	ADVIA-120	18	8.578	0.502	5.85	6.400	10.700	

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020914X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	18	2.156	0.322	14.93	0.200	4.200
2020914X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	18	0.611	0.145	23.71	0	2.600
2020914X	3	ADVIA 120	6 LUC	9	ADVIA-120	17	1.871	0.337	18.04	0	3.900
2020914X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	127	64.024	0.800	1.25	48.000	80.000
2020914X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	128	24.304	1.351	5.56	18.200	30.400
2020914X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	128	5.317	0.669	12.58	3.300	7.300
2020914X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	133	5.939	0.320	5.38	3.900	7.900
2020914X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	128	0.482	0.417	86.49	0	2.500
2020914X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	48	60.646	0.703	1.16	45.500	75.800
2020914X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	49	28.229	0.782	2.77	21.200	35.300
2020914X	5	BECTON D	3 MONOCYTE	29	B.C600,800	49	4.969	0.531	10.69	3.000	7.000
2020914X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	49	6.049	0.298	4.92	4.000	8.000
2020914X	5	BECTON D	5 BASOPHIL	29	B.C600,800	49	0.098	0.038	38.53	0	2.100
2020914X	6	BC Mindray	1 NEUTROPHIL	32	BC-5390	28	54.574	0.979	1.79	40.900	68.200
2020914X	6	BC Mindray	1 NEUTROPHIL	36	BC DXH 500	18	78.406	2.227	2.84	58.800	98.000
2020914X	6	BC Mindray	2 LYMPHOCYTE	32	BC-5390	28	27.593	1.907	6.91	20.700	34.500
2020914X	6	BC Mindray	2 LYMPHOCYTE	36	BC DXH 500	18	10.300	2.220	21.55	7.700	12.900
2020914X	6	BC Mindray	3 MONOCYTE	32	BC-5390	28	8.136	1.236	15.19	6.100	10.200
2020914X	6	BC Mindray	3 MONOCYTE	36	BC DXH 500	18	0.028	0.045	161.25	0	2.000
2020914X	6	BC Mindray	4 EOSINOPHIL	32	BC-5390	28	9.361	0.602	6.43	7.000	11.700
2020914X	6	BC Mindray	4 EOSINOPHIL	36	BC DXH 500	18	11.267	1.506	13.37	8.500	14.100
2020914X	6	BC Mindray	5 BASOPHIL	32	BC-5390	28	28.675	0.241	0.84	21.500	35.800
2020914X	6	BC Mindray	5 BASOPHIL	36	BC DXH 500	18	0.011	0.031	282.84	0	2.000
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2020915X	1	SYSMEX CBC	1 NEUTROPHIL	20	SYSMEX XT-	18	67.159	1.586	2.36	50.400	83.900
2020915X	1	SYSMEX CBC	1 NEUTROPHIL	25	XS SERIES	152	59.481	1.165	1.96	44.600	74.400
2020915X	1	SYSMEX CBC	2 LYMPHOCYTE	20	SYSMEX XT-	16	16.456	2.003	12.17	12.300	20.600
2020915X	1	SYSMEX CBC	2 LYMPHOCYTE	25	XS SERIES	152	14.742	1.039	7.05	11.100	18.400
2020915X	1	SYSMEX CBC	3 MONOCYTE	20	SYSMEX XT-	17	1.435	0.274	19.11	0	3.400
2020915X	1	SYSMEX CBC	3 MONOCYTE	25	XS SERIES	152	0.890	0.276	30.98	0	2.900
2020915X	1	SYSMEX CBC	4 EOSINOPHIL	20	SYSMEX XT-	17	15.106	0.609	4.03	11.300	18.900

### Participant Summary

#### MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020915X	1	SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	150	15.416	0.623	4.04	11.600	19.300	
2020915X	1	SYSMEX CBC	5 BASOPHIL	20 SYSMEX XT-	17	84.519	1.114	1.32	63.400	105.600	
2020915X	1	SYSMEX CBC	5 BASOPHIL	25 XS SERIES	153	9.355	0.959	10.26	7.000	11.700	
2020915X	2	ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	156	51.502	2.581	5.01	38.600	64.400	
2020915X	2	ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	24	48.971	2.261	4.62	36.700	61.200	
2020915X	2	ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	154	33.872	2.190	6.47	25.400	42.300	
2020915X	2	ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	24	32.238	2.528	7.84	24.200	40.300	
2020915X	2	ACT PENTRA	3 MONOCYTE	15 COUL ACT 5	154	6.552	1.633	24.92	4.600	8.600	
2020915X	2	ACT PENTRA	3 MONOCYTE	24 ABX PENTRA	24	7.221	1.854	25.67	5.200	9.200	
2020915X	2	ACT PENTRA	4 EOSINOPHIL	15 COUL ACT 5	154	7.913	0.862	10.90	5.900	9.900	
2020915X	2	ACT PENTRA	4 EOSINOPHIL	24 ABX PENTRA	25	8.338	1.696	20.34	6.300	10.400	
2020915X	2	ACT PENTRA	5 BASOPHIL	15 COUL ACT 5	161	62.828	1.758	2.80	47.100	78.500	
2020915X	2	ACT PENTRA	5 BASOPHIL	24 ABX PENTRA	24	3.078	0.098	3.17	1.100	5.100	
2020915X	3	ADVIA 120	1 NEUTROPHIL	9 ADVIA-120	18	45.706	1.090	2.39	34.300	57.100	
2020915X	3	ADVIA 120	2 LYMPHOCYTE	9 ADVIA-120	18	34.483	1.285	3.73	25.900	43.100	
2020915X	3	ADVIA 120	3 MONOCYTE	9 ADVIA-120	18	14.522	0.816	5.62	10.900	18.200	
2020915X	3	ADVIA 120	4 EOSINOPHIL	9 ADVIA-120	19	1.483	0.350	23.60	0	3.500	
2020915X	3	ADVIA 120	5 BASOPHIL	9 ADVIA-120	18	0.483	0.171	35.33	0	2.500	
2020915X	3	ADVIA 120	6 LUC	9 ADVIA-120	17	3.594	0.654	18.19	1.600	5.600	
2020915X	4	AB- CELL-D	1 NEUTROPHIL	8 CELL-DYN	128	46.417	2.017	4.35	34.800	58.000	
2020915X	4	AB- CELL-D	2 LYMPHOCYTE	8 CELL-DYN	128	41.013	2.471	6.02	30.800	51.300	
2020915X	4	AB- CELL-D	3 MONOCYTE	8 CELL-DYN	128	6.302	1.314	20.85	4.300	8.300	
2020915X	4	AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	127	5.631	0.478	8.49	3.600	7.600	
2020915X	4	AB- CELL-D	5 BASOPHIL	8 CELL-DYN	133	0.710	0.556	78.32	0	2.700	
2020915X	5	BECTON D	1 NEUTROPHIL	29 B.C600,800	49	52.161	0.782	1.50	39.100	65.200	
2020915X	5	BECTON D	2 LYMPHOCYTE	29 B.C600,800	49	35.871	1.454	4.05	26.900	44.800	
2020915X	5	BECTON D	3 MONOCYTE	29 B.C600,800	49	3.878	0.988	25.47	1.900	5.900	
2020915X	5	BECTON D	4 EOSINOPHIL	29 B.C600,800	49	7.876	0.315	4.00	5.900	9.900	
2020915X	5	BECTON D	5 BASOPHIL	29 B.C600,800	49	0.112	0.056	49.73	0	2.100	
2020915X	6	BC Mindray	1 NEUTROPHIL	32 BC-5390	28	48.811	1.033	2.12	36.600	61.000	
2020915X	6	BC Mindray	1 NEUTROPHIL	36 BC DXH 500	18	75.456	5.072	6.72	56.600	94.300	



**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020915X	6 BC Mindray	2 LYMPHOCYTE	32 BC-5390	28	36.737	2.007	5.46	27.600	45.900		
2020915X	6 BC Mindray	2 LYMPHOCYTE	36 BC DXH 500	18	12.517	5.275	42.15	9.400	15.600		
2020915X	6 BC Mindray	3 MONOCYTE	32 BC-5390	29	7.645	1.263	16.52	5.600	9.600		
2020915X	6 BC Mindray	3 MONOCYTE	36 BC DXH 500	18	0.050	0.050	100.00	0	2.100		
2020915X	6 BC Mindray	4 EOSINOPHIL	32 BC-5390	27	6.588	0.759	11.52	4.600	8.600		
2020915X	6 BC Mindray	4 EOSINOPHIL	36 BC DXH 500	18	12.006	1.788	14.89	9.000	15.000		
2020915X	6 BC Mindray	5 BASOPHIL	32 BC-5390	28	26.875	0.419	1.56	20.200	33.600		
2020915X	6 BC Mindray	5 BASOPHIL	36 BC DXH 500	18	0.011	0.031	282.84	0	2.000		

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
185 Diff 6 Parameter Ins						CLIA TEST ID #: 765      EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}					
2020911Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	33	51.219	0.965	1.88	38.400	64.000		
2020911Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	37	48.072	0.968	2.01	36.100	60.100		
2020911Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	33	28.058	1.032	3.68	21.000	35.100		
2020911Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	37	30.311	1.068	3.52	22.700	37.900		
2020911Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	2.403	0.517	21.50	0.400	4.400		
2020911Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	37	1.197	0.201	16.76	0	3.200		
2020911Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	33	13.327	0.576	4.32	10.000	16.700		
2020911Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	37	12.975	0.578	4.45	9.700	16.200		
2020911Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	33	5.106	0.856	16.77	3.100	7.100		
2020911Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	37	7.433	0.651	8.76	5.400	9.400		
2020911Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	33	12.958	0.546	4.21	9.700	16.200		
2020911Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	37	12.992	0.604	4.65	9.700	16.200		
2020911Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	17	0.006	0.024	387.30	0	2.000		
2020912Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	33	61.485	1.769	2.88	46.100	76.900		

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2020912Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	37	57.661	0.974	1.69	43.200	72.100			
2020912Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	33	15.939	0.707	4.44	12.000	19.900			
2020912Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	37	17.600	0.790	4.49	13.200	22.000			
2020912Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	1.536	0.460	29.93	0	3.500			
2020912Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	37	0.611	0.266	43.60	0	2.600			
2020912Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	33	15.718	0.507	3.23	11.800	19.600			
2020912Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	37	15.408	0.633	4.11	11.600	19.300			
2020912Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	33	5.150	1.431	27.78	3.200	7.200			
2020912Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	37	8.719	0.678	7.77	6.500	10.900			
2020912Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	33	15.533	0.658	4.23	11.700	19.400			
2020912Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	37	15.717	0.995	6.33	11.800	19.600			
2020912Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	17	0.056	0.117	208.17	0	2.100			
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020913Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	33	59.406	1.296	2.18	44.600	74.300			
2020913Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	37	55.758	0.982	1.76	41.800	69.700			
2020913Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	33	19.218	0.610	3.17	14.400	24.000			
2020913Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	37	20.336	0.459	2.26	15.300	25.400			
2020913Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	1.652	0.360	21.81	0	3.700			
2020913Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	37	0.681	0.135	19.84	0	2.700			
2020913Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	33	14.718	0.700	4.76	11.000	18.400			
2020913Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	37	14.660	0.504	3.44	11.000	18.300			
2020913Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	33	4.913	0.761	15.50	2.900	6.900			
2020913Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	37	8.558	0.730	8.53	6.400	10.700			
2020913Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	33	14.897	0.653	4.38	11.200	18.600			
2020913Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	37	15.200	0.720	4.73	11.400	19.000			
2020913Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	17	0.756	0.512	67.74	0	2.800			
2020914Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	33	51.231	1.127	2.20	38.400	64.000			
2020914Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	37	48.214	1.015	2.11	36.200	60.300			
2020914Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	33	28.022	0.751	2.68	21.000	35.000			

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 127 DIFF 6 PARAM INSTRUM								
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
2020914Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	37	30.139	1.192	3.96	22.600	37.700			
2020914Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	2.467	0.518	21.00	0.500	4.500			
2020914Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	37	1.236	0.206	16.64	0	3.200			
2020914Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	33	13.124	0.413	3.15	9.800	16.400			
2020914Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	37	12.992	0.641	4.93	9.700	16.200			
2020914Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	33	5.185	1.119	21.59	3.200	7.200			
2020914Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	37	7.419	0.618	8.33	5.400	9.400			
2020914Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	33	13.082	0.624	4.77	9.800	16.400			
2020914Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	37	13.086	0.657	5.02	9.800	16.400			
2020914Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	17	0.013	0.033	264.58	0	2.000			
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2020915Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	33	61.567	1.547	2.51	46.200	77.000			
2020915Y	1 6 PAR REG	1 NEUTROPHIL	35 XN350/450/	37	57.714	1.179	2.04	43.300	72.100			
2020915Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	33	15.942	0.927	5.81	12.000	19.900			
2020915Y	1 6 PAR REG	2 LYMPHOCYTE	35 XN350/450/	37	17.611	1.376	7.81	13.200	22.000			
2020915Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	33	1.518	0.355	23.41	0	3.500			
2020915Y	1 6 PAR REG	3 MONOCYTE	35 XN350/450/	37	0.717	0.219	30.59	0	2.700			
2020915Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	33	15.706	0.691	4.40	11.800	19.600			
2020915Y	1 6 PAR REG	4 EOSINOPHIL	35 XN350/450/	37	15.217	0.724	4.76	11.400	19.000			
2020915Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	33	5.116	1.261	24.65	3.100	7.100			
2020915Y	1 6 PAR REG	5 BASOPHIL	35 XN350/450/	37	8.742	0.715	8.18	6.600	10.900			
2020915Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	33	15.724	0.954	6.07	11.800	19.700			
2020915Y	1 6 PAR REG	6 IMMATURE G	35 XN350/450/	37	15.858	0.853	5.38	11.900	19.800			
2020915Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	17	0.050	0.137	273.86	0	2.100			
SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 128 DIFF 3 PARAMETER INS								

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 128 DIFF 3 PARAMETER INS

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH				
186	DIFF 3 PARAMETER INS	CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25	% OR 2.00	{GREATER}							
2020911Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	35	61.421	0.903	1.47	46.100	76.800				
2020911Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	46	60.200	1.092	1.81	45.200	75.300				
2020911Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	36	13.823	0.756	5.47	10.400	17.300				
2020911Z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	46	10.978	0.741	6.75	8.200	13.700				
2020911Z	8 3ST	3 Neutrophil	1 KX 21 - Sy	36	24.766	0.986	3.98	18.600	31.000				
2020911Z	8 3ST	3 Neutrophil	3 ABBOTT C-D	46	28.900	1.094	3.79	21.700	36.100				
2020911Z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	14.433	0.552	3.83	10.800	18.000				
2020911Z	9 3D	1 Lymphocyte	4 JT - Coult	12	14.258	0.998	7.00	10.700	17.800				
2020911Z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	4.050	0.381	9.41	2.100	6.100				
2020911Z	9 3D	2 Mix M/E/Ba	4 JT - Coult	10	3.810	0.513	13.46	1.800	5.800				
2020911Z	9 3D	3 Neutrophil	2 AcT Diff 2	16	81.693	1.156	1.42	61.300	102.100				
2020911Z	9 3D	3 Neutrophil	4 JT - Coult	10	81.720	1.031	1.26	61.300	102.200				
2020912Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	36	13.114	1.569	11.96	9.800	16.400				
2020912Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	46	15.567	1.417	9.10	11.700	19.500				
2020912Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	36	18.363	1.388	7.56	13.800	23.000				
2020912Z	8 3ST	2 Mix M/E/Ba	3 ABBOTT C-D	46	11.782	1.499	12.72	8.800	14.700				
2020912Z	8 3ST	3 Neutrophil	1 KX 21 - Sy	36	68.523	1.813	2.65	51.400	85.700				
2020912Z	8 3ST	3 Neutrophil	3 ABBOTT C-D	46	72.649	2.365	3.26	54.500	90.800				
2020912Z	9 3D	1 Lymphocyte	2 AcT Diff 2	16	61.387	1.448	2.36	46.000	76.700				
2020912Z	9 3D	1 Lymphocyte	4 JT - Coult	12	60.942	3.413	5.60	45.700	76.200				
2020912Z	9 3D	2 Mix M/E/Ba	2 AcT Diff 2	16	7.240	1.137	15.71	5.200	9.200				
2020912Z	9 3D	2 Mix M/E/Ba	4 JT - Coult	10	5.920	1.725	29.14	3.900	7.900				
2020912Z	9 3D	3 Neutrophil	2 AcT Diff 2	16	31.400	1.312	4.18	23.600	39.300				
2020912Z	9 3D	3 Neutrophil	4 JT - Coult	10	32.070	1.519	4.74	24.100	40.100				
2020913Z	8 3ST	1 Lymphocyte	1 KX 21 - Sy	36	30.337	1.040	3.43	22.800	37.900				
2020913Z	8 3ST	1 Lymphocyte	3 ABBOTT C-D	46	30.482	1.321	4.33	22.900	38.100				
2020913Z	8 3ST	2 Mix M/E/Ba	1 KX 21 - Sy	36	17.143	0.797	4.65	12.900	21.400				

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 128 DIFF 3 PARAMETER INS

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020913Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	46	12.840	1.458	11.36	9.600	16.100		
2020913Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	36	52.514	1.009	1.92	39.400	65.600		
2020913Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	46	56.631	2.016	3.56	42.500	70.800		
2020913Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	33.693	0.688	2.04	25.300	42.100		
2020913Z	9 3D	1	Lymphocyte	4	JT - Coult	12	33.267	2.012	6.05	25.000	41.600		
2020913Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	5.673	0.538	9.49	3.700	7.700		
2020913Z	9 3D	2	Mix M/E/Ba	4	JT - Coult	10	5.110	0.696	13.63	3.100	7.100		
2020913Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	60.660	0.936	1.54	45.500	75.800		
2020913Z	9 3D	3	Neutrophil	4	JT - Coult	10	61.020	1.347	2.21	45.800	76.300		
186 DIFF 3 PARAMETER INS		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2020914Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	36	61.340	0.695	1.13	46.000	76.700		
2020914Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	47	60.167	1.289	2.14	45.100	75.200		
2020914Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	36	13.889	0.635	4.57	10.400	17.400		
2020914Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	46	10.884	0.646	5.93	8.200	13.600		
2020914Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	36	24.771	0.800	3.23	18.600	31.000		
2020914Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	45	29.027	1.043	3.59	21.800	36.300		
2020914Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	14.800	0.438	2.96	11.100	18.500		
2020914Z	9 3D	1	Lymphocyte	4	JT - Coult	12	14.350	0.968	6.75	10.800	17.900		
2020914Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	3.920	0.533	13.60	1.900	5.900		
2020914Z	9 3D	2	Mix M/E/Ba	4	JT - Coult	10	3.830	0.550	14.35	1.800	5.800		
2020914Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	81.280	0.620	0.76	61.000	101.600		
2020914Z	9 3D	3	Neutrophil	4	JT - Coult	10	81.540	0.831	1.02	61.200	101.900		
2020915Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	35	12.432	1.101	8.86	9.300	15.500		
2020915Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	46	15.749	1.536	9.75	11.800	19.700		
2020915Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	36	18.431	1.393	7.56	13.800	23.000		
2020915Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	46	11.878	1.458	12.28	8.900	14.800		
2020915Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	37	68.997	1.769	2.56	51.700	86.200		
2020915Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	46	72.364	2.337	3.23	54.300	90.500		
2020915Z	9 3D	1	Lymphocyte	2	AcT Diff 2	16	61.480	1.330	2.16	46.100	76.900		

### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 128 DIFF 3 PARAMETER INS									
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020915Z	9 3D	1	Lymphocyte	4	JT - Coult	11	61.291	3.491	5.70	46.000	76.600		
2020915Z	9 3D	2	Mix M/E/Ba	2	AcT Diff 2	16	7.227	1.317	18.23	5.200	9.200		
2020915Z	9 3D	2	Mix M/E/Ba	4	JT - Coult	10	6.000	1.210	20.17	4.000	8.000		
2020915Z	9 3D	3	Neutrophil	2	AcT Diff 2	16	31.347	1.424	4.54	23.500	39.200		
2020915Z	9 3D	3	Neutrophil	4	JT - Coult	10	31.860	1.618	5.08	23.900	39.800		

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMPL BLOOD COUNT									
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %									
2020-911	1 Sysmex CBC	98	SYSMEX K-	13	5.688	0.051	0.90	5.350	6.030				
2020-911	1 Sysmex CBC	127	SYSMEX XT	17	4.630	0.059	1.27	4.350	4.910				
2020-911	1 Sysmex CBC	132	SYSMEX XS	151	4.539	0.053	1.18	4.270	4.810				
2020-911	1 Sysmex CBC	141	SYSMEX XN	33	4.553	0.056	1.23	4.280	4.830				
2020-911	1 Sysmex CBC	152	System XN	39	4.549	0.053	1.18	4.280	4.820				
2020-911	2 ACT PENTRA	125	COULTER AC	153	4.556	0.076	1.66	4.280	4.830				
2020-911	2 ACT PENTRA	130	ABX PENTRA	25	4.520	0.071	1.57	4.250	4.790				
2020-911	3 ADVIA 120	116	ADVIA-120	18	4.599	0.093	2.01	4.320	4.880				
2020-911	4 AB-CELL DY	115	ABBOTT C-D	121	4.604	0.133	2.89	4.330	4.880				
2020-911	4 AB-CELL DY	149	BC DXH 500	14	4.618	0.100	2.17	4.340	4.890				
2020-911	5 BECTON D	137	B.C DXH	46	4.130	0.057	1.37	3.880	4.380				
2020-911	6 MR Mindray	145	BC-5390	28	4.420	0.066	1.50	4.150	4.680				
2020-911	8 3 ST	84	ABBOTT C-D	43	5.764	0.137	2.38	5.420	6.110				
2020-911	8 3 ST	98	SYSMEX K-	16	5.729	0.085	1.49	5.380	6.070				
2020-911	9 3 D	33	COULTER JT	12	5.739	0.100	1.74	5.390	6.080				
2020-911	9 3 D	120	COULTER	16	5.740	0.124	2.16	5.400	6.080				
2020-912	1 Sysmex CBC	98	SYSMEX K-	13	2.417	0.026	1.08	2.270	2.560				

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-912	1	Sysmex CBC	127 SYSMEX XT	17	2.545	0.034	1.34	2.390	2.700		
2020-912	1	Sysmex CBC	132 SYSMEX XS	151	2.448	0.029	1.20	2.300	2.600		
2020-912	1	Sysmex CBC	141 SYSMEX XN	33	2.441	0.034	1.41	2.290	2.590		
2020-912	1	Sysmex CBC	152 Sysmex XN	39	2.415	0.024	0.98	2.270	2.560		
2020-912	2	ACT PENTRA	125 COULTER AC	153	4.541	0.067	1.48	4.270	4.810		
2020-912	2	ACT PENTRA	130 ABX PENTRA	25	4.541	0.063	1.40	4.270	4.810		
2020-912	3	ADVIA 120	116 ADVIA-120	18	4.586	0.070	1.54	4.310	4.860		
2020-912	4	AB-CELL DY	115 ABBOTT C-D	121	2.265	0.054	2.40	2.130	2.400		
2020-912	4	AB-CELL DY	149 BC DXH 500	14	2.267	0.047	2.08	2.130	2.400		
2020-912	5	BECTON D	137 B.C DXH	46	2.556	0.038	1.49	2.400	2.710		
2020-912	6	MR Mindray	145 BC-5390	28	2.080	0.039	1.89	1.960	2.210		
2020-912	8	3 ST	84 ABBOTT C-D	43	2.488	0.086	3.46	2.340	2.640		
2020-912	8	3 ST	98 SYSMEX K-	16	2.432	0.036	1.46	2.290	2.580		
2020-912	9	3 D	33 COULTER JT	12	2.420	0.043	1.76	2.270	2.570		
2020-912	9	3 D	120 COULTER	16	2.480	0.046	1.84	2.330	2.630		
549 RED CELL COUNT (RBC)		CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %								
2020-913	1	Sysmex CBC	98 SYSMEX K-	13	4.159	0.056	1.35	3.910	4.410		
2020-913	1	Sysmex CBC	127 SYSMEX XT	17	5.651	0.101	1.78	5.310	5.990		
2020-913	1	Sysmex CBC	132 SYSMEX XS	151	5.588	0.062	1.11	5.250	5.920		
2020-913	1	Sysmex CBC	141 SYSMEX XN	33	5.629	0.076	1.35	5.290	5.970		
2020-913	1	Sysmex CBC	152 Sysmex XN	39	5.638	0.047	0.83	5.300	5.980		
2020-913	2	ACT PENTRA	125 COULTER AC	153	2.271	0.050	2.22	2.140	2.410		
2020-913	2	ACT PENTRA	130 ABX PENTRA	25	2.277	0.046	2.04	2.140	2.410		
2020-913	3	ADVIA 120	116 ADVIA-120	18	2.397	0.033	1.38	2.250	2.540		
2020-913	4	AB-CELL DY	115 ABBOTT C-D	121	5.221	0.159	3.05	4.910	5.530		
2020-913	4	AB-CELL DY	149 BC DXH 500	14	5.204	0.059	1.14	4.890	5.520		
2020-913	5	BECTON D	137 B.C DXH	46	5.380	0.069	1.28	5.060	5.700		
2020-913	6	MR Mindray	145 BC-5390	28	5.071	0.086	1.69	4.770	5.380		
2020-913	8	3 ST	84 ABBOTT C-D	43	4.260	0.116	2.72	4.000	4.520		
2020-913	8	3 ST	98 SYSMEX K-	16	4.159	0.067	1.61	3.910	4.410		
2020-913	9	3 D	33 COULTER JT	12	4.666	0.090	1.92	4.390	4.950		

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART		TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-913	9 3 D	120 COULTER		16	4.608	0.060	1.29	4.330	4.880	
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %							
2020-914	1 Sysmex CBC	98 SYSMEX K-		13	5.685	0.034	0.60	5.340	6.030	
2020-914	1 Sysmex CBC	127 SYSMEX XT		17	4.637	0.078	1.67	4.360	4.920	
2020-914	1 Sysmex CBC	132 SYSMEX XS		151	4.541	0.056	1.24	4.270	4.810	
2020-914	1 Sysmex CBC	141 SYSMEX XN		33	4.558	0.057	1.24	4.280	4.830	
2020-914	1 Sysmex CBC	152 Sysmex XN		39	4.552	0.044	0.97	4.280	4.830	
2020-914	2 ACT PENTRA	125 COULTER AC		153	5.046	0.074	1.48	4.740	5.350	
2020-914	2 ACT PENTRA	130 ABX PENTRA		25	5.002	0.077	1.55	4.700	5.300	
2020-914	3 ADVIA 120	116 ADVIA-120		18	5.267	0.110	2.09	4.950	5.580	
2020-914	4 AB-CELL DY	115 ABBOTT C-D		121	4.608	0.130	2.82	4.330	4.880	
2020-914	4 AB-CELL DY	149 BC DXH 500		14	4.582	0.061	1.33	4.310	4.860	
2020-914	5 BECTON D	137 B.C DXH		46	4.117	0.042	1.01	3.870	4.360	
2020-914	6 MR Mindray	145 BC-5390		28	4.430	0.066	1.50	4.160	4.700	
2020-914	8 3 ST	84 ABBOTT C-D		43	5.757	0.149	2.60	5.410	6.100	
2020-914	8 3 ST	98 SYSMEX K-		16	5.721	0.095	1.67	5.380	6.060	
2020-914	9 3 D	33 COULTER JT		12	5.737	0.131	2.29	5.390	6.080	
2020-914	9 3 D	120 COULTER		16	5.730	0.157	2.75	5.390	6.070	
2020-915	1 Sysmex CBC	98 SYSMEX K-		13	2.418	0.033	1.37	2.270	2.560	
2020-915	1 Sysmex CBC	127 SYSMEX XT		17	2.547	0.036	1.42	2.390	2.700	
2020-915	1 Sysmex CBC	132 SYSMEX XS		151	2.449	0.031	1.25	2.300	2.600	
2020-915	1 Sysmex CBC	141 SYSMEX XN		33	2.441	0.032	1.31	2.290	2.590	
2020-915	1 Sysmex CBC	152 Sysmex XN		39	2.414	0.023	0.97	2.270	2.560	
2020-915	2 ACT PENTRA	125 COULTER AC		153	2.265	0.058	2.55	2.130	2.400	
2020-915	2 ACT PENTRA	130 ABX PENTRA		25	2.262	0.054	2.41	2.130	2.400	
2020-915	3 ADVIA 120	116 ADVIA-120		18	2.403	0.032	1.33	2.260	2.550	
2020-915	4 AB-CELL DY	115 ABBOTT C-D		121	2.264	0.057	2.53	2.130	2.400	
2020-915	4 AB-CELL DY	149 BC DXH 500		14	2.277	0.024	1.04	2.140	2.410	
2020-915	5 BECTON D	137 B.C DXH		46	2.553	0.036	1.40	2.400	2.710	



**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-915	6 MR Mindray	145 BC-5390			28	2.086	0.040	1.94	1.960	2.210	
2020-915	8 3 ST	84 ABBOTT C-D			43	2.471	0.068	2.77	2.320	2.620	
2020-915	8 3 ST	98 SYSMEX K-			16	2.446	0.058	2.36	2.300	2.590	
2020-915	9 3 D	33 COULTER JT			12	2.433	0.046	1.91	2.290	2.580	
2020-915	9 3 D	120 COULTER			16	2.457	0.062	2.53	2.310	2.600	
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785									
2020-911	1 Sysmex CBC	98 SYSMEX K-			13	49.869	0.794	1.59	46.900	52.900	
2020-911	1 Sysmex CBC	127 SYSMEX XT			17	39.600	0.744	1.88	37.200	42.000	
2020-911	1 Sysmex CBC	132 SYSMEX XS			150	38.972	0.720	1.85	36.600	41.300	
2020-911	1 Sysmex CBC	141 SYSMEX XN			35	38.511	0.520	1.35	36.200	40.800	
2020-911	1 Sysmex CBC	152 Sysmex XN			39	38.487	0.687	1.78	36.200	40.800	
2020-911	2 ACT PENTRA	125 COULTER AC			153	37.801	0.925	2.45	35.500	40.100	
2020-911	2 ACT PENTRA	130 ABX PENTRA			25	37.496	0.686	1.83	35.200	39.700	
2020-911	3 ADVIA 120	116 ADVIA-120			18	36.478	0.621	1.70	34.300	38.700	
2020-911	4 AB-CELL DY	115 ABBOTT C-D			122	38.213	1.149	3.01	35.900	40.500	
2020-911	4 AB-CELL DY	149 BC DXH 500			14	41.293	0.793	1.92	38.800	43.800	
2020-911	5 BECTON D	137 B.C DXH			46	38.770	0.530	1.37	36.400	41.100	
2020-911	6 MR Mindray	145 BC-5390			28	38.161	0.704	1.85	35.900	40.500	
2020-911	8 3 ST	84 ABBOTT C-D			43	53.228	1.452	2.73	50.000	56.400	
2020-911	8 3 ST	98 SYSMEX K-			16	50.280	0.818	1.63	47.300	53.300	
2020-911	9 3 D	33 COULTER JT			12	52.250	0.925	1.77	49.100	55.400	
2020-911	9 3 D	120 COULTER			16	52.480	1.244	2.37	49.300	55.600	
2020-912	1 Sysmex CBC	98 SYSMEX K-			13	18.062	0.318	1.76	17.000	19.100	
2020-912	1 Sysmex CBC	127 SYSMEX XT			17	18.735	0.366	1.95	17.600	19.900	
2020-912	1 Sysmex CBC	132 SYSMEX XS			150	18.009	0.508	2.82	16.900	19.100	
2020-912	1 Sysmex CBC	141 SYSMEX XN			35	17.663	0.324	1.84	16.600	18.700	
2020-912	1 Sysmex CBC	152 Sysmex XN			39	17.585	0.289	1.64	16.500	18.600	
2020-912	2 ACT PENTRA	125 COULTER AC			153	37.786	0.689	1.82	35.500	40.100	
2020-912	2 ACT PENTRA	130 ABX PENTRA			25	37.772	0.687	1.82	35.500	40.000	

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-912	3	ADVIA 120	116 ADVIA-120	18	36.383	0.599	1.65	34.200	38.600	
2020-912	4	AB-CELL DY	115 ABBOTT C-D	122	15.720	0.414	2.64	14.800	16.700	
2020-912	4	AB-CELL DY	149 BC DXH 500	14	17.057	0.344	2.02	16.000	18.100	
2020-912	5	BECTON D	137 B.C DXH	46	19.657	0.342	1.74	18.500	20.800	
2020-912	6	MR Mindray	145 BC-5390	28	17.096	0.302	1.77	16.100	18.100	
2020-912	8	3 ST	84 ABBOTT C-D	43	18.949	0.634	3.34	17.800	20.100	
2020-912	8	3 ST	98 SYSMEX K-	16	18.200	0.348	1.91	17.100	19.300	
2020-912	9	3 D	33 COULTER JT	12	16.942	0.364	2.15	15.900	18.000	
2020-912	9	3 D	120 COULTER	16	17.373	0.351	2.02	16.300	18.400	
550 HEMATOCRIT (HCT)				CLIA TEST ID #: 785      EVALUATION CRITERIA: TARGET VALUE 6 %						
2020-913	1	Sysmex CBC	98 SYSMEX K-	13	33.877	0.705	2.08	31.800	35.900	
2020-913	1	Sysmex CBC	127 SYSMEX XT	17	52.635	1.095	2.08	49.500	55.800	
2020-913	1	Sysmex CBC	132 SYSMEX XS	150	52.158	0.923	1.77	49.000	55.300	
2020-913	1	Sysmex CBC	141 SYSMEX XN	35	51.411	0.731	1.42	48.300	54.500	
2020-913	1	Sysmex CBC	152 Sysmex XN	39	51.618	0.674	1.31	48.500	54.700	
2020-913	2	ACT PENTRA	125 COULTER AC	153	18.379	0.359	1.95	17.300	19.500	
2020-913	2	ACT PENTRA	130 ABX PENTRA	25	18.348	0.473	2.58	17.200	19.400	
2020-913	3	ADVIA 120	116 ADVIA-120	18	15.906	0.293	1.84	15.000	16.900	
2020-913	4	AB-CELL DY	115 ABBOTT C-D	122	44.488	1.478	3.32	41.800	47.200	
2020-913	4	AB-CELL DY	149 BC DXH 500	14	48.607	0.865	1.78	45.700	51.500	
2020-913	5	BECTON D	137 B.C DXH	46	55.461	0.833	1.50	52.100	58.800	
2020-913	6	MR Mindray	145 BC-5390	28	47.343	1.001	2.11	44.500	50.200	
2020-913	8	3 ST	84 ABBOTT C-D	43	36.170	1.050	2.90	34.000	38.300	
2020-913	8	3 ST	98 SYSMEX K-	16	33.913	0.564	1.66	31.900	35.900	
2020-913	9	3 D	33 COULTER JT	12	38.967	0.781	2.01	36.600	41.300	
2020-913	9	3 D	120 COULTER	16	38.786	0.816	2.10	36.500	41.100	
2020-914	1	Sysmex CBC	98 SYSMEX K-	13	49.731	0.807	1.62	46.700	52.700	
2020-914	1	Sysmex CBC	127 SYSMEX XT	17	39.688	0.763	1.92	37.300	42.100	
2020-914	1	Sysmex CBC	132 SYSMEX XS	150	38.979	0.746	1.91	36.600	41.300	

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-914	1	Sysmex CBC	141	SYSMEX XN	35	38.549	0.586	1.52	36.200	40.900	
2020-914	1	Sysmex CBC	152	Sysmex XN	39	38.592	0.569	1.47	36.300	40.900	
2020-914	2	ACT PENTRA	125	COULTER AC	153	44.875	0.835	1.86	42.200	47.600	
2020-914	2	ACT PENTRA	130	ABX PENTRA	25	44.596	0.683	1.53	41.900	47.300	
2020-914	3	ADVIA 120	116	ADVIA-120	18	46.956	1.259	2.68	44.100	49.800	
2020-914	4	AB-CELL DY	115	ABBOTT C-D	122	38.333	1.236	3.23	36.000	40.600	
2020-914	4	AB-CELL DY	149	BC DXH 500	14	41.100	0.500	1.22	38.600	43.600	
2020-914	5	BECTON D	137	B.C DXH	46	38.785	0.526	1.36	36.500	41.100	
2020-914	6	MR Mindray	145	BC-5390	28	38.164	0.699	1.83	35.900	40.500	
2020-914	8	3 ST	84	ABBOTT C-D	43	53.144	1.272	2.39	50.000	56.300	
2020-914	8	3 ST	98	SYSMEX K-	16	50.053	0.818	1.63	47.100	53.100	
2020-914	9	3 D	33	COULTER JT	12	52.017	1.025	1.97	48.900	55.100	
2020-914	9	3 D	120	COULTER	16	52.400	1.629	3.11	49.300	55.500	
550 HEMATOCRIT (HCT)			CLIA TEST ID #: 785			EVALUATION CRITERIA: TARGET VALUE 6 %					
2020-915	1	Sysmex CBC	98	SYSMEX K-	13	17.992	0.338	1.88	16.900	19.100	
2020-915	1	Sysmex CBC	127	SYSMEX XT	17	18.729	0.351	1.87	17.600	19.900	
2020-915	1	Sysmex CBC	132	SYSMEX XS	150	18.023	0.347	1.93	16.900	19.100	
2020-915	1	Sysmex CBC	141	SYSMEX XN	35	17.631	0.296	1.68	16.600	18.700	
2020-915	1	Sysmex CBC	152	Sysmex XN	39	17.577	0.313	1.78	16.500	18.600	
2020-915	2	ACT PENTRA	125	COULTER AC	153	18.324	0.472	2.58	17.200	19.400	
2020-915	2	ACT PENTRA	130	ABX PENTRA	25	18.184	0.464	2.55	17.100	19.300	
2020-915	3	ADVIA 120	116	ADVIA-120	18	15.989	0.323	2.02	15.000	16.900	
2020-915	4	AB-CELL DY	115	ABBOTT C-D	122	15.719	0.448	2.85	14.800	16.700	
2020-915	4	AB-CELL DY	149	BC DXH 500	14	17.086	0.185	1.08	16.100	18.100	
2020-915	5	BECTON D	137	B.C DXH	46	19.641	0.317	1.61	18.500	20.800	
2020-915	6	MR Mindray	145	BC-5390	28	17.143	0.376	2.20	16.100	18.200	
2020-915	8	3 ST	84	ABBOTT C-D	43	18.860	0.558	2.96	17.700	20.000	
2020-915	8	3 ST	98	SYSMEX K-	16	18.156	0.377	2.08	17.100	19.200	
2020-915	9	3 D	33	COULTER JT	12	17.033	0.419	2.46	16.000	18.100	
2020-915	9	3 D	120	COULTER	16	17.071	0.474	2.78	16.000	18.100	

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
551	HEMOGLOBIN (HGB)	CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}							
2020-911	1	Sysmex CBC	98	SYSMEX K-	13	17.977	0.172	0.95	16.700	19.200	
2020-911	1	Sysmex CBC	127	SYSMEX XT	17	13.288	0.153	1.15	12.400	14.200	
2020-911	1	Sysmex CBC	132	SYSMEX XS	151	13.172	0.152	1.15	12.300	14.100	
2020-911	1	Sysmex CBC	141	SYSMEX XN	33	13.324	0.121	0.91	12.400	14.300	
2020-911	1	Sysmex CBC	152	Sysmex XN	39	13.300	0.122	0.92	12.400	14.200	
2020-911	2	ACT PENTRA	125	COULTER AC	153	13.641	0.177	1.30	12.700	14.600	
2020-911	2	ACT PENTRA	130	ABX PENTRA	25	13.652	0.200	1.47	12.700	14.600	
2020-911	3	ADVIA 120	116	ADVIA-120	18	13.411	0.262	1.96	12.500	14.300	
2020-911	4	AB-CELL DY	115	ABBOTT C-D	123	13.903	0.300	2.15	12.900	14.900	
2020-911	4	AB-CELL DY	149	BC DXH 500	14	12.357	0.354	2.86	11.500	13.200	
2020-911	5	BECTON D	137	B.C DXH	46	12.148	0.170	1.40	11.300	13.000	
2020-911	6	MR Mindray	145	BC-5390	28	13.107	0.191	1.46	12.200	14.000	
2020-911	8	3 ST	84	ABBOTT C-D	43	18.052	0.319	1.77	16.800	19.300	
2020-911	8	3 ST	98	SYSMEX K-	16	18.160	0.212	1.17	16.900	19.400	
2020-911	9	3 D	33	COULTER JT	12	18.792	0.333	1.77	17.500	20.100	
2020-911	9	3 D	120	COULTER	16	18.147	0.314	1.73	16.900	19.400	
2020-912	1	Sysmex CBC	98	SYSMEX K-	13	6.085	0.129	2.12	5.700	6.500	
2020-912	1	Sysmex CBC	127	SYSMEX XT	17	6.071	0.075	1.23	5.600	6.500	
2020-912	1	Sysmex CBC	132	SYSMEX XS	151	5.908	0.146	2.47	5.500	6.300	
2020-912	1	Sysmex CBC	141	SYSMEX XN	33	6.045	0.074	1.23	5.600	6.500	
2020-912	1	Sysmex CBC	152	Sysmex XN	39	6.026	0.064	1.06	5.600	6.400	
2020-912	2	ACT PENTRA	125	COULTER AC	153	13.639	0.175	1.29	12.700	14.600	
2020-912	2	ACT PENTRA	130	ABX PENTRA	25	13.676	0.188	1.38	12.700	14.600	
2020-912	3	ADVIA 120	116	ADVIA-120	18	13.356	0.231	1.73	12.400	14.300	
2020-912	4	AB-CELL DY	115	ABBOTT C-D	123	5.890	0.175	2.97	5.500	6.300	
2020-912	4	AB-CELL DY	149	BC DXH 500	14	5.036	0.117	2.33	4.700	5.400	
2020-912	5	BECTON D	137	B.C DXH	46	6.052	0.074	1.23	5.600	6.500	
2020-912	6	MR Mindray	145	BC-5390	28	5.804	0.124	2.13	5.400	6.200	

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-912	8 3 ST	84 ABBOTT C-D			43	6.328	0.151	2.39	5.900	6.800	
2020-912	8 3 ST	98 SYSMEX K-			16	6.106	0.139	2.28	5.700	6.500	
2020-912	9 3 D	33 COULTER JT			12	5.700	0.168	2.95	5.300	6.100	
2020-912	9 3 D	120 COULTER			16	5.800	0.146	2.52	5.400	6.200	
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}							
2020-913	1 Sysmex CBC	98 SYSMEX K-			13	12.062	0.155	1.28	11.200	12.900	
2020-913	1 Sysmex CBC	127 SYSMEX XT			17	17.565	0.133	0.75	16.300	18.800	
2020-913	1 Sysmex CBC	132 SYSMEX XS			151	17.605	0.197	1.12	16.400	18.800	
2020-913	1 Sysmex CBC	141 SYSMEX XN			33	17.685	0.154	0.87	16.400	18.900	
2020-913	1 Sysmex CBC	152 Sysmex XN			39	17.649	0.150	0.85	16.400	18.900	
2020-913	2 ACT PENTRA	125 COULTER AC			153	6.495	0.111	1.72	6.000	7.000	
2020-913	2 ACT PENTRA	130 ABX PENTRA			25	6.396	0.106	1.66	5.900	6.800	
2020-913	3 ADVIA 120	116 ADVIA-120			18	5.767	0.137	2.38	5.400	6.200	
2020-913	4 AB-CELL DY	115 ABBOTT C-D			123	16.631	0.355	2.13	15.500	17.800	
2020-913	4 AB-CELL DY	149 BC DXH 500			14	15.143	0.269	1.78	14.100	16.200	
2020-913	5 BECTON D	137 B.C DXH			46	17.074	0.225	1.32	15.900	18.300	
2020-913	6 MR Mindray	145 BC-5390			28	16.357	0.248	1.52	15.200	17.500	
2020-913	8 3 ST	84 ABBOTT C-D			43	12.157	0.217	1.79	11.300	13.000	
2020-913	8 3 ST	98 SYSMEX K-			16	12.093	0.157	1.30	11.200	12.900	
2020-913	9 3 D	33 COULTER JT			12	13.808	0.301	2.18	12.800	14.800	
2020-913	9 3 D	120 COULTER			16	13.367	0.305	2.28	12.400	14.300	
2020-914	1 Sysmex CBC	98 SYSMEX K-			13	18.008	0.173	0.96	16.700	19.300	
2020-914	1 Sysmex CBC	127 SYSMEX XT			17	13.276	0.159	1.20	12.300	14.200	
2020-914	1 Sysmex CBC	132 SYSMEX XS			151	13.195	0.147	1.11	12.300	14.100	
2020-914	1 Sysmex CBC	141 SYSMEX XN			33	13.345	0.121	0.91	12.400	14.300	
2020-914	1 Sysmex CBC	152 Sysmex XN			39	13.326	0.115	0.86	12.400	14.300	
2020-914	2 ACT PENTRA	125 COULTER AC			153	16.197	0.211	1.30	15.100	17.300	
2020-914	2 ACT PENTRA	130 ABX PENTRA			25	16.312	0.253	1.55	15.200	17.500	
2020-914	3 ADVIA 120	116 ADVIA-120			18	17.389	0.354	2.04	16.200	18.600	

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-914	4 AB-CELL DY	115 ABBOTT C-D			123	13.882	0.564	4.06	12.900	14.900	
2020-914	4 AB-CELL DY	149 BC DXH 500			14	12.207	0.237	1.95	11.400	13.100	
2020-914	5 BECTON D	137 B.C DXH			46	12.172	0.164	1.35	11.300	13.000	
2020-914	6 MR Mindray	145 BC-5390			28	13.089	0.168	1.28	12.200	14.000	
2020-914	8 3 ST	84 ABBOTT C-D			43	18.051	0.302	1.67	16.800	19.300	
2020-914	8 3 ST	98 SYSMEX K-			16	18.180	0.190	1.05	16.900	19.500	
2020-914	9 3 D	33 COULTER JT			12	18.883	0.251	1.33	17.600	20.200	
2020-914	9 3 D	120 COULTER			16	18.353	0.316	1.72	17.100	19.600	
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795							EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}		
2020-915	1 Sysmex CBC	98 SYSMEX K-			13	6.075	0.142	2.34	5.600	6.500	
2020-915	1 Sysmex CBC	127 SYSMEX XT			17	6.088	0.068	1.11	5.700	6.500	
2020-915	1 Sysmex CBC	132 SYSMEX XS			151	5.927	0.075	1.26	5.500	6.300	
2020-915	1 Sysmex CBC	141 SYSMEX XN			33	6.048	0.056	0.92	5.600	6.500	
2020-915	1 Sysmex CBC	152 Sysmex XN			39	6.037	0.067	1.10	5.600	6.500	
2020-915	2 ACT PENTRA	125 COULTER AC			153	6.483	0.152	2.34	6.000	6.900	
2020-915	2 ACT PENTRA	130 ABX PENTRA			25	6.356	0.127	1.99	5.900	6.800	
2020-915	3 ADVIA 120	116 ADVIA-120			18	5.794	0.108	1.86	5.400	6.200	
2020-915	4 AB-CELL DY	115 ABBOTT C-D			123	5.871	0.205	3.49	5.500	6.300	
2020-915	4 AB-CELL DY	149 BC DXH 500			14	5.029	0.103	2.05	4.700	5.400	
2020-915	5 BECTON D	137 B.C DXH			46	6.052	0.074	1.23	5.600	6.500	
2020-915	6 MR Mindray	145 BC-5390			28	5.807	0.110	1.89	5.400	6.200	
2020-915	8 3 ST	84 ABBOTT C-D			43	6.314	0.159	2.52	5.900	6.800	
2020-915	8 3 ST	98 SYSMEX K-			16	6.100	0.146	2.39	5.700	6.500	
2020-915	9 3 D	33 COULTER JT			12	5.708	0.171	2.99	5.300	6.100	
2020-915	9 3 D	120 COULTER			16	5.793	0.110	1.90	5.400	6.200	
552 WHITE CELL COUNT(WBC)		CLIA TEST ID #: 805							EVALUATION CRITERIA: TARGET VALUE 15 %		
2020-911	1 Sysmex CBC	98 SYSMEX K-			13	19.631	0.516	2.63	16.700	22.600	
2020-911	1 Sysmex CBC	127 SYSMEX XT			17	7.812	0.293	3.75	6.600	9.000	
2020-911	1 Sysmex CBC	132 SYSMEX XS			149	6.673	0.571	8.56	5.700	7.700	
2020-911	1 Sysmex CBC	141 SYSMEX XN			34	7.344	0.473	6.44	6.200	8.400	

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-911	1	Sysmex CBC	152	Sysmex XN	39	6.624	0.292	4.41	5.600	7.600	
2020-911	2	ACT PENTRA	125	COULTER AC	153	8.072	0.256	3.18	6.900	9.300	
2020-911	2	ACT PENTRA	130	ABX PENTRA	25	7.840	0.294	3.75	6.700	9.000	
2020-911	3	ADVIA 120	116	ADVIA-120	18	7.750	0.455	5.87	6.600	8.900	
2020-911	4	AB-CELL DY	115	ABBOTT C-D	123	7.559	0.236	3.13	6.400	8.700	
2020-911	4	AB-CELL DY	149	BC DXH 500	14	6.521	0.208	3.18	5.500	7.500	
2020-911	5	BECTON D	137	B.C DXH	46	9.387	0.190	2.02	8.000	10.800	
2020-911	6	MR Mindray	145	BC-5390	28	8.196	0.272	3.32	7.000	9.400	
2020-911	8	3 ST	84	ABBOTT C-D	43	21.070	1.193	5.66	17.900	24.200	
2020-911	8	3 ST	98	SYSMEX K-	16	19.620	0.290	1.48	16.700	22.600	
2020-911	9	3 D	33	COULTER JT	12	21.500	0.456	2.12	18.300	24.700	
2020-911	9	3 D	120	COULTER	16	20.327	0.386	1.90	17.300	23.400	
552 WHITE CELL COUNT(WBC		CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %								
2020-912	1	Sysmex CBC	98	SYSMEX K-	13	2.877	0.080	2.78	2.400	3.300	
2020-912	1	Sysmex CBC	127	SYSMEX XT	17	3.953	0.124	3.14	3.400	4.500	
2020-912	1	Sysmex CBC	132	SYSMEX XS	149	3.900	0.149	3.83	3.300	4.500	
2020-912	1	Sysmex CBC	141	SYSMEX XN	34	3.715	0.321	8.64	3.200	4.300	
2020-912	1	Sysmex CBC	152	Sysmex XN	39	3.126	0.202	6.47	2.700	3.600	
2020-912	2	ACT PENTRA	125	COULTER AC	153	8.052	0.228	2.83	6.800	9.300	
2020-912	2	ACT PENTRA	130	ABX PENTRA	25	7.868	0.353	4.49	6.700	9.000	
2020-912	3	ADVIA 120	116	ADVIA-120	18	7.867	0.428	5.44	6.700	9.000	
2020-912	4	AB-CELL DY	115	ABBOTT C-D	123	2.899	0.162	5.58	2.500	3.300	
2020-912	4	AB-CELL DY	149	BC DXH 500	14	2.064	0.111	5.37	1.800	2.400	
2020-912	5	BECTON D	137	B.C DXH	46	4.078	0.146	3.58	3.500	4.700	
2020-912	6	MR Mindray	145	BC-5390	28	3.407	0.131	3.84	2.900	3.900	
2020-912	8	3 ST	84	ABBOTT C-D	43	3.284	0.184	5.61	2.800	3.800	
2020-912	8	3 ST	98	SYSMEX K-	16	2.853	0.088	3.10	2.400	3.300	
2020-912	9	3 D	33	COULTER JT	12	2.025	0.083	4.09	1.700	2.300	
2020-912	9	3 D	120	COULTER	16	2.160	0.108	5.01	1.800	2.500	

## Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
552	WHITE CELL COUNT(WBC	CLIA TEST ID #: 805	EVALUATION CRITERIA: TARGET VALUE 15 %								
2020-913	1 Sysmex CBC	98 SYSMEX K-			13	7.900	0.175	2.22	6.700	9.100	
2020-913	1 Sysmex CBC	127 SYSMEX XT			17	20.847	0.690	3.31	17.700	24.000	
2020-913	1 Sysmex CBC	132 SYSMEX XS			149	20.451	0.510	2.50	17.400	23.500	
2020-913	1 Sysmex CBC	141 SYSMEX XN			34	21.174	0.626	2.96	18.000	24.300	
2020-913	1 Sysmex CBC	152 Sysmex XN			39	20.397	0.367	1.80	17.300	23.500	
2020-913	2 ACT PENTRA	125 COULTER AC			153	2.886	0.116	4.01	2.500	3.300	
2020-913	2 ACT PENTRA	130 ABX PENTRA			25	2.832	0.138	4.86	2.400	3.300	
2020-913	3 ADVIA 120	116 ADVIA-120			18	3.144	0.117	3.71	2.700	3.600	
2020-913	4 AB-CELL DY	115 ABBOTT C-D			123	20.021	0.584	2.91	17.000	23.000	
2020-913	4 AB-CELL DY	149 BC DXH 500			14	18.250	0.472	2.58	15.500	21.000	
2020-913	5 BECTON D	137 B.C DXH			46	20.871	0.322	1.54	17.700	24.000	
2020-913	6 MR Mindray	145 BC-5390			28	18.671	0.437	2.34	15.900	21.500	
2020-913	8 3 ST	84 ABBOTT C-D			43	8.345	0.411	4.93	7.100	9.600	
2020-913	8 3 ST	98 SYSMEX K-			16	7.767	0.221	2.85	6.600	8.900	
2020-913	9 3 D	33 COULTER JT			12	7.775	0.209	2.68	6.600	8.900	
2020-913	9 3 D	120 COULTER			16	7.673	0.169	2.20	6.500	8.800	
2020-914	1 Sysmex CBC	98 SYSMEX K-			13	19.623	0.419	2.14	16.700	22.600	
2020-914	1 Sysmex CBC	127 SYSMEX XT			17	7.729	0.304	3.94	6.600	8.900	
2020-914	1 Sysmex CBC	132 SYSMEX XS			149	6.730	0.585	8.70	5.700	7.700	
2020-914	1 Sysmex CBC	141 SYSMEX XN			34	7.397	0.409	5.53	6.300	8.500	
2020-914	1 Sysmex CBC	152 Sysmex XN			39	6.632	0.319	4.81	5.600	7.600	
2020-914	2 ACT PENTRA	125 COULTER AC			153	18.357	0.568	3.10	15.600	21.100	
2020-914	2 ACT PENTRA	130 ABX PENTRA			25	17.936	0.601	3.35	15.200	20.600	
2020-914	3 ADVIA 120	116 ADVIA-120			18	16.956	1.049	6.19	14.400	19.500	
2020-914	4 AB-CELL DY	115 ABBOTT C-D			123	7.643	0.224	2.94	6.500	8.800	
2020-914	4 AB-CELL DY	149 BC DXH 500			14	6.507	0.144	2.21	5.500	7.500	
2020-914	5 BECTON D	137 B.C DXH			46	9.426	0.189	2.01	8.000	10.800	
2020-914	6 MR Mindray	145 BC-5390			28	8.200	0.222	2.71	7.000	9.400	



### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 132 COMPL BLOOD COUNT

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-914	8 3 ST	84 ABBOTT C-D		43	20.945	1.181	5.64	17.800	24.100		
2020-914	8 3 ST	98 SYSMEX K-		16	19.500	0.365	1.87	16.600	22.400		
2020-914	9 3 D	33 COULTER JT		12	21.658	0.630	2.91	18.400	24.900		
2020-914	9 3 D	120 COULTER		16	20.520	0.425	2.07	17.400	23.600		
552 WHITE CELL COUNT(WBC)		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %							
2020-915	1 Sysmex CBC	98 SYSMEX K-		13	2.885	0.086	2.99	2.500	3.300		
2020-915	1 Sysmex CBC	127 SYSMEX XT		17	3.894	0.176	4.53	3.300	4.500		
2020-915	1 Sysmex CBC	132 SYSMEX XS		149	3.904	0.183	4.68	3.300	4.500		
2020-915	1 Sysmex CBC	141 SYSMEX XN		34	3.729	0.344	9.23	3.200	4.300		
2020-915	1 Sysmex CBC	152 Sysmex XN		39	3.168	0.199	6.28	2.700	3.600		
2020-915	2 ACT PENTRA	125 COULTER AC		153	2.888	0.118	4.07	2.500	3.300		
2020-915	2 ACT PENTRA	130 ABX PENTRA		25	2.844	0.150	5.27	2.400	3.300		
2020-915	3 ADVIA 120	116 ADVIA-120		18	3.194	0.113	3.53	2.700	3.700		
2020-915	4 AB-CELL DY	115 ABBOTT C-D		123	2.900	0.136	4.69	2.500	3.300		
2020-915	4 AB-CELL DY	149 BC DXH 500		14	2.079	0.115	5.51	1.800	2.400		
2020-915	5 BECTON D	137 B.C DXH		46	4.089	0.110	2.69	3.500	4.700		
2020-915	6 MR Mindray	145 BC-5390		28	3.419	0.072	2.11	2.900	3.900		
2020-915	8 3 ST	84 ABBOTT C-D		43	3.309	0.209	6.31	2.800	3.800		
2020-915	8 3 ST	98 SYSMEX K-		16	2.807	0.085	3.04	2.400	3.200		
2020-915	9 3 D	33 COULTER JT		12	2.042	0.119	5.82	1.700	2.300		
2020-915	9 3 D	120 COULTER		16	2.173	0.093	4.27	1.800	2.500		

SPEC: 20 HEMATOLOGY SUB: 70 GENERAL GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-911	1 SYSMEX CBC	39 SYSMEX K-		12	377.250	39.701	10.52	283.000	472.000		
2020-911	1 SYSMEX CBC	75 SYSMEX XT		17	193.765	11.259	5.81	145.000	242.000		
2020-911	1 SYSMEX CBC	79 SYSMEX XS		151	187.714	9.470	5.05	141.000	235.000		

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-911	1	SYSMEX CBC	87 SYSMEX XN	31	192.500	8.053	4.18	144.000	241.000		
2020-911	1	SYSMEX CBC	98 SYSMEX 350	39	192.000	9.315	4.85	144.000	240.000		
2020-911	2	ACT5 PENTR	73 COULTER AC	150	245.718	11.941	4.86	184.000	307.000		
2020-911	2	ACT5 PENTR	78 ABX PENTRA	25	238.400	9.243	3.88	179.000	298.000		
2020-911	3	ADVIA 120	64 ADVIA-120	18	219.588	9.146	4.17	165.000	274.000		
2020-911	4	AB- CELL-D	63 ABBOTT C-D	121	258.076	16.191	6.27	194.000	323.000		
2020-911	4	AB- CELL-D	94 BC DXH 500	15	245.400	11.758	4.79	184.000	307.000		
2020-911	5	BECTON D	83 B.C600,800	46	265.467	7.870	2.96	199.000	332.000		
2020-911	6	MR Mindray	91 BC-5390	28	251.036	10.490	4.18	188.000	314.000		
2020-911	8	3 ST	31 ABBOTT C-D	42	391.405	31.498	8.05	294.000	489.000		
2020-911	8	3 ST	39 SYSMEX K-	16	403.467	17.750	4.40	303.000	504.000		
2020-911	9	3 D	50 COULT.JT,M	11	612.636	24.005	3.92	459.000	766.000		
2020-911	9	3 D	68 COULTERACT	16	594.667	28.217	4.75	446.000	743.000		
553 PLATELET COUNT		CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-912	1	SYSMEX CBC	39 SYSMEX K-	12	62.000	5.759	9.29	47.000	78.000		
2020-912	1	SYSMEX CBC	75 SYSMEX XT	17	55.059	3.134	5.69	41.000	69.000		
2020-912	1	SYSMEX CBC	79 SYSMEX XS	151	50.813	4.538	8.93	38.000	64.000		
2020-912	1	SYSMEX CBC	87 SYSMEX XN	31	45.581	3.670	8.05	34.000	57.000		
2020-912	1	SYSMEX CBC	98 SYSMEX 350	39	46.872	3.625	7.73	35.000	59.000		
2020-912	2	ACT5 PENTR	73 COULTER AC	150	243.691	12.336	5.06	183.000	305.000		
2020-912	2	ACT5 PENTR	78 ABX PENTRA	25	240.600	11.168	4.64	180.000	301.000		
2020-912	3	ADVIA 120	64 ADVIA-120	18	217.294	8.484	3.90	163.000	272.000		
2020-912	4	AB- CELL-D	63 ABBOTT C-D	121	79.235	6.178	7.80	59.000	99.000		
2020-912	4	AB- CELL-D	94 BC DXH 500	15	84.667	4.894	5.78	64.000	106.000		
2020-912	5	BECTON D	83 B.C600,800	46	75.022	2.231	2.97	56.000	94.000		
2020-912	6	MR Mindray	91 BC-5390	28	56.000	2.699	4.82	42.000	70.000		
2020-912	8	3 ST	31 ABBOTT C-D	42	64.929	6.635	10.22	49.000	81.000		
2020-912	8	3 ST	39 SYSMEX K-	16	61.375	5.957	9.71	46.000	77.000		
2020-912	9	3 D	50 COULT.JT,M	11	70.364	4.333	6.16	53.000	88.000		
2020-912	9	3 D	68 COULTERACT	16	69.867	5.084	7.28	52.000	87.000		

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS			LABS	VALUE	DEVIATION	CV	LOW	HIGH
553	PLATELET COUNT	CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-913	1	SYSMEX CBC	39	SYSMEX K-		12	190.250	16.432	8.64	143.000	238.000
2020-913	1	SYSMEX CBC	75	SYSMEX XT		17	414.059	24.537	5.93	311.000	518.000
2020-913	1	SYSMEX CBC	79	SYSMEX XS		151	387.727	15.503	4.00	291.000	485.000
2020-913	1	SYSMEX CBC	87	SYSMEX XN		31	414.000	16.125	3.89	311.000	518.000
2020-913	1	SYSMEX CBC	98	SYSMEX 350		39	415.026	15.440	3.72	311.000	519.000
2020-913	2	ACT5 PENTR	73	COULTER AC		150	66.413	4.545	6.84	50.000	83.000
2020-913	2	ACT5 PENTR	78	ABX PENTRA		25	68.720	3.704	5.39	52.000	86.000
2020-913	3	ADVIA 120	64	ADVIA-120		18	54.111	5.120	9.46	41.000	68.000
2020-913	4	AB- CELL-D	63	ABBOTT C-D		121	487.275	27.949	5.74	365.000	609.000
2020-913	4	AB- CELL-D	94	BC DXH 500		15	476.133	9.244	1.94	357.000	595.000
2020-913	5	BECTON D	83	B.C600,800		46	469.804	24.641	5.25	352.000	587.000
2020-913	6	MR Mindray	91	BC-5390		28	501.357	16.815	3.35	376.000	627.000
2020-913	8	3 ST	31	ABBOTT C-D		42	196.786	18.122	9.21	148.000	246.000
2020-913	8	3 ST	39	SYSMEX K-		16	200.933	5.698	2.84	151.000	251.000
2020-913	9	3 D	50	COULT.JT,M		11	281.545	11.040	3.92	211.000	352.000
2020-913	9	3 D	68	COULTERAcT		16	277.667	15.239	5.49	208.000	347.000
2020-914	1	SYSMEX CBC	39	SYSMEX K-		12	375.750	38.880	10.35	282.000	470.000
2020-914	1	SYSMEX CBC	75	SYSMEX XT		17	194.294	11.513	5.93	146.000	243.000
2020-914	1	SYSMEX CBC	79	SYSMEX XS		151	184.356	13.479	7.31	138.000	230.000
2020-914	1	SYSMEX CBC	87	SYSMEX XN		31	191.200	6.695	3.50	143.000	239.000
2020-914	1	SYSMEX CBC	98	SYSMEX 350		39	191.026	8.766	4.59	143.000	239.000
2020-914	2	ACT5 PENTR	73	COULTER AC		150	483.772	19.780	4.09	363.000	605.000
2020-914	2	ACT5 PENTR	78	ABX PENTRA		25	478.160	20.807	4.35	359.000	598.000
2020-914	3	ADVIA 120	64	ADVIA-120		18	435.000	27.166	6.25	326.000	544.000
2020-914	4	AB- CELL-D	63	ABBOTT C-D		121	255.378	16.025	6.27	192.000	319.000
2020-914	4	AB- CELL-D	94	BC DXH 500		15	243.267	9.862	4.05	182.000	304.000
2020-914	5	BECTON D	83	B.C600,800		46	262.848	9.000	3.42	197.000	329.000
2020-914	6	MR Mindray	91	BC-5390		28	254.036	10.108	3.98	191.000	318.000

### Participant Summary

MONTH 11 OF YEAR 2020

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 134 PLATELET COUNT NEW							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-914	8 3 ST	31 ABBOTT C-D		42	385.119	26.861	6.97	289.000	481.000		
2020-914	8 3 ST	39 SYSMEX K-		16	398.667	13.415	3.36	299.000	498.000		
2020-914	9 3 D	50 COULT.JT,M		11	598.545	31.930	5.33	449.000	748.000		
2020-914	9 3 D	68 COULTERACT		16	599.533	31.494	5.25	450.000	749.000		
553 PLATELET COUNT		CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-915	1 SYSMEX CBC	39 SYSMEX K-		12	61.250	3.370	5.50	46.000	77.000		
2020-915	1 SYSMEX CBC	75 SYSMEX XT		17	55.176	3.347	6.07	41.000	69.000		
2020-915	1 SYSMEX CBC	79 SYSMEX XS		151	50.707	4.506	8.89	38.000	63.000		
2020-915	1 SYSMEX CBC	87 SYSMEX XN		31	47.032	3.188	6.78	35.000	59.000		
2020-915	1 SYSMEX CBC	98 SYSMEX 350		39	47.359	3.655	7.72	36.000	59.000		
2020-915	2 ACT5 PENTR	73 COULTER AC		150	66.369	4.603	6.94	50.000	83.000		
2020-915	2 ACT5 PENTR	78 ABX PENTRA		25	66.880	3.680	5.50	50.000	84.000		
2020-915	3 ADVIA 120	64 ADVIA-120		18	55.944	5.223	9.34	42.000	70.000		
2020-915	4 AB- CELL-D	63 ABBOTT C-D		121	78.286	5.990	7.65	59.000	98.000		
2020-915	4 AB- CELL-D	94 BC DXH 500		15	83.533	5.084	6.09	63.000	104.000		
2020-915	5 BECTON D	83 B.C600,800		46	74.826	2.792	3.73	56.000	94.000		
2020-915	6 MR Mindray	91 BC-5390		28	56.148	2.352	4.19	42.000	70.000		
2020-915	8 3 ST	31 ABBOTT C-D		42	65.381	6.883	10.53	49.000	82.000		
2020-915	8 3 ST	39 SYSMEX K-		16	63.313	6.162	9.73	47.000	79.000		
2020-915	9 3 D	50 COULT.JT,M		11	71.909	3.204	4.46	54.000	90.000		
2020-915	9 3 D	68 COULTERACT		16	68.333	5.211	7.63	51.000	85.000		

SPEC: 20 HEMATOLOGY		SUB: 75 BODY FLUID		GRP: 145 BODY FLUID CELL COU.							
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
159 BF RED CELL COUNT		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {GREATER}								
2020-951				47	0.761	0.889	116.86	0	3.000		
2020-952				47	1059.978	96.520	9.11	770.000	1350.000		

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPEC: 20 HEMATOLOGY		SUB: 75 BODY FLUID		GRP: 145 BODY FLUID CELL COU.						
ANALYTE	SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
159	BF RED CELL COUNT		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3		SD OR 25 %	{GREATER}		
	2020-953				47	0.804	1.191	148.06	0	4.000
	2020-954				47	0.652	0.865	132.66	0	3.000
	2020-955				47	1058.826	96.200	9.09	770.000	1347.000
160	BF WHITE CELL COUNT									
	2020-951				46	5.478	1.874	34.20	0.000	11.000
	2020-952				46	319.444	26.035	8.15	240.000	399.000
	2020-953				46	5.644	2.358	41.78	0	13.000
	2020-954				46	5.717	2.018	35.30	0.000	12.000
	2020-955				46	327.413	26.433	8.07	246.000	409.000

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH		GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
083 ABO GROUP											
2020-961				005 007	005 007		005 - AB	91	87	90	100
2020-961				005 007	005 007		006 - A1B	91	3	90	100
2020-961				005 007	005 007		008 - O	91	1	90	0
2020-962				001 003	001 003		002 - A1	91	4	91	100
2020-962				001 003	001 003		001 - A	91	87	91	100

**Participant Summary**

**MONTH 11 OF YEAR 2020**

SPE: 30 IMMUNOHEMATOLOGY		SUB: 80 ABO-RH				GRP: 150 ABO-RH							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
083 ABO GROUP													
2020-963				004 004	004 004		004 - B	91	91	91	100		
2020-964				001 003	001 003		002 - A1	91	4	91	100		
2020-964				001 003	001 003		001 - A	91	87	91	100		
2020-965				008 008	008 008		008 - O	91	91	91	100		
126 RH TYPE													
2020-961				001 003	001 003		001 - Rh+	91	91	91	100		
2020-962				001 003	001 003		001 - Rh+	91	91	91	100		
2020-963				001 003	001 003		001 - Rh+	91	91	91	100		
2020-964				003 004	003 004		004 - Rh-	91	79	91	100		
2020-964				003 004	003 004		003 - Rh-Du np	91	12	91	100		
2020-965				003 004	003 004		002 - Rh+ Du va.	91	4	87	0		
2020-965				003 004	003 004		003 - Rh-Du np	91	12	87	100		
2020-965				003 004	003 004		004 - Rh-	91	75	87	100		

### Participant Summary

MONTH 11 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFeree LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2020-971				001 001	001 001		001 - NEGATIVE	68	68	68	100
2020-972				001 001	001 001		001 - NEGATIVE	68	68	68	100
2020-973					002 002		002 - POSITIVE	68	61	61	100
2020-973					002 002		001 - NEGATIVE	68	7	61	0
2020-974				002 002	002 002		002 - POSITIVE	68	68	68	100
2020-975				001 001	001 001		001 - NEGATIVE	68	68	68	100
179 DIRECT ANTIG(COOMBS)											
SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFeree LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-651	001 Poly AHG	001 Gel				002 004	003 - 3+	13	5	13	100
2020-651	001 Poly AHG	001 Gel				002 004	004 - 4+	13	8	13	100
2020-651	001 Poly AHG	002 Tubo				002 004	003 - 3+	29	5	29	100
2020-651	001 Poly AHG	002 Tubo				002 004	002 - 2+	29	23	29	100
2020-651	001 Poly AHG	002 Tubo				002 004	004 - 4+	29	1	29	100
2020-651	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	15	12	15	100
2020-651	002 MnoAnt-IgG	002 Tubo		002 004	002 004		003 - 3+	15	3	15	100
2020-652	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	13	13	13	100
2020-652	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	30	30	30	100

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-652	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	15	15	15	100
2020-652	003 Anti-C3b/d	002 Tubo		001 001	001 001	001 001	001 - Negativo	10	10	10	100
179 DIRECT ANTIG(COOMBS)											
2020-653	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	12	12	12	100
2020-653	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	30	30	30	100
2020-653	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	10	10	10	100
2020-653	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	15	15	15	100
2020-653	003 Anti-C3b/d	002 Tubo		001 001	001 001	001 001	001 - Negativo	10	10	10	100
2020-654	001 Poly AHG	001 Gel		001 001	001 001		001 - Negativo	13	13	13	100
2020-654	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	30	30	30	100
2020-654	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	15	15	15	100
2020-654	003 Anti-C3b/d	002 Tubo		001 001	001 001	001 001	001 - Negativo	10	10	10	100
2020-655	001 Poly AHG	001 Gel				002 004	004 - 4+	13	7	13	100
2020-655	001 Poly AHG	001 Gel				002 004	003 - 3+	13	6	13	100
2020-655	001 Poly AHG	002 Tubo				002 004	004 - 4+	30	1	30	100
2020-655	001 Poly AHG	002 Tubo				002 004	003 - 3+	30	9	30	100
2020-655	001 Poly AHG	002 Tubo				002 004	002 - 2+	30	20	30	100
2020-655	002 MnoAnt-IgG	002 Tubo		002 004	002 004		003 - 3+	14	1	14	100
2020-655	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	14	13	14	100



### Participant Summary

MONTH 11 OF YEAR 2020

SPE: 30 IMMUNOHEMATOLOGY			SUB: 100 COMPATIBILITY TESTING				GRP: 170 COMPATIBILITY TEST.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
085 COMPATIBIL. TESTING											
2020-981				001 001	001 001		001 - COMPATIBLE	41	41	41	100
2020-982				001 001	001 001		001 - COMPATIBLE	41	40	40	100
2020-982				001 001	001 001		002 - INCOMPATIB	41	1	40	0
2020-983				002 002	002 002		002 - INCOMPATIB	41	41	41	100
2020-984				001 001	001 001		001 - COMPATIBLE	41	41	41	100
2020-985				002 002	002 002		002 - INCOMPATIB	41	41	41	100
SPE: 30 IMMUNOHEMATOLOGY			SUB: 110 ANTIBODY IDENTIFICATION				GRP: 180 ANTIBODY IDENT.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
086 ANTIBODY IDENT.TEST.											
2020-984					003 003		003 - Anti D	18	9	18	100
2020-984					005 005		005 - Anti C	18	9	18	100

**Participant Summary**  
**MONTH 11 OF YEAR 2020**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 230 HBsAG, ANTI HBC, HIV					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
093 ANTI-HUMAN IMM. (HIV)												
2020-991	055 ALERE DETE			001 001	001 001		001 - NEG/NR	30	30	30	100	
2020-992	055 ALERE DETE			002 002	002 002		002 - POS/R	30	30	30	100	
2020-993	055 ALERE DETE			001 001	001 001		001 - NEG/NR	30	30	30	100	
2020-994	055 ALERE DETE			001 001	001 001		001 - NEG/NR	30	30	30	100	
2020-995	055 ALERE DETE			002 002	002 002		002 - POS/R	30	30	30	100	