

Participant Summary
MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY SUB: 10 ROUTINE GRP: 10 ROUTINE

ANALYTE				NO. PART		TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
001 ALAN AMINO ALT/SGPT		CLIA TEST ID #: 255	EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-781	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	18	250.778	9.710	3.87	201.000	301.000	
2020-781	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	248.500	9.097	3.66	199.000	298.000	
2020-781	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	76	264.827	7.504	2.83	212.000	318.000	
2020-781	11 MOD. IFCC	184 INTEGRA	93 ROCHE	29	243.966	10.196	4.18	195.000	293.000	
2020-781	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	26	223.560	7.890	3.53	179.000	268.000	
2020-781	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	271.345	11.663	4.30	217.000	326.000	
2020-781	11 MOD. IFCC	203 ROCHE	93 ROCHE	27	245.148	8.258	3.37	196.000	294.000	
2020-781	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	10	263.200	8.085	3.07	211.000	316.000	
2020-781	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	252.700	12.877	5.10	202.000	303.000	
2020-781	13 ALA &aKETO	165 VITROS	155 ALTV	126	230.863	5.497	2.38	185.000	277.000	
2020-782	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	18	111.944	3.937	3.52	90.000	134.000	
2020-782	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	114.833	1.908	1.66	92.000	138.000	
2020-782	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	76	120.895	3.633	3.01	97.000	145.000	
2020-782	11 MOD. IFCC	184 INTEGRA	93 ROCHE	29	108.679	3.576	3.29	87.000	130.000	
2020-782	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	26	97.769	3.400	3.48	78.000	117.000	
2020-782	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	123.517	5.923	4.79	99.000	148.000	
2020-782	11 MOD. IFCC	203 ROCHE	93 ROCHE	27	109.148	3.903	3.58	87.000	131.000	
2020-782	11 MOD. IFCC	212 S. D. EXL	143 SIEMENS	10	120.600	4.432	3.67	96.000	145.000	
2020-782	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	117.000	3.975	3.40	94.000	140.000	
2020-782	13 ALA &aKETO	165 VITROS	155 ALTV	126	105.176	2.561	2.44	84.000	126.000	
2020-783	5 ENZ. 37 C	197 ARCHITECT	1 ABBOTT	18	288.778	10.385	3.60	231.000	347.000	
2020-783	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	282.417	8.751	3.10	226.000	339.000	
2020-783	11 MOD. IFCC	38 SIEMENS DI	143 SIEMENS	76	302.895	7.476	2.47	242.000	363.000	
2020-783	11 MOD. IFCC	184 INTEGRA	93 ROCHE	29	278.310	11.195	4.02	223.000	334.000	
2020-783	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	26	254.160	5.357	2.11	203.000	305.000	
2020-783	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	310.267	14.021	4.52	248.000	372.000	
2020-783	11 MOD. IFCC	203 ROCHE	93 ROCHE	27	282.231	7.752	2.75	226.000	339.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-783	11 MOD. IFCC	212	S. D. EXL	143	SIEMENS	10	298.000	24.795	8.32	238.000	358.000		
2020-783	11 MOD. IFCC	214	Envoy 500	150	ELITECH	10	288.500	19.200	6.66	231.000	346.000		
2020-783	13 ALA &aKETO	165	VITROS	155	ALTV	126	263.520	5.665	2.15	211.000	316.000		
001	ALAN AMINO ALT/SGPT	CLIA TEST ID #: 255		EVALUATION CRITERIA: TARGET VALUE 20 %									
2020-784	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	18	358.556	13.574	3.79	287.000	430.000		
2020-784	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	343.417	5.866	1.71	275.000	412.000		
2020-784	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	76	372.541	9.743	2.62	298.000	447.000		
2020-784	11 MOD. IFCC	184	INTEGRA	93	ROCHE	29	341.964	10.831	3.17	274.000	410.000		
2020-784	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	312.577	8.418	2.69	250.000	375.000		
2020-784	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	30	379.600	16.687	4.40	304.000	456.000		
2020-784	11 MOD. IFCC	203	ROCHE	93	ROCHE	27	348.630	13.348	3.83	279.000	418.000		
2020-784	11 MOD. IFCC	212	S. D. EXL	143	SIEMENS	10	372.800	14.905	4.00	298.000	447.000		
2020-784	11 MOD. IFCC	214	Envoy 500	150	ELITECH	10	358.300	24.129	6.73	287.000	430.000		
2020-784	13 ALA &aKETO	165	VITROS	155	ALTV	126	324.096	8.362	2.58	259.000	389.000		
2020-785	5 ENZ. 37 C	197	ARCHITECT	1	ABBOTT	18	172.222	6.562	3.81	138.000	207.000		
2020-785	10 KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	172.500	3.428	1.99	138.000	207.000		
2020-785	11 MOD. IFCC	38	SIEMENS DI	143	SIEMENS	76	184.200	4.540	2.46	147.000	221.000		
2020-785	11 MOD. IFCC	184	INTEGRA	93	ROCHE	29	166.893	4.945	2.96	134.000	200.000		
2020-785	11 MOD. IFCC	186	BECK. COUL	76	AU-BECKMAN	26	152.720	3.231	2.12	122.000	183.000		
2020-785	11 MOD. IFCC	191	RX-DAYTONA	126	RANDOX	30	190.267	10.630	5.59	152.000	228.000		
2020-785	11 MOD. IFCC	203	ROCHE	93	ROCHE	27	168.889	6.172	3.65	135.000	203.000		
2020-785	11 MOD. IFCC	212	S. D. EXL	143	SIEMENS	10	185.500	6.037	3.25	148.000	223.000		
2020-785	11 MOD. IFCC	214	Envoy 500	150	ELITECH	10	181.500	5.869	3.23	145.000	218.000		
2020-785	13 ALA &aKETO	165	VITROS	155	ALTV	126	159.240	3.647	2.29	127.000	191.000		
002	ALBUMIN	CLIA TEST ID #: 265		EVALUATION CRITERIA: TARGET VALUE 10 %									
2020-781	3 DYE BI-B.G	165	VITROS	63	KODAK	146	3.968	0.136	3.43	3.600	4.400		
2020-781	3 DYE BI-B.G	184	INTEGRA	93	ROCHE	35	4.080	0.104	2.54	3.700	4.500		
2020-781	3 DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	26	3.908	0.106	2.70	3.500	4.300		
2020-781	3 DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	3.963	0.114	2.88	3.600	4.400		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-781	3	DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	3.983	0.134	3.37	3.600	4.400
2020-781	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	20	3.890	0.104	2.68	3.500	4.300
2020-781	3	DYE BI-B.G	199	Rx-IMOLA	126	RANDOX	12	4.008	0.064	1.60	3.600	4.400
2020-781	3	DYE BI-B.G	203	ROCHE	93	ROCHE	33	4.106	0.090	2.19	3.700	4.500
2020-781	3	DYE BI-B.G	214	Envoy 500	150	ELITECH	10	4.120	0.154	3.73	3.700	4.500
2020-781	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	79	3.797	0.101	2.65	3.400	4.200
2020-781	4	DYE BI-B.P	212	S. D. EXL	143	SIEMENS	14	3.779	0.161	4.26	3.400	4.200
2020-781	4	DYE BI-B.P	215	B.C DXC	134	BECKM.COUL	11	3.909	0.079	2.03	3.500	4.300
002 ALBUMIN			CLIA TEST ID #: 265		EVALUATION CRITERIA: TARGET VALUE 10 %							
2020-782	3	DYE BI-B.G	165	VITROS	63	KODAK	146	2.581	0.079	3.06	2.300	2.800
2020-782	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	35	2.771	0.082	2.97	2.500	3.000
2020-782	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	26	2.604	0.077	2.97	2.300	2.900
2020-782	3	DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	2.750	0.076	2.78	2.500	3.000
2020-782	3	DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	2.719	0.096	3.55	2.400	3.000
2020-782	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	20	2.640	0.080	3.03	2.400	2.900
2020-782	3	DYE BI-B.G	199	Rx-IMOLA	126	RANDOX	12	2.758	0.064	2.32	2.500	3.000
2020-782	3	DYE BI-B.G	203	ROCHE	93	ROCHE	33	2.803	0.081	2.89	2.500	3.100
2020-782	3	DYE BI-B.G	214	Envoy 500	150	ELITECH	10	2.840	0.092	3.23	2.600	3.100
2020-782	4	DYE BI-B.P	38	SIEMENS DI	143	SIEMENS	79	2.497	0.076	3.05	2.200	2.700
2020-782	4	DYE BI-B.P	212	S. D. EXL	143	SIEMENS	14	2.486	0.125	5.01	2.200	2.700
2020-782	4	DYE BI-B.P	215	B.C DXC	134	BECKM.COUL	11	2.591	0.090	3.47	2.300	2.900
2020-783	3	DYE BI-B.G	165	VITROS	63	KODAK	146	4.346	0.128	2.94	3.900	4.800
2020-783	3	DYE BI-B.G	184	INTEGRA	93	ROCHE	35	4.386	0.125	2.84	3.900	4.800
2020-783	3	DYE BI-B.G	186	BECK. COUL	76	AU-BECKMAN	26	4.215	0.117	2.77	3.800	4.600
2020-783	3	DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	4.280	0.114	2.66	3.900	4.700
2020-783	3	DYE BI-B.G	191	RX-DAYTONA	126	RANDOX	31	4.250	0.126	2.96	3.800	4.700
2020-783	3	DYE BI-B.G	197	ARCHITECT	1	ABBOTT	20	4.210	0.104	2.48	3.800	4.600
2020-783	3	DYE BI-B.G	199	Rx-IMOLA	126	RANDOX	12	4.342	0.095	2.20	3.900	4.800
2020-783	3	DYE BI-B.G	203	ROCHE	93	ROCHE	33	4.425	0.106	2.40	4.000	4.900

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-783	3 DYE BI-B.G	214 Envoy 500	150 ELITECH		10	4.430	0.200	4.52	4.000	4.900	
2020-783	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		79	4.131	0.106	2.57	3.700	4.500	
2020-783	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		14	4.129	0.153	3.71	3.700	4.500	
2020-783	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL		11	4.255	0.089	2.09	3.800	4.700	
002 ALBUMIN		CLIA TEST ID #: 265	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-784	3 DYE BI-B.G	165 VITROS	63 KODAK		146	5.058	0.170	3.35	4.600	5.600	
2020-784	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		35	4.957	0.123	2.47	4.500	5.500	
2020-784	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		26	4.812	0.165	3.43	4.300	5.300	
2020-784	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		31	4.867	0.125	2.56	4.400	5.400	
2020-784	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		31	4.787	0.123	2.57	4.300	5.300	
2020-784	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		20	4.780	0.093	1.94	4.300	5.300	
2020-784	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX		12	4.933	0.094	1.91	4.400	5.400	
2020-784	3 DYE BI-B.G	203 ROCHE	93 ROCHE		33	5.016	0.097	1.94	4.500	5.500	
2020-784	3 DYE BI-B.G	214 Envoy 500	150 ELITECH		10	5.030	0.149	2.96	4.500	5.500	
2020-784	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		79	4.714	0.149	3.16	4.200	5.200	
2020-784	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		14	4.729	0.183	3.87	4.300	5.200	
2020-784	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL		11	4.864	0.088	1.81	4.400	5.400	
2020-785	3 DYE BI-B.G	165 VITROS	63 KODAK		146	3.203	0.105	3.29	2.900	3.500	
2020-785	3 DYE BI-B.G	184 INTEGRA	93 ROCHE		35	3.354	0.091	2.70	3.000	3.700	
2020-785	3 DYE BI-B.G	186 BECK. COUL	76 AU-BECKMAN		26	3.165	0.092	2.90	2.800	3.500	
2020-785	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		31	3.280	0.065	1.99	3.000	3.600	
2020-785	3 DYE BI-B.G	191 RX-DAYTONA	126 RANDOX		31	3.267	0.094	2.89	2.900	3.600	
2020-785	3 DYE BI-B.G	197 ARCHITECT	1 ABBOTT		20	3.200	0.077	2.42	2.900	3.500	
2020-785	3 DYE BI-B.G	199 Rx-IMOLA	126 RANDOX		12	3.325	0.072	2.17	3.000	3.700	
2020-785	3 DYE BI-B.G	203 ROCHE	93 ROCHE		33	3.388	0.082	2.42	3.000	3.700	
2020-785	3 DYE BI-B.G	214 Envoy 500	150 ELITECH		10	3.440	0.102	2.96	3.100	3.800	
2020-785	4 DYE BI-B.P	38 SIEMENS DI	143 SIEMENS		79	3.055	0.068	2.21	2.700	3.400	
2020-785	4 DYE BI-B.P	212 S. D. EXL	143 SIEMENS		14	3.064	0.117	3.82	2.800	3.400	
2020-785	4 DYE BI-B.P	215 B.C DXC	134 BECKM.COUL		11	3.191	0.079	2.48	2.900	3.500	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275	EVALUATION CRITERIA: TARGET VALUE	30 %							
2020-781	3	COLORIMET.	184 INTEGRA	93 ROCHE	31	201.333	10.879	5.40	141.000	262.000	
2020-781	3	COLORIMET.	203 ROCHE	93 ROCHE	24	190.739	5.581	2.93	134.000	248.000	
2020-781	6	KINETIC 37	165 VITROS	63 KODAK	146	157.814	8.097	5.13	110.000	205.000	
2020-781	6	KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	209.200	24.310	11.62	146.000	272.000	
2020-781	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	167.250	11.555	6.91	117.000	217.000	
2020-781	8	BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	170.481	19.215	11.27	119.000	222.000	
2020-781	9	ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	25	175.880	14.503	8.25	123.000	229.000	
2020-781	11	Modified I	38 SIEMENS DI	143 SIEMENS	75	207.013	7.788	3.76	145.000	269.000	
2020-781	12	P NITRO PH	197 ARCHITECT	1 ABBOTT	11	194.000	11.045	5.69	136.000	252.000	
2020-782	3	COLORIMET.	184 INTEGRA	93 ROCHE	31	93.800	5.009	5.34	66.000	122.000	
2020-782	3	COLORIMET.	203 ROCHE	93 ROCHE	24	89.208	3.391	3.80	62.000	116.000	
2020-782	6	KINETIC 37	165 VITROS	63 KODAK	146	99.152	5.246	5.29	69.000	129.000	
2020-782	6	KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	97.700	13.914	14.24	68.000	127.000	
2020-782	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	79.667	4.460	5.60	56.000	104.000	
2020-782	8	BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	79.259	9.567	12.07	55.000	103.000	
2020-782	9	ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	25	83.640	7.076	8.46	59.000	109.000	
2020-782	11	Modified I	38 SIEMENS DI	143 SIEMENS	75	95.827	3.931	4.10	67.000	125.000	
2020-782	12	P NITRO PH	197 ARCHITECT	1 ABBOTT	11	90.727	4.634	5.11	64.000	118.000	
2020-783	3	COLORIMET.	184 INTEGRA	93 ROCHE	31	227.900	10.429	4.58	160.000	296.000	
2020-783	3	COLORIMET.	203 ROCHE	93 ROCHE	24	218.167	7.840	3.59	153.000	284.000	
2020-783	6	KINETIC 37	165 VITROS	63 KODAK	146	165.772	8.842	5.33	116.000	216.000	
2020-783	6	KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	237.500	24.728	10.41	166.000	309.000	
2020-783	6	KINETIC 37	215 B.C DXC	134 BECKM.COUL	12	195.583	13.200	6.75	137.000	254.000	
2020-783	8	BOW. MC-37	186 BECK. COUL	76 AU-BECKMAN	27	193.741	23.202	11.98	136.000	252.000	
2020-783	9	ENZYMAT.37	191 RX-DAYTONA	126 RANDOX	25	201.000	16.361	8.14	141.000	261.000	
2020-783	11	Modified I	38 SIEMENS DI	143 SIEMENS	75	238.297	8.376	3.51	167.000	310.000	
2020-783	12	P NITRO PH	197 ARCHITECT	1 ABBOTT	11	221.909	10.475	4.72	155.000	288.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
003	ALKALINE PHOSPHATE	CLIA TEST ID #: 275				EVALUATION CRITERIA: TARGET VALUE 30 %						
2020-784	3	COLORIMET.	184	INTEGRA	93	ROCHE	31	284.567	12.352	4.34	199.000	370.000
2020-784	3	COLORIMET.	203	ROCHE	93	ROCHE	24	269.375	11.135	4.13	189.000	350.000
2020-784	6	KINETIC 37	165	VITROS	63	KODAK	146	178.655	9.788	5.48	125.000	232.000
2020-784	6	KINETIC 37	194	BIOLIS-24S	154	(IDG)	10	294.600	27.347	9.28	206.000	383.000
2020-784	6	KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	241.583	10.428	4.32	169.000	314.000
2020-784	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	27	240.889	28.148	11.69	169.000	313.000
2020-784	9	ENZYMAT.37	191	RX-DAYTONA	126	RANDOX	25	243.200	19.360	7.96	170.000	316.000
2020-784	11	Modified I	38	SIEMENS DI	143	SIEMENS	75	296.486	9.644	3.25	208.000	385.000
2020-784	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	11	274.091	14.055	5.13	192.000	356.000
2020-785	3	COLORIMET.	184	INTEGRA	93	ROCHE	31	140.267	6.572	4.69	98.000	182.000
2020-785	3	COLORIMET.	203	ROCHE	93	ROCHE	24	133.083	5.530	4.15	93.000	173.000
2020-785	6	KINETIC 37	165	VITROS	63	KODAK	146	131.303	6.860	5.22	92.000	171.000
2020-785	6	KINETIC 37	194	BIOLIS-24S	154	(IDG)	10	144.100	18.436	12.79	101.000	187.000
2020-785	6	KINETIC 37	215	B.C DXC	134	BECKM.COUL	12	117.250	8.176	6.97	82.000	152.000
2020-785	8	BOW. MC-37	186	BECK. COUL	76	AU-BECKMAN	27	118.556	13.439	11.34	83.000	154.000
2020-785	9	ENZYMAT.37	191	RX-DAYTONA	126	RANDOX	25	124.320	9.358	7.53	87.000	162.000
2020-785	11	Modified I	38	SIEMENS DI	143	SIEMENS	75	144.587	5.205	3.60	101.000	188.000
2020-785	12	P NITRO PH	197	ARCHITECT	1	ABBOTT	11	136.091	7.403	5.44	95.000	177.000
004	AMYLASE	CLIA TEST ID #: 285										
2020-781	3	ENZYM. 37C	165	VITROS	63	KODAK	84	138.821	6.634	4.78	97.000	180.000
2020-781	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	12	190.667	3.448	1.81	133.000	248.000
2020-781	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	24	178.083	4.152	2.33	125.000	232.000
2020-781	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	11	182.455	2.189	1.20	128.000	237.000
2020-781	9	ENZ/COL.37	203	ROCHE	93	ROCHE	22	176.045	2.915	1.66	123.000	229.000
2020-781	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	141.118	9.292	6.58	99.000	183.000
2020-782	3	ENZYM. 37C	165	VITROS	63	KODAK	84	56.667	6.081	10.73	40.000	74.000
2020-782	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	12	89.833	2.511	2.80	63.000	117.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-782	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	24	83.375	1.775	2.13	58.000	108.000
2020-782	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	11	86.636	1.150	1.33	61.000	113.000
2020-782	9	ENZ/COL.37	203	ROCHE	93	ROCHE	22	82.909	1.505	1.82	58.000	108.000
2020-782	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	65.706	3.561	5.42	46.000	85.000
004	AMYLASE	CLIA TEST ID #: 285		EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-783	3	ENZYM. 37C	165	VITROS	63	KODAK	84	159.369	6.290	3.95	112.000	207.000
2020-783	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	12	219.083	9.060	4.14	153.000	285.000
2020-783	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	24	204.833	4.625	2.26	143.000	266.000
2020-783	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	11	206.909	3.801	1.84	145.000	269.000
2020-783	9	ENZ/COL.37	203	ROCHE	93	ROCHE	22	201.818	3.786	1.88	141.000	262.000
2020-783	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	160.353	10.128	6.32	112.000	208.000
2020-784	3	ENZYM. 37C	165	VITROS	63	KODAK	84	194.429	7.501	3.86	136.000	253.000
2020-784	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	12	269.333	4.871	1.81	189.000	350.000
2020-784	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	24	251.542	5.058	2.01	176.000	327.000
2020-784	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	11	254.000	5.135	2.02	178.000	330.000
2020-784	9	ENZ/COL.37	203	ROCHE	93	ROCHE	22	248.619	3.000	1.21	174.000	323.000
2020-784	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	196.471	12.472	6.35	138.000	255.000
2020-785	3	ENZYM. 37C	165	VITROS	63	KODAK	84	93.310	6.516	6.98	65.000	121.000
2020-785	3	ENZYM. 37C	215	B.C DXC	134	BECKM.COUL	12	134.083	3.883	2.90	94.000	174.000
2020-785	4	HYDRO ACT.	38	SIEMENS DI	143	SIEMENS	24	124.458	2.380	1.91	87.000	162.000
2020-785	9	ENZ/COL.37	184	INTEGRA	93	ROCHE	11	127.182	1.466	1.15	89.000	165.000
2020-785	9	ENZ/COL.37	203	ROCHE	93	ROCHE	22	123.238	1.925	1.56	86.000	160.000
2020-785	10	CNPG3	186	BECK. COUL	76	AU-BECKMAN	17	98.059	5.641	5.75	69.000	127.000
005	ASP. AMINO AST/SGOT	CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-781	1	ASPAR&KETO	165	VITROS	63	KODAK	139	340.420	11.274	3.31	272.000	409.000
2020-781	1	ASPAR&KETO	194	BIOLIS-24S	154	(IDG)	11	281.273	17.488	6.22	225.000	338.000
2020-781	5	ENZYM. 37C	38	SIEMENS DI	143	SIEMENS	80	277.088	9.305	3.36	222.000	333.000
2020-781	5	ENZYM. 37C	212	S. D. EXL	143	SIEMENS	14	277.786	8.308	2.99	222.000	333.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-781	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	268.727	4.614	1.72	215.000	322.000		
2020-781	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	279.700	9.477	3.39	224.000	336.000		
2020-781	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	27	242.000	9.190	3.80	194.000	290.000		
2020-781	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	303.733	17.208	5.67	243.000	364.000		
2020-781	11 MOD. IFCC	203 ROCHE	93 ROCHE	25	282.240	12.735	4.51	226.000	339.000		
2020-781	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	278.300	9.880	3.55	223.000	334.000		
2020-781	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	11	271.182	6.658	2.46	217.000	325.000		
005 ASP. AMINO AST/SGOT		CLIA TEST ID #: 295	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-782	1 ASPAR&KETO	165 VITROS	63 KODAK	139	148.059	4.098	2.77	118.000	178.000		
2020-782	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	11	142.364	6.892	4.84	114.000	171.000		
2020-782	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	134.025	4.170	3.11	107.000	161.000		
2020-782	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	14	134.929	3.011	2.23	108.000	162.000		
2020-782	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	131.818	3.298	2.50	105.000	158.000		
2020-782	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	136.967	5.030	3.67	110.000	164.000		
2020-782	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	27	118.926	5.824	4.90	95.000	143.000		
2020-782	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	148.500	9.677	6.52	119.000	178.000		
2020-782	11 MOD. IFCC	203 ROCHE	93 ROCHE	25	136.800	5.987	4.38	109.000	164.000		
2020-782	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	137.400	4.341	3.16	110.000	165.000		
2020-782	13 NADH wo P5	197 ARCHITECT	1 ABBOTT	11	132.636	2.672	2.01	106.000	159.000		
2020-783	1 ASPAR&KETO	165 VITROS	63 KODAK	139	396.906	13.484	3.40	318.000	476.000		
2020-783	1 ASPAR&KETO	194 BIOLIS-24S	154 (IDG)	11	315.545	20.843	6.61	252.000	379.000		
2020-783	5 ENZYMA.37C	38 SIEMENS DI	143 SIEMENS	80	315.725	9.303	2.95	253.000	379.000		
2020-783	5 ENZYMA.37C	212 S. D. EXL	143 SIEMENS	14	320.643	8.199	2.56	257.000	385.000		
2020-783	10 KINETIC 37	215 B.C DXC	134 BECKM.COUL	11	305.636	11.609	3.80	245.000	367.000		
2020-783	11 MOD. IFCC	184 INTEGRA	93 ROCHE	30	315.100	10.160	3.22	252.000	378.000		
2020-783	11 MOD. IFCC	186 BECK. COUL	76 AU-BECKMAN	27	273.462	10.778	3.94	219.000	328.000		
2020-783	11 MOD. IFCC	191 RX-DAYTONA	126 RANDOX	30	341.000	17.288	5.07	273.000	409.000		
2020-783	11 MOD. IFCC	203 ROCHE	93 ROCHE	25	320.292	9.834	3.07	256.000	384.000		
2020-783	11 MOD. IFCC	214 Envoy 500	150 ELITECH	10	318.500	16.536	5.19	255.000	382.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-783	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	11	307.727	6.326	2.06	246.000	369.000
005	ASP. AMINO AST/SGOT		CLIA TEST ID #: 295		EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-784	1	ASPAR&KETO		165	VITROS	63	KODAK	139	512.748	20.646	4.03	410.000	615.000
2020-784	1	ASPAR&KETO		194	BIOLIS-24S	154	(IDG)	11	365.636	21.141	5.78	293.000	439.000
2020-784	5	ENZYMA.37C		38	SIEMENS DI	143	SIEMENS	80	387.163	11.484	2.97	310.000	465.000
2020-784	5	ENZYMA.37C		212	S. D. EXL	143	SIEMENS	14	392.286	12.475	3.18	314.000	471.000
2020-784	10	KINETIC 37		215	B.C DXC	134	BECKM.COUL	11	372.000	13.212	3.55	298.000	446.000
2020-784	11	MOD. IFCC		184	INTEGRA	93	ROCHE	30	382.500	11.775	3.08	306.000	459.000
2020-784	11	MOD. IFCC		186	BECK. COUL	76	AU-BECKMAN	27	330.885	11.988	3.62	265.000	397.000
2020-784	11	MOD. IFCC		191	RX-DAYTONA	126	RANDOX	30	412.867	21.950	5.32	330.000	495.000
2020-784	11	MOD. IFCC		203	ROCHE	93	ROCHE	25	387.280	16.965	4.38	310.000	465.000
2020-784	11	MOD. IFCC		214	Envoy 500	150	ELITECH	10	385.700	13.878	3.60	309.000	463.000
2020-784	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	11	372.273	9.762	2.62	298.000	447.000
2020-785	1	ASPAR&KETO		165	VITROS	63	KODAK	139	227.234	6.843	3.01	182.000	273.000
2020-785	1	ASPAR&KETO		194	BIOLIS-24S	154	(IDG)	11	206.273	11.226	5.44	165.000	248.000
2020-785	5	ENZYMA.37C		38	SIEMENS DI	143	SIEMENS	80	197.438	5.733	2.90	158.000	237.000
2020-785	5	ENZYMA.37C		212	S. D. EXL	143	SIEMENS	14	200.214	4.988	2.49	160.000	240.000
2020-785	10	KINETIC 37		215	B.C DXC	134	BECKM.COUL	11	193.000	4.178	2.16	154.000	232.000
2020-785	11	MOD. IFCC		184	INTEGRA	93	ROCHE	30	198.767	5.743	2.89	159.000	239.000
2020-785	11	MOD. IFCC		186	BECK. COUL	76	AU-BECKMAN	27	173.815	8.411	4.84	139.000	209.000
2020-785	11	MOD. IFCC		191	RX-DAYTONA	126	RANDOX	30	217.000	12.391	5.71	174.000	260.000
2020-785	11	MOD. IFCC		203	ROCHE	93	ROCHE	25	200.708	6.374	3.18	161.000	241.000
2020-785	11	MOD. IFCC		214	Envoy 500	150	ELITECH	10	201.500	6.217	3.09	161.000	242.000
2020-785	13	NADH wo P5		197	ARCHITECT	1	ABBOTT	11	191.636	5.613	2.93	153.000	230.000
006	BILIRUBIN, TOTAL		CLIA TEST ID #: 305										
2020-781	4	J-G WITH B		38	SIEMENS DI	143	SIEMENS	79	3.490	0.158	4.53	2.790	4.190
2020-781	4	J-G WITH B		186	BECK. COUL	76	AU-BECKMAN	24	3.469	0.112	3.22	2.780	4.160
2020-781	4	J-G WITH B		191	RX-DAYTONA	126	RANDOX	22	3.813	0.140	3.66	3.050	4.580
2020-781	4	J-G WITH B		212	S. D. EXL	143	SIEMENS	11	3.554	0.184	5.18	2.840	4.260

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-781	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	3.928	0.130	3.32	3.140	4.710		
2020-781	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	3.364	0.184	5.47	2.690	4.040		
2020-781	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	3.467	0.205	5.93	2.770	4.160		
2020-781	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	3.178	0.129	4.05	2.540	3.810		
2020-781	13 AZOBIL.	203 ROCHE	93 ROCHE	23	3.100	0.092	2.96	2.480	3.720		
2020-781	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	11	3.132	0.163	5.20	2.510	3.760		
006 BILIRUBIN, TOTAL		CLIA TEST ID #: 305	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-782	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	1.411	0.069	4.85	1.130	1.690		
2020-782	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	24	1.484	0.049	3.31	1.190	1.780		
2020-782	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	22	1.555	0.134	8.61	1.240	1.870		
2020-782	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	11	1.455	0.099	6.80	1.160	1.750		
2020-782	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	1.738	0.118	6.79	1.390	2.090		
2020-782	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	1.367	0.047	3.45	1.090	1.640		
2020-782	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	1.349	0.098	7.30	1.080	1.620		
2020-782	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	1.304	0.056	4.26	1.040	1.560		
2020-782	13 AZOBIL.	203 ROCHE	93 ROCHE	23	1.263	0.044	3.45	1.010	1.520		
2020-782	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	11	1.301	0.111	8.50	1.040	1.560		
2020-783	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	4.054	0.173	4.26	3.240	4.860		
2020-783	4 J-G WITH B	186 BECK. COUL	76 AU-BECKMAN	24	3.978	0.122	3.07	3.180	4.770		
2020-783	4 J-G WITH B	191 RX-DAYTONA	126 RANDOX	22	4.456	0.192	4.32	3.560	5.350		
2020-783	4 J-G WITH B	212 S. D. EXL	143 SIEMENS	11	4.159	0.209	5.04	3.330	4.990		
2020-783	5 J-G NO BK.	215 B.C DXC	134 BECKM.COUL	12	4.538	0.174	3.84	3.630	5.450		
2020-783	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	3.886	0.196	5.05	3.110	4.660		
2020-783	7 DIAZO-DYPH	165 VITROS	63 KODAK	143	3.867	0.205	5.31	3.090	4.640		
2020-783	13 AZOBIL.	184 INTEGRA	93 ROCHE	18	3.652	0.136	3.72	2.920	4.380		
2020-783	13 AZOBIL.	203 ROCHE	93 ROCHE	23	3.593	0.123	3.41	2.870	4.310		
2020-783	19 DIAZO W BL	197 ARCHITECT	1 ABBOTT	11	3.645	0.156	4.27	2.920	4.370		
2020-784	4 J-G WITH B	38 SIEMENS DI	143 SIEMENS	79	5.145	0.229	4.46	4.120	6.170		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-784	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	24	4.944	0.115	2.33	3.960	5.930
2020-784	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	22	5.695	0.264	4.63	4.560	6.830
2020-784	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	11	5.218	0.223	4.27	4.170	6.260
2020-784	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	5.639	0.175	3.10	4.510	6.770
2020-784	7	DIAZO-DYPH	165	VITROS	63	KODAK	143	4.967	0.205	4.14	3.970	5.960
2020-784	7	DIAZO-DYPH	165	VITROS	63	KODAK	143	4.985	0.260	5.23	3.990	5.980
2020-784	13	AZOBIL.	184	INTEGRA	93	ROCHE	18	4.602	0.177	3.86	3.680	5.520
2020-784	13	AZOBIL.	203	ROCHE	93	ROCHE	23	4.578	0.129	2.81	3.660	5.490
2020-784	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	11	4.582	0.240	5.23	3.670	5.500
006 BILIRUBIN, TOTAL			CLIA TEST ID #: 305			EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-785	4	J-G WITH B	38	SIEMENS DI	143	SIEMENS	79	2.306	0.117	5.05	1.850	2.770
2020-785	4	J-G WITH B	186	BECK. COUL	76	AU-BECKMAN	24	2.368	0.076	3.23	1.890	2.840
2020-785	4	J-G WITH B	191	RX-DAYTONA	126	RANDOX	22	2.579	0.121	4.70	2.060	3.090
2020-785	4	J-G WITH B	212	S. D. EXL	143	SIEMENS	11	2.350	0.118	5.02	1.880	2.820
2020-785	5	J-G NO BK.	215	B.C DXC	134	BECKM.COUL	12	2.644	0.152	5.76	2.120	3.170
2020-785	7	DIAZO-DYPH	165	VITROS	63	KODAK	143	2.220	0.131	5.91	1.780	2.660
2020-785	7	DIAZO-DYPH	165	VITROS	63	KODAK	143	2.233	0.094	4.22	1.790	2.680
2020-785	13	AZOBIL.	184	INTEGRA	93	ROCHE	18	2.118	0.094	4.42	1.690	2.540
2020-785	13	AZOBIL.	203	ROCHE	93	ROCHE	23	2.062	0.061	2.94	1.650	2.470
2020-785	19	DIAZO W BL	197	ARCHITECT	1	ABBOTT	11	2.085	0.113	5.43	1.670	2.500
010 CALCIUM, TOTAL			CLIA TEST ID #: 345			EVALUATION CRITERIA: TARGET VALUE 1.00						
2020-781	3	ARSEN.-III	165	VITROS	63	KODAK	160	9.991	0.175	1.75	9.000	11.000
2020-781	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	10.266	0.162	1.58	9.300	11.300
2020-781	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	34	10.368	0.266	2.57	9.400	11.400
2020-781	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	10.033	0.311	3.10	9.000	11.000
2020-781	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	12	10.200	0.332	3.25	9.200	11.200
2020-781	3	ARSEN.-III	214	Envoy 500	150	ELITECH	10	10.340	0.294	2.84	9.300	11.300
2020-781	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	9.903	0.295	2.98	8.900	10.900
2020-781	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	10.139	0.323	3.19	9.100	11.100
2020-781	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	10.190	0.301	2.96	9.200	11.200

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-781	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	9.943	0.331	3.33	8.900	10.900
2020-781	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	9.867	0.175	1.77	8.900	10.900
2020-781	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	10.213	0.190	1.86	9.200	11.200
2020-781	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	22	10.155	0.221	2.18	9.200	11.200
010	CALCIUM, TOTAL		CLIA TEST ID #: 345		EVALUATION CRITERIA: TARGET VALUE 1.00							
2020-782	3	ARSEN.-III	165	VITROS	63	KODAK	160	7.648	0.162	2.12	6.600	8.600
2020-782	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	7.772	0.160	2.05	6.800	8.800
2020-782	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	34	8.050	0.206	2.56	7.100	9.100
2020-782	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	7.527	0.173	2.30	6.500	8.500
2020-782	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	12	7.967	0.243	3.05	7.000	9.000
2020-782	3	ARSEN.-III	214	Envoy 500	150	ELITECH	10	7.840	0.185	2.37	6.800	8.800
2020-782	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	7.679	0.247	3.21	6.700	8.700
2020-782	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	7.828	0.278	3.56	6.800	8.800
2020-782	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	7.800	0.224	2.87	6.800	8.800
2020-782	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	7.714	0.259	3.35	6.700	8.700
2020-782	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	7.650	0.194	2.53	6.700	8.700
2020-782	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	7.887	0.193	2.44	6.900	8.900
2020-782	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	22	7.809	0.217	2.78	6.800	8.800
2020-783	3	ARSEN.-III	165	VITROS	63	KODAK	160	10.519	0.215	2.05	9.500	11.500
2020-783	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	10.845	0.203	1.87	9.800	11.800
2020-783	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	34	10.915	0.327	2.99	9.900	11.900
2020-783	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	10.613	0.253	2.38	9.600	11.600
2020-783	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	12	10.842	0.307	2.83	9.800	11.800
2020-783	3	ARSEN.-III	214	Envoy 500	150	ELITECH	10	10.910	0.318	2.91	9.900	11.900
2020-783	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	10.479	0.335	3.19	9.500	11.500
2020-783	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	10.683	0.227	2.12	9.700	11.700
2020-783	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	10.710	0.314	2.94	9.700	11.700
2020-783	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	10.550	0.277	2.63	9.600	11.600
2020-783	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	10.450	0.382	3.65	9.500	11.500

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-783	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	10.663	0.206	1.93	9.700	11.700
2020-783	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	22	10.732	0.258	2.41	9.700	11.700
010	CALCIUM, TOTAL		CLIA TEST ID #: 345				EVALUATION CRITERIA: TARGET VALUE 1.00					
2020-784	3	ARSEN.-III	165	VITROS	63	KODAK	160	11.457	0.246	2.15	10.500	12.500
2020-784	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	11.914	0.301	2.53	10.900	12.900
2020-784	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	34	11.876	0.435	3.66	10.900	12.900
2020-784	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	11.807	0.307	2.60	10.800	12.800
2020-784	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	12	11.833	0.461	3.89	10.800	12.800
2020-784	3	ARSEN.-III	214	Envoy 500	150	ELITECH	10	11.990	0.197	1.64	11.000	13.000
2020-784	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	11.506	0.369	3.20	10.500	12.500
2020-784	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	11.656	0.261	2.24	10.700	12.700
2020-784	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	11.700	0.283	2.42	10.700	12.700
2020-784	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	11.557	0.435	3.77	10.600	12.600
2020-784	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	11.392	0.275	2.42	10.400	12.400
2020-784	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	11.769	0.169	1.43	10.800	12.800
2020-784	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	22	11.736	0.323	2.75	10.700	12.700
2020-785	3	ARSEN.-III	165	VITROS	63	KODAK	160	8.690	0.173	1.99	7.700	9.700
2020-785	3	ARSEN.-III	186	BECK. COUL	76	AU-BECKMAN	29	8.797	0.143	1.62	7.800	9.800
2020-785	3	ARSEN.-III	191	RX-DAYTONA	126	RANDOX	34	9.009	0.221	2.46	8.000	10.000
2020-785	3	ARSEN.-III	197	ARCHITECT	1	ABBOTT	15	8.500	0.179	2.10	7.500	9.500
2020-785	3	ARSEN.-III	199	Rx-IMOLA	126	RANDOX	12	8.983	0.336	3.74	8.000	10.000
2020-785	3	ARSEN.-III	214	Envoy 500	150	ELITECH	10	8.890	0.164	1.84	7.900	9.900
2020-785	8	CRESOLPHT.	38	SIEMENS DI	143	SIEMENS	81	8.615	0.239	2.77	7.600	9.600
2020-785	8	CRESOLPHT.	184	INTEGRA	93	ROCHE	18	8.778	0.253	2.88	7.800	9.800
2020-785	8	CRESOLPHT.	203	ROCHE	93	ROCHE	10	8.840	0.269	3.04	7.800	9.800
2020-785	8	CRESOLPHT.	212	S. D. EXL	143	SIEMENS	14	8.700	0.270	3.10	7.700	9.700
2020-785	14	ELECTRODE	215	B.C DXC	134	BECKM.COUL	12	8.608	0.175	2.04	7.600	9.600
2020-785	15	CA-NM-BAPT	184	INTEGRA	93	ROCHE	16	8.838	0.136	1.54	7.800	9.800
2020-785	15	CA-NM-BAPT	203	ROCHE	93	ROCHE	22	8.838	0.170	1.93	7.800	9.800

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
011	CHLORIDE	CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2020-781	3 ISE DILUTE	184 INTEGRA	93 ROCHE		37	109.189	2.534	2.32	104.000	115.000	
2020-781	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	106.069	2.348	2.21	101.000	111.000	
2020-781	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		17	107.000	2.376	2.22	102.000	112.000	
2020-781	3 ISE DILUTE	203 ROCHE	93 ROCHE		35	104.514	1.746	1.67	99.000	110.000	
2020-781	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	106.417	1.935	1.82	101.000	112.000	
2020-781	4 ISE UND-DI	165 VITROS	63 KODAK		129	106.531	1.691	1.59	101.000	112.000	
2020-781	4 ISE UND-DI	165 VITROS	156 Kit 32		29	107.250	1.326	1.24	102.000	113.000	
2020-781	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX		35	106.571	2.453	2.30	101.000	112.000	
2020-781	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX		12	106.750	2.420	2.27	101.000	112.000	
2020-781	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		78	108.688	2.326	2.14	103.000	114.000	
2020-781	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	108.500	1.842	1.70	103.000	114.000	
2020-782	3 ISE DILUTE	184 INTEGRA	93 ROCHE		37	96.472	2.327	2.41	92.000	101.000	
2020-782	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	95.448	1.922	2.01	91.000	100.000	
2020-782	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		17	97.059	1.662	1.71	92.000	102.000	
2020-782	3 ISE DILUTE	203 ROCHE	93 ROCHE		35	92.486	2.430	2.63	88.000	97.000	
2020-782	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	96.167	2.154	2.24	91.000	101.000	
2020-782	4 ISE UND-DI	165 VITROS	63 KODAK		129	95.795	1.569	1.64	91.000	101.000	
2020-782	4 ISE UND-DI	165 VITROS	156 Kit 32		29	96.621	1.243	1.29	92.000	101.000	
2020-782	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX		35	94.714	2.361	2.49	90.000	99.000	
2020-782	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX		12	94.583	2.597	2.75	90.000	99.000	
2020-782	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		78	95.256	1.489	1.56	90.000	100.000	
2020-782	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	95.071	1.751	1.84	90.000	100.000	
2020-783	3 ISE DILUTE	184 INTEGRA	93 ROCHE		37	111.081	3.008	2.71	106.000	117.000	
2020-783	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	107.793	2.369	2.20	102.000	113.000	
2020-783	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		17	109.176	2.307	2.11	104.000	115.000	
2020-783	3 ISE DILUTE	203 ROCHE	93 ROCHE		35	107.286	1.798	1.68	102.000	113.000	
2020-783	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	108.750	2.890	2.66	103.000	114.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-783	4 ISE UND-DI	165 VITROS	63 KODAK	129	108.617	2.054	1.89	103.000	114.000		
2020-783	4 ISE UND-DI	165 VITROS	156 Kit 32	29	109.714	1.708	1.56	104.000	115.000		
2020-783	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	35	108.857	3.034	2.79	103.000	114.000		
2020-783	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	12	109.417	2.753	2.52	104.000	115.000		
2020-783	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	111.545	2.316	2.08	106.000	117.000		
2020-783	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	111.571	1.990	1.78	106.000	117.000		
011 CHLORIDE		CLIA TEST ID #: 355	EVALUATION CRITERIA: TARGET VALUE 5 %								
2020-784	3 ISE DILUTE	184 INTEGRA	93 ROCHE	37	116.111	2.716	2.34	110.000	122.000		
2020-784	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	111.241	2.144	1.93	106.000	117.000		
2020-784	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	17	112.706	2.346	2.08	107.000	118.000		
2020-784	3 ISE DILUTE	203 ROCHE	93 ROCHE	35	111.647	1.998	1.79	106.000	117.000		
2020-784	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	112.833	2.075	1.84	107.000	118.000		
2020-784	4 ISE UND-DI	165 VITROS	63 KODAK	129	112.575	1.974	1.75	107.000	118.000		
2020-784	4 ISE UND-DI	165 VITROS	156 Kit 32	29	113.207	1.606	1.42	108.000	119.000		
2020-784	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	35	113.486	2.556	2.25	108.000	119.000		
2020-784	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	12	114.417	2.871	2.51	109.000	120.000		
2020-784	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	116.481	2.426	2.08	111.000	122.000		
2020-784	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	115.857	2.356	2.03	110.000	122.000		
2020-785	3 ISE DILUTE	184 INTEGRA	93 ROCHE	37	101.730	2.738	2.69	97.000	107.000		
2020-785	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	99.862	1.737	1.74	95.000	105.000		
2020-785	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	17	101.353	1.713	1.69	96.000	106.000		
2020-785	3 ISE DILUTE	203 ROCHE	93 ROCHE	35	97.600	1.945	1.99	93.000	102.000		
2020-785	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	100.667	1.491	1.48	96.000	106.000		
2020-785	4 ISE UND-DI	165 VITROS	63 KODAK	129	100.383	1.572	1.57	95.000	105.000		
2020-785	4 ISE UND-DI	165 VITROS	156 Kit 32	29	101.000	0.910	0.90	96.000	106.000		
2020-785	4 ISE UND-DI	191 RX-DAYTONA	126 RANDOX	35	99.343	2.242	2.26	94.000	104.000		
2020-785	4 ISE UND-DI	199 Rx-IMOLA	126 RANDOX	12	99.750	2.385	2.39	95.000	105.000		
2020-785	9 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	100.705	1.586	1.57	96.000	106.000		
2020-785	9 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	101.143	1.457	1.44	96.000	106.000		

Participant Summary
MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY SUB: 10 ROUTINE GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
012	CHOLESTEROL, TOTAL	CLIA TEST ID #: 365				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-781	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	79	175.641	5.630	3.21	158.000	193.000		
2020-781	1 ENZYMATI	165 VITROS	63 KODAK	149	187.306	5.615	3.00	169.000	206.000		
2020-781	1 ENZYMATI	184 INTEGRA	93 ROCHE	20	182.900	6.204	3.39	165.000	201.000		
2020-781	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	196.711	7.553	3.84	177.000	216.000		
2020-781	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	202.667	5.558	2.74	182.000	223.000		
2020-781	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	185.471	4.948	2.67	167.000	204.000		
2020-781	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	11	192.909	6.639	3.44	174.000	212.000		
2020-781	1 ENZYMATI	203 ROCHE	93 ROCHE	20	185.200	4.812	2.60	167.000	204.000		
2020-781	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	176.615	6.058	3.43	159.000	194.000		
2020-781	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	193.273	6.412	3.32	174.000	213.000		
2020-781	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	186.833	3.436	1.84	168.000	206.000		
2020-781	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	183.680	3.792	2.06	165.000	202.000		
2020-781	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	184.800	2.482	1.34	166.000	203.000		
2020-782	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	79	138.821	4.537	3.27	125.000	153.000		
2020-782	1 ENZYMATI	165 VITROS	63 KODAK	149	138.639	3.856	2.78	125.000	153.000		
2020-782	1 ENZYMATI	184 INTEGRA	93 ROCHE	20	142.474	5.394	3.79	128.000	157.000		
2020-782	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	155.667	7.318	4.70	140.000	171.000		
2020-782	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	153.211	5.863	3.83	138.000	169.000		
2020-782	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	144.765	3.335	2.30	130.000	159.000		
2020-782	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	11	149.273	4.287	2.87	134.000	164.000		
2020-782	1 ENZYMATI	203 ROCHE	93 ROCHE	20	145.150	3.991	2.75	131.000	160.000		
2020-782	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	142.000	8.027	5.65	128.000	156.000		
2020-782	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	149.455	7.203	4.82	135.000	164.000		
2020-782	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	144.250	2.126	1.47	130.000	159.000		
2020-782	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	143.600	2.577	1.79	129.000	158.000		
2020-782	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	141.760	2.534	1.79	128.000	156.000		
2020-783	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	79	183.570	7.129	3.88	165.000	202.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-783	1 ENZYMATI	165 VITROS	63 KODAK	149	199.850	5.829	2.92	180.000	220.000		
2020-783	1 ENZYMATI	184 INTEGRA	93 ROCHE	20	191.500	5.714	2.98	172.000	211.000		
2020-783	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	206.474	9.182	4.45	186.000	227.000		
2020-783	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	217.333	6.128	2.82	196.000	239.000		
2020-783	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	195.235	4.180	2.14	176.000	215.000		
2020-783	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	11	204.000	8.079	3.96	184.000	224.000		
2020-783	1 ENZYMATI	203 ROCHE	93 ROCHE	20	195.050	4.387	2.25	176.000	215.000		
2020-783	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	188.786	11.213	5.94	170.000	208.000		
2020-783	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	203.364	11.850	5.83	183.000	224.000		
2020-783	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	197.750	6.287	3.18	178.000	218.000		
2020-783	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	192.800	4.561	2.37	174.000	212.000		
2020-783	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	192.200	3.655	1.90	173.000	211.000		
012 CHOLESTEROL, TOTAL		CLIA TEST ID #: 365	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-784	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	79	200.101	8.028	4.01	180.000	220.000		
2020-784	1 ENZYMATI	165 VITROS	63 KODAK	149	222.687	6.123	2.75	200.000	245.000		
2020-784	1 ENZYMATI	184 INTEGRA	93 ROCHE	20	208.450	7.025	3.37	188.000	229.000		
2020-784	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	232.333	8.498	3.66	209.000	256.000		
2020-784	1 ENZYMATI	191 RX-DAYTONA	126 RANDOX	38	225.289	9.782	4.34	203.000	248.000		
2020-784	1 ENZYMATI	197 ARCHITECT	1 ABBOTT	17	212.882	5.603	2.63	192.000	234.000		
2020-784	1 ENZYMATI	199 Rx-IMOLA	126 RANDOX	11	224.000	9.274	4.14	202.000	246.000		
2020-784	1 ENZYMATI	203 ROCHE	93 ROCHE	20	214.200	4.996	2.33	193.000	236.000		
2020-784	1 ENZYMATI	212 S. D. EXL	143 SIEMENS	14	203.214	12.066	5.94	183.000	224.000		
2020-784	1 ENZYMATI	214 Envoy 500	150 ELITECH	11	223.545	8.680	3.88	201.000	246.000		
2020-784	1 ENZYMATI	215 B.C DXC	134 BECKM.COUL	12	217.083	3.662	1.69	195.000	239.000		
2020-784	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	213.800	3.124	1.46	192.000	235.000		
2020-784	8 ENZYMATI	186 BECK. COUL	76 AU-BECKMAN	26	210.680	5.357	2.54	190.000	232.000		
2020-785	1 ENZYMATI	38 SIEMENS DI	143 SIEMENS	79	153.392	4.916	3.20	138.000	169.000		
2020-785	1 ENZYMATI	165 VITROS	63 KODAK	149	159.432	4.428	2.78	143.000	175.000		
2020-785	1 ENZYMATI	184 INTEGRA	93 ROCHE	20	159.400	4.994	3.13	143.000	175.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-785	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	38	171.459	6.508	3.80	154.000	189.000
2020-785	1	ENZYMATIC	191	RX-DAYTONA	126	RANDOX	38	183.667	7.318	3.98	165.000	202.000
2020-785	1	ENZYMATIC	197	ARCHITECT	1	ABBOTT	17	162.412	3.881	2.39	146.000	179.000
2020-785	1	ENZYMATIC	199	Rx-IMOLA	126	RANDOX	11	168.727	7.098	4.21	152.000	186.000
2020-785	1	ENZYMATIC	203	ROCHE	93	ROCHE	20	162.150	3.991	2.46	146.000	178.000
2020-785	1	ENZYMATIC	212	S. D. EXL	143	SIEMENS	14	155.286	11.548	7.44	140.000	171.000
2020-785	1	ENZYMATIC	214	Envoy 500	150	ELITECH	11	169.909	7.549	4.44	153.000	187.000
2020-785	1	ENZYMATIC	215	B.C DXC	134	BECKM.COUL	12	161.500	3.617	2.24	145.000	178.000
2020-785	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	26	159.600	2.577	1.61	144.000	176.000
2020-785	8	ENZYMATIC	186	BECK. COUL	76	AU-BECKMAN	26	162.400	3.720	2.29	146.000	179.000
013 CHOLESTEROL, (HDL)			CLIA TEST ID #: 375				EVALUATION CRITERIA: TARGET VALUE 30 %					
2020-781	11	ENC.-COLO.	165	VITROS	63	KODAK	144	67.308	3.488	5.18	47.100	87.500
2020-781	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	31	70.474	3.800	5.39	49.300	91.600
2020-781	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	26	72.158	5.354	7.42	50.500	93.800
2020-781	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	35	65.776	5.227	7.95	46.000	85.500
2020-781	11	ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	11	43.327	3.124	7.21	30.300	56.300
2020-781	11	ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	60.133	6.775	11.27	42.100	78.200
2020-781	11	ENC.-COLO.	203	ROCHE	93	ROCHE	30	69.893	2.660	3.81	48.900	90.900
2020-781	12	HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	17	69.118	1.961	2.84	48.400	89.900
2020-781	12	HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	71.900	3.673	5.11	50.300	93.500
2020-781	12	HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	72.533	3.335	4.60	50.800	94.300
2020-781	13	AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	78	60.374	2.814	4.66	42.300	78.500
2020-781	13	AMIN.W.E.Q	212	S. D. EXL	143	SIEMENS	14	59.786	1.612	2.70	41.900	77.700
2020-782	11	ENC.-COLO.	165	VITROS	63	KODAK	144	54.219	2.898	5.35	38.000	70.500
2020-782	11	ENC.-COLO.	184	INTEGRA	93	ROCHE	31	59.880	3.608	6.02	41.900	77.800
2020-782	11	ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	26	57.008	4.826	8.47	39.900	74.100
2020-782	11	ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	35	52.820	5.632	10.66	37.000	68.700
2020-782	11	ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	11	33.973	2.360	6.95	23.800	44.200
2020-782	11	ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	48.508	5.773	11.90	34.000	63.100

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-782	11 ENC.-COLO.	203	ROCHE	93	ROCHE	30	58.914	2.412	4.09	41.200	76.600	
2020-782	12 HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	17	56.165	1.874	3.34	39.300	73.000	
2020-782	12 HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	57.860	2.034	3.52	40.500	75.200	
2020-782	12 HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	58.225	3.286	5.64	40.800	75.700	
2020-782	13 AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	78	55.397	2.272	4.10	38.800	72.000	
2020-782	13 AMIN.W.E.Q	212	S. D. EXL	143	SIEMENS	14	54.857	1.505	2.74	38.400	71.300	
013	CHOLESTEROL, (HDL)	CLIA	TEST ID #: 375	EVALUATION CRITERIA: TARGET VALUE 30 %								
2020-783	11 ENC.-COLO.	165	VITROS	63	KODAK	144	70.340	3.885	5.52	49.200	91.400	
2020-783	11 ENC.-COLO.	184	INTEGRA	93	ROCHE	31	71.639	3.952	5.52	50.100	93.100	
2020-783	11 ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	26	74.800	5.417	7.24	52.400	97.200	
2020-783	11 ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	35	68.171	6.255	9.18	47.700	88.600	
2020-783	11 ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	11	45.018	3.459	7.68	31.500	58.500	
2020-783	11 ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	64.942	8.750	13.47	45.500	84.400	
2020-783	11 ENC.-COLO.	203	ROCHE	93	ROCHE	30	71.683	3.031	4.23	50.200	93.200	
2020-783	12 HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	17	73.082	2.189	2.99	51.200	95.000	
2020-783	12 HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	75.650	4.062	5.37	53.000	98.300	
2020-783	12 HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	76.558	3.717	4.86	53.600	99.500	
2020-783	13 AMIN.W.E.Q	38	SIEMENS DI	143	SIEMENS	78	61.257	2.463	4.02	42.900	79.600	
2020-783	13 AMIN.W.E.Q	212	S. D. EXL	143	SIEMENS	14	61.286	2.630	4.29	42.900	79.700	
2020-784	11 ENC.-COLO.	165	VITROS	63	KODAK	144	77.337	3.928	5.08	54.100	100.500	
2020-784	11 ENC.-COLO.	184	INTEGRA	93	ROCHE	31	75.806	3.834	5.06	53.100	98.500	
2020-784	11 ENC.-COLO.	186	BECK. COUL	76	AU-BECKMAN	26	81.269	5.507	6.78	56.900	105.700	
2020-784	11 ENC.-COLO.	191	RX-DAYTONA	126	RANDOX	35	75.089	7.377	9.82	52.600	97.600	
2020-784	11 ENC.-COLO.	194	BIOLIS-24S	154	(IDG)	11	49.100	3.523	7.17	34.400	63.800	
2020-784	11 ENC.-COLO.	199	Rx-IMOLA	126	RANDOX	12	71.075	7.319	10.30	49.800	92.400	
2020-784	11 ENC.-COLO.	203	ROCHE	93	ROCHE	30	76.310	3.718	4.87	53.400	99.200	
2020-784	12 HOM-LQ.S.D	197	ARCHITECT	1	ABBOTT	17	78.453	1.821	2.32	54.900	102.000	
2020-784	12 HOM-LQ.S.D	214	Envoy 500	150	ELITECH	10	82.140	3.319	4.04	57.500	106.800	
2020-784	12 HOM-LQ.S.D	215	B.C DXC	134	BECKM.COUL	12	83.108	4.523	5.44	58.200	108.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-784	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	62.563	2.695	4.31	43.800	81.300		
2020-784	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	14	61.857	2.326	3.76	43.300	80.400		
013 CHOLESTEROL, (HDL)		CLIA TEST ID #: 375				EVALUATION CRITERIA: TARGET VALUE 30 %					
2020-785	11 ENC.-COLO.	165 VITROS	63 KODAK	144	59.369	3.063	5.16	41.600	77.200		
2020-785	11 ENC.-COLO.	184 INTEGRA	93 ROCHE	31	63.674	3.396	5.33	44.600	82.800		
2020-785	11 ENC.-COLO.	186 BECK. COUL	76 AU-BECKMAN	26	62.908	4.133	6.57	44.000	81.800		
2020-785	11 ENC.-COLO.	191 RX-DAYTONA	126 RANDOX	35	59.891	5.314	8.87	41.900	77.900		
2020-785	11 ENC.-COLO.	194 BIOLIS-24S	154 (IDG)	11	37.818	1.992	5.27	26.500	49.200		
2020-785	11 ENC.-COLO.	199 Rx-IMOLA	126 RANDOX	12	57.250	4.361	7.62	40.100	74.400		
2020-785	11 ENC.-COLO.	203 ROCHE	93 ROCHE	30	64.087	3.172	4.95	44.900	83.300		
2020-785	12 HOM-LQ.S.D	197 ARCHITECT	1 ABBOTT	17	61.971	1.537	2.48	43.400	80.600		
2020-785	12 HOM-LQ.S.D	214 Envoy 500	150 ELITECH	10	64.820	3.028	4.67	45.400	84.300		
2020-785	12 HOM-LQ.S.D	215 B.C DXC	134 BECKM.COUL	12	64.600	3.794	5.87	45.200	84.000		
2020-785	13 AMIN.W.E.Q	38 SIEMENS DI	143 SIEMENS	78	57.211	2.186	3.82	40.000	74.400		
2020-785	13 AMIN.W.E.Q	212 S. D. EXL	143 SIEMENS	14	57.000	1.927	3.38	39.900	74.100		
014 CREATINE KINASE (CK)		CLIA TEST ID #: 385									
2020-781	3 KINETIC 37	165 VITROS	63 KODAK	76	278.413	20.333	7.30	195.000	362.000		
2020-781	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	332.778	12.461	3.74	233.000	433.000		
2020-781	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	304.100	6.564	2.16	213.000	395.000		
2020-781	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	387.636	11.452	2.95	271.000	504.000		
2020-781	8 Hexokinase	184 INTEGRA	93 ROCHE	12	367.250	12.617	3.44	257.000	477.000		
2020-781	8 Hexokinase	203 ROCHE	93 ROCHE	12	369.667	14.395	3.89	259.000	481.000		
2020-781	9 SIEMENS	38 SIEMENS DI	143 SIEMENS	29	371.448	9.394	2.53	260.000	483.000		
2020-782	3 KINETIC 37	165 VITROS	63 KODAK	76	142.000	10.108	7.12	99.000	185.000		
2020-782	3 KINETIC 37	186 BECK. COUL	76 AU-BECKMAN	18	148.722	5.848	3.93	104.000	193.000		
2020-782	3 KINETIC 37	194 BIOLIS-24S	154 (IDG)	10	137.900	4.110	2.98	97.000	179.000		
2020-782	4 OLIVER ROS	215 B.C DXC	134 BECKM.COUL	11	176.273	5.754	3.26	123.000	229.000		
2020-782	8 Hexokinase	184 INTEGRA	93 ROCHE	12	167.833	6.216	3.70	117.000	218.000		
2020-782	8 Hexokinase	203 ROCHE	93 ROCHE	12	166.750	7.001	4.20	117.000	217.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-782	9 SIEMENS	38	SIEMENS DI	143	SIEMENS	29	168.103	4.751	2.83	118.000	219.000
014	CREATINE KINASE (CK)	CLIA TEST ID #: 385				EVALUATION CRITERIA: TARGET VALUE 30 %					
2020-783	3 KINETIC 37	165	VITROS	63	KODAK	76	308.267	22.323	7.24	216.000	401.000
2020-783	3 KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	382.167	14.908	3.90	268.000	497.000
2020-783	3 KINETIC 37	194	BIOLIS-24S	154	(IDG)	10	346.700	15.369	4.43	243.000	451.000
2020-783	4 OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	447.364	24.182	5.41	313.000	582.000
2020-783	8 Hexokinase	184	INTEGRA	93	ROCHE	12	412.500	17.017	4.13	289.000	536.000
2020-783	8 Hexokinase	203	ROCHE	93	ROCHE	12	423.000	14.172	3.35	296.000	550.000
2020-783	9 SIEMENS	38	SIEMENS DI	143	SIEMENS	29	424.607	9.363	2.21	297.000	552.000
2020-784	3 KINETIC 37	165	VITROS	63	KODAK	76	352.224	27.362	7.77	247.000	458.000
2020-784	3 KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	471.444	15.840	3.36	330.000	613.000
2020-784	3 KINETIC 37	194	BIOLIS-24S	154	(IDG)	10	431.500	12.548	2.91	302.000	561.000
2020-784	4 OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	548.182	16.067	2.93	384.000	713.000
2020-784	8 Hexokinase	184	INTEGRA	93	ROCHE	12	512.417	15.511	3.03	359.000	666.000
2020-784	8 Hexokinase	203	ROCHE	93	ROCHE	12	525.417	16.580	3.16	368.000	683.000
2020-784	9 SIEMENS	38	SIEMENS DI	143	SIEMENS	29	524.552	13.815	2.63	367.000	682.000
2020-785	3 KINETIC 37	165	VITROS	63	KODAK	76	208.608	14.341	6.87	146.000	271.000
2020-785	3 KINETIC 37	186	BECK. COUL	76	AU-BECKMAN	18	228.611	7.761	3.39	160.000	297.000
2020-785	3 KINETIC 37	194	BIOLIS-24S	154	(IDG)	10	210.400	4.409	2.10	147.000	274.000
2020-785	4 OLIVER ROS	215	B.C DXC	134	BECKM.COUL	11	270.636	7.547	2.79	189.000	352.000
2020-785	8 Hexokinase	184	INTEGRA	93	ROCHE	12	252.667	9.086	3.60	177.000	328.000
2020-785	8 Hexokinase	203	ROCHE	93	ROCHE	12	255.000	9.574	3.75	179.000	332.000
2020-785	9 SIEMENS	38	SIEMENS DI	143	SIEMENS	29	255.483	6.704	2.62	179.000	332.000
016	CREATININE	CLIA TEST ID #: 405				EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}					
2020-781	2 AK. PIC. N	191	RX-DAYTONA	126	RANDOX	10	4.740	0.294	6.20	4.030	5.450
2020-781	2 AK. PIC. N	194	BIOLIS-24S	154	(IDG)	11	5.510	0.201	3.65	4.680	6.340
2020-781	3 AK. PIC. L	191	RX-DAYTONA	126	RANDOX	23	4.508	0.360	7.99	3.830	5.180
2020-781	6 KIN. ALK P	184	INTEGRA	93	ROCHE	23	5.388	0.159	2.95	4.580	6.200

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-781	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	29	5.310	0.196	3.70	4.510	6.110
2020-781	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	19	5.531	0.226	4.08	4.700	6.360
2020-781	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	5.370	0.148	2.75	4.560	6.180
2020-781	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	12	5.530	0.133	2.41	4.700	6.360
2020-781	8	IDMS-CREA	165	VITROS	63	KODAK	156	5.336	0.130	2.44	4.540	6.140
2020-781	11	IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	73	5.618	0.162	2.88	4.780	6.460
2020-781	11	IDMS Jaffe	212	S. D. EXL	143	SIEMENS	10	5.573	0.254	4.57	4.740	6.410
016 CREATININE			CLIA TEST ID #: 405	EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}								
2020-782	2	AK. PIC. N	191	RX-DAYTONA	126	RANDOX	10	3.719	0.260	6.98	3.160	4.280
2020-782	2	AK. PIC. N	194	BIOLIS-24S	154	(IDG)	11	4.399	0.151	3.44	3.740	5.060
2020-782	3	AK. PIC. L	191	RX-DAYTONA	126	RANDOX	23	3.600	0.277	7.68	3.060	4.140
2020-782	6	KIN. ALK P	184	INTEGRA	93	ROCHE	23	4.445	0.136	3.07	3.780	5.110
2020-782	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	29	4.252	0.167	3.94	3.610	4.890
2020-782	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	19	4.254	0.158	3.71	3.620	4.890
2020-782	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	4.426	0.129	2.92	3.760	5.090
2020-782	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	12	4.342	0.142	3.28	3.690	4.990
2020-782	8	IDMS-CREA	165	VITROS	63	KODAK	156	4.165	0.098	2.35	3.540	4.790
2020-782	11	IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	73	4.504	0.142	3.14	3.830	5.180
2020-782	11	IDMS Jaffe	212	S. D. EXL	143	SIEMENS	10	4.464	0.190	4.27	3.790	5.130
2020-783	2	AK. PIC. N	191	RX-DAYTONA	126	RANDOX	10	4.870	0.290	5.95	4.140	5.600
2020-783	2	AK. PIC. N	194	BIOLIS-24S	154	(IDG)	11	5.791	0.139	2.41	4.920	6.660
2020-783	3	AK. PIC. L	191	RX-DAYTONA	126	RANDOX	23	4.692	0.365	7.78	3.990	5.400
2020-783	6	KIN. ALK P	184	INTEGRA	93	ROCHE	23	5.520	0.176	3.19	4.690	6.350
2020-783	6	KIN. ALK P	186	BECK. COUL	76	AU-BECKMAN	29	5.540	0.154	2.79	4.710	6.370
2020-783	6	KIN. ALK P	197	ARCHITECT	1	ABBOTT	19	5.860	0.221	3.77	4.980	6.740
2020-783	6	KIN. ALK P	203	ROCHE	93	ROCHE	22	5.603	0.162	2.90	4.760	6.440
2020-783	6	KIN. ALK P	215	B.C DXC	134	BECKM.COUL	12	5.833	0.228	3.90	4.960	6.710
2020-783	8	IDMS-CREA	165	VITROS	63	KODAK	156	5.574	0.142	2.54	4.740	6.410
2020-783	11	IDMS Jaffe	38	SIEMENS DI	143	SIEMENS	73	5.921	0.184	3.10	5.030	6.810

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-783	11	IDMS	Jaffe	212	S. D. EXL	143	SIEMENS	10	5.908	0.214	3.62	5.020	6.790
016	CREATININE	CLIA TEST ID #: 405				EVALUATION CRITERIA: TARGET VALUE 15 % OR 0.30 {GREATER}							
2020-784	2	AK. PIC. N		191	RX-DAYTONA	126	RANDOX	10	5.366	0.262	4.89	4.560	6.170
2020-784	2	AK. PIC. N		194	BIOLIS-24S	154	(IDG)	11	6.299	0.215	3.42	5.350	7.240
2020-784	3	AK. PIC. L		191	RX-DAYTONA	126	RANDOX	23	5.120	0.308	6.02	4.350	5.890
2020-784	6	KIN. ALK P		184	INTEGRA	93	ROCHE	23	5.833	0.225	3.86	4.960	6.710
2020-784	6	KIN. ALK P		186	BECK. COUL	76	AU-BECKMAN	29	5.969	0.244	4.09	5.070	6.860
2020-784	6	KIN. ALK P		197	ARCHITECT	1	ABBOTT	19	6.452	0.232	3.60	5.480	7.420
2020-784	6	KIN. ALK P		203	ROCHE	93	ROCHE	22	5.993	0.147	2.45	5.090	6.890
2020-784	6	KIN. ALK P		215	B.C DXC	134	BECKM.COUL	12	6.317	0.159	2.52	5.370	7.260
2020-784	8	IDMS-CREA		165	VITROS	63	KODAK	156	5.992	0.157	2.62	5.090	6.890
2020-784	11	IDMS Jaffe		38	SIEMENS DI	143	SIEMENS	73	6.459	0.191	2.95	5.490	7.430
2020-784	11	IDMS Jaffe		212	S. D. EXL	143	SIEMENS	10	6.459	0.153	2.37	5.490	7.430
2020-785	2	AK. PIC. N		191	RX-DAYTONA	126	RANDOX	10	4.156	0.273	6.57	3.530	4.780
2020-785	2	AK. PIC. N		194	BIOLIS-24S	154	(IDG)	11	4.861	0.169	3.47	4.130	5.590
2020-785	3	AK. PIC. L		191	RX-DAYTONA	126	RANDOX	23	3.975	0.208	5.23	3.380	4.570
2020-785	6	KIN. ALK P		184	INTEGRA	93	ROCHE	23	4.802	0.163	3.39	4.080	5.520
2020-785	6	KIN. ALK P		186	BECK. COUL	76	AU-BECKMAN	29	4.710	0.136	2.88	4.000	5.420
2020-785	6	KIN. ALK P		197	ARCHITECT	1	ABBOTT	19	4.829	0.180	3.73	4.100	5.550
2020-785	6	KIN. ALK P		203	ROCHE	93	ROCHE	22	4.842	0.123	2.53	4.120	5.570
2020-785	6	KIN. ALK P		215	B.C DXC	134	BECKM.COUL	12	4.899	0.108	2.21	4.160	5.630
2020-785	8	IDMS-CREA		165	VITROS	63	KODAK	156	4.679	0.115	2.47	3.980	5.380
2020-785	11	IDMS Jaffe		38	SIEMENS DI	143	SIEMENS	73	4.974	0.143	2.87	4.230	5.720
2020-785	11	IDMS Jaffe		212	S. D. EXL	143	SIEMENS	10	4.948	0.189	3.83	4.210	5.690
017	GLUCOSE	CLIA TEST ID #: 415				EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}							
2020-781	2	GL. OXI. E		215	B.C DXC	134	BECKM.COUL	12	122.250	2.681	2.19	110.000	134.000
2020-781	3	GL. OXI. C		165	VITROS	63	KODAK	166	124.768	2.773	2.22	112.000	137.000
2020-781	3	GL. OXI. C		213	SELECTRA P	150	ELITECH	12	127.000	5.401	4.25	114.000	140.000
2020-781	3	GL. OXI. C		214	Envoy 500	150	ELITECH	12	131.000	4.340	3.31	118.000	144.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-781	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	81	129.392	3.399	2.63	116.000	142.000
2020-781	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	29	124.069	4.059	3.27	112.000	136.000
2020-781	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	26	131.269	4.063	3.09	118.000	144.000
2020-781	5	HEXOKINASE	203 ROCHE	93	ROCHE	19	124.789	3.019	2.42	112.000	137.000
2020-781	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	129.500	2.228	1.72	117.000	142.000
2020-781	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	11	119.818	3.040	2.54	108.000	132.000
2020-781	12	HEXOKINASE	184 INTEGRA	93	ROCHE	26	127.920	2.741	2.14	115.000	141.000
2020-781	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	18	120.500	2.455	2.04	108.000	133.000
2020-781	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	123.188	4.850	3.94	111.000	136.000
017	GLUCOSE	CLIA TEST ID #: 415	EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}								
2020-782	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	77.333	2.392	3.09	70.000	85.000
2020-782	3	GL. OXI. C	165 VITROS	63	KODAK	166	77.104	2.053	2.66	69.000	85.000
2020-782	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	81.167	3.078	3.79	73.000	89.000
2020-782	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	84.333	2.867	3.40	76.000	93.000
2020-782	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	81	83.100	2.332	2.81	75.000	91.000
2020-782	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	29	79.207	2.310	2.92	71.000	87.000
2020-782	5	HEXOKINASE	191 RX-DAYTONA	126	RANDOX	26	86.000	2.935	3.41	77.000	95.000
2020-782	5	HEXOKINASE	203 ROCHE	93	ROCHE	19	78.684	2.226	2.83	71.000	87.000
2020-782	5	HEXOKINASE	212 S. D. EXL	143	SIEMENS	14	83.286	1.385	1.66	75.000	92.000
2020-782	6	HEXOKI. CO	194 BIOLIS-24S	154	(IDG)	11	77.909	2.429	3.12	70.000	86.000
2020-782	12	HEXOKINASE	184 INTEGRA	93	ROCHE	26	80.692	2.893	3.58	73.000	89.000
2020-782	12	HEXOKINASE	197 ARCHITECT	1	ABBOTT	18	77.000	1.453	1.89	69.000	85.000
2020-782	12	HEXOKINASE	203 ROCHE	93	ROCHE	16	79.063	3.092	3.91	71.000	87.000
2020-783	2	GL. OXI. E	215 B.C DXC	134	BECKM.COUL	12	135.083	5.267	3.90	122.000	149.000
2020-783	3	GL. OXI. C	165 VITROS	63	KODAK	166	136.928	3.510	2.56	123.000	151.000
2020-783	3	GL. OXI. C	213 SELECTRA P	150	ELITECH	12	137.333	4.422	3.22	124.000	151.000
2020-783	3	GL. OXI. C	214 Envoy 500	150	ELITECH	12	142.000	4.778	3.37	128.000	156.000
2020-783	5	HEXOKINASE	38 SIEMENS DI	143	SIEMENS	81	140.691	4.259	3.03	127.000	155.000
2020-783	5	HEXOKINASE	186 BECK. COUL	76	AU-BECKMAN	29	135.759	4.174	3.07	122.000	149.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-783	5	HEXOKINASE	191 RX-DAYTONA	126 RANDOX	26	142.423	4.325	3.04	128.000	157.000	
2020-783	5	HEXOKINASE	203 ROCHE	93 ROCHE	19	136.316	3.908	2.87	123.000	150.000	
2020-783	5	HEXOKINASE	212 S. D. EXL	143 SIEMENS	14	142.214	2.908	2.04	128.000	156.000	
2020-783	6	HEXOKI. CO	194 BIOLIS-24S	154 (IDG)	11	133.727	4.826	3.61	120.000	147.000	
2020-783	12	HEXOKINASE	184 INTEGRA	93 ROCHE	26	138.038	4.155	3.01	124.000	152.000	
2020-783	12	HEXOKINASE	197 ARCHITECT	1 ABBOTT	18	131.667	1.972	1.50	119.000	145.000	
2020-783	12	HEXOKINASE	203 ROCHE	93 ROCHE	16	135.563	4.937	3.64	122.000	149.000	
017	GLUCOSE		CLIA TEST ID #: 415			EVALUATION CRITERIA: TARGET VALUE 10 % OR 6.00 {GREATER}					
2020-784	2	GL. OXI. E	215 B.C DXC	134 BECKM.COUL	12	155.583	3.730	2.40	140.000	171.000	
2020-784	3	GL. OXI. C	165 VITROS	63 KODAK	166	160.783	4.308	2.68	145.000	177.000	
2020-784	3	GL. OXI. C	213 SELECTRA P	150 ELITECH	12	158.500	5.041	3.18	143.000	174.000	
2020-784	3	GL. OXI. C	214 Envoy 500	150 ELITECH	12	165.917	4.499	2.71	149.000	183.000	
2020-784	5	HEXOKINASE	38 SIEMENS DI	143 SIEMENS	81	162.716	4.821	2.96	146.000	179.000	
2020-784	5	HEXOKINASE	186 BECK. COUL	76 AU-BECKMAN	29	156.655	4.971	3.17	141.000	172.000	
2020-784	5	HEXOKINASE	191 RX-DAYTONA	126 RANDOX	26	163.731	4.486	2.74	147.000	180.000	
2020-784	5	HEXOKINASE	203 ROCHE	93 ROCHE	19	157.632	4.726	3.00	142.000	173.000	
2020-784	5	HEXOKINASE	212 S. D. EXL	143 SIEMENS	14	163.214	3.949	2.42	147.000	180.000	
2020-784	6	HEXOKI. CO	194 BIOLIS-24S	154 (IDG)	11	153.000	4.862	3.18	138.000	168.000	
2020-784	12	HEXOKINASE	184 INTEGRA	93 ROCHE	26	160.000	4.961	3.10	144.000	176.000	
2020-784	12	HEXOKINASE	197 ARCHITECT	1 ABBOTT	18	153.167	3.403	2.22	138.000	168.000	
2020-784	12	HEXOKINASE	203 ROCHE	93 ROCHE	16	158.000	5.958	3.77	142.000	174.000	
2020-785	2	GL. OXI. E	215 B.C DXC	134 BECKM.COUL	12	96.500	2.179	2.26	87.000	106.000	
2020-785	3	GL. OXI. C	165 VITROS	63 KODAK	166	97.831	2.531	2.59	88.000	108.000	
2020-785	3	GL. OXI. C	213 SELECTRA P	150 ELITECH	12	100.583	3.639	3.62	91.000	111.000	
2020-785	3	GL. OXI. C	214 Envoy 500	150 ELITECH	12	104.833	3.158	3.01	94.000	115.000	
2020-785	5	HEXOKINASE	38 SIEMENS DI	143 SIEMENS	81	102.750	2.700	2.63	92.000	113.000	
2020-785	5	HEXOKINASE	186 BECK. COUL	76 AU-BECKMAN	29	98.793	2.657	2.69	89.000	109.000	
2020-785	5	HEXOKINASE	191 RX-DAYTONA	126 RANDOX	26	105.577	2.858	2.71	95.000	116.000	
2020-785	5	HEXOKINASE	203 ROCHE	93 ROCHE	19	98.526	1.983	2.01	89.000	108.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-785	5	HEXOKINASE	212	S. D. EXL	143	SIEMENS	14	103.643	2.057	1.98	93.000	114.000
2020-785	6	HEXOKI. CO	194	BIOLIS-24S	154	(IDG)	11	97.000	2.629	2.71	87.000	107.000
2020-785	12	HEXOKINASE	184	INTEGRA	93	ROCHE	26	101.320	2.428	2.40	91.000	111.000
2020-785	12	HEXOKINASE	197	ARCHITECT	1	ABBOTT	18	96.529	1.753	1.82	87.000	106.000
2020-785	12	HEXOKINASE	203	ROCHE	93	ROCHE	16	97.500	3.500	3.59	88.000	107.000
018	IRON, TOTAL		CLIA TEST ID #: 425			EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-781	4	COULOMETR.	165	VITROS	63	KODAK	14	225.071	5.849	2.60	180.000	270.000
2020-782	4	COULOMETR.	165	VITROS	63	KODAK	14	111.571	3.831	3.43	89.000	134.000
2020-783	4	COULOMETR.	165	VITROS	63	KODAK	14	258.071	8.836	3.42	206.000	310.000
2020-784	4	COULOMETR.	165	VITROS	63	KODAK	14	327.786	8.342	2.55	262.000	393.000
2020-785	4	COULOMETR.	165	VITROS	63	KODAK	14	156.714	4.891	3.12	125.000	188.000
019	LACTASE DEHY. (LDH)		CLIA TEST ID #: 435									
2020-781	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	376.222	28.031	7.45	301.000	451.000
2020-781	3	KINETIC-37	194	BIOLIS-24S	154	(IDG)	10	419.400	14.108	3.36	336.000	503.000
2020-781	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	376.000	13.994	3.72	301.000	451.000
2020-781	6	NADH	184	INTEGRA	93	ROCHE	14	476.500	17.488	3.67	381.000	572.000
2020-781	6	NADH	203	ROCHE	93	ROCHE	22	468.318	15.200	3.25	375.000	562.000
2020-781	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	32	448.625	19.931	4.44	359.000	538.000
2020-781	9	LDH1	165	VITROS	63	KODAK	68	552.779	12.505	2.26	442.000	663.000
2020-782	3	KINETIC-37	186	BECK. COUL	76	AU-BECKMAN	18	181.778	13.439	7.39	145.000	218.000
2020-782	3	KINETIC-37	194	BIOLIS-24S	154	(IDG)	10	207.300	7.444	3.59	166.000	249.000
2020-782	3	KINETIC-37	215	B.C DXC	134	BECKM.COUL	11	185.000	6.836	3.69	148.000	222.000
2020-782	6	NADH	184	INTEGRA	93	ROCHE	14	237.857	7.170	3.01	190.000	285.000
2020-782	6	NADH	203	ROCHE	93	ROCHE	22	229.545	6.985	3.04	184.000	275.000
2020-782	7	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	32	218.438	11.316	5.18	175.000	262.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-782	9 LDH1	165 VITROS	63 KODAK	68	278.779	5.333	1.91	223.000	335.000		
019	LACTASE DEHY. (LDH)	CLIA TEST ID #: 435	EVALUATION CRITERIA: TARGET VALUE 20 %								
2020-783	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	18	428.941	28.753	6.70	343.000	515.000		
2020-783	3 KINETIC-37	194 BIOLIS-24S	154 (IDG)	10	475.700	19.427	4.08	381.000	571.000		
2020-783	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	11	428.364	25.737	6.01	343.000	514.000		
2020-783	6 NADH	184 INTEGRA	93 ROCHE	14	534.571	13.783	2.58	428.000	641.000		
2020-783	6 NADH	203 ROCHE	93 ROCHE	22	533.682	19.869	3.72	427.000	640.000		
2020-783	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	32	509.313	20.487	4.02	407.000	611.000		
2020-783	9 LDH1	165 VITROS	63 KODAK	68	625.206	19.286	3.08	500.000	750.000		
2020-784	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	18	521.000	41.409	7.95	417.000	625.000		
2020-784	3 KINETIC-37	194 BIOLIS-24S	154 (IDG)	10	583.600	25.457	4.36	467.000	700.000		
2020-784	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	11	514.636	17.844	3.47	412.000	618.000		
2020-784	6 NADH	184 INTEGRA	93 ROCHE	14	648.214	18.667	2.88	519.000	778.000		
2020-784	6 NADH	203 ROCHE	93 ROCHE	22	652.591	25.544	3.91	522.000	783.000		
2020-784	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	32	623.281	22.726	3.65	499.000	748.000		
2020-784	9 LDH1	165 VITROS	63 KODAK	68	753.250	37.124	4.93	603.000	904.000		
2020-785	3 KINETIC-37	186 BECK. COUL	76 AU-BECKMAN	18	264.389	19.068	7.21	212.000	317.000		
2020-785	3 KINETIC-37	194 BIOLIS-24S	154 (IDG)	10	299.400	10.865	3.63	240.000	359.000		
2020-785	3 KINETIC-37	215 B.C DXC	134 BECKM.COUL	11	270.545	9.365	3.46	216.000	325.000		
2020-785	6 NADH	184 INTEGRA	93 ROCHE	14	340.214	8.938	2.63	272.000	408.000		
2020-785	6 NADH	203 ROCHE	93 ROCHE	22	334.727	11.058	3.30	268.000	402.000		
2020-785	7 SIEMENS 37	38 SIEMENS DI	143 SIEMENS	32	318.188	13.192	4.15	255.000	382.000		
2020-785	9 LDH1	165 VITROS	63 KODAK	68	395.940	8.952	2.26	317.000	475.000		
021	MAGNESIUM	CLIA TEST ID #: 455	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-781	2 CALMAGITE	215 B.C DXC	134 BECKM.COUL	11	4.755	0.108	2.26	3.600	5.900		
2020-781	5 COLOR./K E	165 VITROS	63 KODAK	22	5.309	0.120	2.27	4.000	6.600		
2020-782	2 CALMAGITE	215 B.C DXC	134 BECKM.COUL	11	3.591	0.090	2.51	2.700	4.500		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-782	5	COLOR./K E	165	VITROS	63	KODAK	22	4.032	0.122	3.03	3.000	5.000
021	MAGNESIUM	CLIA TEST ID #: 455				EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-783	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	5.018	0.164	3.27	3.800	6.300
2020-783	5	COLOR./K E	165	VITROS	63	KODAK	22	5.591	0.165	2.95	4.200	7.000
2020-784	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	5.582	0.083	1.49	4.200	7.000
2020-784	5	COLOR./K E	165	VITROS	63	KODAK	22	6.082	0.201	3.31	4.600	7.600
2020-785	2	CALMAGITE	215	B.C DXC	134	BECKM.COUL	11	4.136	0.115	2.78	3.100	5.200
2020-785	5	COLOR./K E	165	VITROS	63	KODAK	22	4.550	0.103	2.27	3.400	5.700
022	POTASSIUM	CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00						
2020-781	3	ISE DILUT.	184	INTEGRA	93	ROCHE	35	5.321	0.087	1.63	4.300	6.300
2020-781	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	29	5.231	0.105	2.01	4.200	6.200
2020-781	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	15	5.247	0.120	2.29	4.200	6.200
2020-781	3	ISE DILUT.	203	ROCHE	93	ROCHE	36	5.336	0.111	2.08	4.300	6.300
2020-781	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	5.258	0.064	1.22	4.300	6.300
2020-781	4	ISE UNDIL.	165	VITROS	63	KODAK	134	5.301	0.085	1.60	4.300	6.300
2020-781	4	ISE UNDIL.	165	VITROS	156	Kit 32	27	5.312	0.085	1.59	4.300	6.300
2020-781	4	ISE UNDIL.	191	RX-DAYTONA	126	RANDOX	33	5.194	0.122	2.36	4.200	6.200
2020-781	4	ISE UNDIL.	199	Rx-IMOLA	126	RANDOX	12	5.167	0.149	2.89	4.200	6.200
2020-781	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	5.226	0.099	1.90	4.200	6.200
2020-781	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	5.236	0.081	1.55	4.200	6.200
2020-782	3	ISE DILUT.	184	INTEGRA	93	ROCHE	35	4.162	0.064	1.54	3.200	5.200
2020-782	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	29	4.117	0.075	1.81	3.100	5.100
2020-782	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	15	4.113	0.088	2.15	3.100	5.100
2020-782	3	ISE DILUT.	203	ROCHE	93	ROCHE	36	4.191	0.113	2.70	3.200	5.200
2020-782	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	4.042	0.086	2.13	3.000	5.000
2020-782	4	ISE UNDIL.	165	VITROS	63	KODAK	134	4.123	0.068	1.66	3.100	5.100
2020-782	4	ISE UNDIL.	165	VITROS	156	Kit 32	27	4.144	0.068	1.65	3.100	5.100

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-782	4	ISE UNLIL.	191	RX-DAYTONA	126	RANLDOX	33	4.081	0.098	2.41	3.100	5.100
2020-782	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLDOX	12	4.075	0.136	3.34	3.100	5.100
2020-782	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	4.101	0.064	1.56	3.100	5.100
2020-782	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	4.100	0.076	1.84	3.100	5.100
022	POTASSIUM	CLIA TEST ID #: 465				EVALUATION CRITERIA: TARGET VALUE 1.00						
2020-783	3	ISE DILUT.	184	INTEGRA	93	ROCHE	35	5.523	0.093	1.68	4.500	6.500
2020-783	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	29	5.503	0.103	1.88	4.500	6.500
2020-783	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	15	5.540	0.108	1.96	4.500	6.500
2020-783	3	ISE DILUT.	203	ROCHE	93	ROCHE	36	5.611	0.117	2.09	4.600	6.600
2020-783	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	5.583	0.207	3.72	4.600	6.600
2020-783	4	ISE UNLIL.	165	VITROS	63	KODAK	134	5.583	0.108	1.93	4.600	6.600
2020-783	4	ISE UNLIL.	165	VITROS	156	Kit 32	27	5.615	0.111	1.98	4.600	6.600
2020-783	4	ISE UNLIL.	191	RX-DAYTONA	126	RANLDOX	33	5.453	0.130	2.38	4.500	6.500
2020-783	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLDOX	12	5.433	0.143	2.64	4.400	6.400
2020-783	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	5.508	0.105	1.90	4.500	6.500
2020-783	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	5.521	0.067	1.22	4.500	6.500
2020-784	3	ISE DILUT.	184	INTEGRA	93	ROCHE	35	6.047	0.081	1.34	5.000	7.000
2020-784	3	ISE DILUT.	186	BECK. COUL	76	AU-BECKMAN	29	6.000	0.123	2.05	5.000	7.000
2020-784	3	ISE DILUT.	197	ARCHITECT	1	ABBOTT	15	6.060	0.114	1.89	5.100	7.100
2020-784	3	ISE DILUT.	203	ROCHE	93	ROCHE	36	6.125	0.123	2.01	5.100	7.100
2020-784	3	ISE DILUT.	215	B.C DXC	134	BECKM.COUL	12	6.117	0.114	1.87	5.100	7.100
2020-784	4	ISE UNLIL.	165	VITROS	63	KODAK	134	6.134	0.118	1.92	5.100	7.100
2020-784	4	ISE UNLIL.	165	VITROS	156	Kit 32	27	6.130	0.101	1.65	5.100	7.100
2020-784	4	ISE UNLIL.	191	RX-DAYTONA	126	RANLDOX	33	6.003	0.105	1.74	5.000	7.000
2020-784	4	ISE UNLIL.	199	Rx-IMOLA	126	RANLDOX	12	5.933	0.239	4.03	4.900	6.900
2020-784	6	QUICK LYTE	38	SIEMENS DI	143	SIEMENS	78	6.032	0.109	1.80	5.000	7.000
2020-784	6	QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	5.971	0.162	2.72	5.000	7.000
2020-785	3	ISE DILUT.	184	INTEGRA	93	ROCHE	35	4.617	0.077	1.68	3.600	5.600

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-785	3 ISE DILUT.	186 BECK. COUL	76 AU-BECKMAN	29	4.590	0.071	1.55	3.600	5.600		
2020-785	3 ISE DILUT.	197 ARCHITECT	1 ABBOTT	15	4.600	0.103	2.25	3.600	5.600		
2020-785	3 ISE DILUT.	203 ROCHE	93 ROCHE	36	4.689	0.107	2.29	3.700	5.700		
2020-785	3 ISE DILUT.	215 B.C DXC	134 BECKM.COUL	12	4.592	0.064	1.39	3.600	5.600		
2020-785	4 ISE UNDIL.	165 VITROS	63 KODAK	134	4.629	0.073	1.58	3.600	5.600		
2020-785	4 ISE UNDIL.	165 VITROS	156 Kit 32	27	4.637	0.067	1.46	3.600	5.600		
2020-785	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX	33	4.563	0.093	2.03	3.600	5.600		
2020-785	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX	12	4.525	0.136	3.01	3.500	5.500		
2020-785	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	78	4.593	0.061	1.34	3.600	5.600		
2020-785	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	4.600	0.085	1.84	3.600	5.600		
023 SODIUM		CLIA TEST ID #: 475	EVALUATION CRITERIA: TARGET VALUE 4.00								
2020-781	3 ISE DILUTE	184 INTEGRA	93 ROCHE	32	154.613	1.826	1.18	151.000	159.000		
2020-781	3 ISE DILUTE	184 INTEGRA	93 ROCHE	32	153.800	4.354	2.83	150.000	158.000		
2020-781	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	155.000	2.841	1.83	151.000	159.000		
2020-781	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	156.400	4.160	2.66	152.000	160.000		
2020-781	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	155.000	1.414	0.91	151.000	159.000		
2020-781	3 ISE DILUTE	203 ROCHE	93 ROCHE	27	155.333	1.866	1.20	151.000	159.000		
2020-781	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL	12	153.583	1.754	1.14	150.000	158.000		
2020-781	4 ISE UNDIL.	165 VITROS	63 KODAK	124	154.138	2.069	1.34	150.000	158.000		
2020-781	4 ISE UNDIL.	165 VITROS	156 Kit 32	29	153.357	1.493	0.97	149.000	157.000		
2020-781	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX	28	153.179	3.423	2.23	149.000	157.000		
2020-781	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX	11	153.909	2.466	1.60	150.000	158.000		
2020-781	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	79	154.359	2.511	1.63	150.000	158.000		
2020-781	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS	79	154.000	0.000	0.00	150.000	158.000		
2020-781	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS	14	154.929	1.831	1.18	151.000	159.000		
2020-782	3 ISE DILUTE	184 INTEGRA	93 ROCHE	32	139.200	2.993	2.15	135.000	143.000		
2020-782	3 ISE DILUTE	184 INTEGRA	93 ROCHE	32	137.806	1.804	1.31	134.000	142.000		
2020-782	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN	29	138.571	1.635	1.18	135.000	143.000		
2020-782	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT	15	137.833	2.267	1.64	134.000	142.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-782	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	139.200	2.535	1.82	135.000	143.000	
2020-782	3 ISE DILUTE	203 ROCHE	93 ROCHE		27	139.000	2.309	1.66	135.000	143.000	
2020-782	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	136.583	2.783	2.04	133.000	141.000	
2020-782	4 ISE UNDIL.	165 VITROS	63 KODAK		124	137.041	1.810	1.32	133.000	141.000	
2020-782	4 ISE UNDIL.	165 VITROS	156 Kit 32		29	136.655	1.372	1.00	133.000	141.000	
2020-782	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		28	136.214	2.610	1.92	132.000	140.000	
2020-782	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		11	137.273	3.107	2.26	133.000	141.000	
2020-782	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	139.667	0.471	0.34	136.000	144.000	
2020-782	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	138.975	1.903	1.37	135.000	143.000	
2020-782	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	138.857	1.846	1.33	135.000	143.000	
023 SODIUM		CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00							
2020-783	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	156.531	2.500	1.60	153.000	161.000	
2020-783	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	156.400	2.332	1.49	152.000	160.000	
2020-783	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	158.172	2.464	1.56	154.000	162.000	
2020-783	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	160.200	3.166	1.98	156.000	164.000	
2020-783	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	159.167	3.625	2.28	155.000	163.000	
2020-783	3 ISE DILUTE	203 ROCHE	93 ROCHE		27	159.296	2.565	1.61	155.000	163.000	
2020-783	3 ISE DILUTE	215 B.C DXC	134 BECKM.COUL		12	158.833	5.520	3.48	155.000	163.000	
2020-783	4 ISE UNDIL.	165 VITROS	63 KODAK		124	157.548	2.384	1.51	154.000	162.000	
2020-783	4 ISE UNDIL.	165 VITROS	156 Kit 32		29	156.893	1.934	1.23	153.000	161.000	
2020-783	4 ISE UNDIL.	191 RX-DAYTONA	126 RANDOX		28	156.607	2.807	1.79	153.000	161.000	
2020-783	4 ISE UNDIL.	199 Rx-IMOLA	126 RANDOX		11	156.909	2.429	1.55	153.000	161.000	
2020-783	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	157.500	2.716	1.72	154.000	162.000	
2020-783	6 QUICK LYTE	38 SIEMENS DI	143 SIEMENS		79	157.000	2.160	1.38	153.000	161.000	
2020-783	6 QUICK LYTE	212 S. D. EXL	143 SIEMENS		14	158.357	1.875	1.18	154.000	162.000	
2020-784	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	162.400	2.417	1.49	158.000	166.000	
2020-784	3 ISE DILUTE	184 INTEGRA	93 ROCHE		32	162.548	1.775	1.09	159.000	167.000	
2020-784	3 ISE DILUTE	186 BECK. COUL	76 AU-BECKMAN		29	163.966	3.168	1.93	160.000	168.000	
2020-784	3 ISE DILUTE	197 ARCHITECT	1 ABBOTT		15	165.000	2.769	1.68	161.000	169.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-784	3 ISE DILUTE	197	ARCHITECT	1	ABBOTT	15	166.200	3.410	2.05	162.000	170.000	
2020-784	3 ISE DILUTE	203	ROCHE	93	ROCHE	27	165.741	2.702	1.63	162.000	170.000	
2020-784	3 ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	164.583	1.847	1.12	161.000	169.000	
2020-784	4 ISE UNDIL.	165	VITROS	63	KODAK	124	164.270	2.564	1.56	160.000	168.000	
2020-784	4 ISE UNDIL.	165	VITROS	156	Kit 32	29	163.310	2.119	1.30	159.000	167.000	
2020-784	4 ISE UNDIL.	191	RX-DAYTONA	126	RANDOX	28	163.357	2.979	1.82	159.000	167.000	
2020-784	4 ISE UNDIL.	199	Rx-IMOLA	126	RANDOX	11	163.455	4.250	2.60	159.000	167.000	
2020-784	6 QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	162.000	0.816	0.50	158.000	166.000	
2020-784	6 QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	162.859	2.453	1.51	159.000	167.000	
2020-784	6 QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	163.357	2.991	1.83	159.000	167.000	
023	SODIUM	CLIA TEST ID #: 475		EVALUATION CRITERIA: TARGET VALUE 4.00								
2020-785	3 ISE DILUTE	184	INTEGRA	93	ROCHE	32	143.938	2.263	1.57	140.000	148.000	
2020-785	3 ISE DILUTE	184	INTEGRA	93	ROCHE	32	143.800	2.315	1.61	140.000	148.000	
2020-785	3 ISE DILUTE	186	BECK. COUL	76	AU-BECKMAN	29	145.517	1.774	1.22	142.000	150.000	
2020-785	3 ISE DILUTE	197	ARCHITECT	1	ABBOTT	15	146.400	2.498	1.71	142.000	150.000	
2020-785	3 ISE DILUTE	197	ARCHITECT	1	ABBOTT	15	145.167	2.409	1.66	141.000	149.000	
2020-785	3 ISE DILUTE	203	ROCHE	93	ROCHE	27	145.963	2.081	1.43	142.000	150.000	
2020-785	3 ISE DILUTE	215	B.C DXC	134	BECKM.COUL	12	144.167	2.115	1.47	140.000	148.000	
2020-785	4 ISE UNDIL.	165	VITROS	63	KODAK	124	144.179	1.967	1.36	140.000	148.000	
2020-785	4 ISE UNDIL.	165	VITROS	156	Kit 32	29	143.621	1.215	0.85	140.000	148.000	
2020-785	4 ISE UNDIL.	191	RX-DAYTONA	126	RANDOX	28	143.857	2.386	1.66	140.000	148.000	
2020-785	4 ISE UNDIL.	199	Rx-IMOLA	126	RANDOX	11	143.818	2.249	1.56	140.000	148.000	
2020-785	6 QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	145.354	1.807	1.24	141.000	149.000	
2020-785	6 QUICK LYTE	38	SIEMENS DI	143	SIEMENS	79	145.333	0.471	0.32	141.000	149.000	
2020-785	6 QUICK LYTE	212	S. D. EXL	143	SIEMENS	14	146.214	1.897	1.30	142.000	150.000	
024	TOTAL PROTEIN	CLIA TEST ID #: 485		EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-781	1 BIURET	38	SIEMENS DI	143	SIEMENS	79	6.123	0.134	2.19	5.500	6.700	
2020-781	1 BIURET	165	VITROS	63	KODAK	144	5.827	0.130	2.22	5.200	6.400	
2020-781	1 BIURET	184	INTEGRA	93	ROCHE	28	5.650	0.170	3.01	5.100	6.200	
2020-781	1 BIURET	186	BECK. COUL	76	AU-BECKMAN	25	5.808	0.162	2.79	5.200	6.400	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-781	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	5.916	0.186	3.14	5.300	6.500
2020-781	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	6.000	0.294	4.91	5.400	6.600
2020-781	1	BIURET	197	ARCHITECT	1	ABBOTT	20	5.905	0.216	3.65	5.300	6.500
2020-781	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	5.873	0.286	4.88	5.300	6.500
2020-781	1	BIURET	203	ROCHE	93	ROCHE	29	5.896	0.124	2.10	5.300	6.500
2020-781	1	BIURET	212	S. D. EXL	143	SIEMENS	14	6.150	0.098	1.60	5.500	6.800
2020-781	1	BIURET	214	Envoy 500	150	ELITECH	10	6.020	0.189	3.13	5.400	6.600
2020-781	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	5.900	0.187	3.17	5.300	6.500
024 TOTAL PROTEIN			CLIA TEST ID #: 485				EVALUATION CRITERIA: TARGET VALUE 10 %					
2020-782	1	BIURET	38	SIEMENS DI	143	SIEMENS	79	4.081	0.097	2.37	3.700	4.500
2020-782	1	BIURET	165	VITROS	63	KODAK	144	3.958	0.108	2.74	3.600	4.400
2020-782	1	BIURET	184	INTEGRA	93	ROCHE	28	3.761	0.150	3.98	3.400	4.100
2020-782	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	25	3.840	0.102	2.66	3.500	4.200
2020-782	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	4.083	0.134	3.29	3.700	4.500
2020-782	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	3.941	0.127	3.23	3.500	4.300
2020-782	1	BIURET	197	ARCHITECT	1	ABBOTT	20	3.921	0.089	2.28	3.500	4.300
2020-782	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	3.973	0.218	5.48	3.600	4.400
2020-782	1	BIURET	203	ROCHE	93	ROCHE	29	3.893	0.084	2.16	3.500	4.300
2020-782	1	BIURET	212	S. D. EXL	143	SIEMENS	14	4.100	0.065	1.60	3.700	4.500
2020-782	1	BIURET	214	Envoy 500	150	ELITECH	10	4.060	0.185	4.57	3.700	4.500
2020-782	1	BIURET	215	B.C DXC	134	BECKM.COUL	12	3.908	0.132	3.38	3.500	4.300
2020-783	1	BIURET	38	SIEMENS DI	143	SIEMENS	79	6.654	0.161	2.41	6.000	7.300
2020-783	1	BIURET	165	VITROS	63	KODAK	144	6.302	0.142	2.25	5.700	6.900
2020-783	1	BIURET	184	INTEGRA	93	ROCHE	28	6.132	0.163	2.65	5.500	6.700
2020-783	1	BIURET	186	BECK. COUL	76	AU-BECKMAN	25	6.272	0.148	2.37	5.600	6.900
2020-783	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	6.481	0.238	3.67	5.800	7.100
2020-783	1	BIURET	191	RX-DAYTONA	126	RANDOX	33	6.583	0.313	4.76	5.900	7.200
2020-783	1	BIURET	197	ARCHITECT	1	ABBOTT	20	6.484	0.160	2.46	5.800	7.100
2020-783	1	BIURET	199	Rx-IMOLA	126	RANDOX	11	6.409	0.394	6.15	5.800	7.100

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-783	1 BIURET	203 ROCHE	93 ROCHE	29	6.389	0.129	2.02	5.800	7.000		
2020-783	1 BIURET	212 S. D. EXL	143 SIEMENS	14	6.707	0.110	1.64	6.000	7.400		
2020-783	1 BIURET	214 Envoy 500	150 ELITECH	10	6.560	0.341	5.20	5.900	7.200		
2020-783	1 BIURET	215 B.C DXC	134 BECKM.COUL	12	6.475	0.245	3.79	5.800	7.100		
024 TOTAL PROTEIN		CLIA TEST ID #: 485	EVALUATION CRITERIA: TARGET VALUE 10 %								
2020-784	1 BIURET	38 SIEMENS DI	143 SIEMENS	79	7.651	0.183	2.39	6.900	8.400		
2020-784	1 BIURET	165 VITROS	63 KODAK	144	7.227	0.181	2.51	6.500	7.900		
2020-784	1 BIURET	184 INTEGRA	93 ROCHE	28	7.011	0.190	2.70	6.300	7.700		
2020-784	1 BIURET	186 BECK. COUL	76 AU-BECKMAN	25	7.216	0.189	2.62	6.500	7.900		
2020-784	1 BIURET	191 RX-DAYTONA	126 RANDOX	33	7.400	0.447	6.04	6.700	8.100		
2020-784	1 BIURET	191 RX-DAYTONA	126 RANDOX	33	7.322	0.327	4.46	6.600	8.100		
2020-784	1 BIURET	197 ARCHITECT	1 ABBOTT	20	7.445	0.254	3.41	6.700	8.200		
2020-784	1 BIURET	199 Rx-IMOLA	126 RANDOX	11	7.309	0.353	4.83	6.600	8.000		
2020-784	1 BIURET	203 ROCHE	93 ROCHE	29	7.307	0.144	1.97	6.600	8.000		
2020-784	1 BIURET	212 S. D. EXL	143 SIEMENS	14	7.686	0.106	1.38	6.900	8.500		
2020-784	1 BIURET	214 Envoy 500	150 ELITECH	10	7.560	0.314	4.15	6.800	8.300		
2020-784	1 BIURET	215 B.C DXC	134 BECKM.COUL	12	7.467	0.209	2.81	6.700	8.200		
2020-785	1 BIURET	38 SIEMENS DI	143 SIEMENS	79	4.960	0.097	1.95	4.500	5.500		
2020-785	1 BIURET	165 VITROS	63 KODAK	144	4.796	0.130	2.71	4.300	5.300		
2020-785	1 BIURET	184 INTEGRA	93 ROCHE	28	4.611	0.150	3.24	4.100	5.100		
2020-785	1 BIURET	186 BECK. COUL	76 AU-BECKMAN	25	4.704	0.096	2.04	4.200	5.200		
2020-785	1 BIURET	191 RX-DAYTONA	126 RANDOX	33	4.797	0.179	3.74	4.300	5.300		
2020-785	1 BIURET	191 RX-DAYTONA	126 RANDOX	33	4.883	0.254	5.21	4.400	5.400		
2020-785	1 BIURET	197 ARCHITECT	1 ABBOTT	20	4.784	0.109	2.28	4.300	5.300		
2020-785	1 BIURET	199 Rx-IMOLA	126 RANDOX	11	4.791	0.239	4.99	4.300	5.300		
2020-785	1 BIURET	203 ROCHE	93 ROCHE	29	4.779	0.161	3.36	4.300	5.300		
2020-785	1 BIURET	212 S. D. EXL	143 SIEMENS	14	5.014	0.106	2.11	4.500	5.500		
2020-785	1 BIURET	214 Envoy 500	150 ELITECH	10	4.960	0.201	4.05	4.500	5.500		
2020-785	1 BIURET	215 B.C DXC	134 BECKM.COUL	12	4.758	0.132	2.77	4.300	5.200		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
025 TRIGLYCERIDES		CLIA TEST ID #: 495	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-781	2 ENZ. COLO.	165 VITROS	63 KODAK		146	200.726	6.611	3.29	151.000	251.000	
2020-781	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	169.235	3.116	1.84	127.000	212.000	
2020-781	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		77	165.440	4.500	2.72	124.000	207.000	
2020-781	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		15	165.071	3.283	1.99	124.000	206.000	
2020-781	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		27	171.926	5.242	3.05	129.000	215.000	
2020-781	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		23	164.652	7.475	4.54	123.000	206.000	
2020-781	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)		10	147.700	8.247	5.58	111.000	185.000	
2020-781	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		11	182.545	4.500	2.47	137.000	228.000	
2020-781	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		25	169.040	5.923	3.50	127.000	211.000	
2020-781	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		12	168.083	8.607	5.12	126.000	210.000	
2020-781	10 ENZYM. GPO	203 ROCHE	93 ROCHE		18	167.941	4.783	2.85	126.000	210.000	
2020-781	10 ENZYM. GPO	214 Envoy 500	150 ELITECH		11	173.091	5.680	3.28	130.000	216.000	
2020-781	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		11	163.455	6.140	3.76	123.000	204.000	
2020-782	2 ENZ. COLO.	165 VITROS	63 KODAK		146	136.938	5.392	3.94	103.000	171.000	
2020-782	2 ENZ. COLO.	203 ROCHE	93 ROCHE		17	120.125	2.870	2.39	90.000	150.000	
2020-782	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS		77	113.400	3.237	2.85	85.000	142.000	
2020-782	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS		15	112.286	2.986	2.66	84.000	140.000	
2020-782	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN		27	119.519	4.246	3.55	90.000	149.000	
2020-782	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX		23	117.783	6.338	5.38	88.000	147.000	
2020-782	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)		10	105.300	6.310	5.99	79.000	132.000	
2020-782	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL		11	125.000	3.742	2.99	94.000	156.000	
2020-782	10 ENZYM. GPO	184 INTEGRA	93 ROCHE		25	120.840	5.143	4.26	91.000	151.000	
2020-782	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX		12	118.667	6.872	5.79	89.000	148.000	
2020-782	10 ENZYM. GPO	203 ROCHE	93 ROCHE		18	120.471	3.183	2.64	90.000	151.000	
2020-782	10 ENZYM. GPO	214 Envoy 500	150 ELITECH		11	121.727	4.202	3.45	91.000	152.000	
2020-782	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT		11	113.818	3.242	2.85	85.000	142.000	
2020-783	2 ENZ. COLO.	165 VITROS	63 KODAK		146	217.151	6.450	2.97	163.000	271.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-783	2 ENZ. COLO.	203 ROCHE	93 ROCHE	17	180.941	4.249	2.35	136.000	226.000		
2020-783	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	77	178.303	4.877	2.74	134.000	223.000		
2020-783	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS	15	177.857	3.701	2.08	133.000	222.000		
2020-783	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN	27	184.815	5.818	3.15	139.000	231.000		
2020-783	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX	23	178.870	10.331	5.78	134.000	224.000		
2020-783	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)	10	157.300	9.424	5.99	118.000	197.000		
2020-783	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL	11	196.182	7.517	3.83	147.000	245.000		
2020-783	10 ENZYM. GPO	184 INTEGRA	93 ROCHE	25	180.720	6.207	3.43	136.000	226.000		
2020-783	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX	12	175.833	8.552	4.86	132.000	220.000		
2020-783	10 ENZYM. GPO	203 ROCHE	93 ROCHE	18	179.706	2.739	1.52	135.000	225.000		
2020-783	10 ENZYM. GPO	214 Envoy 500	150 ELITECH	11	186.727	6.440	3.45	140.000	233.000		
2020-783	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT	11	176.273	4.555	2.58	132.000	220.000		
025 TRIGLYCERIDES		CLIA TEST ID #: 495	EVALUATION CRITERIA: TARGET VALUE 25 %								
2020-784	2 ENZ. COLO.	165 VITROS	63 KODAK	146	249.144	6.921	2.78	187.000	311.000		
2020-784	2 ENZ. COLO.	203 ROCHE	93 ROCHE	17	201.647	4.043	2.00	151.000	252.000		
2020-784	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	77	202.680	5.034	2.48	152.000	253.000		
2020-784	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS	15	201.000	4.766	2.37	151.000	251.000		
2020-784	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN	27	210.000	6.566	3.13	158.000	263.000		
2020-784	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX	23	197.609	10.709	5.42	148.000	247.000		
2020-784	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)	10	177.300	10.149	5.72	133.000	222.000		
2020-784	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL	11	222.455	6.814	3.06	167.000	278.000		
2020-784	10 ENZYM. GPO	184 INTEGRA	93 ROCHE	25	201.880	5.955	2.95	151.000	252.000		
2020-784	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX	12	199.500	7.837	3.93	150.000	249.000		
2020-784	10 ENZYM. GPO	203 ROCHE	93 ROCHE	18	200.647	4.297	2.14	150.000	251.000		
2020-784	10 ENZYM. GPO	214 Envoy 500	150 ELITECH	11	212.727	5.189	2.44	160.000	266.000		
2020-784	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT	11	199.000	4.452	2.24	149.000	249.000		
2020-785	2 ENZ. COLO.	165 VITROS	63 KODAK	146	164.884	5.970	3.62	124.000	206.000		
2020-785	2 ENZ. COLO.	203 ROCHE	93 ROCHE	17	142.176	3.535	2.49	107.000	178.000		
2020-785	4 ENZ. (GPD)	38 SIEMENS DI	143 SIEMENS	77	135.368	3.228	2.38	102.000	169.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-785	4 ENZ. (GPD)	212 S. D. EXL	143 SIEMENS	15	134.357	2.409	1.79	101.000	168.000		
2020-785	9 GPO TRIND.	186 BECK. COUL	76 AU-BECKMAN	27	142.444	3.909	2.74	107.000	178.000		
2020-785	9 GPO TRIND.	191 RX-DAYTONA	126 RANDOX	23	138.217	4.736	3.43	104.000	173.000		
2020-785	9 GPO TRIND.	194 BIOLIS-24S	154 (IDG)	10	125.100	6.759	5.40	94.000	156.000		
2020-785	9 GPO TRIND.	215 B.C DXC	134 BECKM.COUL	11	152.545	7.715	5.06	114.000	191.000		
2020-785	10 ENZYM. GPO	184 INTEGRA	93 ROCHE	25	142.200	4.988	3.51	107.000	178.000		
2020-785	10 ENZYM. GPO	191 RX-DAYTONA	126 RANDOX	12	137.167	6.530	4.76	103.000	171.000		
2020-785	10 ENZYM. GPO	203 ROCHE	93 ROCHE	18	140.824	3.148	2.24	106.000	176.000		
2020-785	10 ENZYM. GPO	214 Envoy 500	150 ELITECH	11	148.000	5.461	3.69	111.000	185.000		
2020-785	11 GLY PO4 OX	197 ARCHITECT	1 ABBOTT	11	135.000	3.593	2.66	101.000	169.000		
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505	EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}								
2020-781	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	81	31.275	0.935	2.99	28.000	34.000		
2020-781	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	37	30.514	1.056	3.46	28.000	33.000		
2020-781	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	31.138	1.332	4.28	28.000	34.000		
2020-781	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	34	29.939	1.127	3.76	27.000	33.000		
2020-781	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	34	30.333	0.471	1.55	28.000	33.000		
2020-781	2 UREAS.GLDH	203 ROCHE	93 ROCHE	36	30.111	1.308	4.34	27.000	33.000		
2020-781	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	14	31.286	0.700	2.24	28.000	34.000		
2020-781	2 UREAS.GLDH	214 Envoy 500	150 ELITECH	10	31.700	1.187	3.75	29.000	35.000		
2020-781	8 UREASE QUI	165 VITROS	63 KODAK	162	28.056	0.733	2.61	26.000	31.000		
2020-781	8 UREASE QUI	165 VITROS	63 KODAK	162	29.000	0.894	3.08	26.000	32.000		
2020-782	2 UREAS.GLDH	38 SIEMENS DI	143 SIEMENS	81	19.663	0.921	4.69	18.000	22.000		
2020-782	2 UREAS.GLDH	184 INTEGRA	93 ROCHE	37	19.216	0.776	4.04	17.000	21.000		
2020-782	2 UREAS.GLDH	186 BECK. COUL	76 AU-BECKMAN	29	19.759	0.677	3.43	18.000	22.000		
2020-782	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	34	19.667	0.471	2.40	18.000	22.000		
2020-782	2 UREAS.GLDH	191 RX-DAYTONA	126 RANDOX	34	19.265	1.093	5.67	17.000	21.000		
2020-782	2 UREAS.GLDH	203 ROCHE	93 ROCHE	36	19.111	1.125	5.89	17.000	21.000		
2020-782	2 UREAS.GLDH	212 S. D. EXL	143 SIEMENS	14	19.929	0.961	4.82	18.000	22.000		
2020-782	2 UREAS.GLDH	214 Envoy 500	150 ELITECH	10	20.200	0.748	3.70	18.000	22.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-782	8	UREASE QUI	165	VITROS	63	KODAK	162	18.800	0.400	2.13	17.000	21.000
2020-782	8	UREASE QUI	165	VITROS	63	KODAK	162	17.451	0.578	3.31	15.000	19.000
026 UREA NITROGEN (BUN)		CLIA TEST ID #: 505				EVALUATION CRITERIA: TARGET VALUE 9 % OR 2.00 {GREATER}						
2020-783	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	81	34.037	1.365	4.01	31.000	37.000
2020-783	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	37	33.216	1.094	3.29	30.000	36.000
2020-783	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	29	34.103	1.269	3.72	31.000	37.000
2020-783	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	34.000	0.000	0.00	31.000	37.000
2020-783	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	32.818	1.623	4.94	30.000	36.000
2020-783	2	UREAS.GLDH	203	ROCHE	93	ROCHE	36	32.833	1.590	4.84	30.000	36.000
2020-783	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	33.857	1.125	3.32	31.000	37.000
2020-783	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	34.400	1.356	3.94	31.000	37.000
2020-783	8	UREASE QUI	165	VITROS	63	KODAK	162	30.843	0.687	2.23	28.000	34.000
2020-783	8	UREASE QUI	165	VITROS	63	KODAK	162	32.800	1.327	4.04	30.000	36.000
2020-784	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	81	39.704	1.543	3.89	36.000	43.000
2020-784	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	37	38.730	1.287	3.32	35.000	42.000
2020-784	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	29	39.345	1.211	3.08	36.000	43.000
2020-784	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	38.333	0.471	1.23	35.000	42.000
2020-784	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	38.121	1.838	4.82	35.000	42.000
2020-784	2	UREAS.GLDH	203	ROCHE	93	ROCHE	36	38.417	1.656	4.31	35.000	42.000
2020-784	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	39.429	0.821	2.08	36.000	43.000
2020-784	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	39.700	1.676	4.22	36.000	43.000
2020-784	8	UREASE QUI	165	VITROS	63	KODAK	162	37.200	0.748	2.01	34.000	41.000
2020-784	8	UREASE QUI	165	VITROS	63	KODAK	162	36.388	0.949	2.61	33.000	40.000
2020-785	2	UREAS.GLDH	38	SIEMENS DI	143	SIEMENS	81	24.613	0.981	3.99	22.000	27.000
2020-785	2	UREAS.GLDH	184	INTEGRA	93	ROCHE	37	24.054	0.868	3.61	22.000	26.000
2020-785	2	UREAS.GLDH	186	BECK. COUL	76	AU-BECKMAN	29	24.655	0.755	3.06	22.000	27.000
2020-785	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	24.182	0.903	3.73	22.000	26.000
2020-785	2	UREAS.GLDH	191	RX-DAYTONA	126	RANDOX	34	24.333	0.471	1.94	22.000	27.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 10 ROUTINE

GRP: 10 ROUTINE

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-785	2	UREAS.GLDH	203	ROCHE	93	ROCHE	36	23.857	1.073	4.50	22.000	26.000
2020-785	2	UREAS.GLDH	212	S. D. EXL	143	SIEMENS	14	24.714	1.161	4.70	22.000	27.000
2020-785	2	UREAS.GLDH	214	Envoy 500	150	ELITECH	10	25.400	1.200	4.72	23.000	28.000
2020-785	8	UREASE QUI	165	VITROS	63	KODAK	162	21.870	0.612	2.80	20.000	24.000
2020-785	8	UREASE QUI	165	VITROS	63	KODAK	162	22.800	0.748	3.28	21.000	25.000
027	URIC ACID	CLIA TEST ID #: 515				EVALUATION CRITERIA: TARGET VALUE 17 %						
2020-781	2	URICASE	38	SIEMENS DI	143	SIEMENS	45	7.882	0.147	1.86	6.500	9.200
2020-781	2	URICASE	165	VITROS	63	KODAK	99	7.832	0.190	2.43	6.500	9.200
2020-781	2	URICASE	184	INTEGRA	93	ROCHE	18	8.083	0.206	2.55	6.700	9.500
2020-781	2	URICASE	197	ARCHITECT	1	ABBOTT	10	7.900	0.195	2.47	6.600	9.200
2020-781	2	URICASE	203	ROCHE	93	ROCHE	23	7.717	0.279	3.62	6.400	9.000
2020-781	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	7.883	0.154	1.95	6.500	9.200
2020-781	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	11	7.364	0.098	1.33	6.100	8.600
2020-782	2	URICASE	38	SIEMENS DI	143	SIEMENS	45	4.758	0.129	2.71	3.900	5.600
2020-782	2	URICASE	165	VITROS	63	KODAK	99	4.757	0.107	2.25	3.900	5.600
2020-782	2	URICASE	184	INTEGRA	93	ROCHE	18	4.894	0.122	2.50	4.100	5.700
2020-782	2	URICASE	197	ARCHITECT	1	ABBOTT	10	4.850	0.067	1.38	4.000	5.700
2020-782	2	URICASE	203	ROCHE	93	ROCHE	23	4.709	0.132	2.79	3.900	5.500
2020-782	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	4.874	0.129	2.65	4.000	5.700
2020-782	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	11	4.609	0.090	1.95	3.800	5.400
2020-783	2	URICASE	38	SIEMENS DI	143	SIEMENS	45	8.698	0.213	2.45	7.200	10.200
2020-783	2	URICASE	165	VITROS	63	KODAK	99	8.634	0.226	2.61	7.200	10.100
2020-783	2	URICASE	184	INTEGRA	93	ROCHE	18	8.872	0.235	2.65	7.400	10.400
2020-783	2	URICASE	197	ARCHITECT	1	ABBOTT	10	8.700	0.161	1.85	7.200	10.200
2020-783	2	URICASE	203	ROCHE	93	ROCHE	23	8.509	0.234	2.75	7.100	10.000
2020-783	5	URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	8.605	0.268	3.12	7.100	10.100
2020-783	5	URICASE 37	215	B.C DXC	134	BECKM.COUL	11	8.082	0.248	3.07	6.700	9.500

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
027	URIC ACID	CLIA TEST ID #: 515				EVALUATION CRITERIA: TARGET VALUE 17 %						
2020-784	2 URICASE	38	SIEMENS DI	143	SIEMENS	45	10.168	0.208	2.04	8.400	11.900	
2020-784	2 URICASE	165	VITROS	63	KODAK	99	10.164	0.273	2.68	8.400	11.900	
2020-784	2 URICASE	184	INTEGRA	93	ROCHE	18	10.372	0.213	2.05	8.600	12.100	
2020-784	2 URICASE	197	ARCHITECT	1	ABBOTT	10	10.210	0.145	1.42	8.500	11.900	
2020-784	2 URICASE	203	ROCHE	93	ROCHE	23	10.004	0.265	2.64	8.300	11.700	
2020-784	5 URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	10.039	0.189	1.88	8.300	11.700	
2020-784	5 URICASE 37	215	B.C DXC	134	BECKM.COUL	11	9.336	0.143	1.53	7.700	10.900	
2020-785	2 URICASE	38	SIEMENS DI	143	SIEMENS	45	6.075	0.119	1.96	5.000	7.100	
2020-785	2 URICASE	165	VITROS	63	KODAK	99	6.084	0.147	2.41	5.000	7.100	
2020-785	2 URICASE	184	INTEGRA	93	ROCHE	18	6.289	0.149	2.36	5.200	7.400	
2020-785	2 URICASE	197	ARCHITECT	1	ABBOTT	10	6.210	0.158	2.54	5.200	7.300	
2020-785	2 URICASE	203	ROCHE	93	ROCHE	23	5.996	0.178	2.97	5.000	7.000	
2020-785	5 URICASE 37	186	BECK. COUL	76	AU-BECKMAN	19	6.216	0.184	2.96	5.200	7.300	
2020-785	5 URICASE 37	215	B.C DXC	134	BECKM.COUL	11	5.836	0.098	1.68	4.800	6.800	
135	GGT (GAMMA GLU. TRA)	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-781	5 KINETIC 37	165	VITROS	63	KODAK	23	396.565	6.927	1.75	317.000	476.000	
2020-782	5 KINETIC 37	165	VITROS	63	KODAK	23	196.000	3.718	1.90	157.000	235.000	
2020-783	5 KINETIC 37	165	VITROS	63	KODAK	23	442.636	9.763	2.21	354.000	531.000	
2020-784	5 KINETIC 37	165	VITROS	63	KODAK	23	525.957	19.694	3.74	421.000	631.000	
2020-785	5 KINETIC 37	165	VITROS	63	KODAK	23	289.696	5.221	1.80	232.000	348.000	
136	LACTIC ACID											
2020-781	1 LACTATE-P.	165	VITROS	63	KODAK	18	4.294	0.054	1.26	3.400	5.200	
2020-781	1 LACTATE-P.	165	VITROS	63	KODAK	18	4.550	0.076	1.68	3.600	5.500	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
136	LACTIC ACID	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-782	1	LACTATE-P.	165	VITROS	63	KODAK	18	2.483	0.069	2.77	2.000	3.000
2020-782	1	LACTATE-P.	165	VITROS	63	KODAK	18	2.294	0.040	1.76	1.800	2.800
2020-783	1	LACTATE-P.	165	VITROS	63	KODAK	18	4.756	0.117	2.45	3.800	5.700
2020-783	1	LACTATE-P.	165	VITROS	63	KODAK	18	5.167	0.094	1.82	4.100	6.200
2020-784	1	LACTATE-P.	165	VITROS	63	KODAK	18	6.067	0.094	1.55	4.900	7.300
2020-784	1	LACTATE-P.	165	VITROS	63	KODAK	18	5.659	0.103	1.82	4.500	6.800
2020-785	1	LACTATE-P.	165	VITROS	63	KODAK	18	3.188	0.047	1.48	2.600	3.800
2020-785	1	LACTATE-P.	165	VITROS	63	KODAK	18	3.417	0.037	1.09	2.700	4.100
137	LIPASE											
2020-781	1	COLORIMET.	203	ROCHE	93	ROCHE	13	55.692	7.770	13.95	45.000	67.000
2020-781	1	COLORIMET.	203	ROCHE	93	ROCHE	13	51.333	8.756	17.06	41.000	62.000
2020-781	2	ENZIMATIC	165	VITROS	63	KODAK	82	579.296	17.752	3.06	463.000	695.000
2020-781	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	61.083	3.148	5.15	49.000	73.000
2020-781	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	64.643	3.108	4.81	52.000	78.000
2020-781	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	55.214	8.187	14.83	44.000	66.000
2020-781	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	51.333	8.756	17.06	41.000	62.000
2020-781	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	232.250	14.754	6.35	186.000	279.000
2020-781	5	SIEMENS 37	38	SIEMENS DI	143	SIEMENS	18	208.167	16.724	8.03	167.000	250.000
2020-782	1	COLORIMET.	203	ROCHE	93	ROCHE	13	32.000	5.312	16.60	26.000	38.000
2020-782	1	COLORIMET.	203	ROCHE	93	ROCHE	13	32.308	3.871	11.98	26.000	39.000
2020-782	2	ENZIMATIC	165	VITROS	63	KODAK	82	321.610	10.623	3.30	257.000	386.000
2020-782	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	41.000	2.769	6.75	33.000	49.000
2020-782	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	35.357	2.158	6.10	28.000	42.000
2020-782	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	32.571	4.609	14.15	26.000	39.000
2020-782	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	32.000	5.312	16.60	26.000	38.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-782	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	122.000	11.542	9.46	98.000	146.000
2020-782	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	127.250	5.804	4.56	102.000	153.000
137	LIPASE	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-783	1	COLORIMET.	203	ROCHE	93	ROCHE	13	62.154	8.857	14.25	50.000	75.000	
2020-783	1	COLORIMET.	203	ROCHE	93	ROCHE	13	56.000	9.440	16.86	45.000	67.000	
2020-783	2	ENZIMATIC	165	VITROS	63	KODAK	82	638.439	20.575	3.22	511.000	766.000	
2020-783	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	65.417	2.465	3.77	52.000	79.000	
2020-783	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	71.714	2.281	3.18	57.000	86.000	
2020-783	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	61.071	8.844	14.48	49.000	73.000	
2020-783	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	56.000	9.440	16.86	45.000	67.000	
2020-783	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	262.750	18.713	7.12	210.000	315.000
2020-783	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	232.333	20.683	8.90	186.000	279.000
2020-784	1	COLORIMET.	203	ROCHE	93	ROCHE	13	66.444	11.006	16.56	53.000	80.000	
2020-784	1	COLORIMET.	203	ROCHE	93	ROCHE	13	72.615	9.402	12.95	58.000	87.000	
2020-784	2	ENZIMATIC	165	VITROS	63	KODAK	82	741.610	25.765	3.47	593.000	890.000	
2020-784	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	74.333	3.543	4.77	59.000	89.000	
2020-784	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	86.214	4.784	5.55	69.000	103.000	
2020-784	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	72.143	10.696	14.83	58.000	87.000	
2020-784	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	66.444	11.006	16.56	53.000	80.000	
2020-784	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	272.500	24.348	8.93	218.000	327.000
2020-784	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	315.000	14.018	4.45	252.000	378.000
2020-785	1	COLORIMET.	203	ROCHE	93	ROCHE	13	42.462	5.500	12.95	34.000	51.000	
2020-785	1	COLORIMET.	203	ROCHE	93	ROCHE	13	38.889	6.437	16.55	31.000	47.000	
2020-785	2	ENZIMATIC	165	VITROS	63	KODAK	82	437.667	12.560	2.87	350.000	525.000	
2020-785	2	ENZIMATIC	215	B.C DXC	134	BECKM.COUL	12	49.417	2.431	4.92	40.000	59.000	
2020-785	4	ENZ/COLOR	186	BECK. COUL	76	AU-BECKMAN	14	48.071	1.944	4.04	38.000	58.000	
2020-785	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	42.429	5.937	13.99	34.000	51.000	
2020-785	4	ENZ/COLOR	203	ROCHE	93	ROCHE	14	38.889	6.437	16.55	31.000	47.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE							
ANALYTE													
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
2020-785	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	175.000	6.481	3.70	140.000	210.000
2020-785	5	SIEMENS	37	38	SIEMENS DI	143	SIEMENS	18	161.167	15.316	9.50	129.000	193.000
138	PHOSPHORUS	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %							
2020-781	1	PHOSPHOMO.		38	SIEMENS DI	143	SIEMENS	42	4.467	0.161	3.61	3.600	5.400
2020-781	1	PHOSPHOMO.		165	VITROS	63	KODAK	86	4.330	0.120	2.77	3.500	5.200
2020-781	1	PHOSPHOMO.		184	INTEGRA	93	ROCHE	16	4.281	0.118	2.77	3.400	5.100
2020-781	1	PHOSPHOMO.		186	BECK. COUL	76	AU-BECKMAN	16	4.244	0.169	3.99	3.400	5.100
2020-781	1	PHOSPHOMO.		203	ROCHE	93	ROCHE	30	4.337	0.143	3.29	3.500	5.200
2020-781	1	PHOSPHOMO.		215	B.C DXC	134	BECKM.COUL	12	4.567	0.217	4.76	3.700	5.500
2020-781	2	END POINT		184	INTEGRA	93	ROCHE	10	4.280	0.209	4.88	3.400	5.100
2020-782	1	PHOSPHOMO.		38	SIEMENS DI	143	SIEMENS	42	2.783	0.099	3.54	2.200	3.300
2020-782	1	PHOSPHOMO.		165	VITROS	63	KODAK	86	2.893	0.098	3.39	2.300	3.500
2020-782	1	PHOSPHOMO.		184	INTEGRA	93	ROCHE	16	2.669	0.092	3.43	2.100	3.200
2020-782	1	PHOSPHOMO.		186	BECK. COUL	76	AU-BECKMAN	16	2.600	0.112	4.30	2.100	3.100
2020-782	1	PHOSPHOMO.		203	ROCHE	93	ROCHE	30	2.687	0.123	4.58	2.100	3.200
2020-782	1	PHOSPHOMO.		215	B.C DXC	134	BECKM.COUL	12	2.767	0.125	4.51	2.200	3.300
2020-782	2	END POINT		184	INTEGRA	93	ROCHE	10	2.740	0.258	9.40	2.200	3.300
2020-783	1	PHOSPHOMO.		38	SIEMENS DI	143	SIEMENS	42	4.883	0.141	2.89	3.900	5.900
2020-783	1	PHOSPHOMO.		165	VITROS	63	KODAK	86	4.714	0.122	2.59	3.800	5.700
2020-783	1	PHOSPHOMO.		184	INTEGRA	93	ROCHE	16	4.700	0.100	2.13	3.800	5.600
2020-783	1	PHOSPHOMO.		186	BECK. COUL	76	AU-BECKMAN	16	4.613	0.165	3.59	3.700	5.500
2020-783	1	PHOSPHOMO.		203	ROCHE	93	ROCHE	30	4.787	0.120	2.51	3.800	5.700
2020-783	1	PHOSPHOMO.		215	B.C DXC	134	BECKM.COUL	12	5.092	0.263	5.16	4.100	6.100
2020-783	2	END POINT		184	INTEGRA	93	ROCHE	10	4.610	0.221	4.80	3.700	5.500
2020-784	1	PHOSPHOMO.		38	SIEMENS DI	143	SIEMENS	42	5.667	0.157	2.77	4.500	6.800
2020-784	1	PHOSPHOMO.		165	VITROS	63	KODAK	86	5.460	0.147	2.70	4.400	6.600
2020-784	1	PHOSPHOMO.		184	INTEGRA	93	ROCHE	16	5.456	0.137	2.51	4.400	6.500

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE				GRP: 10 ROUTINE						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-784	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	16	5.400	0.166	3.07	4.300	6.500
2020-784	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	30	5.590	0.168	3.01	4.500	6.700
2020-784	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	5.967	0.287	4.81	4.800	7.200
2020-784	2	END POINT	184	INTEGRA	93	ROCHE	10	5.410	0.243	4.49	4.300	6.500
138	PHOSPHORUS	CLIA TEST ID #: 0				EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-785	1	PHOSPHOMO.	38	SIEMENS DI	143	SIEMENS	42	3.507	0.126	3.58	2.800	4.200
2020-785	1	PHOSPHOMO.	165	VITROS	63	KODAK	86	3.540	0.102	2.88	2.800	4.200
2020-785	1	PHOSPHOMO.	184	INTEGRA	93	ROCHE	16	3.350	0.100	2.99	2.700	4.000
2020-785	1	PHOSPHOMO.	186	BECK. COUL	76	AU-BECKMAN	16	3.300	0.150	4.55	2.600	4.000
2020-785	1	PHOSPHOMO.	203	ROCHE	93	ROCHE	30	3.413	0.136	3.98	2.700	4.100
2020-785	1	PHOSPHOMO.	215	B.C DXC	134	BECKM.COUL	12	3.575	0.153	4.29	2.900	4.300
2020-785	2	END POINT	184	INTEGRA	93	ROCHE	10	3.310	0.164	4.96	2.600	4.000
169	CO2					EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-781	1	Generic Me	165	VITROS	63	KODAK	22	20.818	1.800	8.65	15.000	26.000
2020-781	1	Generic Me	184	INTEGRA	93	ROCHE	13	21.154	2.143	10.13	15.000	28.000
2020-781	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	22	19.909	1.535	7.71	15.000	25.000
2020-781	1	Generic Me	203	ROCHE	93	ROCHE	15	19.733	1.692	8.57	15.000	25.000
2020-781	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	65	22.692	2.877	12.68	14.000	31.000
2020-781	2	ENZYMATIC	165	VITROS	63	KODAK	100	20.740	1.598	7.70	16.000	26.000
2020-781	2	ENZYMATIC	165	VITROS	156	Kit 32	24	20.208	1.554	7.69	16.000	25.000
2020-781	2	ENZYMATIC	203	ROCHE	93	ROCHE	13	20.385	2.338	11.47	13.000	27.000
2020-782	1	Generic Me	165	VITROS	63	KODAK	22	14.455	1.698	11.75	9.000	20.000
2020-782	1	Generic Me	184	INTEGRA	93	ROCHE	13	14.769	1.527	10.34	10.000	19.000
2020-782	1	Generic Me	186	BECK. COUL	76	AU-BECKMAN	22	14.227	1.379	9.70	10.000	18.000
2020-782	1	Generic Me	203	ROCHE	93	ROCHE	15	14.333	1.193	8.32	11.000	18.000
2020-782	2	ENZYMATIC	38	SIEMENS DI	143	SIEMENS	65	16.908	2.600	15.38	9.000	25.000
2020-782	2	ENZYMATIC	165	VITROS	63	KODAK	100	14.660	1.614	11.01	10.000	20.000
2020-782	2	ENZYMATIC	165	VITROS	156	Kit 32	24	13.250	1.267	9.56	9.000	17.000
2020-782	2	ENZYMATIC	203	ROCHE	93	ROCHE	13	14.615	1.443	9.87	10.000	19.000

Participant Summary
MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE								
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
169	CO2	CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-783	1	Generic Me	165 VITROS	63	KODAK	22	21.727	1.572	7.23	17.000	26.000	
2020-783	1	Generic Me	184 INTEGRA	93	ROCHE	13	22.077	2.526	11.44	15.000	30.000	
2020-783	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	21.591	2.082	9.64	15.000	28.000	
2020-783	1	Generic Me	203 ROCHE	93	ROCHE	15	21.200	2.040	9.62	15.000	27.000	
2020-783	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	65	24.262	2.413	9.95	17.000	32.000	
2020-783	2	ENZYMATIC	165 VITROS	63	KODAK	100	21.850	1.492	6.83	17.000	26.000	
2020-783	2	ENZYMATIC	165 VITROS	156	Kit 32	24	21.458	1.414	6.59	17.000	26.000	
2020-783	2	ENZYMATIC	203 ROCHE	93	ROCHE	13	21.692	2.462	11.35	14.000	29.000	
2020-784	1	Generic Me	165 VITROS	63	KODAK	22	24.227	2.066	8.53	18.000	30.000	
2020-784	1	Generic Me	184 INTEGRA	93	ROCHE	13	24.846	2.381	9.58	18.000	32.000	
2020-784	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	24.955	2.205	8.84	18.000	32.000	
2020-784	1	Generic Me	203 ROCHE	93	ROCHE	15	24.400	1.855	7.60	19.000	30.000	
2020-784	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	65	26.844	2.545	9.48	19.000	34.000	
2020-784	2	ENZYMATIC	165 VITROS	63	KODAK	100	24.434	1.736	7.10	19.000	30.000	
2020-784	2	ENZYMATIC	165 VITROS	156	Kit 32	24	24.125	1.563	6.48	19.000	29.000	
2020-784	2	ENZYMATIC	203 ROCHE	93	ROCHE	13	25.308	3.147	12.44	16.000	35.000	
2020-785	1	Generic Me	165 VITROS	63	KODAK	22	15.909	1.311	8.24	12.000	20.000	
2020-785	1	Generic Me	184 INTEGRA	93	ROCHE	13	16.692	2.197	13.16	10.000	23.000	
2020-785	1	Generic Me	186 BECK. COUL	76	AU-BECKMAN	22	15.773	1.594	10.10	11.000	21.000	
2020-785	1	Generic Me	203 ROCHE	93	ROCHE	15	15.867	1.454	9.17	12.000	20.000	
2020-785	2	ENZYMATIC	38 SIEMENS DI	143	SIEMENS	65	18.344	2.438	13.29	11.000	26.000	
2020-785	2	ENZYMATIC	165 VITROS	63	KODAK	100	16.190	1.566	9.68	11.000	21.000	
2020-785	2	ENZYMATIC	165 VITROS	156	Kit 32	24	15.458	1.079	6.98	12.000	19.000	
2020-785	2	ENZYMATIC	203 ROCHE	93	ROCHE	13	16.462	2.437	14.81	9.000	24.000	
170	DIRECT BILIRUBIN	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}										
2020-781	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	47	1.164	0.091	7.81	0.900	1.400	
2020-781	11	SPECT. WIT	165 VITROS	63	KODAK	60	1.169	0.088	7.49	0.900	1.400	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 10 ROUTINE		GRP: 10 ROUTINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-781	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	1.064	0.081	7.62	0.800	1.300
2020-781	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	1.150	0.065	5.61	0.900	1.400
170	DIRECT BILIRUBIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 20 % {GREATER}							
2020-782	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	47	0.496	0.050	10.16	0.300	0.600
2020-782	11	SPECT. WIT	165 VITROS	63	KODAK	60	0.269	0.103	38.19	0.000	0.600
2020-782	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.486	0.035	7.20	0.400	0.600
2020-782	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	0.509	0.051	10.10	0.400	0.700
2020-783	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	47	1.313	0.077	5.86	1.100	1.600
2020-783	11	SPECT. WIT	165 VITROS	63	KODAK	60	1.388	0.104	7.51	1.100	1.700
2020-783	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	1.214	0.091	7.53	0.900	1.500
2020-783	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	1.317	0.107	8.11	1.000	1.600
2020-784	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	47	1.634	0.093	5.68	1.300	2.000
2020-784	11	SPECT. WIT	165 VITROS	63	KODAK	60	1.766	0.118	6.70	1.400	2.100
2020-784	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	1.479	0.121	8.16	1.100	1.800
2020-784	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	1.617	0.099	6.10	1.300	1.900
2020-785	4	JEND.-GROG	38 SIEMENS DI	143	SIEMENS	47	0.787	0.064	8.13	0.600	1.000
2020-785	11	SPECT. WIT	165 VITROS	63	KODAK	60	0.663	0.095	14.41	0.400	0.900
2020-785	13	AZOBIL.W/D	203 ROCHE	93	ROCHE	15	0.771	0.059	7.64	0.600	0.900
2020-785	16	DIAZO ENDP	215 B.C DXC	134	BECKM.COUL	12	0.800	0.058	7.22	0.600	1.000

SPEC: 10 CHEMISTRY		SUB: 20 BLOOD GASES		GRP: 30 BLOOD GASES							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE 8 % OR 5.00 {GREATER}							
2020-831	265	ROCHE COBA		30	81.250	1.624	2.00	74.800	87.800		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 20 BLOOD GASES			GRP: 30 BLOOD GASES				
ANALYTE										
SAMPLE	METHOD		SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
007	BLOOD GASES-pCO2		CLIA TEST ID #: 335	EVALUATION CRITERIA: TARGET VALUE	8	% OR 5.00	{GREATER}			
2020-832	265 ROCHE COBA				30	53.267	2.377	4.46	48.300	58.300
2020-833	265 ROCHE COBA				30	24.441	0.549	2.25	19.400	29.400
2020-834	265 ROCHE COBA				30	63.383	0.879	1.39	58.300	68.500
2020-835	265 ROCHE COBA				30	24.420	0.629	2.58	19.400	29.400
008	BLOOD GASES-pH		CLIA TEST ID #: 315	EVALUATION CRITERIA: TARGET VALUE	0.04					
2020-831	265 ROCHE COBA				30	7.099	0.007	0.10	7.059	7.139
2020-832	265 ROCHE COBA				30	7.324	0.014	0.20	7.284	7.364
2020-833	265 ROCHE COBA				30	7.601	0.004	0.05	7.561	7.641
2020-834	265 ROCHE COBA				30	7.196	0.005	0.08	7.156	7.236
2020-835	265 ROCHE COBA				30	7.601	0.005	0.07	7.561	7.641
009	BLOOD GASES-pO2		CLIA TEST ID #: 325	EVALUATION CRITERIA: TARGET VALUE	3	SD				
2020-831	265 ROCHE COBA				30	71.082	5.263	7.40	55.300	86.900
2020-832	265 ROCHE COBA				30	123.783	6.173	4.99	105.300	142.300
2020-833	265 ROCHE COBA				30	152.087	4.312	2.84	139.200	165.000
2020-834	265 ROCHE COBA				30	323.650	20.133	6.22	263.300	384.000
2020-835	265 ROCHE COBA				30	150.552	5.038	3.35	135.400	165.700

Participant Summary
MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)					
ANALYTE									
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
029	FOLL. STIM. HOR.-FSH	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD				
2020-811	116 ROCHE COBA			15	28.853	2.623	9.09	21.000	36.700
2020-812	116 ROCHE COBA			15	21.787	1.934	8.88	16.000	27.600
2020-813	116 ROCHE COBA			15	16.540	1.574	9.52	11.800	21.300
2020-814	116 ROCHE COBA			15	24.353	2.074	8.52	18.100	30.600
2020-815	116 ROCHE COBA			15	12.507	1.149	9.18	9.100	16.000
030	FREE THYROXINE	CLIA TEST ID #: 545							
2020-811	115 VITROS ECI			35	6.976	0.064	0.92	6.800	7.200
2020-811	116 ROCHE COBA			43	3.540	0.186	5.25	3.000	4.100
2020-811	125 B.C.ACCESS			43	2.865	0.141	4.93	2.400	3.300
2020-811	149 ARC i2000s			21	2.429	0.178	7.32	1.900	3.000
2020-811	160 SIEMENS EX			41	3.740	0.164	4.39	3.200	4.200
2020-812	115 VITROS ECI			35	6.137	0.155	2.53	5.700	6.600
2020-812	116 ROCHE COBA			43	2.786	0.136	4.87	2.400	3.200
2020-812	125 B.C.ACCESS			43	2.423	0.114	4.70	2.100	2.800
2020-812	149 ARC i2000s			21	2.071	0.139	6.69	1.700	2.500
2020-812	160 SIEMENS EX			41	2.895	0.130	4.50	2.500	3.300
2020-813	115 VITROS ECI			35	5.051	0.175	3.46	4.500	5.600
2020-813	116 ROCHE COBA			43	2.188	0.105	4.80	1.900	2.500
2020-813	125 B.C.ACCESS			43	2.007	0.092	4.61	1.700	2.300
2020-813	149 ARC i2000s			21	1.660	0.080	4.82	1.400	1.900
2020-813	160 SIEMENS EX			41	2.218	0.122	5.51	1.900	2.600

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
030	FREE THYROXINE	CLIA TEST ID #: 545	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-814	115 VITROS ECI			35	6.518	0.129	1.99	6.100	6.900	
2020-814	116 ROCHE COBA			43	3.051	0.148	4.86	2.600	3.500	
2020-814	125 B.C.ACCESS			43	2.633	0.120	4.54	2.300	3.000	
2020-814	149 ARC i2000s			21	2.205	0.140	6.33	1.800	2.600	
2020-814	160 SIEMENS EX			41	3.188	0.125	3.92	2.800	3.600	
2020-815	115 VITROS ECI			35	3.843	0.189	4.91	3.300	4.400	
2020-815	116 ROCHE COBA			43	1.707	0.087	5.12	1.400	2.000	
2020-815	125 B.C.ACCESS			43	1.605	0.068	4.24	1.400	1.800	
2020-815	149 ARC i2000s			21	1.290	0.068	5.30	1.100	1.500	
2020-815	160 SIEMENS EX			41	1.693	0.079	4.65	1.500	1.900	
032	hCG-QUANTITATIVE	CLIA TEST ID #: 555								
2020-811	115 VITROS ECI			18	1946.817	95.426	4.90	1660.500	2233.100	
2020-811	116 ROCHE COBA			19	1192.653	106.769	8.95	872.300	1513.000	
2020-811	125 B.C.ACCESS			11	687.309	30.427	4.43	596.000	778.600	
2020-812	115 VITROS ECI			18	1350.456	64.653	4.79	1156.500	1544.400	
2020-812	116 ROCHE COBA			19	817.232	61.782	7.56	631.900	1002.600	
2020-812	125 B.C.ACCESS			11	530.182	22.742	4.29	462.000	598.400	
2020-813	115 VITROS ECI			18	912.478	51.129	5.60	759.100	1065.900	
2020-813	116 ROCHE COBA			19	540.826	37.857	7.00	427.300	654.400	
2020-813	125 B.C.ACCESS			11	392.464	19.586	4.99	333.700	451.200	
2020-814	115 VITROS ECI			18	1558.833	84.910	5.45	1304.100	1813.600	
2020-814	116 ROCHE COBA			19	918.205	101.500	11.05	613.700	1222.700	
2020-814	125 B.C.ACCESS			11	589.545	20.851	3.54	527.000	652.100	
2020-815	115 VITROS ECI			18	570.917	27.478	4.81	488.500	653.400	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-815	116	ROCHE COBA		19	334.768	24.717	7.38	260.600	408.900	
2020-815	125	B.C.ACCESS		11	265.927	9.690	3.64	236.900	295.000	
034	PROGESTERONE		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD					
2020-811	116	ROCHE COBA		10	26.960	1.255	4.65	23.200	30.700	
2020-812	116	ROCHE COBA		10	20.140	0.705	3.50	18.000	22.300	
2020-813	116	ROCHE COBA		10	14.490	0.474	3.27	13.100	15.900	
2020-814	116	ROCHE COBA		10	22.330	0.723	3.24	20.200	24.500	
2020-815	116	ROCHE COBA		10	9.720	0.483	4.97	8.300	11.200	
035	PROLACTIN									
2020-811	116	ROCHE COBA		17	29.088	1.505	5.17	24.600	33.600	
2020-812	116	ROCHE COBA		17	20.976	1.095	5.22	17.700	24.300	
2020-813	116	ROCHE COBA		17	14.800	0.799	5.40	12.400	17.200	
2020-814	116	ROCHE COBA		17	23.906	1.215	5.08	20.300	27.600	
2020-815	116	ROCHE COBA		17	10.141	0.478	4.71	8.700	11.600	
037	TESTOSTERONE									
2020-811	116	ROCHE COBA		10	14.120	0.668	4.73	12.100	16.100	
2020-812	116	ROCHE COBA		10	11.090	0.745	6.72	8.900	13.300	
2020-813	116	ROCHE COBA		10	93.460	256.180	274.11	0	862.000	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
037	TESTOSTERONE	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD							
2020-814	116 ROCHE COBA			10	12.450	0.754	6.06	10.200	14.700	
2020-815	116 ROCHE COBA			10	5.580	0.485	8.70	4.100	7.000	
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585								
2020-811	32 DADE DIME.			10	5.110	0.713	13.96	3.000	7.300	
2020-811	106 TOSOH AIA			14	5.900	0.327	5.55	4.900	6.900	
2020-811	115 VITROS ECI			40	6.503	0.266	4.09	5.700	7.300	
2020-811	116 ROCHE COBA			51	5.270	0.176	3.34	4.700	5.800	
2020-811	125 B.C.ACCESS			59	5.075	0.284	5.59	4.200	5.900	
2020-811	149 ARC i2000s			23	5.005	0.238	4.76	4.300	5.700	
2020-811	160 SIEMENS EX			67	4.351	0.235	5.39	3.600	5.100	
2020-812	32 DADE DIME.			10	3.790	0.432	11.41	2.500	5.100	
2020-812	106 TOSOH AIA			14	4.200	0.359	8.54	3.100	5.300	
2020-812	115 VITROS ECI			40	4.833	0.162	3.35	4.300	5.300	
2020-812	116 ROCHE COBA			51	4.006	0.141	3.51	3.600	4.400	
2020-812	125 B.C.ACCESS			59	3.660	0.191	5.22	3.100	4.200	
2020-812	149 ARC i2000s			23	3.641	0.164	4.51	3.100	4.100	
2020-812	160 SIEMENS EX			67	3.243	0.170	5.24	2.700	3.800	
2020-813	32 DADE DIME.			10	2.770	0.297	10.72	1.900	3.700	
2020-813	106 TOSOH AIA			14	3.214	0.207	6.43	2.600	3.800	
2020-813	115 VITROS ECI			40	3.521	0.111	3.16	3.200	3.900	
2020-813	116 ROCHE COBA			51	3.031	0.120	3.95	2.700	3.400	
2020-813	125 B.C.ACCESS			59	2.636	0.159	6.05	2.200	3.100	
2020-813	149 ARC i2000s			23	2.630	0.123	4.68	2.300	3.000	
2020-813	160 SIEMENS EX			67	2.416	0.150	6.22	2.000	2.900	
2020-814	32 DADE DIME.			10	4.250	0.508	11.96	2.700	5.800	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 30 ENDOCRINOLOGY

GRP: 40 ENDOCRINOLOGY (GEN.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-814	106	TOSOH AIA		14	4.843	0.329	6.79	3.900	5.800
2020-814	115	VITROS ECI		40	5.430	0.186	3.43	4.900	6.000
2020-814	116	ROCHE COBA		51	4.465	0.143	3.19	4.000	4.900
2020-814	125	B.C.ACCESS		59	4.156	0.220	5.28	3.500	4.800
2020-814	149	ARC i2000s		23	4.074	0.177	4.36	3.500	4.600
2020-814	160	SIEMENS EX		67	3.629	0.194	5.34	3.000	4.200
039	THYRO. STI. HOR.-TSH	CLIA TEST ID #: 585	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-815	32	DADE DIME.		10	1.930	0.200	10.38	1.300	2.500
2020-815	106	TOSOH AIA		14	2.243	0.164	7.29	1.800	2.700
2020-815	115	VITROS ECI		40	2.503	0.080	3.20	2.300	2.700
2020-815	116	ROCHE COBA		51	2.192	0.082	3.74	1.900	2.400
2020-815	125	B.C.ACCESS		59	1.869	0.114	6.09	1.500	2.200
2020-815	149	ARC i2000s		23	1.852	0.088	4.74	1.600	2.100
2020-815	160	SIEMENS EX		67	1.714	0.097	5.65	1.400	2.000
040	THYROXINE (T4)	CLIA TEST ID #: 595	EVALUATION CRITERIA: TARGET VALUE 20 % OR 1.00 {GREATER}						
2020-811	32	DADE DIME.		50	13.625	0.083	0.61	10.900	16.400
2020-811	32	DADE DIME.		50	14.810	0.810	5.47	11.800	17.800
2020-811	106	TOSOH AIA		10	11.460	0.849	7.41	9.200	13.800
2020-811	115	VITROS ECI		35	12.120	0.442	3.65	9.700	14.500
2020-811	116	ROCHE COBA		48	13.625	0.083	0.61	10.900	16.400
2020-811	116	ROCHE COBA		48	12.925	0.791	6.12	10.300	15.500
2020-811	125	B.C.ACCESS		49	12.035	0.806	6.70	9.600	14.400
2020-811	149	ARC i2000s		20	13.625	0.083	0.61	10.900	16.400
2020-811	149	ARC i2000s		20	13.180	1.123	8.52	10.500	15.800
2020-812	32	DADE DIME.		50	10.949	0.640	5.84	8.800	13.100
2020-812	32	DADE DIME.		50	10.700	1.275	11.91	8.600	12.800
2020-812	106	TOSOH AIA		10	9.300	0.643	6.92	7.400	11.200
2020-812	115	VITROS ECI		35	9.374	0.387	4.13	7.500	11.200
2020-812	116	ROCHE COBA		48	9.972	0.579	5.81	8.000	12.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY		GRP: 40 ENDOCRINOLOGY (GEN.)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-815	116	ROCHE COBA		14	2.000	0.000	0.00	1.500	2.500	
2020-815	116	ROCHE COBA		14	2.180	0.605	27.74	1.600	2.700	
2020-815	125	B.C.ACCESS		12	3.025	0.657	21.72	2.300	3.800	
2020-815	149	ARC i2000s		10	3.050	0.297	9.75	2.300	3.800	
142 hLH (HUMAN LUT. HOR)				CLIA TEST ID #: 0 EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-811	116	ROCHE COBA		15	49.380	2.590	5.25	37.000	61.700	
2020-812	116	ROCHE COBA		15	36.200	1.981	5.47	27.200	45.300	
2020-813	116	ROCHE COBA		15	26.120	1.293	4.95	19.600	32.700	
2020-814	116	ROCHE COBA		15	41.007	2.063	5.03	30.800	51.300	
2020-815	116	ROCHE COBA		15	18.160	0.894	4.92	13.600	22.700	

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)						
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFeree	COMMER	REPORTED	PART # LABS	# "SF"	Score		
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
031 hCG-QUALITATIVE												
2020-811	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100
2020-811	142	INSTANT VI		002	002	002	002	002 - POSITIVE	51	51	51	100
2020-811	150	AIM STEP		002	002	002	002	002 - POSITIVE	85	85	85	100
2020-811	164	HCG Alere		002	002	002	002	002 - POSITIVE	67	67	67	100
2020-811	165	Osom HCG C		002	002	002	002	002 - POSITIVE	19	19	19	100
2020-811	166	FaStep		002	002	002	002	002 - POSITIVE	10	10	10	100
2020-811	167	Tanner Sci		002	002	002	002	002 - POSITIVE	15	15	15	100
2020-812	113	BIO-SIGN		002	002	002	002	002 - POSITIVE	21	21	21	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-812	142	INSTANT VI		002 002	002 002		002 - POSITIVE	51	51	51	100	
2020-812	150	AIM STEP		002 002	002 002		002 - POSITIVE	85	85	85	100	
2020-812	164	HCG Alere		002 002	002 002		002 - POSITIVE	67	67	67	100	
2020-812	165	Osom HCG C		002 002	002 002		002 - POSITIVE	19	19	19	100	
2020-812	166	FaStep		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-812	167	Tanner Sci		002 002	002 002		002 - POSITIVE	15	15	15	100	
031 hCG-QUALITATIVE												
2020-813	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-813	142	INSTANT VI		002 002	002 002		002 - POSITIVE	51	51	51	100	
2020-813	150	AIM STEP		002 002	002 002		002 - POSITIVE	85	85	85	100	
2020-813	164	HCG Alere		002 002	002 002		002 - POSITIVE	67	67	67	100	
2020-813	165	Osom HCG C		002 002	002 002		002 - POSITIVE	19	19	19	100	
2020-813	166	FaStep		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-813	167	Tanner Sci		002 002	002 002		002 - POSITIVE	15	15	15	100	
2020-814	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-814	142	INSTANT VI		002 002	002 002		002 - POSITIVE	51	51	51	100	
2020-814	150	AIM STEP		002 002	002 002		001 - NEGATIVE	85	1	84	0	
2020-814	150	AIM STEP		002 002	002 002		002 - POSITIVE	85	84	84	100	
2020-814	164	HCG Alere		002 002	002 002		002 - POSITIVE	67	67	67	100	
2020-814	165	Osom HCG C		002 002	002 002		002 - POSITIVE	19	19	19	100	
2020-814	166	FaStep		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-814	167	Tanner Sci		002 002	002 002		002 - POSITIVE	15	15	15	100	
2020-815	113	BIO-SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-815	142	INSTANT VI		002 002	002 002		002 - POSITIVE	51	51	51	100	
2020-815	150	AIM STEP		002 002	002 002		001 - NEGATIVE	85	1	84	0	
2020-815	150	AIM STEP		002 002	002 002		002 - POSITIVE	85	84	84	100	
2020-815	164	HCG Alere		002 002	002 002		002 - POSITIVE	67	67	67	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 40 ENDOCRINOLOGY (GEN.)					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-815	165	Osom HCG C		002 002	002 002		002 - POSITIVE	19	19	19	100	
2020-815	166	FaStep		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-815	167	Tanner Sci		002 002	002 002		002 - POSITIVE	15	15	15	100	

SPE: 10 CHEMISTRY			SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
151 hCG ONLY QUALITATIVE												
2020-801	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-801	083	STAMBIO		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-801	112	ONE STEP D		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-801	114	INSTANT VI		002 002	002 002		002 - POSITIVE	114	114	114	100	
2020-801	118	AIM STEP		002 002	002 002		002 - POSITIVE	144	144	144	100	
2020-801	130	ALERE HCG		002 002	002 002		002 - POSITIVE	89	89	89	100	
2020-801	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100	
2020-801	133	FaStep		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-801	134	Tanner Sc.		002 002	002 002		001 - NEGATIVE	68	1	67	0	
2020-801	134	Tanner Sc.		002 002	002 002		002 - POSITIVE	68	67	67	100	
2020-802	018	BIO SIGN		002 002	002 002		002 - POSITIVE	21	21	21	100	
2020-802	083	STAMBIO		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-802	112	ONE STEP D		002 002	002 002		002 - POSITIVE	10	10	10	100	
2020-802	114	INSTANT VI		002 002	002 002		002 - POSITIVE	114	114	114	100	
2020-802	118	AIM STEP		002 002	002 002		002 - POSITIVE	144	144	144	100	
2020-802	130	ALERE HCG		002 002	002 002		002 - POSITIVE	89	89	89	100	
2020-802	131	OSOM HCG		002 002	002 002		002 - POSITIVE	25	25	25	100	
2020-802	133	FaStep		002 002	002 002		002 - POSITIVE	21	21	21	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-802	134	Tanner Sc.		002	002	002	002 - POSITIVE	68	68	68	100		
	151 hCG ONLY QUALITATIVE												
2020-803	018	BIO SIGN		001	001	001	001 - NEGATIVE	21	21	21	100		
2020-803	083	STAMBIO		001	001	001	001 - NEGATIVE	10	10	10	100		
2020-803	112	ONE STEP D		001	001	001	001 - NEGATIVE	10	10	10	100		
2020-803	114	INSTANT VI		001	001	001	001 - NEGATIVE	114	114	114	100		
2020-803	118	AIM STEP		001	001	001	001 - NEGATIVE	144	144	144	100		
2020-803	130	ALERE HCG		001	001	001	001 - NEGATIVE	89	88	88	100		
2020-803	130	ALERE HCG		001	001	001	002 - POSITIVE	89	1	88	0		
2020-803	131	OSOM HCG		001	001	001	001 - NEGATIVE	25	25	25	100		
2020-803	133	FaStep		001	001	001	001 - NEGATIVE	21	21	21	100		
2020-803	134	Tanner Sc.		001	001	001	001 - NEGATIVE	68	67	67	100		
2020-803	134	Tanner Sc.		001	001	001	002 - POSITIVE	68	1	67	0		
2020-804	018	BIO SIGN		002	002	002	002 - POSITIVE	21	21	21	100		
2020-804	083	STAMBIO		002	002	002	002 - POSITIVE	10	10	10	100		
2020-804	112	ONE STEP D		002	002	002	002 - POSITIVE	10	10	10	100		
2020-804	114	INSTANT VI		002	002	002	001 - NEGATIVE	114	1	113	0		
2020-804	114	INSTANT VI		002	002	002	002 - POSITIVE	114	113	113	100		
2020-804	118	AIM STEP		002	002	002	002 - POSITIVE	144	144	144	100		
2020-804	130	ALERE HCG		002	002	002	002 - POSITIVE	89	89	89	100		
2020-804	131	OSOM HCG		002	002	002	002 - POSITIVE	25	25	25	100		
2020-804	133	FaStep		002	002	002	002 - POSITIVE	21	21	21	100		
2020-804	134	Tanner Sc.		002	002	002	001 - NEGATIVE	68	1	67	0		
2020-804	134	Tanner Sc.		002	002	002	002 - POSITIVE	68	67	67	100		
2020-805	018	BIO SIGN		001	001	001	001 - NEGATIVE	21	21	21	100		
2020-805	083	STAMBIO		001	001	001	001 - NEGATIVE	10	10	10	100		
2020-805	112	ONE STEP D		001	001	001	001 - NEGATIVE	10	10	10	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 45 ENDOCRINOL.-hCG ONLY						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-805	114	INSTANT VI		001 001	001 001		001 - NEGATIVE	114	112	112	100	
2020-805	114	INSTANT VI		001 001	001 001		002 - POSITIVE	114	2	112	0	
2020-805	118	AIM STEP		001 001	001 001		001 - NEGATIVE	144	144	144	100	
2020-805	130	ALERE HCG		001 001	001 001		001 - NEGATIVE	89	88	88	100	
2020-805	130	ALERE HCG		001 001	001 001		002 - POSITIVE	89	1	88	0	
2020-805	131	OSOM HCG		001 001	001 001		001 - NEGATIVE	25	25	25	100	
2020-805	133	FaStep		001 001	001 001		001 - NEGATIVE	21	21	21	100	
2020-805	134	Tanner Sc.		001 001	001 001		001 - NEGATIVE	68	68	68	100	

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
173 T3 UPTAKE												
CLIA TEST ID #: 565												
EVALUATION CRITERIA: TARGET VALUE 3 SD												
2020-811	1	%U	84 VITROS ECI	1	25-35	28	50.775	1.387	2.73	46.610	54.940	
2020-811	1	%U	91 B.C.ACCESS	2	35-45	31	46.200	2.574	5.57	38.480	53.920	
2020-811	2	TUV	95 Arc i2000s	3	0.7 - 1.2	14	1.019	0.066	6.48	0.820	1.220	
2020-811	4	%	32 DIMENSION	2	35-45	49	42.569	2.108	4.95	36.250	48.890	
2020-811	5	TBI	98 ROCHE COBA	4	0.8-1.3	44	0.535	0.073	13.56	0.320	0.750	
2020-812	1	%U	84 VITROS ECI	1	25-35	28	52.929	1.358	2.57	48.850	57.000	
2020-812	1	%U	91 B.C.ACCESS	2	35-45	31	47.370	2.887	6.09	38.710	56.030	
2020-812	2	TUV	95 Arc i2000s	3	0.7 - 1.2	14	0.921	0.058	6.30	0.750	1.090	
2020-812	4	%	32 DIMENSION	2	35-45	49	41.932	1.903	4.54	36.220	47.640	
2020-812	5	TBI	98 ROCHE COBA	4	0.8-1.3	44	0.628	0.068	10.75	0.430	0.830	
2020-813	1	%U	84 VITROS ECI	1	25-35	28	54.657	1.594	2.92	49.870	59.440	
2020-813	1	%U	91 B.C.ACCESS	2	35-45	31	49.203	2.419	4.92	41.950	56.460	
2020-813	2	TUV	95 Arc i2000s	3	0.7 - 1.2	14	0.808	0.044	5.40	0.680	0.940	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 30 ENDOCRINOLOGY				GRP: 55 ENDO.(T3 UPTAKE)						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2020-813	4 %	32	DIMENSION	2	35-45	49	41.637	1.676	4.03	36.610	46.670	
2020-813	5 TBI	98	ROCHE COBA	4	0.8-1.3	44	0.699	0.065	9.31	0.500	0.890	
173 T3 UPTAKE		CLIA TEST ID #: 565		EVALUATION CRITERIA: TARGET VALUE 3 SD								
2020-814	1 %U	84	VITROS ECI	1	25-35	28	52.100	1.224	2.35	48.430	55.770	
2020-814	1 %U	91	B.C.ACCESS	2	35-45	31	46.867	2.352	5.02	39.810	53.920	
2020-814	2 TUV	95	Arc i2000s	3	0.7 - 1.2	14	0.956	0.051	5.30	0.800	1.110	
2020-814	4 %	32	DIMENSION	2	35-45	49	42.077	1.678	3.99	37.040	47.110	
2020-814	5 TBI	98	ROCHE COBA	4	0.8-1.3	44	0.603	0.074	12.32	0.380	0.830	
2020-815	1 %U	84	VITROS ECI	1	25-35	28	56.393	1.508	2.67	51.870	60.920	
2020-815	1 %U	91	B.C.ACCESS	2	35-45	31	50.213	2.240	4.46	43.490	56.930	
2020-815	2 TUV	95	Arc i2000s	3	0.7 - 1.2	14	0.713	0.038	5.39	0.600	0.830	
2020-815	4 %	32	DIMENSION	2	35-45	49	41.874	1.536	3.67	37.260	46.480	
2020-815	5 TBI	98	ROCHE COBA	4	0.8-1.3	44	0.753	0.057	7.63	0.580	0.930	

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY				GRP: 60 TOXICOLOGY						
ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
043 DIGOXIN		CLIA TEST ID #: 645		EVALUATION CRITERIA: TARGET VALUE 20 % OR 0.20 {GREATER}								
2020-811	50 VITROS					14	1.786	0.196	10.97	1.400	2.100	
2020-812	50 VITROS					14	1.393	0.202	14.48	1.100	1.700	
2020-813	50 VITROS					14	1.036	0.117	11.31	0.800	1.200	
2020-814	50 VITROS					14	1.543	0.159	10.31	1.200	1.900	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 40 TOXICOLOGY

GRP: 60 TOXICOLOGY

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
043 DIGOXIN		CLIA TEST ID #: 645		EVALUATION CRITERIA: TARGET VALUE 20 % OR 0.20 {GREATER}				
2020-815	50 VITROS		14	0.786	0.130	16.56	0.600	1.000
045 FERRITIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 %				
2020-811	50 VITROS		12	132.250	5.688	4.30	99.200	165.300
2020-811	90 ROCHE COBA		12	239.355	11.239	4.70	179.500	299.200
2020-811	90 ROCHE COBA		12	179.360	9.107	5.08	134.500	224.200
2020-812	50 VITROS		12	93.883	3.048	3.25	70.400	117.400
2020-812	90 ROCHE COBA		12	170.127	9.338	5.49	127.600	212.700
2020-812	90 ROCHE COBA		12	133.400	4.087	3.06	100.100	166.800
2020-813	50 VITROS		12	66.700	2.763	4.14	50.000	83.400
2020-813	90 ROCHE COBA		12	121.145	6.499	5.36	90.900	151.400
2020-813	90 ROCHE COBA		12	89.400	6.780	7.58	67.100	111.800
2020-814	50 VITROS		12	107.750	4.815	4.47	80.800	134.700
2020-814	90 ROCHE COBA		12	192.991	8.522	4.42	144.700	241.200
2020-814	90 ROCHE COBA		12	148.060	13.400	9.05	111.000	185.100
2020-815	50 VITROS		12	46.233	1.881	4.07	34.700	57.800
2020-815	90 ROCHE COBA		12	85.327	4.533	5.31	64.000	106.700
2020-815	90 ROCHE COBA		12	62.660	3.180	5.08	47.000	78.300
046 GENTAMICIN		CLIA TEST ID #: 665						
2020-811	50 VITROS		13	13.215	0.535	4.05	9.900	16.500
2020-812	50 VITROS		13	9.075	0.398	4.39	6.800	11.300
2020-813	50 VITROS		13	6.062	0.139	2.29	4.500	7.600

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 40 TOXICOLOGY		GRP: 60 TOXICOLOGY						
ANALYTE										
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
046	GENTAMICIN	CLIA TEST ID #: 665		EVALUATION CRITERIA: TARGET VALUE 25 %						
2020-814	50 VITROS			13	10.292	0.531	5.16	7.700	12.900	
	2020-815	50 VITROS		13	3.823	0.070	1.82	2.900	4.800	
050	PHENYTOIN	CLIA TEST ID #: 695								
	2020-811	50 VITROS		16	24.394	0.942	3.86	18.300	30.500	
	2020-812	50 VITROS		16	16.769	0.705	4.20	12.600	21.000	
	2020-813	50 VITROS		16	11.375	0.416	3.66	8.500	14.200	
	2020-814	50 VITROS		16	19.569	0.644	3.29	14.700	24.500	
	2020-815	50 VITROS		16	6.863	0.316	4.60	5.100	8.600	
054	THEOPHYLLINE	CLIA TEST ID #: 735								
	2020-811	50 VITROS		13	50.192	2.927	5.83	37.600	62.700	
	2020-812	50 VITROS		13	36.669	1.271	3.47	27.500	45.800	
	2020-813	50 VITROS		13	26.208	0.898	3.43	19.700	32.800	
	2020-814	50 VITROS		13	39.969	1.296	3.24	30.000	50.000	
	2020-815	50 VITROS		13	17.138	0.640	3.74	12.900	21.400	
127	VANCOMYCIN	CLIA TEST ID #: 0								
	2020-811	50 VITROS		16	44.819	2.417	5.39	33.600	56.000	
	2020-812	50 VITROS		16	29.153	0.630	2.16	21.900	36.400	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 40 TOXICOLOGY			GRP: 60 TOXICOLOGY					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
127	VANCOMYCIN		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 %							
2020-813	50	VITROS		16	19.625	0.525	2.68	14.700	24.500		
2020-814	50	VITROS		16	34.087	0.605	1.78	25.600	42.600		
2020-815	50	VITROS		16	12.338	0.337	2.73	9.300	15.400		
SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS					
ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
065	URINA-PROTEIN QUANT.			EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}							
2020-841	58	VITROS		40	104.000	5.958	5.73	78.000	130.000		
2020-841	62	DADE DIM.		50	106.900	2.343	2.19	80.200	133.600		
2020-841	79	BECK. COUL		13	99.946	4.396	4.40	75.000	124.900		
2020-841	93	ARCHITECT		17	94.982	3.043	3.20	71.200	118.700		
2020-841	95	INTEGRA		11	85.855	5.251	6.12	64.400	107.300		
2020-841	113	ROCHE COBA		26	89.092	1.611	1.81	66.800	111.400		
2020-841	139	S. EXL.200		11	106.655	2.010	1.88	80.000	133.300		
2020-842	58	VITROS		40	7.333	0.570	7.77	2.300	12.300		
2020-842	62	DADE DIM.		50	14.520	1.568	10.80	9.500	19.500		
2020-842	79	BECK. COUL		13	4.692	0.578	12.33	0	9.700		
2020-842	93	ARCHITECT		17	6.963	0.908	13.04	2.000	12.000		
2020-842	95	INTEGRA		11	3.440	0.395	11.50	0	8.400		
2020-842	113	ROCHE COBA		26	3.980	0.290	7.28	0	9.000		
2020-842	139	S. EXL.200		11	14.645	1.267	8.65	9.600	19.600		
2020-843	58	VITROS		40	9.385	1.028	10.96	4.400	14.400		
2020-843	62	DADE DIM.		50	22.900	1.888	8.24	17.200	28.600		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-843	79	BECK. COUL		13	5.317	0.710	13.36	0.300	10.300
2020-843	93	ARCHITECT		17	7.312	1.568	21.44	2.300	12.300
2020-843	95	INTEGRA		11	2.600	0.622	23.94	0	7.600
2020-843	113	ROCHE COBA		26	3.824	0.501	13.11	0	8.800
2020-843	139	S. EXL.200		11	23.345	1.173	5.03	17.500	29.200
065 URINA-PROTEIN QUANT.				CLIA TEST ID #: 0 EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-844	58	VITROS		40	137.359	8.654	6.30	103.000	171.700
2020-844	62	DADE DIM.		50	112.245	2.171	1.93	84.200	140.300
2020-844	79	BECK. COUL		13	101.146	4.643	4.59	75.900	126.400
2020-844	93	ARCHITECT		17	94.735	3.481	3.67	71.100	118.400
2020-844	95	INTEGRA		11	85.045	5.656	6.65	63.800	106.300
2020-844	113	ROCHE COBA		26	88.404	1.929	2.18	66.300	110.500
2020-844	139	S. EXL.200		11	112.209	1.930	1.72	84.200	140.300
2020-845	58	VITROS		40	146.615	7.909	5.39	110.000	183.300
2020-845	62	DADE DIM.		50	114.284	1.794	1.57	85.700	142.900
2020-845	79	BECK. COUL		13	100.877	4.713	4.67	75.700	126.100
2020-845	93	ARCHITECT		17	98.563	3.442	3.49	73.900	123.200
2020-845	95	INTEGRA		11	89.636	5.969	6.66	67.200	112.000
2020-845	113	ROCHE COBA		26	93.240	1.679	1.80	69.900	116.600
2020-845	139	S. EXL.200		11	113.990	0.774	0.68	85.500	142.500
067 URI-SPECIFIC GRAVITY				EVALUATION CRITERIA: TARGET VALUE 3 SD					
2020-841	11	AMES NM10S		49	1.025	0.003	0.25	1.017	1.033
2020-841	13	BAYER CLI		36	1.023	0.003	0.28	1.015	1.032
2020-841	67	URISCAN		51	1.025	0.004	0.42	1.012	1.038
2020-841	71	ATLAS		18	1.029	0.002	0.16	1.024	1.034
2020-841	72	URS-9-TECO		104	1.022	0.003	0.30	1.013	1.032
2020-841	80	ROCHE (CHE		37	1.016	0.002	0.17	1.011	1.021
2020-841	89	AIMSTRIP		114	1.024	0.003	0.26	1.016	1.031
2020-841	94	CLINITEK		104	1.025	0.002	0.21	1.018	1.031

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-841	102	ACON LAB		171	1.024	0.002	0.23	1.017	1.031
2020-841	131	iCHEM VELO		27	1.025	0.001	0.07	1.023	1.027
2020-841	135	Hybrido AU		13	1.024	0.001	0.14	1.019	1.028
2020-841	138	URS-10T		11	1.021	0.003	0.28	1.012	1.030
2020-841	140	MISSION A		66	1.024	0.002	0.22	1.017	1.030
2020-841	145	CLINITEK		11	1.029	0.001	0.15	1.025	1.034
2020-841	147	Mission120		17	1.023	0.002	0.23	1.016	1.030
2020-841	148	Mission500		17	1.024	0.002	0.16	1.020	1.029
067 URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD					
2020-842	11	AMES NM10S		49	1.023	0.003	0.30	1.014	1.033
2020-842	13	BAYER CLI		36	1.021	0.004	0.38	1.009	1.033
2020-842	67	URISCAN		51	1.022	0.003	0.34	1.012	1.032
2020-842	71	ATLAS		18	1.031	0.001	0.14	1.027	1.035
2020-842	72	URS-9-TECO		104	1.019	0.004	0.39	1.008	1.031
2020-842	80	ROCHE (CHE		37	1.016	0.002	0.22	1.010	1.023
2020-842	89	AIMSTRIP		114	1.022	0.003	0.32	1.012	1.032
2020-842	94	CLINITEK		104	1.023	0.003	0.27	1.015	1.031
2020-842	102	ACON LAB		171	1.023	0.003	0.29	1.014	1.032
2020-842	131	iCHEM VELO		27	1.026	0.001	0.09	1.024	1.029
2020-842	135	Hybrido AU		13	1.026	0.001	0.14	1.022	1.030
2020-842	138	URS-10T		11	1.018	0.006	0.56	1.001	1.035
2020-842	140	MISSION A		66	1.022	0.003	0.26	1.014	1.030
2020-842	145	CLINITEK		11	1.031	0.002	0.22	1.025	1.038
2020-842	147	Mission120		17	1.021	0.002	0.19	1.015	1.027
2020-842	148	Mission500		17	1.024	0.002	0.22	1.017	1.030
2020-843	11	AMES NM10S		49	1.013	0.004	0.36	1.002	1.024
2020-843	13	BAYER CLI		36	1.013	0.003	0.27	1.005	1.021
2020-843	67	URISCAN		51	1.011	0.006	0.57	0.994	1.028
2020-843	71	ATLAS		18	1.030	0.000	0.04	1.029	1.031

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-843	72	URS-9-TECO		104	1.012	0.004	0.35	1.002	1.023
2020-843	80	ROCHE (CHE		37	1.012	0.003	0.33	1.002	1.022
2020-843	89	AIMSTRIP		114	1.014	0.004	0.41	1.002	1.027
2020-843	94	CLINITEK		104	1.013	0.004	0.36	1.002	1.024
2020-843	102	ACON LAB		171	1.016	0.004	0.42	1.003	1.029
2020-843	131	iCHEM VELO		27	1.025	0.001	0.09	1.023	1.028
2020-843	135	Hybrido AU		13	1.025	0.001	0.14	1.020	1.029
2020-843	138	URS-10T		11	1.011	0.003	0.28	1.002	1.020
2020-843	140	MISSION A		66	1.015	0.004	0.43	1.002	1.029
2020-843	145	CLINITEK		11	1.030	0.001	0.06	1.028	1.032
2020-843	147	Mission120		17	1.018	0.002	0.23	1.012	1.025
2020-843	148	Mission500		17	1.019	0.002	0.16	1.014	1.024
067	URI-SPECIFIC GRAVITY		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD					
2020-844	11	AMES NM10S		49	1.020	0.003	0.27	1.012	1.028
2020-844	13	BAYER CLI		36	1.020	0.002	0.24	1.012	1.027
2020-844	67	URISCAN		51	1.017	0.003	0.27	1.009	1.025
2020-844	71	ATLAS		18	1.026	0.001	0.09	1.023	1.029
2020-844	72	URS-9-TECO		104	1.019	0.002	0.21	1.013	1.026
2020-844	80	ROCHE (CHE		37	1.011	0.002	0.21	1.005	1.017
2020-844	89	AIMSTRIP		114	1.018	0.004	0.35	1.008	1.029
2020-844	94	CLINITEK		104	1.020	0.002	0.23	1.013	1.027
2020-844	102	ACON LAB		171	1.019	0.003	0.26	1.011	1.027
2020-844	131	iCHEM VELO		27	1.025	0.001	0.07	1.023	1.028
2020-844	135	Hybrido AU		13	1.025	0.002	0.17	1.020	1.030
2020-844	138	URS-10T		11	1.018	0.003	0.32	1.008	1.028
2020-844	140	MISSION A		66	1.019	0.003	0.25	1.011	1.026
2020-844	145	CLINITEK		11	1.026	0.001	0.09	1.024	1.029
2020-844	147	Mission120		17	1.020	0.000	0.00	1.020	1.020
2020-844	148	Mission500		17	1.020	0.000	0.00	1.020	1.020

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
067	URI-SPECIFIC GRAVITY	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD						
2020-845	11 AMES NM10S			49	1.017	0.003	0.34	1.006	1.027
2020-845	13 BAYER CLI			36	1.017	0.003	0.30	1.007	1.026
2020-845	67 URISCAN			51	1.013	0.003	0.31	1.004	1.022
2020-845	71 ATLAS			18	1.030	0.001	0.06	1.028	1.032
2020-845	72 URS-9-TECO			104	1.017	0.004	0.38	1.005	1.028
2020-845	80 ROCHE (CHE			37	1.012	0.002	0.24	1.005	1.020
2020-845	89 AIMSTRIP			114	1.016	0.004	0.40	1.004	1.028
2020-845	94 CLINITEK			104	1.017	0.003	0.27	1.008	1.025
2020-845	102 ACON LAB			171	1.017	0.004	0.38	1.005	1.029
2020-845	131 iCHEM VELO			27	1.028	0.001	0.08	1.026	1.030
2020-845	135 Hybrido AU			13	1.028	0.001	0.13	1.024	1.032
2020-845	138 URS-10T			11	1.015	0.004	0.36	1.004	1.026
2020-845	140 MISSION A			66	1.016	0.004	0.40	1.004	1.028
2020-845	145 CLINITEK			11	1.030	0.001	0.09	1.027	1.033
2020-845	147 Mission120			17	1.020	0.000	0.00	1.020	1.020
2020-845	148 Mission500			17	1.019	0.002	0.16	1.015	1.024
166	MICROALBUMIN		EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}						
2020-841	58 VITROS			35	835.685	38.587	4.62	626.800	1044.600
2020-841	58 VITROS			35	753.425	62.355	8.28	565.100	941.800
2020-841	62 DADE DIM.			66	825.041	34.525	4.18	618.800	1031.300
2020-841	79 BECK. COUL			18	786.178	48.875	6.22	589.600	982.700
2020-841	85 Rx DAYTONA			18	739.733	121.889	16.48	554.800	924.700
2020-841	85 Rx DAYTONA			18	753.425	62.355	8.28	565.100	941.800
2020-841	93 ARCHITECT			20	813.753	70.389	8.65	610.300	1017.200
2020-841	93 ARCHITECT			20	753.425	62.355	8.28	565.100	941.800
2020-841	95 INTEGRA			23	812.800	31.872	3.92	609.600	1016.000
2020-841	113 ROCHE COBA			25	766.272	38.512	5.03	574.700	957.800
2020-841	139 S. EXL.200			16	753.425	62.355	8.28	565.100	941.800
2020-841	139 S. EXL.200			16	843.413	55.799	6.62	632.600	1054.300

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
166	MICROALBUMIN	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	25	%	OR 5.00	{GREATER}			
2020-842	58	VITROS		35	1.275	2.208	173.21	0	6.300	
2020-842	58	VITROS		35	6.029	0.169	2.80	1.000	11.000	
2020-842	62	DADE DIM.		66	3.840	1.569	40.86	0	8.800	
2020-842	79	BECK. COUL		18	2.522	2.826	112.04	0	7.500	
2020-842	85	Rx DAYTONA		18	1.472	2.018	137.05	0	6.500	
2020-842	85	Rx DAYTONA		18	1.275	2.208	173.21	0	6.300	
2020-842	93	ARCHITECT		20	1.275	2.208	173.21	0	6.300	
2020-842	93	ARCHITECT		20	4.989	0.031	0.62	0.000	10.000	
2020-842	95	INTEGRA		23	2.518	1.511	60.01	0	7.500	
2020-842	113	ROCHE COBA		25	11.175	2.749	24.60	6.200	16.200	
2020-842	139	S. EXL.200		16	3.306	2.004	60.61	0	8.300	
2020-842	139	S. EXL.200		16	1.275	2.208	173.21	0	6.300	
2020-843	58	VITROS		35	6.029	0.169	2.80	1.000	11.000	
2020-843	58	VITROS		35	1.275	2.208	173.21	0	6.300	
2020-843	62	DADE DIM.		66	3.191	1.469	46.03	0	8.200	
2020-843	79	BECK. COUL		18	2.511	2.826	112.53	0	7.500	
2020-843	85	Rx DAYTONA		18	1.500	2.053	136.84	0	6.500	
2020-843	85	Rx DAYTONA		18	1.275	2.208	173.21	0	6.300	
2020-843	93	ARCHITECT		20	4.963	0.113	2.28	0.000	10.000	
2020-843	93	ARCHITECT		20	1.275	2.208	173.21	0	6.300	
2020-843	95	INTEGRA		23	2.441	1.461	59.86	0	7.400	
2020-843	113	ROCHE COBA		25	11.175	2.749	24.60	6.200	16.200	
2020-843	139	S. EXL.200		16	1.275	2.208	173.21	0	6.300	
2020-843	139	S. EXL.200		16	2.313	1.067	46.12	0	7.300	
2020-844	58	VITROS		35	747.650	60.547	8.10	560.700	934.600	
2020-844	58	VITROS		35	831.318	40.249	4.84	623.500	1039.100	
2020-844	62	DADE DIM.		66	839.249	42.707	5.09	629.400	1049.100	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-844	79	BECK. COUL		18	786.383	43.197	5.49	589.800	983.000	
2020-844	85	Rx DAYTONA		18	760.133	94.109	12.38	570.100	950.200	
2020-844	85	Rx DAYTONA		18	747.650	60.547	8.10	560.700	934.600	
2020-844	93	ARCHITECT		20	747.650	60.547	8.10	560.700	934.600	
2020-844	93	ARCHITECT		20	828.395	65.474	7.90	621.300	1035.500	
2020-844	95	INTEGRA		23	811.935	27.597	3.40	609.000	1014.900	
2020-844	113	ROCHE COBA		25	777.356	47.585	6.12	583.000	971.700	
2020-844	139	S. EXL.200		16	843.894	101.520	12.03	632.900	1054.900	
2020-844	139	S. EXL.200		16	747.650	60.547	8.10	560.700	934.600	
166	MICROALBUMIN		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}					
2020-845	58	VITROS		35	816.482	47.147	5.77	612.400	1020.600	
2020-845	58	VITROS		35	734.150	48.001	6.54	550.600	917.700	
2020-845	62	DADE DIM.		66	835.119	41.427	4.96	626.300	1043.900	
2020-845	79	BECK. COUL		18	777.344	48.192	6.20	583.000	971.700	
2020-845	85	Rx DAYTONA		18	738.844	77.986	10.56	554.100	923.600	
2020-845	85	Rx DAYTONA		18	734.150	48.001	6.54	550.600	917.700	
2020-845	93	ARCHITECT		20	800.321	48.005	6.00	600.200	1000.400	
2020-845	93	ARCHITECT		20	734.150	48.001	6.54	550.600	917.700	
2020-845	95	INTEGRA		23	797.409	31.861	4.00	598.100	996.800	
2020-845	113	ROCHE COBA		25	763.968	42.802	5.60	573.000	955.000	
2020-845	139	S. EXL.200		16	734.150	48.001	6.54	550.600	917.700	
2020-845	139	S. EXL.200		16	833.067	53.013	6.36	624.800	1041.300	
167	CREATININE									
2020-841	58	VITROS		48	3.157	0.509	16.11	0	8.200	
2020-841	62	DADE DIM.		58	7.974	5.826	73.06	3.000	13.000	
2020-841	79	BECK. COUL		16	1.600	3.196	199.74	0	6.600	
2020-841	85	Rx DAYTONA		10	1.400	1.982	141.57	0	6.400	
2020-841	93	ARCHITECT		19	4.961	0.116	2.34	0.000	10.000	
2020-841	95	INTEGRA		20	0.532	0.883	166.04	0	5.500	
2020-841	113	ROCHE COBA		27	3.804	0.923	24.28	0	8.800	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 50 URINALYSIS		GRP: 70 URINALYSIS						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-841	139 S. EXL.200			19	7.789	6.098	78.28	2.800	12.800	
167 CREATININE		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 25 % OR 5.00 {GREATER}							
2020-842	58 VITROS			48	3.149	0.514	16.34	0	8.100	
2020-842	62 DADE DIM.			58	7.971	5.830	73.14	3.000	13.000	
2020-842	79 BECK. COUL			16	1.506	3.238	214.94	0	6.500	
2020-842	85 Rx DAYTONA			10	1.250	1.503	120.28	0	6.300	
2020-842	93 ARCHITECT			19	4.961	0.116	2.34	0.000	10.000	
2020-842	95 INTEGRA			20	0.479	1.136	237.15	0	5.500	
2020-842	113 ROCHE COBA			27	3.835	0.809	21.10	0	8.800	
2020-842	139 S. EXL.200			19	7.789	6.098	78.28	2.800	12.800	
2020-843	58 VITROS			48	242.354	12.233	5.05	181.800	302.900	
2020-843	62 DADE DIM.			58	235.391	10.384	4.41	176.500	294.200	
2020-843	79 BECK. COUL			16	216.450	14.521	6.71	162.300	270.600	
2020-843	85 Rx DAYTONA			10	220.640	27.463	12.45	165.500	275.800	
2020-843	93 ARCHITECT			19	229.821	7.826	3.41	172.400	287.300	
2020-843	95 INTEGRA			20	249.420	12.210	4.90	187.100	311.800	
2020-843	113 ROCHE COBA			27	245.744	6.641	2.70	184.300	307.200	
2020-843	139 S. EXL.200			19	243.126	6.423	2.64	182.300	303.900	
2020-844	58 VITROS			48	233.109	8.864	3.80	174.800	291.400	
2020-844	62 DADE DIM.			58	240.832	11.397	4.73	180.600	301.000	
2020-844	79 BECK. COUL			16	219.325	14.532	6.63	164.500	274.200	
2020-844	85 Rx DAYTONA			10	226.980	23.985	10.57	170.200	283.700	
2020-844	93 ARCHITECT			19	231.989	5.243	2.26	174.000	290.000	
2020-844	95 INTEGRA			20	252.935	10.932	4.32	189.700	316.200	
2020-844	113 ROCHE COBA			27	249.481	7.502	3.01	187.100	311.900	
2020-844	139 S. EXL.200			19	248.221	7.186	2.89	186.200	310.300	
2020-845	58 VITROS			48	233.856	8.103	3.46	175.400	292.300	

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY			SUB: 50 URINALYSIS			GRP: 70 URINALYSIS				
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2020-845	62	DADE DIM.		58	230.465	10.640	4.62	172.800	288.100	
2020-845	79	BECK. COUL		16	203.113	15.724	7.74	152.300	253.900	
2020-845	85	Rx DAYTONA		10	224.570	29.764	13.25	168.400	280.700	
2020-845	93	ARCHITECT		19	225.526	6.044	2.68	169.100	281.900	
2020-845	95	INTEGRA		20	243.895	10.638	4.36	182.900	304.900	
2020-845	113	ROCHE COBA		27	242.856	7.001	2.88	182.100	303.600	
2020-845	139	S. EXL.200		19	236.868	6.217	2.62	177.700	296.100	

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST	LEVEL 1	LEVEL 2	LEVEL 3	PART	REFEREE	COMMER	REPORTED	PART	# LABS	# "SF"	Score	
SAMPLE	RESULT CODE	RESULT CODE	RESULT CODE	LOW/HIGH	LOW/HIGH	LOW/HIGH	VALUE	LABS	ANSW	LABS		
057 URINALYSIS-BILIRUBIN												
2020-841	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	49	49	49	100
2020-841	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	37	37	37	100
2020-841	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	49	49	100
2020-841	067	URISCAN		057	057	057	057	058 - TRACES, 1+	50	1	49	0
2020-841	071	ATLAS		057	057	057	057	057 - NEGATIVE	18	18	18	100
2020-841	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-841	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	36	36	36	100
2020-841	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	113	113	113	100
2020-841	094	CLINITEK		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-841	102	ACON LAB		057	057	057	057	057 - NEGATIVE	167	167	167	100
2020-841	131	iCHEM VELO		057	057	057	057	057 - NEGATIVE	26	26	26	100
2020-841	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	13	13	13	100
2020-841	138	URS-10T		057	057	057	057	057 - NEGATIVE	11	11	11	100
2020-841	140	MISSION A		057	057	057	057	057 - NEGATIVE	67	67	67	100
2020-841	145	CLINITEK		057	057	057	057	057 - NEGATIVE	10	10	10	100
2020-841	147	Mission120		057	057	057	057	057 - NEGATIVE	17	17	17	100
2020-841	148	Mission500		057	057	057	057	057 - NEGATIVE	21	21	21	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
057 URINALYSIS-BILIRUBIN												
2020-842	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	49	49	49	100
2020-842	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	37	37	37	100
2020-842	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	50	50	100
2020-842	071	ATLAS		057	057	057	057	057 - NEGATIVE	18	18	18	100
2020-842	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-842	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	36	36	36	100
2020-842	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	113	113	113	100
2020-842	094	CLINITEK		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-842	102	ACON LAB		057	057	057	057	057 - NEGATIVE	167	167	167	100
2020-842	131	ICHEM VELO		057	057	057	057	057 - NEGATIVE	26	26	26	100
2020-842	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	13	13	13	100
2020-842	138	URS-10T		057	057	057	057	057 - NEGATIVE	11	11	11	100
2020-842	140	MISSION A		057	057	057	057	057 - NEGATIVE	67	67	67	100
2020-842	145	CLINITEK		057	057	057	057	057 - NEGATIVE	10	10	10	100
2020-842	147	Mission120		057	057	057	057	057 - NEGATIVE	17	17	17	100
2020-842	148	Mission500		057	057	057	057	057 - NEGATIVE	21	21	21	100
2020-843	011	AMES NM10S		057	057	057	057	057 - NEGATIVE	49	49	49	100
2020-843	013	BAYER CLI		057	057	057	057	057 - NEGATIVE	37	37	37	100
2020-843	067	URISCAN		057	057	057	057	057 - NEGATIVE	50	50	50	100
2020-843	071	ATLAS		057	057	057	057	057 - NEGATIVE	18	18	18	100
2020-843	072	URS-9-TECO		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-843	080	ROCHE (CHE		057	057	057	057	057 - NEGATIVE	36	36	36	100
2020-843	089	AIMSTRIP		057	057	057	057	057 - NEGATIVE	113	113	113	100
2020-843	094	CLINITEK		057	057	057	057	057 - NEGATIVE	104	104	104	100
2020-843	102	ACON LAB		057	057	057	057	057 - NEGATIVE	167	167	167	100
2020-843	131	ICHEM VELO		057	057	057	057	057 - NEGATIVE	26	26	26	100
2020-843	135	Hybrido AU		057	057	057	057	057 - NEGATIVE	13	13	13	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-843	138	URS-10T		057 057	057 057		057 - NEGATIVE	11	11	11	100	
2020-843	140	MISSION A		057 057	057 057		057 - NEGATIVE	67	67	67	100	
2020-843	145	CLINITEK		057 057	057 057		057 - NEGATIVE	10	10	10	100	
2020-843	147	Mission120		057 057	057 057		057 - NEGATIVE	17	17	17	100	
2020-843	148	Mission500		057 057	057 057		057 - NEGATIVE	21	21	21	100	
057 URINALYSIS-BILIRUBIN												
2020-844	011	AMES NM10S		057 057	057 057		057 - NEGATIVE	49	49	49	100	
2020-844	013	BAYER CLI		057 057	057 057		057 - NEGATIVE	37	37	37	100	
2020-844	067	URISCAN		057 057	057 057		057 - NEGATIVE	50	50	50	100	
2020-844	071	ATLAS		057 057	057 057		057 - NEGATIVE	18	18	18	100	
2020-844	072	URS-9-TECO		057 057	057 057		057 - NEGATIVE	104	104	104	100	
2020-844	080	ROCHE (CHE		057 057	057 057		057 - NEGATIVE	36	36	36	100	
2020-844	089	AIMSTRIP		057 057	057 057		057 - NEGATIVE	113	113	113	100	
2020-844	094	CLINITEK		057 057	057 057		057 - NEGATIVE	104	103	103	100	
2020-844	094	CLINITEK		057 057	057 057		059 - 2+	104	1	103	0	
2020-844	102	ACON LAB		057 057	057 057		057 - NEGATIVE	167	166	166	100	
2020-844	102	ACON LAB		057 057	057 057		058 - TRACES, 1+	167	1	166	0	
2020-844	131	iCHEM VELO		057 057	057 057		057 - NEGATIVE	26	26	26	100	
2020-844	135	Hybrido AU		057 057	057 057		057 - NEGATIVE	13	13	13	100	
2020-844	138	URS-10T		057 057	057 057		057 - NEGATIVE	11	11	11	100	
2020-844	140	MISSION A		057 057	057 057		057 - NEGATIVE	67	67	67	100	
2020-844	145	CLINITEK		057 057	057 057		057 - NEGATIVE	10	10	10	100	
2020-844	147	Mission120		057 057	057 057		057 - NEGATIVE	17	17	17	100	
2020-844	148	Mission500		057 057	057 057		057 - NEGATIVE	21	21	21	100	
2020-845	011	AMES NM10S		058 060	057 059		058 - TRACES, 1+	49	7	49	100	
2020-845	011	AMES NM10S		058 060	057 059		059 - 2+	49	36	49	100	
2020-845	011	AMES NM10S		058 060	057 059		060 - 3+	49	6	49	100	
2020-845	013	BAYER CLI		058 060	057 059		058 - TRACES, 1+	37	8	37	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	013	BAYER CLI		058 060	057 059		059 - 2+	37	27	37	100	
2020-845	013	BAYER CLI		058 060	057 059		060 - 3+	37	2	37	100	
2020-845	067	URISCAN		057 059	057 059		057 - NEGATIVE	50	11	49	100	
2020-845	067	URISCAN		057 059	057 059		058 - TRACES, 1+	50	27	49	100	
2020-845	067	URISCAN		057 059	057 059		059 - 2+	50	11	49	100	
2020-845	067	URISCAN		057 059	057 059		060 - 3+	50	1	49	0	
2020-845	071	ATLAS		057 059	057 059		058 - TRACES, 1+	18	18	18	100	
2020-845	072	URS-9-TECO		058 060	057 059		057 - NEGATIVE	104	2	104	100	
2020-845	072	URS-9-TECO		058 060	057 059		058 - TRACES, 1+	104	25	104	100	
2020-845	072	URS-9-TECO		058 060	057 059		059 - 2+	104	50	104	100	
2020-845	072	URS-9-TECO		058 060	057 059		060 - 3+	104	27	104	100	
2020-845	080	ROCHE (CHE		057 059	057 059		058 - TRACES, 1+	36	32	34	100	
2020-845	080	ROCHE (CHE		057 059	057 059		059 - 2+	36	2	34	100	
2020-845	080	ROCHE (CHE		057 059	057 059		060 - 3+	36	2	34	0	
2020-845	089	AIMSTRIP		058 060	057 059		057 - NEGATIVE	113	5	113	100	
2020-845	089	AIMSTRIP		058 060	057 059		058 - TRACES, 1+	113	47	113	100	
2020-845	089	AIMSTRIP		058 060	057 059		059 - 2+	113	52	113	100	
2020-845	089	AIMSTRIP		058 060	057 059		060 - 3+	113	9	113	100	
2020-845	094	CLINITEK		058 060	057 059		057 - NEGATIVE	104	1	104	100	
2020-845	094	CLINITEK		058 060	057 059		058 - TRACES, 1+	104	27	104	100	
2020-845	094	CLINITEK		058 060	057 059		059 - 2+	104	75	104	100	
2020-845	094	CLINITEK		058 060	057 059		060 - 3+	104	1	104	100	
2020-845	102	ACON LAB		057 059	057 059		057 - NEGATIVE	167	4	160	100	
2020-845	102	ACON LAB		057 059	057 059		058 - TRACES, 1+	167	112	160	100	
2020-845	102	ACON LAB		057 059	057 059		059 - 2+	167	44	160	100	
2020-845	102	ACON LAB		057 059	057 059		060 - 3+	167	7	160	0	
2020-845	131	iCHEM VELO		057 057	057 059		057 - NEGATIVE	26	26	26	100	
2020-845	135	Hybrido AU		058 060	057 059		059 - 2+	13	11	13	100	
2020-845	135	Hybrido AU		058 060	057 059		060 - 3+	13	2	13	100	
2020-845	138	URS-10T		058 060	057 059		058 - TRACES, 1+	11	4	11	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	138	URS-10T		058 060	057 059		059 - 2+	11	5	11	100	
2020-845	138	URS-10T		058 060	057 059		060 - 3+	11	2	11	100	
2020-845	140	MISSION A		057 059	057 059		057 - NEGATIVE	67	2	65	100	
2020-845	140	MISSION A		057 059	057 059		058 - TRACES, 1+	67	51	65	100	
2020-845	140	MISSION A		057 059	057 059		059 - 2+	67	12	65	100	
2020-845	140	MISSION A		057 059	057 059		060 - 3+	67	2	65	0	
2020-845	145	CLINITEK		057 057	057 059		057 - NEGATIVE	10	9	10	100	
2020-845	145	CLINITEK		057 057	057 059		058 - TRACES, 1+	10	1	10	100	
2020-845	147	Mission120		057 059	057 059		058 - TRACES, 1+	17	15	17	100	
2020-845	147	Mission120		057 059	057 059		059 - 2+	17	2	17	100	
2020-845	148	Mission500		057 059	057 059		058 - TRACES, 1+	21	18	20	100	
2020-845	148	Mission500		057 059	057 059		059 - 2+	21	2	20	100	
2020-845	148	Mission500		057 059	057 059		060 - 3+	21	1	20	0	
058 URINALYSIS-GLUCOSE												
2020-841	011	AMES NM10S		076 078	076 078		075 - 250mg/dl-+	49	1	48	0	
2020-841	011	AMES NM10S		076 078	076 078		076 - 500mg/dl2+	49	12	48	100	
2020-841	011	AMES NM10S		076 078	076 078		077 - 1000mg/dl	49	34	48	100	
2020-841	011	AMES NM10S		076 078	076 078		078 - 2000mg/dl	49	2	48	100	
2020-841	013	BAYER CLI		076 078	076 078		075 - 250mg/dl-+	36	1	35	0	
2020-841	013	BAYER CLI		076 078	076 078		076 - 500mg/dl2+	36	14	35	100	
2020-841	013	BAYER CLI		076 078	076 078		077 - 1000mg/dl	36	21	35	100	
2020-841	067	URISCAN		077 078	076 078		076 - 500mg/dl2+	50	1	50	100	
2020-841	067	URISCAN		077 078	076 078		077 - 1000mg/dl	50	17	50	100	
2020-841	067	URISCAN		077 078	076 078		078 - 2000mg/dl	50	32	50	100	
2020-841	071	ATLAS		076 078	076 078		077 - 1000mg/dl	18	17	18	100	
2020-841	071	ATLAS		076 078	076 078		078 - 2000mg/dl	18	1	18	100	
2020-841	072	URS-9-TECO		076 078	076 078		076 - 500mg/dl2+	105	29	105	100	
2020-841	072	URS-9-TECO		076 078	076 078		077 - 1000mg/dl	105	64	105	100	
2020-841	072	URS-9-TECO		076 078	076 078		078 - 2000mg/dl	105	12	105	100	
2020-841	080	ROCHE (CHE		076 078	076 078		077 - 1000mg/dl	37	36	37	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-841	080	ROCHE (CHE		076 078	076 078		078 - 2000mg/dl	37	1	37	100
2020-841	089	AIMSTRIP		075 077	076 078		075 - 250mg/dl-+	113	8	113	100
2020-841	089	AIMSTRIP		075 077	076 078		076 - 500mg/dl2+	113	63	113	100
2020-841	089	AIMSTRIP		075 077	076 078		077 - 1000mg/dl	113	40	113	100
2020-841	089	AIMSTRIP		075 077	076 078		078 - 2000mg/dl	113	2	113	100
2020-841	094	CLINITEK		076 078	076 078		076 - 500mg/dl2+	104	36	104	100
2020-841	094	CLINITEK		076 078	076 078		077 - 1000mg/dl	104	64	104	100
2020-841	094	CLINITEK		076 078	076 078		078 - 2000mg/dl	104	4	104	100
2020-841	102	ACON LAB		075 077	076 078		076 - 500mg/dl2+	170	103	170	100
2020-841	102	ACON LAB		075 077	076 078		077 - 1000mg/dl	170	66	170	100
2020-841	102	ACON LAB		075 077	076 078		078 - 2000mg/dl	170	1	170	100
2020-841	131	iCHEM VELO		075 077	076 078		076 - 500mg/dl2+	27	23	27	100
2020-841	131	iCHEM VELO		075 077	076 078		077 - 1000mg/dl	27	4	27	100
2020-841	135	Hybrido AU		076 078	076 078		077 - 1000mg/dl	13	8	13	100
2020-841	135	Hybrido AU		076 078	076 078		078 - 2000mg/dl	13	5	13	100
2020-841	138	URS-10T		076 078	076 078		076 - 500mg/dl2+	11	2	11	100
2020-841	138	URS-10T		076 078	076 078		077 - 1000mg/dl	11	9	11	100
2020-841	140	MISSION A		076 078	076 078		075 - 250mg/dl-+	65	1	64	0
2020-841	140	MISSION A		076 078	076 078		076 - 500mg/dl2+	65	25	64	100
2020-841	140	MISSION A		076 078	076 078		077 - 1000mg/dl	65	36	64	100
2020-841	140	MISSION A		076 078	076 078		078 - 2000mg/dl	65	3	64	100
2020-841	145	CLINITEK		076 078	076 078		077 - 1000mg/dl	10	9	10	100
2020-841	145	CLINITEK		076 078	076 078		078 - 2000mg/dl	10	1	10	100
2020-841	147	Mission120		076 078	076 078		076 - 500mg/dl2+	17	1	17	100
2020-841	147	Mission120		076 078	076 078		077 - 1000mg/dl	17	12	17	100
2020-841	147	Mission120		076 078	076 078		078 - 2000mg/dl	17	4	17	100
2020-841	148	Mission500		076 078	076 078		076 - 500mg/dl2+	21	7	21	100
2020-841	148	Mission500		076 078	076 078		077 - 1000mg/dl	21	13	21	100
2020-841	148	Mission500		076 078	076 078		078 - 2000mg/dl	21	1	21	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
058 URINALYSIS-GLUCOSE													
2020-842	011	AMES NM10S		074	076	074	076	073 - NEGATIVE	49	1	48	0	
2020-842	011	AMES NM10S		074	076	074	076	074 - TRACES	49	22	48	100	
2020-842	011	AMES NM10S		074	076	074	076	075 - 250mg/dl-+	49	25	48	100	
2020-842	011	AMES NM10S		074	076	074	076	076 - 500mg/dl2+	49	1	48	100	
2020-842	013	BAYER CLI		074	076	074	076	073 - NEGATIVE	36	1	35	0	
2020-842	013	BAYER CLI		074	076	074	076	074 - TRACES	36	14	35	100	
2020-842	013	BAYER CLI		074	076	074	076	075 - 250mg/dl-+	36	21	35	100	
2020-842	067	URISCAN		074	076	074	076	074 - TRACES	50	1	43	100	
2020-842	067	URISCAN		074	076	074	076	075 - 250mg/dl-+	50	21	43	100	
2020-842	067	URISCAN		074	076	074	076	076 - 500mg/dl2+	50	21	43	100	
2020-842	067	URISCAN		074	076	074	076	077 - 1000mg/dl	50	4	43	0	
2020-842	067	URISCAN		074	076	074	076	078 - 2000mg/dl	50	3	43	0	
2020-842	071	ATLAS		076	078	074	076	076 - 500mg/dl2+	18	6	18	100	
2020-842	071	ATLAS		076	078	074	076	077 - 1000mg/dl	18	12	18	100	
2020-842	072	URS-9-TECO		074	076	074	076	073 - NEGATIVE	105	10	95	0	
2020-842	072	URS-9-TECO		074	076	074	076	074 - TRACES	105	30	95	100	
2020-842	072	URS-9-TECO		074	076	074	076	075 - 250mg/dl-+	105	58	95	100	
2020-842	072	URS-9-TECO		074	076	074	076	076 - 500mg/dl2+	105	7	95	100	
2020-842	080	ROCHE (CHE		076	078	074	076	075 - 250mg/dl-+	37	2	37	100	
2020-842	080	ROCHE (CHE		076	078	074	076	076 - 500mg/dl2+	37	1	37	100	
2020-842	080	ROCHE (CHE		076	078	074	076	077 - 1000mg/dl	37	33	37	100	
2020-842	080	ROCHE (CHE		076	078	074	076	078 - 2000mg/dl	37	1	37	100	
2020-842	089	AIMSTRIP		074	076	074	076	073 - NEGATIVE	113	1	110	0	
2020-842	089	AIMSTRIP		074	076	074	076	074 - TRACES	113	21	110	100	
2020-842	089	AIMSTRIP		074	076	074	076	075 - 250mg/dl-+	113	89	110	100	
2020-842	089	AIMSTRIP		074	076	074	076	078 - 2000mg/dl	113	2	110	0	
2020-842	094	CLINITEK		074	076	074	076	074 - TRACES	104	50	104	100	
2020-842	094	CLINITEK		074	076	074	076	075 - 250mg/dl-+	104	54	104	100	
2020-842	102	ACON LAB		074	076	074	076	073 - NEGATIVE	170	1	169	0	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-842	102	ACON LAB		074 076	074 076		074 - TRACES	170	26	169	100	
2020-842	102	ACON LAB		074 076	074 076		075 - 250mg/dl-+	170	143	169	100	
2020-842	131	iCHEM VELO		075 077	074 076		076 - 500mg/dl2+	27	23	27	100	
2020-842	131	iCHEM VELO		075 077	074 076		077 - 1000mg/dl	27	4	27	100	
2020-842	135	Hybrido AU		076 078	074 076		077 - 1000mg/dl	13	8	13	100	
2020-842	135	Hybrido AU		076 078	074 076		078 - 2000mg/dl	13	5	13	100	
2020-842	138	URS-10T		073 075	074 076		074 - TRACES	11	6	11	100	
2020-842	138	URS-10T		073 075	074 076		075 - 250mg/dl-+	11	5	11	100	
2020-842	140	MISSION A		074 076	074 076		073 - NEGATIVE	65	1	64	0	
2020-842	140	MISSION A		074 076	074 076		074 - TRACES	65	9	64	100	
2020-842	140	MISSION A		074 076	074 076		075 - 250mg/dl-+	65	54	64	100	
2020-842	140	MISSION A		074 076	074 076		076 - 500mg/dl2+	65	1	64	100	
2020-842	145	CLINITEK		076 078	074 076		077 - 1000mg/dl	10	9	10	100	
2020-842	145	CLINITEK		076 078	074 076		078 - 2000mg/dl	10	1	10	100	
2020-842	147	Mission120		074 076	074 076		074 - TRACES	17	1	17	100	
2020-842	147	Mission120		074 076	074 076		075 - 250mg/dl-+	17	16	17	100	
2020-842	148	Mission500		074 076	074 076		075 - 250mg/dl-+	21	21	21	100	
058 URINALYSIS-GLUCOSE												
2020-843	011	AMES NM10S		076 078	076 078		075 - 250mg/dl-+	49	1	48	0	
2020-843	011	AMES NM10S		076 078	076 078		076 - 500mg/dl2+	49	14	48	100	
2020-843	011	AMES NM10S		076 078	076 078		077 - 1000mg/dl	49	31	48	100	
2020-843	011	AMES NM10S		076 078	076 078		078 - 2000mg/dl	49	3	48	100	
2020-843	013	BAYER CLI		076 078	076 078		075 - 250mg/dl-+	36	1	35	0	
2020-843	013	BAYER CLI		076 078	076 078		076 - 500mg/dl2+	36	13	35	100	
2020-843	013	BAYER CLI		076 078	076 078		077 - 1000mg/dl	36	22	35	100	
2020-843	067	URISCAN		077 078	076 078		075 - 250mg/dl-+	50	1	49	0	
2020-843	067	URISCAN		077 078	076 078		076 - 500mg/dl2+	50	6	49	100	
2020-843	067	URISCAN		077 078	076 078		077 - 1000mg/dl	50	13	49	100	
2020-843	067	URISCAN		077 078	076 078		078 - 2000mg/dl	50	30	49	100	
2020-843	071	ATLAS		075 077	076 078		076 - 500mg/dl2+	18	17	18	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-843	071	ATLAS		075 077	076 078		077 - 1000mg/dl	18	1	18	100	
2020-843	072	URS-9-TECO		076 078	076 078		074 - TRACES	105	1	103	0	
2020-843	072	URS-9-TECO		076 078	076 078		075 - 250mg/dl-+	105	1	103	0	
2020-843	072	URS-9-TECO		076 078	076 078		076 - 500mg/dl2+	105	38	103	100	
2020-843	072	URS-9-TECO		076 078	076 078		077 - 1000mg/dl	105	59	103	100	
2020-843	072	URS-9-TECO		076 078	076 078		078 - 2000mg/dl	105	6	103	100	
2020-843	080	ROCHE (CHE		076 078	076 078		077 - 1000mg/dl	37	36	37	100	
2020-843	080	ROCHE (CHE		076 078	076 078		078 - 2000mg/dl	37	1	37	100	
2020-843	089	AIMSTRIP		075 077	076 078		075 - 250mg/dl-+	113	2	113	100	
2020-843	089	AIMSTRIP		075 077	076 078		076 - 500mg/dl2+	113	66	113	100	
2020-843	089	AIMSTRIP		075 077	076 078		077 - 1000mg/dl	113	43	113	100	
2020-843	089	AIMSTRIP		075 077	076 078		078 - 2000mg/dl	113	2	113	100	
2020-843	094	CLINITEK		076 078	076 078		074 - TRACES	104	2	100	0	
2020-843	094	CLINITEK		076 078	076 078		075 - 250mg/dl-+	104	2	100	0	
2020-843	094	CLINITEK		076 078	076 078		076 - 500mg/dl2+	104	31	100	100	
2020-843	094	CLINITEK		076 078	076 078		077 - 1000mg/dl	104	66	100	100	
2020-843	094	CLINITEK		076 078	076 078		078 - 2000mg/dl	104	3	100	100	
2020-843	102	ACON LAB		075 077	076 078		076 - 500mg/dl2+	170	96	170	100	
2020-843	102	ACON LAB		075 077	076 078		077 - 1000mg/dl	170	73	170	100	
2020-843	102	ACON LAB		075 077	076 078		078 - 2000mg/dl	170	1	170	100	
2020-843	131	iCHEM VELO		075 077	076 078		076 - 500mg/dl2+	27	23	27	100	
2020-843	131	iCHEM VELO		075 077	076 078		077 - 1000mg/dl	27	4	27	100	
2020-843	135	Hybrido AU		076 078	076 078		077 - 1000mg/dl	13	8	13	100	
2020-843	135	Hybrido AU		076 078	076 078		078 - 2000mg/dl	13	5	13	100	
2020-843	138	URS-10T		076 078	076 078		076 - 500mg/dl2+	11	2	11	100	
2020-843	138	URS-10T		076 078	076 078		077 - 1000mg/dl	11	9	11	100	
2020-843	140	MISSION A		076 078	076 078		075 - 250mg/dl-+	65	2	63	0	
2020-843	140	MISSION A		076 078	076 078		076 - 500mg/dl2+	65	23	63	100	
2020-843	140	MISSION A		076 078	076 078		077 - 1000mg/dl	65	34	63	100	
2020-843	140	MISSION A		076 078	076 078		078 - 2000mg/dl	65	6	63	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-843	145	CLINITEK		075 077	076 078		076 - 500mg/dl2+	10	9	10	100		
2020-843	145	CLINITEK		075 077	076 078		077 - 1000mg/dl	10	1	10	100		
2020-843	147	Mission120		076 078	076 078		076 - 500mg/dl2+	17	2	17	100		
2020-843	147	Mission120		076 078	076 078		077 - 1000mg/dl	17	13	17	100		
2020-843	147	Mission120		076 078	076 078		078 - 2000mg/dl	17	2	17	100		
2020-843	148	Mission500		076 078	076 078		076 - 500mg/dl2+	21	9	21	100		
2020-843	148	Mission500		076 078	076 078		077 - 1000mg/dl	21	10	21	100		
2020-843	148	Mission500		076 078	076 078		078 - 2000mg/dl	21	2	21	100		
058 URINALYSIS-GLUCOSE													
2020-844	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	49	49	49	100		
2020-844	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	36	36	36	100		
2020-844	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100		
2020-844	071	ATLAS		073 073	073 073		073 - NEGATIVE	18	18	18	100		
2020-844	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	105	105	105	100		
2020-844	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	37	37	37	100		
2020-844	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	113	113	113	100		
2020-844	094	CLINITEK		073 073	073 073		073 - NEGATIVE	104	104	104	100		
2020-844	102	ACON LAB		073 073	073 073		073 - NEGATIVE	170	170	170	100		
2020-844	131	ICHEM VELO		073 073	073 073		073 - NEGATIVE	27	27	27	100		
2020-844	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	13	13	13	100		
2020-844	138	URS-10T		073 073	073 073		073 - NEGATIVE	11	11	11	100		
2020-844	140	MISSION A		073 073	073 073		073 - NEGATIVE	65	65	65	100		
2020-844	145	CLINITEK		073 073	073 073		073 - NEGATIVE	10	10	10	100		
2020-844	147	Mission120		073 073	073 073		073 - NEGATIVE	17	17	17	100		
2020-844	148	Mission500		073 073	073 073		073 - NEGATIVE	21	21	21	100		
2020-845	011	AMES NM10S		073 073	073 073		073 - NEGATIVE	49	49	49	100		
2020-845	013	BAYER CLI		073 073	073 073		073 - NEGATIVE	36	36	36	100		
2020-845	067	URISCAN		073 073	073 073		073 - NEGATIVE	50	50	50	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	071	ATLAS		073 073	073 073		073 - NEGATIVE	18	18	18	100	
2020-845	072	URS-9-TECO		073 073	073 073		073 - NEGATIVE	105	105	105	100	
2020-845	080	ROCHE (CHE		073 073	073 073		073 - NEGATIVE	37	37	37	100	
2020-845	089	AIMSTRIP		073 073	073 073		073 - NEGATIVE	113	113	113	100	
2020-845	094	CLINITEK		073 073	073 073		073 - NEGATIVE	104	104	104	100	
2020-845	102	ACON LAB		073 073	073 073		073 - NEGATIVE	170	170	170	100	
2020-845	131	iCHEM VELO		073 073	073 073		073 - NEGATIVE	27	27	27	100	
2020-845	135	Hybrido AU		073 073	073 073		073 - NEGATIVE	13	13	13	100	
2020-845	138	URS-10T		073 073	073 073		073 - NEGATIVE	11	11	11	100	
2020-845	140	MISSION A		073 073	073 073		073 - NEGATIVE	65	65	65	100	
2020-845	145	CLINITEK		073 073	073 073		073 - NEGATIVE	10	10	10	100	
2020-845	147	Mission120		073 073	073 073		073 - NEGATIVE	17	17	17	100	
2020-845	148	Mission500		073 073	073 073		073 - NEGATIVE	21	21	21	100	
059 URINALYS.HG.OR BLOOD												
2020-841	011	AMES NM10S		068 069	068 069		067 - TRACES	49	1	48	0	
2020-841	011	AMES NM10S		068 069	068 069		068 - POS.(1-2+)	49	4	48	100	
2020-841	011	AMES NM10S		068 069	068 069		069 - POS.(3+)	49	44	48	100	
2020-841	013	BAYER CLI		068 069	068 069		068 - POS.(1-2+)	36	2	36	100	
2020-841	013	BAYER CLI		068 069	068 069		069 - POS.(3+)	36	34	36	100	
2020-841	067	URISCAN		068 069	068 069		068 - POS.(1-2+)	50	6	50	100	
2020-841	067	URISCAN		068 069	068 069		069 - POS.(3+)	50	44	50	100	
2020-841	071	ATLAS		068 069	068 069		069 - POS.(3+)	18	18	18	100	
2020-841	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+)	105	9	105	100	
2020-841	072	URS-9-TECO		068 069	068 069		069 - POS.(3+)	105	96	105	100	
2020-841	080	ROCHE (CHE		068 069	068 069		067 - TRACES	37	3	34	0	
2020-841	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+)	37	1	34	100	
2020-841	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+)	37	33	34	100	
2020-841	089	AIMSTRIP		068 069	068 069		066 - NEGATIVE	113	1	112	0	
2020-841	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+)	113	4	112	100	
2020-841	089	AIMSTRIP		068 069	068 069		069 - POS.(3+)	113	108	112	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-841	094	CLINITEK		068 069	068 069		068 - POS.(1-2+)	104	2	104	100		
2020-841	094	CLINITEK		068 069	068 069		069 - POS.(3+)	104	102	104	100		
2020-841	102	ACON LAB		068 069	068 069		067 - TRACES	171	1	170	0		
2020-841	102	ACON LAB		068 069	068 069		068 - POS.(1-2+)	171	5	170	100		
2020-841	102	ACON LAB		068 069	068 069		069 - POS.(3+)	171	165	170	100		
2020-841	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+)	27	2	27	100		
2020-841	131	iCHEM VELO		068 069	068 069		069 - POS.(3+)	27	25	27	100		
2020-841	135	Hybrido AU		068 069	068 069		068 - POS.(1-2+)	13	1	13	100		
2020-841	135	Hybrido AU		068 069	068 069		069 - POS.(3+)	13	12	13	100		
2020-841	138	URS-10T		068 069	068 069		068 - POS.(1-2+)	11	1	11	100		
2020-841	138	URS-10T		068 069	068 069		069 - POS.(3+)	11	10	11	100		
2020-841	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	66	4	66	100		
2020-841	140	MISSION A		068 069	068 069		069 - POS.(3+)	66	62	66	100		
2020-841	145	CLINITEK		068 069	068 069		069 - POS.(3+)	10	10	10	100		
2020-841	147	Mission120		068 069	068 069		069 - POS.(3+)	17	17	17	100		
2020-841	148	Mission500		068 069	068 069		068 - POS.(1-2+)	19	1	19	100		
2020-841	148	Mission500		068 069	068 069		069 - POS.(3+)	19	18	19	100		
059 URINALYS.HG.OR BLOOD													
2020-842	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	49	49	49	100		
2020-842	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	36	35	35	100		
2020-842	013	BAYER CLI		066 066	066 066		067 - TRACES	36	1	35	0		
2020-842	067	URISCAN		066 066	066 066		066 - NEGATIVE	50	50	50	100		
2020-842	071	ATLAS		066 066	066 066		066 - NEGATIVE	18	18	18	100		
2020-842	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	105	102	102	100		
2020-842	072	URS-9-TECO		066 066	066 066		067 - TRACES	105	3	102	0		
2020-842	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	37	37	37	100		
2020-842	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	113	113	113	100		
2020-842	094	CLINITEK		066 066	066 066		066 - NEGATIVE	104	104	104	100		
2020-842	102	ACON LAB		066 066	066 066		066 - NEGATIVE	171	171	171	100		
2020-842	131	iCHEM VELO		066 066	066 066		066 - NEGATIVE	27	27	27	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-842	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	13	13	13	100	
2020-842	138	URS-10T		066 066	066 066		066 - NEGATIVE	11	11	11	100	
2020-842	140	MISSION A		066 066	066 066		066 - NEGATIVE	66	66	66	100	
2020-842	145	CLINITEK		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-842	147	Mission120		066 066	066 066		066 - NEGATIVE	17	17	17	100	
2020-842	148	Mission500		066 066	066 066		066 - NEGATIVE	19	19	19	100	
059 URINALYS.HG.OR BLOOD												
2020-843	011	AMES NM10S		066 066	066 066		066 - NEGATIVE	49	47	47	100	
2020-843	011	AMES NM10S		066 066	066 066		068 - POS. (1-2+)	49	1	47	0	
2020-843	011	AMES NM10S		066 066	066 066		069 - POS. (3+)	49	1	47	0	
2020-843	013	BAYER CLI		066 066	066 066		066 - NEGATIVE	36	36	36	100	
2020-843	067	URISCAN		066 066	066 066		066 - NEGATIVE	50	50	50	100	
2020-843	071	ATLAS		066 066	066 066		066 - NEGATIVE	18	18	18	100	
2020-843	072	URS-9-TECO		066 066	066 066		066 - NEGATIVE	105	105	105	100	
2020-843	080	ROCHE (CHE		066 066	066 066		066 - NEGATIVE	37	37	37	100	
2020-843	089	AIMSTRIP		066 066	066 066		066 - NEGATIVE	113	113	113	100	
2020-843	094	CLINITEK		066 066	066 066		066 - NEGATIVE	104	104	104	100	
2020-843	102	ACON LAB		066 066	066 066		066 - NEGATIVE	171	171	171	100	
2020-843	131	ICHEM VELO		066 066	066 066		066 - NEGATIVE	27	27	27	100	
2020-843	135	Hybrido AU		066 066	066 066		066 - NEGATIVE	13	13	13	100	
2020-843	138	URS-10T		066 066	066 066		066 - NEGATIVE	11	11	11	100	
2020-843	140	MISSION A		066 066	066 066		066 - NEGATIVE	66	66	66	100	
2020-843	145	CLINITEK		066 066	066 066		066 - NEGATIVE	10	10	10	100	
2020-843	147	Mission120		066 066	066 066		066 - NEGATIVE	17	17	17	100	
2020-843	148	Mission500		066 066	066 066		066 - NEGATIVE	19	19	19	100	
2020-844	011	AMES NM10S		068 069	068 069		066 - NEGATIVE	49	1	47	0	
2020-844	011	AMES NM10S		068 069	068 069		067 - TRACES	49	1	47	0	
2020-844	011	AMES NM10S		068 069	068 069		068 - POS. (1-2+)	49	1	47	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-844	011	AMES NM10S		068 069	068 069		069 - POS.(3+))	49	46	47	100
2020-844	013	BAYER CLI		068 069	068 069		069 - POS.(3+))	36	36	36	100
2020-844	067	URISCAN		068 069	068 069		068 - POS.(1-2+))	50	4	50	100
2020-844	067	URISCAN		068 069	068 069		069 - POS.(3+))	50	46	50	100
2020-844	071	ATLAS		068 069	068 069		069 - POS.(3+))	18	18	18	100
2020-844	072	URS-9-TECO		068 069	068 069		066 - NEGATIVE	105	3	101	0
2020-844	072	URS-9-TECO		068 069	068 069		067 - TRACES	105	1	101	0
2020-844	072	URS-9-TECO		068 069	068 069		068 - POS.(1-2+))	105	11	101	100
2020-844	072	URS-9-TECO		068 069	068 069		069 - POS.(3+))	105	90	101	100
2020-844	080	ROCHE (CHE		068 069	068 069		066 - NEGATIVE	37	1	36	0
2020-844	080	ROCHE (CHE		068 069	068 069		068 - POS.(1-2+))	37	1	36	100
2020-844	080	ROCHE (CHE		068 069	068 069		069 - POS.(3+))	37	35	36	100
2020-844	089	AIMSTRIP		068 069	068 069		068 - POS.(1-2+))	113	3	113	100
2020-844	089	AIMSTRIP		068 069	068 069		069 - POS.(3+))	113	110	113	100
2020-844	094	CLINITEK		068 069	068 069		068 - POS.(1-2+))	104	1	104	100
2020-844	094	CLINITEK		068 069	068 069		069 - POS.(3+))	104	103	104	100
2020-844	102	ACON LAB		068 069	068 069		068 - POS.(1-2+))	171	2	171	100
2020-844	102	ACON LAB		068 069	068 069		069 - POS.(3+))	171	169	171	100
2020-844	131	iCHEM VELO		068 069	068 069		068 - POS.(1-2+))	27	2	27	100
2020-844	131	iCHEM VELO		068 069	068 069		069 - POS.(3+))	27	25	27	100
2020-844	135	Hybrido AU		068 069	068 069		068 - POS.(1-2+))	13	4	13	100
2020-844	135	Hybrido AU		068 069	068 069		069 - POS.(3+))	13	9	13	100
2020-844	138	URS-10T		068 069	068 069		068 - POS.(1-2+))	11	1	11	100
2020-844	138	URS-10T		068 069	068 069		069 - POS.(3+))	11	10	11	100
2020-844	140	MISSION A		068 069	068 069		068 - POS.(1-2+))	66	2	66	100
2020-844	140	MISSION A		068 069	068 069		069 - POS.(3+))	66	64	66	100
2020-844	145	CLINITEK		068 069	068 069		069 - POS.(3+))	10	10	10	100
2020-844	147	Mission120		068 069	068 069		069 - POS.(3+))	17	17	17	100
2020-844	148	Mission500		068 069	068 069		069 - POS.(3+))	19	19	19	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
059 URINALYS.HG.OR BLOOD													
2020-845	011	AMES NM10S		068	069	068	069	066 - NEGATIVE	49	1	47	0	
2020-845	011	AMES NM10S		068	069	068	069	067 - TRACES	49	1	47	0	
2020-845	011	AMES NM10S		068	069	068	069	068 - POS.(1-2+)	49	15	47	100	
2020-845	011	AMES NM10S		068	069	068	069	069 - POS.(3+))	49	32	47	100	
2020-845	013	BAYER CLI		068	069	068	069	068 - POS.(1-2+)	36	6	36	100	
2020-845	013	BAYER CLI		068	069	068	069	069 - POS.(3+))	36	30	36	100	
2020-845	067	URISCAN		068	069	068	069	068 - POS.(1-2+)	50	10	50	100	
2020-845	067	URISCAN		068	069	068	069	069 - POS.(3+))	50	40	50	100	
2020-845	071	ATLAS		068	069	068	069	068 - POS.(1-2+)	18	6	18	100	
2020-845	071	ATLAS		068	069	068	069	069 - POS.(3+))	18	12	18	100	
2020-845	072	URS-9-TECO		068	069	068	069	066 - NEGATIVE	105	2	102	0	
2020-845	072	URS-9-TECO		068	069	068	069	067 - TRACES	105	1	102	0	
2020-845	072	URS-9-TECO		068	069	068	069	068 - POS.(1-2+)	105	26	102	100	
2020-845	072	URS-9-TECO		068	069	068	069	069 - POS.(3+))	105	76	102	100	
2020-845	080	ROCHE (CHE		068	069	068	069	068 - POS.(1-2+)	37	1	37	100	
2020-845	080	ROCHE (CHE		068	069	068	069	069 - POS.(3+))	37	36	37	100	
2020-845	089	AIMSTRIP		068	069	068	069	068 - POS.(1-2+)	113	8	113	100	
2020-845	089	AIMSTRIP		068	069	068	069	069 - POS.(3+))	113	105	113	100	
2020-845	094	CLINITEK		068	069	068	069	067 - TRACES	104	1	103	0	
2020-845	094	CLINITEK		068	069	068	069	068 - POS.(1-2+)	104	20	103	100	
2020-845	094	CLINITEK		068	069	068	069	069 - POS.(3+))	104	83	103	100	
2020-845	102	ACON LAB		068	069	068	069	068 - POS.(1-2+)	171	6	171	100	
2020-845	102	ACON LAB		068	069	068	069	069 - POS.(3+))	171	165	171	100	
2020-845	131	iCHEM VELO		067	069	068	069	067 - TRACES	27	1	27	100	
2020-845	131	iCHEM VELO		067	069	068	069	068 - POS.(1-2+)	27	24	27	100	
2020-845	131	iCHEM VELO		067	069	068	069	069 - POS.(3+))	27	2	27	100	
2020-845	135	Hybrido AU		067	069	068	069	068 - POS.(1-2+)	13	13	13	100	
2020-845	138	URS-10T		068	069	068	069	068 - POS.(1-2+)	11	3	11	100	
2020-845	138	URS-10T		068	069	068	069	069 - POS.(3+))	11	8	11	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	140	MISSION A		068 069	068 069		068 - POS.(1-2+)	66	2	66	100	
2020-845	140	MISSION A		068 069	068 069		069 - POS.(3+)	66	64	66	100	
2020-845	145	CLINITEK		068 069	068 069		068 - POS.(1-2+)	10	4	10	100	
2020-845	145	CLINITEK		068 069	068 069		069 - POS.(3+)	10	6	10	100	
2020-845	147	Mission120		068 069	068 069		069 - POS.(3+)	17	17	17	100	
2020-845	148	Mission500		068 069	068 069		069 - POS.(3+)	19	19	19	100	
060 URINALYSIS-KETONES												
2020-841	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	49	49	49	100	
2020-841	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	36	36	36	100	
2020-841	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100	
2020-841	071	ATLAS		090 090	090 090		090 - NEGATIVE	18	18	18	100	
2020-841	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	105	105	105	100	
2020-841	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	37	37	37	100	
2020-841	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	113	113	113	100	
2020-841	094	CLINITEK		090 090	090 090		090 - NEGATIVE	104	104	104	100	
2020-841	102	ACON LAB		090 090	090 090		090 - NEGATIVE	170	169	169	100	
2020-841	102	ACON LAB		090 090	090 090		093 - MOD. (2+)	170	1	169	0	
2020-841	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100	
2020-841	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	13	13	13	100	
2020-841	138	URS-10T		090 090	090 090		090 - NEGATIVE	11	11	11	100	
2020-841	140	MISSION A		090 090	090 090		090 - NEGATIVE	66	66	66	100	
2020-841	145	CLINITEK		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-841	147	Mission120		090 090	090 090		090 - NEGATIVE	17	17	17	100	
2020-841	148	Mission500		090 090	090 090		090 - NEGATIVE	20	20	20	100	
2020-842	011	AMES NM10S		092 094	092 094		091 - TRACE	49	1	48	0	
2020-842	011	AMES NM10S		092 094	092 094		092 - SMALL (1+)	49	8	48	100	
2020-842	011	AMES NM10S		092 094	092 094		093 - MOD. (2+)	49	34	48	100	
2020-842	011	AMES NM10S		092 094	092 094		094 - LARGE-3-4+	49	6	48	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-842	013	BAYER CLI		092 094	092 094		091 - TRACE	36	1	35	0
2020-842	013	BAYER CLI		092 094	092 094		092 - SMALL (1+)	36	6	35	100
2020-842	013	BAYER CLI		092 094	092 094		093 - MOD. (2+)	36	26	35	100
2020-842	013	BAYER CLI		092 094	092 094		094 - LARGE-3-4+	36	3	35	100
2020-842	067	URISCAN		092 094	092 094		092 - SMALL (1+)	50	7	50	100
2020-842	067	URISCAN		092 094	092 094		093 - MOD. (2+)	50	37	50	100
2020-842	067	URISCAN		092 094	092 094		094 - LARGE-3-4+	50	6	50	100
2020-842	071	ATLAS		092 094	092 094		092 - SMALL (1+)	18	1	18	100
2020-842	071	ATLAS		092 094	092 094		093 - MOD. (2+)	18	17	18	100
2020-842	072	URS-9-TECO		092 094	092 094		092 - SMALL (1+)	105	11	105	100
2020-842	072	URS-9-TECO		092 094	092 094		093 - MOD. (2+)	105	60	105	100
2020-842	072	URS-9-TECO		092 094	092 094		094 - LARGE-3-4+	105	34	105	100
2020-842	080	ROCHE (CHE		092 094	092 094		092 - SMALL (1+)	37	10	37	100
2020-842	080	ROCHE (CHE		092 094	092 094		093 - MOD. (2+)	37	19	37	100
2020-842	080	ROCHE (CHE		092 094	092 094		094 - LARGE-3-4+	37	8	37	100
2020-842	089	AIMSTRIP		092 094	092 094		091 - TRACE	113	1	112	0
2020-842	089	AIMSTRIP		092 094	092 094		092 - SMALL (1+)	113	13	112	100
2020-842	089	AIMSTRIP		092 094	092 094		093 - MOD. (2+)	113	65	112	100
2020-842	089	AIMSTRIP		092 094	092 094		094 - LARGE-3-4+	113	34	112	100
2020-842	094	CLINITEK		092 094	092 094		092 - SMALL (1+)	104	22	104	100
2020-842	094	CLINITEK		092 094	092 094		093 - MOD. (2+)	104	80	104	100
2020-842	094	CLINITEK		092 094	092 094		094 - LARGE-3-4+	104	2	104	100
2020-842	102	ACON LAB		092 094	092 094		090 - NEGATIVE	170	1	169	0
2020-842	102	ACON LAB		092 094	092 094		092 - SMALL (1+)	170	9	169	100
2020-842	102	ACON LAB		092 094	092 094		093 - MOD. (2+)	170	101	169	100
2020-842	102	ACON LAB		092 094	092 094		094 - LARGE-3-4+	170	59	169	100
2020-842	131	iCHEM VELO		091 093	092 094		092 - SMALL (1+)	27	21	27	100
2020-842	131	iCHEM VELO		091 093	092 094		093 - MOD. (2+)	27	5	27	100
2020-842	131	iCHEM VELO		091 093	092 094		094 - LARGE-3-4+	27	1	27	100
2020-842	135	Hybrido AU		092 094	092 094		092 - SMALL (1+)	13	1	13	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-842	135	Hybrido AU		092 094	092 094		093 - MOD. (2+)	13	12	13	100	
2020-842	138	URS-10T		092 094	092 094		093 - MOD. (2+)	11	9	11	100	
2020-842	138	URS-10T		092 094	092 094		094 - LARGE-3-4+	11	2	11	100	
2020-842	140	MISSION A		093 094	092 094		092 - SMALL (1+)	66	4	66	100	
2020-842	140	MISSION A		093 094	092 094		093 - MOD. (2+)	66	19	66	100	
2020-842	140	MISSION A		093 094	092 094		094 - LARGE-3-4+	66	43	66	100	
2020-842	145	CLINITEK		092 094	092 094		092 - SMALL (1+)	10	1	10	100	
2020-842	145	CLINITEK		092 094	092 094		093 - MOD. (2+)	10	9	10	100	
2020-842	147	Mission120		093 094	092 094		094 - LARGE-3-4+	17	17	17	100	
2020-842	148	Mission500		092 094	092 094		092 - SMALL (1+)	20	1	20	100	
2020-842	148	Mission500		092 094	092 094		093 - MOD. (2+)	20	14	20	100	
2020-842	148	Mission500		092 094	092 094		094 - LARGE-3-4+	20	5	20	100	
060 URINALYSIS-KETONES												
2020-843	011	AMES NM10S		090 090	090 090		090 - NEGATIVE	49	49	49	100	
2020-843	013	BAYER CLI		090 090	090 090		090 - NEGATIVE	36	36	36	100	
2020-843	067	URISCAN		090 090	090 090		090 - NEGATIVE	50	50	50	100	
2020-843	071	ATLAS		090 090	090 090		090 - NEGATIVE	18	18	18	100	
2020-843	072	URS-9-TECO		090 090	090 090		090 - NEGATIVE	105	105	105	100	
2020-843	080	ROCHE (CHE		090 090	090 090		090 - NEGATIVE	37	37	37	100	
2020-843	089	AIMSTRIP		090 090	090 090		090 - NEGATIVE	113	113	113	100	
2020-843	094	CLINITEK		090 090	090 090		090 - NEGATIVE	104	103	103	100	
2020-843	094	CLINITEK		090 090	090 090		094 - LARGE-3-4+	104	1	103	0	
2020-843	102	ACON LAB		090 090	090 090		090 - NEGATIVE	170	170	170	100	
2020-843	131	iCHEM VELO		090 090	090 090		090 - NEGATIVE	27	27	27	100	
2020-843	135	Hybrido AU		090 090	090 090		090 - NEGATIVE	13	13	13	100	
2020-843	138	URS-10T		090 090	090 090		090 - NEGATIVE	11	11	11	100	
2020-843	140	MISSION A		090 090	090 090		090 - NEGATIVE	66	66	66	100	
2020-843	145	CLINITEK		090 090	090 090		090 - NEGATIVE	10	10	10	100	
2020-843	147	Mission120		090 090	090 090		090 - NEGATIVE	17	17	17	100	
2020-843	148	Mission500		090 090	090 090		090 - NEGATIVE	20	20	20	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
060 URINALYSIS-KETONES													
2020-844	011 AMES NM10S			090 090	090 090		090 - NEGATIVE	49	49	49	100		
2020-844	013 BAYER CLI			090 090	090 090		090 - NEGATIVE	36	36	36	100		
2020-844	067 URISCAN			090 090	090 090		090 - NEGATIVE	50	49	49	100		
2020-844	067 URISCAN			090 090	090 090		093 - MOD. (2+)	50	1	49	0		
2020-844	071 ATLAS			090 090	090 090		090 - NEGATIVE	18	18	18	100		
2020-844	072 URS-9-TECO			090 090	090 090		090 - NEGATIVE	105	105	105	100		
2020-844	080 ROCHE (CHE			090 090	090 090		090 - NEGATIVE	37	37	37	100		
2020-844	089 AIMSTRIP			090 090	090 090		090 - NEGATIVE	113	112	112	100		
2020-844	089 AIMSTRIP			090 090	090 090		094 - LARGE-3-4+	113	1	112	0		
2020-844	094 CLINITEK			090 090	090 090		090 - NEGATIVE	104	103	103	100		
2020-844	094 CLINITEK			090 090	090 090		094 - LARGE-3-4+	104	1	103	0		
2020-844	102 ACON LAB			090 090	090 090		090 - NEGATIVE	170	170	170	100		
2020-844	131 iCHEM VELO			090 090	090 090		090 - NEGATIVE	27	27	27	100		
2020-844	135 Hybrido AU			090 090	090 090		090 - NEGATIVE	13	13	13	100		
2020-844	138 URS-10T			090 090	090 090		090 - NEGATIVE	11	11	11	100		
2020-844	140 MISSION A			090 090	090 090		090 - NEGATIVE	66	66	66	100		
2020-844	145 CLINITEK			090 090	090 090		090 - NEGATIVE	10	10	10	100		
2020-844	147 Mission120			090 090	090 090		090 - NEGATIVE	17	17	17	100		
2020-844	148 Mission500			090 090	090 090		090 - NEGATIVE	20	20	20	100		
2020-845	011 AMES NM10S			093 094	093 094		093 - MOD. (2+)	49	1	49	100		
2020-845	011 AMES NM10S			093 094	093 094		094 - LARGE-3-4+	49	48	49	100		
2020-845	013 BAYER CLI			093 094	093 094		093 - MOD. (2+)	36	1	36	100		
2020-845	013 BAYER CLI			093 094	093 094		094 - LARGE-3-4+	36	35	36	100		
2020-845	067 URISCAN			093 094	093 094		093 - MOD. (2+)	50	16	50	100		
2020-845	067 URISCAN			093 094	093 094		094 - LARGE-3-4+	50	34	50	100		
2020-845	071 ATLAS			093 094	093 094		094 - LARGE-3-4+	18	18	18	100		
2020-845	072 URS-9-TECO			093 094	093 094		090 - NEGATIVE	105	2	103	0		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-845	072	URS-9-TECO		093 094	093 094		093 - MOD. (2+)	105	3	103	100		
2020-845	072	URS-9-TECO		093 094	093 094		094 - LARGE-3-4+	105	100	103	100		
2020-845	080	ROCHE (CHE		093 094	093 094		093 - MOD. (2+)	37	4	37	100		
2020-845	080	ROCHE (CHE		093 094	093 094		094 - LARGE-3-4+	37	33	37	100		
2020-845	089	AIMSTRIP		093 094	093 094		093 - MOD. (2+)	113	1	113	100		
2020-845	089	AIMSTRIP		093 094	093 094		094 - LARGE-3-4+	113	112	113	100		
2020-845	094	CLINITEK		093 094	093 094		090 - NEGATIVE	104	1	103	0		
2020-845	094	CLINITEK		093 094	093 094		093 - MOD. (2+)	104	2	103	100		
2020-845	094	CLINITEK		093 094	093 094		094 - LARGE-3-4+	104	101	103	100		
2020-845	102	ACON LAB		093 094	093 094		094 - LARGE-3-4+	170	170	170	100		
2020-845	131	iCHEM VELO		092 094	093 094		093 - MOD. (2+)	27	15	27	100		
2020-845	131	iCHEM VELO		092 094	093 094		094 - LARGE-3-4+	27	12	27	100		
2020-845	135	Hybrido AU		093 094	093 094		094 - LARGE-3-4+	13	13	13	100		
2020-845	138	URS-10T		093 094	093 094		094 - LARGE-3-4+	11	11	11	100		
2020-845	140	MISSION A		093 094	093 094		093 - MOD. (2+)	66	1	66	100		
2020-845	140	MISSION A		093 094	093 094		094 - LARGE-3-4+	66	65	66	100		
2020-845	145	CLINITEK		093 094	093 094		094 - LARGE-3-4+	10	10	10	100		
2020-845	147	Mission120		093 094	093 094		094 - LARGE-3-4+	17	17	17	100		
2020-845	148	Mission500		093 094	093 094		094 - LARGE-3-4+	20	20	20	100		
061 URINALYSIS-NITRATE													
2020-841	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	49	49	49	100		
2020-841	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100		
2020-841	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100		
2020-841	071	ATLAS		070 070	070 070		070 - NEGATIVE	18	18	18	100		
2020-841	072	URS-9-TECO		070 070	070 070		070 - NEGATIVE	104	104	104	100		
2020-841	080	ROCHE (CHE		070 070	070 070		070 - NEGATIVE	37	37	37	100		
2020-841	089	AIMSTRIP		070 070	070 070		070 - NEGATIVE	113	113	113	100		
2020-841	094	CLINITEK		070 070	070 070		070 - NEGATIVE	104	104	104	100		
2020-841	102	ACON LAB		070 070	070 070		070 - NEGATIVE	170	170	170	100		
2020-841	131	iCHEM VELO		070 070	070 070		070 - NEGATIVE	27	27	27	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-841	135	Hybrido AU		070 070	070 070		070 - NEGATIVE	13	13	13	100	
2020-841	138	URS-10T		070 070	070 070		070 - NEGATIVE	11	11	11	100	
2020-841	140	MISSION A		070 070	070 070		070 - NEGATIVE	66	66	66	100	
2020-841	145	CLINITEK		070 070	070 070		070 - NEGATIVE	10	10	10	100	
2020-841	147	Mission120		070 070	070 070		070 - NEGATIVE	17	17	17	100	
2020-841	148	Mission500		070 070	070 070		070 - NEGATIVE	20	20	20	100	
061 URINALYSIS-NITRATE												
2020-842	011	AMES NM10S		071 072	071 072		072 - POSITIVE	49	49	49	100	
2020-842	013	BAYER CLI		071 072	071 072		072 - POSITIVE	37	37	37	100	
2020-842	067	URISCAN		071 072	071 072		072 - POSITIVE	50	50	50	100	
2020-842	071	ATLAS		071 072	071 072		072 - POSITIVE	18	18	18	100	
2020-842	072	URS-9-TECO		071 072	071 072		071 - WEAK POS.	104	1	104	100	
2020-842	072	URS-9-TECO		071 072	071 072		072 - POSITIVE	104	103	104	100	
2020-842	080	ROCHE (CHE		071 072	071 072		072 - POSITIVE	37	37	37	100	
2020-842	089	AIMSTRIP		071 072	071 072		072 - POSITIVE	113	113	113	100	
2020-842	094	CLINITEK		071 072	071 072		072 - POSITIVE	104	104	104	100	
2020-842	102	ACON LAB		071 072	071 072		072 - POSITIVE	170	170	170	100	
2020-842	131	iCHEM VELO		071 072	071 072		072 - POSITIVE	27	27	27	100	
2020-842	135	Hybrido AU		071 072	071 072		072 - POSITIVE	13	13	13	100	
2020-842	138	URS-10T		071 072	071 072		072 - POSITIVE	11	11	11	100	
2020-842	140	MISSION A		071 072	071 072		072 - POSITIVE	66	66	66	100	
2020-842	145	CLINITEK		071 072	071 072		072 - POSITIVE	10	10	10	100	
2020-842	147	Mission120		071 072	071 072		072 - POSITIVE	17	17	17	100	
2020-842	148	Mission500		071 072	071 072		072 - POSITIVE	20	20	20	100	
2020-843	011	AMES NM10S		070 070	070 070		070 - NEGATIVE	49	49	49	100	
2020-843	013	BAYER CLI		070 070	070 070		070 - NEGATIVE	37	37	37	100	
2020-843	067	URISCAN		070 070	070 070		070 - NEGATIVE	50	50	50	100	
2020-843	071	ATLAS		070 070	070 070		070 - NEGATIVE	18	18	18	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-843	072	URS-9-TECO		070	070	070	070	070 - NEGATIVE	104	104	104	100	
2020-843	080	ROCHE (CHE		070	070	070	070	070 - NEGATIVE	37	37	37	100	
2020-843	089	AIMSTRIP		070	070	070	070	070 - NEGATIVE	113	113	113	100	
2020-843	094	CLINITEK		070	070	070	070	070 - NEGATIVE	104	104	104	100	
2020-843	102	ACON LAB		070	070	070	070	070 - NEGATIVE	170	170	170	100	
2020-843	131	ICHEM VELO		070	070	070	070	070 - NEGATIVE	27	27	27	100	
2020-843	135	Hybrido AU		070	070	070	070	070 - NEGATIVE	13	13	13	100	
2020-843	138	URS-10T		070	070	070	070	070 - NEGATIVE	11	11	11	100	
2020-843	140	MISSION A		070	070	070	070	070 - NEGATIVE	66	66	66	100	
2020-843	145	CLINITEK		070	070	070	070	070 - NEGATIVE	10	10	10	100	
2020-843	147	Mission120		070	070	070	070	070 - NEGATIVE	17	17	17	100	
2020-843	148	Mission500		070	070	070	070	070 - NEGATIVE	20	20	20	100	
061 URINALYSIS-NITRATE													
2020-844	011	AMES NM10S		070	070	070	070	070 - NEGATIVE	49	49	49	100	
2020-844	013	BAYER CLI		070	070	070	070	070 - NEGATIVE	37	37	37	100	
2020-844	067	URISCAN		070	070	070	070	070 - NEGATIVE	50	50	50	100	
2020-844	071	ATLAS		070	070	070	070	070 - NEGATIVE	18	18	18	100	
2020-844	072	URS-9-TECO		070	070	070	070	070 - NEGATIVE	104	104	104	100	
2020-844	080	ROCHE (CHE		070	070	070	070	070 - NEGATIVE	37	37	37	100	
2020-844	089	AIMSTRIP		070	070	070	070	070 - NEGATIVE	113	113	113	100	
2020-844	094	CLINITEK		070	070	070	070	070 - NEGATIVE	104	104	104	100	
2020-844	102	ACON LAB		070	070	070	070	070 - NEGATIVE	170	170	170	100	
2020-844	131	ICHEM VELO		070	070	070	070	070 - NEGATIVE	27	27	27	100	
2020-844	135	Hybrido AU		070	070	070	070	070 - NEGATIVE	13	13	13	100	
2020-844	138	URS-10T		070	070	070	070	070 - NEGATIVE	11	11	11	100	
2020-844	140	MISSION A		070	070	070	070	070 - NEGATIVE	66	66	66	100	
2020-844	145	CLINITEK		070	070	070	070	070 - NEGATIVE	10	10	10	100	
2020-844	147	Mission120		070	070	070	070	070 - NEGATIVE	17	17	17	100	
2020-844	148	Mission500		070	070	070	070	070 - NEGATIVE	20	20	20	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
061 URINALYSIS-NITRATE												
2020-845	011	AMES NM10S		070	070	070	070	070 - NEGATIVE	49	49	49	100
2020-845	013	BAYER CLI		070	070	070	070	070 - NEGATIVE	37	37	37	100
2020-845	067	URISCAN		070	070	070	070	070 - NEGATIVE	50	49	49	100
2020-845	067	URISCAN		070	070	070	070	072 - POSITIVE	50	1	49	0
2020-845	071	ATLAS		070	070	070	070	070 - NEGATIVE	18	18	18	100
2020-845	072	URS-9-TECO		070	070	070	070	070 - NEGATIVE	104	103	103	100
2020-845	072	URS-9-TECO		070	070	070	070	072 - POSITIVE	104	1	103	0
2020-845	080	ROCHE (CHE		070	070	070	070	070 - NEGATIVE	37	37	37	100
2020-845	089	AIMSTRIP		070	070	070	070	070 - NEGATIVE	113	112	112	100
2020-845	089	AIMSTRIP		070	070	070	070	072 - POSITIVE	113	1	112	0
2020-845	094	CLINITEK		070	070	070	070	070 - NEGATIVE	104	103	103	100
2020-845	094	CLINITEK		070	070	070	070	072 - POSITIVE	104	1	103	0
2020-845	102	ACON LAB		070	070	070	070	070 - NEGATIVE	170	170	170	100
2020-845	131	iCHEM VELO		070	070	070	070	070 - NEGATIVE	27	27	27	100
2020-845	135	Hybrido AU		070	070	070	070	070 - NEGATIVE	13	13	13	100
2020-845	138	URS-10T		070	070	070	070	070 - NEGATIVE	11	11	11	100
2020-845	140	MISSION A		070	070	070	070	070 - NEGATIVE	66	64	64	100
2020-845	140	MISSION A		070	070	070	070	072 - POSITIVE	66	2	64	0
2020-845	145	CLINITEK		070	070	070	070	070 - NEGATIVE	10	10	10	100
2020-845	147	Mission120		070	070	070	070	070 - NEGATIVE	17	17	17	100
2020-845	148	Mission500		070	070	070	070	070 - NEGATIVE	20	20	20	100
062 URINALYSIS-pH												
2020-841	011	AMES NM10S		082	086	082	086	083 - 5.5	49	2	49	100
2020-841	011	AMES NM10S		082	086	082	086	084 - 6.0	49	45	49	100
2020-841	011	AMES NM10S		082	086	082	086	085 - 6.5	49	2	49	100
2020-841	013	BAYER CLI		082	086	082	086	083 - 5.5	37	3	37	100
2020-841	013	BAYER CLI		082	086	082	086	084 - 6.0	37	33	37	100
2020-841	013	BAYER CLI		082	086	082	086	085 - 6.5	37	1	37	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-841	067	URISCAN		081 085	082 086		082 - 5.0	50	8	50	100
2020-841	067	URISCAN		081 085	082 086		083 - 5.5	50	34	50	100
2020-841	067	URISCAN		081 085	082 086		084 - 6.0	50	8	50	100
2020-841	071	ATLAS		082 086	082 086		084 - 6.0	18	18	18	100
2020-841	072	URS-9-TECO		082 086	082 086		082 - 5.0	104	1	104	100
2020-841	072	URS-9-TECO		082 086	082 086		083 - 5.5	104	2	104	100
2020-841	072	URS-9-TECO		082 086	082 086		084 - 6.0	104	57	104	100
2020-841	072	URS-9-TECO		082 086	082 086		085 - 6.5	104	44	104	100
2020-841	080	ROCHE (CHE		082 086	082 086		084 - 6.0	37	37	37	100
2020-841	089	AIMSTRIP		082 086	082 086		082 - 5.0	113	7	113	100
2020-841	089	AIMSTRIP		082 086	082 086		084 - 6.0	113	103	113	100
2020-841	089	AIMSTRIP		082 086	082 086		085 - 6.5	113	3	113	100
2020-841	094	CLINITEK		082 086	082 086		083 - 5.5	103	3	103	100
2020-841	094	CLINITEK		082 086	082 086		084 - 6.0	103	100	103	100
2020-841	102	ACON LAB		082 086	082 086		083 - 5.5	170	4	170	100
2020-841	102	ACON LAB		082 086	082 086		084 - 6.0	170	162	170	100
2020-841	102	ACON LAB		082 086	082 086		085 - 6.5	170	3	170	100
2020-841	102	ACON LAB		082 086	082 086		082 - 5.0	170	1	170	100
2020-841	131	iCHEM VELO		082 086	082 086		082 - 5.0	27	1	27	100
2020-841	131	iCHEM VELO		082 086	082 086		084 - 6.0	27	26	27	100
2020-841	135	Hybrido AU		082 086	082 086		084 - 6.0	13	13	13	100
2020-841	138	URS-10T		082 086	082 086		084 - 6.0	11	9	11	100
2020-841	138	URS-10T		082 086	082 086		085 - 6.5	11	2	11	100
2020-841	140	MISSION A		082 086	082 086		084 - 6.0	66	64	66	100
2020-841	140	MISSION A		082 086	082 086		083 - 5.5	66	2	66	100
2020-841	145	CLINITEK		082 086	082 086		084 - 6.0	11	11	11	100
2020-841	147	Mission120		082 086	082 086		083 - 5.5	17	2	17	100
2020-841	147	Mission120		082 086	082 086		084 - 6.0	17	15	17	100
2020-841	148	Mission500		082 086	082 086		084 - 6.0	20	20	20	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
062 URINALYSIS-pH											
2020-842	011	AMES NM10S		080 084	080 084		082 - 5.0	49	24	49	100
2020-842	011	AMES NM10S		080 084	080 084		083 - 5.5	49	17	49	100
2020-842	011	AMES NM10S		080 084	080 084		084 - 6.0	49	8	49	100
2020-842	013	BAYER CLI		080 084	080 084		082 - 5.0	37	19	37	100
2020-842	013	BAYER CLI		080 084	080 084		083 - 5.5	37	15	37	100
2020-842	013	BAYER CLI		080 084	080 084		084 - 6.0	37	3	37	100
2020-842	067	URISCAN		080 084	080 084		082 - 5.0	50	43	50	100
2020-842	067	URISCAN		080 084	080 084		083 - 5.5	50	6	50	100
2020-842	067	URISCAN		080 084	080 084		084 - 6.0	50	1	50	100
2020-842	071	ATLAS		080 084	080 084		082 - 5.0	18	14	18	100
2020-842	071	ATLAS		080 084	080 084		083 - 5.5	18	4	18	100
2020-842	072	URS-9-TECO		082 086	080 084		082 - 5.0	104	16	104	100
2020-842	072	URS-9-TECO		082 086	080 084		083 - 5.5	104	19	104	100
2020-842	072	URS-9-TECO		082 086	080 084		084 - 6.0	104	60	104	100
2020-842	072	URS-9-TECO		082 086	080 084		085 - 6.5	104	9	104	100
2020-842	080	ROCHE (CHE		080 084	080 084		082 - 5.0	37	37	37	100
2020-842	089	AIMSTRIP		080 084	080 084		082 - 5.0	113	64	113	100
2020-842	089	AIMSTRIP		080 084	080 084		083 - 5.5	113	32	113	100
2020-842	089	AIMSTRIP		080 084	080 084		084 - 6.0	113	17	113	100
2020-842	094	CLINITEK		081 085	080 084		082 - 5.0	103	26	103	100
2020-842	094	CLINITEK		081 085	080 084		083 - 5.5	103	67	103	100
2020-842	094	CLINITEK		081 085	080 084		084 - 6.0	103	10	103	100
2020-842	102	ACON LAB		082 086	080 084		082 - 5.0	170	53	170	100
2020-842	102	ACON LAB		082 086	080 084		083 - 5.5	170	39	170	100
2020-842	102	ACON LAB		082 086	080 084		084 - 6.0	170	77	170	100
2020-842	102	ACON LAB		082 086	080 084		085 - 6.5	170	1	170	100
2020-842	131	iCHEM VELO		080 084	080 084		082 - 5.0	27	27	27	100
2020-842	135	Hybrido AU		080 084	080 084		082 - 5.0	13	11	13	100
2020-842	135	Hybrido AU		080 084	080 084		083 - 5.5	13	2	13	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-842	138	URS-10T		082 086	080 084		082 - 5.0	11	2	11	100	
2020-842	138	URS-10T		082 086	080 084		083 - 5.5	11	4	11	100	
2020-842	138	URS-10T		082 086	080 084		084 - 6.0	11	5	11	100	
2020-842	140	MISSION A		081 085	080 084		082 - 5.0	66	19	66	100	
2020-842	140	MISSION A		081 085	080 084		083 - 5.5	66	30	66	100	
2020-842	140	MISSION A		081 085	080 084		084 - 6.0	66	17	66	100	
2020-842	145	CLINITEK		080 084	080 084		082 - 5.0	11	10	11	100	
2020-842	145	CLINITEK		080 084	080 084		083 - 5.5	11	1	11	100	
2020-842	147	Mission120		081 085	080 084		082 - 5.0	17	7	17	100	
2020-842	147	Mission120		081 085	080 084		083 - 5.5	17	10	17	100	
2020-842	148	Mission500		082 086	080 084		082 - 5.0	20	1	20	100	
2020-842	148	Mission500		082 086	080 084		083 - 5.5	20	5	20	100	
2020-842	148	Mission500		082 086	080 084		084 - 6.0	20	14	20	100	
062 URINALYSIS-pH												
2020-843	011	AMES NM10S		086 088	086 088		086 - 7.0	49	2	49	100	
2020-843	011	AMES NM10S		086 088	086 088		087 - 7.5	49	4	49	100	
2020-843	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	49	43	49	100	
2020-843	013	BAYER CLI		086 088	086 088		086 - 7.0	37	1	37	100	
2020-843	013	BAYER CLI		086 088	086 088		087 - 7.5	37	6	37	100	
2020-843	013	BAYER CLI		086 088	086 088		088 - 8.0OR MORE	37	30	37	100	
2020-843	067	URISCAN		085 088	086 088		086 - 7.0	50	2	50	100	
2020-843	067	URISCAN		085 088	086 088		087 - 7.5	50	32	50	100	
2020-843	067	URISCAN		085 088	086 088		088 - 8.0OR MORE	50	16	50	100	
2020-843	071	ATLAS		085 088	086 088		087 - 7.5	18	13	18	100	
2020-843	071	ATLAS		085 088	086 088		088 - 8.0OR MORE	18	5	18	100	
2020-843	072	URS-9-TECO		085 088	086 088		086 - 7.0	104	11	104	100	
2020-843	072	URS-9-TECO		085 088	086 088		087 - 7.5	104	62	104	100	
2020-843	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	104	31	104	100	
2020-843	080	ROCHE (CHE		086 088	086 088		086 - 7.0	37	6	37	100	
2020-843	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	37	31	37	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-843	089	AIMSTRIP		086 088	086 088		087 - 7.5	113	19	113	100	
2020-843	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	113	94	113	100	
2020-843	094	CLINITEK		086 088	086 088		086 - 7.0	103	1	103	100	
2020-843	094	CLINITEK		086 088	086 088		087 - 7.5	103	14	103	100	
2020-843	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	103	88	103	100	
2020-843	102	ACON LAB		086 088	086 088		087 - 7.5	170	23	170	100	
2020-843	102	ACON LAB		086 088	086 088		088 - 8.0OR MORE	170	147	170	100	
2020-843	131	iCHEM VELO		086 088	086 088		086 - 7.0	27	5	27	100	
2020-843	131	iCHEM VELO		086 088	086 088		088 - 8.0OR MORE	27	22	27	100	
2020-843	135	Hybrido AU		086 088	086 088		088 - 8.0OR MORE	13	13	13	100	
2020-843	138	URS-10T		085 088	086 088		086 - 7.0	11	1	11	100	
2020-843	138	URS-10T		085 088	086 088		087 - 7.5	11	8	11	100	
2020-843	138	URS-10T		085 088	086 088		088 - 8.0OR MORE	11	2	11	100	
2020-843	140	MISSION A		086 088	086 088		087 - 7.5	66	10	66	100	
2020-843	140	MISSION A		086 088	086 088		088 - 8.0OR MORE	66	56	66	100	
2020-843	145	CLINITEK		086 088	086 088		088 - 8.0OR MORE	11	11	11	100	
2020-843	147	Mission120		086 088	086 088		087 - 7.5	17	1	17	100	
2020-843	147	Mission120		086 088	086 088		088 - 8.0OR MORE	17	16	17	100	
2020-843	148	Mission500		086 088	086 088		087 - 7.5	20	1	20	100	
2020-843	148	Mission500		086 088	086 088		088 - 8.0OR MORE	20	19	20	100	
062 URINALYSIS-pH												
2020-844	011	AMES NM10S		085 088	084 088		086 - 7.0	49	9	49	100	
2020-844	011	AMES NM10S		085 088	084 088		087 - 7.5	49	40	49	100	
2020-844	013	BAYER CLI		085 088	084 088		086 - 7.0	37	10	37	100	
2020-844	013	BAYER CLI		085 088	084 088		087 - 7.5	37	27	37	100	
2020-844	067	URISCAN		084 088	084 088		085 - 6.5	50	3	50	100	
2020-844	067	URISCAN		084 088	084 088		086 - 7.0	50	43	50	100	
2020-844	067	URISCAN		084 088	084 088		087 - 7.5	50	4	50	100	
2020-844	071	ATLAS		084 088	084 088		086 - 7.0	18	16	18	100	
2020-844	071	ATLAS		084 088	084 088		087 - 7.5	18	2	18	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-844	072	URS-9-TECO		084 088	084 088		086 - 7.0	104	95	104	100
2020-844	072	URS-9-TECO		084 088	084 088		087 - 7.5	104	8	104	100
2020-844	072	URS-9-TECO		084 088	084 088		084 - 6.0	104	1	104	100
2020-844	080	ROCHE (CHE		084 088	084 088		086 - 7.0	37	37	37	100
2020-844	089	AIMSTRIP		084 088	084 088		084 - 6.0	113	1	113	100
2020-844	089	AIMSTRIP		084 088	084 088		086 - 7.0	113	66	113	100
2020-844	089	AIMSTRIP		084 088	084 088		087 - 7.5	113	45	113	100
2020-844	089	AIMSTRIP		084 088	084 088		088 - 8.0OR MORE	113	1	113	100
2020-844	094	CLINITEK		085 088	084 088		086 - 7.0	103	17	103	100
2020-844	094	CLINITEK		085 088	084 088		087 - 7.5	103	85	103	100
2020-844	094	CLINITEK		085 088	084 088		088 - 8.0OR MORE	103	1	103	100
2020-844	102	ACON LAB		084 088	084 088		085 - 6.5	170	1	170	100
2020-844	102	ACON LAB		084 088	084 088		086 - 7.0	170	134	170	100
2020-844	102	ACON LAB		084 088	084 088		087 - 7.5	170	35	170	100
2020-844	131	iCHEM VELO		084 088	084 088		086 - 7.0	27	27	27	100
2020-844	135	Hybrido AU		085 088	084 088		086 - 7.0	13	1	13	100
2020-844	135	Hybrido AU		085 088	084 088		087 - 7.5	13	12	13	100
2020-844	138	URS-10T		084 088	084 088		085 - 6.5	11	1	11	100
2020-844	138	URS-10T		084 088	084 088		086 - 7.0	11	9	11	100
2020-844	138	URS-10T		084 088	084 088		087 - 7.5	11	1	11	100
2020-844	140	MISSION A		084 088	084 088		086 - 7.0	66	57	66	100
2020-844	140	MISSION A		084 088	084 088		087 - 7.5	66	9	66	100
2020-844	145	CLINITEK		084 088	084 088		086 - 7.0	11	6	11	100
2020-844	145	CLINITEK		084 088	084 088		087 - 7.5	11	5	11	100
2020-844	147	Mission120		084 088	084 088		086 - 7.0	17	16	17	100
2020-844	147	Mission120		084 088	084 088		087 - 7.5	17	1	17	100
2020-844	148	Mission500		084 088	084 088		086 - 7.0	20	20	20	100
062 URINALYSIS-pH											
2020-845	011	AMES NM10S		086 088	086 088		087 - 7.5	49	9	49	100
2020-845	011	AMES NM10S		086 088	086 088		088 - 8.0OR MORE	49	40	49	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	013	BAYER CLI		086 088	086 088		087 - 7.5	37	5	37	100	
2020-845	013	BAYER CLI		086 088	086 088		088 - 8.0OR MORE	37	32	37	100	
2020-845	067	URISCAN		085 088	086 088		086 - 7.0	50	1	50	100	
2020-845	067	URISCAN		085 088	086 088		087 - 7.5	50	39	50	100	
2020-845	067	URISCAN		085 088	086 088		088 - 8.0OR MORE	50	10	50	100	
2020-845	071	ATLAS		085 088	086 088		087 - 7.5	18	16	18	100	
2020-845	071	ATLAS		085 088	086 088		088 - 8.0OR MORE	18	2	18	100	
2020-845	072	URS-9-TECO		085 088	086 088		086 - 7.0	104	20	104	100	
2020-845	072	URS-9-TECO		085 088	086 088		087 - 7.5	104	66	104	100	
2020-845	072	URS-9-TECO		085 088	086 088		088 - 8.0OR MORE	104	18	104	100	
2020-845	080	ROCHE (CHE		086 088	086 088		086 - 7.0	37	4	37	100	
2020-845	080	ROCHE (CHE		086 088	086 088		088 - 8.0OR MORE	37	33	37	100	
2020-845	089	AIMSTRIP		086 088	086 088		082 - 5.0	113	1	112	0	
2020-845	089	AIMSTRIP		086 088	086 088		087 - 7.5	113	18	112	100	
2020-845	089	AIMSTRIP		086 088	086 088		088 - 8.0OR MORE	113	94	112	100	
2020-845	094	CLINITEK		086 088	086 088		086 - 7.0	103	1	103	100	
2020-845	094	CLINITEK		086 088	086 088		087 - 7.5	103	18	103	100	
2020-845	094	CLINITEK		086 088	086 088		088 - 8.0OR MORE	103	84	103	100	
2020-845	102	ACON LAB		086 088	086 088		087 - 7.5	170	23	170	100	
2020-845	102	ACON LAB		086 088	086 088		088 - 8.0OR MORE	170	147	170	100	
2020-845	131	iCHEM VELO		084 088	086 088		086 - 7.0	27	21	27	100	
2020-845	131	iCHEM VELO		084 088	086 088		088 - 8.0OR MORE	27	6	27	100	
2020-845	135	Hybrido AU		086 088	086 088		088 - 8.0OR MORE	13	13	13	100	
2020-845	138	URS-10T		085 088	086 088		086 - 7.0	11	1	11	100	
2020-845	138	URS-10T		085 088	086 088		087 - 7.5	11	8	11	100	
2020-845	138	URS-10T		085 088	086 088		088 - 8.0OR MORE	11	2	11	100	
2020-845	140	MISSION A		086 088	086 088		087 - 7.5	66	6	66	100	
2020-845	140	MISSION A		086 088	086 088		088 - 8.0OR MORE	66	60	66	100	
2020-845	145	CLINITEK		086 088	086 088		087 - 7.5	11	2	11	100	
2020-845	145	CLINITEK		086 088	086 088		088 - 8.0OR MORE	11	9	11	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	147	Mission120		086 088	086 088		087 - 7.5	17	3	17	100	
2020-845	147	Mission120		086 088	086 088		088 - 8.0OR MORE	17	14	17	100	
2020-845	148	Mission500		086 088	086 088		087 - 7.5	20	3	20	100	
2020-845	148	Mission500		086 088	086 088		088 - 8.0OR MORE	20	17	20	100	
064 URIN-PROTEIN QUALI.												
2020-841	011	AMES NM10S		053 055	053 055		054 - 2+	50	41	50	100	
2020-841	011	AMES NM10S		053 055	053 055		053 - 1+	50	3	50	100	
2020-841	011	AMES NM10S		053 055	053 055		055 - 3+	50	6	50	100	
2020-841	013	BAYER CLI		053 055	053 055		054 - 2+	36	29	36	100	
2020-841	013	BAYER CLI		053 055	053 055		053 - 1+	36	3	36	100	
2020-841	013	BAYER CLI		053 055	053 055		055 - 3+	36	4	36	100	
2020-841	067	URISCAN		053 055	053 055		054 - 2+	50	33	50	100	
2020-841	067	URISCAN		053 055	053 055		053 - 1+	50	16	50	100	
2020-841	067	URISCAN		053 055	053 055		055 - 3+	50	1	50	100	
2020-841	071	ATLAS		053 055	053 055		054 - 2+	18	16	18	100	
2020-841	071	ATLAS		053 055	053 055		053 - 1+	18	2	18	100	
2020-841	072	URS-9-TECO		053 055	053 055		054 - 2+	105	76	104	100	
2020-841	072	URS-9-TECO		053 055	053 055		053 - 1+	105	9	104	100	
2020-841	072	URS-9-TECO		053 055	053 055		055 - 3+	105	19	104	100	
2020-841	072	URS-9-TECO		053 055	053 055		056 - 4+	105	1	104	0	
2020-841	080	ROCHE (CHE		053 055	053 055		054 - 2+	37	27	37	100	
2020-841	080	ROCHE (CHE		053 055	053 055		053 - 1+	37	1	37	100	
2020-841	080	ROCHE (CHE		053 055	053 055		055 - 3+	37	9	37	100	
2020-841	089	AIMSTRIP		053 055	053 055		054 - 2+	112	91	112	100	
2020-841	089	AIMSTRIP		053 055	053 055		053 - 1+	112	10	112	100	
2020-841	089	AIMSTRIP		053 055	053 055		055 - 3+	112	11	112	100	
2020-841	094	CLINITEK		053 055	053 055		054 - 2+	103	88	102	100	
2020-841	094	CLINITEK		053 055	053 055		053 - 1+	103	10	102	100	
2020-841	094	CLINITEK		053 055	053 055		055 - 3+	103	4	102	100	
2020-841	094	CLINITEK		053 055	053 055		056 - 4+	103	1	102	0	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-841	102	ACON LAB		053 055	053 055		054 - 2+	169	143	169	100		
2020-841	102	ACON LAB		053 055	053 055		053 - 1+	169	14	169	100		
2020-841	102	ACON LAB		053 055	053 055		055 - 3+	169	12	169	100		
2020-841	131	iCHEM VELO		053 055	053 055		054 - 2+	27	25	27	100		
2020-841	131	iCHEM VELO		053 055	053 055		053 - 1+	27	1	27	100		
2020-841	131	iCHEM VELO		053 055	053 055		055 - 3+	27	1	27	100		
2020-841	135	Hybrido AU		053 055	053 055		054 - 2+	13	12	13	100		
2020-841	135	Hybrido AU		053 055	053 055		055 - 3+	13	1	13	100		
2020-841	138	URS-10T		053 055	053 055		054 - 2+	11	6	11	100		
2020-841	138	URS-10T		053 055	053 055		053 - 1+	11	1	11	100		
2020-841	138	URS-10T		053 055	053 055		055 - 3+	11	4	11	100		
2020-841	140	MISSION A		053 055	053 055		054 - 2+	66	55	66	100		
2020-841	140	MISSION A		053 055	053 055		053 - 1+	66	7	66	100		
2020-841	140	MISSION A		053 055	053 055		055 - 3+	66	4	66	100		
2020-841	145	CLINITEK		053 055	053 055		054 - 2+	11	10	11	100		
2020-841	145	CLINITEK		053 055	053 055		053 - 1+	11	1	11	100		
2020-841	147	Mission120		053 055	053 055		054 - 2+	17	12	16	100		
2020-841	147	Mission120		053 055	053 055		052 - TRACES	17	1	16	0		
2020-841	147	Mission120		053 055	053 055		053 - 1+	17	4	16	100		
2020-841	148	Mission500		053 055	053 055		054 - 2+	20	16	20	100		
2020-841	148	Mission500		053 055	053 055		053 - 1+	20	3	20	100		
2020-841	148	Mission500		053 055	053 055		055 - 3+	20	1	20	100		
064 URIN-PROTEIN QUALI.													
2020-842	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	50	50	50	100		
2020-842	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	36	36	36	100		
2020-842	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	49	49	100		
2020-842	067	URISCAN		051 051	051 051		052 - TRACES	50	1	49	0		
2020-842	071	ATLAS		051 051	051 051		051 - NEGATIVE	18	18	18	100		
2020-842	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	105	105	105	100		
2020-842	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	37	37	37	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-842	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	112	111	111	100
2020-842	089	AIMSTRIP		051 051	051 051		052 - TRACES	112	1	111	0
2020-842	094	CLINITEK		051 051	051 051		051 - NEGATIVE	103	103	103	100
2020-842	102	ACON LAB		051 051	051 051		051 - NEGATIVE	169	169	169	100
2020-842	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	27	27	27	100
2020-842	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	13	13	13	100
2020-842	138	URS-10T		051 051	051 051		051 - NEGATIVE	11	11	11	100
2020-842	140	MISSION A		051 051	051 051		051 - NEGATIVE	66	66	66	100
2020-842	145	CLINITEK		051 051	051 051		051 - NEGATIVE	11	11	11	100
2020-842	147	Mission120		051 051	051 051		051 - NEGATIVE	17	17	17	100
2020-842	148	Mission500		051 051	051 051		051 - NEGATIVE	20	20	20	100
064 URIN-PROTEIN QUALI.											
2020-843	011	AMES NM10S		051 051	051 051		051 - NEGATIVE	50	50	50	100
2020-843	013	BAYER CLI		051 051	051 051		051 - NEGATIVE	36	36	36	100
2020-843	067	URISCAN		051 051	051 051		051 - NEGATIVE	50	50	50	100
2020-843	071	ATLAS		051 051	051 051		051 - NEGATIVE	18	18	18	100
2020-843	072	URS-9-TECO		051 051	051 051		051 - NEGATIVE	105	104	104	100
2020-843	072	URS-9-TECO		051 051	051 051		052 - TRACES	105	1	104	0
2020-843	080	ROCHE (CHE		051 051	051 051		051 - NEGATIVE	37	37	37	100
2020-843	089	AIMSTRIP		051 051	051 051		051 - NEGATIVE	112	108	108	100
2020-843	089	AIMSTRIP		051 051	051 051		052 - TRACES	112	4	108	0
2020-843	094	CLINITEK		051 051	051 051		051 - NEGATIVE	103	102	102	100
2020-843	094	CLINITEK		051 051	051 051		052 - TRACES	103	1	102	0
2020-843	102	ACON LAB		051 051	051 051		051 - NEGATIVE	169	169	169	100
2020-843	131	iCHEM VELO		051 051	051 051		051 - NEGATIVE	27	27	27	100
2020-843	135	Hybrido AU		051 051	051 051		051 - NEGATIVE	13	13	13	100
2020-843	138	URS-10T		051 051	051 051		051 - NEGATIVE	11	10	10	100
2020-843	138	URS-10T		051 051	051 051		052 - TRACES	11	1	10	0
2020-843	140	MISSION A		051 051	051 051		051 - NEGATIVE	66	66	66	100
2020-843	145	CLINITEK		051 051	051 051		051 - NEGATIVE	11	11	11	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-843	147	Mission120		051 051	051 051		051 - NEGATIVE	17	17	17	100	
2020-843	148	Mission500		051 051	051 051		051 - NEGATIVE	20	19	19	100	
2020-843	148	Mission500		051 051	051 051		052 - TRACES	20	1	19	0	
064 URIN-PROTEIN QUALI.												
2020-844	011	AMES NM10S		053 055	053 055		054 - 2+	50	31	50	100	
2020-844	011	AMES NM10S		053 055	053 055		053 - 1+	50	2	50	100	
2020-844	011	AMES NM10S		053 055	053 055		055 - 3+	50	17	50	100	
2020-844	013	BAYER CLI		053 055	053 055		054 - 2+	36	23	35	100	
2020-844	013	BAYER CLI		053 055	053 055		051 - NEGATIVE	36	1	35	0	
2020-844	013	BAYER CLI		053 055	053 055		055 - 3+	36	12	35	100	
2020-844	067	URISCAN		053 055	053 055		054 - 2+	50	41	49	100	
2020-844	067	URISCAN		053 055	053 055		052 - TRACES	50	1	49	0	
2020-844	067	URISCAN		053 055	053 055		053 - 1+	50	5	49	100	
2020-844	067	URISCAN		053 055	053 055		055 - 3+	50	3	49	100	
2020-844	071	ATLAS		053 055	053 055		054 - 2+	18	15	17	100	
2020-844	071	ATLAS		053 055	053 055		053 - 1+	18	1	17	100	
2020-844	071	ATLAS		053 055	053 055		055 - 3+	18	1	17	100	
2020-844	071	ATLAS		053 055	053 055		056 - 4+	18	1	17	0	
2020-844	072	URS-9-TECO		053 055	053 055		056 - 4+	105	2	103	0	
2020-844	072	URS-9-TECO		053 055	053 055		054 - 2+	105	54	103	100	
2020-844	072	URS-9-TECO		053 055	053 055		053 - 1+	105	2	103	100	
2020-844	072	URS-9-TECO		053 055	053 055		055 - 3+	105	47	103	100	
2020-844	080	ROCHE (CHE		053 055	053 055		054 - 2+	37	25	37	100	
2020-844	080	ROCHE (CHE		053 055	053 055		053 - 1+	37	1	37	100	
2020-844	080	ROCHE (CHE		053 055	053 055		055 - 3+	37	11	37	100	
2020-844	089	AIMSTRIP		053 055	053 055		054 - 2+	112	78	106	100	
2020-844	089	AIMSTRIP		053 055	053 055		051 - NEGATIVE	112	1	106	0	
2020-844	089	AIMSTRIP		053 055	053 055		053 - 1+	112	5	106	100	
2020-844	089	AIMSTRIP		053 055	053 055		055 - 3+	112	23	106	100	
2020-844	089	AIMSTRIP		053 055	053 055		056 - 4+	112	5	106	0	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-844	094	CLINITEK		053 055	053 055		054 - 2+	103	87	102	100		
2020-844	094	CLINITEK		053 055	053 055		053 - 1+	103	7	102	100		
2020-844	094	CLINITEK		053 055	053 055		055 - 3+	103	8	102	100		
2020-844	094	CLINITEK		053 055	053 055		056 - 4+	103	1	102	0		
2020-844	102	ACON LAB		053 055	053 055		054 - 2+	169	118	167	100		
2020-844	102	ACON LAB		053 055	053 055		051 - NEGATIVE	169	1	167	0		
2020-844	102	ACON LAB		053 055	053 055		053 - 1+	169	7	167	100		
2020-844	102	ACON LAB		053 055	053 055		055 - 3+	169	42	167	100		
2020-844	102	ACON LAB		053 055	053 055		056 - 4+	169	1	167	0		
2020-844	131	iCHEM VELO		053 055	053 055		054 - 2+	27	25	27	100		
2020-844	131	iCHEM VELO		053 055	053 055		053 - 1+	27	1	27	100		
2020-844	131	iCHEM VELO		053 055	053 055		055 - 3+	27	1	27	100		
2020-844	135	Hybrido AU		053 055	053 055		054 - 2+	13	12	13	100		
2020-844	135	Hybrido AU		053 055	053 055		055 - 3+	13	1	13	100		
2020-844	138	URS-10T		054 056	053 055		054 - 2+	11	2	11	100		
2020-844	138	URS-10T		054 056	053 055		055 - 3+	11	8	11	100		
2020-844	138	URS-10T		054 056	053 055		056 - 4+	11	1	11	100		
2020-844	140	MISSION A		053 055	053 055		054 - 2+	66	53	65	100		
2020-844	140	MISSION A		053 055	053 055		053 - 1+	66	4	65	100		
2020-844	140	MISSION A		053 055	053 055		055 - 3+	66	8	65	100		
2020-844	140	MISSION A		053 055	053 055		056 - 4+	66	1	65	0		
2020-844	145	CLINITEK		053 055	053 055		054 - 2+	11	7	11	100		
2020-844	145	CLINITEK		053 055	053 055		055 - 3+	11	4	11	100		
2020-844	147	Mission120		053 055	053 055		055 - 3+	17	1	17	100		
2020-844	147	Mission120		053 055	053 055		054 - 2+	17	15	17	100		
2020-844	147	Mission120		053 055	053 055		053 - 1+	17	1	17	100		
2020-844	148	Mission500		053 055	053 055		054 - 2+	20	13	20	100		
2020-844	148	Mission500		053 055	053 055		053 - 1+	20	2	20	100		
2020-844	148	Mission500		053 055	053 055		055 - 3+	20	5	20	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
064 URIN-PROTEIN QUALI.													
2020-845	011	AMES NM10S		053 055	053 055		054 - 2+	50	30	49	100		
2020-845	011	AMES NM10S		053 055	053 055		053 - 1+	50	1	49	100		
2020-845	011	AMES NM10S		053 055	053 055		055 - 3+	50	18	49	100		
2020-845	011	AMES NM10S		053 055	053 055		056 - 4+	50	1	49	0		
2020-845	013	BAYER CLI		053 055	053 055		054 - 2+	36	20	36	100		
2020-845	013	BAYER CLI		053 055	053 055		055 - 3+	36	16	36	100		
2020-845	067	URISCAN		053 055	053 055		054 - 2+	50	41	50	100		
2020-845	067	URISCAN		053 055	053 055		053 - 1+	50	3	50	100		
2020-845	067	URISCAN		053 055	053 055		055 - 3+	50	6	50	100		
2020-845	071	ATLAS		053 055	053 055		054 - 2+	18	15	18	100		
2020-845	071	ATLAS		053 055	053 055		053 - 1+	18	2	18	100		
2020-845	071	ATLAS		053 055	053 055		055 - 3+	18	1	18	100		
2020-845	072	URS-9-TECO		053 055	053 055		054 - 2+	105	54	104	100		
2020-845	072	URS-9-TECO		053 055	053 055		053 - 1+	105	3	104	100		
2020-845	072	URS-9-TECO		053 055	053 055		055 - 3+	105	47	104	100		
2020-845	072	URS-9-TECO		053 055	053 055		056 - 4+	105	1	104	0		
2020-845	080	ROCHE (CHE		053 055	053 055		054 - 2+	37	25	37	100		
2020-845	080	ROCHE (CHE		053 055	053 055		053 - 1+	37	1	37	100		
2020-845	080	ROCHE (CHE		053 055	053 055		055 - 3+	37	11	37	100		
2020-845	089	AIMSTRIP		053 055	053 055		054 - 2+	112	71	106	100		
2020-845	089	AIMSTRIP		053 055	053 055		053 - 1+	112	3	106	100		
2020-845	089	AIMSTRIP		053 055	053 055		055 - 3+	112	32	106	100		
2020-845	089	AIMSTRIP		053 055	053 055		056 - 4+	112	6	106	0		
2020-845	094	CLINITEK		053 055	053 055		054 - 2+	103	82	103	100		
2020-845	094	CLINITEK		053 055	053 055		053 - 1+	103	6	103	100		
2020-845	094	CLINITEK		053 055	053 055		055 - 3+	103	15	103	100		
2020-845	102	ACON LAB		053 055	053 055		054 - 2+	169	106	168	100		
2020-845	102	ACON LAB		053 055	053 055		053 - 1+	169	6	168	100		
2020-845	102	ACON LAB		053 055	053 055		055 - 3+	169	56	168	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	102	ACON LAB		053 055	053 055		056 - 4+	169	1	168	0	
2020-845	131	iCHEM VELO		053 055	053 055		054 - 2+	27	24	27	100	
2020-845	131	iCHEM VELO		053 055	053 055		055 - 3+	27	3	27	100	
2020-845	135	Hybrido AU		053 055	053 055		054 - 2+	13	11	13	100	
2020-845	135	Hybrido AU		053 055	053 055		055 - 3+	13	2	13	100	
2020-845	138	URS-10T		054 056	053 055		054 - 2+	11	2	11	100	
2020-845	138	URS-10T		054 056	053 055		055 - 3+	11	8	11	100	
2020-845	138	URS-10T		054 056	053 055		056 - 4+	11	1	11	100	
2020-845	140	MISSION A		053 055	053 055		054 - 2+	66	51	65	100	
2020-845	140	MISSION A		053 055	053 055		053 - 1+	66	4	65	100	
2020-845	140	MISSION A		053 055	053 055		055 - 3+	66	10	65	100	
2020-845	140	MISSION A		053 055	053 055		056 - 4+	66	1	65	0	
2020-845	145	CLINITEK		054 056	053 055		054 - 2+	11	3	11	100	
2020-845	145	CLINITEK		054 056	053 055		053 - 1+	11	1	11	100	
2020-845	145	CLINITEK		054 056	053 055		055 - 3+	11	7	11	100	
2020-845	147	Mission120		053 055	053 055		054 - 2+	17	15	17	100	
2020-845	147	Mission120		053 055	053 055		053 - 1+	17	1	17	100	
2020-845	147	Mission120		053 055	053 055		055 - 3+	17	1	17	100	
2020-845	148	Mission500		053 055	053 055		054 - 2+	20	12	20	100	
2020-845	148	Mission500		053 055	053 055		053 - 1+	20	1	20	100	
2020-845	148	Mission500		053 055	053 055		055 - 3+	20	7	20	100	
066 URINAL.-UROBILINOGEN												
2020-841	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-841	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2020-841	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-841	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	18	18	18	100	
2020-841	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	105	105	105	100	
2020-841	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2020-841	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	112	112	100	
2020-841	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	104	104	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-841	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	168	168	168	100	
2020-841	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-841	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2020-841	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-841	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	66	66	66	100	
2020-841	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-841	147	Mission120		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
2020-841	148	Mission500		061 061	061 061		061 - NORMAL,NEG	20	20	20	100	
066 URINAL.-UROBILINOGEN												
2020-842	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-842	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2020-842	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-842	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	18	18	18	100	
2020-842	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	105	105	105	100	
2020-842	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2020-842	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	111	111	100	
2020-842	089	AIMSTRIP		061 061	061 061		062 - 2	112	1	111	0	
2020-842	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	104	104	100	
2020-842	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	168	168	168	100	
2020-842	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-842	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2020-842	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-842	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	66	66	66	100	
2020-842	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-842	147	Mission120		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
2020-842	148	Mission500		061 061	061 061		061 - NORMAL,NEG	20	20	20	100	
2020-843	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-843	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-843	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-843	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	18	18	18	100	
2020-843	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	105	105	105	100	
2020-843	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2020-843	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	112	112	100	
2020-843	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	104	104	100	
2020-843	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	168	168	168	100	
2020-843	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-843	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2020-843	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-843	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	66	66	66	100	
2020-843	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-843	147	Mission120		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
2020-843	148	Mission500		061 061	061 061		061 - NORMAL,NEG	20	20	20	100	
066 URINAL.-UROBILINOGEN												
2020-844	011	AMES NM10S		061 061	061 061		061 - NORMAL,NEG	50	50	50	100	
2020-844	013	BAYER CLI		061 061	061 061		061 - NORMAL,NEG	36	36	36	100	
2020-844	067	URISCAN		061 061	061 061		061 - NORMAL,NEG	50	49	49	100	
2020-844	067	URISCAN		061 061	061 061		063 - 4	50	1	49	0	
2020-844	071	ATLAS		061 061	061 061		061 - NORMAL,NEG	18	18	18	100	
2020-844	072	URS-9-TECO		061 061	061 061		061 - NORMAL,NEG	105	105	105	100	
2020-844	080	ROCHE (CHE		061 061	061 061		061 - NORMAL,NEG	37	37	37	100	
2020-844	089	AIMSTRIP		061 061	061 061		061 - NORMAL,NEG	112	112	112	100	
2020-844	094	CLINITEK		061 061	061 061		061 - NORMAL,NEG	104	103	103	100	
2020-844	094	CLINITEK		061 061	061 061		063 - 4	104	1	103	0	
2020-844	102	ACON LAB		061 061	061 061		061 - NORMAL,NEG	168	168	168	100	
2020-844	131	iCHEM VELO		061 061	061 061		061 - NORMAL,NEG	27	27	27	100	
2020-844	135	Hybrido AU		061 061	061 061		061 - NORMAL,NEG	13	13	13	100	
2020-844	138	URS-10T		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-844	140	MISSION A		061 061	061 061		061 - NORMAL,NEG	66	66	66	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-844	145	CLINITEK		061 061	061 061		061 - NORMAL,NEG	11	11	11	100	
2020-844	147	Mission120		061 061	061 061		061 - NORMAL,NEG	17	17	17	100	
2020-844	148	Mission500		061 061	061 061		061 - NORMAL,NEG	20	20	20	100	
066 URINAL.-UROBILINOGEN												
2020-845	011	AMES NM10S		061 063	061 063		061 - NORMAL,NEG	50	8	50	100	
2020-845	011	AMES NM10S		061 063	061 063		062 - 2	50	25	50	100	
2020-845	011	AMES NM10S		061 063	061 063		063 - 4	50	17	50	100	
2020-845	013	BAYER CLI			061 063		061 - NORMAL,NEG	36	18	36	100	
2020-845	013	BAYER CLI			061 063		062 - 2	36	16	36	100	
2020-845	013	BAYER CLI			061 063		063 - 4	36	2	36	100	
2020-845	067	URISCAN			061 063		061 - NORMAL,NEG	50	11	48	100	
2020-845	067	URISCAN			061 063		062 - 2	50	7	48	100	
2020-845	067	URISCAN			061 063		063 - 4	50	30	48	100	
2020-845	067	URISCAN			061 063		064 - 8	50	2	48	0	
2020-845	071	ATLAS		061 063	061 063		062 - 2	18	12	18	100	
2020-845	071	ATLAS		061 063	061 063		063 - 4	18	6	18	100	
2020-845	072	URS-9-TECO		061 063	061 063		061 - NORMAL,NEG	105	22	102	100	
2020-845	072	URS-9-TECO		061 063	061 063		062 - 2	105	54	102	100	
2020-845	072	URS-9-TECO		061 063	061 063		063 - 4	105	26	102	100	
2020-845	072	URS-9-TECO		061 063	061 063		064 - 8	105	3	102	0	
2020-845	080	ROCHE (CHE		061 061	061 063		061 - NORMAL,NEG	37	33	37	100	
2020-845	080	ROCHE (CHE		061 061	061 063		062 - 2	37	4	37	100	
2020-845	089	AIMSTRIP			061 063		061 - NORMAL,NEG	112	67	112	100	
2020-845	089	AIMSTRIP			061 063		062 - 2	112	35	112	100	
2020-845	089	AIMSTRIP			061 063		063 - 4	112	10	112	100	
2020-845	094	CLINITEK		061 063	061 063		061 - NORMAL,NEG	104	22	104	100	
2020-845	094	CLINITEK		061 063	061 063		062 - 2	104	55	104	100	
2020-845	094	CLINITEK		061 063	061 063		063 - 4	104	27	104	100	
2020-845	102	ACON LAB		061 061	061 063		061 - NORMAL,NEG	168	136	168	100	
2020-845	102	ACON LAB		061 061	061 063		062 - 2	168	30	168	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	102	ACON LAB		061 061	061 063		063 - 4	168	2	168	100	
2020-845	131	iCHEM VELO		062 064	061 063		063 - 4	27	27	27	100	
2020-845	135	Hybrido AU		061 063	061 063		062 - 2	13	12	13	100	
2020-845	135	Hybrido AU		061 063	061 063		063 - 4	13	1	13	100	
2020-845	138	URS-10T		061 063	061 063		063 - 4	11	4	11	100	
2020-845	138	URS-10T		061 063	061 063		061 - NORMAL,NEG	11	3	11	100	
2020-845	138	URS-10T		061 063	061 063		062 - 2	11	4	11	100	
2020-845	140	MISSION A		061 061	061 063		061 - NORMAL,NEG	66	53	66	100	
2020-845	140	MISSION A		061 061	061 063		062 - 2	66	11	66	100	
2020-845	140	MISSION A		061 061	061 063		063 - 4	66	2	66	100	
2020-845	145	CLINITEK		061 063	061 063		062 - 2	11	11	11	100	
2020-845	147	Mission120		061 061	061 063		061 - NORMAL,NEG	17	16	17	100	
2020-845	147	Mission120		061 061	061 063		062 - 2	17	1	17	100	
2020-845	148	Mission500		061 061	061 063		061 - NORMAL,NEG	20	18	20	100	
2020-845	148	Mission500		061 061	061 063		062 - 2	20	1	20	100	
2020-845	148	Mission500		061 061	061 063		063 - 4	20	1	20	100	
184 URINALYSIS LEUKOCYTE												
2020-841				001 001	001 001		001 - Negative	22	22	22	100	
2020-841	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100	
2020-841	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100	
2020-841	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100	
2020-841	071	ATLAS		001 001	001 001		001 - Negative	17	17	17	100	
2020-841	072	URS-9-TECO		001 001	001 001		001 - Negative	95	95	95	100	
2020-841	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100	
2020-841	089	AIMSTRIP		001 001	001 001		001 - Negative	107	107	107	100	
2020-841	094	CLINITEK		001 001	001 001		001 - Negative	102	102	102	100	
2020-841	102	ACON LAB		001 001	001 001		001 - Negative	150	150	150	100	
2020-841	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100	
2020-841	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100	
2020-841	138	URS-10T		001 001	001 001		001 - Negative	11	11	11	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-841	140	MISSION A		001 001	001 001		001 - Negative	75	75	75	100	
2020-841	145	CLINITEK		001 001	001 001		001 - Negative	11	11	11	100	
2020-841	147	Mission120		001 001	001 001		001 - Negative	18	18	18	100	
2020-841	148	Mission500		001 001	001 001		001 - Negative	19	19	19	100	
184 URINALYSIS LEUKOCYTE												
2020-842				002 004	003 005		002 - Traces	22	1	22	100	
2020-842				002 004	003 005		003 - Small	22	11	22	100	
2020-842				002 004	003 005		004 - Moderate	22	6	22	100	
2020-842				002 004	003 005		005 - Large	22	4	22	100	
2020-842	011	AMES NM10S		002 004	003 005		002 - Traces	47	4	47	100	
2020-842	011	AMES NM10S		002 004	003 005		003 - Small	47	33	47	100	
2020-842	011	AMES NM10S		002 004	003 005		004 - Moderate	47	9	47	100	
2020-842	011	AMES NM10S		002 004	003 005		005 - Large	47	1	47	100	
2020-842	013	BAYER CLI		002 004	003 005		002 - Traces	36	1	36	100	
2020-842	013	BAYER CLI		002 004	003 005		003 - Small	36	25	36	100	
2020-842	013	BAYER CLI		002 004	003 005		004 - Moderate	36	9	36	100	
2020-842	013	BAYER CLI		002 004	003 005		005 - Large	36	1	36	100	
2020-842	067	URISCAN		003 005	003 005		002 - Traces	49	1	48	0	
2020-842	067	URISCAN		003 005	003 005		003 - Small	49	5	48	100	
2020-842	067	URISCAN		003 005	003 005		004 - Moderate	49	23	48	100	
2020-842	067	URISCAN		003 005	003 005		005 - Large	49	20	48	100	
2020-842	071	ATLAS		003 005	003 005		004 - Moderate	17	14	17	100	
2020-842	071	ATLAS		003 005	003 005		005 - Large	17	3	17	100	
2020-842	072	URS-9-TECO		002 004	003 005		002 - Traces	95	5	95	100	
2020-842	072	URS-9-TECO		002 004	003 005		003 - Small	95	47	95	100	
2020-842	072	URS-9-TECO		002 004	003 005		004 - Moderate	95	43	95	100	
2020-842	080	ROCHE (CHE		004 005	003 005		003 - Small	34	1	34	100	
2020-842	080	ROCHE (CHE		004 005	003 005		004 - Moderate	34	3	34	100	
2020-842	080	ROCHE (CHE		004 005	003 005		005 - Large	34	30	34	100	
2020-842	089	AIMSTRIP		002 004	003 005		001 - Negative	107	1	106	0	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY

SUB: 50 URINALYSIS

GRP: 70 URINALYSIS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2020-842	089	AIMSTRIP		002 004	003 005		002 - Traces	107	6	106	100
2020-842	089	AIMSTRIP		002 004	003 005		003 - Small	107	65	106	100
2020-842	089	AIMSTRIP		002 004	003 005		004 - Moderate	107	31	106	100
2020-842	089	AIMSTRIP		002 004	003 005		005 - Large	107	4	106	100
2020-842	094	CLINITEK		002 004	003 005		002 - Traces	102	5	102	100
2020-842	094	CLINITEK		002 004	003 005		003 - Small	102	89	102	100
2020-842	094	CLINITEK		002 004	003 005		004 - Moderate	102	7	102	100
2020-842	094	CLINITEK		002 004	003 005		005 - Large	102	1	102	100
2020-842	102	ACON LAB		002 004	003 005		002 - Traces	150	4	150	100
2020-842	102	ACON LAB		002 004	003 005		003 - Small	150	82	150	100
2020-842	102	ACON LAB		002 004	003 005		004 - Moderate	150	63	150	100
2020-842	102	ACON LAB		002 004	003 005		005 - Large	150	1	150	100
2020-842	131	iCHEM VELO		004 005	003 005		004 - Moderate	26	2	26	100
2020-842	131	iCHEM VELO		004 005	003 005		005 - Large	26	24	26	100
2020-842	135	Hybrido AU		004 005	003 005		005 - Large	13	13	13	100
2020-842	138	URS-10T		002 004	003 005		003 - Small	11	7	11	100
2020-842	138	URS-10T		002 004	003 005		004 - Moderate	11	4	11	100
2020-842	140	MISSION A		002 004	003 005		003 - Small	75	41	75	100
2020-842	140	MISSION A		002 004	003 005		004 - Moderate	75	27	75	100
2020-842	140	MISSION A		002 004	003 005		005 - Large	75	1	75	100
2020-842	140	MISSION A		002 004	003 005		002 - Traces	75	6	75	100
2020-842	145	CLINITEK		003 005	003 005		003 - Small	11	1	11	100
2020-842	145	CLINITEK		003 005	003 005		004 - Moderate	11	10	11	100
2020-842	147	Mission120		002 004	003 005		002 - Traces	18	1	18	100
2020-842	147	Mission120		002 004	003 005		003 - Small	18	15	18	100
2020-842	147	Mission120		002 004	003 005		004 - Moderate	18	1	18	100
2020-842	147	Mission120		002 004	003 005		005 - Large	18	1	18	100
2020-842	148	Mission500		003 005	003 005		003 - Small	19	8	19	100
2020-842	148	Mission500		003 005	003 005		004 - Moderate	19	11	19	100

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
184 URINALYSIS LEUKOCYTE													
2020-843				001 001	001 001		001 - Negative	22	22	22	100		
2020-843	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100		
2020-843	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100		
2020-843	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100		
2020-843	071	ATLAS		001 001	001 001		001 - Negative	17	17	17	100		
2020-843	072	URS-9-TECO		001 001	001 001		001 - Negative	95	95	95	100		
2020-843	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100		
2020-843	089	AIMSTRIP		001 001	001 001		001 - Negative	107	107	107	100		
2020-843	094	CLINITEK		001 001	001 001		001 - Negative	102	102	102	100		
2020-843	102	ACON LAB		001 001	001 001		001 - Negative	150	149	149	100		
2020-843	102	ACON LAB		001 001	001 001		005 - Large	150	1	149	0		
2020-843	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100		
2020-843	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100		
2020-843	138	URS-10T		001 001	001 001		001 - Negative	11	11	11	100		
2020-843	140	MISSION A		001 001	001 001		001 - Negative	75	75	75	100		
2020-843	145	CLINITEK		001 001	001 001		001 - Negative	11	11	11	100		
2020-843	147	Mission120		001 001	001 001		001 - Negative	18	18	18	100		
2020-843	148	Mission500		001 001	001 001		001 - Negative	19	19	19	100		
2020-844				001 001	001 001		001 - Negative	22	22	22	100		
2020-844	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100		
2020-844	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100		
2020-844	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100		
2020-844	071	ATLAS		001 001	001 001		001 - Negative	17	17	17	100		
2020-844	072	URS-9-TECO		001 001	001 001		001 - Negative	95	93	93	100		
2020-844	072	URS-9-TECO		001 001	001 001		005 - Large	95	2	93	0		
2020-844	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100		
2020-844	089	AIMSTRIP		001 001	001 001		001 - Negative	107	106	106	100		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS				GRP: 70 URINALYSIS							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2020-844	089	AIMSTRIP		001 001	001 001		004 - Moderate	107	1	106	0		
2020-844	094	CLINITEK		001 001	001 001		001 - Negative	102	102	102	100		
2020-844	102	ACON LAB		001 001	001 001		001 - Negative	150	150	150	100		
2020-844	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100		
2020-844	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100		
2020-844	138	URS-10T		001 001	001 001		001 - Negative	11	11	11	100		
2020-844	140	MISSION A		001 001	001 001		001 - Negative	75	75	75	100		
2020-844	145	CLINITEK		001 001	001 001		001 - Negative	11	11	11	100		
2020-844	147	Mission120		001 001	001 001		001 - Negative	18	18	18	100		
2020-844	148	Mission500		001 001	001 001		001 - Negative	19	19	19	100		
184 URINALYSIS LEUKOCYTE													
2020-845				001 001	001 001		001 - Negative	22	22	22	100		
2020-845	011	AMES NM10S		001 001	001 001		001 - Negative	47	47	47	100		
2020-845	013	BAYER CLI		001 001	001 001		001 - Negative	36	36	36	100		
2020-845	067	URISCAN		001 001	001 001		001 - Negative	49	49	49	100		
2020-845	071	ATLAS		001 001	001 001		001 - Negative	17	17	17	100		
2020-845	072	URS-9-TECO		001 001	001 001		001 - Negative	95	94	94	100		
2020-845	072	URS-9-TECO		001 001	001 001		003 - Small	95	1	94	0		
2020-845	080	ROCHE (CHE		001 001	001 001		001 - Negative	34	34	34	100		
2020-845	089	AIMSTRIP		001 001	001 001		001 - Negative	107	106	106	100		
2020-845	089	AIMSTRIP		001 001	001 001		004 - Moderate	107	1	106	0		
2020-845	094	CLINITEK		001 001	001 001		001 - Negative	102	102	102	100		
2020-845	102	ACON LAB		001 001	001 001		001 - Negative	150	148	148	100		
2020-845	102	ACON LAB		001 001	001 001		003 - Small	150	1	148	0		
2020-845	102	ACON LAB		001 001	001 001		005 - Large	150	1	148	0		
2020-845	131	iCHEM VELO		001 001	001 001		001 - Negative	26	26	26	100		
2020-845	135	Hybrido AU		001 001	001 001		001 - Negative	13	13	13	100		
2020-845	138	URS-10T		001 001	001 001		001 - Negative	11	10	10	100		
2020-845	138	URS-10T		001 001	001 001		003 - Small	11	1	10	0		
2020-845	140	MISSION A		001 001	001 001		002 - Traces	75	2	73	0		

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 70 URINALYSIS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2020-845	140	MISSION A		001 001	001 001		001 - Negative	75	73	73	100	
2020-845	145	CLINITEK		001 001	001 001		001 - Negative	11	11	11	100	
2020-845	147	Mission120		001 001	001 001		001 - Negative	18	18	18	100	
2020-845	148	Mission500		001 001	001 001		001 - Negative	19	19	19	100	
<hr/>												
SPE: 10 CHEMISTRY			SUB: 50 URINALYSIS				GRP: 80 PREGNANCY TEST					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
063 URI-PREGNANCY TEST												
2020-841	091	AIM STEP		001 001	001 001		001 - NEGATIVE	34	34	34	100	
2020-841	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2020-842	091	AIM STEP		001 001	001 001		001 - NEGATIVE	34	34	34	100	
2020-842	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2020-843	091	AIM STEP		001 001	001 001		001 - NEGATIVE	34	34	34	100	
2020-843	101	HCG Alere		001 001	001 001		001 - NEGATIVE	23	23	23	100	
2020-844	091	AIM STEP		002 002	002 002		002 - POSITIVE	34	34	34	100	
2020-844	101	HCG Alere		002 002	002 002		002 - POSITIVE	23	23	23	100	
2020-845	091	AIM STEP		002 002	002 002		002 - POSITIVE	34	34	34	100	
2020-845	101	HCG Alere		002 002	002 002		002 - POSITIVE	23	23	23	100	

Participant Summary

MONTH 10 OF YEAR 2020

SPE: 10 CHEMISTRY		SUB: 50 URINALYSIS			GRP: 75 URINALYSIS SEDIMENT						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
150 URINAL.- SEDIMENT											
2020-851					028 028		008 - ERYTHROCYT	871	1	869	0
2020-851					028 028		014 - CA. OXALAT	871	1	869	0
2020-851					028 028		028 - TRIPLE PH.	871	869	869	100
2020-852					038 038		004 - HYALINE C.	868	1	867	0
2020-852					038 038		038 - YEAST/FUN.	868	867	867	100
2020-853					008 008		008 - ERYTHROCYT	868	861	861	100
2020-853					008 008		009 - LEUKOCYTE	868	1	861	0
2020-853					008 008		028 - TRIPLE PH.	868	1	861	0
2020-853					008 008		034 - OIL DROPL.	868	4	861	0
2020-853					008 008		038 - YEAST/FUN.	868	1	861	0
2020-854					004 004		004 - HYALINE C.	867	863	863	100
2020-854					004 004		006 - WAXY CAST	867	3	863	0
2020-854					004 004		038 - YEAST/FUN.	867	1	863	0
2020-855					036 036		008 - ERYTHROCYT	869	1	854	0
2020-855					036 036		026 - CA. CARBO.	869	1	854	0
2020-855					036 036		031 - FIBERS	869	1	854	0
2020-855					036 036		035 - POLLEN GR.	869	12	854	0
2020-855					036 036		036 - STARCH CR.	869	854	854	100

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 90 SPECIAL TEST

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
068	CARCINOE. ANTIG. CEA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-811	29 VITROS			11	25.909	0.592	2.28	24.100	27.700
2020-811	38 ROCHE E			21	24.195	1.466	6.06	19.800	28.600
2020-811	44 BECK.COUL.			17	28.288	2.163	7.65	21.800	34.800
2020-811	50 ARC			12	29.283	1.442	4.92	25.000	33.600
2020-812	29 VITROS			11	17.791	0.491	2.76	16.300	19.300
2020-812	38 ROCHE E			21	16.957	1.357	8.00	12.900	21.000
2020-812	44 BECK.COUL.			17	19.069	1.137	5.96	15.700	22.500
2020-812	50 ARC			12	19.433	0.798	4.11	17.000	21.800
2020-813	29 VITROS			11	12.055	0.350	2.90	11.000	13.100
2020-813	38 ROCHE E			21	11.790	0.900	7.63	9.100	14.500
2020-813	44 BECK.COUL.			17	12.969	0.648	5.00	11.000	14.900
2020-813	50 ARC			12	13.142	0.736	5.60	10.900	15.400
2020-814	29 VITROS			11	20.745	0.654	3.15	18.800	22.700
2020-814	38 ROCHE E			21	19.533	1.601	8.20	14.700	24.300
2020-814	44 BECK.COUL.			17	22.053	1.237	5.61	18.300	25.800
2020-814	50 ARC			12	22.350	1.241	5.55	18.600	26.100
2020-815	29 VITROS			11	7.655	0.178	2.32	7.100	8.200
2020-815	38 ROCHE E			21	7.810	0.623	7.97	5.900	9.700
2020-815	44 BECK.COUL.			17	8.063	0.459	5.70	6.700	9.400
2020-815	50 ARC			12	8.417	0.496	5.90	6.900	9.900
069	PROS. SPEC-ANTIG-PSA								
2020-811	29 VITROS			36	8.063	0.317	3.93	7.100	9.000
2020-811	36 TOSOH AIA			13	9.154	1.062	11.60	6.900	11.400
2020-811	38 ROCHE E			46	11.400	0.705	6.19	9.300	13.500
2020-811	39 DIMENS.RXL			54	10.400	0.449	4.32	9.100	11.700

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-811	44	BECK.COUL.		53	12.175	0.542	4.45	10.600	13.800
2020-811	50	ARC		18	8.747	0.629	7.19	6.900	10.600
2020-811	57	S. EXL 200		17	10.376	0.517	4.99	8.800	11.900
069	PROS. SPEC-ANTIG-PSA	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{LOWER}			
2020-812	29	VITROS		36	5.697	0.253	4.45	4.900	6.500
2020-812	36	TOSOH AIA		13	6.733	0.330	4.90	5.700	7.700
2020-812	38	ROCHE E		46	8.028	0.542	6.75	6.400	9.700
2020-812	39	DIMENS.RXL		54	7.151	0.384	5.38	6.000	8.300
2020-812	44	BECK.COUL.		53	8.387	0.444	5.30	7.100	9.700
2020-812	50	ARC		18	6.061	0.593	9.78	4.500	7.600
2020-812	57	S. EXL 200		17	7.082	0.305	4.31	6.200	8.000
2020-813	29	VITROS		36	4.040	0.164	4.07	3.500	4.500
2020-813	36	TOSOH AIA		13	4.633	0.201	4.35	4.000	5.200
2020-813	38	ROCHE E		46	5.507	0.373	6.78	4.400	6.600
2020-813	39	DIMENS.RXL		54	4.862	0.308	6.33	3.900	5.800
2020-813	44	BECK.COUL.		53	5.717	0.267	4.67	4.900	6.500
2020-813	50	ARC		18	3.950	0.293	7.42	3.100	4.800
2020-813	57	S. EXL 200		17	4.794	0.262	5.47	4.000	5.600
2020-814	29	VITROS		36	6.554	0.290	4.43	5.700	7.400
2020-814	36	TOSOH AIA		13	7.546	0.750	9.94	5.700	9.400
2020-814	38	ROCHE E		46	9.191	0.639	6.95	7.300	11.100
2020-814	39	DIMENS.RXL		54	8.274	0.429	5.19	7.000	9.600
2020-814	44	BECK.COUL.		53	9.785	0.394	4.03	8.600	11.000
2020-814	50	ARC		18	6.883	0.466	6.77	5.500	8.300
2020-814	57	S. EXL 200		17	8.276	0.381	4.61	7.100	9.400
2020-815	29	VITROS		36	2.667	0.100	3.75	2.400	3.000
2020-815	36	TOSOH AIA		13	2.900	0.254	8.77	2.200	3.600

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 90 SPECIAL TEST							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-814	44	BECK.COUL.		29	532.552	33.422	6.28	432.000	633.000		
2020-814	50	ARC		17	717.059	62.511	8.72	538.000	896.000		
2020-814	57	S. EXL 200		11	709.545	51.250	7.22	556.000	863.000		
071 VITAMIN B 12			CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}						
2020-815	29	VITROS		22	397.273	15.413	3.88	351.000	444.000		
2020-815	38	ROCHE E		23	347.565	47.306	13.61	261.000	434.000		
2020-815	38	ROCHE E		23	254.250	11.121	4.37	221.000	288.000		
2020-815	44	BECK.COUL.		29	276.724	18.813	6.80	220.000	333.000		
2020-815	50	ARC		17	397.353	21.804	5.49	332.000	463.000		
2020-815	57	S. EXL 200		11	460.455	36.933	8.02	350.000	571.000		
SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
146 GLYCOHEMOGLOBIN											
2020-821	2	DCA-2000		12	6.350	0.161	2.53	5.900	6.800		
2020-821	12	COBAS/INT.		28	5.896	0.299	5.08	5.000	6.800		
2020-821	14	IMMUNOTUR.		10	6.070	0.456	7.52	4.700	7.400		
2020-821	22	VITROS		40	5.726	0.213	3.73	5.100	6.400		
2020-821	23	IMOLA/RX		10	6.370	0.469	7.36	5.000	7.800		
2020-821	30	DAYTONA		13	6.238	0.282	4.51	5.400	7.100		
2020-821	31	IMMUNOTURB		31	5.743	0.369	6.43	4.600	6.900		
2020-821	40	SIEMENS DI		81	6.114	0.245	4.01	5.400	6.800		
2020-821	46	ARCHITECT		13	5.625	0.092	1.64	5.300	5.900		
2020-822	2	DCA-2000		12	11.233	1.023	9.11	8.400	14.000		
2020-822	12	COBAS/INT.		28	10.661	0.389	3.64	9.500	11.800		
2020-822	14	IMMUNOTUR.		10	11.130	1.348	12.11	8.300	13.900		
2020-822	22	VITROS		40	10.141	0.428	4.22	8.900	11.400		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE					
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-822	23	IMOLA/RX		10	12.520	1.416	11.31	9.400	15.700
2020-822	30	DAYTONA		13	12.038	1.002	8.32	9.000	15.000
2020-822	31	IMMUNOTURB		31	10.667	0.331	3.10	9.700	11.700
2020-822	40	SIEMENS DI		81	10.681	0.290	2.71	9.800	11.500
2020-822	46	ARCHITECT		13	10.215	0.251	2.45	9.500	11.000
146	GLYCOHEMOGLOBIN	CLIA TEST ID #:	0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 25 % {LOWER}					
2020-823	2	DCA-2000		12	6.417	0.285	4.45	5.600	7.300
2020-823	12	COBAS/INT.		28	5.857	0.325	5.54	4.900	6.800
2020-823	14	IMMUNOTUR.		10	6.240	0.539	8.64	4.700	7.800
2020-823	22	VITROS		40	5.705	0.233	4.09	5.000	6.400
2020-823	23	IMOLA/RX		10	6.470	0.420	6.49	5.200	7.700
2020-823	30	DAYTONA		13	6.223	0.353	5.68	5.200	7.300
2020-823	31	IMMUNOTURB		31	5.726	0.369	6.44	4.600	6.800
2020-823	40	SIEMENS DI		81	6.111	0.248	4.06	5.400	6.900
2020-823	46	ARCHITECT		13	5.617	0.099	1.76	5.300	5.900
2020-824	2	DCA-2000		12	7.591	0.162	2.14	7.100	8.100
2020-824	12	COBAS/INT.		28	7.232	0.309	4.28	6.300	8.200
2020-824	14	IMMUNOTUR.		10	7.650	1.011	13.22	5.700	9.600
2020-824	22	VITROS		40	7.033	0.298	4.24	6.100	7.900
2020-824	23	IMOLA/RX		10	7.940	0.577	7.26	6.200	9.700
2020-824	30	DAYTONA		13	7.762	0.431	5.55	6.500	9.100
2020-824	31	IMMUNOTURB		31	7.090	0.277	3.91	6.300	7.900
2020-824	40	SIEMENS DI		81	7.440	0.265	3.57	6.600	8.200
2020-824	46	ARCHITECT		13	6.838	0.198	2.90	6.200	7.400
2020-825	2	DCA-2000		12	7.600	0.183	2.40	7.100	8.100
2020-825	12	COBAS/INT.		28	7.211	0.303	4.20	6.300	8.100
2020-825	14	IMMUNOTUR.		10	7.690	1.000	13.01	5.800	9.600
2020-825	22	VITROS		40	7.082	0.314	4.43	6.100	8.000

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY		SUB: 60 SPECIAL TESTS		GRP: 95 GLYCOHEMOGLOBINE							
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2020-825	23	IMOLA/RX		10	7.820	0.483	6.18	6.400	9.300		
2020-825	30	DAYTONA		13	7.654	0.318	4.15	6.700	8.600		
2020-825	31	IMMUNOTURB		31	7.113	0.242	3.40	6.400	7.800		
2020-825	40	SIEMENS DI		81	7.449	0.276	3.70	6.600	8.300		
2020-825	46	ARCHITECT		13	6.831	0.186	2.72	6.300	7.400		
SPEC: 10 CHEMISTRY						SUB: 60 SPECIAL TESTS		GRP: 99 Vitamin D			
ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
178	Vitamin D	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	20	%						
2020-301	3	VITROS		33	59.406	5.544	9.33	48.000	71.000		
2020-301	5	ROCHE E		31	41.517	3.114	7.50	33.000	50.000		
2020-301	9	BECK.COUL.		45	63.614	4.872	7.66	51.000	76.000		
2020-301	11	ARC		14	58.500	2.557	4.37	47.000	70.000		
2020-301	25	S. DIME		16	46.000	3.082	6.70	37.000	55.000		
2020-302	3	VITROS		33	43.250	4.603	10.64	35.000	52.000		
2020-302	5	ROCHE E		31	31.967	3.420	10.70	26.000	38.000		
2020-302	9	BECK.COUL.		45	47.932	3.639	7.59	38.000	58.000		
2020-302	11	ARC		14	44.857	1.641	3.66	36.000	54.000		
2020-302	25	S. DIME		16	35.188	2.721	7.73	28.000	42.000		
2020-303	3	VITROS		33	82.344	5.898	7.16	66.000	99.000		
2020-303	5	ROCHE E		31	50.267	4.939	9.83	40.000	60.000		
2020-303	9	BECK.COUL.		45	78.273	6.454	8.25	63.000	94.000		
2020-303	11	ARC		14	73.143	2.531	3.46	59.000	88.000		
2020-303	25	S. DIME		16	56.563	3.856	6.82	45.000	68.000		
2020-304	3	VITROS		33	24.394	4.369	17.91	20.000	29.000		

Participant Summary

MONTH 10 OF YEAR 2020

SPEC: 10 CHEMISTRY

SUB: 60 SPECIAL TESTS

GRP: 99 Vitamin D

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2020-304	5	ROCHE E		31	20.900	2.256	10.79	17.000	25.000
2020-304	9	BECK.COUL.		45	29.600	2.568	8.68	24.000	36.000
2020-304	11	ARC		14	26.929	1.534	5.70	22.000	32.000
2020-304	25	S. DIME		16	22.938	1.886	8.22	18.000	28.000
178 Vitamin D		CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 20 %						
2020-305	3	VITROS		33	32.938	4.603	13.97	26.000	40.000
2020-305	5	ROCHE E		31	23.267	2.407	10.35	19.000	28.000
2020-305	9	BECK.COUL.		45	34.000	2.646	7.78	27.000	41.000
2020-305	11	ARC		14	31.286	1.278	4.08	25.000	38.000
2020-305	25	S. DIME		16	26.625	2.342	8.80	21.000	32.000