

**Participant Summary**  
**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 225 HELICOBACTER PYLORI					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
155 HELICOBACTER PYLORI												
2021-771	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	22	22	22	100	
2021-772	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2021-773	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	22	22	22	100	
2021-774	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	22	22	22	100	
2021-775	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	22	22	22	100	

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 210 ANTINUCLEAR ANTIB.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
089 ANTIN. ANT.-ANA-QUAL												
2021-731	196	ENZYMATIC		002 002	002 002		002 - POS.OR ABN	10	10	10	100	
2021-732	196	ENZYMATIC		002 002	002 002		002 - POS.OR ABN	10	10	10	100	
2021-733	196	ENZYMATIC		001 001	001 001		001 - NEG.OR NOR	10	10	10	100	
2021-734	196	ENZYMATIC		002 002	002 002		002 - POS.OR ABN	10	10	10	100	
2021-735	196	ENZYMATIC		001 001	001 001		001 - NEG.OR NOR	10	10	10	100	

SPEC: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
145 CRP - QUANTITATIVE												
			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}								
2021-741	34	DADE DIMEN		12	0.162	0.119	73.83	0	0.520			
2021-741	37	BECKMAN		12	0.388	0.271	69.86	0	1.200			
2021-741	39	ROCHE COBA		12	0.040	0.027	68.26	0	0.120			
2021-741	42	ASI ARLING		27	0.100	0.000	0.00	0	0.400			
2021-741	42	ASI ARLING		27	0.200	0.283	141.42	0	1.050			
2021-741	47	VITROS 560		18	0.500	0.000	0.00	0.500	0.500			
2021-741	47	VITROS 560		18	0.100	0.000	0.00	0	0.400			
2021-742	34	DADE DIMEN		12	1.434	0.174	12.13	0.910	1.960			
2021-742	37	BECKMAN		12	1.324	0.038	2.85	1.210	1.440			
2021-742	39	ROCHE COBA		12	1.158	0.048	4.15	1.010	1.300			
2021-742	42	ASI ARLING		27	2.423	1.159	47.82	0	5.900			

## Participant Summary

MONTH 9 OF YEAR 2021

SPEC: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

ANALYTE			NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-742	42	ASI ARLING	27	1.380	1.000	72.46	0	2.880
2021-742	47	VITROS 560	18	1.380	1.000	72.46	0	2.880
2021-742	47	VITROS 560	18	2.432	0.230	9.45	1.740	3.120
145 CRP - QUANTITATIVE			CLIA TEST ID #: 0      EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}					
2021-743	34	DADE DIMEN	12	3.667	0.151	4.13	3.210	4.120
2021-743	37	BECKMAN	12	3.403	0.155	4.55	2.940	3.870
2021-743	39	ROCHE COBA	12	3.204	0.159	4.97	2.730	3.680
2021-743	42	ASI ARLING	27	3.000	1.000	33.33	1.500	4.500
2021-743	42	ASI ARLING	27	2.931	1.204	41.07	0	6.540
2021-743	47	VITROS 560	18	4.971	0.273	5.50	4.150	5.790
2021-743	47	VITROS 560	18	3.000	1.000	33.33	1.500	4.500
2021-744	34	DADE DIMEN	12	0.113	0.129	114.93	0	0.500
2021-744	37	BECKMAN	12	0.394	0.278	70.48	0	1.230
2021-744	39	ROCHE COBA	12	0.028	0.017	62.05	0	0.080
2021-744	42	ASI ARLING	27	0.200	0.283	141.42	0	1.050
2021-744	42	ASI ARLING	27	0.100	0.000	0.00	0	0.400
2021-744	47	VITROS 560	18	0.100	0.000	0.00	0	0.400
2021-744	47	VITROS 560	18	0.500	0.000	0.00	0.500	0.500
2021-745	34	DADE DIMEN	12	1.610	0.073	4.54	1.390	1.830
2021-745	37	BECKMAN	12	1.392	0.100	7.19	1.090	1.690
2021-745	39	ROCHE COBA	12	1.303	0.055	4.18	1.140	1.470
2021-745	42	ASI ARLING	27	1.500	1.000	66.67	0	3.900
2021-745	42	ASI ARLING	27	1.846	0.598	32.40	0.050	3.640
2021-745	47	VITROS 560	18	2.559	0.183	7.15	2.010	3.110
2021-745	47	VITROS 560	18	1.500	1.000	66.67	0	3.900

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
091 ANTISTREP.O-QUALITA.													
2021-741	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	21	21	21	100		
2021-741	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	24	24	24	100		
2021-742	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	21	21	21	100		
2021-742	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	24	24	24	100		
2021-743	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	21	21	21	100		
2021-743	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	24	24	24	100		
2021-744	378 RHEUMAJET			002 002	002 002		002 - POSITIVE	21	21	21	100		
2021-744	384 ASI A.SCT.			002 002	002 002		002 - POSITIVE	24	24	24	100		
2021-745	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	21	21	21	100		
2021-745	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	24	24	24	100		
092 ANTISTREP.O-SEMI QUA													
2021-741	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	18	18	18	100		
2021-742	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	18	18	18	100		
2021-743	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	18	18	18	100		
2021-744	381 RHEUMAJET			015 019	015 019		016 - 200 IU/ml	18	4	18	100		
2021-744	381 RHEUMAJET			015 019	015 019		017 - 400 IU/ml	18	14	18	100		
2021-745	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	18	18	18	100		

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
103 INF. MONO- QUALITAT.												
2021-741	109	BIOKIT MON		001 001	001 001		001 - NEGATIVE	76	76	76	100	
2021-741	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	127	127	127	100	
2021-741	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	35	35	35	100	
2021-741	405	Acceava		001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-742	109	BIOKIT MON		001 001	001 001		001 - NEGATIVE	76	76	76	100	
2021-742	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	127	127	127	100	
2021-742	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	35	35	35	100	
2021-742	405	Acceava		001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-743	109	BIOKIT MON		001 001	001 001		001 - NEGATIVE	76	76	76	100	
2021-743	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	127	127	127	100	
2021-743	394	TECO DIAG		001 001	001 001		002 - POSITIVE	35	1	34	0	
2021-743	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	35	34	34	100	
2021-743	405	Acceava		001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-744	109	BIOKIT MON		001 001	001 001		001 - NEGATIVE	76	76	76	100	
2021-744	393	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	127	127	127	100	
2021-744	394	TECO DIAG		001 001	001 001		001 - NEGATIVE	35	35	35	100	
2021-744	405	Acceava		001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-745	109	BIOKIT MON		002 002	002 002		002 - POSITIVE	76	76	76	100	
2021-745	393	ASI A.SCT.		002 002	002 002		002 - POSITIVE	127	127	127	100	
2021-745	394	TECO DIAG		002 002	002 002		002 - POSITIVE	35	35	35	100	
2021-745	405	Acceava		002 002	002 002		002 - POSITIVE	14	13	13	100	
2021-745	405	Acceava		002 002	002 002		001 - NEGATIVE	14	1	13	0	

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
105 RF FACTOR(RA)-QUALI.												
2021-741	041	BIOKIT RHE		001	001	001	001	001 - NEGATIVE	15	15	15	100
2021-741	393	RHEUMAJET		001	001	001	001	001 - NEGATIVE	80	80	80	100
2021-741	400	TECO DIAG.		001	001	001	001	001 - NEGATIVE	54	54	54	100
2021-741	402	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	127	126	126	100
2021-741	402	ASI A.SCT.		001	001	001	001	002 - POSITIVE	127	1	126	0
2021-741	406	AIM CRP RF		001	001	001	001	001 - NEGATIVE	31	29	29	100
2021-741	406	AIM CRP RF		001	001	001	001	002 - POSITIVE	31	2	29	0
2021-741	408	STANBIO		001	001	001	001	001 - NEGATIVE	23	23	23	100
2021-742	041	BIOKIT RHE		002	002	002	002	002 - POSITIVE	15	15	15	100
2021-742	393	RHEUMAJET		002	002	002	002	002 - POSITIVE	80	80	80	100
2021-742	400	TECO DIAG.		002	002	002	002	002 - POSITIVE	54	54	54	100
2021-742	402	ASI A.SCT.		002	002	002	002	002 - POSITIVE	127	127	127	100
2021-742	406	AIM CRP RF		002	002	002	002	002 - POSITIVE	31	31	31	100
2021-742	408	STANBIO		002	002	002	002	002 - POSITIVE	23	23	23	100
2021-743	041	BIOKIT RHE		001	001	001	001	001 - NEGATIVE	15	15	15	100
2021-743	393	RHEUMAJET		001	001	001	001	001 - NEGATIVE	80	80	80	100
2021-743	400	TECO DIAG.		001	001	001	001	001 - NEGATIVE	54	54	54	100
2021-743	402	ASI A.SCT.		001	001	001	001	001 - NEGATIVE	127	124	124	100
2021-743	402	ASI A.SCT.		001	001	001	001	002 - POSITIVE	127	3	124	0
2021-743	406	AIM CRP RF		001	001	001	001	001 - NEGATIVE	31	28	28	100
2021-743	406	AIM CRP RF		001	001	001	001	002 - POSITIVE	31	3	28	0
2021-743	408	STANBIO		001	001	001	001	001 - NEGATIVE	23	23	23	100
2021-744	041	BIOKIT RHE		001	001	001	001	001 - NEGATIVE	15	15	15	100
2021-744	393	RHEUMAJET		001	001	001	001	001 - NEGATIVE	80	80	80	100
2021-744	400	TECO DIAG.		001	001	001	001	001 - NEGATIVE	54	53	53	100

### Participant Summary

MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-744	400	TECO DIAG.		001 001	001 001		002 - POSITIVE	54	1	53	0
2021-744	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	127	126	126	100
2021-744	402	ASI A.SCT.		001 001	001 001		002 - POSITIVE	127	1	126	0
2021-744	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	31	31	31	100
2021-744	408	STANBIO		001 001	001 001		001 - NEGATIVE	23	23	23	100
105 RF FACTOR(RA)-QUALI.											
2021-745	041	BIOKIT RHE		002 002	002 002		002 - POSITIVE	15	15	15	100
2021-745	393	RHEUMAJET		002 002	002 002		002 - POSITIVE	80	80	80	100
2021-745	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	54	54	54	100
2021-745	402	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	127	2	125	0
2021-745	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	127	125	125	100
2021-745	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	31	31	31	100
2021-745	408	STANBIO		002 002	002 002		002 - POSITIVE	23	23	23	100
106 RF FAC(RA)-SEMI QUAN											
2021-741	393	RHEUMAJET		001 001	001 001		001 - <8-20 IU/m	41	40	40	100
2021-741	393	RHEUMAJET		001 001	001 001		002 - 8-20 IU/ml	41	1	40	0
2021-741	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	12	11	11	100
2021-741	407	ASI SLIDE		001 001	001 001		006 - 136-320 IU	12	1	11	0
2021-742	393	RHEUMAJET		003 007	003 007		003 - 24-40 IU/m	41	2	41	100
2021-742	393	RHEUMAJET		003 007	003 007		004 - 40-80 IU/m	41	16	41	100
2021-742	393	RHEUMAJET		003 007	003 007		005 - 72-160 IU	41	23	41	100
2021-742	407	ASI SLIDE		004 008	003 007		001 - <8-20 IU/m	12	1	11	0
2021-742	407	ASI SLIDE		004 008	003 007		004 - 40-80 IU/m	12	2	11	100
2021-742	407	ASI SLIDE		004 008	003 007		005 - 72-160 IU	12	2	11	100
2021-742	407	ASI SLIDE		004 008	003 007		006 - 136-320 IU	12	4	11	100
2021-742	407	ASI SLIDE		004 008	003 007		007 - 264-640 IU	12	3	11	100
2021-743	393	RHEUMAJET		001 001	001 001		001 - <8-20 IU/m	41	41	41	100

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-743	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	12	11	11	100	
2021-743	407	ASI SLIDE		001 001	001 001		002 - 8-20 IU/ml	12	1	11	0	
106 RF FAC(RA)-SEMI QUAN												
2021-744	393	RHEUMAJET		001 001	001 001		001 - <8-20 IU/m	41	41	41	100	
2021-744	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	12	12	12	100	
2021-745	393	RHEUMAJET		002 006	002 006		003 - 24-40 IU/m	41	7	41	100	
2021-745	393	RHEUMAJET		002 006	002 006		004 - 40-80 IU/m	41	33	41	100	
2021-745	393	RHEUMAJET		002 006	002 006		005 - 72-160 IU	41	1	41	100	
2021-745	407	ASI SLIDE		002 006	002 006		001 - <8-20 IU/m	12	1	10	0	
2021-745	407	ASI SLIDE		002 006	002 006		002 - 8-20 IU/ml	12	1	10	100	
2021-745	407	ASI SLIDE		002 006	002 006		004 - 40-80 IU/m	12	3	10	100	
2021-745	407	ASI SLIDE		002 006	002 006		005 - 72-160 IU	12	3	10	100	
2021-745	407	ASI SLIDE		002 006	002 006		006 - 136-320 IU	12	3	10	100	
2021-745	407	ASI SLIDE		002 006	002 006		007 - 264-640 IU	12	1	10	0	
144 CRP- QUALITATIVE												
2021-741	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	143	143	143	100	
2021-741	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	42	42	42	100	
2021-741	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	39	39	39	100	
2021-741	041	STANBIO		001 001	001 001		001 - NEGATIVE	13	13	13	100	
2021-742	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	143	143	143	100	
2021-742	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	42	1	41	0	
2021-742	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	42	41	41	100	
2021-742	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	39	39	39	100	
2021-742	041	STANBIO		002 002	002 002		002 - POSITIVE	13	13	13	100	
2021-743	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	143	7	136	0	



**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-743	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	143	136	136	100
2021-743	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	42	3	39	0
2021-743	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	42	39	39	100
2021-743	036	AIM CRP RF		002 002	002 002		001 - NEGATIVE	39	1	38	0
2021-743	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	39	38	38	100
2021-743	041	STANBIO		002 002	002 002		002 - POSITIVE	13	13	13	100
144 CRP- QUALITATIVE											
2021-744	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	143	140	140	100
2021-744	026	ASI A.SCT.		001 001	001 001		002 - POSITIVE	143	3	140	0
2021-744	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	42	42	42	100
2021-744	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	39	38	38	100
2021-744	036	AIM CRP RF		001 001	001 001		002 - POSITIVE	39	1	38	0
2021-744	041	STANBIO		001 001	001 001		001 - NEGATIVE	13	13	13	100
2021-745	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	143	2	141	0
2021-745	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	143	141	141	100
2021-745	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	42	1	41	0
2021-745	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	42	41	41	100
2021-745	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	39	39	39	100
2021-745	041	STANBIO		002 002	002 002		002 - POSITIVE	13	13	13	100

**Participant Summary**  
**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 240 RUBELLA - QUALITAT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
107 RUBELLA-QUALITATIVE												
2021-721	335	BIOSIGN-BI		002	002	002	002	001 - NEG-NOT IM	11	1	10	0
2021-721	335	BIOSIGN-BI		002	002	002	002	002 - POS.OR IM.	11	10	10	100
2021-722	335	BIOSIGN-BI		001	001	001	001	001 - NEG-NOT IM	11	10	10	100
2021-722	335	BIOSIGN-BI		001	001	001	001	002 - POS.OR IM.	11	1	10	0
2021-723	335	BIOSIGN-BI		002	002	002	002	001 - NEG-NOT IM	11	1	10	0
2021-723	335	BIOSIGN-BI		002	002	002	002	002 - POS.OR IM.	11	10	10	100
2021-724	335	BIOSIGN-BI		001	001	001	001	001 - NEG-NOT IM	11	11	11	100
2021-725	335	BIOSIGN-BI		002	002	002	002	001 - NEG-NOT IM	11	1	10	0
2021-725	335	BIOSIGN-BI		002	002	002	002	002 - POS.OR IM.	11	10	10	100

### Participant Summary

MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 247 MYCOPLASMA					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
548 MYCOPLASMA												
2021-791	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	481	481	481	100
2021-792	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	481	5	476	0
2021-792	001	IMMU. CARD		002	002	002	002	002 - Reactive	481	476	476	100
2021-793	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	481	481	481	100
2021-794	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	481	2	479	0
2021-794	001	IMMU. CARD		002	002	002	002	002 - Reactive	481	479	479	100
2021-795	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	481	3	478	0
2021-795	001	IMMU. CARD		002	002	002	002	002 - Reactive	481	478	478	100
554 COVID-19 CORONAVIRUS												
SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-661	001	IgM	001	CORONACHEC	002	002	002 - Positive	16	16	16	100	
2021-661	001	IgM	002	HEALGEN	002	002	002 - Positive	285	285	285	100	
2021-661	001	IgM	007	RIGHT SIGN	002	002	002 - Positive	92	92	92	100	
2021-661	001	IgM	016	CLARITY	002	002	001 - Negative	43	2	41	0	
2021-661	001	IgM	016	CLARITY	002	002	002 - Positive	43	41	41	100	
2021-661	001	IgM	030	FASTEP	002	002	002 - Positive	18	18	18	100	
2021-661	001	IgM	033	BIOHIT SAR	002	002	002 - Positive	10	10	10	100	
2021-661	001	IgM	034	Care Start	002	002	002 - Positive	13	13	13	100	
2021-661	002	IgG	001	CORONACHEC	002	002	002 - Positive	16	16	16	100	

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 248 COVID-19 CORONAVIRUS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-661	002 IgG	002 HEALGEN		002 002			002 - Positive	285	285	285	100
2021-661	002 IgG	007 RIGHT SIGN		002 002			002 - Positive	90	90	90	100
2021-661	002 IgG	016 CLARITY		002 002			002 - Positive	43	43	43	100
2021-661	002 IgG	030 FASTEP		002 002			002 - Positive	18	18	18	100
2021-661	002 IgG	033 BIOHIT SAR		002 002			002 - Positive	10	10	10	100
2021-661	002 IgG	034 Care Start		002 002			002 - Positive	13	13	13	100
2021-661	003 SARS COV 2	021 VITROS SAR		002 002			002 - Positive	13	13	13	100
554 COVID-19 CORONAVIRUS											
2021-662	001 IgM	001 CORONACHEC		001 001			001 - Negative	16	16	16	100
2021-662	001 IgM	002 HEALGEN		001 001			001 - Negative	283	283	283	100
2021-662	001 IgM	007 RIGHT SIGN		001 001			001 - Negative	91	91	91	100
2021-662	001 IgM	016 CLARITY		001 001			001 - Negative	43	43	43	100
2021-662	001 IgM	030 FASTEP		001 001			001 - Negative	18	18	18	100
2021-662	001 IgM	033 BIOHIT SAR		001 001			001 - Negative	10	10	10	100
2021-662	001 IgM	034 Care Start		001 001			001 - Negative	13	13	13	100
2021-662	002 IgG	001 CORONACHEC		001 001			001 - Negative	16	16	16	100
2021-662	002 IgG	002 HEALGEN		001 001			001 - Negative	285	285	285	100
2021-662	002 IgG	007 RIGHT SIGN		001 001			001 - Negative	91	91	91	100
2021-662	002 IgG	016 CLARITY		001 001			001 - Negative	43	43	43	100
2021-662	002 IgG	030 FASTEP		001 001			001 - Negative	18	18	18	100
2021-662	002 IgG	033 BIOHIT SAR		001 001			001 - Negative	10	10	10	100
2021-662	002 IgG	034 Care Start		001 001			001 - Negative	13	13	13	100
2021-662	003 SARS COV 2	021 VITROS SAR		001 001			001 - Negative	13	13	13	100
2021-663	001 IgM	001 CORONACHEC		001 001			001 - Negative	16	16	16	100
2021-663	001 IgM	002 HEALGEN		001 001			001 - Negative	285	285	285	100
2021-663	001 IgM	007 RIGHT SIGN		001 001			001 - Negative	91	91	91	100
2021-663	001 IgM	016 CLARITY		001 001			001 - Negative	43	43	43	100
2021-663	001 IgM	030 FASTEP		001 001			001 - Negative	18	18	18	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 248 COVID-19 CORONAVIRUS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-663	001 IgM	033 BIOHIT SAR		001 001			001 - Negative	10	10	10	100
2021-663	001 IgM	034 Care Start		001 001			001 - Negative	13	13	13	100
2021-663	002 IgG	001 CORONACHEC		001 001			001 - Negative	16	16	16	100
2021-663	002 IgG	002 HEALGEN		001 001			001 - Negative	285	285	285	100
2021-663	002 IgG	007 RIGHT SIGN		001 001			001 - Negative	91	91	91	100
2021-663	002 IgG	016 CLARITY		001 001			001 - Negative	43	43	43	100
2021-663	002 IgG	030 FASTEP		001 001			001 - Negative	18	18	18	100
2021-663	002 IgG	033 BIOHIT SAR		001 001			001 - Negative	10	10	10	100
2021-663	002 IgG	034 Care Start		001 001			001 - Negative	13	13	13	100
2021-663	003 SARS COV 2	021 VITROS SAR		001 001			001 - Negative	13	13	13	100
554 COVID-19 CORONAVIRUS											
2021-664	001 IgM	001 CORONACHEC		002 002			002 - Positive	16	16	16	100
2021-664	001 IgM	002 HEALGEN		002 002			002 - Positive	285	285	285	100
2021-664	001 IgM	007 RIGHT SIGN		002 002			001 - Negative	91	1	90	0
2021-664	001 IgM	007 RIGHT SIGN		002 002			002 - Positive	91	90	90	100
2021-664	001 IgM	016 CLARITY		002 002			001 - Negative	43	2	41	0
2021-664	001 IgM	016 CLARITY		002 002			002 - Positive	43	41	41	100
2021-664	001 IgM	030 FASTEP		002 002			002 - Positive	18	18	18	100
2021-664	001 IgM	033 BIOHIT SAR		002 002			002 - Positive	10	10	10	100
2021-664	001 IgM	034 Care Start		002 002			002 - Positive	13	13	13	100
2021-664	002 IgG	001 CORONACHEC		002 002			002 - Positive	16	16	16	100
2021-664	002 IgG	002 HEALGEN		002 002			002 - Positive	285	285	285	100
2021-664	002 IgG	007 RIGHT SIGN		002 002			002 - Positive	91	91	91	100
2021-664	002 IgG	016 CLARITY		002 002			002 - Positive	43	43	43	100
2021-664	002 IgG	030 FASTEP		002 002			002 - Positive	18	18	18	100
2021-664	002 IgG	033 BIOHIT SAR		002 002			002 - Positive	10	10	10	100
2021-664	002 IgG	034 Care Start		002 002			002 - Positive	13	13	13	100
2021-664	003 SARS COV 2	021 VITROS SAR		002 002			002 - Positive	13	13	13	100

## Participant Summary

### MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
554 COVID-19 CORONAVIRUS											
2021-665	001	IgM	001	CORONACHEC	001	001	001 - Negative	16	16	16	100
2021-665	001	IgM	002	HEALGEN	001	001	001 - Negative	285	285	285	100
2021-665	001	IgM	007	RIGHT SIGN	001	001	001 - Negative	91	91	91	100
2021-665	001	IgM	016	CLARITY	001	001	001 - Negative	43	43	43	100
2021-665	001	IgM	030	FASTEP	001	001	001 - Negative	18	18	18	100
2021-665	001	IgM	033	BIOHIT SAR	001	001	001 - Negative	10	10	10	100
2021-665	001	IgM	034	Care Start	001	001	001 - Negative	13	13	13	100
2021-665	002	IgG	001	CORONACHEC	001	001	001 - Negative	16	16	16	100
2021-665	002	IgG	002	HEALGEN	001	001	001 - Negative	285	285	285	100
2021-665	002	IgG	007	RIGHT SIGN	001	001	001 - Negative	91	91	91	100
2021-665	002	IgG	016	CLARITY	001	001	001 - Negative	42	42	42	100
2021-665	002	IgG	030	FASTEP	001	001	001 - Negative	18	18	18	100
2021-665	002	IgG	033	BIOHIT SAR	001	001	001 - Negative	10	10	10	100
2021-665	002	IgG	034	Care Start	001	001	001 - Negative	13	13	13	100
2021-665	003	SARS COV 2	021	VITROS SAR	001	001	001 - Negative	13	13	13	100

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
109 SYPHILLIS SERO.-QUAL												
2021-711	002	RPR	028	BIOKIT-RPR	001	001	001 001	001 - NR Non Rea	69	69	69	100
2021-711	002	RPR	030	IMMUNOSTIC	001	001	001 001	001 - NR Non Rea	145	145	145	100
2021-711	002	RPR	033	ASI	001	001	001 001	001 - NR Non Rea	397	397	397	100
2021-711	002	RPR	040	TECO DIAG.	001	001	001 001	001 - NR Non Rea	59	59	59	100
2021-711	002	RPR	053	AIM RPR	001	001	001 001	001 - NR Non Rea	55	55	55	100
2021-711	002	RPR	057	Germain La	001	001	001 001	001 - NR Non Rea	19	19	19	100

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY									
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
109 SYPHILLIS SERO.-QUAL																	
2021-712	002	RPR	028	BIOKIT-RPR			004	004	004	004			004 - React RPR	69	69	69	100
2021-712	002	RPR	030	IMMUNOSTIC			004	004	004	004			004 - React RPR	145	145	145	100
2021-712	002	RPR	033	ASI			004	004	004	004			004 - React RPR	397	397	397	100
2021-712	002	RPR	040	TECO DIAG.			004	004	004	004			004 - React RPR	59	59	59	100
2021-712	002	RPR	053	AIM RPR			004	004	004	004			004 - React RPR	55	55	55	100
2021-712	002	RPR	057	Germain La			004	004	004	004			004 - React RPR	19	19	19	100
2021-713	002	RPR	028	BIOKIT-RPR			001	001	001	001			001 - NR Non Rea	69	69	69	100
2021-713	002	RPR	030	IMMUNOSTIC			001	001	001	001			001 - NR Non Rea	145	145	145	100
2021-713	002	RPR	033	ASI			001	001	001	001			001 - NR Non Rea	397	397	397	100
2021-713	002	RPR	040	TECO DIAG.			001	001	001	001			001 - NR Non Rea	59	59	59	100
2021-713	002	RPR	053	AIM RPR			001	001	001	001			001 - NR Non Rea	55	55	55	100
2021-713	002	RPR	057	Germain La			001	001	001	001			001 - NR Non Rea	19	19	19	100
2021-714	002	RPR	028	BIOKIT-RPR			004	004	004	004			004 - React RPR	69	69	69	100
2021-714	002	RPR	030	IMMUNOSTIC			004	004	004	004			004 - React RPR	145	145	145	100
2021-714	002	RPR	033	ASI			004	004	004	004			004 - React RPR	397	397	397	100
2021-714	002	RPR	040	TECO DIAG.			004	004	004	004			004 - React RPR	59	59	59	100
2021-714	002	RPR	053	AIM RPR			004	004	004	004			004 - React RPR	55	55	55	100
2021-714	002	RPR	057	Germain La			004	004	004	004			004 - React RPR	19	19	19	100
2021-715	002	RPR	028	BIOKIT-RPR			004	004	004	004			004 - React RPR	69	69	69	100
2021-715	002	RPR	030	IMMUNOSTIC			004	004	004	004			004 - React RPR	145	145	145	100
2021-715	002	RPR	033	ASI			004	004	004	004			004 - React RPR	397	397	397	100
2021-715	002	RPR	040	TECO DIAG.			004	004	004	004			004 - React RPR	59	59	59	100
2021-715	002	RPR	053	AIM RPR			004	004	004	004			004 - React RPR	55	55	55	100
2021-715	002	RPR	057	Germain La			004	004	004	004			004 - React RPR	19	19	19	100

**Participant Summary**

**MONTH 9 OF YEAR 2021**

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
110 SYPHILLIS SERO-QUANT											
2021-711	002 RPR	028 BLOKIT-RPR		001 001	001 001		001 - NR	53	53	53	100
2021-711	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR	58	58	58	100
2021-711	002 RPR	033 ASI		001 001	001 001		001 - NR	176	176	176	100
2021-711	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR	12	12	12	100
2021-712	002 RPR	028 BLOKIT-RPR		004 006	004 006		004 - R- 2 DILS.	53	15	53	100
2021-712	002 RPR	028 BLOKIT-RPR		004 006	004 006		005 - R- 4 DILS.	53	35	53	100
2021-712	002 RPR	028 BLOKIT-RPR		004 006	004 006		006 - R- 8 DILS.	53	3	53	100
2021-712	002 RPR	030 IMMUNOSTIC		004 006	004 006		003 - R- 1 DIL.	58	1	57	0
2021-712	002 RPR	030 IMMUNOSTIC		004 006	004 006		004 - R- 2 DILS.	58	11	57	100
2021-712	002 RPR	030 IMMUNOSTIC		004 006	004 006		005 - R- 4 DILS.	58	44	57	100
2021-712	002 RPR	030 IMMUNOSTIC		004 006	004 006		006 - R- 8 DILS.	58	2	57	100
2021-712	002 RPR	033 ASI		004 006	004 006		004 - R- 2 DILS.	176	48	176	100
2021-712	002 RPR	033 ASI		004 006	004 006		005 - R- 4 DILS.	176	120	176	100
2021-712	002 RPR	033 ASI		004 006	004 006		006 - R- 8 DILS.	176	8	176	100
2021-712	002 RPR	040 TECO DIAG.		003 005	004 006		004 - R- 2 DILS.	12	8	12	100
2021-712	002 RPR	040 TECO DIAG.		003 005	004 006		005 - R- 4 DILS.	12	4	12	100
2021-713	002 RPR	028 BLOKIT-RPR		001 001	001 001		001 - NR	53	53	53	100
2021-713	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR	58	58	58	100
2021-713	002 RPR	033 ASI		001 001	001 001		001 - NR	176	176	176	100
2021-713	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR	12	12	12	100
2021-714	002 RPR	028 BLOKIT-RPR		004 006	004 006		004 - R- 2 DILS.	53	18	53	100
2021-714	002 RPR	028 BLOKIT-RPR		004 006	004 006		005 - R- 4 DILS.	53	32	53	100
2021-714	002 RPR	028 BLOKIT-RPR		004 006	004 006		006 - R- 8 DILS.	53	3	53	100
2021-714	002 RPR	030 IMMUNOSTIC		004 006	004 006		004 - R- 2 DILS.	58	12	58	100
2021-714	002 RPR	030 IMMUNOSTIC		004 006	004 006		005 - R- 4 DILS.	58	44	58	100



## Participant Summary

### MONTH 9 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 130 SYPHILIS SEROLOGY

GRP: 250 SYPHILIS SEROLOGY

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-714	002 RPR	030 IMMUNOSTIC		004 006	004 006		006 - R- 8 DILS.	58	2	58	100
2021-714	002 RPR	033 ASI		004 006	004 006		003 - R- 1 DIL.	176	1	175	0
2021-714	002 RPR	033 ASI		004 006	004 006		004 - R- 2 DILS.	176	35	175	100
2021-714	002 RPR	033 ASI		004 006	004 006		005 - R- 4 DILS.	176	131	175	100
2021-714	002 RPR	033 ASI		004 006	004 006		006 - R- 8 DILS.	176	9	175	100
2021-714	002 RPR	040 TECO DIAG.		003 005	004 006		003 - R- 1 DIL.	12	1	12	100
2021-714	002 RPR	040 TECO DIAG.		003 005	004 006		004 - R- 2 DILS.	12	7	12	100
2021-714	002 RPR	040 TECO DIAG.		003 005	004 006		005 - R- 4 DILS.	12	4	12	100
110 SYPHILLIS SERO-QUANT											
2021-715	002 RPR	028 BLOKIT-RPR		005 007	005 007		004 - R- 2 DILS.	53	2	51	0
2021-715	002 RPR	028 BLOKIT-RPR		005 007	005 007		005 - R- 4 DILS.	53	15	51	100
2021-715	002 RPR	028 BLOKIT-RPR		005 007	005 007		006 - R- 8 DILS.	53	34	51	100
2021-715	002 RPR	028 BLOKIT-RPR		005 007	005 007		007 - R- 16 DILS	53	2	51	100
2021-715	002 RPR	030 IMMUNOSTIC		005 007	005 007		004 - R- 2 DILS.	58	4	54	0
2021-715	002 RPR	030 IMMUNOSTIC		005 007	005 007		005 - R- 4 DILS.	58	19	54	100
2021-715	002 RPR	030 IMMUNOSTIC		005 007	005 007		006 - R- 8 DILS.	58	35	54	100
2021-715	002 RPR	033 ASI		005 007	005 007		004 - R- 2 DILS.	176	1	175	0
2021-715	002 RPR	033 ASI		005 007	005 007		005 - R- 4 DILS.	176	57	175	100
2021-715	002 RPR	033 ASI		005 007	005 007		006 - R- 8 DILS.	176	116	175	100
2021-715	002 RPR	033 ASI		005 007	005 007		007 - R- 16 DILS	176	2	175	100
2021-715	002 RPR	040 TECO DIAG.		004 006	005 007		004 - R- 2 DILS.	12	1	12	100
2021-715	002 RPR	040 TECO DIAG.		004 006	005 007		005 - R- 4 DILS.	12	7	12	100
2021-715	002 RPR	040 TECO DIAG.		004 006	005 007		006 - R- 8 DILS.	12	4	12	100

### Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 260 BACTER. -GRAM STAIN			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
111 BACTER. GRAM STAIN											
2021-611				002 002	002 002		002 - GRAM POS.	66	66	66	100
2021-612				001 001	001 001		001 - GRAM NEG.	66	66	66	100
2021-613				001 001	001 001		001 - GRAM NEG.	66	65	65	100
2021-613				001 001	001 001		002 - GRAM POS.	66	1	65	0
2021-614				002 002	002 002		002 - GRAM POS.	66	66	66	100
2021-615				002 002	002 002		002 - GRAM POS.	66	66	66	100
112 BACTER. -IDENTIFICAT.											
SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	250 - NO N. gon.	42	1	14	0
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	120 - Neisse. sp	42	7	14	100
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	121 - Neisse. go	42	7	14	100
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	258 - g -bacilli	42	1	14	0
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	259 - g+ cocci	42	1	14	0
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	260 - g - cocci	42	23	14	0
2021-631	001	AEROBIC	001	CONVENCIO.	121	121	191 - Strep. sp.	42	2	14	0
2021-631	001	AEROBIC	034	MICROSCAN	121	121	121 - Neisse. go	18	8	8	100
2021-631	001	AEROBIC	034	MICROSCAN	121	121	207 - Strep. mi.	18	6	8	0
2021-631	001	AEROBIC	034	MICROSCAN	121	121	260 - g - cocci	18	3	8	0
2021-631	001	AEROBIC	034	MICROSCAN	121	121	191 - Strep. sp.	18	1	8	0

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-631	001	AEROBIC	046	MICROSCAN	121	121	250 - NO N. gon.	16	1	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	121 - Neisse. go	16	1	1	100
2021-631	001	AEROBIC	046	MICROSCAN	121	121	207 - Strep. mi.	16	8	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	259 - g+ cocci	16	2	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	188 - Staphy. ep	16	1	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	191 - Strep. sp.	16	1	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	194 - Strp a noD	16	1	1	0
2021-631	001	AEROBIC	046	MICROSCAN	121	121	205 - Strep. bo.	16	1	1	0
2021-631	001	AEROBIC	049	VITEK 2	121	121	121 - Neisse. go	17	9	9	100
2021-631	001	AEROBIC	049	VITEK 2	121	121	207 - Strep. mi.	17	4	9	0
2021-631	001	AEROBIC	049	VITEK 2	121	121	260 - g - cocci	17	4	9	0
112 BACTER.-IDENTIFICAT.											
2021-632	001	AEROBIC	001	CONVENCIO.	188	188	259 - g+ cocci	19	1	16	0
2021-632	001	AEROBIC	001	CONVENCIO.	188	188	188 - Staphy. ep	19	3	16	100
2021-632	001	AEROBIC	001	CONVENCIO.	195	195	195 - Strp bet A	19	7	16	100
2021-632	001	AEROBIC	001	CONVENCIO.	207	207	207 - Strep. mi.	19	1	16	100
2021-632	001	AEROBIC	001	CONVENCIO.	207	207	184 - Staphy. sp	19	1	16	0
2021-632	001	AEROBIC	001	CONVENCIO.	207	207	185 - Staphy. co	19	1	16	0
2021-632	001	AEROBIC	001	CONVENCIO.	207	207	191 - Strep. sp.	19	5	16	100
2021-632	001	AEROBIC	032	MICROSCAN	188	188	188 - Staphy. ep	55	21	53	100
2021-632	001	AEROBIC	032	MICROSCAN	195	195	195 - Strp bet A	55	25	53	100
2021-632	001	AEROBIC	032	MICROSCAN	207	207	207 - Strep. mi.	55	6	53	100
2021-632	001	AEROBIC	032	MICROSCAN	207	207	191 - Strep. sp.	55	1	53	100
2021-632	001	AEROBIC	032	MICROSCAN	207	207	193 - Strep ap v	55	1	53	0
2021-632	001	AEROBIC	032	MICROSCAN	207	207	199 - Strep gr C	55	1	53	0
2021-632	001	AEROBIC	034	MICROSCAN	188	188	259 - g+ cocci	61	2	57	0
2021-632	001	AEROBIC	034	MICROSCAN	188	188	188 - Staphy. ep	61	22	57	100
2021-632	001	AEROBIC	034	MICROSCAN	195	195	187 - Staphy. au	61	1	57	0
2021-632	001	AEROBIC	034	MICROSCAN	195	195	195 - Strp bet A	61	22	57	100
2021-632	001	AEROBIC	034	MICROSCAN	207	207	207 - Strep. mi.	61	13	57	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-632	001 AEROBIC	034 MICROSCAN			207 207		205 - Strep. bo.	61	1	57	0
2021-632	001 AEROBIC	042 VITEK(AMS)			188 188		188 - Staphy. ep	19	7	19	100
2021-632	001 AEROBIC	042 VITEK(AMS)			195 195		195 - Strp bet A	19	8	19	100
2021-632	001 AEROBIC	042 VITEK(AMS)			207 207		207 - Strep. mi.	19	2	19	100
2021-632	001 AEROBIC	042 VITEK(AMS)			207 207		191 - Strep. sp.	19	2	19	100
2021-632	001 AEROBIC	046 MICROSCAN			188 188		188 - Staphy. ep	78	30	75	100
2021-632	001 AEROBIC	046 MICROSCAN			195 195		195 - Strp bet A	78	32	75	100
2021-632	001 AEROBIC	046 MICROSCAN			207 207		207 - Strep. mi.	78	13	75	100
2021-632	001 AEROBIC	046 MICROSCAN			207 207		193 - Strep ap v	78	1	75	0
2021-632	001 AEROBIC	046 MICROSCAN			207 207		194 - Strp a noD	78	1	75	0
2021-632	001 AEROBIC	046 MICROSCAN			207 207		197 - Strp noABD	78	1	75	0
2021-632	001 AEROBIC	049 VITEK 2			188 188		188 - Staphy. ep	102	35	101	100
2021-632	001 AEROBIC	049 VITEK 2			195 195		195 - Strp bet A	102	37	101	100
2021-632	001 AEROBIC	049 VITEK 2			207 207		207 - Strep. mi.	102	25	101	100
2021-632	001 AEROBIC	049 VITEK 2			207 207		191 - Strep. sp.	102	4	101	100
2021-632	001 AEROBIC	049 VITEK 2			207 207		198 - Strep. aga	102	1	101	0
112 BACTER.-IDENTIFICAT.											
2021-633	001 AEROBIC	001 CONVENCIO.			068 068		068 - Escher. co	11	2	11	100
2021-633	001 AEROBIC	001 CONVENCIO.			198 198		191 - Strep. sp.	11	1	11	100
2021-633	001 AEROBIC	001 CONVENCIO.			198 198		198 - Strep. aga	11	8	11	100
2021-633	001 AEROBIC	032 MICROSCAN			068 068		068 - Escher. co	49	24	49	100
2021-633	001 AEROBIC	032 MICROSCAN			198 198		191 - Strep. sp.	49	2	49	100
2021-633	001 AEROBIC	032 MICROSCAN			198 198		198 - Strep. aga	49	23	49	100
2021-633	001 AEROBIC	034 MICROSCAN			068 068		068 - Escher. co	53	26	49	100
2021-633	001 AEROBIC	034 MICROSCAN			198 198		207 - Strep. mi.	53	4	49	0
2021-633	001 AEROBIC	034 MICROSCAN			198 198		198 - Strep. aga	53	23	49	100
2021-633	001 AEROBIC	042 VITEK(AMS)			198 198		207 - Strep. mi.	11	1	10	0
2021-633	001 AEROBIC	042 VITEK(AMS)			198 198		191 - Strep. sp.	11	2	10	100
2021-633	001 AEROBIC	042 VITEK(AMS)			198 198		198 - Strep. aga	11	8	10	100
2021-633	001 AEROBIC	046 MICROSCAN			068 068		068 - Escher. co	71	35	67	100

### Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-633	001 AEROBIC	046 MICROSCAN			198 198		207 - Strep. mi.	71	3	67	0
2021-633	001 AEROBIC	046 MICROSCAN			198 198		194 - Strp a noD	71	1	67	0
2021-633	001 AEROBIC	046 MICROSCAN			198 198		198 - Strep. aga	71	32	67	100
2021-633	001 AEROBIC	049 VITEK 2			068 068		068 - Escher. co	84	38	80	100
2021-633	001 AEROBIC	049 VITEK 2			198 198		207 - Strep. mi.	84	4	80	0
2021-633	001 AEROBIC	049 VITEK 2			198 198		191 - Strep. sp.	84	5	80	100
2021-633	001 AEROBIC	049 VITEK 2			198 198		198 - Strep. aga	84	37	80	100
112 BACTER.-IDENTIFICAT.											
2021-634	001 AEROBIC	032 MICROSCAN			070 070		070 - Entero. ae	25	25	25	100
2021-634	001 AEROBIC	034 MICROSCAN			070 070		069 - Entero. sp	26	1	26	100
2021-634	001 AEROBIC	034 MICROSCAN			070 070		070 - Entero. ae	26	25	26	100
2021-634	001 AEROBIC	046 MICROSCAN			070 070		070 - Entero. ae	35	35	35	100
2021-634	001 AEROBIC	049 VITEK 2			070 070		070 - Entero. ae	38	38	38	100
2021-635	001 AEROBIC	032 MICROSCAN			137 137		137 - Proteu. mi	25	25	25	100
2021-635	001 AEROBIC	034 MICROSCAN			137 137		137 - Proteu. mi	26	26	26	100
2021-635	001 AEROBIC	046 MICROSCAN			137 137		137 - Proteu. mi	35	35	35	100
2021-635	001 AEROBIC	049 VITEK 2			137 137		137 - Proteu. mi	38	38	38	100
SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
113 BACT.ANTIM. SUSCEPT.											
2021-634	004 MICRO SCAN	001 AMIKACIN	000 Not apply		001 001		001 - SUSCEPTIB.	63	63	63	100
2021-634	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		004 004		004 - RESISTANT	31	29	29	100
2021-634	004 MICRO SCAN	003 AMPICILLIN	000 Not apply		004 004		001 - SUSCEPTIB.	31	2	29	0
2021-634	004 MICRO SCAN	007 AZTREONAM	000 Not apply		001 001		001 - SUSCEPTIB.	15	15	15	100
2021-634	004 MICRO SCAN	020 CEFTAZIDIM	000 Not apply		001 001		001 - SUSCEPTIB.	17	17	17	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-634	004	MICRO SCAN	022	CEFTRIAXON	000	Not apply	001 001	001 - SUSCEPTIB.	12	12	12	100
2021-634	004	MICRO SCAN	029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	69	69	69	100
2021-634	004	MICRO SCAN	037	GENTAMICIN	000	Not apply	001 001	001 - SUSCEPTIB.	58	58	58	100
2021-634	004	MICRO SCAN	050	NITROFURAN	000	Not apply	004 004	004 - RESISTANT	23	22	22	100
2021-634	004	MICRO SCAN	050	NITROFURAN	000	Not apply	004 004	001 - SUSCEPTIB.	23	1	22	0
2021-634	004	MICRO SCAN	062	TETRACYCLI	000	Not apply	001 001	001 - SUSCEPTIB.	14	14	14	100
2021-634	004	MICRO SCAN	065	TOBRAMYCIN	000	Not apply	001 001	001 - SUSCEPTIB.	52	52	52	100
2021-634	004	MICRO SCAN	066	TRIMETHOPR	000	Not apply	001 001	001 - SUSCEPTIB.	20	20	20	100
2021-634	004	MICRO SCAN	067	TRIME/SULF	000	Not apply	001 001	001 - SUSCEPTIB.	67	67	67	100
2021-634	004	MICRO SCAN	072	CEFEPIME	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	38	100
2021-634	004	MICRO SCAN	073	PIPER/TAZO	000	Not apply	001 001	001 - SUSCEPTIB.	22	22	22	100
2021-634	004	MICRO SCAN	080	LEVOFLOXA.	000	Not apply	001 001	001 - SUSCEPTIB.	47	47	47	100
2021-634	004	MICRO SCAN	096	ERTAPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	23	23	23	100
2021-634	006	VITEK SYS.	001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	18	18	18	100
2021-634	006	VITEK SYS.	010	CEFAZOLIN	000	Not apply	004 004	004 - RESISTANT	18	17	17	100
2021-634	006	VITEK SYS.	010	CEFAZOLIN	000	Not apply	004 004	001 - SUSCEPTIB.	18	1	17	0
2021-634	006	VITEK SYS.	020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	16	16	16	100
2021-634	006	VITEK SYS.	022	CEFTRIAXON	000	Not apply	001 001	001 - SUSCEPTIB.	15	15	15	100
2021-634	006	VITEK SYS.	029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	44	44	44	100
2021-634	006	VITEK SYS.	037	GENTAMICIN	000	Not apply	001 001	001 - SUSCEPTIB.	34	34	34	100
2021-634	006	VITEK SYS.	038	IMIPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	13	13	13	100
2021-634	006	VITEK SYS.	050	NITROFURAN	000	Not apply	004 004	004 - RESISTANT	12	10	10	100
2021-634	006	VITEK SYS.	050	NITROFURAN	000	Not apply	004 004	002 - INTERMEDI.	12	2	10	0
2021-634	006	VITEK SYS.	065	TOBRAMYCIN	000	Not apply	001 001	001 - SUSCEPTIB.	24	24	24	100
2021-634	006	VITEK SYS.	067	TRIME/SULF	000	Not apply	001 001	001 - SUSCEPTIB.	39	39	39	100
2021-634	006	VITEK SYS.	072	CEFEPIME	000	Not apply	001 001	001 - SUSCEPTIB.	31	31	31	100
2021-634	006	VITEK SYS.	073	PIPER/TAZO	000	Not apply	001 001	001 - SUSCEPTIB.	27	27	27	100
2021-634	006	VITEK SYS.	080	LEVOFLOXA.	000	Not apply	001 001	001 - SUSCEPTIB.	16	16	16	100
2021-634	006	VITEK SYS.	096	ERTAPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	24	24	24	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
113 BACT.ANTIM. SUSCEPT.											
2021-635	004	MICRO SCAN 001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	58	58	58	100
2021-635	004	MICRO SCAN 003	AMPICILLIN	000	Not apply	001 001	001 - SUSCEPTIB.	55	55	55	100
2021-635	004	MICRO SCAN 004	AMPIC/SULB	000	Not apply	001 001	001 - SUSCEPTIB.	22	22	22	100
2021-635	004	MICRO SCAN 007	AZTREONAM	000	Not apply	001 001	001 - SUSCEPTIB.	12	12	12	100
2021-635	004	MICRO SCAN 010	CEFAZOLIN	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	38	100
2021-635	004	MICRO SCAN 020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	17	17	17	100
2021-635	004	MICRO SCAN 022	CEFTRIAXON	000	Not apply	001 001	001 - SUSCEPTIB.	10	10	10	100
2021-635	004	MICRO SCAN 029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	68	68	68	100
2021-635	004	MICRO SCAN 037	GENTAMICIN	000	Not apply	001 001	001 - SUSCEPTIB.	58	58	58	100
2021-635	004	MICRO SCAN 050	NITROFURAN	000	Not apply	004 004	004 - RESISTANT	33	31	31	100
2021-635	004	MICRO SCAN 050	NITROFURAN	000	Not apply	004 004	001 - SUSCEPTIB.	33	2	31	0
2021-635	004	MICRO SCAN 065	TOBRAMYCIN	000	Not apply	001 001	001 - SUSCEPTIB.	48	48	48	100
2021-635	004	MICRO SCAN 067	TRIME/SULF	000	Not apply	001 001	004 - RESISTANT	63	1	62	0
2021-635	004	MICRO SCAN 067	TRIME/SULF	000	Not apply	001 001	001 - SUSCEPTIB.	63	62	62	100
2021-635	004	MICRO SCAN 072	CEFEPIME	000	Not apply	001 001	001 - SUSCEPTIB.	27	27	27	100
2021-635	004	MICRO SCAN 073	PIPER/TAZO	000	Not apply	001 001	001 - SUSCEPTIB.	16	16	16	100
2021-635	004	MICRO SCAN 080	LEVOFLOXA.	000	Not apply	001 001	001 - SUSCEPTIB.	46	46	46	100
2021-635	004	MICRO SCAN 087	MEROPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	10	10	10	100
2021-635	004	MICRO SCAN 096	ERTAPENEM	000	Not apply	001 001	001 - SUSCEPTIB.	21	21	21	100
2021-635	006	VITEK SYS. 001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	18	18	17	100
2021-635	006	VITEK SYS. 001	AMIKACIN	000	Not apply	001 001	001 - SUSCEPTIB.	18	18	17	0
2021-635	006	VITEK SYS. 003	AMPICILLIN	000	Not apply	001 001	001 - SUSCEPTIB.	35	35	35	100
2021-635	006	VITEK SYS. 004	AMPIC/SULB	000	Not apply	001 001	001 - SUSCEPTIB.	15	15	15	100
2021-635	006	VITEK SYS. 010	CEFAZOLIN	000	Not apply	001 001	001 - SUSCEPTIB.	24	24	24	100
2021-635	006	VITEK SYS. 020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	14	14	13	100
2021-635	006	VITEK SYS. 020	CEFTAZIDIM	000	Not apply	001 001	001 - SUSCEPTIB.	14	14	13	0
2021-635	006	VITEK SYS. 022	CEFTRIAXON	000	Not apply	001 001	001 - SUSCEPTIB.	14	14	14	100
2021-635	006	VITEK SYS. 029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	37	100
2021-635	006	VITEK SYS. 029	CIPROFLOXA	000	Not apply	001 001	001 - SUSCEPTIB.	38	38	37	0

### Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 280 BACT. ANTIMIC. SUSC.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-635	006 VITEK SYS.	037 GENTAMICIN	000 Not apply	001	001		001 - SUSCEPTIB.	32	32	31	100	
2021-635	006 VITEK SYS.	037 GENTAMICIN	000 Not apply	001	001		001 - SUSCEPTIB.	32	32	31	0	
2021-635	006 VITEK SYS.	038 IMIPENEM	000 Not apply	001	001		001 - SUSCEPTIB.	12	12	12	100	
2021-635	006 VITEK SYS.	050 NITROFURAN	000 Not apply	004	004		004 - RESISTANT	30	30	30	100	
2021-635	006 VITEK SYS.	065 TOBRAMYCIN	000 Not apply	001	001		001 - SUSCEPTIB.	20	20	20	100	
2021-635	006 VITEK SYS.	067 TRIME/SULF	000 Not apply	001	001		001 - SUSCEPTIB.	38	38	37	100	
2021-635	006 VITEK SYS.	067 TRIME/SULF	000 Not apply	001	001		001 - SUSCEPTIB.	38	38	37	0	
2021-635	006 VITEK SYS.	072 CEFEPIME	000 Not apply	001	001		001 - SUSCEPTIB.	28	28	28	100	
2021-635	006 VITEK SYS.	073 PIPER/TAZO	000 Not apply	001	001		001 - SUSCEPTIB.	24	24	24	100	
2021-635	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply	001	001		001 - SUSCEPTIB.	17	17	17	100	
2021-635	006 VITEK SYS.	096 ERTAPENEM	000 Not apply	001	001		001 - SUSCEPTIB.	21	21	21	100	

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 263 BACTER.- DIRECT ANT.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
147 BACTER.DIRECT ANTIG.												
2021-621				001	001	001	001 - NEGATIVE	10	10	10	100	
2021-622				002	002	002	002 - POSITIVE	10	10	10	100	
2021-623				002	002	002	002 - POSITIVE	10	10	10	100	
2021-624				001	001	001	001 - NEGATIVE	10	10	10	100	
2021-625				001	001	001	001 - NEGATIVE	10	10	10	100	



## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 170 PARASITOLOGY

GRP: 350 ID OF PARASITES

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
120 IDENT OF PARASITES											
2021-701	001	E. histoly	002	CYSTS		001 001	002 - FEW	38	16	38	100
2021-701	001	E. histoly	002	CYSTS		001 001	003 - MODERATE	38	19	38	100
2021-701	001	E. histoly	002	CYSTS		001 001	004 - MANY	38	3	38	100
2021-701	003	E. coli	002	CYSTS		002 004	006 - PAR NOT RE	82	3	79	0
2021-701	003	E. coli	002	CYSTS		002 004	002 - FEW	82	30	79	100
2021-701	003	E. coli	002	CYSTS		002 004	003 - MODERATE	82	31	79	100
2021-701	003	E. coli	002	CYSTS		002 004	004 - MANY	82	18	79	100
2021-701	004	Endo. nana	002	CYSTS			002 - FEW	23	20	0	0
2021-701	004	Endo. nana	002	CYSTS			003 - MODERATE	23	3	0	0
2021-701	021	Taenia s/s	003	OVA OR EGG		002 004	002 - FEW	82	53	82	100
2021-701	021	Taenia s/s	003	OVA OR EGG		002 004	003 - MODERATE	82	22	82	100
2021-701	021	Taenia s/s	003	OVA OR EGG		002 004	004 - MANY	82	7	82	100
2021-702	031	Clonorchis	003	OVA OR EGG		002 004	006 - PAR NOT RE	82	34	48	0
2021-702	031	Clonorchis	003	OVA OR EGG		002 004	002 - FEW	82	46	48	100
2021-702	031	Clonorchis	003	OVA OR EGG		002 004	003 - MODERATE	82	2	48	100
2021-702	050	NPF	005	NOT APPLY			005 - NOT APPLY	31	31	0	0
2021-703	091	No trichro	005	NOT APPLY		001 001	004 - MANY	78	1	78	100
2021-703	091	No trichro	005	NOT APPLY		001 001	005 - NOT APPLY	78	77	78	100
2021-704	024	Hymenolep.	003	OVA OR EGG		002 004	002 - FEW	82	55	82	100
2021-704	024	Hymenolep.	003	OVA OR EGG		002 004	003 - MODERATE	82	24	82	100
2021-704	024	Hymenolep.	003	OVA OR EGG		002 004	004 - MANY	82	3	82	100
2021-705	024	Hymenolep.	003	OVA OR EGG		005 005	006 - PAR NOT RE	82	5	77	0
2021-705	024	Hymenolep.	003	OVA OR EGG		005 005	002 - FEW	82	3	77	100

### Participant Summary

#### MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 350 ID OF PARASITES			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-705	024 Hymenolep.	003 OVA OR EGG			005	005	005 - NOT APPLY	82	74	77	100
SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 360 PRES/ABS OF PARASITE			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
121 PRESENCE OR ABSENCE											
2021-701				002	002	002	002 - Presence	30	30	30	100
2021-702							001 - Absence	29	11	29	0
2021-702							002 - Presence	29	18	29	0
2021-703				003	003	003	003 - NO TRICHRO	28	28	28	100
2021-704				002	002	002	001 - Absence	29	1	28	0
2021-704				002	002	002	002 - Presence	29	28	28	100
2021-705				002	002	002	002 - Presence	29	29	29	100
SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
176 INFLUENZA A&B QUALI											
2021-641	001 INFLU. A	001 BINAX NOW		001	001	001	001 - NEGATIVE	11	11	11	100
2021-641	001 INFLU. A	003 Sofia CLIA		001	001	001	001 - NEGATIVE	101	101	101	100
2021-641	001 INFLU. A	004 Sekisui Os		001	001	001	001 - NEGATIVE	147	147	147	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A & B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score			
2021-641	001	INFLU. A	006	Quidel Flu	001	001	001	001	001	001 - NEGATIVE	43	43	43	100
2021-641	001	INFLU. A	007	BD Veritor	001	001	001	001	001	001 - NEGATIVE	24	24	24	100
2021-641	001	INFLU. A	008	Remel x/p	001	001	001	001	001	001 - NEGATIVE	19	19	19	100
2021-641	001	INFLU. A	010	OraSure	001	001	001	001	001	001 - NEGATIVE	159	159	159	100
2021-641	001	INFLU. A	012	BIOSIG A+B	001	001	001	001	001	001 - NEGATIVE	187	186	186	100
2021-641	001	INFLU. A	012	BIOSIG A+B	001	001	001	001	002	002 - POSITIVE	187	1	186	0
2021-641	002	INFLU. B	001	BINAX NOW	002	002	002	002	002	002 - POSITIVE	12	12	12	100
2021-641	002	INFLU. B	003	Sofia CLIA	002	002	002	002	002	002 - POSITIVE	101	101	101	100
2021-641	002	INFLU. B	004	Sekisui Os	002	002	002	002	002	002 - POSITIVE	146	146	146	100
2021-641	002	INFLU. B	006	Quidel Flu	002	002	002	002	002	002 - POSITIVE	43	43	43	100
2021-641	002	INFLU. B	007	BD Veritor	002	002	002	002	002	002 - POSITIVE	24	24	24	100
2021-641	002	INFLU. B	008	Remel x/p	002	002	002	002	002	002 - POSITIVE	19	19	19	100
2021-641	002	INFLU. B	010	OraSure	002	002	002	002	002	002 - POSITIVE	159	159	159	100
2021-641	002	INFLU. B	012	BIOSIG A+B	002	002	002	002	002	002 - POSITIVE	185	185	185	100
176	INFLUENZA A&B QUALI													
2021-642	001	INFLU. A	001	BINAX NOW	001	001	001	001	001	001 - NEGATIVE	11	11	11	100
2021-642	001	INFLU. A	003	Sofia CLIA	001	001	001	001	001	001 - NEGATIVE	102	101	101	100
2021-642	001	INFLU. A	003	Sofia CLIA	001	001	001	001	002	002 - POSITIVE	102	1	101	0
2021-642	001	INFLU. A	004	Sekisui Os	001	001	001	001	001	001 - NEGATIVE	147	147	147	100
2021-642	001	INFLU. A	006	Quidel Flu	001	001	001	001	001	001 - NEGATIVE	42	42	42	100
2021-642	001	INFLU. A	007	BD Veritor	001	001	001	001	001	001 - NEGATIVE	24	24	24	100
2021-642	001	INFLU. A	008	Remel x/p	001	001	001	001	001	001 - NEGATIVE	19	19	19	100
2021-642	001	INFLU. A	010	OraSure	001	001	001	001	001	001 - NEGATIVE	160	160	160	100
2021-642	001	INFLU. A	012	BIOSIG A+B	001	001	001	001	001	001 - NEGATIVE	186	186	186	100
2021-642	002	INFLU. B	001	BINAX NOW	002	002	002	002	001	001 - NEGATIVE	11	1	10	0
2021-642	002	INFLU. B	001	BINAX NOW	002	002	002	002	002	002 - POSITIVE	11	10	10	100
2021-642	002	INFLU. B	003	Sofia CLIA	002	002	002	002	002	002 - POSITIVE	100	100	100	100
2021-642	002	INFLU. B	004	Sekisui Os	002	002	002	002	002	002 - POSITIVE	147	147	147	100
2021-642	002	INFLU. B	006	Quidel Flu	002	002	002	002	002	002 - POSITIVE	43	43	43	100
2021-642	002	INFLU. B	007	BD Veritor	002	002	002	002	002	002 - POSITIVE	24	24	24	100

## Participant Summary

MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2021-642	002 INFLU. B	008 Remel x/p		002 002	002 002		002 - POSITIVE	19	19	19	100		
2021-642	002 INFLU. B	010 OraSure		002 002	002 002		002 - POSITIVE	159	159	159	100		
2021-642	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		001 - NEGATIVE	186	1	185	0		
2021-642	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	186	185	185	100		
176	INFLUENZA A&B QUALI												
2021-643	001 INFLU. A	001 BINAX NOW		002 002	002 002		002 - POSITIVE	11	11	11	100		
2021-643	001 INFLU. A	003 Sofia CLIA		002 002	002 002		001 - NEGATIVE	102	1	101	0		
2021-643	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	102	101	101	100		
2021-643	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	147	147	147	100		
2021-643	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	43	43	43	100		
2021-643	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100		
2021-643	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	19	19	19	100		
2021-643	001 INFLU. A	010 OraSure		002 002	002 002		001 - NEGATIVE	159	1	158	0		
2021-643	001 INFLU. A	010 OraSure		002 002	002 002		002 - POSITIVE	159	158	158	100		
2021-643	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	186	186	186	100		
2021-643	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	11	11	11	100		
2021-643	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100		
2021-643	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	147	147	147	100		
2021-643	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	43	43	43	100		
2021-643	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100		
2021-643	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	19	19	19	100		
2021-643	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	159	159	159	100		
2021-643	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	186	186	186	100		
2021-644	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	11	11	11	100		
2021-644	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100		
2021-644	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	148	148	148	100		
2021-644	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	43	43	43	100		
2021-644	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100		

## Participant Summary

### MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A &amp; B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-644	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	20	20	20	100
2021-644	001 INFLU. A	010 OraSure		001 001	001 001		001 - NEGATIVE	159	158	158	100
2021-644	001 INFLU. A	010 OraSure		001 001	001 001		002 - POSITIVE	159	1	158	0
2021-644	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	187	187	187	100
2021-644	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	11	11	11	100
2021-644	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100
2021-644	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	146	146	146	100
2021-644	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	43	43	43	100
2021-644	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-644	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-644	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	159	159	159	100
2021-644	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	185	185	185	100
176	INFLUENZA A&B QUALI										
2021-645	001 INFLU. A	001 BINAX NOW		002 002	002 002		002 - POSITIVE	11	11	11	100
2021-645	001 INFLU. A	003 Sofia CLIA		002 002	002 002		001 - NEGATIVE	102	1	101	0
2021-645	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	102	101	101	100
2021-645	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	147	147	147	100
2021-645	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	43	43	43	100
2021-645	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100
2021-645	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	19	19	19	100
2021-645	001 INFLU. A	010 OraSure		002 002	002 002		002 - POSITIVE	160	160	160	100
2021-645	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		001 - NEGATIVE	185	1	184	0
2021-645	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	185	184	184	100
2021-645	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	11	11	11	100
2021-645	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	100	100	100	100
2021-645	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	147	147	147	100
2021-645	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	43	43	43	100
2021-645	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-645	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	19	19	19	100
2021-645	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	159	159	159	100

### Participant Summary

#### MONTH 9 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 400 Influenza A & B				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-645	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	184	184	184	100	
SPE: 50 MICROBIOLOGY				SUB: 180 Virology				GRP: 450 RSV Syncytial Virus				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
555 Resp Syncytial Virus												
2021-641	001 ALERE ABBO			001 001			001 - Neg	26	26	26	100	
2021-641	002 SOFIA RSV			001 001			001 - Neg	24	24	24	100	
2021-642	001 ALERE ABBO			001 001			001 - Neg	26	26	26	100	
2021-642	002 SOFIA RSV			001 001			001 - Neg	24	24	24	100	
2021-643	001 ALERE ABBO			001 001			001 - Neg	26	26	26	100	
2021-643	002 SOFIA RSV			001 001			001 - Neg	24	24	24	100	
2021-644	001 ALERE ABBO			002 002			001 - Neg	26	1	25	0	
2021-644	001 ALERE ABBO			002 002			002 - Pos	26	25	25	100	
2021-644	002 SOFIA RSV			002 002			002 - Pos	24	24	24	100	
2021-645	001 ALERE ABBO			001 001			001 - Neg	26	26	26	100	
2021-645	002 SOFIA RSV			001 001			001 - Neg	24	24	24	100	