

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 225 HELICOBACTER PYLORI					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
155 HELICOBACTER PYLORI												
2021-381	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2021-382	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2021-383	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	22	22	22	100	
2021-384	017	INST.V.ALF		001 001	001 001		001 - NEGATIVE	22	22	22	100	
2021-385	017	INST.V.ALF		002 002	002 002		002 - POSITIVE	22	22	22	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 210 ANTINUCLEAR ANTIB.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
089 ANTIN. ANT.-ANA-QUAL												
2021-341	196	ENZYMATIC		002 002	002 002		002 - POS.OR ABN	10	10	10	100	
2021-342	196	ENZYMATIC		001 001	001 001		001 - NEG.OR NOR	10	10	10	100	
2021-343	196	ENZYMATIC		001 001	001 001		001 - NEG.OR NOR	10	10	10	100	
2021-344	196	ENZYMATIC		002 002	002 002		002 - POS.OR ABN	10	10	10	100	
2021-345	196	ENZYMATIC		001 001	001 001		001 - NEG.OR NOR	10	10	10	100	

SPEC: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
ANALYTE												
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH			
145 CRP - QUANTITATIVE												
			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}								
2021-351	34	DADE DIMEN		14	1.313	0.242	18.44	0.590	2.040			
2021-351	37	BECKMAN		12	1.215	0.080	6.55	0.980	1.450			
2021-351	39	ROCHE COBA		13	1.107	0.079	7.16	0.870	1.340			
2021-351	42	ASI ARLING		24	1.200	0.000	0.00	0.600	1.800			
2021-351	42	ASI ARLING		24	2.922	1.164	39.85	0	6.410			
2021-351	47	VITROS 560		16	1.200	0.000	0.00	0.600	1.800			
2021-351	47	VITROS 560		16	2.101	0.180	8.56	1.560	2.640			
2021-352	34	DADE DIMEN		14	0.070	0.056	80.62	0	0.240			
2021-352	37	BECKMAN		12	0.405	0.279	68.84	0	1.240			
2021-352	39	ROCHE COBA		13	0.025	0.017	68.31	0	0.080			
2021-352	42	ASI ARLING		24	0.183	0.276	151.19	0	1.010			

Participant Summary

MONTH 4 OF YEAR 2021

SPEC: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
2021-352	42	ASI ARLING			24	0.100	0.000	0.00	0	0.400
2021-352	47	VITROS 560			16	0.500	0.000	0.00	0.500	0.500
2021-352	47	VITROS 560			16	0.100	0.000	0.00	0	0.400
145 CRP - QUANTITATIVE			CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3 SD OR 10.00 {LOWER}						
2021-353	34	DADE DIMEN			14	3.050	0.296	9.72	2.160	3.940
2021-353	37	BECKMAN			12	2.638	0.115	4.37	2.290	2.980
2021-353	39	ROCHE COBA			13	2.674	0.282	10.55	1.830	3.520
2021-353	42	ASI ARLING			24	2.500	0.000	0.00	1.600	3.400
2021-353	42	ASI ARLING			24	2.357	0.817	34.67	0	4.810
2021-353	47	VITROS 560			16	2.500	0.000	0.00	1.600	3.400
2021-353	47	VITROS 560			16	3.937	0.267	6.78	3.140	4.740
2021-354	34	DADE DIMEN			14	1.292	0.189	14.65	0.720	1.860
2021-354	37	BECKMAN			12	1.200	0.068	5.65	1.000	1.400
2021-354	39	ROCHE COBA			13	1.117	0.085	7.61	0.860	1.370
2021-354	42	ASI ARLING			24	1.809	1.129	62.42	0	5.200
2021-354	42	ASI ARLING			24	1.230	1.000	81.30	0	3.930
2021-354	47	VITROS 560			16	2.135	0.173	8.10	1.620	2.650
2021-354	47	VITROS 560			16	1.230	1.000	81.30	0	3.930
2021-355	34	DADE DIMEN			14	0.072	0.055	76.92	0	0.240
2021-355	37	BECKMAN			12	0.413	0.286	69.18	0	1.270
2021-355	39	ROCHE COBA			13	0.024	0.017	68.53	0	0.070
2021-355	42	ASI ARLING			24	0.100	0.000	0.00	0	0.700
2021-355	42	ASI ARLING			24	0.183	0.276	151.19	0	1.010
2021-355	47	VITROS 560			16	0.100	0.000	0.00	0	0.700
2021-355	47	VITROS 560			16	0.500	0.000	0.00	0.500	0.500

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
091 ANTISTREP.O-QUALITA.												
2021-351	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	24	24	24	100	
2021-351	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	23	23	23	100	
2021-352	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	24	24	24	100	
2021-352	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	23	23	23	100	
2021-353	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	24	24	24	100	
2021-353	384 ASI A.SCT.			001 001	001 001		002 - POSITIVE	23	2	21	0	
2021-353	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	23	21	21	100	
2021-354	378 RHEUMAJET			001 001	001 001		001 - NEGATIVE	24	24	24	100	
2021-354	384 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	23	23	23	100	
2021-355	378 RHEUMAJET			002 002	002 002		001 - NEGATIVE	24	1	23	0	
2021-355	378 RHEUMAJET			002 002	002 002		002 - POSITIVE	24	23	23	100	
2021-355	384 ASI A.SCT.			002 002	002 002		002 - POSITIVE	23	23	23	100	
092 ANTISTREP.O-SEMI QUA												
2021-351	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	20	20	20	100	
2021-352	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	20	20	20	100	
2021-353	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	20	20	20	100	
2021-354	381 RHEUMAJET			014 014	014 014		014 - <100 IU/ml	20	20	20	100	
2021-355	381 RHEUMAJET			015 018	015 019		016 - 200 IU/ml	20	11	20	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-355	381 RHEUMAJET			015 018	015 019		017 - 400 IU/ml	20	9	20	100	
	103 INF. MONO- QUALITAT.											
2021-351	109 BIODIT MON			001 001	001 001		001 - NEGATIVE	79	79	79	100	
2021-351	393 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	126	126	126	100	
2021-351	394 TECO DIAG			001 001	001 001		001 - NEGATIVE	36	36	36	100	
2021-351	405 Acceava			001 001	001 001		001 - NEGATIVE	13	13	13	100	
2021-352	109 BIODIT MON			001 001	001 001		001 - NEGATIVE	79	79	79	100	
2021-352	393 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	126	126	126	100	
2021-352	394 TECO DIAG			001 001	001 001		001 - NEGATIVE	36	36	36	100	
2021-352	405 Acceava			001 001	001 001		001 - NEGATIVE	13	13	13	100	
2021-353	109 BIODIT MON			001 001	001 001		001 - NEGATIVE	79	79	79	100	
2021-353	393 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	126	126	126	100	
2021-353	394 TECO DIAG			001 001	001 001		001 - NEGATIVE	36	36	36	100	
2021-353	405 Acceava			001 001	001 001		001 - NEGATIVE	13	13	13	100	
2021-354	109 BIODIT MON			002 002	002 002		002 - POSITIVE	79	79	79	100	
2021-354	393 ASI A.SCT.			002 002	002 002		002 - POSITIVE	126	126	126	100	
2021-354	394 TECO DIAG			002 002	002 002		002 - POSITIVE	36	36	36	100	
2021-354	405 Acceava			002 002	002 002		002 - POSITIVE	13	13	13	100	
2021-355	109 BIODIT MON			001 001	001 001		001 - NEGATIVE	79	79	79	100	
2021-355	393 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	126	126	126	100	
2021-355	394 TECO DIAG			001 001	001 001		001 - NEGATIVE	36	36	36	100	
2021-355	405 Acceava			001 001	001 001		001 - NEGATIVE	13	13	13	100	
	105 RF FACTOR(RA)-QUALI.											
2021-351	041 BIODIT RHE			002 002	002 002		002 - POSITIVE	18	18	18	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-351	393	RHEUMAJET		002 002	002 002		002 - POSITIVE	80	80	80	100
2021-351	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	53	53	53	100
2021-351	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	126	126	126	100
2021-351	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	32	32	32	100
2021-351	408	STANBIO		002 002	002 002		002 - POSITIVE	22	22	22	100
105 RF FACTOR(RA)-QUALI.											
2021-352	041	BIOKIT RHE		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-352	393	RHEUMAJET		001 001	001 001		001 - NEGATIVE	80	80	80	100
2021-352	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	53	53	53	100
2021-352	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	126	126	126	100
2021-352	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	32	32	32	100
2021-352	408	STANBIO		001 001	001 001		001 - NEGATIVE	22	22	22	100
2021-353	041	BIOKIT RHE		001 001	001 001		002 - POSITIVE	18	3	15	0
2021-353	041	BIOKIT RHE		001 001	001 001		001 - NEGATIVE	18	15	15	100
2021-353	393	RHEUMAJET		001 001	001 001		001 - NEGATIVE	80	78	78	100
2021-353	393	RHEUMAJET		001 001	001 001		002 - POSITIVE	80	2	78	0
2021-353	400	TECO DIAG.		001 001	001 001		002 - POSITIVE	53	3	50	0
2021-353	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	53	50	50	100
2021-353	402	ASI A.SCT.			001 001		001 - NEGATIVE	126	83	83	100
2021-353	402	ASI A.SCT.			001 001		002 - POSITIVE	126	43	83	0
2021-353	406	AIM CRP RF			001 001		001 - NEGATIVE	32	13	13	100
2021-353	406	AIM CRP RF			001 001		002 - POSITIVE	32	19	13	0
2021-353	408	STANBIO		001 001	001 001		001 - NEGATIVE	22	22	22	100
2021-354	041	BIOKIT RHE		002 002	002 002		002 - POSITIVE	18	18	18	100
2021-354	393	RHEUMAJET		002 002	002 002		002 - POSITIVE	80	80	80	100
2021-354	400	TECO DIAG.		002 002	002 002		002 - POSITIVE	53	53	53	100
2021-354	402	ASI A.SCT.		002 002	002 002		002 - POSITIVE	126	126	126	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-354	406	AIM CRP RF		002 002	002 002		002 - POSITIVE	32	32	32	100	
2021-354	408	STANBIO		002 002	002 002		002 - POSITIVE	22	22	22	100	
105 RF FACTOR(RA)-QUALI.												
2021-355	041	BIOKIT RHE		001 001	001 001		001 - NEGATIVE	18	18	18	100	
2021-355	393	RHEUMAJET		001 001	001 001		001 - NEGATIVE	80	80	80	100	
2021-355	400	TECO DIAG.		001 001	001 001		001 - NEGATIVE	53	53	53	100	
2021-355	402	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	126	126	126	100	
2021-355	406	AIM CRP RF		001 001	001 001		001 - NEGATIVE	32	32	32	100	
2021-355	408	STANBIO		001 001	001 001		001 - NEGATIVE	22	22	22	100	
106 RF FAC(RA)-SEMI QUAN												
2021-351	393	RHEUMAJET		003 007	003 007		005 - 72-160 IU	42	31	41	100	
2021-351	393	RHEUMAJET		003 007	003 007		004 - 40-80 IU/m	42	9	41	100	
2021-351	393	RHEUMAJET		003 007	003 007		002 - 8-20 IU/ml	42	1	41	0	
2021-351	393	RHEUMAJET		003 007	003 007		006 - 136-320 IU	42	1	41	100	
2021-351	407	ASI SLIDE		004 008	003 007		005 - 72-160 IU	11	2	11	100	
2021-351	407	ASI SLIDE		004 008	003 007		007 - 264-640 IU	11	1	11	100	
2021-351	407	ASI SLIDE		004 008	003 007		003 - 24-40 IU/m	11	1	11	100	
2021-351	407	ASI SLIDE		004 008	003 007		006 - 136-320 IU	11	7	11	100	
2021-352	393	RHEUMAJET		001 001	001 001		001 - <8-20 IU/m	42	42	42	100	
2021-352	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	11	11	11	100	
2021-353	393	RHEUMAJET		001 001	001 001		002 - 8-20 IU/ml	42	1	41	0	
2021-353	393	RHEUMAJET		001 001	001 001		001 - <8-20 IU/m	42	41	41	100	
2021-353	407	ASI SLIDE		001 001	001 001		001 - <8-20 IU/m	11	9	9	100	
2021-353	407	ASI SLIDE		001 001	001 001		002 - 8-20 IU/ml	11	2	9	0	
2021-354	393	RHEUMAJET		002 006	002 006		004 - 40-80 IU/m	42	31	42	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY		SUB: 120 GENERAL				GRP: 220 ASOT, RA, MONO, CRP						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-354	393 RHEUMAJET			002 006	002 006		002 - 8-20 IU/ml	42	1	42	100	
2021-354	393 RHEUMAJET			002 006	002 006		003 - 24-40 IU/m	42	9	42	100	
2021-354	393 RHEUMAJET			002 006	002 006		005 - 72-160 IU	42	1	42	100	
2021-354	407 ASI SLIDE			003 007	002 006		006 - 136-320 IU	11	3	11	100	
2021-354	407 ASI SLIDE			003 007	002 006		003 - 24-40 IU/m	11	1	11	100	
2021-354	407 ASI SLIDE			003 007	002 006		005 - 72-160 IU	11	4	11	100	
2021-354	407 ASI SLIDE			003 007	002 006		004 - 40-80 IU/m	11	3	11	100	
106 RF FAC(RA)-SEMI QUAN												
2021-355	393 RHEUMAJET			001 001	001 001		001 - <8-20 IU/m	42	42	42	100	
2021-355	407 ASI SLIDE			001 001	001 001		001 - <8-20 IU/m	11	11	11	100	
144 CRP- QUALITATIVE												
2021-351	001 BIODIT RH.			002 002	002 002		002 - POSITIVE	14	14	14	100	
2021-351	026 ASI A.SCT.			002 002	002 002		002 - POSITIVE	130	130	130	100	
2021-351	030 TECO DIAG.			002 002	002 002		002 - POSITIVE	43	40	40	100	
2021-351	030 TECO DIAG.			002 002	002 002		001 - NEGATIVE	43	3	40	0	
2021-351	036 AIM CRP RF			002 002	002 002		002 - POSITIVE	41	41	41	100	
2021-351	041 STANBIO			002 002	002 002		002 - POSITIVE	11	11	11	100	
2021-352	001 BIODIT RH.			001 001	001 001		001 - NEGATIVE	14	14	14	100	
2021-352	026 ASI A.SCT.			001 001	001 001		001 - NEGATIVE	130	130	130	100	
2021-352	030 TECO DIAG.			001 001	001 001		001 - NEGATIVE	43	43	43	100	
2021-352	036 AIM CRP RF			001 001	001 001		001 - NEGATIVE	41	41	41	100	
2021-352	041 STANBIO			001 001	001 001		001 - NEGATIVE	11	11	11	100	
2021-353	001 BIODIT RH.			002 002	002 002		002 - POSITIVE	14	13	13	100	
2021-353	001 BIODIT RH.			002 002	002 002		001 - NEGATIVE	14	1	13	0	
2021-353	026 ASI A.SCT.			002 002	002 002		001 - NEGATIVE	130	3	127	0	
2021-353	026 ASI A.SCT.			002 002	002 002		002 - POSITIVE	130	127	127	100	
2021-353	030 TECO DIAG.			002 002	002 002		002 - POSITIVE	43	43	43	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 220 ASOT, RA, MONO, CRP

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-353	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	41	40	40	100
2021-353	036	AIM CRP RF		002 002	002 002		001 - NEGATIVE	41	1	40	0
2021-353	041	STANBIO		002 002	002 002		002 - POSITIVE	11	11	11	100
	144	CRP- QUALITATIVE									
2021-354	001	BIOKIT RH.		002 002	002 002		001 - NEGATIVE	14	1	13	0
2021-354	001	BIOKIT RH.		002 002	002 002		002 - POSITIVE	14	13	13	100
2021-354	026	ASI A.SCT.		002 002	002 002		002 - POSITIVE	130	128	128	100
2021-354	026	ASI A.SCT.		002 002	002 002		001 - NEGATIVE	130	2	128	0
2021-354	030	TECO DIAG.		002 002	002 002		001 - NEGATIVE	43	9	34	0
2021-354	030	TECO DIAG.		002 002	002 002		002 - POSITIVE	43	34	34	100
2021-354	036	AIM CRP RF		002 002	002 002		002 - POSITIVE	41	41	41	100
2021-354	041	STANBIO		002 002	002 002		001 - NEGATIVE	11	1	10	0
2021-354	041	STANBIO		002 002	002 002		002 - POSITIVE	11	10	10	100
2021-355	001	BIOKIT RH.		001 001	001 001		001 - NEGATIVE	14	14	14	100
2021-355	026	ASI A.SCT.		001 001	001 001		001 - NEGATIVE	130	122	122	100
2021-355	026	ASI A.SCT.		001 001	001 001		002 - POSITIVE	130	8	122	0
2021-355	030	TECO DIAG.		001 001	001 001		001 - NEGATIVE	43	42	42	100
2021-355	030	TECO DIAG.		001 001	001 001		002 - POSITIVE	43	1	42	0
2021-355	036	AIM CRP RF		001 001	001 001		001 - NEGATIVE	41	39	39	100
2021-355	036	AIM CRP RF		001 001	001 001		002 - POSITIVE	41	2	39	0
2021-355	041	STANBIO		001 001	001 001		001 - NEGATIVE	11	10	10	100
2021-355	041	STANBIO		001 001	001 001		002 - POSITIVE	11	1	10	0

Participant Summary
MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 120 GENERAL				GRP: 240 RUBELLA - QUALITAT.				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
107 RUBELLA-QUALITATIVE											
2021-331	335 BIOSIGN-BI			001 001	001 001		001 - NEG-NOT IM	11	11	11	100
2021-332	335 BIOSIGN-BI			001 001	001 001		001 - NEG-NOT IM	11	11	11	100
2021-333	335 BIOSIGN-BI			002 002	002 002		002 - POS.OR IM.	11	11	11	100
2021-334	335 BIOSIGN-BI			002 002	002 002		002 - POS.OR IM.	11	11	11	100
2021-335	335 BIOSIGN-BI			001 001	001 001		001 - NEG-NOT IM	11	11	11	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 247 MYCOPLASMA				
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
548 MYCOPLASMA												
2021-121	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	492	491	491	100
2021-121	001	IMMU. CARD		001	001	001	001	002 - Reactive	492	1	491	0
2021-122	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	492	3	489	0
2021-122	001	IMMU. CARD		002	002	002	002	002 - Reactive	492	489	489	100
2021-123	001	IMMU. CARD		002	002	002	002	002 - Reactive	492	490	490	100
2021-123	001	IMMU. CARD		002	002	002	002	001 - Non Reacti	492	2	490	0
2021-124	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	492	490	490	100
2021-124	001	IMMU. CARD		001	001	001	001	002 - Reactive	492	2	490	0
2021-125	001	IMMU. CARD		001	001	001	001	001 - Non Reacti	492	490	490	100
2021-125	001	IMMU. CARD		001	001	001	001	002 - Reactive	492	2	490	0

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
554 COVID-19 CORONAVIRUS											
2021-271	001	IgM	001 CORONACHEC	002	002		002 - Positive	20	20	20	100
2021-271	001	IgM	002 HEALGEN	002	002		001 - Negative	337	1	336	0
2021-271	001	IgM	002 HEALGEN	002	002		002 - Positive	337	336	336	100
2021-271	001	IgM	007 RIGHT SIGN	002	002		002 - Positive	78	78	78	100
2021-271	001	IgM	016 CLARITY	002	002		002 - Positive	44	44	44	100
2021-271	001	IgM	030 FASTEP	002	002		002 - Positive	13	13	13	100
2021-271	002	IgG	001 CORONACHEC	002	002		002 - Positive	20	20	20	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 248 COVID-19 CORONAVIRUS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-271	002 IgG	002 HEALGEN		002 002			001 - Negative	337	1	336	0
2021-271	002 IgG	002 HEALGEN		002 002			002 - Positive	337	336	336	100
2021-271	002 IgG	007 RIGHT SIGN		002 002			002 - Positive	78	78	78	100
2021-271	002 IgG	016 CLARITY		002 002			002 - Positive	43	43	43	100
2021-271	002 IgG	030 FASTEP		002 002			002 - Positive	13	13	13	100
2021-271	003 SARS COV 2	024 DIMENSION		002 002			002 - Positive	18	18	18	100
554 COVID-19 CORONAVIRUS											
2021-272	001 IgM	001 CORONACHEC		001 001			001 - Negative	20	20	20	100
2021-272	001 IgM	002 HEALGEN		001 001			001 - Negative	337	337	337	100
2021-272	001 IgM	007 RIGHT SIGN		001 001			001 - Negative	78	78	78	100
2021-272	001 IgM	016 CLARITY		001 001			001 - Negative	44	44	44	100
2021-272	001 IgM	030 FASTEP		001 001			001 - Negative	13	13	13	100
2021-272	002 IgG	001 CORONACHEC		001 001			001 - Negative	20	20	20	100
2021-272	002 IgG	002 HEALGEN		001 001			001 - Negative	337	337	337	100
2021-272	002 IgG	007 RIGHT SIGN		001 001			001 - Negative	78	78	78	100
2021-272	002 IgG	016 CLARITY		001 001			001 - Negative	44	44	44	100
2021-272	002 IgG	030 FASTEP		001 001			001 - Negative	13	13	13	100
2021-272	003 SARS COV 2	024 DIMENSION		001 001			001 - Negative	18	18	18	100
2021-273	001 IgM	001 CORONACHEC		002 002			002 - Positive	20	20	20	100
2021-273	001 IgM	002 HEALGEN		002 002			002 - Positive	337	337	337	100
2021-273	001 IgM	007 RIGHT SIGN		002 002			002 - Positive	78	78	78	100
2021-273	001 IgM	016 CLARITY		002 002			002 - Positive	44	44	44	100
2021-273	001 IgM	030 FASTEP		002 002			002 - Positive	13	13	13	100
2021-273	002 IgG	001 CORONACHEC		002 002			002 - Positive	20	20	20	100
2021-273	002 IgG	002 HEALGEN		002 002			002 - Positive	337	337	337	100
2021-273	002 IgG	007 RIGHT SIGN		002 002			002 - Positive	78	78	78	100
2021-273	002 IgG	016 CLARITY		002 002			002 - Positive	44	44	44	100
2021-273	002 IgG	030 FASTEP		002 002			002 - Positive	13	13	13	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY

SUB: 120 GENERAL

GRP: 248 COVID-19 CORONAVIRUS

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-273	003 SARS COV 2	024 DIMENSION		002 002			002 - Positive	18	18	18	100
554 COVID-19 CORONAVIRUS											
2021-274	001 IgM	001 CORONACHEC		002 002			002 - Positive	20	20	20	100
2021-274	001 IgM	002 HEALGEN		002 002			002 - Positive	338	337	337	100
2021-274	001 IgM	002 HEALGEN		002 002			001 - Negative	338	1	337	0
2021-274	001 IgM	007 RIGHT SIGN		002 002			002 - Positive	78	78	78	100
2021-274	001 IgM	016 CLARITY		002 002			001 - Negative	44	1	43	0
2021-274	001 IgM	016 CLARITY		002 002			002 - Positive	44	43	43	100
2021-274	001 IgM	030 FASTEP		002 002			002 - Positive	13	13	13	100
2021-274	002 IgG	001 CORONACHEC		002 002			002 - Positive	20	20	20	100
2021-274	002 IgG	002 HEALGEN		002 002			002 - Positive	335	334	334	100
2021-274	002 IgG	002 HEALGEN		002 002			001 - Negative	335	1	334	0
2021-274	002 IgG	007 RIGHT SIGN		002 002			002 - Positive	78	78	78	100
2021-274	002 IgG	016 CLARITY		002 002			001 - Negative	43	1	42	0
2021-274	002 IgG	016 CLARITY		002 002			002 - Positive	43	42	42	100
2021-274	002 IgG	030 FASTEP		002 002			002 - Positive	13	13	13	100
2021-274	003 SARS COV 2	024 DIMENSION		002 002			002 - Positive	18	18	18	100
2021-275	001 IgM	001 CORONACHEC		001 001			001 - Negative	20	20	20	100
2021-275	001 IgM	002 HEALGEN		001 001			001 - Negative	337	336	336	100
2021-275	001 IgM	002 HEALGEN		001 001			002 - Positive	337	1	336	0
2021-275	001 IgM	007 RIGHT SIGN		001 001			001 - Negative	78	78	78	100
2021-275	001 IgM	016 CLARITY		001 001			002 - Positive	44	1	43	0
2021-275	001 IgM	016 CLARITY		001 001			001 - Negative	44	43	43	100
2021-275	001 IgM	030 FASTEP		001 001			001 - Negative	13	13	13	100
2021-275	002 IgG	001 CORONACHEC		001 001			001 - Negative	20	20	20	100
2021-275	002 IgG	002 HEALGEN		001 001			001 - Negative	335	334	334	100
2021-275	002 IgG	002 HEALGEN		001 001			002 - Positive	335	1	334	0
2021-275	002 IgG	007 RIGHT SIGN		001 001			001 - Negative	78	78	78	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY				SUB: 120 GENERAL				GRP: 248 COVID-19 CORONAVIRUS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-275	002 IgG	016 CLARITY		001 001			001 - Negative	44	43	43	100
2021-275	002 IgG	016 CLARITY		001 001			002 - Positive	44	1	43	0
2021-275	002 IgG	030 FASTEP		001 001			001 - Negative	13	13	13	100
2021-275	003 SARS COV 2	024 DIMENSION		001 001			001 - Negative	18	18	18	100
SPE: 40 IMMUNOLOGY											
SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY							
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
109 SYPHILLIS SERO.-QUAL											
2021-321	002 RPR	028 BLOKIT-RPR		004 004	004 004		004 - React RPR	74	74	74	100
2021-321	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	154	154	154	100
2021-321	002 RPR	033 ASI		004 004	004 004		001 - NR Non Rea	385	1	384	0
2021-321	002 RPR	033 ASI		004 004	004 004		004 - React RPR	385	384	384	100
2021-321	002 RPR	040 TECO DIAG.		004 004	004 004		004 - React RPR	67	67	67	100
2021-321	002 RPR	053 AIM RPR		004 004	004 004		004 - React RPR	48	48	48	100
2021-321	002 RPR	057 Germain La		004 004	004 004		004 - React RPR	18	18	18	100
2021-322	002 RPR	028 BLOKIT-RPR		001 001	001 001		001 - NR Non Rea	74	74	74	100
2021-322	002 RPR	030 IMMUNOSTIC		001 001	001 001		001 - NR Non Rea	154	154	154	100
2021-322	002 RPR	033 ASI		001 001	001 001		004 - React RPR	385	1	384	0
2021-322	002 RPR	033 ASI		001 001	001 001		001 - NR Non Rea	385	384	384	100
2021-322	002 RPR	040 TECO DIAG.		001 001	001 001		001 - NR Non Rea	67	67	67	100
2021-322	002 RPR	053 AIM RPR		001 001	001 001		001 - NR Non Rea	48	48	48	100
2021-322	002 RPR	057 Germain La		001 001	001 001		001 - NR Non Rea	18	18	18	100
2021-323	002 RPR	028 BLOKIT-RPR		004 004	004 004		004 - React RPR	74	74	74	100
2021-323	002 RPR	030 IMMUNOSTIC		004 004	004 004		004 - React RPR	154	154	154	100
2021-323	002 RPR	033 ASI		004 004	004 004		004 - React RPR	385	385	385	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILLIS SEROLOGY				GRP: 250 SYPHILLIS SEROLOGY									
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-323	002	RPR	040	TECO DIAG.		004	004	004	004				004 - React RPR	67	67	67	100
2021-323	002	RPR	053	AIM RPR		004	004	004	004				004 - React RPR	48	48	48	100
2021-323	002	RPR	057	Germain La		004	004	004	004				004 - React RPR	18	18	18	100
109 SYPHILLIS SERO.-QUAL																	
2021-324	002	RPR	028	BIOKIT-RPR		004	004	004	004				004 - React RPR	74	74	74	100
2021-324	002	RPR	030	IMMUNOSTIC		004	004	004	004				004 - React RPR	154	154	154	100
2021-324	002	RPR	033	ASI		004	004	004	004				001 - NR Non Rea	385	1	384	0
2021-324	002	RPR	033	ASI		004	004	004	004				004 - React RPR	385	384	384	100
2021-324	002	RPR	040	TECO DIAG.		004	004	004	004				004 - React RPR	67	67	67	100
2021-324	002	RPR	053	AIM RPR		004	004	004	004				004 - React RPR	48	48	48	100
2021-324	002	RPR	057	Germain La		004	004	004	004				004 - React RPR	18	18	18	100
110 SYPHILLIS SERO-QUANT																	
2021-321	002	RPR	028	BIOKIT-RPR		004	006	004	006				004 - R- 2 DILS.	57	2	57	100
2021-321	002	RPR	028	BIOKIT-RPR		004	006	004	006				006 - R- 8 DILS.	57	13	57	100
2021-321	002	RPR	028	BIOKIT-RPR		004	006	004	006				005 - R- 4 DILS.	57	42	57	100
2021-321	002	RPR	030	IMMUNOSTIC		004	006	004	006				004 - R- 2 DILS.	61	2	61	100
2021-321	002	RPR	030	IMMUNOSTIC		004	006	004	006				005 - R- 4 DILS.	61	42	61	100
2021-321	002	RPR	030	IMMUNOSTIC		004	006	004	006				006 - R- 8 DILS.	61	17	61	100
2021-321	002	RPR	033	ASI		004	006	004	006				004 - R- 2 DILS.	163	1	162	100
2021-321	002	RPR	033	ASI		004	006	004	006				007 - R- 16 DILS	163	1	162	0
2021-321	002	RPR	033	ASI		004	006	004	006				006 - R- 8 DILS.	163	61	162	100
2021-321	002	RPR	033	ASI		004	006	004	006				005 - R- 4 DILS.	163	100	162	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY				SUB: 130 SYPHILLIS SEROLOGY				GRP: 250 SYPHILLIS SEROLOGY										
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE		PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-321	002	RPR	040	TECO DIAG.			004	006	004	006			005 - R- 4 DILS.	18	13	18	100	
2021-321	002	RPR	040	TECO DIAG.			004	006	004	006			006 - R- 8 DILS.	18	2	18	100	
2021-321	002	RPR	040	TECO DIAG.			004	006	004	006			004 - R- 2 DILS.	18	3	18	100	
110 SYPHILLIS SERO-QUANT																		
2021-322	002	RPR	028	BIOKIT-RPR			001	001	001	001			001 - NR	57	57	57	100	
2021-322	002	RPR	030	IMMUNOSTIC			001	001	001	001			001 - NR	61	61	61	100	
2021-322	002	RPR	033	ASI			001	001	001	001			001 - NR	163	163	163	100	
2021-322	002	RPR	040	TECO DIAG.			001	001	001	001			001 - NR	18	18	18	100	
2021-323	002	RPR	028	BIOKIT-RPR			003	005	004	006			005 - R- 4 DILS.	57	20	57	100	
2021-323	002	RPR	028	BIOKIT-RPR			003	005	004	006			004 - R- 2 DILS.	57	37	57	100	
2021-323	002	RPR	030	IMMUNOSTIC			003	005	004	006			005 - R- 4 DILS.	61	23	61	100	
2021-323	002	RPR	030	IMMUNOSTIC			003	005	004	006			006 - R- 8 DILS.	61	2	61	100	
2021-323	002	RPR	030	IMMUNOSTIC			003	005	004	006			004 - R- 2 DILS.	61	36	61	100	
2021-323	002	RPR	033	ASI			004	006	004	006			005 - R- 4 DILS.	163	81	163	100	
2021-323	002	RPR	033	ASI			004	006	004	006			006 - R- 8 DILS.	163	2	163	100	
2021-323	002	RPR	033	ASI			004	006	004	006			004 - R- 2 DILS.	163	80	163	100	
2021-323	002	RPR	040	TECO DIAG.			003	005	004	006			005 - R- 4 DILS.	18	2	18	100	
2021-323	002	RPR	040	TECO DIAG.			003	005	004	006			003 - R- 1 DIL.	18	1	18	100	
2021-323	002	RPR	040	TECO DIAG.			003	005	004	006			004 - R- 2 DILS.	18	15	18	100	
2021-324	002	RPR	028	BIOKIT-RPR			003	005	003	005			006 - R- 8 DILS.	57	1	56	0	
2021-324	002	RPR	028	BIOKIT-RPR			003	005	003	005			005 - R- 4 DILS.	57	13	56	100	
2021-324	002	RPR	028	BIOKIT-RPR			003	005	003	005			004 - R- 2 DILS.	57	43	56	100	
2021-324	002	RPR	030	IMMUNOSTIC			003	005	003	005			006 - R- 8 DILS.	61	2	59	0	
2021-324	002	RPR	030	IMMUNOSTIC			003	005	003	005			005 - R- 4 DILS.	61	24	59	100	
2021-324	002	RPR	030	IMMUNOSTIC			003	005	003	005			004 - R- 2 DILS.	61	35	59	100	
2021-324	002	RPR	033	ASI			004	006	003	005			004 - R- 2 DILS.	163	80	162	100	
2021-324	002	RPR	033	ASI			004	006	003	005			001 - NR	163	1	162	0	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 40 IMMUNOLOGY			SUB: 130 SYPHILIS SEROLOGY				GRP: 250 SYPHILIS SEROLOGY											
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH		REFEREE LOW/HIGH		COMMER LOW/HIGH		REPORTED VALUE		PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-324	002	RPR	033	ASI			004	006	003	005			005 - R- 4 DILS.	163	82	162	100	
2021-324	002	RPR	040	TECO DIAG.			003	005	003	005			005 - R- 4 DILS.	18	3	18	100	
2021-324	002	RPR	040	TECO DIAG.			003	005	003	005			004 - R- 2 DILS.	18	15	18	100	
110 SYPHILLIS SERO-QUANT																		
2021-325	002	RPR	028	BIOKIT-RPR			001	001	001	001			001 - NR	57	57	57	100	
2021-325	002	RPR	030	IMMUNOSTIC			001	001	001	001			001 - NR	61	61	61	100	
2021-325	002	RPR	033	ASI			001	001	001	001			001 - NR	163	162	162	100	
2021-325	002	RPR	033	ASI			001	001	001	001			005 - R- 4 DILS.	163	1	162	0	
2021-325	002	RPR	040	TECO DIAG.			001	001	001	001			001 - NR	18	18	18	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 260 BACTER. -GRAM STAIN			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
111 BACTER. GRAM STAIN											
2021-221				002 002	002 002		002 - GRAM POS.	64	64	64	100
2021-222				002 002	002 002		002 - GRAM POS.	64	64	64	100
2021-223				002 002	002 002		002 - GRAM POS.	64	64	64	100
2021-224				001 001	001 001		001 - GRAM NEG.	64	63	63	100
2021-224				001 001	001 001		002 - GRAM POS.	64	1	63	0
2021-225				001 001	001 001		001 - GRAM NEG.	64	63	63	100
2021-225				001 001	001 001		002 - GRAM POS.	64	1	63	0
112 BACTER.-IDENTIFICAT.											
SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-241	001 AEROBIC	032 MICROSCAN			183 183		181 - Shigel. fl	24	1	23	0
2021-241	001 AEROBIC	032 MICROSCAN			183 183		179 - Shigel. sp	24	4	23	100
2021-241	001 AEROBIC	032 MICROSCAN			183 183		183 - Shigel. so	24	19	23	100
2021-241	001 AEROBIC	034 MICROSCAN			183 183		183 - Shigel. so	25	24	25	100
2021-241	001 AEROBIC	034 MICROSCAN			183 183		179 - Shigel. sp	25	1	25	100
2021-241	001 AEROBIC	041 VITEK(AMS)			183 183		183 - Shigel. so	10	5	10	100
2021-241	001 AEROBIC	041 VITEK(AMS)			183 183		179 - Shigel. sp	10	5	10	100
2021-241	001 AEROBIC	046 MICROSCAN			183 183		179 - Shigel. sp	35	4	33	100
2021-241	001 AEROBIC	046 MICROSCAN			183 183		075 - Enter faec	35	1	33	0
2021-241	001 AEROBIC	046 MICROSCAN			183 183		183 - Shigel. so	35	29	33	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY			GRP: 270 BACTERIOLOGY IDENT.						
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-241	001	AEROBIC	046	MICROSCAN			183 183		258 - g -bacilli	35	1	33	0
2021-241	001	AEROBIC	049	VITEK 2			183 183		179 - Shigel. sp	38	8	38	100
2021-241	001	AEROBIC	049	VITEK 2			183 183		183 - Shigel. so	38	30	38	100
112 BACTER.-IDENTIFICAT.													
2021-242	001	AEROBIC	001	CONVENCIO.			188 188		185 - Staphy. co	18	3	11	0
2021-242	001	AEROBIC	001	CONVENCIO.			188 188		047 - Clotri. sp	18	1	11	0
2021-242	001	AEROBIC	001	CONVENCIO.			188 188		258 - g -bacilli	18	2	11	0
2021-242	001	AEROBIC	001	CONVENCIO.			188 188		188 - Staphy. ep	18	3	11	100
2021-242	001	AEROBIC	001	CONVENCIO.			187 187		187 - Staphy. au	18	8	11	100
2021-242	001	AEROBIC	001	CONVENCIO.			188 188		257 - g+ bacilli	18	1	11	0
2021-242	001	AEROBIC	032	MICROSCAN			188 188		188 - Staphy. ep	46	22	46	100
2021-242	001	AEROBIC	032	MICROSCAN			187 187		187 - Staphy. au	46	24	46	100
2021-242	001	AEROBIC	034	MICROSCAN			188 188		028 - Bacte. fra	49	1	47	0
2021-242	001	AEROBIC	034	MICROSCAN			188 188		258 - g -bacilli	49	1	47	0
2021-242	001	AEROBIC	034	MICROSCAN			188 188		188 - Staphy. ep	49	23	47	100
2021-242	001	AEROBIC	034	MICROSCAN			187 187		187 - Staphy. au	49	24	47	100
2021-242	001	AEROBIC	042	VITEK(AMS)			188 188		188 - Staphy. ep	18	9	18	100
2021-242	001	AEROBIC	042	VITEK(AMS)			187 187		187 - Staphy. au	18	9	18	100
2021-242	001	AEROBIC	046	MICROSCAN			187 187		187 - Staphy. au	68	35	68	100
2021-242	001	AEROBIC	046	MICROSCAN			188 188		188 - Staphy. ep	68	33	68	100
2021-242	001	AEROBIC	049	VITEK 2			188 188		188 - Staphy. ep	72	36	72	100
2021-242	001	AEROBIC	049	VITEK 2			187 187		187 - Staphy. au	72	36	72	100
2021-242	002	ANAEROBIC	001	CONVENCIO.			028 028		258 - g -bacilli	12	10	1	0
2021-242	002	ANAEROBIC	001	CONVENCIO.			028 028		028 - Bacte. fra	12	1	1	100
2021-242	002	ANAEROBIC	001	CONVENCIO.			028 028		257 - g+ bacilli	12	1	1	0
2021-242	002	ANAEROBIC	049	VITEK 2			028 028		023 - Bacte. sp.	10	1	7	100
2021-242	002	ANAEROBIC	049	VITEK 2			028 028		258 - g -bacilli	10	3	7	0
2021-242	002	ANAEROBIC	049	VITEK 2			028 028		028 - Bacte. fra	10	6	7	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 270 BACTERIOLOGY IDENT.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
112 BACTER.-IDENTIFICAT.											
2021-243	001	AEROBIC	001	CONVENCIO.	188	188	257 - g+ bacilli	16	1	14	0
2021-243	001	AEROBIC	001	CONVENCIO.	188	188	185 - Staphy. co	16	1	14	0
2021-243	001	AEROBIC	001	CONVENCIO.	198	198	198 - Strep. aga	16	8	14	100
2021-243	001	AEROBIC	001	CONVENCIO.	198	198	191 - Strep. sp.	16	1	14	100
2021-243	001	AEROBIC	001	CONVENCIO.	188	188	188 - Staphy. ep	16	5	14	100
2021-243	001	AEROBIC	032	MICROSCAN	188	188	188 - Staphy. ep	39	16	39	100
2021-243	001	AEROBIC	032	MICROSCAN	198	198	198 - Strep. aga	39	23	39	100
2021-243	001	AEROBIC	034	MICROSCAN	188	188	188 - Staphy. ep	38	15	38	100
2021-243	001	AEROBIC	034	MICROSCAN	198	198	198 - Strep. aga	38	23	38	100
2021-243	001	AEROBIC	042	VITEK(AMS)	198	198	198 - Strep. aga	14	9	14	100
2021-243	001	AEROBIC	042	VITEK(AMS)	188	188	188 - Staphy. ep	14	5	14	100
2021-243	001	AEROBIC	046	MICROSCAN	188	188	188 - Staphy. ep	61	25	60	100
2021-243	001	AEROBIC	046	MICROSCAN	188	188	187 - Staphy. au	61	1	60	0
2021-243	001	AEROBIC	046	MICROSCAN	198	198	198 - Strep. aga	61	35	60	100
2021-243	001	AEROBIC	049	VITEK 2	198	198	198 - Strep. aga	63	36	61	100
2021-243	001	AEROBIC	049	VITEK 2	198	198	195 - Strp bet A	63	1	61	0
2021-243	001	AEROBIC	049	VITEK 2	188	188	188 - Staphy. ep	63	25	61	100
2021-243	001	AEROBIC	049	VITEK 2	188	188	189 - Staphy. he	63	1	61	0
2021-243	002	ANAEROBIC	001	CONVENCIO.	051	051	051 - Clostri pe	10	1	2	100
2021-243	002	ANAEROBIC	001	CONVENCIO.	051	051	047 - Clotri. sp	10	1	2	100
2021-243	002	ANAEROBIC	001	CONVENCIO.	051	051	257 - g+ bacilli	10	8	2	0
2021-243	002	ANAEROBIC	049	VITEK 2	051	051	257 - g+ bacilli	11	2	9	0
2021-243	002	ANAEROBIC	049	VITEK 2	051	051	051 - Clostri pe	11	9	9	100
2021-244	001	AEROBIC	032	MICROSCAN	177	177	177 - Serrat. ma	25	25	25	100
2021-244	001	AEROBIC	034	MICROSCAN	177	177	177 - Serrat. ma	24	24	24	100
2021-244	001	AEROBIC	041	VITEK(AMS)	177	177	177 - Serrat. ma	10	10	10	100
2021-244	001	AEROBIC	046	MICROSCAN	177	177	177 - Serrat. ma	36	36	36	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 140 BACTERIOLOGY				GRP: 270 BACTERIOLOGY IDENT.						
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-244	001	AEROBIC	049	VITEK 2			177	177		177 - Serrat. ma	37	37	37	100
112 BACTER.-IDENTIFICAT.														
2021-245	001	AEROBIC	032	MICROSCAN			187	187		187 - Staphy. au	24	24	24	100
2021-245	001	AEROBIC	034	MICROSCAN			187	187		187 - Staphy. au	24	24	24	100
2021-245	001	AEROBIC	046	MICROSCAN			187	187		187 - Staphy. au	35	35	35	100
2021-245	001	AEROBIC	049	VITEK 2			187	187		187 - Staphy. au	36	36	36	100
SPE: 50 MICROBIOLOGY														
SUB: 140 BACTERIOLOGY														
GRP: 280 BACT. ANTIMIC. SUSC.														
TEST SAMPLE	LEVEL 1 RESULT CODE		LEVEL 2 RESULT CODE		LEVEL 3 RESULT CODE		PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
113 BACT.ANTIM. SUSCEPT.														
2021-244	004	MICRO SCAN	001	AMIKACIN	000	Not apply		001	001	001 - SUSCEPTIB.	55	55	55	100
2021-244	004	MICRO SCAN	003	AMPICILLIN	000	Not apply		004	004	001 - SUSCEPTIB.	31	1	30	0
2021-244	004	MICRO SCAN	003	AMPICILLIN	000	Not apply		004	004	004 - RESISTANT	31	30	30	100
2021-244	004	MICRO SCAN	007	AZTREONAM	000	Not apply		001	001	001 - SUSCEPTIB.	14	14	14	100
2021-244	004	MICRO SCAN	020	CEFTAZIDIM	000	Not apply		001	001	001 - SUSCEPTIB.	18	18	18	100
2021-244	004	MICRO SCAN	029	CIPROFLOXA	000	Not apply		001	001	001 - SUSCEPTIB.	66	66	66	100
2021-244	004	MICRO SCAN	037	GENTAMICIN	000	Not apply		001	001	001 - SUSCEPTIB.	58	58	58	100
2021-244	004	MICRO SCAN	050	NITROFURAN	000	Not apply		004	004	004 - RESISTANT	24	22	22	100
2021-244	004	MICRO SCAN	050	NITROFURAN	000	Not apply		004	004	001 - SUSCEPTIB.	24	2	22	0
2021-244	004	MICRO SCAN	065	TOBRAMYCIN	000	Not apply		001	001	001 - SUSCEPTIB.	54	54	54	100
2021-244	004	MICRO SCAN	066	TRIMETHOPR	000	Not apply		001	001	001 - SUSCEPTIB.	10	10	10	100
2021-244	004	MICRO SCAN	067	TRIME/SULF	000	Not apply		001	001	001 - SUSCEPTIB.	65	65	65	100
2021-244	004	MICRO SCAN	072	CEFEPIME	000	Not apply		001	001	001 - SUSCEPTIB.	39	39	39	100
2021-244	004	MICRO SCAN	073	PIPER/TAZO	000	Not apply		001	001	001 - SUSCEPTIB.	17	17	17	100
2021-244	004	MICRO SCAN	080	LEVOFLOXA.	000	Not apply		001	001	001 - SUSCEPTIB.	50	50	50	100
2021-244	004	MICRO SCAN	087	MEROPENEM	000	Not apply		001	001	001 - SUSCEPTIB.	11	11	11	100
2021-244	004	MICRO SCAN	096	ERTAPENEM	000	Not apply		001	001	001 - SUSCEPTIB.	20	20	20	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-244	006 VITEK SYS.	001 AMIKACIN	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	18	100
2021-244	006 VITEK SYS.	010 CEFAZOLIN	000 Not apply		004 004		004 - RESISTANT	19	19	19	100
2021-244	006 VITEK SYS.	020 CEFTAZIDIM	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	18	100
2021-244	006 VITEK SYS.	022 CEFTRIAXON	000 Not apply		001 001		001 - SUSCEPTIB.	14	14	14	100
2021-244	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	39	39	39	100
2021-244	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	35	35	35	100
2021-244	006 VITEK SYS.	050 NITROFURAN	000 Not apply		004 004		004 - RESISTANT	26	24	24	100
2021-244	006 VITEK SYS.	050 NITROFURAN	000 Not apply		004 004		001 - SUSCEPTIB.	26	2	24	0
2021-244	006 VITEK SYS.	065 TOBRAMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	29	29	29	100
2021-244	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	35	35	35	100
2021-244	006 VITEK SYS.	072 CEFEPIME	000 Not apply		001 001		001 - SUSCEPTIB.	39	39	39	100
2021-244	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	17	17	17	100
2021-244	006 VITEK SYS.	096 ERTAPENEM	000 Not apply		001 001		001 - SUSCEPTIB.	31	31	31	100
113 BACT.ANTIM. SUSCEPT.											
2021-245	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	65	65	64	100
2021-245	004 MICRO SCAN	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	65	65	64	0
2021-245	004 MICRO SCAN	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	23	23	23	100
2021-245	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	63	63	62	100
2021-245	004 MICRO SCAN	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	63	63	62	0
2021-245	004 MICRO SCAN	053 OXACILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	42	100
2021-245	004 MICRO SCAN	054 PENICILLIN	000 Not apply		001 001		004 - RESISTANT	26	1	25	0
2021-245	004 MICRO SCAN	054 PENICILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	26	25	25	100
2021-245	004 MICRO SCAN	057 RIFAMPIN	000 Not apply		001 001		001 - SUSCEPTIB.	16	16	16	100
2021-245	004 MICRO SCAN	062 TETRACYCLI	000 Not apply		001 001		001 - SUSCEPTIB.	35	35	35	100
2021-245	004 MICRO SCAN	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	76	76	76	100
2021-245	004 MICRO SCAN	068 VANCOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	50	50	50	100
2021-245	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	33	100
2021-245	004 MICRO SCAN	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	33	0
2021-245	004 MICRO SCAN	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	51	51	50	100
2021-245	004 MICRO SCAN	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	51	51	50	0

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 140 BACTERIOLOGY

GRP: 280 BACT. ANTIMIC. SUSC.

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-245	004 MICRO SCAN	097 DAPTOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	15	15	15	100
2021-245	004 MICRO SCAN	103 CEFTAROLIN	000 Not apply				001 - SUSCEPTIB.	11	11	0	0
2021-245	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	33	100
2021-245	006 VITEK SYS.	029 CIPROFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	34	34	33	0
2021-245	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	17	100
2021-245	006 VITEK SYS.	037 GENTAMICIN	000 Not apply		001 001		001 - SUSCEPTIB.	18	18	17	0
2021-245	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		002 - INTERMEDI.	21	2	19	0
2021-245	006 VITEK SYS.	050 NITROFURAN	000 Not apply		001 001		001 - SUSCEPTIB.	21	19	19	100
2021-245	006 VITEK SYS.	053 OXACILLIN	000 Not apply		001 001		004 - RESISTANT	26	1	24	0
2021-245	006 VITEK SYS.	053 OXACILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	26	25	24	100
2021-245	006 VITEK SYS.	053 OXACILLIN	000 Not apply		001 001		001 - SUSCEPTIB.	26	25	24	0
2021-245	006 VITEK SYS.	057 RIFAMPIN	000 Not apply		001 001		001 - SUSCEPTIB.	14	14	13	100
2021-245	006 VITEK SYS.	057 RIFAMPIN	000 Not apply		001 001		001 - SUSCEPTIB.	14	14	13	0
2021-245	006 VITEK SYS.	062 TETRACYCLI	000 Not apply		001 001		001 - SUSCEPTIB.	33	33	32	100
2021-245	006 VITEK SYS.	062 TETRACYCLI	000 Not apply		001 001		001 - SUSCEPTIB.	33	33	32	0
2021-245	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	100
2021-245	006 VITEK SYS.	067 TRIME/SULF	000 Not apply		001 001		001 - SUSCEPTIB.	42	42	41	0
2021-245	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	38	38	37	100
2021-245	006 VITEK SYS.	068 VANCOMYCIN	000 Not apply		001 001		001 - SUSCEPTIB.	38	38	37	0
2021-245	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	16	16	15	100
2021-245	006 VITEK SYS.	080 LEVOFLOXA.	000 Not apply		001 001		001 - SUSCEPTIB.	16	16	15	0
2021-245	006 VITEK SYS.	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	39	39	38	100
2021-245	006 VITEK SYS.	086 LINEZOLID	000 Not apply		001 001		001 - SUSCEPTIB.	39	39	38	0
2021-245	006 VITEK SYS.	095 MOXIFLOXA	000 Not apply		001 001		001 - SUSCEPTIB.	12	12	12	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY		SUB: 140 BACTERIOLOGY				GRP: 263 BACTER.- DIRECT ANT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
147 BACTER.DIRECT ANTIG.											
2021-231				002 002	002 002		002 - POSITIVE	10	10	10	100
2021-232				001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-233				001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-234				002 002	002 002		002 - POSITIVE	10	10	10	100
2021-235				002 002	002 002		002 - POSITIVE	10	10	10	100
SPE: 50 MICROBIOLOGY		SUB: 170 PARASITOLOGY				GRP: 350 ID OF PARASITES					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
120 IDENT OF PARASITES											
2021-311	050 NPF	005 NOT APPLY			005 005		005 - NOT APPLY	83	83	83	100
2021-312	015 Trichuris	003 OVA OR EGG			002 004		003 - MODERATE	83	37	81	100
2021-312	015 Trichuris	003 OVA OR EGG			002 004		002 - FEW	83	30	81	100
2021-312	015 Trichuris	003 OVA OR EGG			002 004		004 - MANY	83	14	81	100
2021-312	015 Trichuris	003 OVA OR EGG			002 004		006 - PAR NOT RE	83	2	81	0
2021-312	016 Ascaris l.	003 OVA OR EGG			002 004		002 - FEW	83	2	83	100
2021-312	016 Ascaris l.	003 OVA OR EGG			002 004		003 - MODERATE	83	12	83	100
2021-312	016 Ascaris l.	003 OVA OR EGG			002 004		004 - MANY	83	69	83	100
2021-313	009 Giardia l.	002 CYSTS			002 004		004 - MANY	83	48	81	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY			GRP: 350 ID OF PARASITES					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score	
2021-313	009	Giardia l.	002	CYSTS	002	004	006 - PAR NOT RE	83	2	81	0	
2021-313	009	Giardia l.	002	CYSTS	002	004	003 - MODERATE	83	27	81	100	
2021-313	009	Giardia l.	002	CYSTS	002	004	002 - FEW	83	6	81	100	
2021-313	024	Hymenolep.	003	OVA OR EGG	002	004	002 - FEW	83	32	81	100	
2021-313	024	Hymenolep.	003	OVA OR EGG	002	004	003 - MODERATE	83	39	81	100	
2021-313	024	Hymenolep.	003	OVA OR EGG	002	004	004 - MANY	83	10	81	100	
2021-313	024	Hymenolep.	003	OVA OR EGG	002	004	006 - PAR NOT RE	83	2	81	0	
120 IDENT OF PARASITES												
2021-314	050	NPF	005	NOT APPLY	005	005	005 - NOT APPLY	83	83	83	100	
2021-315	015	Trichuris	003	OVA OR EGG	005	005	002 - FEW	83	2	82	100	
2021-315	015	Trichuris	003	OVA OR EGG	005	005	005 - NOT APPLY	83	80	82	100	
2021-315	015	Trichuris	003	OVA OR EGG	005	005	006 - PAR NOT RE	83	1	82	0	
2021-315	016	Ascaris l.	003	OVA OR EGG	005	005	006 - PAR NOT RE	83	2	81	0	
2021-315	016	Ascaris l.	003	OVA OR EGG	005	005	002 - FEW	83	2	81	100	
2021-315	016	Ascaris l.	003	OVA OR EGG	005	005	005 - NOT APPLY	83	79	81	100	
2021-315	017	Hookworm	003	OVA OR EGG	005	005	005 - NOT APPLY	83	80	82	100	
2021-315	017	Hookworm	003	OVA OR EGG	005	005	006 - PAR NOT RE	83	1	82	0	
2021-315	017	Hookworm	003	OVA OR EGG	005	005	002 - FEW	83	2	82	100	

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY				SUB: 170 PARASITOLOGY				GRP: 360 PRES/ABS OF PARASITE			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
121 PRESENCE OR ABSENCE											
2021-311				001 001	001 001		001 - Absence	30	29	29	100
2021-311				001 001	001 001		002 - Presence	30	1	29	0
2021-312				002 002	002 002		002 - Presence	32	32	32	100
2021-313				002 002	002 002		002 - Presence	31	30	30	100
2021-313				002 002	002 002		001 - Absence	31	1	30	0
2021-314				001 001	001 001		001 - Absence	30	30	30	100
2021-315				002 002	002 002		002 - Presence	44	44	44	100
176 INFLUENZA A&B QUALI											
2021-251	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-251	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100
2021-251	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	144	144	100
2021-251	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-251	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-251	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-251	001 INFLU. A	010 OraSure		001 001	001 001		001 - NEGATIVE	167	167	167	100
2021-251	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	186	186	186	100
2021-251	002 INFLU. B	001 BINAX NOW		002 002	002 002		002 - POSITIVE	10	10	10	100
2021-251	002 INFLU. B	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	102	102	102	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A & B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-251	002 INFLU. B	004 Sekisui Os		002 002	002 002		002 - POSITIVE	144	143	143	100
2021-251	002 INFLU. B	004 Sekisui Os		002 002	002 002		001 - NEGATIVE	144	1	143	0
2021-251	002 INFLU. B	006 Quidel Flu		002 002	002 002		002 - POSITIVE	51	50	50	100
2021-251	002 INFLU. B	006 Quidel Flu		002 002	002 002		001 - NEGATIVE	51	1	50	0
2021-251	002 INFLU. B	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100
2021-251	002 INFLU. B	008 Remel x/p		002 002	002 002		002 - POSITIVE	18	18	18	100
2021-251	002 INFLU. B	010 OraSure		002 002	002 002		002 - POSITIVE	167	167	167	100
2021-251	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	185	185	185	100
176	INFLUENZA A&B QUALI										
2021-252	001 INFLU. A	001 BINAX NOW		002 002	002 002		002 - POSITIVE	10	10	10	100
2021-252	001 INFLU. A	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	102	102	102	100
2021-252	001 INFLU. A	004 Sekisui Os		002 002	002 002		002 - POSITIVE	144	144	144	100
2021-252	001 INFLU. A	006 Quidel Flu		002 002	002 002		002 - POSITIVE	51	50	50	100
2021-252	001 INFLU. A	006 Quidel Flu		002 002	002 002		001 - NEGATIVE	51	1	50	0
2021-252	001 INFLU. A	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100
2021-252	001 INFLU. A	008 Remel x/p		002 002	002 002		002 - POSITIVE	18	18	18	100
2021-252	001 INFLU. A	010 OraSure		002 002	002 002		002 - POSITIVE	167	167	167	100
2021-252	001 INFLU. A	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	186	186	186	100
2021-252	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-252	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	102	102	102	100
2021-252	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	144	144	100
2021-252	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	50	50	100
2021-252	002 INFLU. B	006 Quidel Flu		001 001	001 001		002 - POSITIVE	51	1	50	0
2021-252	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-252	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-252	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	167	167	167	100
2021-252	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	185	185	185	100
2021-253	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A & B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-253	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	103	102	102	100
2021-253	001 INFLU. A	003 Sofia CLIA		001 001	001 001		002 - POSITIVE	103	1	102	0
2021-253	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	144	144	100
2021-253	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-253	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-253	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-253	001 INFLU. A	010 OraSure		001 001	001 001		001 - NEGATIVE	166	166	166	100
2021-253	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	185	185	185	100
2021-253	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-253	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100
2021-253	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	144	144	100
2021-253	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-253	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-253	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-253	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	168	168	168	100
2021-253	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	184	184	184	100
176	INFLUENZA A&B QUALI										
2021-254	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-254	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	102	102	102	100
2021-254	001 INFLU. A	004 Sekisui Os		001 001	001 001		002 - POSITIVE	144	1	143	0
2021-254	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	143	143	100
2021-254	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-254	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-254	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-254	001 INFLU. A	010 OraSure		001 001	001 001		001 - NEGATIVE	167	167	167	100
2021-254	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		002 - POSITIVE	186	1	185	0
2021-254	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	186	185	185	100
2021-254	002 INFLU. B	001 BINAX NOW		002 002	002 002		002 - POSITIVE	10	10	10	100
2021-254	002 INFLU. B	003 Sofia CLIA		002 002	002 002		002 - POSITIVE	102	102	102	100
2021-254	002 INFLU. B	004 Sekisui Os		002 002	002 002		002 - POSITIVE	144	143	143	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY

SUB: 180 Virology

GRP: 400 Influenza A & B

TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
2021-254	002 INFLU. B	004 Sekisui Os		002 002	002 002		001 - NEGATIVE	144	1	143	0
2021-254	002 INFLU. B	006 Quidel Flu		002 002	002 002		002 - POSITIVE	51	51	51	100
2021-254	002 INFLU. B	007 BD Veritor		002 002	002 002		002 - POSITIVE	24	24	24	100
2021-254	002 INFLU. B	008 Remel x/p		002 002	002 002		002 - POSITIVE	18	18	18	100
2021-254	002 INFLU. B	010 OraSure		002 002	002 002		002 - POSITIVE	167	167	167	100
2021-254	002 INFLU. B	012 BIOSIG A+B		002 002	002 002		002 - POSITIVE	184	184	184	100
176	INFLUENZA A&B QUALI										
2021-255	001 INFLU. A	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-255	001 INFLU. A	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	103	103	103	100
2021-255	001 INFLU. A	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	144	144	100
2021-255	001 INFLU. A	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-255	001 INFLU. A	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-255	001 INFLU. A	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-255	001 INFLU. A	010 OraSure		001 001	001 001		001 - NEGATIVE	167	167	167	100
2021-255	001 INFLU. A	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	186	186	186	100
2021-255	002 INFLU. B	001 BINAX NOW		001 001	001 001		001 - NEGATIVE	10	10	10	100
2021-255	002 INFLU. B	003 Sofia CLIA		001 001	001 001		001 - NEGATIVE	101	101	101	100
2021-255	002 INFLU. B	004 Sekisui Os		001 001	001 001		001 - NEGATIVE	144	143	143	100
2021-255	002 INFLU. B	004 Sekisui Os		001 001	001 001		002 - POSITIVE	144	1	143	0
2021-255	002 INFLU. B	006 Quidel Flu		001 001	001 001		001 - NEGATIVE	51	51	51	100
2021-255	002 INFLU. B	007 BD Veritor		001 001	001 001		001 - NEGATIVE	24	24	24	100
2021-255	002 INFLU. B	008 Remel x/p		001 001	001 001		001 - NEGATIVE	18	18	18	100
2021-255	002 INFLU. B	010 OraSure		001 001	001 001		001 - NEGATIVE	167	167	167	100
2021-255	002 INFLU. B	012 BIOSIG A+B		001 001	001 001		001 - NEGATIVE	185	185	185	100

Participant Summary

MONTH 4 OF YEAR 2021

SPE: 50 MICROBIOLOGY		SUB: 180 Virology			GRP: 450 RSV Syncytial Virus						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
555 Resp Syncytial Virus											
2021-251	001	ALERE ABBO		001	001		001 - Neg	27	27	27	100
2021-251	002	SOFIA RSV		001	001		001 - Neg	21	21	21	100
2021-252	001	ALERE ABBO		001	001		001 - Neg	27	27	27	100
2021-252	002	SOFIA RSV		001	001		001 - Neg	21	21	21	100
2021-253	001	ALERE ABBO		002	002		002 - Pos	27	27	27	100
2021-253	002	SOFIA RSV		002	002		002 - Pos	21	21	21	100
2021-254	001	ALERE ABBO		001	001		001 - Neg	27	27	27	100
2021-254	002	SOFIA RSV		001	001		002 - Pos	21	1	20	0
2021-254	002	SOFIA RSV		001	001		001 - Neg	21	20	20	100
2021-255	001	ALERE ABBO		002	002		002 - Pos	27	27	27	100
2021-255	002	SOFIA RSV		002	002		002 - Pos	21	21	21	100