

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 140 COAGULATION (HEMAT.)							
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE		
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
080 FIBRINOGEN				CLIA TEST ID #: 825						EVALUATION CRITERIA: TARGET VALUE 20 %	
2024-941	73 ACL 8 TO E	87 HEMOSIL-C		11	528.818	35.941	6.80	423.000	635.000		
2024-941	79 CS-2500	67 DADE		17	430.294	34.441	8.00	344.000	516.000		
2024-942	73 ACL 8 TO E	87 HEMOSIL-C		11	146.273	8.147	5.57	117.000	176.000		
2024-942	79 CS-2500	67 DADE		17	120.235	9.515	7.91	96.000	144.000		
2024-943	73 ACL 8 TO E	87 HEMOSIL-C		11	298.727	13.122	4.39	239.000	358.000		
2024-943	79 CS-2500	67 DADE		17	241.294	17.891	7.41	193.000	290.000		
2024-944	73 ACL 8 TO E	87 HEMOSIL-C		11	291.000	15.527	5.34	233.000	349.000		
2024-944	79 CS-2500	67 DADE		17	238.706	19.164	8.03	191.000	286.000		
2024-945	73 ACL 8 TO E	87 HEMOSIL-C		11	306.273	9.166	2.99	245.000	368.000		
2024-945	79 CS-2500	67 DADE		17	243.000	29.708	12.23	194.000	292.000		
081 PART.THROM. TIME-PTT				CLIA TEST ID #: 835						EVALUATION CRITERIA: TARGET VALUE 15 %	
2024-941	62 DADE SYS.	35 DADE ACTIN		58	27.305	0.986	3.61	23.200	31.400		
2024-941	69 DIAG.STAGO	49 STA-PTT		18	28.371	0.684	2.41	24.100	32.600		
2024-941	73 ACL 8 TO E	56 PTT-HEMO		32	36.810	1.720	4.67	31.300	42.300		
2024-941	76 STAGO STAR	49 STA-PTT		22	28.000	1.337	4.78	23.800	32.200		
2024-941	76 STAGO STAR	49 STA-PTT		22	30.220	1.649	5.46	25.700	34.800		
2024-941	79 CS-2500	35 DADE ACTIN		17	27.241	0.732	2.69	23.200	31.300		
2024-942	62 DADE SYS.	35 DADE ACTIN		58	28.854	1.157	4.01	24.500	33.200		
2024-942	69 DIAG.STAGO	49 STA-PTT		18	33.712	0.946	2.81	28.700	38.800		
2024-942	73 ACL 8 TO E	56 PTT-HEMO		32	29.342	0.438	1.49	24.900	33.700		
2024-942	76 STAGO STAR	49 STA-PTT		22	30.280	2.128	7.03	25.700	34.800		
2024-942	76 STAGO STAR	49 STA-PTT		22	33.232	1.809	5.44	28.200	38.200		
2024-942	79 CS-2500	35 DADE ACTIN		17	29.335	1.030	3.51	24.900	33.700		

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 140 COAGULATION (HEMAT.)

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
081	PART.THROM. TIME-PTT	CLIA TEST ID #: 835	EVALUATION CRITERIA: TARGET VALUE 15 %						
2024-943	62 DADE SYS.	35 DADE ACTIN		58	42.475	1.935	4.56	36.100	48.800
2024-943	69 DIAG.STAGO	49 STA-PTT		18	43.753	1.131	2.59	37.200	50.300
2024-943	73 ACL 8 TO E	56 PTT-HEMO		32	44.232	0.824	1.86	37.600	50.900
2024-943	76 STAGO STAR	49 STA-PTT		22	39.240	1.624	4.14	33.400	45.100
2024-943	76 STAGO STAR	49 STA-PTT		22	42.595	2.901	6.81	36.200	49.000
2024-943	79 CS-2500	35 DADE ACTIN		17	44.406	2.074	4.67	37.700	51.100
2024-944	62 DADE SYS.	35 DADE ACTIN		58	38.528	1.949	5.06	32.700	44.300
2024-944	69 DIAG.STAGO	49 STA-PTT		18	39.882	1.116	2.80	33.900	45.900
2024-944	73 ACL 8 TO E	56 PTT-HEMO		32	38.400	0.600	1.56	32.600	44.200
2024-944	76 STAGO STAR	49 STA-PTT		22	38.932	3.124	8.03	33.100	44.800
2024-944	76 STAGO STAR	49 STA-PTT		22	36.480	0.983	2.69	31.000	42.000
2024-944	79 CS-2500	35 DADE ACTIN		17	39.506	1.585	4.01	33.600	45.400
2024-945	62 DADE SYS.	35 DADE ACTIN		58	29.722	1.229	4.13	25.300	34.200
2024-945	69 DIAG.STAGO	49 STA-PTT		18	30.712	0.891	2.90	26.100	35.300
2024-945	73 ACL 8 TO E	56 PTT-HEMO		32	30.135	0.685	2.27	25.600	34.700
2024-945	76 STAGO STAR	49 STA-PTT		22	30.400	1.049	3.45	25.800	35.000
2024-945	76 STAGO STAR	49 STA-PTT		22	29.345	1.467	5.00	24.900	33.700
2024-945	79 CS-2500	35 DADE ACTIN		17	30.329	1.263	4.17	25.800	34.900
082	PROTHROMBIN TIME (PT)	CLIA TEST ID #: 845							
2024-941	62 DADE SYS.	33 DADE INNOV		57	11.441	0.289	2.53	9.700	13.200
2024-941	73 ACL 8 TO E	41 HEMOSIL RE		28	10.800	0.214	1.98	9.200	12.400
2024-941	76 STAGO STAR	44 STA NEO CI		15	13.780	0.281	2.04	11.700	15.800
2024-941	79 CS-2500	33 DADE INNOV		18	11.711	0.221	1.89	10.000	13.500
2024-942	62 DADE SYS.	33 DADE INNOV		57	12.459	0.376	3.02	10.600	14.300
2024-942	73 ACL 8 TO E	41 HEMOSIL RE		28	13.359	0.436	3.26	11.400	15.400
2024-942	76 STAGO STAR	44 STA NEO CI		15	15.633	0.536	3.43	13.300	18.000

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 140 COAGULATION (HEMAT.)

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2024-942	79 CS-2500	33 DADE INNOV		18	12.939	0.116	0.90	11.000	14.900		
082 PROTHROMBIN TIME (PT)		CLIA TEST ID #: 845	EVALUATION CRITERIA: TARGET VALUE 15 %								
2024-943	62 DADE SYS.	33 DADE INNOV		57	26.060	1.153	4.42	22.200	30.000		
2024-943	73 ACL 8 TO E	41 HEMOSIL RE		28	31.593	1.647	5.21	26.900	36.300		
2024-943	76 STAGO STAR	44 STA NEO CI		15	36.860	1.739	4.72	31.300	42.400		
2024-943	79 CS-2500	33 DADE INNOV		18	26.133	0.514	1.97	22.200	30.100		
2024-944	62 DADE SYS.	33 DADE INNOV		57	19.021	0.660	3.47	16.200	21.900		
2024-944	73 ACL 8 TO E	41 HEMOSIL RE		28	22.363	1.111	4.97	19.000	25.700		
2024-944	76 STAGO STAR	44 STA NEO CI		15	27.513	1.094	3.97	23.400	31.600		
2024-944	79 CS-2500	33 DADE INNOV		18	19.328	0.256	1.32	16.400	22.200		
2024-945	62 DADE SYS.	33 DADE INNOV		57	10.298	0.270	2.62	8.800	11.800		
2024-945	73 ACL 8 TO E	41 HEMOSIL RE		28	11.233	0.260	2.31	9.500	12.900		
2024-945	76 STAGO STAR	44 STA NEO CI		15	13.120	0.423	3.22	11.200	15.100		
2024-945	79 CS-2500	33 DADE INNOV		18	10.865	0.076	0.70	9.200	12.500		

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPE: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 100 HEMA. CELL IDENTIFI.						
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
072 HEMA. CELL IDENTIFI.											
2024-891					036	036	033 - POLYCHROMA	914	1	897	0
2024-891					036	036	036 - BASOPHILIC	914	897	897	100
2024-891					036	036	007 - BASOPHIL,	914	15	897	0
2024-891					036	036	073 - RETICULOCY	914	1	897	0
2024-892					051	051	037 - HEMOGLOBIN	914	1	911	0
2024-892					051	051	049 - FRAGMENTED	914	1	911	0
2024-892					051	051	051 - SICKLE CEL	914	911	911	100
2024-892					051	051	052 - SPHEROCYTE	914	1	911	0
2024-893					008	008	005 - BAND (STAB	914	2	910	0
2024-893					008	008	007 - BASOPHIL,	914	2	910	0
2024-893					008	008	008 - EOSINOPHIL	914	910	910	100
2024-894					020	020	013 - LYMPHOCYTE	914	1	906	0
2024-894					020	020	014 - LYMPHOCYTE	914	1	906	0
2024-894					020	020	017 - MONOCYTE	914	1	906	0
2024-894					020	020	019 - PLATELETS,	914	3	906	0
2024-894					020	020	020 - GIANT PLA	914	906	906	100
2024-894					020	020	033 - POLYCHROMA	914	1	906	0
2024-894					020	020	007 - BASOPHIL,	914	1	906	0
2024-895					017	017	014 - LYMPHOCYTE	914	2	911	0
2024-895					017	017	017 - MONOCYTE	914	911	911	100
2024-895					017	017	020 - GIANT PLA	914	1	911	0

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 137 RETICULOCYTE COUNT						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
149	RETICULOCYTE COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE	3	SD OR 25 %	{GREATER}				
2024-921	7	M.B.MANUAL		11	11.490	2.606	22.68	3.700	19.300	
2024-921	24	Unicel DxH		18	8.218	0.419	5.10	6.200	10.300	
2024-921	25	SYSMEX XN		32	4.556	1.286	28.22	0.700	8.400	
2024-922	7	M.B.MANUAL		11	11.530	2.722	23.61	3.400	19.700	
2024-922	24	Unicel DxH		18	8.318	0.446	5.36	6.200	10.400	
2024-922	25	SYSMEX XN		32	4.447	1.287	28.95	0.600	8.300	
2024-923	7	M.B.MANUAL		11	1.520	0.511	33.65	0.000	3.100	
2024-923	24	Unicel DxH		18	0.772	0.073	9.46	0.600	1.000	
2024-923	25	SYSMEX XN		32	0.819	0.088	10.77	0.600	1.100	
2024-924	7	M.B.MANUAL		11	5.920	1.460	24.66	1.500	10.300	
2024-924	24	Unicel DxH		18	3.728	0.349	9.37	2.700	4.800	
2024-924	25	SYSMEX XN		32	2.006	0.574	28.60	0.300	3.700	
2024-925	7	M.B.MANUAL		11	1.420	0.340	23.94	0.400	2.400	
2024-925	24	Unicel DxH		18	0.728	0.080	11.04	0.500	1.000	
2024-925	25	SYSMEX XN		32	0.822	0.093	11.27	0.500	1.100	

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
153	SED. RATE (ESR)		EVALUATION CRITERIA: TARGET VALUE	3	SD OR 35 %	{GREATER}				
2024-931	4	SEDIPLAST		642	8.842	1.950	22.05	3.000	15.000	
2024-931	4	SEDIPLAST		642	11.125	3.100	27.86	2.000	20.000	
2024-931	10	SEDI RATE		67	9.769	1.187	12.15	6.000	13.000	

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 138 SED. RATE (ESR)							
2024-931	11	SEDIMAT-15		47	9.617	7.379	76.73	0	32.000		
2024-931	19	MINI CUBE		47	16.348	2.478	15.16	9.000	24.000		
153 SED. RATE (ESR)		CLIA TEST ID #: 0		EVALUATION CRITERIA: TARGET VALUE 3 SD OR 35 % {GREATER}							
2024-932	4	SEDIPLAST		642	48.250	11.020	22.84	15.000	81.000		
2024-932	4	SEDIPLAST		642	42.380	7.331	17.30	20.000	64.000		
2024-932	10	SEDI RATE		67	40.955	9.159	22.36	13.000	68.000		
2024-932	11	SEDIMAT-15		47	126.356	17.628	13.95	73.000	179.000		
2024-932	19	MINI CUBE		47	51.809	17.145	33.09	0.000	103.000		
2024-933	4	SEDIPLAST		642	9.000	1.955	21.73	3.000	15.000		
2024-933	4	SEDIPLAST		642	11.125	2.891	25.99	2.000	20.000		
2024-933	10	SEDI RATE		67	9.955	1.224	12.29	6.000	14.000		
2024-933	11	SEDIMAT-15		47	10.348	8.793	84.97	0	37.000		
2024-933	19	MINI CUBE		47	16.739	2.641	15.78	9.000	25.000		
2024-934	4	SEDIPLAST		642	55.500	11.597	20.90	21.000	90.000		
2024-934	4	SEDIPLAST		642	42.117	7.453	17.70	20.000	64.000		
2024-934	10	SEDI RATE		67	40.687	8.740	21.48	14.000	67.000		
2024-934	11	SEDIMAT-15		47	123.283	24.345	19.75	50.000	196.000		
2024-934	19	MINI CUBE		47	55.277	20.605	37.28	0	117.000		
2024-935	4	SEDIPLAST		642	42.134	6.647	15.78	22.000	62.000		
2024-935	4	SEDIPLAST		642	48.500	9.772	20.15	19.000	78.000		
2024-935	10	SEDI RATE		67	40.731	7.469	18.34	18.000	63.000		
2024-935	11	SEDIMAT-15		47	128.422	12.514	9.74	83.000	173.000		
2024-935	19	MINI CUBE		47	59.362	19.474	32.81	1.000	118.000		

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE											
SAMPLE	METHOD	SYST/INST	REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25	% OR 2.00	{GREATER}					
2024911X	1	SYSMEX CBC	1 NEUTROPHIL	25	XS SERIES	125	55.454	1.323	2.38	41.600	69.300
2024911X	1	SYSMEX CBC	2 LYMPHOCYTE	25	XS SERIES	125	20.078	0.932	4.64	15.100	25.100
2024911X	1	SYSMEX CBC	3 MONOCYTE	25	XS SERIES	125	1.257	0.285	22.69	0	3.300
2024911X	1	SYSMEX CBC	4 EOSINOPHIL	25	XS SERIES	125	14.263	0.532	3.73	10.700	17.800
2024911X	1	SYSMEX CBC	5 BASOPHIL	25	XS SERIES	125	8.948	0.778	8.69	6.700	11.200
2024911X	2	ACT PENTRA	1 NEUTROPHIL	15	COUL ACT 5	64	53.937	1.655	3.07	40.500	67.400
2024911X	2	ACT PENTRA	1 NEUTROPHIL	24	ABX PENTRA	32	52.406	1.118	2.13	39.300	65.500
2024911X	2	ACT PENTRA	2 LYMPHOCYTE	15	COUL ACT 5	64	39.914	1.587	3.98	29.900	49.900
2024911X	2	ACT PENTRA	2 LYMPHOCYTE	24	ABX PENTRA	33	38.463	1.120	2.91	28.800	48.100
2024911X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	64	1.844	0.405	21.98	0	3.800
2024911X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	32	1.969	0.369	18.77	0.000	4.000
2024911X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	64	4.352	0.335	7.70	2.400	6.400
2024911X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	31	4.326	0.278	6.43	2.300	6.300
2024911X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	64	56.525	1.228	2.17	42.400	70.700
2024911X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	32	2.828	0.045	1.59	0.800	4.800
2024911X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	17	72.929	0.477	0.65	54.700	91.200
2024911X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	17	15.831	0.342	2.16	11.900	19.800
2024911X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	17	7.665	0.389	5.08	5.700	9.700
2024911X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	17	0.400	0.137	34.30	0	2.400
2024911X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	0.424	0.139	32.93	0	2.400
2024911X	3	ADVIA 120	6 LUC	9	ADVIA-120	16	2.863	0.271	9.48	0.900	4.900
2024911X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	45	76.245	0.731	0.96	57.200	95.300
2024911X	4	AB- CELL-D	1 NEUTROPHIL	34	EMERALD 22	10	73.078	0.555	0.76	54.800	91.300
2024911X	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	43	76.412	0.675	0.88	57.300	95.500
2024911X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	45	15.058	1.020	6.78	11.300	18.800
2024911X	4	AB- CELL-D	2 LYMPHOCYTE	34	EMERALD 22	10	13.778	1.081	7.85	10.300	17.200
2024911X	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	43	14.663	0.691	4.72	11.000	18.300
2024911X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	45	3.743	0.475	12.69	1.700	5.700
2024911X	4	AB- CELL-D	3 MONOCYTE	34	EMERALD 22	10	9.500	0.973	10.24	7.100	11.900
2024911X	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	43	4.112	0.356	8.66	2.100	6.100

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2024911X	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	45	4.171	0.186	4.45	2.200	6.200			
2024911X	4 AB- CELL-D	4 EOSINOPHIL	34 EMERALD 22	10	3.333	0.704	21.12	1.300	5.300			
2024911X	4 AB- CELL-D	4 EOSINOPHIL	37 ABB.CEL RU	43	4.167	0.233	5.59	2.200	6.200			
2024911X	4 AB- CELL-D	5 BASOPHIL	8 CELL-DYN	45	0.722	0.327	45.31	0	2.700			
2024911X	4 AB- CELL-D	5 BASOPHIL	34 EMERALD 22	10	0.300	0.082	27.22	0	2.300			
2024911X	4 AB- CELL-D	5 BASOPHIL	37 ABB.CEL RU	43	0.590	0.189	31.96	0	2.600			
2024911X	5 BECTON D	1 NEUTROPHIL	29 B.C600,800	59	67.377	1.611	2.39	50.500	84.200			
2024911X	5 BECTON D	2 LYMPHOCYTE	29 B.C600,800	59	20.940	1.513	7.23	15.700	26.200			
2024911X	5 BECTON D	3 MONOCYTE	29 B.C600,800	59	6.816	0.820	12.03	4.800	8.800			
2024911X	5 BECTON D	4 EOSINOPHIL	29 B.C600,800	59	4.782	0.322	6.74	2.800	6.800			
2024911X	5 BECTON D	5 BASOPHIL	29 B.C600,800	59	0.000	0.000	0.00	0	2.000			
2024911X	6 MR Mindray	1 NEUTROPHIL	32 BC-5390	64	60.111	1.224	2.04	45.100	75.100			
2024911X	6 MR Mindray	2 LYMPHOCYTE	32 BC-5390	64	18.484	1.010	5.46	13.900	23.100			
2024911X	6 MR Mindray	3 MONOCYTE	32 BC-5390	64	8.156	1.106	13.57	6.100	10.200			
2024911X	6 MR Mindray	4 EOSINOPHIL	32 BC-5390	64	13.251	0.443	3.34	9.900	16.600			
2024911X	6 MR Mindray	5 BASOPHIL	32 BC-5390	64	32.248	0.187	0.58	24.200	40.300			
2024911X	7 DXH	1 NEUTROPHIL	36 BC DXH 500	96	48.502	5.177	10.67	36.400	60.600			
2024911X	7 DXH	2 LYMPHOCYTE	36 BC DXH 500	96	39.347	5.283	13.43	29.500	49.200			
2024911X	7 DXH	3 MONOCYTE	36 BC DXH 500	96	0.178	0.225	126.48	0	2.200			
2024911X	7 DXH	4 EOSINOPHIL	36 BC DXH 500	96	11.754	1.331	11.32	8.800	14.700			
2024911X	7 DXH	5 BASOPHIL	36 BC DXH 500	96	0.053	0.128	242.68	0	2.100			
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2024912X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	125	49.189	1.193	2.43	36.900	61.500			
2024912X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	125	28.488	1.082	3.80	21.400	35.600			
2024912X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	125	1.704	0.413	24.25	0	3.700			
2024912X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	125	12.900	0.584	4.53	9.700	16.100			
2024912X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	125	7.764	0.698	8.99	5.800	9.800			
2024912X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	64	73.071	1.210	1.66	54.800	91.300			
2024912X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	32	70.353	0.921	1.31	52.800	87.900			
2024912X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	64	19.838	1.161	5.85	14.900	24.800			
2024912X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	32	18.603	0.812	4.37	14.000	23.300			



**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2024912X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	64	1.924	0.505	26.26	0	3.900
2024912X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	32	1.981	0.370	18.69	0.000	4.000
2024912X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	64	5.048	0.458	9.08	3.000	7.000
2024912X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	32	5.069	0.292	5.76	3.100	7.100
2024912X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	64	80.321	1.400	1.74	60.200	100.400
2024912X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	32	3.984	0.045	1.12	2.000	6.000
2024912X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	17	43.238	1.928	4.46	32.400	54.000
2024912X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	17	36.144	2.824	7.81	27.100	45.200
2024912X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	17	13.881	1.162	8.37	10.400	17.400
2024912X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	17	0.356	0.154	43.22	0	2.400
2024912X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	1.076	0.298	27.69	0	3.100
2024912X	3	ADVIA 120	6 LUC	9	ADVIA-120	16	5.794	0.699	12.06	3.800	7.800
2024912X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	45	65.633	1.140	1.74	49.200	82.000
2024912X	4	AB- CELL-D	1 NEUTROPHIL	34	EMERALD 22	10	25.133	7.715	30.70	18.900	31.400
2024912X	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	43	64.984	1.653	2.54	48.700	81.200
2024912X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	45	24.036	1.663	6.92	18.000	30.000
2024912X	4	AB- CELL-D	2 LYMPHOCYTE	34	EMERALD 22	10	25.833	4.683	18.13	19.400	32.300
2024912X	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	43	24.765	2.428	9.81	18.600	31.000
2024912X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	45	5.480	0.763	13.93	3.500	7.500
2024912X	4	AB- CELL-D	3 MONOCYTE	34	EMERALD 22	10	20.722	3.160	15.25	15.500	25.900
2024912X	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	43	5.747	1.103	19.20	3.700	7.700
2024912X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	45	3.973	0.263	6.61	2.000	6.000
2024912X	4	AB- CELL-D	4 EOSINOPHIL	34	EMERALD 22	10	27.389	9.809	35.81	20.500	34.200
2024912X	4	AB- CELL-D	4 EOSINOPHIL	37	ABB.CEL RU	43	4.074	0.231	5.68	2.100	6.100
2024912X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	45	0.980	0.426	43.43	0	3.000
2024912X	4	AB- CELL-D	5 BASOPHIL	34	EMERALD 22	10	1.422	0.461	32.44	0	3.400
2024912X	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	43	0.419	0.309	73.69	0	2.400
2024912X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	59	60.821	0.935	1.54	45.600	76.000
2024912X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	59	25.317	1.351	5.34	19.000	31.600
2024912X	5	BECTON D	3 MONOCYTE	29	B.C600,800	59	7.253	0.792	10.92	5.300	9.300
2024912X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	59	6.607	0.323	4.88	4.600	8.600

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2024912X	5	BECTON D	5	BASOPHIL	29	B.C600,800	59	0.000	0.000	0.00	0	2.000
2024912X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	64	55.153	0.728	1.32	41.400	68.900
2024912X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	64	28.581	1.271	4.45	21.400	35.700
2024912X	6	MR Mindray	3	MONOCYTE	32	BC-5390	64	6.647	1.017	15.30	4.600	8.600
2024912X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	64	9.633	0.381	3.96	7.200	12.000
2024912X	6	MR Mindray	5	BASOPHIL	32	BC-5390	64	30.092	0.290	0.96	22.600	37.600
2024912X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	96	67.918	1.861	2.74	50.900	84.900
2024912X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	96	26.991	1.609	5.96	20.200	33.700
2024912X	7	DXH	3	MONOCYTE	36	BC DXH 500	96	0.400	0.371	92.75	0	2.400
2024912X	7	DXH	4	EOSINOPHIL	36	BC DXH 500	96	4.309	0.468	10.86	2.300	6.300
2024912X	7	DXH	5	BASOPHIL	36	BC DXH 500	96	0.413	0.592	143.36	0	2.400
172 DIFF. WBC FOR 5 PAR.			CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2024913X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	125	58.921	1.614	2.74	44.200	73.700
2024913X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	125	15.231	1.643	10.79	11.400	19.000
2024913X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	125	1.196	0.304	25.46	0	3.200
2024913X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	125	15.217	0.556	3.66	11.400	19.000
2024913X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	125	9.327	0.837	8.98	7.000	11.700
2024913X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	64	51.380	2.586	5.03	38.500	64.200
2024913X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	32	51.097	1.509	2.95	38.300	63.900
2024913X	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	64	38.294	2.507	6.55	28.700	47.900
2024913X	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	32	36.088	1.648	4.57	27.100	45.100
2024913X	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	64	3.222	0.786	24.41	1.200	5.200
2024913X	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	32	3.234	0.652	20.15	1.200	5.200
2024913X	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	64	7.114	0.614	8.63	5.100	9.100
2024913X	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	32	6.800	0.828	12.17	4.800	8.800
2024913X	2	ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	64	54.827	1.282	2.34	41.100	68.500
2024913X	2	ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	32	2.766	0.064	2.32	0.800	4.800
2024913X	3	ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	17	62.888	0.680	1.08	47.200	78.600
2024913X	3	ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	17	24.694	0.931	3.77	18.500	30.900
2024913X	3	ADVIA 120	3	MONOCYTE	9	ADVIA-120	17	7.629	0.339	4.45	5.600	9.600
2024913X	3	ADVIA 120	4	EOSINOPHIL	9	ADVIA-120	17	0.318	0.162	50.92	0	2.300

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2024913X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	0.659	0.264	40.01	0	2.700
2024913X	3	ADVIA 120	6 LUC	9	ADVIA-120	16	4.056	0.666	16.42	2.100	6.100
2024913X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	45	52.256	1.174	2.25	39.200	65.300
2024913X	4	AB- CELL-D	1 NEUTROPHIL	34	EMERALD 22	11	43.110	13.292	30.83	32.300	53.900
2024913X	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	43	50.423	1.449	2.87	37.800	63.000
2024913X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	45	36.187	2.739	7.57	27.100	45.200
2024913X	4	AB- CELL-D	2 LYMPHOCYTE	34	EMERALD 22	10	29.467	2.847	9.66	22.100	36.800
2024913X	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	43	37.070	2.620	7.07	27.800	46.300
2024913X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	45	7.275	1.603	22.03	5.300	9.300
2024913X	4	AB- CELL-D	3 MONOCYTE	34	EMERALD 22	10	19.233	2.315	12.04	14.400	24.000
2024913X	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	43	8.488	1.543	18.18	6.400	10.600
2024913X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	45	3.116	0.320	10.28	1.100	5.100
2024913X	4	AB- CELL-D	4 EOSINOPHIL	37	ABB.CEL RU	43	3.098	0.364	11.76	1.100	5.100
2024913X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	45	1.347	0.615	45.66	0	3.300
2024913X	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	43	0.883	0.421	47.64	0	2.900
2024913X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	59	56.786	1.656	2.92	42.600	71.000
2024913X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	59	28.738	2.085	7.26	21.600	35.900
2024913X	5	BECTON D	3 MONOCYTE	29	B.C600,800	59	5.816	0.817	14.05	3.800	7.800
2024913X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	59	8.768	0.489	5.58	6.600	11.000
2024913X	5	BECTON D	5 BASOPHIL	29	B.C600,800	59	0.002	0.013	754.98	0	2.000
2024913X	6	MR Mindray	1 NEUTROPHIL	32	BC-5390	64	47.461	1.065	2.24	35.600	59.300
2024913X	6	MR Mindray	2 LYMPHOCYTE	32	BC-5390	64	38.462	1.285	3.34	28.800	48.100
2024913X	6	MR Mindray	3 MONOCYTE	32	BC-5390	64	6.778	1.087	16.03	4.800	8.800
2024913X	6	MR Mindray	4 EOSINOPHIL	32	BC-5390	64	7.308	0.543	7.43	5.300	9.300
2024913X	6	MR Mindray	5 BASOPHIL	32	BC-5390	64	29.416	0.450	1.53	22.100	36.800
2024913X	7	DXH	1 NEUTROPHIL	36	BC DXH 500	96	47.947	5.398	11.26	36.000	59.900
2024913X	7	DXH	2 LYMPHOCYTE	36	BC DXH 500	96	39.793	5.587	14.04	29.800	49.700
2024913X	7	DXH	3 MONOCYTE	36	BC DXH 500	96	0.135	0.103	76.28	0	2.100
2024913X	7	DXH	4 EOSINOPHIL	36	BC DXH 500	96	11.863	1.106	9.33	8.900	14.800
2024913X	7	DXH	5 BASOPHIL	36	BC DXH 500	96	0.085	0.196	231.06	0	2.100

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765				EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}						
2024914X	1	SYSMEX CBC	1	NEUTROPHIL	25	XS SERIES	125	55.359	1.385	2.50	41.500	69.200
2024914X	1	SYSMEX CBC	2	LYMPHOCYTE	25	XS SERIES	125	19.968	1.097	5.50	15.000	25.000
2024914X	1	SYSMEX CBC	3	MONOCYTE	25	XS SERIES	125	1.240	0.288	23.24	0	3.200
2024914X	1	SYSMEX CBC	4	EOSINOPHIL	25	XS SERIES	125	14.445	0.518	3.58	10.800	18.100
2024914X	1	SYSMEX CBC	5	BASOPHIL	25	XS SERIES	125	9.012	0.763	8.47	6.800	11.300
2024914X	2	ACT PENTRA	1	NEUTROPHIL	15	COUL ACT 5	64	73.131	1.276	1.74	54.800	91.400
2024914X	2	ACT PENTRA	1	NEUTROPHIL	24	ABX PENTRA	32	70.456	0.934	1.33	52.800	88.100
2024914X	2	ACT PENTRA	2	LYMPHOCYTE	15	COUL ACT 5	64	19.852	0.911	4.59	14.900	24.800
2024914X	2	ACT PENTRA	2	LYMPHOCYTE	24	ABX PENTRA	32	18.528	0.838	4.52	13.900	23.200
2024914X	2	ACT PENTRA	3	MONOCYTE	15	COUL ACT 5	64	1.894	0.491	25.94	0	3.900
2024914X	2	ACT PENTRA	3	MONOCYTE	24	ABX PENTRA	32	1.944	0.352	18.10	0	3.900
2024914X	2	ACT PENTRA	4	EOSINOPHIL	15	COUL ACT 5	64	4.963	0.342	6.89	3.000	7.000
2024914X	2	ACT PENTRA	4	EOSINOPHIL	24	ABX PENTRA	32	5.045	0.264	5.23	3.000	7.000
2024914X	2	ACT PENTRA	5	BASOPHIL	15	COUL ACT 5	64	80.114	1.439	1.80	60.100	100.100
2024914X	2	ACT PENTRA	5	BASOPHIL	24	ABX PENTRA	32	3.984	0.044	1.11	2.000	6.000
2024914X	3	ADVIA 120	1	NEUTROPHIL	9	ADVIA-120	17	62.875	0.618	0.98	47.200	78.600
2024914X	3	ADVIA 120	2	LYMPHOCYTE	9	ADVIA-120	17	24.450	0.756	3.09	18.300	30.600
2024914X	3	ADVIA 120	3	MONOCYTE	9	ADVIA-120	17	7.794	0.396	5.08	5.800	9.800
2024914X	3	ADVIA 120	4	EOSINOPHIL	9	ADVIA-120	17	0.294	0.083	28.15	0	2.300
2024914X	3	ADVIA 120	5	BASOPHIL	9	ADVIA-120	17	0.665	0.276	41.58	0	2.700
2024914X	3	ADVIA 120	6	LUC	9	ADVIA-120	16	4.200	0.556	13.23	2.200	6.200
2024914X	4	AB- CELL-D	1	NEUTROPHIL	8	CELL-DYN	45	76.548	0.572	0.75	57.400	95.700
2024914X	4	AB- CELL-D	1	NEUTROPHIL	34	EMERALD 22	10	73.000	0.680	0.93	54.800	91.300
2024914X	4	AB- CELL-D	1	NEUTROPHIL	37	ABB.CEL RU	43	76.447	0.590	0.77	57.300	95.600
2024914X	4	AB- CELL-D	2	LYMPHOCYTE	8	CELL-DYN	45	14.712	0.855	5.81	11.000	18.400
2024914X	4	AB- CELL-D	2	LYMPHOCYTE	34	EMERALD 22	10	13.689	1.094	7.99	10.300	17.100
2024914X	4	AB- CELL-D	2	LYMPHOCYTE	37	ABB.CEL RU	43	14.863	0.697	4.69	11.100	18.600
2024914X	4	AB- CELL-D	3	MONOCYTE	8	CELL-DYN	45	3.764	0.451	11.98	1.800	5.800
2024914X	4	AB- CELL-D	3	MONOCYTE	34	EMERALD 22	10	9.778	0.916	9.37	7.300	12.200
2024914X	4	AB- CELL-D	3	MONOCYTE	37	ABB.CEL RU	43	4.005	0.455	11.36	2.000	6.000

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH			
2024914X	4 AB- CELL-D	4 EOSINOPHIL	8 CELL-DYN	45	4.193	0.194	4.62	2.200	6.200			
2024914X	4 AB- CELL-D	4 EOSINOPHIL	34 EMERALD 22	10	3.267	0.799	24.45	1.300	5.300			
2024914X	4 AB- CELL-D	4 EOSINOPHIL	37 ABB.CEL RU	42	4.154	0.152	3.65	2.200	6.200			
2024914X	4 AB- CELL-D	5 BASOPHIL	8 CELL-DYN	45	0.720	0.310	43.09	0	2.700			
2024914X	4 AB- CELL-D	5 BASOPHIL	34 EMERALD 22	10	0.289	0.074	25.51	0	2.300			
2024914X	4 AB- CELL-D	5 BASOPHIL	37 ABB.CEL RU	43	0.528	0.209	39.68	0	2.500			
2024914X	5 BECTON D	1 NEUTROPHIL	29 B.C600,800	59	67.957	2.573	3.79	51.000	84.900			
2024914X	5 BECTON D	2 LYMPHOCYTE	29 B.C600,800	59	20.526	1.924	9.37	15.400	25.700			
2024914X	5 BECTON D	3 MONOCYTE	29 B.C600,800	59	6.886	0.596	8.66	4.900	8.900			
2024914X	5 BECTON D	4 EOSINOPHIL	29 B.C600,800	59	4.841	0.645	13.33	2.800	6.800			
2024914X	5 BECTON D	5 BASOPHIL	29 B.C600,800	60	0.000	0.000	0.00	0	2.000			
2024914X	6 MR Mindray	1 NEUTROPHIL	32 BC-5390	64	60.436	1.237	2.05	45.300	75.500			
2024914X	6 MR Mindray	2 LYMPHOCYTE	32 BC-5390	64	18.275	0.974	5.33	13.700	22.800			
2024914X	6 MR Mindray	3 MONOCYTE	32 BC-5390	64	8.211	1.047	12.75	6.200	10.300			
2024914X	6 MR Mindray	4 EOSINOPHIL	32 BC-5390	63	13.078	0.384	2.94	9.800	16.300			
2024914X	6 MR Mindray	5 BASOPHIL	32 BC-5390	64	32.244	0.153	0.47	24.200	40.300			
2024914X	7 DXH	1 NEUTROPHIL	36 BC DXH 500	96	79.976	1.213	1.52	60.000	100.000			
2024914X	7 DXH	2 LYMPHOCYTE	36 BC DXH 500	96	12.203	1.237	10.14	9.200	15.300			
2024914X	7 DXH	3 MONOCYTE	36 BC DXH 500	96	0.324	0.230	71.04	0	2.300			
2024914X	7 DXH	4 EOSINOPHIL	36 BC DXH 500	96	7.318	0.467	6.39	5.300	9.300			
2024914X	7 DXH	5 BASOPHIL	36 BC DXH 500	96	0.149	0.242	162.00	0	2.100			
172 DIFF. WBC FOR 5 PAR.		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2024915X	1 SYSMEX CBC	1 NEUTROPHIL	25 XS SERIES	125	48.902	1.147	2.35	36.700	61.100			
2024915X	1 SYSMEX CBC	2 LYMPHOCYTE	25 XS SERIES	125	28.728	1.113	3.88	21.500	35.900			
2024915X	1 SYSMEX CBC	3 MONOCYTE	25 XS SERIES	124	1.636	0.356	21.75	0	3.600			
2024915X	1 SYSMEX CBC	4 EOSINOPHIL	25 XS SERIES	125	12.921	0.587	4.54	9.700	16.200			
2024915X	1 SYSMEX CBC	5 BASOPHIL	25 XS SERIES	125	7.741	0.704	9.09	5.700	9.700			
2024915X	2 ACT PENTRA	1 NEUTROPHIL	15 COUL ACT 5	64	51.683	2.850	5.51	38.800	64.600			
2024915X	2 ACT PENTRA	1 NEUTROPHIL	24 ABX PENTRA	32	50.931	1.704	3.35	38.200	63.700			
2024915X	2 ACT PENTRA	2 LYMPHOCYTE	15 COUL ACT 5	64	38.182	2.281	5.97	28.600	47.700			
2024915X	2 ACT PENTRA	2 LYMPHOCYTE	24 ABX PENTRA	32	36.266	1.566	4.32	27.200	45.300			

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2024915X	2	ACT PENTRA	3 MONOCYTE	15	COUL ACT 5	64	3.092	0.781	25.26	1.100	5.100
2024915X	2	ACT PENTRA	3 MONOCYTE	24	ABX PENTRA	33	3.245	0.602	18.55	1.200	5.200
2024915X	2	ACT PENTRA	4 EOSINOPHIL	15	COUL ACT 5	64	7.145	0.768	10.75	5.100	9.100
2024915X	2	ACT PENTRA	4 EOSINOPHIL	24	ABX PENTRA	32	6.778	0.607	8.95	4.800	8.800
2024915X	2	ACT PENTRA	5 BASOPHIL	15	COUL ACT 5	64	55.062	1.378	2.50	41.300	68.800
2024915X	2	ACT PENTRA	5 BASOPHIL	24	ABX PENTRA	32	2.763	0.048	1.75	0.800	4.800
2024915X	3	ADVIA 120	1 NEUTROPHIL	9	ADVIA-120	17	72.812	0.493	0.68	54.600	91.000
2024915X	3	ADVIA 120	2 LYMPHOCYTE	9	ADVIA-120	17	15.929	0.527	3.31	11.900	19.900
2024915X	3	ADVIA 120	3 MONOCYTE	9	ADVIA-120	17	7.741	0.387	4.99	5.700	9.700
2024915X	3	ADVIA 120	4 EOSINOPHIL	9	ADVIA-120	17	0.441	0.129	29.15	0	2.400
2024915X	3	ADVIA 120	5 BASOPHIL	9	ADVIA-120	17	0.429	0.145	33.67	0	2.400
2024915X	3	ADVIA 120	6 LUC	9	ADVIA-120	16	2.825	0.286	10.13	0.800	4.800
2024915X	4	AB- CELL-D	1 NEUTROPHIL	8	CELL-DYN	45	65.913	0.928	1.41	49.400	82.400
2024915X	4	AB- CELL-D	1 NEUTROPHIL	34	EMERALD 22	10	17.878	4.748	26.56	13.400	22.300
2024915X	4	AB- CELL-D	1 NEUTROPHIL	37	ABB.CEL RU	42	65.057	1.418	2.18	48.800	81.300
2024915X	4	AB- CELL-D	2 LYMPHOCYTE	8	CELL-DYN	45	23.700	1.284	5.42	17.800	29.600
2024915X	4	AB- CELL-D	2 LYMPHOCYTE	34	EMERALD 22	10	26.233	5.856	22.32	19.700	32.800
2024915X	4	AB- CELL-D	2 LYMPHOCYTE	37	ABB.CEL RU	42	25.169	1.583	6.29	18.900	31.500
2024915X	4	AB- CELL-D	3 MONOCYTE	8	CELL-DYN	45	5.445	0.668	12.26	3.400	7.400
2024915X	4	AB- CELL-D	3 MONOCYTE	34	EMERALD 22	10	21.322	4.308	20.20	16.000	26.700
2024915X	4	AB- CELL-D	3 MONOCYTE	37	ABB.CEL RU	43	5.284	0.814	15.41	3.300	7.300
2024915X	4	AB- CELL-D	4 EOSINOPHIL	8	CELL-DYN	45	4.096	0.281	6.87	2.100	6.100
2024915X	4	AB- CELL-D	4 EOSINOPHIL	34	EMERALD 22	10	33.222	9.091	27.37	24.900	41.500
2024915X	4	AB- CELL-D	4 EOSINOPHIL	37	ABB.CEL RU	43	4.174	0.333	7.97	2.200	6.200
2024915X	4	AB- CELL-D	5 BASOPHIL	8	CELL-DYN	45	0.938	0.440	46.96	0	2.900
2024915X	4	AB- CELL-D	5 BASOPHIL	34	EMERALD 22	10	2.078	0.689	33.17	0.100	4.100
2024915X	4	AB- CELL-D	5 BASOPHIL	37	ABB.CEL RU	42	0.300	0.278	92.54	0	2.300
2024915X	5	BECTON D	1 NEUTROPHIL	29	B.C600,800	58	60.660	0.908	1.50	45.500	75.800
2024915X	5	BECTON D	2 LYMPHOCYTE	29	B.C600,800	60	25.083	3.069	12.23	18.800	31.400
2024915X	5	BECTON D	3 MONOCYTE	29	B.C600,800	59	7.184	0.824	11.46	5.200	9.200
2024915X	5	BECTON D	4 EOSINOPHIL	29	B.C600,800	59	6.603	0.372	5.63	4.600	8.600

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 125 DIFF. FOR 5 PAR NEW

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2024915X	5	BECTON D	5	BASOPHIL	29	B.C600,800	59	0.000	0.000	0.00	0	2.000
2024915X	6	MR Mindray	1	NEUTROPHIL	32	BC-5390	64	55.147	0.749	1.36	41.400	68.900
2024915X	6	MR Mindray	2	LYMPHOCYTE	32	BC-5390	64	28.387	1.461	5.15	21.300	35.500
2024915X	6	MR Mindray	3	MONOCYTE	32	BC-5390	64	6.705	0.929	13.86	4.700	8.700
2024915X	6	MR Mindray	4	EOSINOPHIL	32	BC-5390	64	9.755	0.458	4.70	7.300	12.200
2024915X	6	MR Mindray	5	BASOPHIL	32	BC-5390	64	29.992	0.264	0.88	22.500	37.500
2024915X	7	DXH	1	NEUTROPHIL	36	BC DXH 500	96	80.074	1.319	1.65	60.100	100.100
2024915X	7	DXH	2	LYMPHOCYTE	36	BC DXH 500	96	12.166	1.372	11.28	9.100	15.200
2024915X	7	DXH	3	MONOCYTE	36	BC DXH 500	96	0.284	0.245	86.08	0	2.300
2024915X	7	DXH	4	EOSINOPHIL	36	BC DXH 500	96	7.314	0.486	6.65	5.300	9.300
2024915X	7	DXH	5	BASOPHIL	36	BC DXH 500	96	0.112	0.216	192.50	0	2.100

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE												
SAMPLE	METHOD	SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}								
2024911Y	1	6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	53	59.259	1.453	2.45	44.400	74.100
2024911Y	1	6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	221	56.073	1.061	1.89	42.100	70.100
2024911Y	1	6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	53	18.575	0.802	4.32	13.900	23.200
2024911Y	1	6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	221	19.372	1.037	5.35	14.500	24.200
2024911Y	1	6 PAR REG	3	MONOCYTE	31	SYSMEX XN	53	1.916	0.441	23.03	0	3.900
2024911Y	1	6 PAR REG	3	MONOCYTE	35	SYS XNL SR	221	1.074	0.181	16.86	0	3.100
2024911Y	1	6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	53	14.943	0.959	6.42	11.200	18.700
2024911Y	1	6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	221	14.955	0.568	3.80	11.200	18.700
2024911Y	1	6 PAR REG	5	BASOPHIL	31	SYSMEX XN	53	4.960	1.666	33.59	3.000	7.000
2024911Y	1	6 PAR REG	5	BASOPHIL	35	SYS XNL SR	221	8.560	0.673	7.86	6.400	10.700
2024911Y	1	6 PAR REG	6	IMMATURE G	31	SYSMEX XN	53	14.761	1.182	8.01	11.100	18.500

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE								NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST		REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
2024911Y	1 6 PAR REG	6	IMMATURE G	35	SYS XNL SR	220	15.166	0.764	5.04	11.400	19.000		
2024911Y	1 6 PAR REG	7	NRBC	31	SYSMEX XN	27	0.012	0.032	276.89	0	2.000		
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765		EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}									
2024912Y	1 6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	53	52.802	2.185	4.14	39.600	66.000		
2024912Y	1 6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	221	49.589	0.908	1.83	37.200	62.000		
2024912Y	1 6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	53	25.790	1.094	4.24	19.300	32.200		
2024912Y	1 6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	221	27.983	0.626	2.24	21.000	35.000		
2024912Y	1 6 PAR REG	3	MONOCYTE	31	SYSMEX XN	53	3.214	0.657	20.46	1.200	5.200		
2024912Y	1 6 PAR REG	3	MONOCYTE	35	SYS XNL SR	221	1.690	0.285	16.89	0	3.700		
2024912Y	1 6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	53	13.350	0.559	4.19	10.000	16.700		
2024912Y	1 6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	221	13.174	0.552	4.19	9.900	16.500		
2024912Y	1 6 PAR REG	5	BASOPHIL	31	SYSMEX XN	53	5.049	0.663	13.13	3.000	7.000		
2024912Y	1 6 PAR REG	5	BASOPHIL	35	SYS XNL SR	222	7.608	0.650	8.55	5.600	9.600		
2024912Y	1 6 PAR REG	6	IMMATURE G	31	SYSMEX XN	53	13.080	0.939	7.18	9.800	16.400		
2024912Y	1 6 PAR REG	6	IMMATURE G	35	SYS XNL SR	219	13.402	0.663	4.95	10.100	16.800		
2024912Y	1 6 PAR REG	7	NRBC	31	SYSMEX XN	28	0.186	0.185	99.41	0	2.200		
2024913Y	1 6 PAR REG	1	NEUTROPHIL	31	SYSMEX XN	53	61.902	1.549	2.50	46.400	77.400		
2024913Y	1 6 PAR REG	1	NEUTROPHIL	35	SYS XNL SR	221	58.390	1.189	2.04	43.800	73.000		
2024913Y	1 6 PAR REG	2	LYMPHOCYTE	31	SYSMEX XN	53	14.988	0.704	4.70	11.200	18.700		
2024913Y	1 6 PAR REG	2	LYMPHOCYTE	35	SYS XNL SR	221	16.289	0.733	4.50	12.200	20.400		
2024913Y	1 6 PAR REG	3	MONOCYTE	31	SYSMEX XN	53	2.290	0.519	22.67	0.300	4.300		
2024913Y	1 6 PAR REG	3	MONOCYTE	35	SYS XNL SR	221	1.150	0.252	21.93	0	3.200		
2024913Y	1 6 PAR REG	4	EOSINOPHIL	31	SYSMEX XN	53	15.516	0.701	4.52	11.600	19.400		
2024913Y	1 6 PAR REG	4	EOSINOPHIL	35	SYS XNL SR	221	15.141	0.772	5.10	11.400	18.900		
2024913Y	1 6 PAR REG	5	BASOPHIL	31	SYSMEX XN	53	4.975	1.775	35.67	3.000	7.000		
2024913Y	1 6 PAR REG	5	BASOPHIL	35	SYS XNL SR	221	9.043	0.778	8.61	6.800	11.300		
2024913Y	1 6 PAR REG	6	IMMATURE G	31	SYSMEX XN	53	15.638	0.753	4.81	11.700	19.500		
2024913Y	1 6 PAR REG	6	IMMATURE G	35	SYS XNL SR	220	15.931	0.785	4.93	11.900	19.900		
2024913Y	1 6 PAR REG	7	NRBC	31	SYSMEX XN	27	0.141	0.213	151.36	0	2.100		



**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 127 DIFF 6 PARAM INSTRUM

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH		
185 Diff 6 Parameter Ins		CLIA TEST ID #: 765	EVALUATION CRITERIA: TARGET VALUE	25 % OR 2.00	{GREATER}						
2024914Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	53	59.300	1.412	2.38	44.500	74.100		
2024914Y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	221	56.099	1.263	2.25	42.100	70.100		
2024914Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	53	18.660	0.631	3.38	14.000	23.300		
2024914Y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	221	19.323	1.093	5.66	14.500	24.200		
2024914Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	53	1.884	0.311	16.53	0	3.900		
2024914Y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	221	1.070	0.182	17.06	0	3.100		
2024914Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	53	14.927	0.985	6.60	11.200	18.700		
2024914Y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	221	15.012	0.573	3.82	11.300	18.800		
2024914Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	53	5.154	0.982	19.06	3.200	7.200		
2024914Y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	221	8.607	0.697	8.09	6.500	10.800		
2024914Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	53	14.620	1.030	7.04	11.000	18.300		
2024914Y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	220	15.249	0.794	5.21	11.400	19.100		
2024914Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	27	0.019	0.039	204.94	0	2.000		
2024915Y	1 6 PAR REG	1 NEUTROPHIL	31 SYSMEX XN	53	52.324	1.334	2.55	39.200	65.400		
2024915Y	1 6 PAR REG	1 NEUTROPHIL	35 SYS XNL SR	220	49.579	0.949	1.91	37.200	62.000		
2024915Y	1 6 PAR REG	2 LYMPHOCYTE	31 SYSMEX XN	53	25.951	0.833	3.21	19.500	32.400		
2024915Y	1 6 PAR REG	2 LYMPHOCYTE	35 SYS XNL SR	221	27.979	0.707	2.53	21.000	35.000		
2024915Y	1 6 PAR REG	3 MONOCYTE	31 SYSMEX XN	53	3.082	0.528	17.14	1.100	5.100		
2024915Y	1 6 PAR REG	3 MONOCYTE	35 SYS XNL SR	222	1.679	0.281	16.73	0	3.700		
2024915Y	1 6 PAR REG	4 EOSINOPHIL	31 SYSMEX XN	53	13.412	0.482	3.59	10.100	16.800		
2024915Y	1 6 PAR REG	4 EOSINOPHIL	35 SYS XNL SR	221	13.173	0.554	4.20	9.900	16.500		
2024915Y	1 6 PAR REG	5 BASOPHIL	31 SYSMEX XN	53	5.080	0.849	16.72	3.100	7.100		
2024915Y	1 6 PAR REG	5 BASOPHIL	35 SYS XNL SR	221	7.602	0.804	10.58	5.600	9.600		
2024915Y	1 6 PAR REG	6 IMMATURE G	31 SYSMEX XN	52	13.230	0.534	4.03	9.900	16.500		
2024915Y	1 6 PAR REG	6 IMMATURE G	35 SYS XNL SR	220	13.348	0.617	4.62	10.000	16.700		
2024915Y	1 6 PAR REG	7 NRBC	31 SYSMEX XN	29	0.143	0.182	127.48	0	2.100		

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 128 DIFF 3 PARAMETER INS

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 128 DIFF 3 PARAMETER INS

ANALYTE							NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH		
186	DIFF 3 PARAMETER INS	CLIA TEST ID #: 765			EVALUATION CRITERIA: TARGET VALUE 25 % OR 2.00 {GREATER}							
2024911Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	17	28.912	0.776	2.68	21.700	36.100	
2024911Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	19	29.105	1.049	3.60	21.800	36.400	
2024911Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	17	18.735	1.195	6.38	14.100	23.400	
2024911Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	19	12.495	1.084	8.68	9.400	15.600	
2024911Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	17	52.353	1.221	2.33	39.300	65.400	
2024911Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	20	58.225	1.431	2.46	43.700	72.800	
2024912Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	17	29.053	0.677	2.33	21.800	36.300	
2024912Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	19	29.247	0.899	3.07	21.900	36.600	
2024912Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	17	19.059	0.846	4.44	14.300	23.800	
2024912Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	19	12.432	1.128	9.07	9.300	15.500	
2024912Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	17	51.829	0.986	1.90	38.900	64.800	
2024912Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	20	58.130	1.717	2.95	43.600	72.700	
2024913Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	17	61.712	0.790	1.28	46.300	77.100	
2024913Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	19	60.800	0.871	1.43	45.600	76.000	
2024913Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	17	14.118	0.844	5.98	10.600	17.600	
2024913Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	19	10.321	0.998	9.67	7.700	12.900	
2024913Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	17	24.171	0.788	3.26	18.100	30.200	
2024913Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	20	28.860	0.874	3.03	21.600	36.100	
2024914Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	17	28.982	1.021	3.52	21.700	36.200	
2024914Z	8 3ST	1	Lymphocyte	3	ABBOTT C-D	19	29.753	0.956	3.21	22.300	37.200	
2024914Z	8 3ST	2	Mix M/E/Ba	1	KX 21 - Sy	17	19.318	0.920	4.76	14.500	24.100	
2024914Z	8 3ST	2	Mix M/E/Ba	3	ABBOTT C-D	19	12.416	1.386	11.16	9.300	15.500	
2024914Z	8 3ST	3	Neutrophil	1	KX 21 - Sy	17	51.641	1.303	2.52	38.700	64.600	
2024914Z	8 3ST	3	Neutrophil	3	ABBOTT C-D	20	57.625	2.016	3.50	43.200	72.000	
2024915Z	8 3ST	1	Lymphocyte	1	KX 21 - Sy	17	13.194	0.981	7.44	9.900	16.500	

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 128 DIFF 3 PARAMETER INS

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
2024915Z	8	3ST	1	Lymphocyte	3	ABBOTT C-D	19	15.363	1.069	6.96	11.500	19.200
2024915Z	8	3ST	2	Mix M/E/Ba	1	KX 21 - Sy	16	18.381	1.702	9.26	13.800	23.000
2024915Z	8	3ST	2	Mix M/E/Ba	3	ABBOTT C-D	19	11.805	1.211	10.26	8.900	14.800
2024915Z	8	3ST	3	Neutrophil	1	KX 21 - Sy	18	68.435	1.178	1.72	51.300	85.500
2024915Z	8	3ST	3	Neutrophil	3	ABBOTT C-D	20	72.530	2.147	2.96	54.400	90.700

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMP BLOOD COUNT NEW

ANALYTE												
SAMPLE	METHOD		SYST/INST		REAGENTS		NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH
549	RED CELL COUNT (RBC)		CLIA TEST ID #: 775		EVALUATION CRITERIA: TARGET VALUE 6 %							
2024-911	1	Sysmex CBC	132	SYSMEX XS			126	5.640	0.076	1.34	5.300	5.980
2024-911	1	Sysmex CBC	141	SYSMEX XN			55	5.655	0.049	0.87	5.320	5.990
2024-911	1	Sysmex CBC	152	Sysmex XN			223	5.673	0.061	1.08	5.330	6.010
2024-911	2	ACT PENTRA	125	COULTER AC			63	4.414	0.082	1.87	4.150	4.680
2024-911	2	ACT PENTRA	130	ABX PENTRA			32	4.372	0.064	1.47	4.110	4.630
2024-911	3	ADVIA 120	116	ADVIA-120			17	5.361	0.054	1.01	5.040	5.680
2024-911	4	AB-CELL DY	115	ABBOTT C-D			35	5.280	0.066	1.24	4.960	5.600
2024-911	4	AB-CELL DY	142	EMERALD			11	5.062	0.133	2.64	4.760	5.370
2024-911	4	AB-CELL DY	153	AB-CEL-RUB			52	5.406	0.100	1.84	5.080	5.730
2024-911	5	BECTON D	137	B.C DXH			58	5.411	0.069	1.27	5.090	5.740
2024-911	6	MR Mindray	145	BC-5390			64	5.324	0.080	1.51	5.000	5.640
2024-911	7	DXH500 Ser	149	BC DXH 500			99	2.290	0.032	1.39	2.150	2.430
2024-911	8	3 ST	84	ABBOTT C-D			20	4.164	0.102	2.44	3.910	4.410
2024-911	8	3 ST	138	SY KX-21N			13	4.139	0.067	1.63	3.890	4.390
2024-912	1	Sysmex CBC	132	SYSMEX XS			126	4.611	0.062	1.34	4.330	4.890
2024-912	1	Sysmex CBC	141	SYSMEX XN			55	4.627	0.053	1.14	4.350	4.900
2024-912	1	Sysmex CBC	152	Sysmex XN			223	4.608	0.054	1.17	4.330	4.880

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2024-912	2	ACT PENTRA	125 COULTER AC	63	5.075	0.083	1.63	4.770	5.380
2024-912	2	ACT PENTRA	130 ABX PENTRA	32	5.034	0.080	1.58	4.730	5.340
2024-912	3	ADVIA 120	116 ADVIA-120	17	2.156	0.034	1.60	2.030	2.290
2024-912	4	AB-CELL DY	115 ABBOTT C-D	35	4.654	0.090	1.92	4.370	4.930
2024-912	4	AB-CELL DY	142 EMERALD	11	4.506	0.101	2.25	4.240	4.780
2024-912	4	AB-CELL DY	153 AB-CEL-RUB	52	4.773	0.085	1.78	4.490	5.060
2024-912	5	BECTON D	137 B.C DXH	58	3.937	0.059	1.50	3.700	4.170
2024-912	6	MR Mindray	145 BC-5390	64	4.442	0.061	1.37	4.180	4.710
2024-912	7	DXH500 Ser	149 BC DXH 500	99	4.536	0.081	1.78	4.260	4.810
2024-912	8	3 ST	84 ABBOTT C-D	20	4.185	0.103	2.47	3.930	4.440
2024-912	8	3 ST	138 SY KX-21N	13	4.126	0.068	1.66	3.880	4.370
549 RED CELL COUNT (RBC)		CLIA TEST ID #: 775			EVALUATION CRITERIA: TARGET VALUE 6 %				
2024-913	1	Sysmex CBC	132 SYSMEX XS	126	2.312	0.031	1.34	2.170	2.450
2024-913	1	Sysmex CBC	141 SYSMEX XN	55	2.291	0.032	1.38	2.150	2.430
2024-913	1	Sysmex CBC	152 Sysmex XN	223	2.275	0.028	1.23	2.140	2.410
2024-913	2	ACT PENTRA	125 COULTER AC	63	2.300	0.041	1.78	2.160	2.440
2024-913	2	ACT PENTRA	130 ABX PENTRA	32	2.288	0.047	2.03	2.150	2.430
2024-913	3	ADVIA 120	116 ADVIA-120	17	4.399	0.088	1.99	4.140	4.660
2024-913	4	AB-CELL DY	115 ABBOTT C-D	35	2.220	0.105	4.72	2.090	2.350
2024-913	4	AB-CELL DY	142 EMERALD	11	2.072	0.052	2.53	1.950	2.200
2024-913	4	AB-CELL DY	153 AB-CEL-RUB	52	2.196	0.039	1.78	2.060	2.330
2024-913	5	BECTON D	137 B.C DXH	58	2.540	0.037	1.44	2.390	2.690
2024-913	6	MR Mindray	145 BC-5390	64	2.203	0.030	1.37	2.070	2.340
2024-913	7	DXH500 Ser	149 BC DXH 500	99	2.286	0.042	1.85	2.150	2.420
2024-913	8	3 ST	84 ABBOTT C-D	20	5.898	0.167	2.83	5.540	6.250
2024-913	8	3 ST	138 SY KX-21N	13	5.902	0.118	1.99	5.550	6.260
2024-914	1	Sysmex CBC	132 SYSMEX XS	126	5.644	0.076	1.34	5.310	5.980
2024-914	1	Sysmex CBC	141 SYSMEX XN	55	5.656	0.049	0.87	5.320	6.000
2024-914	1	Sysmex CBC	152 Sysmex XN	223	5.674	0.064	1.13	5.330	6.010

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2024-914	2	ACT PENTRA	125 COULTER AC	63	5.062	0.075	1.47	4.760	5.370	
2024-914	2	ACT PENTRA	130 ABX PENTRA	32	5.039	0.061	1.21	4.740	5.340	
2024-914	3	ADVIA 120	116 ADVIA-120	17	4.415	0.100	2.27	4.150	4.680	
2024-914	4	AB-CELL DY	115 ABBOTT C-D	35	5.261	0.065	1.24	4.950	5.580	
2024-914	4	AB-CELL DY	142 EMERALD	11	5.070	0.144	2.84	4.770	5.370	
2024-914	4	AB-CELL DY	153 AB-CEL-RUB	52	5.388	0.087	1.61	5.070	5.710	
2024-914	5	BECTON D	137 B.C DXH	58	5.387	0.069	1.29	5.060	5.710	
2024-914	6	MR Mindray	145 BC-5390	64	5.342	0.088	1.64	5.020	5.660	
2024-914	7	DXH500 Ser	149 BC DXH 500	99	5.217	0.080	1.53	4.900	5.530	
2024-914	8	3 ST	84 ABBOTT C-D	20	4.192	0.109	2.61	3.940	4.440	
2024-914	8	3 ST	138 SY KX-21N	13	4.136	0.066	1.59	3.890	4.380	
549	RED CELL COUNT (RBC)	CLIA TEST ID #: 775	EVALUATION CRITERIA: TARGET VALUE 6 %							
2024-915	1	Sysmex CBC	132 SYSMEX XS	126	4.613	0.059	1.27	4.340	4.890	
2024-915	1	Sysmex CBC	141 SYSMEX XN	55	4.629	0.040	0.87	4.350	4.910	
2024-915	1	Sysmex CBC	152 Sysmex XN	223	4.612	0.053	1.16	4.340	4.890	
2024-915	2	ACT PENTRA	125 COULTER AC	63	2.301	0.041	1.79	2.160	2.440	
2024-915	2	ACT PENTRA	130 ABX PENTRA	32	2.293	0.036	1.57	2.160	2.430	
2024-915	3	ADVIA 120	116 ADVIA-120	17	5.299	0.101	1.91	4.980	5.620	
2024-915	4	AB-CELL DY	115 ABBOTT C-D	35	4.647	0.062	1.34	4.370	4.930	
2024-915	4	AB-CELL DY	142 EMERALD	11	4.513	0.102	2.26	4.240	4.780	
2024-915	4	AB-CELL DY	153 AB-CEL-RUB	52	4.752	0.086	1.80	4.470	5.040	
2024-915	5	BECTON D	137 B.C DXH	58	3.942	0.058	1.48	3.710	4.180	
2024-915	6	MR Mindray	145 BC-5390	64	4.435	0.060	1.36	4.170	4.700	
2024-915	7	DXH500 Ser	149 BC DXH 500	99	5.216	0.070	1.34	4.900	5.530	
2024-915	8	3 ST	84 ABBOTT C-D	20	2.497	0.075	2.99	2.350	2.650	
2024-915	8	3 ST	138 SY KX-21N	13	2.491	0.040	1.60	2.340	2.640	
550	HEMATOCRIT (HCT)	CLIA TEST ID #: 785								
2024-911	1	Sysmex CBC	132 SYSMEX XS	126	51.854	1.347	2.60	48.700	55.000	
2024-911	1	Sysmex CBC	141 SYSMEX XN	55	50.035	0.842	1.68	47.000	53.000	
2024-911	1	Sysmex CBC	152 Sysmex XN	223	50.561	0.884	1.75	47.500	53.600	

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2024-911	2	ACT PENTRA	125 COULTER AC	63	37.149	0.736	1.98	34.900	39.400
2024-911	2	ACT PENTRA	130 ABX PENTRA	32	36.800	0.667	1.81	34.600	39.000
2024-911	3	ADVIA 120	116 ADVIA-120	17	48.365	0.974	2.01	45.500	51.300
2024-911	4	AB-CELL DY	115 ABBOTT C-D	36	42.881	0.967	2.26	40.300	45.500
2024-911	4	AB-CELL DY	142 EMERALD	11	45.609	2.334	5.12	42.900	48.300
2024-911	4	AB-CELL DY	153 AB-CEL-RUB	51	43.953	0.916	2.08	41.300	46.600
2024-911	5	BECTON D	137 B.C DXH	58	57.262	0.899	1.57	53.800	60.700
2024-911	6	MR Mindray	145 BC-5390	64	47.211	1.230	2.60	44.400	50.000
2024-911	7	DXH500 Ser	149 BC DXH 500	99	18.015	0.325	1.80	16.900	19.100
2024-911	8	3 ST	84 ABBOTT C-D	20	34.885	0.801	2.30	32.800	37.000
2024-911	8	3 ST	138 SY KX-21N	13	33.477	0.581	1.73	31.500	35.500
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %					
2024-912	1	Sysmex CBC	132 SYSMEX XS	126	38.880	0.951	2.44	36.500	41.200
2024-912	1	Sysmex CBC	141 SYSMEX XN	55	37.913	0.762	2.01	35.600	40.200
2024-912	1	Sysmex CBC	152 Sysmex XN	223	38.101	0.685	1.80	35.800	40.400
2024-912	2	ACT PENTRA	125 COULTER AC	63	46.616	1.010	2.17	43.800	49.400
2024-912	2	ACT PENTRA	130 ABX PENTRA	32	46.200	0.764	1.65	43.400	49.000
2024-912	3	ADVIA 120	116 ADVIA-120	17	15.812	0.323	2.05	14.900	16.800
2024-912	4	AB-CELL DY	115 ABBOTT C-D	36	37.350	0.981	2.63	35.100	39.600
2024-912	4	AB-CELL DY	142 EMERALD	11	38.855	1.525	3.93	36.500	41.200
2024-912	4	AB-CELL DY	153 AB-CEL-RUB	51	38.498	0.857	2.23	36.200	40.800
2024-912	5	BECTON D	137 B.C DXH	58	37.460	0.558	1.49	35.200	39.700
2024-912	6	MR Mindray	145 BC-5390	64	36.923	1.030	2.79	34.700	39.100
2024-912	7	DXH500 Ser	149 BC DXH 500	99	41.018	0.887	2.16	38.600	43.500
2024-912	8	3 ST	84 ABBOTT C-D	20	35.015	0.708	2.02	32.900	37.100
2024-912	8	3 ST	138 SY KX-21N	13	33.315	0.514	1.54	31.300	35.300
2024-913	1	Sysmex CBC	132 SYSMEX XS	126	17.918	0.405	2.26	16.800	19.000
2024-913	1	Sysmex CBC	141 SYSMEX XN	55	17.087	0.367	2.15	16.100	18.100
2024-913	1	Sysmex CBC	152 Sysmex XN	223	17.241	0.309	1.79	16.200	18.300

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2024-913	2	ACT PENTRA	125 COULTER AC	63	18.022	0.433	2.40	16.900	19.100
2024-913	2	ACT PENTRA	130 ABX PENTRA	32	17.847	0.415	2.33	16.800	18.900
2024-913	3	ADVIA 120	116 ADVIA-120	17	36.100	0.968	2.68	33.900	38.300
2024-913	4	AB-CELL DY	115 ABBOTT C-D	36	15.886	0.331	2.08	14.900	16.800
2024-913	4	AB-CELL DY	142 EMERALD	11	16.282	0.766	4.71	15.300	17.300
2024-913	4	AB-CELL DY	153 AB-CEL-RUB	51	15.870	0.352	2.22	14.900	16.800
2024-913	5	BECTON D	137 B.C DXH	58	21.191	0.346	1.63	19.900	22.500
2024-913	6	MR Mindray	145 BC-5390	64	16.675	0.447	2.68	15.700	17.700
2024-913	7	DXH500 Ser	149 BC DXH 500	99	17.999	0.405	2.25	16.900	19.100
2024-913	8	3 ST	84 ABBOTT C-D	20	53.940	1.168	2.17	50.700	57.200
2024-913	8	3 ST	138 SY KX-21N	13	52.231	1.065	2.04	49.100	55.400
550 HEMATOCRIT (HCT)		CLIA TEST ID #: 785		EVALUATION CRITERIA: TARGET VALUE 6 %					
2024-914	1	Sysmex CBC	132 SYSMEX XS	126	51.778	1.155	2.23	48.700	54.900
2024-914	1	Sysmex CBC	141 SYSMEX XN	55	50.129	0.805	1.61	47.100	53.100
2024-914	1	Sysmex CBC	152 Sysmex XN	223	50.657	0.859	1.69	47.600	53.700
2024-914	2	ACT PENTRA	125 COULTER AC	63	46.476	0.904	1.95	43.700	49.300
2024-914	2	ACT PENTRA	130 ABX PENTRA	32	46.250	0.820	1.77	43.500	49.000
2024-914	3	ADVIA 120	116 ADVIA-120	17	36.276	1.107	3.05	34.100	38.500
2024-914	4	AB-CELL DY	115 ABBOTT C-D	36	42.842	1.000	2.34	40.300	45.400
2024-914	4	AB-CELL DY	142 EMERALD	11	45.464	2.256	4.96	42.700	48.200
2024-914	4	AB-CELL DY	153 AB-CEL-RUB	51	43.939	0.878	2.00	41.300	46.600
2024-914	5	BECTON D	137 B.C DXH	58	56.981	0.940	1.65	53.600	60.400
2024-914	6	MR Mindray	145 BC-5390	64	46.973	1.455	3.10	44.200	49.800
2024-914	7	DXH500 Ser	149 BC DXH 500	99	50.139	1.012	2.02	47.100	53.100
2024-914	8	3 ST	84 ABBOTT C-D	20	35.085	0.705	2.01	33.000	37.200
2024-914	8	3 ST	138 SY KX-21N	13	33.377	0.524	1.57	31.400	35.400
2024-915	1	Sysmex CBC	132 SYSMEX XS	126	38.902	0.875	2.25	36.600	41.200
2024-915	1	Sysmex CBC	141 SYSMEX XN	55	37.853	0.558	1.47	35.600	40.100
2024-915	1	Sysmex CBC	152 Sysmex XN	223	38.121	0.667	1.75	35.800	40.400

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2024-915	2	ACT PENTRA	125 COULTER AC	63	18.015	0.378	2.10	16.900	19.100	
2024-915	2	ACT PENTRA	130 ABX PENTRA	32	17.875	0.397	2.22	16.800	18.900	
2024-915	3	ADVIA 120	116 ADVIA-120	17	47.771	1.244	2.60	44.900	50.600	
2024-915	4	AB-CELL DY	115 ABBOTT C-D	36	37.369	0.806	2.16	35.100	39.600	
2024-915	4	AB-CELL DY	142 EMERALD	11	38.764	1.488	3.84	36.400	41.100	
2024-915	4	AB-CELL DY	153 AB-CEL-RUB	51	38.372	0.849	2.21	36.100	40.700	
2024-915	5	BECTON D	137 B.C DXH	58	37.550	0.621	1.65	35.300	39.800	
2024-915	6	MR Mindray	145 BC-5390	64	36.769	1.062	2.89	34.600	39.000	
2024-915	7	DXH500 Ser	149 BC DXH 500	99	50.231	0.992	1.98	47.200	53.200	
2024-915	8	3 ST	84 ABBOTT C-D	20	18.505	0.539	2.91	17.400	19.600	
2024-915	8	3 ST	138 SY KX-21N	13	18.223	0.338	1.85	17.100	19.300	
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795	EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}							
2024-911	1	Sysmex CBC	132 SYSMEX XS	127	17.483	0.253	1.45	16.300	18.700	
2024-911	1	Sysmex CBC	141 SYSMEX XN	55	17.495	0.176	1.01	16.300	18.700	
2024-911	1	Sysmex CBC	152 Sysmex XN	222	17.473	0.147	0.84	16.300	18.700	
2024-911	2	ACT PENTRA	125 COULTER AC	63	12.887	0.199	1.54	12.000	13.800	
2024-911	2	ACT PENTRA	130 ABX PENTRA	32	12.819	0.113	0.88	11.900	13.700	
2024-911	3	ADVIA 120	116 ADVIA-120	17	17.241	0.293	1.70	16.000	18.400	
2024-911	4	AB-CELL DY	115 ABBOTT C-D	36	16.514	0.379	2.30	15.400	17.700	
2024-911	4	AB-CELL DY	142 EMERALD	11	15.300	0.467	3.05	14.200	16.400	
2024-911	4	AB-CELL DY	153 AB-CEL-RUB	51	16.637	0.269	1.61	15.500	17.800	
2024-911	5	BECTON D	137 B.C DXH	58	17.319	0.207	1.20	16.100	18.500	
2024-911	6	MR Mindray	145 BC-5390	64	16.637	0.220	1.32	15.500	17.800	
2024-911	7	DXH500 Ser	149 BC DXH 500	99	6.123	0.103	1.68	5.700	6.600	
2024-911	8	3 ST	84 ABBOTT C-D	20	12.100	0.182	1.50	11.300	12.900	
2024-911	8	3 ST	138 SY KX-21N	13	12.031	0.246	2.05	11.200	12.900	
2024-912	1	Sysmex CBC	132 SYSMEX XS	127	13.094	0.181	1.39	12.200	14.000	
2024-912	1	Sysmex CBC	141 SYSMEX XN	55	13.231	0.115	0.87	12.300	14.200	
2024-912	1	Sysmex CBC	152 Sysmex XN	222	13.160	0.115	0.87	12.200	14.100	



**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL		GRP: 132 COMP BLOOD COUNT NEW						
ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2024-912	2	ACT PENTRA	125 COULTER AC	63	16.227	0.226	1.39	15.100	17.400	
2024-912	2	ACT PENTRA	130 ABX PENTRA	32	16.272	0.159	0.97	15.100	17.400	
2024-912	3	ADVIA 120	116 ADVIA-120	17	5.706	0.100	1.75	5.300	6.100	
2024-912	4	AB-CELL DY	115 ABBOTT C-D	36	13.336	0.299	2.24	12.400	14.300	
2024-912	4	AB-CELL DY	142 EMERALD	11	11.882	0.449	3.78	11.100	12.700	
2024-912	4	AB-CELL DY	153 AB-CEL-RUB	51	13.400	0.221	1.65	12.500	14.300	
2024-912	5	BECTON D	137 B.C DXH	58	11.684	0.168	1.44	10.900	12.500	
2024-912	6	MR Mindray	145 BC-5390	64	12.894	0.145	1.12	12.000	13.800	
2024-912	7	DXH500 Ser	149 BC DXH 500	99	14.336	0.157	1.09	13.300	15.300	
2024-912	8	3 ST	84 ABBOTT C-D	20	12.185	0.220	1.80	11.300	13.000	
2024-912	8	3 ST	138 SY KX-21N	13	12.083	0.121	1.00	11.200	12.900	
551 HEMOGLOBIN (HGB)			CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}					
2024-913	1	Sysmex CBC	132 SYSMEX XS	127	5.750	0.115	1.99	5.300	6.200	
2024-913	1	Sysmex CBC	141 SYSMEX XN	55	5.822	0.080	1.38	5.400	6.200	
2024-913	1	Sysmex CBC	152 Sysmex XN	222	5.807	0.068	1.17	5.400	6.200	
2024-913	2	ACT PENTRA	125 COULTER AC	63	6.179	0.118	1.92	5.700	6.600	
2024-913	2	ACT PENTRA	130 ABX PENTRA	32	6.075	0.066	1.09	5.600	6.500	
2024-913	3	ADVIA 120	116 ADVIA-120	17	13.041	0.185	1.42	12.100	14.000	
2024-913	4	AB-CELL DY	115 ABBOTT C-D	36	5.917	0.178	3.01	5.500	6.300	
2024-913	4	AB-CELL DY	142 EMERALD	11	4.850	0.081	1.66	4.500	5.200	
2024-913	4	AB-CELL DY	153 AB-CEL-RUB	51	5.710	0.113	1.98	5.300	6.100	
2024-913	5	BECTON D	137 B.C DXH	58	6.634	0.097	1.47	6.200	7.100	
2024-913	6	MR Mindray	145 BC-5390	64	5.600	0.076	1.36	5.200	6.000	
2024-913	7	DXH500 Ser	149 BC DXH 500	99	6.059	0.091	1.49	5.600	6.500	
2024-913	8	3 ST	84 ABBOTT C-D	20	18.600	0.420	2.26	17.300	19.900	
2024-913	8	3 ST	138 SY KX-21N	13	18.608	0.279	1.50	17.300	19.900	
2024-914	1	Sysmex CBC	132 SYSMEX XS	127	17.484	0.266	1.52	16.300	18.700	
2024-914	1	Sysmex CBC	141 SYSMEX XN	55	17.504	0.140	0.80	16.300	18.700	
2024-914	1	Sysmex CBC	152 Sysmex XN	222	17.464	0.148	0.85	16.200	18.700	

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH
2024-914	2	ACT PENTRA	125 COULTER AC	63	16.216	0.218	1.35	15.100	17.400
2024-914	2	ACT PENTRA	130 ABX PENTRA	32	16.291	0.174	1.07	15.200	17.400
2024-914	3	ADVIA 120	116 ADVIA-120	17	13.053	0.200	1.53	12.100	14.000
2024-914	4	AB-CELL DY	115 ABBOTT C-D	36	16.428	0.379	2.31	15.300	17.600
2024-914	4	AB-CELL DY	142 EMERALD	11	15.255	0.431	2.83	14.200	16.300
2024-914	4	AB-CELL DY	153 AB-CEL-RUB	51	16.516	0.275	1.66	15.400	17.700
2024-914	5	BECTON D	137 B.C DXH	58	17.272	0.212	1.23	16.100	18.500
2024-914	6	MR Mindray	145 BC-5390	64	16.679	0.248	1.48	15.500	17.800
2024-914	7	DXH500 Ser	149 BC DXH 500	99	17.553	0.225	1.28	16.300	18.800
2024-914	8	3 ST	84 ABBOTT C-D	20	12.200	0.259	2.12	11.300	13.100
2024-914	8	3 ST	138 SY KX-21N	13	12.054	0.227	1.89	11.200	12.900
551 HEMOGLOBIN (HGB)		CLIA TEST ID #: 795		EVALUATION CRITERIA: TARGET VALUE 7 % OR 2.00 {LOWER}					
2024-915	1	Sysmex CBC	132 SYSMEX XS	127	13.102	0.180	1.38	12.200	14.000
2024-915	1	Sysmex CBC	141 SYSMEX XN	55	13.243	0.113	0.85	12.300	14.200
2024-915	1	Sysmex CBC	152 Sysmex XN	222	13.165	0.122	0.92	12.200	14.100
2024-915	2	ACT PENTRA	125 COULTER AC	63	6.184	0.122	1.97	5.800	6.600
2024-915	2	ACT PENTRA	130 ABX PENTRA	32	6.081	0.047	0.77	5.700	6.500
2024-915	3	ADVIA 120	116 ADVIA-120	17	17.088	0.287	1.68	15.900	18.300
2024-915	4	AB-CELL DY	115 ABBOTT C-D	36	13.350	0.284	2.13	12.400	14.300
2024-915	4	AB-CELL DY	142 EMERALD	11	11.745	0.489	4.16	10.900	12.600
2024-915	4	AB-CELL DY	153 AB-CEL-RUB	51	13.347	0.252	1.89	12.400	14.300
2024-915	5	BECTON D	137 B.C DXH	58	11.705	0.162	1.39	10.900	12.500
2024-915	6	MR Mindray	145 BC-5390	64	12.898	0.175	1.35	12.000	13.800
2024-915	7	DXH500 Ser	149 BC DXH 500	99	17.509	0.195	1.11	16.300	18.700
2024-915	8	3 ST	84 ABBOTT C-D	20	6.500	0.103	1.58	6.000	7.000
2024-915	8	3 ST	138 SY KX-21N	13	6.254	0.206	3.30	5.800	6.700
552 WHITE CELL COUNT(WBC)		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %					
2024-911	1	Sysmex CBC	132 SYSMEX XS	126	21.440	0.720	3.36	18.200	24.700
2024-911	1	Sysmex CBC	141 SYSMEX XN	55	22.000	0.349	1.58	18.700	25.300
2024-911	1	Sysmex CBC	152 Sysmex XN	222	21.647	0.521	2.41	18.400	24.900

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2024-911	2	ACT PENTRA	125 COULTER AC	63	8.882	0.339	3.81	7.500	10.200	
2024-911	2	ACT PENTRA	130 ABX PENTRA	32	8.831	0.193	2.18	7.500	10.200	
2024-911	3	ADVIA 120	116 ADVIA-120	17	16.829	0.480	2.85	14.300	19.400	
2024-911	4	AB-CELL DY	115 ABBOTT C-D	37	20.189	0.511	2.53	17.200	23.200	
2024-911	4	AB-CELL DY	142 EMERALD	11	17.836	1.279	7.17	15.200	20.500	
2024-911	4	AB-CELL DY	153 AB-CEL-RUB	50	19.888	0.471	2.37	16.900	22.900	
2024-911	5	BECTON D	137 B.C DXH	58	21.709	0.464	2.14	18.500	25.000	
2024-911	6	MR Mindray	145 BC-5390	64	18.367	0.493	2.68	15.600	21.100	
2024-911	7	DXH500 Ser	149 BC DXH 500	99	2.343	0.101	4.33	2.000	2.700	
2024-911	8	3 ST	84 ABBOTT C-D	20	8.130	0.315	3.87	6.900	9.300	
2024-911	8	3 ST	138 SY KX-21N	13	8.400	0.124	1.48	7.100	9.700	
552 WHITE CELL COUNT(WBC		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %						
2024-912	1	Sysmex CBC	132 SYSMEX XS	126	8.074	0.393	4.87	6.900	9.300	
2024-912	1	Sysmex CBC	141 SYSMEX XN	55	8.138	0.228	2.81	6.900	9.400	
2024-912	1	Sysmex CBC	152 Sysmex XN	222	7.770	0.170	2.19	6.600	8.900	
2024-912	2	ACT PENTRA	125 COULTER AC	63	19.823	0.702	3.54	16.800	22.800	
2024-912	2	ACT PENTRA	130 ABX PENTRA	32	19.642	0.425	2.16	16.700	22.600	
2024-912	3	ADVIA 120	116 ADVIA-120	17	3.375	0.135	3.99	2.900	3.900	
2024-912	4	AB-CELL DY	115 ABBOTT C-D	37	7.403	0.216	2.92	6.300	8.500	
2024-912	4	AB-CELL DY	142 EMERALD	11	6.445	0.421	6.53	5.500	7.400	
2024-912	4	AB-CELL DY	153 AB-CEL-RUB	50	7.520	0.320	4.26	6.400	8.600	
2024-912	5	BECTON D	137 B.C DXH	58	9.718	0.190	1.96	8.300	11.200	
2024-912	6	MR Mindray	145 BC-5390	64	7.990	0.167	2.09	6.800	9.200	
2024-912	7	DXH500 Ser	149 BC DXH 500	99	7.178	0.233	3.25	6.100	8.300	
2024-912	8	3 ST	84 ABBOTT C-D	20	8.321	0.269	3.23	7.100	9.600	
2024-912	8	3 ST	138 SY KX-21N	13	8.377	0.204	2.44	7.100	9.600	
2024-913	1	Sysmex CBC	132 SYSMEX XS	126	4.317	0.152	3.51	3.700	5.000	
2024-913	1	Sysmex CBC	141 SYSMEX XN	55	4.236	0.096	2.27	3.600	4.900	
2024-913	1	Sysmex CBC	152 Sysmex XN	222	4.076	0.108	2.66	3.500	4.700	

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY      SUB: 70 GENERAL      GRP: 132 COMP BLOOD COUNT NEW

ANALYTE				NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE	
SAMPLE	METHOD	SYST/INST	REAGENTS	LABS	VALUE	DEVIATION	CV	LOW	HIGH	
2024-913	2	ACT PENTRA	125 COULTER AC	63	3.123	0.114	3.66	2.700	3.600	
2024-913	2	ACT PENTRA	130 ABX PENTRA	32	3.119	0.107	3.44	2.700	3.600	
2024-913	3	ADVIA 120	116 ADVIA-120	17	7.576	0.466	6.15	6.400	8.700	
2024-913	4	AB-CELL DY	115 ABBOTT C-D	37	2.851	0.108	3.79	2.400	3.300	
2024-913	4	AB-CELL DY	142 EMERALD	11	2.218	0.262	11.82	1.900	2.600	
2024-913	4	AB-CELL DY	153 AB-CEL-RUB	50	2.934	0.154	5.26	2.500	3.400	
2024-913	5	BECTON D	137 B.C DXH	58	4.390	0.106	2.42	3.700	5.000	
2024-913	6	MR Mindray	145 BC-5390	64	3.316	0.087	2.62	2.800	3.800	
2024-913	7	DXH500 Ser	149 BC DXH 500	99	2.322	0.114	4.92	2.000	2.700	
2024-913	8	3 ST	84 ABBOTT C-D	20	20.410	0.939	4.60	17.300	23.500	
2024-913	8	3 ST	138 SY KX-21N	13	20.915	0.530	2.54	17.800	24.100	
552 WHITE CELL COUNT(WBC		CLIA TEST ID #: 805		EVALUATION CRITERIA: TARGET VALUE 15 %						
2024-914	1	Sysmex CBC	132 SYSMEX XS	126	21.475	0.691	3.22	18.300	24.700	
2024-914	1	Sysmex CBC	141 SYSMEX XN	55	22.052	0.309	1.40	18.700	25.400	
2024-914	1	Sysmex CBC	152 Sysmex XN	222	21.626	0.516	2.39	18.400	24.900	
2024-914	2	ACT PENTRA	125 COULTER AC	63	19.827	0.696	3.51	16.900	22.800	
2024-914	2	ACT PENTRA	130 ABX PENTRA	32	19.725	0.501	2.54	16.800	22.700	
2024-914	3	ADVIA 120	116 ADVIA-120	17	7.700	0.308	4.00	6.500	8.900	
2024-914	4	AB-CELL DY	115 ABBOTT C-D	37	20.050	0.417	2.08	17.000	23.100	
2024-914	4	AB-CELL DY	142 EMERALD	11	17.718	1.233	6.96	15.100	20.400	
2024-914	4	AB-CELL DY	153 AB-CEL-RUB	50	19.841	0.450	2.27	16.900	22.800	
2024-914	5	BECTON D	137 B.C DXH	58	21.653	0.448	2.07	18.400	24.900	
2024-914	6	MR Mindray	145 BC-5390	64	18.380	0.436	2.37	15.600	21.100	
2024-914	7	DXH500 Ser	149 BC DXH 500	99	17.511	0.501	2.86	14.900	20.100	
2024-914	8	3 ST	84 ABBOTT C-D	20	8.315	0.413	4.96	7.100	9.600	
2024-914	8	3 ST	138 SY KX-21N	13	8.462	0.178	2.10	7.200	9.700	
2024-915	1	Sysmex CBC	132 SYSMEX XS	126	8.058	0.424	5.26	6.800	9.300	
2024-915	1	Sysmex CBC	141 SYSMEX XN	55	8.189	0.248	3.02	7.000	9.400	
2024-915	1	Sysmex CBC	152 Sysmex XN	222	7.781	0.164	2.11	6.600	8.900	

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 132 COMP BLOOD COUNT NEW						
ANALYTE											
SAMPLE	METHOD	SYST/INST		REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
2024-915	2	ACT	PENTRA	125	COULTER AC	63	3.129	0.122	3.91	2.700	3.600
2024-915	2	ACT	PENTRA	130	ABX PENTRA	32	3.138	0.093	2.95	2.700	3.600
2024-915	3	ADVIA	120	116	ADVIA-120	17	16.753	0.582	3.48	14.200	19.300
2024-915	4	AB-CELL	DY	115	ABBOTT C-D	37	7.369	0.173	2.35	6.300	8.500
2024-915	4	AB-CELL	DY	142	EMERALD	11	6.310	0.164	2.60	5.400	7.300
2024-915	4	AB-CELL	DY	153	AB-CEL-RUB	50	7.560	0.301	3.99	6.400	8.700
2024-915	5	BECTON	D	137	B.C DXH	58	9.705	0.225	2.32	8.200	11.200
2024-915	6	MR Mindray		145	BC-5390	64	7.975	0.175	2.19	6.800	9.200
2024-915	7	DXH500	Ser	149	BC DXH 500	99	17.535	0.533	3.04	14.900	20.200
2024-915	8	3	ST	84	ABBOTT C-D	20	2.635	0.177	6.71	2.200	3.000
2024-915	8	3	ST	138	SY KX-21N	13	2.554	0.101	3.95	2.200	2.900

SPEC: 20 HEMATOLOGY		SUB: 70 GENERAL			GRP: 134 PLATELET COUNT NEW						
ANALYTE											
SAMPLE	METHOD	SYST/INST		REAGENTS	NO. PART LABS	TARGET VALUE	STANDARD DEVIATION	% CV	ACCEPTABLE LOW	RANGE HIGH	
553	PLATELET COUNT	CLIA TEST ID #: 815		EVALUATION CRITERIA: TARGET VALUE 25 %							
2024-911	1	SYSMEX	CBC	79	SYSMEX XS	126	394.808	16.418	4.16	296.000	494.000
2024-911	1	SYSMEX	CBC	87	SYSMEX XN	57	416.175	16.053	3.86	312.000	520.000
2024-911	1	SYSMEX	CBC	98	SYSMEX 350	221	412.756	14.864	3.60	310.000	516.000
2024-911	2	ACT5	PENTR	73	COULTER AC	61	239.705	10.980	4.58	180.000	300.000
2024-911	2	ACT5	PENTR	78	ABX PENTRA	32	235.469	8.404	3.57	177.000	294.000
2024-911	3	ADVIA	120	64	ADVIA-120	17	449.471	31.862	7.09	337.000	562.000
2024-911	4	AB- CELL-D		63	ABBOTT C-D	35	501.257	23.627	4.71	376.000	627.000
2024-911	4	AB- CELL-D		88	CELL-DYN E	11	495.182	25.012	5.05	371.000	619.000
2024-911	4	AB- CELL-D		100	ABB.RUBY	51	523.333	28.042	5.36	393.000	654.000
2024-911	5	BECTON	D	83	B.C600,800	58	489.052	14.472	2.96	367.000	611.000
2024-911	6	MR Mindray		91	BC-5390	64	525.730	15.296	2.91	394.000	657.000
2024-911	8	3	ST	31	ABBOTT C-D	20	194.350	24.757	12.74	146.000	243.000
2024-911	8	3	ST	84	KX21,300	13	211.846	13.132	6.20	159.000	265.000
2024-911	10	DxH		94	BC DXH 500	99	78.582	4.971	6.33	59.000	98.000

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
553	PLATELET COUNT	CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %								
2024-912	1	SYSMEX CBC	79	SYSMEX XS	126	210.576	8.285	3.93	158.000	263.000	
2024-912	1	SYSMEX CBC	87	SYSMEX XN	57	206.702	8.392	4.06	155.000	258.000	
2024-912	1	SYSMEX CBC	98	SYSMEX 350	221	205.269	8.391	4.09	154.000	257.000	
2024-912	2	ACT5 PENTR	73	COULTER AC	61	467.377	16.145	3.45	351.000	584.000	
2024-912	2	ACT5 PENTR	78	ABX PENTRA	32	468.000	17.425	3.72	351.000	585.000	
2024-912	3	ADVIA 120	64	ADVIA-120	17	53.353	5.235	9.81	40.000	67.000	
2024-912	4	AB- CELL-D	63	ABBOTT C-D	35	289.400	15.045	5.20	217.000	362.000	
2024-912	4	AB- CELL-D	88	CELL-DYN E	11	287.364	19.773	6.88	216.000	359.000	
2024-912	4	AB- CELL-D	100	ABB.RUBY	51	289.608	18.916	6.53	217.000	362.000	
2024-912	5	BECTON D	83	B.C600,800	58	262.069	7.019	2.68	197.000	328.000	
2024-912	6	MR Mindray	91	BC-5390	64	260.609	8.672	3.33	195.000	326.000	
2024-912	8	3 ST	31	ABBOTT C-D	20	197.100	24.249	12.30	148.000	246.000	
2024-912	8	3 ST	84	KX21,300	13	211.154	13.705	6.49	158.000	264.000	
2024-912	10	DxH	94	BC DXH 500	99	250.928	11.393	4.54	188.000	314.000	
2024-913	1	SYSMEX CBC	79	SYSMEX XS	126	55.897	3.477	6.22	42.000	70.000	
2024-913	1	SYSMEX CBC	87	SYSMEX XN	57	48.561	3.140	6.47	36.000	61.000	
2024-913	1	SYSMEX CBC	98	SYSMEX 350	221	49.529	3.061	6.18	37.000	62.000	
2024-913	2	ACT5 PENTR	73	COULTER AC	61	70.689	4.964	7.02	53.000	88.000	
2024-913	2	ACT5 PENTR	78	ABX PENTRA	32	70.313	4.369	6.21	53.000	88.000	
2024-913	3	ADVIA 120	64	ADVIA-120	17	218.353	13.651	6.25	164.000	273.000	
2024-913	4	AB- CELL-D	63	ABBOTT C-D	35	82.143	5.026	6.12	62.000	103.000	
2024-913	4	AB- CELL-D	88	CELL-DYN E	11	95.091	10.808	11.37	71.000	119.000	
2024-913	4	AB- CELL-D	100	ABB.RUBY	51	82.765	6.370	7.70	62.000	103.000	
2024-913	5	BECTON D	83	B.C600,800	58	82.776	2.305	2.78	62.000	103.000	
2024-913	6	MR Mindray	91	BC-5390	64	52.206	2.918	5.59	39.000	65.000	
2024-913	8	3 ST	31	ABBOTT C-D	20	376.500	50.570	13.43	282.000	471.000	
2024-913	8	3 ST	84	KX21,300	13	416.077	23.421	5.63	312.000	520.000	
2024-913	10	DxH	94	BC DXH 500	99	79.412	5.003	6.30	60.000	99.000	

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY

SUB: 70 GENERAL

GRP: 134 PLATELET COUNT NEW

ANALYTE						NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH	
553	PLATELET COUNT	CLIA TEST ID #: 815	EVALUATION CRITERIA: TARGET VALUE 25 %								
2024-914	1	SYSMEX CBC	79	SYSMEX XS	126	395.584	18.198	4.60	297.000	494.000	
2024-914	1	SYSMEX CBC	87	SYSMEX XN	57	414.632	16.877	4.07	311.000	518.000	
2024-914	1	SYSMEX CBC	98	SYSMEX 350	221	412.101	17.419	4.23	309.000	515.000	
2024-914	2	ACT5 PENTR	73	COULTER AC	61	465.750	13.005	2.79	349.000	582.000	
2024-914	2	ACT5 PENTR	78	ABX PENTRA	32	464.156	14.965	3.22	348.000	580.000	
2024-914	3	ADVIA 120	64	ADVIA-120	17	219.941	14.152	6.43	165.000	275.000	
2024-914	4	AB- CELL-D	63	ABBOTT C-D	35	500.171	22.215	4.44	375.000	625.000	
2024-914	4	AB- CELL-D	88	CELL-DYN E	11	495.818	31.397	6.33	372.000	620.000	
2024-914	4	AB- CELL-D	100	ABB.RUBY	51	524.706	26.061	4.97	394.000	656.000	
2024-914	5	BECTON D	83	B.C600,800	58	496.466	18.292	3.68	372.000	621.000	
2024-914	6	MR Mindray	91	BC-5390	64	528.516	17.032	3.22	396.000	661.000	
2024-914	8	3 ST	31	ABBOTT C-D	20	196.550	27.032	13.75	147.000	246.000	
2024-914	8	3 ST	84	KX21,300	13	212.231	14.651	6.90	159.000	265.000	
2024-914	10	DxH	94	BC DXH 500	99	495.443	20.877	4.21	372.000	619.000	
2024-915	1	SYSMEX CBC	79	SYSMEX XS	126	209.913	9.089	4.33	157.000	262.000	
2024-915	1	SYSMEX CBC	87	SYSMEX XN	57	205.930	8.837	4.29	154.000	257.000	
2024-915	1	SYSMEX CBC	98	SYSMEX 350	221	205.823	8.282	4.02	154.000	257.000	
2024-915	2	ACT5 PENTR	73	COULTER AC	61	69.475	5.431	7.82	52.000	87.000	
2024-915	2	ACT5 PENTR	78	ABX PENTRA	32	69.500	4.880	7.02	52.000	87.000	
2024-915	3	ADVIA 120	64	ADVIA-120	17	445.882	30.146	6.76	334.000	557.000	
2024-915	4	AB- CELL-D	63	ABBOTT C-D	35	291.657	15.943	5.47	219.000	365.000	
2024-915	4	AB- CELL-D	88	CELL-DYN E	11	291.182	17.177	5.90	218.000	364.000	
2024-915	4	AB- CELL-D	100	ABB.RUBY	51	295.647	20.609	6.97	222.000	370.000	
2024-915	5	BECTON D	83	B.C600,800	58	261.298	7.878	3.02	196.000	327.000	
2024-915	6	MR Mindray	91	BC-5390	64	259.937	8.621	3.32	195.000	325.000	
2024-915	8	3 ST	31	ABBOTT C-D	20	64.000	6.050	9.45	48.000	80.000	
2024-915	8	3 ST	84	KX21,300	13	64.846	5.021	7.74	49.000	81.000	
2024-915	10	DxH	94	BC DXH 500	99	493.794	22.293	4.51	370.000	617.000	

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPEC: 20 HEMATOLOGY                      SUB: 75 BODY FLUID                      GRP: 145 BODY FLUID CELL COU.

ANALYTE					NO. PART	TARGET	STANDARD	%	ACCEPTABLE	RANGE
SAMPLE	METHOD	SYST/INST	REAGENTS		LABS	VALUE	DEVIATION	CV	LOW	HIGH
159	BF RED CELL COUNT	CLIA TEST ID #: 0	EVALUATION CRITERIA: TARGET VALUE 3	SD OR 25 %	{GREATER}					
	2024-951				39	1030.711	111.592	10.83	696.000	1365.000
	2024-952				39	1064.263	99.301	9.33	766.000	1362.000
	2024-953				39	0.421	0.544	129.30	0	2.000
	2024-954				39	1051.684	97.204	9.24	760.000	1343.000
	2024-955				39	0.421	0.544	129.30	0	2.000
160	BF WHITE CELL COUNT									
	2024-951				38	258.222	31.652	12.26	163.000	353.000
	2024-952				38	262.306	28.789	10.98	176.000	349.000
	2024-953				38	4.108	1.857	45.20	0	10.000
	2024-954				38	273.806	30.744	11.23	182.000	366.000
	2024-955				38	4.676	2.157	46.13	0	11.000



**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPE: 30 IMMUNOHEMATOLOGY												SUB: 80 ABO-RH				GRP: 150 ABO-RH			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score								
083 ABO GROUP																			
2024-961				005 007	005 007		005 - AB	76	76	76	100								
2024-962				001 003	001 003		001 - A	76	75	76	100								
2024-962				001 003	001 003		002 - A1	76	1	76	100								
2024-963				001 003	001 003		001 - A	76	75	76	100								
2024-963				001 003	001 003		002 - A1	76	1	76	100								
2024-964				004 004	004 004		004 - B	76	76	76	100								
2024-965				008 008	008 008		008 - O	76	76	76	100								
126 RH TYPE																			
2024-961				001 003	001 003		001 - Rh+	76	76	76	100								
2024-962				001 003	001 003		001 - Rh+	76	76	76	100								
2024-963				003 004	003 004		003 - Rh-Du np	76	8	76	100								
2024-963				003 004	003 004		004 - Rh-	76	68	76	100								
2024-964				001 003	001 003		001 - Rh+	76	76	76	100								
2024-965				003 004	003 004		003 - Rh-Du np	76	8	76	100								
2024-965				003 004	003 004		004 - Rh-	76	68	76	100								

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 160 Indirect COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
084 INDIRECT COOMBS											
2024-971				001 001	001 001		001 - NEGATIVE	61	61	61	100
2024-972				001 001	001 001		001 - NEGATIVE	61	61	61	100
2024-973				001 001	001 001		001 - NEGATIVE	61	61	61	100
2024-974				002 002	002 002		001 - NEGATIVE	61	1	60	0
2024-974				002 002	002 002		002 - POSITIVE	61	60	60	100
2024-975				002 002	002 002		002 - POSITIVE	61	61	61	100
SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS			
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score
179 DIRECT ANTIG(COOMBS)											
2024-651	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	37	37	37	100
2024-651	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100
2024-651	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	17	17	17	100
2024-651	003 Anti-C3b/d	002 Tubo			001 001	001 001	001 - Negativo	10	10	10	100
2024-652	001 Poly AHG	002 Tubo		002 004	002 004		002 - 2+	37	32	37	100
2024-652	001 Poly AHG	002 Tubo		002 004	002 004		003 - 3+	37	5	37	100
2024-652	002 MnoAnt-IgG	001 Gel			002 004		002 - 2+	11	1	11	100
2024-652	002 MnoAnt-IgG	001 Gel			002 004		003 - 3+	11	5	11	100
2024-652	002 MnoAnt-IgG	001 Gel			002 004		004 - 4+	11	5	11	100

**Participant Summary**  
**MONTH 11 OF YEAR 2024**

SPE: 30 IMMUNOHEMATOLOGY				SUB: 90 COOMBS				GRP: 165 Direct COOMBS					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART LABS	# LABS ANSW	# "SF" LABS	Score		
2024-652	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	17	14	17	100		
2024-652	002 MnoAnt-IgG	002 Tubo		002 004	002 004		003 - 3+	17	3	17	100		
179 DIRECT ANTIG(COOMBS)													
2024-653	001 Poly AHG	002 Tubo			002 004		002 - 2+	37	28	37	100		
2024-653	001 Poly AHG	002 Tubo			002 004		003 - 3+	37	8	37	100		
2024-653	001 Poly AHG	002 Tubo			002 004		004 - 4+	37	1	37	100		
2024-653	002 MnoAnt-IgG	001 Gel			002 004		002 - 2+	11	1	11	100		
2024-653	002 MnoAnt-IgG	001 Gel			002 004		003 - 3+	11	5	11	100		
2024-653	002 MnoAnt-IgG	001 Gel			002 004		004 - 4+	11	5	11	100		
2024-653	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	17	15	17	100		
2024-653	002 MnoAnt-IgG	002 Tubo		002 004	002 004		003 - 3+	17	2	17	100		
2024-654	001 Poly AHG	002 Tubo		002 004	002 004		002 - 2+	37	30	37	100		
2024-654	001 Poly AHG	002 Tubo		002 004	002 004		003 - 3+	37	5	37	100		
2024-654	001 Poly AHG	002 Tubo		002 004	002 004		004 - 4+	37	2	37	100		
2024-654	002 MnoAnt-IgG	001 Gel			002 004		002 - 2+	11	1	11	100		
2024-654	002 MnoAnt-IgG	001 Gel			002 004		003 - 3+	11	5	11	100		
2024-654	002 MnoAnt-IgG	001 Gel			002 004		004 - 4+	11	5	11	100		
2024-654	002 MnoAnt-IgG	002 Tubo		002 004	002 004		002 - 2+	17	14	17	100		
2024-654	002 MnoAnt-IgG	002 Tubo		002 004	002 004		003 - 3+	17	2	17	100		
2024-654	002 MnoAnt-IgG	002 Tubo		002 004	002 004		004 - 4+	17	1	17	100		
2024-655	001 Poly AHG	002 Tubo		001 001	001 001		001 - Negativo	37	37	37	100		
2024-655	002 MnoAnt-IgG	001 Gel		001 001	001 001		001 - Negativo	11	11	11	100		
2024-655	002 MnoAnt-IgG	002 Tubo		001 001	001 001		001 - Negativo	17	17	17	100		
2024-655	003 Anti-C3b/d	002 Tubo			001 001	001 001	001 - Negativo	10	10	10	100		

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPE: 30 IMMUNOHEMATOLOGY												SUB: 100 COMPATIBILITY TESTING												GRP: 170 COMPATIBILITY TEST.											
TEST	LEVEL 1		LEVEL 2		LEVEL 3		PART		REFEREE		COMMER		REPORTED		PART		# LABS		# "SF"		Score														
SAMPLE	RESULT CODE		RESULT CODE		RESULT CODE		LOW/HIGH		LOW/HIGH		LOW/HIGH		VALUE		LABS		ANSW		LABS																
085 COMPATIBIL. TESTING																																			
2024-981							001	001	001	001			001 - COMPATIBLE		39	39	39	39	39	39	100														
2024-982							002	002	002	002			002 - INCOMPATIB		39	39	39	39	39	39	100														
2024-983							001	001	001	001			001 - COMPATIBLE		39	39	39	39	39	39	100														
2024-984							002	002	002	002			002 - INCOMPATIB		39	39	39	39	39	39	100														
2024-985							002	002	002	002			002 - INCOMPATIB		39	39	39	39	39	39	100														

**Participant Summary**

**MONTH 11 OF YEAR 2024**

SPE: 30 IMMUNOHEMATOLOGY		SUB: 110 ANTIBODY IDENTIFICATION				GRP: 180 ANTIBODY IDENT.					
TEST SAMPLE	LEVEL 1 RESULT CODE	LEVEL 2 RESULT CODE	LEVEL 3 RESULT CODE	PART LOW/HIGH	REFEREE LOW/HIGH	COMMER LOW/HIGH	REPORTED VALUE	PART # LABS	# LABS ANSW	# "SF" LABS	Score
086 ANTIBODY IDENT.TEST.											
2024-981				001 001	001 001		001 - UN. ANT.ND	10	10	10	100
2024-982				001 001	001 001		001 - UN. ANT.ND	10	10	10	100
2024-983				001 001	001 001		001 - UN. ANT.ND	10	10	10	100
2024-984				029 029	029 029		029 - Anti s	10	10	10	100
2024-985				003 003	003 003		003 - Anti D	10	10	10	100